

Alcohol Testing Program - Instrument Processing Sheet

Agency: PORT ST LUCIE PD

Instrument Serial Number: 80-001323

Date In: 4/30/2026

DI Completion Date: 5/7/26

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>WKP</u> Date: <u>4/30/2026</u>		Quality Checks By: <u>IAS</u> Date: <u>5/1/26</u>		Flow Adjustment By: _____																
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Instrument arrived with an additional static bag and an agency note in the box. Left side keyboard hook arrived broken. WKP 4/30/2026		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>213</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP102</u> 32 mm <u>.144</u> (.139-.169) 36 mm <u>.164</u> (.156-.190) 53 mm <u>.234</u> (.228-.278) 103 mm <u>.496</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>34419</u> Gauge: <u>1010</u> Instrument: <u>1009</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 3/3/2028	0.080	MP5089	202603B 3/3/2028	0.200	MP5090	202603D 3/4/2028	0.080 DGS	N/A	23825080A1 10/5/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____	
Simulator	Serial #	Lot#/Exp																		
0.050	MP5088	202603A 3/3/2028																		
0.080	MP5089	202603B 3/3/2028																		
0.200	MP5090	202603D 3/4/2028																		
0.080 DGS	N/A	23825080A1 10/5/2027																		
Optical Bench Adjustment By: <u>KTS</u>		Department Inspection By: <u>KTS</u>																		
Barometric Pressure Gauge: <u>1008</u> ID#: <u>28427</u>		Barometric Pressure ID#: <u>34419</u>																		
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1009</u> Instrument: <u>1010</u>																
0.000	MP6289	N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/27</u>																
0.040	MP6290	25090	3/11/27	Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/27</u>																
0.100	MP6291	25282	8/12/27	Simulator	Serial Number															
0.200	MP6292	25020	1/14/27	0.000	MP5086															
0.300	MP6293	25150	5/6/27	Interferent	MP5087															
0.080 DGS	N/A	AG510701	4/17/27	0.050	MP5088															
☒ Post Optical Bench Adjustment Stability Checks				0.080	MP5089															
Simulator	Serial #	Lot #	Expiration	0.200	MP5090															
0.050	MP5088	202603A	3/3/28	Attachments																
0.080	MP5089	202603B	3/3/28	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment ☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☒ Other: <u>AI Note</u>																
0.200	MP5090	202603D	3/4/28																	
0.080 DGS	N/A	23825080A1	10/5/27																	
Gauge ID #: <u>34419</u>																				
Gauge: <u>1011</u> Instrument: <u>1010</u>																				
Notes/Suggested Service: Instrument's keyboard is broken. Emailed AI to order a new one. KTS 5/8/26				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																
				Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.05.09 18:15:15 -04'00' Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.05.15 08:39:29 -04'00'																
				Tech Review Admin Review																

PORT ST LUCIE POLICE DEPARTMENT
AGENCY INSPECTORS
OFFICER K. APPELBAUM & B. THOMPSON
121 SW PORT ST LUCIE BLVD
PORT ST LUCIE, FLORIDA, 34984
772-871-5001

* TOLD TO SEND BACK FOR
INSPECTION

KTS 80-001323

4/30/26

Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L ≤0.003 g/210L of Wet
PORT ST LUCIE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001323 05/01/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 16:16 Control Test 0.048 16:16 Air Blank 0.000 16:17 Control Test 0.048 16:18 Air Blank 0.000 16:18 Control Test 0.048 16:19 Air Blank 0.000 16:20 Control Test Stats Average 0.0480 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>Isabella Smyth on site</i> Operator's Signature	PORT ST LUCIE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001323 05/01/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:45 Control Test 0.077 12:46 Air Blank 0.000 12:47 Control Test 0.076 12:47 Air Blank 0.000 12:48 Control Test 0.076 12:49 Air Blank 0.000 12:49 Control Test Stats Average 0.0763 Std Dev 0.0006 Rel Std Dev(%) 0.7564 <i>Isabella Smyth on site</i> Operator's Signature	PORT ST LUCIE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001323 05/01/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:38 Control Test 0.197 12:38 Air Blank 0.000 12:39 Control Test 0.195 12:40 Air Blank 0.000 12:40 Control Test 0.195 12:41 Air Blank 0.000 12:41 Control Test Stats Average 0.1957 Std Dev 0.0012 Rel Std Dev(%) 0.5911 <i>Isabella Smyth on site</i> Operator's Signature	PORT ST LUCIE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001323 05/01/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 16:11 Control Test 0.078 16:11 Air Blank 0.000 16:12 Control Test 0.079 16:12 Air Blank 0.000 16:12 Control Test 0.079 16:13 Air Blank 0.000 16:13 Control Test Stats Average 0.0787 Std Dev 0.0005 Rel Std Dev(%) 0.7339 <i>Isabella Smyth on site</i> Operator's Signature

Optical Adjustment 80-001323
Aborted due to inputting wrong
ARS control value. *KATZ*

PORT ST LUCIE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001323
05/07/2026 06:52:54

Auto Calibration
Max Power Res Value = 44
Auto Range Res Value = 26

**** AUTO CAL FAIL

Optical Adjustment 80-001323

Katz

PORT ST LUCIE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001323
05/07/2026 06:56:40

Auto Calibration
Max Power Res Value = 44
Auto Range Res Value = 26

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12802, 9um Io = 13650
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1120 (-0.0190)
Sample #2 = 0.1200 (0.0010)
Sample #3 = 0.1100 (0.0220)
Sample #4 = 0.1200 (0.0440)
Avg % Abs = 0.1167 (0.0223)
STD DEV = 0.0058 (0.0215)
REL STD DEV = 4.949 (96.277)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1170 (0.0000)
Sample #2 = 0.1220 (0.0020)
Sample #3 = 0.1100 (0.0070)
Sample #4 = 0.1020 (0.0190)
Avg % Abs = 0.1113 (0.0093)
STD DEV = 0.0101 (0.0087)
REL STD DEV = 9.042 (93.610)

Sol Value = 0.041 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12795, 9um Io = 13657
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8660 (-0.0220)
Sample #2 = 0.8140 (0.0370)
Sample #3 = 0.8300 (0.0370)
Sample #4 = 0.8340 (0.0330)
Avg % Abs = 0.8260 (0.0357)
STD DEV = 0.0106 (0.0023)
REL STD DEV = 1.281 (6.475)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5390 (-0.0160)
Sample #2 = 1.5130 (0.0210)
Sample #3 = 1.5030 (0.0360)
Sample #4 = 1.5150 (0.0240)
Avg % Abs = 1.5103 (0.0270)
STD DEV = 0.0064 (0.0079)
REL STD DEV = 0.426 (29.397)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12788, 9um Io = 13654
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9050 (-0.0190)
Sample #2 = 1.8750 (0.0170)
Sample #3 = 1.8870 (0.0050)
Sample #4 = 1.8730 (0.0190)
Avg % Abs = 1.8783 (0.0137)
STD DEV = 0.0076 (0.0076)
REL STD DEV = 0.403 (55.404)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6080 (0.0020)
Sample #2 = 3.5530 (0.0570)
Sample #3 = 3.5540 (0.0490)
Sample #4 = 3.5480 (0.0640)
Avg % Abs = 3.5517 (0.0567)
STD DEV = 0.0032 (0.0075)
REL STD DEV = 0.091 (13.245)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12790, 9um Io = 13651
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6770 (-0.0140)
Sample #2 = 3.6260 (0.0450)
Sample #3 = 3.5890 (0.0770)
Sample #4 = 3.6070 (0.0670)
Avg % Abs = 3.6073 (0.0630)
STD DEV = 0.0185 (0.0164)
REL STD DEV = 0.513 (25.985)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.9170 (-0.0080)
Sample #2 = 6.8470 (0.0970)
Sample #3 = 6.8000 (0.1240)
Sample #4 = 6.8130 (0.1180)
Avg % Abs = 6.8200 (0.1130)
STD DEV = 0.0243 (0.0142)
REL STD DEV = 0.356 (12.546)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12788, 9um Io = 13647
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.3480 (-0.0030)
Sample #2 = 5.2840 (0.0660)
Sample #3 = 5.2650 (0.0700)
Sample #4 = 5.2510 (0.0840)
Avg % Abs = 5.2667 (0.0733)
STD DEV = 0.0166 (0.0095)
REL STD DEV = 0.314 (12.889)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 10.0080 (-0.0070)
Sample #2 = 9.8980 (0.1260)
Sample #3 = 9.8760 (0.1480)
Sample #4 = 9.8680 (0.1560)
Avg % Abs = 9.8807 (0.1433)
STD DEV = 0.0155 (0.0155)
REL STD DEV = 0.157 (10.838)

***** AUTO CAL DATA *****
<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.117
Std Dev = 0.01 Rel Std Dev = 4.95
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.826
Std Dev = 0.01 Rel Std Dev = 1.28
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.878
Std Dev = 0.01 Rel Std Dev = 0.40
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.607
Std Dev = 0.02 Rel Std Dev = 0.51
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.267
Std Dev = 0.02 Rel Std Dev = 0.31
Zero Order Coef = -299.63
First Order Coef = 2641.77
Second Order Coef = 23.98
Standard Deviation = 13.394496

<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.111
Std Dev = 0.01 Rel Std Dev = 9.04
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.510
Std Dev = 0.01 Rel Std Dev = 0.43
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.552
Std Dev = 0.00 Rel Std Dev = 0.09
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.820
Std Dev = 0.02 Rel Std Dev = 0.36
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.881
Std Dev = 0.02 Rel Std Dev = 0.16
Zero Order Coef = -140.34
First Order Coef = 1330.41
Second Order Coef = 13.05
Standard Deviation = 11.848620

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0002
0.040 0.040 0.0001
0.100 0.100 0.0003
0.200 0.200 -0.0004
0.300 0.300 0.0001

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0002
0.040 0.040 0.0001
0.100 0.100 0.0003
0.200 0.200 -0.0003
0.300 0.300 0.0001


Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3142.00
Sample #2 = 3185.00
Sample #3 = 3167.00
Sample #4 = 3173.00
Average Result = 3175.0000
STD DEV = 9.1652
REL STD DEV = 0.289

***** CHANNEL 2
Sample #1 = 3441.00
Sample #2 = 3501.00
Sample #3 = 3510.00
Sample #4 = 3510.00
Average Result = 3507.0000
STD DEV = 5.1962
REL STD DEV = 0.148

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1008
3 um H2O Adjust (mg/l*10,000) = 634
9 um H2O Adjust (mg/l*10,000) = 302
***** AUTO CAL PASS

Post-Optical

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
PORT ST LUCIE PD Intoxilizer - Alcohol Analyzer Model 8000 05/07/2026 Software: 8100.27 SN 80-001323 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:58</td></tr> <tr><td>Control Test</td><td>0.050</td><td>07:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:00</td></tr> <tr><td>Control Test</td><td>0.050</td><td>08:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:01</td></tr> <tr><td>Control Test</td><td>0.051</td><td>08:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0503</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1471</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	07:58	Control Test	0.050	07:59	Air Blank	0.000	08:00	Control Test	0.050	08:00	Air Blank	0.000	08:01	Control Test	0.051	08:02	Air Blank	0.000	08:02	Control Test Stats			Average	0.0503		Std Dev	0.0006		Rel Std Dev(%)	1.1471		PORT ST LUCIE PD Intoxilizer - Alcohol Analyzer Model 8000 05/07/2026 Software: 8100.27 SN 80-001323 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:50</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:52</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:53</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:54</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	07:50	Control Test	0.080	07:51	Air Blank	0.000	07:52	Control Test	0.080	07:52	Air Blank	0.000	07:53	Control Test	0.079	07:53	Air Blank	0.000	07:54	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		PORT ST LUCIE PD Intoxilizer - Alcohol Analyzer Model 8000 05/07/2026 Software: 8100.27 SN 80-001323 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:05</td></tr> <tr><td>Control Test</td><td>0.202</td><td>08:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:06</td></tr> <tr><td>Control Test</td><td>0.200</td><td>08:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:08</td></tr> <tr><td>Control Test</td><td>0.200</td><td>08:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:09</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2007</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5754</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	08:05	Control Test	0.202	08:06	Air Blank	0.000	08:06	Control Test	0.200	08:07	Air Blank	0.000	08:08	Control Test	0.200	08:08	Air Blank	0.000	08:09	Control Test Stats			Average	0.2007		Std Dev	0.0012		Rel Std Dev(%)	0.5754		PORT ST LUCIE PD Intoxilizer - Alcohol Analyzer Model 8000 05/07/2026 Software: 8100.27 SN 80-001323 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:43</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:44</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:45</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:46</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	07:43	Control Test	0.080	07:44	Air Blank	0.000	07:44	Control Test	0.080	07:44	Air Blank	0.000	07:45	Control Test	0.080	07:45	Air Blank	0.000	07:46	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:58																																																																																																																																																	
Control Test	0.050	07:59																																																																																																																																																	
Air Blank	0.000	08:00																																																																																																																																																	
Control Test	0.050	08:00																																																																																																																																																	
Air Blank	0.000	08:01																																																																																																																																																	
Control Test	0.051	08:02																																																																																																																																																	
Air Blank	0.000	08:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0503																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1471																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:50																																																																																																																																																	
Control Test	0.080	07:51																																																																																																																																																	
Air Blank	0.000	07:52																																																																																																																																																	
Control Test	0.080	07:52																																																																																																																																																	
Air Blank	0.000	07:53																																																																																																																																																	
Control Test	0.079	07:53																																																																																																																																																	
Air Blank	0.000	07:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:05																																																																																																																																																	
Control Test	0.202	08:06																																																																																																																																																	
Air Blank	0.000	08:06																																																																																																																																																	
Control Test	0.200	08:07																																																																																																																																																	
Air Blank	0.000	08:08																																																																																																																																																	
Control Test	0.200	08:08																																																																																																																																																	
Air Blank	0.000	08:09																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2007																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5754																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:43																																																																																																																																																	
Control Test	0.080	07:44																																																																																																																																																	
Air Blank	0.000	07:44																																																																																																																																																	
Control Test	0.080	07:44																																																																																																																																																	
Air Blank	0.000	07:45																																																																																																																																																	
Control Test	0.080	07:45																																																																																																																																																	
Air Blank	0.000	07:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
 Operator's Signature	 Operator's Signature	 Operator's Signature	 Operator's Signature																																																																																																																																																

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST LUCIE PD
Time of Inspection: 11:09

Date of Inspection: 05/07/2026

Serial Number: 80-001323
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#: 202603B 202603B Exp: 03/04/2028 3/3/2028 KTS 5/7/26	0.20g/210L Test (g/210L) Lot#: 202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 23825080A1 Exp: 10/05/2027
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.079	0.200	0.078
0.000	0.050	0.079	0.200	0.078
0.000	0.051	0.080	0.200	0.078
0.000	0.050	0.079	0.200	0.078
0.000	0.050	0.079	0.200	0.079
0.000	0.051	0.079	0.200	0.079
0.000	0.051	0.080	0.200	0.079
0.000	0.051	0.080	0.200	0.079
0.000	0.051	0.079	0.200	0.078

Standard Deviations	0.0005	0.0004	0.0000	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN

Signature and Printed Name

05/07/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001323, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001323</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PORT ST LUCIE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/07/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:09</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/07/2026

Date

Katie T Spearin
KATIE T SPEARIN,

Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1