

Alcohol Testing Program - Instrument Processing Sheet

Agency: PORT RICHEY PD

Instrument Serial Number: 80-000967

Date In: 6/10/2026

DI Completion Date: 6/17/26

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>WKP</u> Date: <u>6/10/2026</u>		Quality Checks By: <u>KTS</u> Date: <u>6/10/26</u>		Flow Adjustment By: <u>KTS</u>																																	
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Agency note about battery change included in the box. WKP 6/10/2026		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>121</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm 0.121 (.139-.169) 36 mm 0.132 (.156-.190) 53 mm 0.210 (.228-.278) 103 mm 0.507 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1011</u> Instrument: <u>1012</u> ☒ Stability Checks		Flow Column #: <u>ATP102</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value: <u>121</u> ☒ Post Adjustment Verification (L/S) Flow Column #: <u>ATP103</u> 32 mm 0.144 (.139-.169) 36 mm 0.164 (.156-.190) 53 mm 0.238 (.228-.278) 103 mm 0.519 (.447-.547)																																	
						<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:15%;">Lot#/Exp</th> <th style="width:50%;">Maintenance By: <u>KTS</u> Date: <u>6/10/26</u></th> </tr> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A 3/3/28</td> <td rowspan="4"> ☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: </td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B 3/3/28</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D 3/4/28</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/27</td> </tr> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: <u>KTS</u> Date: <u>6/10/26</u>	0.050	MP5088	202603A 3/3/28	☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:	0.080	MP5089	202603B 3/3/28	0.200	MP5090	202603D 3/4/28	0.080 DGS	N/A	23825080A1 10/5/27													
								Simulator	Serial #	Lot#/Exp	Maintenance By: <u>KTS</u> Date: <u>6/10/26</u>																										
								0.050	MP5088	202603A 3/3/28	☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:																										
								0.080	MP5089	202603B 3/3/28																											
0.200	MP5090	202603D 3/4/28																																			
0.080 DGS	N/A	23825080A1 10/5/27																																			
Optical Bench Adjustment By: <u>KTS</u>		Department Inspection By: <u>KTS</u>																																			
Barometric Pressure Gauge: <u>1013</u> ID#: <u>34420</u>		Barometric Pressure ID#: <u>28427*</u>																																			
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td>MP6294</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>25090</td> <td>3/11/27</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>25282</td> <td>8/12/27</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>25020</td> <td>1/14/27</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>25150</td> <td>5/6/27</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701</td> <td>4/17/27</td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	3/11/27	0.100	MP6296	25282	8/12/27	0.200	MP6297	25020	1/14/27	0.300	MP6298	25150	5/6/27	0.080 DGS	N/A	AG510701	4/17/27	Gauge: <u>1011 / 1009</u> Instrument: <u>1012 / 1010</u> Mouth Alcohol Solution Lot #: <u>2025-D*</u> Exp: <u>9/25/27</u> Acetone Stock Solution Lot #: <u>2025-B*</u> Exp: <u>9/22/27</u>							
Simulator	Serial #	Lot #	Expiration																																		
0.000	MP6294	N/A	N/A																																		
0.040	MP6295	25090	3/11/27																																		
0.100	MP6296	25282	8/12/27																																		
0.200	MP6297	25020	1/14/27																																		
0.300	MP6298	25150	5/6/27																																		
0.080 DGS	N/A	AG510701	4/17/27																																		
☒ Post Optical Bench Adjustment Stability Checks		<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A</td> <td>3/3/28</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B</td> <td>3/3/28</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D</td> <td>3/4/28</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1</td> <td>10/5/27</td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	3/3/28	0.080	MP5089	202603B	3/3/28	0.200	MP5090	202603D	3/4/28	0.080 DGS	N/A	23825080A1	10/5/27	<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP5086 / MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP5087*</td> </tr> <tr> <td>0.050</td> <td>MP5088*</td> </tr> <tr> <td>0.080</td> <td>MP5089*</td> </tr> <tr> <td>0.200</td> <td>MP5090*</td> </tr> </table>		Simulator	Serial Number	0.000	MP5086 / MP6289	Interferent	MP5087*	0.050	MP5088*	0.080	MP5089*	0.200	MP5090*
Simulator	Serial #	Lot #	Expiration																																		
0.050	MP5088	202603A	3/3/28																																		
0.080	MP5089	202603B	3/3/28																																		
0.200	MP5090	202603D	3/4/28																																		
0.080 DGS	N/A	23825080A1	10/5/27																																		
Simulator	Serial Number																																				
0.000	MP5086 / MP6289																																				
Interferent	MP5087*																																				
0.050	MP5088*																																				
0.080	MP5089*																																				
0.200	MP5090*																																				
Gauge ID #: <u>28427</u> Gauge: <u>1011</u> Instrument: <u>1012</u>		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment		☒ Post-Stability Checks ☒ Flow Adjustment ☐ Form 40 ☒ Other: Agency Note Internal Control Test																																	
Notes/Suggested Service: Changed date and time after battery change to standard time. Changed time to reflect military time in between stability checks. KTS 6/10/26 Conducted DI prior to post-stability checks. Performed post-stability checks and conduct a second DI. KTS 6/17/26 * indicates the same serial number used for equipment for the second department inspection and solution lot numbers. KTS 6/17/26		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																																			
		☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																																			
		☒ Conduct an Agency Inspection Before Evidentiary Use																																			
Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.06.30 14:35:45 -04'00'		LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.07.09 16:46:42 -04'00'																																			
Tech Review		Admin Review																																			

PORT Richey P.D.

BATTERY change

WLP 6/10/20 80-000927

80-000967 Flow Adjustment

fatg 6/10/26

PORT RICHEY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000967
06/10/2026
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 5.195

2: Rate (Liters/min) = 15

SQRT(Diff) = 9.797

3: Rate (Liters/min) = 30

SQRT(Diff) = 17.145

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 815

Rounded Intercept = -569503

Correlation = 0.99986

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L $\leq 0.003 \text{ g/210L of Wet}$																																																																																																																																																
☒ PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000967 06/10/2026 Software: 8100.27 <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>02:58</td></tr> <tr><td>Control Test</td><td>0.048</td><td>02:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>02:59</td></tr> <tr><td>Control Test</td><td>0.048</td><td>03:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>03:01</td></tr> <tr><td>Control Test</td><td>0.048</td><td>03:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>03:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>KATZ</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	02:58	Control Test	0.048	02:59	Air Blank	0.000	02:59	Control Test	0.048	03:00	Air Blank	0.000	03:01	Control Test	0.048	03:01	Air Blank	0.000	03:02	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		☒ PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000967 06/10/2026 Software: 8100.27 <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:20</td></tr> <tr><td>Control Test</td><td>0.076</td><td>15:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:22</td></tr> <tr><td>Control Test</td><td>0.077</td><td>15:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:23</td></tr> <tr><td>Control Test</td><td>0.076</td><td>15:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0763</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7564</td><td></td></tr> </tbody> </table> <i>KATZ</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	15:20	Control Test	0.076	15:21	Air Blank	0.000	15:22	Control Test	0.077	15:22	Air Blank	0.000	15:23	Control Test	0.076	15:24	Air Blank	0.000	15:24	Control Test Stats			Average	0.0763		Std Dev	0.0006		Rel Std Dev(%)	0.7564		☒ PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000967 06/10/2026 Software: 8100.27 <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:28</td></tr> <tr><td>Control Test</td><td>0.194</td><td>15:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:29</td></tr> <tr><td>Control Test</td><td>0.194</td><td>15:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:30</td></tr> <tr><td>Control Test</td><td>0.194</td><td>15:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1940</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>KATZ</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	15:28	Control Test	0.194	15:28	Air Blank	0.000	15:29	Control Test	0.194	15:30	Air Blank	0.000	15:30	Control Test	0.194	15:31	Air Blank	0.000	15:31	Control Test Stats			Average	0.1940		Std Dev	0.0000		Rel Std Dev(%)	0.0000		☒ PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000967 06/10/2026 Software: 8100.27 <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:16</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:17</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:18</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:19</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> <i>KATZ</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	15:16	Control Test	0.079	15:17	Air Blank	0.000	15:17	Control Test	0.079	15:17	Air Blank	0.000	15:18	Control Test	0.080	15:18	Air Blank	0.000	15:19	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	02:58																																																																																																																																																	
Control Test	0.048	02:59																																																																																																																																																	
Air Blank	0.000	02:59																																																																																																																																																	
Control Test	0.048	03:00																																																																																																																																																	
Air Blank	0.000	03:01																																																																																																																																																	
Control Test	0.048	03:01																																																																																																																																																	
Air Blank	0.000	03:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:20																																																																																																																																																	
Control Test	0.076	15:21																																																																																																																																																	
Air Blank	0.000	15:22																																																																																																																																																	
Control Test	0.077	15:22																																																																																																																																																	
Air Blank	0.000	15:23																																																																																																																																																	
Control Test	0.076	15:24																																																																																																																																																	
Air Blank	0.000	15:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0763																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7564																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:28																																																																																																																																																	
Control Test	0.194	15:28																																																																																																																																																	
Air Blank	0.000	15:29																																																																																																																																																	
Control Test	0.194	15:30																																																																																																																																																	
Air Blank	0.000	15:30																																																																																																																																																	
Control Test	0.194	15:31																																																																																																																																																	
Air Blank	0.000	15:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1940																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:16																																																																																																																																																	
Control Test	0.079	15:17																																																																																																																																																	
Air Blank	0.000	15:17																																																																																																																																																	
Control Test	0.079	15:17																																																																																																																																																	
Air Blank	0.000	15:18																																																																																																																																																	
Control Test	0.080	15:18																																																																																																																																																	
Air Blank	0.000	15:19																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		

PORT RICHEY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000967
06/10/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	02:54
Control Test	0.000	02:54
Air Blank	0.000	02:54
Control Test	0.000	02:55
Air Blank	0.000	02:55
Control Test	0.000	02:55
Air Blank	0.000	02:56
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

80-000967 Dry Gas Stability
#1 - Dry Gas Regulator Connection
Cord was ajar. Fully connected
and reran analysis

 6/10/26

80-000967 Cal. Adjustment

6/11/26 *Kaz*

PORT RICHEY PD

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-000967

06/11/2026

08:24:08

Auto Calibration

Max Power Res Value = 23

Auto Range Res Value = 7

Sol Value = 0.000 g/210L ***

Fit value = 0.0000 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12362, 9um Io = 14208

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.0970 (-0.0250)

Sample #2 = 0.1050 (-0.0330)

Sample #3 = 0.0930 (-0.0160)

Sample #4 = 0.0990 (-0.0270)

Avg % Abs = 0.0990 (-0.0253)

STD DEV = 0.0060 (0.0086)

REL STD DEV = 6.061 (34.033)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.0980 (0.0140)

Sample #2 = 0.1220 (-0.0100)

Sample #3 = 0.1060 (0.0100)

Sample #4 = 0.1050 (0.0000)

Avg % Abs = 0.1110 (0.0000)

STD DEV = 0.0095 (0.0100)

REL STD DEV = 8.594 (0.000)

Sol Value = 0.040 g/210L ***

Fit value = 0.1905 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12361, 9um Io = 14205

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.8510 (-0.0240)

Sample #2 = 0.8740 (-0.0170)

Sample #3 = 0.8570 (-0.0190)

Sample #4 = 0.8690 (-0.0100)

Avg % Abs = 0.8667 (-0.0153)

STD DEV = 0.0087 (0.0047)

REL STD DEV = 1.008 (30.821)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 1.5350 (-0.0040)

Sample #2 = 1.5380 (0.0030)

Sample #3 = 1.5590 (-0.0040)

Sample #4 = 1.5650 (-0.0060)

Avg % Abs = 1.5540 (-0.0023)

STD DEV = 0.0142 (0.0047)

REL STD DEV = 0.912 (202.535)

Sol Value = 0.100 g/210L ***

Fit value = 0.4762 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12358, 9um Io = 14203

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 1.9530 (-0.0130)

Sample #2 = 1.9400 (-0.0210)

Sample #3 = 1.9640 (-0.0140)

Sample #4 = 1.9550 (-0.0300)

Avg % Abs = 1.9530 (-0.0217)

STD DEV = 0.0121 (0.0080)

REL STD DEV = 0.621 (37.019)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 3.6410 (-0.0160)

Sample #2 = 3.6020 (-0.0010)

Sample #3 = 3.6100 (0.0010)

Sample #4 = 3.6440 (-0.0190)

Avg % Abs = 3.6187 (-0.0063)

STD DEV = 0.0223 (0.0110)

REL STD DEV = 0.616 (173.923)

Sol Value = 0.200 g/210L ***

Fit value = 0.9524 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12361, 9um Io = 14204

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 3.7750 (-0.0060)

Sample #2 = 3.7680 (0.0010)

Sample #3 = 3.7700 (0.0290)

Sample #4 = 3.7660 (0.0360)

Avg % Abs = 3.7680 (0.0220)

STD DEV = 0.0020 (0.0185)

REL STD DEV = 0.053 (84.183)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 6.9750 (-0.0160)

Sample #2 = 6.9360 (0.0140)

Sample #3 = 6.9550 (0.0140)

Sample #4 = 6.9440 (0.0250)

Avg % Abs = 6.9450 (0.0177)

STD DEV = 0.0095 (0.0064)

REL STD DEV = 0.137 (35.948)

Sol Value = 0.300 g/210L ***

Fit value = 1.4286 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12358, 9um Io = 14202

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 5.5210 (-0.0350)

Sample #2 = 5.5090 (0.0060)

Sample #3 = 5.5060 (-0.0120)

Sample #4 = 5.5150 (-0.0210)

Avg % Abs = 5.5100 (-0.0090)

STD DEV = 0.0046 (0.0137)

REL STD DEV = 0.083 (152.753)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 10.0830 (0.0000)

Sample #2 = 10.0770 (0.0200)

Sample #3 = 10.0940 (0.0030)

Sample #4 = 10.0810 (0.0060)

Avg % Abs = 10.0840 (0.0097)

STD DEV = 0.0089 (0.0091)

REL STD DEV = 0.088 (93.867)

***** AUTO CAL DATA *****

<<<< CHANNEL 1 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.099

Std Dev = 0.01 Rel Std Dev = 6.06

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.867

Std Dev = 0.01 Rel Std Dev = 1.01

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 1.953

Std Dev = 0.01 Rel Std Dev = 0.62

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 3.768

Std Dev = 0.00 Rel Std Dev = 0.05

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 5.510

Std Dev = 0.00 Rel Std Dev = 0.08

Zero Order Coef = -260.66

First Order Coef = 2513.26

Second Order Coef = 22.91

Standard Deviation = 21.636940

<<<< CHANNEL 2 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.111

Std Dev = 0.01 Rel Std Dev = 8.59

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.554

Std Dev = 0.01 Rel Std Dev = 0.91

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.619

Std Dev = 0.02 Rel Std Dev = 0.62

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 6.945

Std Dev = 0.01 Rel Std Dev = 0.14

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 10.084

Std Dev = 0.01 Rel Std Dev = 0.09

Zero Order Coef = -149.46

First Order Coef = 1310.57

Second Order Coef = 11.97

Standard Deviation = 9.076319

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0002
0.040	0.041	-0.0006
0.100	0.099	0.0006
0.200	0.200	-0.0002
0.300	0.300	0.0001

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0001
0.040	0.040	-0.0002
0.100	0.100	0.0003
0.200	0.200	-0.0001
0.300	0.300	0.0000

Sol Value = 0.080 g/210L ***

Fit value = 0.3810 mg/l %%%

Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 3220.00

Sample #2 = 3144.00

Sample #3 = 3100.00

Sample #4 = 3138.00

Average Result = 3127.3333

STD DEV = 23.8607

REL STD DEV = 0.763

***** CHANNEL 2

Sample #1 = 3441.00

Sample #2 = 3428.00

Sample #3 = 3397.00

Sample #4 = 3406.00

Average Result = 3410.3333

STD DEV = 15.9478

REL STD DEV = 0.468

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1013

3 um H2O Adjust (mg/l*10,000) = 682

9 um H2O Adjust (mg/l*10,000) = 399

**** AUTO CAL PASS

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT RICHEY PD
Time of Inspection: 12:20

Date of Inspection: 06/11/2026

Serial Number: 80-000967
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.050	0.078	0.198	0.081
0.000	0.050	0.078	0.199	0.080
0.000	0.050	0.078	0.198	0.081
0.000	0.050	0.078	0.198	0.081
0.000	0.050	0.078	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.078	0.198	0.081
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.078	0.198	0.080
0.000	0.051	0.079	0.198	0.080


Standard Deviations	0.0003	0.0004	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN

Signature and Printed Name


06/11/2026
Date

Post-Optical

Adjustments

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																																																																																																																												
PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 06/12/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>08:38</td></tr><tr><td>Control Test</td><td>0.049</td><td>08:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:39</td></tr><tr><td>Control Test</td><td>0.049</td><td>08:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:41</td></tr><tr><td>Control Test</td><td>0.050</td><td>08:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:42</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr></tbody></table> PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 06/12/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>08:36</td></tr><tr><td>Control Test</td><td>0.049</td><td>08:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:39</td></tr><tr><td>Control Test</td><td>0.049</td><td>08:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:41</td></tr><tr><td>Control Test</td><td>0.050</td><td>08:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:42</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	08:38	Control Test	0.049	08:39	Air Blank	0.000	08:39	Control Test	0.049	08:40	Air Blank	0.000	08:41	Control Test	0.050	08:41	Air Blank	0.000	08:42	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		Test	g/210L	Time	Air Blank	0.000	08:36	Control Test	0.049	08:39	Air Blank	0.000	08:39	Control Test	0.049	08:40	Air Blank	0.000	08:41	Control Test	0.050	08:41	Air Blank	0.000	08:42	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 06/12/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>08:43</td></tr><tr><td>Control Test</td><td>0.079</td><td>08:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:44</td></tr><tr><td>Control Test</td><td>0.078</td><td>08:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:45</td></tr><tr><td>Control Test</td><td>0.079</td><td>08:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:46</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>1.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.7339</td><td></td></tr></tbody></table> PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 06/12/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>08:43</td></tr><tr><td>Control Test</td><td>0.079</td><td>08:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:44</td></tr><tr><td>Control Test</td><td>0.078</td><td>08:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:45</td></tr><tr><td>Control Test</td><td>0.079</td><td>08:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:46</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>1.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.7339</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	08:43	Control Test	0.079	08:43	Air Blank	0.000	08:44	Control Test	0.078	08:45	Air Blank	0.000	08:45	Control Test	0.079	08:46	Air Blank	0.000	08:46	Control Test Stats			Average	0.0787		Std Dev	1.0006		Rel Std Dev(%)	1.7339		Test	g/210L	Time	Air Blank	0.000	08:43	Control Test	0.079	08:43	Air Blank	0.000	08:44	Control Test	0.078	08:45	Air Blank	0.000	08:45	Control Test	0.079	08:46	Air Blank	0.000	08:46	Control Test Stats			Average	0.0787		Std Dev	1.0006		Rel Std Dev(%)	1.7339		PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 06/12/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>08:48</td></tr><tr><td>Control Test</td><td>0.198</td><td>08:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:49</td></tr><tr><td>Control Test</td><td>0.199</td><td>08:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:50</td></tr><tr><td>Control Test</td><td>0.198</td><td>08:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:51</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1983</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr></tbody></table> PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 06/12/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>08:27</td></tr><tr><td>Control Test</td><td>0.080</td><td>08:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:27</td></tr><tr><td>Control Test</td><td>0.080</td><td>08:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:28</td></tr><tr><td>Control Test</td><td>0.079</td><td>08:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:29</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	08:48	Control Test	0.198	08:48	Air Blank	0.000	08:49	Control Test	0.199	08:50	Air Blank	0.000	08:50	Control Test	0.198	08:51	Air Blank	0.000	08:51	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		Test	g/210L	Time	Air Blank	0.000	08:27	Control Test	0.080	08:27	Air Blank	0.000	08:27	Control Test	0.080	08:28	Air Blank	0.000	08:28	Control Test	0.079	08:29	Air Blank	0.000	08:29	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		PORT RICHEY PD Intoxilyzer - Alcohol Analyzer Model 8000 06/12/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>08:27</td></tr><tr><td>Control Test</td><td>0.080</td><td>08:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:27</td></tr><tr><td>Control Test</td><td>0.080</td><td>08:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:28</td></tr><tr><td>Control Test</td><td>0.079</td><td>08:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:29</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	08:27	Control Test	0.080	08:27	Air Blank	0.000	08:27	Control Test	0.080	08:28	Air Blank	0.000	08:28	Control Test	0.079	08:29	Air Blank	0.000	08:29	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																																																																																																																													
Air Blank	0.000	08:38																																																																																																																																																																																																																																																													
Control Test	0.049	08:39																																																																																																																																																																																																																																																													
Air Blank	0.000	08:39																																																																																																																																																																																																																																																													
Control Test	0.049	08:40																																																																																																																																																																																																																																																													
Air Blank	0.000	08:41																																																																																																																																																																																																																																																													
Control Test	0.050	08:41																																																																																																																																																																																																																																																													
Air Blank	0.000	08:42																																																																																																																																																																																																																																																													
Control Test Stats																																																																																																																																																																																																																																																															
Average	0.0493																																																																																																																																																																																																																																																														
Std Dev	0.0006																																																																																																																																																																																																																																																														
Rel Std Dev(%)	1.1703																																																																																																																																																																																																																																																														
Test	g/210L	Time																																																																																																																																																																																																																																																													
Air Blank	0.000	08:36																																																																																																																																																																																																																																																													
Control Test	0.049	08:39																																																																																																																																																																																																																																																													
Air Blank	0.000	08:39																																																																																																																																																																																																																																																													
Control Test	0.049	08:40																																																																																																																																																																																																																																																													
Air Blank	0.000	08:41																																																																																																																																																																																																																																																													
Control Test	0.050	08:41																																																																																																																																																																																																																																																													
Air Blank	0.000	08:42																																																																																																																																																																																																																																																													
Control Test Stats																																																																																																																																																																																																																																																															
Average	0.0493																																																																																																																																																																																																																																																														
Std Dev	0.0006																																																																																																																																																																																																																																																														
Rel Std Dev(%)	1.1703																																																																																																																																																																																																																																																														
Test	g/210L	Time																																																																																																																																																																																																																																																													
Air Blank	0.000	08:43																																																																																																																																																																																																																																																													
Control Test	0.079	08:43																																																																																																																																																																																																																																																													
Air Blank	0.000	08:44																																																																																																																																																																																																																																																													
Control Test	0.078	08:45																																																																																																																																																																																																																																																													
Air Blank	0.000	08:45																																																																																																																																																																																																																																																													
Control Test	0.079	08:46																																																																																																																																																																																																																																																													
Air Blank	0.000	08:46																																																																																																																																																																																																																																																													
Control Test Stats																																																																																																																																																																																																																																																															
Average	0.0787																																																																																																																																																																																																																																																														
Std Dev	1.0006																																																																																																																																																																																																																																																														
Rel Std Dev(%)	1.7339																																																																																																																																																																																																																																																														
Test	g/210L	Time																																																																																																																																																																																																																																																													
Air Blank	0.000	08:43																																																																																																																																																																																																																																																													
Control Test	0.079	08:43																																																																																																																																																																																																																																																													
Air Blank	0.000	08:44																																																																																																																																																																																																																																																													
Control Test	0.078	08:45																																																																																																																																																																																																																																																													
Air Blank	0.000	08:45																																																																																																																																																																																																																																																													
Control Test	0.079	08:46																																																																																																																																																																																																																																																													
Air Blank	0.000	08:46																																																																																																																																																																																																																																																													
Control Test Stats																																																																																																																																																																																																																																																															
Average	0.0787																																																																																																																																																																																																																																																														
Std Dev	1.0006																																																																																																																																																																																																																																																														
Rel Std Dev(%)	1.7339																																																																																																																																																																																																																																																														
Test	g/210L	Time																																																																																																																																																																																																																																																													
Air Blank	0.000	08:48																																																																																																																																																																																																																																																													
Control Test	0.198	08:48																																																																																																																																																																																																																																																													
Air Blank	0.000	08:49																																																																																																																																																																																																																																																													
Control Test	0.199	08:50																																																																																																																																																																																																																																																													
Air Blank	0.000	08:50																																																																																																																																																																																																																																																													
Control Test	0.198	08:51																																																																																																																																																																																																																																																													
Air Blank	0.000	08:51																																																																																																																																																																																																																																																													
Control Test Stats																																																																																																																																																																																																																																																															
Average	0.1983																																																																																																																																																																																																																																																														
Std Dev	0.0006																																																																																																																																																																																																																																																														
Rel Std Dev(%)	0.2911																																																																																																																																																																																																																																																														
Test	g/210L	Time																																																																																																																																																																																																																																																													
Air Blank	0.000	08:27																																																																																																																																																																																																																																																													
Control Test	0.080	08:27																																																																																																																																																																																																																																																													
Air Blank	0.000	08:27																																																																																																																																																																																																																																																													
Control Test	0.080	08:28																																																																																																																																																																																																																																																													
Air Blank	0.000	08:28																																																																																																																																																																																																																																																													
Control Test	0.079	08:29																																																																																																																																																																																																																																																													
Air Blank	0.000	08:29																																																																																																																																																																																																																																																													
Control Test Stats																																																																																																																																																																																																																																																															
Average	0.0797																																																																																																																																																																																																																																																														
Std Dev	0.0006																																																																																																																																																																																																																																																														
Rel Std Dev(%)	0.7247																																																																																																																																																																																																																																																														
Test	g/210L	Time																																																																																																																																																																																																																																																													
Air Blank	0.000	08:27																																																																																																																																																																																																																																																													
Control Test	0.080	08:27																																																																																																																																																																																																																																																													
Air Blank	0.000	08:27																																																																																																																																																																																																																																																													
Control Test	0.080	08:28																																																																																																																																																																																																																																																													
Air Blank	0.000	08:28																																																																																																																																																																																																																																																													
Control Test	0.079	08:29																																																																																																																																																																																																																																																													
Air Blank	0.000	08:29																																																																																																																																																																																																																																																													
Control Test Stats																																																																																																																																																																																																																																																															
Average	0.0797																																																																																																																																																																																																																																																														
Std Dev	0.0006																																																																																																																																																																																																																																																														
Rel Std Dev(%)	0.7247																																																																																																																																																																																																																																																														

 Operator's Signature

 Operator's Signature

 Operator's Signature

 Operator's Signature

80-000967 Internal Control Test

Inadvertently hit '1' instead
of 's' for stability check.

[Signature] 6/12/26

PORT RICHEY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000967
06/12/2026
Software: 8100.27

ITP CONTROL TEST

Test	g/210L	Time
Air Blank	ABT*	08:37
Air Blank	0.000	08:37

*Sequence Aborted

[Signature]
Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT RICHEY PD
Time of Inspection: 12:29

Date of Inspection: 06/17/2026

Serial Number: 80-000967
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.078	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.079	0.198	0.080


Standard Deviations	0.0004	0.0003	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN

Signature and Printed Name

06/17/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000967, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000967</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>PORT RICHEY PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/17/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:29</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/17/2026

Date

KATIE T SPEARIN,

Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality