

Alcohol Testing Program - Instrument Processing Sheet

Agency: Santa Rosa CO SO

Instrument Serial Number: 80-000772

Date In: 7/17/2026

DI Completion Date: 7/23/2026

☐ Ship

☐ P/U

☐ H/D

☒ CMI

☐ EE

Intake By: <u>IAS</u> Date: <u>7/17/2026</u>		Quality Checks By: <u>WKP</u> Date: <u>7/22/2026</u>		Flow Adjustment By: _____																																							
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: AI Note Attached. IAS 7/17/26		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>216</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.148</u> (.139-.169) 36 mm <u>0.160</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.500</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>34419</u> Gauge: <u>1012</u> Instrument: <u>1008</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																							
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6291	202603A 3/3/2028	0.080	MP6292	202603B 3/3/2028	0.200	MP6293	202603D 3/4/2028	0.080 DGS	N/A	23825080A1 10/5/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____																				
						Simulator	Serial #	Lot#/Exp																																			
						0.050	MP6291	202603A 3/3/2028																																			
						0.080	MP6292	202603B 3/3/2028																																			
0.200	MP6293	202603D 3/4/2028																																									
0.080 DGS	N/A	23825080A1 10/5/2027																																									
Optical Bench Adjustment By: _____ Department Inspection By: <u>WKP</u> Barometric Pressure Gauge: _____ ID#: _____ Barometric Pressure ID#: <u>28427</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1014</u> Instrument: <u>1012</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP6291</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																								
0.000		N/A	N/A																																								
0.040																																											
0.100																																											
0.200																																											
0.300																																											
0.080 DGS	N/A																																										
Simulator	Serial Number																																										
0.000	MP6289																																										
Interferent	MP6290																																										
0.050	MP6291																																										
0.080	MP6292																																										
0.200	MP6293																																										
☐ Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			☐ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment		☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☒ Other: AI Note Return Material Authorization																			
Simulator	Serial #	Lot #	Expiration																																								
0.050																																											
0.080																																											
0.200																																											
0.080 DGS	N/A																																										
Gauge ID #: _____ Gauge: _____ Instrument: _____ Notes/Suggested Service: Upon beginning stability checks I visually confirmed that the modem port was pushed into the instrument and needs to go to repair. WKP 7/22/2026 Admin review: Added zip code to the bill to address. WKP 7/31/2026		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																									
		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.07.27 13:04:01 -04'00' Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.07.31 10:44:25 -04'00' Tech Review Admin Review																																									



Santa Rosa Sheriff's Office
Sheriff Bob Johnson
Patrol Division

TO: Florida Department of Law Enforcement Alcohol Testing Program
FROM: Sergeant Raymond Spencer
DATE: July 13, 2026
RE: Intoxilyzer S/N 80-000772

Enclosed is Intox S/N 80-000772 that is being shipped for its yearly inspection. This instrument data port has broken and I am unable to upload the monthly inspection report. If the port can't be fixed by FDLE, can this be shipped to CMI for repair.

Thanks,

A handwritten signature in blue ink that reads "Raymond Spencer".

Sergeant Raymond Spencer

IAS 7/17/26

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
SANTA ROSA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/22/2006 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:14 Control Test 0.049 12:15 Air Blank 0.000 12:16 Control Test 0.049 12:16 Air Blank 0.000 12:17 Control Test 0.049 12:18 Air Blank 0.000 12:19 Control Test 0.049 12:19 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>Wenji</i>	SANTA ROSA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/22/2006 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:21 Control Test 0.079 12:21 Air Blank 0.000 12:22 Control Test 0.077 12:23 Air Blank 0.000 12:23 Control Test 0.078 12:24 Air Blank 0.000 12:24 Control Test 0.077 12:24 Control Test Stats Average 0.0777 Std Dev 0.0006 Rel Std Dev(%) 0.7434 Operator's Signature <i>Wenji</i>	SANTA ROSA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/22/2006 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:28 Control Test 0.197 12:29 Air Blank 0.000 12:30 Control Test 0.196 12:30 Air Blank 0.000 12:31 Control Test 0.196 12:32 Air Blank 0.000 12:32 Control Test Stats Average 0.1963 Std Dev 0.0006 Rel Std Dev(%) 0.2941 Operator's Signature <i>Wenji</i>	SANTA ROSA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/22/2006 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:39 Control Test 0.080 12:39 Air Blank 0.000 12:39 Control Test 0.081 12:39 Air Blank 0.000 12:40 Control Test 0.081 12:40 Air Blank 0.000 12:41 Control Test 0.081 12:41 Control Test Stats Average 0.0807 Std Dev 0.0006 Rel Std Dev(%) 0.7157 Operator's Signature <i>Wenji</i>

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SANTA ROSA COUNTY SO
Time of Inspection: 12:14

Date of Inspection: 07/23/2026

Serial Number: 80-000772
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.048	0.078	0.195	0.081
0.000	0.049	0.078	0.196	0.081
0.000	0.049	0.078	0.197	0.081
0.000	0.049	0.078	0.195	0.082
0.000	0.049	0.077	0.196	0.081
0.000	0.049	0.078	0.196	0.082
0.000	0.049	0.078	0.197	0.082
0.000	0.049	0.078	0.196	0.081
0.000	0.049	0.078	0.196	0.082
0.000	0.049	0.078	0.197	0.081
Standard Deviations	0.0003	0.0003	0.0007	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WEN-CHI K. PIERSON

Signature and Printed Name

07/23/2026
Date

Return Material Authorization

Ship to:☒ CMI, Inc.☐ Enforcement ElectronicsShipment to repair facility authorized by: Raymond Spencer on 7/13/2026Items Returned: ☒ Instrument ☐ Supplies ☐ Other Describe: Intoxilyzer 8000Instrument Model: 80-000772 Serial Number: 80-000772

Bill To Address:

Attn: Sergeant Raymond Spencer
5755 E Milton Rd Milton, FL 32583

Ship to Address:

FDLE Off-Site Mail Facility
c/o Florida Department of Law Enforcement
Alcohol Testing Program
813 B Lake Bradford Road
Tallahassee, FL 32304

Reason for Return:

Modem port is pushed inside of the instrument.

I require an estimate **BEFORE** any repairs will be authorized and/or conductedPlease contact: Raymond Spencer Phone #: (850) 454-7668Email: rspencer@srsos.net

ATP Contact Name: Wen-Chi Pierson

ATP Email: Wen-ChiPierson@fdle.state.fl.us



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000772, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000772</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>SANTA ROSA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/23/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:14</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi

Chi Pierson

Date: 2026.07.23
15:14:10 -04'00'

Pierson

07/23/2026

Date

WEN-CHI K PIERSON,

Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality