

Alcohol Testing Program - Instrument Processing Sheet

Agency: Florida Highway Patrol

Instrument Serial Number: 80-006630

Date In: 7/28/2026

DI Completion Date: 8/4/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>TDG</u> Date: <u>8/3/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>8/4/2026</u>		Flow Adjustment By: _____																											
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>249</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm 0.164 (.139-.169) 36 mm 0.175 (.156-.190) 53 mm 0.250 (.228-.278) 103 mm 0.511 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1016</u> Instrument: <u>1013</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																											
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>28424080A3 11/5/2026</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6286	202603A 3/3/2028	0.080	MP6287	202603B 3/3/2028	0.200	MP6288	202603D 3/4/2028	0.080 DGS	N/A	28424080A3 11/5/2026	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:								
						Simulator	Serial #	Lot#/Exp																							
						0.050	MP6286	202603A 3/3/2028																							
						0.080	MP6287	202603B 3/3/2028																							
0.200	MP6288	202603D 3/4/2028																													
0.080 DGS	N/A	28424080A3 11/5/2026																													
Optical Bench Adjustment By: _____		Department Inspection By: <u>TDG</u>																													
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>33363</u>																													
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1017</u> Instrument: <u>1014</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>	
Simulator	Serial #	Lot #	Expiration																												
0.000		N/A	N/A																												
0.040																															
0.100																															
0.200																															
0.300																															
0.080 DGS	N/A																														
☐ Post Optical Bench Adjustment Stability Checks		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 50%;">Simulator</th> <th style="width: 50%;">Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																
Simulator	Serial Number																														
0.000	MP6284																														
Interferent	MP6285																														
0.050	MP6286																														
0.080	MP6287																														
0.200	MP6288																														
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment									
Simulator	Serial #	Lot #	Expiration																												
0.050																															
0.080																															
0.200																															
0.080 DGS	N/A																														
Gauge ID #: _____ Gauge: _____ Instrument: _____		☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☒ Other: Calculation Sheet																													
Notes/Suggested Service:		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																													
		☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																													
		☒ Conduct an Agency Inspection Before Evidentiary Use																													
LeAndra Higginbotham		Kaitlyn Spearin																													
Digitally signed by LeAndra Higginbotham Date: 2026.08.06 11:03:42 -04'00'		Digitally signed by Kaitlyn Spearin Date: 2026.08.06 11:34:07 -04'00'																													
Tech Review		Admin Review																													

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤0.003 of Wet																																																																																																																																																
FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006630 08/04/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:27</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:27	Control Test	0.050	11:28	Air Blank	0.000	11:29	Control Test	0.049	11:29	Air Blank	0.000	11:30	Control Test	0.050	11:30	Air Blank	0.000	11:31	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006630 08/04/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:12</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:13</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:14</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:12	Control Test	0.079	11:12	Air Blank	0.000	11:13	Control Test	0.079	11:14	Air Blank	0.000	11:14	Control Test	0.079	11:15	Air Blank	0.000	11:15	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006630 08/04/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr> <tr><td>Control Test</td><td>0.197</td><td>11:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1977</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2921</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:18	Control Test	0.198	11:18	Air Blank	0.000	11:19	Control Test	0.197	11:20	Air Blank	0.000	11:20	Control Test	0.198	11:21	Air Blank	0.000	11:21	Control Test Stats			Average	0.1977		Std Dev	0.0006		Rel Std Dev(%)	0.2921		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006630 08/04/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:01</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:03</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0773</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7466</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:01	Control Test	0.078	11:02	Air Blank	0.000	11:02	Control Test	0.077	11:02	Air Blank	0.000	11:03	Control Test	0.077	11:03	Air Blank	0.000	11:04	Control Test Stats			Average	0.0773		Std Dev	0.0006		Rel Std Dev(%)	0.7466	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:27																																																																																																																																																	
Control Test	0.050	11:28																																																																																																																																																	
Air Blank	0.000	11:29																																																																																																																																																	
Control Test	0.049	11:29																																																																																																																																																	
Air Blank	0.000	11:30																																																																																																																																																	
Control Test	0.050	11:30																																																																																																																																																	
Air Blank	0.000	11:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0497																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1625																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:12																																																																																																																																																	
Control Test	0.079	11:12																																																																																																																																																	
Air Blank	0.000	11:13																																																																																																																																																	
Control Test	0.079	11:14																																																																																																																																																	
Air Blank	0.000	11:14																																																																																																																																																	
Control Test	0.079	11:15																																																																																																																																																	
Air Blank	0.000	11:15																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:18																																																																																																																																																	
Control Test	0.198	11:18																																																																																																																																																	
Air Blank	0.000	11:19																																																																																																																																																	
Control Test	0.197	11:20																																																																																																																																																	
Air Blank	0.000	11:20																																																																																																																																																	
Control Test	0.198	11:21																																																																																																																																																	
Air Blank	0.000	11:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1977																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2921																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:01																																																																																																																																																	
Control Test	0.078	11:02																																																																																																																																																	
Air Blank	0.000	11:02																																																																																																																																																	
Control Test	0.077	11:02																																																																																																																																																	
Air Blank	0.000	11:03																																																																																																																																																	
Control Test	0.077	11:03																																																																																																																																																	
Air Blank	0.000	11:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0773																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7466																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 15:12

Date of Inspection: 08/04/2026

Serial Number: 80-006630
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28424080A3 Exp: 11/05/2026
0.000	0.049	0.078	0.196	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.050	0.078	0.197	0.078
0.000	0.050	0.078	0.197	0.077
0.000	0.050	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.050	0.078	0.197	0.077
0.000	0.050	0.078	0.197	0.077
0.000	0.050	0.079	0.197	0.077
0.000	0.050	0.078	0.197	0.077

Standard Deviations	0.0004	0.0003	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

08/04/2026
Date

Florida Department of Law Enforcement
Alcohol Testing Program - Average Calculations for Calibration Certificate

Calculations By: TDG

Serial Number: 80-006630

Date: 8/6/2026

Solution:

0.05 g/210 L Test: 0.497 / 10 = 0.050 g/210L
(SUM) (AVERAGE)

0.08 g/210L Test: 0.781 / 10 = 0.078 g/210L
(SUM) (AVERAGE)

0.20 g/210L Test: 1.969 / 10 = 0.197 g/210L
(SUM) (AVERAGE)

0.08 g/210L Dry Gas Std Test: 0.775 / 10 = 0.078 g/210L
(SUM) (AVERAGE)

Comments:

Average = (Sum of observations) / (Total number of observations)

All results are reported in g/210L



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006630, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006630</u>	Average of $y \pm$ Uncertainty*	
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>08/04/2026</u>	0.078 g/210 L	0.004
Calibration Time:	<u>15:12</u>	0.197 g/ 210 L	0.008
		0.078 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/04/2026

Date

Taylor Gutschow

Digitally signed by Taylor
Gutschow
Date: 2026.08.06 10:06:02 -04'00'

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 August 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1