

Return Material Authorization

Ship to: CMI, Inc.
 Enforcement Electronics

Shipment to repair facility authorized by: Ofc. Matthew Rasmussen on 9/12/2025

Items Returned: Instrument Supplies Other Describe: _____
Instrument Model: Intoxilyzer 8000 Serial Number: 80-001649

Bill To Address:
Titusville Police Department

1100 John Glenn Blvd.

Titusville, Florida, 32780

Ship to Address:
FDLE Tallahassee

Reason for Return:
DSP fail on initialization diagnostic check.

Please choose one of the following options:

1. I _____, authorize all repairs.

2. I _____, authorize repairs up to \$_____.

3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Ofc. Matthew Rasmussen
Phone #: 321-576-6067 Email: Matthew.Rasmussen@titusville.com

ATP Contact Name: LeAndra Higginbotham ATP Email: LeAndraHigginbotham@fdle.state.fl.us