

Alcohol Testing Program - Instrument Processing Sheet

Agency: Suwannee County SO

Instrument Serial Number: 80-000768

Date In: 5/4/2026

DI Completion Date: 05/18/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>SLH</u> Date: <u>5/4/2026</u>		Quality Checks By: <u>SLH</u> Date: <u>5/8/2026</u>		Flow Adjustment By: _____																													
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>213</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP105</u> 32 mm <u>0.152</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.250</u> (.228-.278) 103 mm <u>0.527</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1010</u> Instrument: <u>1012</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																													
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A 03/03/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B 03/03/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D 03/04/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/05/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 03/03/2028	0.080	MP5089	202603B 03/03/2028	0.200	MP5090	202603D 03/04/2028	0.080 DGS	N/A	23825080A1 10/05/2027	Maintenance By: <u>SLH</u> Date: <u>5/6/2026</u> ☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:														
Simulator	Serial #	Lot#/Exp																															
0.050	MP5088	202603A 03/03/2028																															
0.080	MP5089	202603B 03/03/2028																															
0.200	MP5090	202603D 03/04/2028																															
0.080 DGS	N/A	23825080A1 10/05/2027																															
Optical Bench Adjustment By: <u>WKP</u> Barometric Pressure Gauge: <u>1015</u> ID#: <u>34420</u>		Department Inspection By: <u>SLH</u> Barometric Pressure ID#: <u>28662</u>																															
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>25090</td> <td>3/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>25282</td> <td>8/12/2027</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>25020</td> <td>1/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>25150</td> <td>5/6/2027</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701</td> <td>4/17/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	3/11/2027	0.100	MP6296	25282	8/12/2027	0.200	MP6297	25020	1/14/2027	0.300	MP6298	25150	5/6/2027	0.080 DGS	N/A	AG510701	4/17/2027	Gauge: <u>1019</u> Instrument: <u>1019</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u>			
Simulator	Serial #	Lot #	Expiration																														
0.000	MP6294	N/A	N/A																														
0.040	MP6295	25090	3/11/2027																														
0.100	MP6296	25282	8/12/2027																														
0.200	MP6297	25020	1/14/2027																														
0.300	MP6298	25150	5/6/2027																														
0.080 DGS	N/A	AG510701	4/17/2027																														
☒ Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A</td> <td>3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B</td> <td>3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D</td> <td>3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23725080A1</td> <td>10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	3/3/2028	0.080	MP5089	202603B	3/3/2028	0.200	MP5090	202603D	3/4/2028	0.080 DGS	N/A	23725080A1	10/5/2027	<table border="1" style="width: 100%; border-collapse: collapse;"> <tbody> <tr> <td style="width: 50%;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment </td> <td style="width: 50%;"> ☒ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other: </td> </tr> </tbody> </table>				☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment	☒ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other:						
Simulator	Serial #	Lot #	Expiration																														
0.050	MP5088	202603A	3/3/2028																														
0.080	MP5089	202603B	3/3/2028																														
0.200	MP5090	202603D	3/4/2028																														
0.080 DGS	N/A	23725080A1	10/5/2027																														
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment	☒ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other:																																
Gauge ID #: <u>28427</u> Gauge: <u>1014</u> Instrument: <u>1014</u>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																															
Notes/Suggested Service:		<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 50%; text-align: center;"> Digitally signed by Taylor Gutschow Date: 2026.05.20 15:42:08 -04'00' </td> <td style="width: 50%; text-align: center;"> Digitally signed by Kaitlyn Spearin Date: 2026.05.22 13:15:40 -04'00' </td> </tr> </table>				Digitally signed by Taylor Gutschow Date: 2026.05.20 15:42:08 -04'00'	Digitally signed by Kaitlyn Spearin Date: 2026.05.22 13:15:40 -04'00'																										
Digitally signed by Taylor Gutschow Date: 2026.05.20 15:42:08 -04'00'	Digitally signed by Kaitlyn Spearin Date: 2026.05.22 13:15:40 -04'00'																																
		Tech Review Admin Review																															

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUWANNEE CO SO
Time of Inspection: 13:25

Date of Inspection: 05/08/2026

Serial Number: 80-000768
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Number of Simulators Used: _____

Remarks:

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.




LEANDRA HIGGINBOTHAM
Signature and Printed Name

05/08/2026
Date

SUWANNEE CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000768
05/08/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:48
Control Test	0.048	13:49
Air Blank	0.000	13:49
Control Test	0.048	13:50
Air Blank	0.000	13:50
Control Test	0.048	13:51
Air Blank	0.000	13:52
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

SUWANNEE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000768
 05/08/2026
 Software: 8100.27

NO user or
 equipment
 error -

need

OBA

OBA = Optical Bench Adjustment

Sut

Test	g/210L	Time
Air Blank	0.000	13:53
Control Test	0.076	13:54
Air Blank	0.000	13:54
Control Test	0.076	13:55
Air Blank	0.000	13:55
Control Test	0.076	13:56
Air Blank	0.000	13:57
Control Test Stats		
Average	0.0760 X	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

SUWANNEE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000768
 05/08/2026
 Software: 8100.27


 Operator's Signature

Test	g/210L	Time
Air Blank	0.000	13:58
Control Test	0.082	13:59
Air Blank	0.000	13:59
Control Test	0.081	13:59
Air Blank	0.000	14:00
Control Test	0.082	14:00
Air Blank	0.000	14:01
Control Test Stats		
Average	0.0817 X	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	


 Operator's Signature

SUWANNEE CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000768
05/08/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:43
Control Test	0.198	13:44
Air Blank	0.000	13:44
Control Test	0.197	13:45
Air Blank	0.000	13:45
Control Test	0.197	13:46
Air Blank	0.000	13:47
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	



Operator's Signature

SUNANEE CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000768
05/15/2026 11:19:49

Auto Calibration
Max Power Res Value = 33
Auto Range Res Value = 7

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12635, Sum Io = 13835

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.1400 (0.0040)
Sample #2 = 0.1210 (0.0760)
Sample #3 = 0.1620 (0.1120)
Sample #4 = 0.1470 (0.1580)
Avg % Abs = 0.1433 (0.1153)
STD DEV = 0.0207 (0.0411)
REL STD DEV = 14.473 (35.637)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 0.1590 (-0.0080)
Sample #2 = 0.1330 (0.0180)
Sample #3 = 0.1810 (-0.0010)
Sample #4 = 0.1550 (0.0270)
Avg % Abs = 0.1563 (0.0147)
STD DEV = 0.0240 (0.0143)
REL STD DEV = 15.370 (97.463)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12604, Sum Io = 13830

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.8520 (-0.0020)
Sample #2 = 0.8860 (0.0150)
Sample #3 = 0.8680 (0.0500)
Sample #4 = 0.8400 (0.0960)
Avg % Abs = 0.8647 (0.0570)
STD DEV = 0.0232 (0.0406)
REL STD DEV = 2.681 (71.199)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.6210 (0.0020)
Sample #2 = 1.6050 (0.0360)
Sample #3 = 1.6100 (0.0530)
Sample #4 = 1.5830 (0.0720)
Avg % Abs = 1.5993 (0.0537)
STD DEV = 0.0144 (0.0180)
REL STD DEV = 0.898 (33.558)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12591, Sum Io = 13821

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 1.9520 (-0.0070)
Sample #2 = 1.9490 (0.0260)
Sample #3 = 1.9270 (0.0660)
Sample #4 = 1.9240 (0.0770)
Avg % Abs = 1.9313 (0.0563)
STD DEV = 0.0102 (0.0268)
REL STD DEV = 0.529 (47.643)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 3.7060 (-0.0070)
Sample #2 = 3.6960 (0.0300)
Sample #3 = 3.6920 (0.0430)
Sample #4 = 3.6990 (0.0450)
Avg % Abs = 3.6957 (0.0393)
STD DEV = 0.0035 (0.0081)
REL STD DEV = 0.095 (20.706)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12574, Sum Io = 13802

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 3.6860 (0.0040)
Sample #2 = 3.6340 (0.0500)
Sample #3 = 3.6330 (0.0780)
Sample #4 = 3.6660 (0.0680)
Avg % Abs = 3.6443 (0.0653)
STD DEV = 0.0188 (0.0142)
REL STD DEV = 0.515 (21.718)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 7.0230 (-0.0030)
Sample #2 = 6.9320 (0.0900)
Sample #3 = 6.9180 (0.1090)
Sample #4 = 6.9300 (0.1050)
Avg % Abs = 6.9267 (0.1013)
STD DEV = 0.0076 (0.0100)
REL STD DEV = 0.109 (9.865)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12573, Sum Io = 13807

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 5.4410 (-0.0090)
Sample #2 = 5.4400 (0.0340)
Sample #3 = 5.4530 (0.0500)
Sample #4 = 5.4320 (0.0710)
Avg % Abs = 5.4417 (0.0517)
STD DEV = 0.0106 (0.0186)
REL STD DEV = 0.195 (35.915)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 10.2870 (-0.0280)
Sample #2 = 10.2370 (0.0350)
Sample #3 = 10.2210 (0.0430)
Sample #4 = 10.2340 (0.0430)
Avg % Abs = 10.2307 (0.0403)
STD DEV = 0.0085 (0.0046)
REL STD DEV = 0.083 (11.452)

***** AUTO CAL DATA *****

Channel 1
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.143
Std Dev = 0.02 Rel Std Dev = 14.47
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.665
Std Dev = 0.02 Rel Std Dev = 2.68
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.931
Std Dev = 0.01 Rel Std Dev = 0.53
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.644
Std Dev = 0.02 Rel Std Dev = 0.52
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.442
Std Dev = 0.01 Rel Std Dev = 0.19
Zero Order Coef = -422.51
First Order Coef = 2717.69
Second Order Coef = -1.94
Standard Deviation = 49.865204

Channel 2
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.156
Std Dev = 0.02 Rel Std Dev = 15.37
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.599
Std Dev = 0.01 Rel Std Dev = 0.90
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.696
Std Dev = 0.00 Rel Std Dev = 0.10
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.927
Std Dev = 0.01 Rel Std Dev = 0.11
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.231
Std Dev = 0.01 Rel Std Dev = 0.08
Zero Order Coef = -246.51
First Order Coef = 1346.29
Second Order Coef = 7.52
Standard Deviation = 60.184425

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.001 0.0007
0.040 0.040 -0.0004
0.100 0.101 -0.0012
0.200 0.199 -0.0014
0.300 0.300 -0.0005

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.001 0.0008
0.040 0.040 -0.0004
0.100 0.101 -0.0015
0.200 0.198 0.0018
0.300 0.301 -0.0006

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %
Samples Taken = 4, Discarded = 1
Channel 1
Sample #1 = 3091.00
Sample #2 = 3119.00
Sample #3 = 3105.00
Sample #4 = 3127.00
Average Result = 3117.0000
STD DEV = 11.1355
REL STD DEV = 0.357

Channel 2
Sample #1 = 3382.00
Sample #2 = 3385.00
Sample #3 = 3371.00
Sample #4 = 3398.00
Average Result = 3384.6667
STD DEV = 13.5031
REL STD DEV = 0.399

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1015
3 um H2O Adjust (mg/l*10,000) = 692
9 um H2O Adjust (mg/l*10,000) = 425
**** AUTO CAL PASS

optical Bench
Calibration
adjustment
wkp
wkp 5/15/26

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053 ✓	0.08g/210L 0.077 to 0.083 ✓	0.20g/210L 0.194 to 0.206 ✓	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																																																																
SUWANNEE CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000768 05/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:12</td></tr> <tr><td>Control Test</td><td>0.051</td><td>12:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:13</td></tr> <tr><td>Control Test</td><td>0.051</td><td>12:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:14</td></tr> <tr><td>Control Test</td><td>0.051</td><td>12:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0510</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	12:12	Control Test	0.051	12:12	Air Blank	0.000	12:13	Control Test	0.051	12:14	Air Blank	0.000	12:14	Control Test	0.051	12:15	Air Blank	0.000	12:16	Control Test Stats			Average	0.0510		Std Dev	0.0000		Rel Std Dev(%)	0.0000		SUWANNEE CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000768 05/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:17</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:18</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:20</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	12:17	Control Test	0.081	12:18	Air Blank	0.000	12:18	Control Test	0.081	12:19	Air Blank	0.000	12:20	Control Test	0.081	12:20	Air Blank	0.000	12:21	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000		SUWANNEE CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000768 05/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:22</td></tr> <tr><td>Control Test</td><td>0.202</td><td>12:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:23</td></tr> <tr><td>Control Test</td><td>0.201</td><td>12:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr> <tr><td>Control Test</td><td>0.201</td><td>12:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.2013</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2868</td><td></td></tr> </tbody> </table> Operator's Signature <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	12:22	Control Test	0.202	12:23	Air Blank	0.000	12:23	Control Test	0.201	12:24	Air Blank	0.000	12:25	Control Test	0.201	12:25	Air Blank	0.000	12:26	Control Test Stats			Average	0.2013		Std Dev	0.0006		Rel Std Dev(%)	0.2868		SUWANNEE CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000768 05/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:27</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:28</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:30</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table> Operator's Signature <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	12:27	Control Test	0.079	12:27	Air Blank	0.000	12:28	Control Test	0.080	12:28	Air Blank	0.000	12:29	Control Test	0.080	12:29	Air Blank	0.000	12:30	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:12																																																																																																																																																	
Control Test	0.051	12:12																																																																																																																																																	
Air Blank	0.000	12:13																																																																																																																																																	
Control Test	0.051	12:14																																																																																																																																																	
Air Blank	0.000	12:14																																																																																																																																																	
Control Test	0.051	12:15																																																																																																																																																	
Air Blank	0.000	12:16																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0510																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:17																																																																																																																																																	
Control Test	0.081	12:18																																																																																																																																																	
Air Blank	0.000	12:18																																																																																																																																																	
Control Test	0.081	12:19																																																																																																																																																	
Air Blank	0.000	12:20																																																																																																																																																	
Control Test	0.081	12:20																																																																																																																																																	
Air Blank	0.000	12:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:22																																																																																																																																																	
Control Test	0.202	12:23																																																																																																																																																	
Air Blank	0.000	12:23																																																																																																																																																	
Control Test	0.201	12:24																																																																																																																																																	
Air Blank	0.000	12:25																																																																																																																																																	
Control Test	0.201	12:25																																																																																																																																																	
Air Blank	0.000	12:26																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2013																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2868																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:27																																																																																																																																																	
Control Test	0.079	12:27																																																																																																																																																	
Air Blank	0.000	12:28																																																																																																																																																	
Control Test	0.080	12:28																																																																																																																																																	
Air Blank	0.000	12:29																																																																																																																																																	
Control Test	0.080	12:29																																																																																																																																																	
Air Blank	0.000	12:30																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SUWANNEE CO SO
Time of Inspection: 12:30

Date of Inspection: 05/18/2026

Serial Number: 80-000768
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.050	0.080	0.202	0.080
0.000	0.051	0.081	0.203	0.080
0.000	0.050	0.081	0.204	0.080
0.000	0.050	0.081	0.203	0.080
0.000	0.050	0.081	0.203	0.080
0.000	0.050	0.081	0.204	0.079
0.000	0.050	0.081	0.203	0.080
0.000	0.050	0.082	0.203	0.079
0.000	0.050	0.081	0.203	0.079
0.000	0.050	0.082	0.203	0.080


Standard Deviations	0.0003	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

05/18/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000768, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000768</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>SUWANNEE CO SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>05/18/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:30</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Date

05/18/2026
Leandra Higginbotham
LEANDRA HIGGINBOTHAM,

Department Inspector

Service • Integrity • Respect • Quality