

Alcohol Testing Program - Instrument Processing Sheet

Agency: DELAND PD

Instrument Serial Number: 80-001114

Date In: 6/24/2026

DI Completion Date: 7/9/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>WKP</u> Date: <u>6/24/2026</u>		Quality Checks By: <u>SLH</u> Date: <u>6/30/2026</u>		Flow Adjustment By: _____																																															
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>216</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.160</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.250</u> (.228-.278) 103 mm <u>0.507</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1011</u> Instrument: <u>1014</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																															
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot#/Exp</th></tr></thead><tbody><tr><td rowspan="2">0.050</td><td rowspan="2">MP5088</td><td>202603A</td></tr><tr><td>03/03/2028</td></tr><tr><td rowspan="2">0.080</td><td rowspan="2">MP5089</td><td>202603B</td></tr><tr><td>03/03/2028</td></tr><tr><td rowspan="2">0.200</td><td rowspan="2">MP5090</td><td>202603D</td></tr><tr><td>03/04/2028</td></tr><tr><td rowspan="2">0.080 DGS</td><td rowspan="2">N/A</td><td>23825080A1</td></tr><tr><td>10/5/2027</td></tr></tbody></table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A	03/03/2028	0.080	MP5089	202603B	03/03/2028	0.200	MP5090	202603D	03/04/2028	0.080 DGS	N/A	23825080A1	10/5/2027	<table border="1"><thead><tr><th>Maintenance By:</th><th>Date:</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Tank Sensor Tare</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other:</td><td></td></tr></tbody></table>		Maintenance By:	Date:	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Tank Sensor Tare		☐ Breath Tube Replacement		☐ Other:																
Simulator	Serial #	Lot#/Exp																																																	
0.050	MP5088	202603A																																																	
		03/03/2028																																																	
0.080	MP5089	202603B																																																	
		03/03/2028																																																	
0.200	MP5090	202603D																																																	
		03/04/2028																																																	
0.080 DGS	N/A	23825080A1																																																	
		10/5/2027																																																	
Maintenance By:	Date:																																																		
☐ Battery Replacement																																																			
☐ Dry Gas Regulator Replacement																																																			
☐ Tank Sensor Tare																																																			
☐ Breath Tube Replacement																																																			
☐ Other:																																																			
Optical Bench Adjustment By: <u>SLH</u>		Department Inspection By: <u>SLH</u>																																																	
Barometric Pressure Gauge: <u>1014</u> ID#: <u>28662</u>		Barometric Pressure ID#: <u>34419</u>																																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>25090</td><td>3/11/2027</td></tr><tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/2026</td></tr><tr><td>0.200</td><td>MP6297</td><td>25020</td><td>1/14/2027</td></tr><tr><td>0.300</td><td>MP6298</td><td>25150</td><td>5/6/2027</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG526603</td><td>09/23/2027</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	3/11/2027	0.100	MP6296	24390	10/29/2026	0.200	MP6297	25020	1/14/2027	0.300	MP6298	25150	5/6/2027	0.080 DGS	N/A	AG526603	09/23/2027	<table border="1"><thead><tr><th>Gauge: <u>1017</u></th><th>Instrument: <u>1016</u></th></tr></thead><tbody><tr><td>Mouth Alcohol Solution Lot #: <u>2025-D</u></td><td>Exp: <u>09/25/2027</u></td></tr><tr><td>Acetone Stock Solution Lot #: <u>2025-B</u></td><td>Exp: <u>09/22/2027</u></td></tr></tbody></table>		Gauge: <u>1017</u>	Instrument: <u>1016</u>	Mouth Alcohol Solution Lot #: <u>2025-D</u>	Exp: <u>09/25/2027</u>	Acetone Stock Solution Lot #: <u>2025-B</u>	Exp: <u>09/22/2027</u>	<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																																
0.000	MP6294	N/A	N/A																																																
0.040	MP6295	25090	3/11/2027																																																
0.100	MP6296	24390	10/29/2026																																																
0.200	MP6297	25020	1/14/2027																																																
0.300	MP6298	25150	5/6/2027																																																
0.080 DGS	N/A	AG526603	09/23/2027																																																
Gauge: <u>1017</u>	Instrument: <u>1016</u>																																																		
Mouth Alcohol Solution Lot #: <u>2025-D</u>	Exp: <u>09/25/2027</u>																																																		
Acetone Stock Solution Lot #: <u>2025-B</u>	Exp: <u>09/22/2027</u>																																																		
Simulator	Serial Number																																																		
0.000	MP5086																																																		
Interferent	MP5087																																																		
0.050	MP5088																																																		
0.080	MP5089																																																		
0.200	MP5090																																																		
☒ Post Optical Bench Adjustment Stability Checks		<table border="1"><thead><tr><th>Attachment</th><th>Post-Stability Checks</th></tr></thead><tbody><tr><td>☒ Form 41</td><td>☒ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Adjustment</td></tr><tr><td>☒ Calibration Certificate</td><td>☒ Form 40</td></tr><tr><td>☒ Optical Bench Adjustment</td><td>☐ Other:</td></tr></tbody></table>		Attachment	Post-Stability Checks	☒ Form 41	☒ Post-Stability Checks	☒ Stability Checks	☐ Flow Adjustment	☒ Calibration Certificate	☒ Form 40	☒ Optical Bench Adjustment	☐ Other:																																						
Attachment	Post-Stability Checks																																																		
☒ Form 41	☒ Post-Stability Checks																																																		
☒ Stability Checks	☐ Flow Adjustment																																																		
☒ Calibration Certificate	☒ Form 40																																																		
☒ Optical Bench Adjustment	☐ Other:																																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202603A</td><td>03/03/2028</td></tr><tr><td>0.080</td><td>MP5089</td><td>202603B</td><td>03/03/2028</td></tr><tr><td>0.200</td><td>MP5090</td><td>202603D</td><td>03/04/2028</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>23825080A1</td><td>10/5/2027</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	03/03/2028	0.080	MP5089	202603B	03/03/2028	0.200	MP5090	202603D	03/04/2028	0.080 DGS	N/A	23825080A1	10/5/2027																														
Simulator	Serial #	Lot #	Expiration																																																
0.050	MP5088	202603A	03/03/2028																																																
0.080	MP5089	202603B	03/03/2028																																																
0.200	MP5090	202603D	03/04/2028																																																
0.080 DGS	N/A	23825080A1	10/5/2027																																																
Gauge ID #: <u>34419</u> Gauge: <u>1015</u> Instrument: <u>1014</u>																																																			
Notes/Suggested Service: Stabilities performed 7/1/2026.SLH 7/1/2026 Post stabilities performed 2 separate days. SLH 7/7/2026		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																	
		<table border="1"><tr><td>Wen-Chi Pierson</td><td>Digitally signed by Wen-Chi Pierson Date: 2026.07.10 08:57:11 -04'00'</td><td>Kaitlyn Spearin</td><td>Digitally signed by Kaitlyn Spearin Date: 2026.07.14 09:50:09 -04'00'</td></tr></table>				Wen-Chi Pierson	Digitally signed by Wen-Chi Pierson Date: 2026.07.10 08:57:11 -04'00'	Kaitlyn Spearin	Digitally signed by Kaitlyn Spearin Date: 2026.07.14 09:50:09 -04'00'																																										
Wen-Chi Pierson	Digitally signed by Wen-Chi Pierson Date: 2026.07.10 08:57:11 -04'00'	Kaitlyn Spearin	Digitally signed by Kaitlyn Spearin Date: 2026.07.14 09:50:09 -04'00'																																																
		<table border="1"><tr><td>Tech Review</td><td>Admin Review</td></tr></table>				Tech Review	Admin Review																																												
Tech Review	Admin Review																																																		

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DELAND PD

Time of Inspection: 09:18

Date of Inspection: 07/01/2026

Serial Number: 80-001114

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

07/01/2026
Date

Stability Checks

80-00 1114

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L ☒	0.077 to 0.083 g/210L ☒	0.194 to 0.206 g/210L ☒	0.077 to 0.083 g/210L 50.003 g/210L of Wet ☒
DELAND PD Intoxilyzer - Alcohol Analyzer Model 8000 07/01/2026 SN 80-001114 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:25 Control Test 0.048 09:26 Air Blank 0.000 09:26 Control Test 0.049 09:27 Air Blank 0.000 09:27 Control Test 0.049 09:28 Air Blank 0.000 09:29 Control Test Status Average 0.0487 Std Dev 1.0006 Rel Std Dev(%) 1.1863 Operator's Signature <i>Michael</i>	DELAND PD Intoxilyzer - Alcohol Analyzer Model 8000 07/01/2026 SN 80-001114 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:20 Control Test 0.077 09:21 Air Blank 1.000 09:21 Control Test 0.077 09:22 Air Blank 0.000 09:22 Control Test 0.077 09:23 Air Blank 0.000 09:24 Control Test Status Average 0.0770 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>Michael</i>	DELAND PD Intoxilyzer - Alcohol Analyzer Model 8000 07/01/2026 SN 80-001114 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:33 Control Test 0.197 09:33 Air Blank 0.000 09:34 Control Test 0.157 09:35 Air Blank 0.000 09:35 Control Test 0.157 09:36 Air Blank 0.000 09:36 Control Test Status Average 0.1570 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>Michael</i>	DELAND PD Intoxilyzer - Alcohol Analyzer Model 8000 07/01/2026 SN 80-001114 Software: 8100.27 Test g/210L Time Air Blank 1.000 09:49 Control Test 1.081 09:50 Air Blank 0.000 09:50 Control Test 0.081 09:51 Air Blank 0.000 09:51 Control Test 0.081 09:52 Air Blank 0.000 09:52 Control Test Status Average 0.0810 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>Michael</i>

OBA=optical bench adjustment

needs OBA

SW

Verified instrument was able to read gas pressure.



DELAND PD

Intoxilyzer - Alcohol Analyzer

Model 8000

07/07/2026

SN 80-001114

13:38:34

SUB

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.1470	(-0.0150)	0.1430	(-0.0150)
Sample #2	0.1450	(0.0310)	0.1420	(0.0000)
Sample #3	0.1460	(0.0530)	0.1290	(0.0090)
Sample #4	0.1340	(0.0860)	0.1320	(0.0210)
Avg % Abs	0.1417	(0.0567)	0.1343	(0.0100)
STD DEV	0.0067	(0.0277)	0.0068	(0.0105)
REL STD DEV	4.700	(48.852)	5.067	(105.357)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.8620	(-0.0140)	1.5600	(-0.0040)
Sample #2	0.8610	(0.0070)	1.5560	(0.0060)
Sample #3	0.8730	(0.0190)	1.5640	(0.0100)
Sample #4	0.8770	(0.0240)	1.5580	(0.0070)
Avg % Abs	0.8703	(0.0167)	1.5593	(0.0077)
STD DEV	0.0083	(0.0087)	0.0042	(0.0021)
REL STD DEV	0.957	(52.421)	0.267	(27.152)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	1.9710	(-0.0090)	3.6420	(-0.0130)
Sample #2	1.9460	(0.0220)	3.6270	(0.0140)
Sample #3	1.9540	(0.0350)	3.6330	(0.0150)
Sample #4	1.9430	(0.0370)	3.6590	(0.0000)
Avg % Abs	1.9477	(0.0313)	3.6397	(0.0097)
STD DEV	0.0057	(0.0081)	0.0170	(0.0084)
REL STD DEV	0.292	(25.993)	0.467	(86.757)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	3.7370	(-0.0220)	6.9610	(-0.0080)
Sample #2	3.7360	(-0.0120)	6.9440	(0.0040)
Sample #3	3.7480	(-0.0090)	6.9610	(-0.0080)
Sample #4	3.7290	(-0.0050)	6.9620	(-0.0020)
Avg % Abs	3.7377	(-0.0087)	6.9557	(-0.0020)
STD DEV	0.0096	(0.0035)	0.0101	(0.0060)
REL STD DEV	0.257	(40.522)	0.145	(300.000)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.4290	(-0.0100)	10.1040	(-0.0130)
Sample #2	5.4220	(0.0180)	10.0940	(0.0110)
Sample #3	5.4050	(0.0260)	10.0910	(0.0200)
Sample #4	5.4140	(0.0230)	10.0620	(0.0260)
Avg % Abs	5.4137	(0.0223)	10.0823	(0.0190)
STD DEV	0.0085	(0.0040)	0.0177	(0.0075)
REL STD DEV	0.157	(18.096)	0.175	(39.736)

DELAND PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
07/07/2026 13:38:34

Auto Calibration

pg 2 of 2

<<<< 3um >>>>

Zero Order Coef -342.10
First Order Coef 2545.90
Second Order Coef 28.28

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0004
0.040	0.040	0.0002
0.100	0.099	0.0008
0.200	0.201	-0.0009
0.300	0.300	0.0003

<<<< 9um >>>>

-171.31
1309.09
12.34

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0001
0.040	0.040	0.0001
0.100	0.100	0.0001
0.200	0.200	-0.0002
0.300	0.300	0.0001

<<<< 3um >>>> <<<< 9um >>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1 2852.00 3250.00
Sample #2 2890.00 3262.00
Sample #3 2874.00 3267.00
Sample #4 2882.00 3246.00
Avg 2882.0000 3258.3333
STD DEV 8.0000 10.9697
REL STD DEV 0.278 0.337
H2O adjust (mg/l*10k) 927 551

Barometric Pressure = 1014

*****CALIBRATION SUCCESSFUL*****

SWA

Post Optical Bench Adjustment Stability Checks

80-001114
SWA

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L 50.003 g/210L of Wet
DELAND PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001114 07/09/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:12 Control Test 0.050 09:12 Air Blank 0.000 09:13 Control Test 0.050 09:14 Air Blank 0.000 09:14 Control Test 0.050 09:15 Air Blank 0.000 09:15 Control Test 0.050 09:15 Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>M. Sigel</i>	DELAND PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001114 07/07/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:30 Control Test 0.078 14:30 Air Blank 0.000 14:31 Control Test 0.078 14:31 Air Blank 0.000 14:32 Control Test 0.079 14:32 Air Blank 0.000 14:33 Control Test 0.000 14:33 Control Test Stats Average 0.0783 Std Dev 0.0006 Rel Std Dev(%) 0.7370 Operator's Signature <i>M. Sigel</i>	DELAND PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001114 07/09/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:05 Control Test 0.195 09:05 Air Blank 0.000 09:06 Control Test 0.196 09:07 Air Blank 0.000 09:07 Control Test 0.197 09:08 Air Blank 0.000 09:09 Control Test 0.196 09:09 Std Dev 0.0010 Rel Std Dev(%) 0.5102 Control Test Stats Average 0.1960 Std Dev 0.0010 Rel Std Dev(%) 0.5102 Operator's Signature <i>M. Sigel</i>	DELAND PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001114 07/17/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:40 Control Test 0.081 14:40 Air Blank 0.000 14:41 Control Test 0.080 14:41 Air Blank 0.000 14:42 Control Test 0.080 14:42 Air Blank 0.000 14:43 Control Test 0.080 14:43 Control Test Stats Average 0.0803 Std Dev 0.0006 Rel Std Dev(%) 0.7167 Operator's Signature <i>M. Sigel</i>

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DELAND PD
Time of Inspection: 12:05

Date of Inspection: 07/09/2026

Serial Number: 80-001114
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.078	0.196	0.081
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.078	0.195	0.080
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.078	0.196	0.080


Standard Deviations	0.0000	0.0000	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5


Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

07/09/2026
Date



ANAB
ANSI National Accreditation Board
ACCREDITED
FORENSIC CALIBRATION
LABORATORY

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001114, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001114</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>DELAND PD</u>	0.050 g/210 L 0.004
Calibration Date:	<u>07/09/2026</u>	0.080 g/210 L 0.004
Calibration Time:	<u>12:05</u>	0.200 g/210 L 0.008
		0.080 g/210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACCS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/09/2026
Date

LEANDRA HIGGINBOTHAM,
Department Inspector