

Alcohol Testing Program - Instrument Processing Sheet

Agency: Edgewater PD Instrument Serial Number: 80-001243  
 Date In: 12/23/2025 DI Completion Date: 2/12/2026  Ship  P/U  H/D  CMI  EE

Intake By: <u>WKP</u> Date: <u>12/23/2025</u>		Quality Checks By: <u>SLH</u> Date: <u>2/11/2026</u>		Flow Adjustment By: <u>SLH</u>			
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input checked="" type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: AI could not induce INT* exception message. See correspondence from 12/22/25. WKP 12/23/2025		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>184</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 103</u> 32 mm <u>0.136</u> (.139-.169) 36 mm <u>0.148</u> (.156-.190) 53 mm <u>0.226</u> (.228-.278) 103 mm <u>0.507</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28421</u> Gauge: <u>1019</u> Instrument: <u>1014</u> <input checked="" type="checkbox"/> Stability Checks		Flow Column #: <u>ATP 102</u> <input checked="" type="checkbox"/> 5L/min – 17mm <input checked="" type="checkbox"/> 15L/min – 53mm <input checked="" type="checkbox"/> 30L/min – 103mm <input checked="" type="checkbox"/> R-Value: <u>184</u> <input checked="" type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: <u>ATP 103</u> 32 mm <u>0.140</u> (.139-.169) 36 mm <u>0.156</u> (.156-.190) 53 mm <u>0.234</u> (.228-.278) 103 mm <u>0.519</u> (.447-.547)		Maintenance By: _____ Date: _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other: _____	
<b>Optical Bench Adjustment</b> By: _____		<b>Department Inspection</b> By: <u>SLH</u>					
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>34418</u>					
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1015</u>	Instrument: <u>1010</u>		
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u>	Exp: <u>09/25/2027</u>		
0.040				Acetone Stock Solution Lot #: <u>2025-B</u>	Exp: <u>09/22/2027</u>		
0.100				Simulator	Serial Number		
0.200				0.000	MP5086		
0.300				Interferent	MP6290		
0.080 DGS	N/A			0.050	MP6291		
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks				0.080	MP6292		
Simulator	Serial #	Lot #	Expiration	0.200	MP6293		
0.050				<b>Attachments</b>			
0.080				<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks		
0.200				<input checked="" type="checkbox"/> Stability Checks for acetone soln	<input checked="" type="checkbox"/> Flow Adjustment		
0.080 DGS	N/A			<input checked="" type="checkbox"/> Calibration Certificate	<input checked="" type="checkbox"/> Form 40		
Gauge ID #: _____				<input type="checkbox"/> Optical Bench Adjustment	<input checked="" type="checkbox"/> Other: INT Detect evaluation		
Gauge: _____ Instrument: _____				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC			
Notes/Suggested Service: Performed 3 replicates of acetone solution (soln) to trigger Interferent Detect and was successful. SLH 12/29/25				<input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use			
Stability checks were performed on 2/12/2026. SLH 2/12/26				<input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use			
Digitally signed by Shayla Platt Date: 2026.02.20 22:53:12 -05'00'		Digitally signed by Shayla Platt Date: 2026.02.20 22:53:12 -05'00'		Digitally signed by Shayla Platt Date: 2026.02.20 22:53:59 -05'00'			
Tech Review		Shayla Platt		Admin Review			

80-001243

SUB 12/29/25

was able to  
trigger INT  
exception.

SUB

EDGEWATER PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001243  
12/29/2025  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:54
Control Test	INT*	12:55
Air Blank	0.000	12:56
Control Test	INT*	12:56
Air Blank	0.000	12:57
Control Test	INT*	12:58
Air Blank	0.000	12:58
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

\*Interferent Detect

*Miggelott*  
Operator Signature

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EDGEWATER PD

Time of Inspection: 10:37

Date of Inspection: 02/11/2026

Serial Number: 80-001243

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: \_\_\_\_\_

**Remarks:**

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

02/11/2026  
Date

80-001243

SJA 2/11/2026

EDGEWATER PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001243  
02/11/2026  
Software: 8100.27

Flow Rate Calibration\*\*\*\*\*

1: Rate (Liters/min) = 5  
SQRT(Diff) = 7.348

2: Rate (Liters/min) = 15  
SQRT(Diff) = 11.445

3: Rate (Liters/min) = 30  
SQRT(Diff) = 20.172

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 747

Rounded Intercept = -818324

Correlation = 0.99591

# Stability Checks

Sat 2/12/26  
80-001243

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																																
<p>EDGEWATER PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001243 02/12/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:10</td></tr> <tr><td>Control Test</td><td>0.148</td><td>10:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:11</td></tr> <tr><td>Control Test</td><td>0.148</td><td>10:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:12</td></tr> <tr><td>Control Test</td><td>0.148</td><td>10:13</td></tr> <tr><td>Air Blank</td><td>0.100</td><td>10:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1450</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	10:10	Control Test	0.148	10:11	Air Blank	0.000	10:11	Control Test	0.148	10:12	Air Blank	0.000	10:12	Control Test	0.148	10:13	Air Blank	0.100	10:14	Control Test Stats			Average	0.1450		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>EDGEWATER PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001243 02/12/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:01</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:02</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:03</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:01	Control Test	0.079	12:01	Air Blank	0.000	12:02	Control Test	0.079	12:03	Air Blank	0.000	12:03	Control Test	0.079	12:04	Air Blank	0.000	12:04	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>EDGEWATER PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001243 02/12/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:56</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:57</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:59</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2906</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	10:56	Control Test	0.199	10:57	Air Blank	0.000	10:57	Control Test	0.198	10:58	Air Blank	0.000	10:59	Control Test	0.199	10:59	Air Blank	0.000	11:00	Control Test Stats			Average	0.1987		Std Dev	0.0006		Rel Std Dev(%)	0.2906		<p>DGS</p> <p>EDGEWATER PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001243 02/12/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:17</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:18</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:19</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	09:17	Control Test	0.078	09:18	Air Blank	0.000	09:18	Control Test	0.079	09:18	Air Blank	0.000	09:19	Control Test	0.078	09:19	Air Blank	0.000	09:20	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:10																																																																																																																																																	
Control Test	0.148	10:11																																																																																																																																																	
Air Blank	0.000	10:11																																																																																																																																																	
Control Test	0.148	10:12																																																																																																																																																	
Air Blank	0.000	10:12																																																																																																																																																	
Control Test	0.148	10:13																																																																																																																																																	
Air Blank	0.100	10:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1450																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:01																																																																																																																																																	
Control Test	0.079	12:01																																																																																																																																																	
Air Blank	0.000	12:02																																																																																																																																																	
Control Test	0.079	12:03																																																																																																																																																	
Air Blank	0.000	12:03																																																																																																																																																	
Control Test	0.079	12:04																																																																																																																																																	
Air Blank	0.000	12:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:56																																																																																																																																																	
Control Test	0.199	10:57																																																																																																																																																	
Air Blank	0.000	10:57																																																																																																																																																	
Control Test	0.198	10:58																																																																																																																																																	
Air Blank	0.000	10:59																																																																																																																																																	
Control Test	0.199	10:59																																																																																																																																																	
Air Blank	0.000	11:00																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1987																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2906																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:17																																																																																																																																																	
Control Test	0.078	09:18																																																																																																																																																	
Air Blank	0.000	09:18																																																																																																																																																	
Control Test	0.079	09:18																																																																																																																																																	
Air Blank	0.000	09:19																																																																																																																																																	
Control Test	0.078	09:19																																																																																																																																																	
Air Blank	0.000	09:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
<p><i>M. S. G. S.</i> Operator's Signature</p>	<p><i>M. S. G. S.</i> Operator's Signature</p>	<p><i>M. S. G. S.</i> Operator's Signature</p>	<p><i>M. S. G. S.</i> Operator's Signature</p>																																																																																																																																																

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: EDGEWATER PD  
Time of Inspection: 18:23

Date of Inspection: 02/12/2026

Serial Number: 80-001243  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000 / 0.000	0.048	0.078	0.197	0.079
0.000 / 0.000	0.048	0.078	0.198	0.079
0.000 / 0.000	0.048	0.078	0.198	0.079
0.000 / 0.000	0.048	0.079	0.198	0.079
0.000 / 0.000	0.048	0.078	0.198	0.079
0.000 / 0.000	0.048	0.079	0.198	0.079
0.000 / 0.000	0.048	0.079	0.198	0.079
0.000 / 0.000	0.048	0.078	0.198	0.079
0.000 / 0.000	0.048	0.079	0.198	0.079
INT / 0.000	0.048	0.079	0.199	0.080

Standard Deviations	0.0000	0.0005	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5


**Remarks:**

00: Interferent Detect. CLEARED AREA WITH FAN



SUB

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 LEANDRA HIGGINBOTHAM  
 Signature and Printed Name

02/12/2026  
 Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001243, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001243</u>	UNCERTAINTY* ±	
Owning Agency:	<u>EDGEWATER PD</u>	0.050 g/210 L	0.004
Calibration Date:	<u>02/12/2026</u>	0.080 g/210 L	0.004
Calibration Time:	<u>18:23</u>	0.200 g/210 L	0.008
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3)

The instrument results before and after any adjustment are found in the associated pre and post stability checks

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS), ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 January 2026  
Issuing Authority: Alcohol Testing Program

**LEANDRA HUGGINBOTHAM,**  
Department Inspector

02/12/2026

Date

Service • Integrity • Respect • Quality