

Alcohol Testing Program - Instrument Processing Sheet

Agency: HILLSBOROUGH CO SO

Instrument Serial Number: 80-007158

Date In: 7/31/2026

DI Completion Date: N/A

☒ Ship☐ P/U☐ H/D☒ CMI☐ EE

Intake By: <u>KTS</u> Date: <u>8/3/2026</u>		Quality Checks By: <u>KTS</u> Date: <u>8/6/26</u>		Flow Adjustment By: _____																																					
☒ Annual ☐ Dropped Off ☒ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>178</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.167</u> (.139-.169) 36 mm <u>0.179</u> (.156-.190) 53 mm <u>0.250</u> (.228-.278) 103 mm <u>0.507</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1015</u> Instrument: <u>1017</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																					
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot#/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202603A 3/3/28</td></tr><tr><td>0.080</td><td>MP5089</td><td>202603B 3/3/28</td></tr><tr><td>0.200</td><td>MP5090</td><td>202603D 3/4/28</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>23825080A1 10/5/27</td></tr></tbody></table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 3/3/28	0.080	MP5089	202603B 3/3/28	0.200	MP5090	202603D 3/4/28	0.080 DGS	N/A	23825080A1 10/5/27	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:																						
Simulator	Serial #	Lot#/Exp																																							
0.050	MP5088	202603A 3/3/28																																							
0.080	MP5089	202603B 3/3/28																																							
0.200	MP5090	202603D 3/4/28																																							
0.080 DGS	N/A	23825080A1 10/5/27																																							
Optical Bench Adjustment By: <u>IAS</u>		Department Inspection By: <u>KTS</u>																																							
Barometric Pressure Gauge: <u>1018</u> ID#: <u>34420</u>		Barometric Pressure ID#: <u>28427</u>																																							
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>25090</td><td>3/11/27</td></tr><tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/26</td></tr><tr><td>0.200</td><td>MP6297</td><td>25020</td><td>1/14/27</td></tr><tr><td>0.300</td><td>MP6298</td><td>24430</td><td>12/10/26</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG526603</td><td>9/23/27</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	3/11/27	0.100	MP6296	24390	10/29/26	0.200	MP6297	25020	1/14/27	0.300	MP6298	24430	12/10/26	0.080 DGS	N/A	AG526603	9/23/27	Gauge: <u>1015</u> Instrument: <u>1016</u> Mouth Alcohol Solution Lot #: _____ Exp: _____ Acetone Stock Solution Lot #: _____ Exp: _____											
Simulator	Serial #	Lot #	Expiration																																						
0.000	MP6294	N/A	N/A																																						
0.040	MP6295	25090	3/11/27																																						
0.100	MP6296	24390	10/29/26																																						
0.200	MP6297	25020	1/14/27																																						
0.300	MP6298	24430	12/10/26																																						
0.080 DGS	N/A	AG526603	9/23/27																																						
☒ Post Optical Bench Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202603A</td><td>3/3/28</td></tr><tr><td>0.080</td><td>MP5089</td><td>202603B</td><td>3/3/28</td></tr><tr><td>0.200</td><td>MP5090</td><td>202603D</td><td>3/4/28</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>23725080A1</td><td>10/5/27</td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	3/3/28	0.080	MP5089	202603B	3/3/28	0.200	MP5090	202603D	3/4/28	0.080 DGS	N/A	23725080A1	10/5/27	<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></tbody></table>				Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200	
Simulator	Serial #	Lot #	Expiration																																						
0.050	MP5088	202603A	3/3/28																																						
0.080	MP5089	202603B	3/3/28																																						
0.200	MP5090	202603D	3/4/28																																						
0.080 DGS	N/A	23725080A1	10/5/27																																						
Simulator	Serial Number																																								
0.000																																									
Interferent																																									
0.050																																									
0.080																																									
0.200																																									
Gauge ID #: <u>28427</u> Gauge: <u>1014</u> Instrument: <u>1016</u>		Attachments ☒ Form 41 ☒ Stability Checks ☐ Calibration Certificate ☒ Optical Bench Adjustment ☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☒ Other: Return Material Authorization																																							
Notes/Suggested Service: Tech review: added barometric pressure. KTS 8/14/26		☐ Instrument Complies with Chapter 11D-8, FAC ☒ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use S. LeAndra Higginbotham Date: 2026.08.14 15:07:37 -04'00'																																							
		Tech Review		Admin Review																																					

Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																																																
0.047 to 0.053 g/210L ☒	0.077 to 0.083 g/210L ☒	0.194 to 0.206 g/210L ☒	0.077 to 0.083 g/210L ☒ ≤ 0.003 g/210L of Wet ☒																																																																																																																																																
HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 08/06/2026 Software: 8100.27 SN 80-007158	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 08/06/2026 Software: 8100.27 SN 80-007158	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 08/06/2026 Software: 8100.27 SN 80-007158	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 08/06/2026 Software: 8100.27 SN 80-007158																																																																																																																																																
<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:54</td></tr> <tr><td>Control Test</td><td>0.054</td><td>06:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:55</td></tr> <tr><td>Control Test</td><td>0.054</td><td>06:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:56</td></tr> <tr><td>Control Test</td><td>0.053</td><td>06:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:58</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0537</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.0758</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:54	Control Test	0.054	06:55	Air Blank	0.000	06:55	Control Test	0.054	06:56	Air Blank	0.000	06:56	Control Test	0.053	06:57	Air Blank	0.000	06:58	Control Test Stats			Average	0.0537		Std Dev	0.0006		Rel Std Dev(%)	1.0758		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:45</td></tr> <tr><td>Control Test</td><td>0.081</td><td>06:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:46</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:47</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:48</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:45	Control Test	0.081	06:46	Air Blank	0.000	06:46	Control Test	0.080	06:46	Air Blank	0.000	06:47	Control Test	0.080	06:47	Air Blank	0.000	06:48	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:04</td></tr> <tr><td>Control Test</td><td>0.205</td><td>07:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:05</td></tr> <tr><td>Control Test</td><td>0.203</td><td>07:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:06</td></tr> <tr><td>Control Test</td><td>0.203</td><td>07:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2037</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5670</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	07:04	Control Test	0.205	07:04	Air Blank	0.000	07:05	Control Test	0.203	07:06	Air Blank	0.000	07:06	Control Test	0.203	07:07	Air Blank	0.000	07:08	Control Test Stats			Average	0.2037		Std Dev	0.0012		Rel Std Dev(%)	0.5670		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:59</td></tr> <tr><td>Control Test</td><td>0.082</td><td>06:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:00</td></tr> <tr><td>Control Test</td><td>0.082</td><td>07:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:01</td></tr> <tr><td>Control Test</td><td>0.082</td><td>07:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0820</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	06:59	Control Test	0.082	06:59	Air Blank	0.000	07:00	Control Test	0.082	07:01	Air Blank	0.000	07:01	Control Test	0.082	07:02	Air Blank	0.000	07:02	Control Test Stats			Average	0.0820		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:54																																																																																																																																																	
Control Test	0.054	06:55																																																																																																																																																	
Air Blank	0.000	06:55																																																																																																																																																	
Control Test	0.054	06:56																																																																																																																																																	
Air Blank	0.000	06:56																																																																																																																																																	
Control Test	0.053	06:57																																																																																																																																																	
Air Blank	0.000	06:58																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0537																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.0758																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:45																																																																																																																																																	
Control Test	0.081	06:46																																																																																																																																																	
Air Blank	0.000	06:46																																																																																																																																																	
Control Test	0.080	06:46																																																																																																																																																	
Air Blank	0.000	06:47																																																																																																																																																	
Control Test	0.080	06:47																																																																																																																																																	
Air Blank	0.000	06:48																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:04																																																																																																																																																	
Control Test	0.205	07:04																																																																																																																																																	
Air Blank	0.000	07:05																																																																																																																																																	
Control Test	0.203	07:06																																																																																																																																																	
Air Blank	0.000	07:06																																																																																																																																																	
Control Test	0.203	07:07																																																																																																																																																	
Air Blank	0.000	07:08																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2037																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5670																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:59																																																																																																																																																	
Control Test	0.082	06:59																																																																																																																																																	
Air Blank	0.000	07:00																																																																																																																																																	
Control Test	0.082	07:01																																																																																																																																																	
Air Blank	0.000	07:01																																																																																																																																																	
Control Test	0.082	07:02																																																																																																																																																	
Air Blank	0.000	07:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0820																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
 Operator's Signature	 Operator's Signature	 Operator's Signature	 Operator's Signature																																																																																																																																																

HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-007158
 08/06/2026 08:55:10

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>>			<<<<< 9um >>>>>		
Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.1410	(-0.0050)	0.2090	(-0.0190)	
Sample #2	0.1560	(-0.0040)	0.1830	(0.0050)	
Sample #3	0.1390	(0.0130)	0.2090	(-0.0020)	
Sample #4	0.1450	(0.0150)	0.2090	(0.0060)	
Avg % Abs	0.1467	(0.0080)	0.2003	(0.0030)	
STD DEV	0.0086	(0.0104)	0.0150	(0.0044)	
REL STD DEV	5.878	(130.504)	7.493	(145.297)	
Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.9020	(-0.0090)	1.5200	(-0.0110)	
Sample #2	0.8980	(0.0050)	1.4910	(0.0210)	
Sample #3	0.8870	(0.0140)	1.5010	(0.0060)	
Sample #4	0.8940	(0.0150)	1.5060	(0.0240)	
Avg % Abs	0.8930	(0.0113)	1.4993	(0.0170)	
STD DEV	0.0056	(0.0055)	0.0076	(0.0096)	
REL STD DEV	0.623	(48.596)	0.509	(56.727)	
Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	2.0050	(-0.0070)	3.4620	(-0.0160)	
Sample #2	1.9870	(0.0060)	3.3720	(0.0250)	
Sample #3	1.9800	(0.0130)	3.3920	(0.0260)	
Sample #4	1.9930	(0.0170)	3.3770	(0.0310)	
Avg % Abs	1.9867	(0.0120)	3.3803	(0.0273)	
STD DEV	0.0065	(0.0056)	0.0104	(0.0032)	
REL STD DEV	0.328	(46.398)	0.308	(11.761)	
Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	3.8210	(-0.0040)	6.5210	(0.0000)	
Sample #2	3.7820	(0.0110)	6.4060	(0.0290)	
Sample #3	3.7740	(0.0110)	6.4020	(0.0360)	
Sample #4	3.7830	(0.0150)	6.3600	(0.0450)	
Avg % Abs	3.7797	(0.0123)	6.3893	(0.0367)	
STD DEV	0.0049	(0.0023)	0.0255	(0.0080)	
REL STD DEV	0.131	(18.725)	0.399	(21.875)	
Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	5.6000	(-0.0170)	9.4380	(0.0100)	
Sample #2	5.5070	(0.0120)	9.2800	(0.0450)	
Sample #3	5.4960	(0.0150)	9.2450	(0.0630)	
Sample #4	5.5330	(0.0150)	9.2860	(0.0570)	
Avg % Abs	5.5120	(0.0140)	9.2703	(0.0550)	
STD DEV	0.0190	(0.0017)	0.0221	(0.0092)	
REL STD DEV	0.345	(12.372)	0.239	(16.664)	

IAS

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007158
08/06/2026 08:55:10

Auto Calibration

pg 2 of 2

<<<< 3um >>>>

Zero Order Coef -372.57
First Order Coef 2535.37
Second Order Coef 22.41

<<<< 9um >>>>

-296.95
1453.55
12.91

Act Fit Residual
(g/210L) (g/210L) (g/210L)
0.000 -0.000 0.0000
0.040 0.040 -0.0001
0.100 0.100 0.0002
0.200 0.200 -0.0001
0.300 0.300 0.0001

Act Fit Residual
(g/210L) (g/210L) (g/210L)
0.000 -0.000 0.0001
0.040 0.040 -0.0001
0.100 0.100 -0.0000
0.200 0.200 0.0001
0.300 0.300 -0.0000

<<<< 3um >>>> <<<< 9um >>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1 3082.00 3193.00
Sample #2 3057.00 3181.00
Sample #3 3064.00 3179.00
Sample #4 3039.00 3164.00
Avg 3053.3333 3174.6667
STD DEV 12.8970 9.2916
REL STD DEV 0.422 0.293
H2O adjust (mg/l*10k) 756 635

Barometric Pressure = 1018

*****CALIBRATION SUCCESSFUL***** IAS

Post-Optical

80-007158

Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																																																
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 08/06/2026 Software: 8100.27 SN 80-007158 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:02</td></tr><tr><td>Control Test</td><td>0.053</td><td>12:03</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:04</td></tr><tr><td>Control Test</td><td>0.051</td><td>12:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:05</td></tr><tr><td>Control Test</td><td>0.051</td><td>12:06</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:06</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0517</td><td></td></tr><tr><td>Std Dev</td><td>0.0012</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>2.2349</td><td></td></tr></table> Operator's Signature <i>Isabella Sumpfauer</i>	Test	g/210L	Time	Air Blank	0.000	12:02	Control Test	0.053	12:03	Air Blank	0.000	12:04	Control Test	0.051	12:04	Air Blank	0.000	12:05	Control Test	0.051	12:06	Air Blank	0.000	12:06	Control Test Stats			Average	0.0517		Std Dev	0.0012		Rel Std Dev(%)	2.2349		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 08/06/2026 Software: 8100.27 SN 80-007158 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:20</td></tr><tr><td>Control Test</td><td>0.081</td><td>12:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:22</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:24</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2500</td><td></td></tr></table> Operator's Signature <i>Isabella Sumpfauer</i>	Test	g/210L	Time	Air Blank	0.000	12:20	Control Test	0.081	12:21	Air Blank	0.000	12:21	Control Test	0.080	12:22	Air Blank	0.000	12:22	Control Test	0.079	12:23	Air Blank	0.000	12:24	Control Test Stats			Average	0.0800		Std Dev	0.0010		Rel Std Dev(%)	1.2500		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 08/06/2026 Software: 8100.27 SN 80-007158 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:35</td></tr><tr><td>Control Test</td><td>0.201</td><td>12:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:36</td></tr><tr><td>Control Test</td><td>0.199</td><td>12:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:38</td></tr><tr><td>Control Test</td><td>0.198</td><td>12:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:39</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1993</td><td></td></tr><tr><td>Std Dev</td><td>0.0015</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7663</td><td></td></tr></table> Operator's Signature <i>Isabella Sumpfauer</i>	Test	g/210L	Time	Air Blank	0.000	12:35	Control Test	0.201	12:36	Air Blank	0.000	12:36	Control Test	0.199	12:37	Air Blank	0.000	12:38	Control Test	0.198	12:38	Air Blank	0.000	12:39	Control Test Stats			Average	0.1993		Std Dev	0.0015		Rel Std Dev(%)	0.7663		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 08/16/2026 Software: 8100.27 SN 80-007158 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:56</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:57</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:58</td></tr><tr><td>Control Test</td><td>0.078</td><td>12:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1787</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></table> Operator's Signature <i>Isabella Sumpfauer</i>	Test	g/210L	Time	Air Blank	0.000	12:56	Control Test	0.079	12:56	Air Blank	0.000	12:57	Control Test	0.079	12:57	Air Blank	0.000	12:58	Control Test	0.078	12:58	Air Blank	0.000	12:59	Control Test Stats			Average	0.1787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:02																																																																																																																																																	
Control Test	0.053	12:03																																																																																																																																																	
Air Blank	0.000	12:04																																																																																																																																																	
Control Test	0.051	12:04																																																																																																																																																	
Air Blank	0.000	12:05																																																																																																																																																	
Control Test	0.051	12:06																																																																																																																																																	
Air Blank	0.000	12:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0517																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	2.2349																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:20																																																																																																																																																	
Control Test	0.081	12:21																																																																																																																																																	
Air Blank	0.000	12:21																																																																																																																																																	
Control Test	0.080	12:22																																																																																																																																																	
Air Blank	0.000	12:22																																																																																																																																																	
Control Test	0.079	12:23																																																																																																																																																	
Air Blank	0.000	12:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2500																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:35																																																																																																																																																	
Control Test	0.201	12:36																																																																																																																																																	
Air Blank	0.000	12:36																																																																																																																																																	
Control Test	0.199	12:37																																																																																																																																																	
Air Blank	0.000	12:38																																																																																																																																																	
Control Test	0.198	12:38																																																																																																																																																	
Air Blank	0.000	12:39																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1993																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	0.7663																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:56																																																																																																																																																	
Control Test	0.079	12:56																																																																																																																																																	
Air Blank	0.000	12:57																																																																																																																																																	
Control Test	0.079	12:57																																																																																																																																																	
Air Blank	0.000	12:58																																																																																																																																																	
Control Test	0.078	12:58																																																																																																																																																	
Air Blank	0.000	12:59																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 06:42

Date of Inspection: 08/10/2026

Serial Number: 80-007158
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK		No	Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK		No	Barometric Pressure Sensor Check: OK		No
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

[illegible]

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:

Non-compliance: .

Instrument produced an Interferent Detect exception message for the alcohol-free breath sample. Repeated analysis and got an alcohol measurement of 0.007g/210L. Did not repeat analysis a third time- sending instrument to CMI. KTS 8/11/26

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

KATIE T SPEARIN

Signature and Printed Name

08/10/2026
Date

Return Material Authorization

Ship to:☒ CMI, Inc.☐ Enforcement ElectronicsShipment to repair facility authorized by: Jenny Hammond on 8/11/2026Items Returned: ☒ Instrument ☐ Supplies ☐ Other Describe: _____Instrument Model: Intoxilyzer 8000 Serial Number: 80-007158

Bill To Address:

Kevin Cole

Hillsborough County SO

1201 Orient Road, Tampa, FL, 33619

Ship to Address:

FL Dept of Law Enforcement

813 B Lake Bradford Rd, Tallahassee, FL 32304

Reason for Return:

Instrument produced interferent detect message during the alcohol-free breath sample and when repeated, measured a 0.007g/210L alcohol measurement.

I require an estimate **BEFORE** any repairs will be authorized and/or conducted

Please contact: Kevin Brown Phone #: 863-247-8360Email: kcbrown@teamhcsso.comATP Contact Name: Katie SpearinATP Email: KaitlynSpearin@fdle.state.fl.us