Alcohol Testing Program - Instrument Processing Sheet

Agency: SANTA ROSA COUNTY SO **Instrument Serial Number:** 80-001302

Date In: 7/21/2026 **DI Completion Date:** 7/23/2026 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: KTS Date: 7/21/2026

☒ Annual
☐ Dropped Off
☐ Registration
☒ Return from CMI / EE
☐ Training Instrument

Visual Inspection

☒ Case ☒ Handle
☒ Keyboard ☒ Dry Gas Shelf
☒ Feet ☒ Breath Tube
☒ Ports ☒ Screws Tight

Other Equipment/Accessories

☒ Power Cord ☐ Printer Cable
☐ Static Bag ☐ 12V DC Cable

Notes:

Quality Checks By: WKP Date: 7/22/2026

☒ Breath Tube Screen
☒ Replace External O-Rings
☒ Instrument Set Up Verified
☒ R-Value: 206
☒ Flow Verification (L/s)
Flow Column #: ATP 105
32 mm 0.148 (.139-.169)
36 mm 0.164 (.156-.190)
53 mm 0.230 (.228-.278)
103 mm 0.503 (.447-.547)
☒ Barometric Pressure Check
Gauge ID #: 34419
Gauge: 1012 Instrument: 1013
☒ Stability Checks

Simulator	Serial #	Lot#/Exp
0.050	MP6291	202603A
		3/3/2028
0.080	MP6292	202603B
		3/3/2028
0.200	MP6293	202603D
		3/4/2028
0.080 DGS	N/A	23825080A1
		10/5/2027

Flow Adjustment By:

Flow Column #:

☐ 5L/min – 17mm
☐ 15L/min – 53mm
☐ 30L/min – 103mm

☐ R-Value:
☐ Post Adjustment Verification (L/S)
Flow Column #:
32 mm (.139-.169)
36 mm (.156-.190)
53 mm (.228-.278)
103 mm (.447-.547)

Maintenance By: Date:

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Tank Sensor Tare
☐ Breath Tube Replacement
☐ Other:

Optical Bench Adjustment By:

Barometric Pressure Gauge: ID#:

Simulator	Serial #	Lot #	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Optical Bench Adjustment Stability Checks

Simulator	Serial #	Lot #	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Gauge ID #:
Gauge: Instrument:

Notes/Suggested Service:

Department Inspection By: WKP

Barometric Pressure ID#: 28427
Gauge: 1014 Instrument: 1018
Mouth Alcohol Solution Lot #: 2025-D Exp: 9/25/2027
Acetone Stock Solution Lot #: 2025-B Exp: 9/22/2027

Simulator	Serial Number
0.000	MP6289
Interferent	MP6290
0.050	MP6291
0.080	MP6292
0.200	MP6293

Attachments

☒ Form 41
☒ Stability Checks
☒ Calibration Certificate
☐ Optical Bench Adjustment
☐ Post-Stability Checks
☐ Flow Adjustment
☐ Form 40
☐ Other:

☒ **Instrument Complies with Chapter 11D-8, FAC**
☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**

☒ **Return to/Place into Evidentiary Use**
☐ **Remain Out of Evidentiary Use**

☒ **Conduct an Agency Inspection Before Evidentiary Use**

Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.07.27 15:03:10 -04'00' — **Tech Review**

LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.07.31 09:17:13 -04'00' — **Admin Review**

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ✓	*0.077 to 0.083* ✓	*0.194 to 0.206* ✓	*0.077 to 0.083* ✓ *≤0.003 of Wet* ✓

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001302
07/22/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:47
Control Test	0.050	12:48
Air Blank	0.000	12:49
Control Test	0.050	12:49
Air Blank	0.000	12:50
Control Test	0.049	12:51
Air Blank	0.000	12:51

Control Test Stats

Average	0.0497
Std Dev	0.0006
Rel Std Dev(%)	1.1625

Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001302
07/22/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:52
Control Test	0.078	12:53
Air Blank	0.000	12:53
Control Test	0.077	12:54
Air Blank	0.000	12:55
Control Test	0.078	12:55
Air Blank	0.000	12:56

Control Test Stats

Average	0.0777
Std Dev	0.0006
Rel Std Dev(%)	0.7434

Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001302
07/22/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:04
Control Test	0.197	13:05
Air Blank	0.000	13:05
Control Test	0.199	13:06
Air Blank	0.000	13:07
Control Test	0.199	13:07
Air Blank	0.000	13:08

Control Test Stats

Average	0.1983
Std Dev	0.0012
Rel Std Dev(%)	0.5822

Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001302
07/22/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:10
Control Test	0.079	13:10
Air Blank	0.000	13:10
Control Test	0.079	13:11
Air Blank	0.000	13:11
Control Test	0.080	13:12
Air Blank	0.000	13:12

Control Test Stats

Average	0.0793
Std Dev	0.0006
Rel Std Dev(%)	0.7277

Operator's Signature

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SANTA ROSA COUNTY SO Serial Number: 80-001302
Time of Inspection:12:10 Date of Inspection:07/23/2026 Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#: 202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 23825080A1 Exp: 10/05/2027
0.000	0.049	0.078	0.202	0.080
0.000	0.049	0.078	0.202	0.080
0.000	0.049	0.078	0.202	0.079
0.000	0.049	0.079	0.202	0.080
0.000	0.049	0.079	0.202	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.049	0.079	0.202	0.080
0.000	0.050	0.078	0.202	0.080
0.000	0.049	0.078	0.201	0.080
0.000	0.050	0.079	0.202	0.080
Standard Deviations	0.0004	0.0005	0.0003	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:
Int Det: RFI Detect. Passed by instrument with a cell phone in proximity. Removed the cell phone from the area and repeated. WKP 7/23/2026

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Wen-Chi Pierson
WEN-CHI K PIERSON
Signature and Printed Name

07/23/2026
Date

FDLE/ATP Form 41 -- Revised August 2005

Reprinted From Database

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001302 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001302
Owning Agency:	SANTA ROSA COUNTY SO
Calibration Date:	07/23/2026
Calibration Time:	12:10

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.008
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full,
without written approval of the Florida Department of
Law Enforcement Alcohol Testing Program.

07/23/2026
Date

Wen-Chi Pierson
Digitally signed by Wen-Chi Pierson
Date: 2026.07.23 14:43:54 -04'00'

WEN-CHI K PIERSON,
Department Inspector

Alcohol Testing Program - Instrument Processing Sheet

Agency: SANTA ROSA COUNTY SO **Instrument Serial Number:** 80-001302

Date In: 6/9/2026 **DI Completion Date:** N/A ☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake By: SLH Date: 6/9/2026

☒ Annual
☐ Dropped Off
☐ Registration
☐ Return from CMI / EE
☐ Training Instrument

Visual Inspection

☒ Case ☒ Handle
☒ Keyboard ☒ Dry Gas Shelf
☒ Feet ☒ Breath Tube
☒ Ports ☒ Screws Tight

Other Equipment/Accessories

☐ Power Cord ☐ Printer Cable
☒ Static Bag ☐ 12V DC Cable

Notes:

Quality Checks By: IAS Date: 6/11/2026

☒ Breath Tube Screen
☒ Replace External O-Rings
☒ Instrument Set Up Verified
☒ R-Value: 111
☒ Flow Verification (L/s)
Flow Column #: ATP102

32 mm	.152	(.139-.169)
36 mm	.160	(.156-.190)
53 mm	.242	(.228-.278)
103 mm	.523	(.447-.547)

☒ Barometric Pressure Check
Gauge ID #: 28427
Gauge: 1011 Instrument: 1012
☒ Stability Checks

Simulator	Serial #	Lot#/Exp
0.050	MP5088	202603A
		3/3/2028
0.080	MP6292	202603B
		3/3/2028
0.200		
0.080 DGS	N/A	23825080A1
		10/5/2027

Flow Adjustment By:

Flow Column #:

☐ 5L/min – 17mm
☐ 15L/min – 53mm
☐ 30L/min – 103mm

☐ R-Value:
☐ Post Adjustment Verification (L/S)
Flow Column #:

32 mm		(.139-.169)
36 mm		(.156-.190)
53 mm		(.228-.278)
103 mm		(.447-.547)

Maintenance By: Date:

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Tank Sensor Tare
☐ Breath Tube Replacement
☐ Other:

Optical Bench Adjustment By:

Barometric Pressure Gauge: ID#:

Simulator	Serial #	Lot #	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Optical Bench Adjustment Stability Checks

Simulator	Serial #	Lot #	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Gauge ID #:
Gauge: Instrument:

Department Inspection By:

Barometric Pressure ID#:
Gauge: Instrument:
Mouth Alcohol Solution Lot #: Exp:
Acetone Stock Solution Lot #: Exp:

Simulator	Serial Number
0.000	
Interferent	
0.050	
0.080	
0.200	

Attachments

☐ Form 41
☒ Stability Checks
☐ Calibration Certificate
☐ Optical Bench Adjustment
☐ Post-Stability Checks
☐ Flow Adjustment
☐ Form 40
☒ Other: Return Material Authorization

Notes/Suggested Service:

☐ **Instrument Complies with Chapter 11D-8, FAC**
☒ **Instrument Does Not Comply with Chapter 11D-8, FAC**

☐ **Return to/Place into Evidentiary Use**
☒ **Remain Out of Evidentiary Use**

☐ **Conduct an Agency Inspection Before Evidentiary Use**

LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.06.18 18:29:05 -04'00'
Tech Review

Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.06.19 05:55:26 -04'00'
Admin Review

Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L [X]	*0.077 to 0.083 g/210L* [X]	*0.194 to 0.206 g/210L* [ ]	*0.077 to 0.083 g/210L* [✓] *≤0.003 g/210L of Wet* [X]

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001302
06/11/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:57
Control Test	0.080	11:57
Air Blank	0.000	11:58
Control Test	0.080	11:58
Air Blank	0.000	11:59
Control Test	0.080	12:00
Air Blank	0.000	12:00

Control Test Stats

Average	0.0800
Std Dev	0.0000
Rel Std Dev(%)	0.0000

Isabella Szymkowski
Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001302
06/11/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:19
Control Test	0.080	12:20
Air Blank	0.000	12:20
Control Test	0.080	12:21
Air Blank	0.000	12:21
Control Test	0.080	12:22
Air Blank	0.000	12:22

Control Test Stats

Average	0.0800
Std Dev	0.0000
Rel Std Dev(%)	0.0000

Isabella Szymkowski
Operator's Signature

Did not run due to instrument needing repair. IAS 6/12/26

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001302
06/11/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:26
Control Test	0.079	11:26
Air Blank	0.000	11:27
Control Test	0.078	11:27
Air Blank	0.000	11:27
Control Test	0.079	11:28
Air Blank	0.001	11:28

Control Test Stats

Average	0.0787
Std Dev	0.0016
Rel Std Dev(%)	0.7339

Isabella Szymkowski
Operator's Signature

#2 run of .05g/210L
Wet Stability

Ran another Stability check to confirm the correct sample type was selected. IAS 6/12/26

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001302
06/11/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:06
Control Test	0.000	12:07
Air Blank	0.000	12:07
Control Test	RFI*	12:08
Air Blank	0.000	12:08

*RFI Detect

Isabella Swyzkowski
Operator's Signature

#3 run of .05g/210L
Wet Stability

Ran another stability check due to RFI exception message in Run #2. IAS 6/12/26

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001302
06/11/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:09
Control Test	0.000	12:10
Air Blank	0.000	12:10
Control Test	0.000	12:11
Air Blank	0.000	12:11
Control Test	0.000	12:12
Air Blank	0.000	12:12

Control Test Stats

Average	0.0000
Std Dev	0.0000
Rel Std Dev(%)	0.0000

Isabella Swyzkowski
Operator's Signature

Return Material Authorization

Ship to: ☒ CMI, Inc. ☐ Enforcement Electronics

Shipment to repair facility authorized by: Duron Nelson on 6/15/2026

Items Returned: ☒ Instrument ☐ Supplies ☐ Other Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001302

Bill To Address:	Ship to Address:
Santa Rosa County Sheriff Office ATTN: Sgt Raymond Spencer 5755 East Milton Rd Milton, FL 32565	FDLE Offsite Mail Facility c/o FDLE HQ 813 B Lake Bradford Rd Tallahassee, FL 32304

Reason for Return:

The instrument was reading the alcohol reference solution concentration as zero during the stability checks and intermittently flagged radio frequency interference.

I require an estimate **BEFORE** any repairs will be authorized and/or conducted

Please contact: Raymond Spencer Phone #: 8504547668

Email: rspencer@srso.net

ATP Contact Name: Wen-Chi Pierson ATP Email: Wen-ChiPierson@fdle.state.fl.us