

Alcohol Testing Program - Instrument Processing Sheet

Agency: INDIALANTIC PD

Instrument Serial Number: 80-001006

Date In: 6/8/2026

DI Completion Date: 6/12/2026

☒ Ship☐ P/U☐ H/D☐ CMI☐ EE

Intake By: KTS Date: 6/8/2026		Quality Checks By: IAS Date: 6/11/2026		Flow Adjustment By:	
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: 181 ☒ Flow Verification (L/s) Flow Column #: ATP102 32 mm .164 (.139-.169) 36 mm .175 (.156-.190) 53 mm .246 (.228-.278) 103 mm .515 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: 28427 Gauge: 1011 Instrument: 1011 ☒ Stability Checks		Flow Column #: ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: ☐ Post Adjustment Verification (L/S) Flow Column #: 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)	
		Simulator		Serial #	Lot#/Exp
		0.050		MP6291	202603A 3/3/2028
		0.080		MP5089	202603B 3/3/2028
		0.200		MP5090	202603D 3/4/2028
		0.080 DGS		N/A	23825080A1 10/5/2027
Optical Bench Adjustment By:		Department Inspection		By: SLH	
Barometric Pressure Gauge: ID#:		Barometric Pressure ID#: 28427			
Simulator	Serial #	Lot #	Expiration	Gauge: 1010	Instrument: 1011
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: 2025-D	Exp: 09/25/2027
0.040				Acetone Stock Solution Lot #: 2025-B	Exp: 09/22/2027
0.100				Simulator	Serial Number
0.200				0.000	MP6294
0.300				Interferent	MP6290
0.080 DGS	N/A			0.050	MP6291
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP5089
Simulator	Serial #	Lot #	Expiration	0.200	MP5090
0.050				Attachments	
0.080				☒ Form 41	
0.200				☒ Stability Checks	
0.080 DGS	N/A			☒ Calibration Certificate	
Gauge ID #: _____ Gauge: _____ Instrument: _____				☐ Optical Bench Adjustment	
Notes/Suggested Service: Agency indicated alcohol free controls had alcohol results and interferent detect results which were not able to be replicated. SLH 6/12/26				☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other:	
				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use	
Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.06.19 14:27:12 -04'00'				Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.06.22 09:53:01 -04'00'	
Tech Review				Admin Review	

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIALANTIC PD
Time of Inspection: 10:48

Date of Inspection: 06/11/2026

Serial Number: 80-001006
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____


Number of Simulators Used: _____

Remarks:

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISABELLA A SMYKOWSKI

Signature and Printed Name

06/11/2026
Date

Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L 50.003 g/210L of Wet
INDUSTRAL PD Intoxilyzer - Alcohol Analyzer Model 8000 16/11/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:54 Control Test 0.052 12:54 Air Blank 0.000 12:55 Control Test 0.049 12:56 Air Blank 0.000 12:56 Control Test 0.049 12:57 Air Blank 0.000 12:57 Control Test Stats Average 0.0500 Std Dev 0.0017 Rel Std Dev(%) 3.4541	INDUSTRAL PD Intoxilyzer - Alcohol Analyzer Model 8000 16/11/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 13:49 Control Test 0.079 13:49 Air Blank 0.000 13:50 Control Test 0.078 13:51 Air Blank 0.000 13:51 Control Test 0.078 13:52 Air Blank 0.000 13:53 Control Test Stats Average 0.0783 Std Dev 0.0006 Rel Std Dev(%) 0.7370	INDUSTRAL PD Intoxilyzer - Alcohol Analyzer Model 8000 16/11/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 13:54 Control Test 0.198 13:55 Air Blank 0.000 13:56 Control Test 0.197 13:56 Air Blank 0.000 13:57 Control Test 0.197 13:58 Air Blank 0.000 13:58 Control Test Stats Average 0.1973 Std Dev 0.0006 Rel Std Dev(%) 0.2926	INDUSTRAL PD Intoxilyzer - Alcohol Analyzer Model 8000 16/11/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:46 Control Test 0.081 11:46 Air Blank 0.000 11:46 Control Test 0.081 11:47 Air Blank 0.000 11:47 Control Test 0.081 11:48 Air Blank 0.000 11:48 Control Test Stats Average 0.0810 Std Dev 0.0000 Rel Std Dev(%) 0.0000
<i>Isabelle Szymanski</i> Operator's Signature	<i>Isabelle Szymanski</i> Operator's Signature	<i>Isabelle Szymanski</i> Operator's Signature	<i>Isabelle Szymanski</i> Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIALANTIC PD
Time of Inspection: 16:56

Date of Inspection: 06/12/2026

Serial Number: 80-001006
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.050	0.078	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.081
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.051	0.081	0.200	0.080


Standard Deviations	0.0004	0.0008	0.0006	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.



I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

LEANDRA HIGGINBOTHAM

06/12/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001006 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001006</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>INDIALANTIC PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/12/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>16:56</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Date

06/12/2026

LEANDRA HIGGINBOTHAM,

Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality