

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKALOOSA COUNTY SO
Time of Inspection: 11:48

Date of Inspection: 04/14/2021

Serial Number: 80-000763
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.046	0.080	0.194	
0.000	0.047	0.081	0.200	
0.000	0.046	0.081	0.201	

Number of Simulators Used: 5

Remarks: During the inspection I finished the OZA solution checks, I reviewed the solution checks then hit ESCAPE so the instrument would take me to the next step which would be the 0.08 Dry Gas Std Test.

When I hit ESCAPE I WAS TAKEN PAST THE DRY GAS STD TEST AND THE DIAGNOSTIC CHECK AND STRAIGHT TO THE REVIEW REMARKS PART OF THE INSP. FDLE DEPARTMENT INSPECTOR SGT. ADVISED ME TO CONDUCT ANOTHER INSPECTION.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Michael W Gallagher
Signature and Printed Name

MICHAEL W GALLAGHER

04/14/2021
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKALOOSA COUNTY SO
Time of Inspection: 13:16

Date of Inspection: 04/14/2021

Serial Number: 80-000763
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1349789 Exp: 07/15/2023
0.000	0.047	0.080	0.193	0.081
0.000	0.048	0.081	0.201	0.081
0.000	0.048	0.082	0.202	0.081

Number of Simulators Used: 5

Remarks:

FDLE ADVISED THIS 2ND INSP BE CONDUCTED

Previous inspection on this date failed to allow 08 Dry Gas Std Test and Diagnostic test. During this inspection instrument tested O.K.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D/S Michael W Gallagher MICHAEL W GALLAGHER
Signature and Printed Name

04/14/2021
Date