



INSTRUMENT PROCESSING SHEET

Agency WALTON COUNTY SOS/N 80-001312Florida Department of
Law EnforcementDate In 9/3/25DI Completion Date 09/04/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By <u>KTS</u>	Date <u>9/3/25</u>	Quality Checks	By <u>SLH</u>	Date <u>9/3/25</u>	Flow Calibration	By _____	Date _____																																																																	
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Sue dropped off in Pelican case.</u>			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>223</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.460</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																																			
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 06/19/2026	0.080	MP5089	202406L 06/19/2026	0.200	MP5090	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	<table border="1"><thead><tr><th colspan="2">Maintenance By _____ Date _____</th></tr><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr></thead></table>			Maintenance By _____ Date _____		☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____																																									
Simulator	Serial #	Lot #/Exp																																																																							
0.050	MP5088	202406K 06/19/2026																																																																							
0.080	MP5089	202406L 06/19/2026																																																																							
0.200	MP5090	202406N 06/20/2026																																																																							
0.080 DGS	N/A	AG429602 10/22/2026																																																																							
Maintenance By _____ Date _____																																																																									
☐ Battery Replacement																																																																									
☐ Dry Gas Regulator Replacement																																																																									
☐ Breath Tube Replacement																																																																									
☐ Other _____																																																																									
Calibration Adjustment By _____			Department Inspection By <u>SLH</u>																																																																						
Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28421</u> Gauge <u>1011</u> Instrument <u>1009</u> Mouth Alcohol Solution Lot # <u>2025-A</u> Acetone Stock Solution Lot # <u>2024-B</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table> Attachments <table border="1"><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other _____</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6289	Interferent	MP5087	0.050	MP5088	0.080	MP6292	0.200	MP6293	☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☐ Other _____
Simulator	Serial #	Lot #	Expiration																																																																						
0.000		N/A	N/A																																																																						
0.040																																																																									
0.100																																																																									
0.200																																																																									
0.300																																																																									
0.080 DGS	N/A																																																																								
Simulator	Serial #	Lot #	Expiration																																																																						
0.050																																																																									
0.080																																																																									
0.200																																																																									
0.080 DGS	N/A																																																																								
Simulator	Serial Number																																																																								
0.000	MP6289																																																																								
Interferent	MP5087																																																																								
0.050	MP5088																																																																								
0.080	MP6292																																																																								
0.200	MP6293																																																																								
☒ Form 41	☐ Post-Stability Checks																																																																								
☒ Stability Checks	☐ Flow Calibration																																																																								
☒ Calibration Certificate	☐ Form 40																																																																								
☐ Calibration Adjustment	☐ Other _____																																																																								
Notes/Suggested Service: _____ _____ _____ _____ _____ _____			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.09.07 22:45:57 -0400 Shayla Platt Digitally signed by Shayla Platt Date: 2025.09.17 12:43:40 -0400 Tech Review / Date _____ Admin Review / Date _____																																																																						

Stability Checks

80-001312 Sub 9/3/2025

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L 50.003 g/210L of Wet
WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 09/03/2025 Software: 8100.27 SN 80-001312	WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 09/03/2025 Software: 8100.27 SN 81-001312	WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 09/03/2025 Software: 8100.27 SN 80-001312	WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 09/03/2025 Software: 8100.27 SN 80-001312
Performed Root Case Analysis Test g/210L Time Air Blank 0.000 14:54 Control Test 0.047 14:55 Air Blank 0.000 14:56 Control Test 0.047 14:56 Air Blank 0.000 14:57 Control Test 0.047 14:58 Air Blank 0.000 14:58 Control Test Stats Average 0.0473 Std Dev 0.0003 Rel Std Dev(%) 0.0003 Operator's Signature <i>M. Sigel</i>	Performed Root Case Analysis Test g/210L Time Air Blank 0.000 15:04 Control Test 0.077 15:05 Air Blank 0.000 15:05 Control Test 0.078 15:06 Air Blank 0.000 15:06 Control Test 0.078 15:07 Air Blank 0.000 15:08 Control Test Stats Average 0.0777 Std Dev 0.0006 Rel Std Dev(%) 0.0006 Operator's Signature <i>M. Sigel</i>	Performed Root Case Analysis Test g/210L Time Air Blank 0.000 14:59 Control Test 0.196 15:00 Air Blank 0.000 15:00 Control Test 0.196 15:01 Air Blank 0.000 15:02 Control Test 0.197 15:02 Air Blank 0.000 15:03 Control Test Stats Average 0.1963 Std Dev 0.0006 Rel Std Dev(%) 0.0006 Operator's Signature <i>M. Sigel</i>	Performed Root Case Analysis Test g/210L Time Air Blank 0.000 15:12 Control Test 0.078 15:12 Air Blank 0.000 15:13 Control Test 0.078 15:13 Air Blank 0.000 15:14 Control Test 0.078 15:14 Air Blank 0.000 15:14 Control Test Stats Average 0.0780 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>M. Sigel</i>

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO

Serial Number: 80-001312

Time of Inspection: 14:13

Date of Inspection: 09/04/2025

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#: 202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#: 202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG429602 Exp: 10/22/2026
0.000	0.047	0.077	0.197	0.079
0.000	0.047	0.077	0.197	0.079
0.000	0.047	0.078	0.197	0.079
0.000	0.047	0.078	0.198	0.078
0.000	0.047	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.078
0.000	0.047	0.077	0.197	0.079
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.078


Standard Deviations	0.0005	0.0004	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.




LEANDRA HIGGINBOTHAM

Signature and Printed Name

09/04/2025

Date

- noted - printer did not work
SW



ANAB
ANSI National Accreditation Board
ACCREDITED
ISO/IEC 17025
FORENSIC CALIBRATION
LABORATORY

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001312, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001312</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>WALTON COUNTY SO</u>	0.050 g/210 L	0.004
Calibration Date:	<u>09/04/2025</u>	0.080 g/210 L	0.004
Calibration Time:	<u>14:13</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

09/04/2025

Date


LEANDRA HIGGINBOTHAM,

Department Inspector

Service • Integrity • Respect • Quality