

Alcohol Testing Program - Instrument Processing Sheet

Agency: SEMINOLE COUNTY SO

Instrument Serial Number: 80-001272

Date In: 4/30/2026

DI Completion Date: 05/01/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: WKP Date: 4/30/2026		Quality Checks By: SLH Date: 04/30/2026		Flow Adjustment By: _____																
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: 216 ☒ Flow Verification (L/s) Flow Column #: ATP 105 32 mm 0.160 (.139-.169) 36 mm 0.171 (.156-.190) 53 mm 0.250 (.228-.278) 103 mm 0.511 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: 28427 Gauge: 1006 Instrument: 1008 ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																
		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202406K 06/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202603B 03/03/2028</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202406N 06/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701 04/17/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6291	202406K 06/19/2026	0.080	MP6292	202603B 03/03/2028	0.200	MP6293	202406N 06/20/2026	0.080 DGS	N/A	AG510701 04/17/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:	
Simulator	Serial #	Lot#/Exp																		
0.050	MP6291	202406K 06/19/2026																		
0.080	MP6292	202603B 03/03/2028																		
0.200	MP6293	202406N 06/20/2026																		
0.080 DGS	N/A	AG510701 04/17/2027																		
Optical Bench Adjustment By: _____		Department Inspection By: SLH																		
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: 28427																		
Simulator	Serial #	Lot #	Expiration	Gauge: 1008	Instrument: 1011															
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: 2025-D	Exp: 09/25/2027															
0.040				Acetone Stock Solution Lot #: 2025-B	Exp: 09/22/2027															
0.100				Simulator	Serial Number															
0.200				0.000	MP5086															
0.300				Interferent	MP6290															
0.080 DGS	N/A			0.050	MP6291															
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP6292															
Simulator	Serial #	Lot #	Expiration	0.200	MP6293															
0.050				Attachments																
0.080				☒ Form 41	☐ Post-Stability Checks															
0.200				☒ Stability Checks	☐ Flow Adjustment															
0.080 DGS	N/A			☒ Calibration Certificate	☐ Form 40															
Gauge ID #: _____				☐ Optical Bench Adjustment	☐ Other:															
Gauge: _____ Instrument: _____																				
Notes/Suggested Service:				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																
				Digitally signed by Taylor Gutschow Date: 2026.05.09 17:38:32 -04'00' Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.05.18 11:51:43 -04'00'																
				Tech Review Admin Review																

Stability Checks

80-001272 SUH
4/30/2026

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																																
SEMINOLE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001272 04/30/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:16</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:17</td></tr> <tr><td>Control Test</td><td>0.048</td><td>15:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:19</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Miguel</i>	Test	g/210L	Time	Air Blank	0.000	15:16	Control Test	0.049	15:17	Air Blank	0.000	15:17	Control Test	0.048	15:18	Air Blank	0.000	15:19	Control Test	0.049	15:19	Air Blank	0.000	15:20	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		SEMINOLE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001272 04/30/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:25</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:26</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:27</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:29</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Miguel</i>	Test	g/210L	Time	Air Blank	0.000	15:25	Control Test	0.079	15:25	Air Blank	0.000	15:26	Control Test	0.078	15:27	Air Blank	0.000	15:27	Control Test	0.079	15:28	Air Blank	0.000	15:29	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		SEMINOLE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001272 04/30/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:40</td></tr> <tr><td>Control Test</td><td>0.196</td><td>15:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:41</td></tr> <tr><td>Control Test</td><td>0.197</td><td>15:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:42</td></tr> <tr><td>Control Test</td><td>0.196</td><td>15:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:44</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2941</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Miguel</i>	Test	g/210L	Time	Air Blank	0.000	15:40	Control Test	0.196	15:41	Air Blank	0.000	15:41	Control Test	0.197	15:42	Air Blank	0.000	15:42	Control Test	0.196	15:43	Air Blank	0.000	15:44	Control Test Stats			Average	0.1963		Std Dev	0.0006		Rel Std Dev(%)	0.2941		DGS SEMINOLE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001272 04/30/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:31</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:32</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:33</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:34</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Miguel</i>	Test	g/210L	Time	Air Blank	0.000	15:31	Control Test	0.080	15:32	Air Blank	0.000	15:32	Control Test	0.080	15:33	Air Blank	0.000	15:33	Control Test	0.080	15:33	Air Blank	0.000	15:34	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:16																																																																																																																																																	
Control Test	0.049	15:17																																																																																																																																																	
Air Blank	0.000	15:17																																																																																																																																																	
Control Test	0.048	15:18																																																																																																																																																	
Air Blank	0.000	15:19																																																																																																																																																	
Control Test	0.049	15:19																																																																																																																																																	
Air Blank	0.000	15:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:25																																																																																																																																																	
Control Test	0.079	15:25																																																																																																																																																	
Air Blank	0.000	15:26																																																																																																																																																	
Control Test	0.078	15:27																																																																																																																																																	
Air Blank	0.000	15:27																																																																																																																																																	
Control Test	0.079	15:28																																																																																																																																																	
Air Blank	0.000	15:29																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:40																																																																																																																																																	
Control Test	0.196	15:41																																																																																																																																																	
Air Blank	0.000	15:41																																																																																																																																																	
Control Test	0.197	15:42																																																																																																																																																	
Air Blank	0.000	15:42																																																																																																																																																	
Control Test	0.196	15:43																																																																																																																																																	
Air Blank	0.000	15:44																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1963																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2941																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:31																																																																																																																																																	
Control Test	0.080	15:32																																																																																																																																																	
Air Blank	0.000	15:32																																																																																																																																																	
Control Test	0.080	15:33																																																																																																																																																	
Air Blank	0.000	15:33																																																																																																																																																	
Control Test	0.080	15:33																																																																																																																																																	
Air Blank	0.000	15:34																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SEMINOLE COUNTY SO
Time of Inspection: 14:40

Date of Inspection: 05/01/2026

Serial Number: 80-001272
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.078	0.197	0.079
0.000	0.050	0.079	0.197	0.080
0.000	0.050	0.078	0.197	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.050	0.079	0.197	0.080
0.000	0.050	0.079	0.197	0.080


Standard Deviations	0.0005	0.0005	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

05/01/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001272, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001272</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>SEMINOLE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/01/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:40</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/01/2026

Date


LEANDRA HIGGINBOTHAM,
Department Inspector