

Alcohol Testing Program - Instrument Processing Sheet

Agency: Fort Lauderdale Police Department

Instrument Serial Number: 80-006939

Date In: 5/7/2026

DI Completion Date: 7/22/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>TDG</u> Date: <u>5/18/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>7/22/2026</u>		Flow Adjustment By: _____																								
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Regular was detached from the dry gas shelf but still connected to the back of the instrument. No screws were included. Missing F2 key on keyboard.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>214</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm 0.148 (.139-.169) 36 mm 0.167 (.156-.190) 53 mm 0.238 (.228-.278) 103 mm 0.507 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1016</u> Instrument: <u>1012</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																								
						<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP6286</td> <td>202603A</td> </tr> <tr> <td>3/3/2028</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP6287</td> <td>202603B</td> </tr> <tr> <td>3/3/2028</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP6288</td> <td>202603D</td> </tr> <tr> <td>3/4/2028</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>28424080A3</td> </tr> <tr> <td>11/5/2026</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6286	202603A	3/3/2028	0.080	MP6287	202603B	3/3/2028	0.200	MP6288	202603D	3/4/2028	0.080 DGS	N/A	28424080A3	11/5/2026	Maintenance By: <u>TDG</u> Date: <u>5/18/2026</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☒ Other: Added dry gas regulator plate and four screws. Secured the regulator to the shelf.	
						Simulator	Serial #	Lot#/Exp																				
						0.050	MP6286	202603A																				
								3/3/2028																				
0.080	MP6287	202603B																										
		3/3/2028																										
0.200	MP6288	202603D																										
		3/4/2028																										
0.080 DGS	N/A	28424080A3																										
		11/5/2026																										
Optical Bench Adjustment By: _____		Department Inspection By: <u>TDG</u>																										
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>33363 (x2)</u>																										
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1016 / 1016</u> Instrument: <u>1012 / 1012</u>																								
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>																								
0.040				Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>																								
0.100				Simulator	Serial Number																							
0.200				0.000	MP6284																							
0.300				Interferent	MP6285																							
0.080 DGS	N/A			0.050	MP6286																							
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP6287																							
Simulator	Serial #	Lot #	Expiration	0.200	MP6288																							
0.050				Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment ☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other:																								
0.080																												
0.200																												
0.080 DGS	N/A																											
Gauge ID #: _____ Gauge: _____ Instrument: _____																												
Notes/Suggested Service: Called and spoke to the Agency Inspector about the unscrewed regulator, which was similarly unscrewed during the December 2025 inspection. He advised the screws were coming out in the car, and the operator may have removed them. I told him I would add a plate between the screws and shelf, and I instructed him to tell the operator not to remove the screws. If the screws come out again with the plate attached, I told him to contact me to discuss. (TDG 5/18/26)				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																								
				☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																								
				☒ Conduct an Agency Inspection Before Evidentiary Use																								
				LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.08.03 12:28:54 -04'00' Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.08.04 08:58:03 -04'00'																								
Tech Review		Admin Review																										

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT LAUDERDALE PD

Time of Inspection: 12:11

Date of Inspection: 07/22/2026

Serial Number: 80-006939

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.




TAYLOR D. GUTSCHOW

Signature and Printed Name

07/22/2026

Date

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤ 0.003 of Wet																																																																																																																																																
FORT LAUDERDALE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006939 07/22/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:55</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:57</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:53	Control Test	0.049	12:54	Air Blank	0.000	12:54	Control Test	0.049	12:55	Air Blank	0.000	12:55	Control Test	0.049	12:56	Air Blank	0.000	12:57	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FORT LAUDERDALE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006939 07/22/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:47</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:49</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:50</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:51</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7677</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:47	Control Test	0.079	12:48	Air Blank	0.000	12:49	Control Test	0.079	12:49	Air Blank	0.000	12:50	Control Test	0.080	12:50	Air Blank	0.000	12:51	Control Test Stats			Average	0.0793		Std Dev	0.0016		Rel Std Dev(%)	0.7677		FORT LAUDERDALE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006939 07/22/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:42</td></tr> <tr><td>Control Test</td><td>0.199</td><td>12:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:43</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:44</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:45</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.1983</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:42	Control Test	0.199	12:42	Air Blank	0.000	12:43	Control Test	0.198	12:44	Air Blank	0.000	12:44	Control Test	0.198	12:45	Air Blank	0.000	12:45	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		FORT LAUDERDALE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006939 07/22/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:13</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:14</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:15</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:13	Control Test	0.078	12:14	Air Blank	0.000	12:14	Control Test	0.078	12:14	Air Blank	0.000	12:15	Control Test	0.078	12:15	Air Blank	0.000	12:16	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:53																																																																																																																																																	
Control Test	0.049	12:54																																																																																																																																																	
Air Blank	0.000	12:54																																																																																																																																																	
Control Test	0.049	12:55																																																																																																																																																	
Air Blank	0.000	12:55																																																																																																																																																	
Control Test	0.049	12:56																																																																																																																																																	
Air Blank	0.000	12:57																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:47																																																																																																																																																	
Control Test	0.079	12:48																																																																																																																																																	
Air Blank	0.000	12:49																																																																																																																																																	
Control Test	0.079	12:49																																																																																																																																																	
Air Blank	0.000	12:50																																																																																																																																																	
Control Test	0.080	12:50																																																																																																																																																	
Air Blank	0.000	12:51																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel Std Dev(%)	0.7677																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:42																																																																																																																																																	
Control Test	0.199	12:42																																																																																																																																																	
Air Blank	0.000	12:43																																																																																																																																																	
Control Test	0.198	12:44																																																																																																																																																	
Air Blank	0.000	12:44																																																																																																																																																	
Control Test	0.198	12:45																																																																																																																																																	
Air Blank	0.000	12:45																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1983																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:13																																																																																																																																																	
Control Test	0.078	12:14																																																																																																																																																	
Air Blank	0.000	12:14																																																																																																																																																	
Control Test	0.078	12:14																																																																																																																																																	
Air Blank	0.000	12:15																																																																																																																																																	
Control Test	0.078	12:15																																																																																																																																																	
Air Blank	0.000	12:16																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT LAUDERDALE PD
Time of Inspection: 13:23

Date of Inspection: 07/22/2026

Serial Number: 80-006939
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:

A F / M A: .

*Inadvertently did not provide mouth alcohol sample. Inadvertently
aborted instead of repeating the test. Will repeat the inspection.
MG 7/22/26*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

07/22/2026
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT LAUDERDALE PD
Time of Inspection: 15:30

Date of Inspection: 07/22/2026

Serial Number: 80-006939
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28424080A3 Exp: 11/05/2026
0.000	0.050	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.050	0.078	0.197	0.078
0.000	0.050	0.079	0.197	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.050	0.078	0.197	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.050	0.078	0.197	0.078

Standard Deviations	0.0005	0.0004	0.0000	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 TAYLOR D GUTSCHOW
Signature and Printed Name

07/22/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006939, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006939</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FORT LAUDERDALE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/22/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:30</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/22/2026

Date

**Taylor
Gutschow**

Digitally signed by Taylor
Gutschow
Date: 2026.07.22 15:55:52
+04'00'

**TAYLOR D GUTSCHOW,
Department Inspector**

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1