

Alcohol Testing Program - Instrument Processing Sheet

Agency: JACKSONVILLE SO

Instrument Serial Number: 80-001069

Date In: 6/23/2026

DI Completion Date: 06/25/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: KTS Date: 6/23/2026		Quality Checks By: SLH Date: 6/23/2026		Flow Adjustment By: _____																															
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: 232 ☒ Flow Verification (L/s) Flow Column #: ATP 105 32 mm 0.156 (.139-.169) 36 mm 0.167 (.156-.190) 53 mm 0.238 (.228-.278) 103 mm 0.503 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: 34419 Gauge: 1016 Instrument: 1013 ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																															
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP6289</td> <td>202603B</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202603D</td> <td>☐ Other:</td> </tr> <tr> <td>3/4/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23825080A1</td> <td></td> </tr> <tr> <td>10/5/2027</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP5088	202603A	☐ Battery Replacement	3/3/2028	☐ Dry Gas Regulator Replacement	0.080	MP6289	202603B	☐ Tank Sensor Tare	3/3/2028	☐ Breath Tube Replacement	0.200	MP5090	202603D	☐ Other:	3/4/2028		0.080 DGS	N/A	23825080A1		10/5/2027	
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																										
						0.050	MP5088	202603A	☐ Battery Replacement																										
								3/3/2028	☐ Dry Gas Regulator Replacement																										
0.080	MP6289	202603B	☐ Tank Sensor Tare																																
		3/3/2028	☐ Breath Tube Replacement																																
0.200	MP5090	202603D	☐ Other:																																
		3/4/2028																																	
0.080 DGS	N/A	23825080A1																																	
		10/5/2027																																	
Optical Bench Adjustment By: SLH Department Inspection By: SLH Barometric Pressure Gauge: 1013 ID#: 28662 Barometric Pressure ID#: 34419																																			
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>25090</td> <td>3/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>24390</td> <td>10/29/2026</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>25020</td> <td>1/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>24430</td> <td>12/10/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603</td> <td>9/23/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	3/11/2027	0.100	MP6296	24390	10/29/2026	0.200	MP6297	25020	1/14/2027	0.300	MP6298	24430	12/10/2026	0.080 DGS	N/A	AG526603	9/23/2027	Gauge: 1014 Instrument: 1013 Mouth Alcohol Solution Lot #: 2025-D Exp: 09/25/2027 Acetone Stock Solution Lot #: 2025-B Exp: 09/22/2027					
Simulator	Serial #	Lot #	Expiration																																
0.000	MP6294	N/A	N/A																																
0.040	MP6295	25090	3/11/2027																																
0.100	MP6296	24390	10/29/2026																																
0.200	MP6297	25020	1/14/2027																																
0.300	MP6298	24430	12/10/2026																																
0.080 DGS	N/A	AG526603	9/23/2027																																
☒ Post Optical Bench Adjustment Stability Checks		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A</td> <td>3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP6289</td> <td>202603B</td> <td>3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D</td> <td>3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1</td> <td>10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	3/3/2028	0.080	MP6289	202603B	3/3/2028	0.200	MP5090	202603D	3/4/2028	0.080 DGS	N/A	23825080A1	10/5/2027												
Simulator	Serial #	Lot #	Expiration																																
0.050	MP5088	202603A	3/3/2028																																
0.080	MP6289	202603B	3/3/2028																																
0.200	MP5090	202603D	3/4/2028																																
0.080 DGS	N/A	23825080A1	10/5/2027																																
Gauge ID #: 34419 Gauge: 1013 Instrument: 1012		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment																																	
Notes/Suggested Service: Tech Review- Corrected the 0.080g/210L lot number on Form 41 to 202603B. SLH 7/1/2026		☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																																	
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																																	
		☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																	
Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.07.01 15:56:31 -04'00'		Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.07.09 09:50:25 -04'00'																																	
Tech Review		Admin Review																																	

Stability Checks

80-00 1069

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L 50.003 g/210L of Wet
JACKSONVILLE SD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001069 06/23/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:17 Control Test 0.050 14:17 Air Blank 0.000 14:18 Control Test 0.048 14:19 Air Blank 0.000 14:19 Control Test 0.048 14:20 Air Blank 0.000 14:20 Control Test Stats Average 0.0487 Std Dev 0.0012 Rel Std Dev(%) 2.3727 Operator Signature <i>M. Sigal</i>	JACKSONVILLE SD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001069 06/23/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:58 Control Test 0.077 14:59 Air Blank 0.000 14:59 Control Test 0.077 15:00 Air Blank 0.000 15:01 Control Test 0.077 15:01 Air Blank 0.000 15:02 Control Test Stats Average 0.0770 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator Signature <i>M. Sigal</i>	JACKSONVILLE SD Intoxilizer - Alcohol Analyzer Model 8000 SN 30-001069 06/23/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:23 Control Test 0.195 14:24 Air Blank 0.000 14:24 Control Test 0.197 14:25 Air Blank 0.000 14:26 Control Test 0.197 14:26 Air Blank 0.000 14:27 Control Test Stats Average 0.1963 Std Dev 0.0112 Rel Std Dev(%) 0.5581 Operator Signature <i>M. Sigal</i>	JACKSONVILLE SD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001069 06/23/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:29 Control Test 0.081 14:29 Air Blank 0.000 14:30 Control Test 0.081 14:30 Air Blank 0.000 14:31 Control Test 0.082 14:31 Air Blank 0.000 14:32 Control Test Stats Average 0.0813 Std Dev 0.0006 Rel Std Dev(%) 0.7099 Operator Signature <i>M. Sigal</i>

DGS

Verified instrument was able to read gas pressure.



JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001069
06/23/2026 15:53:41

SJA

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.0900	(-0.0070)	0.1980	(-0.0050)
Sample #2	0.0930	(-0.0310)	0.2090	(-0.0310)
Sample #3	0.0960	(0.0060)	0.1750	(0.0050)
Sample #4	0.0970	(-0.0030)	0.2010	(-0.0180)
Avg % Abs	0.0953	(-0.0093)	0.1950	(-0.0147)
STD DEV	0.0021	(0.0193)	0.0178	(0.0182)
REL STD DEV	2.184	(206.742)	9.116	(124.296)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.8550	(-0.0050)	1.6470	(-0.0180)
Sample #2	0.8670	(0.0010)	1.6440	(-0.0030)
Sample #3	0.8700	(-0.0040)	1.6210	(-0.0070)
Sample #4	0.8570	(0.0060)	1.6170	(0.0000)
Avg % Abs	0.8647	(0.0010)	1.6273	(-0.0033)
STD DEV	0.0068	(0.0050)	0.0146	(0.0035)
REL STD DEV	0.787	(500.000)	0.895	(105.357)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	2.0020	(-0.0240)	3.7520	(-0.0390)
Sample #2	2.0000	(-0.0090)	3.7440	(-0.0060)
Sample #3	2.0160	(-0.0010)	3.7420	(0.0070)
Sample #4	1.9950	(0.0360)	3.7380	(0.0150)
Avg % Abs	2.0037	(0.0087)	3.7413	(0.0053)
STD DEV	0.0110	(0.0240)	0.0031	(0.0106)
REL STD DEV	0.547	(277.003)	0.082	(198.726)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	3.8350	(-0.0150)	7.0520	(0.0010)
Sample #2	3.8480	(0.0190)	7.0580	(0.0250)
Sample #3	3.8040	(0.0320)	7.0480	(0.0470)
Sample #4	3.8400	(0.0040)	7.0520	(0.0280)
Avg % Abs	3.8307	(0.0183)	7.0527	(0.0333)
STD DEV	0.0234	(0.0140)	0.0050	(0.0119)
REL STD DEV	0.612	(76.429)	0.071	(35.791)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.6270	(-0.0170)	10.2610	(-0.0230)
Sample #2	5.6320	(0.0170)	10.2190	(0.0340)
Sample #3	5.6120	(0.0080)	10.2490	(0.0260)
Sample #4	5.6340	(0.0330)	10.2290	(0.0430)
Avg % Abs	5.6260	(0.0193)	10.2323	(0.0343)
STD DEV	0.0122	(0.0127)	0.0153	(0.0085)
REL STD DEV	0.216	(65.495)	0.149	(24.772)

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001069
06/23/2026 15:53:41

Auto Calibration

pg 2 of 2

<<<< 3um >>>>

Zero Order Coef -240.23
First Order Coef 2461.89
Second Order Coef 21.52

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0001
0.040	0.040	0.0000
0.100	0.100	-0.0004
0.200	0.200	0.0004
0.300	0.300	-0.0001

<<<< 9um >>>>

-254.05
1302.05
11.66

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0000
0.040	0.040	0.0002
0.100	0.100	-0.0004
0.200	0.200	0.0003
0.300	0.300	-0.0001

<<<< 3um >>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample

Sample #1	3205.00
Sample #2	3282.00
Sample #3	3232.00
Sample #4	3256.00
Avg	3256.6667
STD DEV	25.0067
REL STD DEV	0.768
H2O adjust (mg/l*10k)	553

<<<< 9um >>>>

3301.00
3301.00
3299.00
3332.00
3310.6667
18.5023
0.559
499

Barometric Pressure = 1012

*****CALIBRATION SUCCESSFUL*****

SWS

Post Optical Bench Adjustment Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L 50.003 g/210L of Wet
☒ JACKSONVILLE 50 Intoxilyzer - Alcotest Analyzer Model 810C 06/23/2026 SN 80-001069 Software: 810C.27 Test g/210L Time Air Blank 0.000 17:37 Control Test 0.050 17:37 Air Blank 0.000 17:38 Control Test 0.050 17:38 Air Blank 0.000 17:39 Control Test 0.049 17:40 Air Blank 0.000 17:40 Control Test Stats Average 0.0497 Std Dev 0.0006 Rel Std Dev(%) 1.1625 Operator's Signature <i>Michael</i>	☒ JACKSONVILLE 50 Intoxilyzer - Alcotest Analyzer Model 810C 06/23/2026 SN 80-001069 Software: 810C.27 Test g/210L Time Air Blank 0.000 17:51 Control Test 0.079 17:52 Air Blank 0.000 17:52 Control Test 0.079 17:53 Air Blank 0.000 17:54 Control Test 0.073 17:54 Air Blank 0.000 17:55 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>Michael</i>	☒ JACKSONVILLE 50 Intoxilyzer - Alcotest Analyzer Model 810C 06/23/2026 SN 80-001069 Software: 810C.27 Test g/210L Time Air Blank 0.000 17:42 Control Test 0.199 17:43 Air Blank 0.000 17:43 Control Test 0.198 17:44 Air Blank 0.000 17:44 Control Test 0.199 17:45 Air Blank 0.000 17:46 Control Test Stats Average 0.1987 Std Dev 0.0006 Rel Std Dev(%) 0.2906 Operator's Signature <i>Michael</i>	☒ JACKSONVILLE 50 Intoxilyzer - Alcotest Analyzer Model 810C 06/23/2026 SN 80-001069 Software: 810C.27 Test g/210L Time Air Blank 0.000 17:58 Control Test 0.081 17:58 Air Blank 0.000 17:59 Control Test 0.080 17:59 Air Blank 0.000 18:00 Control Test 0.080 18:00 Air Blank 0.000 18:00 Control Test Stats Average 0.0803 Std Dev 0.0006 Rel Std Dev(%) 0.7187 Operator's Signature <i>Michael</i>

80-001069 SUT

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 12:47

Date of Inspection: 06/25/2026

Serial Number: 80-001069
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.052	0.079	0.198	0.081
0.000	0.051	0.079	0.198	0.081
0.000	0.051	0.080	0.198	0.080
0.000	0.051	0.079	0.199	0.080
0.000	0.051	0.080	0.199	0.080
0.000	0.051	0.080	0.199	0.081
0.000	0.051	0.079	0.199	0.081
0.000	0.052	0.079	0.200	0.080
0.000	0.051	0.080	0.199	0.080
0.000	0.052	0.081	0.199	0.081


Standard Deviations	0.0004	0.0006	0.0006	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 LEANDRA HIGGINBOTHAM
Signature and Printed Name

06/25/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001069, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001069</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>JACKSONVILLE SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/25/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:47</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.


Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/25/2026

Date


LEANDRA HIGGINBOTHAM,
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1