

Instrument Processing Sheet

Agency: St. Lucie County Sheriff's Office Instrument Serial Number: 80-000788
 Date In: 10/16/2025 DI Completion Date: 12/10/2025 Ship P/U H/D CMI EE

Intake By: <u>TDG</u> Date: <u>12/9/25</u>	Quality Checks By: <u>TDG</u> Date: <u>12/9/25</u>	Flow Adjustment By: _____ Date: _____															
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>167</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP101</u> 32 mm <u>0.160</u> (.139-.169) 36 mm <u>0.175</u> (.156-.190) 53 mm <u>0.253</u> (.228-.278) 103 mm <u>0.519</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>33364</u> Gauge: <u>1016</u> Instrument: <u>1015</u> <input checked="" type="checkbox"/> Stability Checks	Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)															
	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202406K 6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202406L 6/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603 9/23/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	0.050	MP6286	202406K 6/19/2026	0.080	MP6287	202406L 6/19/2026	0.200	MP6288	202406N 6/20/2026	0.080 DGS	N/A	AG526603 9/23/2027	Maintenance By: _____ Date: _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:
Simulator	Serial #	Lot#/Exp															
0.050	MP6286	202406K 6/19/2026															
0.080	MP6287	202406L 6/19/2026															
0.200	MP6288	202406N 6/20/2026															
0.080 DGS	N/A	AG526603 9/23/2027															

Optical Bench Adjustment By: <u>TDG</u>	Department Inspection By: <u>TDG</u>																																								
Barometric Pressure Gauge: <u>1015</u> ID#: <u>26932</u>	Barometric Pressure ID#: <u>33364</u>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5097</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5098</td> <td>25090</td> <td>3/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP5099</td> <td>24110</td> <td>3/5/2026</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>25020</td> <td>1/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>24430</td> <td>12/10/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>28424080A3</td> <td>11/15/2026</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	25090	3/11/2027	0.100	MP5099	24110	3/5/2026	0.200	MP5100	25020	1/14/2027	0.300	MP5101	24430	12/10/2026	0.080 DGS	N/A	28424080A3	11/15/2026	Gauge: <u>1017</u> Instrument: <u>1016</u> Mouth Alcohol Solution Lot #: <u>2025-C</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2024-B</u> Exp: <u>7/19/2026</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																						
0.000	MP5097	N/A	N/A																																						
0.040	MP5098	25090	3/11/2027																																						
0.100	MP5099	24110	3/5/2026																																						
0.200	MP5100	25020	1/14/2027																																						
0.300	MP5101	24430	12/10/2026																																						
0.080 DGS	N/A	28424080A3	11/15/2026																																						
Simulator	Serial Number																																								
0.000	MP6284																																								
Interferent	MP6285																																								
0.050	MP6286																																								
0.080	MP6287																																								
0.200	MP6288																																								
<input checked="" type="checkbox"/> Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202406K</td> <td>6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202406L</td> <td>6/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202406N</td> <td>6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603</td> <td>9/23/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050	MP6286	202406K	6/19/2026	0.080	MP6287	202406L	6/19/2026	0.200	MP6288	202406N	6/20/2026	0.080 DGS	N/A	AG526603	9/23/2027	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Optical Bench Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Other:																				
Simulator	Serial #	Lot #	Expiration																																						
0.050	MP6286	202406K	6/19/2026																																						
0.080	MP6287	202406L	6/19/2026																																						
0.200	MP6288	202406N	6/20/2026																																						
0.080 DGS	N/A	AG526603	9/23/2027																																						
Barometric Pressure Gauge: <u>1015</u> ID#: <u>26932</u>																																									

Notes/Suggested Service:	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between; margin-top: 10px;"> <div style="text-align: center;"> Digitally signed by Shayla Platt Date: 2025.12.14 13:42:55 -05'00' </div> <div style="text-align: center;"> Digitally signed by LeAndra Higginbotham Date: 2025.12.15 09:17:48 -05'00' </div> </div>
	Tech Review: _____ Admin Review: _____

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ST LUCIE COUNTY SO
Time of Inspection: 13:06

Date of Inspection: 12/09/2025

Serial Number: 80-000788
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

12/09/2025
Date

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/> ≤0.003 of Wet <input checked="" type="checkbox"/>																																																																																																												
<p>ST. LUISIE COUNTY SO Intoxilizer - Alcohol Analyzer Model: 8000 12/09/2025 Software: 8.001.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:39</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:40</td></tr> <tr><td>Control Test</td><td>0.047</td><td>13:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:42</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:43</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0477</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.2112</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	13:39	Control Test	0.048	13:40	Air Blank	0.000	13:40	Control Test	0.047	13:41	Air Blank	0.000	13:42	Control Test	0.048	13:42	Air Blank	0.000	13:43	Control Test Stats			Average	0.0477		Std Dev	0.0006		Rel. Std Dev(%)	1.2112		<p>ST. LUISIE COUNTY SO Intoxilizer - Alcohol Analyzer Model: 8000 12/09/2025 Software: 8.001.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:22</td></tr> <tr><td>Control Test</td><td>0.082</td><td>13:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:23</td></tr> <tr><td>Control Test</td><td>0.083</td><td>13:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:24</td></tr> <tr><td>Control Test</td><td>0.082</td><td>13:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0823</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7012</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	13:22	Control Test	0.082	13:22	Air Blank	0.000	13:23	Control Test	0.083	13:23	Air Blank	0.000	13:24	Control Test	0.082	13:25	Air Blank	0.000	13:25	Control Test Stats			Average	0.0823		Std Dev	0.0006		Rel. Std Dev(%)	0.7012		<p>ST. LUISIE COUNTY SO Intoxilizer - Alcohol Analyzer Model: 8000 12/09/2025 Software: 8.001.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:30</td></tr> <tr><td>Control Test</td><td>0.203</td><td>13:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:31</td></tr> <tr><td>Control Test</td><td>0.202</td><td>13:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:32</td></tr> <tr><td>Control Test</td><td>0.205</td><td>13:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:33</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2033</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7512</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	13:30	Control Test	0.203	13:30	Air Blank	0.000	13:31	Control Test	0.202	13:32	Air Blank	0.000	13:32	Control Test	0.205	13:33	Air Blank	0.000	13:33	Control Test Stats			Average	0.2033		Std Dev	0.0015		Rel. Std Dev(%)	0.7512		<p>Printed to external printer. Will attach. <i>ML</i> 12/9/25</p>
Test	g/210L	Time																																																																																																													
Air Blank	0.000	13:39																																																																																																													
Control Test	0.048	13:40																																																																																																													
Air Blank	0.000	13:40																																																																																																													
Control Test	0.047	13:41																																																																																																													
Air Blank	0.000	13:42																																																																																																													
Control Test	0.048	13:42																																																																																																													
Air Blank	0.000	13:43																																																																																																													
Control Test Stats																																																																																																															
Average	0.0477																																																																																																														
Std Dev	0.0006																																																																																																														
Rel. Std Dev(%)	1.2112																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	13:22																																																																																																													
Control Test	0.082	13:22																																																																																																													
Air Blank	0.000	13:23																																																																																																													
Control Test	0.083	13:23																																																																																																													
Air Blank	0.000	13:24																																																																																																													
Control Test	0.082	13:25																																																																																																													
Air Blank	0.000	13:25																																																																																																													
Control Test Stats																																																																																																															
Average	0.0823																																																																																																														
Std Dev	0.0006																																																																																																														
Rel. Std Dev(%)	0.7012																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	13:30																																																																																																													
Control Test	0.203	13:30																																																																																																													
Air Blank	0.000	13:31																																																																																																													
Control Test	0.202	13:32																																																																																																													
Air Blank	0.000	13:32																																																																																																													
Control Test	0.205	13:33																																																																																																													
Air Blank	0.000	13:33																																																																																																													
Control Test Stats																																																																																																															
Average	0.2033																																																																																																														
Std Dev	0.0015																																																																																																														
Rel. Std Dev(%)	0.7512																																																																																																														

ST LUCIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000788
12/09/2025
Software: 8100.27

DGS

Test	g/210L	Time
Air Blank	0.000	13:15
Control Test	0.079	13:15
Air Blank	0.000	13:16
Control Test	0.078	13:16
Air Blank	0.000	13:17
Control Test	0.079	13:17
Air Blank	0.000	13:17
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

ML

Operator's Signature

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																			
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																			
<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>																																																																																																																																																			
<input checked="" type="checkbox"/> ≤0.003 of Wet																																																																																																																																																						
<p>ST. JOHNS COUNTY SC Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000788 12/19/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:53</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:54</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:55</td></tr> <tr><td>Control Test</td><td>0.051</td><td>14:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:56</td></tr> <tr><td>Control Test</td><td>0.051</td><td>14:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel. Std Dev.(%)</td><td>2.3249</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	14:53	Control Test	0.049	14:53	Air Blank	0.000	14:54	Control Test	0.049	14:55	Air Blank	0.000	14:55	Control Test	0.051	14:56	Air Blank	0.000	14:56	Control Test	0.051	14:56	Control Test Stats			Average	0.0497		Std Dev	0.0012		Rel. Std Dev.(%)	2.3249		<p>ST. JOHNS COUNTY SC Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000788 12/19/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:00</td></tr> <tr><td>Control Test</td><td>0.081</td><td>15:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:01</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:02</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:03</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel. Std Dev.(%)</td><td>1.9255</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:00	Control Test	0.081	15:00	Air Blank	0.000	15:01	Control Test	0.079	15:01	Air Blank	0.000	15:02	Control Test	0.080	15:03	Air Blank	0.000	15:03	Control Test Stats			Average	0.0793		Std Dev	0.0015		Rel. Std Dev.(%)	1.9255		<p>ST. JOHNS COUNTY SC Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000788 12/19/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:07</td></tr> <tr><td>Control Test</td><td>0.203</td><td>15:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:08</td></tr> <tr><td>Control Test</td><td>0.200</td><td>15:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:09</td></tr> <tr><td>Control Test</td><td>0.200</td><td>15:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2010</td><td></td></tr> <tr><td>Std Dev</td><td>0.0017</td><td></td></tr> <tr><td>Rel. Std Dev.(%)</td><td>0.8657</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:07	Control Test	0.203	15:07	Air Blank	0.000	15:08	Control Test	0.200	15:08	Air Blank	0.000	15:09	Control Test	0.200	15:10	Air Blank	0.000	15:10	Control Test Stats			Average	0.2010		Std Dev	0.0017		Rel. Std Dev.(%)	0.8657		<p>ST. JOHNS COUNTY SC Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000788 12/19/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:47</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:48</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:49</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:50</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel. Std Dev.(%)</td><td>1.4555</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	14:47	Control Test	0.080	14:48	Air Blank	0.000	14:48	Control Test	0.078	14:48	Air Blank	0.000	14:49	Control Test	0.080	14:49	Air Blank	0.000	14:50	Control Test Stats			Average	0.0793		Std Dev	0.0012		Rel. Std Dev.(%)	1.4555	
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	14:53																																																																																																																																																				
Control Test	0.049	14:53																																																																																																																																																				
Air Blank	0.000	14:54																																																																																																																																																				
Control Test	0.049	14:55																																																																																																																																																				
Air Blank	0.000	14:55																																																																																																																																																				
Control Test	0.051	14:56																																																																																																																																																				
Air Blank	0.000	14:56																																																																																																																																																				
Control Test	0.051	14:56																																																																																																																																																				
Control Test Stats																																																																																																																																																						
Average	0.0497																																																																																																																																																					
Std Dev	0.0012																																																																																																																																																					
Rel. Std Dev.(%)	2.3249																																																																																																																																																					
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	15:00																																																																																																																																																				
Control Test	0.081	15:00																																																																																																																																																				
Air Blank	0.000	15:01																																																																																																																																																				
Control Test	0.079	15:01																																																																																																																																																				
Air Blank	0.000	15:02																																																																																																																																																				
Control Test	0.080	15:03																																																																																																																																																				
Air Blank	0.000	15:03																																																																																																																																																				
Control Test Stats																																																																																																																																																						
Average	0.0793																																																																																																																																																					
Std Dev	0.0015																																																																																																																																																					
Rel. Std Dev.(%)	1.9255																																																																																																																																																					
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	15:07																																																																																																																																																				
Control Test	0.203	15:07																																																																																																																																																				
Air Blank	0.000	15:08																																																																																																																																																				
Control Test	0.200	15:08																																																																																																																																																				
Air Blank	0.000	15:09																																																																																																																																																				
Control Test	0.200	15:10																																																																																																																																																				
Air Blank	0.000	15:10																																																																																																																																																				
Control Test Stats																																																																																																																																																						
Average	0.2010																																																																																																																																																					
Std Dev	0.0017																																																																																																																																																					
Rel. Std Dev.(%)	0.8657																																																																																																																																																					
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	14:47																																																																																																																																																				
Control Test	0.080	14:48																																																																																																																																																				
Air Blank	0.000	14:48																																																																																																																																																				
Control Test	0.078	14:48																																																																																																																																																				
Air Blank	0.000	14:49																																																																																																																																																				
Control Test	0.080	14:49																																																																																																																																																				
Air Blank	0.000	14:50																																																																																																																																																				
Control Test Stats																																																																																																																																																						
Average	0.0793																																																																																																																																																					
Std Dev	0.0012																																																																																																																																																					
Rel. Std Dev.(%)	1.4555																																																																																																																																																					
Operator's Signature 	Operator's Signature 	Operator's Signature 	Operator's Signature 																																																																																																																																																			

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ST LUCIE COUNTY SO
Time of Inspection: 13:06

Date of Inspection: 12/10/2025

Serial Number: 80-000788
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.047	0.081	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.048	0.081	0.201	0.079
0.000	0.049	0.080	0.200	0.080
0.000	0.048	0.079	0.202	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.048	0.080	0.198	0.079
0.000	0.047	0.080	0.198	0.079
0.000	0.049	0.081	0.199	0.080

Standard Deviations	0.0008	0.0008	0.0016	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0009 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

12/10/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000788, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000788</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>ST LUCIE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/10/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:06</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.12.10 15:12:18
-05'00'

12/10/2025 Date
Taylor D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality