



Agency Polk County Sheriff's Office

S/N 80-001157

Date In 08/06/2025 DI Completion Date 09/22/2025

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By TDG Date 08/22/2025

- ☒ Annual
- ☐ Registration
- ☐ Return from CMI / EE

- Case
- Keyboard
- Feet
- Ports
- Handle
- Dry Gas Shelf
- Breath Tube
- Screws Tight

- ☐ Power cord
- ☐ Printer Cable
- ☒ Static Bag
- ☐ 12V DC Cable

Notes:

Quality Checks By TDG Date 09/22/2025

☒ Breath Tube Screen
☒ Replace External O-Rings
☒ Instrument Set Up Verified
☒ R-Value 206

☒ Flow Verification (L/s)
 Flow Column # ATP104

 32 mm 0.156 (.139 - .169)
 36 mm 0.167 (.156 - .190)
 53 mm 0.242 (.228 - .278)
 103 mm 0.507 (.447 - .547)

■ Barometric Pressure Check
Gauge ID # 28199
■ Stability Checks

Flow Calibration By_____ Date_____	
------------------------------------	--

Flow Column # _____

☐ 5L/min – 17mm

☐ 15L/min – 53mm

☐ 30L/min – 103mm

☐ R-Value _____

☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm	_____	(.139 - .169)
36 mm	_____	(.156 - .190)
53 mm	_____	(.228 - .278)
103 mm	_____	(.447 - .547)

Maintenance By _____ Date _____

☐ Battery Replacement

☐ Dry Gas Regulator Replacement

☐ Breath Tube Replacement

☐ Other _____

Calibration Adjustment	By _____	Department Inspection	By <u>TDG</u>
------------------------	----------	-----------------------	---------------

Barometric Pressure Gauge _____ ID # _____

☐ Post Calibration Adjustment Stability Checks

Department Inspection By TDG

Barometric Pressure ID# 28199
Gauge 1016 Instrument 1021
Mouth Alcohol Solution Lot # 2024-A
Acetone Stock Solution Lot # 2024-B

Attachments

☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____
---	--

Notes/Suggested Service: _____

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC

- ☒ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use

- ☒ Conduct an Agency Inspection Before Evidentiary Use

Digitally signed by Charles Platt

Phil Nicodemo
Digitally signed by Phil Nicodemo
Date: 2025.10.08 12:37:11

Shayla Platt
Digitally signed by Shayla Platt
Date: 2025.10.15

Task Review / Date _____ Admin Review / Date _____

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK COUNTY SO

Time of Inspection: 11:31

Date of Inspection: 09/22/2025

Serial Number: 80-001157

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

09/22/2025
Date

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																												
POLK COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 09/22/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:48</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:49</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:51</td></tr> <tr><td>Control Test</td><td>0.048</td><td>11:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	11:48	Control Test	0.049	11:49	Air Blank	0.000	11:49	Control Test	0.049	11:50	Air Blank	0.000	11:51	Control Test	0.048	11:51	Air Blank	0.000	11:52	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		POLK COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 09/22/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:40</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:41</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:43</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:43</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	11:40	Control Test	0.079	11:40	Air Blank	0.000	11:41	Control Test	0.078	11:42	Air Blank	0.000	11:43	Control Test	0.079	11:43	Air Blank	0.000	11:43	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		POLK COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 09/22/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:04</td></tr> <tr><td>Control Test</td><td>0.196</td><td>12:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:05</td></tr> <tr><td>Control Test</td><td>0.195</td><td>12:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:06</td></tr> <tr><td>Control Test</td><td>0.196</td><td>12:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1957</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2951</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	12:04	Control Test	0.196	12:04	Air Blank	0.000	12:05	Control Test	0.195	12:06	Air Blank	0.000	12:06	Control Test	0.196	12:07	Air Blank	0.000	12:07	Control Test Stats			Average	0.1957		Std Dev	0.0006		Rel Std Dev(%)	0.2951		DGS 0.08g/210L 0.077 to 0.083 Printed to external printer and attached. ML 9/22/25
Test	g/210L	Time																																																																																																													
Air Blank	0.000	11:48																																																																																																													
Control Test	0.049	11:49																																																																																																													
Air Blank	0.000	11:49																																																																																																													
Control Test	0.049	11:50																																																																																																													
Air Blank	0.000	11:51																																																																																																													
Control Test	0.048	11:51																																																																																																													
Air Blank	0.000	11:52																																																																																																													
Control Test Stats																																																																																																															
Average	0.0487																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	1.1863																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	11:40																																																																																																													
Control Test	0.079	11:40																																																																																																													
Air Blank	0.000	11:41																																																																																																													
Control Test	0.078	11:42																																																																																																													
Air Blank	0.000	11:43																																																																																																													
Control Test	0.079	11:43																																																																																																													
Air Blank	0.000	11:43																																																																																																													
Control Test Stats																																																																																																															
Average	0.0787																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	0.7339																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	12:04																																																																																																													
Control Test	0.196	12:04																																																																																																													
Air Blank	0.000	12:05																																																																																																													
Control Test	0.195	12:06																																																																																																													
Air Blank	0.000	12:06																																																																																																													
Control Test	0.196	12:07																																																																																																													
Air Blank	0.000	12:07																																																																																																													
Control Test Stats																																																																																																															
Average	0.1957																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	0.2951																																																																																																														

POLK COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001157
09/22/2025
Software: 8100.27

0.08
D65

Test	g/210L	Time
Air Blank	0.000	11:34
Control Test	0.077	11:34
Air Blank	0.000	11:34
Control Test	0.077	11:35
Air Blank	0.000	11:35
Control Test	0.078	11:36
Air Blank	0.000	11:36
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

MC

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK COUNTY SO
Time of Inspection: 14:00

Date of Inspection: 09/22/2025

Serial Number: 80-001157
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.198	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.079
0.000	0.048	0.079	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.196	0.078


Standard Deviations	0.0004	0.0004	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

09/22/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001157, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001157</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>POLK COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>09/22/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:00</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.09.23 13:10:21
-04'00'

09/22/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality