

Alcohol Testing Program - Instrument Processing Sheet

Agency: Fort Myers Police Department Instrument Serial Number: 80-005655

Date In: 6/19/2026 DI Completion Date: 6/30/2026 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>TDG</u> Date: <u>6/24/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>6/29/2026</u>		Flow Adjustment By: _____																																							
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: No box. Missing back right foot.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>169</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm 0.148 (.139-.169) 36 mm 0.164 (.156-.190) 53 mm 0.234 (.228-.278) 103 mm 0.507 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1019</u> Instrument: <u>1018</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																																							
						<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603 9/23/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6286	202603A 3/3/2028	0.080	MP6287	202603B 3/3/2028	0.200	MP6288	202603D 3/4/2028	0.080 DGS	N/A	AG526603 9/23/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____																				
						Simulator	Serial #	Lot#/Exp																																			
						0.050	MP6286	202603A 3/3/2028																																			
						0.080	MP6287	202603B 3/3/2028																																			
0.200	MP6288	202603D 3/4/2028																																									
0.080 DGS	N/A	AG526603 9/23/2027																																									
Optical Bench Adjustment By: <u>TDG</u> Department Inspection By: <u>TDG</u> Barometric Pressure Gauge: <u>1019</u> ID#: <u>33364</u> Barometric Pressure ID#: <u>33363</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5097</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5098</td> <td>25090</td> <td>3/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP5099</td> <td>24390</td> <td>10/29/2026</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>25020</td> <td>1/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>24430</td> <td>12/10/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>28424080A3</td> <td>11/5/2026</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	25090	3/11/2027	0.100	MP5099	24390	10/29/2026	0.200	MP5100	25020	1/14/2027	0.300	MP5101	24430	12/10/2026	0.080 DGS	N/A	28424080A3	11/5/2026	Gauge: <u>1017</u> Instrument: <u>1017</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																								
0.000	MP5097	N/A	N/A																																								
0.040	MP5098	25090	3/11/2027																																								
0.100	MP5099	24390	10/29/2026																																								
0.200	MP5100	25020	1/14/2027																																								
0.300	MP5101	24430	12/10/2026																																								
0.080 DGS	N/A	28424080A3	11/5/2026																																								
Simulator	Serial Number																																										
0.000	MP6284																																										
Interferent	MP6285																																										
0.050	MP6286																																										
0.080	MP6287																																										
0.200	MP6288																																										
☒ Post Optical Bench Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202603A</td> <td>3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202603B</td> <td>3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202603D</td> <td>3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603</td> <td>9/23/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050	MP6286	202603A	3/3/2028	0.080	MP6287	202603B	3/3/2028	0.200	MP6288	202603D	3/4/2028	0.080 DGS	N/A	AG526603	9/23/2027	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment		☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other: _____																			
Simulator	Serial #	Lot #	Expiration																																								
0.050	MP6286	202603A	3/3/2028																																								
0.080	MP6287	202603B	3/3/2028																																								
0.200	MP6288	202603D	3/4/2028																																								
0.080 DGS	N/A	AG526603	9/23/2027																																								
Gauge ID #: <u>33363</u> Gauge: <u>1017</u> Instrument: <u>1017</u> Notes/Suggested Service:		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																									
		<table style="width:100%;"> <tr> <td style="width:50%; text-align: center;"> Digitally signed by Wen-Chi Pierson Date: 2026.07.01 08:58:37 -04'00' </td> <td style="width:50%; text-align: center;"> Digitally signed by Kaitlyn Spearin Date: 2026.07.09 10:08:51 -04'00' </td> </tr> </table>				Digitally signed by Wen-Chi Pierson Date: 2026.07.01 08:58:37 -04'00'	Digitally signed by Kaitlyn Spearin Date: 2026.07.09 10:08:51 -04'00'																																				
		Digitally signed by Wen-Chi Pierson Date: 2026.07.01 08:58:37 -04'00'	Digitally signed by Kaitlyn Spearin Date: 2026.07.09 10:08:51 -04'00'																																								
<table style="width:100%;"> <tr> <td style="width:50%; text-align: center;">Tech Review</td> <td style="width:50%; text-align: center;">Admin Review</td> </tr> </table>				Tech Review	Admin Review																																						
Tech Review	Admin Review																																										

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
FOR: MYERS PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-005655 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:04</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:05</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:06</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	13:04	Control Test	0.048	13:04	Air Blank	0.000	13:05	Control Test	0.049	13:05	Air Blank	0.000	13:06	Control Test	0.048	13:07	Air Blank	0.000	13:07	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		FOR: MYERS PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-005655 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:04</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:09</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	13:04	Control Test	0.079	13:08	Air Blank	0.000	13:09	Control Test	0.078	13:09	Air Blank	0.000	13:10	Control Test	0.078	13:10	Air Blank	0.000	13:11	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		FOR: MYERS PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-005655 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:57</td></tr> <tr><td>Control Test</td><td>0.194</td><td>12:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:58</td></tr> <tr><td>Control Test</td><td>0.192</td><td>12:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr> <tr><td>Control Test</td><td>0.192</td><td>13:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1927</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5993</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	12:57	Control Test	0.194	12:58	Air Blank	0.000	12:58	Control Test	0.192	12:59	Air Blank	0.000	12:59	Control Test	0.192	13:00	Air Blank	0.000	13:01	Control Test Stats			Average	0.1927		Std Dev	0.0012		Rel Std Dev(%)	0.5993		FOR: MYERS PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-005655 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:57</td></tr> <tr><td>Control Test</td><td>0.194</td><td>12:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:58</td></tr> <tr><td>Control Test</td><td>0.192</td><td>12:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr> <tr><td>Control Test</td><td>0.192</td><td>13:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1927</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5993</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	12:57	Control Test	0.194	12:58	Air Blank	0.000	12:58	Control Test	0.192	12:59	Air Blank	0.000	12:59	Control Test	0.192	13:00	Air Blank	0.000	13:01	Control Test Stats			Average	0.1927		Std Dev	0.0012		Rel Std Dev(%)	0.5993	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:04																																																																																																																																																	
Control Test	0.048	13:04																																																																																																																																																	
Air Blank	0.000	13:05																																																																																																																																																	
Control Test	0.049	13:05																																																																																																																																																	
Air Blank	0.000	13:06																																																																																																																																																	
Control Test	0.048	13:07																																																																																																																																																	
Air Blank	0.000	13:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:04																																																																																																																																																	
Control Test	0.079	13:08																																																																																																																																																	
Air Blank	0.000	13:09																																																																																																																																																	
Control Test	0.078	13:09																																																																																																																																																	
Air Blank	0.000	13:10																																																																																																																																																	
Control Test	0.078	13:10																																																																																																																																																	
Air Blank	0.000	13:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:57																																																																																																																																																	
Control Test	0.194	12:58																																																																																																																																																	
Air Blank	0.000	12:58																																																																																																																																																	
Control Test	0.192	12:59																																																																																																																																																	
Air Blank	0.000	12:59																																																																																																																																																	
Control Test	0.192	13:00																																																																																																																																																	
Air Blank	0.000	13:01																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1927																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5993																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:57																																																																																																																																																	
Control Test	0.194	12:58																																																																																																																																																	
Air Blank	0.000	12:58																																																																																																																																																	
Control Test	0.192	12:59																																																																																																																																																	
Air Blank	0.000	12:59																																																																																																																																																	
Control Test	0.192	13:00																																																																																																																																																	
Air Blank	0.000	13:01																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1927																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5993																																																																																																																																																		

***** AUTO CAL DATA *****
Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 0.000
0.000 0.000 0.000
0.040 0.040 0.0003
0.100 0.100 0.0001
0.200 0.200 0.0000
0.300 0.300 0.0000

***** CHANNEL 1 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.075
Std Dev = 0.03 Rel Std Dev = 38.80
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.765
Std Dev = 1.02 Rel Std Dev = 3.12
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.794
Std Dev = 0.03 Rel Std Dev = 1.42
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.531
Std Dev = 0.01 Rel Std Dev = 0.23
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.152
Std Dev = 0.03 Rel Std Dev = 0.56
Zero Order Coef = -173.15
First Order Coef = 2679.35
Second Order Coef = 23.85
Standard Deviation = 45.116074

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 6.9450 (-0.0120)
Sample #2 = 6.9420 (-0.0070)
Sample #3 = 6.8000 (-0.0200)
Sample #4 = 6.8240 (-0.0190)
Avg % Abs = 6.8247 (-0.0073)
Std Dev = 0.0170 (-0.0136)
REL STD DEV = 0.249 (185.140)

***** CHANNEL 1 *****
Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12747, Sum Io = 13740
***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 5.1500 (-0.0160)
Sample #2 = 5.1850 (-0.0200)
Sample #3 = 5.1350 (-0.0120)
Sample #4 = 5.1350 (-0.0120)
Avg % Abs = 5.1517 (-0.0013)
Std Dev = 0.0289 (-0.0185)
REL STD DEV = 0.560 (1385.640)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.5890 (-0.0210)
Sample #2 = 1.5230 (-0.0090)
Sample #3 = 1.5240 (-0.0110)
Sample #4 = 1.5460 (-0.0030)
Avg % Abs = 1.5310 (-0.0077)
Std Dev = 0.0130 (-0.0042)
REL STD DEV = 0.849 (54.304)

Auto Calibration
Max Power Res Value = 27
Auto Range Res Value = 15
Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12755, Sum Io = 13748
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.8000 (-0.0120)
Sample #2 = 1.7860 (-0.0100)
Sample #3 = 1.8220 (-0.0080)
Sample #4 = 1.7730 (-0.0170)
Avg % Abs = 1.7937 (-0.0063)
Std Dev = 0.0254 (-0.0129)
REL STD DEV = 1.415 (203.637)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.5620 (-0.0110)
Sample #2 = 3.5510 (-0.0140)
Sample #3 = 3.5950 (-0.0060)
Sample #4 = 3.5830 (-0.0040)
Avg % Abs = 3.5763 (-0.0040)
Std Dev = 0.0227 (-0.0100)
REL STD DEV = 0.636 (250.000)

***** CHANNEL 1 *****
Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12747, Sum Io = 13742
***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.8000 (-0.0120)
Sample #2 = 1.7860 (-0.0100)
Sample #3 = 1.8220 (-0.0080)
Sample #4 = 1.7730 (-0.0170)
Avg % Abs = 1.7937 (-0.0063)
Std Dev = 0.0254 (-0.0129)
REL STD DEV = 1.415 (203.637)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.1960 (-0.0100)
Sample #2 = 0.1420 (-0.0120)
Sample #3 = 0.1520 (-0.0050)
Sample #4 = 0.1580 (-0.0110)
Avg % Abs = 0.1507 (-0.0033)
Std Dev = 0.0081 (-0.0038)
REL STD DEV = 5.365 (40.564)

***** CHANNEL 1 *****
Sol Value = 0.140 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12751, Sum Io = 13744
***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.8140 (-0.0150)
Sample #2 = 0.7480 (-0.0120)
Sample #3 = 0.7540 (-0.0140)
Sample #4 = 0.7920 (-0.0080)
Avg % Abs = 0.7647 (-0.0060)
Std Dev = 0.0239 (-0.0122)
REL STD DEV = 3.120 (202.759)

***** CHANNEL 1 *****
Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 2 *****
Sample #1 = 3354.00
Sample #2 = 3380.00
Sample #3 = 3365.00
Sample #4 = 3369.00
Average Result = 3371.3333
STD DEV = 7.7675
REL STD DEV = 0.230

***** CHANNEL 2 *****
Sample #1 = 3335.00
Sample #2 = 3331.00
Sample #3 = 3348.00
Sample #4 = 3321.00
Average Result = 3333.3333
STD DEV = 13.5504
REL STD DEV = 0.410

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 9.9260 (-0.0070)
Sample #2 = 9.9340 (-0.0150)
Sample #3 = 9.9160 (-0.0200)
Sample #4 = 9.9190 (-0.0080)
Avg % Abs = 9.9197 (-0.0143)
Std Dev = 0.0140 (-0.0060)
REL STD DEV = 0.141 (42.054)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.4870 (-0.0100)
Sample #2 = 3.5360 (-0.0140)
Sample #3 = 3.5220 (-0.0000)
Sample #4 = 3.5360 (-0.0120)
Avg % Abs = 3.5313 (-0.0087)
Std Dev = 0.0081 (-0.0076)
REL STD DEV = 0.229 (87.358)

***** CHANNEL 1 *****
Sol Value = 0.140 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12751, Sum Io = 13744
***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.8140 (-0.0150)
Sample #2 = 0.7480 (-0.0120)
Sample #3 = 0.7540 (-0.0140)
Sample #4 = 0.7920 (-0.0080)
Avg % Abs = 0.7647 (-0.0060)
Std Dev = 0.0239 (-0.0122)
REL STD DEV = 3.120 (202.759)

***** CHANNEL 2 *****
Sample #1 = 3335.00
Sample #2 = 3331.00
Sample #3 = 3348.00
Sample #4 = 3321.00
Average Result = 3333.3333
STD DEV = 13.5504
REL STD DEV = 0.410

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 9.9260 (-0.0070)
Sample #2 = 9.9340 (-0.0150)
Sample #3 = 9.9160 (-0.0200)
Sample #4 = 9.9190 (-0.0080)
Avg % Abs = 9.9197 (-0.0143)
Std Dev = 0.0140 (-0.0060)
REL STD DEV = 0.141 (42.054)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.4870 (-0.0100)
Sample #2 = 3.5360 (-0.0140)
Sample #3 = 3.5220 (-0.0000)
Sample #4 = 3.5360 (-0.0120)
Avg % Abs = 3.5313 (-0.0087)
Std Dev = 0.0081 (-0.0076)
REL STD DEV = 0.229 (87.358)

***** CHANNEL 1 *****
Sol Value = 0.140 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12751, Sum Io = 13744
***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.8140 (-0.0150)
Sample #2 = 0.7480 (-0.0120)
Sample #3 = 0.7540 (-0.0140)
Sample #4 = 0.7920 (-0.0080)
Avg % Abs = 0.7647 (-0.0060)
Std Dev = 0.0239 (-0.0122)
REL STD DEV = 3.120 (202.759)

***** CHANNEL 2 *****
Sample #1 = 3335.00
Sample #2 = 3331.00
Sample #3 = 3348.00
Sample #4 = 3321.00
Average Result = 3333.3333
STD DEV = 13.5504
REL STD DEV = 0.410

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 9.9260 (-0.0070)
Sample #2 = 9.9340 (-0.0150)
Sample #3 = 9.9160 (-0.0200)
Sample #4 = 9.9190 (-0.0080)
Avg % Abs = 9.9197 (-0.0143)
Std Dev = 0.0140 (-0.0060)
REL STD DEV = 0.141 (42.054)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.4870 (-0.0100)
Sample #2 = 3.5360 (-0.0140)
Sample #3 = 3.5220 (-0.0000)
Sample #4 = 3.5360 (-0.0120)
Avg % Abs = 3.5313 (-0.0087)
Std Dev = 0.0081 (-0.0076)
REL STD DEV = 0.229 (87.358)

***** CHANNEL 1 *****
Sol Value = 0.140 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12751, Sum Io = 13744
***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.8140 (-0.0150)
Sample #2 = 0.7480 (-0.0120)
Sample #3 = 0.7540 (-0.0140)
Sample #4 = 0.7920 (-0.0080)
Avg % Abs = 0.7647 (-0.0060)
Std Dev = 0.0239 (-0.0122)
REL STD DEV = 3.120 (202.759)

***** CHANNEL 2 *****
Sample #1 = 3335.00
Sample #2 = 3331.00
Sample #3 = 3348.00
Sample #4 = 3321.00
Average Result = 3333.3333
STD DEV = 13.5504
REL STD DEV = 0.410

***** CHANNEL 1 *****
Sol Value = 0.140 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12751, Sum Io = 13744
***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.8140 (-0.0150)
Sample #2 = 0.7480 (-0.0120)
Sample #3 = 0.7540 (-0.0140)
Sample #4 = 0.7920 (-0.0080)
Avg % Abs = 0.7647 (-0.0060)
Std Dev = 0.0239 (-0.0122)
REL STD DEV = 3.120 (202.759)

Optical Calibration
Adjustment
By: TDG

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005655 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:01</td></tr> <tr><td>Control Test</td><td>0.050</td><td>15:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:02</td></tr> <tr><td>Control Test</td><td>0.050</td><td>15:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:03</td></tr> <tr><td>Control Test</td><td>0.050</td><td>15:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:01	Control Test	0.050	15:02	Air Blank	0.000	15:02	Control Test	0.050	15:03	Air Blank	0.000	15:03	Control Test	0.050	15:04	Air Blank	0.000	15:05	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005655 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:15</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:16</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:17</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:18</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:15	Control Test	0.079	15:15	Air Blank	0.000	15:16	Control Test	0.078	15:16	Air Blank	0.000	15:17	Control Test	0.079	15:18	Air Blank	0.000	15:18	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005655 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:24</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:25</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:26</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:28</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:24	Control Test	0.198	15:25	Air Blank	0.000	15:25	Control Test	0.198	15:26	Air Blank	0.000	15:26	Control Test	0.198	15:27	Air Blank	0.000	15:28	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005655 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:53</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:54</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:55</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	14:53	Control Test	0.079	14:54	Air Blank	0.000	14:54	Control Test	0.078	14:54	Air Blank	0.000	14:55	Control Test	0.079	14:55	Air Blank	0.000	14:56	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:01																																																																																																																																																	
Control Test	0.050	15:02																																																																																																																																																	
Air Blank	0.000	15:02																																																																																																																																																	
Control Test	0.050	15:03																																																																																																																																																	
Air Blank	0.000	15:03																																																																																																																																																	
Control Test	0.050	15:04																																																																																																																																																	
Air Blank	0.000	15:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:15																																																																																																																																																	
Control Test	0.079	15:15																																																																																																																																																	
Air Blank	0.000	15:16																																																																																																																																																	
Control Test	0.078	15:16																																																																																																																																																	
Air Blank	0.000	15:17																																																																																																																																																	
Control Test	0.079	15:18																																																																																																																																																	
Air Blank	0.000	15:18																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:24																																																																																																																																																	
Control Test	0.198	15:25																																																																																																																																																	
Air Blank	0.000	15:25																																																																																																																																																	
Control Test	0.198	15:26																																																																																																																																																	
Air Blank	0.000	15:26																																																																																																																																																	
Control Test	0.198	15:27																																																																																																																																																	
Air Blank	0.000	15:28																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:53																																																																																																																																																	
Control Test	0.079	14:54																																																																																																																																																	
Air Blank	0.000	14:54																																																																																																																																																	
Control Test	0.078	14:54																																																																																																																																																	
Air Blank	0.000	14:55																																																																																																																																																	
Control Test	0.079	14:55																																																																																																																																																	
Air Blank	0.000	14:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
 Operator's Signature	 Operator's Signature	 Operator's Signature	 Operator's Signature																																																																																																																																																

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT MYERS PD
Time of Inspection: 12:53

Date of Inspection: 06/30/2026

Serial Number: 80-005655
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.079	0.199	0.078
0.000	0.050	0.079	0.199	0.078
0.000	0.050	0.080	0.200	0.077
0.000	0.050	0.079	0.200	0.077
0.000	0.051	0.079	0.200	0.077
0.000	0.050	0.079	0.200	0.078
0.000	0.050	0.079	0.200	0.078
0.000	0.051	0.079	0.199	0.078
0.000	0.051	0.079	0.200	0.077


Standard Deviations	0.0004	0.0004	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/30/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005655, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005655</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FORT MYERS PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/30/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:53</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2026.06.30 14:36:06
-04'00'

06/30/2026

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality