Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEBRETON,QUINN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LYNCH,DANIEL
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: JANKOWSKI

DATE/TIME OF TEST: 04/02/2025 23:56:32

Diagnostics Check		OK	23:49
Air Blank		0.000	23:50
Control Test		0.082	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.187	23:51
Breath Volume		1.289	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.178	23:55
Breath Volume		1.363	23:55
Air Blank		0.000	23:55
Control Test		0.082	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 04/11/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOWERS,SHEHARA,D
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: MCNAMARA,KYLE
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: RHODES

DATE/TIME OF TEST: 04/04/2025 22:15:04

Diagnostics Check		OK	22:08
Air Blank		0.000	22:09
Control Test		0.081	22:09
Air Blank		0.000	22:09
Subject Sample	# 1	0.000	22:10
Breath Volume		2.175	22:10
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 2	0.000	22:13
Breath Volume		2.062	22:13
Air Blank		0.000	22:14
Control Test		0.082	22:14
Air Blank		0.000	22:14
Diagnostics Check		OK	22:15

Data Download Date/Time: 04/11/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MODERSON,C,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: NIMMO,O
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: BONVENTRE

DATE/TIME OF TEST: 04/15/2025 20:27:12

Diagnostics Check		OK	20:24	
Air Blank		0.000	20:24	
Control Test		0.081	20:25	
Air Blank		0.000	20:25	
Subject Sample	# 1		20:25	Subject Test Refused
Breath Volume		0.000	20:25	
Air Blank		0.000	20:26	
Control Test		0.081	20:26	
Air Blank		0.000	20:27	
Diagnostics Check		OK	20:27	

Data Download Date/Time: 05/22/2025 20:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,JAMES,B
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: MCNAMARA,KYLE
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 04/20/2025 20:26:53

Diagnostics Check		OK	20:23	
Air Blank		0.000	20:24	
Control Test		0.082	20:24	
Air Blank		0.000	20:24	
Subject Sample	# 1		20:25	Subject Test Refused
Breath Volume		0.000	20:25	
Air Blank		0.000	20:25	
Control Test		0.082	20:26	
Air Blank		0.000	20:26	
Diagnostics Check		OK	20:26	

Data Download Date/Time: 05/22/2025 20:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOWERS,SHEHARA,D
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: ROMO,SERGIO
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: HOANG

DATE/TIME OF TEST: 04/21/2025 03:38:04

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:30	
Control Test		0.082	03:30	
Air Blank		0.000	03:30	
Subject Sample	# 1		03:34	No Sample Provided
Breath Volume		0.000	03:34	
Air Blank		0.000	03:34	
Air Blank		0.000	03:36	
Subject Sample	# 2		03:36	Subject Test Refused
Breath Volume		0.000	03:36	
Air Blank		0.000	03:37	
Control Test		0.081	03:37	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 05/22/2025 20:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MODERSON,C,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: ZAYAS,A
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BYRNE

DATE/TIME OF TEST: 04/25/2025 23:11:43

Diagnostics Check		OK	23:05
Air Blank		0.000	23:05
Control Test		0.081	23:06
Air Blank		0.000	23:06
Subject Sample	# 1	0.232	23:07
Breath Volume		3.363	23:07
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.223	23:10
Breath Volume		2.371	23:10
Air Blank		0.000	23:10
Control Test		0.081	23:11
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 05/22/2025 20:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: BRACKIN,KENNETH
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:16

SUBJECT'S LAST NAME: CORONA

DATE/TIME OF TEST: 04/26/2025 13:47:32

Diagnostics Check		OK	13:36	
Air Blank		0.000	13:37	
Control Test		0.081	13:37	
Air Blank		0.000	13:38	
Subject Sample	# 1		13:41	No Sample Provided
Breath Volume		0.000	13:41	
Air Blank		0.000	13:41	
Air Blank		0.000	13:43	
Subject Sample	# 2		13:45	Subject Test Refused
Breath Volume		0.015	13:45	
Air Blank		0.000	13:46	
Control Test		0.082	13:46	
Air Blank		0.000	13:47	
Diagnostics Check		OK	13:47	

Data Download Date/Time: 05/22/2025 20:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 04/04/2025 01:10:50

Diagnostics Check		OK	01:08	
Air Blank		0.000	01:08	
Control Test		0.081	01:08	
Air Blank		0.000	01:09	
Subject Sample	# 1		01:09	RFI Detect
Breath Volume		0.000	01:09	
Air Blank		0.000	01:09	
Control Test		0.082	01:10	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 04/20/2025 19:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 04/04/2025 01:18:03

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.082	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.077	01:13
Breath Volume		2.679	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.075	01:16
Breath Volume		2.730	01:16
Air Blank		0.000	01:17
Control Test		0.080	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:18

Data Download Date/Time: 04/20/2025 19:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: STACY,COREY
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: SANFILIPPO

DATE/TIME OF TEST: 04/11/2025 00:48:09

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.081	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.038	00:41
Breath Volume		2.636	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.033	00:46
Breath Volume		1.324	00:46
Air Blank		0.000	00:47
Control Test		0.079	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 04/20/2025 19:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MALAGON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: NATHAN

DATE/TIME OF TEST: 04/11/2025 18:07:52

Diagnostics Check		OK	17:58
Air Blank		0.000	17:58
Control Test		0.080	17:59
Air Blank		0.000	17:59
Subject Sample	# 1	0.268	18:00
Breath Volume		1.320	18:00
Air Blank		0.000	18:01
Air Blank		0.000	18:02
Subject Sample	# 2	0.290	18:03
Breath Volume		1.871	18:03
Air Blank		0.000	18:04
Air Blank		0.000	18:05
Subject Sample	# 3	0.279	18:06
Breath Volume		1.781	18:06
Air Blank		0.000	18:06
Control Test		0.080	18:07
Air Blank		0.000	18:07
Diagnostics Check		OK	18:07

Data Download Date/Time: 04/20/2025 19:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCCRANN,SEAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/20/2025 03:21:29

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.081	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.092	03:16
Breath Volume		1.699	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.088	03:19
Breath Volume		1.406	03:19
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 04/20/2025 19:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,COREY,D
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: CAPLE,CAMERON
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 04/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 04/29/2025 23:15:34

Diagnostics Check		OK	22:59	
Air Blank		0.000	22:59	
Control Test		0.081	23:00	
Air Blank		0.000	23:00	
Subject Sample	# 1	0.210	23:03	
Breath Volume		1.453	23:03	
Air Blank		0.000	23:03	
Air Blank		0.000	23:05	
Subject Sample	# 2	0.193	23:08	Volume Not Met
Breath Volume		0.402	23:08	
Air Blank		0.000	23:09	
Air Blank		0.000	23:11	
Subject Sample	# 3	0.209	23:13	
Breath Volume		1.968	23:13	
Air Blank		0.000	23:14	
Control Test		0.076	23:14	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 05/23/2025 18:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANDY,JAMIE,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: JOHNSON,DARIUS
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: REYES MARTINEZ

DATE/TIME OF TEST: 04/19/2025 05:56:29

Diagnostics Check		OK	05:49
Air Blank		0.000	05:50
Control Test		0.078	05:50
Air Blank		0.000	05:50
Subject Sample	# 1	0.223	05:51
Breath Volume		2.886	05:51
Air Blank		0.000	05:52
Air Blank		0.000	05:54
Subject Sample	# 2	0.222	05:54
Breath Volume		2.828	05:54
Air Blank		0.000	05:55
Control Test		0.077	05:55
Air Blank		0.000	05:56
Diagnostics Check		OK	05:56

Data Download Date/Time: 04/29/2025 10:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SORRELL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 04/26/2025 22:52:28

Diagnostics Check		OK	22:45
Air Blank		0.000	22:46
Control Test		0.077	22:46
Air Blank		0.000	22:47
Subject Sample	# 1	0.000	22:47
Breath Volume		2.996	22:47
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.000	22:51
Breath Volume		1.628	22:51
Air Blank		0.000	22:51
Control Test		0.076	22:51
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 04/29/2025 10:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: FELICIANO

DATE/TIME OF TEST: 04/01/2025 01:13:21

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.082	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.000	01:08
Breath Volume		2.785	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.000	01:11
Breath Volume		2.925	01:11
Air Blank		0.000	01:12
Control Test		0.082	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 04/23/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARQUEZ CARRILLO,JOSE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: KRIZAN

DATE/TIME OF TEST: 04/01/2025 16:34:27

Diagnostics Check		OK	16:31	
Air Blank		0.000	16:32	
Control Test		0.081	16:32	
Air Blank		0.000	16:32	
Subject Sample	# 1		16:33	Subject Test Refused
Breath Volume		0.000	16:33	
Air Blank		0.000	16:33	
Control Test		0.081	16:33	
Air Blank		0.000	16:34	
Diagnostics Check		OK	16:34	

Data Download Date/Time: 04/23/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: DIEHL

DATE/TIME OF TEST: 04/03/2025 00:47:54

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:45	
Control Test		0.082	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 1		00:46	Subject Test Refused
Breath Volume		0.000	00:46	
Air Blank		0.000	00:47	
Control Test		0.082	00:47	
Air Blank		0.000	00:47	
Diagnostics Check		OK	00:47	

Data Download Date/Time: 04/23/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUZMAN,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CUSHING

DATE/TIME OF TEST: 04/03/2025 03:27:16

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.111	03:22
Breath Volume		3.539	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.120	03:25
Breath Volume		3.843	03:25
Air Blank		0.000	03:26
Control Test		0.081	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 04/23/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HERNANDEZ,LUIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: WENDELL

DATE/TIME OF TEST: 04/04/2025 00:09:59

Diagnostics Check		OK	00:03
Air Blank		0.000	00:04
Control Test		0.079	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.154	00:05
Breath Volume		1.312	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.152	00:08
Breath Volume		1.367	00:08
Air Blank		0.000	00:09
Control Test		0.083	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 04/23/2025 09:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAMMOCK,ANDREW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HAHN

DATE/TIME OF TEST: 04/04/2025 03:30:56

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.084	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.117	03:26
Breath Volume		3.292	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.119	03:29
Breath Volume		3.398	03:29
Air Blank		0.000	03:30
Control Test		0.082	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 04/23/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: LUCCA RIVERA

DATE/TIME OF TEST: 04/05/2025 03:06:52

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.082	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.007	03:02
Breath Volume		1.339	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.000	03:05
Breath Volume		1.492	03:05
Air Blank		0.000	03:05
Control Test		0.081	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 04/23/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RODRIGUEZ,R
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 04/05/2025 03:57:25

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.082	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.139	03:52
Breath Volume		2.687	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.132	03:55
Breath Volume		2.601	03:55
Air Blank		0.000	03:56
Control Test		0.082	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 04/23/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 11:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:34

SUBJECT'S LAST NAME: DURHAM

DATE/TIME OF TEST: 04/05/2025 13:02:13

Diagnostics Check		OK	12:56
Air Blank		0.000	12:56
Control Test		0.082	12:56
Air Blank		0.000	12:57
Subject Sample	# 1	0.185	12:57
Breath Volume		1.496	12:57
Air Blank		0.000	12:58
Air Blank		0.000	13:00
Subject Sample	# 2	0.180	13:00
Breath Volume		1.289	13:00
Air Blank		0.000	13:01
Control Test		0.082	13:01
Air Blank		0.000	13:02
Diagnostics Check		OK	13:02

Data Download Date/Time: 04/23/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSEN,DUSTIN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANSEN,DUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: BUI
ARREST TIME: 14:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: SAUVAGEOT

DATE/TIME OF TEST: 04/05/2025 16:19:22

Diagnostics Check		OK	16:16	
Air Blank		0.000	16:17	
Control Test		0.082	16:17	
Air Blank		0.000	16:17	
Subject Sample	# 1		16:18	Subject Test Refused
Breath Volume		0.000	16:18	
Air Blank		0.000	16:18	
Control Test		0.082	16:18	
Air Blank		0.000	16:19	
Diagnostics Check		OK	16:19	

Data Download Date/Time: 04/23/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACHADO,LAURA,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MACHADO,LAURA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: VALENTIN CINTRON

DATE/TIME OF TEST: 04/05/2025 21:35:20

Diagnostics Check		OK	21:32	
Air Blank		0.000	21:32	
Control Test		0.082	21:33	
Air Blank		0.000	21:33	
Subject Sample	# 1		21:33	Subject Test Refused
Breath Volume		0.000	21:33	
Air Blank		0.000	21:34	
Control Test		0.082	21:34	
Air Blank		0.000	21:35	
Diagnostics Check		OK	21:35	

Data Download Date/Time: 04/23/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLIPIN,A,
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: KLIPIN,A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/05/2025 22:23:23

Diagnostics Check		OK	22:10	
Air Blank		0.000	22:11	
Control Test		0.082	22:11	
Air Blank		0.000	22:12	
Subject Sample	# 1	0.180	22:15	Volume Not Met
Breath Volume		0.445	22:15	
Air Blank		0.000	22:15	
Air Blank		0.000	22:17	
Subject Sample	# 2	0.184	22:18	
Breath Volume		1.523	22:18	
Air Blank		0.000	22:19	
Air Blank		0.000	22:20	
Subject Sample	# 3	0.183	22:21	
Breath Volume		2.683	22:21	
Air Blank		0.000	22:22	
Control Test		0.081	22:22	
Air Blank		0.000	22:23	
Diagnostics Check		OK	22:23	

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DELAY,A
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/05/2025 23:04:15

Diagnostics Check		OK	22:56
Air Blank		0.000	22:57
Control Test		0.081	22:57
Air Blank		0.000	22:58
Subject Sample	# 1	0.264	22:59
Breath Volume		1.906	22:59
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.265	23:02
Breath Volume		2.335	23:02
Air Blank		0.000	23:03
Control Test		0.081	23:03
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BRODEUR,K
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: BOND

DATE/TIME OF TEST: 04/05/2025 23:59:10

Diagnostics Check		OK	23:56	
Air Blank		0.000	23:56	
Control Test		0.082	23:56	
Air Blank		0.000	23:57	
Subject Sample	# 1		23:57	Subject Test Refused
Breath Volume		0.000	23:57	
Air Blank		0.000	23:58	
Control Test		0.082	23:58	
Air Blank		0.000	23:59	
Diagnostics Check		OK	23:59	

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEISER,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GEISER,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MEAD

DATE/TIME OF TEST: 04/06/2025 00:07:49

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:05	
Control Test		0.082	00:05	
Air Blank		0.000	00:06	
Subject Sample	# 1		00:06	Subject Test Refused
Breath Volume		0.000	00:06	
Air Blank		0.000	00:06	
Control Test		0.082	00:07	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MESSENGER,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 04/06/2025 04:16:01

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:13	
Control Test		0.082	04:14	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:14	Subject Test Refused
Breath Volume		0.000	04:14	
Air Blank		0.000	04:15	
Control Test		0.082	04:15	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:16	

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EXTINE,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 04/06/2025 22:10:19

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.082	22:05
Air Blank		0.000	22:05
Subject Sample	# 1	0.000	22:06
Breath Volume		1.964	22:06
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.000	22:08
Breath Volume		1.320	22:08
Air Blank		0.000	22:09
Control Test		0.081	22:09
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PASKALAKIS,BRIANNA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 04/06/2025 23:44:43

Diagnostics Check		OK	23:41	
Air Blank		0.000	23:42	
Control Test		0.081	23:42	
Air Blank		0.000	23:43	
Subject Sample	# 1		23:43	Subject Test Refused
Breath Volume		0.000	23:43	
Air Blank		0.000	23:43	
Control Test		0.081	23:44	
Air Blank		0.000	23:44	
Diagnostics Check		OK	23:44	

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STOFFAN,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 04/08/2025 21:31:39

Diagnostics Check		OK	21:25
Air Blank		0.000	21:26
Control Test		0.081	21:26
Air Blank		0.000	21:26
Subject Sample	# 1	0.249	21:27
Breath Volume		2.886	21:27
Air Blank		0.000	21:27
Air Blank		0.000	21:29
Subject Sample	# 2	0.253	21:30
Breath Volume		2.156	21:30
Air Blank		0.000	21:30
Control Test		0.082	21:31
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARRIZA,DAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 04/09/2025 02:24:06

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.080	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.122	02:19
Breath Volume		1.617	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:21
Subject Sample	# 2	0.130	02:22
Breath Volume		1.535	02:22
Air Blank		0.000	02:23
Control Test		0.081	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: BLANKS

DATE/TIME OF TEST: 04/09/2025 22:59:01

Diagnostics Check		OK	22:52
Air Blank		0.000	22:53
Control Test		0.082	22:53
Air Blank		0.000	22:53
Subject Sample	# 1	0.210	22:54
Breath Volume		1.191	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.201	22:57
Breath Volume		1.707	22:57
Air Blank		0.000	22:58
Control Test		0.082	22:58
Air Blank		0.000	22:58
Diagnostics Check		OK	22:59

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GOODPASTER,KALEB
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: MCCANTS

DATE/TIME OF TEST: 04/11/2025 21:35:34

Diagnostics Check		OK	21:29
Air Blank		0.000	21:29
Control Test		0.082	21:30
Air Blank		0.000	21:30
Subject Sample	# 1	0.134	21:31
Breath Volume		1.449	21:31
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.130	21:34
Breath Volume		1.199	21:34
Air Blank		0.000	21:34
Control Test		0.080	21:34
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GOODPASTER,KALEB
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: AVILA JR

DATE/TIME OF TEST: 04/12/2025 01:52:03

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.138	01:47
Breath Volume		3.773	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.141	01:50
Breath Volume		3.335	01:50
Air Blank		0.000	01:51
Control Test		0.081	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:52

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHANFISH,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SIMMS

DATE/TIME OF TEST: 04/13/2025 03:54:07

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.081	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.119	03:49
Breath Volume		1.687	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.121	03:52
Breath Volume		2.097	03:52
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAMMOCK,ANDREW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 04/13/2025 05:45:50

Diagnostics Check		OK	05:39
Air Blank		0.000	05:40
Control Test		0.082	05:40
Air Blank		0.000	05:41
Subject Sample	# 1	0.000	05:41
Breath Volume		1.968	05:41
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.000	05:44
Breath Volume		1.914	05:44
Air Blank		0.000	05:44
Control Test		0.083	05:45
Air Blank		0.000	05:45
Diagnostics Check		OK	05:45

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOLITOR,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: BA

DATE/TIME OF TEST: 04/13/2025 18:41:23

Diagnostics Check		OK	18:28	
Air Blank		0.000	18:29	
Control Test		0.082	18:29	
Air Blank		0.000	18:29	
Subject Sample	# 1	0.255	18:31	
Breath Volume		1.175	18:31	
Air Blank		0.000	18:31	
Air Blank		0.000	18:33	
Subject Sample	# 2	0.192	18:36	Volume Not Met
Breath Volume		0.121	18:36	
Air Blank		0.000	18:37	
Air Blank		0.000	18:39	
Subject Sample	# 3	0.269	18:39	
Breath Volume		1.664	18:39	
Air Blank		0.000	18:40	
Control Test		0.080	18:40	
Air Blank		0.000	18:41	
Diagnostics Check		OK	18:41	

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHANFISH,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: CULVER

DATE/TIME OF TEST: 04/13/2025 23:46:37

Diagnostics Check		OK	23:43	
Air Blank		0.000	23:44	
Control Test		0.083	23:44	
Air Blank		0.000	23:44	
Subject Sample	# 1		23:45	Subject Test Refused
Breath Volume		0.000	23:45	
Air Blank		0.000	23:45	
Control Test		0.082	23:46	
Air Blank		0.000	23:46	
Diagnostics Check		OK	23:46	

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,S,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAZIME,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: SEIGLER

DATE/TIME OF TEST: 04/14/2025 00:50:37

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:48	
Control Test		0.082	00:48	
Air Blank		0.000	00:48	
Subject Sample	# 1		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:49	
Control Test		0.082	00:50	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CONNELL,CHASON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 04/14/2025 07:48:40

Diagnostics Check		OK	07:42
Air Blank		0.000	07:43
Control Test		0.083	07:43
Air Blank		0.000	07:44
Subject Sample	# 1	0.000	07:44
Breath Volume		1.808	07:44
Air Blank		0.000	07:44
Air Blank		0.000	07:46
Subject Sample	# 2	0.000	07:47
Breath Volume		2.136	07:47
Air Blank		0.000	07:47
Control Test		0.081	07:48
Air Blank		0.000	07:48
Diagnostics Check		OK	07:48

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MAMCZUR

DATE/TIME OF TEST: 04/15/2025 02:10:19

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:07	
Control Test		0.082	02:08	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Control Test		0.083	02:09	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNISTER,STEPHEN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BANNISTER,STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 04/15/2025 19:35:41

Diagnostics Check		OK	19:28
Air Blank		0.000	19:29
Control Test		0.082	19:29
Air Blank		0.000	19:30
Subject Sample	# 1	0.187	19:30
Breath Volume		1.984	19:30
Air Blank		0.000	19:31
Air Blank		0.000	19:33
Subject Sample	# 2	0.200	19:34
Breath Volume		1.660	19:34
Air Blank		0.000	19:34
Control Test		0.081	19:35
Air Blank		0.000	19:35
Diagnostics Check		OK	19:35

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONOHUE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: AMEEN

DATE/TIME OF TEST: 04/16/2025 09:19:00

Diagnostics Check		OK	09:12
Air Blank		0.000	09:13
Control Test		0.082	09:13
Air Blank		0.000	09:14
Subject Sample	# 1	0.000	09:14
Breath Volume		1.679	09:14
Air Blank		0.000	09:15
Air Blank		0.000	09:17
Subject Sample	# 2	0.000	09:17
Breath Volume		1.300	09:17
Air Blank		0.000	09:18
Control Test		0.082	09:18
Air Blank		0.000	09:18
Diagnostics Check		OK	09:19

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: LAMBE

DATE/TIME OF TEST: 04/16/2025 21:42:22

Diagnostics Check		OK	21:35
Air Blank		0.000	21:35
Control Test		0.082	21:35
Air Blank		0.000	21:36
Subject Sample	# 1	0.000	21:37
Breath Volume		3.125	21:37
Air Blank		0.000	21:38
Air Blank		0.000	21:40
Subject Sample	# 2	0.000	21:41
Breath Volume		2.503	21:41
Air Blank		0.000	21:41
Control Test		0.082	21:41
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 04/23/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HEYMANS

DATE/TIME OF TEST: 04/18/2025 00:23:06

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.082	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.000	00:18
Breath Volume		3.062	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.000	00:21
Breath Volume		3.320	00:21
Air Blank		0.000	00:22
Control Test		0.083	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RONDOS,R
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 07:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:58

SUBJECT'S LAST NAME: CASTEN

DATE/TIME OF TEST: 04/18/2025 08:38:00

Diagnostics Check		OK	08:35	
Air Blank		0.000	08:35	
Control Test		0.082	08:35	
Air Blank		0.000	08:36	
Subject Sample	# 1		08:36	Subject Test Refused
Breath Volume		0.000	08:36	
Air Blank		0.000	08:37	
Control Test		0.083	08:37	
Air Blank		0.000	08:37	
Diagnostics Check		OK	08:38	

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERYANT,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: DOWD

DATE/TIME OF TEST: 04/18/2025 20:16:43

Diagnostics Check		OK	20:10
Air Blank		0.000	20:10
Control Test		0.082	20:11
Air Blank		0.000	20:11
Subject Sample	# 1	0.202	20:12
Breath Volume		1.226	20:12
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.209	20:15
Breath Volume		1.308	20:15
Air Blank		0.000	20:15
Control Test		0.082	20:16
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: FAVERO

DATE/TIME OF TEST: 04/19/2025 03:03:32

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.083	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.117	02:59
Breath Volume		1.296	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.121	03:02
Breath Volume		1.617	03:02
Air Blank		0.000	03:02
Control Test		0.082	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEWS,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: JUSTO CABANAS

DATE/TIME OF TEST: 04/19/2025 05:21:16

Diagnostics Check		OK	05:14
Air Blank		0.000	05:15
Control Test		0.083	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.151	05:16
Breath Volume		2.285	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.151	05:19
Breath Volume		1.671	05:19
Air Blank		0.000	05:20
Control Test		0.082	05:20
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEISER,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GEISER,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DINDA

DATE/TIME OF TEST: 04/19/2025 21:37:04

Diagnostics Check		OK	21:34	
Air Blank		0.000	21:34	
Control Test		0.083	21:34	
Air Blank		0.000	21:35	
Subject Sample	# 1		21:35	Subject Test Refused
Breath Volume		0.000	21:35	
Air Blank		0.000	21:36	
Control Test		0.082	21:36	
Air Blank		0.000	21:36	
Diagnostics Check		OK	21:37	

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEISER,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRYANT,RENARD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CHIRICO

DATE/TIME OF TEST: 04/19/2025 21:43:18

Diagnostics Check		OK	21:40	
Air Blank		0.000	21:40	
Control Test		0.082	21:41	
Air Blank		0.000	21:41	
Subject Sample	# 1		21:41	Subject Test Refused
Breath Volume		0.000	21:41	
Air Blank		0.000	21:42	
Control Test		0.082	21:42	
Air Blank		0.000	21:43	
Diagnostics Check		OK	21:43	

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: BRANDEL

DATE/TIME OF TEST: 04/20/2025 00:17:19

Diagnostics Check		OK	00:14	
Air Blank		0.000	00:14	
Control Test		0.083	00:15	
Air Blank		0.000	00:15	
Subject Sample	# 1		00:15	Subject Test Refused
Breath Volume		0.000	00:15	
Air Blank		0.000	00:16	
Control Test		0.082	00:16	
Air Blank		0.000	00:17	
Diagnostics Check		OK	00:17	

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEWTON,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: ANDINO

DATE/TIME OF TEST: 04/20/2025 00:57:17

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:54	
Control Test		0.082	00:55	
Air Blank		0.000	00:55	
Subject Sample	# 1		00:55	Subject Test Refused
Breath Volume		0.000	00:55	
Air Blank		0.000	00:56	
Control Test		0.082	00:56	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: SENA

DATE/TIME OF TEST: 04/20/2025 04:17:32

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.082	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.106	04:12
Breath Volume		1.339	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.104	04:16
Breath Volume		1.312	04:16
Air Blank		0.000	04:16
Control Test		0.082	04:16
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LOCKE,ALEXIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: SCHOOLEY

DATE/TIME OF TEST: 04/20/2025 22:35:04

Diagnostics Check		OK	22:32	
Air Blank		0.000	22:32	
Control Test		0.083	22:32	
Air Blank		0.000	22:33	
Subject Sample	# 1		22:33	Subject Test Refused
Breath Volume		0.000	22:33	
Air Blank		0.000	22:34	
Control Test		0.083	22:34	
Air Blank		0.000	22:34	
Diagnostics Check		OK	22:35	

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EXTINE,ALYSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: CASTRO ALTAMIRANO

DATE/TIME OF TEST: 04/20/2025 22:47:44

Diagnostics Check		OK	22:44	
Air Blank		0.000	22:45	
Control Test		0.082	22:45	
Air Blank		0.000	22:46	
Subject Sample	# 1		22:46	Subject Test Refused
Breath Volume		0.000	22:46	
Air Blank		0.000	22:46	
Control Test		0.082	22:47	
Air Blank		0.000	22:47	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: MOON

DATE/TIME OF TEST: 04/22/2025 01:02:01

Diagnostics Check		OK	00:59	
Air Blank		0.000	00:59	
Control Test		0.082	00:59	
Air Blank		0.000	01:00	
Subject Sample	# 1		01:00	Subject Test Refused
Breath Volume		0.000	01:00	
Air Blank		0.000	01:01	
Control Test		0.082	01:01	
Air Blank		0.000	01:01	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HERNANDEZ,LUIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: LACELLE

DATE/TIME OF TEST: 04/23/2025 00:44:27

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:40	
Control Test		0.082	00:41	
Air Blank		0.000	00:41	
Subject Sample	# 1		00:43	Subject Test Refused
Breath Volume		0.000	00:43	
Air Blank		0.000	00:43	
Control Test		0.082	00:43	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: HUTFILZ

DATE/TIME OF TEST: 04/23/2025 01:46:07

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.083	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.150	01:40
Breath Volume		1.417	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.167	01:44
Breath Volume		1.421	01:44
Air Blank		0.000	01:45
Control Test		0.081	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:44

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 04/23/2025 07:16:11

Diagnostics Check		OK	07:04	
Air Blank		0.000	07:05	
Control Test		0.081	07:05	
Air Blank		0.000	07:05	
Subject Sample	# 1	0.158	07:09	Volume Not Met
Breath Volume		0.671	07:09	
Air Blank		0.000	07:09	
Air Blank		0.000	07:11	
Subject Sample	# 2	0.140	07:14	Volume Not Met
Breath Volume		0.363	07:14	
Air Blank		0.000	07:15	
Control Test		0.081	07:15	
Air Blank		0.000	07:16	
Diagnostics Check		OK	07:16	

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:44

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 04/23/2025 07:28:10

Diagnostics Check		OK	07:19
Air Blank		0.000	07:20
Control Test		0.081	07:20
Air Blank		0.000	07:21
Subject Sample	# 1	0.177	07:22
Breath Volume		1.179	07:22
Air Blank		0.000	07:22
Air Blank		0.000	07:24
Subject Sample	# 2	0.175	07:26
Breath Volume		1.433	07:26
Air Blank		0.000	07:27
Control Test		0.081	07:27
Air Blank		0.000	07:28
Diagnostics Check		OK	07:28

Data Download Date/Time: 04/23/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SALZMAN,BLAKE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:11

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/23/2025 10:35:34

Diagnostics Check		OK	10:32	
Air Blank		0.000	10:33	
Control Test		0.082	10:33	
Air Blank		0.000	10:34	
Subject Sample	# 1		10:34	Subject Test Refused
Breath Volume		0.000	10:34	
Air Blank		0.000	10:34	
Control Test		0.082	10:34	
Air Blank		0.000	10:35	
Diagnostics Check		OK	10:35	

Data Download Date/Time: 05/21/2025 09:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARDEN,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 13:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: TOMLINSON

DATE/TIME OF TEST: 04/23/2025 15:06:15

Diagnostics Check		OK	15:03	
Air Blank		0.000	15:03	
Control Test		0.082	15:04	
Air Blank		0.000	15:04	
Subject Sample	# 1		15:04	Subject Test Refused
Breath Volume		0.000	15:04	
Air Blank		0.000	15:05	
Control Test		0.082	15:05	
Air Blank		0.000	15:06	
Diagnostics Check		OK	15:06	

Data Download Date/Time: 05/21/2025 09:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EXCAT,ROSEBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: MARSH-DELK

DATE/TIME OF TEST: 04/23/2025 22:15:24

Diagnostics Check		OK	22:12	
Air Blank		0.000	22:12	
Control Test		0.082	22:13	
Air Blank		0.000	22:13	
Subject Sample	# 1		22:14	Subject Test Refused
Breath Volume		0.000	22:14	
Air Blank		0.000	22:14	
Control Test		0.083	22:14	
Air Blank		0.000	22:15	
Diagnostics Check		OK	22:15	

Data Download Date/Time: 05/21/2025 09:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STACY,CONNER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/23/2025 22:51:03

Diagnostics Check		OK	22:44
Air Blank		0.000	22:45
Control Test		0.082	22:45
Air Blank		0.000	22:46
Subject Sample	# 1	0.127	22:46
Breath Volume		1.765	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:48
Subject Sample	# 2	0.135	22:49
Breath Volume		2.039	22:49
Air Blank		0.000	22:50
Control Test		0.082	22:50
Air Blank		0.000	22:50
Diagnostics Check		OK	22:51

Data Download Date/Time: 05/21/2025 09:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JUNE,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MCCLURE

DATE/TIME OF TEST: 04/26/2025 00:07:31

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.083	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.214	00:02
Breath Volume		1.933	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.211	00:05
Breath Volume		1.941	00:05
Air Blank		0.000	00:06
Control Test		0.081	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 05/21/2025 09:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUZMAN,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: VEIMAU

DATE/TIME OF TEST: 04/26/2025 01:01:37

Diagnostics Check		OK	00:58	
Air Blank		0.000	00:59	
Control Test		0.082	00:59	
Air Blank		0.000	01:00	
Subject Sample	# 1		01:00	Subject Test Refused
Breath Volume		0.000	01:00	
Air Blank		0.000	01:00	
Control Test		0.082	01:01	
Air Blank		0.000	01:01	
Diagnostics Check		OK	01:01	

Data Download Date/Time: 05/21/2025 09:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SEVOSTJANOV,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MENDOZA-BOSCHETTI

DATE/TIME OF TEST: 04/26/2025 02:55:45

Diagnostics Check		OK	02:52	
Air Blank		0.000	02:53	
Control Test		0.082	02:53	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:54	Subject Test Refused
Breath Volume		0.000	02:54	
Air Blank		0.000	02:54	
Control Test		0.081	02:55	
Air Blank		0.000	02:55	
Diagnostics Check		OK	02:55	

Data Download Date/Time: 05/21/2025 09:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKALKO,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: LEO

DATE/TIME OF TEST: 04/26/2025 04:09:55

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.082	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.021	04:05
Breath Volume		1.671	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:07
Subject Sample	# 2	0.020	04:08
Breath Volume		1.750	04:08
Air Blank		0.000	04:09
Control Test		0.082	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 05/21/2025 09:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NURY,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: PEREZ-MARTINEZ

DATE/TIME OF TEST: 04/26/2025 04:41:33

Diagnostics Check		OK	04:38	
Air Blank		0.000	04:39	
Control Test		0.082	04:39	
Air Blank		0.000	04:40	
Subject Sample	# 1		04:40	Subject Test Refused
Breath Volume		0.000	04:40	
Air Blank		0.000	04:40	
Control Test		0.082	04:40	
Air Blank		0.000	04:41	
Diagnostics Check		OK	04:41	

Data Download Date/Time: 05/21/2025 09:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DALTON,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: AFFENDAKES

DATE/TIME OF TEST: 04/26/2025 22:00:11

Diagnostics Check		OK	21:53
Air Blank		0.000	21:54
Control Test		0.082	21:54
Air Blank		0.000	21:55
Subject Sample	# 1	0.241	21:55
Breath Volume		2.031	21:55
Air Blank		0.000	21:56
Air Blank		0.000	21:58
Subject Sample	# 2	0.233	21:58
Breath Volume		1.289	21:58
Air Blank		0.000	21:59
Control Test		0.081	21:59
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 05/21/2025 09:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCFERRIN,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GOMEZ LOPEZ

DATE/TIME OF TEST: 04/27/2025 01:04:06

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.082	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.165	00:59
Breath Volume		1.488	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.164	01:02
Breath Volume		1.437	01:02
Air Blank		0.000	01:03
Control Test		0.082	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 05/21/2025 09:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILSON,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: LUX

DATE/TIME OF TEST: 04/27/2025 02:40:10

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.082	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.239	02:35
Breath Volume		2.410	02:35
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.248	02:38
Breath Volume		2.835	02:38
Air Blank		0.000	02:39
Control Test		0.081	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 05/21/2025 09:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: MARZEC

DATE/TIME OF TEST: 04/27/2025 18:59:18

Diagnostics Check		OK	18:53
Air Blank		0.000	18:53
Control Test		0.082	18:53
Air Blank		0.000	18:54
Subject Sample	# 1	0.000	18:54
Breath Volume		2.234	18:54
Air Blank		0.000	18:55
Air Blank		0.000	18:57
Subject Sample	# 2	0.000	18:57
Breath Volume		2.128	18:57
Air Blank		0.000	18:58
Control Test		0.081	18:58
Air Blank		0.000	18:59
Diagnostics Check		OK	18:59

Data Download Date/Time: 05/21/2025 09:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RIBIC,SENAHID
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/28/2025 00:10:21

Diagnostics Check		OK	00:07	
Air Blank		0.000	00:07	
Control Test		0.083	00:08	
Air Blank		0.000	00:08	
Subject Sample	# 1		00:08	Subject Test Refused
Breath Volume		0.000	00:08	
Air Blank		0.000	00:09	
Control Test		0.082	00:09	
Air Blank		0.000	00:10	
Diagnostics Check		OK	00:10	

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WRIGHT,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: LUDWIG

DATE/TIME OF TEST: 04/28/2025 21:04:34

Diagnostics Check		OK	20:55	
Air Blank		0.000	20:55	
Control Test		0.082	20:55	
Air Blank		0.000	20:56	
Subject Sample	# 1	0.330	20:58	
Breath Volume		1.996	20:58	
Air Blank		0.000	20:59	
Air Blank		0.000	21:01	
Subject Sample	# 2		21:03	Subject Test Refused
Breath Volume		0.402	21:03	
Air Blank		0.000	21:03	
Control Test		0.082	21:03	
Air Blank		0.000	21:04	
Diagnostics Check		OK	21:04	

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BOATENG,DERRICK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: MINOR

DATE/TIME OF TEST: 04/28/2025 22:10:46

Diagnostics Check		OK	22:07	
Air Blank		0.000	22:08	
Control Test		0.082	22:08	
Air Blank		0.000	22:08	
Subject Sample	# 1		22:09	Subject Test Refused
Breath Volume		0.000	22:09	
Air Blank		0.000	22:09	
Control Test		0.082	22:10	
Air Blank		0.000	22:10	
Diagnostics Check		OK	22:10	

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 04/29/2025 03:58:15

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:50	
Control Test		0.083	03:50	
Air Blank		0.000	03:51	
Subject Sample	# 1		03:54	No Sample Provided
Breath Volume		0.000	03:54	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 2		03:56	Subject Test Refused
Breath Volume		0.000	03:56	
Air Blank		0.000	03:57	
Control Test		0.082	03:57	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:58	

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ALVAREZ,JOSE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: GARZON ARDILA

DATE/TIME OF TEST: 04/29/2025 22:25:12

Diagnostics Check		OK	22:19
Air Blank		0.000	22:19
Control Test		0.082	22:19
Air Blank		0.000	22:20
Subject Sample	# 1	0.232	22:20
Breath Volume		2.484	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.237	22:23
Breath Volume		2.222	22:23
Air Blank		0.000	22:24
Control Test		0.082	22:24
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENTLEY,TYLER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BONFERRARO,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 11:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:51

SUBJECT'S LAST NAME: HAMEL

DATE/TIME OF TEST: 04/02/2025 13:18:36

Diagnostics Check		OK	13:12
Air Blank		0.000	13:12
Control Test		0.079	13:12
Air Blank		0.000	13:13
Subject Sample	# 1	0.000	13:14
Breath Volume		4.710	13:14
Air Blank		0.000	13:14
Air Blank		0.000	13:16
Subject Sample	# 2	0.000	13:17
Breath Volume		4.761	13:17
Air Blank		0.000	13:17
Control Test		0.079	13:18
Air Blank		0.000	13:18
Diagnostics Check		OK	13:18

Data Download Date/Time: 04/23/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/03/2025 05:17:50

Diagnostics Check		OK	05:11
Air Blank		0.000	05:11
Control Test		0.079	05:11
Air Blank		0.000	05:12
Subject Sample	# 1	0.175	05:12
Breath Volume		3.855	05:12
Air Blank		0.000	05:13
Air Blank		0.000	05:15
Subject Sample	# 2	0.166	05:16
Breath Volume		2.226	05:16
Air Blank		0.000	05:16
Control Test		0.079	05:17
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 04/23/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: GAMBOA

DATE/TIME OF TEST: 04/05/2025 03:33:50

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.205	03:29
Breath Volume		1.828	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.204	03:32
Breath Volume		1.281	03:32
Air Blank		0.000	03:32
Control Test		0.079	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 04/23/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEELER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TACIA,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 13:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: ZIPFEL

DATE/TIME OF TEST: 04/05/2025 14:16:44

Diagnostics Check		OK	14:12	
Air Blank		0.000	14:13	
Control Test		0.079	14:13	
Air Blank		0.000	14:14	
Subject Sample	# 1		14:15	Subject Test Refused
Breath Volume		0.000	14:15	
Air Blank		0.000	14:15	
Control Test		0.079	14:16	
Air Blank		0.000	14:16	
Diagnostics Check		OK	14:16	

Data Download Date/Time: 04/23/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MASON,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 04/05/2025 19:23:28

Diagnostics Check		OK	19:17
Air Blank		0.000	19:17
Control Test		0.079	19:18
Air Blank		0.000	19:18
Subject Sample	# 1	0.165	19:18
Breath Volume		2.289	19:18
Air Blank		0.000	19:19
Air Blank		0.000	19:21
Subject Sample	# 2	0.154	19:21
Breath Volume		2.761	19:21
Air Blank		0.000	19:22
Control Test		0.078	19:22
Air Blank		0.000	19:23
Diagnostics Check		OK	19:23

Data Download Date/Time: 04/23/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BJURHOLM,JACOB,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BJURHOLM,JACOB
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 04/06/2025 00:08:02

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:05	
Control Test		0.079	00:06	
Air Blank		0.000	00:06	
Subject Sample	# 1		00:06	Subject Test Refused
Breath Volume		0.000	00:06	
Air Blank		0.000	00:07	
Control Test		0.080	00:07	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:08	

Data Download Date/Time: 04/23/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BOATENG,DERRICK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 04/06/2025 04:19:08

Diagnostics Check		OK	04:16	
Air Blank		0.000	04:16	
Control Test		0.079	04:17	
Air Blank		0.000	04:17	
Subject Sample	# 1		04:17	Subject Test Refused
Breath Volume		0.000	04:17	
Air Blank		0.000	04:18	
Control Test		0.079	04:18	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 04/23/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENTLEY,TYLER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LARSON,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 09:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: VLIEK

DATE/TIME OF TEST: 04/07/2025 10:04:18

Diagnostics Check		OK	10:01	
Air Blank		0.000	10:01	
Control Test		0.080	10:02	
Air Blank		0.000	10:02	
Subject Sample	# 1		10:02	Subject Test Refused
Breath Volume		0.000	10:02	
Air Blank		0.000	10:03	
Control Test		0.080	10:03	
Air Blank		0.000	10:04	
Diagnostics Check		OK	10:04	

Data Download Date/Time: 04/23/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: WALTER

DATE/TIME OF TEST: 04/08/2025 05:07:31

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.223	05:00
Breath Volume		2.515	05:00
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.226	05:06
Breath Volume		2.765	05:06
Air Blank		0.000	05:06
Control Test		0.080	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 04/23/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONZON GARCIA,ADEL,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/09/2025 05:08:01

Diagnostics Check		OK	04:54	
Air Blank		0.000	04:54	
Control Test		0.079	04:54	
Air Blank		0.000	04:55	
Subject Sample	# 1	0.007	04:57	
Breath Volume		1.863	04:57	
Air Blank		0.000	04:58	
Air Blank		0.000	05:00	
Subject Sample	# 2	0.007	05:03	Volume Not Met
Breath Volume		0.035	05:03	
Air Blank		0.000	05:04	
Air Blank		0.000	05:06	
Subject Sample	# 3	0.000	05:06	
Breath Volume		1.550	05:06	
Air Blank		0.000	05:07	
Control Test		0.079	05:07	
Air Blank		0.000	05:07	
Diagnostics Check		OK	05:08	

Data Download Date/Time: 04/23/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: BURNSIDE

DATE/TIME OF TEST: 04/10/2025 21:08:39

Diagnostics Check		OK	20:59	
Air Blank		0.000	20:59	
Control Test		0.080	20:59	
Air Blank		0.000	21:00	
Subject Sample	# 1	0.000	21:03	
Breath Volume		1.351	21:03	
Air Blank		0.000	21:03	
Air Blank		0.000	21:05	
Subject Sample	# 2		21:07	Subject Test Refused
Breath Volume		0.000	21:07	
Air Blank		0.000	21:07	
Control Test		0.079	21:08	
Air Blank		0.000	21:08	
Diagnostics Check		OK	21:08	

Data Download Date/Time: 04/23/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAMMOCK,ANDREW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: RIDENOUR

DATE/TIME OF TEST: 04/12/2025 00:36:28

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.080	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.186	00:32
Breath Volume		1.855	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.191	00:34
Breath Volume		1.773	00:34
Air Blank		0.000	00:35
Control Test		0.079	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 04/23/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCFERRIN,JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: CASEY

DATE/TIME OF TEST: 04/12/2025 03:35:19

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:32	
Control Test		0.079	03:32	
Air Blank		0.000	03:33	
Subject Sample	# 1		03:33	Subject Test Refused
Breath Volume		0.000	03:33	
Air Blank		0.000	03:34	
Control Test		0.080	03:34	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 04/23/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENTLEY,TYLER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: OSSIAN,MARK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 13:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:28

SUBJECT'S LAST NAME: CUEVAS

DATE/TIME OF TEST: 04/12/2025 14:53:29

Diagnostics Check		OK	14:48	
Air Blank		0.000	14:49	
Control Test		0.079	14:49	
Air Blank		0.000	14:49	
Subject Sample	# 1		14:52	Subject Test Refused
Breath Volume		0.000	14:52	
Air Blank		0.000	14:52	
Control Test		0.079	14:52	
Air Blank		0.000	14:53	
Diagnostics Check		OK	14:53	

Data Download Date/Time: 04/23/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 04/15/2025 01:06:57

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:04	
Control Test		0.079	01:04	
Air Blank		0.000	01:05	
Subject Sample	# 1		01:05	Subject Test Refused
Breath Volume		0.000	01:05	
Air Blank		0.000	01:06	
Control Test		0.080	01:06	
Air Blank		0.000	01:06	
Diagnostics Check		OK	01:06	

Data Download Date/Time: 04/23/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 04/15/2025 02:53:31

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:51	
Control Test		0.079	02:51	
Air Blank		0.000	02:51	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:52	
Control Test		0.079	02:52	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 04/23/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: WHITTLETON

DATE/TIME OF TEST: 04/18/2025 00:46:10

Diagnostics Check		OK	00:37	
Air Blank		0.000	00:38	
Control Test		0.080	00:38	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:42	No Sample Provided
Breath Volume		0.000	00:42	
Air Blank		0.000	00:42	
Air Blank		0.000	00:44	
Subject Sample	# 2		00:44	Subject Test Refused
Breath Volume		0.000	00:44	
Air Blank		0.000	00:45	
Control Test		0.080	00:45	
Air Blank		0.000	00:46	
Diagnostics Check		OK	00:46	

Data Download Date/Time: 04/23/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOGLER-TORTORA,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:17

SUBJECT'S LAST NAME: NEVINS

DATE/TIME OF TEST: 04/18/2025 11:44:46

Diagnostics Check		OK	11:37
Air Blank		0.000	11:37
Control Test		0.080	11:38
Air Blank		0.000	11:38
Subject Sample	# 1	0.096	11:40
Breath Volume		1.535	11:40
Air Blank		0.000	11:40
Air Blank		0.000	11:42
Subject Sample	# 2	0.097	11:43
Breath Volume		1.554	11:43
Air Blank		0.000	11:43
Control Test		0.079	11:44
Air Blank		0.000	11:44
Diagnostics Check		OK	11:44

Data Download Date/Time: 04/23/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REIDY,CPL,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REIDY,CPL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CUMBO

DATE/TIME OF TEST: 04/20/2025 01:11:30

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.078	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.080	01:06
Breath Volume		1.511	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.087	01:10
Breath Volume		1.816	01:10
Air Blank		0.000	01:10
Control Test		0.080	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 04/23/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MESSENGER,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: MARTINEZ-GARCIA

DATE/TIME OF TEST: 04/20/2025 04:41:23

Diagnostics Check		OK	04:33
Air Blank		0.000	04:34
Control Test		0.080	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.184	04:36
Breath Volume		1.804	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:38
Subject Sample	# 2	0.184	04:39
Breath Volume		1.800	04:39
Air Blank		0.000	04:40
Control Test		0.080	04:40
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 04/23/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 04/24/2025 20:29:38

Diagnostics Check		OK	20:26	
Air Blank		0.000	20:26	
Control Test		0.079	20:27	
Air Blank		0.000	20:27	
Subject Sample	# 1		20:28	Subject Test Refused
Breath Volume		0.000	20:28	
Air Blank		0.000	20:28	
Control Test		0.079	20:29	
Air Blank		0.000	20:29	
Diagnostics Check		OK	20:29	

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,SCOTT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOODY,ANIKA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 07:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:23

SUBJECT'S LAST NAME: DOWNING

DATE/TIME OF TEST: 04/26/2025 09:01:53

Diagnostics Check		OK	08:55
Air Blank		0.000	08:55
Control Test		0.079	08:56
Air Blank		0.000	08:56
Subject Sample	# 1	0.000	08:57
Breath Volume		2.156	08:57
Air Blank		0.000	08:57
Air Blank		0.000	08:59
Subject Sample	# 2	0.000	09:00
Breath Volume		1.679	09:00
Air Blank		0.000	09:00
Control Test		0.079	09:01
Air Blank		0.000	09:01
Diagnostics Check		OK	09:01

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 04/26/2025 20:43:52

Diagnostics Check		OK	20:41	
Air Blank		0.000	20:41	
Control Test		0.078	20:41	
Air Blank		0.000	20:42	
Subject Sample	# 1		20:42	Subject Test Refused
Breath Volume		0.000	20:42	
Air Blank		0.000	20:42	
Control Test		0.079	20:43	
Air Blank		0.000	20:43	
Diagnostics Check		OK	20:43	

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: MCABEE

DATE/TIME OF TEST: 04/27/2025 01:13:19

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.079	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.149	01:08
Breath Volume		3.847	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.145	01:11
Breath Volume		3.378	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTINO,ERIK,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CUTINO,ERIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 04/27/2025 03:03:32

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.079	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.132	02:58
Breath Volume		1.558	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.128	03:02
Breath Volume		1.375	03:02
Air Blank		0.000	03:02
Control Test		0.078	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 04/28/2025 05:17:11

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.079	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.178	05:12
Breath Volume		1.210	05:12
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.179	05:15
Breath Volume		1.851	05:15
Air Blank		0.000	05:16
Control Test		0.079	05:16
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: SAUSEDA,THORNE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: DEAL V

DATE/TIME OF TEST: 04/07/2025 00:02:53

Diagnostics Check		OK	23:56
Air Blank		0.000	23:57
Control Test		0.081	23:57
Air Blank		0.000	23:57
Subject Sample	# 2	0.105	23:58
Breath Volume		1.363	23:58
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.121	00:01
Breath Volume		2.707	00:01
Air Blank		0.000	00:01
Control Test		0.079	00:02
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 04/19/2025 15:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 04/07/2025 02:24:53

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.081	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.137	02:20
Breath Volume		2.683	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.139	02:23
Breath Volume		1.570	02:23
Air Blank		0.000	02:23
Control Test		0.080	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 04/19/2025 15:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CAMPAGONE,ALEXANDER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: DIGIOVANNI

DATE/TIME OF TEST: 04/12/2025 03:37:02

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.081	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.143	03:32
Breath Volume		2.917	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:34
Subject Sample	# 2	0.142	03:35
Breath Volume		2.031	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:37

Data Download Date/Time: 04/19/2025 15:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: O'CONNELL,WILLIAM
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 04/13/2025 02:51:27

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.081	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.104	02:46
Breath Volume		1.308	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.096	02:49
Breath Volume		1.558	02:49
Air Blank		0.000	02:50
Control Test		0.077	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 04/19/2025 15:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MCCARRELL,KYLEE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 04/13/2025 21:50:01

Diagnostics Check		OK	21:43
Air Blank		0.000	21:44
Control Test		0.082	21:44
Air Blank		0.000	21:44
Subject Sample	# 1	0.207	21:45
Breath Volume		3.078	21:45
Air Blank		0.000	21:46
Air Blank		0.000	21:47
Subject Sample	# 2	0.195	21:48
Breath Volume		3.136	21:48
Air Blank		0.000	21:49
Control Test		0.077	21:49
Air Blank		0.000	21:49
Diagnostics Check		OK	21:50

Data Download Date/Time: 04/19/2025 15:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MURPHY,KEVIN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DALRYMPLE

DATE/TIME OF TEST: 04/18/2025 02:55:23

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.082	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.114	02:50
Breath Volume		1.234	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.121	02:53
Breath Volume		1.789	02:53
Air Blank		0.000	02:54
Control Test		0.078	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 04/19/2025 15:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: HOPKINS,ROBERT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: SPARACELLO

DATE/TIME OF TEST: 04/19/2025 23:06:17

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.079	23:00
Air Blank		0.000	23:00
Subject Sample	# 1	0.296	23:01
Breath Volume		1.269	23:01
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.293	23:04
Breath Volume		1.472	23:04
Air Blank		0.000	23:05
Control Test		0.079	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 05/04/2025 07:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,JUSTIN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CARROLL,STEVEN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: CADET

DATE/TIME OF TEST: 04/20/2025 13:30:22

Diagnostics Check		OK	13:23
Air Blank		0.000	13:24
Control Test		0.080	13:24
Air Blank		0.000	13:25
Subject Sample	# 1	0.142	13:25
Breath Volume		3.375	13:25
Air Blank		0.000	13:26
Air Blank		0.000	13:28
Subject Sample	# 2	0.147	13:28
Breath Volume		3.265	13:28
Air Blank		0.000	13:29
Control Test		0.080	13:29
Air Blank		0.000	13:30
Diagnostics Check		OK	13:30

Data Download Date/Time: 05/04/2025 07:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: PROVENCHER,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DADISMAN

DATE/TIME OF TEST: 04/21/2025 21:30:27

Diagnostics Check		OK	21:23
Air Blank		0.000	21:24
Control Test		0.081	21:24
Air Blank		0.000	21:25
Subject Sample	# 1	0.213	21:25
Breath Volume		3.558	21:25
Air Blank		0.000	21:26
Air Blank		0.000	21:28
Subject Sample	# 2	0.208	21:28
Breath Volume		3.406	21:28
Air Blank		0.000	21:29
Control Test		0.079	21:29
Air Blank		0.000	21:30
Diagnostics Check		OK	21:30

Data Download Date/Time: 05/04/2025 07:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,DYLAN,T
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MIXON,GERAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/26/2025 03:23:30

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.081	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.153	03:18
Breath Volume		4.023	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.140	03:21
Breath Volume		1.929	03:21
Air Blank		0.000	03:22
Control Test		0.078	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 05/04/2025 07:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: RICHARDS,JAIDON
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 04/27/2025 02:48:02

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:38	
Control Test		0.079	02:38	
Air Blank		0.000	02:39	
Subject Sample	# 1	0.184	02:42	Volume Not Met
Breath Volume		0.035	02:42	
Air Blank		0.000	02:43	
Air Blank		0.000	02:44	
Subject Sample	# 2		02:46	Subject Test Refused
Breath Volume		0.000	02:46	
Air Blank		0.000	02:47	
Control Test		0.077	02:47	
Air Blank		0.000	02:47	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 05/04/2025 07:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: BELK,ROMONE
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 04/28/2025 00:34:31

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:26	
Control Test		0.081	00:26	
Air Blank		0.000	00:27	
Subject Sample	# 1		00:30	No Sample Provided
Breath Volume		0.000	00:30	
Air Blank		0.000	00:30	
Air Blank		0.000	00:32	
Subject Sample	# 2		00:33	Subject Test Refused
Breath Volume		0.000	00:33	
Air Blank		0.000	00:33	
Control Test		0.081	00:33	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 05/04/2025 07:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEBRETON,QUINN,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: ALFORD,RODNEY
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: DESROSIER

DATE/TIME OF TEST: 04/03/2025 01:27:20

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:18	
Control Test		0.083	01:18	
Air Blank		0.000	01:19	
Subject Sample	# 1	0.097	01:22	Volume Not Met
Breath Volume		0.089	01:22	
Air Blank		0.000	01:23	
Air Blank		0.000	01:25	
Subject Sample	# 2		01:25	Subject Test Refused
Breath Volume		0.000	01:25	
Air Blank		0.000	01:26	
Control Test		0.080	01:26	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 04/16/2025 08:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEBRETON,QUINN,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: ALFORD,RODNEY
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: DESROSIER

DATE/TIME OF TEST: 04/03/2025 01:54:50

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:40	
Control Test		0.082	01:41	
Air Blank		0.000	01:41	
Subject Sample	# 1	0.167	01:44	Volume Not Met
Breath Volume		0.050	01:44	
Air Blank		0.000	01:45	
Air Blank		0.000	01:47	
Subject Sample	# 2	0.156	01:47	
Breath Volume		1.406	01:47	
Air Blank		0.000	01:48	
Air Blank		0.000	01:50	
Subject Sample	# 3	0.141	01:53	Volume Not Met
Breath Volume		0.066	01:53	
Air Blank		0.000	01:53	
Control Test		0.080	01:54	
Air Blank		0.000	01:54	
Diagnostics Check		OK	01:54	

Data Download Date/Time: 04/16/2025 08:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,BRADLEY,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: DAVIS,A
ARRESTING OFFICER AGENCY: ROCKLEDGE PD
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: COMER

DATE/TIME OF TEST: 04/11/2025 01:22:19

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.082	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.100	01:17
Breath Volume		1.226	01:17
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.104	01:20
Breath Volume		1.835	01:20
Air Blank		0.000	01:21
Control Test		0.082	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 04/16/2025 08:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEBRETON,QUINN,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: LEMOINE,TEGAN
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: BERNIER

DATE/TIME OF TEST: 04/14/2025 04:49:30

Diagnostics Check		OK	04:39
Air Blank		0.000	04:40
Control Test		0.083	04:40
Air Blank		0.000	04:41
Subject Sample	# 1	0.129	04:42
Breath Volume		1.957	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.126	04:47
Breath Volume		1.933	04:47
Air Blank		0.000	04:48
Control Test		0.081	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 04/16/2025 08:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: DELACAMPA,B
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/03/2025 00:25:48

Diagnostics Check		OK	00:19
Air Blank		0.000	00:20
Control Test		0.080	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.073	00:21
Breath Volume		2.558	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:23
Subject Sample	# 2	0.073	00:24
Breath Volume		2.429	00:24
Air Blank		0.000	00:24
Control Test		0.080	00:25
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 04/28/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: RINGNALDA

DATE/TIME OF TEST: 04/03/2025 01:23:46

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:21	
Control Test		0.080	01:21	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:22	
Control Test		0.080	01:23	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 04/28/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DONALD,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: LUKIANOV,ALEXANDER
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/12/2025 02:16:42

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.079	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.206	02:08
Breath Volume		2.257	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:10
Subject Sample	# 2	0.170	02:11
Breath Volume		2.445	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 3	0.193	02:15
Breath Volume		1.816	02:15
Air Blank		0.000	02:15
Control Test		0.080	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 04/28/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFEY,KYLE,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: LUKIANOV,ALEXANDER
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BENECOMO

DATE/TIME OF TEST: 04/12/2025 03:10:26

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.081	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.000	03:06
Breath Volume		1.687	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.000	03:09
Breath Volume		2.195	03:09
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 04/28/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: LEHOCZKY

DATE/TIME OF TEST: 04/12/2025 03:46:45

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:44	
Control Test		0.080	03:44	
Air Blank		0.000	03:45	
Subject Sample	# 1		03:45	Subject Test Refused
Breath Volume		0.000	03:45	
Air Blank		0.000	03:45	
Control Test		0.081	03:46	
Air Blank		0.000	03:46	
Diagnostics Check		OK	03:46	

Data Download Date/Time: 04/28/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,ROBERT,J
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: LUKIANOV,ALEXANDER
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 04/12/2025 07:19:13

Diagnostics Check		OK	07:11
Air Blank		0.000	07:11
Control Test		0.080	07:12
Air Blank		0.000	07:12
Subject Sample	# 1	0.144	07:13
Breath Volume		3.148	07:13
Air Blank		0.000	07:14
Air Blank		0.000	07:16
Subject Sample	# 2	0.137	07:17
Breath Volume		2.664	07:17
Air Blank		0.000	07:18
Control Test		0.081	07:18
Air Blank		0.000	07:19
Diagnostics Check		OK	07:19

Data Download Date/Time: 04/28/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 04/02/2025 03:22:54

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.082	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.145	03:17
Breath Volume		1.210	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:19
Subject Sample	# 2	0.148	03:20
Breath Volume		1.574	03:20
Air Blank		0.000	03:21
Control Test		0.078	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: GOVERGO

DATE/TIME OF TEST: 04/02/2025 20:46:33

Diagnostics Check		OK	20:43	
Air Blank		0.000	20:44	
Control Test		0.081	20:44	
Air Blank		0.000	20:44	
Subject Sample	# 1		20:45	Subject Test Refused
Breath Volume		0.000	20:45	
Air Blank		0.000	20:45	
Control Test		0.082	20:45	
Air Blank		0.000	20:46	
Diagnostics Check		OK	20:46	

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THREADFORD,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: BRILL

DATE/TIME OF TEST: 04/04/2025 00:43:58

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.082	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.059	00:39
Breath Volume		3.171	00:39
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.056	00:42
Breath Volume		2.644	00:42
Air Blank		0.000	00:43
Control Test		0.081	00:43
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 04/04/2025 05:09:38

Diagnostics Check		OK	05:06	
Air Blank		0.000	05:07	
Control Test		0.082	05:07	
Air Blank		0.000	05:08	
Subject Sample	# 1		05:08	Subject Test Refused
Breath Volume		0.000	05:08	
Air Blank		0.000	05:08	
Control Test		0.082	05:09	
Air Blank		0.000	05:09	
Diagnostics Check		OK	05:09	

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: BUNTING

DATE/TIME OF TEST: 04/04/2025 23:17:06

Diagnostics Check		OK	23:09
Air Blank		0.000	23:09
Control Test		0.082	23:10
Air Blank		0.000	23:10
Subject Sample	# 1	0.024	23:11
Breath Volume		1.582	23:11
Air Blank		0.000	23:11
Air Blank		0.000	23:13
Subject Sample	# 2	0.024	23:15
Breath Volume		1.699	23:15
Air Blank		0.000	23:16
Control Test		0.081	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: LIGHTSEY

DATE/TIME OF TEST: 04/05/2025 03:59:53

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.081	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.164	03:54
Breath Volume		1.984	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.151	03:58
Breath Volume		1.503	03:58
Air Blank		0.000	03:58
Control Test		0.078	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALLACE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: SPARKS

DATE/TIME OF TEST: 04/06/2025 02:17:47

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.082	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.081	02:13
Breath Volume		1.875	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.075	02:16
Breath Volume		2.578	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 04/06/2025 07:43:47

Diagnostics Check		OK	07:40	
Air Blank		0.000	07:41	
Control Test		0.082	07:41	
Air Blank		0.000	07:42	
Subject Sample	# 1		07:42	Subject Test Refused
Breath Volume		0.000	07:42	
Air Blank		0.000	07:42	
Control Test		0.081	07:43	
Air Blank		0.000	07:43	
Diagnostics Check		OK	07:43	

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: FOREMAN

DATE/TIME OF TEST: 04/08/2025 01:48:45

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:46	
Control Test		0.081	01:46	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:47	Subject Test Refused
Breath Volume		0.000	01:47	
Air Blank		0.000	01:47	
Control Test		0.082	01:48	
Air Blank		0.000	01:48	
Diagnostics Check		OK	01:48	

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BORCHARDT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:23

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 04/08/2025 11:46:44

Diagnostics Check		OK	11:43	
Air Blank		0.000	11:44	
Control Test		0.082	11:44	
Air Blank		0.000	11:45	
Subject Sample	# 1		11:45	Subject Test Refused
Breath Volume		0.000	11:45	
Air Blank		0.000	11:45	
Control Test		0.082	11:46	
Air Blank		0.000	11:46	
Diagnostics Check		OK	11:46	

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 04/09/2025 01:54:18

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:51	
Control Test		0.082	01:52	
Air Blank		0.000	01:52	
Subject Sample	# 1		01:52	Subject Test Refused
Breath Volume		0.000	01:52	
Air Blank		0.000	01:53	
Control Test		0.081	01:53	
Air Blank		0.000	01:54	
Diagnostics Check		OK	01:54	

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/09/2025 05:16:08

Diagnostics Check		OK	05:09
Air Blank		0.000	05:10
Control Test		0.081	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.000	05:11
Breath Volume		2.042	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.000	05:14
Breath Volume		2.578	05:14
Air Blank		0.000	05:15
Control Test		0.081	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: SPRUIELL

DATE/TIME OF TEST: 04/10/2025 23:51:08

Diagnostics Check		OK	23:44
Air Blank		0.000	23:45
Control Test		0.082	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.101	23:46
Breath Volume		2.238	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:48
Subject Sample	# 2	0.095	23:49
Breath Volume		1.488	23:49
Air Blank		0.000	23:50
Control Test		0.080	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 04/11/2025 04:23:34

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:21	
Control Test		0.082	04:21	
Air Blank		0.000	04:22	
Subject Sample	# 1		04:22	Subject Test Refused
Breath Volume		0.000	04:22	
Air Blank		0.000	04:22	
Control Test		0.081	04:22	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 04/11/2025 05:13:52

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:11	
Control Test		0.082	05:11	
Air Blank		0.000	05:12	
Subject Sample	# 1		05:12	Subject Test Refused
Breath Volume		0.000	05:12	
Air Blank		0.000	05:12	
Control Test		0.081	05:13	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: WASSNER

DATE/TIME OF TEST: 04/11/2025 20:21:40

Diagnostics Check		OK	20:15
Air Blank		0.000	20:15
Control Test		0.082	20:15
Air Blank		0.000	20:16
Subject Sample	# 1	0.272	20:16
Breath Volume		2.718	20:16
Air Blank		0.000	20:17
Air Blank		0.000	20:19
Subject Sample	# 2	0.262	20:19
Breath Volume		2.480	20:19
Air Blank		0.000	20:20
Control Test		0.078	20:21
Air Blank		0.000	20:21
Diagnostics Check		OK	20:21

Data Download Date/Time: 04/15/2025 18:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: TRAVIS

DATE/TIME OF TEST: 04/12/2025 04:37:34

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.082	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.174	04:32
Breath Volume		1.628	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.166	04:35
Breath Volume		1.609	04:35
Air Blank		0.000	04:36
Control Test		0.081	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 04/15/2025 18:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MITCHEM

DATE/TIME OF TEST: 04/12/2025 19:59:54

Diagnostics Check		OK	19:53
Air Blank		0.000	19:53
Control Test		0.082	19:54
Air Blank		0.000	19:54
Subject Sample	# 1	0.196	19:55
Breath Volume		3.175	19:55
Air Blank		0.000	19:55
Air Blank		0.000	19:57
Subject Sample	# 2	0.188	19:58
Breath Volume		3.195	19:58
Air Blank		0.000	19:58
Control Test		0.078	19:59
Air Blank		0.000	19:59
Diagnostics Check		OK	19:59

Data Download Date/Time: 04/15/2025 18:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KELLY,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: SINCLAIR VASALLO

DATE/TIME OF TEST: 04/12/2025 23:57:03

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.081	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.189	23:52
Breath Volume		1.515	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.184	23:55
Breath Volume		1.679	23:55
Air Blank		0.000	23:56
Control Test		0.080	23:56
Air Blank		0.000	23:56
Diagnostics Check		OK	23:57

Data Download Date/Time: 04/15/2025 18:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 04/13/2025 04:13:50

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:10	
Control Test		0.082	04:10	
Air Blank		0.000	04:10	
Subject Sample	# 1		04:12	Subject Test Refused
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	
Control Test		0.081	04:13	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 04/15/2025 18:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: CALMO MENDEZ

DATE/TIME OF TEST: 04/13/2025 05:50:25

Diagnostics Check		OK	05:47	
Air Blank		0.000	05:48	
Control Test		0.081	05:48	
Air Blank		0.000	05:48	
Subject Sample	# 1		05:49	Subject Test Refused
Breath Volume		0.000	05:49	
Air Blank		0.000	05:49	
Control Test		0.081	05:49	
Air Blank		0.000	05:50	
Diagnostics Check		OK	05:50	

Data Download Date/Time: 04/15/2025 18:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: THANOS

DATE/TIME OF TEST: 04/13/2025 23:44:58

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:32	
Control Test		0.081	23:33	
Air Blank		0.000	23:33	
Subject Sample	# 1	0.231	23:35	
Breath Volume		2.066	23:35	
Air Blank		0.000	23:36	
Air Blank		0.000	23:37	
Subject Sample	# 2	0.201	23:39	
Breath Volume		1.316	23:39	
Air Blank		0.000	23:40	
Air Blank		0.000	23:41	
Subject Sample	# 3	0.216	23:43	
Breath Volume		1.804	23:43	
Air Blank		0.000	23:44	
Control Test		0.074	23:44	Control Outside Tolerance
Air Blank		0.000	23:44	

Data Download Date/Time: 04/15/2025 18:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: THANOS

DATE/TIME OF TEST: 04/14/2025 00:24:05

Diagnostics Check		OK	00:09	
Air Blank		0.000	00:09	
Control Test		0.080	00:10	
Air Blank		0.000	00:10	
Subject Sample	# 1	0.217	00:11	
Breath Volume		1.539	00:11	
Air Blank		0.000	00:12	
Air Blank		0.000	00:13	
Subject Sample	# 2		00:17	No Sample Provided
Breath Volume		0.000	00:17	
Air Blank		0.000	00:17	
Air Blank		0.000	00:19	
Subject Sample	# 3		00:22	No Sample Provided
Breath Volume		0.000	00:22	
Air Blank		0.000	00:23	
Control Test		0.081	00:23	
Air Blank		0.000	00:24	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 04/15/2025 18:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/14/2025 04:31:09

Diagnostics Check		OK	04:28	
Air Blank		0.000	04:28	
Control Test		0.081	04:29	
Air Blank		0.000	04:29	
Subject Sample	# 1		04:29	Subject Test Refused
Breath Volume		0.000	04:29	
Air Blank		0.000	04:30	
Control Test		0.081	04:30	
Air Blank		0.000	04:31	
Diagnostics Check		OK	04:31	

Data Download Date/Time: 04/15/2025 18:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MUSTAFIC

DATE/TIME OF TEST: 04/15/2025 04:55:00

Diagnostics Check		OK	04:45
Air Blank		0.000	04:45
Control Test		0.081	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.241	04:47
Breath Volume		1.464	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:49
Subject Sample	# 2	0.186	04:50
Breath Volume		1.792	04:50
Air Blank		0.000	04:51
Air Blank		0.000	04:52
Subject Sample	# 3	0.221	04:53
Breath Volume		2.832	04:53
Air Blank		0.000	04:54
Control Test		0.077	04:54
Air Blank		0.000	04:54
Diagnostics Check		OK	04:55

Data Download Date/Time: 04/15/2025 18:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: MILAM

DATE/TIME OF TEST: 04/16/2025 04:48:34

Diagnostics Check		OK	04:45	
Air Blank		0.000	04:46	
Control Test		0.082	04:46	
Air Blank		0.000	04:46	
Subject Sample	# 1		04:47	Subject Test Refused
Breath Volume		0.000	04:47	
Air Blank		0.000	04:47	
Control Test		0.081	04:47	
Air Blank		0.000	04:48	
Diagnostics Check		OK	04:48	

Data Download Date/Time: 05/13/2025 14:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 04/18/2025 02:34:34

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:32	
Control Test		0.082	02:32	
Air Blank		0.000	02:33	
Subject Sample	# 1		02:33	Subject Test Refused
Breath Volume		0.000	02:33	
Air Blank		0.000	02:33	
Control Test		0.082	02:33	
Air Blank		0.000	02:34	
Diagnostics Check		OK	02:34	

Data Download Date/Time: 05/13/2025 14:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RAY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/18/2025 05:11:23

Diagnostics Check		OK	05:03
Air Blank		0.000	05:04
Control Test		0.082	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.115	05:06
Breath Volume		1.257	05:06
Air Blank		0.000	05:06
Air Blank		0.000	05:08
Subject Sample	# 2	0.116	05:09
Breath Volume		1.726	05:09
Air Blank		0.000	05:10
Control Test		0.079	05:10
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,PETER,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/20/2025 02:31:05

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.082	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.129	02:26
Breath Volume		1.222	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:28
Subject Sample	# 2	0.128	02:29
Breath Volume		1.816	02:29
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SEARS,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/20/2025 04:46:18

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:43	
Control Test		0.082	04:44	
Air Blank		0.000	04:44	
Subject Sample	# 1		04:44	Subject Test Refused
Breath Volume		0.000	04:44	
Air Blank		0.000	04:45	
Control Test		0.082	04:45	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: GRIGGS

DATE/TIME OF TEST: 04/22/2025 05:03:12

Diagnostics Check		OK	05:00	
Air Blank		0.000	05:00	
Control Test		0.082	05:01	
Air Blank		0.000	05:01	
Subject Sample	# 1		05:01	Subject Test Refused
Breath Volume		0.000	05:01	
Air Blank		0.000	05:02	
Control Test		0.082	05:02	
Air Blank		0.000	05:03	
Diagnostics Check		OK	05:03	

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 04/24/2025 21:24:11

Diagnostics Check		OK	21:17
Air Blank		0.000	21:18
Control Test		0.082	21:18
Air Blank		0.000	21:19
Subject Sample	# 1	0.063	21:19
Breath Volume		1.714	21:19
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.063	21:22
Breath Volume		1.316	21:22
Air Blank		0.000	21:23
Control Test		0.081	21:23
Air Blank		0.000	21:24
Diagnostics Check		OK	21:24

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALLACE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 04/26/2025 23:24:11

Diagnostics Check		OK	23:16
Air Blank		0.000	23:16
Control Test		0.081	23:17
Air Blank		0.000	23:17
Subject Sample	# 1	0.147	23:19
Breath Volume		1.562	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:21
Subject Sample	# 2	0.144	23:22
Breath Volume		1.671	23:22
Air Blank		0.000	23:23
Control Test		0.077	23:23
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA',GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 04/27/2025 01:23:37

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:21	
Control Test		0.082	01:21	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:22	
Control Test		0.081	01:23	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEFFERSON,JUWAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ECHEVERRIAVIDAL,X
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 04/27/2025 05:04:47

Diagnostics Check		OK	04:57
Air Blank		0.000	04:58
Control Test		0.081	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.194	04:59
Breath Volume		3.046	04:59
Air Blank		0.000	05:01
Air Blank		0.000	05:02
Subject Sample	# 2	0.193	05:02
Breath Volume		2.519	05:02
Air Blank		0.000	05:03
Control Test		0.078	05:04
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 04/28/2025 01:29:13

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.082	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.077	01:24
Breath Volume		1.363	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:26
Subject Sample	# 2	0.070	01:27
Breath Volume		1.527	01:27
Air Blank		0.000	01:28
Control Test		0.077	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RONRUILLE,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: SIBILLE

DATE/TIME OF TEST: 04/28/2025 02:13:49

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:11	
Control Test		0.081	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:12	
Control Test		0.082	02:13	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: SANABRIA BARRERA

DATE/TIME OF TEST: 04/29/2025 18:49:06

Diagnostics Check		OK	18:42
Air Blank		0.000	18:43
Control Test		0.082	18:43
Air Blank		0.000	18:43
Subject Sample	# 1	0.093	18:44
Breath Volume		3.097	18:44
Air Blank		0.000	18:45
Air Blank		0.000	18:46
Subject Sample	# 2	0.091	18:47
Breath Volume		3.246	18:47
Air Blank		0.000	18:48
Control Test		0.080	18:48
Air Blank		0.000	18:49
Diagnostics Check		OK	18:49

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: BRAMLITT

DATE/TIME OF TEST: 04/01/2025 01:56:22

Diagnostics Check		OK	01:47	
Air Blank		0.000	01:48	
Control Test		0.079	01:48	
Air Blank		0.000	01:49	
Subject Sample	# 1		01:52	No Sample Provided
Breath Volume		0.000	01:52	
Air Blank		0.000	01:52	
Air Blank		0.000	01:54	
Subject Sample	# 2		01:55	Subject Test Refused
Breath Volume		0.000	01:55	
Air Blank		0.000	01:55	
Control Test		0.079	01:55	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:56	

Data Download Date/Time: 04/24/2025 09:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: KIGHT,W
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: VERDADERO

DATE/TIME OF TEST: 04/03/2025 03:36:43

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.079	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.107	03:32
Breath Volume		1.347	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.110	03:35
Breath Volume		1.324	03:35
Air Blank		0.000	03:35
Control Test		0.079	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 04/24/2025 09:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: HICKEN

DATE/TIME OF TEST: 04/05/2025 03:41:29

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.080	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.155	03:36
Breath Volume		1.941	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.147	03:40
Breath Volume		2.332	03:40
Air Blank		0.000	03:40
Control Test		0.080	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 04/24/2025 09:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 04/06/2025 05:08:30

Diagnostics Check		OK	05:00	
Air Blank		0.000	05:00	
Control Test		0.079	05:00	
Air Blank		0.000	05:01	
Subject Sample	# 1		05:04	No Sample Provided
Breath Volume		0.000	05:04	
Air Blank		0.000	05:04	
Air Blank		0.000	05:06	
Subject Sample	# 2		05:07	Subject Test Refused
Breath Volume		0.000	05:07	
Air Blank		0.000	05:07	
Control Test		0.079	05:07	
Air Blank		0.000	05:08	
Diagnostics Check		OK	05:08	

Data Download Date/Time: 04/24/2025 09:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: LAST

DATE/TIME OF TEST: 04/07/2025 03:04:00

Diagnostics Check		OK	02:49	
Air Blank		0.000	02:50	
Control Test		0.079	02:50	
Air Blank		0.000	02:51	
Subject Sample	# 1	0.174	02:54	Volume Not Met
Breath Volume		0.671	02:54	
Air Blank		0.000	02:54	
Air Blank		0.000	02:56	
Subject Sample	# 2	0.192	02:58	
Breath Volume		1.863	02:58	
Air Blank		0.000	02:59	
Air Blank		0.000	03:01	
Subject Sample	# 3	0.179	03:02	
Breath Volume		1.625	03:02	
Air Blank		0.000	03:03	
Control Test		0.078	03:03	
Air Blank		0.000	03:03	
Diagnostics Check		OK	03:04	

Data Download Date/Time: 04/24/2025 09:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: KRAUSE,B
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 04/09/2025 22:33:37

Diagnostics Check		OK	22:20	
Air Blank		0.000	22:20	
Control Test		0.080	22:20	
Air Blank		0.000	22:21	
Subject Sample	# 1		22:24	No Sample Provided
Breath Volume		0.000	22:24	
Air Blank		0.000	22:24	
Air Blank		0.000	22:26	
Subject Sample	# 2	0.236	22:28	
Breath Volume		1.273	22:28	
Air Blank		0.000	22:29	
Air Blank		0.000	22:31	
Subject Sample	# 3	0.241	22:32	
Breath Volume		2.375	22:32	
Air Blank		0.000	22:32	
Control Test		0.079	22:33	
Air Blank		0.000	22:33	
Diagnostics Check		OK	22:33	

Data Download Date/Time: 04/24/2025 09:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MARTIN,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: ASCHENBECK

DATE/TIME OF TEST: 04/10/2025 23:34:39

Diagnostics Check		OK	23:25
Air Blank		0.000	23:25
Control Test		0.080	23:25
Air Blank		0.000	23:26
Subject Sample	# 1	0.161	23:29
Breath Volume		2.613	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:31
Subject Sample	# 2	0.159	23:33
Breath Volume		2.324	23:33
Air Blank		0.000	23:33
Control Test		0.080	23:34
Air Blank		0.000	23:34
Diagnostics Check		OK	23:34

Data Download Date/Time: 04/24/2025 09:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MCDADE

DATE/TIME OF TEST: 04/11/2025 03:33:13

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.079	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.159	03:27
Breath Volume		1.507	03:27
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.152	03:31
Breath Volume		1.742	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 04/24/2025 09:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,J,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: NEWTON,N
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 04/11/2025 19:39:25

Diagnostics Check		OK	19:33
Air Blank		0.000	19:33
Control Test		0.079	19:33
Air Blank		0.000	19:34
Subject Sample	# 1	0.210	19:35
Breath Volume		1.402	19:35
Air Blank		0.000	19:35
Air Blank		0.000	19:37
Subject Sample	# 2	0.218	19:37
Breath Volume		1.375	19:37
Air Blank		0.000	19:38
Control Test		0.079	19:38
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 04/24/2025 09:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MARTIN,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: RIOFRIO

DATE/TIME OF TEST: 04/15/2025 05:29:09

Diagnostics Check		OK	05:20	
Air Blank		0.000	05:20	
Control Test		0.080	05:21	
Air Blank		0.000	05:21	
Subject Sample	# 1		05:24	No Sample Provided
Breath Volume		0.000	05:24	
Air Blank		0.000	05:25	
Air Blank		0.000	05:27	
Subject Sample	# 2		05:27	Subject Test Refused
Breath Volume		0.000	05:27	
Air Blank		0.000	05:28	
Control Test		0.080	05:28	
Air Blank		0.000	05:29	
Diagnostics Check		OK	05:29	

Data Download Date/Time: 04/24/2025 09:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SEARS,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 04/16/2025 03:14:55

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.080	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.137	03:09
Breath Volume		1.273	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.126	03:13
Breath Volume		1.378	03:13
Air Blank		0.000	03:14
Control Test		0.079	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 04/24/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MCGRORY

DATE/TIME OF TEST: 04/16/2025 22:20:46

Diagnostics Check		OK	22:13
Air Blank		0.000	22:13
Control Test		0.080	22:13
Air Blank		0.000	22:14
Subject Sample	# 1	0.250	22:15
Breath Volume		1.164	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.247	22:19
Breath Volume		1.742	22:19
Air Blank		0.000	22:19
Control Test		0.080	22:20
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 04/24/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/18/2025 03:26:51

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.080	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.143	03:21
Breath Volume		3.289	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:23
Subject Sample	# 2	0.137	03:25
Breath Volume		2.804	03:25
Air Blank		0.000	03:25
Control Test		0.080	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 04/24/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ROY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BUI,L
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ORTIZ DE JESUS

DATE/TIME OF TEST: 04/18/2025 03:49:15

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.190	03:44
Breath Volume		2.929	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.180	03:47
Breath Volume		2.019	03:47
Air Blank		0.000	03:48
Control Test		0.080	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 04/24/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 04/19/2025 00:57:19

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:49	
Control Test		0.079	00:49	
Air Blank		0.000	00:50	
Subject Sample	# 1		00:53	No Sample Provided
Breath Volume		0.000	00:53	
Air Blank		0.000	00:53	
Air Blank		0.000	00:55	
Subject Sample	# 2		00:55	Subject Test Refused
Breath Volume		0.000	00:55	
Air Blank		0.000	00:56	
Control Test		0.080	00:56	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 04/24/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 04/20/2025 01:13:03

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:04	
Control Test		0.079	01:05	
Air Blank		0.000	01:05	
Subject Sample	# 1		01:08	No Sample Provided
Breath Volume		0.000	01:08	
Air Blank		0.000	01:09	
Air Blank		0.000	01:11	
Subject Sample	# 2		01:11	Subject Test Refused
Breath Volume		0.000	01:11	
Air Blank		0.000	01:12	
Control Test		0.080	01:12	
Air Blank		0.000	01:12	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 04/24/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MARTIN,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MESSER

DATE/TIME OF TEST: 04/20/2025 04:04:25

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:56	
Control Test		0.079	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 1		04:00	No Sample Provided
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Air Blank		0.000	04:02	
Subject Sample	# 2		04:03	Subject Test Refused
Breath Volume		0.000	04:03	
Air Blank		0.000	04:03	
Control Test		0.079	04:03	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 04/24/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ERNEST,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GREEN,S
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:08

SUBJECT'S LAST NAME: REDUBLO

DATE/TIME OF TEST: 04/20/2025 15:59:17

Diagnostics Check		OK	15:47	
Air Blank		0.000	15:47	
Control Test		0.080	15:47	
Air Blank		0.000	15:48	
Subject Sample	# 1	0.322	15:49	
Breath Volume		1.218	15:49	
Air Blank		0.000	15:50	
Air Blank		0.000	15:52	
Subject Sample	# 2	0.291	15:55	
Breath Volume		1.519	15:55	
Air Blank		0.000	15:55	
Air Blank		0.000	15:57	
			15:57	Improper Sample
Air Blank		0.000	15:58	
Control Test		0.079	15:58	
Air Blank		0.000	15:59	
Diagnostics Check		OK	15:59	

Data Download Date/Time: 04/24/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: TECH

DATE/TIME OF TEST: 04/21/2025 03:25:35

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:17	
Control Test		0.080	03:17	
Air Blank		0.000	03:18	
Subject Sample	# 1		03:21	No Sample Provided
Breath Volume		0.000	03:21	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 2		03:24	Subject Test Refused
Breath Volume		0.000	03:24	
Air Blank		0.000	03:24	
Control Test		0.080	03:25	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 04/24/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 04/21/2025 21:20:38

Diagnostics Check		OK	21:13
Air Blank		0.000	21:14
Control Test		0.081	21:14
Air Blank		0.000	21:15
Subject Sample	# 1	0.000	21:16
Breath Volume		1.480	21:16
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.000	21:19
Breath Volume		1.574	21:19
Air Blank		0.000	21:19
Control Test		0.080	21:20
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

Data Download Date/Time: 04/24/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MAYO

DATE/TIME OF TEST: 04/25/2025 03:50:00

Diagnostics Check		OK	03:41	
Air Blank		0.000	03:41	
Control Test		0.080	03:42	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:45	No Sample Provided
Breath Volume		0.000	03:45	
Air Blank		0.000	03:46	
Air Blank		0.000	03:48	
Subject Sample	# 2		03:48	Subject Test Refused
Breath Volume		0.000	03:48	
Air Blank		0.000	03:49	
Control Test		0.080	03:49	
Air Blank		0.000	03:49	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 05/25/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ROY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: STEEL

DATE/TIME OF TEST: 04/27/2025 04:31:49

Diagnostics Check		OK	04:23	
Air Blank		0.000	04:23	
Control Test		0.080	04:24	
Air Blank		0.000	04:24	
Subject Sample	# 1		04:27	No Sample Provided
Breath Volume		0.000	04:27	
Air Blank		0.000	04:28	
Air Blank		0.000	04:30	
Subject Sample	# 2		04:30	Subject Test Refused
Breath Volume		0.000	04:30	
Air Blank		0.000	04:30	
Control Test		0.080	04:31	
Air Blank		0.000	04:31	
Diagnostics Check		OK	04:31	

Data Download Date/Time: 05/25/2025 11:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: EDMOND

DATE/TIME OF TEST: 04/27/2025 22:56:55

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.080	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.000	22:52
Breath Volume		1.605	22:52
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.000	22:55
Breath Volume		1.535	22:55
Air Blank		0.000	22:56
Control Test		0.080	22:56
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 05/25/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 04/28/2025 22:21:16

Diagnostics Check		OK	22:14
Air Blank		0.000	22:14
Control Test		0.080	22:15
Air Blank		0.000	22:15
Subject Sample	# 1	0.204	22:16
Breath Volume		4.179	22:16
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.210	22:19
Breath Volume		2.378	22:19
Air Blank		0.000	22:20
Control Test		0.079	22:20
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 05/25/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MARTIN,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: VILLALONA

DATE/TIME OF TEST: 04/29/2025 04:32:48

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.080	04:27
Air Blank		0.000	04:27
Subject Sample	# 1	0.085	04:28
Breath Volume		2.816	04:28
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.074	04:31
Breath Volume		1.214	04:31
Air Blank		0.000	04:31
Control Test		0.080	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 05/25/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SPRUILL

DATE/TIME OF TEST: 04/29/2025 19:29:46

Diagnostics Check		OK	19:23
Air Blank		0.000	19:23
Control Test		0.080	19:24
Air Blank		0.000	19:24
Subject Sample	# 1	0.254	19:25
Breath Volume		1.785	19:25
Air Blank		0.000	19:25
Air Blank		0.000	19:27
Subject Sample	# 2	0.257	19:28
Breath Volume		2.535	19:28
Air Blank		0.000	19:28
Control Test		0.080	19:29
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

Data Download Date/Time: 05/25/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: WOODARD

DATE/TIME OF TEST: 04/30/2025 03:25:16

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.118	03:20
Breath Volume		2.582	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.115	03:23
Breath Volume		1.671	03:23
Air Blank		0.000	03:24
Control Test		0.080	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 05/25/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,ALEXANDER,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: VAVAL,LUC
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:36

SUBJECT'S LAST NAME: BENNING

DATE/TIME OF TEST: 04/06/2025 11:09:06

Diagnostics Check		OK	11:02
Air Blank		0.000	11:03
Control Test		0.078	11:03
Air Blank		0.000	11:04
Subject Sample	# 1	0.170	11:04
Breath Volume		2.562	11:04
Air Blank		0.000	11:05
Air Blank		0.000	11:07
Subject Sample	# 2	0.165	11:07
Breath Volume		3.332	11:07
Air Blank		0.000	11:08
Control Test		0.078	11:08
Air Blank		0.000	11:09
Diagnostics Check		OK	11:09

Data Download Date/Time: 05/02/2025 13:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,MATTHEW,C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: BRENT,MARK
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 04/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402649340
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: IRIZARRI

DATE/TIME OF TEST: 04/20/2025 23:08:16

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.081	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.223	23:03
Breath Volume		1.917	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.228	23:06
Breath Volume		1.988	23:06
Air Blank		0.000	23:07
Control Test		0.080	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 05/27/2025 12:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 04/02/2025 01:46:08

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.081	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.000	01:41
Breath Volume		1.371	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.000	01:44
Breath Volume		1.843	01:44
Air Blank		0.000	01:45
Control Test		0.081	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 04/17/2025 21:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: CHEATHAM,T
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: VASQUEZ SANCHEZ

DATE/TIME OF TEST: 04/12/2025 04:55:33

Diagnostics Check		OK	04:44	
Air Blank		0.000	04:44	
Control Test		0.079	04:44	
Air Blank		0.000	04:45	
Subject Sample	# 1	0.197	04:48	Volume Not Met
Breath Volume		0.210	04:48	
Air Blank		0.000	04:49	
Air Blank		0.000	04:51	
Subject Sample	# 2		04:54	No Sample Provided
Breath Volume		0.000	04:54	
Air Blank		0.000	04:54	
Control Test		0.076	04:54	
Air Blank		0.000	04:55	
Diagnostics Check		OK	04:55	

Data Download Date/Time: 04/17/2025 21:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: CHEATHAM,T
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 03/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: VASQUEZ SANCHEZ

DATE/TIME OF TEST: 04/12/2025 05:09:14

Diagnostics Check		OK	04:57	
Air Blank		0.000	04:58	
Control Test		0.078	04:58	
Air Blank		0.000	04:59	
Subject Sample	# 1		05:02	No Sample Provided
Breath Volume		0.000	05:02	
Air Blank		0.000	05:02	
Air Blank		0.000	05:04	
Subject Sample	# 2		05:07	No Sample Provided
Breath Volume		0.000	05:07	
Air Blank		0.000	05:08	
Control Test		0.080	05:08	
Air Blank		0.000	05:09	
Diagnostics Check		OK	05:09	

Data Download Date/Time: 04/17/2025 21:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPANA,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: FIVECOAT,JUSTIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: CAIRO

DATE/TIME OF TEST: 04/19/2025 17:00:46

Diagnostics Check		OK	16:54
Air Blank		0.000	16:54
Control Test		0.081	16:55
Air Blank		0.000	16:55
Subject Sample	# 1	0.000	16:56
Breath Volume		2.511	16:56
Air Blank		0.000	16:56
Air Blank		0.000	16:58
Subject Sample	# 2	0.000	16:59
Breath Volume		3.492	16:59
Air Blank		0.000	16:59
Control Test		0.080	17:00
Air Blank		0.000	17:00
Diagnostics Check		OK	17:00

Data Download Date/Time: 05/11/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: PARKER,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: BUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: HERRING

DATE/TIME OF TEST: 04/26/2025 21:24:26

Diagnostics Check		OK	21:17
Air Blank		0.000	21:17
Control Test		0.080	21:18
Air Blank		0.000	21:18
Subject Sample	# 1	0.123	21:20
Breath Volume		3.367	21:20
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.134	21:22
Breath Volume		3.296	21:22
Air Blank		0.000	21:23
Control Test		0.079	21:23
Air Blank		0.000	21:24
Diagnostics Check		OK	21:24

Data Download Date/Time: 05/11/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LISS,K,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAND,D
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: NASH

DATE/TIME OF TEST: 04/26/2025 23:51:51

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:46
Subject Sample	# 1	0.022	23:47
Breath Volume		2.199	23:47
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.018	23:50
Breath Volume		2.421	23:50
Air Blank		0.000	23:50
Control Test		0.080	23:51
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 05/11/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEWS,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: KERSEY

DATE/TIME OF TEST: 04/01/2025 23:34:23

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:28	
Control Test		0.079	23:29	
Air Blank		0.000	23:29	
Subject Sample	# 1	0.162	23:29	
Breath Volume		2.121	23:29	
Air Blank		0.000	23:30	
Air Blank		0.000	23:32	
Subject Sample	# 2		23:32	Subject Test Refused
Breath Volume		0.000	23:32	
Air Blank		0.000	23:33	
Control Test		0.079	23:33	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 04/23/2025 11:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,STEPHAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,STEPHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: PILLION

DATE/TIME OF TEST: 04/05/2025 19:18:55

Diagnostics Check		OK	19:16	
Air Blank		0.000	19:16	
Control Test		0.079	19:16	
Air Blank		0.000	19:17	
Subject Sample	# 1		19:17	Subject Test Refused
Breath Volume		0.000	19:17	
Air Blank		0.000	19:17	
Control Test		0.078	19:18	
Air Blank		0.000	19:18	
Diagnostics Check		OK	19:18	

Data Download Date/Time: 04/23/2025 11:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: RITTER

DATE/TIME OF TEST: 04/05/2025 22:54:30

Diagnostics Check		OK	22:51	
Air Blank		0.000	22:51	
Control Test		0.079	22:52	
Air Blank		0.000	22:52	
Subject Sample	# 1		22:53	Subject Test Refused
Breath Volume		0.000	22:53	
Air Blank		0.000	22:53	
Control Test		0.080	22:53	
Air Blank		0.000	22:54	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 04/23/2025 11:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHRAMM,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 04/05/2025 23:01:26

Diagnostics Check		OK	22:58	
Air Blank		0.000	22:59	
Control Test		0.080	22:59	
Air Blank		0.000	22:59	
Subject Sample	# 1		23:00	Subject Test Refused
Breath Volume		0.000	23:00	
Air Blank		0.000	23:00	
Control Test		0.080	23:00	
Air Blank		0.000	23:01	
Diagnostics Check		OK	23:01	

Data Download Date/Time: 04/23/2025 11:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: RITTER

DATE/TIME OF TEST: 04/05/2025 23:52:18

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:47	
Control Test		0.079	23:48	
Air Blank		0.000	23:48	
Subject Sample	# 1		23:50	Subject Test Refused
Breath Volume		0.000	23:50	
Air Blank		0.000	23:51	
Control Test		0.079	23:51	
Air Blank		0.000	23:52	
Diagnostics Check		OK	23:52	

Data Download Date/Time: 04/23/2025 11:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: KOZAKEWICH

DATE/TIME OF TEST: 04/11/2025 02:26:24

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:24	
Control Test		0.079	02:24	
Air Blank		0.000	02:24	
Subject Sample	# 1		02:25	Subject Test Refused
Breath Volume		0.000	02:25	
Air Blank		0.000	02:25	
Control Test		0.080	02:25	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 04/23/2025 11:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 04/16/2025 04:59:11

Diagnostics Check		OK	04:56	
Air Blank		0.000	04:56	
Control Test		0.079	04:57	
Air Blank		0.000	04:57	
Subject Sample	# 1		04:57	Subject Test Refused
Breath Volume		0.000	04:57	
Air Blank		0.000	04:58	
Control Test		0.080	04:58	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 04/23/2025 11:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/19/2025 22:45:47

Diagnostics Check		OK	22:36
Air Blank		0.000	22:37
Control Test		0.080	22:37
Air Blank		0.000	22:37
Subject Sample	# 1	0.262	22:38
Breath Volume		1.906	22:38
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.241	22:41
Breath Volume		1.300	22:41
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 3	0.240	22:44
Breath Volume		1.609	22:44
Air Blank		0.000	22:44
Control Test		0.080	22:45
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 04/23/2025 11:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: RUIZ FRIAS

DATE/TIME OF TEST: 04/04/2025 19:12:13

Diagnostics Check		OK	19:05
Air Blank		0.000	19:06
Control Test		0.081	19:06
Air Blank		0.000	19:06
Subject Sample	# 1	0.193	19:07
Breath Volume		2.527	19:07
Air Blank		0.000	19:08
Air Blank		0.000	19:10
Subject Sample	# 2	0.199	19:10
Breath Volume		2.000	19:10
Air Blank		0.000	19:11
Control Test		0.081	19:11
Air Blank		0.000	19:12
Diagnostics Check		OK	19:12

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHANDLER,RYAN,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CHANDLER,RYAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 04/05/2025 04:38:10

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.081	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.086	04:33
Breath Volume		1.710	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.084	04:36
Breath Volume		1.988	04:36
Air Blank		0.000	04:37
Control Test		0.080	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHANDLER,RYAN,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: TIMBROOK,RICKY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: SUPER

DATE/TIME OF TEST: 04/05/2025 04:55:03

Diagnostics Check		OK	04:48
Air Blank		0.000	04:49
Control Test		0.081	04:49
Air Blank		0.000	04:50
Subject Sample	# 1	0.137	04:50
Breath Volume		2.917	04:50
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.141	04:53
Breath Volume		2.667	04:53
Air Blank		0.000	04:54
Control Test		0.081	04:54
Air Blank		0.000	04:54
Diagnostics Check		OK	04:55

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: STUPPI JR

DATE/TIME OF TEST: 04/05/2025 21:00:34

Diagnostics Check		OK	20:54
Air Blank		0.000	20:54
Control Test		0.081	20:54
Air Blank		0.000	20:55
Subject Sample	# 1	0.025	20:56
Breath Volume		2.125	20:56
Air Blank		0.000	20:56
Air Blank		0.000	20:58
Subject Sample	# 2	0.023	20:59
Breath Volume		1.332	20:59
Air Blank		0.000	20:59
Control Test		0.081	20:59
Air Blank		0.000	21:00
Diagnostics Check		OK	21:00

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: ROJAS SANTIAGO

DATE/TIME OF TEST: 04/06/2025 03:14:04

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.081	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.099	03:09
Breath Volume		1.542	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.100	03:12
Breath Volume		1.773	03:12
Air Blank		0.000	03:13
Control Test		0.081	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:14

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: CRAYTON JR

DATE/TIME OF TEST: 04/09/2025 22:41:38

Diagnostics Check		OK	22:35
Air Blank		0.000	22:35
Control Test		0.082	22:36
Air Blank		0.000	22:36
Subject Sample	# 1	0.027	22:37
Breath Volume		1.160	22:37
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.023	22:40
Breath Volume		1.605	22:40
Air Blank		0.000	22:40
Control Test		0.081	22:41
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: KOSIANOWSKI

DATE/TIME OF TEST: 04/12/2025 21:19:01

Diagnostics Check		OK	21:12
Air Blank		0.000	21:12
Control Test		0.081	21:13
Air Blank		0.000	21:13
Subject Sample	# 1	0.263	21:14
Breath Volume		1.503	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:16
Subject Sample	# 2	0.247	21:17
Breath Volume		1.128	21:17
Air Blank		0.000	21:18
Control Test		0.081	21:18
Air Blank		0.000	21:18
Diagnostics Check		OK	21:19

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CORRALES,C
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: PEKARSKI

DATE/TIME OF TEST: 04/13/2025 05:57:04

Diagnostics Check		OK	05:50
Air Blank		0.000	05:50
Control Test		0.081	05:51
Air Blank		0.000	05:51
Subject Sample	# 1	0.100	05:52
Breath Volume		3.277	05:52
Air Blank		0.000	05:52
Air Blank		0.000	05:54
Subject Sample	# 2	0.090	05:55
Breath Volume		1.429	05:55
Air Blank		0.000	05:56
Control Test		0.081	05:56
Air Blank		0.000	05:56
Diagnostics Check		OK	05:57

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: LOPES

DATE/TIME OF TEST: 04/14/2025 01:03:29

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.081	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.197	00:58
Breath Volume		1.449	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.189	01:01
Breath Volume		1.421	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/18/2025 02:56:38

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.081	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.205	02:51
Breath Volume		2.039	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.203	02:55
Breath Volume		1.492	02:55
Air Blank		0.000	02:55
Control Test		0.081	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: DONU-GRANADINO

DATE/TIME OF TEST: 04/20/2025 04:06:09

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.154	04:01
Breath Volume		1.761	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.164	04:04
Breath Volume		2.542	04:04
Air Blank		0.000	04:05
Control Test		0.082	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MORALES,LAUREN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/20/2025 04:37:27

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.082	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.042	04:32
Breath Volume		1.898	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.042	04:35
Breath Volume		1.433	04:35
Air Blank		0.000	04:36
Control Test		0.082	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: CAMARGO

DATE/TIME OF TEST: 04/25/2025 22:02:09

Diagnostics Check		OK	21:55
Air Blank		0.000	21:55
Control Test		0.081	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.159	21:57
Breath Volume		3.199	21:57
Air Blank		0.000	21:57
Air Blank		0.000	21:59
Subject Sample	# 2	0.152	22:00
Breath Volume		3.796	22:00
Air Blank		0.000	22:01
Control Test		0.082	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: PERZ-YERBAFRIA

DATE/TIME OF TEST: 04/26/2025 23:41:24

Diagnostics Check		OK	23:31
Air Blank		0.000	23:31
Control Test		0.081	23:32
Air Blank		0.000	23:32
Subject Sample	# 1	0.144	23:35
Breath Volume		1.589	23:35
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.140	23:39
Breath Volume		1.851	23:39
Air Blank		0.000	23:40
Control Test		0.083	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LIGHTFOOT,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: OTHER
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 04/28/2025 01:06:57

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.082	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.000	01:02
Breath Volume		2.605	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.000	01:05
Breath Volume		1.781	01:05
Air Blank		0.000	01:06
Control Test		0.081	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 05/03/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,N,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 04/01/2025 02:55:45

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.082	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.150	02:49
Breath Volume		1.132	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.149	02:54
Breath Volume		1.457	02:54
Air Blank		0.000	02:54
Control Test		0.081	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 04/15/2025 20:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: BAHR

DATE/TIME OF TEST: 04/02/2025 01:07:11

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:04	
Control Test		0.081	01:05	
Air Blank		0.000	01:05	
Subject Sample	# 1		01:05	Subject Test Refused
Breath Volume		0.000	01:05	
Air Blank		0.000	01:06	
Control Test		0.081	01:06	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:07	

Data Download Date/Time: 04/15/2025 20:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: COTHREN

DATE/TIME OF TEST: 04/02/2025 03:25:49

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:23	
Control Test		0.081	03:23	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:24	Subject Test Refused
Breath Volume		0.000	03:24	
Air Blank		0.000	03:24	
Control Test		0.080	03:25	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 04/15/2025 20:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROSSI,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 04/03/2025 05:47:18

Diagnostics Check		OK	05:37
Air Blank		0.000	05:38
Control Test		0.082	05:38
Air Blank		0.000	05:38
Subject Sample	# 1	0.134	05:39
Breath Volume		2.105	05:39
Air Blank		0.000	05:40
Air Blank		0.000	05:42
Subject Sample	# 2	0.130	05:45
Breath Volume		1.203	05:45
Air Blank		0.000	05:46
Control Test		0.081	05:46
Air Blank		0.000	05:47
Diagnostics Check		OK	05:47

Data Download Date/Time: 04/15/2025 20:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNIGHT,ADAM,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 04/04/2025 05:42:30

Diagnostics Check		OK	05:33
Air Blank		0.000	05:34
Control Test		0.082	05:34
Air Blank		0.000	05:35
Subject Sample	# 1	0.144	05:36
Breath Volume		2.156	05:36
Air Blank		0.000	05:37
Air Blank		0.000	05:39
Subject Sample	# 2	0.145	05:40
Breath Volume		1.890	05:40
Air Blank		0.000	05:41
Control Test		0.079	05:41
Air Blank		0.000	05:42
Diagnostics Check		OK	05:42

Data Download Date/Time: 04/15/2025 20:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: SINGER

DATE/TIME OF TEST: 04/04/2025 17:05:40

Diagnostics Check		OK	16:51	
Air Blank		0.000	16:51	
Control Test		0.081	16:52	
Air Blank		0.000	16:52	
Subject Sample	# 1	0.156	16:55	Volume Not Met
Breath Volume		0.781	16:55	
Air Blank		0.000	16:56	
Air Blank		0.000	16:58	
Subject Sample	# 2	0.156	16:59	
Breath Volume		1.296	16:59	
Air Blank		0.000	16:59	
Air Blank		0.000	17:01	
Subject Sample	# 3	0.151	17:03	
Breath Volume		1.402	17:03	
Air Blank		0.000	17:04	
Control Test		0.079	17:04	
Air Blank		0.000	17:05	
Diagnostics Check		OK	17:05	

Data Download Date/Time: 04/15/2025 20:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,I
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: POOLE

DATE/TIME OF TEST: 04/05/2025 03:52:03

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.081	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.208	03:46
Breath Volume		1.101	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.217	03:50
Breath Volume		1.250	03:50
Air Blank		0.000	03:51
Control Test		0.080	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:52

Data Download Date/Time: 04/15/2025 20:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: ZILINSKI

DATE/TIME OF TEST: 04/05/2025 21:47:38

Diagnostics Check		OK	21:39
Air Blank		0.000	21:40
Control Test		0.081	21:40
Air Blank		0.000	21:41
Subject Sample	# 1	0.185	21:42
Breath Volume		2.265	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.185	21:45
Breath Volume		1.761	21:45
Air Blank		0.000	21:46
Control Test		0.079	21:47
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 04/15/2025 20:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 04/06/2025 00:08:27

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.081	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.096	00:03
Breath Volume		2.031	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.091	00:06
Breath Volume		2.113	00:06
Air Blank		0.000	00:07
Control Test		0.080	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 04/15/2025 20:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEIS,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: BOGAN

DATE/TIME OF TEST: 04/06/2025 01:11:09

Diagnostics Check		OK	01:08	
Air Blank		0.000	01:08	
Control Test		0.081	01:09	
Air Blank		0.000	01:09	
Subject Sample	# 1		01:09	Subject Test Refused
Breath Volume		0.000	01:09	
Air Blank		0.000	01:10	
Control Test		0.080	01:10	
Air Blank		0.000	01:11	
Diagnostics Check		OK	01:11	

Data Download Date/Time: 04/15/2025 20:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROSSI,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/06/2025 04:37:36

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.126	04:32
Breath Volume		1.593	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.136	04:35
Breath Volume		1.480	04:35
Air Blank		0.000	04:36
Control Test		0.080	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 04/15/2025 20:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: BARTHOLOMEW

DATE/TIME OF TEST: 04/06/2025 21:41:25

Diagnostics Check		OK	21:35
Air Blank		0.000	21:35
Control Test		0.081	21:35
Air Blank		0.000	21:36
Subject Sample	# 1	0.180	21:36
Breath Volume		2.437	21:36
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.182	21:39
Breath Volume		2.535	21:39
Air Blank		0.000	21:40
Control Test		0.080	21:40
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 04/15/2025 20:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEIS,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: TUNSILL

DATE/TIME OF TEST: 04/07/2025 00:29:09

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.081	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.127	00:24
Breath Volume		1.503	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:26
Subject Sample	# 2	0.129	00:27
Breath Volume		1.136	00:27
Air Blank		0.000	00:28
Control Test		0.079	00:28
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 04/15/2025 20:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROSADO,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 04/07/2025 06:48:19

Diagnostics Check		OK	06:41
Air Blank		0.000	06:42
Control Test		0.081	06:42
Air Blank		0.000	06:42
Subject Sample	# 1	0.123	06:43
Breath Volume		2.011	06:43
Air Blank		0.000	06:44
Air Blank		0.000	06:46
Subject Sample	# 2	0.125	06:46
Breath Volume		1.660	06:46
Air Blank		0.000	06:47
Control Test		0.081	06:47
Air Blank		0.000	06:48
Diagnostics Check		OK	06:48

Data Download Date/Time: 04/15/2025 20:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: PATTILLO

DATE/TIME OF TEST: 04/07/2025 19:21:29

Diagnostics Check		OK	19:13
Air Blank		0.000	19:14
Control Test		0.081	19:14
Air Blank		0.000	19:15
Subject Sample	# 1	0.000	19:16
Breath Volume		1.425	19:16
Air Blank		0.000	19:17
Air Blank		0.000	19:19
Subject Sample	# 2	0.000	19:20
Breath Volume		1.332	19:20
Air Blank		0.000	19:20
Control Test		0.080	19:20
Air Blank		0.000	19:21
Diagnostics Check		OK	19:21

Data Download Date/Time: 04/15/2025 20:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: BOWENS

DATE/TIME OF TEST: 04/11/2025 03:45:38

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.082	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.169	03:40
Breath Volume		1.171	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.176	03:43
Breath Volume		1.359	03:43
Air Blank		0.000	03:44
Control Test		0.080	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 04/15/2025 20:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRAMS,JADA,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BYRAMS,JADA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: CRANE

DATE/TIME OF TEST: 04/01/2025 21:45:25

Diagnostics Check		OK	21:31	
Air Blank		0.000	21:31	
Control Test		0.080	21:32	
Air Blank		0.000	21:32	
Subject Sample	# 1	0.193	21:35	Volume Not Met
Breath Volume		0.242	21:35	
Air Blank		0.000	21:36	
Air Blank		0.000	21:38	
Subject Sample	# 2	0.190	21:40	
Breath Volume		1.613	21:40	
Air Blank		0.000	21:41	
Air Blank		0.000	21:43	
Subject Sample	# 3	0.191	21:43	
Breath Volume		1.363	21:43	
Air Blank		0.000	21:44	
Control Test		0.078	21:44	
Air Blank		0.000	21:45	
Diagnostics Check		OK	21:45	

Data Download Date/Time: 04/30/2025 10:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: EMBURY

DATE/TIME OF TEST: 04/01/2025 23:40:09

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:37	
Control Test		0.077	23:38	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:38	RFI Detect
Breath Volume		0.000	23:38	
Air Blank		0.000	23:39	
Control Test		0.077	23:39	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 04/30/2025 10:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: EMBURY

DATE/TIME OF TEST: 04/01/2025 23:49:04

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.078	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.257	23:43
Breath Volume		2.332	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.258	23:47
Breath Volume		1.570	23:47
Air Blank		0.000	23:48
Control Test		0.079	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:49

Data Download Date/Time: 04/30/2025 10:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,WILLIAM,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: WATSON,WILLIAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/03/2025 23:51:53

Diagnostics Check		OK	23:45
Air Blank		0.000	23:45
Control Test		0.080	23:46
Air Blank		0.000	23:46
Subject Sample	# 1	0.174	23:47
Breath Volume		1.941	23:47
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.185	23:50
Breath Volume		2.179	23:50
Air Blank		0.000	23:50
Control Test		0.079	23:51
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 04/30/2025 10:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONS,EVAN,N
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FREDRICKSON,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 04/05/2025 04:15:10

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.149	04:10
Breath Volume		2.093	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.153	04:13
Breath Volume		1.542	04:13
Air Blank		0.000	04:14
Control Test		0.077	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 04/30/2025 10:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,WILLIAM,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HOWARD,JESSE
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/06/2025 18:16:29

Diagnostics Check		OK	18:09
Air Blank		0.000	18:10
Control Test		0.080	18:10
Air Blank		0.000	18:11
Subject Sample	# 1	0.192	18:12
Breath Volume		2.523	18:12
Air Blank		0.000	18:12
Air Blank		0.000	18:14
Subject Sample	# 2	0.198	18:14
Breath Volume		2.812	18:14
Air Blank		0.000	18:15
Control Test		0.079	18:15
Air Blank		0.000	18:16
Diagnostics Check		OK	18:16

Data Download Date/Time: 04/30/2025 10:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,R,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 04/15/2025 23:59:49

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.080	23:54
Air Blank		0.000	23:54
Subject Sample	# 1	0.234	23:55
Breath Volume		2.402	23:55
Air Blank		0.000	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.235	23:58
Breath Volume		1.511	23:58
Air Blank		0.000	23:58
Control Test		0.079	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 04/30/2025 10:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAINRIGHT,DAVID,T
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: OLIVA,PETER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 04/05/2025 01:10:37

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.081	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.191	01:06
Breath Volume		1.539	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.174	01:09
Breath Volume		1.109	01:09
Air Blank		0.000	01:09
Control Test		0.081	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 04/23/2025 11:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALDONADO,SOLOMON,
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: AROCHO,J
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 04/07/2025 07:55:20

Diagnostics Check		OK	07:49
Air Blank		0.000	07:49
Control Test		0.080	07:50
Air Blank		0.000	07:50
Subject Sample	# 1	0.000	07:51
Breath Volume		2.367	07:51
Air Blank		0.000	07:51
Air Blank		0.000	07:53
Subject Sample	# 2	0.000	07:53
Breath Volume		2.222	07:53
Air Blank		0.000	07:54
Control Test		0.080	07:54
Air Blank		0.000	07:55
Diagnostics Check		OK	07:55

Data Download Date/Time: 04/23/2025 11:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: FLANAGAN

DATE/TIME OF TEST: 04/03/2025 02:44:40

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:40	
Control Test		0.080	02:41	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:43	Subject Test Refused
Breath Volume		0.000	02:43	
Air Blank		0.000	02:43	
Control Test		0.080	02:44	
Air Blank		0.000	02:44	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 04/03/2025 02:51:27

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:47	
Control Test		0.080	02:47	
Air Blank		0.000	02:48	
Subject Sample	# 1		02:50	Subject Test Refused
Breath Volume		0.000	02:50	
Air Blank		0.000	02:50	
Control Test		0.080	02:50	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEUA,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:23

SUBJECT'S LAST NAME: STORIE

DATE/TIME OF TEST: 04/03/2025 11:53:05

Diagnostics Check		OK	11:47
Air Blank		0.000	11:47
Control Test		0.081	11:47
Air Blank		0.000	11:48
Subject Sample	# 1	0.000	11:48
Breath Volume		2.097	11:48
Air Blank		0.000	11:49
Air Blank		0.000	11:51
Subject Sample	# 2	0.000	11:51
Breath Volume		1.742	11:51
Air Blank		0.000	11:52
Control Test		0.080	11:52
Air Blank		0.000	11:52
Diagnostics Check		OK	11:53

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:26

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 04/03/2025 14:49:23

Diagnostics Check		OK	14:46	
Air Blank		0.000	14:46	
Control Test		0.081	14:47	
Air Blank		0.000	14:47	
Subject Sample	# 1		14:48	Subject Test Refused
Breath Volume		0.000	14:48	
Air Blank		0.000	14:48	
Control Test		0.081	14:48	
Air Blank		0.000	14:49	
Diagnostics Check		OK	14:49	

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GADDIS,J
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: SCHROEDER

DATE/TIME OF TEST: 04/04/2025 02:19:49

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.081	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.000	02:15
Breath Volume		2.257	02:15
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.000	02:18
Breath Volume		2.007	02:18
Air Blank		0.000	02:18
Control Test		0.080	02:19
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 04/04/2025 03:26:13

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:18	
Control Test		0.080	03:18	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:22	No Sample Provided
Breath Volume		0.000	03:22	
Air Blank		0.000	03:22	
Air Blank		0.000	03:24	
Subject Sample	# 2		03:24	Subject Test Refused
Breath Volume		0.000	03:24	
Air Blank		0.000	03:25	
Control Test		0.080	03:25	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: FRICKE

DATE/TIME OF TEST: 04/04/2025 22:46:01

Diagnostics Check		OK	22:32	
Air Blank		0.000	22:33	
Control Test		0.082	22:33	
Air Blank		0.000	22:34	
Subject Sample	# 1		22:37	No Sample Provided
Breath Volume		0.000	22:37	
Air Blank		0.000	22:37	
Air Blank		0.000	22:39	
Subject Sample	# 2	0.186	22:41	
Breath Volume		1.597	22:41	
Air Blank		0.000	22:41	
Air Blank		0.000	22:43	
Subject Sample	# 3	0.184	22:44	
Breath Volume		1.578	22:44	
Air Blank		0.000	22:45	
Control Test		0.079	22:45	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/06/2025 02:03:03

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:00	
Control Test		0.080	02:00	
Air Blank		0.000	02:01	
Subject Sample	# 1		02:01	Subject Test Refused
Breath Volume		0.000	02:01	
Air Blank		0.000	02:02	
Control Test		0.080	02:02	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: KELSEY

DATE/TIME OF TEST: 04/06/2025 02:32:57

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.000	02:28
Breath Volume		2.253	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.000	02:31
Breath Volume		2.085	02:31
Air Blank		0.000	02:32
Control Test		0.079	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTI,JORDAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LUCEY,CONOR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: NIYONSABA

DATE/TIME OF TEST: 04/07/2025 22:17:36

Diagnostics Check		OK	22:08	
Air Blank		0.000	22:09	
Control Test		0.081	22:09	
Air Blank		0.000	22:09	
Subject Sample	# 1		22:13	No Sample Provided
Breath Volume		0.000	22:13	
Air Blank		0.000	22:13	
Air Blank		0.000	22:15	
Subject Sample	# 2		22:16	Subject Test Refused
Breath Volume		0.000	22:16	
Air Blank		0.000	22:16	
Control Test		0.080	22:16	
Air Blank		0.000	22:17	
Diagnostics Check		OK	22:17	

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/09/2025 01:20:29

Diagnostics Check		OK	01:12	
Air Blank		0.000	01:12	
Control Test		0.081	01:12	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:16	No Sample Provided
Breath Volume		0.000	01:16	
Air Blank		0.000	01:16	
Air Blank		0.000	01:18	
Subject Sample	# 2		01:19	Subject Test Refused
Breath Volume		0.000	01:19	
Air Blank		0.000	01:19	
Control Test		0.080	01:19	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:20	

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 04/09/2025 01:29:32

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.081	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.000	01:25
Breath Volume		2.082	01:25
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.000	01:28
Breath Volume		1.464	01:28
Air Blank		0.000	01:28
Control Test		0.081	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 04/09/2025 03:01:17

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.216	02:56
Breath Volume		2.058	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.210	02:59
Breath Volume		2.031	02:59
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BISHOP,OFC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: RESTREPO

DATE/TIME OF TEST: 04/09/2025 18:42:47

Diagnostics Check		OK	18:35
Air Blank		0.000	18:36
Control Test		0.080	18:36
Air Blank		0.000	18:37
Subject Sample	# 1	0.227	18:38
Breath Volume		2.007	18:38
Air Blank		0.000	18:38
Air Blank		0.000	18:40
Subject Sample	# 2	0.241	18:41
Breath Volume		1.160	18:41
Air Blank		0.000	18:41
Control Test		0.080	18:42
Air Blank		0.000	18:42
Diagnostics Check		OK	18:42

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 04/10/2025 01:54:02

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.081	01:48
Air Blank		0.000	01:48
Subject Sample	# 1	0.000	01:49
Breath Volume		2.964	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.000	01:52
Breath Volume		2.640	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:54

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 04/10/2025 03:45:26

Diagnostics Check		OK	03:39	
Air Blank		0.000	03:40	
Control Test		0.081	03:40	
Air Blank		0.000	03:41	
Subject Sample	# 1		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:44	
Control Test		0.080	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 04/10/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: CORONADO

DATE/TIME OF TEST: 04/10/2025 22:10:17

Diagnostics Check		OK	22:03
Air Blank		0.000	22:04
Control Test		0.081	22:04
Air Blank		0.000	22:04
Subject Sample	# 1	0.342	22:05
Breath Volume		3.175	22:05
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.360	22:08
Breath Volume		3.058	22:08
Air Blank		0.000	22:09
Control Test		0.079	22:09
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 05/01/2025 14:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/11/2025 00:28:19

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:24	
Control Test		0.080	00:24	
Air Blank		0.000	00:25	
Subject Sample	# 1		00:26	Subject Test Refused
Breath Volume		0.000	00:26	
Air Blank		0.000	00:27	
Control Test		0.080	00:27	
Air Blank		0.000	00:28	
Diagnostics Check		OK	00:28	

Data Download Date/Time: 05/01/2025 14:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: VENTURA

DATE/TIME OF TEST: 04/11/2025 22:04:20

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:56	
Control Test		0.080	21:56	
Air Blank		0.000	21:56	
Subject Sample	# 1		22:00	No Sample Provided
Breath Volume		0.000	22:00	
Air Blank		0.000	22:00	
Air Blank		0.000	22:02	
Subject Sample	# 2		22:02	Subject Test Refused
Breath Volume		0.000	22:02	
Air Blank		0.000	22:03	
Control Test		0.080	22:03	
Air Blank		0.000	22:04	
Diagnostics Check		OK	22:04	

Data Download Date/Time: 05/01/2025 14:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ARNAUD

DATE/TIME OF TEST: 04/12/2025 01:58:02

Diagnostics Check		OK	01:51
Air Blank		0.000	01:51
Control Test		0.081	01:51
Air Blank		0.000	01:52
Subject Sample	# 1	0.149	01:53
Breath Volume		1.914	01:53
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.137	01:56
Breath Volume		1.664	01:56
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:57
Diagnostics Check		OK	01:58

Data Download Date/Time: 05/01/2025 14:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 04/12/2025 02:47:29

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.080	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.238	02:43
Breath Volume		2.347	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.227	02:45
Breath Volume		2.578	02:45
Air Blank		0.000	02:46
Control Test		0.080	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 05/01/2025 14:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ELFTMANN

DATE/TIME OF TEST: 04/12/2025 02:54:34

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:50	
Control Test		0.079	02:50	
Air Blank		0.000	02:51	
Subject Sample	# 1		02:53	Subject Test Refused
Breath Volume		0.000	02:53	
Air Blank		0.000	02:53	
Control Test		0.080	02:53	
Air Blank		0.000	02:54	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 05/01/2025 14:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/12/2025 03:48:52

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.080	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.103	03:44
Breath Volume		1.550	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.101	03:47
Breath Volume		1.664	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 05/01/2025 14:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DAUGHETEE,NICHOLAS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:43

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 04/15/2025 14:14:46

Diagnostics Check		OK	14:08
Air Blank		0.000	14:09
Control Test		0.081	14:09
Air Blank		0.000	14:09
Subject Sample	# 1	0.000	14:10
Breath Volume		2.054	14:10
Air Blank		0.000	14:10
Air Blank		0.000	14:12
Subject Sample	# 2	0.000	14:13
Breath Volume		1.433	14:13
Air Blank		0.000	14:13
Control Test		0.080	14:14
Air Blank		0.000	14:14
Diagnostics Check		OK	14:14

Data Download Date/Time: 05/01/2025 14:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: HAMMERLE

DATE/TIME OF TEST: 04/17/2025 02:35:47

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:32	
Control Test		0.080	02:32	
Air Blank		0.000	02:32	
Subject Sample	# 1		02:34	Subject Test Refused
Breath Volume		0.000	02:34	
Air Blank		0.000	02:34	
Control Test		0.081	02:35	
Air Blank		0.000	02:35	
Diagnostics Check		OK	02:35	

Data Download Date/Time: 05/01/2025 14:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: SIEWRUK

DATE/TIME OF TEST: 04/17/2025 03:10:20

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:56	
Control Test		0.080	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 1	0.100	03:00	Volume Not Met
Breath Volume		0.226	03:00	
Air Blank		0.000	03:00	
Air Blank		0.000	03:02	
Subject Sample	# 2	0.137	03:04	
Breath Volume		1.781	03:04	
Air Blank		0.000	03:04	
Air Blank		0.000	03:06	
Subject Sample	# 3	0.127	03:08	
Breath Volume		1.726	03:08	
Air Blank		0.000	03:09	
Control Test		0.079	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 05/01/2025 14:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: BEECKEN

DATE/TIME OF TEST: 04/18/2025 01:24:07

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.080	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.153	01:19
Breath Volume		1.449	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.159	01:22
Breath Volume		2.136	01:22
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 05/01/2025 14:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: KERSHAW

DATE/TIME OF TEST: 04/18/2025 23:14:43

Diagnostics Check		OK	23:09	
Air Blank		0.000	23:09	
Control Test		0.081	23:09	
Air Blank		0.000	23:10	
Subject Sample	# 1		23:13	Subject Test Refused
Breath Volume		0.000	23:13	
Air Blank		0.000	23:13	
Control Test		0.081	23:14	
Air Blank		0.000	23:14	
Diagnostics Check		OK	23:14	

Data Download Date/Time: 05/01/2025 14:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: SCHARFY

DATE/TIME OF TEST: 04/19/2025 01:34:44

Diagnostics Check		OK	01:29	
Air Blank		0.000	01:30	
Control Test		0.081	01:30	
Air Blank		0.000	01:31	
Subject Sample	# 1		01:33	Subject Test Refused
Breath Volume		0.000	01:33	
Air Blank		0.000	01:33	
Control Test		0.080	01:34	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 05/01/2025 14:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: LESKO

DATE/TIME OF TEST: 04/19/2025 03:00:35

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:48	
Control Test		0.081	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 1	0.144	02:52	Volume Not Met
Breath Volume		0.871	02:52	
Air Blank		0.000	02:52	
Air Blank		0.000	02:54	
Subject Sample	# 2	0.154	02:55	
Breath Volume		2.175	02:55	
Air Blank		0.000	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 3	0.145	02:59	
Breath Volume		1.597	02:59	
Air Blank		0.000	02:59	
Control Test		0.080	02:59	
Air Blank		0.000	03:00	
Diagnostics Check		OK	03:00	

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: DYSARD

DATE/TIME OF TEST: 04/19/2025 04:00:21

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.080	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.164	03:55
Breath Volume		3.113	03:55
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.160	03:58
Breath Volume		2.699	03:58
Air Blank		0.000	03:59
Control Test		0.080	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 04/20/2025 03:11:23

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.120	03:06
Breath Volume		1.226	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.116	03:09
Breath Volume		1.285	03:09
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: MOND

DATE/TIME OF TEST: 04/20/2025 04:19:32

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.080	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.134	04:14
Breath Volume		1.992	04:14
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.133	04:18
Breath Volume		1.640	04:18
Air Blank		0.000	04:18
Control Test		0.081	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ATTAWAY

DATE/TIME OF TEST: 04/23/2025 01:47:42

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.081	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.240	01:43
Breath Volume		1.652	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.233	01:46
Breath Volume		1.445	01:46
Air Blank		0.000	01:46
Control Test		0.079	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: AYA JIMENEZ

DATE/TIME OF TEST: 04/24/2025 03:50:16

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:38	
Control Test		0.080	03:38	
Air Blank		0.000	03:39	
Subject Sample	# 1		03:42	No Sample Provided
Breath Volume		0.000	03:42	
Air Blank		0.000	03:43	
Air Blank		0.000	03:45	
Subject Sample	# 2	0.146	03:45	
Breath Volume		2.375	03:45	
Air Blank		0.000	03:46	
Air Blank		0.000	03:48	
Subject Sample	# 3	0.137	03:48	
Breath Volume		1.429	03:48	
Air Blank		0.000	03:49	
Control Test		0.080	03:49	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: MAZURE

DATE/TIME OF TEST: 04/24/2025 04:26:34

Diagnostics Check		OK	04:22	
Air Blank		0.000	04:22	
Control Test		0.079	04:23	
Air Blank		0.000	04:23	
Subject Sample	# 1		04:25	Subject Test Refused
Breath Volume		0.000	04:25	
Air Blank		0.000	04:25	
Control Test		0.079	04:25	
Air Blank		0.000	04:26	
Diagnostics Check		OK	04:26	

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 04/24/2025 19:27:11

Diagnostics Check		OK	19:18
Air Blank		0.000	19:19
Control Test		0.080	19:19
Air Blank		0.000	19:20
Subject Sample	# 1	0.250	19:22
Breath Volume		2.296	19:22
Air Blank		0.000	19:23
Air Blank		0.000	19:24
Subject Sample	# 2	0.265	19:25
Breath Volume		2.921	19:25
Air Blank		0.000	19:26
Control Test		0.078	19:26
Air Blank		0.000	19:27
Diagnostics Check		OK	19:27

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: TYSON

DATE/TIME OF TEST: 04/24/2025 20:40:25

Diagnostics Check		OK	20:33
Air Blank		0.000	20:33
Control Test		0.080	20:34
Air Blank		0.000	20:34
Subject Sample	# 1	0.000	20:36
Breath Volume		2.421	20:36
Air Blank		0.000	20:36
Air Blank		0.000	20:38
Subject Sample	# 2	0.000	20:39
Breath Volume		2.550	20:39
Air Blank		0.000	20:39
Control Test		0.080	20:39
Air Blank		0.000	20:40
Diagnostics Check		OK	20:40

Data Download Date/Time: 05/01/2025 14:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BELLAMY

DATE/TIME OF TEST: 04/25/2025 00:48:29

Diagnostics Check		OK	00:41
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.024	00:44
Breath Volume		1.597	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.024	00:47
Breath Volume		1.273	00:47
Air Blank		0.000	00:47
Control Test		0.080	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BISHOP,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/25/2025 18:15:08

Diagnostics Check		OK	18:11	
Air Blank		0.000	18:11	
Control Test		0.080	18:11	
Air Blank		0.000	18:12	
Subject Sample	# 1		18:13	Subject Test Refused
Breath Volume		0.000	18:13	
Air Blank		0.000	18:14	
Control Test		0.080	18:14	
Air Blank		0.000	18:15	
Diagnostics Check		OK	18:15	

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: MEHEDI

DATE/TIME OF TEST: 04/25/2025 21:14:47

Diagnostics Check		OK	21:10	
Air Blank		0.000	21:11	
Control Test		0.080	21:11	
Air Blank		0.000	21:11	
Subject Sample	# 1		21:13	Subject Test Refused
Breath Volume		0.000	21:13	
Air Blank		0.000	21:13	
Control Test		0.080	21:14	
Air Blank		0.000	21:14	
Diagnostics Check		OK	21:14	

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/26/2025 01:48:00

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.080	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.188	01:43
Breath Volume		1.648	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.185	01:46
Breath Volume		1.390	01:46
Air Blank		0.000	01:47
Control Test		0.080	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:48

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/26/2025 04:11:40

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:56	
Control Test		0.080	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 1	0.144	04:00	No .020 agreement
Breath Volume		2.160	04:00	
Air Blank		0.000	04:01	
Air Blank		0.000	04:03	
Subject Sample	# 2	0.126	04:06	Volume Not Met
Breath Volume		0.363	04:06	
Air Blank		0.000	04:06	
Air Blank		0.000	04:08	
Subject Sample	# 3	0.123	04:10	No .020 agreement
Breath Volume		1.988	04:10	
Air Blank		0.000	04:10	
Control Test		0.079	04:11	
Air Blank		0.000	04:11	
Diagnostics Check		OK	04:11	

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/26/2025 04:22:30

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.079	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.127	04:17
Breath Volume		1.562	04:17
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.140	04:21
Breath Volume		2.253	04:21
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: SWINEY

DATE/TIME OF TEST: 04/26/2025 21:55:32

Diagnostics Check		OK	21:49
Air Blank		0.000	21:49
Control Test		0.080	21:49
Air Blank		0.000	21:50
Subject Sample	# 1	0.148	21:51
Breath Volume		1.179	21:51
Air Blank		0.000	21:51
Air Blank		0.000	21:53
Subject Sample	# 2	0.156	21:54
Breath Volume		1.359	21:54
Air Blank		0.000	21:54
Control Test		0.081	21:54
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JULIO,C
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MASLE,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: KAISER

DATE/TIME OF TEST: 04/28/2025 04:03:39

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.080	03:54
Air Blank		0.000	03:55
Subject Sample	# 1	0.119	03:57
Breath Volume		1.351	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.125	04:02
Breath Volume		1.457	04:02
Air Blank		0.000	04:02
Control Test		0.079	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JULIO,C
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: KENNY

DATE/TIME OF TEST: 04/30/2025 02:28:38

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:17	
Control Test		0.080	02:18	
Air Blank		0.000	02:18	
Subject Sample	# 1		02:21	No Sample Provided
Breath Volume		0.000	02:21	
Air Blank		0.000	02:22	
Air Blank		0.000	02:24	
Subject Sample	# 2		02:27	No Sample Provided
Breath Volume		0.000	02:27	
Air Blank		0.000	02:27	
Control Test		0.080	02:28	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 05/01/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGSON,ROLAND,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: LOUX,SAGE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: CURTISS

DATE/TIME OF TEST: 04/03/2025 18:10:38

Diagnostics Check		OK	18:02
Air Blank		0.000	18:02
Control Test		0.080	18:03
Air Blank		0.000	18:03
Subject Sample	# 1	0.012	18:05
Breath Volume		1.566	18:05
Air Blank		0.000	18:06
Air Blank		0.000	18:08
Subject Sample	# 2	0.012	18:09
Breath Volume		2.148	18:09
Air Blank		0.000	18:09
Control Test		0.080	18:10
Air Blank		0.000	18:10
Diagnostics Check		OK	18:10

Data Download Date/Time: 04/17/2025 13:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: STOGSDILL,MADISON
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: GLOGG

DATE/TIME OF TEST: 04/12/2025 00:52:32

Diagnostics Check		OK	00:49	
Air Blank		0.000	00:50	
Control Test		0.081	00:50	
Air Blank		0.000	00:50	
Subject Sample	# 1		00:51	Subject Test Refused
Breath Volume		0.000	00:51	
Air Blank		0.000	00:51	
Control Test		0.080	00:51	
Air Blank		0.000	00:52	
Diagnostics Check		OK	00:52	

Data Download Date/Time: 04/17/2025 13:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTIEN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: ACAHUA MACUIXTLE

DATE/TIME OF TEST: 04/13/2025 01:31:59

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.081	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.171	01:27
Breath Volume		2.304	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:29
Subject Sample	# 2	0.174	01:30
Breath Volume		1.453	01:30
Air Blank		0.000	01:31
Control Test		0.080	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 04/17/2025 13:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALAMEDA,JORGE,L
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: LOUX,SAGE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 04/16/2025 18:50:46

Diagnostics Check		OK	18:46	
Air Blank		0.000	18:46	
Control Test		0.081	18:46	
Air Blank		0.000	18:47	
Subject Sample	# 1		18:49	Subject Test Refused
Breath Volume		0.000	18:49	
Air Blank		0.000	18:49	
Control Test		0.081	18:50	
Air Blank		0.000	18:50	
Diagnostics Check		OK	18:50	

Data Download Date/Time: 04/17/2025 13:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: LOUX,SAGE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/17/2025 20:19:52

Diagnostics Check		OK	20:16	
Air Blank		0.000	20:17	
Control Test		0.081	20:17	
Air Blank		0.000	20:18	
Subject Sample	# 1		20:18	Subject Test Refused
Breath Volume		0.000	20:18	
Air Blank		0.000	20:18	
Control Test		0.080	20:19	
Air Blank		0.000	20:19	
Diagnostics Check		OK	20:19	

Data Download Date/Time: 05/26/2025 10:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 04/19/2025 19:42:25

Diagnostics Check		OK	19:35
Air Blank		0.000	19:36
Control Test		0.080	19:36
Air Blank		0.000	19:37
Subject Sample	# 1	0.137	19:38
Breath Volume		2.929	19:38
Air Blank		0.000	19:38
Air Blank		0.000	19:40
Subject Sample	# 2	0.136	19:40
Breath Volume		2.601	19:40
Air Blank		0.000	19:41
Control Test		0.080	19:41
Air Blank		0.000	19:42
Diagnostics Check		OK	19:42

Data Download Date/Time: 05/26/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: OCHOA SALES

DATE/TIME OF TEST: 04/26/2025 19:55:15

Diagnostics Check		OK	19:46
Air Blank		0.000	19:47
Control Test		0.080	19:47
Air Blank		0.000	19:48
Subject Sample	# 1	0.233	19:50
Breath Volume		1.621	19:50
Air Blank		0.000	19:50
Air Blank		0.000	19:52
Subject Sample	# 2	0.230	19:53
Breath Volume		2.117	19:53
Air Blank		0.000	19:54
Control Test		0.080	19:54
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 05/26/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DYE,BRANDON
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: GARTMOND

DATE/TIME OF TEST: 04/02/2025 03:51:11

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:36	
Control Test		0.079	03:37	
Air Blank		0.000	03:37	
Subject Sample	# 1	0.116	03:39	
Breath Volume		1.160	03:39	
Air Blank		0.000	03:40	
Air Blank		0.000	03:42	
Subject Sample	# 2	0.103	03:45	Volume Not Met
Breath Volume		0.531	03:45	
Air Blank		0.000	03:46	
Air Blank		0.000	03:47	
Subject Sample	# 3	0.113	03:49	
Breath Volume		1.171	03:49	
Air Blank		0.000	03:50	
Control Test		0.079	03:50	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 04/22/2025 09:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ELQREISH

DATE/TIME OF TEST: 04/03/2025 01:23:12

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:12	
Control Test		0.080	01:12	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:16	No Sample Provided
Breath Volume		0.000	01:16	
Air Blank		0.000	01:16	
Air Blank		0.000	01:18	
Subject Sample	# 2		01:21	No Sample Provided
Breath Volume		0.000	01:21	
Air Blank		0.000	01:22	
Control Test		0.080	01:22	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 04/22/2025 09:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: COLDING

DATE/TIME OF TEST: 04/04/2025 23:47:29

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.081	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.104	23:43
Breath Volume		2.296	23:43
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.100	23:45
Breath Volume		1.855	23:45
Air Blank		0.000	23:46
Control Test		0.079	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 04/22/2025 09:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: DIAZ CARDONA

DATE/TIME OF TEST: 04/05/2025 01:13:28

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.079	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.298	01:08
Breath Volume		2.347	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.289	01:11
Breath Volume		2.578	01:11
Air Blank		0.000	01:12
Control Test		0.078	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 04/22/2025 09:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIELDS,NIKOLAUS,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 04/05/2025 05:14:11

Diagnostics Check		OK	05:07
Air Blank		0.000	05:08
Control Test		0.078	05:08
Air Blank		0.000	05:09
Subject Sample	# 1	0.107	05:09
Breath Volume		1.593	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.106	05:12
Breath Volume		1.757	05:12
Air Blank		0.000	05:13
Control Test		0.079	05:13
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 04/22/2025 09:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: CARRASQUILLO GOMEZ

DATE/TIME OF TEST: 04/05/2025 05:37:57

Diagnostics Check		OK	05:26	
Air Blank		0.000	05:26	
Control Test		0.080	05:27	
Air Blank		0.000	05:27	
Subject Sample	# 1		05:30	No Sample Provided
Breath Volume		0.000	05:30	
Air Blank		0.000	05:31	
Air Blank		0.000	05:33	
Subject Sample	# 2		05:36	No Sample Provided
Breath Volume		0.000	05:36	
Air Blank		0.000	05:37	
Control Test		0.081	05:37	
Air Blank		0.000	05:37	
Diagnostics Check		OK	05:37	

Data Download Date/Time: 04/22/2025 09:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIELDS,NIKOLAUS,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HERNANDEZ,EFRAIN
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:08

SUBJECT'S LAST NAME: ORGA-ARAUJO

DATE/TIME OF TEST: 04/05/2025 08:06:02

Diagnostics Check		OK	07:54	
Air Blank		0.000	07:55	
Control Test		0.081	07:55	
Air Blank		0.000	07:55	
Subject Sample	# 1		07:59	No Sample Provided
Breath Volume		0.000	07:59	
Air Blank		0.000	07:59	
Air Blank		0.000	08:01	
Subject Sample	# 2		08:04	No Sample Provided
Breath Volume		0.000	08:04	
Air Blank		0.000	08:05	
Control Test		0.081	08:05	
Air Blank		0.000	08:05	
Diagnostics Check		OK	08:06	

Data Download Date/Time: 04/22/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: HAMES

DATE/TIME OF TEST: 04/06/2025 00:19:22

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.079	00:12
Air Blank		0.000	00:13
Subject Sample	# 1	0.140	00:14
Breath Volume		1.511	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.148	00:17
Breath Volume		1.140	00:17
Air Blank		0.000	00:18
Control Test		0.079	00:18
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 04/22/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATINO,ALONZO,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CONTRERA DIAZ,EDUARDO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: PATINO

DATE/TIME OF TEST: 04/06/2025 19:54:37

Diagnostics Check		OK	19:47
Air Blank		0.000	19:48
Control Test		0.080	19:48
Air Blank		0.000	19:49
Subject Sample	# 1	0.148	19:49
Breath Volume		1.187	19:49
Air Blank		0.000	19:50
Air Blank		0.000	19:52
Subject Sample	# 2	0.168	19:53
Breath Volume		2.437	19:53
Air Blank		0.000	19:53
Control Test		0.080	19:54
Air Blank		0.000	19:54
Diagnostics Check		OK	19:54

Data Download Date/Time: 04/22/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: AZUA DE ZAMORA

DATE/TIME OF TEST: 04/06/2025 23:44:09

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:43	
Control Test	0.183	23:43	Control Outside Tolerance
Air Blank	0.000	23:44	

Data Download Date/Time: 04/22/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: AZUA DE ZAMORA

DATE/TIME OF TEST: 04/06/2025 23:52:46

Diagnostics Check		OK	23:46
Air Blank		0.000	23:46
Control Test		0.078	23:47
Air Blank		0.000	23:47
Subject Sample	# 1	0.187	23:48
Breath Volume		1.628	23:48
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.177	23:51
Breath Volume		1.593	23:51
Air Blank		0.000	23:51
Control Test		0.080	23:52
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 04/22/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: SMITH JR

DATE/TIME OF TEST: 04/08/2025 01:17:23

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:11	
Control Test		0.079	01:12	
Air Blank		0.000	01:12	
Subject Sample	# 1	0.203	01:13	
Breath Volume		1.386	01:13	
Air Blank		0.000	01:14	
Air Blank		0.000	01:15	
Subject Sample	# 2	0.219	01:16	
Breath Volume		1.617	01:16	
Air Blank			01:16	RFI Detect
Air Blank		0.000	01:17	

Data Download Date/Time: 04/22/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: SMITH JR

DATE/TIME OF TEST: 04/08/2025 01:27:47

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.077	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.213	01:20
Breath Volume		2.027	01:20
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.220	01:25
Breath Volume		2.636	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 04/22/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREWS,TYREE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: VICTOR

DATE/TIME OF TEST: 04/08/2025 06:32:36

Diagnostics Check		OK	06:25
Air Blank		0.000	06:26
Control Test		0.080	06:26
Air Blank		0.000	06:27
Subject Sample	# 1	0.000	06:28
Breath Volume		1.472	06:28
Air Blank		0.000	06:28
Air Blank		0.000	06:30
Subject Sample	# 2	0.000	06:31
Breath Volume		1.453	06:31
Air Blank		0.000	06:31
Control Test		0.078	06:32
Air Blank		0.000	06:32
Diagnostics Check		OK	06:32

Data Download Date/Time: 04/22/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MCCOSCO,LANDON
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: LANDOCQUE

DATE/TIME OF TEST: 04/08/2025 20:41:48

Diagnostics Check		OK	20:35
Air Blank		0.000	20:35
Control Test		0.080	20:36
Air Blank		0.000	20:36
Subject Sample	# 1	0.160	20:37
Breath Volume		1.398	20:37
Air Blank		0.000	20:37
Air Blank		0.000	20:39
Subject Sample	# 2	0.155	20:40
Breath Volume		1.339	20:40
Air Blank		0.000	20:40
Control Test		0.078	20:41
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 04/22/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: DIAZ-LOZADA

DATE/TIME OF TEST: 04/10/2025 02:38:35

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.204	02:34
Breath Volume		2.011	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.203	02:36
Breath Volume		1.914	02:36
Air Blank		0.000	02:37
Control Test		0.077	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 04/22/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DYE,BRANDON
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 04/11/2025 01:06:30

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:55	
Control Test		0.080	00:55	
Air Blank		0.000	00:56	
Subject Sample	# 1	0.074	00:59	Volume Not Met
Breath Volume		0.566	00:59	
Air Blank		0.000	01:00	
Air Blank		0.000	01:01	
Subject Sample	# 2	0.081	01:04	Volume Not Met
Breath Volume		0.601	01:04	
Air Blank		0.000	01:05	
Control Test		0.078	01:05	
Air Blank		0.000	01:06	
Diagnostics Check		OK	01:06	

Data Download Date/Time: 04/22/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DYE,BRANDON
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 04/11/2025 03:02:53

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:46	
Control Test		0.079	02:46	
Air Blank		0.000	02:47	
Subject Sample	# 1	0.063	02:50	Volume Not Met
Breath Volume		0.503	02:50	
Air Blank		0.000	02:51	
Air Blank		0.000	02:53	
Subject Sample	# 2	0.062	02:55	
Breath Volume		1.109	02:55	
Air Blank		0.000	02:56	
Air Blank		0.000	02:58	
Subject Sample	# 3	0.056	03:01	Volume Not Met
Breath Volume		1.031	03:01	
Air Blank		0.000	03:01	
Control Test		0.078	03:02	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 04/22/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO,DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: VILLARREAL

DATE/TIME OF TEST: 04/11/2025 04:04:33

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:53	
Control Test		0.079	03:53	
Air Blank		0.000	03:53	
Subject Sample	# 1	0.165	03:57	Volume Not Met
Breath Volume		0.730	03:57	
Air Blank		0.000	03:58	
Air Blank		0.000	03:59	
Subject Sample	# 2	0.145	04:02	Volume Not Met
Breath Volume		0.242	04:02	
Air Blank		0.000	04:03	
Control Test		0.078	04:03	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 04/22/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO,DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: VILLARREAL

DATE/TIME OF TEST: 04/11/2025 04:16:02

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.077	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.173	04:11
Breath Volume		1.785	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:13
Subject Sample	# 2	0.163	04:14
Breath Volume		1.730	04:14
Air Blank		0.000	04:15
Control Test		0.079	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:16

Data Download Date/Time: 04/22/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ARIK,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: JONES,ARIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/11/2025 22:32:30

Diagnostics Check	OK	22:31	
Air Blank		22:32	RFI Detect
Air Blank	0.000	22:32	

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ARIK,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: JONES,ARIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/11/2025 22:41:58

Diagnostics Check		OK	22:34
Air Blank		0.000	22:34
Control Test		0.080	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.162	22:36
Breath Volume		1.410	22:36
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.156	22:40
Breath Volume		1.277	22:40
Air Blank		0.000	22:41
Control Test		0.078	22:41
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: PEREZ AMARO

DATE/TIME OF TEST: 04/11/2025 22:55:28

Diagnostics Check		OK	22:48
Air Blank		0.000	22:48
Control Test		0.078	22:49
Air Blank		0.000	22:49
Subject Sample	# 1	0.185	22:50
Breath Volume		2.542	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.192	22:53
Breath Volume		2.339	22:53
Air Blank		0.000	22:54
Control Test		0.078	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MORRIS JR

DATE/TIME OF TEST: 04/12/2025 01:27:59

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.263	01:22
Breath Volume		1.492	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.280	01:26
Breath Volume		1.730	01:26
Air Blank		0.000	01:27
Control Test		0.079	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ARIK,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: JONES,ARIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: VICENTE RAMOS

DATE/TIME OF TEST: 04/12/2025 17:12:01

Diagnostics Check		OK	17:05
Air Blank		0.000	17:06
Control Test		0.081	17:06
Air Blank		0.000	17:07
Subject Sample	# 1	0.175	17:07
Breath Volume		3.550	17:07
Air Blank		0.000	17:08
Air Blank		0.000	17:10
Subject Sample	# 2	0.156	17:10
Breath Volume		2.605	17:10
Air Blank		0.000	17:11
Control Test		0.078	17:11
Air Blank		0.000	17:11
Diagnostics Check		OK	17:12

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: CHANCY

DATE/TIME OF TEST: 04/13/2025 02:41:52

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.080	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.191	02:37
Breath Volume		2.078	02:37
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.179	02:40
Breath Volume		1.378	02:40
Air Blank		0.000	02:40
Control Test		0.080	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOTEN,NATHANIEL,K
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WOOTEN,NATHANIEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: KIM

DATE/TIME OF TEST: 04/13/2025 04:28:39

Diagnostics Check		OK	04:17	
Air Blank		0.000	04:17	
Control Test		0.079	04:18	
Air Blank		0.000	04:18	
Subject Sample	# 1		04:21	No Sample Provided
Breath Volume		0.000	04:21	
Air Blank		0.000	04:22	
Air Blank		0.000	04:24	
Subject Sample	# 2		04:27	No Sample Provided
Breath Volume		0.000	04:27	
Air Blank		0.000	04:27	
Control Test		0.079	04:28	
Air Blank		0.000	04:28	
Diagnostics Check		OK	04:28	

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: LLUBERES

DATE/TIME OF TEST: 04/14/2025 02:11:40

Diagnostics Check		OK	01:59	
Air Blank		0.000	02:00	
Control Test		0.080	02:00	
Air Blank		0.000	02:00	
Subject Sample	# 1	0.241	02:04	Volume Not Met
Breath Volume		0.269	02:04	
Air Blank		0.000	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 2	0.242	02:09	Volume Not Met
Breath Volume		0.347	02:09	
Air Blank		0.000	02:10	
Control Test		0.077	02:11	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: ALLENDES DONOSO

DATE/TIME OF TEST: 04/14/2025 05:48:36

Diagnostics Check		OK	05:37	
Air Blank		0.000	05:37	
Control Test		0.080	05:37	
Air Blank		0.000	05:38	
Subject Sample	# 1		05:41	No Sample Provided
Breath Volume		0.000	05:41	
Air Blank		0.000	05:42	
Air Blank		0.000	05:44	
Subject Sample	# 2		05:47	No Sample Provided
Breath Volume		0.000	05:47	
Air Blank		0.000	05:47	
Control Test		0.081	05:47	
Air Blank		0.000	05:48	
Diagnostics Check		OK	05:48	

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONTRERA DIAZ,EDUARDO,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FORSBERG,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: HERNANDEZ FARELO

DATE/TIME OF TEST: 04/14/2025 19:10:11

Diagnostics Check		OK	19:03
Air Blank		0.000	19:04
Control Test		0.079	19:04
Air Blank		0.000	19:05
Subject Sample	# 1	0.154	19:05
Breath Volume		2.238	19:05
Air Blank		0.000	19:06
Air Blank		0.000	19:08
Subject Sample	# 2	0.167	19:08
Breath Volume		3.335	19:08
Air Blank		0.000	19:09
Control Test		0.079	19:09
Air Blank		0.000	19:10
Diagnostics Check		OK	19:10

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: CASSIDY

DATE/TIME OF TEST: 04/15/2025 02:27:31

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.080	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.077	02:23
Breath Volume		2.585	02:23
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.076	02:25
Breath Volume		2.199	02:25
Air Blank		0.000	02:26
Control Test		0.080	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: VUBERT

DATE/TIME OF TEST: 04/15/2025 23:28:39

Diagnostics Check		OK	23:21
Air Blank		0.000	23:22
Control Test		0.080	23:22
Air Blank		0.000	23:23
Subject Sample	# 1	0.024	23:24
Breath Volume		1.628	23:24
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.022	23:27
Breath Volume		2.437	23:27
Air Blank		0.000	23:27
Control Test		0.079	23:28
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:19

SUBJECT'S LAST NAME: VINTON

DATE/TIME OF TEST: 04/16/2025 14:49:32

Diagnostics Check		OK	14:41
Air Blank		0.000	14:42
Control Test		0.081	14:42
Air Blank		0.000	14:43
Subject Sample	# 1	0.188	14:44
Breath Volume		1.988	14:44
Air Blank		0.000	14:45
Air Blank		0.000	14:47
Subject Sample	# 2	0.174	14:47
Breath Volume		1.281	14:47
Air Blank		0.000	14:48
Control Test		0.079	14:48
Air Blank		0.000	14:49
Diagnostics Check		OK	14:49

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 04/16/2025 18:26:42

Diagnostics Check		OK	18:20
Air Blank		0.000	18:21
Control Test		0.081	18:21
Air Blank		0.000	18:22
Subject Sample	# 1	0.000	18:22
Breath Volume		1.738	18:22
Air Blank		0.000	18:22
Air Blank		0.000	18:24
Subject Sample	# 2	0.000	18:25
Breath Volume		1.437	18:25
Air Blank		0.000	18:25
Control Test		0.081	18:26
Air Blank		0.000	18:26
Diagnostics Check		OK	18:26

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREWS,TYREE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WRIGHT,CORY
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: GLENN

DATE/TIME OF TEST: 04/17/2025 07:07:33

Diagnostics Check		OK	07:01
Air Blank		0.000	07:01
Control Test		0.080	07:02
Air Blank		0.000	07:02
Subject Sample	# 1	0.000	07:03
Breath Volume		1.636	07:03
Air Blank		0.000	07:03
Air Blank		0.000	07:05
Subject Sample	# 2	0.000	07:06
Breath Volume		1.605	07:06
Air Blank		0.000	07:06
Control Test		0.079	07:06
Air Blank		0.000	07:07
Diagnostics Check		OK	07:07

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONCION,CHRISTIAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MONCION,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 04/18/2025 02:01:07

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:58	
Control Test		0.080	01:59	
Air Blank		0.000	01:59	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	02:00	
Control Test		0.081	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 04/22/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 04/18/2025 04:29:08

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.080	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.141	04:24
Breath Volume		1.101	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.153	04:27
Breath Volume		1.687	04:27
Air Blank		0.000	04:28
Control Test		0.081	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 04/22/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: GUILLERMO-GONZALEZ

DATE/TIME OF TEST: 04/20/2025 00:14:57

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.082	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.100	00:10
Breath Volume		2.167	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.097	00:13
Breath Volume		2.121	00:13
Air Blank		0.000	00:14
Control Test		0.080	00:14
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 04/22/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 04/20/2025 01:45:04

Diagnostics Check		OK	01:33	
Air Blank		0.000	01:34	
Control Test		0.080	01:34	
Air Blank		0.000	01:34	
Subject Sample	# 1		01:38	No Sample Provided
Breath Volume		0.000	01:38	
Air Blank		0.000	01:38	
Air Blank		0.000	01:40	
Subject Sample	# 2		01:43	No Sample Provided
Breath Volume		0.000	01:43	
Air Blank		0.000	01:44	
Control Test		0.081	01:44	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 04/22/2025 09:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIELDS,NIKOLAUS,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 04/20/2025 06:20:51

Diagnostics Check		OK	06:14
Air Blank		0.000	06:15
Control Test		0.079	06:15
Air Blank		0.000	06:15
Subject Sample	# 1	0.102	06:16
Breath Volume		1.425	06:16
Air Blank		0.000	06:16
Air Blank		0.000	06:18
Subject Sample	# 2	0.110	06:19
Breath Volume		2.304	06:19
Air Blank		0.000	06:19
Control Test		0.080	06:20
Air Blank		0.000	06:20
Diagnostics Check		OK	06:20

Data Download Date/Time: 04/22/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,LEONARDO,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DE LA CRUZ,LEONARDO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: LUGO ACOSTA

DATE/TIME OF TEST: 04/21/2025 01:33:32

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.080	01:27
Air Blank		0.000	01:27
Subject Sample	# 1	0.113	01:28
Breath Volume		1.605	01:28
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.115	01:31
Breath Volume		3.234	01:31
Air Blank		0.000	01:32
Control Test		0.079	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 04/22/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: AGRAMONTE RONDON

DATE/TIME OF TEST: 04/21/2025 22:47:45

Diagnostics Check		OK	22:36	
Air Blank		0.000	22:36	
Control Test		0.080	22:37	
Air Blank		0.000	22:37	
Subject Sample	# 1		22:40	No Sample Provided
Breath Volume		0.000	22:40	
Air Blank		0.000	22:41	
Air Blank		0.000	22:43	
Subject Sample	# 2		22:46	No Sample Provided
Breath Volume		0.000	22:46	
Air Blank		0.000	22:46	
Control Test		0.080	22:47	
Air Blank		0.000	22:47	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 04/22/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 04/23/2025 02:30:01

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.080	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.178	02:25
Breath Volume		1.312	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.162	02:28
Breath Volume		1.875	02:28
Air Blank		0.000	02:29
Control Test		0.078	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:30

Data Download Date/Time: 05/20/2025 05:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: FLINSPACH

DATE/TIME OF TEST: 04/23/2025 03:16:34

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.080	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.094	03:11
Breath Volume		2.191	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.096	03:14
Breath Volume		2.589	03:14
Air Blank		0.000	03:15
Control Test		0.079	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 05/20/2025 05:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,LEONARDO,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DE LA CRUZ,LEONARDO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: BEACHAM

DATE/TIME OF TEST: 04/25/2025 05:43:06

Diagnostics Check		OK	05:36
Air Blank		0.000	05:36
Control Test		0.080	05:36
Air Blank		0.000	05:37
Subject Sample	# 1	0.125	05:38
Breath Volume		2.738	05:38
Air Blank		0.000	05:38
Air Blank		0.000	05:40
Subject Sample	# 2	0.122	05:41
Breath Volume		1.421	05:41
Air Blank		0.000	05:42
Control Test		0.078	05:42
Air Blank		0.000	05:43
Diagnostics Check		OK	05:43

Data Download Date/Time: 05/20/2025 05:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BAGGETT,TYLER
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: TUPOU

DATE/TIME OF TEST: 04/26/2025 04:53:57

Diagnostics Check		OK	04:46
Air Blank		0.000	04:47
Control Test		0.080	04:47
Air Blank		0.000	04:47
Subject Sample	# 1	0.207	04:48
Breath Volume		2.375	04:48
Air Blank		0.000	04:49
Air Blank		0.000	04:51
Subject Sample	# 2	0.204	04:52
Breath Volume		1.687	04:52
Air Blank		0.000	04:53
Control Test		0.078	04:53
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 05/20/2025 05:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORESTEIRE,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FORESTEIRE,JOSEPH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/26/2025 23:28:11

Diagnostics Check		OK	23:20
Air Blank		0.000	23:21
Control Test		0.080	23:21
Air Blank		0.000	23:21
Subject Sample	# 1	0.105	23:23
Breath Volume		1.132	23:23
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.101	23:26
Breath Volume		1.437	23:26
Air Blank		0.000	23:27
Control Test		0.078	23:27
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 05/20/2025 05:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: GALLIAN

DATE/TIME OF TEST: 04/27/2025 02:09:48

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.080	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.189	02:05
Breath Volume		2.214	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:07
Subject Sample	# 2	0.188	02:08
Breath Volume		1.878	02:08
Air Blank		0.000	02:08
Control Test		0.079	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 05/20/2025 05:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,WALDRICK,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,WALDRICK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: SIRIGIRI

DATE/TIME OF TEST: 04/27/2025 10:26:44

Diagnostics Check	OK	10:26	
Air Blank		10:26	RFI Detect
Air Blank		10:26	RFI Detect

Data Download Date/Time: 05/20/2025 05:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,WALDRICK,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,WALDRICK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: SIRIGIRI

DATE/TIME OF TEST: 04/27/2025 10:29:59

Diagnostics Check	OK	10:29	
Air Blank		10:29	RFI Detect
Air Blank		10:29	RFI Detect

Data Download Date/Time: 05/20/2025 05:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,WALDRICK,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,WALDRICK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: SIRIGIRI

DATE/TIME OF TEST: 04/27/2025 10:48:19

Diagnostics Check		OK	10:36	
Air Blank		0.000	10:36	
Control Test		0.079	10:37	
Air Blank		0.000	10:37	
Subject Sample	# 1	0.195	10:40	Volume Not Met
Breath Volume		0.031	10:40	
Air Blank		0.000	10:41	
Air Blank		0.000	10:43	
Subject Sample	# 2	0.185	10:46	Volume Not Met
Breath Volume		0.050	10:46	
Air Blank		0.000	10:47	
Control Test		0.076	10:47	
Air Blank		0.000	10:48	
Diagnostics Check		OK	10:48	

Data Download Date/Time: 05/20/2025 05:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: MARQUART

DATE/TIME OF TEST: 04/27/2025 20:49:29

Diagnostics Check		OK	20:39
Air Blank		0.000	20:40
Control Test		0.080	20:40
Air Blank		0.000	20:41
Subject Sample	# 1	0.259	20:41
Breath Volume		2.570	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:43
Subject Sample	# 2	0.282	20:44
Breath Volume		1.441	20:44
Air Blank		0.000	20:45
Air Blank		0.000	20:47
Subject Sample	# 3	0.256	20:47
Breath Volume		1.894	20:47
Air Blank		0.000	20:48
Control Test		0.080	20:48
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 05/20/2025 05:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,MALACHI,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RANDOLPH,MALACHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: MOULE

DATE/TIME OF TEST: 04/27/2025 21:03:23

Diagnostics Check		OK	20:56
Air Blank		0.000	20:57
Control Test		0.080	20:57
Air Blank		0.000	20:57
Subject Sample	# 1	0.115	20:58
Breath Volume		1.789	20:58
Air Blank		0.000	20:59
Air Blank		0.000	21:01
Subject Sample	# 2	0.121	21:01
Breath Volume		1.691	21:01
Air Blank		0.000	21:02
Control Test		0.080	21:02
Air Blank		0.000	21:03
Diagnostics Check		OK	21:03

Data Download Date/Time: 05/20/2025 05:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 04/28/2025 04:49:22

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.080	04:44
Air Blank		0.000	04:44
Subject Sample	# 1	0.000	04:45
Breath Volume		1.925	04:45
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.000	04:47
Breath Volume		1.699	04:47
Air Blank		0.000	04:48
Control Test		0.081	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 05/20/2025 05:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,LEONARDO,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DE LA CRUZ,LEONARDO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: VELEZ VAZQUEZ

DATE/TIME OF TEST: 04/28/2025 05:06:33

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:00
Subject Sample	# 1	0.014	05:01
Breath Volume		1.332	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.011	05:05
Breath Volume		1.343	05:05
Air Blank		0.000	05:05
Control Test		0.079	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 05/20/2025 05:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DAVIS,SHANIA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: ROJAS MENDOZA

DATE/TIME OF TEST: 04/29/2025 23:32:40

Diagnostics Check		OK	23:24
Air Blank		0.000	23:24
Control Test		0.079	23:25
Air Blank		0.000	23:25
Subject Sample	# 1	0.223	23:27
Breath Volume		1.417	23:27
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.210	23:31
Breath Volume		1.445	23:31
Air Blank		0.000	23:31
Control Test		0.077	23:32
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 05/20/2025 05:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZIZ,D,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: AZIZ,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 04/05/2025 19:06:36

Diagnostics Check		OK	19:00
Air Blank		0.000	19:01
Control Test		0.080	19:01
Air Blank		0.000	19:02
Subject Sample	# 1	0.000	19:02
Breath Volume		3.105	19:02
Air Blank		0.000	19:02
Air Blank		0.000	19:04
Subject Sample	# 2	0.000	19:05
Breath Volume		3.656	19:05
Air Blank		0.000	19:05
Control Test		0.079	19:06
Air Blank		0.000	19:06
Diagnostics Check		OK	19:06

Data Download Date/Time: 04/30/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 04/06/2025 01:54:30

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.163	01:49
Breath Volume		1.167	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.154	01:52
Breath Volume		2.484	01:52
Air Blank		0.000	01:53
Control Test		0.079	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 04/30/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIZZO,HAYLEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RIZZO,HAYLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 04/06/2025 15:07:24

Diagnostics Check		OK	15:00
Air Blank		0.000	15:00
Control Test		0.080	15:01
Air Blank		0.000	15:01
Subject Sample	# 1	0.000	15:02
Breath Volume		2.167	15:02
Air Blank		0.000	15:03
Air Blank		0.000	15:05
Subject Sample	# 2	0.000	15:06
Breath Volume		3.671	15:06
Air Blank		0.000	15:06
Control Test		0.080	15:06
Air Blank		0.000	15:07
Diagnostics Check		OK	15:07

Data Download Date/Time: 04/30/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,ADRIANNE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SNYDER,ADRIANNE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: FRIENSEN

DATE/TIME OF TEST: 04/08/2025 19:01:18

Diagnostics Check		OK	18:54
Air Blank		0.000	18:55
Control Test		0.080	18:55
Air Blank		0.000	18:56
Subject Sample	# 1	0.314	18:56
Breath Volume		1.902	18:56
Air Blank		0.000	18:57
Air Blank		0.000	18:59
Subject Sample	# 2	0.304	18:59
Breath Volume		1.707	18:59
Air Blank		0.000	19:00
Control Test		0.079	19:00
Air Blank		0.000	19:01
Diagnostics Check		OK	19:01

Data Download Date/Time: 04/30/2025 11:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CARABALLO,D
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ALBRIGHT

DATE/TIME OF TEST: 04/15/2025 23:57:27

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.080	23:51
Air Blank		0.000	23:52
Subject Sample	# 1	0.000	23:52
Breath Volume		3.296	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.000	23:56
Breath Volume		3.152	23:56
Air Blank		0.000	23:56
Control Test		0.079	23:56
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 04/30/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: RASTOGI

DATE/TIME OF TEST: 04/18/2025 04:19:23

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.080	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.176	04:14
Breath Volume		1.160	04:14
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.171	04:17
Breath Volume		1.207	04:17
Air Blank		0.000	04:18
Control Test		0.078	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 04/30/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRACCICA,C.,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BRANCH,S.
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ARDITO

DATE/TIME OF TEST: 04/19/2025 01:18:09

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.225	01:13
Breath Volume		3.054	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:15
Subject Sample	# 2	0.229	01:16
Breath Volume		4.144	01:16
Air Blank		0.000	01:17
Control Test		0.078	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 04/30/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,R,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 04/19/2025 01:51:20

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.080	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.193	01:46
Breath Volume		2.984	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.193	01:49
Breath Volume		2.601	01:49
Air Blank		0.000	01:50
Control Test		0.079	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 04/30/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: REXROTH

DATE/TIME OF TEST: 04/19/2025 04:29:35

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.080	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.170	04:23
Breath Volume		1.917	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:25
Subject Sample	# 2	0.164	04:27
Breath Volume		2.171	04:27
Air Blank		0.000	04:28
Control Test		0.079	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 04/30/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,WILLIAM,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GRISMER,GARRETT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:14

SUBJECT'S LAST NAME: ORLANDO

DATE/TIME OF TEST: 04/19/2025 14:52:24

Diagnostics Check		OK	14:42
Air Blank		0.000	14:43
Control Test		0.078	14:43
Air Blank		0.000	14:44
Subject Sample	# 1	0.268	14:44
Breath Volume		1.738	14:44
Air Blank		0.000	14:45
Air Blank		0.000	14:47
Subject Sample	# 2	0.306	14:47
Breath Volume		2.039	14:47
Air Blank		0.000	14:48
Air Blank		0.000	14:50
Subject Sample	# 3	0.282	14:50
Breath Volume		1.742	14:50
Air Blank		0.000	14:51
Control Test		0.079	14:51
Air Blank		0.000	14:52
Diagnostics Check		OK	14:52

Data Download Date/Time: 04/30/2025 11:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONS,DEPUTY,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SYMONS,DEPUTY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/20/2025 05:57:47

Diagnostics Check		OK	05:51
Air Blank		0.000	05:51
Control Test		0.080	05:51
Air Blank		0.000	05:52
Subject Sample	# 1	0.135	05:53
Breath Volume		2.214	05:53
Air Blank		0.000	05:53
Air Blank		0.000	05:55
Subject Sample	# 2	0.138	05:56
Breath Volume		2.601	05:56
Air Blank		0.000	05:56
Control Test		0.079	05:57
Air Blank		0.000	05:57
Diagnostics Check		OK	05:57

Data Download Date/Time: 04/30/2025 11:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: RANDHAWA

DATE/TIME OF TEST: 04/24/2025 01:46:23

Diagnostics Check		OK	01:34	
Air Blank		0.000	01:34	
Control Test		0.081	01:35	
Air Blank		0.000	01:35	
Subject Sample	# 1	0.110	01:38	Volume Not Met
Breath Volume		0.554	01:38	
Air Blank		0.000	01:39	
Air Blank		0.000	01:41	
Subject Sample	# 2	0.135	01:41	
Breath Volume		1.894	01:41	
Air Blank		0.000	01:42	
Air Blank		0.000	01:44	
Subject Sample	# 3	0.135	01:44	
Breath Volume		2.910	01:44	
Air Blank		0.000	01:45	
Control Test		0.079	01:45	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 04/30/2025 11:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRACCICA,CARMEN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GERARDI,JOHN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MANSCO

DATE/TIME OF TEST: 04/24/2025 02:48:25

Diagnostics Check		OK	02:41
Air Blank		0.000	02:42
Control Test		0.080	02:42
Air Blank		0.000	02:43
Subject Sample	# 1	0.164	02:43
Breath Volume		1.675	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.163	02:46
Breath Volume		1.246	02:46
Air Blank		0.000	02:47
Control Test		0.080	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 04/30/2025 11:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GIULIANI

DATE/TIME OF TEST: 04/24/2025 04:19:16

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.080	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.113	04:14
Breath Volume		2.871	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.112	04:17
Breath Volume		2.101	04:17
Air Blank		0.000	04:18
Control Test		0.079	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 04/30/2025 11:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRACCICA,CARMEN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: AKERS,DYLAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 04/29/2025 01:58:46

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.190	01:50
Breath Volume		2.308	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.169	01:54
Breath Volume		1.492	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 3	0.169	01:57
Breath Volume		1.714	01:57
Air Blank		0.000	01:57
Control Test		0.078	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 04/30/2025 11:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEAUREGARD,JEFF,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: MAGANA,E
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/01/2025 20:47:11

Diagnostics Check		OK	20:38
Air Blank		0.000	20:39
Control Test		0.078	20:39
Air Blank		0.000	20:39
Subject Sample	# 1	0.214	20:40
Breath Volume		1.332	20:40
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.207	20:45
Breath Volume		1.992	20:45
Air Blank		0.000	20:46
Control Test		0.075	20:46
Air Blank		0.000	20:47
Diagnostics Check		OK	20:47

Data Download Date/Time: 04/30/2025 14:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,DESTINEE,R
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: SMITH,JONATHAN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/05/2025 22:44:59

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.079	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.251	22:39
Breath Volume		1.261	22:39
Air Blank		0.000	22:40
Air Blank		0.000	22:41
Subject Sample	# 2	0.231	22:43
Breath Volume		1.222	22:43
Air Blank		0.000	22:44
Control Test		0.076	22:44
Air Blank		0.000	22:44
Diagnostics Check		OK	22:44

Data Download Date/Time: 04/30/2025 14:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,DESTINEE,R
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: GOMEZ,DESTINEE
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/06/2025 05:26:34

Diagnostics Check		OK	05:18
Air Blank		0.000	05:19
Control Test		0.077	05:19
Air Blank		0.000	05:20
Subject Sample	# 1	0.096	05:21
Breath Volume		2.644	05:21
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.093	05:24
Breath Volume		2.574	05:24
Air Blank		0.000	05:25
Control Test		0.078	05:25
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 04/30/2025 14:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,DESTINEE,R
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: GOMEZ,DESTINEE
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 04/18/2025 19:43:21

Diagnostics Check		OK	19:36
Air Blank		0.000	19:36
Control Test		0.078	19:36
Air Blank		0.000	19:37
Subject Sample	# 1	0.157	19:38
Breath Volume		1.410	19:38
Air Blank		0.000	19:39
Air Blank		0.000	19:40
Subject Sample	# 2	0.156	19:41
Breath Volume		1.785	19:41
Air Blank		0.000	19:42
Control Test		0.078	19:42
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 04/30/2025 14:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: BOYLES

DATE/TIME OF TEST: 04/16/2025 02:16:52

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.079	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.078	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: SLAUGHTER

DATE/TIME OF TEST: 04/17/2025 00:56:34

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.078	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.099	00:52
Breath Volume		2.312	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.105	00:55
Breath Volume		2.546	00:55
Air Blank		0.000	00:55
Control Test		0.078	00:55
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: WHIGHAM

DATE/TIME OF TEST: 04/17/2025 05:36:35

Diagnostics Check		OK	05:30
Air Blank		0.000	05:30
Control Test		0.078	05:30
Air Blank		0.000	05:31
Subject Sample	# 1	0.152	05:32
Breath Volume		2.132	05:32
Air Blank		0.000	05:32
Air Blank		0.000	05:34
Subject Sample	# 2	0.161	05:35
Breath Volume		2.484	05:35
Air Blank		0.000	05:35
Control Test		0.078	05:35
Air Blank		0.000	05:36
Diagnostics Check		OK	05:36

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: JENSEN

DATE/TIME OF TEST: 04/17/2025 23:50:32

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:48	
Control Test		0.078	23:48	
Air Blank		0.000	23:49	
Subject Sample	# 1		23:49	Subject Test Refused
Breath Volume		0.000	23:49	
Air Blank		0.000	23:49	
Control Test		0.078	23:49	
Air Blank		0.000	23:50	
Diagnostics Check		OK	23:50	

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEIS,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: LAMBERSON

DATE/TIME OF TEST: 04/18/2025 01:49:37

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:47	
Control Test		0.079	01:47	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:48	Subject Test Refused
Breath Volume		0.000	01:48	
Air Blank		0.000	01:48	
Control Test		0.078	01:49	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARTIN,T
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/18/2025 04:12:43

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.078	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.011	04:08
Breath Volume		4.101	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.012	04:11
Breath Volume		4.671	04:11
Air Blank		0.000	04:11
Control Test		0.079	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLADDEN,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:56

SUBJECT'S LAST NAME: KAROBIA

DATE/TIME OF TEST: 04/24/2025 09:30:04

Diagnostics Check		OK	09:16	
Air Blank		0.000	09:17	
Control Test		0.078	09:17	
Air Blank		0.000	09:17	
Subject Sample	# 1	0.184	09:21	Volume Not Met
Breath Volume		0.714	09:21	
Air Blank		0.000	09:21	
Air Blank		0.000	09:23	
Subject Sample	# 2	0.198	09:25	
Breath Volume		1.152	09:25	
Air Blank		0.000	09:26	
Air Blank		0.000	09:27	
Subject Sample	# 3	0.205	09:28	
Breath Volume		1.617	09:28	
Air Blank		0.000	09:29	
Control Test		0.077	09:29	
Air Blank		0.000	09:29	
Diagnostics Check		OK	09:30	

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEIS,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/25/2025 00:38:21

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.078	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.118	00:33
Breath Volume		2.132	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.111	00:36
Breath Volume		2.160	00:36
Air Blank		0.000	00:37
Control Test		0.078	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROSADO,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 04/25/2025 05:27:41

Diagnostics Check		OK	05:21
Air Blank		0.000	05:21
Control Test		0.078	05:22
Air Blank		0.000	05:22
Subject Sample	# 1	0.170	05:23
Breath Volume		2.183	05:23
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.179	05:26
Breath Volume		2.566	05:26
Air Blank		0.000	05:26
Control Test		0.079	05:27
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: SAPIENZA

DATE/TIME OF TEST: 04/25/2025 18:24:23

Diagnostics Check		OK	18:21	
Air Blank		0.000	18:21	
Control Test		0.077	18:22	
Air Blank		0.000	18:22	
Subject Sample	# 1		18:23	Subject Test Refused
Breath Volume		0.000	18:23	
Air Blank		0.000	18:23	
Control Test		0.078	18:23	
Air Blank		0.000	18:24	
Diagnostics Check		OK	18:24	

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEIS,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: LUNDQUIST

DATE/TIME OF TEST: 04/26/2025 03:19:04

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.019	03:14
Breath Volume		1.675	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.018	03:17
Breath Volume		2.125	03:17
Air Blank		0.000	03:18
Control Test		0.078	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:19

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/26/2025 04:28:44

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.078	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.174	04:23
Breath Volume		1.949	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.178	04:27
Breath Volume		2.585	04:27
Air Blank		0.000	04:27
Control Test		0.078	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLADDEN,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRY,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:56

SUBJECT'S LAST NAME: BREED

DATE/TIME OF TEST: 04/27/2025 09:23:45

Diagnostics Check		OK	09:16
Air Blank		0.000	09:17
Control Test		0.079	09:17
Air Blank		0.000	09:18
Subject Sample	# 1	0.118	09:19
Breath Volume		3.101	09:19
Air Blank		0.000	09:19
Air Blank		0.000	09:21
Subject Sample	# 2	0.115	09:22
Breath Volume		2.949	09:22
Air Blank		0.000	09:22
Control Test		0.077	09:23
Air Blank		0.000	09:23
Diagnostics Check		OK	09:23

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 04/27/2025 22:27:17

Diagnostics Check		OK	22:20
Air Blank		0.000	22:21
Control Test		0.078	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.073	22:22
Breath Volume		2.496	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.068	22:25
Breath Volume		1.230	22:25
Air Blank		0.000	22:26
Control Test		0.078	22:26
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 04/27/2025 23:54:05

Diagnostics Check		OK	23:48
Air Blank		0.000	23:48
Control Test		0.078	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.000	23:49
Breath Volume		4.703	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.000	23:52
Breath Volume		4.105	23:52
Air Blank		0.000	23:53
Control Test		0.078	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: CAMACHO

DATE/TIME OF TEST: 04/28/2025 00:54:12

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.079	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.120	00:49
Breath Volume		1.316	00:49
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.117	00:52
Breath Volume		1.152	00:52
Air Blank		0.000	00:53
Control Test		0.078	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 05/13/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ECHEVERRIAVIDAL,X
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 04/28/2025 02:16:35

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:08	
Control Test		0.078	02:08	
Air Blank		0.000	02:09	
Subject Sample	# 1	0.118	02:12	Volume Not Met
Breath Volume		0.097	02:12	
Air Blank		0.000	02:13	
Air Blank		0.000	02:15	
Subject Sample	# 2		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.076	02:15	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 05/13/2025 15:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: FLUGENCE

DATE/TIME OF TEST: 04/28/2025 04:15:05

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:12	
Control Test		0.078	04:13	
Air Blank		0.000	04:13	
Subject Sample	# 1		04:13	Subject Test Refused
Breath Volume		0.000	04:13	
Air Blank		0.000	04:14	
Control Test		0.077	04:14	
Air Blank		0.000	04:14	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 05/13/2025 15:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: GREENE,DEVIN
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: LOZA-JIMENEZ

DATE/TIME OF TEST: 04/20/2025 00:44:47

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:42	
Control Test		0.080	00:42	
Air Blank		0.000	00:43	
Subject Sample	# 1		00:43	Subject Test Refused
Breath Volume		0.000	00:43	
Air Blank		0.000	00:43	
Control Test		0.081	00:44	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 04/30/2025 16:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,LAMAR,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: GARDNER,JEVON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 13:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:07

SUBJECT'S LAST NAME: STIEREN

DATE/TIME OF TEST: 04/23/2025 14:35:18

Diagnostics Check		OK	14:28
Air Blank		0.000	14:29
Control Test		0.081	14:29
Air Blank		0.000	14:29
Subject Sample	# 1	0.276	14:30
Breath Volume		3.203	14:30
Air Blank		0.000	14:31
Air Blank		0.000	14:33
Subject Sample	# 2	0.275	14:33
Breath Volume		3.457	14:33
Air Blank		0.000	14:34
Control Test		0.079	14:34
Air Blank		0.000	14:35
Diagnostics Check		OK	14:35

Data Download Date/Time: 04/30/2025 16:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO MELENDEZ,JOSE,A
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: CRISERA,M
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: OCASIO MELENDEZ

DATE/TIME OF TEST: 04/27/2025 19:07:07

Diagnostics Check		OK	19:01
Air Blank		0.000	19:01
Control Test		0.081	19:01
Air Blank		0.000	19:02
Subject Sample	# 1	0.068	19:02
Breath Volume		2.546	19:02
Air Blank		0.000	19:03
Air Blank		0.000	19:05
Subject Sample	# 2	0.066	19:05
Breath Volume		2.394	19:05
Air Blank		0.000	19:06
Control Test		0.080	19:06
Air Blank		0.000	19:07
Diagnostics Check		OK	19:07

Data Download Date/Time: 04/30/2025 16:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEES,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: PLANT,ROBERT
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402635303
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 10:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:21

SUBJECT'S LAST NAME: AJAO

DATE/TIME OF TEST: 04/09/2025 13:53:52

Diagnostics Check		OK	13:47
Air Blank		0.000	13:47
Control Test		0.081	13:47
Air Blank		0.000	13:48
Subject Sample	# 1	0.000	13:49
Breath Volume		3.347	13:49
Air Blank		0.000	13:49
Air Blank		0.000	13:51
Subject Sample	# 2	0.000	13:52
Breath Volume		2.390	13:52
Air Blank		0.000	13:52
Control Test		0.081	13:53
Air Blank		0.000	13:53
Diagnostics Check		OK	13:53

Data Download Date/Time: 04/10/2025 10:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUSKES,ARTHUR,A
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: BARRIOS,HECTOR
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402635303
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: ADMIN
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: HUEGERICH

DATE/TIME OF TEST: 04/28/2025 09:18:10

Diagnostics Check		OK	09:12
Air Blank		0.000	09:12
Control Test		0.080	09:12
Air Blank		0.000	09:13
Subject Sample	# 1	0.038	09:13
Breath Volume		2.023	09:13
Air Blank		0.000	09:14
Air Blank		0.000	09:16
Subject Sample	# 2	0.036	09:16
Breath Volume		2.640	09:16
Air Blank		0.000	09:17
Control Test		0.079	09:17
Air Blank		0.000	09:18
Diagnostics Check		OK	09:18

Data Download Date/Time: 05/20/2025 08:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: OGLE,SHAWN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: CALABRESE

DATE/TIME OF TEST: 04/27/2025 00:00:18

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:57	
Control Test		0.079	23:58	
Air Blank		0.000	23:58	
Subject Sample	# 1		23:58	Subject Test Refused
Breath Volume		0.000	23:58	
Air Blank		0.000	23:59	
Control Test		0.079	23:59	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 04/28/2025 14:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.