Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGDON,TODD,S
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HODGDON,TODD
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 05/10/2025 22:54:53

Diagnostics Check		OK	22:48
Air Blank		0.000	22:49
Control Test		0.081	22:49
Air Blank		0.000	22:49
Subject Sample	# 1	0.178	22:50
Breath Volume		1.417	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:52
Subject Sample	# 2	0.175	22:53
Breath Volume		1.792	22:53
Air Blank		0.000	22:53
Control Test		0.080	22:54
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 05/23/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIARA,KENDALL,R
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/13/2025 17:43:11

Diagnostics Check		OK	17:36
Air Blank		0.000	17:36
Control Test		0.080	17:37
Air Blank		0.000	17:37
Subject Sample	# 1	0.304	17:38
Breath Volume		2.851	17:38
Air Blank		0.000	17:39
Air Blank		0.000	17:40
Subject Sample	# 2	0.305	17:41
Breath Volume		2.531	17:41
Air Blank		0.000	17:42
Control Test		0.079	17:42
Air Blank		0.000	17:43
Diagnostics Check		OK	17:43

Data Download Date/Time: 05/23/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNETT,JAMES,C
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: YUERGENS,AARON
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: CAMPOS BALTODANO

DATE/TIME OF TEST: 05/14/2025 00:49:15

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.079	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.203	00:44
Breath Volume		1.250	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.187	00:47
Breath Volume		1.316	00:47
Air Blank		0.000	00:48
Control Test		0.079	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 05/23/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMOR,CHELO,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: 0458,CHELO ZAMOR
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: TIMKO

DATE/TIME OF TEST: 05/17/2025 05:56:15

Diagnostics Check		OK	05:49
Air Blank		0.000	05:49
Control Test		0.080	05:50
Air Blank		0.000	05:50
Subject Sample	# 1	0.285	05:51
Breath Volume		3.984	05:51
Air Blank		0.000	05:52
Air Blank		0.000	05:53
Subject Sample	# 2	0.284	05:54
Breath Volume		4.101	05:54
Air Blank		0.000	05:55
Control Test		0.077	05:55
Air Blank		0.000	05:56
Diagnostics Check		OK	05:56

Data Download Date/Time: 05/23/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTSON,BRANDON,K
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: JENKINSON,THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: JIMENEZ ROBLES

DATE/TIME OF TEST: 05/19/2025 03:43:34

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.000	03:39
Breath Volume		1.691	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.000	03:42
Breath Volume		2.218	03:42
Air Blank		0.000	03:42
Control Test		0.080	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 05/23/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/01/2025 02:05:54

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.130	02:01
Breath Volume		2.613	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:03
Subject Sample	# 2	0.128	02:04
Breath Volume		2.011	02:04
Air Blank		0.000	02:04
Control Test		0.078	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/02/2025 03:49:05

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.078	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.176	03:44
Breath Volume		2.121	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.167	03:47
Breath Volume		2.652	03:47
Air Blank		0.000	03:48
Control Test		0.078	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: KILIAN,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: MILEY

DATE/TIME OF TEST: 05/03/2025 22:30:56

Diagnostics Check		OK	22:24
Air Blank		0.000	22:25
Control Test		0.078	22:25
Air Blank		0.000	22:26
Subject Sample	# 1	0.219	22:26
Breath Volume		1.984	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:28
Subject Sample	# 2	0.220	22:29
Breath Volume		1.792	22:29
Air Blank		0.000	22:30
Control Test		0.077	22:30
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAYSON,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SMITH,JAYSON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: STEARNS

DATE/TIME OF TEST: 05/04/2025 17:13:22

Diagnostics Check		OK	17:10	
Air Blank		0.000	17:10	
Control Test		0.079	17:11	
Air Blank		0.000	17:11	
Subject Sample	# 1		17:12	Subject Test Refused
Breath Volume		0.000	17:12	
Air Blank		0.000	17:12	
Control Test		0.079	17:12	
Air Blank		0.000	17:13	
Diagnostics Check		OK	17:13	

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ZVOSECHZ

DATE/TIME OF TEST: 05/07/2025 03:55:35

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:53	
Control Test		0.079	03:53	
Air Blank		0.000	03:53	
Subject Sample	# 1		03:54	Subject Test Refused
Breath Volume		0.000	03:54	
Air Blank		0.000	03:54	
Control Test		0.078	03:54	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: CREVELLO

DATE/TIME OF TEST: 05/09/2025 22:31:32

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.078	22:26
Air Blank		0.000	22:26
Subject Sample	# 1	0.000	22:27
Breath Volume		1.593	22:27
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.000	22:30
Breath Volume		1.402	22:30
Air Blank		0.000	22:30
Control Test		0.078	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: GOTT

DATE/TIME OF TEST: 05/10/2025 03:12:30

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:09	
Control Test		0.079	03:10	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:11	Subject Test Refused
Breath Volume		0.000	03:11	
Air Blank		0.000	03:11	
Control Test		0.078	03:11	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: STARRY

DATE/TIME OF TEST: 05/11/2025 01:25:09

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.078	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.087	01:20
Breath Volume		1.140	01:20
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.085	01:23
Breath Volume		1.589	01:23
Air Blank		0.000	01:24
Control Test		0.077	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: COLVIN

DATE/TIME OF TEST: 05/14/2025 22:19:41

Diagnostics Check		OK	22:13
Air Blank		0.000	22:13
Control Test		0.078	22:14
Air Blank		0.000	22:14
Subject Sample	# 1	0.000	22:15
Breath Volume		2.425	22:15
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.000	22:18
Breath Volume		1.789	22:18
Air Blank		0.000	22:18
Control Test		0.078	22:19
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: SNELL

DATE/TIME OF TEST: 05/16/2025 19:56:43

Diagnostics Check		OK	19:50
Air Blank		0.000	19:50
Control Test		0.079	19:51
Air Blank		0.000	19:51
Subject Sample	# 1	0.023	19:52
Breath Volume		2.636	19:52
Air Blank		0.000	19:52
Air Blank		0.000	19:54
Subject Sample	# 2	0.020	19:55
Breath Volume		2.320	19:55
Air Blank		0.000	19:55
Control Test		0.078	19:56
Air Blank		0.000	19:56
Diagnostics Check		OK	19:56

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 05/16/2025 23:41:35

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.079	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.165	23:36
Breath Volume		1.851	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.167	23:39
Breath Volume		1.792	23:39
Air Blank		0.000	23:40
Control Test		0.077	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SIMPKINS,BENJAMIN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 05/17/2025 01:15:19

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:01	
Control Test		0.079	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:05	No Sample Provided
Breath Volume		0.000	01:05	
Air Blank		0.000	01:06	
Air Blank		0.000	01:08	
Subject Sample	# 2	0.147	01:10	
Breath Volume		2.738	01:10	
Air Blank		0.000	01:11	
Air Blank		0.000	01:13	
Subject Sample	# 3	0.141	01:13	
Breath Volume		2.593	01:13	
Air Blank		0.000	01:14	
Control Test		0.077	01:14	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REDZEPI,IMAN,
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: TRIEGLAFF,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: FERGERSON

DATE/TIME OF TEST: 05/18/2025 04:33:26

Diagnostics Check		OK	04:27	
Air Blank		0.000	04:28	
Control Test		0.078	04:28	
Air Blank		0.000	04:29	
Subject Sample	# 1		04:32	Subject Test Refused
Breath Volume		0.000	04:32	
Air Blank		0.000	04:32	
Control Test		0.077	04:32	
Air Blank		0.000	04:33	
Diagnostics Check		OK	04:33	

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REDZEPI,IMAN,
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: TRIEGLAFF,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: FERGERSON

DATE/TIME OF TEST: 05/18/2025 04:48:23

Diagnostics Check		OK	04:36	
Air Blank		0.000	04:37	
Control Test		0.078	04:37	
Air Blank		0.000	04:38	
Subject Sample	# 1		04:41	No Sample Provided
Breath Volume		0.000	04:41	
Air Blank		0.000	04:41	
Air Blank		0.000	04:43	
Subject Sample	# 2	0.000	04:46	Volume Not Met
Breath Volume		0.070	04:46	
Air Blank		0.000	04:47	
Control Test		0.078	04:47	
Air Blank		0.000	04:48	
Diagnostics Check		OK	04:48	

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: LYTTLE

DATE/TIME OF TEST: 05/21/2025 02:27:34

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.078	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.146	02:22
Breath Volume		1.941	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.144	02:25
Breath Volume		1.406	02:25
Air Blank		0.000	02:26
Control Test		0.078	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: SPRAGUE

DATE/TIME OF TEST: 05/21/2025 23:25:51

Diagnostics Check		OK	23:19
Air Blank		0.000	23:20
Control Test		0.078	23:20
Air Blank		0.000	23:20
Subject Sample	# 1	0.178	23:21
Breath Volume		3.238	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:23
Subject Sample	# 2	0.174	23:24
Breath Volume		2.859	23:24
Air Blank		0.000	23:24
Control Test		0.078	23:25
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESSLER,GREYSON,L
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: RESSLER,GREYSON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: PEREZ ZEPEDA

DATE/TIME OF TEST: 05/24/2025 22:23:01

Diagnostics Check		OK	22:16
Air Blank		0.000	22:17
Control Test		0.078	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.185	22:18
Breath Volume		2.687	22:18
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.178	22:21
Breath Volume		2.386	22:21
Air Blank		0.000	22:22
Control Test		0.077	22:22
Air Blank		0.000	22:22
Diagnostics Check		OK	22:23

Data Download Date/Time: 06/04/2025 17:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOX,GRAHAM,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ALNWEIRAN,OMAR
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 05/26/2025 20:07:28

Diagnostics Check		OK	20:04	
Air Blank		0.000	20:04	
Control Test		0.078	20:04	
Air Blank		0.000	20:05	
Subject Sample	# 1		20:06	Subject Test Refused
Breath Volume		0.000	20:06	
Air Blank		0.000	20:06	
Control Test		0.078	20:06	
Air Blank		0.000	20:07	
Diagnostics Check		OK	20:07	

Data Download Date/Time: 06/04/2025 17:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:33

SUBJECT'S LAST NAME: LANPHEAR

DATE/TIME OF TEST: 05/27/2025 11:07:20

Diagnostics Check		OK	10:59
Air Blank		0.000	10:59
Control Test		0.079	10:59
Air Blank		0.000	11:00
Subject Sample	# 1	0.098	11:02
Breath Volume		2.000	11:02
Air Blank		0.000	11:02
Air Blank		0.000	11:04
Subject Sample	# 2	0.090	11:05
Breath Volume		1.257	11:05
Air Blank		0.000	11:06
Control Test		0.078	11:06
Air Blank		0.000	11:07
Diagnostics Check		OK	11:07

Data Download Date/Time: 06/04/2025 17:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAYSON,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SMITH,JAYSON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: GOODE

DATE/TIME OF TEST: 05/28/2025 22:43:36

Diagnostics Check		OK	22:40	
Air Blank		0.000	22:41	
Control Test		0.079	22:41	
Air Blank		0.000	22:42	
Subject Sample	# 1		22:42	Subject Test Refused
Breath Volume		0.000	22:42	
Air Blank		0.000	22:42	
Control Test		0.079	22:43	
Air Blank		0.000	22:43	
Diagnostics Check		OK	22:43	

Data Download Date/Time: 06/04/2025 17:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAYSON,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SMITH,JAYSON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 05/31/2025 20:18:31

Diagnostics Check		OK	20:12
Air Blank		0.000	20:12
Control Test		0.078	20:13
Air Blank		0.000	20:13
Subject Sample	# 1	0.180	20:14
Breath Volume		2.871	20:14
Air Blank		0.000	20:14
Air Blank		0.000	20:16
Subject Sample	# 2	0.184	20:16
Breath Volume		2.964	20:16
Air Blank		0.000	20:17
Control Test		0.078	20:17
Air Blank		0.000	20:18
Diagnostics Check		OK	20:18

Data Download Date/Time: 06/04/2025 17:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LIGHTFOOT,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: GOMEZ PEREZ

DATE/TIME OF TEST: 05/01/2025 23:30:46

Diagnostics Check		OK	23:23
Air Blank		0.000	23:24
Control Test		0.079	23:24
Air Blank		0.000	23:25
Subject Sample	# 1	0.053	23:26
Breath Volume		1.839	23:26
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.051	23:29
Breath Volume		1.800	23:29
Air Blank		0.000	23:29
Control Test		0.080	23:30
Air Blank		0.000	23:30
Diagnostics Check		OK	23:30

Data Download Date/Time: 05/03/2025 00:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASCKO,J,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: DODSON,H
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:46

SUBJECT'S LAST NAME: VICKERY

DATE/TIME OF TEST: 05/02/2025 13:16:26

Diagnostics Check		OK	13:09
Air Blank		0.000	13:10
Control Test		0.081	13:10
Air Blank		0.000	13:10
Subject Sample	# 1	0.000	13:12
Breath Volume		1.992	13:12
Air Blank		0.000	13:12
Air Blank		0.000	13:14
Subject Sample	# 2	0.000	13:15
Breath Volume		1.859	13:15
Air Blank		0.000	13:15
Control Test		0.080	13:15
Air Blank		0.000	13:16
Diagnostics Check		OK	13:16

Data Download Date/Time: 05/03/2025 00:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/02/2025 21:30:52

Diagnostics Check		OK	21:24
Air Blank		0.000	21:25
Control Test		0.080	21:25
Air Blank		0.000	21:25
Subject Sample	# 1	0.000	21:26
Breath Volume		2.750	21:26
Air Blank		0.000	21:26
Air Blank		0.000	21:28
Subject Sample	# 2	0.000	21:29
Breath Volume		1.484	21:29
Air Blank		0.000	21:29
Control Test		0.084	21:30
Air Blank		0.000	21:30
Diagnostics Check		OK	21:30

Data Download Date/Time: 05/03/2025 00:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: NILES

DATE/TIME OF TEST: 05/02/2025 23:36:46

Diagnostics Check		OK	23:28
Air Blank		0.000	23:29
Control Test		0.081	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.224	23:32
Breath Volume		3.394	23:32
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.206	23:35
Breath Volume		1.882	23:35
Air Blank		0.000	23:35
Control Test		0.078	23:36
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 05/03/2025 00:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/02/2025 23:46:56

Diagnostics Check		OK	23:39
Air Blank		0.000	23:40
Control Test		0.079	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.145	23:42
Breath Volume		2.031	23:42
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.144	23:45
Breath Volume		1.699	23:45
Air Blank		0.000	23:45
Control Test		0.079	23:46
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 05/03/2025 00:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GIARDINO,C
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: PLUMADORE-LIPPERSON

DATE/TIME OF TEST: 05/06/2025 02:22:22

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.083	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.089	02:17
Breath Volume		2.019	02:17
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.083	02:20
Breath Volume		2.097	02:20
Air Blank		0.000	02:21
Control Test		0.077	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 05/31/2025 03:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REICH,OWEN,S
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REICH,OWEN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:28

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 05/07/2025 15:55:53

Diagnostics Check		OK	15:49
Air Blank		0.000	15:50
Control Test		0.081	15:50
Air Blank		0.000	15:51
Subject Sample	# 1	0.000	15:51
Breath Volume		1.992	15:51
Air Blank		0.000	15:52
Air Blank		0.000	15:54
Subject Sample	# 2	0.000	15:54
Breath Volume		1.925	15:54
Air Blank		0.000	15:54
Control Test		0.080	15:55
Air Blank		0.000	15:55
Diagnostics Check		OK	15:55

Data Download Date/Time: 05/31/2025 03:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIE,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CAPONE,M
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/08/2025 17:58:28

Diagnostics Check		OK	17:51
Air Blank		0.000	17:51
Control Test		0.081	17:51
Air Blank		0.000	17:52
Subject Sample	# 1	0.144	17:53
Breath Volume		2.003	17:53
Air Blank		0.000	17:54
Air Blank		0.000	17:56
Subject Sample	# 2	0.145	17:56
Breath Volume		1.585	17:56
Air Blank		0.000	17:57
Control Test		0.078	17:57
Air Blank		0.000	17:58
Diagnostics Check		OK	17:58

Data Download Date/Time: 05/31/2025 03:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: ANDRUS

DATE/TIME OF TEST: 05/16/2025 23:29:38

Diagnostics Check		OK	23:26	
Air Blank		0.000	23:26	
Control Test		0.080	23:26	
Air Blank		0.000	23:27	
Subject Sample	# 1		23:28	Subject Test Refused
Breath Volume		0.000	23:28	
Air Blank		0.000	23:28	
Control Test		0.081	23:29	
Air Blank		0.000	23:29	
Diagnostics Check		OK	23:29	

Data Download Date/Time: 05/31/2025 03:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MORALES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MAXFIELD

DATE/TIME OF TEST: 05/17/2025 00:38:37

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.081	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.121	00:34
Breath Volume		2.226	00:34
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.121	00:37
Breath Volume		1.890	00:37
Air Blank		0.000	00:37
Control Test		0.081	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 05/31/2025 03:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: GODWIN THORNTON

DATE/TIME OF TEST: 05/17/2025 03:18:01

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.191	03:12
Breath Volume		1.289	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.176	03:16
Breath Volume		1.171	03:16
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:18

Data Download Date/Time: 05/31/2025 03:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIE,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: WOODIE,L
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: TAJAH

DATE/TIME OF TEST: 05/18/2025 01:23:21

Diagnostics Check		OK	01:14	
Air Blank		0.000	01:14	
Control Test		0.081	01:15	
Air Blank		0.000	01:15	
Subject Sample	# 1	0.193	01:18	Volume Not Met
Breath Volume		0.070	01:18	
Air Blank		0.000	01:19	
Air Blank		0.000	01:21	
Subject Sample	# 2		01:21	Subject Test Refused
Breath Volume		0.000	01:21	
Air Blank		0.000	01:22	
Control Test		0.076	01:22	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 05/31/2025 03:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ZOFNAS

DATE/TIME OF TEST: 05/24/2025 23:57:57

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.081	23:51
Air Blank		0.000	23:52
Subject Sample	# 1	0.155	23:52
Breath Volume		1.378	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.166	23:56
Breath Volume		1.101	23:56
Air Blank		0.000	23:57
Control Test		0.080	23:57
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 05/31/2025 03:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JARROD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: DURRANCE,PAUL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:54

SUBJECT'S LAST NAME: SANTIAGO MERCED

DATE/TIME OF TEST: 05/25/2025 13:22:23

Diagnostics Check		OK	13:14
Air Blank		0.000	13:15
Control Test		0.081	13:15
Air Blank		0.000	13:16
Subject Sample	# 1	0.000	13:18
Breath Volume		2.058	13:18
Air Blank		0.000	13:18
Air Blank		0.000	13:20
Subject Sample	# 2	0.000	13:21
Breath Volume		3.664	13:21
Air Blank		0.000	13:21
Control Test		0.081	13:21
Air Blank		0.000	13:22
Diagnostics Check		OK	13:22

Data Download Date/Time: 05/31/2025 03:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: DE JESUS

DATE/TIME OF TEST: 05/25/2025 23:15:36

Diagnostics Check		OK	23:05
Air Blank		0.000	23:05
Control Test		0.081	23:06
Air Blank		0.000	23:06
Subject Sample	# 1	0.148	23:09
Breath Volume		1.398	23:09
Air Blank		0.000	23:10
Air Blank		0.000	23:12
Subject Sample	# 2	0.150	23:13
Breath Volume		1.117	23:13
Air Blank		0.000	23:14
Control Test		0.078	23:14
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 05/31/2025 03:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DEUEN,COLE,C
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VAN DEUSEN,COLE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: FONTANEZ

DATE/TIME OF TEST: 05/26/2025 00:08:07

Diagnostics Check		OK	00:01
Air Blank		0.000	00:02
Control Test		0.080	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.000	00:03
Breath Volume		2.132	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.000	00:06
Breath Volume		2.050	00:06
Air Blank		0.000	00:07
Control Test		0.080	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 05/31/2025 03:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: RODRIGUEZ RIVERA

DATE/TIME OF TEST: 05/26/2025 03:38:17

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.079	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.263	03:33
Breath Volume		2.746	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:35
Subject Sample	# 2	0.273	03:36
Breath Volume		2.363	03:36
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 05/31/2025 03:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SANCHEZ-HERNANDEZ

DATE/TIME OF TEST: 05/29/2025 23:41:32

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:38	
Control Test		0.082	23:39	
Air Blank		0.000	23:39	
Subject Sample	# 1		23:39	Slope Not Met
Breath Volume		0.820	23:39	
Air Blank		0.000	23:40	
Control Test		0.081	23:40	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 05/31/2025 03:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SANCHEZ-HERNANDEZ

DATE/TIME OF TEST: 05/29/2025 23:51:59

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.080	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.237	23:46
Breath Volume		1.574	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.236	23:50
Breath Volume		1.824	23:50
Air Blank		0.000	23:51
Control Test		0.077	23:51
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 05/31/2025 03:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PAULSON,HEATH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:03

SUBJECT'S LAST NAME: YERKES

DATE/TIME OF TEST: 05/03/2025 12:54:59

Diagnostics Check		OK	12:48
Air Blank		0.000	12:49
Control Test		0.081	12:49
Air Blank		0.000	12:49
Subject Sample	# 1	0.000	12:50
Breath Volume		3.066	12:50
Air Blank		0.000	12:51
Air Blank		0.000	12:53
Subject Sample	# 2	0.000	12:53
Breath Volume		2.992	12:53
Air Blank		0.000	12:54
Control Test		0.081	12:54
Air Blank		0.000	12:54
Diagnostics Check		OK	12:54

Data Download Date/Time: 05/09/2025 14:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/04/2025 00:19:51

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.081	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.231	00:15
Breath Volume		2.085	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.230	00:18
Breath Volume		1.796	00:18
Air Blank		0.000	00:18
Control Test		0.080	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 05/09/2025 14:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SANCHEZ PINEDA

DATE/TIME OF TEST: 05/04/2025 02:39:23

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.081	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.106	02:34
Breath Volume		2.328	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.108	02:37
Breath Volume		2.054	02:37
Air Blank		0.000	02:38
Control Test		0.081	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 05/09/2025 14:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DICKERSON,JOHN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: CASTANEDA MEDINA

DATE/TIME OF TEST: 05/05/2025 03:25:45

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.081	03:20
Air Blank		0.000	03:20
Subject Sample	# 1	0.205	03:21
Breath Volume		2.082	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.204	03:24
Breath Volume		2.347	03:24
Air Blank		0.000	03:24
Control Test		0.081	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 05/09/2025 14:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: SAQUIC SUY

DATE/TIME OF TEST: 05/05/2025 05:15:17

Diagnostics Check		OK	05:08
Air Blank		0.000	05:08
Control Test		0.082	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.169	05:10
Breath Volume		2.007	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.173	05:13
Breath Volume		1.980	05:13
Air Blank		0.000	05:14
Control Test		0.081	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 05/09/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: VENKAT

DATE/TIME OF TEST: 05/06/2025 00:37:09

Diagnostics Check		OK	00:28
Air Blank		0.000	00:29
Control Test		0.082	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.126	00:32
Breath Volume		1.160	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.134	00:35
Breath Volume		1.472	00:35
Air Blank		0.000	00:36
Control Test		0.080	00:36
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 05/09/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIER,MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: CHAVIS

DATE/TIME OF TEST: 05/06/2025 03:59:18

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.081	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.000	03:54
Breath Volume		3.460	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.000	03:57
Breath Volume		4.289	03:57
Air Blank		0.000	03:58
Control Test		0.081	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 05/09/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: DIAZ ESCALANTE

DATE/TIME OF TEST: 05/10/2025 00:48:21

Diagnostics Check		OK	00:41
Air Blank		0.000	00:42
Control Test		0.081	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.242	00:43
Breath Volume		1.589	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.248	00:46
Breath Volume		1.929	00:46
Air Blank		0.000	00:47
Control Test		0.080	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 06/20/2025 16:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: MONTOYA MORA

DATE/TIME OF TEST: 05/10/2025 02:16:31

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.081	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.158	02:12
Breath Volume		1.996	02:12
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.156	02:14
Breath Volume		1.894	02:14
Air Blank		0.000	02:15
Control Test		0.080	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 06/20/2025 16:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 05/10/2025 23:22:52

Diagnostics Check		OK	23:16
Air Blank		0.000	23:16
Control Test		0.081	23:16
Air Blank		0.000	23:17
Subject Sample	# 1	0.029	23:18
Breath Volume		1.617	23:18
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.027	23:21
Breath Volume		1.140	23:21
Air Blank		0.000	23:21
Control Test		0.081	23:22
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 06/20/2025 16:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MILORD,JHAMIL
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: BVUNZAWABAYA

DATE/TIME OF TEST: 05/10/2025 23:47:40

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.081	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.185	23:43
Breath Volume		2.199	23:43
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.184	23:46
Breath Volume		1.187	23:46
Air Blank		0.000	23:46
Control Test		0.080	23:47
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 06/20/2025 16:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: PEREZ MAZARIEGOS

DATE/TIME OF TEST: 05/11/2025 21:21:56

Diagnostics Check		OK	21:15	
Air Blank		0.000	21:16	
Control Test		0.081	21:16	
Air Blank		0.000	21:16	
Subject Sample	# 1	0.290	21:17	
Breath Volume		2.082	21:17	
Air Blank		0.000	21:18	
Air Blank		0.000	21:19	
Subject Sample	# 2	0.293	21:20	
Breath Volume		2.042	21:20	
Air Blank		0.000	21:20	
Control Test		0.087	21:21	Control Outside Tolerance
Air Blank		0.000	21:21	

Data Download Date/Time: 06/20/2025 16:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: PEREZ MAZARIEGOS

DATE/TIME OF TEST: 05/11/2025 21:49:32

Diagnostics Check		OK	21:43
Air Blank		0.000	21:43
Control Test		0.082	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.268	21:44
Breath Volume		2.457	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.274	21:47
Breath Volume		2.699	21:47
Air Blank		0.000	21:48
Control Test		0.080	21:48
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 06/20/2025 16:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 05/11/2025 22:07:34

Diagnostics Check		OK	22:00
Air Blank		0.000	22:00
Control Test		0.080	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.244	22:02
Breath Volume		1.808	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:05
Subject Sample	# 2	0.236	22:05
Breath Volume		2.218	22:05
Air Blank		0.000	22:06
Control Test		0.079	22:06
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 06/20/2025 16:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RODRIGUEZ,ANDRES
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: SALMERON

DATE/TIME OF TEST: 05/12/2025 02:44:20

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.081	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.178	02:39
Breath Volume		4.007	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.187	02:42
Breath Volume		3.035	02:42
Air Blank		0.000	02:43
Control Test		0.081	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 06/20/2025 16:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: REED,CHARLES
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: REMONVIL

DATE/TIME OF TEST: 05/14/2025 00:33:34

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:17	
Control Test		0.080	00:18	
Air Blank		0.000	00:18	
Subject Sample	# 1	0.240	00:20	
Breath Volume		1.558	00:20	
Air Blank		0.000	00:21	
Air Blank		0.000	00:23	
Subject Sample	# 2	0.212	00:26	Volume Not Met
Breath Volume		0.382	00:26	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 3		00:32	No Sample Provided
Breath Volume		0.000	00:32	
Air Blank		0.000	00:32	
Control Test		0.077	00:32	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 06/20/2025 16:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: REED,CHARLES
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: REMONVIL

DATE/TIME OF TEST: 05/14/2025 02:49:52

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.082	02:43
Air Blank		0.000	02:44
Subject Sample	# 1	0.223	02:44
Breath Volume		1.882	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.212	02:48
Breath Volume		2.339	02:48
Air Blank		0.000	02:48
Control Test		0.081	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 06/20/2025 16:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: FELIX JR

DATE/TIME OF TEST: 05/15/2025 02:39:06

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.081	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.219	02:34
Breath Volume		1.457	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.206	02:37
Breath Volume		1.273	02:37
Air Blank		0.000	02:38
Control Test		0.081	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 06/20/2025 16:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: EKLUND,HARRISON
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 05/15/2025 19:40:27

Diagnostics Check		OK	19:34
Air Blank		0.000	19:34
Control Test		0.081	19:35
Air Blank		0.000	19:35
Subject Sample	# 1	0.329	19:36
Breath Volume		1.515	19:36
Air Blank		0.000	19:36
Air Blank		0.000	19:38
Subject Sample	# 2	0.333	19:38
Breath Volume		1.277	19:38
Air Blank		0.000	19:39
Control Test		0.080	19:39
Air Blank		0.000	19:40
Diagnostics Check		OK	19:40

Data Download Date/Time: 06/20/2025 16:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: VICENTE ELIAS

DATE/TIME OF TEST: 05/15/2025 23:25:52

Diagnostics Check		OK	23:19
Air Blank		0.000	23:19
Control Test		0.081	23:20
Air Blank		0.000	23:20
Subject Sample	# 1	0.155	23:21
Breath Volume		3.121	23:21
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.156	23:24
Breath Volume		3.089	23:24
Air Blank		0.000	23:24
Control Test		0.080	23:25
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ARTEMIO

DATE/TIME OF TEST: 05/17/2025 03:06:23

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.081	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.187	03:01
Breath Volume		2.863	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.188	03:04
Breath Volume		2.750	03:04
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: GONZALEZ ESCALANTE

DATE/TIME OF TEST: 05/17/2025 23:24:46

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.081	23:18
Air Blank		0.000	23:19
Subject Sample	# 1	0.199	23:20
Breath Volume		1.941	23:20
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.197	23:23
Breath Volume		1.453	23:23
Air Blank		0.000	23:23
Control Test		0.081	23:24
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: PEREZ BORGES

DATE/TIME OF TEST: 05/18/2025 01:46:14

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.081	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.149	01:41
Breath Volume		1.703	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.149	01:44
Breath Volume		2.113	01:44
Air Blank		0.000	01:45
Control Test		0.081	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MAIER,MARTINA
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: BRAVO VASQUEZ

DATE/TIME OF TEST: 05/18/2025 23:45:03

Diagnostics Check		OK	23:38
Air Blank		0.000	23:39
Control Test		0.080	23:39
Air Blank		0.000	23:39
Subject Sample	# 1	0.351	23:40
Breath Volume		2.671	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:42
Subject Sample	# 2	0.353	23:43
Breath Volume		2.777	23:43
Air Blank		0.000	23:44
Control Test		0.080	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:45

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RONQUILLO,DANIEL
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: SANDOMENO

DATE/TIME OF TEST: 05/20/2025 18:45:02

Diagnostics Check		OK	18:38
Air Blank		0.000	18:39
Control Test		0.081	18:39
Air Blank		0.000	18:39
Subject Sample	# 1	0.305	18:40
Breath Volume		1.906	18:40
Air Blank		0.000	18:41
Air Blank		0.000	18:43
Subject Sample	# 2	0.298	18:43
Breath Volume		1.597	18:43
Air Blank		0.000	18:44
Control Test		0.080	18:44
Air Blank		0.000	18:44
Diagnostics Check		OK	18:45

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ANDRE,GAETGENS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 05/21/2025 00:01:26

Diagnostics Check		OK	23:55
Air Blank		0.000	23:55
Control Test		0.081	23:55
Air Blank		0.000	23:56
Subject Sample	# 1	0.169	23:56
Breath Volume		1.414	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 2	0.167	23:59
Breath Volume		1.992	23:59
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 05/21/2025 02:14:55

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:12	
Control Test		0.081	02:12	
Air Blank		0.000	02:13	
Subject Sample	# 1		02:13	Subject Test Refused
Breath Volume		0.000	02:13	
Air Blank		0.000	02:13	
Control Test		0.081	02:14	
Air Blank		0.000	02:14	
Diagnostics Check		OK	02:14	

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 05/21/2025 02:31:02

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.081	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.180	02:26
Breath Volume		3.277	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.176	02:29
Breath Volume		1.386	02:29
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:31

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 05/21/2025 23:51:30

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.081	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.212	23:46
Breath Volume		2.019	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:48
Subject Sample	# 2	0.213	23:49
Breath Volume		1.296	23:49
Air Blank		0.000	23:50
Control Test		0.080	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: PLATT

DATE/TIME OF TEST: 05/23/2025 04:07:33

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.081	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.114	04:03
Breath Volume		1.105	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.111	04:06
Breath Volume		1.710	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: ELIZALDE MARTINEZ

DATE/TIME OF TEST: 05/23/2025 23:21:57

Diagnostics Check		OK	23:14
Air Blank		0.000	23:15
Control Test		0.081	23:15
Air Blank		0.000	23:16
Subject Sample	# 1	0.313	23:17
Breath Volume		1.523	23:17
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.323	23:20
Breath Volume		2.027	23:20
Air Blank		0.000	23:20
Control Test		0.080	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: LUCATERO SAGRERO

DATE/TIME OF TEST: 05/24/2025 02:28:50

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.082	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.207	02:23
Breath Volume		2.308	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.204	02:27
Breath Volume		2.964	02:27
Air Blank		0.000	02:27
Control Test		0.080	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DORIVAL,ADLER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/24/2025 02:54:00

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.081	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.169	02:49
Breath Volume		1.562	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:51
Subject Sample	# 2	0.173	02:52
Breath Volume		1.414	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:53
Diagnostics Check		OK	02:54

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BANEGAS,CAMERON
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 05/25/2025 03:48:16

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.081	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.219	03:43
Breath Volume		1.351	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.224	03:46
Breath Volume		2.640	03:46
Air Blank		0.000	03:47
Control Test		0.081	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,BROOKE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 05/26/2025 03:23:24

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.081	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.233	03:17
Breath Volume		1.675	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.238	03:21
Breath Volume		1.992	03:21
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MONTOYA MAZANARES

DATE/TIME OF TEST: 05/27/2025 01:28:00

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.081	01:22
Air Blank		0.000	01:22
Subject Sample	# 1	0.249	01:23
Breath Volume		1.367	01:23
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.265	01:26
Breath Volume		2.476	01:26
Air Blank		0.000	01:27
Control Test		0.080	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:28

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LONDONO,JULIAN
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: PAGLIARELLA JR

DATE/TIME OF TEST: 05/27/2025 02:08:09

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.081	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.000	02:03
Breath Volume		3.261	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.000	02:06
Breath Volume		2.503	02:06
Air Blank		0.000	02:07
Control Test		0.081	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 05/28/2025 00:20:20

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.081	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.165	00:15
Breath Volume		1.210	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.160	00:18
Breath Volume		1.609	00:18
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MAROTO,VALERY
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/28/2025 01:45:17

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.081	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.165	01:40
Breath Volume		2.402	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.163	01:43
Breath Volume		2.851	01:43
Air Blank		0.000	01:44
Control Test		0.080	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DEITCH

DATE/TIME OF TEST: 05/29/2025 04:02:34

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.082	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.011	03:58
Breath Volume		1.789	03:58
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.009	04:01
Breath Volume		2.175	04:01
Air Blank		0.000	04:01
Control Test		0.081	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ,ALINA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: BLOCK

DATE/TIME OF TEST: 05/29/2025 22:55:55

Diagnostics Check		OK	22:49
Air Blank		0.000	22:50
Control Test		0.081	22:50
Air Blank		0.000	22:50
Subject Sample	# 1	0.203	22:51
Breath Volume		2.148	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:53
Subject Sample	# 2	0.198	22:54
Breath Volume		1.183	22:54
Air Blank		0.000	22:55
Control Test		0.081	22:55
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: DIEUJUSTE

DATE/TIME OF TEST: 05/30/2025 01:24:05

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.081	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.188	01:19
Breath Volume		1.375	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:21
Subject Sample	# 2	0.190	01:22
Breath Volume		1.375	01:22
Air Blank		0.000	01:23
Control Test		0.081	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/30/2025 02:55:11

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.081	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.196	02:50
Breath Volume		2.171	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.212	02:53
Breath Volume		1.800	02:53
Air Blank		0.000	02:54
Control Test		0.081	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 05/31/2025 01:30:04

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.082	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.208	01:25
Breath Volume		1.433	01:25
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 2	0.212	01:28
Breath Volume		1.500	01:28
Air Blank		0.000	01:29
Control Test		0.081	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:30

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BELMONTE ABZUETA

DATE/TIME OF TEST: 05/31/2025 03:16:53

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.081	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.168	03:12
Breath Volume		2.867	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:14
Subject Sample	# 2	0.172	03:15
Breath Volume		3.058	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LONDONO,JULIAN
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 05/31/2025 22:21:25

Diagnostics Check		OK	22:13
Air Blank		0.000	22:14
Control Test		0.081	22:14
Air Blank		0.000	22:15
Subject Sample	# 1	0.188	22:17
Breath Volume		1.460	22:17
Air Blank		0.000	22:17
Air Blank		0.000	22:19
Subject Sample	# 2	0.195	22:19
Breath Volume		1.179	22:19
Air Blank		0.000	22:20
Control Test		0.081	22:20
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: URBINA PLATA

DATE/TIME OF TEST: 05/31/2025 23:20:22

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.081	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.102	23:15
Breath Volume		2.207	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.100	23:18
Breath Volume		1.402	23:18
Air Blank		0.000	23:19
Control Test		0.082	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,LOGAN,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FARES,RAMEZ
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/17/2025 00:33:58

Diagnostics Check		OK	00:27
Air Blank		0.000	00:28
Control Test		0.078	00:28
Air Blank		0.000	00:28
Subject Sample	# 1	0.124	00:29
Breath Volume		2.320	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:31
Subject Sample	# 2	0.126	00:32
Breath Volume		1.917	00:32
Air Blank		0.000	00:33
Control Test		0.077	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,LOGAN,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VERYCKEN,MATTHEW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: CARRILLO-MATIAS

DATE/TIME OF TEST: 05/17/2025 18:34:01

Diagnostics Check		OK	18:26
Air Blank		0.000	18:27
Control Test		0.079	18:27
Air Blank		0.000	18:27
Subject Sample	# 1	0.197	18:29
Breath Volume		2.078	18:29
Air Blank		0.000	18:30
Air Blank		0.000	18:31
Subject Sample	# 2	0.189	18:32
Breath Volume		2.597	18:32
Air Blank		0.000	18:33
Control Test		0.076	18:33
Air Blank		0.000	18:33
Diagnostics Check		OK	18:34

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,LOGAN,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHITE,THOMAS
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: RUANO-GONZALEZ

DATE/TIME OF TEST: 05/17/2025 21:40:25

Diagnostics Check		OK	21:31
Air Blank		0.000	21:32
Control Test		0.079	21:32
Air Blank		0.000	21:33
Subject Sample	# 1	0.170	21:34
Breath Volume		2.207	21:34
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 2	0.167	21:38
Breath Volume		2.496	21:38
Air Blank		0.000	21:39
Control Test		0.076	21:39
Air Blank		0.000	21:40
Diagnostics Check		OK	21:40

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LUNGU

DATE/TIME OF TEST: 05/19/2025 00:09:34

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:06	
Control Test		0.079	00:06	
Air Blank		0.000	00:07	
Subject Sample	# 1	0.124	00:07	
Breath Volume		3.578	00:07	
Air Blank			00:08	RFI Detect
Air Blank		0.000	00:08	
Control Test		0.079	00:08	
Air Blank		0.000	00:09	
Diagnostics Check		OK	00:09	

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LUNGU

DATE/TIME OF TEST: 05/19/2025 00:18:42

Diagnostics Check		OK	00:11
Air Blank		0.000	00:12
Control Test		0.079	00:12
Air Blank		0.000	00:13
Subject Sample	# 1	0.120	00:13
Breath Volume		1.574	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.119	00:17
Breath Volume		3.210	00:17
Air Blank		0.000	00:17
Control Test		0.079	00:18
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,COLIN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CALDWELL,COLIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CABALLERO

DATE/TIME OF TEST: 05/25/2025 04:26:43

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.079	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.159	04:22
Breath Volume		1.410	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.169	04:25
Breath Volume		1.230	04:25
Air Blank		0.000	04:25
Control Test		0.079	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELACRUZ,JONATHAN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VERYCKEN,MATTHEW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HERNANDEZ-ORTIZ

DATE/TIME OF TEST: 05/31/2025 00:17:04

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.078	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.149	00:12
Breath Volume		3.160	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:14
Subject Sample	# 2	0.134	00:15
Breath Volume		2.992	00:15
Air Blank		0.000	00:16
Control Test		0.077	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:17

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARES,RAMEZ,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FARES,RAMEZ
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 05/31/2025 01:16:52

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.079	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.196	01:11
Breath Volume		1.152	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.192	01:15
Breath Volume		1.496	01:15
Air Blank		0.000	01:15
Control Test		0.077	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: MEJIA RECINOS

DATE/TIME OF TEST: 05/31/2025 01:26:05

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.077	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.164	01:21
Breath Volume		2.187	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:23
Subject Sample	# 2	0.155	01:24
Breath Volume		1.148	01:24
Air Blank		0.000	01:25
Control Test		0.075	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 05/01/2025 05:53:09

Diagnostics Check		OK	05:43	
Air Blank		0.000	05:43	
Control Test		0.081	05:43	
Air Blank		0.000	05:44	
Subject Sample	# 1	0.197	05:47	Volume Not Met
Breath Volume		0.523	05:47	
Air Blank		0.000	05:48	
Air Blank		0.000	05:50	
Subject Sample	# 2		05:51	Subject Test Refused
Breath Volume		0.035	05:51	
Air Blank		0.000	05:52	
Control Test		0.081	05:52	
Air Blank		0.000	05:53	
Diagnostics Check		OK	05:53	

Data Download Date/Time: 05/13/2025 12:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: JAMISON

DATE/TIME OF TEST: 05/02/2025 00:40:13

Diagnostics Check		OK	00:37	
Air Blank		0.000	00:37	
Control Test		0.081	00:38	
Air Blank		0.000	00:38	
Subject Sample	# 1		00:38	Subject Test Refused
Breath Volume		0.000	00:38	
Air Blank		0.000	00:39	
Control Test		0.080	00:39	
Air Blank		0.000	00:40	
Diagnostics Check		OK	00:40	

Data Download Date/Time: 05/13/2025 12:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: DEMPS

DATE/TIME OF TEST: 05/02/2025 03:10:12

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.082	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.135	03:05
Breath Volume		2.386	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.129	03:08
Breath Volume		1.671	03:08
Air Blank		0.000	03:09
Control Test		0.080	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 05/13/2025 12:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 05/02/2025 04:59:24

Diagnostics Check		OK	04:55	
Air Blank		0.000	04:56	
Control Test		0.081	04:56	
Air Blank		0.000	04:56	
Subject Sample	# 1		04:58	Subject Test Refused
Breath Volume		0.000	04:58	
Air Blank		0.000	04:58	
Control Test		0.081	04:58	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 05/13/2025 12:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: QUESADA

DATE/TIME OF TEST: 05/03/2025 01:46:45

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.081	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.090	01:41
Breath Volume		1.296	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:43
Subject Sample	# 2	0.093	01:45
Breath Volume		1.710	01:45
Air Blank		0.000	01:45
Control Test		0.081	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 05/13/2025 12:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MIRKOWSKI

DATE/TIME OF TEST: 05/03/2025 03:27:45

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.128	03:23
Breath Volume		1.851	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.126	03:26
Breath Volume		1.558	03:26
Air Blank		0.000	03:26
Control Test		0.081	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 05/13/2025 12:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: PUJOL

DATE/TIME OF TEST: 05/03/2025 21:15:42

Diagnostics Check		OK	21:09
Air Blank		0.000	21:09
Control Test		0.082	21:10
Air Blank		0.000	21:10
Subject Sample	# 1	0.192	21:11
Breath Volume		2.515	21:11
Air Blank		0.000	21:11
Air Blank		0.000	21:13
Subject Sample	# 2	0.200	21:14
Breath Volume		2.292	21:14
Air Blank		0.000	21:14
Control Test		0.082	21:15
Air Blank		0.000	21:15
Diagnostics Check		OK	21:15

Data Download Date/Time: 05/13/2025 12:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SOROKA,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 05/03/2025 23:24:49

Diagnostics Check		OK	23:21	
Air Blank		0.000	23:22	
Control Test		0.081	23:22	
Air Blank		0.000	23:23	
Subject Sample	# 1		23:23	Subject Test Refused
Breath Volume		0.000	23:23	
Air Blank		0.000	23:23	
Control Test		0.081	23:24	
Air Blank		0.000	23:24	
Diagnostics Check		OK	23:24	

Data Download Date/Time: 05/13/2025 12:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 05/03/2025 23:59:02

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.082	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.102	23:54
Breath Volume		2.066	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.103	23:57
Breath Volume		2.394	23:57
Air Blank		0.000	23:58
Control Test		0.081	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:59

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Neptune Beach PD
ARRESTING OFFICER: RICHARDSON,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MIDANI

DATE/TIME OF TEST: 05/04/2025 03:28:26

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.081	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.169	03:24
Breath Volume		1.863	03:24
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.165	03:26
Breath Volume		1.511	03:26
Air Blank		0.000	03:27
Control Test		0.081	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COREY,H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: DUNHAM

DATE/TIME OF TEST: 05/04/2025 04:12:14

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.082	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.161	04:07
Breath Volume		1.800	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.163	04:10
Breath Volume		1.804	04:10
Air Blank		0.000	04:11
Control Test		0.081	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/04/2025 18:15:47

Diagnostics Check		OK	18:09
Air Blank		0.000	18:09
Control Test		0.082	18:09
Air Blank		0.000	18:10
Subject Sample	# 1	0.424	18:11
Breath Volume		2.710	18:11
Air Blank		0.000	18:11
Air Blank		0.000	18:13
Subject Sample	# 2	0.443	18:14
Breath Volume		3.398	18:14
Air Blank		0.000	18:14
Control Test		0.081	18:15
Air Blank		0.000	18:15
Diagnostics Check		OK	18:15

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 05/04/2025 21:36:06

Diagnostics Check		OK	21:33	
Air Blank		0.000	21:33	
Control Test		0.082	21:34	
Air Blank		0.000	21:34	
Subject Sample	# 1		21:34	Subject Test Refused
Breath Volume		0.000	21:34	
Air Blank		0.000	21:35	
Control Test		0.081	21:35	
Air Blank		0.000	21:36	
Diagnostics Check		OK	21:36	

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JACKSON,O
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/05/2025 02:02:44

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.081	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.153	01:58
Breath Volume		1.484	01:58
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.163	02:01
Breath Volume		1.730	02:01
Air Blank		0.000	02:01
Control Test		0.081	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROSSI,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/05/2025 03:08:24

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:06	
Control Test		0.082	03:06	
Air Blank		0.000	03:06	
Subject Sample	# 1		03:07	Subject Test Refused
Breath Volume		0.000	03:07	
Air Blank		0.000	03:07	
Control Test		0.081	03:07	
Air Blank		0.000	03:08	
Diagnostics Check		OK	03:08	

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: DARLINGTON

DATE/TIME OF TEST: 05/05/2025 20:04:24

Diagnostics Check		OK	19:57
Air Blank		0.000	19:57
Control Test		0.081	19:58
Air Blank		0.000	19:58
Subject Sample	# 1	0.327	19:59
Breath Volume		1.546	19:59
Air Blank		0.000	20:00
Air Blank		0.000	20:02
Subject Sample	# 2	0.337	20:02
Breath Volume		1.593	20:02
Air Blank		0.000	20:03
Control Test		0.081	20:03
Air Blank		0.000	20:04
Diagnostics Check		OK	20:04

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: LESESNE

DATE/TIME OF TEST: 05/05/2025 23:09:14

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:06	
Control Test		0.081	23:07	
Air Blank		0.000	23:07	
Subject Sample	# 1		23:07	Subject Test Refused
Breath Volume		0.000	23:07	
Air Blank		0.000	23:08	
Control Test		0.081	23:08	
Air Blank		0.000	23:09	
Diagnostics Check		OK	23:09	

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: JULIANELLE

DATE/TIME OF TEST: 05/06/2025 01:18:08

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.000	01:13
Breath Volume		1.148	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.000	01:16
Breath Volume		1.105	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 05/06/2025 02:18:07

Diagnostics Check		OK	02:11
Air Blank		0.000	02:12
Control Test		0.081	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.117	02:13
Breath Volume		1.625	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.111	02:16
Breath Volume		1.476	02:16
Air Blank		0.000	02:17
Control Test		0.082	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 05/06/2025 03:30:08

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:27	
Control Test		0.081	03:28	
Air Blank		0.000	03:28	
Subject Sample	# 1		03:28	Subject Test Refused
Breath Volume		0.000	03:28	
Air Blank		0.000	03:29	
Control Test		0.081	03:29	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WREN

DATE/TIME OF TEST: 05/07/2025 02:02:22

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.080	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.172	01:57
Breath Volume		2.437	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.158	02:00
Breath Volume		1.601	02:00
Air Blank		0.000	02:01
Control Test		0.081	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/07/2025 23:41:35

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:38	
Control Test		0.081	23:38	
Air Blank		0.000	23:39	
Subject Sample	# 1		23:40	Interferent Detect
Breath Volume		1.734	23:40	
Air Blank		0.000	23:40	
Control Test		0.081	23:40	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/07/2025 23:49:12

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:46	
Control Test		0.081	23:46	
Air Blank		0.000	23:47	
Subject Sample	# 1		23:47	Interferent Detect
Breath Volume		1.707	23:47	
Air Blank		0.000	23:48	
Control Test		0.081	23:48	
Air Blank		0.000	23:49	
Diagnostics Check		OK	23:49	

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 05/08/2025 02:28:46

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:25	
Control Test		0.080	02:26	
Air Blank		0.000	02:26	
Subject Sample	# 1		02:27	Subject Test Refused
Breath Volume		0.000	02:27	
Air Blank		0.000	02:27	
Control Test		0.082	02:28	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 05/08/2025 23:03:54

Diagnostics Check		OK	23:01	
Air Blank		0.000	23:01	
Control Test		0.081	23:01	
Air Blank		0.000	23:02	
Subject Sample	# 1		23:02	Subject Test Refused
Breath Volume		0.000	23:02	
Air Blank		0.000	23:02	
Control Test		0.081	23:03	
Air Blank		0.000	23:03	
Diagnostics Check		OK	23:03	

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: THIES

DATE/TIME OF TEST: 05/09/2025 02:30:34

Diagnostics Check		OK	02:27	
Air Blank		0.000	02:28	
Control Test		0.082	02:28	
Air Blank		0.000	02:28	
Subject Sample	# 1		02:29	Subject Test Refused
Breath Volume		0.000	02:29	
Air Blank		0.000	02:29	
Control Test		0.082	02:29	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 05/13/2025 12:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/09/2025 05:00:15

Diagnostics Check		OK	04:57	
Air Blank		0.000	04:57	
Control Test		0.083	04:58	
Air Blank		0.000	04:58	
Subject Sample	# 1		04:58	Subject Test Refused
Breath Volume		0.000	04:58	
Air Blank		0.000	04:59	
Control Test		0.082	04:59	
Air Blank		0.000	05:00	
Diagnostics Check		OK	05:00	

Data Download Date/Time: 05/13/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,I
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 05/10/2025 02:34:41

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.082	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.056	02:29
Breath Volume		1.144	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.057	02:33
Breath Volume		3.257	02:33
Air Blank		0.000	02:33
Control Test		0.082	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 05/13/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRERA,R,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/10/2025 04:55:44

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:53	
Control Test		0.081	04:53	
Air Blank		0.000	04:54	
Subject Sample	# 1		04:54	Subject Test Refused
Breath Volume		0.000	04:54	
Air Blank		0.000	04:54	
Control Test		0.081	04:55	
Air Blank		0.000	04:55	
Diagnostics Check		OK	04:55	

Data Download Date/Time: 05/13/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 05/10/2025 23:47:24

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.082	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.041	23:42
Breath Volume		1.476	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.041	23:45
Breath Volume		1.335	23:45
Air Blank		0.000	23:46
Control Test		0.082	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 05/13/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAKER,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: EBRI

DATE/TIME OF TEST: 05/11/2025 06:34:55

Diagnostics Check		OK	06:27
Air Blank		0.000	06:27
Control Test		0.081	06:28
Air Blank		0.000	06:28
Subject Sample	# 1	0.085	06:30
Breath Volume		1.320	06:30
Air Blank		0.000	06:31
Air Blank		0.000	06:32
Subject Sample	# 2	0.087	06:33
Breath Volume		1.191	06:33
Air Blank		0.000	06:33
Control Test		0.081	06:34
Air Blank		0.000	06:34
Diagnostics Check		OK	06:34

Data Download Date/Time: 05/13/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KLING,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: MCINTYRE-ROGERS

DATE/TIME OF TEST: 05/13/2025 02:49:51

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:47	
Control Test		0.081	02:47	
Air Blank		0.000	02:48	
Subject Sample	# 1		02:48	Subject Test Refused
Breath Volume		0.000	02:48	
Air Blank		0.000	02:48	
Control Test		0.082	02:49	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 05/13/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRAY,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: GARCIA OSORIO

DATE/TIME OF TEST: 05/13/2025 07:41:30

Diagnostics Check		OK	07:35
Air Blank		0.000	07:35
Control Test		0.081	07:35
Air Blank		0.000	07:36
Subject Sample	# 1	0.108	07:37
Breath Volume		2.394	07:37
Air Blank		0.000	07:37
Air Blank		0.000	07:39
Subject Sample	# 2	0.106	07:39
Breath Volume		2.300	07:39
Air Blank		0.000	07:40
Control Test		0.082	07:40
Air Blank		0.000	07:41
Diagnostics Check		OK	07:41

Data Download Date/Time: 05/13/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,D,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: MCCREA

DATE/TIME OF TEST: 05/20/2025 01:15:46

Diagnostics Check		OK	01:12	
Air Blank		0.000	01:13	
Control Test		0.082	01:13	
Air Blank		0.000	01:14	
Subject Sample	# 1		01:14	Subject Test Refused
Breath Volume		0.000	01:14	
Air Blank		0.000	01:14	
Control Test		0.082	01:15	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 06/11/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: OJO

DATE/TIME OF TEST: 05/24/2025 03:48:36

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.080	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.241	03:44
Breath Volume		1.988	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.227	03:47
Breath Volume		1.628	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 06/11/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 05/25/2025 05:00:56

Diagnostics Check		OK	04:58	
Air Blank		0.000	04:58	
Control Test		0.082	04:58	
Air Blank		0.000	04:59	
Subject Sample	# 1		04:59	Subject Test Refused
Breath Volume		0.000	04:59	
Air Blank		0.000	05:00	
Control Test		0.081	05:00	
Air Blank		0.000	05:00	
Diagnostics Check		OK	05:00	

Data Download Date/Time: 06/11/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WEST,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: KIFLE

DATE/TIME OF TEST: 05/27/2025 07:57:14

Diagnostics Check		OK	07:50
Air Blank		0.000	07:51
Control Test		0.081	07:51
Air Blank		0.000	07:52
Subject Sample	# 1	0.000	07:52
Breath Volume		2.777	07:52
Air Blank		0.000	07:53
Air Blank		0.000	07:55
Subject Sample	# 2	0.000	07:55
Breath Volume		2.929	07:55
Air Blank		0.000	07:56
Control Test		0.082	07:56
Air Blank		0.000	07:57
Diagnostics Check		OK	07:57

Data Download Date/Time: 06/11/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONROE,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: CHERRY

DATE/TIME OF TEST: 05/27/2025 13:23:06

Diagnostics Check		OK	13:20	
Air Blank		0.000	13:20	
Control Test		0.081	13:21	
Air Blank		0.000	13:21	
Subject Sample	# 1		13:21	Subject Test Refused
Breath Volume		0.000	13:21	
Air Blank		0.000	13:22	
Control Test		0.081	13:22	
Air Blank		0.000	13:23	
Diagnostics Check		OK	13:23	

Data Download Date/Time: 06/11/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,D,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 05/30/2025 04:49:23

Diagnostics Check		OK	04:46	
Air Blank		0.000	04:46	
Control Test		0.081	04:47	
Air Blank		0.000	04:47	
Subject Sample	# 1		04:48	Subject Test Refused
Breath Volume		0.000	04:48	
Air Blank		0.000	04:48	
Control Test		0.081	04:48	
Air Blank		0.000	04:49	
Diagnostics Check		OK	04:49	

Data Download Date/Time: 06/11/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: MORRISSEY

DATE/TIME OF TEST: 05/31/2025 23:29:30

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.081	23:24
Air Blank		0.000	23:24
Subject Sample	# 1	0.030	23:25
Breath Volume		1.968	23:25
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.030	23:28
Breath Volume		2.058	23:28
Air Blank		0.000	23:28
Control Test		0.081	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 06/11/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCARDI,A,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: BROWN,MICHAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CESPEDES ORTIZ

DATE/TIME OF TEST: 05/01/2025 01:55:38

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.081	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.195	01:50
Breath Volume		2.269	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.200	01:54
Breath Volume		3.167	01:54
Air Blank		0.000	01:54
Control Test		0.079	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 05/30/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPELBAUM,KEITH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: DUARTE ESCOBAR

DATE/TIME OF TEST: 05/01/2025 22:25:52

Diagnostics Check		OK	22:19
Air Blank		0.000	22:20
Control Test		0.084	22:20
Air Blank		0.000	22:21
Subject Sample	# 1	0.193	22:21
Breath Volume		1.824	22:21
Air Blank		0.000	22:22
Air Blank		0.000	22:24
Subject Sample	# 2	0.199	22:24
Breath Volume		2.289	22:24
Air Blank		0.000	22:24
Control Test		0.082	22:25
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 05/30/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPELBAUM,KEITH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 05/02/2025 22:52:00

Diagnostics Check		OK	22:46
Air Blank		0.000	22:46
Control Test		0.082	22:46
Air Blank		0.000	22:47
Subject Sample	# 1	0.035	22:47
Breath Volume		1.507	22:47
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.040	22:50
Breath Volume		1.660	22:50
Air Blank		0.000	22:51
Control Test		0.082	22:51
Air Blank		0.000	22:51
Diagnostics Check		OK	22:52

Data Download Date/Time: 05/30/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JUSTIN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: YARISH,DYAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: HUNSINGER

DATE/TIME OF TEST: 05/04/2025 23:55:39

Diagnostics Check		OK	23:42	
Air Blank		0.000	23:43	
Control Test		0.081	23:43	
Air Blank		0.000	23:44	
Subject Sample	# 1	0.183	23:45	
Breath Volume		1.246	23:45	
Air Blank		0.000	23:46	
Air Blank		0.000	23:47	
Subject Sample	# 2	0.174	23:51	Volume Not Met
Breath Volume		0.117	23:51	
Air Blank		0.000	23:51	
Air Blank		0.000	23:53	
Subject Sample	# 3	0.177	23:54	
Breath Volume		1.148	23:54	
Air Blank		0.000	23:54	
Control Test		0.079	23:55	
Air Blank		0.000	23:55	
Diagnostics Check		OK	23:55	

Data Download Date/Time: 05/30/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JUSTIN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: DAIGNEAULT,CODY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: TREVOR

DATE/TIME OF TEST: 05/05/2025 23:53:46

Diagnostics Check		OK	23:47
Air Blank		0.000	23:48
Control Test		0.081	23:48
Air Blank		0.000	23:48
Subject Sample	# 1	0.152	23:49
Breath Volume		4.574	23:49
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.144	23:52
Breath Volume		2.515	23:52
Air Blank		0.000	23:52
Control Test		0.080	23:53
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 05/30/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETO,JOSHUA,W
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MAXIDE
BOUQUETTE,LUTENCIA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male

OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: VALLEJOS

DATE/TIME OF TEST: 05/07/2025 15:11:48

Diagnostics Check		OK	15:09	
Air Blank		0.000	15:09	
Control Test		0.081	15:09	
Air Blank		0.000	15:10	
Subject Sample	# 1		15:10	RFI Detect
Breath Volume		0.000	15:10	
Air Blank		0.000	15:11	
			15:11	RFI Detect
Air Blank		0.000	15:11	

Data Download Date/Time: 05/30/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETO,JOSHUA,W
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MAXIDE BOUQUETTE,LUTENCIA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: VALLEJOS

DATE/TIME OF TEST: 05/07/2025 15:15:20

Diagnostics Check	OK	15:14	
Air Blank		15:14	RFI Detect
Air Blank	0.000	15:15	

Data Download Date/Time: 05/30/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETO,JOSHUA,W
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MAXIDE
BOUQUETTE,LUTENCIA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male

OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: VALLEJOS

DATE/TIME OF TEST: 05/07/2025 15:20:40

Diagnostics Check	OK	15:19	
Air Blank	0.000	15:19	
Control Test	0.082	15:19	
Air Blank		15:20	RFI Detect
Air Blank	0.000	15:20	

Data Download Date/Time: 05/30/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETO,JOSHUA,W
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MAXIDE BOUQUETTE,LUTENCIA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male

OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: VALLEJOS

DATE/TIME OF TEST: 05/07/2025 15:38:10

Diagnostics Check		OK	15:31
Air Blank		0.000	15:31
Control Test		0.082	15:32
Air Blank		0.000	15:32
Subject Sample	# 1	0.338	15:33
Breath Volume		1.824	15:33
Air Blank		0.000	15:34
Air Blank		0.000	15:35
Subject Sample	# 2	0.341	15:36
Breath Volume		2.335	15:36
Air Blank		0.000	15:37
Control Test		0.081	15:37
Air Blank		0.000	15:38
Diagnostics Check		OK	15:38

Data Download Date/Time: 05/30/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPELBAUM,KEITH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: PIERRE-TOUSSAINT

DATE/TIME OF TEST: 05/07/2025 19:03:27

Diagnostics Check		OK	18:56
Air Blank		0.000	18:56
Control Test		0.082	18:57
Air Blank		0.000	18:57
Subject Sample	# 1	0.174	18:58
Breath Volume		2.242	18:58
Air Blank		0.000	18:59
Air Blank		0.000	19:00
Subject Sample	# 2	0.173	19:01
Breath Volume		2.542	19:01
Air Blank		0.000	19:02
Control Test		0.082	19:02
Air Blank		0.000	19:03
Diagnostics Check		OK	19:03

Data Download Date/Time: 05/30/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMCZYK,NIKKI,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TOMCZYK,STENLY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: APPLING

DATE/TIME OF TEST: 05/10/2025 01:25:05

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.082	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.116	01:20
Breath Volume		2.398	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.111	01:23
Breath Volume		2.175	01:23
Air Blank		0.000	01:24
Control Test		0.082	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:25

Data Download Date/Time: 05/30/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHIP,STEVEN,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: PITTS,CHRISTIAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: BARONE

DATE/TIME OF TEST: 05/14/2025 21:57:00

Diagnostics Check		OK	21:50
Air Blank		0.000	21:50
Control Test		0.082	21:51
Air Blank		0.000	21:51
Subject Sample	# 1	0.133	21:52
Breath Volume		1.996	21:52
Air Blank		0.000	21:53
Air Blank		0.000	21:54
Subject Sample	# 2	0.124	21:55
Breath Volume		1.687	21:55
Air Blank		0.000	21:56
Control Test		0.083	21:56
Air Blank		0.000	21:56
Diagnostics Check		OK	21:57

Data Download Date/Time: 05/30/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS,RODNESHIA,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: QUAGLIA,ERIC
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 05/22/2025 20:54:14

Diagnostics Check		OK	20:42	
Air Blank		0.000	20:43	
Control Test		0.082	20:43	
Air Blank		0.000	20:44	
Subject Sample	# 1	0.232	20:47	Volume Not Met
Breath Volume		0.792	20:47	
Air Blank		0.000	20:47	
Air Blank		0.000	20:49	
Subject Sample	# 2	0.227	20:52	Volume Not Met
Breath Volume		0.699	20:52	
Air Blank		0.000	20:53	
Control Test		0.080	20:53	
Air Blank		0.000	20:54	
Diagnostics Check		OK	20:54	

Data Download Date/Time: 05/30/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS,RODNESHIA,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: QUAGLIA,ERIC
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 05/22/2025 21:10:23

Diagnostics Check		OK	20:59	
Air Blank		0.000	20:59	
Control Test		0.078	20:59	
Air Blank		0.000	21:00	
Subject Sample	# 1	0.210	21:03	Volume Not Met
Breath Volume		0.566	21:03	
Air Blank		0.000	21:04	
Air Blank		0.000	21:05	
Subject Sample	# 2		21:09	No Sample Provided
Breath Volume		0.000	21:09	
Air Blank		0.000	21:09	
Control Test		0.080	21:09	
Air Blank		0.000	21:10	
Diagnostics Check		OK	21:10	

Data Download Date/Time: 05/30/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JUSTIN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TEMINO,RYAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 05/28/2025 02:49:00

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:37	
Control Test		0.081	02:38	
Air Blank		0.000	02:38	
Subject Sample	# 1	0.304	02:41	Volume Not Met
Breath Volume		0.316	02:41	
Air Blank		0.000	02:42	
Air Blank		0.000	02:44	
Subject Sample	# 2	0.254	02:47	Volume Not Met
Breath Volume		0.140	02:47	
Air Blank		0.000	02:48	
Control Test		0.078	02:48	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 05/30/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JUSTIN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TEMINO,RYAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 05/28/2025 03:04:16

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:51	
Control Test		0.078	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1	0.280	02:54	
Breath Volume		1.679	02:54	
Air Blank		0.000	02:54	
Air Blank		0.000	02:56	
Subject Sample	# 2	0.238	02:59	Volume Not Met
Breath Volume		0.425	02:59	
Air Blank		0.000	03:00	
Air Blank		0.000	03:02	
Subject Sample	# 3	0.279	03:02	
Breath Volume		1.914	03:02	
Air Blank		0.000	03:03	
Control Test		0.077	03:03	
Air Blank		0.000	03:04	
Diagnostics Check		OK	03:04	

Data Download Date/Time: 05/30/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: LOSH

DATE/TIME OF TEST: 05/29/2025 00:26:49

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.081	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.198	00:22
Breath Volume		2.511	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.201	00:25
Breath Volume		2.390	00:25
Air Blank		0.000	00:25
Control Test		0.080	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 05/30/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEGER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: FEGER,JOSHUA
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 05/06/2025 04:28:36

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:26	
Control Test		0.081	04:26	
Air Blank		0.000	04:26	
Subject Sample	# 1	0.159	04:27	
Breath Volume		1.410	04:27	
Air Blank			04:28	RFI Detect
Air Blank			04:28	RFI Detect
Air Blank			04:28	RFI Detect

Data Download Date/Time: 05/27/2025 08:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEGER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: FEGER,JOSHUA
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 05/06/2025 04:31:06

Diagnostics Check	OK	04:30	
Air Blank		04:30	RFI Detect
Air Blank		04:31	RFI Detect

Data Download Date/Time: 05/27/2025 08:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEGER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: FEGER,JOSHUA
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 05/06/2025 04:40:02

Diagnostics Check		OK	04:33
Air Blank		0.000	04:33
Control Test		0.081	04:34
Air Blank		0.000	04:34
Subject Sample	# 1	0.154	04:35
Breath Volume		1.382	04:35
Air Blank		0.000	04:36
Air Blank		0.000	04:37
Subject Sample	# 2	0.165	04:38
Breath Volume		1.367	04:38
Air Blank		0.000	04:39
Control Test		0.080	04:39
Air Blank		0.000	04:39
Diagnostics Check		OK	04:40

Data Download Date/Time: 05/27/2025 08:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERRICO,SIERRA,L
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: FLANAGAN,ROY
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: DAMBAUGH

DATE/TIME OF TEST: 05/07/2025 01:13:13

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:04	
Control Test		0.081	01:04	
Air Blank		0.000	01:05	
Subject Sample	# 1	0.062	01:06	
Breath Volume		1.574	01:06	
Air Blank		0.000	01:07	
Air Blank		0.000	01:09	
Subject Sample	# 2		01:12	No Sample Provided
Breath Volume		0.000	01:12	
Air Blank			01:12	RFI Detect
Air Blank			01:12	RFI Detect
Air Blank		0.000	01:13	

Data Download Date/Time: 05/27/2025 08:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTEZ,GABRIEL,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: CORTEZ,GABRIEL
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: LOURES

DATE/TIME OF TEST: 05/22/2025 15:27:39

Diagnostics Check		OK	15:20
Air Blank		0.000	15:20
Control Test		0.080	15:21
Air Blank		0.000	15:21
Subject Sample	# 1	0.115	15:23
Breath Volume		2.140	15:23
Air Blank		0.000	15:23
Air Blank		0.000	15:25
Subject Sample	# 2	0.121	15:26
Breath Volume		1.636	15:26
Air Blank		0.000	15:26
Control Test		0.080	15:27
Air Blank		0.000	15:27
Diagnostics Check		OK	15:27

Data Download Date/Time: 05/27/2025 08:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIERA,SCOTT,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: RIERA,SCOTT
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:51

SUBJECT'S LAST NAME: VARGO

DATE/TIME OF TEST: 05/24/2025 15:45:05

Diagnostics Check		OK	15:37	
Air Blank		0.000	15:38	
Control Test		0.080	15:38	
Air Blank		0.000	15:39	
Subject Sample	# 1	0.226	15:39	
Breath Volume		1.699	15:39	
Air Blank		0.000	15:40	
Air Blank		0.000	15:42	
Subject Sample	# 2	0.222	15:43	
Breath Volume		1.769	15:43	
Air Blank			15:44	Purge Fail
Air Blank		0.000	15:45	

Data Download Date/Time: 05/27/2025 08:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,EMMANUEL,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: TORRES,EMMANUEL
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: SANCHEZ IGLESIAS

DATE/TIME OF TEST: 05/07/2025 02:57:39

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.078	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.177	02:52
Breath Volume		1.804	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.188	02:55
Breath Volume		2.179	02:55
Air Blank		0.000	02:56
Control Test		0.077	02:57
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 05/08/2025 10:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEJEDA,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: GUIFARRO,JONATAN
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 05/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: FERREY

DATE/TIME OF TEST: 05/17/2025 22:39:31

Diagnostics Check		OK	22:36	
Air Blank		0.000	22:37	
Control Test		0.081	22:37	
Air Blank		0.000	22:37	
Subject Sample	# 1		22:38	Subject Test Refused
Breath Volume		0.000	22:38	
Air Blank		0.000	22:38	
Control Test		0.080	22:38	
Air Blank		0.000	22:39	
Diagnostics Check		OK	22:39	

Data Download Date/Time: 06/04/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIALLO MORELL,RAUL,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: HERNANDEZ,OSDANY
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 05/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: FERNANDEZ MARTINEZ

DATE/TIME OF TEST: 05/19/2025 00:05:50

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:01	
Control Test		0.081	00:01	
Air Blank		0.000	00:02	
Subject Sample	# 1		00:04	Subject Test Refused
Breath Volume		0.000	00:04	
Air Blank		0.000	00:04	
Control Test		0.080	00:05	
Air Blank		0.000	00:05	
Diagnostics Check		OK	00:05	

Data Download Date/Time: 06/04/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MAGAHA,J
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 05/23/2025 21:05:28

Diagnostics Check		OK	20:59
Air Blank		0.000	20:59
Control Test		0.079	21:00
Air Blank		0.000	21:00
Subject Sample	# 1	0.310	21:01
Breath Volume		1.375	21:01
Air Blank		0.000	21:01
Air Blank		0.000	21:03
Subject Sample	# 2	0.304	21:03
Breath Volume		1.136	21:03
Air Blank		0.000	21:04
Control Test		0.077	21:04
Air Blank		0.000	21:05
Diagnostics Check		OK	21:05

Data Download Date/Time: 06/19/2025 23:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,P
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: DAGUE

DATE/TIME OF TEST: 05/24/2025 04:41:55

Diagnostics Check		OK	04:30	
Air Blank		0.000	04:30	
Control Test		0.080	04:31	
Air Blank		0.000	04:31	
Subject Sample	# 1	0.000	04:34	Volume Not Met
Breath Volume		0.000	04:34	
Air Blank		0.000	04:35	
Air Blank		0.000	04:37	
Subject Sample	# 2	0.000	04:40	Volume Not Met
Breath Volume		0.019	04:40	
Air Blank		0.000	04:41	
Control Test		0.077	04:41	
Air Blank		0.000	04:41	
Diagnostics Check		OK	04:41	

Data Download Date/Time: 06/19/2025 23:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OWENS,J
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MAIDA

DATE/TIME OF TEST: 05/25/2025 01:54:12

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.079	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.203	01:49
Breath Volume		2.171	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.196	01:52
Breath Volume		1.972	01:52
Air Blank		0.000	01:53
Control Test		0.077	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 06/19/2025 23:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 05/29/2025 00:20:29

Diagnostics Check		OK	00:08	
Air Blank		0.000	00:09	
Control Test		0.080	00:09	
Air Blank		0.000	00:09	
Subject Sample	# 1	0.209	00:10	
Breath Volume		1.128	00:10	
Air Blank		0.000	00:11	
Air Blank		0.000	00:12	
Subject Sample	# 2	0.189	00:15	Volume Not Met
Breath Volume		0.128	00:15	
Air Blank		0.000	00:17	
Air Blank		0.000	00:18	
Subject Sample	# 3	0.202	00:18	
Breath Volume		2.417	00:18	
Air Blank		0.000	00:19	
Control Test		0.074	00:19	Control Outside Tolerance
Air Blank		0.000	00:20	

Data Download Date/Time: 06/19/2025 23:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 05/29/2025 00:30:16

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.076	00:24
Air Blank		0.000	00:25
Subject Sample	# 1	0.206	00:25
Breath Volume		1.835	00:25
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.211	00:28
Breath Volume		1.863	00:28
Air Blank		0.000	00:29
Control Test		0.080	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 06/19/2025 23:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/31/2025 03:05:04

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.081	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.165	03:00
Breath Volume		1.496	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.162	03:03
Breath Volume		1.382	03:03
Air Blank		0.000	03:04
Control Test		0.077	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:05

Data Download Date/Time: 06/19/2025 23:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 05/31/2025 21:27:28

Diagnostics Check		OK	21:21	
Air Blank		0.000	21:21	
Control Test		0.081	21:22	
Air Blank		0.000	21:22	
Subject Sample	# 1	0.097	21:23	
Breath Volume		1.601	21:23	
Air Blank		0.000	21:23	
Air Blank		0.000	21:25	
Subject Sample	# 2	0.099	21:25	
Breath Volume		1.457	21:25	
Air Blank		0.000	21:26	
Control Test		0.077	21:26	
Air Blank			21:27	RFI Detect
Air Blank		0.000	21:27	

Data Download Date/Time: 06/19/2025 23:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 05/31/2025 21:37:22

Diagnostics Check		OK	21:31
Air Blank		0.000	21:31
Control Test		0.081	21:32
Air Blank		0.000	21:32
Subject Sample	# 1	0.095	21:33
Breath Volume		1.578	21:33
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.089	21:35
Breath Volume		1.562	21:35
Air Blank		0.000	21:36
Control Test		0.078	21:36
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 06/19/2025 23:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARROQUIN,RICHARD,J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: MARROQUIN,RICHARD
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 05/20/2025 20:49:53

Diagnostics Check		OK	20:43
Air Blank		0.000	20:44
Control Test		0.079	20:44
Air Blank		0.000	20:44
Subject Sample	# 1	0.136	20:45
Breath Volume		4.957	20:45
Air Blank		0.000	20:46
Air Blank		0.000	20:47
Subject Sample	# 2	0.133	20:48
Breath Volume		4.968	20:48
Air Blank		0.000	20:48
Control Test		0.078	20:49
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 05/31/2025 15:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARROQUIN,RICHARD,J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: MCGAREY,TIMOTHY
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: BUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: MURILLO

DATE/TIME OF TEST: 05/24/2025 19:30:54

Diagnostics Check		OK	19:23
Air Blank		0.000	19:24
Control Test		0.079	19:24
Air Blank		0.000	19:24
Subject Sample	# 1	0.159	19:25
Breath Volume		1.261	19:25
Air Blank		0.000	19:25
Air Blank		0.000	19:27
Subject Sample	# 2	0.143	19:29
Breath Volume		1.152	19:29
Air Blank		0.000	19:30
Control Test		0.079	19:30
Air Blank		0.000	19:30
Diagnostics Check		OK	19:30

Data Download Date/Time: 05/31/2025 15:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HANSARD,JOSHUA
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402635303
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: BUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 05/17/2025 18:59:52

Diagnostics Check		OK	18:53
Air Blank		0.000	18:54
Control Test		0.077	18:54
Air Blank		0.000	18:54
Subject Sample	# 1	0.150	18:55
Breath Volume		3.027	18:55
Air Blank		0.000	18:56
Air Blank		0.000	18:57
Subject Sample	# 2	0.136	18:58
Breath Volume		1.738	18:58
Air Blank		0.000	18:58
Control Test		0.077	18:59
Air Blank		0.000	18:59
Diagnostics Check		OK	18:59

Data Download Date/Time: 05/29/2025 15:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCI,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: TUCCI,JOSE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: OTHER
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 05/15/2025 13:30:39

Diagnostics Check		OK	13:24	
Air Blank		0.000	13:25	
Control Test		0.079	13:25	
Air Blank		0.000	13:26	
Subject Sample	# 1	0.000	13:26	
Breath Volume		2.566	13:26	
Air Blank		0.000	13:27	
Air Blank		0.000	13:29	
			13:29	RFI Detect
Air Blank		0.000	13:29	
Control Test		0.078	13:29	
Air Blank			13:30	RFI Detect
Air Blank		0.000	13:30	

Data Download Date/Time: 05/22/2025 11:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 05/20/2025 18:54:01

Diagnostics Check		OK	18:43
Air Blank		0.000	18:44
Control Test		0.080	18:44
Air Blank		0.000	18:45
Subject Sample	# 1	0.107	18:46
Breath Volume		1.390	18:46
Air Blank		0.000	18:47
Air Blank		0.000	18:49
Subject Sample	# 2	0.101	18:52
Breath Volume		1.199	18:52
Air Blank		0.000	18:53
Control Test		0.079	18:53
Air Blank		0.000	18:53
Diagnostics Check		OK	18:54

Data Download Date/Time: 05/31/2025 19:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORD,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TRENT,ANDREW
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 04/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BLANTON

DATE/TIME OF TEST: 05/23/2025 04:12:38

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.197	04:07
Breath Volume		1.125	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.184	04:11
Breath Volume		1.152	04:11
Air Blank		0.000	04:11
Control Test		0.080	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 05/31/2025 19:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITTINGTON,CAROLINE,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: DANFORTH

DATE/TIME OF TEST: 05/26/2025 02:10:18

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:07	
Control Test		0.080	02:08	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Control Test		0.080	02:09	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 05/31/2025 19:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ROBERT,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: C636770921101

DATE/TIME OF TEST: 05/26/2025 20:37:04

Diagnostics Check		OK	20:34	
Air Blank		0.000	20:34	
Control Test		0.080	20:34	
Air Blank		0.000	20:35	
Subject Sample	# 1		20:35	RFI Detect
Breath Volume		0.000	20:35	
Air Blank		0.000	20:36	
Control Test		0.080	20:36	
Air Blank		0.000	20:36	
Diagnostics Check		OK	20:37	

Data Download Date/Time: 05/31/2025 19:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ROBERT,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/26/2025 20:43:46

Diagnostics Check		OK	20:39	
Air Blank		0.000	20:40	
Control Test		0.080	20:40	
Air Blank		0.000	20:41	
Subject Sample	# 1		20:43	RFI Detect
Breath Volume		0.066	20:43	
Air Blank			20:43	RFI Detect
Air Blank			20:43	RFI Detect

Data Download Date/Time: 05/31/2025 19:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ROBERT,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/26/2025 20:58:03

Diagnostics Check		OK	20:55	
Air Blank		0.000	20:55	
Control Test		0.078	20:56	
Air Blank		0.000	20:56	
Subject Sample	# 1		20:56	Subject Test Refused
Breath Volume		0.000	20:56	
Air Blank		0.000	20:57	
Control Test		0.079	20:57	
Air Blank		0.000	20:57	
Diagnostics Check		OK	20:58	

Data Download Date/Time: 05/31/2025 19:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TOTAJ,DARDAN
ARRESTING OFFICER AGENCY: Chiefland PD
LAST AGENCY INSPECTION: 04/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: ROBIE JR

DATE/TIME OF TEST: 05/30/2025 02:40:21

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:37	
Control Test		0.080	02:38	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:38	Subject Test Refused
Breath Volume		0.000	02:38	
Air Blank		0.000	02:39	
Control Test		0.081	02:39	
Air Blank		0.000	02:40	
Diagnostics Check		OK	02:40	

Data Download Date/Time: 05/31/2025 19:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 05/09/2025 03:19:12

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.173	03:14
Breath Volume		1.675	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.166	03:17
Breath Volume		2.265	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 05/12/2025 05:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 05/10/2025 02:38:25

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.079	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.145	02:34
Breath Volume		2.246	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.150	02:36
Breath Volume		2.101	02:36
Air Blank		0.000	02:37
Control Test		0.079	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 05/12/2025 05:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: THOMAS,BRANDEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: BASSAURE

DATE/TIME OF TEST: 05/10/2025 04:35:01

Diagnostics Check	OK	04:34	
Air Blank		04:34	RFI Detect
Air Blank	0.000	04:35	

Data Download Date/Time: 05/12/2025 05:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: THOMAS,BRANDEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: BASSAURE

DATE/TIME OF TEST: 05/10/2025 04:41:03

Diagnostics Check		OK	04:37	
Air Blank		0.000	04:37	
Control Test		0.080	04:37	
Air Blank		0.000	04:38	
Subject Sample	# 1		04:40	Subject Test Refused
Breath Volume		0.000	04:40	
Air Blank			04:40	RFI Detect
Air Blank		0.000	04:41	

Data Download Date/Time: 05/12/2025 05:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: THOMAS,BRANDEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 05/15/2025 04:16:52

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.149	04:12
Breath Volume		2.210	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:14
Subject Sample	# 2	0.152	04:15
Breath Volume		1.203	04:15
Air Blank		0.000	04:15
Control Test		0.080	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 06/11/2025 21:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: CRIBBS

DATE/TIME OF TEST: 05/16/2025 03:33:16

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.202	03:28
Breath Volume		1.398	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.194	03:31
Breath Volume		2.867	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 06/11/2025 21:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: FERRARO

DATE/TIME OF TEST: 05/24/2025 02:36:44

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.078	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.087	02:32
Breath Volume		1.394	02:32
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.091	02:35
Breath Volume		1.281	02:35
Air Blank		0.000	02:35
Control Test		0.079	02:36
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 06/11/2025 21:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: LIPCIUS

DATE/TIME OF TEST: 05/25/2025 01:37:17

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.192	01:32
Breath Volume		3.742	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:34
Subject Sample	# 2	0.196	01:35
Breath Volume		3.757	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 06/11/2025 21:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: THOMAS,BRANDEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MIKE

DATE/TIME OF TEST: 05/26/2025 00:45:30

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.079	00:36	
Air Blank		0.000	00:36	
Subject Sample	# 1	0.207	00:39	Volume Not Met
Breath Volume		0.218	00:39	
Air Blank		0.000	00:40	
Air Blank		0.000	00:42	
Subject Sample	# 2		00:43	Subject Test Refused
Breath Volume		0.000	00:43	
Air Blank		0.000	00:44	
Control Test		0.077	00:44	
Air Blank		0.000	00:44	
Diagnostics Check		Fail	00:45	Temperature Regulation Test
Air Blank		0.000	00:45	

Data Download Date/Time: 06/11/2025 21:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MELENDEZ ENCARNACION

DATE/TIME OF TEST: 05/29/2025 03:32:14

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.078	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.000	03:27
Breath Volume		1.855	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.000	03:30
Breath Volume		1.566	03:30
Air Blank		0.000	03:31
Control Test		0.079	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 06/11/2025 21:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVER,ISAIAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SILVER,ISAIAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: DEROMA

DATE/TIME OF TEST: 05/05/2025 17:07:16

Diagnostics Check		OK	17:03	
Air Blank		0.000	17:04	
Control Test		0.079	17:04	
Air Blank		0.000	17:05	
Subject Sample	# 1	0.206	17:05	
Breath Volume		3.046	17:05	
Air Blank			17:06	Purge Fail
Air Blank		0.000	17:07	

Data Download Date/Time: 05/28/2025 11:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVER,ISAIAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SILVER,ISAIAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: DEROMA

DATE/TIME OF TEST: 05/05/2025 17:19:31

Diagnostics Check		OK	17:12
Air Blank		0.000	17:13
Control Test		0.080	17:13
Air Blank		0.000	17:14
Subject Sample	# 1	0.199	17:15
Breath Volume		2.414	17:15
Air Blank		0.000	17:15
Air Blank		0.000	17:17
Subject Sample	# 2	0.190	17:18
Breath Volume		1.785	17:18
Air Blank		0.000	17:18
Control Test		0.080	17:18
Air Blank		0.000	17:19
Diagnostics Check		OK	17:19

Data Download Date/Time: 05/28/2025 11:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVER,ISAIAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MITCHELL,SHONQUEZ
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: KRUCZALAK

DATE/TIME OF TEST: 05/14/2025 20:51:50

Diagnostics Check		OK	20:45	
Air Blank		0.000	20:46	
Control Test		0.078	20:46	
Air Blank		0.000	20:47	
Subject Sample	# 1	0.107	20:47	
Breath Volume		3.550	20:47	
Air Blank		0.000	20:48	
Air Blank		0.000	20:50	
Subject Sample	# 2	0.106	20:50	
Breath Volume		3.015	20:50	
Air Blank			20:51	RFI Detect
Air Blank		0.000	20:51	

Data Download Date/Time: 05/28/2025 11:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GONZALEZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SALINAS

DATE/TIME OF TEST: 05/11/2025 03:50:41

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.079	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.091	03:46
Breath Volume		2.281	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.093	03:49
Breath Volume		2.031	03:49
Air Blank		0.000	03:49
Control Test		0.079	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 05/27/2025 03:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOPEZ,FRANCISCO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MARTINEZ MORALES

DATE/TIME OF TEST: 05/01/2025 01:17:25

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.079	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.112	01:12
Breath Volume		3.144	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.113	01:15
Breath Volume		2.863	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 05/14/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AULTMAN,SHAWN,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HERRADA,JORGE
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: TORRES-YURET

DATE/TIME OF TEST: 05/03/2025 04:33:25

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.161	04:28
Breath Volume		1.464	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.169	04:31
Breath Volume		2.113	04:31
Air Blank		0.000	04:32
Control Test		0.076	04:32
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 05/14/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 05/04/2025 04:07:25

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.075	04:02
Breath Volume		1.843	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.073	04:05
Breath Volume		2.890	04:05
Air Blank		0.000	04:06
Control Test		0.078	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 05/14/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AULTMAN,SHAWN,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: AULTMAN,SHAWN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: BERGET

DATE/TIME OF TEST: 05/04/2025 19:47:03

Diagnostics Check		OK	19:41	
Air Blank		0.000	19:41	
Control Test		0.079	19:42	
Air Blank		0.000	19:42	
Subject Sample	# 1	0.277	19:45	Volume Not Met
Breath Volume		0.542	19:45	
Air Blank			19:46	Purge Fail
Air Blank		0.000	19:47	

Data Download Date/Time: 05/14/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AULTMAN,SHAWN,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: AULTMAN,SHAWN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: BERGET

DATE/TIME OF TEST: 05/04/2025 20:03:27

Diagnostics Check		OK	19:49	
Air Blank		0.000	19:50	
Control Test		0.076	19:50	
Air Blank		0.000	19:50	
Subject Sample	# 1	0.302	19:51	
Breath Volume		2.007	19:51	
Air Blank		0.000	19:52	
Air Blank		0.000	19:54	
Subject Sample	# 2	0.270	19:57	Volume Not Met
Breath Volume		0.324	19:57	
Air Blank		0.000	19:58	
Air Blank		0.000	20:00	
Subject Sample	# 3	0.289	20:01	
Breath Volume		2.273	20:01	
Air Blank		0.000	20:02	
Control Test		0.075	20:02	
Air Blank		0.000	20:03	
Diagnostics Check		OK	20:03	

Data Download Date/Time: 05/14/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONTANEZ,YOMARI,I
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HARRIS,CASEY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: REDICAN

DATE/TIME OF TEST: 05/04/2025 22:03:12

Diagnostics Check		OK	21:51	
Air Blank		0.000	21:52	
Control Test		0.078	21:52	
Air Blank		0.000	21:52	
Subject Sample	# 1	0.105	21:56	Volume Not Met
Breath Volume		0.386	21:56	
Air Blank		0.000	21:56	
Air Blank		0.000	21:58	
Subject Sample	# 2	0.098	22:01	Volume Not Met
Breath Volume		0.406	22:01	
Air Blank		0.000	22:02	
Control Test		0.077	22:02	
Air Blank		0.000	22:03	
Diagnostics Check		OK	22:03	

Data Download Date/Time: 05/14/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,DERICK,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTEGA,MANUEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 05/06/2025 03:55:45

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.165	03:50
Breath Volume		1.164	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.165	03:54
Breath Volume		1.183	03:54
Air Blank		0.000	03:54
Control Test		0.080	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 05/14/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLEGAS,BRIAN,A
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: LOPEZ FIGUEROA,JOMAR
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: RIVERA VAZQUEZ

DATE/TIME OF TEST: 05/09/2025 04:19:41

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:12	
Control Test		0.079	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1		04:16	No Sample Provided
Breath Volume		0.000	04:16	
Air Blank		0.000	04:17	
Air Blank			04:19	Ambient Fail
Air Blank		0.000	04:19	

Data Download Date/Time: 05/14/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLEGAS,BRIAN,A
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: LOPEZ FIGUEROA,JOMAR
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: RIVERA VAZQUEZ

DATE/TIME OF TEST: 05/09/2025 04:34:51

Diagnostics Check		OK	04:23	
Air Blank		0.000	04:23	
Control Test		0.079	04:23	
Air Blank		0.000	04:24	
Subject Sample	# 1		04:27	No Sample Provided
Breath Volume		0.000	04:27	
Air Blank		0.000	04:28	
Air Blank		0.000	04:30	
Subject Sample	# 2	0.194	04:33	Volume Not Met
Breath Volume		0.101	04:33	
Air Blank		0.000	04:33	
Control Test		0.078	04:34	
Air Blank		0.000	04:34	
Diagnostics Check		OK	04:34	

Data Download Date/Time: 05/14/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELLER,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: DIAZ,GENESIS
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: CARCONE

DATE/TIME OF TEST: 05/09/2025 23:15:15

Diagnostics Check		OK	23:08
Air Blank		0.000	23:09
Control Test		0.080	23:09
Air Blank		0.000	23:10
Subject Sample	# 1	0.200	23:10
Breath Volume		2.660	23:10
Air Blank		0.000	23:11
Air Blank		0.000	23:13
Subject Sample	# 2	0.196	23:13
Breath Volume		2.722	23:13
Air Blank		0.000	23:14
Control Test		0.081	23:14
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 05/14/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AL-ANBKI,HAEDER,N
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMACHO,FRANCES
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: BOLANOS

DATE/TIME OF TEST: 05/10/2025 22:37:10

Diagnostics Check		OK	22:30
Air Blank		0.000	22:31
Control Test		0.079	22:31
Air Blank		0.000	22:31
Subject Sample	# 1	0.183	22:32
Breath Volume		3.148	22:32
Air Blank		0.000	22:33
Air Blank		0.000	22:35
Subject Sample	# 2	0.171	22:35
Breath Volume		2.285	22:35
Air Blank		0.000	22:36
Control Test		0.078	22:36
Air Blank		0.000	22:37
Diagnostics Check		OK	22:37

Data Download Date/Time: 05/14/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,JACOB,S
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCDONALD,KIRK
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: VILORIO

DATE/TIME OF TEST: 05/11/2025 04:06:29

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.146	04:01
Breath Volume		1.863	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.146	04:04
Breath Volume		1.863	04:04
Air Blank		0.000	04:05
Control Test		0.079	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 05/14/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCANLAN,JAMES,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SCANLAN,JAMES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: COWAN

DATE/TIME OF TEST: 05/13/2025 02:08:23

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.046	02:03
Breath Volume		2.488	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.040	02:06
Breath Volume		1.531	02:06
Air Blank		0.000	02:07
Control Test		0.078	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 05/14/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,MORGAN,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CASON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 05/10/2025 00:58:37

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:53	
Control Test		0.080	00:54	
Air Blank		0.000	00:54	
Subject Sample	# 1		00:57	Subject Test Refused
Breath Volume		0.000	00:57	
Air Blank		0.000	00:57	
Control Test		0.080	00:58	
Air Blank		0.000	00:58	
Diagnostics Check		OK	00:58	

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORDOQUI,DANIEL,G
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MOONEY,MORGAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: URBINA MORALES

DATE/TIME OF TEST: 05/10/2025 03:24:29

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.147	03:19
Breath Volume		2.546	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.146	03:22
Breath Volume		1.339	03:22
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAPEE,WYATT,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LOPEZ COVELLA,TRACY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: ZETINO RETANA

DATE/TIME OF TEST: 05/12/2025 05:49:26

Diagnostics Check		OK	05:43
Air Blank		0.000	05:43
Control Test		0.080	05:44
Air Blank		0.000	05:44
Subject Sample	# 1	0.160	05:45
Breath Volume		2.289	05:45
Air Blank		0.000	05:45
Air Blank		0.000	05:47
Subject Sample	# 2	0.165	05:47
Breath Volume		3.218	05:47
Air Blank		0.000	05:48
Control Test		0.080	05:48
Air Blank		0.000	05:49
Diagnostics Check		OK	05:49

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEHRENS,JACOB,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BEHRENS,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: EAGERTON

DATE/TIME OF TEST: 05/13/2025 03:36:53

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.102	03:32
Breath Volume		2.312	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.106	03:35
Breath Volume		2.175	03:35
Air Blank		0.000	03:35
Control Test		0.080	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,MORGAN,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MOONEY,MORGAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 05/16/2025 22:23:46

Diagnostics Check		OK	22:20	
Air Blank		0.000	22:20	
Control Test		0.079	22:20	
Air Blank		0.000	22:21	
Subject Sample	# 1		22:22	Subject Test Refused
Breath Volume		0.000	22:22	
Air Blank		0.000	22:22	
Control Test		0.079	22:23	
Air Blank		0.000	22:23	
Diagnostics Check		OK	22:23	

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENNIS,TIMOTHY,N
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BIERNACKI,CHAD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: SNELLGROVE

DATE/TIME OF TEST: 05/16/2025 23:37:29

Diagnostics Check		OK	23:24	
Air Blank		0.000	23:25	
Control Test		0.079	23:25	
Air Blank		0.000	23:26	
Subject Sample	# 1	0.212	23:27	
Breath Volume		2.195	23:27	
Air Blank		0.000	23:28	
Air Blank		0.000	23:29	
Subject Sample	# 2	0.182	23:33	Volume Not Met
Breath Volume		0.386	23:33	
Air Blank		0.000	23:33	
Air Blank		0.000	23:35	
Subject Sample	# 3	0.185	23:35	
Breath Volume		2.855	23:35	
Air Blank		0.000	23:36	
Control Test		0.073	23:36	Control Outside Tolerance
Air Blank		0.000	23:37	

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENNIS,TIMOTHY,N
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BIERNACKI,CHAD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: SNELLGROVE

DATE/TIME OF TEST: 05/16/2025 23:43:41

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:40	
Control Test		0.078	23:41	
Air Blank		0.000	23:41	
Subject Sample	# 1	0.183	23:42	
Breath Volume		2.820	23:42	
Air Blank			23:43	Purge Fail
Air Blank		0.000	23:43	

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENNIS,TIMOTHY,N
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BIERNACKI,CHAD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: SNELLGROVE

DATE/TIME OF TEST: 05/16/2025 23:52:58

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.083	23:46
Air Blank		0.000	23:47
Subject Sample	# 1	0.186	23:47
Breath Volume		2.324	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:49
Subject Sample	# 2	0.190	23:51
Breath Volume		2.253	23:51
Air Blank		0.000	23:52
Control Test		0.079	23:52
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,RITCHIE,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: RIVERA,RITCHIE
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: TROCHEZ PALMA

DATE/TIME OF TEST: 05/17/2025 03:48:42

Diagnostics Check		OK	03:41	
Air Blank		0.000	03:42	
Control Test		0.080	03:42	
Air Blank		0.000	03:43	
Subject Sample	# 1	0.000	03:44	
Breath Volume		3.500	03:44	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 2		03:47	Subject Test Refused
Breath Volume		0.000	03:47	
Air Blank		0.000	03:47	
Control Test		0.080	03:48	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,RITCHIE,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: RIVERA,RITCHIE
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: TROCHEZ PALMA

DATE/TIME OF TEST: 05/17/2025 04:01:52

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:50	
Control Test		0.080	03:51	
Air Blank		0.000	03:51	
Subject Sample	# 1		03:54	No Sample Provided
Breath Volume		0.000	03:54	
Air Blank		0.000	03:55	
Air Blank		0.000	03:57	
Subject Sample	# 2		04:00	No Sample Provided
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Control Test		0.080	04:01	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORDOQUI,DANIEL,G
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ORDOQUI,DANIEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: LOCKE

DATE/TIME OF TEST: 05/18/2025 02:39:54

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:31	
Control Test		0.080	02:32	
Air Blank		0.000	02:32	
Subject Sample	# 1		02:35	No Sample Provided
Breath Volume		0.000	02:35	
Air Blank		0.000	02:36	
Air Blank		0.000	02:38	
Subject Sample	# 2		02:38	Subject Test Refused
Breath Volume		0.000	02:38	
Air Blank		0.000	02:38	
Control Test		0.080	02:39	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORDOQUI,DANIEL,G
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WAGNER,MICHELLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: RAMOS RAMOS

DATE/TIME OF TEST: 05/20/2025 06:04:52

Diagnostics Check		OK	05:57	
Air Blank		0.000	05:57	
Control Test		0.080	05:57	
Air Blank		0.000	05:58	
Subject Sample	# 1	0.000	06:00	
Breath Volume		1.960	06:00	
Air Blank		0.000	06:01	
Air Blank		0.000	06:03	
Subject Sample	# 2		06:03	Subject Test Refused
Breath Volume		0.000	06:03	
Air Blank		0.000	06:03	
Control Test		0.081	06:04	
Air Blank		0.000	06:04	
Diagnostics Check		OK	06:04	

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORDOQUI,DANIEL,G
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WAGNER,MICHELLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: RAMOS RAMOS

DATE/TIME OF TEST: 05/20/2025 06:14:24

Diagnostics Check		OK	06:08
Air Blank		0.000	06:08
Control Test		0.080	06:09
Air Blank		0.000	06:09
Subject Sample	# 1	0.164	06:09
Breath Volume		1.816	06:09
Air Blank		0.000	06:10
Air Blank		0.000	06:12
Subject Sample	# 2	0.148	06:12
Breath Volume		1.921	06:12
Air Blank		0.000	06:13
Control Test		0.080	06:13
Air Blank		0.000	06:14
Diagnostics Check		OK	06:14

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOTT,JASMIN,H
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: VAZQUEZCRUZ,JORDAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/21/2025 09:16:41

Diagnostics Check		OK	09:10
Air Blank		0.000	09:10
Control Test		0.079	09:11
Air Blank		0.000	09:11
Subject Sample	# 1	0.291	09:12
Breath Volume		2.949	09:12
Air Blank		0.000	09:12
Air Blank		0.000	09:14
Subject Sample	# 2	0.293	09:15
Breath Volume		4.226	09:15
Air Blank		0.000	09:15
Control Test		0.079	09:16
Air Blank		0.000	09:16
Diagnostics Check		OK	09:16

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,RITCHIE,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ROSENBURG,DUSTIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: GILLS

DATE/TIME OF TEST: 05/22/2025 00:22:51

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:11	
Control Test		0.079	00:12	
Air Blank		0.000	00:12	
Subject Sample	# 1		00:15	No Sample Provided
Breath Volume		0.000	00:15	
Air Blank		0.000	00:16	
Air Blank		0.000	00:18	
Subject Sample	# 2		00:21	No Sample Provided
Breath Volume		0.000	00:21	
Air Blank		0.000	00:21	
Control Test		0.080	00:22	
Air Blank		0.000	00:22	
Diagnostics Check		OK	00:22	

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRMANS,ABNER,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: PEYTON,KYLEE
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 05/24/2025 18:37:56

Diagnostics Check		OK	18:34	
Air Blank		0.000	18:34	
Control Test		0.080	18:34	
Air Blank		0.000	18:35	
Subject Sample	# 1		18:36	RFI Detect
Breath Volume		0.000	18:36	
Air Blank		0.000	18:37	
Control Test		0.080	18:37	
Air Blank		0.000	18:37	
Diagnostics Check		OK	18:37	

Data Download Date/Time: 05/31/2025 22:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRMANS,ABNER,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: PEYTON,KYLEE
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 05/24/2025 18:47:31

Diagnostics Check		OK	18:42	
Air Blank		0.000	18:42	
Control Test		0.080	18:43	
Air Blank		0.000	18:43	
Subject Sample	# 1		18:46	Subject Test Refused
Breath Volume		0.000	18:46	
Air Blank		0.000	18:46	
Control Test		0.080	18:46	
Air Blank		0.000	18:47	
Diagnostics Check		OK	18:47	

Data Download Date/Time: 05/31/2025 22:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,SETH,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WILLIAMS,SETH
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: STAPLETON

DATE/TIME OF TEST: 05/25/2025 05:28:57

Diagnostics Check		OK	05:21
Air Blank		0.000	05:22
Control Test		0.081	05:22
Air Blank		0.000	05:23
Subject Sample	# 1	0.205	05:24
Breath Volume		1.375	05:24
Air Blank		0.000	05:24
Air Blank		0.000	05:26
Subject Sample	# 2	0.213	05:27
Breath Volume		2.527	05:27
Air Blank		0.000	05:28
Control Test		0.080	05:28
Air Blank		0.000	05:28
Diagnostics Check		OK	05:28

Data Download Date/Time: 05/31/2025 22:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCRUGGS,JADEN,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SCRUGGS,JADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: CONE

DATE/TIME OF TEST: 05/26/2025 00:04:08

Diagnostics Check		OK	00:01	
Air Blank		0.000	00:01	
Control Test		0.080	00:02	
Air Blank		0.000	00:02	
Subject Sample	# 1		00:02	Subject Test Refused
Breath Volume		0.000	00:02	
Air Blank		0.000	00:03	
Control Test		0.080	00:03	
Air Blank		0.000	00:04	
Diagnostics Check		OK	00:04	

Data Download Date/Time: 05/31/2025 22:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHAGAN,ANDREW,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GEOGHAGAN,ANDREW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: TSOI

DATE/TIME OF TEST: 05/27/2025 02:21:56

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:08	
Control Test		0.080	02:08	
Air Blank		0.000	02:09	
Subject Sample	# 1	0.129	02:12	Volume Not Met
Breath Volume		0.242	02:12	
Air Blank		0.000	02:12	
Air Blank		0.000	02:14	
Subject Sample	# 2	0.158	02:15	
Breath Volume		2.207	02:15	
Air Blank		0.000	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 3	0.144	02:20	
Breath Volume		2.121	02:20	
Air Blank		0.000	02:21	
Control Test		0.078	02:21	RFI Detect
Air Blank			02:21	RFI Detect

Data Download Date/Time: 05/31/2025 22:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHAGAN,ANDREW,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GEOGHAGAN,ANDREW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: TSOI

DATE/TIME OF TEST: 05/27/2025 02:32:49

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.078	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.147	02:28
Breath Volume		1.226	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.131	02:31
Breath Volume		1.718	02:31
Air Blank		0.000	02:31
Control Test		0.080	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 05/31/2025 22:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,RITCHIE,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: RIVERA,RITCHIE
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: WORKMAN

DATE/TIME OF TEST: 05/29/2025 20:39:34

Diagnostics Check		OK	20:33
Air Blank		0.000	20:33
Control Test		0.080	20:34
Air Blank		0.000	20:34
Subject Sample	# 1	0.000	20:35
Breath Volume		3.000	20:35
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.000	20:38
Breath Volume		2.714	20:38
Air Blank		0.000	20:38
Control Test		0.080	20:38
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 05/31/2025 22:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,RITCHIE,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: OROURKE,ROCKY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: PERRYMAN

DATE/TIME OF TEST: 05/30/2025 20:17:05

Diagnostics Check		OK	20:10
Air Blank		0.000	20:10
Control Test		0.079	20:11
Air Blank		0.000	20:11
Subject Sample	# 1	0.178	20:12
Breath Volume		2.148	20:12
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.174	20:15
Breath Volume		3.933	20:15
Air Blank		0.000	20:16
Control Test		0.079	20:16
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

Data Download Date/Time: 05/31/2025 22:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRY,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: GREENE,JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: PEOPLES

DATE/TIME OF TEST: 05/10/2025 15:36:49

Diagnostics Check		OK	15:30
Air Blank		0.000	15:31
Control Test		0.084	15:31
Air Blank		0.000	15:32
Subject Sample	# 1	0.000	15:32
Breath Volume		2.320	15:32
Air Blank		0.000	15:32
Air Blank		0.000	15:34
Subject Sample	# 2	0.000	15:35
Breath Volume		2.804	15:35
Air Blank		0.000	15:35
Control Test		0.083	15:36
Air Blank		0.000	15:36
Diagnostics Check		OK	15:36

Data Download Date/Time: 05/16/2025 10:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,MATHEW,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: PHILLIPS,MATHEW
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: LITTON

DATE/TIME OF TEST: 05/25/2025 22:30:36

Diagnostics Check		OK	22:27	
Air Blank		0.000	22:28	
Control Test		0.083	22:28	
Air Blank		0.000	22:29	
Subject Sample	# 1		22:29	Subject Test Refused
Breath Volume		0.000	22:29	
Air Blank		0.000	22:29	
Control Test		0.083	22:30	
Air Blank		0.000	22:30	
Diagnostics Check		OK	22:30	

Data Download Date/Time: 06/21/2025 18:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,MATHEW,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: PHILLIPS,MATHEW
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: WHITAKER

DATE/TIME OF TEST: 05/30/2025 00:25:40

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.084	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.281	00:20
Breath Volume		1.925	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.261	00:23
Breath Volume		1.238	00:23
Air Blank		0.000	00:24
Control Test		0.082	00:25
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 06/21/2025 18:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: WINTERS

DATE/TIME OF TEST: 05/13/2025 19:36:15

Diagnostics Check		OK	19:30
Air Blank		0.000	19:30
Control Test		0.082	19:31
Air Blank		0.000	19:31
Subject Sample	# 1	0.000	19:32
Breath Volume		3.136	19:32
Air Blank		0.000	19:32
Air Blank		0.000	19:34
Subject Sample	# 2	0.000	19:34
Breath Volume		3.421	19:34
Air Blank		0.000	19:35
Control Test		0.082	19:35
Air Blank		0.000	19:36
Diagnostics Check		OK	19:36

Data Download Date/Time: 05/22/2025 07:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVIS,MILTON
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 04/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: MOULTON

DATE/TIME OF TEST: 05/16/2025 20:38:01

Diagnostics Check		OK	20:31
Air Blank		0.000	20:32
Control Test		0.080	20:32
Air Blank		0.000	20:33
Subject Sample	# 1	0.243	20:33
Breath Volume		2.101	20:33
Air Blank		0.000	20:34
Air Blank		0.000	20:36
Subject Sample	# 2	0.249	20:36
Breath Volume		2.250	20:36
Air Blank		0.000	20:37
Control Test		0.080	20:37
Air Blank		0.000	20:37
Diagnostics Check		OK	20:38

Data Download Date/Time: 05/22/2025 07:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: REGO,T
ARRESTING OFFICER AGENCY: Bay Harbor Island PD
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SOZA

DATE/TIME OF TEST: 05/10/2025 02:19:11

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.148	02:14
Breath Volume		1.378	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:16
Subject Sample	# 2	0.166	02:17
Breath Volume		1.710	02:17
Air Blank		0.000	02:18
Control Test		0.081	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 05/26/2025 22:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: TURNER,KENYON
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 05/11/2025 06:26:46

Diagnostics Check		OK	06:19
Air Blank		0.000	06:19
Control Test		0.080	06:19
Air Blank		0.000	06:20
Subject Sample	# 1	0.000	06:21
Breath Volume		2.660	06:21
Air Blank		0.000	06:21
Air Blank		0.000	06:23
Subject Sample	# 2	0.000	06:25
Breath Volume		2.781	06:25
Air Blank		0.000	06:25
Control Test		0.081	06:26
Air Blank		0.000	06:26
Diagnostics Check		OK	06:26

Data Download Date/Time: 05/26/2025 22:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: TURNER,KENYON
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: APOLLON

DATE/TIME OF TEST: 05/17/2025 07:47:39

Diagnostics Check	OK	07:47	
Air Blank		07:47	RFI Detect
Air Blank		07:47	RFI Detect

Data Download Date/Time: 05/26/2025 22:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: TURNER,KENYON
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:47

SUBJECT'S LAST NAME: APOLLON

DATE/TIME OF TEST: 05/17/2025 08:16:25

Diagnostics Check		OK	08:07
Air Blank		0.000	08:07
Control Test		0.079	08:08
Air Blank		0.000	08:08
Subject Sample	# 1	0.200	08:11
Breath Volume		1.714	08:11
Air Blank		0.000	08:11
Air Blank		0.000	08:13
Subject Sample	# 2	0.213	08:14
Breath Volume		2.117	08:14
Air Blank		0.000	08:15
Control Test		0.075	08:15
Air Blank		0.000	08:16
Diagnostics Check		OK	08:16

Data Download Date/Time: 05/26/2025 22:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: SOLOMON,BERNARD
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: LOPEZ SEGURA

DATE/TIME OF TEST: 05/23/2025 23:38:15

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.080	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 1		23:36	Subject Test Refused
Breath Volume		0.000	23:36	
Air Blank		0.000	23:37	
Control Test		0.081	23:37	
Air Blank		0.000	23:38	
Diagnostics Check		OK	23:38	

Data Download Date/Time: 05/26/2025 22:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: FUCHECK

DATE/TIME OF TEST: 05/01/2025 17:43:54

Diagnostics Check		OK	17:40	
Air Blank		0.000	17:41	
Control Test		0.080	17:41	
Air Blank		0.000	17:42	
Subject Sample	# 1		17:42	Subject Test Refused
Breath Volume		0.000	17:42	
Air Blank		0.000	17:43	
Control Test		0.080	17:43	
Air Blank		0.000	17:43	
Diagnostics Check		OK	17:43	

Data Download Date/Time: 05/31/2025 18:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GREER,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/04/2025 04:20:43

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.078	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.162	04:15
Breath Volume		1.847	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:17
Subject Sample	# 2	0.158	04:19
Breath Volume		2.296	04:19
Air Blank		0.000	04:19
Control Test		0.078	04:20
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 05/31/2025 18:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: GNZALEZ-VILLEGAS

DATE/TIME OF TEST: 05/04/2025 06:48:45

Diagnostics Check		OK	06:42
Air Blank		0.000	06:42
Control Test		0.078	06:43
Air Blank		0.000	06:43
Subject Sample	# 1	0.131	06:44
Breath Volume		2.367	06:44
Air Blank		0.000	06:44
Air Blank		0.000	06:46
Subject Sample	# 2	0.133	06:47
Breath Volume		2.015	06:47
Air Blank		0.000	06:47
Control Test		0.078	06:48
Air Blank		0.000	06:48
Diagnostics Check		OK	06:48

Data Download Date/Time: 05/31/2025 18:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: RUCIO

DATE/TIME OF TEST: 05/05/2025 03:25:20

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.078	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.224	03:20
Breath Volume		2.601	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.214	03:23
Breath Volume		1.742	03:23
Air Blank		0.000	03:24
Control Test		0.077	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 05/31/2025 18:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIAGA,JULIO,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,JULIO
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: MURGOLO

DATE/TIME OF TEST: 05/05/2025 16:03:51

Diagnostics Check		OK	16:01	
Air Blank		0.000	16:01	
Control Test		0.079	16:01	
Air Blank		0.000	16:02	
Subject Sample	# 1		16:02	Subject Test Refused
Breath Volume		0.000	16:02	
Air Blank		0.000	16:02	
Control Test		0.079	16:03	
Air Blank		0.000	16:03	
Diagnostics Check		OK	16:03	

Data Download Date/Time: 05/31/2025 18:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: KINTZING

DATE/TIME OF TEST: 05/08/2025 21:21:43

Diagnostics Check		OK	21:18	
Air Blank		0.000	21:18	
Control Test		0.079	21:19	
Air Blank		0.000	21:19	
Subject Sample	# 1		21:20	Subject Test Refused
Breath Volume		0.000	21:20	
Air Blank		0.000	21:20	
Control Test		0.079	21:21	
Air Blank		0.000	21:21	
Diagnostics Check		OK	21:21	

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACCHIONE,VINCENT,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FERRO,T
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: TOLLNER

DATE/TIME OF TEST: 05/08/2025 22:31:53

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.079	22:26
Air Blank		0.000	22:26
Subject Sample	# 1	0.123	22:27
Breath Volume		2.558	22:27
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.122	22:30
Breath Volume		2.699	22:30
Air Blank		0.000	22:30
Control Test		0.079	22:31
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: SCHULTZ

DATE/TIME OF TEST: 05/13/2025 00:15:57

Diagnostics Check		OK	00:07
Air Blank		0.000	00:08
Control Test		0.079	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.173	00:09
Breath Volume		1.152	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.168	00:14
Breath Volume		1.125	00:14
Air Blank		0.000	00:15
Control Test		0.078	00:15
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,R,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CORDERO RETANA

DATE/TIME OF TEST: 05/14/2025 01:47:29

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.078	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.157	01:42
Breath Volume		2.746	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.156	01:45
Breath Volume		2.308	01:45
Air Blank		0.000	01:46
Control Test		0.077	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIAGA,JULIO,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COURINGTON,K
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: MILLETTE

DATE/TIME OF TEST: 05/15/2025 14:11:37

Diagnostics Check		OK	14:08	
Air Blank		0.000	14:09	
Control Test		0.078	14:09	
Air Blank		0.000	14:10	
Subject Sample	# 1		14:10	Subject Test Refused
Breath Volume		0.000	14:10	
Air Blank		0.000	14:10	
Control Test		0.077	14:11	
Air Blank		0.000	14:11	
Diagnostics Check		OK	14:11	

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BEST

DATE/TIME OF TEST: 05/16/2025 01:27:02

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.077	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.209	01:22
Breath Volume		2.628	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:24
Subject Sample	# 2	0.208	01:25
Breath Volume		1.597	01:25
Air Blank		0.000	01:26
Control Test		0.078	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:27

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: MENDEZ HERNANDEZ

DATE/TIME OF TEST: 05/16/2025 15:43:02

Diagnostics Check		OK	15:35
Air Blank		0.000	15:36
Control Test		0.078	15:36
Air Blank		0.000	15:37
Subject Sample	# 1	0.147	15:38
Breath Volume		2.457	15:38
Air Blank		0.000	15:39
Air Blank		0.000	15:41
Subject Sample	# 2	0.153	15:41
Breath Volume		1.921	15:41
Air Blank		0.000	15:42
Control Test		0.079	15:42
Air Blank		0.000	15:42
Diagnostics Check		OK	15:43

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/16/2025 22:40:50

Diagnostics Check		OK	22:29	
Air Blank		0.000	22:29	
Control Test		0.079	22:30	
Air Blank		0.000	22:30	
Subject Sample	# 1		22:33	No Sample Provided
Breath Volume		0.000	22:33	
Air Blank		0.000	22:34	
Air Blank		0.000	22:36	
Subject Sample	# 2		22:39	No Sample Provided
Breath Volume		0.000	22:39	
Air Blank		0.000	22:39	
Control Test		0.079	22:40	
Air Blank		0.000	22:40	
Diagnostics Check		OK	22:40	

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACCHIONE,VINCENT,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HOLLIFIELD,H
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: SCHICCHI

DATE/TIME OF TEST: 05/17/2025 02:08:27

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:05	
Control Test		0.079	02:05	
Air Blank		0.000	02:05	
Subject Sample	# 1		02:07	Subject Test Refused
Breath Volume		0.000	02:07	
Air Blank		0.000	02:07	
Control Test		0.079	02:07	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FERRO,T
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 05/17/2025 03:17:15

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.079	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:15	Subject Test Refused
Breath Volume		0.000	03:15	
Air Blank		0.000	03:16	
Control Test		0.079	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: TEROVOLAS

DATE/TIME OF TEST: 05/17/2025 04:19:51

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.079	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.098	04:15
Breath Volume		2.250	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.097	04:18
Breath Volume		2.093	04:18
Air Blank		0.000	04:18
Control Test		0.079	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: JENTZSCH

DATE/TIME OF TEST: 05/18/2025 02:04:59

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:02	
Control Test		0.079	02:02	
Air Blank		0.000	02:03	
Subject Sample	# 1		02:03	Subject Test Refused
Breath Volume		0.000	02:03	
Air Blank		0.000	02:04	
Control Test		0.078	02:04	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HILDEBRANDT

DATE/TIME OF TEST: 05/18/2025 04:17:09

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.078	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.155	04:12
Breath Volume		2.023	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:14
Subject Sample	# 2	0.140	04:15
Breath Volume		1.789	04:15
Air Blank		0.000	04:16
Control Test		0.078	04:16
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIAGA,JULIO,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,JULIO
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 05/19/2025 17:13:50

Diagnostics Check		OK	17:11	
Air Blank		0.000	17:11	
Control Test		0.078	17:11	
Air Blank		0.000	17:12	
Subject Sample	# 1		17:12	Subject Test Refused
Breath Volume		0.000	17:12	
Air Blank		0.000	17:12	
Control Test		0.078	17:13	
Air Blank		0.000	17:13	
Diagnostics Check		OK	17:13	

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:58

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 05/21/2025 15:27:27

Diagnostics Check		OK	15:18
Air Blank		0.000	15:18
Control Test		0.078	15:19
Air Blank		0.000	15:19
Subject Sample	# 1	0.133	15:21
Breath Volume		1.789	15:21
Air Blank		0.000	15:22
Air Blank		0.000	15:24
Subject Sample	# 2	0.122	15:25
Breath Volume		1.488	15:25
Air Blank		0.000	15:26
Control Test		0.076	15:26
Air Blank		0.000	15:27
Diagnostics Check		OK	15:27

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOJNOWSKY,IWONA,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: ROUSSEAU

DATE/TIME OF TEST: 05/23/2025 19:37:38

Diagnostics Check		OK	19:30
Air Blank		0.000	19:31
Control Test		0.079	19:31
Air Blank		0.000	19:32
Subject Sample	# 1	0.229	19:33
Breath Volume		2.621	19:33
Air Blank		0.000	19:33
Air Blank		0.000	19:35
Subject Sample	# 2	0.222	19:36
Breath Volume		2.875	19:36
Air Blank		0.000	19:36
Control Test		0.078	19:37
Air Blank		0.000	19:37
Diagnostics Check		OK	19:37

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOJNOWSKY,IWONA,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LOCKE,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MARTINEZ FONTAN

DATE/TIME OF TEST: 05/23/2025 22:38:29

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:36	
Control Test		0.079	22:36	
Air Blank		0.000	22:36	
Subject Sample	# 1		22:37	Subject Test Refused
Breath Volume		0.000	22:37	
Air Blank		0.000	22:37	
Control Test		0.079	22:37	
Air Blank		0.000	22:38	
Diagnostics Check		OK	22:38	

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: GIZZI

DATE/TIME OF TEST: 05/25/2025 04:44:51

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.079	04:37
Air Blank		0.000	04:38
Subject Sample	# 1	0.090	04:39
Breath Volume		2.234	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.100	04:43
Breath Volume		2.242	04:43
Air Blank		0.000	04:43
Control Test		0.081	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: PETRU,A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: HADDIX

DATE/TIME OF TEST: 05/25/2025 19:33:22

Diagnostics Check		OK	19:29	
Air Blank		0.000	19:30	
Control Test		0.079	19:30	
Air Blank		0.000	19:31	
Subject Sample	# 1		19:31	RFI Detect
Breath Volume		1.582	19:31	
Air Blank		0.000	19:32	
Control Test		0.079	19:32	
Air Blank		0.000	19:33	
Diagnostics Check		OK	19:33	

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: PETRU,A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HADDIX

DATE/TIME OF TEST: 05/25/2025 19:43:33

Diagnostics Check		OK	19:40	
Air Blank		0.000	19:41	
Control Test		0.079	19:41	
Air Blank		0.000	19:41	
Subject Sample	# 1		19:42	RFI Detect
Breath Volume		0.000	19:42	
Air Blank		0.000	19:42	
Control Test		0.079	19:42	
Air Blank		0.000	19:43	
Diagnostics Check		OK	19:43	

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,C,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: PETRU,A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HADDIX

DATE/TIME OF TEST: 05/25/2025 19:51:38

Diagnostics Check		OK	19:45
Air Blank		0.000	19:45
Control Test		0.079	19:46
Air Blank		0.000	19:46
Subject Sample	# 1	0.050	19:47
Breath Volume		2.468	19:47
Air Blank		0.000	19:47
Air Blank		0.000	19:49
Subject Sample	# 2	0.048	19:50
Breath Volume		2.488	19:50
Air Blank		0.000	19:50
Control Test		0.079	19:51
Air Blank		0.000	19:51
Diagnostics Check		OK	19:51

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,R,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: AMENDOLARE

DATE/TIME OF TEST: 05/26/2025 18:20:40

Diagnostics Check		OK	18:13
Air Blank		0.000	18:14
Control Test		0.079	18:14
Air Blank		0.000	18:15
Subject Sample	# 1	0.000	18:15
Breath Volume		1.496	18:15
Air Blank		0.000	18:16
Air Blank		0.000	18:18
Subject Sample	# 2	0.000	18:19
Breath Volume		2.046	18:19
Air Blank		0.000	18:19
Control Test		0.079	18:20
Air Blank		0.000	18:20
Diagnostics Check		OK	18:20

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,R,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: PITA

DATE/TIME OF TEST: 05/30/2025 18:34:46

Diagnostics Check		OK	18:24
Air Blank		0.000	18:24
Control Test		0.078	18:24
Air Blank		0.000	18:25
Subject Sample	# 1	0.170	18:27
Breath Volume		1.195	18:27
Air Blank		0.000	18:28
Air Blank		0.000	18:30
Subject Sample	# 2	0.172	18:33
Breath Volume		2.550	18:33
Air Blank		0.000	18:33
Control Test		0.077	18:34
Air Blank		0.000	18:34
Diagnostics Check		OK	18:34

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/31/2025 00:34:39

Diagnostics Check		OK	00:28
Air Blank		0.000	00:28
Control Test		0.078	00:29
Air Blank		0.000	00:29
Subject Sample	# 1	0.054	00:30
Breath Volume		3.457	00:30
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.055	00:33
Breath Volume		3.980	00:33
Air Blank		0.000	00:33
Control Test		0.078	00:34
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 05/31/2025 19:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: RIVERA MOLINA

DATE/TIME OF TEST: 05/31/2025 02:04:00

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:52	
Control Test		0.078	01:53	
Air Blank		0.000	01:53	
Subject Sample	# 1	0.134	01:56	Volume Not Met
Breath Volume		1.089	01:56	
Air Blank		0.000	01:57	
Air Blank		0.000	01:59	
Subject Sample	# 2	0.121	02:02	Volume Not Met
Breath Volume		0.250	02:02	
Air Blank		0.000	02:03	
Control Test		0.077	02:03	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 05/31/2025 19:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: WADDELL

DATE/TIME OF TEST: 05/05/2025 19:48:04

Diagnostics Check		OK	19:41
Air Blank		0.000	19:42
Control Test		0.080	19:42
Air Blank		0.000	19:42
Subject Sample	# 1	0.187	19:43
Breath Volume		1.332	19:43
Air Blank		0.000	19:44
Air Blank		0.000	19:45
Subject Sample	# 2	0.191	19:46
Breath Volume		1.269	19:46
Air Blank		0.000	19:47
Control Test		0.079	19:47
Air Blank		0.000	19:47
Diagnostics Check		OK	19:48

Data Download Date/Time: 05/21/2025 10:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KIRCHGRABER,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 05/05/2025 20:53:06

Diagnostics Check		OK	20:46
Air Blank		0.000	20:47
Control Test		0.080	20:47
Air Blank		0.000	20:47
Subject Sample	# 1	0.236	20:48
Breath Volume		1.582	20:48
Air Blank		0.000	20:49
Air Blank		0.000	20:50
Subject Sample	# 2	0.226	20:51
Breath Volume		1.800	20:51
Air Blank		0.000	20:52
Control Test		0.079	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 05/21/2025 10:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ROJAS TEJADA

DATE/TIME OF TEST: 05/09/2025 21:43:53

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:41	
Control Test		0.080	21:41	
Air Blank		0.000	21:42	
Subject Sample	# 1		21:42	Subject Test Refused
Breath Volume		0.000	21:42	
Air Blank		0.000	21:42	
Control Test		0.078	21:43	
Air Blank		0.000	21:43	
Diagnostics Check		OK	21:43	

Data Download Date/Time: 05/21/2025 10:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: GUTIERREZ RESENDIZ

DATE/TIME OF TEST: 05/13/2025 05:18:54

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.079	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.133	05:14
Breath Volume		2.296	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:16
Subject Sample	# 2	0.129	05:17
Breath Volume		1.433	05:17
Air Blank		0.000	05:17
Control Test		0.079	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 05/21/2025 10:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARINELLI,MATTHEW
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 05/14/2025 02:15:36

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.022	02:11
Breath Volume		2.265	02:11
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.021	02:14
Breath Volume		1.753	02:14
Air Blank		0.000	02:14
Control Test		0.079	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 05/21/2025 10:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: DIMARTINO

DATE/TIME OF TEST: 05/17/2025 02:09:45

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.080	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.000	02:05
Breath Volume		1.250	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.000	02:08
Breath Volume		1.828	02:08
Air Blank		0.000	02:08
Control Test		0.080	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 05/21/2025 10:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,JOHN,S
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: SPRADLEY

DATE/TIME OF TEST: 05/23/2025 19:16:31

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:15	
Control Test	0.000	19:15	Control Outside Tolerance
Air Blank	0.000	19:16	

Data Download Date/Time: 05/28/2025 12:05

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,JOHN,S
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: SPRADLEY

DATE/TIME OF TEST: 05/23/2025 19:28:49

Diagnostics Check	OK	19:27	
Air Blank	0.000	19:27	
Control Test	0.000	19:28	Control Outside Tolerance
Air Blank	0.000	19:28	

Data Download Date/Time: 05/28/2025 12:05

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,D
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: RICE,JOHN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: ADMIN
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 05/23/2025 22:55:49

Diagnostics Check		OK	22:52	
Air Blank		0.000	22:53	
Control Test		0.081	22:53	
Air Blank		0.000	22:53	
Subject Sample	# 1		22:54	RFI Detect
Breath Volume		0.000	22:54	
Air Blank		0.000	22:54	
Control Test		0.081	22:55	
Air Blank		0.000	22:55	
Diagnostics Check		OK	22:55	

Data Download Date/Time: 05/28/2025 12:05

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 05/02/2025 00:03:26

Diagnostics Check		OK	23:59	
Air Blank		0.000	23:59	
Control Test		0.081	00:00	
Air Blank		0.000	00:00	
Subject Sample	# 1		00:02	Subject Test Refused
Breath Volume		0.000	00:02	
Air Blank		0.000	00:02	
Control Test		0.082	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 06/11/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 05/02/2025 05:10:40

Diagnostics Check		OK	05:07	
Air Blank		0.000	05:07	
Control Test		0.082	05:07	
Air Blank		0.000	05:08	
Subject Sample	# 1		05:09	Subject Test Refused
Breath Volume		0.000	05:09	
Air Blank		0.000	05:09	
Control Test		0.081	05:10	
Air Blank		0.000	05:10	
Diagnostics Check		OK	05:10	

Data Download Date/Time: 06/11/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/02/2025 20:18:15

Diagnostics Check		OK	20:10
Air Blank		0.000	20:10
Control Test		0.081	20:10
Air Blank		0.000	20:11
Subject Sample	# 1	0.153	20:13
Breath Volume		2.226	20:13
Air Blank		0.000	20:14
Air Blank		0.000	20:16
Subject Sample	# 2	0.154	20:16
Breath Volume		2.328	20:16
Air Blank		0.000	20:17
Control Test		0.081	20:17
Air Blank		0.000	20:18
Diagnostics Check		OK	20:18

Data Download Date/Time: 06/11/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: FILSON

DATE/TIME OF TEST: 05/03/2025 01:57:08

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.081	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.156	01:51
Breath Volume		1.332	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.142	01:55
Breath Volume		1.210	01:55
Air Blank		0.000	01:56
Control Test		0.080	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 06/11/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/03/2025 03:22:46

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.081	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 1		03:21	Subject Test Refused
Breath Volume		0.000	03:21	
Air Blank		0.000	03:21	
Control Test		0.081	03:22	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 05/04/2025 17:48:03

Diagnostics Check		OK	17:41
Air Blank		0.000	17:42
Control Test		0.081	17:42
Air Blank		0.000	17:43
Subject Sample	# 1	0.135	17:43
Breath Volume		3.050	17:43
Air Blank		0.000	17:44
Air Blank		0.000	17:46
Subject Sample	# 2	0.137	17:46
Breath Volume		2.726	17:46
Air Blank		0.000	17:47
Control Test		0.080	17:47
Air Blank		0.000	17:47
Diagnostics Check		OK	17:48

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 05/04/2025 22:00:29

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:56	
Control Test		0.081	21:56	
Air Blank		0.000	21:57	
Subject Sample	# 1		21:59	Subject Test Refused
Breath Volume		0.000	21:59	
Air Blank		0.000	21:59	
Control Test		0.081	21:59	
Air Blank		0.000	22:00	
Diagnostics Check		OK	22:00	

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 05/04/2025 22:45:55

Diagnostics Check		OK	22:39
Air Blank		0.000	22:39
Control Test		0.081	22:40
Air Blank		0.000	22:40
Subject Sample	# 1	0.120	22:41
Breath Volume		1.832	22:41
Air Blank		0.000	22:42
Air Blank		0.000	22:43
Subject Sample	# 2	0.120	22:44
Breath Volume		1.128	22:44
Air Blank		0.000	22:45
Control Test		0.080	22:45
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: GJINECI

DATE/TIME OF TEST: 05/05/2025 04:46:26

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:42	
Control Test		0.082	04:43	
Air Blank		0.000	04:43	
Subject Sample	# 1		04:45	Subject Test Refused
Breath Volume		0.000	04:45	
Air Blank		0.000	04:45	
Control Test		0.082	04:45	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 05/05/2025 21:46:14

Diagnostics Check		OK	21:34	
Air Blank		0.000	21:35	
Control Test		0.082	21:35	
Air Blank		0.000	21:36	
Subject Sample	# 1		21:39	No Sample Provided
Breath Volume		0.000	21:39	
Air Blank		0.000	21:39	
Air Blank		0.000	21:41	
Subject Sample	# 2		21:44	No Sample Provided
Breath Volume		0.000	21:44	
Air Blank		0.000	21:45	
Control Test		0.081	21:45	
Air Blank		0.000	21:46	
Diagnostics Check		OK	21:46	

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/07/2025 19:08:37

Diagnostics Check		OK	18:56	
Air Blank		0.000	18:56	
Control Test		0.081	18:57	
Air Blank		0.000	18:57	
Subject Sample	# 1		19:00	No Sample Provided
Breath Volume		0.000	19:00	
Air Blank		0.000	19:01	
Air Blank		0.000	19:03	
Subject Sample	# 2	0.205	19:03	
Breath Volume		1.664	19:03	
Air Blank		0.000	19:04	
Air Blank		0.000	19:06	
Subject Sample	# 3	0.213	19:07	
Breath Volume		1.187	19:07	
Air Blank		0.000	19:07	
Control Test		0.081	19:08	
Air Blank		0.000	19:08	
Diagnostics Check		OK	19:08	

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 05/08/2025 03:51:09

Diagnostics Check		OK	03:47	
Air Blank		0.000	03:48	
Control Test		0.081	03:48	
Air Blank		0.000	03:49	
Subject Sample	# 1		03:49	Subject Test Refused
Breath Volume		0.000	03:49	
Air Blank		0.000	03:50	
Control Test		0.081	03:50	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: HELFRICH

DATE/TIME OF TEST: 05/09/2025 01:17:52

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:14	
Control Test		0.081	01:14	
Air Blank		0.000	01:15	
Subject Sample	# 1		01:16	Subject Test Refused
Breath Volume		0.000	01:16	
Air Blank		0.000	01:16	
Control Test		0.082	01:17	
Air Blank		0.000	01:17	
Diagnostics Check		OK	01:17	

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: LERMOND

DATE/TIME OF TEST: 05/09/2025 22:42:23

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.081	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.201	22:37
Breath Volume		2.640	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.206	22:40
Breath Volume		2.007	22:40
Air Blank		0.000	22:41
Control Test		0.081	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 05/10/2025 04:57:46

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:53	
Control Test		0.081	04:53	
Air Blank		0.000	04:53	
Subject Sample	# 1		04:56	Subject Test Refused
Breath Volume		0.000	04:56	
Air Blank		0.000	04:56	
Control Test		0.082	04:57	
Air Blank		0.000	04:57	
Diagnostics Check		OK	04:57	

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORRIGAN,MATTHEW,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: SWEARENGIN

DATE/TIME OF TEST: 05/12/2025 00:01:04

Diagnostics Check		OK	23:56	
Air Blank		0.000	23:56	
Control Test		0.081	23:57	
Air Blank		0.000	23:57	
Subject Sample	# 1		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	00:00	
Control Test		0.082	00:00	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 05/12/2025 20:51:27

Diagnostics Check		OK	20:48	
Air Blank		0.000	20:48	
Control Test		0.081	20:48	
Air Blank		0.000	20:49	
Subject Sample	# 1		20:50	Subject Test Refused
Breath Volume		0.000	20:50	
Air Blank		0.000	20:50	
Control Test		0.081	20:50	
Air Blank		0.000	20:51	
Diagnostics Check		OK	20:51	

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/13/2025 16:45:43

Diagnostics Check		OK	16:35
Air Blank		0.000	16:35
Control Test		0.080	16:35
Air Blank		0.000	16:36
Subject Sample	# 1	0.136	16:37
Breath Volume		1.976	16:37
Air Blank		0.000	16:38
Air Blank		0.000	16:40
Subject Sample	# 2	0.163	16:41
Breath Volume		1.187	16:41
Air Blank		0.000	16:42
Air Blank		0.000	16:43
Subject Sample	# 3	0.151	16:44
Breath Volume		2.222	16:44
Air Blank		0.000	16:44
Control Test		0.078	16:45
Air Blank		0.000	16:45
Diagnostics Check		OK	16:45

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JULIO,C
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZMITH,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: JIMENEZ DURAN

DATE/TIME OF TEST: 05/15/2025 02:31:11

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.082	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.235	02:26
Breath Volume		1.296	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.235	02:29
Breath Volume		1.718	02:29
Air Blank		0.000	02:30
Control Test		0.081	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 06/11/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/15/2025 03:23:31

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:20	
Control Test		0.082	03:21	
Air Blank		0.000	03:21	
Subject Sample	# 1		03:22	Subject Test Refused
Breath Volume		0.000	03:22	
Air Blank		0.000	03:22	
Control Test		0.082	03:22	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/15/2025 23:10:15

Diagnostics Check		OK	23:01	
Air Blank		0.000	23:02	
Control Test		0.081	23:02	
Air Blank		0.000	23:03	
Subject Sample	# 1		23:06	No Sample Provided
Breath Volume		0.000	23:06	
Air Blank		0.000	23:06	
Air Blank		0.000	23:08	
Subject Sample	# 2		23:08	Subject Test Refused
Breath Volume		0.000	23:08	
Air Blank		0.000	23:09	
Control Test		0.081	23:09	
Air Blank		0.000	23:10	
Diagnostics Check		OK	23:10	

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BARTON

DATE/TIME OF TEST: 05/16/2025 02:26:53

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:23	
Control Test		0.081	02:24	
Air Blank		0.000	02:24	
Subject Sample	# 1		02:25	Subject Test Refused
Breath Volume		0.000	02:25	
Air Blank		0.000	02:25	
Control Test		0.082	02:26	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 05/16/2025 20:47:37

Diagnostics Check		OK	20:41
Air Blank		0.000	20:41
Control Test		0.082	20:42
Air Blank		0.000	20:42
Subject Sample	# 1	0.039	20:43
Breath Volume		2.121	20:43
Air Blank		0.000	20:43
Air Blank		0.000	20:45
Subject Sample	# 2	0.038	20:46
Breath Volume		1.312	20:46
Air Blank		0.000	20:46
Control Test		0.082	20:47
Air Blank		0.000	20:47
Diagnostics Check		OK	20:47

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RAUDA RAUDA

DATE/TIME OF TEST: 05/16/2025 23:49:55

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.081	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.113	23:45
Breath Volume		2.109	23:45
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.116	23:48
Breath Volume		1.871	23:48
Air Blank		0.000	23:49
Control Test		0.079	23:49
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: COCKERELL

DATE/TIME OF TEST: 05/17/2025 03:37:02

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:33	
Control Test		0.082	03:34	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:35	Subject Test Refused
Breath Volume		0.000	03:35	
Air Blank		0.000	03:36	
Control Test		0.082	03:36	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: PRIZER

DATE/TIME OF TEST: 05/17/2025 22:06:57

Diagnostics Check		OK	22:01	
Air Blank		0.000	22:02	
Control Test		0.082	22:02	
Air Blank		0.000	22:03	
Subject Sample	# 1		22:05	Subject Test Refused
Breath Volume		0.000	22:05	
Air Blank		0.000	22:06	
Control Test		0.082	22:06	
Air Blank		0.000	22:06	
Diagnostics Check		OK	22:06	

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: SMILEY

DATE/TIME OF TEST: 05/18/2025 02:23:16

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.081	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.134	02:18
Breath Volume		2.875	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.134	02:21
Breath Volume		2.906	02:21
Air Blank		0.000	02:22
Control Test		0.081	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 05/18/2025 22:06:13

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:03	
Control Test		0.082	22:03	
Air Blank		0.000	22:04	
Subject Sample	# 1		22:04	Subject Test Refused
Breath Volume		0.000	22:04	
Air Blank		0.000	22:05	
Control Test		0.082	22:05	
Air Blank		0.000	22:06	
Diagnostics Check		OK	22:06	

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JULIO,C
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MASLE,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SPICER

DATE/TIME OF TEST: 05/20/2025 03:40:37

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.081	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.129	03:36
Breath Volume		1.609	03:36
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.125	03:39
Breath Volume		2.359	03:39
Air Blank		0.000	03:39
Control Test		0.081	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,CARLOS,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MASLE,ADAM
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: CLARKSON

DATE/TIME OF TEST: 05/21/2025 00:45:28

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:42	
Control Test		0.081	00:42	
Air Blank		0.000	00:43	
Subject Sample	# 1		00:44	Subject Test Refused
Breath Volume		0.000	00:44	
Air Blank		0.000	00:44	
Control Test		0.082	00:44	
Air Blank		0.000	00:45	
Diagnostics Check		OK	00:45	

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,OFC,C
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 05/21/2025 15:22:45

Diagnostics Check		OK	15:14
Air Blank		0.000	15:15
Control Test		0.082	15:15
Air Blank		0.000	15:15
Subject Sample	# 1	0.000	15:18
Breath Volume		1.648	15:18
Air Blank		0.000	15:18
Air Blank		0.000	15:20
Subject Sample	# 2	0.000	15:21
Breath Volume		1.597	15:21
Air Blank		0.000	15:21
Control Test		0.080	15:22
Air Blank		0.000	15:22
Diagnostics Check		OK	15:22

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,CLAYTON,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 05/21/2025 21:16:54

Diagnostics Check		OK	21:06
Air Blank		0.000	21:06
Control Test		0.081	21:06
Air Blank		0.000	21:07
Subject Sample	# 1	0.121	21:10
Breath Volume		1.625	21:10
Air Blank		0.000	21:11
Air Blank		0.000	21:13
Subject Sample	# 2	0.125	21:15
Breath Volume		1.644	21:15
Air Blank		0.000	21:15
Control Test		0.079	21:16
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: ONEILL

DATE/TIME OF TEST: 05/24/2025 02:55:37

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.082	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.000	02:51
Breath Volume		1.378	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.000	02:54
Breath Volume		1.378	02:54
Air Blank		0.000	02:54
Control Test		0.082	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: FITZ

DATE/TIME OF TEST: 05/24/2025 23:06:59

Diagnostics Check		OK	23:02	
Air Blank		0.000	23:02	
Control Test		0.082	23:03	
Air Blank		0.000	23:03	
Subject Sample	# 1		23:05	Subject Test Refused
Breath Volume		0.000	23:05	
Air Blank		0.000	23:06	
Control Test		0.083	23:06	
Air Blank		0.000	23:06	
Diagnostics Check		OK	23:06	

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: CABRAL

DATE/TIME OF TEST: 05/26/2025 00:23:23

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.082	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.166	00:18
Breath Volume		1.558	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:20
Subject Sample	# 2	0.175	00:21
Breath Volume		1.269	00:21
Air Blank		0.000	00:22
Control Test		0.079	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 06/11/2025 09:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 05/26/2025 01:08:33

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.081	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.163	01:03
Breath Volume		2.195	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.159	01:06
Breath Volume		1.972	01:06
Air Blank		0.000	01:07
Control Test		0.081	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,CARLOS,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: SPECHT

DATE/TIME OF TEST: 05/27/2025 04:07:23

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:04	
Control Test		0.082	04:04	
Air Blank		0.000	04:05	
Subject Sample	# 1		04:06	Subject Test Refused
Breath Volume		0.000	04:06	
Air Blank		0.000	04:06	
Control Test		0.082	04:06	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:57

SUBJECT'S LAST NAME: PLAUCHE

DATE/TIME OF TEST: 05/27/2025 15:26:17

Diagnostics Check		OK	15:18
Air Blank		0.000	15:18
Control Test		0.082	15:19
Air Blank		0.000	15:19
Subject Sample	# 1	0.000	15:20
Breath Volume		1.367	15:20
Air Blank		0.000	15:20
Air Blank		0.000	15:22
Subject Sample	# 2	0.000	15:24
Breath Volume		1.964	15:24
Air Blank		0.000	15:25
Control Test		0.082	15:25
Air Blank		0.000	15:26
Diagnostics Check		OK	15:26

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: BALSON

DATE/TIME OF TEST: 05/27/2025 19:19:39

Diagnostics Check		OK	19:15	
Air Blank		0.000	19:16	
Control Test		0.082	19:16	
Air Blank		0.000	19:17	
Subject Sample	# 1		19:18	Subject Test Refused
Breath Volume		0.000	19:18	
Air Blank		0.000	19:18	
Control Test		0.082	19:19	
Air Blank		0.000	19:19	
Diagnostics Check		OK	19:19	

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BOBB

DATE/TIME OF TEST: 05/29/2025 22:15:45

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:11	
Control Test		0.082	22:11	
Air Blank		0.000	22:12	
Subject Sample	# 1		22:14	Subject Test Refused
Breath Volume		0.000	22:14	
Air Blank		0.000	22:14	
Control Test		0.081	22:15	
Air Blank		0.000	22:15	
Diagnostics Check		OK	22:15	

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/30/2025 03:33:46

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:26	
Control Test		0.082	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 1	0.032	03:30	
Breath Volume		1.480	03:30	
Air Blank		0.000	03:30	
Air Blank			03:32	RFI Detect
Air Blank		0.000	03:32	
Control Test		0.083	03:33	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:33	

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/30/2025 03:42:37

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.081	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.033	03:37
Breath Volume		1.101	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.030	03:41
Breath Volume		2.050	03:41
Air Blank		0.000	03:41
Control Test		0.080	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 05/30/2025 03:55:57

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:52	
Control Test		0.080	03:53	
Air Blank		0.000	03:53	
Subject Sample	# 1		03:54	Subject Test Refused
Breath Volume		0.000	03:54	
Air Blank		0.000	03:55	
Control Test		0.081	03:55	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCALL,J
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: TEMPLETON

DATE/TIME OF TEST: 05/30/2025 11:23:41

Diagnostics Check		OK	11:17
Air Blank		0.000	11:17
Control Test		0.082	11:18
Air Blank		0.000	11:18
Subject Sample	# 1	0.000	11:19
Breath Volume		3.019	11:19
Air Blank		0.000	11:19
Air Blank		0.000	11:21
Subject Sample	# 2	0.000	11:22
Breath Volume		2.265	11:22
Air Blank		0.000	11:22
Control Test		0.082	11:23
Air Blank		0.000	11:23
Diagnostics Check		OK	11:23

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:47

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 05/30/2025 15:13:37

Diagnostics Check		OK	15:07
Air Blank		0.000	15:08
Control Test		0.081	15:08
Air Blank		0.000	15:09
Subject Sample	# 1	0.000	15:09
Breath Volume		1.507	15:09
Air Blank		0.000	15:09
Air Blank		0.000	15:11
Subject Sample	# 2	0.000	15:12
Breath Volume		2.058	15:12
Air Blank		0.000	15:12
Control Test		0.079	15:13
Air Blank		0.000	15:13
Diagnostics Check		OK	15:13

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 05/30/2025 23:55:28

Diagnostics Check		OK	23:50	
Air Blank		0.000	23:50	
Control Test		0.081	23:50	
Air Blank		0.000	23:51	
Subject Sample	# 1		23:54	Subject Test Refused
Breath Volume		0.000	23:54	
Air Blank		0.000	23:54	
Control Test		0.082	23:54	
Air Blank		0.000	23:55	
Diagnostics Check		OK	23:55	

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BICKING

DATE/TIME OF TEST: 05/31/2025 02:43:25

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.081	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.128	02:38
Breath Volume		2.417	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.124	02:41
Breath Volume		1.933	02:41
Air Blank		0.000	02:42
Control Test		0.081	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 05/31/2025 04:16:08

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:12	
Control Test		0.081	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1		04:14	Subject Test Refused
Breath Volume		0.000	04:14	
Air Blank		0.000	04:15	
Control Test		0.081	04:15	
Air Blank		0.000	04:16	
Diagnostics Check		OK	04:16	

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: FEASTER

DATE/TIME OF TEST: 05/31/2025 05:20:54

Diagnostics Check		OK	05:05	
Air Blank		0.000	05:06	
Control Test		0.081	05:06	
Air Blank		0.000	05:07	
Subject Sample	# 1	0.157	05:08	
Breath Volume		1.308	05:08	
Air Blank		0.000	05:08	
Air Blank		0.000	05:10	
Subject Sample	# 2	0.142	05:13	Volume Not Met
Breath Volume		0.546	05:13	
Air Blank		0.000	05:14	
Air Blank		0.000	05:16	
Subject Sample	# 3	0.037	05:19	Volume Not Met
Breath Volume		0.492	05:19	
Air Blank		0.000	05:19	
Control Test		0.080	05:20	
Air Blank		0.000	05:20	
Diagnostics Check		OK	05:20	

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALENDA,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: BRADEN,CODY
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: KRESGE

DATE/TIME OF TEST: 05/14/2025 20:37:36

Diagnostics Check		OK	20:31
Air Blank		0.000	20:31
Control Test		0.078	20:31
Air Blank		0.000	20:32
Subject Sample	# 1	0.149	20:32
Breath Volume		1.296	20:32
Air Blank		0.000	20:33
Air Blank		0.000	20:35
Subject Sample	# 2	0.153	20:36
Breath Volume		1.671	20:36
Air Blank		0.000	20:36
Control Test		0.079	20:37
Air Blank		0.000	20:37
Diagnostics Check		OK	20:37

Data Download Date/Time: 05/30/2025 17:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALENDA,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: PAULL,JAYDON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: LEONE

DATE/TIME OF TEST: 05/21/2025 00:51:30

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.077	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.179	00:47
Breath Volume		3.855	00:47
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.176	00:49
Breath Volume		3.636	00:49
Air Blank		0.000	00:50
Control Test		0.079	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 05/30/2025 17:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JASON,A
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MCBRIDE,CRYSTA
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 13:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: URQUHART

DATE/TIME OF TEST: 05/23/2025 15:27:01

Diagnostics Check		OK	15:21
Air Blank		0.000	15:21
Control Test		0.079	15:21
Air Blank		0.000	15:22
Subject Sample	# 1	0.000	15:22
Breath Volume		2.234	15:22
Air Blank		0.000	15:23
Air Blank		0.000	15:25
Subject Sample	# 2	0.000	15:25
Breath Volume		2.320	15:25
Air Blank		0.000	15:26
Control Test		0.077	15:26
Air Blank		0.000	15:26
Diagnostics Check		OK	15:27

Data Download Date/Time: 05/30/2025 17:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALENDA,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: CALENDA,NICHOLAS
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: KLEMM

DATE/TIME OF TEST: 05/27/2025 01:18:00

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.078	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.034	01:13
Breath Volume		1.488	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:15
Subject Sample	# 2	0.033	01:16
Breath Volume		2.183	01:16
Air Blank		0.000	01:17
Control Test		0.077	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:18

Data Download Date/Time: 05/30/2025 17:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZELL,BENJAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FELLOWS,P
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 04/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: DIGUER

DATE/TIME OF TEST: 05/30/2025 00:28:44

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.079	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.000	00:24
Breath Volume		1.679	00:24
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.000	00:27
Breath Volume		2.179	00:27
Air Blank		0.000	00:27
Control Test		0.079	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 05/31/2025 08:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SENECAL,JOSHUA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: SYPOLT

DATE/TIME OF TEST: 05/03/2025 23:03:33

Diagnostics Check		OK	22:56
Air Blank		0.000	22:57
Control Test		0.082	22:57
Air Blank		0.000	22:57
Subject Sample	# 1	0.248	22:58
Breath Volume		3.558	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.257	23:01
Breath Volume		3.937	23:01
Air Blank		0.000	23:02
Control Test		0.082	23:02
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 05/30/2025 09:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SANTOS,KEELAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/10/2025 18:01:24

Diagnostics Check		OK	17:54
Air Blank		0.000	17:55
Control Test		0.083	17:55
Air Blank		0.000	17:55
Subject Sample	# 1	0.248	17:56
Breath Volume		2.308	17:56
Air Blank		0.000	17:57
Air Blank		0.000	17:59
Subject Sample	# 2	0.257	17:59
Breath Volume		2.230	17:59
Air Blank		0.000	18:00
Control Test		0.082	18:00
Air Blank		0.000	18:01
Diagnostics Check		OK	18:01

Data Download Date/Time: 05/30/2025 09:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSH,WESLEY,T
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BOWMAN,MAXWELL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 05/12/2025 00:17:24

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.082	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.212	00:12
Breath Volume		2.832	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.229	00:15
Breath Volume		2.312	00:15
Air Blank		0.000	00:16
Control Test		0.081	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 05/30/2025 09:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,KEELAN,D
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SANTOS,KEELAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/12/2025 21:37:18

Diagnostics Check		OK	21:30
Air Blank		0.000	21:31
Control Test		0.082	21:31
Air Blank		0.000	21:32
Subject Sample	# 1	0.309	21:32
Breath Volume		2.660	21:32
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.318	21:35
Breath Volume		2.484	21:35
Air Blank		0.000	21:36
Control Test		0.082	21:36
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 05/30/2025 09:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,RYAN,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TORRES,L
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:31

SUBJECT'S LAST NAME: SIRONEN

DATE/TIME OF TEST: 05/15/2025 17:02:44

Diagnostics Check		OK	16:56	
Air Blank		0.000	16:57	
Control Test		0.082	16:57	
Air Blank		0.000	16:57	
Subject Sample	# 1	0.169	16:58	
Breath Volume		1.363	16:58	
Air Blank		0.000	16:59	
Air Blank		0.000	17:01	
Subject Sample	# 2		17:01	RFI Detect
Breath Volume		0.000	17:01	
Air Blank		0.000	17:01	
Control Test		0.083	17:02	
Air Blank		0.000	17:02	
Diagnostics Check		OK	17:02	

Data Download Date/Time: 05/30/2025 09:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,RYAN,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TORRES,L
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:31

SUBJECT'S LAST NAME: SIRONEN

DATE/TIME OF TEST: 05/15/2025 17:15:32

Diagnostics Check		OK	17:06
Air Blank		0.000	17:07
Control Test		0.082	17:07
Air Blank		0.000	17:08
Subject Sample	# 1	0.164	17:09
Breath Volume		1.335	17:09
Air Blank		0.000	17:10
Air Blank		0.000	17:11
Subject Sample	# 2	0.163	17:13
Breath Volume		1.734	17:13
Air Blank		0.000	17:14
Control Test		0.082	17:14
Air Blank		0.000	17:15
Diagnostics Check		OK	17:15

Data Download Date/Time: 05/30/2025 09:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,KEELAN,D
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: FOSTER,TYLER
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: SCHOCK

DATE/TIME OF TEST: 05/17/2025 01:28:25

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.083	01:22
Air Blank		0.000	01:23
Subject Sample	# 1	0.190	01:23
Breath Volume		1.527	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.185	01:26
Breath Volume		2.343	01:26
Air Blank		0.000	01:27
Control Test		0.082	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 05/30/2025 09:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,KEELAN,D
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SANTOS,KEELAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 05/21/2025 17:37:34

Diagnostics Check		OK	17:31
Air Blank		0.000	17:31
Control Test		0.083	17:31
Air Blank		0.000	17:32
Subject Sample	# 1	0.000	17:32
Breath Volume		1.542	17:32
Air Blank		0.000	17:33
Air Blank		0.000	17:35
Subject Sample	# 2	0.000	17:36
Breath Volume		1.398	17:36
Air Blank		0.000	17:36
Control Test		0.082	17:36
Air Blank		0.000	17:37
Diagnostics Check		OK	17:37

Data Download Date/Time: 05/30/2025 09:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,KEELAN,D
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SPE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: DOBBINS

DATE/TIME OF TEST: 05/21/2025 23:24:37

Diagnostics Check		OK	23:21	
Air Blank		0.000	23:21	
Control Test		0.082	23:22	
Air Blank		0.000	23:22	
Subject Sample	# 1		23:23	Subject Test Refused
Breath Volume		0.000	23:23	
Air Blank		0.000	23:23	
Control Test		0.081	23:24	
Air Blank		0.000	23:24	
Diagnostics Check		OK	23:24	

Data Download Date/Time: 05/30/2025 09:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,KEELAN,D
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SEPE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DOBBINS

DATE/TIME OF TEST: 05/21/2025 23:47:37

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.082	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.234	23:42
Breath Volume		1.863	23:42
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.219	23:45
Breath Volume		1.613	23:45
Air Blank		0.000	23:46
Control Test		0.081	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 05/30/2025 09:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,KEELAN,D
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SENECAL,JOSHUA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: TUMBLIN

DATE/TIME OF TEST: 05/22/2025 03:24:40

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.082	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.097	03:19
Breath Volume		1.406	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.092	03:23
Breath Volume		1.164	03:23
Air Blank		0.000	03:23
Control Test		0.082	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 05/30/2025 09:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: LLINAS,JOEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MAZZA

DATE/TIME OF TEST: 05/26/2025 21:12:09

Diagnostics Check		OK	21:05
Air Blank		0.000	21:06
Control Test		0.083	21:06
Air Blank		0.000	21:06
Subject Sample	# 1	0.183	21:07
Breath Volume		2.546	21:07
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.179	21:10
Breath Volume		2.671	21:10
Air Blank		0.000	21:11
Control Test		0.083	21:11
Air Blank		0.000	21:12
Diagnostics Check		OK	21:12

Data Download Date/Time: 05/30/2025 09:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SENECAL,JOSHUA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 05/27/2025 03:38:11

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.082	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.000	03:33
Breath Volume		4.523	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.000	03:36
Breath Volume		4.820	03:36
Air Blank		0.000	03:37
Control Test		0.083	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 05/30/2025 09:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: AYALA,JUAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 05/29/2025 00:33:18

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:20	
Control Test		0.082	00:21	
Air Blank		0.000	00:21	
Subject Sample	# 1		00:24	No Sample Provided
Breath Volume		0.000	00:24	
Air Blank		0.000	00:25	
Air Blank		0.000	00:27	
Subject Sample	# 2	0.233	00:28	
Breath Volume		2.757	00:28	
Air Blank		0.000	00:29	
Air Blank		0.000	00:30	
Subject Sample	# 3	0.214	00:31	
Breath Volume		2.484	00:31	
Air Blank		0.000	00:32	
Control Test		0.080	00:32	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 05/30/2025 09:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: AYALA,JUAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 05/29/2025 00:45:20

Diagnostics Check		OK	00:38
Air Blank		0.000	00:38
Control Test		0.082	00:39
Air Blank		0.000	00:39
Subject Sample	# 1	0.219	00:40
Breath Volume		3.007	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:42
Subject Sample	# 2	0.221	00:43
Breath Volume		3.003	00:43
Air Blank		0.000	00:44
Control Test		0.082	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 05/30/2025 09:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: LANGLEY

DATE/TIME OF TEST: 05/05/2025 19:36:29

Diagnostics Check		OK	19:33	
Air Blank		0.000	19:33	
Control Test		0.081	19:34	
Air Blank		0.000	19:34	
Subject Sample	# 1		19:35	Subject Test Refused
Breath Volume		0.000	19:35	
Air Blank		0.000	19:35	
Control Test		0.080	19:35	
Air Blank		0.000	19:36	
Diagnostics Check		OK	19:36	

Data Download Date/Time: 05/29/2025 17:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 05/06/2025 03:38:00

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:35	
Control Test		0.081	03:35	
Air Blank		0.000	03:36	
Subject Sample	# 1	0.169	03:36	
Breath Volume		3.121	03:36	
Air Blank			03:37	Purge Fail
Air Blank		0.000	03:38	

Data Download Date/Time: 05/29/2025 17:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 05/06/2025 03:41:46

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.074	03:41	Control Outside Tolerance
Air Blank	0.000	03:41	

Data Download Date/Time: 05/29/2025 17:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 05/06/2025 03:49:44

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.079	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.179	03:44
Breath Volume		3.253	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.173	03:47
Breath Volume		3.023	03:47
Air Blank		0.000	03:48
Control Test		0.080	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 05/29/2025 17:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: HOERL

DATE/TIME OF TEST: 05/09/2025 03:40:10

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.081	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.170	03:35
Breath Volume		2.062	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.170	03:38
Breath Volume		1.968	03:38
Air Blank		0.000	03:39
Control Test		0.080	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 05/29/2025 17:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: DRAKUS

DATE/TIME OF TEST: 05/15/2025 19:29:48

Diagnostics Check		OK	19:21
Air Blank		0.000	19:21
Control Test		0.081	19:21
Air Blank		0.000	19:22
Subject Sample	# 1	0.077	19:22
Breath Volume		1.644	19:22
Air Blank		0.000	19:23
Air Blank		0.000	19:25
Subject Sample	# 2	0.083	19:27
Breath Volume		3.515	19:27
Air Blank		0.000	19:28
Control Test		0.081	19:29
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

Data Download Date/Time: 05/29/2025 17:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/03/2025 01:29:23

Diagnostics Check		OK	01:21
Air Blank		0.000	01:22
Control Test		0.077	01:22
Air Blank		0.000	01:23
Subject Sample	# 1	0.152	01:23
Breath Volume		1.796	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.157	01:27
Breath Volume		1.218	01:27
Air Blank		0.000	01:28
Control Test		0.078	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 06/05/2025 02:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GARCIA,A
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: GOMEZ VALENCIA

DATE/TIME OF TEST: 05/03/2025 04:45:58

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:31	
Control Test		0.080	04:31	
Air Blank		0.000	04:32	
Subject Sample	# 1	0.130	04:35	Volume Not Met
Breath Volume		0.425	04:35	
Air Blank		0.000	04:36	
Air Blank		0.000	04:37	
Subject Sample	# 2	0.134	04:39	
Breath Volume		1.375	04:39	
Air Blank		0.000	04:40	
Air Blank		0.000	04:41	
Subject Sample	# 3	0.125	04:44	
Breath Volume		1.664	04:44	
Air Blank		0.000	04:45	
Control Test		0.080	04:45	
Air Blank		0.000	04:45	
Diagnostics Check		OK	04:45	

Data Download Date/Time: 06/05/2025 02:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: IBANEZ,H
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: VELAZQUEZ TOIRAC

DATE/TIME OF TEST: 05/03/2025 05:14:35

Diagnostics Check		OK	05:11	
Air Blank		0.000	05:12	
Control Test		0.077	05:12	
Air Blank		0.000	05:12	
Subject Sample	# 1		05:13	Subject Test Refused
Breath Volume		0.000	05:13	
Air Blank		0.000	05:13	
Control Test		0.079	05:13	
Air Blank		0.000	05:14	
Diagnostics Check		OK	05:14	

Data Download Date/Time: 06/05/2025 02:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GONZALEZ,L
ARRESTING OFFICER AGENCY: WEST MIAMI PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: TORREZ ROCA

DATE/TIME OF TEST: 05/24/2025 04:00:16

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:55
Subject Sample	# 1	0.191	03:55
Breath Volume		2.226	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.181	03:58
Breath Volume		2.210	03:58
Air Blank		0.000	03:59
Control Test		0.077	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 06/05/2025 02:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: GUTIERREZ HERNANDEZ

DATE/TIME OF TEST: 05/24/2025 05:38:22

Diagnostics Check		OK	05:35	
Air Blank		0.000	05:35	
Control Test		0.079	05:36	
Air Blank		0.000	05:36	
Subject Sample	# 1		05:36	Subject Test Refused
Breath Volume		0.000	05:36	
Air Blank		0.000	05:37	
Control Test		0.079	05:37	
Air Blank		0.000	05:38	
Diagnostics Check		OK	05:38	

Data Download Date/Time: 06/05/2025 02:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: SANDOVAL LANDEROS

DATE/TIME OF TEST: 05/31/2025 00:25:54

Diagnostics Check		OK	00:22	
Air Blank		0.000	00:23	
Control Test		0.079	00:23	
Air Blank		0.000	00:24	
Subject Sample	# 1		00:24	Subject Test Refused
Breath Volume		0.000	00:24	
Air Blank		0.000	00:24	
Control Test		0.079	00:25	
Air Blank		0.000	00:25	
Diagnostics Check		OK	00:25	

Data Download Date/Time: 06/05/2025 02:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: CHURCH

DATE/TIME OF TEST: 05/10/2025 03:34:53

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.079	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.050	03:30
Breath Volume		1.796	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:32
Subject Sample	# 2	0.049	03:33
Breath Volume		1.906	03:33
Air Blank		0.000	03:33
Control Test		0.079	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 05/27/2025 11:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 05/11/2025 05:49:11

Diagnostics Check		OK	05:42
Air Blank		0.000	05:43
Control Test		0.080	05:43
Air Blank		0.000	05:44
Subject Sample	# 1	0.034	05:44
Breath Volume		1.503	05:44
Air Blank		0.000	05:45
Air Blank		0.000	05:46
Subject Sample	# 2	0.036	05:47
Breath Volume		2.035	05:47
Air Blank		0.000	05:48
Control Test		0.079	05:48
Air Blank		0.000	05:49
Diagnostics Check		OK	05:49

Data Download Date/Time: 05/27/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ASTACIO ORTIZ

DATE/TIME OF TEST: 05/24/2025 03:50:05

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.079	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.186	03:45
Breath Volume		2.062	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:47
Subject Sample	# 2	0.183	03:48
Breath Volume		1.718	03:48
Air Blank		0.000	03:49
Control Test		0.077	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 05/27/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTO,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: RESTO,LUIS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 05/01/2025 01:15:33

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.076	01:10
Breath Volume		4.144	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.068	01:13
Breath Volume		2.070	01:13
Air Blank		0.000	01:14
Control Test		0.077	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 05/15/2025 20:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTAMARIA,LEONARDO,T
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SANTAMARIA,LEONARDO
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:24

SUBJECT'S LAST NAME: MIGLIORINO

DATE/TIME OF TEST: 05/03/2025 07:52:35

Diagnostics Check		OK	07:45
Air Blank		0.000	07:46
Control Test		0.079	07:46
Air Blank		0.000	07:47
Subject Sample	# 1	0.207	07:47
Breath Volume		2.011	07:47
Air Blank		0.000	07:48
Air Blank		0.000	07:50
Subject Sample	# 2	0.198	07:50
Breath Volume		1.789	07:50
Air Blank		0.000	07:51
Control Test		0.080	07:51
Air Blank		0.000	07:52
Diagnostics Check		OK	07:52

Data Download Date/Time: 05/15/2025 20:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 05/04/2025 01:21:33

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:10	
Control Test		0.081	01:10	
Air Blank		0.000	01:11	
Subject Sample	# 1		01:14	No Sample Provided
Breath Volume		0.000	01:14	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 2		01:20	No Sample Provided
Breath Volume		0.000	01:20	
Air Blank		0.000	01:20	
Control Test		0.081	01:20	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 05/15/2025 20:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 05/09/2025 17:56:56

Diagnostics Check		OK	17:50
Air Blank		0.000	17:50
Control Test		0.079	17:51
Air Blank		0.000	17:51
Subject Sample	# 1	0.162	17:52
Breath Volume		3.148	17:52
Air Blank		0.000	17:53
Air Blank		0.000	17:54
Subject Sample	# 2	0.154	17:55
Breath Volume		2.503	17:55
Air Blank		0.000	17:56
Control Test		0.079	17:56
Air Blank		0.000	17:56
Diagnostics Check		OK	17:56

Data Download Date/Time: 05/15/2025 20:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: NIZ BARTOLON

DATE/TIME OF TEST: 05/10/2025 00:51:09

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.080	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.252	00:46
Breath Volume		1.765	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:48
Subject Sample	# 2	0.256	00:49
Breath Volume		1.343	00:49
Air Blank		0.000	00:50
Control Test		0.077	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 05/15/2025 20:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,DILLAN,R
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MCLEMORE,GARRETT
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 05/10/2025 21:00:12

Diagnostics Check		OK	20:51	
Air Blank		0.000	20:51	
Control Test		0.080	20:51	
Air Blank		0.000	20:52	
Subject Sample	# 1		20:55	No Sample Provided
Breath Volume		0.000	20:55	
Air Blank		0.000	20:56	
Air Blank		0.000	20:58	
Subject Sample	# 2		20:58	Subject Test Refused
Breath Volume		0.000	20:58	
Air Blank		0.000	20:59	
Control Test		0.079	20:59	
Air Blank		0.000	21:00	
Diagnostics Check		OK	21:00	

Data Download Date/Time: 05/15/2025 20:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: PETTIT

DATE/TIME OF TEST: 05/10/2025 22:34:26

Diagnostics Check		OK	22:31	
Air Blank		0.000	22:31	
Control Test		0.080	22:32	
Air Blank		0.000	22:32	
Subject Sample	# 1		22:33	Subject Test Refused
Breath Volume		0.000	22:33	
Air Blank		0.000	22:33	
Control Test		0.080	22:33	
Air Blank		0.000	22:34	
Diagnostics Check		OK	22:34	

Data Download Date/Time: 05/15/2025 20:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/11/2025 09:13:30

Diagnostics Check		OK	09:10	
Air Blank		0.000	09:11	
Control Test		0.080	09:11	
Air Blank		0.000	09:11	
Subject Sample	# 1		09:12	Subject Test Refused
Breath Volume		0.000	09:12	
Air Blank		0.000	09:12	
Control Test		0.079	09:12	
Air Blank		0.000	09:13	
Diagnostics Check		OK	09:13	

Data Download Date/Time: 05/15/2025 20:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUFFNER,PRESTON,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: KUFFNER,PRESTON
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: MARTUCCI

DATE/TIME OF TEST: 05/13/2025 23:05:48

Diagnostics Check		OK	22:58
Air Blank		0.000	22:59
Control Test		0.080	22:59
Air Blank		0.000	23:00
Subject Sample	# 1	0.198	23:00
Breath Volume		1.886	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.181	23:04
Breath Volume		2.800	23:04
Air Blank		0.000	23:04
Control Test		0.079	23:05
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 05/15/2025 20:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: DAYMON

DATE/TIME OF TEST: 05/14/2025 21:11:37

Diagnostics Check		OK	21:08	
Air Blank		0.000	21:09	
Control Test		0.080	21:09	
Air Blank		0.000	21:10	
Subject Sample	# 1		21:10	Subject Test Refused
Breath Volume		0.000	21:10	
Air Blank		0.000	21:10	
Control Test		0.079	21:11	
Air Blank		0.000	21:11	
Diagnostics Check		OK	21:11	

Data Download Date/Time: 05/15/2025 20:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATINI,BRYAN,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MARTINI,BRYAN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: THIBEAULT

DATE/TIME OF TEST: 05/14/2025 22:59:03

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.080	22:50
Air Blank		0.000	22:51
Subject Sample	# 1	0.179	22:54
Breath Volume		1.558	22:54
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.188	22:57
Breath Volume		1.792	22:57
Air Blank		0.000	22:58
Control Test		0.078	22:58
Air Blank		0.000	22:58
Diagnostics Check		OK	22:59

Data Download Date/Time: 05/15/2025 20:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: GONZALEZ CABALLERO

DATE/TIME OF TEST: 05/18/2025 03:07:17

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:56	
Control Test		0.080	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 1	0.133	03:00	Volume Not Met
Breath Volume		0.468	03:00	
Air Blank		0.000	03:00	
Air Blank		0.000	03:02	
Subject Sample	# 2		03:05	No Sample Provided
Breath Volume		0.000	03:05	
Air Blank		0.000	03:06	
Control Test		0.079	03:06	
Air Blank		0.000	03:07	
Diagnostics Check		OK	03:07	

Data Download Date/Time: 06/22/2025 18:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: BEGOVICH

DATE/TIME OF TEST: 05/23/2025 16:41:58

Diagnostics Check		OK	16:38	
Air Blank		0.000	16:39	
Control Test		0.080	16:39	
Air Blank		0.000	16:39	
Subject Sample	# 1		16:40	Slope Not Met
Breath Volume		1.406	16:40	
Air Blank		0.000	16:41	
Control Test		0.080	16:41	
Air Blank		0.000	16:41	
Diagnostics Check		OK	16:41	

Data Download Date/Time: 06/22/2025 18:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: BEGOVICH

DATE/TIME OF TEST: 05/23/2025 16:50:59

Diagnostics Check		OK	16:43
Air Blank		0.000	16:44
Control Test		0.080	16:44
Air Blank		0.000	16:45
Subject Sample	# 1	0.173	16:46
Breath Volume		2.386	16:46
Air Blank		0.000	16:46
Air Blank		0.000	16:48
Subject Sample	# 2	0.171	16:49
Breath Volume		1.792	16:49
Air Blank		0.000	16:50
Control Test		0.080	16:50
Air Blank		0.000	16:50
Diagnostics Check		OK	16:50

Data Download Date/Time: 06/22/2025 18:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOUSSAINT,WESNES,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TOUSSAINT,WESNES
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: SANTOS PEDRO

DATE/TIME OF TEST: 05/24/2025 17:02:16

Diagnostics Check		OK	16:54
Air Blank		0.000	16:55
Control Test		0.080	16:55
Air Blank		0.000	16:55
Subject Sample	# 1	0.145	16:57
Breath Volume		4.414	16:57
Air Blank		0.000	16:58
Air Blank		0.000	17:00
Subject Sample	# 2	0.146	17:00
Breath Volume		4.242	17:00
Air Blank		0.000	17:01
Control Test		0.080	17:01
Air Blank		0.000	17:02
Diagnostics Check		OK	17:02

Data Download Date/Time: 06/22/2025 18:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 05/25/2025 03:59:28

Diagnostics Check		OK	03:52
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.137	03:54
Breath Volume		3.292	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.129	03:57
Breath Volume		2.613	03:57
Air Blank		0.000	03:58
Control Test		0.080	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 06/22/2025 18:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: SAYLES

DATE/TIME OF TEST: 05/25/2025 20:53:36

Diagnostics Check		OK	20:46
Air Blank		0.000	20:46
Control Test		0.080	20:47
Air Blank		0.000	20:47
Subject Sample	# 1	0.187	20:48
Breath Volume		1.222	20:48
Air Blank		0.000	20:48
Air Blank		0.000	20:50
Subject Sample	# 2	0.194	20:51
Breath Volume		1.699	20:51
Air Blank		0.000	20:52
Control Test		0.079	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 06/22/2025 18:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOURNE,RUSSELL,L
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: BURKE,LOGAN
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ANAYA BANUELOS

DATE/TIME OF TEST: 05/26/2025 01:41:02

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.007	01:36
Breath Volume		3.898	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.000	01:39
Breath Volume		3.859	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:41

Data Download Date/Time: 06/22/2025 18:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: PASCUAL BERNABE

DATE/TIME OF TEST: 05/26/2025 20:50:14

Diagnostics Check		OK	20:43
Air Blank		0.000	20:44
Control Test		0.080	20:44
Air Blank		0.000	20:45
Subject Sample	# 1	0.247	20:45
Breath Volume		3.246	20:45
Air Blank		0.000	20:46
Air Blank		0.000	20:48
Subject Sample	# 2	0.252	20:48
Breath Volume		1.652	20:48
Air Blank		0.000	20:49
Control Test		0.079	20:49
Air Blank		0.000	20:50
Diagnostics Check		OK	20:50

Data Download Date/Time: 06/22/2025 18:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTHOLOMEW,MARK,A
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BARTHOLOMEW,MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 05/29/2025 20:59:12

Diagnostics Check		OK	20:52
Air Blank		0.000	20:53
Control Test		0.078	20:53
Air Blank		0.000	20:54
Subject Sample	# 1	0.170	20:54
Breath Volume		2.882	20:54
Air Blank		0.000	20:55
Air Blank		0.000	20:57
Subject Sample	# 2	0.151	20:57
Breath Volume		1.851	20:57
Air Blank		0.000	20:58
Control Test		0.077	20:58
Air Blank		0.000	20:59
Diagnostics Check		OK	20:59

Data Download Date/Time: 06/22/2025 18:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: VICENTE PEREZ

DATE/TIME OF TEST: 05/31/2025 16:13:10

Diagnostics Check		OK	16:09	
Air Blank		0.000	16:10	
Control Test		0.081	16:10	
Air Blank		0.000	16:11	
Subject Sample	# 1		16:11	Slope Not Met
Breath Volume		1.265	16:11	
Air Blank		0.000	16:12	
Control Test		0.080	16:12	
Air Blank		0.000	16:13	
Diagnostics Check		OK	16:13	

Data Download Date/Time: 06/22/2025 18:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: VICENTE PEREZ

DATE/TIME OF TEST: 05/31/2025 16:20:24

Diagnostics Check		OK	16:17	
Air Blank		0.000	16:17	
Control Test		0.080	16:17	
Air Blank		0.000	16:18	
Subject Sample	# 1		16:19	Subject Test Refused
Breath Volume		0.000	16:19	
Air Blank		0.000	16:19	
Control Test		0.081	16:19	
Air Blank		0.000	16:20	
Diagnostics Check		OK	16:20	

Data Download Date/Time: 06/22/2025 18:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.