Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RODRIGUEZ,E
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 05/06/2025 02:23:46

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.079	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.198	02:19
Breath Volume		1.992	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.190	02:22
Breath Volume		1.855	02:22
Air Blank		0.000	02:22
Control Test		0.081	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 05/08/2025 00:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ADAMS,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/06/2025 03:22:55

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.131	03:18
Breath Volume		1.136	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.131	03:21
Breath Volume		1.894	03:21
Air Blank		0.000	03:22
Control Test		0.081	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 05/08/2025 00:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CASTRO,A
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: JARRETT

DATE/TIME OF TEST: 05/06/2025 04:04:25

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.081	03:58
Air Blank		0.000	03:58
Subject Sample	# 1	0.141	03:59
Breath Volume		3.218	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.122	04:02
Breath Volume		2.019	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 05/08/2025 00:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DEVERNOE,M
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 05/07/2025 03:27:46

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.125	03:23
Breath Volume		3.031	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.127	03:26
Breath Volume		2.789	03:26
Air Blank		0.000	03:26
Control Test		0.080	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 05/08/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ADAMS,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: MANJI

DATE/TIME OF TEST: 05/07/2025 04:28:56

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:21	
Control Test		0.080	04:21	
Air Blank		0.000	04:21	
Subject Sample	# 1		04:24	No Sample Provided
Breath Volume		0.000	04:24	
Air Blank		0.000	04:25	
Air Blank		0.000	04:27	
Subject Sample	# 2		04:27	Subject Test Refused
Breath Volume		0.000	04:27	
Air Blank		0.000	04:28	
Control Test		0.081	04:28	
Air Blank		0.000	04:28	
Diagnostics Check		OK	04:28	

Data Download Date/Time: 05/08/2025 00:15

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FARRELL,S
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 05/07/2025 05:20:39

Diagnostics Check		OK	05:13
Air Blank		0.000	05:14
Control Test		0.081	05:14
Air Blank		0.000	05:14
Subject Sample	# 1	0.064	05:16
Breath Volume		1.570	05:16
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.067	05:19
Breath Volume		1.542	05:19
Air Blank		0.000	05:19
Control Test		0.081	05:20
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 05/08/2025 00:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOWERS,SHEHARA,D
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: MARKS,JAMES
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MCDANEL

DATE/TIME OF TEST: 05/09/2025 05:10:47

Diagnostics Check		OK	05:04
Air Blank		0.000	05:05
Control Test		0.081	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.208	05:06
Breath Volume		2.773	05:06
Air Blank		0.000	05:07
Air Blank		0.000	05:08
Subject Sample	# 2	0.206	05:09
Breath Volume		1.933	05:09
Air Blank		0.000	05:09
Control Test		0.080	05:10
Air Blank		0.000	05:10
Diagnostics Check		OK	05:10

Data Download Date/Time: 05/22/2025 20:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOREHEAD,AVERY,F
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: LYNCH,DANIEL
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: MONTOUTE

DATE/TIME OF TEST: 05/15/2025 21:06:22

Diagnostics Check		OK	20:53	
Air Blank		0.000	20:53	
Control Test		0.081	20:54	
Air Blank		0.000	20:54	
Subject Sample	# 1	0.115	20:57	Volume Not Met
Breath Volume		0.386	20:57	
Air Blank		0.000	20:58	
Air Blank		0.000	21:00	
Subject Sample	# 2	0.136	21:01	
Breath Volume		1.812	21:01	
Air Blank		0.000	21:02	
Air Blank		0.000	21:03	
Subject Sample	# 3	0.130	21:04	
Breath Volume		2.437	21:04	
Air Blank		0.000	21:05	
Control Test		0.079	21:05	
Air Blank		0.000	21:06	
Diagnostics Check		OK	21:06	

Data Download Date/Time: 05/22/2025 20:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,JAMES,B
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: MCNAMARA,KYLE
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: GERONIMO-CARDONA

DATE/TIME OF TEST: 05/18/2025 00:04:57

Diagnostics Check		OK	00:02	
Air Blank		0.000	00:02	
Control Test		0.082	00:02	
Air Blank		0.000	00:03	
Subject Sample	# 1		00:03	Subject Test Refused
Breath Volume		0.000	00:03	
Air Blank		0.000	00:04	
Control Test		0.081	00:04	
Air Blank		0.000	00:04	
Diagnostics Check		OK	00:04	

Data Download Date/Time: 05/22/2025 20:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: PARDO,H
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: BRANNON

DATE/TIME OF TEST: 05/30/2025 22:02:42

Diagnostics Check		OK	21:59	
Air Blank		0.000	22:00	
Control Test		0.081	22:00	
Air Blank		0.000	22:01	
Subject Sample	# 1		22:01	Subject Test Refused
Breath Volume		0.000	22:01	
Air Blank		0.000	22:01	
Control Test		0.081	22:02	
Air Blank		0.000	22:02	
Diagnostics Check		OK	22:02	

Data Download Date/Time: 06/03/2025 20:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: ABRAHAMSEN,ERIK
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 04/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: CONDEE

DATE/TIME OF TEST: 05/03/2025 04:16:07

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.080	04:04	
Air Blank		0.000	04:04	
Subject Sample	# 1		04:07	No Sample Provided
Breath Volume		0.000	04:07	
Air Blank		0.000	04:08	
Air Blank		0.000	04:10	
Subject Sample	# 2	0.160	04:11	
Breath Volume		2.082	04:11	
Air Blank		0.000	04:11	
Air Blank		0.000	04:13	
Subject Sample	# 3	0.173	04:14	
Breath Volume		2.761	04:14	
Air Blank		0.000	04:15	
Control Test		0.080	04:15	
Air Blank		0.000	04:16	
Diagnostics Check		OK	04:16	

Data Download Date/Time: 05/23/2025 18:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: SHERWOOD,JOSEPH
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 04/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: DUCE

DATE/TIME OF TEST: 05/10/2025 02:01:00

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.145	01:56
Breath Volume		1.945	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:58
Subject Sample	# 2	0.134	01:59
Breath Volume		1.601	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:00
Diagnostics Check		OK	02:01

Data Download Date/Time: 05/23/2025 18:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCCRANN,SEAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 05/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: SZYMANSKI

DATE/TIME OF TEST: 05/31/2025 05:07:25

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.081	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.155	05:02
Breath Volume		1.328	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:04
Subject Sample	# 2	0.160	05:05
Breath Volume		1.207	05:05
Air Blank		0.000	05:06
Control Test		0.079	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 06/23/2025 03:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/01/2025 04:32:13

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:29	
Control Test		0.081	04:30	
Air Blank		0.000	04:30	
Subject Sample	# 1		04:30	Subject Test Refused
Breath Volume		0.000	04:30	
Air Blank		0.000	04:31	
Control Test		0.082	04:31	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHIRIO,V
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 05/02/2025 23:15:17

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:11	
Control Test		0.082	23:11	
Air Blank		0.000	23:12	
Subject Sample	# 1		23:13	Subject Test Refused
Breath Volume		0.000	23:13	
Air Blank		0.000	23:14	
Control Test		0.082	23:14	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTINO,ERIK,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CUTINO,ERIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 05/03/2025 03:42:33

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.081	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.157	03:38
Breath Volume		2.460	03:38
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.154	03:41
Breath Volume		1.539	03:41
Air Blank		0.000	03:41
Control Test		0.082	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: PANDO

DATE/TIME OF TEST: 05/03/2025 04:02:48

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:59	
Control Test		0.082	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 1		04:01	Subject Test Refused
Breath Volume		0.000	04:01	
Air Blank		0.000	04:01	
Control Test		0.081	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: RAMOS RESTO

DATE/TIME OF TEST: 05/03/2025 23:49:22

Diagnostics Check		OK	23:46	
Air Blank		0.000	23:47	
Control Test		0.083	23:47	
Air Blank		0.000	23:47	
Subject Sample	# 1		23:48	Subject Test Refused
Breath Volume		0.000	23:48	
Air Blank		0.000	23:48	
Control Test		0.082	23:48	
Air Blank		0.000	23:49	
Diagnostics Check		OK	23:49	

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACHADO,LAURA,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MACHADO,LAURA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PATON

DATE/TIME OF TEST: 05/04/2025 01:03:05

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.082	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.092	00:57
Breath Volume		2.140	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.093	01:01
Breath Volume		2.339	01:01
Air Blank		0.000	01:02
Control Test		0.082	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:03

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZIMBA,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: ESCAMILLA

DATE/TIME OF TEST: 05/04/2025 01:32:28

Diagnostics Check		OK	01:29	
Air Blank		0.000	01:30	
Control Test		0.081	01:30	
Air Blank		0.000	01:30	
Subject Sample	# 1		01:31	Subject Test Refused
Breath Volume		0.000	01:31	
Air Blank		0.000	01:31	
Control Test		0.081	01:31	
Air Blank		0.000	01:32	
Diagnostics Check		OK	01:32	

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERNER,BLAKE
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: RICHBURG

DATE/TIME OF TEST: 05/04/2025 02:37:46

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.082	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.289	02:33
Breath Volume		1.160	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.290	02:36
Breath Volume		1.148	02:36
Air Blank		0.000	02:36
Control Test		0.080	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: ERAZO

DATE/TIME OF TEST: 05/06/2025 01:49:59

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.080	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.182	01:45
Breath Volume		2.789	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:47
Subject Sample	# 2	0.182	01:48
Breath Volume		2.824	01:48
Air Blank		0.000	01:49
Control Test		0.081	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 05/07/2025 07:29:12

Diagnostics Check		OK	07:26	
Air Blank		0.000	07:26	
Control Test		0.082	07:27	
Air Blank		0.000	07:27	
Subject Sample	# 1		07:27	Subject Test Refused
Breath Volume		0.000	07:27	
Air Blank		0.000	07:28	
Control Test		0.082	07:28	
Air Blank		0.000	07:29	
Diagnostics Check		OK	07:29	

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BJURHOLM,JACOB,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BJURHOLM,JACOB
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: ATKINSON

DATE/TIME OF TEST: 05/08/2025 02:02:15

Diagnostics Check		OK	01:59	
Air Blank		0.000	01:59	
Control Test		0.082	02:00	
Air Blank		0.000	02:00	
Subject Sample	# 1		02:00	Subject Test Refused
Breath Volume		0.000	02:00	
Air Blank		0.000	02:01	
Control Test		0.081	02:01	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORGAN,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 14:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: MCCAFFERTY

DATE/TIME OF TEST: 05/08/2025 15:19:04

Diagnostics Check		OK	15:13
Air Blank		0.000	15:13
Control Test		0.082	15:13
Air Blank		0.000	15:14
Subject Sample	# 1	0.000	15:14
Breath Volume		2.019	15:14
Air Blank		0.000	15:15
Air Blank		0.000	15:17
Subject Sample	# 2	0.000	15:17
Breath Volume		1.332	15:17
Air Blank		0.000	15:18
Control Test		0.082	15:18
Air Blank		0.000	15:18
Diagnostics Check		OK	15:19

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 05/08/2025 22:07:34

Diagnostics Check		OK	22:01
Air Blank		0.000	22:01
Control Test		0.082	22:02
Air Blank		0.000	22:02
Subject Sample	# 1	0.158	22:03
Breath Volume		1.390	22:03
Air Blank		0.000	22:03
Air Blank		0.000	22:05
Subject Sample	# 2	0.169	22:06
Breath Volume		1.828	22:06
Air Blank		0.000	22:06
Control Test		0.082	22:06
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: MCKELLAR

DATE/TIME OF TEST: 05/09/2025 03:25:46

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.082	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.000	03:21
Breath Volume		1.671	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.000	03:24
Breath Volume		1.156	03:24
Air Blank		0.000	03:24
Control Test		0.082	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: RINGLER

DATE/TIME OF TEST: 05/10/2025 22:16:09

Diagnostics Check		OK	22:13	
Air Blank		0.000	22:13	
Control Test		0.082	22:14	
Air Blank		0.000	22:14	
Subject Sample	# 1		22:14	Subject Test Refused
Breath Volume		0.000	22:14	
Air Blank		0.000	22:15	
Control Test		0.082	22:15	
Air Blank		0.000	22:16	
Diagnostics Check		OK	22:16	

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAMILTON,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: DESMOND

DATE/TIME OF TEST: 05/11/2025 01:27:06

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.083	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.183	01:22
Breath Volume		1.914	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.164	01:25
Breath Volume		1.421	01:25
Air Blank		0.000	01:26
Control Test		0.081	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLLOWELL,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 05/11/2025 03:25:57

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.082	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.130	03:21
Breath Volume		2.644	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:23
Subject Sample	# 2	0.126	03:24
Breath Volume		2.250	03:24
Air Blank		0.000	03:25
Control Test		0.082	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KING,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/11/2025 04:43:02

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:40	
Control Test		0.082	04:40	
Air Blank		0.000	04:41	
Subject Sample	# 1		04:41	Subject Test Refused
Breath Volume		0.000	04:41	
Air Blank		0.000	04:42	
Control Test		0.082	04:42	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:43	

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GOODPASTER,KALEB
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: LAKE

DATE/TIME OF TEST: 05/11/2025 21:25:50

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:23	
Control Test		0.082	21:23	
Air Blank		0.000	21:24	
Subject Sample	# 1		21:24	Subject Test Refused
Breath Volume		0.000	21:24	
Air Blank		0.000	21:24	
Control Test		0.082	21:25	
Air Blank		0.000	21:25	
Diagnostics Check		OK	21:25	

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PASKALAKIS,BRIANNA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/13/2025 05:15:26

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.082	05:09
Air Blank		0.000	05:10
Subject Sample	# 1	0.171	05:11
Breath Volume		2.066	05:11
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.184	05:13
Breath Volume		2.523	05:13
Air Blank		0.000	05:14
Control Test		0.083	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 05/21/2025 09:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SLOANE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: KARAHALIS

DATE/TIME OF TEST: 05/13/2025 19:15:44

Diagnostics Check		OK	19:09
Air Blank		0.000	19:09
Control Test		0.083	19:10
Air Blank		0.000	19:10
Subject Sample	# 1	0.197	19:11
Breath Volume		2.527	19:11
Air Blank		0.000	19:11
Air Blank		0.000	19:13
Subject Sample	# 2	0.194	19:14
Breath Volume		2.707	19:14
Air Blank		0.000	19:14
Control Test		0.082	19:15
Air Blank		0.000	19:15
Diagnostics Check		OK	19:15

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LAZZAR,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 05/15/2025 03:32:16

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:29	
Control Test		0.082	03:30	
Air Blank		0.000	03:30	
Subject Sample	# 1		03:30	Subject Test Refused
Breath Volume		0.000	03:30	
Air Blank		0.000	03:31	
Control Test		0.082	03:31	
Air Blank		0.000	03:32	
Diagnostics Check		OK	03:32	

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES,MARK,O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FERGUSON,ROBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 06:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:51

SUBJECT'S LAST NAME: CUSICK

DATE/TIME OF TEST: 05/15/2025 07:23:56

Diagnostics Check		OK	07:11	
Air Blank		0.000	07:11	
Control Test		0.082	07:12	
Air Blank		0.000	07:12	
Subject Sample	# 1	0.108	07:15	Volume Not Met
Breath Volume		0.179	07:15	
Air Blank		0.000	07:16	
Air Blank		0.000	07:18	
Subject Sample	# 2	0.092	07:18	
Breath Volume		1.695	07:18	
Air Blank		0.000	07:19	
Air Blank		0.000	07:21	
Subject Sample	# 3	0.100	07:22	
Breath Volume		1.769	07:22	
Air Blank		0.000	07:23	
Control Test		0.082	07:23	
Air Blank		0.000	07:23	
Diagnostics Check		OK	07:23	

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JUNE,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: BELLON

DATE/TIME OF TEST: 05/15/2025 20:40:22

Diagnostics Check		OK	20:37	
Air Blank		0.000	20:37	
Control Test		0.082	20:38	
Air Blank		0.000	20:38	
Subject Sample	# 1		20:39	Subject Test Refused
Breath Volume		0.000	20:39	
Air Blank		0.000	20:39	
Control Test		0.082	20:39	
Air Blank		0.000	20:40	
Diagnostics Check		OK	20:40	

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: AVERBECK

DATE/TIME OF TEST: 05/16/2025 04:13:34

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:11	
Control Test		0.082	04:11	
Air Blank		0.000	04:12	
Subject Sample	# 1		04:12	Subject Test Refused
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	
Control Test		0.082	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEWS,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: HURD

DATE/TIME OF TEST: 05/16/2025 23:27:50

Diagnostics Check		OK	23:24	
Air Blank		0.000	23:25	
Control Test		0.083	23:25	
Air Blank		0.000	23:26	
Subject Sample	# 1		23:26	Subject Test Refused
Breath Volume		0.000	23:26	
Air Blank		0.000	23:26	
Control Test		0.082	23:27	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MILLAN

DATE/TIME OF TEST: 05/17/2025 01:53:17

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.082	01:48
Air Blank		0.000	01:48
Subject Sample	# 1	0.114	01:48
Breath Volume		1.488	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.110	01:51
Breath Volume		1.390	01:51
Air Blank		0.000	01:52
Control Test		0.082	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SALZMAN,BLAKE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 14:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:04

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/17/2025 14:30:36

Diagnostics Check		OK	14:24
Air Blank		0.000	14:24
Control Test		0.082	14:25
Air Blank		0.000	14:25
Subject Sample	# 1	0.108	14:26
Breath Volume		1.464	14:26
Air Blank		0.000	14:26
Air Blank		0.000	14:28
Subject Sample	# 2	0.111	14:29
Breath Volume		2.113	14:29
Air Blank		0.000	14:29
Control Test		0.082	14:30
Air Blank		0.000	14:30
Diagnostics Check		OK	14:30

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GISCHEL,M
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BUELL

DATE/TIME OF TEST: 05/17/2025 20:54:41

Diagnostics Check		OK	20:48
Air Blank		0.000	20:48
Control Test		0.082	20:48
Air Blank		0.000	20:49
Subject Sample	# 1	0.085	20:50
Breath Volume		1.140	20:50
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.087	20:53
Breath Volume		1.605	20:53
Air Blank		0.000	20:53
Control Test		0.082	20:54
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KILIAN,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 05/17/2025 22:20:18

Diagnostics Check		OK	22:17	
Air Blank		0.000	22:17	
Control Test		0.082	22:18	
Air Blank		0.000	22:18	
Subject Sample	# 1		22:18	Subject Test Refused
Breath Volume		0.000	22:18	
Air Blank		0.000	22:19	
Control Test		0.082	22:19	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:20	

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSEN,DUSTIN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANSEN,DUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: BUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: VANDERBENT

DATE/TIME OF TEST: 05/17/2025 23:38:04

Diagnostics Check		OK	23:30
Air Blank		0.000	23:31
Control Test		0.082	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.084	23:33
Breath Volume		1.539	23:33
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.080	23:36
Breath Volume		1.285	23:36
Air Blank		0.000	23:37
Control Test		0.082	23:37
Air Blank		0.000	23:37
Diagnostics Check		OK	23:38

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEGRON,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: BELLOISE

DATE/TIME OF TEST: 05/18/2025 02:07:24

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.082	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.100	02:02
Breath Volume		1.136	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.099	02:05
Breath Volume		1.230	02:05
Air Blank		0.000	02:06
Control Test		0.081	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BRISENO,N
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 05/18/2025 02:47:09

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.082	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.160	02:42
Breath Volume		2.207	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:44
Subject Sample	# 2	0.166	02:45
Breath Volume		2.082	02:45
Air Blank		0.000	02:46
Control Test		0.081	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORRIS,BRIDGETT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 05/18/2025 05:07:41

Diagnostics Check		OK	05:01
Air Blank		0.000	05:01
Control Test		0.082	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.137	05:03
Breath Volume		2.136	05:03
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.135	05:06
Breath Volume		2.265	05:06
Air Blank		0.000	05:06
Control Test		0.082	05:07
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KILIAN,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: HIMES

DATE/TIME OF TEST: 05/18/2025 05:46:46

Diagnostics Check		OK	05:43	
Air Blank		0.000	05:44	
Control Test		0.082	05:44	
Air Blank		0.000	05:45	
Subject Sample	# 1		05:45	Subject Test Refused
Breath Volume		0.000	05:45	
Air Blank		0.000	05:45	
Control Test		0.082	05:46	
Air Blank		0.000	05:46	
Diagnostics Check		OK	05:46	

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FULLER,JOY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: NANGLE

DATE/TIME OF TEST: 05/18/2025 20:41:30

Diagnostics Check		OK	20:38	
Air Blank		0.000	20:39	
Control Test		0.082	20:39	
Air Blank		0.000	20:39	
Subject Sample	# 1		20:40	Subject Test Refused
Breath Volume		0.000	20:40	
Air Blank		0.000	20:40	
Control Test		0.082	20:40	
Air Blank		0.000	20:41	
Diagnostics Check		OK	20:41	

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSHANS,DEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MESSENGER,SEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/18/2025 22:39:44

Diagnostics Check		OK	22:32
Air Blank		0.000	22:32
Control Test		0.082	22:33
Air Blank		0.000	22:33
Subject Sample	# 1	0.139	22:35
Breath Volume		1.375	22:35
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.138	22:38
Breath Volume		1.507	22:38
Air Blank		0.000	22:38
Control Test		0.082	22:39
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NURY,KELSIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 05/20/2025 22:49:54

Diagnostics Check		OK	22:43
Air Blank		0.000	22:43
Control Test		0.082	22:44
Air Blank		0.000	22:44
Subject Sample	# 1	0.308	22:45
Breath Volume		2.687	22:45
Air Blank		0.000	22:46
Air Blank		0.000	22:47
Subject Sample	# 2	0.297	22:48
Breath Volume		2.234	22:48
Air Blank		0.000	22:48
Control Test		0.081	22:49
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 05/21/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: AYUSO-BURGOS

DATE/TIME OF TEST: 05/21/2025 22:50:26

Diagnostics Check		OK	22:44
Air Blank		0.000	22:44
Control Test		0.082	22:45
Air Blank		0.000	22:45
Subject Sample	# 1	0.000	22:46
Breath Volume		1.550	22:46
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.000	22:49
Breath Volume		1.265	22:49
Air Blank		0.000	22:49
Control Test		0.082	22:49
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 06/18/2025 10:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PASKALAKIS,BRIANNA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: MROZEK

DATE/TIME OF TEST: 05/22/2025 00:34:37

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:32	
Control Test		0.082	00:32	
Air Blank		0.000	00:32	
Subject Sample	# 1		00:33	Subject Test Refused
Breath Volume		0.000	00:33	
Air Blank		0.000	00:33	
Control Test		0.082	00:34	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MESSENGER,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: KRAMER

DATE/TIME OF TEST: 05/23/2025 01:19:03

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:16	
Control Test		0.083	01:16	
Air Blank		0.000	01:17	
Subject Sample	# 1		01:17	Subject Test Refused
Breath Volume		0.000	01:17	
Air Blank		0.000	01:18	
Control Test		0.083	01:18	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 05/23/2025 04:15:59

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.083	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.152	04:11
Breath Volume		1.992	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.154	04:14
Breath Volume		1.570	04:14
Air Blank		0.000	04:15
Control Test		0.082	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSEN,DUSTIN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANSEN,DUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: BUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 05/24/2025 18:11:57

Diagnostics Check		OK	18:09	
Air Blank		0.000	18:09	
Control Test		0.082	18:09	
Air Blank		0.000	18:10	
Subject Sample	# 1		18:10	Subject Test Refused
Breath Volume		0.000	18:10	
Air Blank		0.000	18:11	
Control Test		0.083	18:11	
Air Blank		0.000	18:11	
Diagnostics Check		OK	18:11	

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUZMAN,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BLISS

DATE/TIME OF TEST: 05/24/2025 21:58:27

Diagnostics Check		OK	21:50
Air Blank		0.000	21:50
Control Test		0.083	21:51
Air Blank		0.000	21:51
Subject Sample	# 1	0.215	21:53
Breath Volume		2.414	21:53
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.227	21:56
Breath Volume		2.136	21:56
Air Blank		0.000	21:57
Control Test		0.081	21:57
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARRITT,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: PAWLOWSKI

DATE/TIME OF TEST: 05/24/2025 22:45:43

Diagnostics Check		OK	22:38
Air Blank		0.000	22:38
Control Test		0.082	22:39
Air Blank		0.000	22:39
Subject Sample	# 1	0.136	22:40
Breath Volume		1.289	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.146	22:44
Breath Volume		1.750	22:44
Air Blank		0.000	22:44
Control Test		0.083	22:45
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: PURVIS

DATE/TIME OF TEST: 05/25/2025 03:09:06

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:04	
Control Test		0.083	03:05	
Air Blank		0.000	03:05	
Subject Sample	# 1		03:07	Subject Test Refused
Breath Volume		0.000	03:07	
Air Blank		0.000	03:08	
Control Test		0.083	03:08	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKALKO,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/25/2025 03:32:01

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.083	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.093	03:27
Breath Volume		1.488	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.092	03:30
Breath Volume		1.429	03:30
Air Blank		0.000	03:31
Control Test		0.082	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:32

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,JOHN,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: PHILPOTT

DATE/TIME OF TEST: 05/25/2025 06:20:54

Diagnostics Check		OK	06:13
Air Blank		0.000	06:14
Control Test		0.083	06:14
Air Blank		0.000	06:15
Subject Sample	# 1	0.077	06:16
Breath Volume		1.363	06:16
Air Blank		0.000	06:17
Air Blank		0.000	06:18
Subject Sample	# 2	0.076	06:19
Breath Volume		1.531	06:19
Air Blank		0.000	06:20
Control Test		0.083	06:20
Air Blank		0.000	06:20
Diagnostics Check		OK	06:20

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: CURRAN

DATE/TIME OF TEST: 05/26/2025 22:07:36

Diagnostics Check		OK	22:04	
Air Blank		0.000	22:05	
Control Test		0.083	22:05	
Air Blank		0.000	22:06	
Subject Sample	# 1		22:06	Subject Test Refused
Breath Volume		0.000	22:06	
Air Blank		0.000	22:06	
Control Test		0.084	22:07	
Air Blank		0.000	22:07	
Diagnostics Check		OK	22:07	

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZIMBA,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 05/27/2025 01:18:58

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.083	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.192	01:12
Breath Volume		1.285	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.204	01:17
Breath Volume		2.167	01:17
Air Blank		0.000	01:18
Control Test		0.083	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: HERN

DATE/TIME OF TEST: 05/27/2025 23:59:15

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.084	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.000	23:54
Breath Volume		3.222	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.000	23:57
Breath Volume		3.394	23:57
Air Blank		0.000	23:58
Control Test		0.084	23:58
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KIMBERLY,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: REDMAN

DATE/TIME OF TEST: 05/28/2025 19:14:04

Diagnostics Check		OK	19:10	
Air Blank		0.000	19:11	
Control Test		0.082	19:11	
Air Blank		0.000	19:12	
Subject Sample	# 1		19:12	Interferent Detect
Breath Volume		2.433	19:12	
Air Blank		0.000	19:13	
Control Test		0.081	19:13	
Air Blank		0.000	19:13	
Diagnostics Check		OK	19:14	

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 05/28/2025 21:57:58

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:54	
Control Test		0.083	21:54	
Air Blank		0.000	21:54	
Subject Sample	# 1		21:56	Subject Test Refused
Breath Volume		0.000	21:56	
Air Blank		0.000	21:57	
Control Test		0.082	21:57	
Air Blank		0.000	21:57	
Diagnostics Check		OK	21:57	

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JUNE,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: MITCHEM

DATE/TIME OF TEST: 05/30/2025 02:02:46

Diagnostics Check		OK	01:59	
Air Blank		0.000	02:00	
Control Test		0.082	02:00	
Air Blank		0.000	02:01	
Subject Sample	# 1		02:01	Subject Test Refused
Breath Volume		0.000	02:01	
Air Blank		0.000	02:01	
Control Test		0.081	02:02	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SPARKS

DATE/TIME OF TEST: 05/30/2025 03:02:46

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.082	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.000	02:58
Breath Volume		1.328	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.000	03:01
Breath Volume		1.257	03:01
Air Blank		0.000	03:01
Control Test		0.082	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: URF,CALEB
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 11:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:41

SUBJECT'S LAST NAME: HOOK

DATE/TIME OF TEST: 05/30/2025 13:07:58

Diagnostics Check		OK	13:02
Air Blank		0.000	13:02
Control Test		0.081	13:02
Air Blank		0.000	13:03
Subject Sample	# 1	0.164	13:03
Breath Volume		2.566	13:03
Air Blank		0.000	13:04
Air Blank		0.000	13:06
Subject Sample	# 2	0.166	13:06
Breath Volume		1.855	13:06
Air Blank		0.000	13:07
Control Test		0.082	13:07
Air Blank		0.000	13:07
Diagnostics Check		OK	13:07

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHASE,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: KILIAN

DATE/TIME OF TEST: 05/30/2025 20:31:16

Diagnostics Check		OK	20:28	
Air Blank		0.000	20:28	
Control Test		0.082	20:29	
Air Blank		0.000	20:29	
Subject Sample	# 1		20:29	Subject Test Refused
Breath Volume		0.000	20:29	
Air Blank		0.000	20:30	
Control Test		0.082	20:30	
Air Blank		0.000	20:31	
Diagnostics Check		OK	20:31	

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHASE,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: KILIAN

DATE/TIME OF TEST: 05/30/2025 20:49:54

Diagnostics Check		OK	20:42
Air Blank		0.000	20:43
Control Test		0.082	20:43
Air Blank		0.000	20:43
Subject Sample	# 1	0.114	20:45
Breath Volume		1.117	20:45
Air Blank		0.000	20:45
Air Blank		0.000	20:47
Subject Sample	# 2	0.111	20:48
Breath Volume		1.210	20:48
Air Blank		0.000	20:48
Control Test		0.082	20:49
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PASKALAKIS,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 05/31/2025 03:47:10

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.082	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.153	03:42
Breath Volume		2.195	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:44
Subject Sample	# 2	0.158	03:45
Breath Volume		2.125	03:45
Air Blank		0.000	03:46
Control Test		0.082	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LAMBIRD,DAVID
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: TREMBLAY

DATE/TIME OF TEST: 05/31/2025 17:21:04

Diagnostics Check		OK	17:15
Air Blank		0.000	17:15
Control Test		0.082	17:15
Air Blank		0.000	17:16
Subject Sample	# 1	0.187	17:16
Breath Volume		2.773	17:16
Air Blank		0.000	17:17
Air Blank		0.000	17:19
Subject Sample	# 2	0.176	17:19
Breath Volume		1.753	17:19
Air Blank		0.000	17:20
Control Test		0.081	17:20
Air Blank		0.000	17:20
Diagnostics Check		OK	17:21

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KNOPICK,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/31/2025 22:54:43

Diagnostics Check		OK	22:51	
Air Blank		0.000	22:52	
Control Test		0.083	22:52	
Air Blank		0.000	22:52	
Subject Sample	# 1		22:53	Subject Test Refused
Breath Volume		0.000	22:53	
Air Blank		0.000	22:53	
Control Test		0.083	22:54	
Air Blank		0.000	22:54	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 06/18/2025 10:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 05/03/2025 03:52:37

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.165	03:48
Breath Volume		1.546	03:48
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.175	03:51
Breath Volume		1.652	03:51
Air Blank		0.000	03:51
Control Test		0.079	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEWTON,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: DENSON

DATE/TIME OF TEST: 05/03/2025 04:12:03

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.183	04:07
Breath Volume		2.585	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:09
Subject Sample	# 2	0.178	04:10
Breath Volume		1.683	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:12

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LOCKE,REBECCA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: TODD

DATE/TIME OF TEST: 05/04/2025 00:03:04

Diagnostics Check		OK	23:55
Air Blank		0.000	23:56
Control Test		0.079	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.155	23:57
Breath Volume		1.414	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.152	00:01
Breath Volume		1.671	00:01
Air Blank		0.000	00:02
Control Test		0.079	00:02
Air Blank		0.000	00:02
Diagnostics Check		OK	00:03

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CZENSZAK,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: LACINA

DATE/TIME OF TEST: 05/04/2025 14:10:22

Diagnostics Check	OK	14:08	
Air Blank	0.000	14:09	
Control Test	0.079	14:09	
Air Blank		14:09	RFI Detect
Air Blank	0.000	14:10	

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CZENSZAK,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: LACINA

DATE/TIME OF TEST: 05/04/2025 14:19:04

Diagnostics Check		OK	14:13
Air Blank		0.000	14:13
Control Test		0.079	14:13
Air Blank		0.000	14:14
Subject Sample	# 1	0.366	14:14
Breath Volume		3.164	14:14
Air Blank		0.000	14:15
Air Blank		0.000	14:17
Subject Sample	# 2	0.353	14:17
Breath Volume		2.953	14:17
Air Blank		0.000	14:18
Control Test		0.077	14:18
Air Blank		0.000	14:18
Diagnostics Check		OK	14:19

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 05/04/2025 19:10:59

Diagnostics Check		OK	19:04
Air Blank		0.000	19:05
Control Test		0.080	19:05
Air Blank		0.000	19:06
Subject Sample	# 1	0.000	19:06
Breath Volume		1.679	19:06
Air Blank		0.000	19:06
Air Blank		0.000	19:08
Subject Sample	# 2	0.000	19:09
Breath Volume		1.542	19:09
Air Blank		0.000	19:10
Control Test		0.079	19:10
Air Blank		0.000	19:10
Diagnostics Check		OK	19:10

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SWIER,HAYDEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: MCLEOD

DATE/TIME OF TEST: 05/07/2025 09:36:48

Diagnostics Check		OK	09:30
Air Blank		0.000	09:31
Control Test		0.079	09:31
Air Blank		0.000	09:32
Subject Sample	# 1	0.000	09:32
Breath Volume		1.468	09:32
Air Blank		0.000	09:33
Air Blank		0.000	09:35
Subject Sample	# 2	0.000	09:35
Breath Volume		1.898	09:35
Air Blank		0.000	09:35
Control Test		0.080	09:36
Air Blank		0.000	09:36
Diagnostics Check		OK	09:36

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURDUE,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PURDUE,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 05/07/2025 19:07:42

Diagnostics Check		OK	19:01
Air Blank		0.000	19:01
Control Test		0.080	19:01
Air Blank		0.000	19:02
Subject Sample	# 1	0.211	19:02
Breath Volume		1.667	19:02
Air Blank		0.000	19:03
Air Blank		0.000	19:05
Subject Sample	# 2	0.201	19:06
Breath Volume		1.660	19:06
Air Blank		0.000	19:06
Control Test		0.080	19:07
Air Blank		0.000	19:07
Diagnostics Check		OK	19:07

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROWBACK,ASHLEY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: BONNER

DATE/TIME OF TEST: 05/09/2025 13:22:03

Diagnostics Check		OK	13:15
Air Blank		0.000	13:15
Control Test		0.080	13:15
Air Blank		0.000	13:16
Subject Sample	# 1	0.241	13:17
Breath Volume		2.867	13:17
Air Blank		0.000	13:18
Air Blank		0.000	13:19
Subject Sample	# 2	0.226	13:20
Breath Volume		3.449	13:20
Air Blank		0.000	13:21
Control Test		0.080	13:21
Air Blank		0.000	13:21
Diagnostics Check		OK	13:22

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MEARDY

DATE/TIME OF TEST: 05/09/2025 18:23:20

Diagnostics Check		OK	18:20	
Air Blank		0.000	18:20	
Control Test		0.080	18:21	
Air Blank		0.000	18:21	
Subject Sample	# 1		18:21	Subject Test Refused
Breath Volume		0.000	18:21	
Air Blank		0.000	18:22	
Control Test		0.080	18:22	
Air Blank		0.000	18:23	
Diagnostics Check		OK	18:23	

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/10/2025 01:15:05

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:03	
Subject Sample	# 1	0.147	01:04	
Breath Volume		1.472	01:04	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 2		01:10	No Sample Provided
Breath Volume		0.000	01:10	
Air Blank		0.000	01:10	
Air Blank		0.000	01:12	
Subject Sample	# 3		01:13	Subject Test Refused
Breath Volume		0.000	01:13	
Air Blank		0.000	01:14	
Control Test		0.080	01:14	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: VANDENBERG,MADISON
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 09:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: DEANGELO

DATE/TIME OF TEST: 05/13/2025 11:42:48

Diagnostics Check		OK	11:36
Air Blank		0.000	11:37
Control Test		0.080	11:37
Air Blank		0.000	11:38
Subject Sample	# 1	0.000	11:38
Breath Volume		2.265	11:38
Air Blank		0.000	11:39
Air Blank		0.000	11:41
Subject Sample	# 2	0.000	11:41
Breath Volume		2.257	11:41
Air Blank		0.000	11:41
Control Test		0.080	11:42
Air Blank		0.000	11:42
Diagnostics Check		OK	11:42

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORING,BLAINE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LORING,BLAINE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 09:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:17

SUBJECT'S LAST NAME: WEXLER

DATE/TIME OF TEST: 05/15/2025 10:45:14

Diagnostics Check		OK	10:39
Air Blank		0.000	10:39
Control Test		0.079	10:40
Air Blank		0.000	10:40
Subject Sample	# 1	0.000	10:41
Breath Volume		1.664	10:41
Air Blank		0.000	10:41
Air Blank		0.000	10:43
Subject Sample	# 2	0.000	10:43
Breath Volume		1.285	10:43
Air Blank		0.000	10:44
Control Test		0.079	10:44
Air Blank		0.000	10:45
Diagnostics Check		OK	10:45

Data Download Date/Time: 05/21/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SPICER,NICHOLAS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 05/15/2025 19:58:37

Diagnostics Check		OK	19:46	
Air Blank		0.000	19:46	
Control Test		0.080	19:47	
Air Blank		0.000	19:47	
Subject Sample	# 1	0.095	19:50	Volume Not Met
Breath Volume		0.589	19:50	
Air Blank		0.000	19:51	
Air Blank		0.000	19:53	
Subject Sample	# 2	0.103	19:53	
Breath Volume		1.289	19:53	
Air Blank		0.000	19:54	
Air Blank		0.000	19:56	
Subject Sample	# 3	0.097	19:57	
Breath Volume		1.152	19:57	
Air Blank		0.000	19:57	
Control Test		0.078	19:58	
Air Blank		0.000	19:58	
Diagnostics Check		OK	19:58	

Data Download Date/Time: 05/21/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LOCKE,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/17/2025 02:09:57

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:07	
Control Test		0.080	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Control Test		0.079	02:09	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 05/21/2025 10:01

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/17/2025 03:39:49

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.080	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.099	03:35
Breath Volume		1.324	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.103	03:38
Breath Volume		1.472	03:38
Air Blank		0.000	03:38
Control Test		0.080	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 05/21/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLIPIN,A,
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: KLIPIN,A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GRADZKI

DATE/TIME OF TEST: 05/17/2025 04:47:06

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.079	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.182	04:42
Breath Volume		1.675	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:44
Subject Sample	# 2	0.184	04:45
Breath Volume		3.101	04:45
Air Blank		0.000	04:46
Control Test		0.079	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 05/21/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEIB,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: COLLING

DATE/TIME OF TEST: 05/18/2025 19:17:59

Diagnostics Check		OK	19:11
Air Blank		0.000	19:11
Control Test		0.080	19:12
Air Blank		0.000	19:12
Subject Sample	# 1	0.271	19:13
Breath Volume		1.386	19:13
Air Blank		0.000	19:13
Air Blank		0.000	19:15
Subject Sample	# 2	0.253	19:16
Breath Volume		1.210	19:16
Air Blank		0.000	19:17
Control Test		0.079	19:17
Air Blank		0.000	19:17
Diagnostics Check		OK	19:17

Data Download Date/Time: 05/21/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 05/19/2025 03:55:47

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:53	
Control Test		0.080	03:53	
Air Blank		0.000	03:54	
Subject Sample	# 1		03:54	Subject Test Refused
Breath Volume		0.000	03:54	
Air Blank		0.000	03:54	
Control Test		0.080	03:55	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 05/21/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARRIZA,DANIEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 05/20/2025 23:04:50

Diagnostics Check		OK	22:56	
Air Blank		0.000	22:56	
Control Test		0.080	22:57	
Air Blank		0.000	22:57	
Subject Sample	# 1		23:00	No Sample Provided
Breath Volume		0.000	23:00	
Air Blank		0.000	23:01	
Air Blank		0.000	23:03	
Subject Sample	# 2		23:03	Subject Test Refused
Breath Volume		0.000	23:03	
Air Blank		0.000	23:03	
Control Test		0.080	23:04	
Air Blank		0.000	23:04	
Diagnostics Check		OK	23:04	

Data Download Date/Time: 05/21/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAINES,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: OLANG

DATE/TIME OF TEST: 05/21/2025 06:33:19

Diagnostics Check		OK	06:25
Air Blank		0.000	06:26
Control Test		0.080	06:26
Air Blank		0.000	06:27
Subject Sample	# 1	0.207	06:28
Breath Volume		1.332	06:28
Air Blank		0.000	06:28
Air Blank		0.000	06:30
Subject Sample	# 2	0.202	06:31
Breath Volume		1.152	06:31
Air Blank		0.000	06:32
Control Test		0.080	06:32
Air Blank		0.000	06:33
Diagnostics Check		OK	06:33

Data Download Date/Time: 05/21/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARQUEZ CARRILLO,JOSE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: REXROAD

DATE/TIME OF TEST: 05/21/2025 17:20:58

Diagnostics Check		OK	17:11
Air Blank		0.000	17:11
Control Test		0.079	17:12
Air Blank		0.000	17:12
Subject Sample	# 1	0.236	17:13
Breath Volume		2.937	17:13
Air Blank		0.000	17:13
Air Blank		0.000	17:15
Subject Sample	# 2	0.202	17:16
Breath Volume		1.871	17:16
Air Blank		0.000	17:17
Air Blank		0.000	17:18
Subject Sample	# 3	0.237	17:19
Breath Volume		2.886	17:19
Air Blank		0.000	17:20
Control Test		0.079	17:20
Air Blank		0.000	17:20
Diagnostics Check		OK	17:20

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CONSTANT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:44

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 05/23/2025 16:43:05

Diagnostics Check		OK	16:36
Air Blank		0.000	16:36
Control Test		0.079	16:37
Air Blank		0.000	16:37
Subject Sample	# 1	0.133	16:38
Breath Volume		1.718	16:38
Air Blank		0.000	16:38
Air Blank		0.000	16:40
Subject Sample	# 2	0.133	16:41
Breath Volume		1.792	16:41
Air Blank		0.000	16:42
Control Test		0.079	16:42
Air Blank		0.000	16:43
Diagnostics Check		OK	16:43

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEAL,LOUIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 05/25/2025 20:57:48

Diagnostics Check		OK	20:54	
Air Blank		0.000	20:55	
Control Test		0.079	20:55	
Air Blank		0.000	20:56	
Subject Sample	# 1		20:56	Subject Test Refused
Breath Volume		0.000	20:56	
Air Blank		0.000	20:56	
Control Test		0.080	20:57	
Air Blank		0.000	20:57	
Diagnostics Check		OK	20:57	

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: DELGADO TORRES

DATE/TIME OF TEST: 05/26/2025 06:29:00

Diagnostics Check		OK	06:24	
Air Blank		0.000	06:25	
Control Test		0.079	06:25	
Air Blank		0.000	06:25	
Subject Sample	# 1		06:27	RFI Detect
Breath Volume		0.085	06:27	
Air Blank		0.000	06:28	
Control Test		0.079	06:28	
Air Blank		0.000	06:28	
Diagnostics Check		OK	06:29	

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: DELGADO TORRES

DATE/TIME OF TEST: 05/26/2025 06:34:41

Diagnostics Check		OK	06:30	
Air Blank		0.000	06:31	
Control Test		0.079	06:31	
Air Blank		0.000	06:32	
Subject Sample	# 1		06:33	Subject Test Refused
Breath Volume		0.000	06:33	
Air Blank		0.000	06:33	
Control Test		0.079	06:34	
Air Blank		0.000	06:34	
Diagnostics Check		OK	06:34	

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEWTON III,THOMAS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: ALVES

DATE/TIME OF TEST: 05/26/2025 20:42:42

Diagnostics Check		OK	20:39	
Air Blank		0.000	20:40	
Control Test		0.079	20:40	
Air Blank		0.000	20:41	
Subject Sample	# 1		20:41	Subject Test Refused
Breath Volume		0.000	20:41	
Air Blank		0.000	20:41	
Control Test		0.079	20:42	
Air Blank		0.000	20:42	
Diagnostics Check		OK	20:42	

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZIMBA,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 05/27/2025 00:45:54

Diagnostics Check		OK	00:33
Air Blank		0.000	00:33
Control Test		0.080	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.231	00:37
Breath Volume		1.207	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.210	00:40
Breath Volume		1.656	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 3	0.217	00:44
Breath Volume		1.683	00:44
Air Blank		0.000	00:44
Control Test		0.079	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YANICK,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: BOPP

DATE/TIME OF TEST: 05/27/2025 01:45:50

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:43	
Control Test		0.079	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:44	
Control Test		0.079	01:45	
Air Blank		0.000	01:45	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACHADO,LAURA,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BANNISTER,STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CAUDELL

DATE/TIME OF TEST: 05/27/2025 22:47:25

Diagnostics Check		OK	22:40	
Air Blank		0.000	22:40	
Control Test		0.078	22:41	
Air Blank		0.000	22:41	
Subject Sample	# 1	0.061	22:42	
Breath Volume		1.898	22:42	
Air Blank		0.000	22:42	
Air Blank		0.000	22:44	
Subject Sample	# 2	0.061	22:45	
Breath Volume		1.953	22:45	
Air Blank		0.000	22:46	
Control Test		0.080	22:46	
Air Blank			22:46	RFI Detect
Air Blank		0.000	22:47	

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LOCKE,REBECCA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BECHO

DATE/TIME OF TEST: 05/28/2025 01:51:38

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.000	01:47
Breath Volume		1.230	01:47
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.000	01:50
Breath Volume		1.417	01:50
Air Blank		0.000	01:50
Control Test		0.079	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES,MARK,O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,HOWIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 05/28/2025 13:17:22

Diagnostics Check		OK	13:11
Air Blank		0.000	13:11
Control Test		0.079	13:11
Air Blank		0.000	13:12
Subject Sample	# 1	0.222	13:12
Breath Volume		3.867	13:12
Air Blank		0.000	13:13
Air Blank		0.000	13:15
Subject Sample	# 2	0.236	13:15
Breath Volume		3.710	13:15
Air Blank		0.000	13:16
Control Test		0.079	13:16
Air Blank		0.000	13:17
Diagnostics Check		OK	13:17

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KIMBERLY,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: REDMAN

DATE/TIME OF TEST: 05/28/2025 19:24:14

Diagnostics Check		OK	19:20	
Air Blank		0.000	19:20	
Control Test		0.079	19:21	
Air Blank		0.000	19:21	
Subject Sample	# 1		19:22	Interferent Detect
Breath Volume		2.312	19:22	
Air Blank		0.000	19:23	
Control Test		0.079	19:23	
Air Blank		0.000	19:24	
Diagnostics Check		OK	19:24	

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: AGOSTINE

DATE/TIME OF TEST: 05/29/2025 01:02:04

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.223	00:57
Breath Volume		2.429	00:57
Air Blank		0.000	00:58
Air Blank		0.000	00:59
Subject Sample	# 2	0.229	01:00
Breath Volume		2.992	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:02

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/29/2025 04:42:11

Diagnostics Check		OK	04:35
Air Blank		0.000	04:36
Control Test		0.079	04:36
Air Blank		0.000	04:37
Subject Sample	# 1	0.061	04:37
Breath Volume		1.699	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.063	04:40
Breath Volume		2.335	04:40
Air Blank		0.000	04:41
Control Test		0.079	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTINO,ERIK,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CUTINO,ERIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 05/30/2025 01:24:53

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.177	01:19
Breath Volume		1.796	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.172	01:23
Breath Volume		1.949	01:23
Air Blank		0.000	01:23
Control Test		0.079	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 06/18/2025 11:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACHADO,LAURA,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MACHADO,LAURA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 05/31/2025 03:52:17

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.079	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.183	03:47
Breath Volume		3.917	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.169	03:50
Breath Volume		2.656	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 06/18/2025 11:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WRIGHT,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: DEMAREST

DATE/TIME OF TEST: 05/31/2025 21:16:43

Diagnostics Check		OK	21:07
Air Blank		0.000	21:07
Control Test		0.079	21:07
Air Blank		0.000	21:08
Subject Sample	# 1	0.323	21:08
Breath Volume		3.898	21:08
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.344	21:12
Breath Volume		3.890	21:12
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 3	0.343	21:15
Breath Volume		4.015	21:15
Air Blank		0.000	21:15
Control Test		0.079	21:16
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 06/18/2025 11:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: KELLEY,CHAISE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 05/05/2025 00:04:37

Diagnostics Check		OK	00:02	
Air Blank		0.000	00:02	
Control Test		0.080	00:02	
Air Blank		0.000	00:03	
Subject Sample	# 1		00:03	RFI Detect
Breath Volume		0.000	00:03	
Air Blank		0.000	00:04	
			00:04	RFI Detect
Air Blank		0.000	00:04	

Data Download Date/Time: 06/10/2025 13:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: KELLEY,CHAISE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 05/05/2025 00:10:10

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:07	
Control Test		0.081	00:07	
Air Blank		0.000	00:08	
Subject Sample	# 1		00:08	Interferent Detect
Breath Volume		5.019	00:08	
Air Blank		0.000	00:09	
Control Test		0.081	00:09	
Air Blank		0.000	00:10	
Diagnostics Check		OK	00:10	

Data Download Date/Time: 06/10/2025 13:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: KELLEY,CHAISE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 05/05/2025 00:15:29

Diagnostics Check		OK	00:12	
Air Blank		0.000	00:12	
Control Test		0.076	00:13	
Air Blank		0.000	00:13	
Subject Sample	# 1		00:13	Interferent Detect
Breath Volume		3.644	00:13	
Air Blank		0.000	00:14	
Control Test		0.081	00:14	
Air Blank		0.000	00:15	
Diagnostics Check		OK	00:15	

Data Download Date/Time: 06/10/2025 13:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: KELLEY,CHAISE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 05/05/2025 00:29:31

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:26	
Control Test		0.079	00:26	
Air Blank		0.000	00:27	
Subject Sample	# 1		00:27	Interferent Detect
Breath Volume		3.675	00:27	
Air Blank		0.000	00:28	
Control Test		0.081	00:28	
Air Blank		0.000	00:29	
Diagnostics Check		OK	00:29	

Data Download Date/Time: 06/10/2025 13:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CAMPAGNONE,ALEXANDER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 05/06/2025 01:27:06

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.081	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.181	01:22
Breath Volume		2.105	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.172	01:25
Breath Volume		2.355	01:25
Air Blank		0.000	01:26
Control Test		0.078	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 06/10/2025 13:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,DYLAN,T
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ARNUDTS,GEORGE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: GOBERT

DATE/TIME OF TEST: 05/16/2025 00:51:22

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.080	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.058	00:46
Breath Volume		1.664	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.057	00:49
Breath Volume		1.750	00:49
Air Blank		0.000	00:50
Control Test		0.080	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 06/10/2025 13:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWAB,JOSEPH,S
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: UMMEL,JOSHUA
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: MCGRIFF

DATE/TIME OF TEST: 05/20/2025 13:38:39

Diagnostics Check		OK	13:32
Air Blank		0.000	13:33
Control Test		0.081	13:33
Air Blank		0.000	13:34
Subject Sample	# 1	0.000	13:34
Breath Volume		1.421	13:34
Air Blank		0.000	13:34
Air Blank		0.000	13:36
Subject Sample	# 2	0.000	13:37
Breath Volume		1.414	13:37
Air Blank		0.000	13:37
Control Test		0.081	13:38
Air Blank		0.000	13:38
Diagnostics Check		OK	13:38

Data Download Date/Time: 06/10/2025 13:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: KOUTROS,PATRICK
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: SHEFFIELD

DATE/TIME OF TEST: 05/23/2025 22:56:00

Diagnostics Check		OK	22:43	
Air Blank		0.000	22:44	
Control Test		0.082	22:44	
Air Blank		0.000	22:45	
Subject Sample	# 1		22:48	No Sample Provided
Breath Volume		0.000	22:48	
Air Blank		0.000	22:48	
Air Blank		0.000	22:50	
Subject Sample	# 2	0.056	22:51	
Breath Volume		3.156	22:51	
Air Blank		0.000	22:52	
Air Blank		0.000	22:53	
Subject Sample	# 3	0.052	22:54	
Breath Volume		4.203	22:54	
Air Blank		0.000	22:55	
Control Test		0.079	22:55	
Air Blank		0.000	22:55	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 06/10/2025 13:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: RICHARDS,JAIDON
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 05/24/2025 01:46:01

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:43	
Control Test		0.081	01:43	
Air Blank		0.000	01:43	
Subject Sample	# 1		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:45	
Control Test		0.080	01:45	
Air Blank		0.000	01:45	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 06/10/2025 13:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: RICHARDS,JAIDON
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 05/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: VAQUEDANO AVILA

DATE/TIME OF TEST: 05/25/2025 04:57:04

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:43	
Control Test		0.081	04:43	
Air Blank		0.000	04:44	
Subject Sample	# 1	0.149	04:47	Volume Not Met
Breath Volume		0.152	04:47	
Air Blank		0.000	04:48	
Air Blank		0.000	04:49	
Subject Sample	# 2	0.137	04:50	
Breath Volume		1.738	04:50	
Air Blank		0.000	04:51	
Air Blank		0.000	04:53	
Subject Sample	# 3	0.143	04:55	
Breath Volume		2.660	04:55	
Air Blank		0.000	04:56	
Control Test		0.077	04:56	
Air Blank		0.000	04:56	
Diagnostics Check		OK	04:57	

Data Download Date/Time: 06/10/2025 13:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: WEIMER

DATE/TIME OF TEST: 05/24/2025 19:35:46

Diagnostics Check		OK	19:29
Air Blank		0.000	19:30
Control Test		0.079	19:30
Air Blank		0.000	19:30
Subject Sample	# 1	0.184	19:31
Breath Volume		1.976	19:31
Air Blank		0.000	19:31
Air Blank		0.000	19:33
Subject Sample	# 2	0.194	19:34
Breath Volume		1.804	19:34
Air Blank		0.000	19:34
Control Test		0.079	19:35
Air Blank		0.000	19:35
Diagnostics Check		OK	19:35

Data Download Date/Time: 06/26/2025 15:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ARGUDO,L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:38

SUBJECT'S LAST NAME: BARRICK

DATE/TIME OF TEST: 05/13/2025 13:05:26

Diagnostics Check		OK	12:58
Air Blank		0.000	12:58
Control Test		0.078	12:59
Air Blank		0.000	12:59
Subject Sample	# 1	0.162	13:00
Breath Volume		1.156	13:00
Air Blank		0.000	13:01
Air Blank		0.000	13:03
Subject Sample	# 2	0.161	13:03
Breath Volume		1.480	13:03
Air Blank		0.000	13:04
Control Test		0.076	13:04
Air Blank		0.000	13:05
Diagnostics Check		OK	13:05

Data Download Date/Time: 06/11/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/13/2025 21:01:30

Diagnostics Check		OK	20:58	
Air Blank		0.000	20:58	
Control Test		0.079	20:59	
Air Blank		0.000	20:59	
Subject Sample	# 1		21:00	Subject Test Refused
Breath Volume		0.000	21:00	
Air Blank		0.000	21:00	
Control Test		0.080	21:00	
Air Blank		0.000	21:01	
Diagnostics Check		OK	21:01	

Data Download Date/Time: 06/11/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: FOGLE

DATE/TIME OF TEST: 05/14/2025 03:00:25

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.078	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.137	02:55
Breath Volume		2.414	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.134	02:58
Breath Volume		1.914	02:58
Air Blank		0.000	02:59
Control Test		0.078	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 06/11/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: CASTILLO ANTONIO

DATE/TIME OF TEST: 05/14/2025 05:13:30

Diagnostics Check		OK	05:07
Air Blank		0.000	05:07
Control Test		0.078	05:07
Air Blank		0.000	05:08
Subject Sample	# 1	0.086	05:08
Breath Volume		2.496	05:08
Air Blank		0.000	05:09
Air Blank		0.000	05:11
Subject Sample	# 2	0.084	05:11
Breath Volume		1.996	05:11
Air Blank		0.000	05:12
Control Test		0.079	05:12
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 06/11/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: CHEWNING

DATE/TIME OF TEST: 05/14/2025 22:38:01

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:35	
Control Test		0.079	22:35	
Air Blank		0.000	22:36	
Subject Sample	# 1		22:36	Subject Test Refused
Breath Volume		0.000	22:36	
Air Blank		0.000	22:37	
Control Test		0.079	22:37	
Air Blank		0.000	22:37	
Diagnostics Check		OK	22:38	

Data Download Date/Time: 06/11/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 05/15/2025 04:13:10

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:10	
Control Test		0.078	04:11	
Air Blank		0.000	04:11	
Subject Sample	# 1		04:11	Subject Test Refused
Breath Volume		0.000	04:11	
Air Blank		0.000	04:12	
Control Test		0.079	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 06/11/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: O'RILEY

DATE/TIME OF TEST: 05/15/2025 23:16:30

Diagnostics Check		OK	23:10
Air Blank		0.000	23:10
Control Test		0.079	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.000	23:12
Breath Volume		3.214	23:12
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.000	23:15
Breath Volume		3.503	23:15
Air Blank		0.000	23:15
Control Test		0.079	23:15
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 06/11/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BATEH,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 05/16/2025 00:27:17

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:24	
Control Test		0.079	00:25	
Air Blank		0.000	00:25	
Subject Sample	# 1		00:25	Subject Test Refused
Breath Volume		0.000	00:25	
Air Blank		0.000	00:26	
Control Test		0.079	00:26	
Air Blank		0.000	00:27	
Diagnostics Check		OK	00:27	

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RICHARDSON,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 05/16/2025 23:09:07

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.079	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.222	23:03
Breath Volume		1.609	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.216	23:07
Breath Volume		1.535	23:07
Air Blank		0.000	23:08
Control Test		0.079	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: BENDETTI

DATE/TIME OF TEST: 05/17/2025 01:04:24

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.267	00:58
Breath Volume		1.203	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:00
Subject Sample	# 2	0.269	01:02
Breath Volume		1.464	01:02
Air Blank		0.000	01:03
Control Test		0.077	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 05/17/2025 05:41:18

Diagnostics Check		OK	05:34
Air Blank		0.000	05:35
Control Test		0.079	05:35
Air Blank		0.000	05:36
Subject Sample	# 1	0.272	05:36
Breath Volume		1.839	05:36
Air Blank		0.000	05:37
Air Blank		0.000	05:39
Subject Sample	# 2	0.276	05:39
Breath Volume		1.570	05:39
Air Blank		0.000	05:40
Control Test		0.078	05:40
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: BASKERVILLE

DATE/TIME OF TEST: 05/17/2025 23:50:45

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.079	23:44
Air Blank		0.000	23:44
Subject Sample	# 1	0.130	23:46
Breath Volume		1.757	23:46
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.130	23:49
Breath Volume		1.355	23:49
Air Blank		0.000	23:49
Control Test		0.078	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: DORISCA

DATE/TIME OF TEST: 05/18/2025 03:36:22

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.078	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.164	03:31
Breath Volume		1.683	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.160	03:34
Breath Volume		1.480	03:34
Air Blank		0.000	03:35
Control Test		0.078	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COLON,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 05/18/2025 05:48:24

Diagnostics Check		OK	05:42
Air Blank		0.000	05:42
Control Test		0.079	05:43
Air Blank		0.000	05:43
Subject Sample	# 1	0.090	05:43
Breath Volume		1.828	05:43
Air Blank		0.000	05:44
Air Blank		0.000	05:46
Subject Sample	# 2	0.090	05:46
Breath Volume		1.699	05:46
Air Blank		0.000	05:47
Control Test		0.079	05:47
Air Blank		0.000	05:48
Diagnostics Check		OK	05:48

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: SAWA

DATE/TIME OF TEST: 05/19/2025 00:55:47

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:53	
Control Test		0.080	00:53	
Air Blank		0.000	00:54	
Subject Sample	# 1		00:54	Subject Test Refused
Breath Volume		0.000	00:54	
Air Blank		0.000	00:54	
Control Test		0.079	00:55	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COSTA,F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: VASQUEZ LOPEZ

DATE/TIME OF TEST: 05/19/2025 03:33:33

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:31	
Control Test		0.079	03:31	
Air Blank		0.000	03:31	
Subject Sample	# 1		03:32	Subject Test Refused
Breath Volume		0.000	03:32	
Air Blank		0.000	03:32	
Control Test		0.079	03:32	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:33	

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: SWAN

DATE/TIME OF TEST: 05/19/2025 21:50:15

Diagnostics Check		OK	21:47	
Air Blank		0.000	21:47	
Control Test		0.079	21:48	
Air Blank		0.000	21:48	
Subject Sample	# 1		21:48	Subject Test Refused
Breath Volume		0.000	21:48	
Air Blank		0.000	21:49	
Control Test		0.079	21:49	
Air Blank		0.000	21:50	
Diagnostics Check		OK	21:50	

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: ZYBERAJ

DATE/TIME OF TEST: 05/20/2025 20:41:53

Diagnostics Check		OK	20:35
Air Blank		0.000	20:35
Control Test		0.080	20:36
Air Blank		0.000	20:36
Subject Sample	# 1	0.085	20:37
Breath Volume		2.503	20:37
Air Blank		0.000	20:37
Air Blank		0.000	20:39
Subject Sample	# 2	0.080	20:40
Breath Volume		1.792	20:40
Air Blank		0.000	20:40
Control Test		0.079	20:41
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRERA,R,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: ALZURIN

DATE/TIME OF TEST: 05/21/2025 00:55:51

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.079	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.081	00:51
Breath Volume		1.910	00:51
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.080	00:54
Breath Volume		1.886	00:54
Air Blank		0.000	00:54
Control Test		0.079	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: VALLS MADRIGAL

DATE/TIME OF TEST: 05/21/2025 04:08:49

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.080	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.234	04:04
Breath Volume		1.546	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:06
Subject Sample	# 2	0.223	04:07
Breath Volume		2.355	04:07
Air Blank		0.000	04:07
Control Test		0.079	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRERA,R,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SEARS,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: BIRTS

DATE/TIME OF TEST: 05/22/2025 01:10:12

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.032	01:05
Breath Volume		1.203	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.030	01:08
Breath Volume		1.335	01:08
Air Blank		0.000	01:09
Control Test		0.079	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 05/22/2025 20:43:14

Diagnostics Check		OK	20:36
Air Blank		0.000	20:36
Control Test		0.079	20:37
Air Blank		0.000	20:37
Subject Sample	# 1	0.083	20:38
Breath Volume		1.718	20:38
Air Blank		0.000	20:39
Air Blank		0.000	20:41
Subject Sample	# 2	0.086	20:41
Breath Volume		2.031	20:41
Air Blank		0.000	20:42
Control Test		0.079	20:42
Air Blank		0.000	20:43
Diagnostics Check		OK	20:43

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEIS,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 05/23/2025 01:57:10

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:54	
Control Test		0.079	01:54	
Air Blank		0.000	01:55	
Subject Sample	# 1		01:55	Subject Test Refused
Breath Volume		0.000	01:55	
Air Blank		0.000	01:56	
Control Test		0.079	01:56	
Air Blank		0.000	01:57	
Diagnostics Check		OK	01:57	

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: GERACI

DATE/TIME OF TEST: 05/23/2025 04:17:39

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.079	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.143	04:13
Breath Volume		3.222	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.142	04:15
Breath Volume		3.441	04:15
Air Blank		0.000	04:16
Control Test		0.079	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: JOHNSEN

DATE/TIME OF TEST: 05/23/2025 19:54:08

Diagnostics Check		OK	19:47
Air Blank		0.000	19:48
Control Test		0.079	19:48
Air Blank		0.000	19:49
Subject Sample	# 1	0.000	19:49
Breath Volume		2.496	19:49
Air Blank		0.000	19:50
Air Blank		0.000	19:52
Subject Sample	# 2	0.000	19:52
Breath Volume		2.835	19:52
Air Blank		0.000	19:53
Control Test		0.079	19:53
Air Blank		0.000	19:54
Diagnostics Check		OK	19:54

Data Download Date/Time: 06/11/2025 12:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: TORRENCE

DATE/TIME OF TEST: 05/23/2025 21:14:56

Diagnostics Check		OK	21:08
Air Blank		0.000	21:09
Control Test		0.080	21:09
Air Blank		0.000	21:09
Subject Sample	# 1	0.141	21:10
Breath Volume		1.679	21:10
Air Blank		0.000	21:11
Air Blank		0.000	21:12
Subject Sample	# 2	0.134	21:13
Breath Volume		1.402	21:13
Air Blank		0.000	21:13
Control Test		0.078	21:14
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ECHEVERRIAVIDAL,X
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: RAILEY

DATE/TIME OF TEST: 05/23/2025 22:05:06

Diagnostics Check		OK	21:58
Air Blank		0.000	21:59
Control Test		0.079	21:59
Air Blank		0.000	22:00
Subject Sample	# 1	0.151	22:00
Breath Volume		2.617	22:00
Air Blank		0.000	22:01
Air Blank		0.000	22:03
Subject Sample	# 2	0.155	22:03
Breath Volume		2.589	22:03
Air Blank		0.000	22:04
Control Test		0.078	22:04
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: HOLLIDAY

DATE/TIME OF TEST: 05/24/2025 02:04:40

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:02	
Control Test		0.079	02:02	
Air Blank		0.000	02:03	
Subject Sample	# 1		02:03	Subject Test Refused
Breath Volume		0.000	02:03	
Air Blank		0.000	02:03	
Control Test		0.079	02:04	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 06/11/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SNEAD

DATE/TIME OF TEST: 05/24/2025 03:30:26

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:27	
Control Test		0.079	03:28	
Air Blank		0.000	03:28	
Subject Sample	# 1		03:29	Subject Test Refused
Breath Volume		0.000	03:29	
Air Blank		0.000	03:29	
Control Test		0.079	03:29	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 06/11/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ECHEVERRIAVIDAL,X
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 05/24/2025 04:23:02

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.079	04:17
Air Blank		0.000	04:17
Subject Sample	# 1	0.248	04:18
Breath Volume		3.242	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:20
Subject Sample	# 2	0.237	04:21
Breath Volume		2.902	04:21
Air Blank		0.000	04:22
Control Test		0.078	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:23

Data Download Date/Time: 06/11/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: COREY,H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 05/24/2025 20:53:12

Diagnostics Check		OK	20:45
Air Blank		0.000	20:46
Control Test		0.080	20:46
Air Blank		0.000	20:47
Subject Sample	# 1	0.140	20:47
Breath Volume		1.261	20:47
Air Blank		0.000	20:48
Air Blank		0.000	20:50
Subject Sample	# 2	0.128	20:51
Breath Volume		1.613	20:51
Air Blank		0.000	20:52
Control Test		0.079	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 06/11/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MARTIN,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: CARDOSA GARCIA

DATE/TIME OF TEST: 05/25/2025 02:36:03

Diagnostics Check		OK	02:29
Air Blank		0.000	02:30
Control Test		0.079	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.075	02:31
Breath Volume		1.855	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.073	02:34
Breath Volume		1.976	02:34
Air Blank		0.000	02:35
Control Test		0.079	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:36

Data Download Date/Time: 06/11/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: MAGANA

DATE/TIME OF TEST: 05/25/2025 05:03:14

Diagnostics Check		OK	05:00	
Air Blank		0.000	05:00	
Control Test		0.079	05:01	
Air Blank		0.000	05:01	
Subject Sample	# 1		05:01	Subject Test Refused
Breath Volume		0.000	05:01	
Air Blank		0.000	05:02	
Control Test		0.079	05:02	
Air Blank		0.000	05:03	
Diagnostics Check		OK	05:03	

Data Download Date/Time: 06/11/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALLACE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: DESUE

DATE/TIME OF TEST: 05/26/2025 00:51:14

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:48	
Control Test		0.081	00:49	
Air Blank		0.000	00:49	
Subject Sample	# 1		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:50	
Control Test		0.079	00:50	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 06/11/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BARRY,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 05/26/2025 01:40:33

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.079	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.120	01:35
Breath Volume		2.425	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.120	01:38
Breath Volume		2.160	01:38
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 06/11/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 05/26/2025 03:31:04

Diagnostics Check		OK	03:24
Air Blank		0.000	03:24
Control Test		0.079	03:25
Air Blank		0.000	03:25
Subject Sample	# 1	0.208	03:26
Breath Volume		3.515	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:28
Subject Sample	# 2	0.206	03:29
Breath Volume		3.441	03:29
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:31

Data Download Date/Time: 06/11/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Neptune Beach PD
ARRESTING OFFICER: TANNER,M
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: DIVOVIC

DATE/TIME OF TEST: 05/26/2025 04:57:25

Diagnostics Check		OK	04:45	
Air Blank		0.000	04:46	
Control Test		0.079	04:46	
Air Blank		0.000	04:47	
Subject Sample	# 1		04:50	No Sample Provided
Breath Volume		0.000	04:50	
Air Blank		0.000	04:50	
Air Blank		0.000	04:52	
Subject Sample	# 2		04:56	No Sample Provided
Breath Volume		0.000	04:56	
Air Blank		0.000	04:56	
Control Test		0.079	04:56	
Air Blank		0.000	04:57	
Diagnostics Check		OK	04:57	

Data Download Date/Time: 06/11/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: NAPIER

DATE/TIME OF TEST: 05/27/2025 02:23:38

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.080	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.081	02:18
Breath Volume		1.433	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.077	02:22
Breath Volume		1.398	02:22
Air Blank		0.000	02:22
Control Test		0.080	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 06/11/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: GILMER

DATE/TIME OF TEST: 05/29/2025 03:27:21

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.154	03:22
Breath Volume		2.000	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:24
Subject Sample	# 2	0.153	03:25
Breath Volume		1.984	03:25
Air Blank		0.000	03:26
Control Test		0.080	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 06/11/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 05/30/2025 23:46:05

Diagnostics Check		OK	23:39
Air Blank		0.000	23:39
Control Test		0.079	23:40
Air Blank		0.000	23:40
Subject Sample	# 1	0.144	23:41
Breath Volume		4.445	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.143	23:44
Breath Volume		4.449	23:44
Air Blank		0.000	23:45
Control Test		0.079	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 06/11/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEIS,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 05/31/2025 23:09:24

Diagnostics Check		OK	23:02
Air Blank		0.000	23:03
Control Test		0.079	23:03
Air Blank		0.000	23:04
Subject Sample	# 1	0.137	23:04
Breath Volume		2.656	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.135	23:07
Breath Volume		2.281	23:07
Air Blank		0.000	23:08
Control Test		0.079	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 06/11/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: PERESTRELLO

DATE/TIME OF TEST: 05/31/2025 23:59:09

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.079	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.111	23:54
Breath Volume		1.437	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.116	23:57
Breath Volume		1.429	23:57
Air Blank		0.000	23:58
Control Test		0.079	23:58
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 06/11/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPINA,C,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 05/03/2025 23:56:11

Diagnostics Check		OK	23:53	
Air Blank		0.000	23:53	
Control Test		0.082	23:53	
Air Blank		0.000	23:54	
Subject Sample	# 1		23:54	Subject Test Refused
Breath Volume		0.000	23:54	
Air Blank		0.000	23:55	
Control Test		0.082	23:55	
Air Blank		0.000	23:56	
Diagnostics Check		OK	23:56	

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: PALACIOS CRUZ

DATE/TIME OF TEST: 05/05/2025 00:17:47

Diagnostics Check		OK	00:14	
Air Blank		0.000	00:15	
Control Test		0.082	00:15	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:16	Subject Test Refused
Breath Volume		0.000	00:16	
Air Blank		0.000	00:16	
Control Test		0.081	00:17	
Air Blank		0.000	00:17	
Diagnostics Check		OK	00:17	

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 05/09/2025 02:01:22

Diagnostics Check		OK	01:53
Air Blank		0.000	01:54
Control Test		0.082	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.175	01:56
Breath Volume		2.316	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:58
Subject Sample	# 2	0.159	01:59
Breath Volume		2.195	01:59
Air Blank		0.000	02:00
Control Test		0.078	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 05/09/2025 03:07:35

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:05	
Control Test		0.082	03:05	
Air Blank		0.000	03:06	
Subject Sample	# 1		03:06	Subject Test Refused
Breath Volume		0.000	03:06	
Air Blank		0.000	03:06	
Control Test		0.081	03:06	
Air Blank		0.000	03:07	
Diagnostics Check		OK	03:07	

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 05/10/2025 03:30:07

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.082	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.180	03:25
Breath Volume		2.128	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.172	03:28
Breath Volume		2.136	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROSSI,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: RAINS

DATE/TIME OF TEST: 05/11/2025 04:45:07

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:42	
Control Test		0.082	04:43	
Air Blank		0.000	04:43	
Subject Sample	# 1		04:43	Subject Test Refused
Breath Volume		0.000	04:43	
Air Blank		0.000	04:44	
Control Test		0.082	04:44	
Air Blank		0.000	04:45	
Diagnostics Check		OK	04:45	

Data Download Date/Time: 05/13/2025 14:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,C,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: PAQUIN

DATE/TIME OF TEST: 05/11/2025 17:09:02

Diagnostics Check		OK	16:58
Air Blank		0.000	16:59
Control Test		0.081	16:59
Air Blank		0.000	17:00
Subject Sample	# 1	0.063	17:03
Breath Volume		1.449	17:03
Air Blank		0.000	17:04
Air Blank		0.000	17:05
Subject Sample	# 2	0.059	17:07
Breath Volume		1.265	17:07
Air Blank		0.000	17:08
Control Test		0.079	17:08
Air Blank		0.000	17:08
Diagnostics Check		OK	17:09

Data Download Date/Time: 05/13/2025 14:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,D,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: DALLAS

DATE/TIME OF TEST: 05/13/2025 02:09:34

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:07	
Control Test		0.082	02:07	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:08	
Control Test		0.082	02:08	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 05/13/2025 14:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SEARS,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 05/14/2025 01:08:56

Diagnostics Check		OK	01:06	
Air Blank		0.000	01:06	
Control Test		0.082	01:06	
Air Blank		0.000	01:07	
Subject Sample	# 1		01:07	Subject Test Refused
Breath Volume		0.000	01:07	
Air Blank		0.000	01:08	
Control Test		0.081	01:08	
Air Blank		0.000	01:08	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARRILLO,P
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: HEMINGWAY

DATE/TIME OF TEST: 05/15/2025 05:34:49

Diagnostics Check		OK	05:28
Air Blank		0.000	05:29
Control Test		0.081	05:29
Air Blank		0.000	05:29
Subject Sample	# 1	0.150	05:30
Breath Volume		1.867	05:30
Air Blank		0.000	05:31
Air Blank		0.000	05:32
Subject Sample	# 2	0.148	05:33
Breath Volume		1.230	05:33
Air Blank		0.000	05:33
Control Test		0.080	05:34
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: BEATTY

DATE/TIME OF TEST: 05/16/2025 02:05:46

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.082	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.129	02:01
Breath Volume		2.718	02:01
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.131	02:04
Breath Volume		2.554	02:04
Air Blank		0.000	02:04
Control Test		0.081	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: VASQUEZ GALDAMEZ

DATE/TIME OF TEST: 05/17/2025 00:22:36

Diagnostics Check		OK	00:19	
Air Blank		0.000	00:20	
Control Test		0.082	00:20	
Air Blank		0.000	00:21	
Subject Sample	# 1		00:21	Subject Test Refused
Breath Volume		0.000	00:21	
Air Blank		0.000	00:21	
Control Test		0.082	00:21	
Air Blank		0.000	00:22	
Diagnostics Check		OK	00:22	

Data Download Date/Time: 06/11/2025 14:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: OWENS,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: SCULLY

DATE/TIME OF TEST: 05/18/2025 01:21:23

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.082	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.162	01:16
Breath Volume		1.515	01:16
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.166	01:19
Breath Volume		2.070	01:19
Air Blank		0.000	01:20
Control Test		0.080	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 05/18/2025 04:27:44

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.082	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.217	04:22
Breath Volume		2.214	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.215	04:25
Breath Volume		2.042	04:25
Air Blank		0.000	04:26
Control Test		0.080	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARIN,O
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 05/18/2025 06:06:38

Diagnostics Check		OK	06:03	
Air Blank		0.000	06:04	
Control Test		0.082	06:04	
Air Blank		0.000	06:04	
Subject Sample	# 1		06:05	Subject Test Refused
Breath Volume		0.000	06:05	
Air Blank		0.000	06:05	
Control Test		0.082	06:06	
Air Blank		0.000	06:06	
Diagnostics Check		OK	06:06	

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: NEWTON,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/18/2025 23:28:11

Diagnostics Check		OK	23:19
Air Blank		0.000	23:19
Control Test		0.082	23:20
Air Blank		0.000	23:20
Subject Sample	# 1	0.145	23:22
Breath Volume		2.015	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.159	23:26
Breath Volume		1.996	23:26
Air Blank		0.000	23:27
Control Test		0.078	23:27
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:38

SUBJECT'S LAST NAME: TELEMACQUE JR

DATE/TIME OF TEST: 05/19/2025 07:05:19

Diagnostics Check		OK	06:58
Air Blank		0.000	06:59
Control Test		0.082	06:59
Air Blank		0.000	07:00
Subject Sample	# 1	0.000	07:00
Breath Volume		4.773	07:00
Air Blank		0.000	07:01
Air Blank		0.000	07:03
Subject Sample	# 2	0.000	07:03
Breath Volume		4.683	07:03
Air Blank		0.000	07:04
Control Test		0.082	07:04
Air Blank		0.000	07:05
Diagnostics Check		OK	07:05

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 05/19/2025 21:58:09

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:55	
Control Test		0.082	21:55	
Air Blank		0.000	21:56	
Subject Sample	# 1		21:56	Subject Test Refused
Breath Volume		0.000	21:56	
Air Blank		0.000	21:57	
Control Test		0.081	21:57	
Air Blank		0.000	21:58	
Diagnostics Check		OK	21:58	

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: PRENDERGAST

DATE/TIME OF TEST: 05/20/2025 03:29:45

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.081	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.089	03:24
Breath Volume		1.968	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.088	03:28
Breath Volume		2.347	03:28
Air Blank		0.000	03:28
Control Test		0.081	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAYGAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 05/22/2025 03:38:54

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:36	
Control Test		0.082	03:36	
Air Blank		0.000	03:37	
Subject Sample	# 1		03:37	Subject Test Refused
Breath Volume		0.000	03:37	
Air Blank		0.000	03:37	
Control Test		0.082	03:38	
Air Blank		0.000	03:38	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: LUMER

DATE/TIME OF TEST: 05/23/2025 01:10:50

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.081	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.160	01:06
Breath Volume		1.382	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.163	01:09
Breath Volume		1.226	01:09
Air Blank		0.000	01:09
Control Test		0.080	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHR,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROSSI,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 05/23/2025 03:18:58

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.082	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:17	Subject Test Refused
Breath Volume		0.000	03:17	
Air Blank		0.000	03:18	
Control Test		0.082	03:18	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:18	

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/25/2025 01:57:26

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.081	01:51
Air Blank		0.000	01:52
Subject Sample	# 1	0.139	01:52
Breath Volume		1.578	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.149	01:55
Breath Volume		1.953	01:55
Air Blank		0.000	01:56
Control Test		0.079	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 05/25/2025 04:49:40

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.082	04:44
Air Blank		0.000	04:44
Subject Sample	# 1	0.174	04:45
Breath Volume		2.511	04:45
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.165	04:47
Breath Volume		1.140	04:47
Air Blank		0.000	04:48
Control Test		0.079	04:49
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: TEGEGN

DATE/TIME OF TEST: 05/25/2025 22:52:25

Diagnostics Check		OK	22:49	
Air Blank		0.000	22:49	
Control Test		0.082	22:50	
Air Blank		0.000	22:50	
Subject Sample	# 1		22:51	Subject Test Refused
Breath Volume		0.000	22:51	
Air Blank		0.000	22:51	
Control Test		0.081	22:51	
Air Blank		0.000	22:52	
Diagnostics Check		OK	22:52	

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MIXSON

DATE/TIME OF TEST: 05/26/2025 01:52:09

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.082	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.013	01:47
Breath Volume		1.640	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.013	01:50
Breath Volume		1.175	01:50
Air Blank		0.000	01:51
Control Test		0.081	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: BEKELE

DATE/TIME OF TEST: 05/26/2025 03:05:17

Diagnostics Check		OK	03:02	
Air Blank		0.000	03:02	
Control Test		0.082	03:03	
Air Blank		0.000	03:03	
Subject Sample	# 1		03:03	Subject Test Refused
Breath Volume		0.000	03:03	
Air Blank		0.000	03:04	
Control Test		0.082	03:04	
Air Blank		0.000	03:05	
Diagnostics Check		OK	03:05	

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 05/27/2025 00:51:21

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.081	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.208	00:46
Breath Volume		1.285	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.214	00:49
Breath Volume		2.281	00:49
Air Blank		0.000	00:50
Control Test		0.080	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CROSBY,H
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: MCCLELLAN

DATE/TIME OF TEST: 05/28/2025 05:46:31

Diagnostics Check		OK	05:38
Air Blank		0.000	05:38
Control Test		0.082	05:39
Air Blank		0.000	05:39
Subject Sample	# 1	0.141	05:41
Breath Volume		2.675	05:41
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.145	05:44
Breath Volume		2.023	05:44
Air Blank		0.000	05:45
Control Test		0.076	05:45
Air Blank		0.000	05:46
Diagnostics Check		OK	05:46

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/29/2025 22:16:41

Diagnostics Check		OK	22:09
Air Blank		0.000	22:10
Control Test		0.081	22:10
Air Blank		0.000	22:10
Subject Sample	# 1	0.131	22:11
Breath Volume		2.457	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.122	22:14
Breath Volume		3.113	22:14
Air Blank		0.000	22:15
Control Test		0.080	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 06/11/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITSON,G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/30/2025 19:41:20

Diagnostics Check		OK	19:38	
Air Blank		0.000	19:38	
Control Test		0.082	19:39	
Air Blank		0.000	19:39	
Subject Sample	# 1		19:39	Subject Test Refused
Breath Volume		0.000	19:39	
Air Blank		0.000	19:40	
Control Test		0.082	19:40	
Air Blank		0.000	19:41	
Diagnostics Check		OK	19:41	

Data Download Date/Time: 06/11/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 05/31/2025 01:12:45

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.082	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.190	01:07
Breath Volume		1.976	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:09
Subject Sample	# 2	0.186	01:10
Breath Volume		2.355	01:10
Air Blank		0.000	01:11
Control Test		0.078	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 06/11/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JORDAN,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: CHINAPANA

DATE/TIME OF TEST: 05/01/2025 02:11:23

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.174	02:05
Breath Volume		1.578	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.161	02:09
Breath Volume		1.476	02:09
Air Blank		0.000	02:10
Control Test		0.079	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 05/25/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: STECKLEY

DATE/TIME OF TEST: 05/01/2025 05:13:38

Diagnostics Check		OK	05:05	
Air Blank		0.000	05:05	
Control Test		0.080	05:05	
Air Blank		0.000	05:06	
Subject Sample	# 1		05:09	No Sample Provided
Breath Volume		0.000	05:09	
Air Blank		0.000	05:09	
Air Blank		0.000	05:12	
Subject Sample	# 2		05:12	Subject Test Refused
Breath Volume		0.000	05:12	
Air Blank		0.000	05:12	
Control Test		0.080	05:13	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 05/25/2025 11:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ROY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MILLER,B
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 05/02/2025 04:09:21

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:01	
Control Test		0.080	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:05	No Sample Provided
Breath Volume		0.000	04:05	
Air Blank		0.000	04:05	
Air Blank		0.000	04:07	
Subject Sample	# 2		04:07	Subject Test Refused
Breath Volume		0.000	04:07	
Air Blank		0.000	04:08	
Control Test		0.080	04:08	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 05/25/2025 11:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: FAVALORA

DATE/TIME OF TEST: 05/03/2025 19:03:47

Diagnostics Check		OK	18:56
Air Blank		0.000	18:56
Control Test		0.080	18:56
Air Blank		0.000	18:57
Subject Sample	# 1	0.188	18:59
Breath Volume		2.945	18:59
Air Blank		0.000	18:59
Air Blank		0.000	19:01
Subject Sample	# 2	0.194	19:02
Breath Volume		2.503	19:02
Air Blank		0.000	19:02
Control Test		0.081	19:03
Air Blank		0.000	19:03
Diagnostics Check		OK	19:03

Data Download Date/Time: 05/25/2025 11:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: THIBEAUX

DATE/TIME OF TEST: 05/04/2025 02:19:14

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.081	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:18	RFI Detect
Breath Volume		0.000	02:18	
Air Blank			02:18	RFI Detect
Air Blank		0.000	02:19	

Data Download Date/Time: 05/25/2025 11:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: THIBEAUX

DATE/TIME OF TEST: 05/04/2025 02:30:47

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:22	
Control Test		0.080	02:23	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:26	No Sample Provided
Breath Volume		0.000	02:26	
Air Blank		0.000	02:27	
Air Blank		0.000	02:29	
Subject Sample	# 2		02:29	Subject Test Refused
Breath Volume		0.000	02:29	
Air Blank		0.000	02:29	
Control Test		0.080	02:30	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 05/25/2025 11:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ROY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 05/06/2025 00:21:15

Diagnostics Check		OK	00:13	
Air Blank		0.000	00:13	
Control Test		0.079	00:13	
Air Blank		0.000	00:14	
Subject Sample	# 1	0.129	00:16	
Breath Volume		2.707	00:16	
Air Blank		0.000	00:17	
Air Blank		0.000	00:19	
Subject Sample	# 2		00:19	Subject Test Refused
Breath Volume		0.000	00:19	
Air Blank		0.000	00:20	
Control Test		0.079	00:20	
Air Blank		0.000	00:21	
Diagnostics Check		OK	00:21	

Data Download Date/Time: 05/25/2025 11:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: STUDY

DATE/TIME OF TEST: 05/08/2025 20:17:25

Diagnostics Check		OK	20:10
Air Blank		0.000	20:11
Control Test		0.079	20:11
Air Blank		0.000	20:12
Subject Sample	# 1	0.208	20:12
Breath Volume		3.078	20:12
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.210	20:15
Breath Volume		3.859	20:15
Air Blank		0.000	20:16
Control Test		0.079	20:16
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

Data Download Date/Time: 05/25/2025 11:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MCCLENDON

DATE/TIME OF TEST: 05/10/2025 03:24:09

Diagnostics Check		OK	03:12	
Air Blank		0.000	03:12	
Control Test		0.080	03:12	
Air Blank		0.000	03:13	
Subject Sample	# 1	0.106	03:16	Volume Not Met
Breath Volume		0.425	03:16	
Air Blank		0.000	03:17	
Air Blank		0.000	03:18	
Subject Sample	# 2	0.154	03:19	
Breath Volume		1.718	03:19	
Air Blank		0.000	03:20	
Air Blank		0.000	03:22	
Subject Sample	# 3	0.166	03:22	
Breath Volume		1.589	03:22	
Air Blank		0.000	03:23	
Control Test		0.080	03:23	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 05/25/2025 11:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CHILIKI,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: DEESE

DATE/TIME OF TEST: 05/10/2025 05:13:34

Diagnostics Check		OK	05:02	
Air Blank		0.000	05:02	
Control Test		0.080	05:02	
Air Blank		0.000	05:03	
Subject Sample	# 1	0.135	05:06	Volume Not Met
Breath Volume		0.238	05:06	
Air Blank		0.000	05:07	
Air Blank		0.000	05:08	
Subject Sample	# 2	0.134	05:12	Volume Not Met
Breath Volume		0.000	05:12	
Air Blank		0.000	05:12	
Control Test		0.077	05:12	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 05/25/2025 11:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CHILIKI,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: DEESE

DATE/TIME OF TEST: 05/10/2025 05:44:09

Diagnostics Check		OK	05:37
Air Blank		0.000	05:37
Control Test		0.079	05:38
Air Blank		0.000	05:38
Subject Sample	# 1	0.123	05:39
Breath Volume		2.070	05:39
Air Blank		0.000	05:39
Air Blank		0.000	05:41
Subject Sample	# 2	0.126	05:42
Breath Volume		1.781	05:42
Air Blank		0.000	05:43
Control Test		0.079	05:43
Air Blank		0.000	05:44
Diagnostics Check		OK	05:44

Data Download Date/Time: 05/25/2025 11:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: GOINS

DATE/TIME OF TEST: 05/14/2025 04:03:06

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.181	03:58
Breath Volume		2.203	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.189	04:01
Breath Volume		1.730	04:01
Air Blank		0.000	04:02
Control Test		0.079	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 05/25/2025 11:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MORENO,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: WILTON

DATE/TIME OF TEST: 05/16/2025 01:13:32

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.080	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.180	01:08
Breath Volume		1.894	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.171	01:12
Breath Volume		1.671	01:12
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 05/25/2025 11:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ROY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: RAGSDALE

DATE/TIME OF TEST: 05/17/2025 03:15:54

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:07	
Control Test		0.080	03:08	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:11	No Sample Provided
Breath Volume		0.000	03:11	
Air Blank		0.000	03:12	
Air Blank		0.000	03:14	
Subject Sample	# 2		03:14	Subject Test Refused
Breath Volume		0.000	03:14	
Air Blank		0.000	03:15	
Control Test		0.080	03:15	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 05/25/2025 11:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BOIVIN,N
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: FRITH

DATE/TIME OF TEST: 05/17/2025 22:56:28

Diagnostics Check		OK	22:46	
Air Blank		0.000	22:46	
Control Test		0.080	22:47	
Air Blank		0.000	22:47	
Subject Sample	# 1		22:50	No Sample Provided
Breath Volume		0.000	22:50	
Air Blank		0.000	22:51	
Air Blank		0.000	22:53	
Subject Sample	# 2		22:55	Subject Test Refused
Breath Volume		0.000	22:55	
Air Blank		0.000	22:55	
Control Test		0.080	22:55	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 05/25/2025 11:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/18/2025 04:29:39

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:21	
Control Test		0.079	04:21	
Air Blank		0.000	04:22	
Subject Sample	# 1		04:25	No Sample Provided
Breath Volume		0.000	04:25	
Air Blank		0.000	04:25	
Air Blank		0.000	04:27	
Subject Sample	# 2		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:28	
Control Test		0.080	04:29	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 05/25/2025 11:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JORDAN,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/20/2025 22:39:03

Diagnostics Check		OK	22:32
Air Blank		0.000	22:32
Control Test		0.079	22:32
Air Blank		0.000	22:33
Subject Sample	# 1	0.210	22:34
Breath Volume		2.128	22:34
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 2	0.220	22:37
Breath Volume		2.000	22:37
Air Blank		0.000	22:38
Control Test		0.078	22:38
Air Blank		0.000	22:38
Diagnostics Check		OK	22:39

Data Download Date/Time: 05/25/2025 11:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: COUNTS

DATE/TIME OF TEST: 05/21/2025 22:45:13

Diagnostics Check		OK	22:36	
Air Blank		0.000	22:37	
Control Test		0.080	22:37	
Air Blank		0.000	22:38	
Subject Sample	# 1		22:41	No Sample Provided
Breath Volume		0.000	22:41	
Air Blank		0.000	22:41	
Air Blank		0.000	22:43	
Subject Sample	# 2		22:43	Subject Test Refused
Breath Volume		0.000	22:43	
Air Blank		0.000	22:44	
Control Test		0.080	22:44	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 05/25/2025 11:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: BOSTICK

DATE/TIME OF TEST: 05/22/2025 19:43:58

Diagnostics Check		OK	19:35	
Air Blank		0.000	19:35	
Control Test		0.080	19:36	
Air Blank		0.000	19:36	
Subject Sample	# 1		19:39	No Sample Provided
Breath Volume		0.000	19:39	
Air Blank		0.000	19:40	
Air Blank		0.000	19:42	
Subject Sample	# 2		19:42	Subject Test Refused
Breath Volume		0.000	19:42	
Air Blank		0.000	19:43	
Control Test		0.080	19:43	
Air Blank		0.000	19:43	
Diagnostics Check		OK	19:43	

Data Download Date/Time: 05/25/2025 11:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HAUSLER,T
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: BUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/22/2025 23:29:36

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.081	23:24
Air Blank		0.000	23:24
Subject Sample	# 1	0.064	23:25
Breath Volume		2.945	23:25
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.059	23:28
Breath Volume		3.468	23:28
Air Blank		0.000	23:28
Control Test		0.080	23:29
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 05/25/2025 11:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,BRYCE,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:49

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/23/2025 17:25:31

Diagnostics Check		OK	17:14	
Air Blank		0.000	17:14	
Control Test		0.080	17:14	
Air Blank		0.000	17:15	
Subject Sample	# 1		17:18	No Sample Provided
Breath Volume		0.000	17:18	
Air Blank		0.000	17:19	
Air Blank		0.000	17:21	
Subject Sample	# 2		17:24	No Sample Provided
Breath Volume		0.000	17:24	
Air Blank		0.000	17:24	
Control Test		0.080	17:24	
Air Blank		0.000	17:25	
Diagnostics Check		OK	17:25	

Data Download Date/Time: 05/25/2025 11:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ROY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: CERNIK

DATE/TIME OF TEST: 05/23/2025 21:59:41

Diagnostics Check		OK	21:51	
Air Blank		0.000	21:51	
Control Test		0.080	21:51	
Air Blank		0.000	21:52	
Subject Sample	# 1		21:55	No Sample Provided
Breath Volume		0.000	21:55	
Air Blank		0.000	21:56	
Air Blank		0.000	21:58	
Subject Sample	# 2		21:58	Subject Test Refused
Breath Volume		0.000	21:58	
Air Blank		0.000	21:58	
Control Test		0.080	21:59	
Air Blank		0.000	21:59	
Diagnostics Check		OK	21:59	

Data Download Date/Time: 05/25/2025 11:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,BRYCE,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RODECK,H
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:14

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/25/2025 15:48:55

Diagnostics Check		OK	15:37	
Air Blank		0.000	15:38	
Control Test		0.080	15:38	
Air Blank		0.000	15:38	
Subject Sample	# 1		15:41	No Sample Provided
Breath Volume		0.000	15:41	
Air Blank		0.000	15:42	
Air Blank		0.000	15:44	
Subject Sample	# 2		15:47	No Sample Provided
Breath Volume		0.000	15:47	
Air Blank		0.000	15:48	
Control Test		0.079	15:48	
Air Blank		0.000	15:48	
Diagnostics Check		OK	15:48	

Data Download Date/Time: 06/28/2025 11:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ROY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BASOV,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MCANINCH

DATE/TIME OF TEST: 05/25/2025 23:36:05

Diagnostics Check		OK	23:27	
Air Blank		0.000	23:28	
Control Test		0.080	23:28	
Air Blank		0.000	23:28	
Subject Sample	# 1		23:32	No Sample Provided
Breath Volume		0.000	23:32	
Air Blank		0.000	23:32	
Air Blank		0.000	23:34	
Subject Sample	# 2		23:34	Subject Test Refused
Breath Volume		0.000	23:34	
Air Blank		0.000	23:35	
Control Test		0.080	23:35	
Air Blank		0.000	23:36	
Diagnostics Check		OK	23:36	

Data Download Date/Time: 06/28/2025 11:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HAUSLER,T
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: BUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 05/26/2025 19:22:05

Diagnostics Check		OK	19:19	
Air Blank		0.000	19:19	
Control Test		0.080	19:20	
Air Blank		0.000	19:20	
Subject Sample	# 1		19:20	Subject Test Refused
Breath Volume		0.000	19:20	
Air Blank		0.000	19:21	
Control Test		0.080	19:21	
Air Blank		0.000	19:22	
Diagnostics Check		OK	19:22	

Data Download Date/Time: 06/28/2025 11:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: KINGSLEY,H
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: BUI
ARREST TIME: 17:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 05/26/2025 20:32:26

Diagnostics Check		OK	20:26
Air Blank		0.000	20:26
Control Test		0.080	20:27
Air Blank		0.000	20:27
Subject Sample	# 1	0.176	20:28
Breath Volume		1.226	20:28
Air Blank		0.000	20:28
Air Blank		0.000	20:30
Subject Sample	# 2	0.165	20:30
Breath Volume		1.726	20:30
Air Blank		0.000	20:31
Control Test		0.080	20:31
Air Blank		0.000	20:32
Diagnostics Check		OK	20:32

Data Download Date/Time: 06/28/2025 11:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WITMER,B
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: BUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: STRATTON

DATE/TIME OF TEST: 05/26/2025 20:59:45

Diagnostics Check		OK	20:53
Air Blank		0.000	20:54
Control Test		0.080	20:54
Air Blank		0.000	20:54
Subject Sample	# 1	0.118	20:55
Breath Volume		2.160	20:55
Air Blank		0.000	20:55
Air Blank		0.000	20:57
Subject Sample	# 2	0.114	20:58
Breath Volume		2.058	20:58
Air Blank		0.000	20:58
Control Test		0.079	20:59
Air Blank		0.000	20:59
Diagnostics Check		OK	20:59

Data Download Date/Time: 06/28/2025 11:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/27/2025 01:10:26

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:03	
Subject Sample	# 1		01:06	No Sample Provided
Breath Volume		0.000	01:06	
Air Blank		0.000	01:06	
Air Blank		0.000	01:08	
Subject Sample	# 2		01:09	Subject Test Refused
Breath Volume		0.000	01:09	
Air Blank		0.000	01:09	
Control Test		0.080	01:09	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 06/28/2025 11:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 05/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MCCORKLE

DATE/TIME OF TEST: 05/30/2025 23:28:41

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:20	
Control Test		0.080	23:21	
Air Blank		0.000	23:21	
Subject Sample	# 1		23:24	No Sample Provided
Breath Volume		0.000	23:24	
Air Blank		0.000	23:25	
Air Blank		0.000	23:27	
Subject Sample	# 2		23:27	Subject Test Refused
Breath Volume		0.000	23:27	
Air Blank		0.000	23:27	
Control Test		0.081	23:28	
Air Blank		0.000	23:28	
Diagnostics Check		OK	23:28	

Data Download Date/Time: 06/28/2025 11:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DESMORE

DATE/TIME OF TEST: 05/31/2025 03:22:16

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.087	03:17
Breath Volume		1.406	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.090	03:20
Breath Volume		1.667	03:20
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 06/28/2025 11:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 05/31/2025 22:51:31

Diagnostics Check		OK	22:43	
Air Blank		0.000	22:43	
Control Test		0.080	22:43	
Air Blank		0.000	22:44	
Subject Sample	# 1		22:47	No Sample Provided
Breath Volume		0.000	22:47	
Air Blank		0.000	22:47	
Air Blank		0.000	22:49	
Subject Sample	# 2		22:50	Subject Test Refused
Breath Volume		0.000	22:50	
Air Blank		0.000	22:50	
Control Test		0.081	22:50	
Air Blank		0.000	22:51	
Diagnostics Check		OK	22:51	

Data Download Date/Time: 06/28/2025 11:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: HOLLOWELL HOWLE

DATE/TIME OF TEST: 05/24/2025 01:40:02

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.079	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.077	01:35
Breath Volume		3.414	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.078	01:38
Breath Volume		3.726	01:38
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:40

Data Download Date/Time: 06/25/2025 17:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,FELIX,J
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: CASTILLO,FELIX
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: ROSSIGNOL

DATE/TIME OF TEST: 05/24/2025 02:44:01

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:41	
Control Test		0.078	02:41	
Air Blank		0.000	02:42	
Subject Sample	# 1		02:42	Subject Test Refused
Breath Volume		0.000	02:42	
Air Blank		0.000	02:43	
Control Test		0.078	02:43	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 06/25/2025 17:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: AKRIDGE

DATE/TIME OF TEST: 05/24/2025 04:02:26

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.078	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.104	03:58
Breath Volume		1.445	03:58
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.109	04:00
Breath Volume		2.316	04:00
Air Blank		0.000	04:01
Control Test		0.079	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 06/25/2025 17:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 05/25/2025 03:16:22

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.078	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.137	03:11
Breath Volume		2.437	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.139	03:14
Breath Volume		2.636	03:14
Air Blank		0.000	03:15
Control Test		0.079	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 06/25/2025 17:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: DONEGAN

DATE/TIME OF TEST: 05/26/2025 03:11:45

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.078	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.000	03:07
Breath Volume		4.570	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.000	03:10
Breath Volume		3.941	03:10
Air Blank		0.000	03:10
Control Test		0.079	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 06/25/2025 17:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: PARKER,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 05/01/2025 00:14:47

Diagnostics Check		OK	00:01	
Air Blank		0.000	00:02	
Control Test		0.080	00:02	
Air Blank		0.000	00:03	
Subject Sample	# 1	0.213	00:04	
Breath Volume		2.058	00:04	
Air Blank		0.000	00:05	
Air Blank		0.000	00:07	
Subject Sample	# 2	0.209	00:10	Slope Not Level
Breath Volume		1.570	00:10	
Air Blank		0.000	00:10	
Air Blank		0.000	00:12	
Subject Sample	# 3	0.215	00:13	
Breath Volume		1.437	00:13	
Air Blank		0.000	00:13	
Control Test		0.079	00:14	
Air Blank		0.000	00:14	
Diagnostics Check		OK	00:14	

Data Download Date/Time: 05/11/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPANA,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: CAMPANA,NICHOLAS
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 05/03/2025 00:39:37

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.080	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.126	00:34
Breath Volume		2.398	00:34
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.122	00:38
Breath Volume		3.121	00:38
Air Blank		0.000	00:38
Control Test		0.079	00:39
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 05/11/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LISS,K,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: CURBELO,A
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: MAPLE

DATE/TIME OF TEST: 05/04/2025 01:09:23

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:56	
Control Test		0.080	00:56	
Air Blank		0.000	00:57	
Subject Sample	# 1		01:00	No Sample Provided
Breath Volume		0.000	01:00	
Air Blank		0.000	01:00	
Air Blank		0.000	01:02	
Subject Sample	# 2	0.213	01:04	
Breath Volume		1.808	01:04	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 3	0.210	01:07	
Breath Volume		1.179	01:07	
Air Blank		0.000	01:08	
Control Test		0.077	01:08	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

Data Download Date/Time: 05/11/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: BOLOGNA,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 05/04/2025 22:40:10

Diagnostics Check		OK	22:33
Air Blank		0.000	22:33
Control Test		0.080	22:34
Air Blank		0.000	22:34
Subject Sample	# 1	0.093	22:35
Breath Volume		1.500	22:35
Air Blank		0.000	22:36
Air Blank		0.000	22:37
Subject Sample	# 2	0.100	22:38
Breath Volume		2.996	22:38
Air Blank		0.000	22:39
Control Test		0.080	22:39
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 05/11/2025 22:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPANA,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: SHALA,ARBNOR
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: ZISCHKE

DATE/TIME OF TEST: 05/05/2025 15:57:35

Diagnostics Check		OK	15:51
Air Blank		0.000	15:51
Control Test		0.080	15:52
Air Blank		0.000	15:52
Subject Sample	# 1	0.000	15:53
Breath Volume		1.550	15:53
Air Blank		0.000	15:53
Air Blank		0.000	15:55
Subject Sample	# 2	0.000	15:56
Breath Volume		2.925	15:56
Air Blank		0.000	15:56
Control Test		0.080	15:56
Air Blank		0.000	15:57
Diagnostics Check		OK	15:57

Data Download Date/Time: 05/11/2025 22:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: PARKER,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HOOKER

DATE/TIME OF TEST: 05/10/2025 00:13:27

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.081	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.037	00:08
Breath Volume		2.562	00:08
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.034	00:11
Breath Volume		2.109	00:11
Air Blank		0.000	00:12
Control Test		0.079	00:12
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 05/11/2025 22:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: CURBELO,A
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 05/10/2025 00:46:57

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.120	00:42
Breath Volume		1.859	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:44
Subject Sample	# 2	0.123	00:45
Breath Volume		2.187	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 05/11/2025 22:05

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EXLINE,ANDREW,M
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: WILSON,ALLEN
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: TOSCHES

DATE/TIME OF TEST: 05/17/2025 11:21:55

Diagnostics Check		OK	11:15
Air Blank		0.000	11:16
Control Test		0.080	11:16
Air Blank		0.000	11:17
Subject Sample	# 1	0.168	11:17
Breath Volume		3.484	11:17
Air Blank		0.000	11:18
Air Blank		0.000	11:19
Subject Sample	# 2	0.168	11:20
Breath Volume		4.054	11:20
Air Blank		0.000	11:21
Control Test		0.079	11:21
Air Blank		0.000	11:21
Diagnostics Check		OK	11:21

Data Download Date/Time: 07/11/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS COLLAZO,OFFICER,J
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: BERRIOS COLLAZO,OFFICER
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 05/18/2025 03:10:03

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.204	03:05
Breath Volume		1.746	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.201	03:08
Breath Volume		1.472	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:10

Data Download Date/Time: 07/11/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPANA,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: CAMPANA,NICHOLAS
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: BARTLETT

DATE/TIME OF TEST: 05/24/2025 16:36:22

Diagnostics Check		OK	16:30
Air Blank		0.000	16:30
Control Test		0.080	16:31
Air Blank		0.000	16:31
Subject Sample	# 1	0.000	16:32
Breath Volume		2.644	16:32
Air Blank		0.000	16:32
Air Blank		0.000	16:34
Subject Sample	# 2	0.000	16:35
Breath Volume		2.898	16:35
Air Blank		0.000	16:35
Control Test		0.080	16:35
Air Blank		0.000	16:36
Diagnostics Check		OK	16:36

Data Download Date/Time: 07/11/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: FRAGA,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: QUITTER

DATE/TIME OF TEST: 05/25/2025 03:14:56

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.080	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.273	03:10
Breath Volume		1.550	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:12
Subject Sample	# 2	0.274	03:13
Breath Volume		1.203	03:13
Air Blank		0.000	03:14
Control Test		0.079	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 07/11/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HELLER,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SCHORDEN

DATE/TIME OF TEST: 05/26/2025 03:50:26

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.080	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.128	03:46
Breath Volume		2.519	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.122	03:48
Breath Volume		2.425	03:48
Air Blank		0.000	03:49
Control Test		0.078	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 07/11/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: PARKER,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: PEDRAZA-ALMARAZ

DATE/TIME OF TEST: 05/29/2025 20:43:06

Diagnostics Check		OK	20:40	
Air Blank		0.000	20:40	
Control Test		0.080	20:40	
Air Blank		0.000	20:41	
Subject Sample	# 1		20:41	RFI Detect
Breath Volume		0.000	20:41	
Air Blank		0.000	20:42	
Control Test		0.080	20:42	
Air Blank		0.000	20:43	
Diagnostics Check		OK	20:43	

Data Download Date/Time: 07/11/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAKER,J,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: PARKER,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: PEDRAZA-ALMARAZ

DATE/TIME OF TEST: 05/29/2025 20:56:45

Diagnostics Check		OK	20:45
Air Blank		0.000	20:45
Control Test		0.080	20:46
Air Blank		0.000	20:46
Subject Sample	# 1	0.215	20:49
Breath Volume		2.089	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:51
Subject Sample	# 2	0.190	20:52
Breath Volume		1.121	20:52
Air Blank		0.000	20:53
Air Blank		0.000	20:54
Subject Sample	# 3	0.212	20:55
Breath Volume		1.609	20:55
Air Blank		0.000	20:55
Control Test		0.079	20:56
Air Blank		0.000	20:56
Diagnostics Check		OK	20:56

Data Download Date/Time: 07/11/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: FRAGA

DATE/TIME OF TEST: 05/05/2025 22:58:15

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:55	
Control Test		0.080	22:56	
Air Blank		0.000	22:56	
Subject Sample	# 1		22:56	Subject Test Refused
Breath Volume		0.000	22:56	
Air Blank		0.000	22:57	
Control Test		0.080	22:57	
Air Blank		0.000	22:58	
Diagnostics Check		OK	22:58	

Data Download Date/Time: 05/21/2025 12:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: CITY OF MIRAMAR POL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OMORUYI,T,
BREATH TEST OPERATOR AGENCY: Miramar PD
ARRESTING OFFICER: OMORUYI,T
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 05/31/2025 03:18:08

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.076	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.204	03:13
Breath Volume		1.898	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.205	03:16
Breath Volume		1.148	03:16
Air Blank		0.000	03:17
Control Test		0.076	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 06/04/2025 11:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: CALIHUA PANZO

DATE/TIME OF TEST: 05/04/2025 00:31:13

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.081	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.189	00:26
Breath Volume		1.949	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.175	00:29
Breath Volume		1.875	00:29
Air Blank		0.000	00:30
Control Test		0.081	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 05/31/2025 03:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/04/2026
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 05/10/2025 20:02:35

Diagnostics Check		OK	19:55
Air Blank		0.000	19:56
Control Test		0.081	19:56
Air Blank		0.000	19:57
Subject Sample	# 1	0.105	19:57
Breath Volume		1.917	19:57
Air Blank		0.000	19:58
Air Blank		0.000	20:00
Subject Sample	# 2	0.100	20:01
Breath Volume		1.875	20:01
Air Blank		0.000	20:01
Control Test		0.081	20:01
Air Blank		0.000	20:02
Diagnostics Check		OK	20:02

Data Download Date/Time: 05/31/2025 03:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIE,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MILILLO,N
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: DEMPS

DATE/TIME OF TEST: 05/31/2025 20:31:11

Diagnostics Check		OK	20:23
Air Blank		0.000	20:24
Control Test		0.081	20:24
Air Blank		0.000	20:25
Subject Sample	# 1	0.325	20:26
Breath Volume		1.968	20:26
Air Blank		0.000	20:26
Air Blank		0.000	20:28
Subject Sample	# 2	0.330	20:29
Breath Volume		2.023	20:29
Air Blank		0.000	20:30
Control Test		0.081	20:30
Air Blank		0.000	20:31
Diagnostics Check		OK	20:31

Data Download Date/Time: 07/04/2025 00:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRINGTON,CODY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FARRINGTON,CODY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 05/04/2025 05:50:29

Diagnostics Check		OK	05:43
Air Blank		0.000	05:44
Control Test		0.080	05:44
Air Blank		0.000	05:44
Subject Sample	# 1	0.086	05:45
Breath Volume		1.878	05:45
Air Blank		0.000	05:46
Air Blank		0.000	05:48
Subject Sample	# 2	0.082	05:49
Breath Volume		1.410	05:49
Air Blank		0.000	05:49
Control Test		0.080	05:49
Air Blank		0.000	05:50
Diagnostics Check		OK	05:50

Data Download Date/Time: 06/03/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,ROBERT,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SNYDER,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MADSEN

DATE/TIME OF TEST: 05/07/2025 02:12:27

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.081	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.000	02:07
Breath Volume		4.039	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.000	02:11
Breath Volume		3.023	02:11
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 06/03/2025 11:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,WILLIAM,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: WATSON,WILLIAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: PEELE

DATE/TIME OF TEST: 05/11/2025 02:05:47

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.080	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.108	02:00
Breath Volume		1.375	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.112	02:04
Breath Volume		1.316	02:04
Air Blank		0.000	02:04
Control Test		0.080	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 06/03/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: ZAWADA

DATE/TIME OF TEST: 05/20/2025 19:42:39

Diagnostics Check		OK	19:36
Air Blank		0.000	19:36
Control Test		0.079	19:36
Air Blank		0.000	19:37
Subject Sample	# 1	0.244	19:37
Breath Volume		2.351	19:37
Air Blank		0.000	19:38
Air Blank		0.000	19:40
Subject Sample	# 2	0.237	19:41
Breath Volume		2.082	19:41
Air Blank		0.000	19:41
Control Test		0.080	19:42
Air Blank		0.000	19:42
Diagnostics Check		OK	19:42

Data Download Date/Time: 05/27/2025 10:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MELVIN,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: CONN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:23

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 05/09/2025 06:53:13

Diagnostics Check		OK	06:45
Air Blank		0.000	06:46
Control Test		0.079	06:46
Air Blank		0.000	06:47
Subject Sample	# 1	0.059	06:48
Breath Volume		2.914	06:48
Air Blank		0.000	06:49
Air Blank		0.000	06:51
Subject Sample	# 2	0.050	06:51
Breath Volume		1.437	06:51
Air Blank		0.000	06:52
Control Test		0.077	06:52
Air Blank		0.000	06:53
Diagnostics Check		OK	06:53

Data Download Date/Time: 05/30/2025 12:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BAUTISTA,JAVIER
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 05/14/2025 21:04:56

Diagnostics Check		OK	21:02	
Air Blank		0.000	21:02	
Control Test		0.079	21:02	
Air Blank		0.000	21:03	
Subject Sample	# 1		21:03	Subject Test Refused
Breath Volume		0.000	21:03	
Air Blank		0.000	21:04	
Control Test		0.078	21:04	
Air Blank		0.000	21:04	
Diagnostics Check		OK	21:04	

Data Download Date/Time: 05/30/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SUSTAITA,CRISTIAN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GARDUNO

DATE/TIME OF TEST: 05/18/2025 03:58:53

Diagnostics Check		OK	03:52
Air Blank		0.000	03:53
Control Test		0.079	03:53
Air Blank		0.000	03:54
Subject Sample	# 1	0.166	03:54
Breath Volume		2.203	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:56
Subject Sample	# 2	0.162	03:57
Breath Volume		3.042	03:57
Air Blank		0.000	03:57
Control Test		0.078	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 05/30/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,REX,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BRANDEBERRY,JEREMY
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: GUTIERREZ-PATLAN

DATE/TIME OF TEST: 05/21/2025 14:27:48

Diagnostics Check		OK	14:24	
Air Blank		0.000	14:24	
Control Test		0.078	14:25	
Air Blank		0.000	14:25	
Subject Sample	# 1		14:26	Subject Test Refused
Breath Volume		0.000	14:26	
Air Blank		0.000	14:26	
Control Test		0.078	14:27	
Air Blank		0.000	14:27	
Diagnostics Check		OK	14:27	

Data Download Date/Time: 05/30/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SORRELL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CISNEROS

DATE/TIME OF TEST: 05/24/2025 03:30:19

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:27	
Control Test		0.079	03:28	
Air Blank		0.000	03:28	
Subject Sample	# 1		03:28	Subject Test Refused
Breath Volume		0.000	03:28	
Air Blank		0.000	03:29	
Control Test		0.079	03:29	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 05/30/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOVEL,ADAM
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 05/24/2025 21:38:00

Diagnostics Check		OK	21:31
Air Blank		0.000	21:31
Control Test		0.080	21:31
Air Blank		0.000	21:32
Subject Sample	# 1	0.093	21:33
Breath Volume		2.000	21:33
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.093	21:36
Breath Volume		2.351	21:36
Air Blank		0.000	21:37
Control Test		0.078	21:37
Air Blank		0.000	21:37
Diagnostics Check		OK	21:38

Data Download Date/Time: 05/30/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOVEL,ADAM
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 12:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: PALACIOS

DATE/TIME OF TEST: 05/25/2025 00:58:33

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.079	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.206	00:54
Breath Volume		2.484	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.202	00:56
Breath Volume		2.410	00:56
Air Blank		0.000	00:57
Control Test		0.077	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 05/30/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MELVIN,F
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: CONN,CHRIS
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: WINTER

DATE/TIME OF TEST: 05/28/2025 09:09:29

Diagnostics Check		OK	09:00
Air Blank		0.000	09:00
Control Test		0.079	09:01
Air Blank		0.000	09:01
Subject Sample	# 1	0.101	09:04
Breath Volume		2.621	09:04
Air Blank		0.000	09:05
Air Blank		0.000	09:06
Subject Sample	# 2	0.083	09:07
Breath Volume		1.257	09:07
Air Blank		0.000	09:08
Control Test		0.077	09:08
Air Blank		0.000	09:09
Diagnostics Check		OK	09:09

Data Download Date/Time: 05/30/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTIEN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BRAMMER

DATE/TIME OF TEST: 05/01/2025 03:58:14

Diagnostics Check		OK	03:48	
Air Blank		0.000	03:49	
Control Test		0.080	03:49	
Air Blank		0.000	03:49	
Subject Sample	# 1		03:52	No Sample Provided
Breath Volume		0.000	03:52	
Air Blank		0.000	03:53	
Air Blank		0.000	03:55	
Subject Sample	# 2		03:56	Subject Test Refused
Breath Volume		0.000	03:56	
Air Blank		0.000	03:57	
Control Test		0.080	03:57	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:58	

Data Download Date/Time: 05/26/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGSON,ROLAND,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: LOUX,SAGE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: SALVAGNI

DATE/TIME OF TEST: 05/02/2025 18:04:37

Diagnostics Check		OK	17:57
Air Blank		0.000	17:58
Control Test		0.081	17:58
Air Blank		0.000	17:59
Subject Sample	# 1	0.000	18:00
Breath Volume		1.773	18:00
Air Blank		0.000	18:00
Air Blank		0.000	18:02
Subject Sample	# 2	0.000	18:03
Breath Volume		1.839	18:03
Air Blank		0.000	18:03
Control Test		0.080	18:04
Air Blank		0.000	18:04
Diagnostics Check		OK	18:04

Data Download Date/Time: 05/26/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGSON,ROLAND,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SINGSON,ROLAND
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: WIAND

DATE/TIME OF TEST: 05/03/2025 19:11:22

Diagnostics Check		OK	19:08	
Air Blank		0.000	19:08	
Control Test		0.081	19:09	
Air Blank		0.000	19:09	
Subject Sample	# 1		19:09	Subject Test Refused
Breath Volume		0.000	19:09	
Air Blank		0.000	19:10	
Control Test		0.080	19:10	
Air Blank		0.000	19:11	
Diagnostics Check		OK	19:11	

Data Download Date/Time: 05/26/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGSON,ROLAND,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SINGSON,ROLAND
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: WIAND

DATE/TIME OF TEST: 05/03/2025 19:31:54

Diagnostics Check		OK	19:25
Air Blank		0.000	19:26
Control Test		0.080	19:26
Air Blank		0.000	19:27
Subject Sample	# 1	0.000	19:27
Breath Volume		2.968	19:27
Air Blank		0.000	19:28
Air Blank		0.000	19:30
Subject Sample	# 2	0.000	19:30
Breath Volume		3.382	19:30
Air Blank		0.000	19:30
Control Test		0.080	19:31
Air Blank		0.000	19:31
Diagnostics Check		OK	19:31

Data Download Date/Time: 05/26/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 05/07/2025 22:13:42

Diagnostics Check		OK	22:07
Air Blank		0.000	22:08
Control Test		0.080	22:08
Air Blank		0.000	22:08
Subject Sample	# 1	0.213	22:09
Breath Volume		1.941	22:09
Air Blank		0.000	22:09
Air Blank		0.000	22:11
Subject Sample	# 2	0.219	22:12
Breath Volume		1.429	22:12
Air Blank		0.000	22:12
Control Test		0.080	22:13
Air Blank		0.000	22:13
Diagnostics Check		OK	22:13

Data Download Date/Time: 05/26/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: HANDY

DATE/TIME OF TEST: 05/08/2025 00:51:25

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:48	
Control Test		0.080	00:49	
Air Blank		0.000	00:49	
Subject Sample	# 1		00:50	Subject Test Refused
Breath Volume		0.000	00:50	
Air Blank		0.000	00:50	
Control Test		0.080	00:50	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 05/26/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERMAK,OFC,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: LOUX,SAGE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 05/10/2025 22:20:16

Diagnostics Check		OK	22:13
Air Blank		0.000	22:14
Control Test		0.080	22:14
Air Blank		0.000	22:15
Subject Sample	# 1	0.151	22:15
Breath Volume		1.187	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.152	22:18
Breath Volume		2.113	22:18
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 05/26/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERMAK,OFC,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: MARTINEZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: LORENZO

DATE/TIME OF TEST: 05/17/2025 04:02:57

Diagnostics Check		OK	03:59	
Air Blank		0.000	04:00	
Control Test		0.081	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 1		04:01	Subject Test Refused
Breath Volume		0.000	04:01	
Air Blank		0.000	04:02	
Control Test		0.080	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 05/26/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UPCHURCH,JERRY,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: EARLE,MARINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/25/2025 06:45:19

Diagnostics Check		OK	06:41	
Air Blank		0.000	06:41	
Control Test		0.080	06:42	
Air Blank		0.000	06:42	
Subject Sample	# 1		06:43	Subject Test Refused
Breath Volume		0.000	06:43	
Air Blank		0.000	06:44	
Control Test		0.080	06:44	
Air Blank		0.000	06:45	
Diagnostics Check		OK	06:45	

Data Download Date/Time: 05/26/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARDNER,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 04/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: WILKENING

DATE/TIME OF TEST: 05/25/2025 23:45:36

Diagnostics Check		OK	23:39
Air Blank		0.000	23:39
Control Test		0.080	23:40
Air Blank		0.000	23:40
Subject Sample	# 1	0.190	23:41
Breath Volume		2.128	23:41
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.180	23:44
Breath Volume		1.875	23:44
Air Blank		0.000	23:44
Control Test		0.080	23:45
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 05/26/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARDNER,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MARQUARDT

DATE/TIME OF TEST: 05/28/2025 04:25:44

Diagnostics Check		OK	04:22	
Air Blank		0.000	04:22	
Control Test		0.080	04:22	
Air Blank		0.000	04:23	
Subject Sample	# 1		04:24	Subject Test Refused
Breath Volume		0.000	04:24	
Air Blank		0.000	04:24	
Control Test		0.080	04:25	
Air Blank		0.000	04:25	
Diagnostics Check		OK	04:25	

Data Download Date/Time: 06/16/2025 09:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CASTANEDA CASTRO

DATE/TIME OF TEST: 05/01/2025 02:35:14

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.118	02:30
Breath Volume		2.394	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.110	02:33
Breath Volume		1.730	02:33
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 05/20/2025 05:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JANIAK,MALIKHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:56

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/01/2025 11:32:48

Diagnostics Check		OK	11:21	
Air Blank		0.000	11:21	
Control Test		0.080	11:22	
Air Blank		0.000	11:22	
Subject Sample	# 1		11:25	No Sample Provided
Breath Volume		0.000	11:25	
Air Blank		0.000	11:26	
Air Blank		0.000	11:28	
Subject Sample	# 2		11:31	No Sample Provided
Breath Volume		0.000	11:31	
Air Blank		0.000	11:31	
Control Test		0.081	11:32	
Air Blank		0.000	11:32	
Diagnostics Check		OK	11:32	

Data Download Date/Time: 05/20/2025 05:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: TIMBERLAKE

DATE/TIME OF TEST: 05/02/2025 05:45:35

Diagnostics Check		OK	05:38
Air Blank		0.000	05:38
Control Test		0.081	05:39
Air Blank		0.000	05:39
Subject Sample	# 1	0.116	05:40
Breath Volume		1.167	05:40
Air Blank		0.000	05:41
Air Blank		0.000	05:43
Subject Sample	# 2	0.116	05:43
Breath Volume		1.148	05:43
Air Blank		0.000	05:44
Control Test		0.079	05:44
Air Blank		0.000	05:45
Diagnostics Check		OK	05:45

Data Download Date/Time: 05/20/2025 05:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DYE,BRANDON
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 05/03/2025 01:36:08

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:25	
Control Test		0.081	01:25	
Air Blank		0.000	01:26	
Subject Sample	# 1		01:29	No Sample Provided
Breath Volume		0.000	01:29	
Air Blank		0.000	01:29	
Air Blank		0.000	01:31	
Subject Sample	# 2		01:34	No Sample Provided
Breath Volume		0.000	01:34	
Air Blank		0.000	01:35	
Control Test		0.082	01:35	
Air Blank		0.000	01:36	
Diagnostics Check		OK	01:36	

Data Download Date/Time: 05/20/2025 05:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DELMORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 05/03/2025 03:10:59

Diagnostics Check		OK	02:59	
Air Blank		0.000	03:00	
Control Test		0.081	03:00	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:04	No Sample Provided
Breath Volume		0.000	03:04	
Air Blank		0.000	03:04	
Air Blank		0.000	03:06	
Subject Sample	# 2		03:09	No Sample Provided
Breath Volume		0.000	03:09	
Air Blank		0.000	03:10	
Control Test		0.081	03:10	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 05/20/2025 05:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 05/03/2025 05:18:49

Diagnostics Check		OK	05:09
Air Blank		0.000	05:10
Control Test		0.081	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.097	05:13
Breath Volume		1.800	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.095	05:17
Breath Volume		1.718	05:17
Air Blank		0.000	05:17
Control Test		0.077	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 05/20/2025 05:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,RANDY,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WILSON,RANDY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 05/03/2025 06:30:51

Diagnostics Check	OK	06:29	
Air Blank	0.000	06:29	
Control Test	0.080	06:29	
Air Blank		06:30	RFI Detect
Air Blank	0.000	06:30	

Data Download Date/Time: 05/20/2025 05:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,RANDY,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WILSON,RANDY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 05/03/2025 06:39:38

Diagnostics Check		OK	06:33
Air Blank		0.000	06:33
Control Test		0.081	06:33
Air Blank		0.000	06:34
Subject Sample	# 1	0.000	06:34
Breath Volume		1.695	06:34
Air Blank		0.000	06:35
Air Blank		0.000	06:37
Subject Sample	# 2	0.000	06:38
Breath Volume		3.214	06:38
Air Blank		0.000	06:38
Control Test		0.080	06:39
Air Blank		0.000	06:39
Diagnostics Check		OK	06:39

Data Download Date/Time: 05/20/2025 05:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RIVERA-AMARO,JOSE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/05/2025 05:22:47

Diagnostics Check		OK	05:14
Air Blank		0.000	05:14
Control Test		0.080	05:15
Air Blank		0.000	05:15
Subject Sample	# 1	0.160	05:16
Breath Volume		1.648	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.163	05:21
Breath Volume		1.156	05:21
Air Blank		0.000	05:21
Control Test		0.080	05:22
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 05/20/2025 05:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,MALACHI,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RANDOLPH,MALACHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: CAMPBELL JR

DATE/TIME OF TEST: 05/05/2025 20:56:22

Diagnostics Check		OK	20:43	
Air Blank		0.000	20:43	
Control Test		0.081	20:43	
Air Blank		0.000	20:44	
Subject Sample	# 1	0.181	20:47	Volume Not Met
Breath Volume		0.945	20:47	
Air Blank		0.000	20:48	
Air Blank		0.000	20:50	
Subject Sample	# 2	0.172	20:51	
Breath Volume		1.296	20:51	
Air Blank		0.000	20:52	
Air Blank		0.000	20:54	
Subject Sample	# 3	0.173	20:54	
Breath Volume		1.289	20:54	
Air Blank		0.000	20:55	
Control Test		0.077	20:55	
Air Blank		0.000	20:56	
Diagnostics Check		OK	20:56	

Data Download Date/Time: 05/20/2025 05:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: JIRON GALEANO

DATE/TIME OF TEST: 05/05/2025 21:10:28

Diagnostics Check		OK	21:03
Air Blank		0.000	21:03
Control Test		0.079	21:03
Air Blank		0.000	21:04
Subject Sample	# 1	0.198	21:05
Breath Volume		2.000	21:05
Air Blank		0.000	21:05
Air Blank		0.000	21:07
Subject Sample	# 2	0.194	21:08
Breath Volume		2.222	21:08
Air Blank		0.000	21:09
Control Test		0.080	21:09
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 05/20/2025 05:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS IV,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: LAMOUR

DATE/TIME OF TEST: 05/06/2025 03:47:34

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.081	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.000	03:43
Breath Volume		1.156	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.000	03:46
Breath Volume		1.457	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 05/20/2025 05:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 05/06/2025 03:58:12

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.081	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.000	03:53
Breath Volume		2.023	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.000	03:56
Breath Volume		1.816	03:56
Air Blank		0.000	03:57
Control Test		0.082	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 05/20/2025 05:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 05/07/2025 01:14:06

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:03	
Control Test		0.080	01:03	
Air Blank		0.000	01:04	
Subject Sample	# 1		01:07	No Sample Provided
Breath Volume		0.000	01:07	
Air Blank		0.000	01:07	
Air Blank		0.000	01:09	
Subject Sample	# 2		01:12	No Sample Provided
Breath Volume		0.000	01:12	
Air Blank		0.000	01:13	
Control Test		0.080	01:13	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 05/20/2025 05:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: GUAMBANA MERCHAN

DATE/TIME OF TEST: 05/07/2025 23:12:34

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.080	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.056	23:08
Breath Volume		1.429	23:08
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.054	23:11
Breath Volume		1.300	23:11
Air Blank		0.000	23:11
Control Test		0.079	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 05/20/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JANIAK,MALIKHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:52

SUBJECT'S LAST NAME: WITTE

DATE/TIME OF TEST: 05/08/2025 14:22:11

Diagnostics Check		OK	14:15
Air Blank		0.000	14:15
Control Test		0.080	14:16
Air Blank		0.000	14:16
Subject Sample	# 1	0.168	14:17
Breath Volume		1.410	14:17
Air Blank		0.000	14:18
Air Blank		0.000	14:19
Subject Sample	# 2	0.157	14:20
Breath Volume		1.566	14:20
Air Blank		0.000	14:21
Control Test		0.079	14:21
Air Blank		0.000	14:22
Diagnostics Check		OK	14:22

Data Download Date/Time: 05/20/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MORALES SIMON

DATE/TIME OF TEST: 05/09/2025 01:07:49

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.184	01:03
Breath Volume		1.617	01:03
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.164	01:06
Breath Volume		1.390	01:06
Air Blank		0.000	01:06
Control Test		0.080	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 05/20/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BAMACA LOPEZ

DATE/TIME OF TEST: 05/09/2025 02:39:13

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.081	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.195	02:34
Breath Volume		1.382	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.198	02:37
Breath Volume		1.710	02:37
Air Blank		0.000	02:38
Control Test		0.080	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 05/20/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: CAMPO RONDON

DATE/TIME OF TEST: 05/10/2025 02:19:06

Diagnostics Check		OK	02:11
Air Blank		0.000	02:12
Control Test		0.080	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.177	02:14
Breath Volume		2.234	02:14
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.172	02:17
Breath Volume		1.820	02:17
Air Blank		0.000	02:18
Control Test		0.080	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 05/20/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 05/10/2025 03:03:55

Diagnostics Check		OK	02:52	
Air Blank		0.000	02:52	
Control Test		0.081	02:53	
Air Blank		0.000	02:53	
Subject Sample	# 1		02:56	No Sample Provided
Breath Volume		0.000	02:56	
Air Blank		0.000	02:57	
Air Blank		0.000	02:59	
Subject Sample	# 2		03:02	No Sample Provided
Breath Volume		0.000	03:02	
Air Blank		0.000	03:03	
Control Test		0.080	03:03	
Air Blank		0.000	03:03	
Diagnostics Check		OK	03:03	

Data Download Date/Time: 05/20/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/11/2025 02:02:01

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.077	01:57
Breath Volume		1.859	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.076	02:00
Breath Volume		2.042	02:00
Air Blank		0.000	02:01
Control Test		0.080	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:02

Data Download Date/Time: 05/20/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: RIVERA PENA

DATE/TIME OF TEST: 05/11/2025 05:18:55

Diagnostics Check		OK	05:12
Air Blank		0.000	05:13
Control Test		0.081	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.136	05:14
Breath Volume		1.128	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:16
Subject Sample	# 2	0.136	05:17
Breath Volume		1.890	05:17
Air Blank		0.000	05:17
Control Test		0.080	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 05/20/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VON PREYSING,RICHARD,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: VON PREYSING,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: CRUZ MARIN

DATE/TIME OF TEST: 05/11/2025 08:12:10

Diagnostics Check		OK	08:05
Air Blank		0.000	08:06
Control Test		0.080	08:06
Air Blank		0.000	08:07
Subject Sample	# 1	0.145	08:07
Breath Volume		2.117	08:07
Air Blank		0.000	08:08
Air Blank		0.000	08:10
Subject Sample	# 2	0.140	08:10
Breath Volume		2.292	08:10
Air Blank		0.000	08:11
Control Test		0.078	08:11
Air Blank		0.000	08:12
Diagnostics Check		OK	08:12

Data Download Date/Time: 05/20/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:01

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/12/2025 14:34:28

Diagnostics Check		OK	14:26
Air Blank		0.000	14:26
Control Test		0.080	14:26
Air Blank		0.000	14:27
Subject Sample	# 1	0.170	14:29
Breath Volume		1.449	14:29
Air Blank		0.000	14:30
Air Blank		0.000	14:31
Subject Sample	# 2	0.169	14:32
Breath Volume		1.625	14:32
Air Blank		0.000	14:33
Control Test		0.078	14:33
Air Blank		0.000	14:34
Diagnostics Check		OK	14:34

Data Download Date/Time: 05/20/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: STEWART,TRISTAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: SAMMONS

DATE/TIME OF TEST: 05/13/2025 04:04:33

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:53	
Control Test		0.080	03:53	
Air Blank		0.000	03:54	
Subject Sample	# 1		03:57	No Sample Provided
Breath Volume		0.000	03:57	
Air Blank		0.000	03:58	
Air Blank		0.000	04:00	
Subject Sample	# 2		04:03	No Sample Provided
Breath Volume		0.000	04:03	
Air Blank		0.000	04:03	
Control Test		0.080	04:03	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 05/20/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 05/13/2025 17:10:57

Diagnostics Check		OK	17:04
Air Blank		0.000	17:05
Control Test		0.080	17:05
Air Blank		0.000	17:05
Subject Sample	# 1	0.225	17:06
Breath Volume		2.996	17:06
Air Blank		0.000	17:07
Air Blank		0.000	17:08
Subject Sample	# 2	0.224	17:09
Breath Volume		2.714	17:09
Air Blank		0.000	17:10
Control Test		0.079	17:10
Air Blank		0.000	17:10
Diagnostics Check		OK	17:10

Data Download Date/Time: 05/20/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: SISCO JR

DATE/TIME OF TEST: 05/13/2025 18:55:10

Diagnostics Check		OK	18:48
Air Blank		0.000	18:49
Control Test		0.081	18:49
Air Blank		0.000	18:50
Subject Sample	# 1	0.027	18:50
Breath Volume		3.023	18:50
Air Blank		0.000	18:51
Air Blank		0.000	18:53
Subject Sample	# 2	0.028	18:53
Breath Volume		3.453	18:53
Air Blank		0.000	18:54
Control Test		0.081	18:54
Air Blank		0.000	18:55
Diagnostics Check		OK	18:55

Data Download Date/Time: 05/20/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: RYBURN

DATE/TIME OF TEST: 05/14/2025 22:59:47

Diagnostics Check		OK	22:53
Air Blank		0.000	22:54
Control Test		0.080	22:54
Air Blank		0.000	22:55
Subject Sample	# 1	0.000	22:55
Breath Volume		1.980	22:55
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.000	22:58
Breath Volume		1.656	22:58
Air Blank		0.000	22:58
Control Test		0.081	22:59
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 05/20/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: JEAN BAPTISTE

DATE/TIME OF TEST: 05/15/2025 03:29:49

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:18	
Control Test		0.081	03:19	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:22	No Sample Provided
Breath Volume		0.000	03:22	
Air Blank		0.000	03:23	
Air Blank		0.000	03:25	
Subject Sample	# 2		03:28	No Sample Provided
Breath Volume		0.000	03:28	
Air Blank		0.000	03:28	
Control Test		0.080	03:29	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 05/20/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: TOOLE

DATE/TIME OF TEST: 05/16/2025 00:01:19

Diagnostics Check		OK	23:54
Air Blank		0.000	23:54
Control Test		0.081	23:55
Air Blank		0.000	23:55
Subject Sample	# 1	0.029	23:56
Breath Volume		1.890	23:56
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.027	23:59
Breath Volume		1.636	23:59
Air Blank		0.000	00:00
Control Test		0.081	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 05/20/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/16/2025 02:26:27

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.081	02:15	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:19	No Sample Provided
Breath Volume		0.000	02:19	
Air Blank		0.000	02:19	
Air Blank		0.000	02:21	
Subject Sample	# 2		02:25	No Sample Provided
Breath Volume		0.000	02:25	
Air Blank		0.000	02:25	
Control Test		0.081	02:25	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 05/20/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: HAMELINE

DATE/TIME OF TEST: 05/16/2025 05:39:53

Diagnostics Check		OK	05:33
Air Blank		0.000	05:33
Control Test		0.081	05:33
Air Blank		0.000	05:34
Subject Sample	# 1	0.229	05:35
Breath Volume		2.207	05:35
Air Blank		0.000	05:35
Air Blank		0.000	05:37
Subject Sample	# 2	0.236	05:38
Breath Volume		1.562	05:38
Air Blank		0.000	05:38
Control Test		0.080	05:39
Air Blank		0.000	05:39
Diagnostics Check		OK	05:39

Data Download Date/Time: 05/20/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DYE,BRANDON
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 05/17/2025 01:54:08

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.101	01:48
Breath Volume		2.710	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.102	01:52
Breath Volume		1.449	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 05/20/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,EDGAR,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: APONTE,EDGAR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 05/17/2025 21:17:03

Diagnostics Check		OK	21:10
Air Blank		0.000	21:10
Control Test		0.081	21:11
Air Blank		0.000	21:11
Subject Sample	# 1	0.143	21:12
Breath Volume		4.496	21:12
Air Blank		0.000	21:13
Air Blank		0.000	21:14
Subject Sample	# 2	0.145	21:15
Breath Volume		4.730	21:15
Air Blank		0.000	21:16
Control Test		0.079	21:16
Air Blank		0.000	21:16
Diagnostics Check		OK	21:17

Data Download Date/Time: 05/20/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: FELICIANO GERBACIO

DATE/TIME OF TEST: 05/18/2025 02:23:24

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.081	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.169	02:18
Breath Volume		1.484	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.169	02:21
Breath Volume		1.664	02:21
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 05/20/2025 05:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDRON,DAVID,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 05/18/2025 06:22:15

Diagnostics Check		OK	06:15
Air Blank		0.000	06:16
Control Test		0.081	06:16
Air Blank		0.000	06:17
Subject Sample	# 1	0.146	06:17
Breath Volume		2.335	06:17
Air Blank		0.000	06:18
Air Blank		0.000	06:20
Subject Sample	# 2	0.141	06:20
Breath Volume		1.144	06:20
Air Blank		0.000	06:21
Control Test		0.080	06:21
Air Blank		0.000	06:22
Diagnostics Check		OK	06:22

Data Download Date/Time: 05/20/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: DEL TORO MONTANO

DATE/TIME OF TEST: 05/18/2025 20:29:34

Diagnostics Check		OK	20:18	
Air Blank		0.000	20:18	
Control Test		0.079	20:18	
Air Blank		0.000	20:19	
Subject Sample	# 1		20:22	No Sample Provided
Breath Volume		0.000	20:22	
Air Blank		0.000	20:23	
Air Blank		0.000	20:25	
Subject Sample	# 2		20:28	No Sample Provided
Breath Volume		0.000	20:28	
Air Blank		0.000	20:28	
Control Test		0.080	20:28	
Air Blank		0.000	20:29	
Diagnostics Check		OK	20:29	

Data Download Date/Time: 05/20/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOTEN,NATHANIEL,K
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WOOTEN,NATHANIEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: CANDELARIO MAYSONET

DATE/TIME OF TEST: 05/19/2025 06:02:57

Diagnostics Check		OK	05:51	
Air Blank		0.000	05:52	
Control Test		0.080	05:52	
Air Blank		0.000	05:52	
Subject Sample	# 1		05:55	No Sample Provided
Breath Volume		0.000	05:55	
Air Blank		0.000	05:56	
Air Blank		0.000	05:58	
Subject Sample	# 2		06:01	No Sample Provided
Breath Volume		0.000	06:01	
Air Blank		0.000	06:02	
Control Test		0.081	06:02	
Air Blank		0.000	06:02	
Diagnostics Check		OK	06:02	

Data Download Date/Time: 05/20/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/20/2025 22:25:30

Diagnostics Check		OK	22:19
Air Blank		0.000	22:19
Control Test		0.081	22:20
Air Blank		0.000	22:20
Subject Sample	# 1	0.340	22:20
Breath Volume		1.777	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.335	22:23
Breath Volume		1.660	22:23
Air Blank		0.000	22:24
Control Test		0.079	22:24
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 06/26/2025 05:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: VALENCIA

DATE/TIME OF TEST: 05/21/2025 01:26:54

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.000	01:22
Breath Volume		1.648	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.000	01:25
Breath Volume		1.804	01:25
Air Blank		0.000	01:25
Control Test		0.080	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 06/26/2025 05:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: DAVEY

DATE/TIME OF TEST: 05/23/2025 01:33:34

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.081	01:27
Air Blank		0.000	01:28
Subject Sample	# 1	0.094	01:28
Breath Volume		2.000	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.093	01:32
Breath Volume		1.875	01:32
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 06/26/2025 05:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: ANTONIO CRUZ

DATE/TIME OF TEST: 05/24/2025 02:59:48

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.081	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.171	02:54
Breath Volume		1.644	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.180	02:58
Breath Volume		2.078	02:58
Air Blank		0.000	02:58
Control Test		0.080	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 06/26/2025 05:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS IV,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 05/24/2025 04:46:06

Diagnostics Check		OK	04:34	
Air Blank		0.000	04:35	
Control Test		0.080	04:35	
Air Blank		0.000	04:36	
Subject Sample	# 1		04:39	No Sample Provided
Breath Volume		0.000	04:39	
Air Blank		0.000	04:39	
Air Blank		0.000	04:41	
Subject Sample	# 2		04:44	No Sample Provided
Breath Volume		0.000	04:44	
Air Blank		0.000	04:45	
Control Test		0.080	04:45	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 06/26/2025 05:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ESQUIVEL,STEPHANIE
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 05/24/2025 07:11:26

Diagnostics Check		OK	07:05
Air Blank		0.000	07:05
Control Test		0.080	07:05
Air Blank		0.000	07:06
Subject Sample	# 1	0.096	07:06
Breath Volume		2.062	07:06
Air Blank		0.000	07:07
Air Blank		0.000	07:09
Subject Sample	# 2	0.095	07:09
Breath Volume		4.035	07:09
Air Blank		0.000	07:10
Control Test		0.080	07:10
Air Blank		0.000	07:11
Diagnostics Check		OK	07:11

Data Download Date/Time: 06/26/2025 05:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: VICTORIANO HERNANDEZ

DATE/TIME OF TEST: 05/24/2025 22:38:43

Diagnostics Check		OK	22:32
Air Blank		0.000	22:32
Control Test		0.080	22:32
Air Blank		0.000	22:33
Subject Sample	# 1	0.260	22:34
Breath Volume		1.449	22:34
Air Blank		0.000	22:35
Air Blank		0.000	22:36
Subject Sample	# 2	0.251	22:37
Breath Volume		1.363	22:37
Air Blank		0.000	22:37
Control Test		0.079	22:38
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 06/26/2025 05:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,MALACHI,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RANDOLPH,MALACHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: SUMMERS

DATE/TIME OF TEST: 05/24/2025 23:15:32

Diagnostics Check		OK	23:04	
Air Blank		0.000	23:04	
Control Test		0.079	23:04	
Air Blank		0.000	23:05	
Subject Sample	# 1		23:08	No Sample Provided
Breath Volume		0.000	23:08	
Air Blank		0.000	23:09	
Air Blank		0.000	23:11	
Subject Sample	# 2		23:14	No Sample Provided
Breath Volume		0.000	23:14	
Air Blank		0.000	23:14	
Control Test		0.079	23:14	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 06/26/2025 05:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS IV,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: PABON

DATE/TIME OF TEST: 05/25/2025 03:00:05

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.054	02:55
Breath Volume		1.554	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.053	02:58
Breath Volume		1.496	02:58
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 06/26/2025 05:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 05/25/2025 22:45:29

Diagnostics Check		OK	22:29	
Air Blank		0.000	22:29	
Control Test		0.081	22:29	
Air Blank		0.000	22:30	
Subject Sample	# 1	0.097	22:32	
Breath Volume		1.175	22:32	
Air Blank		0.000	22:33	
Air Blank		0.000	22:35	
Subject Sample	# 2	0.089	22:38	Volume Not Met
Breath Volume		0.921	22:38	
Air Blank		0.000	22:39	
Air Blank		0.000	22:40	
Subject Sample	# 3	0.084	22:43	Volume Not Met
Breath Volume		0.226	22:43	
Air Blank		0.000	22:44	
Control Test		0.078	22:44	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 06/26/2025 05:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS IV,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MILLINGTON

DATE/TIME OF TEST: 05/25/2025 23:55:48

Diagnostics Check		OK	23:49
Air Blank		0.000	23:50
Control Test		0.079	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.299	23:51
Breath Volume		2.625	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:53
Subject Sample	# 2	0.302	23:54
Breath Volume		2.687	23:54
Air Blank		0.000	23:54
Control Test		0.079	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 06/26/2025 05:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ARMENTA

DATE/TIME OF TEST: 05/26/2025 00:16:12

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:05	
Control Test		0.079	00:05	
Air Blank		0.000	00:06	
Subject Sample	# 1		00:09	No Sample Provided
Breath Volume		0.000	00:09	
Air Blank		0.000	00:09	
Air Blank		0.000	00:11	
Subject Sample	# 2		00:14	No Sample Provided
Breath Volume		0.000	00:14	
Air Blank		0.000	00:15	
Control Test		0.080	00:15	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 06/26/2025 05:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: CARRILLO-CHAVEZ

DATE/TIME OF TEST: 05/26/2025 17:11:17

Diagnostics Check		OK	16:59	
Air Blank		0.000	17:00	
Control Test		0.080	17:00	
Air Blank		0.000	17:01	
Subject Sample	# 1		17:04	No Sample Provided
Breath Volume		0.000	17:04	
Air Blank		0.000	17:04	
Air Blank		0.000	17:06	
Subject Sample	# 2		17:09	No Sample Provided
Breath Volume		0.000	17:09	
Air Blank		0.000	17:10	
Control Test		0.080	17:10	
Air Blank		0.000	17:11	
Diagnostics Check		OK	17:11	

Data Download Date/Time: 06/26/2025 05:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: PHUONG

DATE/TIME OF TEST: 05/27/2025 01:46:55

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.081	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.192	01:41
Breath Volume		1.394	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.200	01:45
Breath Volume		1.593	01:45
Air Blank		0.000	01:45
Control Test		0.078	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 06/26/2025 05:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/28/2025 17:44:00

Diagnostics Check		OK	17:36
Air Blank		0.000	17:36
Control Test		0.079	17:36
Air Blank		0.000	17:37
Subject Sample	# 1	0.221	17:39
Breath Volume		1.964	17:39
Air Blank		0.000	17:39
Air Blank		0.000	17:41
Subject Sample	# 2	0.224	17:42
Breath Volume		1.753	17:42
Air Blank		0.000	17:43
Control Test		0.079	17:43
Air Blank		0.000	17:43
Diagnostics Check		OK	17:44

Data Download Date/Time: 06/26/2025 05:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 05/29/2025 02:07:22

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.141	02:03
Breath Volume		1.828	02:03
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.136	02:05
Breath Volume		1.582	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 06/26/2025 05:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,PHILIP,P
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: KING,PHILIP
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BRINSON

DATE/TIME OF TEST: 05/30/2025 00:40:05

Diagnostics Check		OK	00:29	
Air Blank		0.000	00:30	
Control Test		0.080	00:30	
Air Blank		0.000	00:30	
Subject Sample	# 1		00:34	No Sample Provided
Breath Volume		0.000	00:34	
Air Blank		0.000	00:34	
Air Blank		0.000	00:36	
Subject Sample	# 2		00:38	Subject Test Refused
Breath Volume		0.000	00:38	
Air Blank		0.000	00:39	
Control Test		0.080	00:39	
Air Blank		0.000	00:40	
Diagnostics Check		OK	00:40	

Data Download Date/Time: 06/26/2025 05:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: GATLIN

DATE/TIME OF TEST: 05/30/2025 03:54:13

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.081	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.165	03:49
Breath Volume		2.019	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.174	03:52
Breath Volume		1.949	03:52
Air Blank		0.000	03:53
Control Test		0.077	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 06/26/2025 05:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MARANO,ANTHONY
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: AVILES SALGADO

DATE/TIME OF TEST: 05/30/2025 20:42:53

Diagnostics Check		OK	20:36
Air Blank		0.000	20:37
Control Test		0.080	20:37
Air Blank		0.000	20:38
Subject Sample	# 1	0.000	20:38
Breath Volume		2.437	20:38
Air Blank		0.000	20:38
Air Blank		0.000	20:40
Subject Sample	# 2	0.000	20:41
Breath Volume		2.292	20:41
Air Blank		0.000	20:41
Control Test		0.080	20:42
Air Blank		0.000	20:42
Diagnostics Check		OK	20:42

Data Download Date/Time: 06/26/2025 05:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO,DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: GOMEZ-PEREZ

DATE/TIME OF TEST: 05/31/2025 00:02:29

Diagnostics Check		OK	23:50	
Air Blank		0.000	23:50	
Control Test		0.081	23:51	
Air Blank		0.000	23:51	
Subject Sample	# 2	0.133	23:54	Volume Not Met
Breath Volume		0.097	23:54	
Air Blank		0.000	23:55	
Air Blank		0.000	23:57	
Subject Sample	# 3	0.173	23:57	
Breath Volume		1.335	23:57	
Air Blank		0.000	23:58	
Air Blank		0.000	00:00	
Subject Sample	# 1	0.173	00:00	
Breath Volume		1.578	00:00	
Air Blank		0.000	00:01	
Control Test		0.078	00:01	
Air Blank		0.000	00:02	
Diagnostics Check		OK	00:02	

Data Download Date/Time: 06/26/2025 05:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/31/2025 04:24:43

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.081	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.020	04:20
Breath Volume		1.523	04:20
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.017	04:23
Breath Volume		1.300	04:23
Air Blank		0.000	04:23
Control Test		0.079	04:24
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 06/26/2025 05:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO,DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: TREJO

DATE/TIME OF TEST: 05/31/2025 05:21:01

Diagnostics Check		OK	05:08	
Air Blank		0.000	05:09	
Control Test		0.081	05:09	
Air Blank		0.000	05:10	
Subject Sample	# 1	0.097	05:10	
Breath Volume		1.230	05:10	
Air Blank		0.000	05:11	
Air Blank		0.000	05:13	
Subject Sample	# 2	0.085	05:16	Volume Not Met
Breath Volume		1.093	05:16	
Air Blank		0.000	05:17	
Air Blank		0.000	05:18	
Subject Sample	# 3	0.092	05:19	
Breath Volume		1.535	05:19	
Air Blank		0.000	05:20	
Control Test		0.078	05:20	
Air Blank		0.000	05:20	
Diagnostics Check		OK	05:21	

Data Download Date/Time: 06/26/2025 05:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 05/31/2025 16:48:24

Diagnostics Check		OK	16:36	
Air Blank		0.000	16:37	
Control Test		0.080	16:37	
Air Blank		0.000	16:38	
Subject Sample	# 1		16:41	No Sample Provided
Breath Volume		0.000	16:41	
Air Blank		0.000	16:41	
Air Blank		0.000	16:43	
Subject Sample	# 2		16:47	No Sample Provided
Breath Volume		0.000	16:47	
Air Blank		0.000	16:47	
Control Test		0.080	16:47	
Air Blank		0.000	16:48	
Diagnostics Check		OK	16:48	

Data Download Date/Time: 06/26/2025 05:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: GARCIA PEREZ

DATE/TIME OF TEST: 05/31/2025 22:03:19

Diagnostics Check		OK	21:51
Air Blank		0.000	21:52
Control Test		0.081	21:52
Air Blank		0.000	21:53
Subject Sample	# 1	0.253	21:54
Breath Volume		1.285	21:54
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.220	21:58
Breath Volume		1.339	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:00
Subject Sample	# 3	0.241	22:01
Breath Volume		1.585	22:01
Air Blank		0.000	22:02
Control Test		0.079	22:02
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 06/26/2025 05:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRANDO,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRANDO,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: TOUMBLARES

DATE/TIME OF TEST: 05/01/2025 22:00:16

Diagnostics Check		OK	21:57	
Air Blank		0.000	21:57	
Control Test		0.080	21:58	
Air Blank		0.000	21:58	
Subject Sample	# 1		21:58	Subject Test Refused
Breath Volume		0.000	21:58	
Air Blank		0.000	21:59	
Control Test		0.080	21:59	
Air Blank		0.000	22:00	
Diagnostics Check		OK	22:00	

Data Download Date/Time: 06/03/2025 08:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 05/02/2025 02:17:51

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.118	02:13
Breath Volume		1.664	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.121	02:16
Breath Volume		1.921	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 06/03/2025 08:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: ZELTMAN

DATE/TIME OF TEST: 05/04/2025 00:06:03

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:00
Subject Sample	# 1	0.217	00:01
Breath Volume		1.917	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:03
Subject Sample	# 2	0.227	00:04
Breath Volume		2.195	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:06

Data Download Date/Time: 06/03/2025 08:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ-GUILLEN,MARCOS,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LOPEZ-GUILLEN,MARCOS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:07

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 05/04/2025 12:27:05

Diagnostics Check		OK	12:21
Air Blank		0.000	12:21
Control Test		0.081	12:21
Air Blank		0.000	12:22
Subject Sample	# 1	0.000	12:22
Breath Volume		1.464	12:22
Air Blank		0.000	12:23
Air Blank		0.000	12:25
Subject Sample	# 2	0.000	12:25
Breath Volume		2.765	12:25
Air Blank		0.000	12:26
Control Test		0.080	12:26
Air Blank		0.000	12:27
Diagnostics Check		OK	12:27

Data Download Date/Time: 06/03/2025 08:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 05/05/2025 01:19:19

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.190	01:14
Breath Volume		1.484	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:16
Subject Sample	# 2	0.201	01:17
Breath Volume		1.320	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 06/03/2025 08:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOGELE,KEVIN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HERON,JULIEN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: PARRA

DATE/TIME OF TEST: 05/05/2025 01:43:26

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.170	01:38
Breath Volume		1.457	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:40
Subject Sample	# 2	0.156	01:41
Breath Volume		1.707	01:41
Air Blank		0.000	01:42
Control Test		0.077	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 06/03/2025 08:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: HARTZ

DATE/TIME OF TEST: 05/06/2025 18:53:11

Diagnostics Check		OK	18:46
Air Blank		0.000	18:47
Control Test		0.080	18:47
Air Blank		0.000	18:48
Subject Sample	# 1	0.158	18:48
Breath Volume		2.621	18:48
Air Blank		0.000	18:49
Air Blank		0.000	18:51
Subject Sample	# 2	0.146	18:51
Breath Volume		2.246	18:51
Air Blank		0.000	18:52
Control Test		0.080	18:52
Air Blank		0.000	18:53
Diagnostics Check		OK	18:53

Data Download Date/Time: 06/03/2025 08:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: DUNNING

DATE/TIME OF TEST: 05/12/2025 03:30:15

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.080	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.198	03:25
Breath Volume		3.214	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:27
Subject Sample	# 2	0.194	03:28
Breath Volume		2.835	03:28
Air Blank		0.000	03:29
Control Test		0.078	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 06/03/2025 08:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ-GUILLEN,MARCOS,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LOPEZ-GUILLEN,MARCOS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:46

SUBJECT'S LAST NAME: LINDSAY

DATE/TIME OF TEST: 05/16/2025 11:15:29

Diagnostics Check		OK	11:09
Air Blank		0.000	11:09
Control Test		0.079	11:10
Air Blank		0.000	11:10
Subject Sample	# 1	0.000	11:11
Breath Volume		5.691	11:11
Air Blank		0.000	11:11
Air Blank		0.000	11:13
Subject Sample	# 2	0.000	11:14
Breath Volume		3.726	11:14
Air Blank		0.000	11:14
Control Test		0.079	11:14
Air Blank		0.000	11:15
Diagnostics Check		OK	11:15

Data Download Date/Time: 06/03/2025 08:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOGELE,KEVIN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LEWIS,FRIEDRICH
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: CAPDEVILA

DATE/TIME OF TEST: 05/17/2025 03:56:35

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.093	03:51
Breath Volume		2.656	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.097	03:55
Breath Volume		2.261	03:55
Air Blank		0.000	03:55
Control Test		0.080	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 06/03/2025 08:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOGELE,KEVIN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: VITO,COURTNEY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 05/18/2025 02:10:17

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.081	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.135	02:05
Breath Volume		2.867	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.122	02:08
Breath Volume		2.363	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 06/03/2025 08:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 05/21/2025 02:28:06

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.194	02:22
Breath Volume		2.222	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.192	02:26
Breath Volume		2.261	02:26
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 06/03/2025 08:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNISTER,STEPHEN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BANNISTER,STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: KINGSLAND

DATE/TIME OF TEST: 05/23/2025 10:30:47

Diagnostics Check		OK	10:24
Air Blank		0.000	10:24
Control Test		0.081	10:24
Air Blank		0.000	10:25
Subject Sample	# 1	0.000	10:25
Breath Volume		1.941	10:25
Air Blank		0.000	10:26
Air Blank		0.000	10:28
Subject Sample	# 2	0.000	10:29
Breath Volume		1.894	10:29
Air Blank		0.000	10:29
Control Test		0.081	10:30
Air Blank		0.000	10:30
Diagnostics Check		OK	10:30

Data Download Date/Time: 06/03/2025 08:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: CERRATO

DATE/TIME OF TEST: 05/25/2025 22:29:07

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.080	22:23
Air Blank		0.000	22:23
Subject Sample	# 1	0.064	22:24
Breath Volume		1.617	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.063	22:27
Breath Volume		1.250	22:27
Air Blank		0.000	22:28
Control Test		0.080	22:28
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 06/03/2025 08:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JEFFERSON,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LONG,JEFFERSON
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 05/26/2025 04:05:32

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.080	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.149	03:59
Breath Volume		2.101	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.164	04:03
Breath Volume		2.441	04:03
Air Blank		0.000	04:04
Control Test		0.078	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 06/03/2025 08:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZIZ,D,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: AZIZ,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: DUVALL

DATE/TIME OF TEST: 05/26/2025 15:43:58

Diagnostics Check		OK	15:41	
Air Blank		0.000	15:41	
Control Test		0.080	15:42	
Air Blank		0.000	15:42	
Subject Sample	# 1		15:43	Sequence Aborted
Breath Volume		0.000	15:43	
Air Blank		0.000	15:43	

Data Download Date/Time: 06/03/2025 11:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZIZ,D,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: AZIZ,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: DUVALL

DATE/TIME OF TEST: 05/26/2025 15:59:21

Diagnostics Check		OK	15:53
Air Blank		0.000	15:53
Control Test		0.081	15:54
Air Blank		0.000	15:54
Subject Sample	# 1	0.000	15:55
Breath Volume		2.082	15:55
Air Blank		0.000	15:55
Air Blank		0.000	15:57
Subject Sample	# 2	0.000	15:57
Breath Volume		1.894	15:57
Air Blank		0.000	15:58
Control Test		0.079	15:58
Air Blank		0.000	15:59
Diagnostics Check		OK	15:59

Data Download Date/Time: 06/03/2025 11:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONS,DEPUTY,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SYMONS,E
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: J232085237047

DATE/TIME OF TEST: 05/27/2025 01:36:53

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.080	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.036	01:32
Breath Volume		1.656	01:32
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.035	01:35
Breath Volume		1.320	01:35
Air Blank		0.000	01:35
Control Test		0.080	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 06/03/2025 11:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOGELE,KEVIN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HERON,JULIEN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: CALLAHAN

DATE/TIME OF TEST: 05/31/2025 22:15:31

Diagnostics Check		OK	22:09
Air Blank		0.000	22:09
Control Test		0.080	22:10
Air Blank		0.000	22:10
Subject Sample	# 1	0.152	22:11
Breath Volume		1.804	22:11
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 2	0.155	22:13
Breath Volume		1.527	22:13
Air Blank		0.000	22:14
Control Test		0.078	22:14
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 06/03/2025 11:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,DESTINEE,R
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: GOMEZ,DESTINEE
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ANTOINE

DATE/TIME OF TEST: 05/22/2025 03:42:32

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.078	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.232	03:32
Breath Volume		1.753	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.203	03:36
Breath Volume		1.445	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 3	0.221	03:40
Breath Volume		1.222	03:40
Air Blank		0.000	03:41
Control Test		0.077	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 05/30/2025 10:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENSON,MAX,C
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: BEAUREGARD,JEFF
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: MIRAS

DATE/TIME OF TEST: 05/31/2025 19:07:10

Diagnostics Check		OK	19:00	
Air Blank		0.000	19:00	
Control Test		0.077	19:00	
Air Blank		0.000	19:01	
Subject Sample	# 1	0.136	19:02	
Breath Volume		1.929	19:02	
Air Blank		0.000	19:03	
Air Blank		0.000	19:04	
Subject Sample	# 2	0.136	19:05	
Breath Volume		1.503	19:05	
Air Blank		0.000	19:06	
Control Test		0.074	19:06	Control Outside Tolerance
Air Blank		0.000	19:07	

Data Download Date/Time: 06/30/2025 19:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENSON,MAX,C
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: BEAUREGARD,JEFF
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: MIRAS

DATE/TIME OF TEST: 05/31/2025 19:16:41

Diagnostics Check		OK	19:10
Air Blank		0.000	19:10
Control Test		0.075	19:11
Air Blank		0.000	19:11
Subject Sample	# 1	0.143	19:12
Breath Volume		1.613	19:12
Air Blank		0.000	19:12
Air Blank		0.000	19:14
Subject Sample	# 2	0.139	19:14
Breath Volume		1.625	19:14
Air Blank		0.000	19:15
Control Test		0.077	19:16
Air Blank		0.000	19:16
Diagnostics Check		OK	19:16

Data Download Date/Time: 06/30/2025 19:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROSSI,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: BLOUNT

DATE/TIME OF TEST: 05/02/2025 03:56:20

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.132	03:51
Breath Volume		1.550	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.138	03:54
Breath Volume		2.164	03:54
Air Blank		0.000	03:55
Control Test		0.078	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 05/13/2025 15:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 05/03/2025 00:43:49

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:41	
Control Test		0.079	00:41	
Air Blank		0.000	00:42	
Subject Sample	# 1		00:42	Subject Test Refused
Breath Volume		0.000	00:42	
Air Blank		0.000	00:42	
Control Test		0.078	00:43	
Air Blank		0.000	00:43	
Diagnostics Check		OK	00:43	

Data Download Date/Time: 05/13/2025 15:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/05/2025 23:43:34

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:41	
Control Test		0.078	23:41	
Air Blank		0.000	23:41	
Subject Sample	# 1		23:42	Subject Test Refused
Breath Volume		0.000	23:42	
Air Blank		0.000	23:42	
Control Test		0.078	23:42	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 05/13/2025 15:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SHIRK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/08/2025 05:58:18

Diagnostics Check		OK	05:50
Air Blank		0.000	05:50
Control Test		0.079	05:51
Air Blank		0.000	05:51
Subject Sample	# 1	0.097	05:53
Breath Volume		1.406	05:53
Air Blank		0.000	05:54
Air Blank		0.000	05:56
Subject Sample	# 2	0.094	05:56
Breath Volume		1.527	05:56
Air Blank		0.000	05:57
Control Test		0.078	05:57
Air Blank		0.000	05:58
Diagnostics Check		OK	05:58

Data Download Date/Time: 05/13/2025 15:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/11/2025 01:34:38

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.078	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 1		01:30	No Sample Provided
Breath Volume		0.000	01:30	
Air Blank		0.000	01:30	
Air Blank		0.000	01:32	
Subject Sample	# 2		01:33	Subject Test Refused
Breath Volume		0.000	01:33	
Air Blank		0.000	01:33	
Control Test		0.078	01:34	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 05/13/2025 15:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BURT,A
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 05/11/2025 02:54:51

Diagnostics Check		OK	02:52	
Air Blank		0.000	02:52	
Control Test		0.079	02:52	
Air Blank		0.000	02:53	
Subject Sample	# 1		02:53	Subject Test Refused
Breath Volume		0.000	02:53	
Air Blank		0.000	02:53	
Control Test		0.079	02:54	
Air Blank		0.000	02:54	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 05/13/2025 15:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 05/11/2025 05:56:57

Diagnostics Check		OK	05:48	
Air Blank		0.000	05:48	
Control Test		0.079	05:49	
Air Blank		0.000	05:49	
Subject Sample	# 1		05:52	No Sample Provided
Breath Volume		0.000	05:52	
Air Blank		0.000	05:53	
Air Blank		0.000	05:55	
Subject Sample	# 2		05:55	Subject Test Refused
Breath Volume		0.000	05:55	
Air Blank		0.000	05:56	
Control Test		0.078	05:56	
Air Blank		0.000	05:56	
Diagnostics Check		OK	05:56	

Data Download Date/Time: 05/13/2025 15:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: LANHAM

DATE/TIME OF TEST: 05/12/2025 01:24:39

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:22	
Control Test		0.078	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:23	
Control Test		0.079	01:24	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 05/13/2025 15:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: NEWKIRK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: CRAIN

DATE/TIME OF TEST: 05/12/2025 14:59:57

Diagnostics Check		OK	14:57	
Air Blank		0.000	14:57	
Control Test		0.079	14:57	
Air Blank		0.000	14:58	
Subject Sample	# 1		14:58	Subject Test Refused
Breath Volume		0.000	14:58	
Air Blank		0.000	14:59	
Control Test		0.078	14:59	
Air Blank		0.000	14:59	
Diagnostics Check		OK	14:59	

Data Download Date/Time: 05/13/2025 15:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: BATTIBULLI

DATE/TIME OF TEST: 05/14/2025 20:08:03

Diagnostics Check		OK	20:01
Air Blank		0.000	20:02
Control Test		0.078	20:02
Air Blank		0.000	20:03
Subject Sample	# 1	0.304	20:03
Breath Volume		2.367	20:03
Air Blank		0.000	20:04
Air Blank		0.000	20:05
Subject Sample	# 2	0.311	20:06
Breath Volume		2.351	20:06
Air Blank		0.000	20:07
Control Test		0.078	20:07
Air Blank		0.000	20:07
Diagnostics Check		OK	20:08

Data Download Date/Time: 06/08/2025 11:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: NUCKOLLS

DATE/TIME OF TEST: 05/24/2025 19:46:02

Diagnostics Check		OK	19:39
Air Blank		0.000	19:39
Control Test		0.079	19:40
Air Blank		0.000	19:40
Subject Sample	# 1	0.188	19:41
Breath Volume		2.078	19:41
Air Blank		0.000	19:41
Air Blank		0.000	19:43
Subject Sample	# 2	0.197	19:44
Breath Volume		2.558	19:44
Air Blank		0.000	19:45
Control Test		0.079	19:45
Air Blank		0.000	19:45
Diagnostics Check		OK	19:46

Data Download Date/Time: 06/08/2025 11:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ACQUAFREEDA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/14/2025 06:37:34

Diagnostics Check		OK	06:31
Air Blank		0.000	06:31
Control Test		0.081	06:32
Air Blank		0.000	06:32
Subject Sample	# 1	0.178	06:32
Breath Volume		2.429	06:32
Air Blank		0.000	06:33
Air Blank		0.000	06:35
Subject Sample	# 2	0.179	06:36
Breath Volume		2.960	06:36
Air Blank		0.000	06:36
Control Test		0.081	06:36
Air Blank		0.000	06:37
Diagnostics Check		OK	06:37

Data Download Date/Time: 05/30/2025 09:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHADL,CAMERON,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SCHADL,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: MAURER

DATE/TIME OF TEST: 05/19/2025 20:16:44

Diagnostics Check		OK	20:09
Air Blank		0.000	20:10
Control Test		0.082	20:10
Air Blank		0.000	20:11
Subject Sample	# 1	0.219	20:12
Breath Volume		2.917	20:12
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.219	20:15
Breath Volume		1.191	20:15
Air Blank		0.000	20:15
Control Test		0.081	20:16
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 05/30/2025 09:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACQUAFREDDA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ACQUAFREDDA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: BARTON

DATE/TIME OF TEST: 05/21/2025 05:10:34

Diagnostics Check		OK	05:07	
Air Blank		0.000	05:07	
Control Test		0.081	05:08	
Air Blank		0.000	05:08	
Subject Sample	# 1	0.000	05:09	
Breath Volume		3.902	05:09	
Air Blank			05:09	Improper Sample
Air Blank		0.000	05:09	
Control Test		0.081	05:09	
Air Blank		0.000	05:10	
Diagnostics Check		OK	05:10	

Data Download Date/Time: 05/30/2025 09:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACQUAFREDDA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ACQUAFREDDA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: COURTNEY

DATE/TIME OF TEST: 05/27/2025 00:54:42

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.082	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.090	00:50
Breath Volume		2.679	00:50
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.089	00:53
Breath Volume		2.531	00:53
Air Blank		0.000	00:53
Control Test		0.081	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 05/30/2025 09:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/13/2025 01:06:57

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.080	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.184	01:02
Breath Volume		3.019	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.166	01:05
Breath Volume		2.203	01:05
Air Blank		0.000	01:06
Control Test		0.079	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 06/05/2025 10:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEENEY,ERIC,K
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: FERPES,JAKE
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 04/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402635303
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: BAILEY-BROWN

DATE/TIME OF TEST: 05/10/2025 01:50:05

Diagnostics Check		OK	01:43
Air Blank		0.000	01:44
Control Test		0.078	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.173	01:45
Breath Volume		2.730	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.175	01:48
Breath Volume		2.726	01:48
Air Blank		0.000	01:49
Control Test		0.079	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:50

Data Download Date/Time: 05/20/2025 08:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,KELVIN,
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: VEGA,KELVIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 05/23/2025 05:39:07

Diagnostics Check		OK	05:32
Air Blank		0.000	05:32
Control Test		0.080	05:33
Air Blank		0.000	05:33
Subject Sample	# 1	0.231	05:34
Breath Volume		3.296	05:34
Air Blank		0.000	05:35
Air Blank		0.000	05:36
Subject Sample	# 2	0.221	05:37
Breath Volume		3.750	05:37
Air Blank		0.000	05:38
Control Test		0.078	05:38
Air Blank		0.000	05:39
Diagnostics Check		OK	05:39

Data Download Date/Time: 05/30/2025 21:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFOUR,HALEY,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SIMAS,MICHAEL
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 05/02/2025 03:14:35

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.080	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.231	03:09
Breath Volume		1.414	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:11
Subject Sample	# 2	0.231	03:13
Breath Volume		2.683	03:13
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 05/27/2025 14:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGECOMB,STEVEN,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: VICTOR,DERNESHA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: HUPPE

DATE/TIME OF TEST: 05/28/2025 00:35:44

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:28	
Control Test		0.080	00:28	
Air Blank		0.000	00:29	
Subject Sample	# 1	0.200	00:31	
Breath Volume		1.101	00:31	
Air Blank		0.000	00:31	
Air Blank		0.000	00:33	
Subject Sample	# 2	0.208	00:34	
Breath Volume		1.105	00:34	
Air Blank		0.000	00:34	
Control Test		0.077	00:34	
Air Blank			00:35	RFI Detect
Air Blank		0.000	00:35	

Data Download Date/Time: 06/30/2025 15:31

COBRA ©1997-2014 CMI, Inc.