Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAYSON,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SMITH,JAYSON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:42

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 06/01/2025 15:09:24

Diagnostics Check		OK	15:03
Air Blank		0.000	15:03
Control Test		0.079	15:04
Air Blank		0.000	15:04
Subject Sample	# 1	0.000	15:05
Breath Volume		2.039	15:05
Air Blank		0.000	15:05
Air Blank		0.000	15:07
Subject Sample	# 2	0.000	15:08
Breath Volume		2.203	15:08
Air Blank		0.000	15:08
Control Test		0.078	15:08
Air Blank		0.000	15:09
Diagnostics Check		OK	15:09

Data Download Date/Time: 06/04/2025 17:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEEHAN,MATTHEW,T
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEEHAN,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: CAROSH

DATE/TIME OF TEST: 06/04/2025 01:08:00

Diagnostics Check		OK	00:59	
Air Blank		0.000	00:59	
Control Test		0.078	01:00	
Air Blank		0.000	01:00	
Subject Sample	# 1		01:03	No Sample Provided
Breath Volume		0.000	01:03	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 2		01:06	Subject Test Refused
Breath Volume		0.000	01:06	
Air Blank		0.000	01:07	
Control Test		0.078	01:07	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 06/04/2025 17:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: RUIZ LARA

DATE/TIME OF TEST: 06/08/2025 03:23:39

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.079	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.259	03:19
Breath Volume		2.011	03:19
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.251	03:22
Breath Volume		2.085	03:22
Air Blank		0.000	03:22
Control Test		0.078	03:23
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 08/01/2025 07:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOX,GRAHAM,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SAEZ,VICTOR
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 06/10/2025 06:16:49

Diagnostics Check		OK	06:08
Air Blank		0.000	06:09
Control Test		0.079	06:09
Air Blank		0.000	06:10
Subject Sample	# 1	0.117	06:12
Breath Volume		3.898	06:12
Air Blank		0.000	06:12
Air Blank		0.000	06:14
Subject Sample	# 2	0.114	06:15
Breath Volume		3.921	06:15
Air Blank		0.000	06:15
Control Test		0.078	06:16
Air Blank		0.000	06:16
Diagnostics Check		OK	06:16

Data Download Date/Time: 08/01/2025 07:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REDZEPI,IMAN,
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: TRIEGLAFF,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 06/11/2025 04:16:48

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:14	
Control Test		0.079	04:14	
Air Blank		0.000	04:15	
Subject Sample	# 1		04:15	Subject Test Refused
Breath Volume		0.000	04:15	
Air Blank		0.000	04:15	
Control Test		0.079	04:16	
Air Blank		0.000	04:16	
Diagnostics Check		OK	04:16	

Data Download Date/Time: 08/01/2025 07:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: FARCHIONE

DATE/TIME OF TEST: 06/11/2025 21:38:49

Diagnostics Check		OK	21:32
Air Blank		0.000	21:32
Control Test		0.079	21:33
Air Blank		0.000	21:33
Subject Sample	# 1	0.071	21:34
Breath Volume		1.457	21:34
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 2	0.070	21:37
Breath Volume		1.488	21:37
Air Blank		0.000	21:37
Control Test		0.079	21:38
Air Blank		0.000	21:38
Diagnostics Check		OK	21:38

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/12/2025 01:36:57

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.135	01:32
Breath Volume		1.437	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:34
Subject Sample	# 2	0.138	01:35
Breath Volume		1.226	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: RODENBAUGH

DATE/TIME OF TEST: 06/13/2025 02:11:19

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:58	
Control Test		0.079	01:59	
Air Blank		0.000	01:59	
Subject Sample	# 1		02:02	No Sample Provided
Breath Volume		0.000	02:02	
Air Blank		0.000	02:03	
Air Blank		0.000	02:05	
Subject Sample	# 2	0.132	02:06	
Breath Volume		1.882	02:06	
Air Blank		0.000	02:06	
Air Blank		0.000	02:08	
Subject Sample	# 3	0.129	02:09	
Breath Volume		1.671	02:09	
Air Blank		0.000	02:10	
Control Test		0.078	02:10	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOOTH,MADISON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: VALENTINE

DATE/TIME OF TEST: 06/13/2025 22:50:21

Diagnostics Check	OK	22:49	
Air Blank		22:49	RFI Detect
Air Blank	0.000	22:50	

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOOTH,MADISON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: VALENTINE

DATE/TIME OF TEST: 06/13/2025 22:58:46

Diagnostics Check		OK	22:51
Air Blank		0.000	22:52
Control Test		0.079	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.000	22:53
Breath Volume		1.289	22:53
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.000	22:57
Breath Volume		1.996	22:57
Air Blank		0.000	22:57
Control Test		0.079	22:58
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOOTH,MADISON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: COLBASSANI

DATE/TIME OF TEST: 06/20/2025 03:56:14

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:53	
Control Test		0.078	03:54	
Air Blank		0.000	03:54	
Subject Sample	# 1		03:54	RFI Detect
Breath Volume		0.000	03:54	
Air Blank		0.000	03:55	
Control Test		0.079	03:55	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOOTH,MADISON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: COLBASSANI

DATE/TIME OF TEST: 06/20/2025 04:04:46

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.079	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.150	03:59
Breath Volume		1.207	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.159	04:03
Breath Volume		1.960	04:03
Air Blank		0.000	04:03
Control Test		0.077	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: CHAMBERLAIN

DATE/TIME OF TEST: 06/21/2025 01:22:27

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:20	
Control Test		0.080	01:20	
Air Blank		0.000	01:20	
Subject Sample	# 1		01:21	Subject Test Refused
Breath Volume		0.000	01:21	
Air Blank		0.000	01:21	
Control Test		0.080	01:21	
Air Blank		0.000	01:22	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 06/21/2025 03:52:55

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:50	
Control Test		0.078	03:50	
Air Blank		0.000	03:51	
Subject Sample	# 1		03:51	Subject Test Refused
Breath Volume		0.000	03:51	
Air Blank		0.000	03:52	
Control Test		0.079	03:52	
Air Blank		0.000	03:52	
Diagnostics Check		OK	03:52	

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 06/25/2025 18:52:03

Diagnostics Check		OK	18:46
Air Blank		0.000	18:46
Control Test		0.079	18:46
Air Blank		0.000	18:47
Subject Sample	# 1	0.000	18:47
Breath Volume		2.464	18:47
Air Blank		0.000	18:48
Air Blank		0.000	18:50
Subject Sample	# 2	0.000	18:50
Breath Volume		2.332	18:50
Air Blank		0.000	18:51
Control Test		0.078	18:51
Air Blank		0.000	18:51
Diagnostics Check		OK	18:52

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/26/2025 00:14:56

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.077	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.156	00:10
Breath Volume		1.460	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:12
Subject Sample	# 2	0.142	00:13
Breath Volume		1.515	00:13
Air Blank		0.000	00:14
Control Test		0.079	00:14
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOOTH,MADISON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SHARPE

DATE/TIME OF TEST: 06/28/2025 00:31:14

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:28	
Control Test		0.078	00:28	
Air Blank		0.000	00:29	
Subject Sample	# 1		00:29	Subject Test Refused
Breath Volume		0.000	00:29	
Air Blank		0.000	00:30	
Control Test		0.078	00:30	
Air Blank		0.000	00:31	
Diagnostics Check		OK	00:31	

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOOTH,MADISON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SHARPE

DATE/TIME OF TEST: 06/28/2025 00:59:54

Diagnostics Check		OK	00:53
Air Blank		0.000	00:53
Control Test		0.078	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.074	00:55
Breath Volume		1.335	00:55
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.073	00:58
Breath Volume		1.281	00:58
Air Blank		0.000	00:58
Control Test		0.079	00:59
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOOTH,MADISON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: HENKLE

DATE/TIME OF TEST: 06/29/2025 01:44:19

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.206	01:39
Breath Volume		1.757	01:39
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.200	01:42
Breath Volume		1.804	01:42
Air Blank		0.000	01:43
Control Test		0.076	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REDZEPI,IMAN,
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: KILIAN,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 06/30/2025 00:41:12

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:38	
Control Test		0.079	00:39	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:39	Subject Test Refused
Breath Volume		0.000	00:39	
Air Blank		0.000	00:40	
Control Test		0.079	00:40	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 08/01/2025 07:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASCKO,J,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:08

SUBJECT'S LAST NAME: INDERMAUR

DATE/TIME OF TEST: 06/01/2025 07:37:10

Diagnostics Check		OK	07:30
Air Blank		0.000	07:31
Control Test		0.081	07:31
Air Blank		0.000	07:31
Subject Sample	# 1	0.241	07:32
Breath Volume		3.101	07:32
Air Blank		0.000	07:33
Air Blank		0.000	07:34
Subject Sample	# 2	0.243	07:35
Breath Volume		2.894	07:35
Air Blank		0.000	07:36
Control Test		0.078	07:36
Air Blank		0.000	07:37
Diagnostics Check		OK	07:37

Data Download Date/Time: 07/04/2025 00:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHANDLER,RYAN,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CHANDLER,RYAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: KOEGER

DATE/TIME OF TEST: 06/10/2025 18:10:48

Diagnostics Check		OK	18:03
Air Blank		0.000	18:04
Control Test		0.082	18:04
Air Blank		0.000	18:05
Subject Sample	# 1	0.175	18:05
Breath Volume		1.593	18:05
Air Blank		0.000	18:06
Air Blank		0.000	18:08
Subject Sample	# 2	0.187	18:09
Breath Volume		1.628	18:09
Air Blank		0.000	18:09
Control Test		0.080	18:10
Air Blank		0.000	18:10
Diagnostics Check		OK	18:10

Data Download Date/Time: 07/04/2025 00:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,BRIAN,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: FROST,BRIAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: AIKEN

DATE/TIME OF TEST: 06/11/2025 15:45:32

Diagnostics Check		OK	15:38
Air Blank		0.000	15:39
Control Test		0.084	15:39
Air Blank		0.000	15:40
Subject Sample	# 1	0.000	15:41
Breath Volume		1.320	15:41
Air Blank		0.000	15:41
Air Blank		0.000	15:43
Subject Sample	# 2	0.000	15:44
Breath Volume		1.601	15:44
Air Blank		0.000	15:44
Control Test		0.082	15:44
Air Blank		0.000	15:45
Diagnostics Check		OK	15:45

Data Download Date/Time: 07/04/2025 00:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:29

SUBJECT'S LAST NAME: FRIEND

DATE/TIME OF TEST: 06/14/2025 17:56:02

Diagnostics Check		OK	17:49
Air Blank		0.000	17:50
Control Test		0.081	17:50
Air Blank		0.000	17:51
Subject Sample	# 1	0.074	17:51
Breath Volume		4.113	17:51
Air Blank		0.000	17:52
Air Blank		0.000	17:54
Subject Sample	# 2	0.069	17:54
Breath Volume		3.812	17:54
Air Blank		0.000	17:55
Control Test		0.081	17:55
Air Blank		0.000	17:55
Diagnostics Check		OK	17:56

Data Download Date/Time: 07/04/2025 00:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIE,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: WOODIE,L
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BUCK

DATE/TIME OF TEST: 06/14/2025 22:30:28

Diagnostics Check		OK	22:22	
Air Blank		0.000	22:23	
Control Test		0.083	22:23	
Air Blank		0.000	22:24	
Subject Sample	# 1	0.243	22:25	
Breath Volume		1.785	22:25	
Air Blank		0.000	22:26	
Air Blank		0.000	22:28	
Subject Sample	# 2	0.239	22:28	
Breath Volume		1.234	22:28	
Air Blank		0.000	22:29	
Control Test		0.074	22:29	Control Outside Tolerance
Air Blank		0.000	22:30	

Data Download Date/Time: 07/04/2025 00:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DEUSEN,COLE,C
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VAN DEUSEN,COLE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 06/21/2025 00:39:10

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:28	
Control Test		0.082	00:28	
Air Blank		0.000	00:29	
Subject Sample	# 1	0.194	00:32	Volume Not Met
Breath Volume		0.085	00:32	
Air Blank		0.000	00:32	
Air Blank		0.000	00:34	
Subject Sample	# 2		00:37	No Sample Provided
Breath Volume		0.000	00:37	
Air Blank		0.000	00:38	
Control Test		0.077	00:38	
Air Blank		0.000	00:39	
Diagnostics Check		OK	00:39	

Data Download Date/Time: 07/04/2025 00:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DEUSEN,COLE,C
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VAN DEUSEN,COLE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 06/21/2025 01:09:42

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:54	
Control Test		0.080	00:54	
Air Blank		0.000	00:55	
Subject Sample	# 1	0.207	00:58	Volume Not Met
Breath Volume		0.957	00:58	
Air Blank		0.000	00:59	
Air Blank		0.000	01:01	
Subject Sample	# 2	0.211	01:02	
Breath Volume		2.093	01:02	
Air Blank		0.000	01:03	
Air Blank		0.000	01:04	
Subject Sample	# 3	0.133	01:07	Volume Not Met
Breath Volume		0.347	01:07	
Air Blank		0.000	01:08	
Control Test		0.074	01:09	Control Outside Tolerance
Air Blank		0.000	01:09	

Data Download Date/Time: 07/04/2025 00:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROBINSON,DAVIS
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: DAWLEY

DATE/TIME OF TEST: 06/01/2025 01:48:49

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.081	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.105	01:43
Breath Volume		1.855	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.106	01:47
Breath Volume		1.972	01:47
Air Blank		0.000	01:47
Control Test		0.081	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 06/01/2025 02:30:46

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:27	
Control Test		0.081	02:27	
Air Blank		0.000	02:27	
Subject Sample	# 1		02:29	Subject Test Refused
Breath Volume		0.000	02:29	
Air Blank		0.000	02:29	
Control Test		0.082	02:30	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DICKERSON,JOHN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 06/01/2025 21:03:57

Diagnostics Check		OK	20:57
Air Blank		0.000	20:57
Control Test		0.082	20:58
Air Blank		0.000	20:58
Subject Sample	# 1	0.171	20:59
Breath Volume		2.156	20:59
Air Blank		0.000	21:00
Air Blank		0.000	21:01
Subject Sample	# 2	0.167	21:02
Breath Volume		2.269	21:02
Air Blank		0.000	21:03
Control Test		0.080	21:03
Air Blank		0.000	21:03
Diagnostics Check		OK	21:03

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: MERDIA

DATE/TIME OF TEST: 06/01/2025 23:46:49

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.147	23:42
Breath Volume		1.296	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:44
Subject Sample	# 2	0.143	23:45
Breath Volume		1.433	23:45
Air Blank		0.000	23:45
Control Test		0.080	23:46
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,BROOKE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: BLANCHARD

DATE/TIME OF TEST: 06/02/2025 02:39:10

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.130	02:34
Breath Volume		2.195	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.133	02:37
Breath Volume		2.472	02:37
Air Blank		0.000	02:38
Control Test		0.081	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GERMUSKA,LAWRENCE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: PECCERELLI

DATE/TIME OF TEST: 06/02/2025 06:42:25

Diagnostics Check		OK	06:34
Air Blank		0.000	06:34
Control Test		0.081	06:35
Air Blank		0.000	06:35
Subject Sample	# 1	0.179	06:37
Breath Volume		1.984	06:37
Air Blank		0.000	06:37
Air Blank		0.000	06:39
Subject Sample	# 2	0.177	06:40
Breath Volume		2.378	06:40
Air Blank		0.000	06:41
Control Test		0.080	06:41
Air Blank		0.000	06:42
Diagnostics Check		OK	06:42

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: TERRAZA MARCOS

DATE/TIME OF TEST: 06/02/2025 22:36:27

Diagnostics Check		OK	22:30
Air Blank		0.000	22:30
Control Test		0.081	22:31
Air Blank		0.000	22:31
Subject Sample	# 1	0.026	22:32
Breath Volume		2.824	22:32
Air Blank		0.000	22:32
Air Blank		0.000	22:34
Subject Sample	# 2	0.026	22:34
Breath Volume		2.816	22:34
Air Blank		0.000	22:35
Control Test		0.081	22:35
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: ARELLANO

DATE/TIME OF TEST: 06/02/2025 23:32:36

Diagnostics Check		OK	23:26
Air Blank		0.000	23:26
Control Test		0.081	23:27
Air Blank		0.000	23:27
Subject Sample	# 1	0.017	23:28
Breath Volume		1.167	23:28
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.018	23:31
Breath Volume		1.589	23:31
Air Blank		0.000	23:31
Control Test		0.081	23:32
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: GIRALDO

DATE/TIME OF TEST: 06/03/2025 00:25:06

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.082	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.000	00:20
Breath Volume		3.039	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.000	00:23
Breath Volume		3.722	00:23
Air Blank		0.000	00:24
Control Test		0.081	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: DE PALMA

DATE/TIME OF TEST: 06/03/2025 02:55:55

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.081	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.014	02:50
Breath Volume		1.382	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.013	02:54
Breath Volume		2.406	02:54
Air Blank		0.000	02:55
Control Test		0.082	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: JIMENEZ PEREZ

DATE/TIME OF TEST: 06/05/2025 23:59:33

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.081	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.200	23:55
Breath Volume		1.660	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.190	23:57
Breath Volume		2.359	23:57
Air Blank		0.000	23:58
Control Test		0.081	23:58
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: ANTONELLI

DATE/TIME OF TEST: 06/06/2025 01:59:26

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.081	01:53
Air Blank		0.000	01:54
Subject Sample	# 1	0.188	01:54
Breath Volume		1.378	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.199	01:57
Breath Volume		2.156	01:57
Air Blank		0.000	01:58
Control Test		0.081	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIER,MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 06/06/2025 04:26:06

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.081	04:20
Air Blank		0.000	04:20
Subject Sample	# 1	0.270	04:21
Breath Volume		3.320	04:21
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.268	04:24
Breath Volume		2.773	04:24
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: ENNIS

DATE/TIME OF TEST: 06/06/2025 18:42:33

Diagnostics Check		OK	18:36
Air Blank		0.000	18:36
Control Test		0.081	18:36
Air Blank		0.000	18:37
Subject Sample	# 1	0.233	18:38
Breath Volume		1.863	18:38
Air Blank		0.000	18:38
Air Blank		0.000	18:40
Subject Sample	# 2	0.238	18:40
Breath Volume		2.035	18:40
Air Blank		0.000	18:41
Control Test		0.080	18:41
Air Blank		0.000	18:42
Diagnostics Check		OK	18:42

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: RAFAEL GASPAR

DATE/TIME OF TEST: 06/06/2025 23:37:22

Diagnostics Check		OK	23:30
Air Blank		0.000	23:31
Control Test		0.081	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.189	23:32
Breath Volume		1.265	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.183	23:35
Breath Volume		1.503	23:35
Air Blank		0.000	23:36
Control Test		0.081	23:36
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXANDER,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CHAJ CHAJ

DATE/TIME OF TEST: 06/07/2025 00:14:30

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.081	00:07
Air Blank		0.000	00:07
Subject Sample	# 1	0.153	00:09
Breath Volume		1.515	00:09
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.164	00:12
Breath Volume		2.519	00:12
Air Blank		0.000	00:13
Control Test		0.080	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: CHAJ CHAJ

DATE/TIME OF TEST: 06/07/2025 00:42:19

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.080	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.197	00:36
Breath Volume		1.804	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.207	00:40
Breath Volume		2.367	00:40
Air Blank		0.000	00:41
Control Test		0.081	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: ALONZO PASCUAL

DATE/TIME OF TEST: 06/07/2025 00:57:59

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.081	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.122	00:53
Breath Volume		1.445	00:53
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.128	00:56
Breath Volume		1.695	00:56
Air Blank		0.000	00:57
Control Test		0.080	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZIMMERMAN,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: VELASQUEZ NAVARRO

DATE/TIME OF TEST: 06/07/2025 05:21:18

Diagnostics Check		OK	05:14
Air Blank		0.000	05:15
Control Test		0.081	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.147	05:16
Breath Volume		3.156	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.140	05:19
Breath Volume		2.929	05:19
Air Blank		0.000	05:20
Control Test		0.081	05:20
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ,ALINA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 06/07/2025 21:00:05

Diagnostics Check		OK	20:53
Air Blank		0.000	20:53
Control Test		0.081	20:53
Air Blank		0.000	20:54
Subject Sample	# 1	0.275	20:55
Breath Volume		2.511	20:55
Air Blank		0.000	20:56
Air Blank		0.000	20:57
Subject Sample	# 2	0.267	20:58
Breath Volume		2.441	20:58
Air Blank		0.000	20:59
Control Test		0.080	20:59
Air Blank		0.000	20:59
Diagnostics Check		OK	21:00

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 06/08/2025 22:31:44

Diagnostics Check		OK	22:25
Air Blank		0.000	22:26
Control Test		0.081	22:26
Air Blank		0.000	22:27
Subject Sample	# 1	0.000	22:27
Breath Volume		1.269	22:27
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.000	22:30
Breath Volume		2.199	22:30
Air Blank		0.000	22:30
Control Test		0.082	22:31
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: LERA

DATE/TIME OF TEST: 06/09/2025 01:30:58

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.081	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.286	01:21
Breath Volume		1.992	01:21
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.230	01:25
Breath Volume		1.382	01:25
Air Blank		0.000	01:26
Air Blank		0.000	01:27
Subject Sample	# 3	0.270	01:29
Breath Volume		1.847	01:29
Air Blank		0.000	01:30
Control Test		0.078	01:30
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: LOUIDOR

DATE/TIME OF TEST: 06/09/2025 01:44:52

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.000	01:40
Breath Volume		2.066	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.000	01:43
Breath Volume		1.746	01:43
Air Blank		0.000	01:43
Control Test		0.080	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JIMENEZ,LLEN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: SAAVEDRA MARTINEZ

DATE/TIME OF TEST: 06/09/2025 19:15:28

Diagnostics Check		OK	19:08
Air Blank		0.000	19:08
Control Test		0.081	19:09
Air Blank		0.000	19:09
Subject Sample	# 1	0.110	19:09
Breath Volume		1.121	19:09
Air Blank		0.000	19:10
Air Blank		0.000	19:12
Subject Sample	# 2	0.105	19:13
Breath Volume		1.453	19:13
Air Blank		0.000	19:14
Control Test		0.081	19:14
Air Blank		0.000	19:15
Diagnostics Check		OK	19:15

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DICKERSON,JOHN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: BERRIOS RUIZ

DATE/TIME OF TEST: 06/11/2025 02:12:13

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.082	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.178	02:05
Breath Volume		1.140	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.177	02:10
Breath Volume		1.617	02:10
Air Blank		0.000	02:11
Control Test		0.081	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COOPER,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: LINO

DATE/TIME OF TEST: 06/12/2025 03:47:16

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.081	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.111	03:40
Breath Volume		1.273	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.115	03:45
Breath Volume		1.535	03:45
Air Blank		0.000	03:46
Control Test		0.081	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SHOWALTER

DATE/TIME OF TEST: 06/13/2025 02:25:55

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.081	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.155	02:21
Breath Volume		2.988	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.149	02:24
Breath Volume		1.902	02:24
Air Blank		0.000	02:25
Control Test		0.081	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PAXTOR VASQUEZ

DATE/TIME OF TEST: 06/13/2025 04:05:28

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.080	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.181	04:00
Breath Volume		1.648	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:02
Subject Sample	# 2	0.180	04:03
Breath Volume		2.085	04:03
Air Blank		0.000	04:04
Control Test		0.080	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: LIVERMAN

DATE/TIME OF TEST: 06/13/2025 23:56:30

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.081	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.089	23:52
Breath Volume		1.886	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.087	23:54
Breath Volume		1.781	23:54
Air Blank		0.000	23:55
Control Test		0.081	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ,ALINA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: VILLASANA

DATE/TIME OF TEST: 06/14/2025 23:30:50

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:20	
Control Test		0.082	23:21	
Air Blank		0.000	23:21	
Subject Sample	# 1	0.113	23:24	Volume Not Met
Breath Volume		0.804	23:24	
Air Blank		0.000	23:25	
Air Blank		0.000	23:27	
Subject Sample	# 2		23:29	Subject Test Refused
Breath Volume		0.894	23:29	
Air Blank		0.000	23:29	
Control Test		0.081	23:30	
Air Blank		0.000	23:30	
Diagnostics Check		OK	23:30	

Data Download Date/Time: 06/20/2025 16:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/15/2025 02:20:24

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:09	
Control Test		0.081	02:09	
Air Blank		0.000	02:10	
Subject Sample	# 1	0.155	02:13	Volume Not Met
Breath Volume		0.351	02:13	
Air Blank		0.000	02:13	
Air Blank		0.000	02:15	
Subject Sample	# 2	0.146	02:18	Volume Not Met
Breath Volume		0.101	02:18	
Air Blank		0.000	02:19	
Control Test		0.079	02:19	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:20	

Data Download Date/Time: 06/20/2025 16:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ALI

DATE/TIME OF TEST: 06/15/2025 03:56:36

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.082	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.125	03:52
Breath Volume		1.773	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.130	03:55
Breath Volume		2.617	03:55
Air Blank		0.000	03:55
Control Test		0.081	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 06/20/2025 16:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: CASILLAS HERNANDEZ

DATE/TIME OF TEST: 06/15/2025 23:11:05

Diagnostics Check		OK	23:04
Air Blank		0.000	23:04
Control Test		0.082	23:04
Air Blank		0.000	23:05
Subject Sample	# 1	0.294	23:06
Breath Volume		1.187	23:06
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.278	23:09
Breath Volume		1.191	23:09
Air Blank		0.000	23:10
Control Test		0.080	23:10
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 06/20/2025 16:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LONDONO,JULIAN
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: PETIT FRERE

DATE/TIME OF TEST: 06/16/2025 00:13:57

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.081	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.226	00:09
Breath Volume		2.035	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:11
Subject Sample	# 2	0.230	00:12
Breath Volume		1.988	00:12
Air Blank		0.000	00:13
Control Test		0.080	00:13
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 06/20/2025 16:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RINI,LAWRENCE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: WOOLF

DATE/TIME OF TEST: 06/17/2025 18:36:11

Diagnostics Check		OK	18:30
Air Blank		0.000	18:30
Control Test		0.082	18:30
Air Blank		0.000	18:31
Subject Sample	# 1	0.105	18:31
Breath Volume		1.945	18:31
Air Blank		0.000	18:32
Air Blank		0.000	18:34
Subject Sample	# 2	0.107	18:34
Breath Volume		3.417	18:34
Air Blank		0.000	18:35
Control Test		0.081	18:35
Air Blank		0.000	18:36
Diagnostics Check		OK	18:36

Data Download Date/Time: 06/20/2025 16:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,BROOKE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 06/18/2025 00:21:28

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.081	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.357	00:16
Breath Volume		1.480	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:18
Subject Sample	# 2	0.344	00:19
Breath Volume		1.273	00:19
Air Blank		0.000	00:20
Control Test		0.080	00:20
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 06/20/2025 16:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 06/18/2025 22:12:39

Diagnostics Check		OK	22:06
Air Blank		0.000	22:06
Control Test		0.081	22:07
Air Blank		0.000	22:07
Subject Sample	# 1	0.000	22:08
Breath Volume		3.191	22:08
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.000	22:11
Breath Volume		3.308	22:11
Air Blank		0.000	22:11
Control Test		0.082	22:12
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 06/20/2025 16:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAPPIN,KEVIN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 06/19/2025 00:23:43

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.082	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.000	00:19
Breath Volume		1.457	00:19
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.000	00:22
Breath Volume		2.292	00:22
Air Blank		0.000	00:22
Control Test		0.082	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 06/20/2025 16:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KENNY,SHANE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:09

SUBJECT'S LAST NAME: DIPIERNO

DATE/TIME OF TEST: 06/19/2025 16:43:56

Diagnostics Check		OK	16:37
Air Blank		0.000	16:37
Control Test		0.081	16:38
Air Blank		0.000	16:38
Subject Sample	# 1	0.116	16:39
Breath Volume		1.808	16:39
Air Blank		0.000	16:39
Air Blank		0.000	16:41
Subject Sample	# 2	0.119	16:42
Breath Volume		3.035	16:42
Air Blank		0.000	16:43
Control Test		0.081	16:43
Air Blank		0.000	16:43
Diagnostics Check		OK	16:43

Data Download Date/Time: 06/20/2025 16:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MAIER,MARTINA
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MARTINEZ JUAREZ

DATE/TIME OF TEST: 06/19/2025 22:10:12

Diagnostics Check		OK	22:02
Air Blank		0.000	22:03
Control Test		0.081	22:03
Air Blank		0.000	22:04
Subject Sample	# 1	0.220	22:05
Breath Volume		2.093	22:05
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.217	22:08
Breath Volume		1.496	22:08
Air Blank		0.000	22:09
Control Test		0.080	22:09
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 06/20/2025 16:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 06/20/2025 05:35:36

Diagnostics Check		OK	05:29
Air Blank		0.000	05:29
Control Test		0.081	05:29
Air Blank		0.000	05:30
Subject Sample	# 1	0.130	05:31
Breath Volume		1.484	05:31
Air Blank		0.000	05:31
Air Blank		0.000	05:33
Subject Sample	# 2	0.129	05:34
Breath Volume		1.917	05:34
Air Blank		0.000	05:34
Control Test		0.081	05:34
Air Blank		0.000	05:35
Diagnostics Check		OK	05:35

Data Download Date/Time: 06/20/2025 16:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARES,RAMEZ,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: BERNAL FIERRO

DATE/TIME OF TEST: 06/01/2025 00:44:13

Diagnostics Check		OK	00:29	
Air Blank		0.000	00:30	
Control Test		0.079	00:30	
Air Blank		0.000	00:30	
Subject Sample	# 1	0.065	00:34	Volume Not Met
Breath Volume		0.304	00:34	
Air Blank		0.000	00:34	
Air Blank		0.000	00:36	
Subject Sample	# 2	0.066	00:37	
Breath Volume		1.703	00:37	
Air Blank		0.000	00:38	
Air Blank		0.000	00:40	
Subject Sample	# 3	0.070	00:42	
Breath Volume		1.320	00:42	
Air Blank		0.000	00:43	
Control Test		0.074	00:43	Control Outside Tolerance
Air Blank		0.000	00:44	

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARES,RAMEZ,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: BERNAL FIERRO

DATE/TIME OF TEST: 06/01/2025 01:04:01

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:50	
Control Test		0.078	00:51	
Air Blank		0.000	00:51	
Subject Sample	# 1	0.060	00:54	Volume Not Met
Breath Volume		0.531	00:54	
Air Blank		0.000	00:55	
Air Blank		0.000	00:57	
Subject Sample	# 2	0.070	00:57	
Breath Volume		1.152	00:57	
Air Blank		0.000	00:58	
Air Blank		0.000	01:00	
Subject Sample	# 3	0.061	01:02	
Breath Volume		1.144	01:02	
Air Blank		0.000	01:03	
Control Test		0.075	01:03	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:04	

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,LOGAN,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MINCA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PACHECO-ARITA

DATE/TIME OF TEST: 06/01/2025 03:47:40

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:36	
Control Test		0.079	03:36	
Air Blank		0.000	03:36	
Subject Sample	# 1	0.190	03:40	Volume Not Met
Breath Volume		0.277	03:40	
Air Blank		0.000	03:40	
Air Blank		0.000	03:42	
Subject Sample	# 2	0.226	03:45	Volume Not Met
Breath Volume		0.355	03:45	
Air Blank		0.000	03:46	
Control Test		0.075	03:46	
Air Blank		0.000	03:47	
Diagnostics Check		OK	03:47	

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,LOGAN,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: OLIVARES,ASHLEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: RIVERA-ZAVALA

DATE/TIME OF TEST: 06/01/2025 04:52:08

Diagnostics Check		OK	04:45
Air Blank		0.000	04:46
Control Test		0.078	04:46
Air Blank		0.000	04:47
Subject Sample	# 1	0.152	04:47
Breath Volume		1.699	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.148	04:50
Breath Volume		2.179	04:50
Air Blank		0.000	04:51
Control Test		0.077	04:51
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/02/2025 21:29:41

Diagnostics Check		OK	21:22
Air Blank		0.000	21:23
Control Test		0.079	21:23
Air Blank		0.000	21:24
Subject Sample	# 1	0.206	21:25
Breath Volume		2.332	21:25
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.206	21:28
Breath Volume		2.222	21:28
Air Blank		0.000	21:28
Control Test		0.077	21:29
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,LOGAN,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: SEBAST

DATE/TIME OF TEST: 06/04/2025 22:56:07

Diagnostics Check		OK	22:49
Air Blank		0.000	22:49
Control Test		0.079	22:50
Air Blank		0.000	22:50
Subject Sample	# 1	0.221	22:51
Breath Volume		3.250	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:53
Subject Sample	# 2	0.208	22:54
Breath Volume		3.226	22:54
Air Blank		0.000	22:55
Control Test		0.075	22:55
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: BROACH

DATE/TIME OF TEST: 06/07/2025 00:26:55

Diagnostics Check		OK	00:19
Air Blank		0.000	00:20
Control Test		0.080	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.182	00:21
Breath Volume		1.628	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.166	00:25
Breath Volume		2.152	00:25
Air Blank		0.000	00:25
Control Test		0.078	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,LOGAN,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FARES,RAMEZ
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MEYERS

DATE/TIME OF TEST: 06/11/2025 00:52:31

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.147	00:47
Breath Volume		2.792	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.141	00:50
Breath Volume		2.035	00:50
Air Blank		0.000	00:51
Control Test		0.078	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARES,RAMEZ,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SCHAEFER

DATE/TIME OF TEST: 06/14/2025 02:37:44

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.144	02:32
Breath Volume		1.449	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.134	02:35
Breath Volume		1.730	02:35
Air Blank		0.000	02:36
Control Test		0.078	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,LOGAN,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: DELACRUZ,JONATHAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: FIEFIA

DATE/TIME OF TEST: 06/16/2025 00:17:29

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.079	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.170	00:12
Breath Volume		1.625	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.160	00:15
Breath Volume		1.480	00:15
Air Blank		0.000	00:16
Control Test		0.078	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 06/20/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: BENITEZ FLORES

DATE/TIME OF TEST: 06/20/2025 23:21:17

Diagnostics Check		OK	23:14
Air Blank		0.000	23:15
Control Test		0.080	23:15
Air Blank		0.000	23:15
Subject Sample	# 1	0.085	23:16
Breath Volume		1.140	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.083	23:19
Breath Volume		1.160	23:19
Air Blank		0.000	23:20
Control Test		0.078	23:20
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 07/15/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNUTSON,TRISTAN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KNUTSON,TRISTAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: CONVERSANO

DATE/TIME OF TEST: 06/22/2025 01:28:13

Diagnostics Check		OK	01:19
Air Blank		0.000	01:20
Control Test		0.079	01:20
Air Blank		0.000	01:20
Subject Sample	# 1	0.092	01:23
Breath Volume		1.226	01:23
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.101	01:26
Breath Volume		2.070	01:26
Air Blank		0.000	01:27
Control Test		0.078	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 07/15/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JONATHAN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FARES,RAMEZ
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: PEREZ BAIL

DATE/TIME OF TEST: 06/23/2025 22:32:30

Diagnostics Check		OK	22:25
Air Blank		0.000	22:26
Control Test		0.079	22:26
Air Blank		0.000	22:27
Subject Sample	# 1	0.157	22:27
Breath Volume		1.464	22:27
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.158	22:30
Breath Volume		1.437	22:30
Air Blank		0.000	22:31
Control Test		0.077	22:31
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 07/15/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERYCKEN,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VERYCKEN,MATTHEW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: KARPENKO

DATE/TIME OF TEST: 06/25/2025 00:31:53

Diagnostics Check	OK	00:31	
Air Blank		00:31	RFI Detect
Air Blank	0.000	00:31	

Data Download Date/Time: 07/15/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERYCKEN,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VERYCKEN,MATTHEW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: KARPENKO

DATE/TIME OF TEST: 06/25/2025 00:39:27

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.118	00:34
Breath Volume		1.753	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.106	00:37
Breath Volume		1.621	00:37
Air Blank		0.000	00:38
Control Test		0.078	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 07/15/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERYCKEN,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BUNDY

DATE/TIME OF TEST: 06/25/2025 02:27:46

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:18	
Control Test		0.079	02:18	
Air Blank		0.000	02:19	
Subject Sample	# 1	0.000	02:21	
Breath Volume		1.628	02:21	
Air Blank		0.000	02:21	
Air Blank		0.000	02:23	
Subject Sample	# 2	0.000	02:26	
Breath Volume		1.503	02:26	
Air Blank		0.000	02:26	
Control Test		0.076	02:26	
Air Blank			02:27	RFI Detect
Air Blank		0.000	02:27	

Data Download Date/Time: 07/15/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,LOGAN,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: ABERNATHY

DATE/TIME OF TEST: 06/27/2025 23:21:38

Diagnostics Check		OK	23:15
Air Blank		0.000	23:15
Control Test		0.080	23:16
Air Blank		0.000	23:16
Subject Sample	# 1	0.097	23:17
Breath Volume		1.304	23:17
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.088	23:19
Breath Volume		1.472	23:19
Air Blank		0.000	23:20
Control Test		0.077	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 07/15/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARES,RAMEZ,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FARES,RAMEZ
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MURILLO

DATE/TIME OF TEST: 06/28/2025 02:58:19

Diagnostics Check		OK	02:52	
Air Blank		0.000	02:53	
Control Test		0.080	02:53	
Air Blank		0.000	02:53	
Subject Sample	# 1		02:56	Subject Test Refused
Breath Volume		0.000	02:56	
Air Blank		0.000	02:57	
Control Test		0.080	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 07/15/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHILDRESS,TIMOTHY,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CHILDRESS,TIMOTHY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 10:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: MAGGI

DATE/TIME OF TEST: 06/28/2025 12:16:49

Diagnostics Check		OK	12:10
Air Blank		0.000	12:11
Control Test		0.079	12:11
Air Blank		0.000	12:11
Subject Sample	# 1	0.182	12:12
Breath Volume		2.988	12:12
Air Blank		0.000	12:13
Air Blank		0.000	12:14
Subject Sample	# 2	0.173	12:15
Breath Volume		1.667	12:15
Air Blank		0.000	12:15
Control Test		0.078	12:16
Air Blank		0.000	12:16
Diagnostics Check		OK	12:16

Data Download Date/Time: 07/15/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARES,RAMEZ,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FARES,RAMEZ
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: GROVES

DATE/TIME OF TEST: 06/28/2025 20:26:01

Diagnostics Check		OK	20:16
Air Blank		0.000	20:17
Control Test		0.079	20:17
Air Blank		0.000	20:18
Subject Sample	# 1	0.160	20:19
Breath Volume		1.683	20:19
Air Blank		0.000	20:20
Air Blank		0.000	20:22
Subject Sample	# 2	0.146	20:24
Breath Volume		1.437	20:24
Air Blank		0.000	20:25
Control Test		0.076	20:25
Air Blank		0.000	20:25
Diagnostics Check		OK	20:26

Data Download Date/Time: 07/15/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELACRUZ,JONATHAN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VERYCKEN,MATTHEW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: EATON

DATE/TIME OF TEST: 06/28/2025 22:52:53

Diagnostics Check		OK	22:45
Air Blank		0.000	22:46
Control Test		0.078	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.228	22:48
Breath Volume		1.750	22:48
Air Blank		0.000	22:49
Air Blank		0.000	22:50
Subject Sample	# 2	0.226	22:51
Breath Volume		1.164	22:51
Air Blank		0.000	22:51
Control Test		0.076	22:52
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 07/15/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERYCKEN,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: DELACRUZ,JONATHAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: METZ

DATE/TIME OF TEST: 06/29/2025 00:07:53

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.058	00:03
Breath Volume		2.441	00:03
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.056	00:06
Breath Volume		2.289	00:06
Air Blank		0.000	00:06
Control Test		0.078	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 07/15/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARES,RAMEZ,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/29/2025 01:40:00

Diagnostics Check		OK	01:33	
Air Blank		0.000	01:34	
Control Test		0.079	01:34	
Air Blank		0.000	01:35	
Subject Sample	# 1	0.202	01:35	
Breath Volume		2.296	01:35	
Air Blank		0.000	01:36	
Air Blank		0.000	01:38	
Subject Sample	# 2	0.205	01:38	
Breath Volume		2.847	01:38	
Air Blank			01:39	RFI Detect
Air Blank		0.000	01:40	

Data Download Date/Time: 07/15/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARES,RAMEZ,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/29/2025 01:49:17

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.079	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.203	01:44
Breath Volume		3.195	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.203	01:47
Breath Volume		2.796	01:47
Air Blank		0.000	01:48
Control Test		0.077	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 07/15/2025 12:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/02/2025 02:53:57

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:51	
Control Test		0.081	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:53	
Control Test		0.081	02:53	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 06/11/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,C,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRY,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:06

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 06/07/2025 07:34:18

Diagnostics Check		OK	07:26
Air Blank		0.000	07:27
Control Test		0.081	07:27
Air Blank		0.000	07:27
Subject Sample	# 1	0.191	07:29
Breath Volume		1.921	07:29
Air Blank		0.000	07:30
Air Blank		0.000	07:31
Subject Sample	# 2	0.194	07:32
Breath Volume		2.023	07:32
Air Blank		0.000	07:33
Control Test		0.080	07:33
Air Blank		0.000	07:34
Diagnostics Check		OK	07:34

Data Download Date/Time: 06/11/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 06/10/2025 03:35:47

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:33	
Control Test		0.081	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:34	Subject Test Refused
Breath Volume		0.000	03:34	
Air Blank		0.000	03:34	
Control Test		0.081	03:35	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 06/11/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEIS,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: WILDES

DATE/TIME OF TEST: 06/11/2025 23:39:43

Diagnostics Check		OK	23:33
Air Blank		0.000	23:34
Control Test		0.082	23:34
Air Blank		0.000	23:34
Subject Sample	# 1	0.008	23:35
Breath Volume		2.996	23:35
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.007	23:38
Breath Volume		3.085	23:38
Air Blank		0.000	23:38
Control Test		0.082	23:39
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 07/10/2025 13:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLADDEN,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAYES,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:21

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/12/2025 10:48:04

Diagnostics Check		OK	10:41
Air Blank		0.000	10:42
Control Test		0.081	10:42
Air Blank		0.000	10:43
Subject Sample	# 1	0.366	10:43
Breath Volume		1.804	10:43
Air Blank		0.000	10:44
Air Blank		0.000	10:46
Subject Sample	# 2	0.359	10:46
Breath Volume		2.871	10:46
Air Blank		0.000	10:47
Control Test		0.080	10:47
Air Blank		0.000	10:47
Diagnostics Check		OK	10:48

Data Download Date/Time: 07/10/2025 13:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: SMOLYAKOVA

DATE/TIME OF TEST: 06/13/2025 04:39:54

Diagnostics Check		OK	04:33
Air Blank		0.000	04:33
Control Test		0.082	04:33
Air Blank		0.000	04:34
Subject Sample	# 1	0.130	04:35
Breath Volume		1.425	04:35
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.131	04:38
Breath Volume		2.136	04:38
Air Blank		0.000	04:38
Control Test		0.082	04:39
Air Blank		0.000	04:39
Diagnostics Check		OK	04:39

Data Download Date/Time: 07/10/2025 13:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LITTLE,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/13/2025 05:30:43

Diagnostics Check		OK	05:24
Air Blank		0.000	05:24
Control Test		0.081	05:25
Air Blank		0.000	05:25
Subject Sample	# 1	0.085	05:26
Breath Volume		1.195	05:26
Air Blank		0.000	05:27
Air Blank		0.000	05:28
Subject Sample	# 2	0.084	05:29
Breath Volume		1.265	05:29
Air Blank		0.000	05:29
Control Test		0.082	05:30
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 07/10/2025 13:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLADDEN,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RICHARDSON,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:19

SUBJECT'S LAST NAME: ADDISON

DATE/TIME OF TEST: 06/13/2025 08:47:50

Diagnostics Check		OK	08:40
Air Blank		0.000	08:40
Control Test		0.082	08:40
Air Blank		0.000	08:41
Subject Sample	# 1	0.131	08:42
Breath Volume		1.347	08:42
Air Blank		0.000	08:43
Air Blank		0.000	08:45
Subject Sample	# 2	0.132	08:46
Breath Volume		1.535	08:46
Air Blank		0.000	08:46
Control Test		0.082	08:47
Air Blank		0.000	08:47
Diagnostics Check		OK	08:47

Data Download Date/Time: 07/10/2025 13:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: NYE

DATE/TIME OF TEST: 06/14/2025 00:24:55

Diagnostics Check		OK	00:22	
Air Blank		0.000	00:22	
Control Test		0.082	00:22	
Air Blank		0.000	00:23	
Subject Sample	# 1		00:23	Subject Test Refused
Breath Volume		0.000	00:23	
Air Blank		0.000	00:23	
Control Test		0.081	00:24	
Air Blank		0.000	00:24	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 07/10/2025 13:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/15/2025 00:17:50

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.081	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.182	00:12
Breath Volume		1.246	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.191	00:16
Breath Volume		1.738	00:16
Air Blank		0.000	00:16
Control Test		0.081	00:17
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 07/10/2025 13:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,R,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: GRANADOS

DATE/TIME OF TEST: 06/15/2025 02:59:20

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.082	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.211	02:54
Breath Volume		3.289	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.209	02:57
Breath Volume		3.378	02:57
Air Blank		0.000	02:58
Control Test		0.081	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: THEOPHILE

DATE/TIME OF TEST: 06/15/2025 22:40:24

Diagnostics Check		OK	22:34
Air Blank		0.000	22:34
Control Test		0.081	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.214	22:35
Breath Volume		2.406	22:35
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.219	22:38
Breath Volume		2.753	22:38
Air Blank		0.000	22:39
Control Test		0.081	22:39
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/16/2025 02:12:46

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.082	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.129	02:07
Breath Volume		1.347	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.120	02:11
Breath Volume		1.605	02:11
Air Blank		0.000	02:11
Control Test		0.081	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: HOLENA

DATE/TIME OF TEST: 06/19/2025 04:45:51

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:43	
Control Test		0.081	04:43	
Air Blank		0.000	04:44	
Subject Sample	# 1		04:44	Subject Test Refused
Breath Volume		0.000	04:44	
Air Blank		0.000	04:44	
Control Test		0.081	04:45	
Air Blank		0.000	04:45	
Diagnostics Check		OK	04:45	

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 06/20/2025 02:00:30

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:58	
Control Test		0.081	01:58	
Air Blank		0.000	01:58	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	01:59	
Control Test		0.081	01:59	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGAN,D,O
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:09

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/22/2025 09:37:15

Diagnostics Check		OK	09:30
Air Blank		0.000	09:31
Control Test		0.081	09:31
Air Blank		0.000	09:32
Subject Sample	# 1	0.067	09:32
Breath Volume		2.511	09:32
Air Blank		0.000	09:33
Air Blank		0.000	09:35
Subject Sample	# 2	0.065	09:35
Breath Volume		2.441	09:35
Air Blank		0.000	09:36
Control Test		0.081	09:36
Air Blank		0.000	09:37
Diagnostics Check		OK	09:37

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGAN,D,O
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/22/2025 13:33:14

Diagnostics Check		OK	13:26
Air Blank		0.000	13:26
Control Test		0.082	13:26
Air Blank		0.000	13:27
Subject Sample	# 1	0.126	13:28
Breath Volume		1.347	13:28
Air Blank		0.000	13:29
Air Blank		0.000	13:31
Subject Sample	# 2	0.129	13:31
Breath Volume		2.414	13:31
Air Blank		0.000	13:32
Control Test		0.081	13:32
Air Blank		0.000	13:33
Diagnostics Check		OK	13:33

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: FERREIRA SOARES

DATE/TIME OF TEST: 06/23/2025 02:46:48

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:43	
Control Test		0.081	02:43	
Air Blank		0.000	02:44	
Subject Sample	# 1		02:45	Subject Test Refused
Breath Volume		0.000	02:45	
Air Blank		0.000	02:45	
Control Test		0.081	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRY,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 06/23/2025 06:20:15

Diagnostics Check		OK	06:17	
Air Blank		0.000	06:17	
Control Test		0.080	06:18	
Air Blank		0.000	06:18	
Subject Sample	# 1		06:18	Subject Test Refused
Breath Volume		0.000	06:18	
Air Blank		0.000	06:19	
Control Test		0.081	06:19	
Air Blank		0.000	06:20	
Diagnostics Check		OK	06:20	

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BISCEGLIO

DATE/TIME OF TEST: 06/27/2025 01:13:34

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.081	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.208	01:08
Breath Volume		1.308	01:08
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.196	01:11
Breath Volume		1.253	01:11
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BROWNLEE

DATE/TIME OF TEST: 06/27/2025 04:13:27

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:10	
Control Test		0.081	04:11	
Air Blank		0.000	04:11	
Subject Sample	# 1		04:12	Subject Test Refused
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	
Control Test		0.081	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: O'CONNELL,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 06/28/2025 03:48:47

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:46	
Control Test		0.082	03:46	
Air Blank		0.000	03:47	
Subject Sample	# 1		03:47	Subject Test Refused
Breath Volume		0.000	03:47	
Air Blank		0.000	03:47	
Control Test		0.081	03:48	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 06/28/2025 04:39:53

Diagnostics Check		OK	04:37	
Air Blank		0.000	04:37	
Control Test		0.082	04:37	
Air Blank		0.000	04:38	
Subject Sample	# 1		04:38	Subject Test Refused
Breath Volume		0.000	04:38	
Air Blank		0.000	04:38	
Control Test		0.081	04:39	
Air Blank		0.000	04:39	
Diagnostics Check		OK	04:39	

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: PAR SINTO

DATE/TIME OF TEST: 06/28/2025 23:38:41

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:36	
Control Test		0.081	23:36	
Air Blank		0.000	23:37	
Subject Sample	# 1		23:37	Subject Test Refused
Breath Volume		0.000	23:37	
Air Blank		0.000	23:37	
Control Test		0.082	23:38	
Air Blank		0.000	23:38	
Diagnostics Check		OK	23:38	

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SEARS,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GARCIA AJQUI

DATE/TIME OF TEST: 06/29/2025 03:38:51

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.081	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.236	03:33
Breath Volume		1.621	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.239	03:37
Breath Volume		1.847	03:37
Air Blank		0.000	03:37
Control Test		0.081	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CORNELIOUS

DATE/TIME OF TEST: 06/30/2025 04:27:34

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.082	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.063	04:22
Breath Volume		1.593	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.059	04:26
Breath Volume		1.375	04:26
Air Blank		0.000	04:26
Control Test		0.081	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGAN,D,O
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:07

SUBJECT'S LAST NAME: KANIA

DATE/TIME OF TEST: 06/30/2025 18:33:58

Diagnostics Check		OK	18:27
Air Blank		0.000	18:28
Control Test		0.081	18:28
Air Blank		0.000	18:28
Subject Sample	# 1	0.000	18:29
Breath Volume		2.996	18:29
Air Blank		0.000	18:30
Air Blank		0.000	18:32
Subject Sample	# 2	0.000	18:32
Breath Volume		4.300	18:32
Air Blank		0.000	18:33
Control Test		0.081	18:33
Air Blank		0.000	18:33
Diagnostics Check		OK	18:33

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRY,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 06/30/2025 21:53:47

Diagnostics Check		OK	21:47
Air Blank		0.000	21:48
Control Test		0.081	21:48
Air Blank		0.000	21:49
Subject Sample	# 1	0.000	21:49
Breath Volume		2.417	21:49
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.000	21:52
Breath Volume		2.085	21:52
Air Blank		0.000	21:52
Control Test		0.081	21:53
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 07/10/2025 13:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEURTSEN,ANTHONY,P
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: COMPIERCHIO,ANTHONY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: KOPLIN

DATE/TIME OF TEST: 06/26/2025 03:34:48

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.081	03:34	
Air Blank		03:34	RFI Detect
Air Blank		03:34	RFI Detect

Data Download Date/Time: 07/04/2025 11:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEURTSEN,ANTHONY,P
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: COMPIERCHIO,ANTHONY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: KOPLIN

DATE/TIME OF TEST: 06/26/2025 03:38:42

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:37	
Control Test		0.080	03:37	
Air Blank		0.000	03:38	
Subject Sample	# 1		03:38	RFI Detect
Breath Volume		0.000	03:38	
Air Blank			03:38	RFI Detect
Air Blank			03:38	RFI Detect

Data Download Date/Time: 07/04/2025 11:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEURTSEN,ANTHONY,P
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: COMPIERCHIO,ANTHONY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: KOPLIN

DATE/TIME OF TEST: 06/26/2025 03:44:20

Diagnostics Check		OK	03:41	
Air Blank		0.000	03:42	
Control Test		0.080	03:42	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:43	RFI Detect
Breath Volume		0.113	03:43	
Air Blank			03:44	RFI Detect
Air Blank			03:44	RFI Detect

Data Download Date/Time: 07/04/2025 11:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEURTSEN,ANTHONY,P
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: COMPIERCHIO,ANTHONY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: KOPLIN

DATE/TIME OF TEST: 06/26/2025 03:48:40

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:46	
Control Test		0.080	03:47	
Air Blank		0.000	03:47	
Subject Sample	# 1		03:48	RFI Detect
Breath Volume		0.000	03:48	
Air Blank			03:48	RFI Detect
Air Blank		0.000	03:48	

Data Download Date/Time: 07/04/2025 11:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEURTSEN,ANTHONY,P
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: COMPIERCHIO,ANTHONY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: KOPLIN

DATE/TIME OF TEST: 06/26/2025 03:59:47

Diagnostics Check		OK	03:52
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:54
Subject Sample	# 1	0.167	03:55
Breath Volume		2.125	03:55
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.164	03:58
Breath Volume		2.019	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 07/04/2025 11:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,EMMANUEL,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: FRANTZ,HUNTER
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 05/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ESTUPINANARIAS

DATE/TIME OF TEST: 06/01/2025 04:59:12

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.080	04:52
Air Blank		0.000	04:53
Subject Sample	# 1	0.187	04:54
Breath Volume		1.675	04:54
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.168	04:57
Breath Volume		2.121	04:57
Air Blank		0.000	04:58
Control Test		0.078	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 06/04/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 06/05/2025 23:38:50

Diagnostics Check		OK	23:32
Air Blank		0.000	23:33
Control Test		0.080	23:33
Air Blank		0.000	23:34
Subject Sample	# 1	0.132	23:34
Breath Volume		3.785	23:34
Air Blank		0.000	23:35
Air Blank		0.000	23:36
Subject Sample	# 2	0.125	23:37
Breath Volume		2.105	23:37
Air Blank		0.000	23:37
Control Test		0.079	23:38
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 06/19/2025 23:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: CATO

DATE/TIME OF TEST: 06/12/2025 00:46:36

Diagnostics Check		OK	00:40
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.000	00:42
Breath Volume		2.042	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.000	00:45
Breath Volume		2.007	00:45
Air Blank		0.000	00:45
Control Test		0.076	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 06/19/2025 23:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 06/22/2025 00:50:19

Diagnostics Check		OK	00:44
Air Blank		0.000	00:44
Control Test		0.081	00:44
Air Blank		0.000	00:45
Subject Sample	# 1	0.139	00:46
Breath Volume		2.187	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.141	00:48
Breath Volume		1.773	00:48
Air Blank		0.000	00:49
Control Test		0.078	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 07/23/2025 22:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: COSBY

DATE/TIME OF TEST: 06/24/2025 02:44:07

Diagnostics Check		OK	02:32	
Air Blank		0.000	02:33	
Control Test		0.078	02:33	
Air Blank		0.000	02:33	
Subject Sample	# 1	0.109	02:36	Volume Not Met
Breath Volume		0.367	02:36	
Air Blank		0.000	02:37	
Air Blank		0.000	02:39	
Subject Sample	# 2	0.090	02:42	Volume Not Met
Breath Volume		0.042	02:42	
Air Blank		0.000	02:43	
Control Test		0.076	02:43	
Air Blank		0.000	02:44	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 07/23/2025 22:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/27/2025 03:40:15

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.080	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.137	03:35
Breath Volume		1.453	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.138	03:38
Breath Volume		1.531	03:38
Air Blank		0.000	03:39
Control Test		0.079	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 07/23/2025 22:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 06/28/2025 01:41:52

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.136	01:37
Breath Volume		2.117	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.134	01:40
Breath Volume		1.835	01:40
Air Blank		0.000	01:40
Control Test		0.077	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 07/23/2025 22:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GLOWACKI

DATE/TIME OF TEST: 06/29/2025 03:34:06

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.082	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.160	03:29
Breath Volume		1.707	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:31
Subject Sample	# 2	0.160	03:32
Breath Volume		2.105	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 07/23/2025 22:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GREEN,O
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: NEUGEBAUER

DATE/TIME OF TEST: 06/28/2025 22:22:57

Diagnostics Check		OK	22:16
Air Blank		0.000	22:16
Control Test		0.080	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.000	22:18
Breath Volume		3.425	22:18
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.000	22:21
Breath Volume		1.683	22:21
Air Blank		0.000	22:22
Control Test		0.080	22:22
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

Data Download Date/Time: 06/29/2025 22:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MEGAN,A
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: THOMAS,MEGAN
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: BUI
ARREST TIME: 16:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: WOODLARD

DATE/TIME OF TEST: 06/06/2025 17:04:08

Diagnostics Check		OK	16:57
Air Blank		0.000	16:57
Control Test		0.078	16:57
Air Blank		0.000	16:58
Subject Sample	# 1	0.093	16:59
Breath Volume		1.601	16:59
Air Blank		0.000	17:00
Air Blank		0.000	17:02
Subject Sample	# 2	0.086	17:02
Breath Volume		1.781	17:02
Air Blank		0.000	17:03
Control Test		0.077	17:03
Air Blank		0.000	17:04
Diagnostics Check		OK	17:04

Data Download Date/Time: 06/28/2025 15:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA,ARMANDO,R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: DOZIER,TERRELL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08425080A2
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MUNOZ CARVAJAL

DATE/TIME OF TEST: 06/07/2025 23:58:05

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.080	23:52
Air Blank		0.000	23:52
Subject Sample	# 1	0.187	23:53
Breath Volume		1.785	23:53
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.191	23:56
Breath Volume		1.445	23:56
Air Blank		0.000	23:57
Control Test		0.080	23:57
Air Blank		0.000	23:57
Diagnostics Check		OK	23:58

Data Download Date/Time: 06/22/2025 21:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: RIGGS-MEADOR

DATE/TIME OF TEST: 06/03/2025 02:03:08

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:56	
Control Test		0.080	01:57	
Air Blank		0.000	01:57	
Subject Sample	# 1	0.177	01:58	
Breath Volume		1.140	01:58	
Air Blank		0.000	01:58	
Air Blank		0.000	02:00	
Subject Sample	# 2		02:02	RFI Detect
Breath Volume		0.429	02:02	
Air Blank			02:02	RFI Detect
Air Blank		0.000	02:03	

Data Download Date/Time: 06/11/2025 21:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: RIGGS-MEADOR

DATE/TIME OF TEST: 06/03/2025 02:12:03

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.186	02:07
Breath Volume		2.328	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:09
Subject Sample	# 2	0.199	02:10
Breath Volume		2.375	02:10
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:12

Data Download Date/Time: 06/11/2025 21:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: NAVARRETE

DATE/TIME OF TEST: 06/09/2025 03:33:07

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.079	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.123	03:28
Breath Volume		1.667	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.120	03:31
Breath Volume		1.617	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 06/11/2025 21:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: CABRERA TAVERAS

DATE/TIME OF TEST: 06/09/2025 20:28:11

Diagnostics Check		OK	20:21
Air Blank		0.000	20:22
Control Test		0.078	20:22
Air Blank		0.000	20:23
Subject Sample	# 1	0.179	20:23
Breath Volume		2.316	20:23
Air Blank		0.000	20:24
Air Blank		0.000	20:26
Subject Sample	# 2	0.182	20:26
Breath Volume		2.027	20:26
Air Blank		0.000	20:27
Control Test		0.079	20:27
Air Blank		0.000	20:28
Diagnostics Check		OK	20:28

Data Download Date/Time: 06/11/2025 21:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/21/2025 01:22:44

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.078	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.146	01:18
Breath Volume		3.558	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:20
Subject Sample	# 2	0.147	01:21
Breath Volume		2.562	01:21
Air Blank		0.000	01:21
Control Test		0.079	01:22
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 07/14/2025 21:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: THOMAS,BRANDEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: DALY

DATE/TIME OF TEST: 06/21/2025 05:47:54

Diagnostics Check		OK	05:41
Air Blank		0.000	05:42
Control Test		0.079	05:42
Air Blank		0.000	05:42
Subject Sample	# 1	0.163	05:43
Breath Volume		1.625	05:43
Air Blank		0.000	05:43
Air Blank		0.000	05:45
Subject Sample	# 2	0.175	05:46
Breath Volume		2.035	05:46
Air Blank		0.000	05:46
Control Test		0.080	05:47
Air Blank		0.000	05:47
Diagnostics Check		OK	05:47

Data Download Date/Time: 07/14/2025 21:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 06/22/2025 03:14:09

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:02	
Control Test		0.079	03:02	
Air Blank		0.000	03:02	
Subject Sample	# 1	0.171	03:06	Volume Not Met
Breath Volume		0.035	03:06	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 2	0.196	03:09	No .020 agreement
Breath Volume		1.253	03:09	
Air Blank		0.000	03:09	
Air Blank		0.000	03:11	
Subject Sample	# 3	0.174	03:12	No .020 agreement
Breath Volume		2.300	03:12	
Air Blank		0.000	03:13	
Control Test		0.078	03:13	
Air Blank		0.000	03:14	
Diagnostics Check		OK	03:14	

Data Download Date/Time: 07/14/2025 21:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ZECHARIAH,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MITCHELL,SHONQUEZ
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: REES

DATE/TIME OF TEST: 06/07/2025 04:03:10

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.182	03:58
Breath Volume		1.886	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.176	04:01
Breath Volume		2.195	04:01
Air Blank		0.000	04:02
Control Test		0.081	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 06/19/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,SHONQUEZ,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MITCHELL,SHONQUEZ
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TALLMAN

DATE/TIME OF TEST: 06/09/2025 03:39:30

Diagnostics Check	OK	03:38	
Air Blank		03:39	RFI Detect
Air Blank	0.000	03:39	

Data Download Date/Time: 06/19/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,SHONQUEZ,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MITCHELL,SHONQUEZ
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TALLMAN

DATE/TIME OF TEST: 06/09/2025 03:43:15

Diagnostics Check	OK	03:42	
Air Blank		03:43	Improper Sample
Air Blank		03:43	Improper Sample

Data Download Date/Time: 06/19/2025 10:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,SHONQUEZ,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MITCHELL,SHONQUEZ
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TALLMAN

DATE/TIME OF TEST: 06/09/2025 03:48:31

Diagnostics Check	OK	03:47	
Air Blank		03:48	Ambient Fail
Air Blank	0.000	03:48	

Data Download Date/Time: 06/19/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,SHONQUEZ,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MITCHELL,SHONQUEZ
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TALLMAN

DATE/TIME OF TEST: 06/09/2025 03:53:47

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:53	
Control Test	0.079	03:53	RFI Detect
Air Blank	0.000	03:53	

Data Download Date/Time: 06/19/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,SHONQUEZ,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MITCHELL,SHONQUEZ
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TALLMAN

DATE/TIME OF TEST: 06/09/2025 04:03:45

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.081	03:58
Air Blank		0.000	03:58
Subject Sample	# 1	0.244	03:59
Breath Volume		3.578	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:01
Subject Sample	# 2	0.238	04:02
Breath Volume		4.488	04:02
Air Blank		0.000	04:02
Control Test		0.080	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 06/19/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ZECHARIAH,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BROWN,ZECHARIAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: ASHLEY

DATE/TIME OF TEST: 06/15/2025 05:03:10

Diagnostics Check		OK	04:56
Air Blank		0.000	04:56
Control Test		0.080	04:57
Air Blank		0.000	04:57
Subject Sample	# 1	0.191	04:58
Breath Volume		2.464	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.179	05:01
Breath Volume		1.914	05:01
Air Blank		0.000	05:02
Control Test		0.081	05:02
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 06/19/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ZECHARIAH,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BROWN,ZECHARIAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: RADTKA

DATE/TIME OF TEST: 06/29/2025 03:09:19

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:06	
Control Test		0.080	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 1	0.209	03:07	
Breath Volume		4.347	03:07	
Air Blank			03:08	Purge Fail
Air Blank		0.000	03:09	

Data Download Date/Time: 07/29/2025 09:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ZECHARIAH,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BROWN,ZECHARIAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: RADTKA

DATE/TIME OF TEST: 06/29/2025 03:23:21

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.080	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.213	03:15
Breath Volume		3.699	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.186	03:18
Breath Volume		2.058	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 3	0.200	03:21
Breath Volume		2.343	03:21
Air Blank		0.000	03:22
Control Test		0.082	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 07/29/2025 09:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,MATHEW,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: RUNGE,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: KEEN

DATE/TIME OF TEST: 06/08/2025 01:04:36

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.084	01:00	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:03	Subject Test Refused
Breath Volume		0.000	01:03	
Air Blank		0.000	01:03	
Control Test		0.084	01:04	
Air Blank		0.000	01:04	
Diagnostics Check		OK	01:04	

Data Download Date/Time: 06/21/2025 18:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,MATHEW,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: RUNGE,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: KEEN

DATE/TIME OF TEST: 06/08/2025 01:13:08

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.084	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.213	01:07
Breath Volume		3.074	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.205	01:11
Breath Volume		3.355	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 06/21/2025 18:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMP,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LAMP,JONATHAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 06/09/2025 03:03:25

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.083	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.205	02:58
Breath Volume		4.167	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.204	03:01
Breath Volume		2.863	03:01
Air Blank		0.000	03:02
Control Test		0.083	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 06/21/2025 18:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRY,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: SPRY,NICHOLAS
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CHESSHER

DATE/TIME OF TEST: 06/14/2025 01:43:14

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:40	
Control Test		0.084	01:41	
Air Blank		0.000	01:41	
Subject Sample	# 1		01:41	Subject Test Refused
Breath Volume		0.000	01:41	
Air Blank		0.000	01:42	
Control Test		0.084	01:42	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 06/21/2025 18:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: GREENE,JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 06/21/2025 00:31:06

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:28	
Control Test		0.084	00:29	
Air Blank		0.000	00:29	
Subject Sample	# 1		00:29	Subject Test Refused
Breath Volume		0.000	00:29	
Air Blank		0.000	00:30	
Control Test		0.084	00:30	
Air Blank		0.000	00:31	
Diagnostics Check		OK	00:31	

Data Download Date/Time: 06/21/2025 18:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: SCHACHTER,SOLANGE
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 05/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HERNANDEZ CARABALLO

DATE/TIME OF TEST: 06/15/2025 01:56:09

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:52	
Control Test		0.080	01:53	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:54	Subject Test Refused
Breath Volume		0.000	01:54	
Air Blank		0.000	01:55	
Control Test		0.081	01:55	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:56	

Data Download Date/Time: 06/17/2025 21:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: TURNER,KENYON
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 05/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: THOMPSON HENRY

DATE/TIME OF TEST: 06/15/2025 02:18:08

Diagnostics Check		OK	02:11
Air Blank		0.000	02:12
Control Test		0.080	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.142	02:13
Breath Volume		2.757	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.150	02:16
Breath Volume		2.679	02:16
Air Blank		0.000	02:17
Control Test		0.079	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 06/17/2025 21:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: BRUDERMANN

DATE/TIME OF TEST: 06/02/2025 01:24:46

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.146	01:19
Breath Volume		1.519	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.151	01:23
Breath Volume		1.878	01:23
Air Blank		0.000	01:23
Control Test		0.078	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: ADAMSON

DATE/TIME OF TEST: 06/03/2025 02:03:37

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.079	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.092	01:58
Breath Volume		2.722	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:00
Subject Sample	# 2	0.096	02:02
Breath Volume		2.832	02:02
Air Blank		0.000	02:02
Control Test		0.079	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOJNOWSKY,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LEOTA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/04/2025 01:23:55

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:21	
Control Test		0.079	01:21	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:22	
Control Test		0.079	01:23	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIAGA,JULIO,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,JULIO
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: HERNANDEZ MERCADO

DATE/TIME OF TEST: 06/06/2025 19:26:42

Diagnostics Check		OK	19:20
Air Blank		0.000	19:20
Control Test		0.079	19:21
Air Blank		0.000	19:21
Subject Sample	# 1	0.164	19:22
Breath Volume		1.574	19:22
Air Blank		0.000	19:22
Air Blank		0.000	19:24
Subject Sample	# 2	0.168	19:25
Breath Volume		1.425	19:25
Air Blank		0.000	19:25
Control Test		0.078	19:26
Air Blank		0.000	19:26
Diagnostics Check		OK	19:26

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 06/07/2025 02:58:55

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:56	
Control Test		0.079	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 1		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:58	
Control Test		0.078	02:58	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIAGA,JULIO,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,JULIO
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 06/07/2025 21:13:32

Diagnostics Check		OK	21:10	
Air Blank		0.000	21:11	
Control Test		0.079	21:11	
Air Blank		0.000	21:11	
Subject Sample	# 1		21:12	Subject Test Refused
Breath Volume		0.000	21:12	
Air Blank		0.000	21:12	
Control Test		0.079	21:12	
Air Blank		0.000	21:13	
Diagnostics Check		OK	21:13	

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOJNOWSKY,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BROOKS,Z
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 06/13/2025 03:05:34

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:51	
Control Test		0.079	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1	0.164	02:54	
Breath Volume		3.136	02:54	
Air Blank		0.000	02:55	
Air Blank		0.000	02:57	
Subject Sample	# 2	0.161	03:00	Volume Not Met
Breath Volume		0.050	03:00	
Air Blank		0.000	03:01	
Air Blank		0.000	03:03	
Subject Sample	# 3	0.159	03:03	
Breath Volume		1.933	03:03	
Air Blank		0.000	03:04	
Control Test		0.077	03:04	
Air Blank		0.000	03:05	
Diagnostics Check		OK	03:05	

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: WETZEL

DATE/TIME OF TEST: 06/14/2025 22:15:46

Diagnostics Check		OK	22:09
Air Blank		0.000	22:10
Control Test		0.078	22:10
Air Blank		0.000	22:10
Subject Sample	# 1	0.000	22:11
Breath Volume		3.250	22:11
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 2	0.000	22:14
Breath Volume		1.636	22:14
Air Blank		0.000	22:14
Control Test		0.079	22:15
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/15/2025 00:02:50

Diagnostics Check		OK	23:54
Air Blank		0.000	23:55
Control Test		0.078	23:55
Air Blank		0.000	23:56
Subject Sample	# 2	0.232	23:57
Breath Volume		1.578	23:57
Air Blank		0.000	23:58
Air Blank		0.000	23:59
Subject Sample	# 1	0.249	00:01
Breath Volume		1.339	00:01
Air Blank		0.000	00:01
Control Test		0.078	00:02
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,R,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 06/15/2025 01:25:48

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.079	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.114	01:21
Breath Volume		2.203	01:21
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.114	01:24
Breath Volume		1.746	01:24
Air Blank		0.000	01:24
Control Test		0.079	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FERRO,T
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: TOZZI

DATE/TIME OF TEST: 06/15/2025 04:19:44

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.079	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.248	04:15
Breath Volume		2.816	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.251	04:18
Breath Volume		2.226	04:18
Air Blank		0.000	04:18
Control Test		0.077	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACCHIONE,VINCENT,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FABY,ROBERT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: BOWLES

DATE/TIME OF TEST: 06/18/2025 22:47:01

Diagnostics Check		OK	22:44	
Air Blank		0.000	22:44	
Control Test		0.079	22:44	
Air Blank		0.000	22:45	
Subject Sample	# 1		22:45	Subject Test Refused
Breath Volume		0.000	22:45	
Air Blank		0.000	22:46	
Control Test		0.079	22:46	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIAGA,JULIO,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ALIAGA,JULIO
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: STANEK

DATE/TIME OF TEST: 06/20/2025 19:19:31

Diagnostics Check		OK	19:13
Air Blank		0.000	19:13
Control Test		0.080	19:14
Air Blank		0.000	19:14
Subject Sample	# 1	0.084	19:15
Breath Volume		3.527	19:15
Air Blank		0.000	19:15
Air Blank		0.000	19:17
Subject Sample	# 2	0.076	19:18
Breath Volume		3.621	19:18
Air Blank		0.000	19:18
Control Test		0.079	19:18
Air Blank		0.000	19:19
Diagnostics Check		OK	19:19

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOJNOWSKY,IWONA,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: EDWARDS,R
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 06/23/2025 01:38:56

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:36	
Control Test		0.079	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:37	Subject Test Refused
Breath Volume		0.000	01:37	
Air Blank		0.000	01:38	
Control Test		0.079	01:38	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POULIN,D,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: POULIN,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/26/2025 02:43:05

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:40	
Control Test		0.080	02:41	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:41	RFI Detect
Breath Volume		0.000	02:41	
Air Blank		0.000	02:42	
Control Test		0.080	02:42	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POULIN',D,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: POULIN,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/26/2025 02:48:24

Diagnostics Check		OK	02:44	
Air Blank		0.000	02:45	
Control Test		0.080	02:45	
Air Blank		0.000	02:45	
Subject Sample	# 1		02:47	Subject Test Refused
Breath Volume		0.000	02:47	
Air Blank		0.000	02:47	
Control Test		0.080	02:47	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOJNOWSKY,IWONA,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KELLY,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BARR

DATE/TIME OF TEST: 06/26/2025 03:18:30

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.230	03:13
Breath Volume		1.175	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.222	03:16
Breath Volume		1.550	03:16
Air Blank		0.000	03:17
Control Test		0.080	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOJNOWSKY,IWONA,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: TORRE MEDINA

DATE/TIME OF TEST: 06/27/2025 00:14:25

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:11	
Control Test		0.079	00:11	
Air Blank		0.000	00:12	
Subject Sample	# 1		00:13	Subject Test Refused
Breath Volume		0.000	00:13	
Air Blank		0.000	00:13	
Control Test		0.080	00:13	
Air Blank		0.000	00:14	
Diagnostics Check		OK	00:14	

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 06/28/2025 00:26:39

Diagnostics Check		OK	00:23	
Air Blank		0.000	00:24	
Control Test		0.079	00:24	
Air Blank		0.000	00:24	
Subject Sample	# 1		00:25	Subject Test Refused
Breath Volume		0.000	00:25	
Air Blank		0.000	00:25	
Control Test		0.079	00:26	
Air Blank		0.000	00:26	
Diagnostics Check		OK	00:26	

Data Download Date/Time: 06/28/2025 02:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: LEHOTSKY

DATE/TIME OF TEST: 06/28/2025 15:33:27

Diagnostics Check		OK	15:26
Air Blank		0.000	15:26
Control Test		0.078	15:27
Air Blank		0.000	15:27
Subject Sample	# 1	0.136	15:28
Breath Volume		2.750	15:28
Air Blank		0.000	15:29
Air Blank		0.000	15:30
Subject Sample	# 2	0.135	15:31
Breath Volume		1.750	15:31
Air Blank		0.000	15:32
Control Test		0.078	15:32
Air Blank		0.000	15:33
Diagnostics Check		OK	15:33

Data Download Date/Time: 07/31/2025 19:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: STRAND

DATE/TIME OF TEST: 06/29/2025 03:55:37

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:52	
Control Test		0.079	03:52	
Air Blank		0.000	03:53	
Subject Sample	# 1		03:54	Subject Test Refused
Breath Volume		0.000	03:54	
Air Blank		0.000	03:54	
Control Test		0.079	03:55	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 07/31/2025 19:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WAKELAND,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/29/2025 04:32:47

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.080	04:27
Air Blank		0.000	04:27
Subject Sample	# 1	0.000	04:28
Breath Volume		1.300	04:28
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.000	04:31
Breath Volume		1.789	04:31
Air Blank		0.000	04:31
Control Test		0.079	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 07/31/2025 19:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HORRIGAN,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/01/2025 18:19:15

Diagnostics Check		OK	18:16	
Air Blank		0.000	18:16	
Control Test		0.081	18:16	
Air Blank		0.000	18:17	
Subject Sample	# 1		18:17	Subject Test Refused
Breath Volume		0.000	18:17	
Air Blank		0.000	18:18	
Control Test		0.081	18:18	
Air Blank		0.000	18:19	
Diagnostics Check		OK	18:19	

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: LARKINS

DATE/TIME OF TEST: 06/05/2025 04:31:07

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.082	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.157	04:26
Breath Volume		2.296	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.163	04:29
Breath Volume		1.824	04:29
Air Blank		0.000	04:30
Control Test		0.081	04:30
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/05/2025 04:45:52

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:33	
Control Test		0.081	04:34	
Air Blank		0.000	04:34	
Subject Sample	# 1	0.144	04:37	Volume Not Met
Breath Volume		0.015	04:37	
Air Blank		0.000	04:38	
Air Blank		0.000	04:40	
Subject Sample	# 2	0.146	04:41	
Breath Volume		1.707	04:41	
Air Blank		0.000	04:41	
Air Blank		0.000	04:43	
Subject Sample	# 3	0.142	04:44	
Breath Volume		1.593	04:44	
Air Blank		0.000	04:44	
Control Test		0.081	04:45	
Air Blank		0.000	04:45	
Diagnostics Check		OK	04:45	

Data Download Date/Time: 06/11/2025 09:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: BERNARDO

DATE/TIME OF TEST: 06/06/2025 22:17:27

Diagnostics Check		OK	22:13	
Air Blank		0.000	22:13	
Control Test		0.082	22:14	
Air Blank		0.000	22:14	
Subject Sample	# 1		22:16	Subject Test Refused
Breath Volume		0.000	22:16	
Air Blank		0.000	22:16	
Control Test		0.082	22:16	
Air Blank		0.000	22:17	
Diagnostics Check		OK	22:17	

Data Download Date/Time: 06/11/2025 09:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: SHEDRICK

DATE/TIME OF TEST: 06/07/2025 04:20:11

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.082	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.157	04:15
Breath Volume		1.554	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.176	04:18
Breath Volume		1.292	04:18
Air Blank		0.000	04:19
Control Test		0.082	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 06/11/2025 09:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: SPECE

DATE/TIME OF TEST: 06/08/2025 04:04:58

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.082	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.157	04:00
Breath Volume		1.539	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.159	04:03
Breath Volume		1.546	04:03
Air Blank		0.000	04:04
Control Test		0.082	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 06/11/2025 09:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: FLUELLEN

DATE/TIME OF TEST: 06/11/2025 01:41:49

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:36	
Control Test		0.082	01:37	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:40	Subject Test Refused
Breath Volume		0.000	01:40	
Air Blank		0.000	01:40	
Control Test		0.082	01:41	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 06/11/2025 09:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: FLUELLEN

DATE/TIME OF TEST: 06/11/2025 01:47:45

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:43	
Control Test		0.082	01:44	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:46	Subject Test Refused
Breath Volume		0.000	01:46	
Air Blank		0.000	01:46	
Control Test		0.082	01:47	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 06/11/2025 09:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/11/2025 03:01:35

Diagnostics Check		OK	02:54
Air Blank		0.000	02:55
Control Test		0.083	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.000	02:57
Breath Volume		2.750	02:57
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.000	03:00
Breath Volume		2.156	03:00
Air Blank		0.000	03:00
Control Test		0.082	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 06/11/2025 09:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: MOLOFSKY

DATE/TIME OF TEST: 06/12/2025 04:01:50

Diagnostics Check		OK	03:48	
Air Blank		0.000	03:48	
Control Test		0.082	03:49	
Air Blank		0.000	03:49	
Subject Sample	# 1		03:52	No Sample Provided
Breath Volume		0.000	03:52	
Air Blank		0.000	03:53	
Air Blank		0.000	03:55	
Subject Sample	# 2	0.190	03:56	
Breath Volume		2.257	03:56	
Air Blank		0.000	03:57	
Air Blank		0.000	03:59	
Subject Sample	# 3	0.196	04:00	
Breath Volume		1.660	04:00	
Air Blank		0.000	04:00	
Control Test		0.082	04:01	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: HALULI

DATE/TIME OF TEST: 06/13/2025 01:21:50

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.083	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.137	01:17
Breath Volume		2.187	01:17
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.134	01:20
Breath Volume		1.734	01:20
Air Blank		0.000	01:20
Control Test		0.080	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: LININGER

DATE/TIME OF TEST: 06/13/2025 04:35:20

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:31	
Control Test		0.082	04:31	
Air Blank		0.000	04:32	
Subject Sample	# 1		04:33	Subject Test Refused
Breath Volume		0.000	04:33	
Air Blank		0.000	04:34	
Control Test		0.083	04:34	
Air Blank		0.000	04:35	
Diagnostics Check		OK	04:35	

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: CHURCH

DATE/TIME OF TEST: 06/13/2025 04:58:27

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.082	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.230	04:53
Breath Volume		1.261	04:53
Air Blank		0.000	04:54
Air Blank		0.000	04:55
Subject Sample	# 2	0.218	04:56
Breath Volume		1.398	04:56
Air Blank		0.000	04:57
Control Test		0.079	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD,MITCHELL,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: DAMBROSIO

DATE/TIME OF TEST: 06/13/2025 21:27:33

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:23	
Control Test		0.082	21:23	
Air Blank		0.000	21:24	
Subject Sample	# 1		21:26	Subject Test Refused
Breath Volume		0.000	21:26	
Air Blank		0.000	21:26	
Control Test		0.082	21:26	
Air Blank		0.000	21:27	
Diagnostics Check		OK	21:27	

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: KOPP

DATE/TIME OF TEST: 06/14/2025 22:57:35

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.082	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.196	22:53
Breath Volume		2.902	22:53
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.189	22:56
Breath Volume		2.199	22:56
Air Blank		0.000	22:56
Control Test		0.082	22:56
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 06/16/2025 00:31:35

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:20	
Control Test		0.082	00:20	
Air Blank		0.000	00:21	
Subject Sample	# 1		00:24	No Sample Provided
Breath Volume		0.000	00:24	
Air Blank		0.000	00:25	
Air Blank		0.000	00:27	
Subject Sample	# 2		00:30	No Sample Provided
Breath Volume		0.000	00:30	
Air Blank		0.000	00:30	
Control Test		0.082	00:30	
Air Blank		0.000	00:31	
Diagnostics Check		OK	00:31	

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:01

SUBJECT'S LAST NAME: CHILDRESS

DATE/TIME OF TEST: 06/18/2025 15:39:18

Diagnostics Check		OK	15:33
Air Blank		0.000	15:33
Control Test		0.082	15:34
Air Blank		0.000	15:34
Subject Sample	# 1	0.000	15:35
Breath Volume		1.898	15:35
Air Blank		0.000	15:35
Air Blank		0.000	15:37
Subject Sample	# 2	0.000	15:37
Breath Volume		1.500	15:37
Air Blank		0.000	15:38
Control Test		0.082	15:38
Air Blank		0.000	15:39
Diagnostics Check		OK	15:39

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 06/18/2025 18:49:13

Diagnostics Check		OK	18:41
Air Blank		0.000	18:42
Control Test		0.082	18:42
Air Blank		0.000	18:43
Subject Sample	# 1	0.113	18:44
Breath Volume		1.675	18:44
Air Blank		0.000	18:45
Air Blank		0.000	18:46
Subject Sample	# 2	0.106	18:47
Breath Volume		2.601	18:47
Air Blank		0.000	18:48
Control Test		0.082	18:48
Air Blank		0.000	18:49
Diagnostics Check		OK	18:49

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: EL OTMANI

DATE/TIME OF TEST: 06/20/2025 03:34:27

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:30	
Control Test		0.082	03:30	
Air Blank		0.000	03:31	
Subject Sample	# 1		03:33	Subject Test Refused
Breath Volume		0.000	03:33	
Air Blank		0.000	03:33	
Control Test		0.083	03:33	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD,MITCHELL,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: JAMES,STEVEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: GRANADOS

DATE/TIME OF TEST: 06/20/2025 23:51:06

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.082	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.120	23:46
Breath Volume		1.734	23:46
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.122	23:49
Breath Volume		2.109	23:49
Air Blank		0.000	23:50
Control Test		0.082	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD,MITCHELL,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: HOLVOET

DATE/TIME OF TEST: 06/21/2025 04:37:53

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.082	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.160	04:33
Breath Volume		2.238	04:33
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.158	04:36
Breath Volume		1.863	04:36
Air Blank		0.000	04:36
Control Test		0.081	04:37
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD,MITCHELL,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: MAYES

DATE/TIME OF TEST: 06/22/2025 02:00:45

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:57	
Control Test		0.082	01:57	
Air Blank		0.000	01:58	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	01:59	
Control Test		0.082	02:00	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD,MITCHELL,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 06/22/2025 02:35:15

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.081	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.139	02:30
Breath Volume		1.824	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 2	0.143	02:33
Breath Volume		1.925	02:33
Air Blank		0.000	02:34
Control Test		0.081	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,R
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SCHOREJS

DATE/TIME OF TEST: 06/22/2025 03:04:20

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:56	
Control Test		0.081	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 1		03:00	No Sample Provided
Breath Volume		0.000	03:00	
Air Blank		0.000	03:00	
Air Blank		0.000	03:02	
Subject Sample	# 2		03:02	Subject Test Refused
Breath Volume		0.000	03:02	
Air Blank		0.000	03:03	
Control Test		0.080	03:03	
Air Blank		0.000	03:04	
Diagnostics Check		OK	03:04	

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 06/23/2025 02:17:40

Diagnostics Check		OK	02:11
Air Blank		0.000	02:12
Control Test		0.083	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.082	02:13
Breath Volume		1.597	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.087	02:16
Breath Volume		1.109	02:16
Air Blank		0.000	02:16
Control Test		0.082	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHETEE,NICHOLAS,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DAUGHETEE,NICHOLAS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:03

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 06/23/2025 11:30:45

Diagnostics Check		OK	11:23	
Air Blank		0.000	11:23	
Control Test		0.082	11:23	
Air Blank		0.000	11:24	
Subject Sample	# 1	0.000	11:27	Volume Not Met
Breath Volume		0.949	11:27	
Air Blank		0.000	11:28	
Air Blank		0.000	11:30	
Subject Sample	# 2		11:30	Sequence Aborted
Breath Volume		0.000	11:30	
Air Blank		0.000	11:30	

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:18

SUBJECT'S LAST NAME: HASAN

DATE/TIME OF TEST: 06/25/2025 13:44:15

Diagnostics Check		OK	13:38
Air Blank		0.000	13:38
Control Test		0.082	13:38
Air Blank		0.000	13:39
Subject Sample	# 1	0.133	13:39
Breath Volume		2.464	13:39
Air Blank		0.000	13:40
Air Blank		0.000	13:42
Subject Sample	# 2	0.126	13:42
Breath Volume		2.144	13:42
Air Blank		0.000	13:43
Control Test		0.082	13:43
Air Blank		0.000	13:44
Diagnostics Check		OK	13:44

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD,MITCHELL,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: GALAMBOS

DATE/TIME OF TEST: 06/26/2025 00:13:11

Diagnostics Check		OK	00:09	
Air Blank		0.000	00:09	
Control Test		0.082	00:09	
Air Blank		0.000	00:10	
Subject Sample	# 1		00:11	Subject Test Refused
Breath Volume		0.000	00:11	
Air Blank		0.000	00:12	
Control Test		0.082	00:12	
Air Blank		0.000	00:13	
Diagnostics Check		OK	00:13	

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ORIANO RAMIREZ

DATE/TIME OF TEST: 06/26/2025 03:28:41

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.082	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.083	03:24
Breath Volume		2.269	03:24
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.080	03:27
Breath Volume		2.144	03:27
Air Blank		0.000	03:27
Control Test		0.080	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:26

SUBJECT'S LAST NAME: POLANCO

DATE/TIME OF TEST: 06/26/2025 11:05:28

Diagnostics Check		OK	10:56	
Air Blank		0.000	10:57	
Control Test		0.082	10:57	
Air Blank		0.000	10:57	
Subject Sample	# 1		11:01	No Sample Provided
Breath Volume		0.000	11:01	
Air Blank		0.000	11:01	
Air Blank		0.000	11:03	
Subject Sample	# 2		11:04	Subject Test Refused
Breath Volume		0.000	11:04	
Air Blank		0.000	11:04	
Control Test		0.082	11:04	
Air Blank		0.000	11:05	
Diagnostics Check		OK	11:05	

Data Download Date/Time: 07/08/2025 21:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 06/27/2025 02:01:23

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.082	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.085	01:56
Breath Volume		2.593	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.081	01:59
Breath Volume		2.285	01:59
Air Blank		0.000	02:00
Control Test		0.083	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 07/08/2025 21:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD,MITCHELL,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: SPARKMAN

DATE/TIME OF TEST: 06/27/2025 02:53:55

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:49	
Control Test		0.081	02:49	
Air Blank		0.000	02:50	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:53	
Control Test		0.082	02:53	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 07/08/2025 21:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: BRINKLEY

DATE/TIME OF TEST: 06/27/2025 03:19:52

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.081	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.000	03:15
Breath Volume		2.324	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.000	03:18
Breath Volume		2.128	03:18
Air Blank		0.000	03:18
Control Test		0.082	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 07/08/2025 21:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:28

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 06/27/2025 08:54:35

Diagnostics Check		OK	08:48
Air Blank		0.000	08:49
Control Test		0.082	08:49
Air Blank		0.000	08:49
Subject Sample	# 1	0.000	08:50
Breath Volume		4.265	08:50
Air Blank		0.000	08:50
Air Blank		0.000	08:52
Subject Sample	# 2	0.000	08:53
Breath Volume		4.324	08:53
Air Blank		0.000	08:53
Control Test		0.080	08:53
Air Blank		0.000	08:54
Diagnostics Check		OK	08:54

Data Download Date/Time: 07/08/2025 21:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: STINE,S
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:28

SUBJECT'S LAST NAME: MUTRIE

DATE/TIME OF TEST: 06/27/2025 10:57:25

Diagnostics Check		OK	10:51
Air Blank		0.000	10:51
Control Test		0.082	10:52
Air Blank		0.000	10:52
Subject Sample	# 1	0.128	10:53
Breath Volume		1.191	10:53
Air Blank		0.000	10:53
Air Blank		0.000	10:55
Subject Sample	# 2	0.121	10:55
Breath Volume		1.527	10:55
Air Blank		0.000	10:56
Control Test		0.081	10:56
Air Blank		0.000	10:57
Diagnostics Check		OK	10:57

Data Download Date/Time: 07/08/2025 21:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: FORTUNATO

DATE/TIME OF TEST: 06/27/2025 22:05:35

Diagnostics Check		OK	21:57
Air Blank		0.000	21:58
Control Test		0.083	21:58
Air Blank		0.000	21:59
Subject Sample	# 1	0.146	22:00
Breath Volume		2.191	22:00
Air Blank		0.000	22:01
Air Blank		0.000	22:03
Subject Sample	# 2	0.143	22:04
Breath Volume		2.554	22:04
Air Blank		0.000	22:04
Control Test		0.081	22:04
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 07/08/2025 21:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: TAVERAS,S
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/28/2025 02:44:01

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.082	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.186	02:34
Breath Volume		2.320	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:36
Subject Sample	# 2	0.160	02:39
Breath Volume		2.042	02:39
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 3	0.174	02:42
Breath Volume		1.660	02:42
Air Blank		0.000	02:43
Control Test		0.080	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:44

Data Download Date/Time: 07/08/2025 21:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD,MITCHELL,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/28/2025 03:59:19

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.082	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.033	03:53
Breath Volume		1.160	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.029	03:57
Breath Volume		1.535	03:57
Air Blank		0.000	03:58
Control Test		0.080	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 07/08/2025 21:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: GRAMLEY

DATE/TIME OF TEST: 06/28/2025 18:39:28

Diagnostics Check		OK	18:29	
Air Blank		0.000	18:29	
Control Test		0.082	18:30	
Air Blank		0.000	18:30	
Subject Sample	# 1		18:33	No Sample Provided
Breath Volume		0.000	18:33	
Air Blank		0.000	18:34	
Air Blank		0.000	18:36	
Subject Sample	# 2		18:38	Subject Test Refused
Breath Volume		0.000	18:38	
Air Blank		0.000	18:38	
Control Test		0.082	18:38	
Air Blank		0.000	18:39	
Diagnostics Check		OK	18:39	

Data Download Date/Time: 07/08/2025 21:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLINE,NICOLE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BISHOP,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: HANNEY

DATE/TIME OF TEST: 06/28/2025 23:22:53

Diagnostics Check		OK	23:18	
Air Blank		0.000	23:18	
Control Test		0.082	23:18	
Air Blank		0.000	23:19	
Subject Sample	# 1		23:21	Subject Test Refused
Breath Volume		0.000	23:21	
Air Blank		0.000	23:21	
Control Test		0.082	23:22	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 07/08/2025 21:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KUSHNIR,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: KUTSYPAK

DATE/TIME OF TEST: 06/29/2025 21:26:20

Diagnostics Check		OK	21:14	
Air Blank		0.000	21:15	
Control Test		0.081	21:15	
Air Blank		0.000	21:16	
Subject Sample	# 1		21:19	No Sample Provided
Breath Volume		0.000	21:19	
Air Blank		0.000	21:19	
Air Blank		0.000	21:21	
Subject Sample	# 2		21:24	No Sample Provided
Breath Volume		0.000	21:24	
Air Blank		0.000	21:25	
Control Test		0.082	21:25	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:26	

Data Download Date/Time: 07/08/2025 21:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDOLO,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: MIDOLO,MICHAEL
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 06/11/2025 17:48:09

Diagnostics Check		OK	17:35	
Air Blank		0.000	17:35	
Control Test		0.080	17:36	
Air Blank		0.000	17:36	
Subject Sample	# 1	0.000	17:39	Volume Not Met
Breath Volume		0.039	17:39	
Air Blank		0.000	17:40	
Air Blank		0.000	17:42	
Subject Sample	# 2	0.000	17:43	
Breath Volume		2.558	17:43	
Air Blank		0.000	17:44	
Air Blank		0.000	17:46	
Subject Sample	# 3	0.000	17:46	
Breath Volume		1.871	17:46	
Air Blank		0.000	17:47	
Control Test		0.079	17:47	
Air Blank		0.000	17:48	
Diagnostics Check		OK	17:48	

Data Download Date/Time: 06/27/2025 11:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: REED,ELIJAH
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: TOMLIN

DATE/TIME OF TEST: 06/01/2025 22:00:39

Diagnostics Check		OK	21:53
Air Blank		0.000	21:54
Control Test		0.083	21:54
Air Blank		0.000	21:55
Subject Sample	# 1	0.288	21:55
Breath Volume		1.511	21:55
Air Blank		0.000	21:56
Air Blank		0.000	21:58
Subject Sample	# 2	0.295	21:58
Breath Volume		2.187	21:58
Air Blank		0.000	21:59
Control Test		0.082	22:00
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 06/29/2025 09:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SENECAL,JOSHUA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: SWEARENGEN

DATE/TIME OF TEST: 06/04/2025 23:20:49

Diagnostics Check		OK	23:17	
Air Blank		0.000	23:18	
Control Test		0.082	23:18	
Air Blank		0.000	23:19	
Subject Sample	# 1		23:19	Subject Test Refused
Breath Volume		0.000	23:19	
Air Blank		0.000	23:19	
Control Test		0.081	23:20	
Air Blank		0.000	23:20	
Diagnostics Check		OK	23:20	

Data Download Date/Time: 06/29/2025 09:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISSA,JAMES,D
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BOWMAN,MAXWELL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: HURTADO

DATE/TIME OF TEST: 06/06/2025 23:05:54

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.083	23:00
Air Blank		0.000	23:00
Subject Sample	# 1	0.132	23:01
Breath Volume		1.984	23:01
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.123	23:04
Breath Volume		1.656	23:04
Air Blank		0.000	23:04
Control Test		0.082	23:05
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 06/29/2025 09:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,RYAN,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: STACY,N
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/07/2025 06:47:32

Diagnostics Check		OK	06:40
Air Blank		0.000	06:40
Control Test		0.083	06:41
Air Blank		0.000	06:41
Subject Sample	# 1	0.077	06:42
Breath Volume		2.421	06:42
Air Blank		0.000	06:43
Air Blank		0.000	06:44
Subject Sample	# 2	0.078	06:45
Breath Volume		1.789	06:45
Air Blank		0.000	06:46
Control Test		0.082	06:46
Air Blank		0.000	06:47
Diagnostics Check		OK	06:47

Data Download Date/Time: 06/29/2025 09:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSH,WESLEY,T
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BROWN,ZECHARIAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: SMYTHURST

DATE/TIME OF TEST: 06/12/2025 04:19:37

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.082	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.000	04:15
Breath Volume		1.777	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.000	04:18
Breath Volume		1.136	04:18
Air Blank		0.000	04:18
Control Test		0.082	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 06/29/2025 09:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: FROST,JOSEPH
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: BOSWORTH

DATE/TIME OF TEST: 06/14/2025 01:16:42

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:14	
Control Test		0.082	01:14	
Air Blank		0.000	01:15	
Subject Sample	# 1		01:15	Subject Test Refused
Breath Volume		0.000	01:15	
Air Blank		0.000	01:15	
Control Test		0.081	01:16	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:16	

Data Download Date/Time: 06/29/2025 09:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: AVIS,JENNIFER
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PUELLO

DATE/TIME OF TEST: 06/14/2025 02:56:05

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.082	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.206	02:51
Breath Volume		3.558	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.195	02:54
Breath Volume		2.925	02:54
Air Blank		0.000	02:55
Control Test		0.081	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 06/29/2025 09:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: AVIS,JENNIFER
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 06/20/2025 00:27:28

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.082	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.239	00:22
Breath Volume		2.781	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:24
Subject Sample	# 2	0.228	00:25
Breath Volume		2.828	00:25
Air Blank		0.000	00:26
Control Test		0.081	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 06/29/2025 09:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DUNNE,JUSTIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: ADMIN
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: MANAF

DATE/TIME OF TEST: 06/24/2025 11:06:01

Diagnostics Check		OK	10:58
Air Blank		0.000	10:59
Control Test		0.083	10:59
Air Blank		0.000	10:59
Subject Sample	# 1	0.023	11:01
Breath Volume		1.265	11:01
Air Blank		0.000	11:01
Air Blank		0.000	11:03
Subject Sample	# 2	0.023	11:04
Breath Volume		1.429	11:04
Air Blank		0.000	11:05
Control Test		0.082	11:05
Air Blank		0.000	11:05
Diagnostics Check		OK	11:06

Data Download Date/Time: 06/29/2025 09:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SALZA,GUENTHER'HENRY
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: DYER

DATE/TIME OF TEST: 06/25/2025 20:39:19

Diagnostics Check		OK	20:32
Air Blank		0.000	20:33
Control Test		0.082	20:33
Air Blank		0.000	20:34
Subject Sample	# 1	0.164	20:34
Breath Volume		1.601	20:34
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.162	20:37
Breath Volume		1.746	20:37
Air Blank		0.000	20:38
Control Test		0.081	20:38
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 06/29/2025 09:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GARCIA,A
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: CUARTAS

DATE/TIME OF TEST: 06/01/2025 01:09:10

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.076	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.216	01:04
Breath Volume		1.613	01:04
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.215	01:07
Breath Volume		1.156	01:07
Air Blank		0.000	01:08
Control Test		0.078	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 06/05/2025 02:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MALTEZ LOPEZ

DATE/TIME OF TEST: 06/07/2025 02:29:20

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.080	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.122	02:24
Breath Volume		1.507	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.115	02:27
Breath Volume		1.207	02:27
Air Blank		0.000	02:28
Control Test		0.078	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 07/10/2025 03:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: CUERVO FERRERA

DATE/TIME OF TEST: 06/26/2025 04:32:44

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:30	
Control Test		0.075	04:30	
Air Blank		0.000	04:31	
Subject Sample	# 1		04:31	Subject Test Refused
Breath Volume		0.000	04:31	
Air Blank		0.000	04:31	
Control Test		0.075	04:32	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 07/10/2025 03:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LAPE,RYAN
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 06/27/2025 22:45:54

Diagnostics Check		OK	22:43	
Air Blank		0.000	22:43	
Control Test		0.075	22:43	
Air Blank		0.000	22:44	
Subject Sample	# 1		22:44	Subject Test Refused
Breath Volume		0.000	22:44	
Air Blank		0.000	22:44	
Control Test		0.077	22:45	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 07/10/2025 03:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: JIMENEZ-RODRIGUEZ

DATE/TIME OF TEST: 06/14/2025 20:34:47

Diagnostics Check		OK	20:27
Air Blank		0.000	20:27
Control Test		0.078	20:28
Air Blank		0.000	20:28
Subject Sample	# 1	0.040	20:29
Breath Volume		1.191	20:29
Air Blank		0.000	20:29
Air Blank		0.000	20:31
Subject Sample	# 2	0.045	20:33
Breath Volume		1.429	20:33
Air Blank		0.000	20:33
Control Test		0.079	20:34
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 07/10/2025 03:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: VAZQUEZ-VILLEGAS

DATE/TIME OF TEST: 06/21/2025 00:56:05

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:40	
Control Test		0.079	00:41	
Air Blank		0.000	00:41	
Subject Sample	# 1	0.243	00:43	
Breath Volume		1.488	00:43	
Air Blank		0.000	00:43	
Air Blank		0.000	00:45	
Subject Sample	# 2	0.219	00:48	Volume Not Met
Breath Volume		0.082	00:48	
Air Blank		0.000	00:49	
Air Blank		0.000	00:51	
Subject Sample	# 3	0.222	00:54	Volume Not Met
Breath Volume		0.000	00:54	
Air Blank		0.000	00:55	
Control Test		0.076	00:55	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 07/10/2025 03:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: VAZQUEZ-VILLEGAS

DATE/TIME OF TEST: 06/21/2025 01:06:44

Diagnostics Check		OK	01:03	
Air Blank		0.000	01:04	
Control Test		0.077	01:04	
Air Blank		0.000	01:05	
Subject Sample	# 1		01:05	Subject Test Refused
Breath Volume		0.000	01:05	
Air Blank		0.000	01:05	
Control Test		0.077	01:06	
Air Blank		0.000	01:06	
Diagnostics Check		OK	01:06	

Data Download Date/Time: 07/10/2025 03:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/21/2025 04:32:45

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.079	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.060	04:27
Breath Volume		2.074	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:29
Subject Sample	# 2	0.057	04:31
Breath Volume		1.667	04:31
Air Blank		0.000	04:31
Control Test		0.077	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 07/10/2025 03:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/21/2025 05:00:26

Diagnostics Check		OK	04:53
Air Blank		0.000	04:54
Control Test		0.078	04:54
Air Blank		0.000	04:55
Subject Sample	# 1	0.114	04:55
Breath Volume		2.011	04:55
Air Blank		0.000	04:56
Air Blank		0.000	04:57
Subject Sample	# 2	0.114	04:58
Breath Volume		2.214	04:58
Air Blank		0.000	04:59
Control Test		0.078	04:59
Air Blank		0.000	05:00
Diagnostics Check		OK	05:00

Data Download Date/Time: 07/10/2025 03:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CALLEJAS

DATE/TIME OF TEST: 06/22/2025 03:25:53

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:17	
Control Test		0.080	03:18	
Air Blank		0.000	03:18	
Subject Sample	# 1		03:21	No Sample Provided
Breath Volume		0.000	03:21	
Air Blank		0.000	03:22	
Air Blank		0.000	03:24	
Subject Sample	# 2		03:24	Subject Test Refused
Breath Volume		0.000	03:24	
Air Blank		0.000	03:24	
Control Test		0.079	03:25	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 07/10/2025 03:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OZETE-FONTE,BARBARA
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 06/28/2025 00:58:57

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.077	00:53
Air Blank		0.000	00:54
Subject Sample	# 1	0.106	00:54
Breath Volume		2.671	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.109	00:57
Breath Volume		1.898	00:57
Air Blank		0.000	00:58
Control Test		0.077	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 07/10/2025 03:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: FOMON

DATE/TIME OF TEST: 06/01/2025 02:15:04

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:06	
Control Test		0.079	02:06	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:10	No Sample Provided
Breath Volume		0.000	02:10	
Air Blank		0.000	02:10	
Air Blank		0.000	02:12	
Subject Sample	# 2		02:13	Subject Test Refused
Breath Volume		0.000	02:13	
Air Blank		0.000	02:14	
Control Test		0.080	02:14	
Air Blank		0.000	02:14	
Diagnostics Check		OK	02:15	

Data Download Date/Time: 06/05/2025 02:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 06/06/2025 02:27:23

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:24	
Control Test		0.079	02:24	
Air Blank		0.000	02:25	
Subject Sample	# 1		02:25	Subject Test Refused
Breath Volume		0.000	02:25	
Air Blank		0.000	02:26	
Control Test		0.079	02:26	
Air Blank		0.000	02:27	
Diagnostics Check		OK	02:27	

Data Download Date/Time: 07/10/2025 04:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: MIRO

DATE/TIME OF TEST: 06/13/2025 00:22:35

Diagnostics Check		OK	00:19	
Air Blank		0.000	00:20	
Control Test		0.079	00:20	
Air Blank		0.000	00:20	
Subject Sample	# 1		00:21	Subject Test Refused
Breath Volume		0.000	00:21	
Air Blank		0.000	00:21	
Control Test		0.081	00:21	
Air Blank		0.000	00:22	
Diagnostics Check		OK	00:22	

Data Download Date/Time: 07/10/2025 04:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: VARGAS-FLORES

DATE/TIME OF TEST: 06/14/2025 02:07:18

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.077	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.141	02:02
Breath Volume		1.203	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.139	02:05
Breath Volume		1.757	02:05
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 07/10/2025 04:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 06/15/2025 00:55:56

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.081	00:49
Air Blank		0.000	00:50
Subject Sample	# 1	0.127	00:50
Breath Volume		1.296	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.127	00:54
Breath Volume		1.738	00:54
Air Blank		0.000	00:54
Control Test		0.082	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 07/10/2025 04:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: EKMEIRO-TROCONIS

DATE/TIME OF TEST: 06/27/2025 02:12:05

Diagnostics Check		OK	02:09	
Air Blank		0.000	02:09	
Control Test		0.079	02:09	
Air Blank		0.000	02:10	
Subject Sample	# 1		02:10	Subject Test Refused
Breath Volume		0.000	02:10	
Air Blank		0.000	02:11	
Control Test		0.079	02:11	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:12	

Data Download Date/Time: 07/10/2025 04:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: BRADFORD

DATE/TIME OF TEST: 06/27/2025 23:17:48

Diagnostics Check		OK	23:08
Air Blank		0.000	23:09
Control Test		0.079	23:09
Air Blank		0.000	23:09
Subject Sample	# 1	0.292	23:10
Breath Volume		2.402	23:10
Air Blank		0.000	23:11
Air Blank		0.000	23:12
Subject Sample	# 2	0.268	23:13
Breath Volume		1.750	23:13
Air Blank		0.000	23:14
Air Blank		0.000	23:15
Subject Sample	# 3	0.282	23:16
Breath Volume		2.148	23:16
Air Blank		0.000	23:16
Control Test		0.078	23:17
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 07/10/2025 13:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/28/2025 02:30:21

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.250	02:22
Breath Volume		1.656	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.229	02:25
Breath Volume		1.691	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 3	0.239	02:28
Breath Volume		1.507	02:28
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 07/10/2025 13:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 06/06/2025 22:43:12

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.078	22:37
Air Blank		0.000	22:38
Subject Sample	# 1	0.000	22:38
Breath Volume		2.402	22:38
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.000	22:41
Breath Volume		3.332	22:41
Air Blank		0.000	22:42
Control Test		0.079	22:42
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 06/20/2025 14:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: HEWITT

DATE/TIME OF TEST: 06/07/2025 04:51:34

Diagnostics Check		OK	04:43
Air Blank		0.000	04:44
Control Test		0.078	04:44
Air Blank		0.000	04:45
Subject Sample	# 1	0.184	04:46
Breath Volume		1.261	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.191	04:49
Breath Volume		1.292	04:49
Air Blank		0.000	04:50
Control Test		0.078	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 06/20/2025 14:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARKER,JEFF
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: MCDOWELL

DATE/TIME OF TEST: 06/09/2025 01:39:12

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.079	01:32
Air Blank		0.000	01:33
Subject Sample	# 1	0.185	01:34
Breath Volume		1.625	01:34
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.179	01:37
Breath Volume		1.355	01:37
Air Blank		0.000	01:38
Control Test		0.079	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 06/20/2025 14:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RICCI,MATTHEW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 06/10/2025 00:43:18

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.078	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.143	00:38
Breath Volume		1.582	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.143	00:41
Breath Volume		1.824	00:41
Air Blank		0.000	00:42
Control Test		0.078	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 06/20/2025 14:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: WOOLVERTON

DATE/TIME OF TEST: 06/13/2025 01:05:06

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.079	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.253	01:00
Breath Volume		2.882	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:02
Subject Sample	# 2	0.247	01:03
Breath Volume		2.019	01:03
Air Blank		0.000	01:04
Control Test		0.080	01:04
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 06/20/2025 14:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 06/13/2025 04:01:28

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:58	
Control Test		0.079	03:58	
Air Blank		0.000	03:59	
Subject Sample	# 1		04:00	Subject Test Refused
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Control Test		0.080	04:00	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 06/20/2025 14:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: SCHWAB

DATE/TIME OF TEST: 06/14/2025 00:24:48

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.079	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.000	00:20
Breath Volume		2.699	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.000	00:23
Breath Volume		2.792	00:23
Air Blank		0.000	00:23
Control Test		0.079	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 06/20/2025 14:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 06/14/2025 05:52:31

Diagnostics Check		OK	05:49	
Air Blank		0.000	05:50	
Control Test		0.079	05:50	
Air Blank		0.000	05:50	
Subject Sample	# 1		05:51	Subject Test Refused
Breath Volume		0.000	05:51	
Air Blank		0.000	05:51	
Control Test		0.079	05:51	
Air Blank		0.000	05:52	
Diagnostics Check		OK	05:52	

Data Download Date/Time: 06/20/2025 14:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 06/14/2025 20:47:32

Diagnostics Check		OK	20:39	
Air Blank		0.000	20:39	
Control Test		0.080	20:39	
Air Blank		0.000	20:40	
Subject Sample	# 1		20:43	No Sample Provided
Breath Volume		0.000	20:43	
Air Blank		0.000	20:43	
Air Blank		0.000	20:45	
Subject Sample	# 2		20:46	Subject Test Refused
Breath Volume		0.000	20:46	
Air Blank		0.000	20:46	
Control Test		0.079	20:46	
Air Blank		0.000	20:47	
Diagnostics Check		OK	20:47	

Data Download Date/Time: 06/20/2025 14:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 06/15/2025 03:27:08

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.079	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.053	03:22
Breath Volume		2.132	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.051	03:25
Breath Volume		2.972	03:25
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 06/20/2025 14:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: LOGAN

DATE/TIME OF TEST: 06/18/2025 22:44:38

Diagnostics Check		OK	22:38
Air Blank		0.000	22:38
Control Test		0.080	22:39
Air Blank		0.000	22:39
Subject Sample	# 1	0.000	22:40
Breath Volume		1.710	22:40
Air Blank		0.000	22:40
Air Blank		0.000	22:42
Subject Sample	# 2	0.000	22:43
Breath Volume		2.089	22:43
Air Blank		0.000	22:43
Control Test		0.079	22:44
Air Blank		0.000	22:44
Diagnostics Check		OK	22:44

Data Download Date/Time: 06/20/2025 14:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: SAAVEDRA MARTINEZ

DATE/TIME OF TEST: 06/19/2025 04:17:44

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.079	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.150	04:13
Breath Volume		1.699	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.156	04:16
Breath Volume		1.328	04:16
Air Blank		0.000	04:16
Control Test		0.079	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 06/20/2025 14:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RODRIGUEZ,ANDRES
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ANDRES-RAMIREZ

DATE/TIME OF TEST: 06/21/2025 01:58:50

Diagnostics Check		OK	01:51
Air Blank		0.000	01:52
Control Test		0.080	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.189	01:54
Breath Volume		1.328	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.191	01:57
Breath Volume		1.433	01:57
Air Blank		0.000	01:57
Control Test		0.080	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 07/11/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: HERNANDEZ ZEPEDA

DATE/TIME OF TEST: 06/21/2025 03:36:09

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.164	03:31
Breath Volume		3.687	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.159	03:34
Breath Volume		3.222	03:34
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 07/11/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE,RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: MCGINN

DATE/TIME OF TEST: 06/21/2025 22:04:08

Diagnostics Check		OK	21:57
Air Blank		0.000	21:58
Control Test		0.080	21:58
Air Blank		0.000	21:59
Subject Sample	# 1	0.120	21:59
Breath Volume		1.292	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.120	22:02
Breath Volume		1.722	22:02
Air Blank		0.000	22:03
Control Test		0.080	22:03
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 07/11/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: MATIAS LOPEZ

DATE/TIME OF TEST: 06/21/2025 22:43:34

Diagnostics Check		OK	22:35
Air Blank		0.000	22:35
Control Test		0.078	22:36
Air Blank		0.000	22:36
Subject Sample	# 1	0.139	22:38
Breath Volume		2.097	22:38
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.134	22:42
Breath Volume		1.652	22:42
Air Blank		0.000	22:42
Control Test		0.078	22:42
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 07/11/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: LIAN

DATE/TIME OF TEST: 06/22/2025 00:15:33

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.079	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.177	00:11
Breath Volume		1.378	00:11
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.183	00:14
Breath Volume		2.089	00:14
Air Blank		0.000	00:14
Control Test		0.079	00:14
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 07/11/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: GRAMAJO DELEON

DATE/TIME OF TEST: 06/22/2025 00:40:51

Diagnostics Check		OK	00:33
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.194	00:35
Breath Volume		1.152	00:35
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.194	00:39
Breath Volume		1.140	00:39
Air Blank		0.000	00:39
Control Test		0.079	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 07/11/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIER,MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: DELEO

DATE/TIME OF TEST: 06/22/2025 02:56:19

Diagnostics Check		OK	02:46
Air Blank		0.000	02:47
Control Test		0.079	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.201	02:48
Breath Volume		2.183	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.180	02:51
Breath Volume		2.156	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 3	0.187	02:54
Breath Volume		1.929	02:54
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 07/11/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: GUTIERREZ DE LA CRUZ

DATE/TIME OF TEST: 06/22/2025 04:51:42

Diagnostics Check		OK	04:44
Air Blank		0.000	04:44
Control Test		0.079	04:45
Air Blank		0.000	04:45
Subject Sample	# 1	0.168	04:46
Breath Volume		1.640	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:48
Subject Sample	# 2	0.174	04:50
Breath Volume		1.878	04:50
Air Blank		0.000	04:50
Control Test		0.080	04:51
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 07/11/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 06/23/2025 01:03:27

Diagnostics Check		OK	00:56
Air Blank		0.000	00:57
Control Test		0.080	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.161	00:58
Breath Volume		1.437	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.168	01:01
Breath Volume		3.378	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 07/11/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: EVINS

DATE/TIME OF TEST: 06/24/2025 02:24:20

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.080	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.000	02:19
Breath Volume		2.273	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.000	02:22
Breath Volume		4.640	02:22
Air Blank		0.000	02:23
Control Test		0.080	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 07/11/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 06/24/2025 03:30:14

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.080	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.000	03:25
Breath Volume		1.164	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.000	03:28
Breath Volume		1.234	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 07/11/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: MAIDEN

DATE/TIME OF TEST: 06/24/2025 20:15:14

Diagnostics Check		OK	20:08
Air Blank		0.000	20:09
Control Test		0.080	20:09
Air Blank		0.000	20:10
Subject Sample	# 1	0.268	20:10
Breath Volume		2.347	20:10
Air Blank		0.000	20:11
Air Blank		0.000	20:13
Subject Sample	# 2	0.263	20:13
Breath Volume		2.523	20:13
Air Blank		0.000	20:14
Control Test		0.078	20:14
Air Blank		0.000	20:15
Diagnostics Check		OK	20:15

Data Download Date/Time: 07/11/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: RUIZ AGUILAR

DATE/TIME OF TEST: 06/24/2025 23:28:46

Diagnostics Check		OK	23:20
Air Blank		0.000	23:21
Control Test		0.081	23:21
Air Blank		0.000	23:22
Subject Sample	# 1	0.178	23:24
Breath Volume		1.492	23:24
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.186	23:27
Breath Volume		1.480	23:27
Air Blank		0.000	23:27
Control Test		0.078	23:28
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 07/11/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PALMER,NEIL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: BENEDICT

DATE/TIME OF TEST: 06/25/2025 04:02:17

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.170	03:57
Breath Volume		1.550	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.173	04:00
Breath Volume		1.507	04:00
Air Blank		0.000	04:01
Control Test		0.079	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 07/11/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: EUSEBIO

DATE/TIME OF TEST: 06/26/2025 23:25:00

Diagnostics Check		OK	23:19
Air Blank		0.000	23:19
Control Test		0.083	23:19
Air Blank		0.000	23:20
Subject Sample	# 1	0.144	23:20
Breath Volume		2.078	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.150	23:23
Breath Volume		2.187	23:23
Air Blank		0.000	23:24
Control Test		0.080	23:24
Air Blank		0.000	23:24
Diagnostics Check		OK	23:25

Data Download Date/Time: 07/11/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SALES-LOPEZ,KEVIN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: NAJERA HERNANDEZ

DATE/TIME OF TEST: 06/28/2025 02:34:55

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.216	02:30
Breath Volume		1.269	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 2	0.198	02:33
Breath Volume		1.695	02:33
Air Blank		0.000	02:34
Control Test		0.080	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 07/11/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: ENCARNACION

DATE/TIME OF TEST: 06/28/2025 04:38:16

Diagnostics Check		OK	04:31
Air Blank		0.000	04:32
Control Test		0.079	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.077	04:33
Breath Volume		2.835	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.081	04:36
Breath Volume		2.675	04:36
Air Blank		0.000	04:37
Control Test		0.080	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 07/11/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LONDONO,JULIAN
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: TELFORD

DATE/TIME OF TEST: 06/28/2025 22:45:27

Diagnostics Check		OK	22:37
Air Blank		0.000	22:38
Control Test		0.079	22:38
Air Blank		0.000	22:39
Subject Sample	# 1	0.159	22:40
Breath Volume		1.914	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.166	22:43
Breath Volume		2.136	22:43
Air Blank		0.000	22:44
Control Test		0.078	22:44
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 07/11/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: URBINA TORREZ

DATE/TIME OF TEST: 06/28/2025 23:45:42

Diagnostics Check		OK	23:39
Air Blank		0.000	23:39
Control Test		0.079	23:40
Air Blank		0.000	23:40
Subject Sample	# 1	0.217	23:41
Breath Volume		1.648	23:41
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.207	23:44
Breath Volume		2.738	23:44
Air Blank		0.000	23:44
Control Test		0.080	23:45
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 07/11/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CURRERI,ETHAN
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: FEINSTEIN

DATE/TIME OF TEST: 06/29/2025 00:45:00

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.079	00:39
Air Blank		0.000	00:39
Subject Sample	# 1	0.231	00:40
Breath Volume		1.679	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:42
Subject Sample	# 2	0.215	00:43
Breath Volume		1.125	00:43
Air Blank		0.000	00:44
Control Test		0.079	00:44
Air Blank		0.000	00:44
Diagnostics Check		OK	00:45

Data Download Date/Time: 07/11/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: BRUNOT

DATE/TIME OF TEST: 06/29/2025 01:58:41

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.080	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.000	01:54
Breath Volume		1.191	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.000	01:57
Breath Volume		1.453	01:57
Air Blank		0.000	01:57
Control Test		0.080	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 07/11/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: ELFRARD

DATE/TIME OF TEST: 06/29/2025 05:42:28

Diagnostics Check		OK	05:30	
Air Blank		0.000	05:31	
Control Test		0.079	05:31	
Air Blank		0.000	05:32	
Subject Sample	# 1	0.109	05:35	Volume Not Met
Breath Volume		0.957	05:35	
Air Blank		0.000	05:35	
Air Blank		0.000	05:37	
Subject Sample	# 2	0.105	05:40	Volume Not Met
Breath Volume		0.218	05:40	
Air Blank		0.000	05:41	
Control Test		0.079	05:41	
Air Blank		0.000	05:42	
Diagnostics Check		OK	05:42	

Data Download Date/Time: 07/11/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: ELFRARD

DATE/TIME OF TEST: 06/29/2025 05:53:51

Diagnostics Check		OK	05:45
Air Blank		0.000	05:46
Control Test		0.079	05:46
Air Blank		0.000	05:47
Subject Sample	# 1	0.105	05:49
Breath Volume		1.113	05:49
Air Blank		0.000	05:49
Air Blank		0.000	05:51
Subject Sample	# 2	0.104	05:52
Breath Volume		1.281	05:52
Air Blank		0.000	05:52
Control Test		0.078	05:53
Air Blank		0.000	05:53
Diagnostics Check		OK	05:53

Data Download Date/Time: 07/11/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 06/30/2025 04:32:37

Diagnostics Check		OK	04:17	
Air Blank		0.000	04:17	
Control Test		0.080	04:17	
Air Blank		0.000	04:18	
Subject Sample	# 1	0.176	04:21	Volume Not Met
Breath Volume		0.105	04:21	
Air Blank		0.000	04:22	
Air Blank		0.000	04:24	
Subject Sample	# 2	0.191	04:25	
Breath Volume		1.417	04:25	
Air Blank		0.000	04:26	
Air Blank		0.000	04:27	
Subject Sample	# 3	0.194	04:30	
Breath Volume		1.500	04:30	
Air Blank		0.000	04:31	
Control Test		0.079	04:31	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 07/11/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: GOUSSE

DATE/TIME OF TEST: 06/30/2025 19:34:29

Diagnostics Check		OK	19:27
Air Blank		0.000	19:28
Control Test		0.079	19:28
Air Blank		0.000	19:29
Subject Sample	# 1	0.000	19:30
Breath Volume		1.371	19:30
Air Blank		0.000	19:30
Air Blank		0.000	19:32
Subject Sample	# 2	0.000	19:33
Breath Volume		1.199	19:33
Air Blank		0.000	19:33
Control Test		0.079	19:33
Air Blank		0.000	19:34
Diagnostics Check		OK	19:34

Data Download Date/Time: 07/11/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: JEFFRIES

DATE/TIME OF TEST: 06/30/2025 23:28:17

Diagnostics Check		OK	23:22
Air Blank		0.000	23:22
Control Test		0.080	23:22
Air Blank		0.000	23:23
Subject Sample	# 1	0.209	23:23
Breath Volume		1.335	23:23
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.214	23:26
Breath Volume		1.503	23:26
Air Blank		0.000	23:27
Control Test		0.079	23:27
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 07/11/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,ERIC,C
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LOTT,ERIC
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: VENTULAN

DATE/TIME OF TEST: 06/01/2025 23:28:41

Diagnostics Check		OK	23:15	
Air Blank		0.000	23:15	
Control Test		0.079	23:15	
Air Blank		0.000	23:16	
Subject Sample	# 1	0.196	23:19	Volume Not Met
Breath Volume		0.085	23:19	
Air Blank		0.000	23:20	
Air Blank		0.000	23:22	
Subject Sample	# 2	0.189	23:23	
Breath Volume		2.035	23:23	
Air Blank		0.000	23:23	
Air Blank		0.000	23:25	
Subject Sample	# 3	0.183	23:27	
Breath Volume		1.132	23:27	
Air Blank		0.000	23:27	
Control Test		0.079	23:28	
Air Blank		0.000	23:28	
Diagnostics Check		OK	23:28	

Data Download Date/Time: 06/22/2025 18:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CADWELL,WILLIAM,M
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:23

SUBJECT'S LAST NAME: CADWELL

DATE/TIME OF TEST: 06/02/2025 14:53:38

Diagnostics Check		OK	14:46	
Air Blank		0.000	14:47	
Control Test		0.081	14:47	
Air Blank		0.000	14:48	
Subject Sample	# 1	0.105	14:49	
Breath Volume		3.175	14:49	
Air Blank		0.000	14:49	
Air Blank		0.000	14:51	
Subject Sample	# 2	0.107	14:51	
Breath Volume		3.230	14:51	
Air Blank		0.000	14:52	
Control Test		0.088	14:52	Control Outside Tolerance
Air Blank		0.000	14:53	

Data Download Date/Time: 06/22/2025 18:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOURNE,RUSSELL,L
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: BURKE,LOGAN
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: REINER

DATE/TIME OF TEST: 06/03/2025 03:46:31

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.080	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.000	03:41
Breath Volume		2.800	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.000	03:45
Breath Volume		2.253	03:45
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 06/22/2025 18:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE,ADAM,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE,ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GILSON

DATE/TIME OF TEST: 06/04/2025 03:25:11

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.143	03:20
Breath Volume		1.714	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.147	03:23
Breath Volume		1.191	03:23
Air Blank		0.000	03:24
Control Test		0.080	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 06/22/2025 18:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/06/2025 17:47:29

Diagnostics Check		OK	17:41
Air Blank		0.000	17:41
Control Test		0.080	17:41
Air Blank		0.000	17:42
Subject Sample	# 1	0.126	17:42
Breath Volume		2.218	17:42
Air Blank		0.000	17:43
Air Blank		0.000	17:45
Subject Sample	# 2	0.122	17:45
Breath Volume		1.953	17:45
Air Blank		0.000	17:46
Control Test		0.080	17:46
Air Blank		0.000	17:47
Diagnostics Check		OK	17:47

Data Download Date/Time: 06/22/2025 18:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:09

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/07/2025 11:36:48

Diagnostics Check		OK	11:30
Air Blank		0.000	11:30
Control Test		0.081	11:31
Air Blank		0.000	11:31
Subject Sample	# 1	0.000	11:32
Breath Volume		2.121	11:32
Air Blank		0.000	11:32
Air Blank		0.000	11:34
Subject Sample	# 2	0.000	11:35
Breath Volume		2.117	11:35
Air Blank		0.000	11:35
Control Test		0.080	11:36
Air Blank		0.000	11:36
Diagnostics Check		OK	11:36

Data Download Date/Time: 06/22/2025 18:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: ROYAL

DATE/TIME OF TEST: 06/08/2025 17:45:16

Diagnostics Check		OK	17:38
Air Blank		0.000	17:39
Control Test		0.080	17:39
Air Blank		0.000	17:40
Subject Sample	# 1	0.012	17:40
Breath Volume		3.144	17:40
Air Blank		0.000	17:41
Air Blank		0.000	17:43
Subject Sample	# 2	0.012	17:43
Breath Volume		3.632	17:43
Air Blank		0.000	17:44
Control Test		0.081	17:44
Air Blank		0.000	17:45
Diagnostics Check		OK	17:45

Data Download Date/Time: 06/22/2025 18:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: HAGBERG

DATE/TIME OF TEST: 06/08/2025 20:53:15

Diagnostics Check		OK	20:46
Air Blank		0.000	20:47
Control Test		0.080	20:47
Air Blank		0.000	20:48
Subject Sample	# 1	0.077	20:48
Breath Volume		4.468	20:48
Air Blank		0.000	20:49
Air Blank		0.000	20:51
Subject Sample	# 2	0.079	20:51
Breath Volume		4.335	20:51
Air Blank		0.000	20:52
Control Test		0.079	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 06/22/2025 18:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEBLAKER,JOSHUA,N
BREATH TEST OPERATOR AGENCY: Sewall's Point PD
ARRESTING OFFICER: DEBLAKER,JOSHUA
ARRESTING OFFICER AGENCY: Sewall's Point PD
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: POUCHIE

DATE/TIME OF TEST: 06/09/2025 20:57:56

Diagnostics Check		OK	20:51
Air Blank		0.000	20:52
Control Test		0.079	20:52
Air Blank		0.000	20:52
Subject Sample	# 1	0.147	20:53
Breath Volume		2.632	20:53
Air Blank		0.000	20:54
Air Blank		0.000	20:55
Subject Sample	# 2	0.144	20:56
Breath Volume		2.898	20:56
Air Blank		0.000	20:56
Control Test		0.078	20:57
Air Blank		0.000	20:57
Diagnostics Check		OK	20:57

Data Download Date/Time: 06/22/2025 18:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,BRYAN,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MARTINI,BRYAN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:52

SUBJECT'S LAST NAME: SINASAC

DATE/TIME OF TEST: 06/11/2025 14:35:28

Diagnostics Check		OK	14:28
Air Blank		0.000	14:29
Control Test		0.080	14:29
Air Blank		0.000	14:30
Subject Sample	# 1	0.204	14:30
Breath Volume		2.667	14:30
Air Blank		0.000	14:31
Air Blank		0.000	14:33
Subject Sample	# 2	0.204	14:33
Breath Volume		2.765	14:33
Air Blank		0.000	14:34
Control Test		0.079	14:34
Air Blank		0.000	14:35
Diagnostics Check		OK	14:35

Data Download Date/Time: 06/22/2025 18:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: RUBIO-FACIO

DATE/TIME OF TEST: 06/12/2025 01:05:08

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.000	01:00
Breath Volume		3.410	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.000	01:03
Breath Volume		1.695	01:03
Air Blank		0.000	01:04
Control Test		0.080	01:04
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 06/22/2025 18:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: HERMANSKY

DATE/TIME OF TEST: 06/12/2025 18:30:23

Diagnostics Check		OK	18:23
Air Blank		0.000	18:24
Control Test		0.080	18:24
Air Blank		0.000	18:24
Subject Sample	# 1	0.000	18:25
Breath Volume		2.597	18:25
Air Blank		0.000	18:25
Air Blank		0.000	18:27
Subject Sample	# 2	0.000	18:29
Breath Volume		2.347	18:29
Air Blank		0.000	18:29
Control Test		0.080	18:29
Air Blank		0.000	18:30
Diagnostics Check		OK	18:30

Data Download Date/Time: 06/22/2025 18:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: CRUZ DIEGO

DATE/TIME OF TEST: 06/14/2025 21:01:00

Diagnostics Check		OK	20:54
Air Blank		0.000	20:54
Control Test		0.081	20:54
Air Blank		0.000	20:55
Subject Sample	# 1	0.224	20:55
Breath Volume		1.699	20:55
Air Blank		0.000	20:56
Air Blank		0.000	20:58
Subject Sample	# 2	0.214	20:59
Breath Volume		1.906	20:59
Air Blank		0.000	21:00
Control Test		0.081	21:00
Air Blank		0.000	21:00
Diagnostics Check		OK	21:01

Data Download Date/Time: 06/22/2025 18:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUFFNER,PRESTON,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: KUFFNER,PRESTON
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: STAUDER

DATE/TIME OF TEST: 06/16/2025 02:08:55

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.080	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.181	02:04
Breath Volume		1.332	02:04
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.181	02:07
Breath Volume		2.015	02:07
Air Blank		0.000	02:07
Control Test		0.079	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 06/22/2025 18:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: VILLATORO LOPEZ

DATE/TIME OF TEST: 06/16/2025 02:46:16

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.081	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.142	02:41
Breath Volume		1.285	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:43
Subject Sample	# 2	0.153	02:44
Breath Volume		2.542	02:44
Air Blank		0.000	02:45
Control Test		0.080	02:45
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 06/22/2025 18:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: WALCK

DATE/TIME OF TEST: 06/16/2025 22:23:56

Diagnostics Check		OK	22:17
Air Blank		0.000	22:17
Control Test		0.080	22:18
Air Blank		0.000	22:18
Subject Sample	# 1	0.304	22:19
Breath Volume		1.890	22:19
Air Blank		0.000	22:20
Air Blank		0.000	22:21
Subject Sample	# 2	0.303	22:22
Breath Volume		1.746	22:22
Air Blank		0.000	22:23
Control Test		0.079	22:23
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 06/22/2025 18:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: BRAVO CINTO

DATE/TIME OF TEST: 06/17/2025 22:27:21

Diagnostics Check		OK	22:20
Air Blank		0.000	22:20
Control Test		0.079	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.262	22:22
Breath Volume		2.292	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:24
Subject Sample	# 2	0.273	22:25
Breath Volume		2.644	22:25
Air Blank		0.000	22:26
Control Test		0.080	22:26
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 06/22/2025 18:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.