

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,JAKOB,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LANIER,Z
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: WEIMER

DATE/TIME OF TEST: 06/01/2025 01:19:36

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.077	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.091	01:14
Breath Volume	1.660	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.092	01:17
Breath Volume	3.332	01:17
Air Blank	0.000	01:18
Control Test	0.077	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 06/27/2025 06:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWMAN,T,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KURTH,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:47

SUBJECT'S LAST NAME: DECKER

DATE/TIME OF TEST: 06/04/2025 14:15:14

Diagnostics Check	OK	14:07
Air Blank	0.000	14:07
Control Test	0.077	14:07
Air Blank	0.000	14:08
Subject Sample # 1	0.232	14:10
Breath Volume	1.253	14:10
Air Blank	0.000	14:10
Air Blank	0.000	14:12
Subject Sample # 2	0.238	14:13
Breath Volume	2.093	14:13
Air Blank	0.000	14:14
Control Test	0.076	14:14
Air Blank	0.000	14:15
Diagnostics Check	OK	14:15

Data Download Date/Time: 06/27/2025 06:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD1619,A,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MURTO,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 06/07/2025 01:00:38

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.078	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.102	00:55
Breath Volume	2.218	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.101	00:59
Breath Volume	3.753	00:59
Air Blank	0.000	00:59
Control Test	0.078	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 06/27/2025 06:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNGO, WILLIAM, F
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HILLIARD, N
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: MIDYETTE

DATE/TIME OF TEST: 06/12/2025 06:41:52

Diagnostics Check	OK	06:35
Air Blank	0.000	06:35
Control Test	0.077	06:36
Air Blank	0.000	06:36
Subject Sample # 1	0.000	06:37
Breath Volume	1.921	06:37
Air Blank	0.000	06:37
Air Blank	0.000	06:39
Subject Sample # 2	0.000	06:40
Breath Volume	2.000	06:40
Air Blank	0.000	06:40
Control Test	0.078	06:41
Air Blank	0.000	06:41
Diagnostics Check	OK	06:41

Data Download Date/Time: 06/27/2025 06:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,H,G
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: THOMPSON,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: KOVACS

DATE/TIME OF TEST: 06/21/2025 00:25:31

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.077	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.294	00:19
Breath Volume	1.429	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.294	00:23
Breath Volume	1.468	00:23
Air Blank	0.000	00:24
Control Test	0.077	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 06/27/2025 06:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES,JOSEAS,I
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HERRINGTON,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 06/25/2025 02:59:07

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:55	
Control Test	0.079	02:55	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:58	
Control Test	0.079	02:58	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 06/27/2025 06:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,A,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KEENEY,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/26/2025 01:42:14

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.078	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.201	01:37
Breath Volume	1.328	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.202	01:40
Breath Volume	1.746	01:40
Air Blank	0.000	01:41
Control Test	0.077	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 06/27/2025 06:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEAL,JACOB,N

BREATH TEST OPERATOR AGENCY: Nassau County SO

ARRESTING OFFICER: PERALTA,C

ARRESTING OFFICER AGENCY: Nassau County SO

LAST AGENCY INSPECTION: 06/24/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 04924080A4

DRY GAS EXPIRY DATE: 03/05/2026

VIOLATION CODE: DUI

ARREST TIME: 16:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 06/27/2025 17:39:12

Diagnostics Check	OK	17:32	
Air Blank	0.000	17:33	
Control Test	0.077	17:33	
Air Blank	0.000	17:34	
Subject Sample	# 1 0.325	17:34	
Breath Volume	1.152	17:34	
Air Blank	0.000	17:35	
Air Blank	0.000	17:37	
Subject Sample	# 2 0.308	17:37	
Breath Volume	1.683	17:37	
Air Blank	0.000	17:38	
Control Test	0.071	17:38	Control Outside Tolerance
Air Blank	0.000	17:39	

Data Download Date/Time: 07/27/2025 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAL,JACOB,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: PERALTA,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 06/27/2025 17:50:45

Diagnostics Check	OK	17:41	
Air Blank	0.000	17:42	
Control Test	0.078	17:42	
Air Blank	0.000	17:43	
Subject Sample # 1	0.290	17:44	
Breath Volume	1.867	17:44	
Air Blank	0.000	17:45	
Air Blank	0.000	17:46	
Subject Sample # 2	0.290	17:48	
Breath Volume	1.867	17:48	
Air Blank	0.000	17:49	
Control Test	0.073	17:50	Control Outside Tolerance
Air Blank	0.000	17:50	

Data Download Date/Time: 07/27/2025 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DAVIS,B,A

BREATH TEST OPERATOR AGENCY: Nassau County SO

ARRESTING OFFICER: KEENEY,T

ARRESTING OFFICER AGENCY: Nassau County SO

LAST AGENCY INSPECTION: 06/24/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 04924080A4

DRY GAS EXPIRY DATE: 03/05/2026

VIOLATION CODE: DUI

ARREST TIME: 22:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: BURDETTE

DATE/TIME OF TEST: 06/30/2025 22:49:27

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:46	
Control Test	0.077	22:47	
Air Blank	0.000	22:47	
Subject Sample	# 1	22:48	Subject Test Refused
Breath Volume	0.000	22:48	
Air Blank	0.000	22:48	
Control Test	0.078	22:48	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 07/27/2025 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GASSEN,STEVE,
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: GASSEN,STEVE
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: REZAEVA

DATE/TIME OF TEST: 06/11/2025 15:13:32

Diagnostics Check	OK	15:05
Air Blank	0.000	15:05
Control Test	0.078	15:06
Air Blank	0.000	15:06
Subject Sample # 1	0.200	15:08
Breath Volume	1.339	15:08
Air Blank	0.000	15:09
Air Blank	0.000	15:11
Subject Sample # 2	0.206	15:12
Breath Volume	1.960	15:12
Air Blank	0.000	15:12
Control Test	0.078	15:12
Air Blank	0.000	15:13
Diagnostics Check	OK	15:13

Data Download Date/Time: 06/11/2025 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN,J,F

BREATH TEST OPERATOR AGENCY: Tarpon Springs PD

ARRESTING OFFICER: BROWN,J

ARRESTING OFFICER AGENCY: Tarpon Springs PD

LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31124080A1

DRY GAS EXPIRY DATE: 12/05/2026

VIOLATION CODE: DUI

ARREST TIME: 23:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: ZAPOTE

DATE/TIME OF TEST: 06/15/2025 00:12:43

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.222	00:08
Breath Volume	2.183	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.208	00:11
Breath Volume	1.300	00:11
Air Blank	0.000	00:11
Control Test	0.079	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 07/09/2025 17:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN,J,F

BREATH TEST OPERATOR AGENCY: Tarpon Springs PD

ARRESTING OFFICER: BRISENO,N

ARRESTING OFFICER AGENCY: Tarpon Springs PD

LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31124080A1

DRY GAS EXPIRY DATE: 12/05/2026

VIOLATION CODE: DUI

ARREST TIME: 01:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 06/15/2025 02:09:44

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.080	02:06	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:08	
Control Test	0.080	02:09	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 07/09/2025 17:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,L,N
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: PARKER,D
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: NEYER

DATE/TIME OF TEST: 06/20/2025 21:57:12

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.081	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.117	21:52
Breath Volume	3.488	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.117	21:55
Breath Volume	3.273	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 07/09/2025 17:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,J,F
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: MARTIN-KEMP,D
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: LOVATO

DATE/TIME OF TEST: 06/24/2025 11:29:41

Diagnostics Check	OK	11:23
Air Blank	0.000	11:23
Control Test	0.081	11:24
Air Blank	0.000	11:24
Subject Sample # 1	0.000	11:25
Breath Volume	2.238	11:25
Air Blank	0.000	11:25
Air Blank	0.000	11:27
Subject Sample # 2	0.000	11:28
Breath Volume	3.328	11:28
Air Blank	0.000	11:28
Control Test	0.081	11:29
Air Blank	0.000	11:29
Diagnostics Check	OK	11:29

Data Download Date/Time: 07/09/2025 17:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNGE,JACOB,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RUNGE,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #:]
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: KEEN

DATE/TIME OF TEST: 06/08/2025 00:22:56

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:18	
Control Test	0.077	00:18	
Air Blank	0.000	00:19	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:22	
Control Test	0.077	00:22	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 06/27/2025 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,MICHELLE,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CLARK,ETHAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: RAHILLY

DATE/TIME OF TEST: 06/21/2025 06:53:07

Diagnostics Check	OK	06:50	
Air Blank	0.000	06:50	
Control Test	0.077	06:51	
Air Blank	0.000	06:51	
Subject Sample	# 1	06:51	RFI Detect
Breath Volume	0.000	06:51	
Air Blank	0.000	06:52	
Control Test	0.077	06:52	
Air Blank	0.000	06:53	
Diagnostics Check	OK	06:53	

Data Download Date/Time: 06/27/2025 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,MICHELLE,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CLARK,ETHAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: RAHILLY

DATE/TIME OF TEST: 06/21/2025 07:02:34

Diagnostics Check	OK	06:56	
Air Blank	0.000	06:57	
Control Test	0.077	06:57	
Air Blank	0.000	06:57	
Subject Sample	# 1 0.192	06:58	
Breath Volume	2.535	06:58	
Air Blank	0.000	06:59	
Air Blank	0.000	07:01	
Subject Sample	# 2 0.191	07:01	
Breath Volume	2.058	07:01	
Air Blank		07:02	RFI Detect
Air Blank		07:02	RFI Detect

Data Download Date/Time: 06/27/2025 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,MICHELLE,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MORENO,MAKENNA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 05:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: JIGME

DATE/TIME OF TEST: 06/21/2025 07:41:19

Diagnostics Check	OK	07:35	
Air Blank	0.000	07:35	
Control Test	0.076	07:36	
Air Blank	0.000	07:36	
Subject Sample	# 1 0.128	07:36	
Breath Volume	3.125	07:36	
Air Blank	0.000	07:37	
Air Blank	0.000	07:39	
Subject Sample	# 2 0.114	07:39	
Breath Volume	2.277	07:39	
Air Blank	0.000	07:40	
Control Test	0.073	07:40	Control Outside Tolerance
Air Blank	0.000	07:41	

Data Download Date/Time: 06/27/2025 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEEBECK,BRANDON,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SEEBECK,BRANDON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: SAGE

DATE/TIME OF TEST: 06/21/2025 17:42:45

Diagnostics Check	OK	17:36
Air Blank	0.000	17:36
Control Test	0.076	17:37
Air Blank	0.000	17:37
Subject Sample # 1	0.116	17:38
Breath Volume	4.265	17:38
Air Blank	0.000	17:38
Air Blank	0.000	17:40
Subject Sample # 2	0.113	17:41
Breath Volume	5.105	17:41
Air Blank	0.000	17:41
Control Test	0.075	17:42
Air Blank	0.000	17:42
Diagnostics Check	OK	17:42

Data Download Date/Time: 06/27/2025 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALIEL,AMANDA,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PAPPAS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 06/22/2025 03:57:35

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.077	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.210	03:51
Breath Volume	1.585	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.201	03:55
Breath Volume	1.410	03:55
Air Blank	0.000	03:56
Control Test	0.077	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 06/27/2025 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHEINHARDT,JERALD,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RHEINHARDT,JERALD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: LAIN

DATE/TIME OF TEST: 06/22/2025 05:16:38

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:14	
Control Test	0.077	05:14	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:15	Subject Test Refused
Breath Volume	0.000	05:15	
Air Blank	0.000	05:15	
Control Test	0.077	05:16	
Air Blank	0.000	05:16	
Diagnostics Check	OK	05:16	

Data Download Date/Time: 06/27/2025 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALIEL,AMANDA,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAUN,JOEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: FURNESS

DATE/TIME OF TEST: 06/27/2025 05:00:21

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:57	
Control Test	0.076	04:58	
Air Blank	0.000	04:58	
Subject Sample	# 1	04:58	Subject Test Refused
Breath Volume	0.000	04:58	
Air Blank	0.000	04:59	
Control Test	0.077	04:59	
Air Blank	0.000	05:00	
Diagnostics Check	OK	05:00	

Data Download Date/Time: 07/31/2025 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JERRY,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MARTINEZ,CARLOS
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: VELASQUEZ PEREZ

DATE/TIME OF TEST: 06/27/2025 23:43:10

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:38	
Control Test	0.077	23:38	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:41	Subject Test Refused
Breath Volume	0.000	23:41	
Air Blank	0.000	23:42	
Control Test	0.078	23:42	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 07/31/2025 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JERRY,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: O'ROURKE,ROCKY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: BRAZZELL

DATE/TIME OF TEST: 06/27/2025 23:56:27

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:54	
Control Test	0.076	23:54	
Air Blank	0.000	23:55	
Subject Sample	# 1	23:55	Subject Test Refused
Breath Volume	0.000	23:55	
Air Blank	0.000	23:55	
		23:55	RFI Detect
Air Blank	0.000	23:56	

Data Download Date/Time: 07/31/2025 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,RITCHIE,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: RIVERA,RITCHIE
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: GAINER

DATE/TIME OF TEST: 06/18/2025 02:34:03

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.078	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:30	No Sample Provided
Breath Volume	0.000	02:30	
Air Blank	0.000	02:30	
Air Blank	0.000	02:32	
Subject Sample	# 2	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Control Test	0.079	02:33	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 07/04/2025 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 240383
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: LYTLE

DATE/TIME OF TEST: 06/09/2025 05:22:17

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.079	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.153	05:17
Breath Volume	2.531	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.154	05:20
Breath Volume	2.726	05:20
Air Blank	0.000	05:21
Control Test	0.079	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 06/28/2025 20:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAMER,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: GONZALEZ,JOSE
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: CORTES PEREZ

DATE/TIME OF TEST: 06/01/2025 00:32:17

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.215	00:27
Breath Volume	1.503	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.218	00:30
Breath Volume	2.687	00:30
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 06/02/2025 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LAWSON,SKYLAR
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/06/2025 02:50:46

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.081	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.154	02:46
Breath Volume	1.570	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.155	02:49
Breath Volume	1.355	02:49
Air Blank	0.000	02:49
Control Test	0.081	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 07/08/2025 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS, RONALD, J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: EVANS, RONALD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 06/07/2025 02:36:39

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.125	02:32
Breath Volume	4.089	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.112	02:35
Breath Volume	2.058	02:35
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 07/08/2025 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,DILLON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: COPPOLA,COREY
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: BOLES

DATE/TIME OF TEST: 06/07/2025 07:25:18

Diagnostics Check	OK	07:18
Air Blank	0.000	07:19
Control Test	0.080	07:19
Air Blank	0.000	07:19
Subject Sample # 1	0.272	07:20
Breath Volume	1.375	07:20
Air Blank	0.000	07:21
Air Blank	0.000	07:23
Subject Sample # 2	0.277	07:23
Breath Volume	2.570	07:23
Air Blank	0.000	07:24
Control Test	0.079	07:24
Air Blank	0.000	07:25
Diagnostics Check	OK	07:25

Data Download Date/Time: 07/08/2025 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS, RONALD, J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: EVANS, RONALD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: THEODORE

DATE/TIME OF TEST: 06/08/2025 03:54:04

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.189	03:48
Breath Volume	1.648	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.188	03:52
Breath Volume	2.914	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:54

Data Download Date/Time: 07/08/2025 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAMER,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: SOSA,KYLE
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: BOFINGER

DATE/TIME OF TEST: 06/11/2025 00:32:38

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:30	
Control Test	0.081	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 07/08/2025 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS, RONALD, J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: EVANS, RONALD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/11/2025 03:33:24

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.081	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.284	03:28
Breath Volume	4.566	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.297	03:31
Breath Volume	3.984	03:31
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 07/08/2025 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAMER,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: STAFFORD,ROBERT
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: BARBOSA

DATE/TIME OF TEST: 06/11/2025 05:24:40

Diagnostics Check	OK	05:17
Air Blank	0.000	05:17
Control Test	0.081	05:18
Air Blank	0.000	05:18
Subject Sample # 1	0.340	05:19
Breath Volume	1.109	05:19
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.332	05:22
Breath Volume	1.210	05:22
Air Blank	0.000	05:23
Control Test	0.079	05:23
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

Data Download Date/Time: 07/08/2025 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MAHARAJ,KRISHNA
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: GADSON

DATE/TIME OF TEST: 06/14/2025 01:34:30

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.015	01:29
Breath Volume	1.437	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.012	01:33
Breath Volume	1.250	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 07/08/2025 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: WILSON,DREW
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: VILLEDA GARCIA

DATE/TIME OF TEST: 06/14/2025 19:20:05

Diagnostics Check	OK	19:13
Air Blank	0.000	19:13
Control Test	0.081	19:13
Air Blank	0.000	19:14
Subject Sample # 1	0.150	19:15
Breath Volume	1.472	19:15
Air Blank	0.000	19:15
Air Blank	0.000	19:17
Subject Sample # 2	0.168	19:18
Breath Volume	1.304	19:18
Air Blank	0.000	19:19
Control Test	0.081	19:19
Air Blank	0.000	19:20
Diagnostics Check	OK	19:20

Data Download Date/Time: 07/08/2025 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REUL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LAWSON,SKYLER
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:33

SUBJECT'S LAST NAME: WHITEHURST

DATE/TIME OF TEST: 06/15/2025 06:56:02

Diagnostics Check	OK	06:53	
Air Blank	0.000	06:53	
Control Test	0.080	06:53	
Air Blank	0.000	06:54	
Subject Sample	# 1	06:54	Subject Test Refused
Breath Volume	0.000	06:54	
Air Blank	0.000	06:55	
Control Test	0.080	06:55	
Air Blank	0.000	06:55	
Diagnostics Check	OK	06:56	

Data Download Date/Time: 07/08/2025 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: JAMES,ANTONIO
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: BECHTOLD

DATE/TIME OF TEST: 06/15/2025 21:05:36

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.081	20:59
Air Blank	0.000	20:59
Subject Sample # 1	0.137	21:00
Breath Volume	1.417	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.137	21:03
Breath Volume	1.128	21:03
Air Blank	0.000	21:04
Control Test	0.080	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 07/08/2025 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,JOSHUA,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILSON,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: CRUZ CRUZ

DATE/TIME OF TEST: 06/19/2025 22:52:46

Diagnostics Check	OK	22:44	
Air Blank	0.000	22:44	
Control Test	0.081	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 1	22:48	No Sample Provided
Breath Volume	0.000	22:48	
Air Blank	0.000	22:48	
Air Blank	0.000	22:50	
Subject Sample	# 2	22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:51	
Control Test	0.081	22:52	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 07/08/2025 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,KALEB,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: SAPUNDJIEFF,THOMAS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 06/21/2025 01:52:31

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:49	
Control Test	0.081	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 07/08/2025 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SOWARDS,JACOB
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 06/21/2025 02:21:11

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:18	
Control Test	0.081	02:19	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:20	
Control Test	0.081	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 07/08/2025 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: JIN,KEVIN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: ENOCH

DATE/TIME OF TEST: 06/21/2025 02:39:49

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:37	
Control Test	0.082	02:37	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:38	Subject Test Refused
Breath Volume	0.000	02:38	
Air Blank	0.000	02:38	
Control Test	0.081	02:39	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 07/08/2025 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,KALEB,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MORRIS,ERIC
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: TERCERO RAFAEL

DATE/TIME OF TEST: 06/21/2025 04:08:04

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.082	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.135	04:03
Breath Volume	3.074	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.136	04:06
Breath Volume	2.367	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:08

Data Download Date/Time: 07/08/2025 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,SKYLER,O
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: WOODARD,JACOB
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: JOSHI

DATE/TIME OF TEST: 06/25/2025 03:05:16

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.081	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.177	03:00
Breath Volume	2.003	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.173	03:03
Breath Volume	1.156	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 07/08/2025 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/26/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 06/08/2025 21:03:33

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.081	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.090	20:58
Breath Volume	1.914	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.098	21:01
Breath Volume	1.628	21:01
Air Blank	0.000	21:02
Control Test	0.080	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 06/23/2025 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUKE,JUSTIN',B
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: DUKE,JUSTIN
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 05/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: SCRIBNER

DATE/TIME OF TEST: 06/13/2025 05:07:09

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:05	
Control Test	0.080	05:05	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:06	Subject Test Refused
Breath Volume	0.000	05:06	
Air Blank	0.000	05:06	
Control Test	0.038	05:07	RFI Detect
Air Blank		05:07	RFI Detect

Data Download Date/Time: 06/23/2025 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY, MATHEW, C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: NAUTA, KYLE
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 06/13/2025 21:24:42

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.080	21:18
Air Blank	0.000	21:18
Subject Sample # 1	0.206	21:19
Breath Volume	1.769	21:19
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.212	21:22
Breath Volume	1.722	21:22
Air Blank	0.000	21:23
Control Test	0.079	21:24
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 06/23/2025 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,MIA,Y
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: GILBERT,ERIC
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:49

SUBJECT'S LAST NAME: MARCHI

DATE/TIME OF TEST: 06/21/2025 11:11:02

Diagnostics Check	OK	11:09	
Air Blank	0.000	11:10	
Control Test	0.080	11:10	
Air Blank		11:10	Improper Sample
Air Blank		11:11	Improper Sample

Data Download Date/Time: 06/23/2025 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,MIA,Y
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: GILBERT,ERIC
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:49

SUBJECT'S LAST NAME: MARCHI

DATE/TIME OF TEST: 06/21/2025 11:14:49

Diagnostics Check	OK	11:14	
Air Blank		11:14	Improper Sample
Air Blank		11:14	Improper Sample

Data Download Date/Time: 06/23/2025 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,MIA,Y
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: GILBERT,ERIC
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:49

SUBJECT'S LAST NAME: MARCHI

DATE/TIME OF TEST: 06/21/2025 11:19:45

Diagnostics Check	OK	11:19	
Air Blank		11:19	Improper Sample
Air Blank		11:19	Improper Sample

Data Download Date/Time: 06/23/2025 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,MIA,Y
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: GILBERT,ERIC
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:49

SUBJECT'S LAST NAME: MARCHI

DATE/TIME OF TEST: 06/21/2025 11:22:46

Diagnostics Check	OK	11:22	
Air Blank		11:22	Improper Sample
Air Blank	0.000	11:22	

Data Download Date/Time: 06/23/2025 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JABLONSKI, MARK, A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JABLONSKI, MARK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 06/01/2025 05:40:10

Diagnostics Check	OK	05:33
Air Blank	0.000	05:33
Control Test	0.082	05:34
Air Blank	0.000	05:34
Subject Sample # 1	0.000	05:35
Breath Volume	1.449	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.000	05:38
Breath Volume	1.246	05:38
Air Blank	0.000	05:39
Control Test	0.083	05:39
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 06/06/2025 05:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODFREY,BRANDON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GODFREY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 06/01/2025 09:16:28

Diagnostics Check	OK	09:11	
Air Blank	0.000	09:11	
Control Test	0.082	09:12	
Air Blank	0.000	09:12	
Subject Sample	# 1	09:15	Subject Test Refused
Breath Volume	0.000	09:15	
Air Blank	0.000	09:15	
Control Test	0.082	09:15	
Air Blank	0.000	09:16	
Diagnostics Check	OK	09:16	

Data Download Date/Time: 06/06/2025 05:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYCHUNAS,REIDE,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WYCHUNAS,REIDE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/01/2025 23:47:04

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:44	
Control Test	0.083	23:45	
Air Blank	0.000	23:45	
Subject Sample	# 1	23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank	0.000	23:46	
Control Test	0.082	23:46	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:47	

Data Download Date/Time: 06/06/2025 05:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,STANLEY,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SLICK,LOREN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/04/2025 21:17:22

Diagnostics Check	OK	21:11
Air Blank	0.000	21:11
Control Test	0.082	21:11
Air Blank	0.000	21:12
Subject Sample # 1	0.159	21:12
Breath Volume	1.964	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.153	21:15
Breath Volume	1.710	21:15
Air Blank	0.000	21:16
Control Test	0.081	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 06/06/2025 05:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYCHUNAS,REIDE,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ODDO,ISAIH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 06/05/2025 04:24:43

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:22	
Control Test	0.082	04:22	
Air Blank	0.000	04:23	
Subject Sample	# 1	04:23	Subject Test Refused
Breath Volume	0.000	04:23	
Air Blank	0.000	04:23	
Control Test	0.082	04:24	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 06/06/2025 05:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,BOBBY,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HARRISON,BOBBY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DELOACH

DATE/TIME OF TEST: 06/06/2025 02:13:45

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.082	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.165	02:08
Breath Volume	1.410	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample # 2	0.156	02:12
Breath Volume	1.101	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 06/06/2025 05:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

INSTRUMENT REGISTERED TO: LAFAYETTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/01/2025 05:32:28

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:30	
Control Test	0.080	05:30	
Air Blank	0.000	05:30	
Subject Sample	# 1	05:31	Subject Test Refused
Breath Volume	0.000	05:31	
Air Blank	0.000	05:31	
Control Test	0.080	05:31	
Air Blank	0.000	05:32	
Diagnostics Check	OK	05:32	

Data Download Date/Time: 06/26/2025 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACLAREN,DEREK,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACLAREN,DEREK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: BAIL

DATE/TIME OF TEST: 06/20/2025 19:03:04

Diagnostics Check	OK	18:55
Air Blank	0.000	18:55
Control Test	0.078	18:55
Air Blank	0.000	18:56
Subject Sample # 1	0.144	18:58
Breath Volume	2.042	18:58
Air Blank	0.000	18:59
Air Blank	0.000	19:00
Subject Sample # 2	0.141	19:01
Breath Volume	1.679	19:01
Air Blank	0.000	19:02
Control Test	0.076	19:02
Air Blank	0.000	19:02
Diagnostics Check	OK	19:03

Data Download Date/Time: 07/01/2025 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,B,L
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: RECCA,MICHAEL
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 06/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: JELKS

DATE/TIME OF TEST: 06/28/2025 02:10:51

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.081	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.000	02:06
Breath Volume	3.492	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.000	02:09
Breath Volume	4.347	02:09
Air Blank	0.000	02:09
Control Test	0.082	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 07/27/2025 08:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 06/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: FILSELL

DATE/TIME OF TEST: 06/28/2025 05:02:08

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.081	04:56
Air Blank	0.000	04:57
Subject Sample # 1	0.260	04:57
Breath Volume	2.562	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.250	05:00
Breath Volume	2.640	05:00
Air Blank	0.000	05:01
Control Test	0.082	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 07/27/2025 08:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 06/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: HELM

DATE/TIME OF TEST: 06/29/2025 04:06:05

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.082	04:04	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Control Test	0.082	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 07/27/2025 08:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MALONEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 06/07/2025 19:22:11

Diagnostics Check	OK	19:18	
Air Blank	0.000	19:19	
Control Test	0.079	19:19	
Air Blank	0.000	19:20	
Subject Sample	# 1	19:21	Sequence Aborted
Breath Volume	0.000	19:21	
Air Blank	0.000	19:22	

Data Download Date/Time: 07/01/2025 15:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MALONEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 06/07/2025 19:29:22

Diagnostics Check	OK	19:25	
Air Blank	0.000	19:26	
Control Test	0.080	19:26	
Air Blank	0.000	19:27	
Subject Sample	# 1	19:28	Subject Test Refused
Breath Volume	0.000	19:28	
Air Blank	0.000	19:28	
Control Test	0.080	19:28	
Air Blank	0.000	19:29	
Diagnostics Check	OK	19:29	

Data Download Date/Time: 07/01/2025 15:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,MATTHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,SAWYER
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/11/2025 01:32:17

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.081	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1 0.150	01:25	Volume Not Met
Breath Volume	0.910	01:25	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 2 0.136	01:30	Volume Not Met
Breath Volume	0.636	01:30	
Air Blank	0.000	01:31	
Control Test	0.081	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 07/01/2025 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 06/01/2025 15:59:48

Diagnostics Check	OK	15:52
Air Blank	0.000	15:53
Control Test	0.081	15:53
Air Blank	0.000	15:54
Subject Sample # 1	0.000	15:55
Breath Volume	1.320	15:55
Air Blank	0.000	15:55
Air Blank	0.000	15:57
Subject Sample # 2	0.000	15:58
Breath Volume	1.304	15:58
Air Blank	0.000	15:58
Control Test	0.080	15:59
Air Blank	0.000	15:59
Diagnostics Check	OK	15:59

Data Download Date/Time: 08/05/2025 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 06/28/2025 02:27:08

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:24	
Control Test	0.079	02:25	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:25	Subject Test Refused
Breath Volume	0.000	02:25	
Air Blank	0.000	02:26	
Control Test	0.078	02:26	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

Data Download Date/Time: 08/05/2025 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: CAREY

DATE/TIME OF TEST: 06/29/2025 01:23:15

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:15	
Control Test	0.079	01:15	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:19	No Sample Provided
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Air Blank	0.000	01:21	
Subject Sample	# 2	01:21	Subject Test Refused
Breath Volume	0.000	01:21	
Air Blank	0.000	01:22	
Control Test	0.080	01:22	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 08/05/2025 08:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,JOHN,H
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: HUNTER,JOHN
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MEEH

DATE/TIME OF TEST: 06/01/2025 01:17:18

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.129	01:12
Breath Volume	2.628	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.129	01:15
Breath Volume	1.429	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 07/01/2025 02:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTER,DYLAN,F
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BONNIN,BRYCE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: BROMLEY

DATE/TIME OF TEST: 06/01/2025 03:19:23

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.081	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.093	03:14
Breath Volume	2.855	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.100	03:17
Breath Volume	2.898	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 07/01/2025 02:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,JOHN,H
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: HUNTER,JOHN
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: AUDIFFRED

DATE/TIME OF TEST: 06/01/2025 20:34:58

Diagnostics Check	OK	20:31	
Air Blank	0.000	20:32	
Control Test	0.081	20:32	
Air Blank	0.000	20:32	
Subject Sample	# 1	20:34	Sequence Aborted
Breath Volume	0.000	20:34	
Air Blank	0.000	20:34	

Data Download Date/Time: 07/01/2025 02:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUALE,BRANDON,P
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: DUALE,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: ZAGER

DATE/TIME OF TEST: 06/07/2025 00:22:33

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.080	00:21	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:22	Sequence Aborted
Breath Volume	0.000	00:22	
Air Blank	0.000	00:22	

Data Download Date/Time: 07/01/2025 02:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUALE,BRANDON,P
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: DUALE,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: ZAGER

DATE/TIME OF TEST: 06/07/2025 00:57:19

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.136	00:52
Breath Volume	1.738	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.153	00:55
Breath Volume	5.074	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 07/01/2025 02:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,DYLAN,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SELLERS,DONALD
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CASH

DATE/TIME OF TEST: 06/07/2025 03:14:23

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.079	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:09	No Sample Provided
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Air Blank	0.000	03:12	
Subject Sample	# 2	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank		03:14	RFI Detect
Air Blank		03:14	RFI Detect

Data Download Date/Time: 07/01/2025 02:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTER,DYLAN,F
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SABATINE,MATTHEW
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 06/14/2025 19:25:29

Diagnostics Check	OK	19:18
Air Blank	0.000	19:19
Control Test	0.081	19:19
Air Blank	0.000	19:20
Subject Sample # 1	0.176	19:21
Breath Volume	3.023	19:21
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.171	19:23
Breath Volume	2.070	19:23
Air Blank	0.000	19:24
Control Test	0.079	19:24
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 07/01/2025 02:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,RICHARD,D
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BALDRIDGE,GAUGE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: KOON

DATE/TIME OF TEST: 06/18/2025 00:12:40

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:08	
Control Test	0.079	00:08	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:11	
Control Test	0.080	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 07/01/2025 02:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,DONALD,R
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SABATINE,MATTHEW
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: MIHAI

DATE/TIME OF TEST: 06/18/2025 16:17:55

Diagnostics Check	OK	16:11
Air Blank	0.000	16:12
Control Test	0.080	16:12
Air Blank	0.000	16:13
Subject Sample # 1	0.085	16:13
Breath Volume	1.562	16:13
Air Blank	0.000	16:14
Air Blank	0.000	16:15
Subject Sample # 2	0.079	16:16
Breath Volume	1.582	16:16
Air Blank	0.000	16:17
Control Test	0.080	16:17
Air Blank	0.000	16:17
Diagnostics Check	OK	16:17

Data Download Date/Time: 07/01/2025 02:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,DYLAN,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BALDRIDGE,GAUGE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/22/2025 21:10:52

Diagnostics Check	OK	21:04
Air Blank	0.000	21:04
Control Test	0.080	21:05
Air Blank	0.000	21:05
Subject Sample # 1	0.098	21:06
Breath Volume	2.246	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:08
Subject Sample # 2	0.098	21:09
Breath Volume	2.601	21:09
Air Blank	0.000	21:09
Control Test	0.080	21:10
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 07/01/2025 02:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUALE,BRANDON,P
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BALDRIDGE,GAUGE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: CARRINGTON

DATE/TIME OF TEST: 06/27/2025 00:12:51

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.226	00:08
Breath Volume	1.937	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:10
Subject Sample # 2	0.211	00:11
Breath Volume	1.429	00:11
Air Blank	0.000	00:11
Control Test	0.077	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 07/01/2025 02:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 05/26/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 17:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: DRAYTON

DATE/TIME OF TEST: 06/04/2025 18:57:10

Diagnostics Check	OK	18:50
Air Blank	0.000	18:51
Control Test	0.080	18:51
Air Blank	0.000	18:52
Subject Sample # 1	0.145	18:52
Breath Volume	2.613	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.145	18:55
Breath Volume	1.894	18:55
Air Blank	0.000	18:56
Control Test	0.079	18:56
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 06/16/2025 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: ALOMA,ANDRES

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 05/26/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 21:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: ESPINO

DATE/TIME OF TEST: 06/04/2025 22:16:15

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.080	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.066	22:11
Breath Volume	3.796	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.067	22:14
Breath Volume	3.769	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 06/16/2025 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GARDNER, ANTHONY, R

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: CHESNEY, JOSHUA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 06/16/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 01:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 06/23/2025 02:11:47

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.080	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.195	02:06
Breath Volume	2.835	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.202	02:10
Breath Volume	2.726	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 07/23/2025 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SINGSON,ROLAND

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 06/16/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 20:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: GREATHOUSE

DATE/TIME OF TEST: 06/28/2025 21:45:28

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:35	
Control Test	0.080	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1 0.199	21:39	Volume Not Met
Breath Volume	0.109	21:39	
Air Blank	0.000	21:40	
Air Blank	0.000	21:42	
Subject Sample	# 2	21:43	Subject Test Refused
Breath Volume	0.128	21:43	
Air Blank	0.000	21:44	
Control Test	0.078	21:44	
Air Blank	0.000	21:45	
Diagnostics Check	OK	21:45	

Data Download Date/Time: 07/23/2025 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,SHANE,C
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: MARSHALL,JARED
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 06/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: CUEVAS URQUIA

DATE/TIME OF TEST: 06/29/2025 08:08:12

Diagnostics Check	OK	07:56	
Air Blank	0.000	07:57	
Control Test	0.080	07:57	
Air Blank	0.000	07:57	
Subject Sample	# 1 0.115	08:01	Volume Not Met
Breath Volume	0.054	08:01	
Air Blank	0.000	08:01	
Air Blank	0.000	08:03	
Subject Sample	# 2 0.152	08:06	Volume Not Met
Breath Volume	0.000	08:06	
Air Blank	0.000	08:07	
Control Test	0.079	08:07	
Air Blank	0.000	08:08	
Diagnostics Check	OK	08:08	

Data Download Date/Time: 07/23/2025 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIGG,JACOB,M
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: PIGG,JACOB
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/02/2025 03:12:19

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.177	03:07
Breath Volume	1.195	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.174	03:10
Breath Volume	1.175	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 06/28/2025 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERTIN,SKYLAR,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GUERTIN,SKYLAR
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 06/07/2025 00:58:01

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.082	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.063	00:53
Breath Volume	1.542	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.061	00:56
Breath Volume	2.207	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:58

Data Download Date/Time: 06/28/2025 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCATO,SHAWN,V
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: BROCATO,SHAWN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: VALLEJO

DATE/TIME OF TEST: 06/07/2025 03:01:20

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.082	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.254	02:56
Breath Volume	2.269	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:58
Subject Sample # 2	0.237	02:59
Breath Volume	1.546	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 06/28/2025 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEMON,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: MEMON,MATTHEW
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: MITRA ROLDAN

DATE/TIME OF TEST: 06/07/2025 22:44:24

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.081	22:39	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:43	
Control Test	0.082	22:43	
Air Blank	0.000	22:44	
Diagnostics Check	OK	22:44	

Data Download Date/Time: 06/28/2025 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: DOZIER, TERRELL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MUNOZ CARVAJAL

DATE/TIME OF TEST: 06/08/2025 01:10:30

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.081	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.176	01:05
Breath Volume	1.558	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:07
Subject Sample # 2	0.186	01:08
Breath Volume	1.609	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 06/28/2025 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIERKOWSKI,CHRISTIAN,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: GAFFNEY

DATE/TIME OF TEST: 06/12/2025 01:20:24

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.156	01:15
Breath Volume	2.554	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.159	01:18
Breath Volume	2.703	01:18
Air Blank	0.000	01:19
Control Test	0.082	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 06/28/2025 19:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIERKOWSKI,CHRISTIAN,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: FOLEY

DATE/TIME OF TEST: 06/13/2025 01:06:41

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.081	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.082	01:01
Breath Volume	2.175	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.082	01:05
Breath Volume	2.160	01:05
Air Blank	0.000	01:05
Control Test	0.081	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 06/28/2025 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMCZYK,NIKKI,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TOMCZYK,NIKKI
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: PERRUSO

DATE/TIME OF TEST: 06/13/2025 20:28:56

Diagnostics Check	OK	20:28	
Air Blank		20:28	RFI Detect
Air Blank	0.000	20:28	

Data Download Date/Time: 06/28/2025 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMCZYK, NIKKI,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TOMCZYK, NIKKI
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: PERRUSO

DATE/TIME OF TEST: 06/13/2025 20:36:55

Diagnostics Check	OK	20:30
Air Blank	0.000	20:30
Control Test	0.081	20:31
Air Blank	0.000	20:31
Subject Sample # 1	0.228	20:32
Breath Volume	1.464	20:32
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample # 2	0.220	20:35
Breath Volume	1.320	20:35
Air Blank	0.000	20:36
Control Test	0.080	20:36
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 06/28/2025 19:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: COLUMBUS,DYLAN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: GABLE

DATE/TIME OF TEST: 06/14/2025 21:36:48

Diagnostics Check	OK	21:29	
Air Blank	0.000	21:29	
Control Test	0.081	21:30	
Air Blank	0.000	21:30	
Subject Sample	# 1	21:33	No Sample Provided
Breath Volume	0.000	21:33	
Air Blank	0.000	21:34	
Air Blank	0.000	21:36	
Subject Sample	# 2	21:36	Sequence Aborted
Breath Volume	0.000	21:36	
Air Blank	0.000	21:36	

Data Download Date/Time: 06/28/2025 19:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: CRUZ ORTIZ

DATE/TIME OF TEST: 06/19/2025 00:59:25

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.081	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.212	00:54
Breath Volume	1.761	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:56
Subject Sample # 2	0.215	00:57
Breath Volume	2.429	00:57
Air Blank	0.000	00:58
Control Test	0.081	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 06/28/2025 19:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIERKOWSKI,CHRISTIAN,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BRADY,ZACHARY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 06/19/2025 02:48:22

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.014	02:43
Breath Volume	3.144	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.013	02:46
Breath Volume	3.148	02:46
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 06/28/2025 19:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMCZYK, NIKKI,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TOMCZYK, NIKKI
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/20/2025 21:16:53

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.081	21:10
Air Blank	0.000	21:10
Subject Sample # 1	0.185	21:12
Breath Volume	2.703	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:14
Subject Sample # 2	0.193	21:15
Breath Volume	4.289	21:15
Air Blank	0.000	21:15
Control Test	0.079	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 06/28/2025 19:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIERKOWSKI,CHRISTIAN,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SWIERKOWSKI,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/21/2025 03:34:41

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.081	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.186	03:30
Breath Volume	4.601	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.183	03:33
Breath Volume	4.308	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 06/28/2025 19:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: SANCHEZ JR

DATE/TIME OF TEST: 06/21/2025 04:11:15

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.154	04:06
Breath Volume	1.562	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:08
Subject Sample # 2	0.151	04:09
Breath Volume	1.785	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 06/28/2025 19:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: BULLEY,RYAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HAMNER

DATE/TIME OF TEST: 06/22/2025 02:27:27

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.081	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.000	02:23
Breath Volume	2.824	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.000	02:26
Breath Volume	2.734	02:26
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 06/28/2025 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCATO,SHAWN,V
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: BROCATO,SHAWN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: DAUGHTRY

DATE/TIME OF TEST: 06/27/2025 03:46:36

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.082	03:42	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:45	Subject Test Refused
Breath Volume	0.000	03:45	
Air Blank	0.000	03:45	
Control Test	0.082	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 06/28/2025 19:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,TERANCE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: MACKIN

DATE/TIME OF TEST: 06/28/2025 21:02:20

Diagnostics Check	OK	20:55
Air Blank	0.000	20:56
Control Test	0.081	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.094	20:57
Breath Volume	1.570	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.095	21:00
Breath Volume	1.671	21:00
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 07/18/2025 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KICKLIGHTER,KEVIN,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ANCEWICZ,VINCENT
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/28/2025 23:30:19

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:24	
Control Test	0.081	23:24	
Air Blank	0.000	23:25	
Subject Sample	# 1 0.197	23:25	
Breath Volume	3.257	23:25	
Air Blank	0.000	23:26	
Air Blank	0.000	23:28	
Subject Sample	# 2	23:28	Slope Not Met
Breath Volume	0.871	23:28	
Air Blank	0.000	23:29	
Control Test	0.081	23:29	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 07/18/2025 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KICKLIGHTER,KEVIN,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ANCEWICZ,VINCENT
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/28/2025 23:40:07

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.192	23:35
Breath Volume	3.429	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.185	23:38
Breath Volume	2.742	23:38
Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 07/18/2025 18:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KICKLIGHTER,KEVIN,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LANIER,BRANDON
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: RUIZ-SAN JUAN

DATE/TIME OF TEST: 06/29/2025 01:33:15

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.144	01:28
Breath Volume	1.757	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.145	01:31
Breath Volume	2.074	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 07/18/2025 18:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCATO,SHAWN,V
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: BROCATO,SHAWN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: GIRALDO ROLDAN

DATE/TIME OF TEST: 06/29/2025 03:29:13

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:12	
Control Test	0.081	03:13	
Air Blank	0.000	03:13	
Subject Sample	# 1 0.176	03:16	Volume Not Met
Breath Volume	0.671	03:16	
Air Blank	0.000	03:17	
Air Blank	0.000	03:19	
Subject Sample	# 2 0.186	03:21	
Breath Volume	1.121	03:21	
Air Blank	0.000	03:22	
Air Blank	0.000	03:24	
Subject Sample	# 3 0.180	03:27	Volume Not Met
Breath Volume	0.203	03:27	
Air Blank	0.000	03:28	
Control Test	0.073	03:28	Control Outside Tolerance
Air Blank	0.000	03:29	

Data Download Date/Time: 07/18/2025 18:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCATO,SHAWN,V
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: BROCATO,SHAWN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: GIRALDO ROLDAN

DATE/TIME OF TEST: 06/29/2025 03:40:32

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.202	03:35
Breath Volume	2.199	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:37
Subject Sample # 2	0.190	03:38
Breath Volume	2.046	03:38
Air Blank	0.000	03:39
Control Test	0.076	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 07/18/2025 18:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KICKLIGHTER,KEVIN,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: PANIAGUA,REY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: CANO PUERTA

DATE/TIME OF TEST: 06/29/2025 04:17:22

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:09	
Control Test	0.079	04:09	
Air Blank	0.000	04:10	
Subject Sample	# 1	04:13	No Sample Provided
Breath Volume	0.000	04:13	
Air Blank	0.000	04:13	
Air Blank	0.000	04:15	
Subject Sample	# 2	04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:16	
Control Test	0.079	04:16	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

Data Download Date/Time: 07/18/2025 18:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: EHRITZ

DATE/TIME OF TEST: 06/01/2025 00:07:02

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.207	00:02
Breath Volume	3.187	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.202	00:05
Breath Volume	3.312	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:07

Data Download Date/Time: 06/13/2025 17:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 06/04/2025 15:29:33

Diagnostics Check	OK	15:23
Air Blank	0.000	15:23
Control Test	0.081	15:24
Air Blank	0.000	15:24
Subject Sample # 1	0.000	15:25
Breath Volume	1.179	15:25
Air Blank	0.000	15:25
Air Blank	0.000	15:27
Subject Sample # 2	0.000	15:28
Breath Volume	1.996	15:28
Air Blank	0.000	15:28
Control Test	0.081	15:28
Air Blank	0.000	15:29
Diagnostics Check	OK	15:29

Data Download Date/Time: 06/13/2025 17:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 06/05/2025 08:41:22

Diagnostics Check	OK	08:35
Air Blank	0.000	08:36
Control Test	0.082	08:36
Air Blank	0.000	08:36
Subject Sample # 1	0.000	08:37
Breath Volume	2.625	08:37
Air Blank	0.000	08:37
Air Blank	0.000	08:39
Subject Sample # 2	0.000	08:40
Breath Volume	2.613	08:40
Air Blank	0.000	08:40
Control Test	0.079	08:40
Air Blank	0.000	08:41
Diagnostics Check	OK	08:41

Data Download Date/Time: 06/13/2025 17:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 06/05/2025 09:06:40

Diagnostics Check	OK	09:00
Air Blank	0.000	09:01
Control Test	0.081	09:01
Air Blank	0.000	09:02
Subject Sample # 1	0.000	09:02
Breath Volume	2.519	09:02
Air Blank	0.000	09:02
Air Blank	0.000	09:04
Subject Sample # 2	0.000	09:05
Breath Volume	2.507	09:05
Air Blank	0.000	09:05
Control Test	0.080	09:06
Air Blank	0.000	09:06
Diagnostics Check	OK	09:06

Data Download Date/Time: 06/13/2025 17:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: MANZANO MELENDEZ

DATE/TIME OF TEST: 06/05/2025 09:26:59

Diagnostics Check	OK	09:21
Air Blank	0.000	09:21
Control Test	0.081	09:21
Air Blank	0.000	09:22
Subject Sample # 1	0.000	09:22
Breath Volume	2.984	09:22
Air Blank	0.000	09:23
Air Blank	0.000	09:25
Subject Sample # 2	0.000	09:25
Breath Volume	2.843	09:25
Air Blank	0.000	09:26
Control Test	0.080	09:26
Air Blank	0.000	09:26
Diagnostics Check	OK	09:26

Data Download Date/Time: 06/13/2025 17:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 06/05/2025 09:41:23

Diagnostics Check	OK	09:34
Air Blank	0.000	09:35
Control Test	0.081	09:35
Air Blank	0.000	09:35
Subject Sample # 1	0.000	09:37
Breath Volume	2.675	09:37
Air Blank	0.000	09:37
Air Blank	0.000	09:39
Subject Sample # 2	0.000	09:40
Breath Volume	2.386	09:40
Air Blank	0.000	09:40
Control Test	0.079	09:40
Air Blank	0.000	09:41
Diagnostics Check	OK	09:41

Data Download Date/Time: 06/13/2025 17:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 06/05/2025 10:01:59

Diagnostics Check	OK	09:55
Air Blank	0.000	09:56
Control Test	0.080	09:56
Air Blank	0.000	09:57
Subject Sample # 1	0.000	09:57
Breath Volume	1.734	09:57
Air Blank	0.000	09:58
Air Blank	0.000	10:00
Subject Sample # 2	0.000	10:00
Breath Volume	1.820	10:00
Air Blank	0.000	10:01
Control Test	0.079	10:01
Air Blank	0.000	10:01
Diagnostics Check	OK	10:01

Data Download Date/Time: 06/13/2025 17:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: RIAN

DATE/TIME OF TEST: 06/05/2025 10:35:31

Diagnostics Check	OK	10:29
Air Blank	0.000	10:30
Control Test	0.081	10:30
Air Blank	0.000	10:30
Subject Sample # 1	0.000	10:31
Breath Volume	1.664	10:31
Air Blank	0.000	10:31
Air Blank	0.000	10:33
Subject Sample # 2	0.000	10:34
Breath Volume	1.140	10:34
Air Blank	0.000	10:34
Control Test	0.079	10:34
Air Blank	0.000	10:35
Diagnostics Check	OK	10:35

Data Download Date/Time: 06/13/2025 17:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: BARTEL

DATE/TIME OF TEST: 06/05/2025 10:52:04

Diagnostics Check	OK	10:46
Air Blank	0.000	10:46
Control Test	0.080	10:46
Air Blank	0.000	10:47
Subject Sample # 1	0.000	10:47
Breath Volume	2.062	10:47
Air Blank	0.000	10:48
Air Blank	0.000	10:50
Subject Sample # 2	0.000	10:50
Breath Volume	2.070	10:50
Air Blank	0.000	10:51
Control Test	0.080	10:51
Air Blank	0.000	10:51
Diagnostics Check	OK	10:52

Data Download Date/Time: 06/13/2025 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: LEITAO

DATE/TIME OF TEST: 06/05/2025 13:24:27

Diagnostics Check	OK	13:18
Air Blank	0.000	13:18
Control Test	0.077	13:19
Air Blank	0.000	13:19
Subject Sample # 1	0.000	13:20
Breath Volume	1.117	13:20
Air Blank	0.000	13:20
Air Blank	0.000	13:22
Subject Sample # 2	0.000	13:23
Breath Volume	1.167	13:23
Air Blank	0.000	13:23
Control Test	0.079	13:23
Air Blank	0.000	13:24
Diagnostics Check	OK	13:24

Data Download Date/Time: 06/13/2025 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:03

SUBJECT'S LAST NAME: TOSADO MERCADO

DATE/TIME OF TEST: 06/05/2025 13:41:28

Diagnostics Check	OK	13:35
Air Blank	0.000	13:35
Control Test	0.080	13:36
Air Blank	0.000	13:36
Subject Sample # 1	0.000	13:37
Breath Volume	1.664	13:37
Air Blank	0.000	13:37
Air Blank	0.000	13:39
Subject Sample # 2	0.000	13:40
Breath Volume	1.566	13:40
Air Blank	0.000	13:40
Control Test	0.079	13:40
Air Blank	0.000	13:41
Diagnostics Check	OK	13:41

Data Download Date/Time: 06/13/2025 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:09

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 06/05/2025 13:58:40

Diagnostics Check	OK	13:51
Air Blank	0.000	13:52
Control Test	0.081	13:52
Air Blank	0.000	13:53
Subject Sample # 1	0.000	13:54
Breath Volume	2.328	13:54
Air Blank	0.000	13:54
Air Blank	0.000	13:56
Subject Sample # 2	0.000	13:57
Breath Volume	1.117	13:57
Air Blank	0.000	13:57
Control Test	0.079	13:58
Air Blank	0.000	13:58
Diagnostics Check	OK	13:58

Data Download Date/Time: 06/13/2025 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:31

SUBJECT'S LAST NAME: COIL

DATE/TIME OF TEST: 06/05/2025 15:11:41

Diagnostics Check	OK	15:05
Air Blank	0.000	15:05
Control Test	0.082	15:06
Air Blank	0.000	15:06
Subject Sample # 1	0.000	15:07
Breath Volume	2.187	15:07
Air Blank	0.000	15:07
Air Blank	0.000	15:09
Subject Sample # 2	0.000	15:10
Breath Volume	1.714	15:10
Air Blank	0.000	15:10
Control Test	0.081	15:11
Air Blank	0.000	15:11
Diagnostics Check	OK	15:11

Data Download Date/Time: 06/13/2025 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/05/2025 15:27:45

Diagnostics Check	OK	15:21
Air Blank	0.000	15:21
Control Test	0.081	15:21
Air Blank	0.000	15:22
Subject Sample # 1	0.000	15:23
Breath Volume	1.519	15:23
Air Blank	0.000	15:23
Air Blank	0.000	15:25
Subject Sample # 2	0.000	15:26
Breath Volume	1.210	15:26
Air Blank	0.000	15:26
Control Test	0.079	15:27
Air Blank	0.000	15:27
Diagnostics Check	OK	15:27

Data Download Date/Time: 06/13/2025 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 15:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: CAMPOS MEDINA

DATE/TIME OF TEST: 06/05/2025 15:46:18

Diagnostics Check	OK	15:40
Air Blank	0.000	15:40
Control Test	0.081	15:41
Air Blank	0.000	15:41
Subject Sample # 1	0.000	15:42
Breath Volume	1.355	15:42
Air Blank	0.000	15:42
Air Blank	0.000	15:44
Subject Sample # 2	0.000	15:44
Breath Volume	1.312	15:44
Air Blank	0.000	15:45
Control Test	0.081	15:45
Air Blank	0.000	15:46
Diagnostics Check	OK	15:46

Data Download Date/Time: 06/13/2025 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: HUGHEY

DATE/TIME OF TEST: 06/06/2025 13:59:25

Diagnostics Check	OK	13:51
Air Blank	0.000	13:52
Control Test	0.082	13:52
Air Blank	0.000	13:53
Subject Sample # 1	0.000	13:54
Breath Volume	1.160	13:54
Air Blank	0.000	13:54
Air Blank	0.000	13:56
Subject Sample # 2	0.000	13:58
Breath Volume	1.187	13:58
Air Blank	0.000	13:58
Control Test	0.080	13:58
Air Blank	0.000	13:59
Diagnostics Check	OK	13:59

Data Download Date/Time: 06/13/2025 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: NAZARIO OLIVERAS

DATE/TIME OF TEST: 06/06/2025 14:12:47

Diagnostics Check	OK	14:06
Air Blank	0.000	14:07
Control Test	0.080	14:07
Air Blank	0.000	14:08
Subject Sample # 1	0.000	14:08
Breath Volume	1.378	14:08
Air Blank	0.000	14:09
Air Blank	0.000	14:11
Subject Sample # 2	0.000	14:11
Breath Volume	1.175	14:11
Air Blank	0.000	14:11
Control Test	0.081	14:12
Air Blank	0.000	14:12
Diagnostics Check	OK	14:12

Data Download Date/Time: 06/13/2025 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 06/06/2025 14:32:48

Diagnostics Check	OK	14:25
Air Blank	0.000	14:25
Control Test	0.081	14:26
Air Blank	0.000	14:26
Subject Sample # 1	0.000	14:28
Breath Volume	1.136	14:28
Air Blank	0.000	14:28
Air Blank	0.000	14:30
Subject Sample # 2	0.000	14:31
Breath Volume	1.183	14:31
Air Blank	0.000	14:31
Control Test	0.079	14:32
Air Blank	0.000	14:32
Diagnostics Check	OK	14:32

Data Download Date/Time: 06/13/2025 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 06/07/2025 00:06:38

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:57	
Control Test	0.082	23:57	
Air Blank	0.000	23:57	
Subject Sample	# 1 0.133	00:00	Volume Not Met
Breath Volume	0.031	00:00	
Air Blank	0.000	00:01	
Air Blank	0.000	00:03	
Subject Sample	# 2	00:05	Subject Test Refused
Breath Volume	0.000	00:05	
Air Blank	0.000	00:05	
Control Test	0.080	00:06	
Air Blank	0.000	00:06	
Diagnostics Check	OK	00:06	

Data Download Date/Time: 06/13/2025 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/08/2025 19:15:36

Diagnostics Check	OK	19:10	
Air Blank	0.000	19:11	
Control Test	0.081	19:11	
Air Blank	0.000	19:12	
Subject Sample	# 1	19:14	Subject Test Refused
Breath Volume	0.000	19:14	
Air Blank	0.000	19:14	
Control Test	0.081	19:15	
Air Blank	0.000	19:15	
Diagnostics Check	OK	19:15	

Data Download Date/Time: 06/13/2025 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: DOMINGES,ESMERALDA
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: CHUDGAR

DATE/TIME OF TEST: 06/09/2025 01:28:58

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.081	01:20
Air Blank	0.000	01:20
Subject Sample # 1	0.196	01:22
Breath Volume	1.394	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.176	01:27
Breath Volume	1.285	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 06/13/2025 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/02/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:32

SUBJECT'S LAST NAME: ARBELAEZ

DATE/TIME OF TEST: 06/09/2025 13:02:23

Diagnostics Check	OK	12:56
Air Blank	0.000	12:56
Control Test	0.082	12:57
Air Blank	0.000	12:57
Subject Sample # 1	0.000	12:57
Breath Volume	1.546	12:57
Air Blank	0.000	12:58
Air Blank	0.000	13:00
Subject Sample # 2	0.000	13:01
Breath Volume	1.988	13:01
Air Blank	0.000	13:01
Control Test	0.080	13:01
Air Blank	0.000	13:02
Diagnostics Check	OK	13:02

Data Download Date/Time: 06/13/2025 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 06/17/2025 04:43:37

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.082	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.000	04:39
Breath Volume	2.496	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.000	04:42
Breath Volume	2.101	04:42
Air Blank	0.000	04:42
Control Test	0.082	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 07/07/2025 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: BUELL

DATE/TIME OF TEST: 06/17/2025 21:08:42

Diagnostics Check	OK	20:57	
Air Blank	0.000	20:57	
Control Test	0.081	20:58	
Air Blank	0.000	20:58	
Subject Sample	# 1	21:01	No Sample Provided
Breath Volume	0.000	21:01	
Air Blank	0.000	21:02	
Air Blank	0.000	21:04	
Subject Sample	# 2	21:07	No Sample Provided
Breath Volume	0.000	21:07	
Air Blank	0.000	21:07	
Control Test	0.082	21:08	
Air Blank	0.000	21:08	
Diagnostics Check	OK	21:08	

Data Download Date/Time: 07/07/2025 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/20/2025 08:40:22

Diagnostics Check	OK	08:34
Air Blank	0.000	08:34
Control Test	0.082	08:35
Air Blank	0.000	08:35
Subject Sample # 1	0.000	08:36
Breath Volume	1.183	08:36
Air Blank	0.000	08:36
Air Blank	0.000	08:38
Subject Sample # 2	0.000	08:39
Breath Volume	1.433	08:39
Air Blank	0.000	08:39
Control Test	0.080	08:39
Air Blank	0.000	08:40
Diagnostics Check	OK	08:40

Data Download Date/Time: 07/07/2025 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 06/23/2025 02:52:24

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.080	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.060	02:47
Breath Volume	1.894	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.062	02:50
Breath Volume	1.597	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 07/07/2025 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:47

SUBJECT'S LAST NAME: LEAL HERNANDEZ

DATE/TIME OF TEST: 06/23/2025 11:13:14

Diagnostics Check	OK	11:07
Air Blank	0.000	11:07
Control Test	0.081	11:08
Air Blank	0.000	11:08
Subject Sample # 1	0.000	11:09
Breath Volume	1.328	11:09
Air Blank	0.000	11:09
Air Blank	0.000	11:11
Subject Sample # 2	0.000	11:11
Breath Volume	1.667	11:11
Air Blank	0.000	11:12
Control Test	0.081	11:12
Air Blank	0.000	11:13
Diagnostics Check	OK	11:13

Data Download Date/Time: 07/07/2025 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: ORTIZ RAMOS

DATE/TIME OF TEST: 06/27/2025 01:16:38

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.081	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.172	01:11
Breath Volume	2.296	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.178	01:15
Breath Volume	2.574	01:15
Air Blank	0.000	01:15
Control Test	0.081	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 07/07/2025 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 06/27/2025 09:41:43

Diagnostics Check	OK	09:35
Air Blank	0.000	09:35
Control Test	0.081	09:36
Air Blank	0.000	09:36
Subject Sample # 1	0.000	09:37
Breath Volume	2.039	09:37
Air Blank	0.000	09:37
Air Blank	0.000	09:39
Subject Sample # 2	0.000	09:40
Breath Volume	2.335	09:40
Air Blank	0.000	09:40
Control Test	0.080	09:41
Air Blank	0.000	09:41
Diagnostics Check	OK	09:41

Data Download Date/Time: 07/07/2025 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: PAYANO

DATE/TIME OF TEST: 06/28/2025 03:31:44

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.160	03:27
Breath Volume	1.710	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.143	03:30
Breath Volume	1.839	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 07/07/2025 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Hamilton County SO
ARRESTING OFFICER: WATSON, JOSEPH
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 06/12/2025 04:43:08

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:40	
Control Test	0.084	04:40	
Air Blank	0.000	04:41	
Subject Sample	# 1	04:41	Subject Test Refused
Breath Volume	0.000	04:41	
Air Blank	0.000	04:42	
Control Test	0.083	04:42	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

Data Download Date/Time: 06/26/2025 01:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,JOSEPH,R
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: HOWELL,DANIEL
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: LARGAESPADA

DATE/TIME OF TEST: 06/27/2025 05:26:16

Diagnostics Check	OK	05:16
Air Blank	0.000	05:17
Control Test	0.083	05:17
Air Blank	0.000	05:18
Subject Sample # 1	0.247	05:18
Breath Volume	1.218	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.222	05:21
Breath Volume	1.148	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 3	0.242	05:24
Breath Volume	1.679	05:24
Air Blank	0.000	05:25
Control Test	0.083	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 07/30/2025 01:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS,KODEY,L
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: BRAGDON,JESSE
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:47

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 06/29/2025 15:10:08

Diagnostics Check	OK	15:07	
Air Blank	0.000	15:07	
Control Test	0.082	15:08	
Air Blank	0.000	15:08	
Subject Sample	# 1	15:08	Subject Test Refused
Breath Volume	0.000	15:08	
Air Blank	0.000	15:09	
Control Test	0.082	15:09	
Air Blank	0.000	15:10	
Diagnostics Check	OK	15:10	

Data Download Date/Time: 07/30/2025 01:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,D
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: NORRIS,JONATHAN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/24/2025 00:14:54

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.000	00:10
Breath Volume	2.617	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.000	00:13
Breath Volume	2.976	00:13
Air Blank	0.000	00:14
Control Test	0.078	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 06/26/2025 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRER,FRANCISCO,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: FERRER,FRANCISCO
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: SHIRLEY

DATE/TIME OF TEST: 06/01/2025 17:22:40

Diagnostics Check	OK	17:22	
Air Blank		17:22	RFI Detect
Air Blank		17:22	RFI Detect

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRER,FRANCISCO,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: FERRER,FRANCISCO
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: SHIRLEY

DATE/TIME OF TEST: 06/01/2025 17:32:17

Diagnostics Check	OK	17:31	
Air Blank		17:32	RFI Detect
Air Blank		17:32	RFI Detect

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRER,FRANCISCO,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: FERRER,FRANCISCO
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: SHIRLEY

DATE/TIME OF TEST: 06/01/2025 17:40:34

Diagnostics Check	OK	17:34
Air Blank	0.000	17:34
Control Test	0.078	17:34
Air Blank	0.000	17:35
Subject Sample # 1	0.000	17:36
Breath Volume	2.707	17:36
Air Blank	0.000	17:36
Air Blank	0.000	17:38
Subject Sample # 2	0.000	17:39
Breath Volume	2.437	17:39
Air Blank	0.000	17:39
Control Test	0.079	17:39
Air Blank	0.000	17:40
Diagnostics Check	OK	17:40

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOEHM, ADAM, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON, T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 06/05/2025 20:07:29

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.078	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.164	20:03
Breath Volume	1.875	20:03
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.145	20:05
Breath Volume	1.496	20:05
Air Blank	0.000	20:06
Control Test	0.077	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JAMES,R
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: DAVIS,JAMES
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: WARKE

DATE/TIME OF TEST: 06/11/2025 22:59:54

Diagnostics Check	OK	22:59	
Air Blank		22:59	RFI Detect
Air Blank	0.000	22:59	

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JAMES,R
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: DAVIS,JAMES
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: WARKE

DATE/TIME OF TEST: 06/11/2025 23:05:25

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:02	
Control Test	0.079	23:03	
Air Blank	0.000	23:03	
Subject Sample	# 1	23:04	Sequence Aborted
Breath Volume	0.000	23:04	
Air Blank	0.000	23:05	

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JAMES,R
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: DAVIS,JAMES
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: WARKE

DATE/TIME OF TEST: 06/11/2025 23:12:35

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:09	
Control Test	0.079	23:09	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:11	RFI Detect
Breath Volume	0.000	23:11	
Air Blank	0.000	23:11	
Control Test	0.079	23:11	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: JOHNSON,DALE
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/13/2025 18:17:41

Diagnostics Check	OK	18:11
Air Blank	0.000	18:11
Control Test	0.078	18:12
Air Blank	0.000	18:12
Subject Sample # 1	0.158	18:13
Breath Volume	2.726	18:13
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.143	18:16
Breath Volume	2.464	18:16
Air Blank	0.000	18:16
Control Test	0.078	18:17
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: MARTIN,ALEXANDER
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:02

SUBJECT'S LAST NAME: HERNERYD

DATE/TIME OF TEST: 06/14/2025 16:23:06

Diagnostics Check	OK	16:20	
Air Blank	0.000	16:20	
Control Test	0.079	16:21	
Air Blank	0.000	16:21	
Subject Sample	# 1	16:21	Subject Test Refused
Breath Volume	0.000	16:21	
Air Blank	0.000	16:22	
Control Test	0.078	16:22	
Air Blank	0.000	16:23	
Diagnostics Check	OK	16:23	

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOEHM, ADAM, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DAVIS, J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 06/14/2025 23:39:27

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.079	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.151	23:34
Breath Volume	2.031	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 2	0.150	23:37
Breath Volume	2.113	23:37
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOEHM, ADAM, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BALAFAS, J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/19/2025 00:31:52

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.079	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.180	00:27
Breath Volume	1.644	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.183	00:30
Breath Volume	1.359	00:30
Air Blank	0.000	00:30
Control Test	0.078	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DAVIS,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: BULLOCK

DATE/TIME OF TEST: 06/21/2025 03:10:05

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:58	
Control Test	0.079	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 1	03:02	No Sample Provided
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Air Blank	0.000	03:04	
Subject Sample	# 2	0.225 03:05	
Breath Volume	2.878	03:05	
Air Blank	0.000	03:06	
Air Blank	0.000	03:08	
Subject Sample	# 3	0.231 03:08	
Breath Volume	3.121	03:08	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TROY,E
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: WASYLYK,TYLER
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:06

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/21/2025 07:42:45

Diagnostics Check	OK	07:41	
Air Blank	0.000	07:41	
Control Test	0.077	07:42	
Air Blank		07:42	RFI Detect
Air Blank		07:42	Improper Sample

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TROY,E
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GETZ,KEVIN
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/21/2025 16:27:01

Diagnostics Check	OK	16:26	
Air Blank		16:26	Improper Sample
Air Blank		16:27	Improper Sample

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOEHM, ADAM, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DAVIS, J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PHOENIX

DATE/TIME OF TEST: 06/28/2025 22:43:19

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:40	
Control Test	0.080	22:41	
Air Blank	0.000	22:41	
Subject Sample	# 1	22:41	Subject Test Refused
Breath Volume	0.000	22:41	
Air Blank	0.000	22:42	
Control Test	0.079	22:42	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

Data Download Date/Time: 06/30/2025 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JAMES,R
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/30/2025 17:54:41

Diagnostics Check	OK	17:43	
Air Blank	0.000	17:44	
Control Test	0.078	17:44	
Air Blank	0.000	17:44	
Subject Sample	# 1	17:47	No Sample Provided
Breath Volume	0.000	17:47	
Air Blank	0.000	17:48	
Air Blank	0.000	17:50	
Subject Sample	# 2	17:53	RFI Detect
Breath Volume	0.000	17:53	
Air Blank	0.000	17:53	
Control Test	0.078	17:54	
Air Blank	0.000	17:54	
Diagnostics Check	OK	17:54	

Data Download Date/Time: 07/31/2025 20:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:51

SUBJECT'S LAST NAME: SVIHLIK

DATE/TIME OF TEST: 06/24/2025 18:19:22

Diagnostics Check	OK	18:11
Air Blank	0.000	18:12
Control Test	0.079	18:12
Air Blank	0.000	18:13
Subject Sample # 1	0.106	18:13
Breath Volume	1.332	18:13
Air Blank	0.000	18:14
Air Blank	0.000	18:16
Subject Sample # 2	0.106	18:17
Breath Volume	1.574	18:17
Air Blank	0.000	18:18
Control Test	0.079	18:18
Air Blank	0.000	18:19
Diagnostics Check	OK	18:19

Data Download Date/Time: 07/25/2025 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 06/29/2025 02:29:06

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:24	
Control Test	0.079	02:24	
Air Blank	0.000	02:25	
Subject Sample	# 1 0.218	02:25	
Breath Volume	1.941	02:25	
Air Blank	0.000	02:26	
Air Blank	0.000	02:28	
Subject Sample	# 2	02:28	RFI Detect
Breath Volume	0.035	02:28	
Air Blank		02:28	RFI Detect
Air Blank	0.000	02:29	

Data Download Date/Time: 07/25/2025 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 06/29/2025 02:32:04

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:30	
Control Test	0.079	02:31	
Air Blank		02:31	RFI Detect
Air Blank	0.000	02:32	

Data Download Date/Time: 07/25/2025 16:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,A,
BREATH TEST OPERATOR AGENCY: Mascotte PD
ARRESTING OFFICER: JOHNSON,M
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: SOTO HERNADEZ

DATE/TIME OF TEST: 06/08/2025 00:04:37

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:02	
Control Test	0.080	00:02	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:03	
Control Test	0.079	00:04	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 06/25/2025 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIMER,J,A
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LATIMER,J
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/28/2025 03:10:05

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:07	
Control Test	0.080	03:07	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 07/07/2025 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIMER,J,A
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LATIMER,J
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: HUTH JR

DATE/TIME OF TEST: 06/29/2025 04:38:44

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:36	
Control Test	0.080	04:36	
Air Blank	0.000	04:37	
Subject Sample	# 1	04:37	Subject Test Refused
Breath Volume	0.000	04:37	
Air Blank	0.000	04:37	
Control Test	0.080	04:38	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

Data Download Date/Time: 07/07/2025 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: NETH

DATE/TIME OF TEST: 06/01/2025 23:15:42

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.080	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.064	23:11
Breath Volume	1.675	23:11
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.061	23:14
Breath Volume	1.917	23:14
Air Blank	0.000	23:14
Control Test	0.079	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 06/12/2025 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: SIMONENKO

DATE/TIME OF TEST: 06/15/2025 21:02:15

Diagnostics Check	OK	20:55
Air Blank	0.000	20:56
Control Test	0.081	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.149	20:57
Breath Volume	2.355	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.136	21:00
Breath Volume	2.109	21:00
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 07/04/2025 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT',R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: SWINDLE

DATE/TIME OF TEST: 06/16/2025 00:10:55

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.054	00:06
Breath Volume	2.867	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.049	00:09
Breath Volume	2.156	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 07/04/2025 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: SERGENT

DATE/TIME OF TEST: 06/24/2025 00:59:37

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.081	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.000	00:55
Breath Volume	2.742	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.000	00:58
Breath Volume	2.078	00:58
Air Blank	0.000	00:58
Control Test	0.081	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 07/04/2025 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: BITUIN

DATE/TIME OF TEST: 06/24/2025 22:54:55

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.081	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.000	22:50
Breath Volume	2.394	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.000	22:53
Breath Volume	1.730	22:53
Air Blank	0.000	22:54
Control Test	0.081	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 07/04/2025 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: AGNEW

DATE/TIME OF TEST: 06/28/2025 22:32:45

Diagnostics Check	OK	22:26
Air Blank	0.000	22:27
Control Test	0.081	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.000	22:28
Breath Volume	2.820	22:28
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.000	22:31
Breath Volume	2.484	22:31
Air Blank	0.000	22:31
Control Test	0.081	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 07/04/2025 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAGUADOG,SHAUN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MAGUADOG,SHAUN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 14:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:51

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 06/05/2025 15:19:47

Diagnostics Check	OK	15:11
Air Blank	0.000	15:11
Control Test	0.080	15:12
Air Blank	0.000	15:12
Subject Sample # 1	0.251	15:14
Breath Volume	1.941	15:14
Air Blank	0.000	15:14
Air Blank	0.000	15:16
Subject Sample # 2	0.248	15:18
Breath Volume	1.675	15:18
Air Blank	0.000	15:18
Control Test	0.079	15:19
Air Blank	0.000	15:19
Diagnostics Check	OK	15:19

Data Download Date/Time: 07/03/2025 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: BLAKELY

DATE/TIME OF TEST: 06/07/2025 03:36:37

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.198	03:32
Breath Volume	2.167	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.187	03:35
Breath Volume	1.835	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 07/03/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: SEARS

DATE/TIME OF TEST: 06/08/2025 02:14:31

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.170	02:10
Breath Volume	2.882	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.177	02:13
Breath Volume	2.742	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 07/03/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: SPANNBAUER

DATE/TIME OF TEST: 06/08/2025 19:50:33

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:35	
Control Test	0.081	19:35	
Air Blank	0.000	19:36	
Subject Sample	# 1 0.089	19:39	Volume Not Met
Breath Volume	0.246	19:39	
Air Blank	0.000	19:40	
Air Blank	0.000	19:41	
Subject Sample	# 2 0.101	19:44	
Breath Volume	1.984	19:44	
Air Blank	0.000	19:44	
Air Blank	0.000	19:46	
Subject Sample	# 3 0.091	19:49	
Breath Volume	1.539	19:49	
Air Blank	0.000	19:49	
Control Test	0.080	19:49	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:50	

Data Download Date/Time: 07/03/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 06/09/2025 00:44:25

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:32	
Control Test	0.081	00:32	
Air Blank	0.000	00:33	
Subject Sample	# 1 0.201	00:36	Volume Not Met
Breath Volume	0.980	00:36	
Air Blank	0.000	00:36	
Air Blank	0.000	00:38	
Subject Sample	# 2 0.238	00:39	
Breath Volume	2.035	00:39	
Air Blank	0.000	00:40	
Air Blank	0.000	00:42	
Subject Sample	# 3 0.233	00:42	
Breath Volume	1.367	00:42	
Air Blank	0.000	00:43	
Control Test	0.081	00:43	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 07/03/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALNAKER, TEDDY, L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MIDYETTE, COREY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: SEYMOUR

DATE/TIME OF TEST: 06/11/2025 17:34:34

Diagnostics Check	OK	17:28
Air Blank	0.000	17:28
Control Test	0.082	17:29
Air Blank	0.000	17:29
Subject Sample # 1	0.132	17:30
Breath Volume	1.914	17:30
Air Blank	0.000	17:30
Air Blank	0.000	17:32
Subject Sample # 2	0.136	17:33
Breath Volume	2.367	17:33
Air Blank	0.000	17:33
Control Test	0.081	17:33
Air Blank	0.000	17:34
Diagnostics Check	OK	17:34

Data Download Date/Time: 07/03/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOUTH,DAMIAN,W
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SHAVER,COLTON
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: ASHBY

DATE/TIME OF TEST: 06/12/2025 20:19:08

Diagnostics Check	OK	20:12
Air Blank	0.000	20:13
Control Test	0.081	20:13
Air Blank	0.000	20:14
Subject Sample # 1	0.000	20:14
Breath Volume	2.867	20:14
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.000	20:17
Breath Volume	2.949	20:17
Air Blank	0.000	20:18
Control Test	0.082	20:18
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 07/03/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONAHAN,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SMITH,DAVON
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 06/15/2025 17:52:04

Diagnostics Check	OK	17:45
Air Blank	0.000	17:45
Control Test	0.081	17:45
Air Blank	0.000	17:46
Subject Sample # 1	0.154	17:47
Breath Volume	2.695	17:47
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample # 2	0.153	17:50
Breath Volume	3.363	17:50
Air Blank	0.000	17:51
Control Test	0.080	17:51
Air Blank	0.000	17:51
Diagnostics Check	OK	17:52

Data Download Date/Time: 07/03/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 06/18/2025 01:28:09

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.081	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.000	01:23
Breath Volume	2.617	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.000	01:26
Breath Volume	2.421	01:26
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 07/03/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAY,JON,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: IRONS,MASON
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: BLACKBURN

DATE/TIME OF TEST: 06/20/2025 02:02:18

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.081	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.200	01:57
Breath Volume	1.363	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.199	02:00
Breath Volume	1.808	02:00
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 07/03/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAY,JON,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: TURNER,MACKENZIE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/21/2025 04:50:54

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.081	04:45
Air Blank	0.000	04:45
Subject Sample # 1	0.108	04:46
Breath Volume	2.542	04:46
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.115	04:49
Breath Volume	2.488	04:49
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 07/03/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAY,JON,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HOLBERT,TYLER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/22/2025 00:50:17

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.081	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.176	00:45
Breath Volume	2.628	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.172	00:48
Breath Volume	1.945	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 07/03/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAK,STEVEN,M
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MORENO,JACOB
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: FAULK

DATE/TIME OF TEST: 06/27/2025 01:27:34

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:24	
Control Test	0.081	01:25	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:26	
Control Test	0.080	01:26	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

Data Download Date/Time: 07/03/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAK,STEVEN,M
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: STONE,SYDNEY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 06/27/2025 04:55:18

Diagnostics Check	OK	04:47
Air Blank	0.000	04:48
Control Test	0.081	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.243	04:50
Breath Volume	1.167	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.262	04:53
Breath Volume	1.210	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 07/03/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: MILEWSKI

DATE/TIME OF TEST: 06/28/2025 23:53:20

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.082	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.186	23:48
Breath Volume	4.386	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.192	23:51
Breath Volume	3.371	23:51
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 07/03/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 06/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 10:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:16

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 06/11/2025 10:39:30

Diagnostics Check	OK	10:38	
Air Blank		10:39	RFI Detect
Air Blank	0.000	10:39	

Data Download Date/Time: 07/02/2025 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 06/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 10:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:16

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 06/11/2025 10:47:32

Diagnostics Check	OK	10:41
Air Blank	0.000	10:42
Control Test	0.083	10:42
Air Blank	0.000	10:42
Subject Sample # 1	0.000	10:43
Breath Volume	3.246	10:43
Air Blank	0.000	10:43
Air Blank	0.000	10:45
Subject Sample # 2	0.000	10:46
Breath Volume	3.214	10:46
Air Blank	0.000	10:46
Control Test	0.083	10:46
Air Blank	0.000	10:47
Diagnostics Check	OK	10:47

Data Download Date/Time: 07/02/2025 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 06/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 06/11/2025 17:46:32

Diagnostics Check	OK	17:43	
Air Blank	0.000	17:44	
Control Test	0.081	17:44	
Air Blank	0.000	17:44	
Subject Sample	# 1	17:45	Subject Test Refused
Breath Volume	0.000	17:45	
Air Blank	0.000	17:45	
Control Test	0.080	17:45	
Air Blank	0.000	17:46	
Diagnostics Check	OK	17:46	

Data Download Date/Time: 07/02/2025 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 06/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 06/17/2025 06:39:00

Diagnostics Check	OK	06:32
Air Blank	0.000	06:33
Control Test	0.083	06:33
Air Blank	0.000	06:34
Subject Sample # 1	0.000	06:34
Breath Volume	3.386	06:34
Air Blank	0.000	06:35
Air Blank	0.000	06:37
Subject Sample # 2	0.000	06:37
Breath Volume	2.851	06:37
Air Blank	0.000	06:38
Control Test	0.082	06:38
Air Blank	0.000	06:38
Diagnostics Check	OK	06:39

Data Download Date/Time: 07/02/2025 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 06/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 06/20/2025 06:24:55

Diagnostics Check	OK	06:18
Air Blank	0.000	06:19
Control Test	0.083	06:19
Air Blank	0.000	06:20
Subject Sample # 1	0.000	06:20
Breath Volume	2.945	06:20
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample # 2	0.000	06:23
Breath Volume	2.949	06:23
Air Blank	0.000	06:24
Control Test	0.082	06:24
Air Blank	0.000	06:24
Diagnostics Check	OK	06:24

Data Download Date/Time: 07/02/2025 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 06/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 06/25/2025 06:40:17

Diagnostics Check	OK	06:34
Air Blank	0.000	06:34
Control Test	0.083	06:35
Air Blank	0.000	06:35
Subject Sample # 1	0.000	06:36
Breath Volume	3.304	06:36
Air Blank	0.000	06:36
Air Blank	0.000	06:38
Subject Sample # 2	0.000	06:38
Breath Volume	2.562	06:38
Air Blank	0.000	06:39
Control Test	0.082	06:39
Air Blank	0.000	06:40
Diagnostics Check	OK	06:40

Data Download Date/Time: 07/02/2025 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 06/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 06:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 06/26/2025 06:36:21

Diagnostics Check	OK	06:30
Air Blank	0.000	06:30
Control Test	0.083	06:31
Air Blank	0.000	06:31
Subject Sample # 1	0.000	06:32
Breath Volume	2.718	06:32
Air Blank	0.000	06:32
Air Blank	0.000	06:34
Subject Sample # 2	0.000	06:34
Breath Volume	2.906	06:34
Air Blank	0.000	06:35
Control Test	0.082	06:35
Air Blank	0.000	06:36
Diagnostics Check	OK	06:36

Data Download Date/Time: 07/02/2025 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 06/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 06:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 06/30/2025 06:47:36

Diagnostics Check	OK	06:41
Air Blank	0.000	06:42
Control Test	0.083	06:42
Air Blank	0.000	06:42
Subject Sample # 1	0.000	06:43
Breath Volume	3.140	06:43
Air Blank	0.000	06:43
Air Blank	0.000	06:45
Subject Sample # 2	0.000	06:46
Breath Volume	3.132	06:46
Air Blank	0.000	06:46
Control Test	0.082	06:47
Air Blank	0.000	06:47
Diagnostics Check	OK	06:47

Data Download Date/Time: 07/02/2025 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: GALLEGOS

DATE/TIME OF TEST: 06/01/2025 21:56:53

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:41	
Control Test	0.079	21:41	
Air Blank	0.000	21:42	
Subject Sample	# 1 0.098	21:45	Volume Not Met
Breath Volume	0.242	21:45	
Air Blank	0.000	21:46	
Air Blank	0.000	21:48	
Subject Sample	# 2 0.103	21:49	
Breath Volume	1.652	21:49	
Air Blank	0.000	21:50	
Air Blank	0.000	21:52	
Subject Sample	# 3 0.105	21:55	
Breath Volume	1.875	21:55	
Air Blank	0.000	21:55	
Control Test	0.079	21:56	
Air Blank	0.000	21:56	
Diagnostics Check	OK	21:56	

Data Download Date/Time: 06/09/2025 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CORTES,CHRISTIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: BROWNING

DATE/TIME OF TEST: 06/01/2025 23:19:14

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:16	
Control Test	0.079	23:17	
Air Blank	0.000	23:17	
Subject Sample	# 1	23:17	Subject Test Refused
Breath Volume	0.000	23:17	
Air Blank	0.000	23:18	
Control Test	0.079	23:18	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 06/09/2025 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTYRE,DELLVON,B
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RANCEL,RANDY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 06/03/2025 03:38:31

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.118	03:32
Breath Volume	2.144	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.105	03:37
Breath Volume	1.988	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 06/09/2025 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/07/2025 01:38:21

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.078	01:36	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.079	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 06/09/2025 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HANNA,WARDELL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/08/2025 21:48:24

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:45	
Control Test	0.079	21:45	
Air Blank	0.000	21:45	
Subject Sample	# 1	21:47	Subject Test Refused
Breath Volume	0.000	21:47	
Air Blank	0.000	21:47	
Control Test	0.078	21:47	
Air Blank	0.000	21:48	
Diagnostics Check	OK	21:48	

Data Download Date/Time: 06/09/2025 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 06/11/2025 01:48:08

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:45	
Control Test	0.079	01:45	
Air Blank	0.000	01:46	
Subject Sample	# 1	01:46	Subject Test Refused
Breath Volume	0.000	01:46	
Air Blank	0.000	01:47	
Control Test	0.079	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 07/17/2025 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RANCEL,RANDY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: PEREZ OLIVA

DATE/TIME OF TEST: 06/13/2025 00:32:32

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:30	
Control Test	0.079	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:31	
Control Test	0.079	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 07/17/2025 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,OSCAR,D
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARIN,MARIO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ARGUDELO CAMARGO

DATE/TIME OF TEST: 06/13/2025 21:56:46

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:41	
Control Test	0.079	21:41	
Air Blank	0.000	21:42	
Subject Sample	# 1 0.095	21:45	Volume Not Met
Breath Volume	0.132	21:45	
Air Blank	0.000	21:45	
Air Blank	0.000	21:47	
Subject Sample	# 2 0.127	21:49	
Breath Volume	1.792	21:49	
Air Blank	0.000	21:50	
Air Blank	0.000	21:52	
Subject Sample	# 3 0.087	21:55	Volume Not Met
Breath Volume	0.195	21:55	
Air Blank	0.000	21:55	
Control Test	0.078	21:56	
Air Blank	0.000	21:56	
Diagnostics Check	OK	21:56	

Data Download Date/Time: 07/17/2025 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARIN,MARIO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: JNO-PIERRE

DATE/TIME OF TEST: 06/15/2025 02:02:35

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.079	01:48	
Air Blank	0.000	01:49	
Subject Sample	# 1 0.145	01:50	
Breath Volume	1.117	01:50	
Air Blank	0.000	01:51	
Air Blank	0.000	01:53	
Subject Sample	# 2 0.139	01:56	Volume Not Met
Breath Volume	0.375	01:56	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 3 0.140	02:01	
Breath Volume	1.578	02:01	
Air Blank	0.000	02:01	
Control Test	0.077	02:01	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 07/17/2025 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTYRE,DELLVON,B
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MCINTYRE,DELLVON
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: LEONTI

DATE/TIME OF TEST: 06/18/2025 03:30:04

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.281	03:25
Breath Volume	1.863	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.286	03:28
Breath Volume	1.812	03:28
Air Blank	0.000	03:29
Control Test	0.078	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:30

Data Download Date/Time: 07/17/2025 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 06/18/2025 16:47:55

Diagnostics Check	OK	16:44	
Air Blank	0.000	16:45	
Control Test	0.079	16:45	
Air Blank	0.000	16:45	
Subject Sample	# 1	16:46	Subject Test Refused
Breath Volume	0.000	16:46	
Air Blank	0.000	16:47	
Control Test	0.078	16:47	
Air Blank	0.000	16:47	
Diagnostics Check	OK	16:47	

Data Download Date/Time: 07/17/2025 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/18/2025 23:30:42

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.079	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.081	23:26
Breath Volume	1.460	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.079	23:29
Breath Volume	1.402	23:29
Air Blank	0.000	23:29
Control Test	0.079	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 07/17/2025 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,MATTHEW
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 06/24/2025 01:04:28

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.078	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.249	00:59
Breath Volume	1.917	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.240	01:02
Breath Volume	1.226	01:02
Air Blank	0.000	01:03
Control Test	0.078	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 07/17/2025 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: GARCIA-MARTINEZ

DATE/TIME OF TEST: 06/25/2025 00:49:59

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.079	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Control Test	0.079	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 07/17/2025 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDON DEL RIO, RODOLFO, E
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: RONDON DEL RIO, RODOLFO
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 10:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: VECIC

DATE/TIME OF TEST: 06/26/2025 23:31:52

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.214	23:26
Breath Volume	1.437	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.216	23:30
Breath Volume	1.789	23:30
Air Blank	0.000	23:30
Control Test	0.077	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 07/17/2025 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: GONZALEZ-QUEVEDO

DATE/TIME OF TEST: 06/27/2025 23:01:02

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:50	
Control Test	0.079	22:50	
Air Blank	0.000	22:50	
Subject Sample	# 1 0.000	22:54	Volume Not Met
Breath Volume	0.070	22:54	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample	# 2	22:59	No Sample Provided
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.079	23:00	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 07/17/2025 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CORTES,CHRISTIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 06/28/2025 04:29:15

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:24	
Control Test	0.078	04:24	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:27	Subject Test Refused
Breath Volume	0.000	04:27	
Air Blank	0.000	04:28	
Control Test	0.078	04:28	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

Data Download Date/Time: 07/17/2025 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CASTRO,SHANE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: NAVARRETE CAPAZAN

DATE/TIME OF TEST: 06/29/2025 01:50:51

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.060	01:46
Breath Volume	1.613	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.058	01:49
Breath Volume	1.906	01:49
Air Blank	0.000	01:49
Control Test	0.078	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 07/17/2025 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 06/29/2025 20:47:42

Diagnostics Check	OK	20:39	
Air Blank	0.000	20:39	
Control Test	0.079	20:40	
Air Blank	0.000	20:40	
Subject Sample	# 1	20:43	No Sample Provided
Breath Volume	0.000	20:43	
Air Blank	0.000	20:44	
Air Blank	0.000	20:46	
Subject Sample	# 2	20:46	Subject Test Refused
Breath Volume	0.000	20:46	
Air Blank	0.000	20:46	
Control Test	0.078	20:47	
Air Blank	0.000	20:47	
Diagnostics Check	OK	20:47	

Data Download Date/Time: 07/17/2025 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MATTHEW,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,MATTHEW
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: WYNE

DATE/TIME OF TEST: 06/29/2025 22:01:49

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.078	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.342	21:56
Breath Volume	2.292	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.340	22:00
Breath Volume	2.167	22:00
Air Blank	0.000	22:00
Control Test	0.078	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 07/17/2025 13:25