

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRAW, KAELEE, R
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRERO-MARTINEZ, KEVEN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 06/15/2025 00:14:57

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.080	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.179	00:10
Breath Volume	2.218	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:12
Subject Sample # 2	0.181	00:13
Breath Volume	2.445	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 07/17/2025 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 06/17/2025 23:32:22

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.043	23:27
Breath Volume	2.664	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.045	23:30
Breath Volume	2.738	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 07/17/2025 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,LEANN,E
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: LEON,KEVIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/22/2025 04:27:21

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.080	04:25	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:26	
Control Test	0.079	04:26	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 07/17/2025 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRERO-MARTINEZ, KEVEN,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ZITO, ANTHONY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 06/27/2025 19:54:30

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:50	
Control Test	0.079	19:50	
Air Blank	0.000	19:51	
Subject Sample	# 1	19:53	Subject Test Refused
Breath Volume	0.000	19:53	
Air Blank	0.000	19:53	
Control Test	0.079	19:53	
Air Blank	0.000	19:54	
Diagnostics Check	OK	19:54	

Data Download Date/Time: 07/17/2025 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRAW,KAELEE,R
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CASANOVA,AVRIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: YARBROUGH

DATE/TIME OF TEST: 06/30/2025 02:05:43

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.080	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.080	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 07/17/2025 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: PALOMINO

DATE/TIME OF TEST: 06/28/2025 05:06:45

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.080	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.174	05:02
Breath Volume	1.617	05:02
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.174	05:05
Breath Volume	1.335	05:05
Air Blank	0.000	05:05
Control Test	0.079	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 08/05/2025 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKER,ALEXA,C
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: GARCIA,DASHIELL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: SOLANKI

DATE/TIME OF TEST: 06/16/2025 18:39:33

Diagnostics Check	OK	18:32
Air Blank	0.000	18:32
Control Test	0.080	18:33
Air Blank	0.000	18:33
Subject Sample # 1	0.182	18:34
Breath Volume	1.257	18:34
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.178	18:37
Breath Volume	1.296	18:37
Air Blank	0.000	18:38
Control Test	0.078	18:38
Air Blank	0.000	18:39
Diagnostics Check	OK	18:39

Data Download Date/Time: 06/27/2025 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: HERRERA HERNANDEZ

DATE/TIME OF TEST: 06/22/2025 01:53:43

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.079	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.079	01:53	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 06/27/2025 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: WHITE,AVERY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402732434

DRY GAS EXPIRY DATE: 04/28/2026

VIOLATION CODE: DUI

ARREST TIME: 01:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PEREZ VELASQUEZ

DATE/TIME OF TEST: 06/23/2025 02:37:05

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.267	02:32
Breath Volume	1.230	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.253	02:35
Breath Volume	1.753	02:35
Air Blank	0.000	02:36
Control Test	0.078	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:37

Data Download Date/Time: 06/27/2025 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MELANIE,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PEREZ,MELISSA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: LINARTE

DATE/TIME OF TEST: 06/23/2025 05:14:25

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:07	
Control Test	0.079	05:08	
Air Blank	0.000	05:08	
Subject Sample	# 1 0.235	05:10	
Breath Volume	2.125	05:10	
Air Blank	0.000	05:11	
Air Blank	0.000	05:12	
Subject Sample	# 2 0.229	05:12	
Breath Volume	1.679	05:12	
Air Blank	0.000	05:13	
Control Test	0.072	05:13	Control Outside Tolerance
Air Blank	0.000	05:14	

Data Download Date/Time: 06/27/2025 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 15:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 06/03/2025 16:05:39

Diagnostics Check	OK	15:59
Air Blank	0.000	16:00
Control Test	0.082	16:00
Air Blank	0.000	16:01
Subject Sample # 1	0.000	16:01
Breath Volume	2.546	16:01
Air Blank	0.000	16:01
Air Blank	0.000	16:03
Subject Sample # 2	0.000	16:04
Breath Volume	2.500	16:04
Air Blank	0.000	16:04
Control Test	0.082	16:05
Air Blank	0.000	16:05
Diagnostics Check	OK	16:05

Data Download Date/Time: 06/20/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 06/06/2025 06:07:51

Diagnostics Check	OK	06:01
Air Blank	0.000	06:02
Control Test	0.081	06:02
Air Blank	0.000	06:03
Subject Sample # 1	0.000	06:03
Breath Volume	3.246	06:03
Air Blank	0.000	06:04
Air Blank	0.000	06:06
Subject Sample # 2	0.000	06:06
Breath Volume	3.324	06:06
Air Blank	0.000	06:06
Control Test	0.082	06:07
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 06/20/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,PAUL,G
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: YOUNG,PAUL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/16/2025 20:15:51

Diagnostics Check	OK	20:09
Air Blank	0.000	20:10
Control Test	0.081	20:10
Air Blank	0.000	20:10
Subject Sample # 1	0.122	20:11
Breath Volume	1.789	20:11
Air Blank	0.000	20:12
Air Blank	0.000	20:13
Subject Sample # 2	0.107	20:14
Breath Volume	1.785	20:14
Air Blank	0.000	20:14
Control Test	0.078	20:15
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 06/20/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES, EMMANUEL,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: MARIN, ALEJANDRO
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: HERNANDEZ GONZALEZ

DATE/TIME OF TEST: 06/08/2025 05:01:05

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:58	
Control Test	0.082	04:58	
Air Blank	0.000	04:59	
Subject Sample	# 1	04:59	Subject Test Refused
Breath Volume	0.000	04:59	
Air Blank	0.000	05:00	
Control Test	0.082	05:00	
Air Blank	0.000	05:00	
Diagnostics Check	OK	05:01	

Data Download Date/Time: 07/02/2025 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DAVID,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: ORFILA,A
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: SUAREZ LEMUS

DATE/TIME OF TEST: 06/15/2025 08:33:32

Diagnostics Check	OK	08:30	
Air Blank	0.000	08:31	
Control Test	0.083	08:31	
Air Blank	0.000	08:31	
Subject Sample	# 1	08:32	Subject Test Refused
Breath Volume	0.000	08:32	
Air Blank	0.000	08:32	
Control Test	0.082	08:32	
Air Blank	0.000	08:33	
Diagnostics Check	OK	08:33	

Data Download Date/Time: 07/02/2025 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEJEDA, CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: BORROTO, JOSEPH
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MILANES-FONSECA

DATE/TIME OF TEST: 06/15/2025 23:48:21

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:45	
Control Test	0.083	23:46	
Air Blank	0.000	23:46	
Subject Sample	# 1	23:46	Subject Test Refused
Breath Volume	0.000	23:46	
Air Blank	0.000	23:47	
Control Test	0.082	23:47	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 07/02/2025 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVA,JUAN,C
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: OLIVA,JUAN
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: TRAVIESO

DATE/TIME OF TEST: 06/21/2025 19:05:31

Diagnostics Check	OK	19:02	
Air Blank	0.000	19:03	
Control Test	0.083	19:03	
Air Blank	0.000	19:03	
Subject Sample	# 1	19:04	Subject Test Refused
Breath Volume	0.000	19:04	
Air Blank	0.000	19:04	
Control Test	0.083	19:04	
Air Blank	0.000	19:05	
Diagnostics Check	OK	19:05	

Data Download Date/Time: 07/02/2025 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEJEDA,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: CRUZ,MARCOS
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: HERNANDEZ-JIMENEZ

DATE/TIME OF TEST: 06/21/2025 21:40:58

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:38	
Control Test	0.083	21:38	
Air Blank	0.000	21:39	
Subject Sample	# 1	21:39	Subject Test Refused
Breath Volume	0.000	21:39	
Air Blank	0.000	21:40	
Control Test	0.083	21:40	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:40	

Data Download Date/Time: 07/02/2025 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JOSE,C
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: OJEDA,RICHARD
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 12:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:03

SUBJECT'S LAST NAME: TIELES

DATE/TIME OF TEST: 06/30/2025 13:39:15

Diagnostics Check	OK	13:31
Air Blank	0.000	13:32
Control Test	0.082	13:32
Air Blank	0.000	13:33
Subject Sample # 1	0.197	13:33
Breath Volume	1.691	13:33
Air Blank	0.000	13:34
Air Blank	0.000	13:36
Subject Sample # 2	0.198	13:37
Breath Volume	2.371	13:37
Air Blank	0.000	13:38
Control Test	0.083	13:38
Air Blank	0.000	13:39
Diagnostics Check	OK	13:39

Data Download Date/Time: 07/02/2025 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JOSE,C
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: SALVATRE,ERCOLE
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 06/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/30/2025 20:34:50

Diagnostics Check	OK	20:31	
Air Blank	0.000	20:32	
Control Test	0.082	20:32	
Air Blank	0.000	20:33	
Subject Sample	# 1	20:33	Subject Test Refused
Breath Volume	0.000	20:33	
Air Blank	0.000	20:33	
Control Test	0.082	20:34	
Air Blank	0.000	20:34	
Diagnostics Check	OK	20:34	

Data Download Date/Time: 07/02/2025 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRANDON,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MILLER,BRANDON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SVENDBLAD

DATE/TIME OF TEST: 06/01/2025 00:03:20

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.077	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.246	23:58
Breath Volume	1.605	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.253	00:01
Breath Volume	1.804	00:01
Air Blank	0.000	00:02
Control Test	0.077	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 06/12/2025 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRERA TERRY,MELANIE,M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE,RENATO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: KOSTADINOVIC

DATE/TIME OF TEST: 06/02/2025 03:00:32

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.079	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.237	02:54
Breath Volume	3.156	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.247	02:58
Breath Volume	3.175	02:58
Air Blank	0.000	02:59
Control Test	0.076	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 06/12/2025 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAZARYAN,LILIT,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ALVAREZ,DANNY
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: SOLISMENENDEZ

DATE/TIME OF TEST: 06/02/2025 13:41:40

Diagnostics Check	OK	13:35
Air Blank	0.000	13:36
Control Test	0.078	13:36
Air Blank	0.000	13:37
Subject Sample # 1	0.000	13:37
Breath Volume	3.210	13:37
Air Blank	0.000	13:37
Air Blank	0.000	13:39
Subject Sample # 2	0.000	13:40
Breath Volume	2.832	13:40
Air Blank	0.000	13:40
Control Test	0.077	13:41
Air Blank	0.000	13:41
Diagnostics Check	OK	13:41

Data Download Date/Time: 06/12/2025 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARBALLO PEREZ,JOSE,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CARBALLO PEREZ,JOSE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: MORK

DATE/TIME OF TEST: 06/02/2025 19:45:00

Diagnostics Check	OK	19:37
Air Blank	0.000	19:38
Control Test	0.078	19:38
Air Blank	0.000	19:39
Subject Sample # 1	0.000	19:39
Breath Volume	1.808	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.000	19:43
Breath Volume	1.734	19:43
Air Blank	0.000	19:44
Control Test	0.078	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:45

Data Download Date/Time: 06/12/2025 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE,RENATO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: GUTIERREZ GALVEZ

DATE/TIME OF TEST: 06/06/2025 01:26:24

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:23	
Control Test	0.078	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:24	Subject Test Refused
Breath Volume	0.179	01:24	
Air Blank	0.000	01:25	
Control Test	0.077	01:25	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 06/12/2025 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: SIOJRIN

DATE/TIME OF TEST: 06/07/2025 04:57:55

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:55	
Control Test	0.078	04:55	
Air Blank	0.000	04:56	
Subject Sample	# 1	04:56	Subject Test Refused
Breath Volume	0.000	04:56	
Air Blank	0.000	04:57	
Control Test	0.078	04:57	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 06/12/2025 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAZARYAN,LILIT,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: BARRETT,DEVON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: MITLO

DATE/TIME OF TEST: 06/07/2025 21:47:57

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.079	21:42
Air Blank	0.000	21:42
Subject Sample # 1	0.025	21:43
Breath Volume	2.890	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:45
Subject Sample # 2	0.025	21:46
Breath Volume	1.332	21:46
Air Blank	0.000	21:47
Control Test	0.077	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 06/12/2025 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CUAREZMA,MICHAEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/13/2025 21:04:19

Diagnostics Check	OK	20:58	
Air Blank	0.000	20:59	
Control Test	0.078	20:59	
Air Blank	0.000	21:00	
Subject Sample	# 1	21:02	Subject Test Refused
Breath Volume	0.000	21:02	
Air Blank	0.000	21:03	
Control Test	0.078	21:03	
Air Blank	0.000	21:04	
Diagnostics Check	OK	21:04	

Data Download Date/Time: 07/17/2025 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRIDO,JEYNY,S
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE,R
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: CALLERI

DATE/TIME OF TEST: 06/15/2025 05:03:07

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.051	04:58
Breath Volume	2.222	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.054	05:01
Breath Volume	3.714	05:01
Air Blank	0.000	05:02
Control Test	0.078	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 07/17/2025 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,KELVIN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CUAREZMA,MICHAEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MARCO

DATE/TIME OF TEST: 06/19/2025 03:11:26

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.078	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.145	03:06
Breath Volume	1.921	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.145	03:09
Breath Volume	2.207	03:09
Air Blank	0.000	03:10
Control Test	0.078	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 07/17/2025 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STENBACK,AUSTIN,B
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: STENBACK,AUSTIN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: BUESTAN

DATE/TIME OF TEST: 06/20/2025 03:18:54

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.078	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.178	03:14
Breath Volume	1.351	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.171	03:17
Breath Volume	2.765	03:17
Air Blank	0.000	03:18
Control Test	0.078	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 07/17/2025 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAURE CASCARET, RENATO, A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE CASCARET, RENATO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CHACON GUTIERREZ

DATE/TIME OF TEST: 06/20/2025 23:01:39

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.198	22:57
Breath Volume	2.375	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.204	23:00
Breath Volume	2.253	23:00
Air Blank	0.000	23:00
Control Test	0.078	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 07/17/2025 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODSTOCK,KENDRA,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: NUNEZ,MARYJOSE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: SAFIULLINA

DATE/TIME OF TEST: 06/21/2025 20:49:07

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:46	
Control Test	0.078	20:47	
Air Blank	0.000	20:47	
Subject Sample	# 1	20:47	Subject Test Refused
Breath Volume	0.000	20:47	
Air Blank	0.000	20:48	
Control Test	0.078	20:48	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

Data Download Date/Time: 07/17/2025 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISNEY,WHITNEY,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SISNEY,WHITNEY
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: WALDROP

DATE/TIME OF TEST: 06/23/2025 22:19:12

Diagnostics Check	OK	22:16	
Air Blank	0.000	22:16	
Control Test	0.078	22:17	
Air Blank	0.000	22:17	
Subject Sample	# 1	22:17	Subject Test Refused
Breath Volume	0.000	22:17	
Air Blank	0.000	22:18	
Control Test	0.079	22:18	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

Data Download Date/Time: 07/17/2025 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUNEZ,JORGE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 06/28/2025 04:35:30

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.242	04:30
Breath Volume	2.261	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.231	04:33
Breath Volume	1.351	04:33
Air Blank	0.000	04:34
Control Test	0.077	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 07/17/2025 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRANDON,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MILLER,BRANDON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/29/2025 03:26:46

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.078	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.150	03:22
Breath Volume	1.332	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.146	03:25
Breath Volume	1.328	03:25
Air Blank	0.000	03:25
Control Test	0.078	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 07/17/2025 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FANJUL,MAXIMO,D
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: FANJUL,MAXIMO
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: MARUSIN

DATE/TIME OF TEST: 06/29/2025 09:16:40

Diagnostics Check	OK	09:10
Air Blank	0.000	09:11
Control Test	0.078	09:11
Air Blank	0.000	09:11
Subject Sample # 1	0.081	09:12
Breath Volume	4.007	09:12
Air Blank	0.000	09:12
Air Blank	0.000	09:14
Subject Sample # 2	0.082	09:15
Breath Volume	2.722	09:15
Air Blank	0.000	09:15
Control Test	0.078	09:16
Air Blank	0.000	09:16
Diagnostics Check	OK	09:16

Data Download Date/Time: 07/17/2025 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ZARATE,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 06/04/2025 00:03:55

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.080	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.203	23:58
Breath Volume	1.800	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.205	00:02
Breath Volume	1.253	00:02
Air Blank	0.000	00:03
Control Test	0.080	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 06/12/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 06/10/2025 23:30:43

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:28	
Control Test	0.080	23:28	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:29	
Control Test	0.079	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 06/12/2025 02:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/12/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: EUSTACHE

DATE/TIME OF TEST: 06/15/2025 04:06:39

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1 0.189	03:56	Volume Not Met
Breath Volume	0.949	03:56	
Air Blank	0.000	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 2 0.200	04:00	
Breath Volume	1.937	04:00	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample	# 3 0.182	04:05	
Breath Volume	1.550	04:05	
Air Blank	0.000	04:05	
Control Test	0.080	04:06	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 06/20/2025 01:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INCER,OSCAR,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MATHIS,E
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:46

SUBJECT'S LAST NAME: KLEE

DATE/TIME OF TEST: 06/03/2025 12:12:43

Diagnostics Check	OK	12:09	
Air Blank	0.000	12:10	
Control Test	0.081	12:10	
Air Blank	0.000	12:11	
Subject Sample	# 1	12:11	Subject Test Refused
Breath Volume	0.000	12:11	
Air Blank	0.000	12:11	
Control Test	0.080	12:12	
Air Blank	0.000	12:12	
Diagnostics Check	OK	12:12	

Data Download Date/Time: 06/05/2025 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAPE,RYAN,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GARCIA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 06/20/2025 20:06:02

Diagnostics Check	OK	19:57
Air Blank	0.000	19:58
Control Test	0.082	19:58
Air Blank	0.000	19:59
Subject Sample # 1	0.303	20:01
Breath Volume	3.207	20:01
Air Blank	0.000	20:02
Air Blank	0.000	20:03
Subject Sample # 2	0.295	20:04
Breath Volume	3.281	20:04
Air Blank	0.000	20:05
Control Test	0.078	20:05
Air Blank	0.000	20:05
Diagnostics Check	OK	20:06

Data Download Date/Time: 07/10/2025 03:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,WILFREDO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GATTORNO,ALEXANDER
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: GARCIA ALVAREZ

DATE/TIME OF TEST: 06/01/2025 07:30:34

Diagnostics Check	OK	07:27	
Air Blank	0.000	07:28	
Control Test	0.081	07:28	
Air Blank	0.000	07:28	
Subject Sample	# 1	07:29	Subject Test Refused
Breath Volume	0.000	07:29	
Air Blank	0.000	07:29	
Control Test	0.082	07:29	
Air Blank	0.000	07:30	
Diagnostics Check	OK	07:30	

Data Download Date/Time: 06/05/2025 02:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTON III, EMANUEL,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: WALLTON II, EMANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/07/2025 20:58:14

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:49	
Control Test	0.081	20:50	
Air Blank	0.000	20:50	
Subject Sample	# 1	20:53	No Sample Provided
Breath Volume	0.000	20:53	
Air Blank	0.000	20:54	
Air Blank	0.000	20:56	
Subject Sample	# 2	20:56	Subject Test Refused
Breath Volume	0.000	20:56	
Air Blank	0.000	20:57	
Control Test	0.081	20:57	
Air Blank	0.000	20:58	
Diagnostics Check	OK	20:58	

Data Download Date/Time: 07/10/2025 03:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ORTEGA,WILLFREDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: SAULEDA

DATE/TIME OF TEST: 06/12/2025 09:04:12

Diagnostics Check	OK	09:01	
Air Blank	0.000	09:01	
Control Test	0.082	09:02	
Air Blank	0.000	09:02	
Subject Sample	# 1	09:02	Subject Test Refused
Breath Volume	0.000	09:02	
Air Blank	0.000	09:03	
Control Test	0.082	09:03	
Air Blank	0.000	09:04	
Diagnostics Check	OK	09:04	

Data Download Date/Time: 07/10/2025 03:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEANDRE,GERALD,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LEANDRE,GERALD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: JEUDI

DATE/TIME OF TEST: 06/16/2025 02:25:01

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:22	
Control Test	0.082	02:22	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:23	Subject Test Refused
Breath Volume	0.000	02:23	
Air Blank	0.000	02:24	
Control Test	0.082	02:24	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 07/10/2025 03:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ-LOPEZ,RICHARD,A
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: GOMEZ-LOPEZ,RICHARD
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 06/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 06/10/2025 03:37:16

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.206	03:32
Breath Volume	1.757	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.213	03:35
Breath Volume	1.882	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/11/2025 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ADSIT,OFFICER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: ROBLEDO

DATE/TIME OF TEST: 06/26/2025 20:32:50

Diagnostics Check	OK	20:25
Air Blank	0.000	20:26
Control Test	0.079	20:26
Air Blank	0.000	20:27
Subject Sample # 1	0.130	20:28
Breath Volume	1.636	20:28
Air Blank	0.000	20:29
Air Blank	0.000	20:30
Subject Sample # 2	0.123	20:31
Breath Volume	1.156	20:31
Air Blank	0.000	20:31
Control Test	0.079	20:32
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 07/09/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: SEFTON

DATE/TIME OF TEST: 06/27/2025 04:20:46

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:18	
Control Test	0.079	04:18	
Air Blank	0.000	04:19	
Subject Sample	# 1	04:19	RFI Detect
Breath Volume	0.000	04:19	
Air Blank	0.000	04:19	
Control Test	0.078	04:20	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 07/09/2025 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MURSCH,THEODORE,R

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: BROOKS,JOSEPH

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 04/24/2023

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 15421080A1

DRY GAS EXPIRY DATE: 06/16/2025

VIOLATION CODE: BUI

ARREST TIME: 16:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: BREAUX

DATE/TIME OF TEST: 06/16/2025 16:42:34

Diagnostics Check	OK	16:39	
Air Blank	0.000	16:40	
Control Test	0.081	16:40	
Air Blank	0.000	16:40	
Subject Sample	# 1	16:41	Subject Test Refused
Breath Volume	0.000	16:41	
Air Blank	0.000	16:41	
Control Test	0.083	16:41	
Air Blank	0.000	16:42	
Diagnostics Check	OK	16:42	

Data Download Date/Time: 05/15/2023 16:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,DANIEL,S

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: WILLIAMS,DANIEL

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403207508

DRY GAS EXPIRY DATE: 12/02/2027

VIOLATION CODE: DUI

ARREST TIME: 06:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: HILLMAN

DATE/TIME OF TEST: 06/05/2025 07:31:37

Diagnostics Check	OK	07:24
Air Blank	0.000	07:25
Control Test	0.081	07:25
Air Blank	0.000	07:26
Subject Sample # 1	0.249	07:26
Breath Volume	1.640	07:26
Air Blank	0.000	07:27
Air Blank	0.000	07:29
Subject Sample # 2	0.251	07:30
Breath Volume	2.500	07:30
Air Blank	0.000	07:30
Control Test	0.079	07:31
Air Blank	0.000	07:31
Diagnostics Check	OK	07:31

Data Download Date/Time: 06/25/2025 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 403207508
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S	DRY GAS EXPIRY DATE: 12/02/2027
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: WILLIAMS,DANIEL	ARREST TIME: 03:04
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 05/29/2025	OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: LENAHAN	DATE/TIME OF TEST: 06/14/2025 03:48:26
------------------------------	--

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.082	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.171	03:43
Breath Volume	3.406	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.170	03:46
Breath Volume	2.710	03:46
Air Blank	0.000	03:47
Control Test	0.082	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 06/25/2025 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 403207508
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S	DRY GAS EXPIRY DATE: 12/02/2027
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: WILLIAMS,DANIEL	ARREST TIME: 19:55
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Female
LAST AGENCY INSPECTION: 05/29/2025	OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: BORITZKI	DATE/TIME OF TEST: 06/14/2025 20:25:47
-------------------------------	--

Diagnostics Check	OK	20:17	
Air Blank	0.000	20:17	
Control Test	0.082	20:17	
Air Blank	0.000	20:18	
Subject Sample	# 1	20:21	No Sample Provided
Breath Volume	0.000	20:21	
Air Blank	0.000	20:21	
Air Blank	0.000	20:23	
Subject Sample	# 2	20:24	Subject Test Refused
Breath Volume	0.000	20:24	
Air Blank	0.000	20:24	
Control Test	0.082	20:25	
Air Blank	0.000	20:25	
Diagnostics Check	OK	20:25	

Data Download Date/Time: 06/25/2025 17:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REEVES,JAMES,F

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: REEVES,JAMES

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403207508

DRY GAS EXPIRY DATE: 06/27/2026

VIOLATION CODE: BUI

ARREST TIME: 17:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: SPAFFORD

DATE/TIME OF TEST: 06/16/2025 18:00:35

Diagnostics Check	OK	17:56	
Air Blank	0.000	17:57	
Control Test	0.080	17:57	
Air Blank	0.000	17:58	
Subject Sample	# 1	17:59	Subject Test Refused
Breath Volume	0.000	17:59	
Air Blank	0.000	17:59	
Control Test	0.082	17:59	
Air Blank	0.000	18:00	
Diagnostics Check	OK	18:00	

Data Download Date/Time: 06/25/2025 17:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REEVES,JAMES,F

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: REEVES,JAMES

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403207508

DRY GAS EXPIRY DATE: 06/27/2026

VIOLATION CODE: BUI

ARREST TIME: 18:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: IBRAHIM

DATE/TIME OF TEST: 06/20/2025 18:42:07

Diagnostics Check	OK	18:36
Air Blank	0.000	18:36
Control Test	0.081	18:36
Air Blank	0.000	18:37
Subject Sample	# 1 0.158	18:37
Breath Volume	1.906	18:37
Air Blank	0.000	18:38
Air Blank	0.000	18:40
Subject Sample	# 2 0.164	18:40
Breath Volume	3.214	18:40
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

Data Download Date/Time: 06/25/2025 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITCHARD, THOMAS, D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: PRITCHARD, THOMAS
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: LEFEBVRE

DATE/TIME OF TEST: 06/21/2025 22:42:19

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:38	
Control Test	0.080	22:39	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:40	Subject Test Refused
Breath Volume	0.000	22:40	
Air Blank	0.000	22:41	
Control Test	0.081	22:41	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:42	

Data Download Date/Time: 06/25/2025 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 403207508
BREATH TEST OPERATOR: WARF,KYLE,J	DRY GAS EXPIRY DATE: 06/27/2026
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: WARF,KYLE	ARREST TIME: 00:25
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 06/25/2025	OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: KENDRICK	DATE/TIME OF TEST: 06/26/2025 01:15:25
-------------------------------	--

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.082	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.193	01:11
Breath Volume	2.390	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.203	01:13
Breath Volume	1.156	01:13
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 07/09/2025 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 403207508
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S	DRY GAS EXPIRY DATE: 12/02/2027
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: WILLIAMS,DANIEL	ARREST TIME: 02:53
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Female
LAST AGENCY INSPECTION: 06/25/2025	OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ANES RIVERA	DATE/TIME OF TEST: 06/29/2025 03:36:19
----------------------------------	--

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.143	03:30
Breath Volume	1.929	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.139	03:34
Breath Volume	1.875	03:34
Air Blank	0.000	03:35
Control Test	0.082	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 07/09/2025 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULASKI,DAMON,J
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PADGETT,ORIE
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/22/2025 01:21:58

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.078	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.111	01:17
Breath Volume	1.609	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.113	01:20
Breath Volume	1.167	01:20
Air Blank	0.000	01:21
Control Test	0.077	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 07/19/2025 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRARO,PAUL,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: FERRARO,PAUL
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: ALL

DATE/TIME OF TEST: 06/01/2025 20:27:38

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.084	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.238	20:22
Breath Volume	2.769	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.229	20:26
Breath Volume	3.195	20:26
Air Blank	0.000	20:26
Control Test	0.083	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON, TORI, D
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WOOD, R
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: ALTAMIRANO

DATE/TIME OF TEST: 06/02/2025 17:51:44

Diagnostics Check	OK	17:50	
Air Blank	0.000	17:50	
Control Test	0.085	17:51	
Air Blank		17:51	RFI Detect
Air Blank	0.000	17:51	

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,TORL,D
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WOOD,R
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: ALTAMIRANO

DATE/TIME OF TEST: 06/02/2025 17:55:07

Diagnostics Check	OK	17:53	
Air Blank	0.000	17:54	
Control Test	0.085	17:54	
Air Blank		17:54	RFI Detect
Air Blank	0.000	17:55	

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,ROHAN,G
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SAINT JEAN,STALENS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: PARISE

DATE/TIME OF TEST: 06/02/2025 23:33:52

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.084	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.177	23:29
Breath Volume	3.699	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.175	23:32
Breath Volume	3.593	23:32
Air Blank	0.000	23:32
Control Test	0.084	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRAGO,JAMES,A
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: SERRAGO,JAMES
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: TANGEL

DATE/TIME OF TEST: 06/07/2025 11:46:28

Diagnostics Check	OK	11:39
Air Blank	0.000	11:40
Control Test	0.085	11:40
Air Blank	0.000	11:40
Subject Sample # 1	0.066	11:41
Breath Volume	1.320	11:41
Air Blank	0.000	11:42
Air Blank	0.000	11:44
Subject Sample # 2	0.062	11:44
Breath Volume	1.691	11:44
Air Blank	0.000	11:45
Control Test	0.084	11:45
Air Blank	0.000	11:46
Diagnostics Check	OK	11:46

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES DE OCA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MATERA,DEREK
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: WETZEL

DATE/TIME OF TEST: 06/08/2025 19:21:15

Diagnostics Check	OK	19:14
Air Blank	0.000	19:15
Control Test	0.084	19:15
Air Blank	0.000	19:16
Subject Sample # 1	0.269	19:16
Breath Volume	3.855	19:16
Air Blank	0.000	19:17
Air Blank	0.000	19:19
Subject Sample # 2	0.272	19:19
Breath Volume	3.660	19:19
Air Blank	0.000	19:20
Control Test	0.085	19:20
Air Blank	0.000	19:21
Diagnostics Check	OK	19:21

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,ROHAN,G
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: GODINEZ,OSVALDO
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: PASTOR SARAT

DATE/TIME OF TEST: 06/10/2025 02:56:06

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.083	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.204	02:50
Breath Volume	1.824	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.206	02:54
Breath Volume	2.246	02:54
Air Blank	0.000	02:55
Control Test	0.084	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZUNIGA,S,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: ZUNIGA,S
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: SALVADOR LUIS

DATE/TIME OF TEST: 06/15/2025 02:44:13

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.084	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.163	02:39
Breath Volume	2.246	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.160	02:42
Breath Volume	2.601	02:42
Air Blank	0.000	02:43
Control Test	0.085	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MALLONE,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: GENDRON

DATE/TIME OF TEST: 06/15/2025 03:35:15

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.083	03:31	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.031	03:33	
Air Blank	0.000	03:34	
Control Test	0.084	03:34	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: FORSE

DATE/TIME OF TEST: 06/15/2025 03:47:40

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.082	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.149	03:43
Breath Volume	1.441	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.154	03:46
Breath Volume	2.425	03:46
Air Blank	0.000	03:46
Control Test	0.083	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FANEGO, DAVID,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FANEGO, DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: GARCIA OSORIO

DATE/TIME OF TEST: 06/23/2025 04:40:50

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.084	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.109	04:36
Breath Volume	2.601	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.104	04:39
Breath Volume	1.664	04:39
Air Blank	0.000	04:39
Control Test	0.084	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 07/22/2025 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES DE OCA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MONTES DE OCA,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:54

SUBJECT'S LAST NAME: DELEON

DATE/TIME OF TEST: 06/25/2025 15:24:30

Diagnostics Check	OK	15:17
Air Blank	0.000	15:18
Control Test	0.083	15:18
Air Blank	0.000	15:18
Subject Sample # 1	0.275	15:20
Breath Volume	2.539	15:20
Air Blank	0.000	15:20
Air Blank	0.000	15:22
Subject Sample # 2	0.272	15:22
Breath Volume	2.750	15:22
Air Blank	0.000	15:23
Control Test	0.082	15:23
Air Blank	0.000	15:24
Diagnostics Check	OK	15:24

Data Download Date/Time: 07/22/2025 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,JOSHUA,E
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: NORRIS,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: WHITACRE

DATE/TIME OF TEST: 06/27/2025 12:02:28

Diagnostics Check	OK	11:56
Air Blank	0.000	11:56
Control Test	0.085	11:56
Air Blank	0.000	11:57
Subject Sample # 1	0.000	11:58
Breath Volume	1.386	11:58
Air Blank	0.000	11:58
Air Blank	0.000	12:00
Subject Sample # 2	0.000	12:01
Breath Volume	1.828	12:01
Air Blank	0.000	12:01
Control Test	0.084	12:01
Air Blank	0.000	12:02
Diagnostics Check	OK	12:02

Data Download Date/Time: 07/22/2025 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: LARA

DATE/TIME OF TEST: 06/28/2025 03:27:39

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.082	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.069	03:23
Breath Volume	2.343	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.067	03:26
Breath Volume	2.355	03:26
Air Blank	0.000	03:26
Control Test	0.084	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 07/22/2025 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PEREZ MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: RODRIGUEZ MARTINEZ

DATE/TIME OF TEST: 06/05/2025 02:16:54

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.179	02:12
Breath Volume	1.656	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.174	02:15
Breath Volume	1.726	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 06/06/2025 04:38:13

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.000	04:34
Breath Volume	1.617	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.000	04:36
Breath Volume	2.058	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: MARTINEZ PULGAR

DATE/TIME OF TEST: 06/07/2025 06:28:27

Diagnostics Check	OK	06:21
Air Blank	0.000	06:21
Control Test	0.080	06:21
Air Blank	0.000	06:22
Subject Sample # 1	0.151	06:23
Breath Volume	1.296	06:23
Air Blank	0.000	06:24
Air Blank	0.000	06:26
Subject Sample # 2	0.164	06:26
Breath Volume	1.464	06:26
Air Blank	0.000	06:27
Control Test	0.079	06:27
Air Blank	0.000	06:28
Diagnostics Check	OK	06:28

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ROJASCABELLO

DATE/TIME OF TEST: 06/11/2025 01:48:07

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.194	01:43
Breath Volume	3.402	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.196	01:46
Breath Volume	3.910	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MILLANDSANTIAGO

DATE/TIME OF TEST: 06/13/2025 02:37:37

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.145	02:32
Breath Volume	1.601	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.151	02:36
Breath Volume	1.632	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/13/2025 04:37:52

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.051	04:33
Breath Volume	2.371	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.051	04:36
Breath Volume	1.574	04:36
Air Blank	0.000	04:36
Control Test	0.079	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PERAZARECINOS

DATE/TIME OF TEST: 06/14/2025 03:04:42

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.226	02:59
Breath Volume	1.101	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:01
Subject Sample # 2	0.211	03:02
Breath Volume	2.820	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: SHAFFER

DATE/TIME OF TEST: 06/15/2025 03:31:03

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.134	03:24
Breath Volume	1.160	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.139	03:29
Breath Volume	2.136	03:29
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:31

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: CONWAY

DATE/TIME OF TEST: 06/19/2025 02:45:21

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:35	
Control Test	0.080	02:35	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:39	No Sample Provided
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Air Blank	0.000	02:41	
Subject Sample	# 2	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Control Test	0.081	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: ESCALONA,JUSTIN
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: HENRIQUEZ OCHOA

DATE/TIME OF TEST: 06/21/2025 05:06:39

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.080	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.151	05:02
Breath Volume	2.984	05:02
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.154	05:05
Breath Volume	3.621	05:05
Air Blank	0.000	05:05
Control Test	0.081	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 06/21/2025 05:20:14

Diagnostics Check	OK	05:13
Air Blank	0.000	05:13
Control Test	0.079	05:14
Air Blank	0.000	05:14
Subject Sample # 1	0.183	05:15
Breath Volume	2.394	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:17
Subject Sample # 2	0.189	05:18
Breath Volume	2.023	05:18
Air Blank	0.000	05:19
Control Test	0.080	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENNING,ASHLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HENNING,ASHLEY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: SACKS

DATE/TIME OF TEST: 06/21/2025 18:14:32

Diagnostics Check	OK	17:58	
Air Blank	0.000	17:58	
Control Test	0.080	17:59	
Air Blank	0.000	17:59	
Subject Sample	# 1 0.252	18:01	
Breath Volume	1.371	18:01	
Air Blank	0.000	18:02	
Air Blank	0.000	18:03	
Subject Sample	# 2 0.213	18:06	Volume Not Met
Breath Volume	0.460	18:06	
Air Blank	0.000	18:07	
Air Blank	0.000	18:09	
Subject Sample	# 3 0.238	18:12	Volume Not Met
Breath Volume	0.171	18:12	
Air Blank	0.000	18:13	
Control Test	0.076	18:13	
Air Blank	0.000	18:14	
Diagnostics Check	OK	18:14	

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENNING,ASHLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HENNING,ASHLEY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: SACKS

DATE/TIME OF TEST: 06/21/2025 18:29:59

Diagnostics Check	OK	18:19	
Air Blank	0.000	18:20	
Control Test	0.078	18:20	
Air Blank	0.000	18:21	
Subject Sample	# 1 0.239	18:24	Volume Not Met
Breath Volume	0.066	18:24	
Air Blank	0.000	18:25	
Air Blank	0.000	18:26	
Subject Sample	# 2	18:28	Subject Test Refused
Breath Volume	0.226	18:28	
Air Blank	0.000	18:29	
Control Test	0.075	18:29	
Air Blank	0.000	18:29	
Diagnostics Check	OK	18:29	

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SENA

DATE/TIME OF TEST: 06/21/2025 23:27:01

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.079	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.265	23:22
Breath Volume	2.921	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:24
Subject Sample # 2	0.254	23:25
Breath Volume	1.265	23:25
Air Blank	0.000	23:26
Control Test	0.079	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:27

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 06/22/2025 02:27:11

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.148	02:22
Breath Volume	2.519	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.150	02:25
Breath Volume	1.570	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULVEY,KEVIN,S
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MULVEY,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: BARAHONA

DATE/TIME OF TEST: 06/23/2025 04:21:05

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:19	
Control Test	0.081	04:19	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:20	Sequence Aborted
Breath Volume	0.000	04:20	
Air Blank	0.000	04:21	

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSALES,ANIBAL,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: SUAREZ MORALES

DATE/TIME OF TEST: 06/23/2025 04:31:36

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.245	04:27
Breath Volume	2.316	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.251	04:30
Breath Volume	2.281	04:30
Air Blank	0.000	04:30
Control Test	0.080	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: FONTE GARCIA

DATE/TIME OF TEST: 06/28/2025 01:56:28

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.090	01:51
Breath Volume	2.128	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.086	01:54
Breath Volume	2.742	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: LUKES

DATE/TIME OF TEST: 06/29/2025 03:30:42

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.081	03:18	
Air Blank	0.000	03:18	
Subject Sample	# 1 0.168	03:21	Volume Not Met
Breath Volume	0.070	03:21	
Air Blank	0.000	03:22	
Air Blank	0.000	03:24	
Subject Sample	# 2 0.170	03:25	
Breath Volume	1.644	03:25	
Air Blank	0.000	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 3 0.161	03:28	
Breath Volume	1.375	03:28	
Air Blank	0.000	03:29	
Control Test	0.077	03:30	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: MEDINILLA

DATE/TIME OF TEST: 06/29/2025 06:00:55

Diagnostics Check	OK	05:57	
Air Blank	0.000	05:58	
Control Test	0.081	05:58	
Air Blank	0.000	05:59	
Subject Sample	# 1	05:59	Subject Test Refused
Breath Volume	0.000	05:59	
Air Blank	0.000	06:00	
Control Test	0.080	06:00	
Air Blank	0.000	06:00	
Diagnostics Check	OK	06:00	

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: RODRIGUEZ HERRERA

DATE/TIME OF TEST: 06/30/2025 03:34:33

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.081	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 07/22/2025 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: DELEON

DATE/TIME OF TEST: 06/03/2025 19:13:30

Diagnostics Check	OK	19:07
Air Blank	0.000	19:07
Control Test	0.083	19:08
Air Blank	0.000	19:08
Subject Sample # 1	0.000	19:09
Breath Volume	3.964	19:09
Air Blank	0.000	19:09
Air Blank	0.000	19:11
Subject Sample # 2	0.000	19:12
Breath Volume	3.671	19:12
Air Blank	0.000	19:12
Control Test	0.082	19:12
Air Blank	0.000	19:13
Diagnostics Check	OK	19:13

Data Download Date/Time: 06/06/2025 02:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MIKESH

DATE/TIME OF TEST: 06/12/2025 00:33:15

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:29	
Control Test	0.083	00:29	
Air Blank	0.000	00:30	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:32	
Control Test	0.081	00:32	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: GARRIDO BRITO

DATE/TIME OF TEST: 06/12/2025 03:35:31

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.082	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.179	03:30
Breath Volume	2.253	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.178	03:33
Breath Volume	1.230	03:33
Air Blank	0.000	03:34
Control Test	0.081	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BRIZENDINE

DATE/TIME OF TEST: 06/13/2025 23:58:49

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.083	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.077	23:54
Breath Volume	3.406	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.077	23:57
Breath Volume	3.753	23:57
Air Blank	0.000	23:57
Control Test	0.082	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 06/14/2025 17:35:11

Diagnostics Check	OK	17:28
Air Blank	0.000	17:29
Control Test	0.082	17:29
Air Blank	0.000	17:30
Subject Sample # 1	0.103	17:30
Breath Volume	3.011	17:30
Air Blank	0.000	17:31
Air Blank	0.000	17:33
Subject Sample # 2	0.100	17:33
Breath Volume	2.457	17:33
Air Blank	0.000	17:34
Control Test	0.081	17:34
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: STRYBEL, B
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: VILLANUEVA CASTRO

DATE/TIME OF TEST: 06/19/2025 00:51:40

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:48	
Control Test	0.082	00:49	
Air Blank	0.000	00:49	
Subject Sample	# 1	00:50	Slope Not Met
Breath Volume	1.683	00:50	
Air Blank	0.000	00:50	
Control Test	0.081	00:51	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: STRYBEL, B
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: VILLANUEVA CASTRO

DATE/TIME OF TEST: 06/19/2025 01:18:49

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.081	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.150	01:13
Breath Volume	1.949	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.145	01:17
Breath Volume	1.597	01:17
Air Blank	0.000	01:17
Control Test	0.083	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 06/19/2025 02:30:03

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.220	02:25
Breath Volume	2.445	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:27
Subject Sample # 2	0.216	02:28
Breath Volume	1.183	02:28
Air Blank	0.000	02:29
Control Test	0.082	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:30

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINTER,ERIC,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: VEGA,JUAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CHAMORRO

DATE/TIME OF TEST: 06/22/2025 02:29:28

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.082	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.142	02:24
Breath Volume	2.390	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.132	02:27
Breath Volume	1.921	02:27
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANKOSKY,D,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: STANKOSKY,D
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: BOSWORTH

DATE/TIME OF TEST: 06/23/2025 03:05:52

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.082	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.181	03:00
Breath Volume	2.515	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.181	03:04
Breath Volume	1.492	03:04
Air Blank	0.000	03:04
Control Test	0.081	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINTER,ERIC,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,CHRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: AGUILAR ANDRES

DATE/TIME OF TEST: 06/26/2025 02:38:12

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.082	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.169	02:33
Breath Volume	1.152	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.166	02:36
Breath Volume	1.132	02:36
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRITZ, TYLER], S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: WHEELER, ZACHARY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:07

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 06/27/2025 08:34:04

Diagnostics Check	OK	08:27
Air Blank	0.000	08:28
Control Test	0.082	08:28
Air Blank	0.000	08:29
Subject Sample # 1	0.135	08:29
Breath Volume	1.976	08:29
Air Blank	0.000	08:30
Air Blank	0.000	08:31
Subject Sample # 2	0.131	08:32
Breath Volume	3.320	08:32
Air Blank	0.000	08:33
Control Test	0.081	08:33
Air Blank	0.000	08:33
Diagnostics Check	OK	08:34

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: KERCE

DATE/TIME OF TEST: 06/27/2025 19:36:32

Diagnostics Check	OK	19:30
Air Blank	0.000	19:30
Control Test	0.082	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.000	19:32
Breath Volume	3.273	19:32
Air Blank	0.000	19:32
Air Blank	0.000	19:34
Subject Sample # 2	0.000	19:35
Breath Volume	3.140	19:35
Air Blank	0.000	19:35
Control Test	0.082	19:35
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNIS,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: STRYBEL,BRIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: FALOCCO

DATE/TIME OF TEST: 06/28/2025 01:40:07

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.082	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.214	01:35
Breath Volume	1.691	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.223	01:38
Breath Volume	1.996	01:38
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ARD,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: WHIDBY

DATE/TIME OF TEST: 06/28/2025 03:02:44

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.082	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.000	02:58
Breath Volume	2.882	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.000	03:01
Breath Volume	2.917	03:01
Air Blank	0.000	03:01
Control Test	0.082	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOULIHAN,RILEY,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: STRYBEL,BRIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: DAMON

DATE/TIME OF TEST: 06/29/2025 03:02:42

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.083	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.000	02:58
Breath Volume	2.671	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.000	03:01
Breath Volume	3.332	03:01
Air Blank	0.000	03:01
Control Test	0.082	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOULIHAN,RILEY,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MORALES,HECTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: LEMONDE-COLEMAN

DATE/TIME OF TEST: 06/29/2025 03:37:27

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.082	03:34	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:36	RFI Detect
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOULIHAN,RILEY,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MORALES,HECTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: LEMONDE-COLEMAN

DATE/TIME OF TEST: 06/29/2025 03:48:16

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.082	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.131	03:43
Breath Volume	1.894	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.125	03:46
Breath Volume	1.750	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLOUGHEN,KORY,I
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MORALES,HECTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: OROZCO-SALGUERO

DATE/TIME OF TEST: 06/30/2025 00:22:19

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.082	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.218	00:17
Breath Volume	2.117	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.213	00:20
Breath Volume	3.082	00:20
Air Blank	0.000	00:21
Control Test	0.082	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 07/09/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,MARCUS,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ENRRIQUEZ,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: JUANAGAPITO

DATE/TIME OF TEST: 06/15/2025 23:28:28

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.081	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.184	23:23
Breath Volume	1.824	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.189	23:26
Breath Volume	1.695	23:26
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 06/28/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,GATLIN,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MARTINEZ,VERONICA
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: CORETOMARTINEZ

DATE/TIME OF TEST: 06/21/2025 04:39:18

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.148	04:34
Breath Volume	3.027	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.141	04:37
Breath Volume	2.667	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 06/28/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTER,LANNY,C
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: RESTER,LANNY
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 06/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 06/30/2025 23:55:21

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:52	
Control Test	0.080	23:52	
Air Blank	0.000	23:52	
Subject Sample	# 1 0.232	23:53	
Breath Volume	1.578	23:53	
Air Blank		23:54	Purge Fail
Air Blank	0.000	23:55	

Data Download Date/Time: 07/24/2025 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MARKHAM,JAMARI
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 06/01/2025 05:59:29

Diagnostics Check	OK	05:58	
Air Blank		05:59	RFI Detect
Air Blank	0.000	05:59	

Data Download Date/Time: 06/10/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MARKHAM,JAMARI
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 06/01/2025 06:14:21

Diagnostics Check	OK	06:02	
Air Blank	0.000	06:03	
Control Test	0.080	06:03	
Air Blank	0.000	06:04	
Subject Sample	# 1	06:07	No Sample Provided
Breath Volume	0.000	06:07	
Air Blank	0.000	06:07	
Air Blank	0.000	06:09	
Subject Sample	# 2	06:12	No Sample Provided
Breath Volume	0.000	06:12	
Air Blank	0.000	06:13	
Control Test	0.081	06:13	
Air Blank	0.000	06:14	
Diagnostics Check	OK	06:14	

Data Download Date/Time: 06/10/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OWENS,JONATHAN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 06/07/2025 04:45:33

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.148	04:40
Breath Volume	2.769	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.152	04:44
Breath Volume	2.648	04:44
Air Blank	0.000	04:44
Control Test	0.081	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 06/10/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: SMITH,ALEXANDER
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/11/2025 03:25:06

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.082	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.169	03:20
Breath Volume	1.875	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:22
Subject Sample # 2	0.171	03:23
Breath Volume	2.277	03:23
Air Blank	0.000	03:24
Control Test	0.082	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 07/14/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,NATHAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MARTINA,RYAN
ARRESTING OFFICER AGENCY: FHP Headquarters
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:52

SUBJECT'S LAST NAME: GALVEZ

DATE/TIME OF TEST: 06/14/2025 14:38:42

Diagnostics Check	OK	14:32
Air Blank	0.000	14:32
Control Test	0.081	14:32
Air Blank	0.000	14:33
Subject Sample # 1	0.067	14:34
Breath Volume	3.773	14:34
Air Blank	0.000	14:34
Air Blank	0.000	14:36
Subject Sample # 2	0.066	14:37
Breath Volume	3.359	14:37
Air Blank	0.000	14:37
Control Test	0.081	14:38
Air Blank	0.000	14:38
Diagnostics Check	OK	14:38

Data Download Date/Time: 07/14/2025 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERSON,CAMERON,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ROGERSON,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 06/15/2025 05:29:25

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:27	
Control Test	0.081	05:27	
Air Blank	0.000	05:27	
Subject Sample	# 1	05:28	Subject Test Refused
Breath Volume	0.000	05:28	
Air Blank	0.000	05:28	
Control Test	0.081	05:28	
Air Blank	0.000	05:29	
Diagnostics Check	OK	05:29	

Data Download Date/Time: 07/14/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,THOMAS,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: YOUNG,THOMAS
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/16/2025 00:16:36

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.080	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.202	00:11
Breath Volume	1.722	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.186	00:15
Breath Volume	1.410	00:15
Air Blank	0.000	00:15
Control Test	0.081	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 07/14/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: SMITH,ALEXANDER
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: GILBERTO

DATE/TIME OF TEST: 06/16/2025 01:10:14

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.082	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.174	01:05
Breath Volume	3.925	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.173	01:08
Breath Volume	4.214	01:08
Air Blank	0.000	01:09
Control Test	0.081	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 07/14/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OWENS,JOHN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: GHIAIE

DATE/TIME OF TEST: 06/17/2025 23:59:57

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.082	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.236	23:54
Breath Volume	1.804	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.218	23:58
Breath Volume	1.101	23:58
Air Blank	0.000	23:59
Control Test	0.081	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 07/14/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BIDDLE,MICAH
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/18/2025 03:35:52

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.082	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.211	03:31
Breath Volume	1.445	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.209	03:34
Breath Volume	1.402	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 07/14/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,THOMAS,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: YOUNG,THOMAS
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MCCLLOUD

DATE/TIME OF TEST: 06/21/2025 02:11:19

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.082	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.026	02:06
Breath Volume	1.648	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.026	02:09
Breath Volume	1.898	02:09
Air Blank	0.000	02:10
Control Test	0.082	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 07/14/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELABERT,NOAH,B
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: GERGEN,GERGEN
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: GELABERT

DATE/TIME OF TEST: 06/21/2025 04:55:17

Diagnostics Check	OK	04:54	
Air Blank		04:55	RFI Detect
Air Blank		04:55	RFI Detect

Data Download Date/Time: 07/14/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,DAMION,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MORALES,DAMION
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ALBEURQUERQUE YEOVA

DATE/TIME OF TEST: 06/22/2025 03:35:42

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.081	03:21	
Air Blank	0.000	03:21	
Subject Sample	# 1 0.193	03:23	
Breath Volume	2.277	03:23	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample	# 2 0.177	03:29	Volume Not Met
Breath Volume	0.031	03:29	
Air Blank	0.000	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 3 0.175	03:34	
Breath Volume	1.500	03:34	
Air Blank	0.000	03:34	
Control Test	0.079	03:35	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 07/14/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,THOMAS,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: THAMES,ADRIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 06/22/2025 05:08:33

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.081	05:03
Air Blank	0.000	05:03
Subject Sample # 1	0.058	05:03
Breath Volume	1.742	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.062	05:07
Breath Volume	1.640	05:07
Air Blank	0.000	05:07
Control Test	0.081	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 07/14/2025 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OWENS,JOHN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/24/2025 03:38:01

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.248	03:30
Breath Volume	1.273	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.227	03:33
Breath Volume	2.070	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 3	0.239	03:36
Breath Volume	2.250	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:38

Data Download Date/Time: 07/14/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HALL,NOAH
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/27/2025 04:20:29

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.081	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.180	04:15
Breath Volume	1.972	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.181	04:18
Breath Volume	1.316	04:18
Air Blank	0.000	04:19
Control Test	0.082	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 07/14/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,R,R
BREATH TEST OPERATOR AGENCY: Midway PD
ARRESTING OFFICER: CRUZ,R
ARRESTING OFFICER AGENCY: Midway PD
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/28/2025 22:44:45

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.131	22:40
Breath Volume	1.664	22:40
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.142	22:43
Breath Volume	1.531	22:43
Air Blank	0.000	22:43
Control Test	0.078	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 07/17/2025 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARCIA,ROGERS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WESTBROOK

DATE/TIME OF TEST: 06/29/2025 01:43:02

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.188	01:38
Breath Volume	1.261	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.190	01:41
Breath Volume	2.304	01:41
Air Blank	0.000	01:42
Control Test	0.076	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:43

Data Download Date/Time: 07/16/2025 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCMORROW,RYAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: STEPHENSON

DATE/TIME OF TEST: 06/01/2025 02:17:42

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.080	02:15	
Air Blank	0.000	02:15	
Subject Sample	# 1	02:16	Subject Test Refused
Breath Volume	0.000	02:16	
Air Blank	0.000	02:16	
Control Test	0.080	02:17	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CONTE,COREY
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: RAJU

DATE/TIME OF TEST: 06/08/2025 02:21:38

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.165	02:17
Breath Volume	3.890	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.150	02:20
Breath Volume	2.113	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 06/25/2025 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO,EDWIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/08/2025 03:23:29

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.133	03:19
Breath Volume	3.531	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.131	03:22
Breath Volume	4.039	03:22
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 06/25/2025 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HOEBBEL

DATE/TIME OF TEST: 06/14/2025 02:49:52

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.243	02:42
Breath Volume	1.652	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:44
Subject Sample # 2	0.270	02:45
Breath Volume	1.230	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample # 3	0.262	02:48
Breath Volume	2.527	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 06/25/2025 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ESPAILLAT,LIANA
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ROBLEROGONZALEZ

DATE/TIME OF TEST: 06/21/2025 00:52:25

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.244	00:48
Breath Volume	1.945	00:48
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.247	00:50
Breath Volume	2.640	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 06/25/2025 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONS,DEPUTY,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GERARDI,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: VIVERETTE

DATE/TIME OF TEST: 06/06/2025 03:02:42

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.077	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.167	02:58
Breath Volume	2.226	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.164	03:01
Breath Volume	2.292	03:01
Air Blank	0.000	03:01
Control Test	0.078	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONNEMANN, MARK, W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HOWARD, J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: BALSAMELLO

DATE/TIME OF TEST: 06/07/2025 02:02:53

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.077	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.129	01:58
Breath Volume	1.941	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.131	02:01
Breath Volume	1.851	02:01
Air Blank	0.000	02:01
Control Test	0.077	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/10/2025 00:35:51

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.077	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.000	00:31
Breath Volume	2.898	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.000	00:34
Breath Volume	3.222	00:34
Air Blank	0.000	00:34
Control Test	0.078	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIVECOAT,JUSTIN,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HOWARD,JESSIE
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 05:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: MCCONN

DATE/TIME OF TEST: 06/10/2025 07:03:39

Diagnostics Check	OK	06:57
Air Blank	0.000	06:57
Control Test	0.076	06:58
Air Blank	0.000	06:58
Subject Sample # 1	0.120	06:59
Breath Volume	1.394	06:59
Air Blank	0.000	06:59
Air Blank	0.000	07:01
Subject Sample # 2	0.115	07:02
Breath Volume	1.324	07:02
Air Blank	0.000	07:02
Control Test	0.077	07:03
Air Blank	0.000	07:03
Diagnostics Check	OK	07:03

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: RODRIGUEZ,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 06/13/2025 19:12:26

Diagnostics Check	OK	19:06
Air Blank	0.000	19:06
Control Test	0.076	19:06
Air Blank	0.000	19:07
Subject Sample # 1	0.000	19:08
Breath Volume	2.535	19:08
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 2	0.000	19:11
Breath Volume	2.113	19:11
Air Blank	0.000	19:11
Control Test	0.076	19:11
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONS,DEPUTY,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SYMONS,E
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: DUARTE

DATE/TIME OF TEST: 06/14/2025 02:59:52

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.077	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.168	02:55
Breath Volume	1.640	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.170	02:58
Breath Volume	1.160	02:58
Air Blank	0.000	02:58
Control Test	0.076	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONS,DEPUTY,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SYMONS,E
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: LESTOURGEON

DATE/TIME OF TEST: 06/15/2025 01:26:16

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.077	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 1 0.094	01:17	Volume Not Met
Breath Volume	0.050	01:17	
Air Blank	0.000	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 2 0.120	01:21	
Breath Volume	2.113	01:21	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample	# 3 0.118	01:24	
Breath Volume	1.710	01:24	
Air Blank	0.000	01:25	
Control Test	0.076	01:25	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON, WILLIAM, M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: WATSON, WILLIAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: REINHARDT

DATE/TIME OF TEST: 06/17/2025 20:51:52

Diagnostics Check	OK	20:45
Air Blank	0.000	20:46
Control Test	0.076	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.186	20:47
Breath Volume	3.167	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:49
Subject Sample # 2	0.184	20:50
Breath Volume	3.015	20:50
Air Blank	0.000	20:50
Control Test	0.075	20:51
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,CORPORAL,J
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: BERRIOS,CORPORAL
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 06/23/2025 00:59:54

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.078	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.148	00:54
Breath Volume	2.843	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.149	00:58
Breath Volume	2.550	00:58
Air Blank	0.000	00:58
Control Test	0.077	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONNEMANN,M,W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: KENNY,E
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: VEGAS

DATE/TIME OF TEST: 06/23/2025 02:57:03

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.077	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.135	02:52
Breath Volume	3.128	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.138	02:55
Breath Volume	4.117	02:55
Air Blank	0.000	02:56
Control Test	0.078	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:57

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: LAWTON

DATE/TIME OF TEST: 06/24/2025 00:34:50

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.078	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.000	00:30
Breath Volume	2.500	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.000	00:33
Breath Volume	2.320	00:33
Air Blank	0.000	00:33
Control Test	0.077	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 06/29/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,JORDIN,K
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HALL,JORDIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 06/24/2025 19:27:30

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.077	19:20
Air Blank	0.000	19:21
Subject Sample # 1	0.129	19:22
Breath Volume	1.957	19:22
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.123	19:25
Breath Volume	1.933	19:25
Air Blank	0.000	19:26
Control Test	0.077	19:26
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

Data Download Date/Time: 06/29/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BABCOCK, EDWARD
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: CARMACK

DATE/TIME OF TEST: 06/24/2025 21:21:38

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.076	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.225	21:15
Breath Volume	1.621	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:17
Subject Sample # 2	0.205	21:20
Breath Volume	1.121	21:20
Air Blank	0.000	21:20
Control Test	0.077	21:21
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 06/29/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DRUMM,R
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 06/24/2025 23:21:51

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.077	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.106	23:17
Breath Volume	1.535	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.106	23:20
Breath Volume	2.035	23:20
Air Blank	0.000	23:20
Control Test	0.077	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 06/29/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPANA,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: CAMPANA,NICHOLAS
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 06/26/2025 01:38:32

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.077	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.185	01:32
Breath Volume	1.613	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.181	01:36
Breath Volume	1.207	01:36
Air Blank	0.000	01:37
Control Test	0.077	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 06/29/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LISS,K,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: DRUMM,R
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SANDVIK

DATE/TIME OF TEST: 06/28/2025 01:24:00

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.077	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.297	01:18
Breath Volume	2.125	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.307	01:22
Breath Volume	1.539	01:22
Air Blank	0.000	01:23
Control Test	0.076	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:24

Data Download Date/Time: 06/29/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: ISHIDA,M
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 06/28/2025 04:17:45

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.077	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.140	04:13
Breath Volume	1.585	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.140	04:16
Breath Volume	1.156	04:16
Air Blank	0.000	04:16
Control Test	0.077	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 06/29/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON, WILLIAM, M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: WATSON, WILLIAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/28/2025 19:05:01

Diagnostics Check	OK	18:56
Air Blank	0.000	18:57
Control Test	0.077	18:57
Air Blank	0.000	18:58
Subject Sample # 1	0.181	18:59
Breath Volume	1.242	18:59
Air Blank	0.000	18:59
Air Blank	0.000	19:01
Subject Sample # 2	0.192	19:03
Breath Volume	1.644	19:03
Air Blank	0.000	19:04
Control Test	0.077	19:04
Air Blank	0.000	19:04
Diagnostics Check	OK	19:05

Data Download Date/Time: 06/29/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GARBER

DATE/TIME OF TEST: 06/29/2025 00:42:25

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.077	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.000	00:38
Breath Volume	1.738	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.000	00:41
Breath Volume	1.925	00:41
Air Blank	0.000	00:41
Control Test	0.077	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 06/29/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TRAN,BRYAN
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/02/2025 04:07:38

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.078	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.219	04:02
Breath Volume	1.261	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.215	04:05
Breath Volume	1.464	04:05
Air Blank	0.000	04:06
Control Test	0.078	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWEENEY,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SWEENEY,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 14:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:46

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 06/02/2025 15:32:58

Diagnostics Check	OK	15:26
Air Blank	0.000	15:27
Control Test	0.079	15:27
Air Blank	0.000	15:27
Subject Sample # 1	0.000	15:28
Breath Volume	1.957	15:28
Air Blank	0.000	15:29
Air Blank	0.000	15:31
Subject Sample # 2	0.000	15:31
Breath Volume	2.511	15:31
Air Blank	0.000	15:32
Control Test	0.079	15:32
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 06/04/2025 19:51:35

Diagnostics Check	OK	19:44
Air Blank	0.000	19:45
Control Test	0.079	19:45
Air Blank	0.000	19:46
Subject Sample # 1	0.000	19:47
Breath Volume	1.960	19:47
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.000	19:50
Breath Volume	2.039	19:50
Air Blank	0.000	19:50
Control Test	0.081	19:51
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO-PEREZ,MICHAEL,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ROMERO-PEREZ,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 06/06/2025 23:33:52

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.050	23:28
Breath Volume	1.703	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.048	23:32
Breath Volume	2.214	23:32
Air Blank	0.000	23:32
Control Test	0.078	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWEENEY,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEREZROSADO,HECTOR
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:23

SUBJECT'S LAST NAME: ORTIZLEON

DATE/TIME OF TEST: 06/07/2025 06:50:14

Diagnostics Check	OK	06:43
Air Blank	0.000	06:44
Control Test	0.079	06:44
Air Blank	0.000	06:44
Subject Sample # 1	0.103	06:45
Breath Volume	1.742	06:45
Air Blank	0.000	06:46
Air Blank	0.000	06:47
Subject Sample # 2	0.097	06:48
Breath Volume	3.906	06:48
Air Blank	0.000	06:49
Control Test	0.078	06:49
Air Blank	0.000	06:50
Diagnostics Check	OK	06:50

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES, CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: MESTRE, EZEQUIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: ORELLANA

DATE/TIME OF TEST: 06/08/2025 05:01:53

Diagnostics Check	OK	04:54
Air Blank	0.000	04:54
Control Test	0.079	04:55
Air Blank	0.000	04:55
Subject Sample # 1	0.220	04:57
Breath Volume	2.468	04:57
Air Blank	0.000	04:58
Air Blank	0.000	04:59
Subject Sample # 2	0.212	05:00
Breath Volume	1.722	05:00
Air Blank	0.000	05:00
Control Test	0.077	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SUAREZ,JAVIER
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: MAFLA NIEVA

DATE/TIME OF TEST: 06/08/2025 23:26:27

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.079	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.193	23:21
Breath Volume	2.253	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:23
Subject Sample # 2	0.186	23:24
Breath Volume	2.289	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: TEADT

DATE/TIME OF TEST: 06/10/2025 21:42:56

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:38	
Control Test	0.079	21:38	
Air Blank	0.000	21:39	
Subject Sample	# 1	21:41	Subject Test Refused
Breath Volume	0.000	21:41	
Air Blank	0.000	21:42	
Control Test	0.079	21:42	
Air Blank	0.000	21:42	
Diagnostics Check	OK	21:42	

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,DERICK,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOPEZ,DERICK
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: COLLAZO

DATE/TIME OF TEST: 06/12/2025 03:14:32

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.088	03:10
Breath Volume	2.675	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.080	03:13
Breath Volume	2.601	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO-PEREZ,MICHAEL,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCDONALD,KIRK
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: VARGAS AGRONT

DATE/TIME OF TEST: 06/12/2025 23:19:04

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.137	23:14
Breath Volume	1.519	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.139	23:17
Breath Volume	1.281	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:19

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARZYGnat,DESTINY,P
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GANESH,BRANDON
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: BASKETT

DATE/TIME OF TEST: 06/14/2025 06:28:18

Diagnostics Check	OK	06:21
Air Blank	0.000	06:21
Control Test	0.080	06:22
Air Blank	0.000	06:22
Subject Sample # 1	0.137	06:23
Breath Volume	3.527	06:23
Air Blank	0.000	06:24
Air Blank	0.000	06:26
Subject Sample # 2	0.137	06:26
Breath Volume	3.578	06:26
Air Blank	0.000	06:27
Control Test	0.078	06:27
Air Blank	0.000	06:28
Diagnostics Check	OK	06:28

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: COSME SILVA

DATE/TIME OF TEST: 06/19/2025 23:02:51

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.000	22:58
Breath Volume	1.386	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.000	23:01
Breath Volume	1.789	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ROSADO LOPEZ

DATE/TIME OF TEST: 06/20/2025 02:18:40

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.079	02:13
Breath Volume	2.140	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.077	02:17
Breath Volume	1.605	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELLER,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HELLER,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: BATISTAMARQUEZ

DATE/TIME OF TEST: 06/21/2025 04:24:01

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.132	04:19
Breath Volume	2.019	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.134	04:22
Breath Volume	1.578	04:22
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:24

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SOSA ZAMBRANO

DATE/TIME OF TEST: 06/22/2025 01:44:42

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.130	01:40
Breath Volume	3.058	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.131	01:43
Breath Volume	1.371	01:43
Air Blank	0.000	01:43
Control Test	0.079	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO-PEREZ,MICHAEL,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: COUCH,DILLON
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HAIR

DATE/TIME OF TEST: 06/23/2025 02:18:59

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.194	02:14
Breath Volume	2.515	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample # 2	0.187	02:17
Breath Volume	2.824	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARZYGnat,DESTINY,P
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCARLEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 07:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 06/24/2025 09:28:22

Diagnostics Check	OK	09:22
Air Blank	0.000	09:22
Control Test	0.080	09:23
Air Blank	0.000	09:23
Subject Sample # 1	0.000	09:24
Breath Volume	2.015	09:24
Air Blank	0.000	09:24
Air Blank	0.000	09:26
Subject Sample # 2	0.000	09:27
Breath Volume	1.984	09:27
Air Blank	0.000	09:27
Control Test	0.080	09:27
Air Blank	0.000	09:28
Diagnostics Check	OK	09:28

Data Download Date/Time: 07/24/2025 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCDONALD,AVERY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 06/26/2025 00:58:49

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.079	00:48	
Air Blank	0.000	00:48	
Subject Sample	# 1 0.165	00:51	Volume Not Met
Breath Volume	0.035	00:51	
Air Blank	0.000	00:52	
Air Blank	0.000	00:54	
Subject Sample	# 2	00:57	No Sample Provided
Breath Volume	0.000	00:57	
Air Blank	0.000	00:57	
Control Test	0.079	00:58	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 07/24/2025 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELLER,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HELLER,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/26/2025 02:33:15

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.000	02:29
Breath Volume	1.128	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.000	02:31
Breath Volume	1.230	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 07/24/2025 12:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELLER,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HELLER,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: WALLS

DATE/TIME OF TEST: 06/27/2025 19:07:29

Diagnostics Check	OK	19:01
Air Blank	0.000	19:02
Control Test	0.080	19:02
Air Blank	0.000	19:02
Subject Sample # 1	0.023	19:03
Breath Volume	2.593	19:03
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.024	19:06
Breath Volume	2.445	19:06
Air Blank	0.000	19:06
Control Test	0.079	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 07/24/2025 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HERRADA, J
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 06/27/2025 22:04:35

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.206	21:59
Breath Volume	2.179	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:01
Subject Sample # 2	0.212	22:02
Breath Volume	2.156	22:02
Air Blank	0.000	22:03
Control Test	0.078	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 07/24/2025 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 06/28/2025 21:54:18

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.081	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.196	21:49
Breath Volume	2.902	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.184	21:52
Breath Volume	1.738	21:52
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 07/24/2025 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: RICKS

DATE/TIME OF TEST: 06/29/2025 00:12:36

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:10	
Control Test	0.079	00:10	
Air Blank	0.000	00:10	
Subject Sample	# 1	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:11	
Control Test	0.079	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 07/24/2025 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: RICKS

DATE/TIME OF TEST: 06/29/2025 00:36:20

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.000	00:32
Breath Volume	2.898	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.000	00:34
Breath Volume	2.757	00:34
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 07/24/2025 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: HELLER

DATE/TIME OF TEST: 06/30/2025 03:03:31

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.116	02:59
Breath Volume	1.328	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.115	03:01
Breath Volume	1.191	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 07/24/2025 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: KERR

DATE/TIME OF TEST: 06/06/2025 04:31:47

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.079	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.000	04:27
Breath Volume	2.675	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.000	04:30
Breath Volume	1.812	04:30
Air Blank	0.000	04:30
Control Test	0.079	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 06/13/2025 17:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,DEMETRIUS,L
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: CLANTON,JESSICA
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 09:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:44

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/06/2025 10:11:15

Diagnostics Check	OK	10:05
Air Blank	0.000	10:05
Control Test	0.079	10:05
Air Blank	0.000	10:06
Subject Sample # 1	0.000	10:06
Breath Volume	2.816	10:06
Air Blank	0.000	10:07
Air Blank	0.000	10:09
Subject Sample # 2	0.000	10:09
Breath Volume	2.492	10:09
Air Blank	0.000	10:10
Control Test	0.080	10:10
Air Blank	0.000	10:11
Diagnostics Check	OK	10:11

Data Download Date/Time: 06/13/2025 17:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/08/2025 02:46:04

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.211	02:39
Breath Volume	1.261	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.209	02:44
Breath Volume	1.621	02:44
Air Blank	0.000	02:45
Control Test	0.078	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:46

Data Download Date/Time: 06/13/2025 17:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLSEN, IAN, N
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: TEAT, MATTHEW
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 12:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:51

SUBJECT'S LAST NAME: ACQUAFREDDA

DATE/TIME OF TEST: 06/09/2025 13:22:03

Diagnostics Check	OK	13:14
Air Blank	0.000	13:14
Control Test	0.079	13:15
Air Blank	0.000	13:15
Subject Sample # 1	0.000	13:16
Breath Volume	2.191	13:16
Air Blank	0.000	13:16
Air Blank	0.000	13:18
Subject Sample # 2	0.000	13:20
Breath Volume	1.710	13:20
Air Blank	0.000	13:21
Control Test	0.080	13:21
Air Blank	0.000	13:21
Diagnostics Check	OK	13:22

Data Download Date/Time: 06/13/2025 17:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: LLOVIO

DATE/TIME OF TEST: 06/13/2025 02:31:45

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.116	02:27
Breath Volume	2.039	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.113	02:30
Breath Volume	1.562	02:30
Air Blank	0.000	02:30
Control Test	0.078	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 06/13/2025 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARIV,ALI,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ZARIV,ALI
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: PINEIRO

DATE/TIME OF TEST: 06/14/2025 22:30:53

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.078	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.281	22:23
Breath Volume	4.074	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.239	22:26
Breath Volume	2.062	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:28
Subject Sample # 3	0.275	22:29
Breath Volume	3.128	22:29
Air Blank	0.000	22:30
Control Test	0.078	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 07/14/2025 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BARNES,NOAH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: IRBY

DATE/TIME OF TEST: 06/22/2025 00:44:26

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.079	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.274	00:40
Breath Volume	2.453	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.256	00:42
Breath Volume	1.921	00:42
Air Blank	0.000	00:43
Control Test	0.078	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 07/14/2025 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: GIVANS,GARRETH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: TSCHERNIA

DATE/TIME OF TEST: 06/22/2025 22:02:18

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.194	21:57
Breath Volume	4.015	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.190	22:00
Breath Volume	4.039	22:00
Air Blank	0.000	22:01
Control Test	0.078	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 07/14/2025 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: STAHL

DATE/TIME OF TEST: 06/27/2025 02:13:27

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.155	02:08
Breath Volume	1.777	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample # 2	0.142	02:12
Breath Volume	1.441	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 07/14/2025 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARIV,ALI,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ZARIV,ALI
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 06/29/2025 03:05:11

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.094	02:59
Breath Volume	1.394	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.085	03:03
Breath Volume	1.117	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 07/14/2025 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARIV,ALI,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ZARIV,ALI
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PANTOJA

DATE/TIME OF TEST: 06/29/2025 03:25:38

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.080	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:18	No Sample Provided
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Air Blank	0.000	03:20	
Subject Sample	# 2	03:24	Volume Not Met
Breath Volume	0.816	03:24	
Air Blank	0.000	03:24	
Control Test	0.081	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 07/14/2025 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARIV,ALI,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BARNES,NOAH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PANTOJA

DATE/TIME OF TEST: 06/29/2025 03:33:02

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:30	
Control Test	0.078	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:32	Sequence Aborted
Breath Volume	0.117	03:32	
Air Blank	0.000	03:33	

Data Download Date/Time: 07/14/2025 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: RASMUSSEN,MATTHEW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: EDGE

DATE/TIME OF TEST: 06/03/2025 22:48:33

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:41	
Control Test	0.081	22:41	
Air Blank	0.000	22:42	
Subject Sample # 1	0.000	22:43	
Breath Volume	2.378	22:43	
Air Blank	0.000	22:43	
Air Blank	0.000	22:45	
Subject Sample # 2	0.000	22:46	
Breath Volume	2.156	22:46	
Air Blank	0.000	22:47	
Control Test	0.082	22:47	
Air Blank	0.000	22:47	
Diagnostics Check	Fail	22:48	Temperature Regulation Test
Air Blank	0.000	22:48	

Data Download Date/Time: 06/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: RASMUSSEN,MATTHEW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 06/07/2025 20:56:23

Diagnostics Check	OK	20:48	
Air Blank	0.000	20:49	
Control Test	0.081	20:49	
Air Blank	0.000	20:49	
Subject Sample	# 1 0.139	20:51	
Breath Volume	2.644	20:51	
Air Blank	0.000	20:51	
Air Blank	0.000	20:53	
Subject Sample	# 2 0.130	20:54	
Breath Volume	2.558	20:54	
Air Blank	0.000	20:54	
Control Test	0.080	20:55	
Air Blank	0.000	20:55	
Diagnostics Check	Fail	20:55	Temperature Regulation Test
Air Blank	0.000	20:56	

Data Download Date/Time: 06/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: RASMUSSEN,MATTHEW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: HADDIX

DATE/TIME OF TEST: 06/07/2025 23:56:42

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.081	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.032	23:52
Breath Volume	4.359	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.031	23:55
Breath Volume	4.109	23:55
Air Blank	0.000	23:55
Control Test	0.081	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 06/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: RASMUSSEN,MATTHEW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: CARDOZA

DATE/TIME OF TEST: 06/09/2025 03:42:55

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.214	03:37
Breath Volume	1.261	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.225	03:41
Breath Volume	1.542	03:41
Air Blank	0.000	03:42
Control Test	0.081	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 06/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: CARUSO,JOSEPH
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: FAYSON

DATE/TIME OF TEST: 06/17/2025 00:13:50

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:02	
Control Test	0.082	00:02	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:06	No Sample Provided
Breath Volume	0.000	00:06	
Air Blank	0.000	00:06	
Air Blank	0.000	00:08	
Subject Sample	# 2	00:11	No Sample Provided
Breath Volume	0.000	00:11	
Air Blank	0.000	00:12	
Control Test	0.081	00:12	
Air Blank	0.000	00:13	
Diagnostics Check	Fail	00:13	Temperature Regulation Test
Air Blank	0.000	00:13	

Data Download Date/Time: 06/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SILVESTRI,CODY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/20/2025 02:13:49

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.081	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.100	02:09
Breath Volume	1.707	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.108	02:12
Breath Volume	1.816	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 06/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DOMINGUEZ,ANDREW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 06/22/2025 00:47:53

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.082	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.220	00:42
Breath Volume	3.453	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.211	00:46
Breath Volume	2.800	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 06/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DOMINGUEZ,ANDREW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CHECCHIA

DATE/TIME OF TEST: 06/22/2025 04:10:49

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:07	
Control Test	0.081	04:07	
Air Blank	0.000	04:08	
Subject Sample	# 1	04:09	RFI Detect
Breath Volume	0.000	04:09	
Air Blank	0.000	04:09	
Control Test	0.081	04:10	
Air Blank	0.000	04:10	
Diagnostics Check	OK	04:10	

Data Download Date/Time: 06/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DOMINGUEZ,ANDREW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CHECCHIA

DATE/TIME OF TEST: 06/22/2025 04:25:39

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:13	
Control Test	0.081	04:13	
Air Blank	0.000	04:14	
Subject Sample	# 1 0.144	04:17	Volume Not Met
Breath Volume	0.105	04:17	
Air Blank	0.000	04:17	
Air Blank	0.000	04:19	
Subject Sample	# 2 0.133	04:20	
Breath Volume	1.250	04:20	
Air Blank	0.000	04:21	
Air Blank	0.000	04:23	
Subject Sample	# 3 0.138	04:24	
Breath Volume	1.332	04:24	
Air Blank	0.000	04:24	
Control Test	0.080	04:25	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 06/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: CARUSO,JOSEPH
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: USELTON

DATE/TIME OF TEST: 06/22/2025 21:45:37

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:38	
Control Test	0.082	21:38	
Air Blank	0.000	21:38	
Subject Sample	# 1 0.000	21:40	
Breath Volume	3.226	21:40	
Air Blank	0.000	21:41	
Air Blank	0.000	21:43	
Subject Sample	# 2 0.000	21:43	
Breath Volume	2.699	21:43	
Air Blank	0.000	21:44	
Control Test	0.081	21:44	
Air Blank	0.000	21:45	
Diagnostics Check	Fail	21:45	Temperature Regulation Test
Air Blank	0.000	21:45	

Data Download Date/Time: 06/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: BELFORD,DALEN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 06/27/2025 21:17:08

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.082	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.415	21:12
Breath Volume	1.953	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:14
Subject Sample # 2	0.403	21:15
Breath Volume	2.281	21:15
Air Blank	0.000	21:16
Control Test	0.079	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 06/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SHORTT,BRONSON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: TOVAR

DATE/TIME OF TEST: 06/30/2025 22:22:44

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.081	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.124	22:17
Breath Volume	1.832	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.125	22:21
Breath Volume	1.566	22:21
Air Blank	0.000	22:21
Control Test	0.081	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 07/29/2025 20:27