

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,BRYAN,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MARTINI,BRYAN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: SALERA

DATE/TIME OF TEST: 07/01/2025 18:38:21

Diagnostics Check	OK	18:23	
Air Blank	0.000	18:24	
Control Test	0.080	18:24	
Air Blank	0.000	18:25	
Subject Sample # 1	0.000	18:25	
Breath Volume	3.289	18:25	
Air Blank	0.000	18:26	
Air Blank	0.000	18:28	
Subject Sample # 2		18:31	No Sample Provided
Breath Volume	0.000	18:31	
Air Blank	0.000	18:31	
Air Blank	0.000	18:33	
Subject Sample # 3		18:36	Subject Test Refused
Breath Volume	0.000	18:36	
Air Blank	0.000	18:37	
Control Test	0.079	18:37	
Air Blank	0.000	18:38	
Diagnostics Check	OK	18:38	

Data Download Date/Time: 07/24/2025 22:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/01/2025 20:43:24

Diagnostics Check	OK	20:37
Air Blank	0.000	20:37
Control Test	0.079	20:38
Air Blank	0.000	20:38
Subject Sample # 1	0.000	20:39
Breath Volume	2.335	20:39
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.000	20:41
Breath Volume	2.507	20:41
Air Blank	0.000	20:42
Control Test	0.079	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 07/24/2025 22:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOUSSAINT,WESNES,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TOUSSAINT,WESNES
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 07/02/2025 17:32:20

Diagnostics Check	OK	17:20	
Air Blank	0.000	17:21	
Control Test	0.078	17:21	
Air Blank	0.000	17:22	
Subject Sample	# 1	17:25	No Sample Provided
Breath Volume	0.000	17:25	
Air Blank	0.000	17:25	
Air Blank	0.000	17:27	
Subject Sample	# 2	17:30	No Sample Provided
Breath Volume	0.000	17:30	
Air Blank	0.000	17:31	
Control Test	0.079	17:31	
Air Blank	0.000	17:32	
Diagnostics Check	OK	17:32	

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LAWTON, BRENNAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: REAM

DATE/TIME OF TEST: 07/03/2025 19:15:33

Diagnostics Check	OK	19:08
Air Blank	0.000	19:08
Control Test	0.079	19:09
Air Blank	0.000	19:09
Subject Sample # 1	0.152	19:10
Breath Volume	1.835	19:10
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample # 2	0.155	19:14
Breath Volume	2.281	19:14
Air Blank	0.000	19:14
Control Test	0.078	19:14
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUFFNER,PRESTON,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: KUFFNER,PRESTON
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: EVERAGE

DATE/TIME OF TEST: 07/03/2025 23:46:31

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.079	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.000	23:42
Breath Volume	1.933	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.000	23:45
Breath Volume	2.058	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,ERIC,C
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LOTT,ERIC
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 07/04/2025 00:11:56

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.092	00:07
Breath Volume	2.281	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.093	00:10
Breath Volume	2.050	00:10
Air Blank	0.000	00:10
Control Test	0.079	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUFFNER,PRESTON,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: KUFFNER,PRESTON
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 07/04/2025 03:18:35

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.208	03:14
Breath Volume	4.351	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.208	03:17
Breath Volume	3.972	03:17
Air Blank	0.000	03:17
Control Test	0.078	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: FERNINDEZ-ZALDIVAR

DATE/TIME OF TEST: 07/04/2025 22:53:59

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.079	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.202	22:49
Breath Volume	3.007	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.212	22:52
Breath Volume	2.011	22:52
Air Blank	0.000	22:53
Control Test	0.078	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:31

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 07/05/2025 17:07:34

Diagnostics Check	OK	17:00
Air Blank	0.000	17:00
Control Test	0.079	17:00
Air Blank	0.000	17:01
Subject Sample # 1	0.305	17:02
Breath Volume	1.816	17:02
Air Blank	0.000	17:03
Air Blank	0.000	17:05
Subject Sample # 2	0.309	17:06
Breath Volume	1.292	17:06
Air Blank	0.000	17:06
Control Test	0.078	17:06
Air Blank	0.000	17:07
Diagnostics Check	OK	17:07

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN, CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: RODRIGUEZ PINEDA

DATE/TIME OF TEST: 07/06/2025 01:21:45

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.079	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.348	01:16
Breath Volume	1.136	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.338	01:20
Breath Volume	1.937	01:20
Air Blank	0.000	01:20
Control Test	0.077	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORMAN,CHASE,J
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: BURKE,LOGAN
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: PELTON

DATE/TIME OF TEST: 07/06/2025 04:27:25

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.079	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.000	04:23
Breath Volume	2.417	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.000	04:26
Breath Volume	2.320	04:26
Air Blank	0.000	04:26
Control Test	0.078	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTO,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: RESTO,LUIS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: DAVENPORT

DATE/TIME OF TEST: 07/09/2025 18:55:15

Diagnostics Check	OK	18:49
Air Blank	0.000	18:49
Control Test	0.080	18:49
Air Blank	0.000	18:50
Subject Sample # 1	0.038	18:50
Breath Volume	4.347	18:50
Air Blank	0.000	18:51
Air Blank	0.000	18:53
Subject Sample # 2	0.042	18:53
Breath Volume	6.417	18:53
Air Blank	0.000	18:54
Control Test	0.078	18:54
Air Blank	0.000	18:55
Diagnostics Check	OK	18:55

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: KUERZ

DATE/TIME OF TEST: 07/10/2025 23:47:48

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.354	23:43
Breath Volume	1.687	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.339	23:46
Breath Volume	2.074	23:46
Air Blank	0.000	23:46
Control Test	0.078	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRADY,REBECCA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:19

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 07/11/2025 18:49:34

Diagnostics Check	OK	18:46	
Air Blank	0.000	18:47	
Control Test	0.080	18:47	
Air Blank	0.000	18:47	
Subject Sample	# 1	18:48	Subject Test Refused
Breath Volume	0.000	18:48	
Air Blank	0.000	18:48	
Control Test	0.079	18:48	
Air Blank	0.000	18:49	
Diagnostics Check	OK	18:49	

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRADY,REBECCA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:09

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 07/12/2025 12:45:58

Diagnostics Check	OK	12:34	
Air Blank	0.000	12:34	
Control Test	0.080	12:35	
Air Blank	0.000	12:35	
Subject Sample	# 1	12:38	No Sample Provided
Breath Volume	0.000	12:38	
Air Blank	0.000	12:39	
Air Blank	0.000	12:41	
Subject Sample	# 2	12:44	No Sample Provided
Breath Volume	0.000	12:44	
Air Blank	0.000	12:45	
Control Test	0.080	12:45	
Air Blank	0.000	12:45	
Diagnostics Check	OK	12:45	

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOUSSAINT,WESNES,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TOUSSAINT,WESNES
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: DUCOTE

DATE/TIME OF TEST: 07/12/2025 18:02:58

Diagnostics Check	OK	17:49	
Air Blank	0.000	17:49	
Control Test	0.080	17:50	
Air Blank	0.000	17:50	
Subject Sample	# 1	17:53	No Sample Provided
Breath Volume	0.000	17:53	
Air Blank	0.000	17:54	
Air Blank	0.000	17:56	
Subject Sample	# 2	0.264	17:57
Breath Volume	4.058	17:57	
Air Blank	0.000	17:58	
Air Blank	0.000	18:00	
Subject Sample	# 3	0.252	18:01
Breath Volume	4.488	18:01	
Air Blank	0.000	18:02	
Control Test	0.079	18:02	
Air Blank	0.000	18:02	
Diagnostics Check	OK	18:02	

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUFFNER,PRESTON,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: KUFFNER,PRESTON
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: FATOVICH

DATE/TIME OF TEST: 07/12/2025 22:02:35

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.079	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.301	21:58
Breath Volume	3.554	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.293	22:01
Breath Volume	2.558	22:01
Air Blank	0.000	22:01
Control Test	0.079	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: SALENETRI

DATE/TIME OF TEST: 07/13/2025 19:45:00

Diagnostics Check	OK	19:38
Air Blank	0.000	19:38
Control Test	0.080	19:39
Air Blank	0.000	19:39
Subject Sample # 1	0.000	19:40
Breath Volume	3.191	19:40
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.000	19:43
Breath Volume	2.992	19:43
Air Blank	0.000	19:44
Control Test	0.079	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:45

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 07/15/2025 22:51:31

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.233	22:47
Breath Volume	3.183	22:47
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.227	22:49
Breath Volume	3.078	22:49
Air Blank	0.000	22:50
Control Test	0.079	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: RAMIREZ NOLASCO

DATE/TIME OF TEST: 07/16/2025 20:42:04

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.080	20:36
Air Blank	0.000	20:36
Subject Sample # 1	0.203	20:37
Breath Volume	1.164	20:37
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.209	20:40
Breath Volume	1.945	20:40
Air Blank	0.000	20:41
Control Test	0.078	20:41
Air Blank	0.000	20:41
Diagnostics Check	OK	20:42

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTEUCCI,DANIELLE,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: MATTEUCCI,DANIELLE
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: SALZBERG

DATE/TIME OF TEST: 07/17/2025 22:05:35

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:52	
Control Test	0.079	21:52	
Air Blank	0.000	21:53	
Subject Sample	# 1 0.137	21:56	Volume Not Met
Breath Volume	0.679	21:56	
Air Blank	0.000	21:56	
Air Blank	0.000	21:58	
Subject Sample	# 2 0.160	22:01	
Breath Volume	1.289	22:01	
Air Blank	0.000	22:01	
Air Blank	0.000	22:03	
Subject Sample	# 3 0.155	22:04	
Breath Volume	1.480	22:04	
Air Blank	0.000	22:04	
Control Test	0.074	22:04	Control Outside Tolerance
Air Blank	0.000	22:05	

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTEUCCI,DANIELLE,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: MATTEUCCI,DANIELLE
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: SALZBERG

DATE/TIME OF TEST: 07/17/2025 22:18:08

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.078	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.151	22:12
Breath Volume	1.125	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.152	22:16
Breath Volume	1.644	22:16
Air Blank	0.000	22:17
Control Test	0.078	22:17
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 07/24/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: COMAS

DATE/TIME OF TEST: 07/19/2025 03:04:49

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.180	03:00
Breath Volume	2.644	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.180	03:03
Breath Volume	2.890	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 07/24/2025 22:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORMAN,CHASE,J
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: BURKE,LOGAN
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: SADDLER

DATE/TIME OF TEST: 07/19/2025 23:37:45

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.073	23:33
Breath Volume	1.859	23:33
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.064	23:36
Breath Volume	2.468	23:36
Air Blank	0.000	23:36
Control Test	0.078	23:37
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 07/24/2025 22:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ST. THOMAS,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ST. THOMAS,JOSHUA
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:24

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 07/20/2025 10:55:59

Diagnostics Check	OK	10:49
Air Blank	0.000	10:50
Control Test	0.080	10:50
Air Blank	0.000	10:50
Subject Sample # 1	0.147	10:51
Breath Volume	1.703	10:51
Air Blank	0.000	10:51
Air Blank	0.000	10:53
Subject Sample # 2	0.144	10:54
Breath Volume	3.007	10:54
Air Blank	0.000	10:55
Control Test	0.079	10:55
Air Blank	0.000	10:55
Diagnostics Check	OK	10:55

Data Download Date/Time: 07/24/2025 22:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORMAN,CHASE,J
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: BURKE,LOGAN
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: REEDER

DATE/TIME OF TEST: 07/24/2025 05:12:00

Diagnostics Check	OK	05:05
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.044	05:07
Breath Volume	2.089	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:09
Subject Sample # 2	0.041	05:10
Breath Volume	2.058	05:10
Air Blank	0.000	05:11
Control Test	0.079	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:12

Data Download Date/Time: 07/24/2025 22:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 07/24/2025 21:00:34

Diagnostics Check	OK	20:53
Air Blank	0.000	20:54
Control Test	0.080	20:54
Air Blank	0.000	20:55
Subject Sample # 1	0.242	20:56
Breath Volume	1.613	20:56
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.249	20:59
Breath Volume	1.511	20:59
Air Blank	0.000	20:59
Control Test	0.078	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 07/24/2025 22:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: ALLFREY JR

DATE/TIME OF TEST: 07/26/2025 20:01:06

Diagnostics Check	OK	19:54
Air Blank	0.000	19:54
Control Test	0.079	19:55
Air Blank	0.000	19:55
Subject Sample # 1	0.219	19:56
Breath Volume	2.875	19:56
Air Blank	0.000	19:57
Air Blank	0.000	19:58
Subject Sample # 2	0.214	19:59
Breath Volume	2.527	19:59
Air Blank	0.000	20:00
Control Test	0.078	20:00
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 08/21/2025 00:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUFFNER,PRESTON,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: KUFFNER,PRESTON
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/27/2025 01:45:25

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.181	01:40
Breath Volume	2.828	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.183	01:43
Breath Volume	2.824	01:43
Air Blank	0.000	01:44
Control Test	0.078	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 08/21/2025 00:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEBLAKER,JOSHUA,N
BREATH TEST OPERATOR AGENCY: Sewall's Point PD
ARRESTING OFFICER: DEBLAKER,JOSHUA
ARRESTING OFFICER AGENCY: Sewall's Point PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: ASPRELLI

DATE/TIME OF TEST: 07/28/2025 01:19:13

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.079	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.176	01:11
Breath Volume	2.796	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.155	01:14
Breath Volume	1.789	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 3	0.163	01:17
Breath Volume	1.984	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 08/21/2025 00:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIGG,JACOB,M
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: PIGG,JACOB
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 07/28/2025 03:10:40

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:03	No Sample Provided
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Air Blank	0.000	03:06	
Subject Sample	# 2	03:09	No Sample Provided
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.079	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 08/21/2025 00:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: PIQUETTE

DATE/TIME OF TEST: 07/28/2025 18:13:48

Diagnostics Check	OK	18:06
Air Blank	0.000	18:06
Control Test	0.079	18:06
Air Blank	0.000	18:07
Subject Sample # 1	0.231	18:08
Breath Volume	1.593	18:08
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.221	18:12
Breath Volume	2.175	18:12
Air Blank	0.000	18:12
Control Test	0.078	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 08/21/2025 00:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: RODRIGUEZ-RODRIGUEZ

DATE/TIME OF TEST: 07/29/2025 23:16:02

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.175	23:11
Breath Volume	3.312	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:13
Subject Sample # 2	0.169	23:14
Breath Volume	3.398	23:14
Air Blank	0.000	23:15
Control Test	0.079	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:16

Data Download Date/Time: 08/21/2025 00:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOUSSAINT,WESNES,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: W.TOUSSAINT,WESNES
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: SHOEMAKER

DATE/TIME OF TEST: 07/30/2025 09:12:16

Diagnostics Check	OK	09:04
Air Blank	0.000	09:04
Control Test	0.079	09:05
Air Blank	0.000	09:05
Subject Sample # 1	0.000	09:06
Breath Volume	4.585	09:06
Air Blank	0.000	09:07
Air Blank	0.000	09:09
Subject Sample # 2	0.000	09:10
Breath Volume	2.734	09:10
Air Blank	0.000	09:11
Control Test	0.079	09:11
Air Blank	0.000	09:12
Diagnostics Check	OK	09:12

Data Download Date/Time: 08/21/2025 00:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUFFNER,PRESTON,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: KUFFNER,PRESTON
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: POST

DATE/TIME OF TEST: 07/31/2025 20:54:43

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.079	20:48
Air Blank	0.000	20:49
Subject Sample # 1	0.147	20:49
Breath Volume	3.300	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.153	20:53
Breath Volume	3.492	20:53
Air Blank	0.000	20:53
Control Test	0.078	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 08/21/2025 00:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,A,
BREATH TEST OPERATOR AGENCY: Mascotte PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 06/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: LUPIAN

DATE/TIME OF TEST: 07/01/2025 21:51:57

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.000	21:47
Breath Volume	3.203	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.000	21:50
Breath Volume	1.902	21:50
Air Blank	0.000	21:51
Control Test	0.078	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 07/07/2025 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,STEPHEN,V
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: SANTOS,MANUAL
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 06/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: HAMMONDS

DATE/TIME OF TEST: 07/04/2025 03:32:08

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.079	03:28	
Air Blank	0.000	03:29	
Subject Sample	# 1 0.263	03:30	
Breath Volume	1.933	03:30	
Air Blank		03:31	Purge Fail
Air Blank	0.000	03:32	

Data Download Date/Time: 07/07/2025 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,STEPHEN,V
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: SANTOS,MANUAL
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 06/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: HAMMONDS

DATE/TIME OF TEST: 07/04/2025 03:49:08

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:39	
Control Test	0.077	03:39	
Air Blank	0.000	03:40	
Subject Sample	# 1 0.256	03:42	
Breath Volume	1.804	03:42	
Air Blank	0.000	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 2 0.255	03:47	
Breath Volume	1.835	03:47	
Air Blank		03:48	Purge Fail
Air Blank	0.000	03:49	

Data Download Date/Time: 07/07/2025 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP, STEVE, A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BARNES, MONIQUE
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 07/04/2025 15:22:05

Diagnostics Check	OK	15:18	
Air Blank	0.000	15:18	
Control Test	0.079	15:18	
Air Blank	0.000	15:19	
Subject Sample	# 1	15:20	Subject Test Refused
Breath Volume	0.000	15:20	
Air Blank	0.000	15:21	
Control Test	0.078	15:21	
Air Blank	0.000	15:22	
Diagnostics Check	OK	15:22	

Data Download Date/Time: 07/07/2025 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,A,
BREATH TEST OPERATOR AGENCY: Mascotte PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 06/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: DONALDSON

DATE/TIME OF TEST: 07/05/2025 21:43:58

Diagnostics Check	OK	21:37
Air Blank	0.000	21:38
Control Test	0.079	21:38
Air Blank	0.000	21:38
Subject Sample # 1	0.079	21:39
Breath Volume	2.542	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:41
Subject Sample # 2	0.078	21:42
Breath Volume	2.613	21:42
Air Blank	0.000	21:43
Control Test	0.078	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 07/07/2025 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ GARCIA,HERNAN,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DIBERNARDINO,BRANDON
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: GARCIA SOREQUE

DATE/TIME OF TEST: 07/10/2025 22:41:49

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.082	22:31
Air Blank	0.000	22:32
Subject Sample # 1	0.188	22:33
Breath Volume	2.128	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.210	22:36
Breath Volume	3.156	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 3	0.217	22:39
Breath Volume	2.394	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 07/14/2025 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIBERNARDINO,BRANDON,M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DIBERNARDINO,BRANDON
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: STOOPS

DATE/TIME OF TEST: 07/13/2025 03:20:32

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.082	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.195	03:15
Breath Volume	1.605	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.183	03:18
Breath Volume	1.988	03:18
Air Blank	0.000	03:19
Control Test	0.081	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 07/14/2025 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/01/2025 23:40:02

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.056	23:35
Breath Volume	3.699	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.055	23:38
Breath Volume	3.160	23:38
Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:40

Data Download Date/Time: 07/15/2025 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,SHELBY,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GARDNER,SHELBY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: MALDONADO MALDONADO

DATE/TIME OF TEST: 07/05/2025 18:21:24

Diagnostics Check	OK	18:15	
Air Blank	0.000	18:15	
Control Test	0.079	18:15	
Air Blank	0.000	18:16	
Subject Sample	# 1	18:19	No Sample Provided
Breath Volume	0.000	18:19	
Air Blank	0.000	18:20	
Air Blank	0.000	18:21	

Data Download Date/Time: 07/15/2025 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,SHELBY,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GARDNER,SHELBY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: GIRON

DATE/TIME OF TEST: 07/05/2025 18:30:10

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.079	18:24
Air Blank	0.000	18:24
Subject Sample # 1	0.176	18:25
Breath Volume	2.265	18:25
Air Blank	0.000	18:26
Air Blank	0.000	18:28
Subject Sample # 2	0.165	18:28
Breath Volume	2.167	18:28
Air Blank	0.000	18:29
Control Test	0.078	18:29
Air Blank	0.000	18:30
Diagnostics Check	OK	18:30

Data Download Date/Time: 07/15/2025 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,JAIME
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: PEREZ DOMINGO

DATE/TIME OF TEST: 07/03/2025 01:36:57

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.241	01:32
Breath Volume	1.992	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.249	01:35
Breath Volume	3.257	01:35
Air Blank	0.000	01:36
Control Test	0.077	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL, KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: TORRES BELLO

DATE/TIME OF TEST: 07/04/2025 01:32:25

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:27	
Control Test	0.079	01:28	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HECTOR, BRADY
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: DANIELS III

DATE/TIME OF TEST: 07/04/2025 20:58:22

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.079	20:52
Air Blank	0.000	20:52
Subject Sample # 1	0.040	20:53
Breath Volume	1.187	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.043	20:56
Breath Volume	1.246	20:56
Air Blank	0.000	20:57
Control Test	0.079	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BYRO

DATE/TIME OF TEST: 07/05/2025 00:21:37

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.213	00:15
Breath Volume	1.125	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.204	00:20
Breath Volume	1.457	00:20
Air Blank	0.000	00:20
Control Test	0.079	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/05/2025 02:36:35

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.115	02:32
Breath Volume	2.140	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.118	02:35
Breath Volume	2.011	02:35
Air Blank	0.000	02:35
Control Test	0.078	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 07/05/2025 04:14:16

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.095	04:09
Breath Volume	1.738	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.102	04:12
Breath Volume	2.203	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ONG,ANDY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 07/05/2025 08:36:08

Diagnostics Check	OK	08:29
Air Blank	0.000	08:29
Control Test	0.079	08:30
Air Blank	0.000	08:30
Subject Sample # 1	0.115	08:31
Breath Volume	2.449	08:31
Air Blank	0.000	08:32
Air Blank	0.000	08:34
Subject Sample # 2	0.118	08:34
Breath Volume	2.472	08:34
Air Blank	0.000	08:35
Control Test	0.078	08:35
Air Blank	0.000	08:36
Diagnostics Check	OK	08:36

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: GALEOTE DURAN

DATE/TIME OF TEST: 07/06/2025 02:16:26

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.202	02:11
Breath Volume	2.179	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.204	02:14
Breath Volume	2.367	02:14
Air Blank	0.000	02:15
Control Test	0.078	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: ESCALANTE

DATE/TIME OF TEST: 07/06/2025 05:16:28

Diagnostics Check	OK	05:08
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.140	05:11
Breath Volume	1.699	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.136	05:14
Breath Volume	2.050	05:14
Air Blank	0.000	05:15
Control Test	0.079	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARTEE, JESSICA
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: POHARESKY

DATE/TIME OF TEST: 07/06/2025 09:39:40

Diagnostics Check	OK	09:33
Air Blank	0.000	09:34
Control Test	0.079	09:34
Air Blank	0.000	09:35
Subject Sample # 1	0.000	09:35
Breath Volume	1.718	09:35
Air Blank	0.000	09:35
Air Blank	0.000	09:37
Subject Sample # 2	0.000	09:38
Breath Volume	1.890	09:38
Air Blank	0.000	09:38
Control Test	0.080	09:39
Air Blank	0.000	09:39
Diagnostics Check	OK	09:39

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLASCO, LUIS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: MALDONADO DOMINGO

DATE/TIME OF TEST: 07/06/2025 16:19:21

Diagnostics Check	OK	16:12
Air Blank	0.000	16:13
Control Test	0.079	16:13
Air Blank	0.000	16:14
Subject Sample # 1	0.192	16:14
Breath Volume	1.562	16:14
Air Blank	0.000	16:15
Air Blank	0.000	16:17
Subject Sample # 2	0.191	16:17
Breath Volume	2.175	16:17
Air Blank	0.000	16:18
Control Test	0.079	16:18
Air Blank	0.000	16:19
Diagnostics Check	OK	16:19

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KOBITKA, NICHOLAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: BRAZEL

DATE/TIME OF TEST: 07/06/2025 17:25:56

Diagnostics Check	OK	17:19
Air Blank	0.000	17:20
Control Test	0.078	17:20
Air Blank	0.000	17:21
Subject Sample # 1	0.166	17:21
Breath Volume	2.414	17:21
Air Blank	0.000	17:22
Air Blank	0.000	17:23
Subject Sample # 2	0.167	17:24
Breath Volume	2.687	17:24
Air Blank	0.000	17:25
Control Test	0.079	17:25
Air Blank	0.000	17:25
Diagnostics Check	OK	17:25

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/09/2025 19:47:21

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.080	19:40
Air Blank	0.000	19:41
Subject Sample # 1	0.194	19:42
Breath Volume	1.156	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:44
Subject Sample # 2	0.201	19:45
Breath Volume	1.148	19:45
Air Blank	0.000	19:46
Control Test	0.078	19:46
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: KRZEMIEN

DATE/TIME OF TEST: 07/09/2025 23:43:58

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.000	23:39
Breath Volume	3.089	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.000	23:42
Breath Volume	3.070	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS, WESLEY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 07/10/2025 20:38:29

Diagnostics Check	OK	20:32
Air Blank	0.000	20:32
Control Test	0.079	20:33
Air Blank	0.000	20:33
Subject Sample # 1	0.046	20:34
Breath Volume	1.644	20:34
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.046	20:36
Breath Volume	2.804	20:36
Air Blank	0.000	20:37
Control Test	0.079	20:37
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: KOCSIS

DATE/TIME OF TEST: 07/10/2025 21:00:48

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.080	20:55
Air Blank	0.000	20:55
Subject Sample # 1	0.213	20:56
Breath Volume	2.179	20:56
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.209	20:59
Breath Volume	2.214	20:59
Air Blank	0.000	20:59
Control Test	0.078	21:00
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 07/11/2025 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: ORTIZ NUNEZ

DATE/TIME OF TEST: 07/11/2025 23:29:29

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.171	23:24
Breath Volume	1.382	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.164	23:27
Breath Volume	1.429	23:27
Air Blank	0.000	23:28
Control Test	0.078	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 08/15/2025 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: ALTAMIRANO

DATE/TIME OF TEST: 07/11/2025 23:57:26

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.101	23:52
Breath Volume	1.214	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.104	23:55
Breath Volume	1.121	23:55
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 08/15/2025 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CUSHNIE,MEAGHAN
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BERMINGHAM

DATE/TIME OF TEST: 07/12/2025 04:22:16

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.183	04:17
Breath Volume	1.296	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.193	04:20
Breath Volume	1.894	04:20
Air Blank	0.000	04:21
Control Test	0.079	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 08/15/2025 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: JENNE

DATE/TIME OF TEST: 07/12/2025 17:40:11

Diagnostics Check	OK	17:37	
Air Blank	0.000	17:37	
Control Test	0.080	17:38	
Air Blank	0.000	17:38	
Subject Sample	# 1	17:38	Subject Test Refused
Breath Volume	0.000	17:38	
Air Blank	0.000	17:39	
Control Test	0.079	17:39	
Air Blank	0.000	17:40	
Diagnostics Check	OK	17:40	

Data Download Date/Time: 08/15/2025 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLUER, ROBINSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MENDEZ JIMENEZ

DATE/TIME OF TEST: 07/12/2025 23:07:13

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.098	23:02
Breath Volume	2.718	23:02
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.099	23:05
Breath Volume	2.660	23:05
Air Blank	0.000	23:06
Control Test	0.078	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 08/15/2025 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZARBA, JEFFREY
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: YESION ESVIN

DATE/TIME OF TEST: 07/13/2025 03:48:17

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.148	03:43
Breath Volume	2.421	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.139	03:46
Breath Volume	1.789	03:46
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 08/15/2025 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: NANDLAL

DATE/TIME OF TEST: 07/13/2025 21:37:05

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.080	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.217	21:32
Breath Volume	2.277	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.221	21:35
Breath Volume	2.339	21:35
Air Blank	0.000	21:36
Control Test	0.078	21:36
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 08/15/2025 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: ENCINAS GONZALES

DATE/TIME OF TEST: 07/13/2025 22:57:00

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.079	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.261	22:52
Breath Volume	2.105	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:54
Subject Sample # 2	0.251	22:55
Breath Volume	2.304	22:55
Air Blank	0.000	22:56
Control Test	0.078	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:57

Data Download Date/Time: 08/15/2025 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MURILLO ALFARO

DATE/TIME OF TEST: 07/14/2025 02:14:17

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.198	02:09
Breath Volume	2.625	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.199	02:12
Breath Volume	2.742	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 08/15/2025 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RENTERIA,ELIZABETH
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MELENDEZ MONTEVERDE

DATE/TIME OF TEST: 07/14/2025 02:33:14

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.000	02:28
Breath Volume	3.562	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.000	02:31
Breath Volume	3.332	02:31
Air Blank	0.000	02:32
Control Test	0.079	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 08/15/2025 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: STEARNS

DATE/TIME OF TEST: 07/14/2025 04:37:46

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.134	04:33
Breath Volume	1.535	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.134	04:36
Breath Volume	1.933	04:36
Air Blank	0.000	04:36
Control Test	0.079	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 08/15/2025 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: STOUPIS

DATE/TIME OF TEST: 07/14/2025 18:50:33

Diagnostics Check	OK	18:46	
Air Blank	0.000	18:46	
Control Test	0.079	18:47	
Air Blank	0.000	18:47	
Subject Sample	# 1	18:49	Subject Test Refused
Breath Volume	0.128	18:49	
Air Blank	0.000	18:49	
Control Test	0.079	18:49	
Air Blank	0.000	18:50	
Diagnostics Check	OK	18:50	

Data Download Date/Time: 08/15/2025 16:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: CASTRO OROZCO

DATE/TIME OF TEST: 07/15/2025 22:00:21

Diagnostics Check	OK	21:52
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.153	21:55
Breath Volume	1.566	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.162	21:58
Breath Volume	1.464	21:58
Air Blank	0.000	21:59
Control Test	0.078	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 08/15/2025 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZIMMERMAN, JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: KESSLER

DATE/TIME OF TEST: 07/16/2025 23:21:30

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:10	
Control Test	0.080	23:10	
Air Blank	0.000	23:11	
Subject Sample	# 1 0.188	23:14	Volume Not Met
Breath Volume	0.414	23:14	
Air Blank	0.000	23:15	
Air Blank	0.000	23:17	
Subject Sample	# 2	23:19	Subject Test Refused
Breath Volume	0.546	23:19	
Air Blank	0.000	23:20	
Control Test	0.077	23:20	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:21	

Data Download Date/Time: 08/15/2025 16:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOMAS ANTONIO, ANIBAL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:26

SUBJECT'S LAST NAME: UGA

DATE/TIME OF TEST: 07/17/2025 13:57:21

Diagnostics Check	OK	13:51
Air Blank	0.000	13:51
Control Test	0.080	13:52
Air Blank	0.000	13:52
Subject Sample # 1	0.000	13:53
Breath Volume	1.718	13:53
Air Blank	0.000	13:53
Air Blank	0.000	13:55
Subject Sample # 2	0.000	13:55
Breath Volume	3.437	13:55
Air Blank	0.000	13:56
Control Test	0.079	13:56
Air Blank	0.000	13:57
Diagnostics Check	OK	13:57

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: CZAJKOSKI

DATE/TIME OF TEST: 07/17/2025 23:57:08

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:50
Subject Sample # 1	0.366	23:51
Breath Volume	1.285	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.385	23:55
Breath Volume	1.605	23:55
Air Blank	0.000	23:56
Control Test	0.077	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMIGELSKI, TREVOR
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/18/2025 23:37:15

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.079	23:32
Breath Volume	1.789	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.079	23:35
Breath Volume	1.898	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: JIMENEZ DIAZ

DATE/TIME OF TEST: 07/19/2025 00:39:32

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.216	00:34
Breath Volume	1.914	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.206	00:37
Breath Volume	1.289	00:37
Air Blank	0.000	00:38
Control Test	0.079	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KOZLOWSKI,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: PROFIT

DATE/TIME OF TEST: 07/19/2025 03:31:06

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.138	03:25
Breath Volume	1.296	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.136	03:29
Breath Volume	1.574	03:29
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: LUJAN SANCHEZ

DATE/TIME OF TEST: 07/19/2025 06:54:55

Diagnostics Check	OK	06:48
Air Blank	0.000	06:48
Control Test	0.079	06:48
Air Blank	0.000	06:49
Subject Sample # 1	0.186	06:50
Breath Volume	2.312	06:50
Air Blank	0.000	06:50
Air Blank	0.000	06:52
Subject Sample # 2	0.192	06:53
Breath Volume	2.351	06:53
Air Blank	0.000	06:54
Control Test	0.079	06:54
Air Blank	0.000	06:54
Diagnostics Check	OK	06:54

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BROWN, MEKHI
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: STEYN

DATE/TIME OF TEST: 07/20/2025 01:47:05

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.193	01:42
Breath Volume	2.750	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.193	01:45
Breath Volume	1.597	01:45
Air Blank	0.000	01:46
Control Test	0.078	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLASCO,LUIS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: MAX STEEVE

DATE/TIME OF TEST: 07/20/2025 10:09:48

Diagnostics Check	OK	10:00
Air Blank	0.000	10:01
Control Test	0.080	10:01
Air Blank	0.000	10:01
Subject Sample # 1	0.163	10:05
Breath Volume	2.484	10:05
Air Blank	0.000	10:05
Air Blank	0.000	10:07
Subject Sample # 2	0.168	10:08
Breath Volume	3.277	10:08
Air Blank	0.000	10:08
Control Test	0.078	10:09
Air Blank	0.000	10:09
Diagnostics Check	OK	10:09

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALKER, MASON
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: DEVEDORPH

DATE/TIME OF TEST: 07/20/2025 22:19:01

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.079	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.368	22:10
Breath Volume	2.125	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:12
Subject Sample # 2	0.345	22:14
Breath Volume	1.648	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:16
Subject Sample # 3	0.362	22:17
Breath Volume	2.378	22:17
Air Blank	0.000	22:18
Control Test	0.077	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:19

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJADA,ALBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: INDALECIO

DATE/TIME OF TEST: 07/22/2025 08:40:04

Diagnostics Check	OK	08:33
Air Blank	0.000	08:34
Control Test	0.079	08:34
Air Blank	0.000	08:35
Subject Sample # 1	0.277	08:35
Breath Volume	1.621	08:35
Air Blank	0.000	08:36
Air Blank	0.000	08:38
Subject Sample # 2	0.278	08:38
Breath Volume	2.675	08:38
Air Blank	0.000	08:39
Control Test	0.078	08:39
Air Blank	0.000	08:39
Diagnostics Check	OK	08:40

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FARDELLA,BLAKE
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: ROMANO

DATE/TIME OF TEST: 07/23/2025 19:00:30

Diagnostics Check	OK	18:54
Air Blank	0.000	18:54
Control Test	0.079	18:54
Air Blank	0.000	18:55
Subject Sample # 1	0.059	18:55
Breath Volume	1.503	18:55
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.055	18:58
Breath Volume	1.652	18:58
Air Blank	0.000	18:59
Control Test	0.079	18:59
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/24/2025 02:30:05

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.126	02:23
Breath Volume	1.339	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.122	02:28
Breath Volume	1.496	02:28
Air Blank	0.000	02:29
Control Test	0.078	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROWE,RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: MANESH

DATE/TIME OF TEST: 07/25/2025 21:18:27

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.079	21:12
Air Blank	0.000	21:12
Subject Sample # 1	0.124	21:13
Breath Volume	1.984	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.116	21:16
Breath Volume	2.125	21:16
Air Blank	0.000	21:17
Control Test	0.079	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: MORALES ANTON

DATE/TIME OF TEST: 07/25/2025 23:35:45

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.080	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.156	23:31
Breath Volume	1.460	23:31
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.156	23:34
Breath Volume	1.621	23:34
Air Blank	0.000	23:34
Control Test	0.079	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMIGELSKI,TREVOR
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: JARAMILLO SAMBRANO

DATE/TIME OF TEST: 07/26/2025 05:50:31

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.079	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.000	05:46
Breath Volume	1.781	05:46
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.000	05:49
Breath Volume	3.980	05:49
Air Blank	0.000	05:49
Control Test	0.079	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: LEIVA PEREZ

DATE/TIME OF TEST: 07/26/2025 22:40:35

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.172	22:35
Breath Volume	1.332	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.176	22:39
Breath Volume	2.601	22:39
Air Blank	0.000	22:39
Control Test	0.080	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SASOV,ALEX
ARRESTING OFFICER AGENCY: Ocean Ridge PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: GRANADOS GONZALEZ

DATE/TIME OF TEST: 07/27/2025 01:29:58

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.158	01:25
Breath Volume	2.433	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.149	01:28
Breath Volume	2.207	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: REED,CHARLES
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: TANKSLEY

DATE/TIME OF TEST: 07/27/2025 03:09:00

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.000	03:04
Breath Volume	1.671	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.000	03:07
Breath Volume	2.312	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:09

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CAMPBELL, GEORGE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 07/27/2025 03:57:32

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.079	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.186	03:52
Breath Volume	1.667	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.185	03:55
Breath Volume	1.792	03:55
Air Blank	0.000	03:56
Control Test	0.077	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMIGELSKI,TREVOR
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: ESCALANTE-MORALES

DATE/TIME OF TEST: 07/27/2025 04:34:32

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.078	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.133	04:29
Breath Volume	1.609	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.133	04:32
Breath Volume	1.703	04:32
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROWE,RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: ALGUACIL PEREZ

DATE/TIME OF TEST: 07/27/2025 05:51:25

Diagnostics Check	OK	05:44
Air Blank	0.000	05:45
Control Test	0.079	05:45
Air Blank	0.000	05:46
Subject Sample # 1	0.254	05:46
Breath Volume	2.937	05:46
Air Blank	0.000	05:47
Air Blank	0.000	05:49
Subject Sample # 2	0.257	05:49
Breath Volume	2.773	05:49
Air Blank	0.000	05:50
Control Test	0.079	05:50
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORTO,EVELYN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:02

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 07/27/2025 08:40:37

Diagnostics Check	OK	08:34
Air Blank	0.000	08:34
Control Test	0.079	08:34
Air Blank	0.000	08:35
Subject Sample # 1	0.098	08:36
Breath Volume	1.355	08:36
Air Blank	0.000	08:36
Air Blank	0.000	08:38
Subject Sample # 2	0.103	08:39
Breath Volume	2.078	08:39
Air Blank	0.000	08:39
Control Test	0.079	08:40
Air Blank	0.000	08:40
Diagnostics Check	OK	08:40

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SALES-LOPEZ, KEVIN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: DEVARONA

DATE/TIME OF TEST: 07/27/2025 20:19:50

Diagnostics Check	OK	20:13
Air Blank	0.000	20:14
Control Test	0.080	20:14
Air Blank	0.000	20:14
Subject Sample # 1	0.108	20:15
Breath Volume	2.691	20:15
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.112	20:18
Breath Volume	3.398	20:18
Air Blank	0.000	20:18
Control Test	0.079	20:19
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: RODRIGUEZ ALCON

DATE/TIME OF TEST: 07/28/2025 02:57:58

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.166	02:53
Breath Volume	1.519	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.163	02:56
Breath Volume	1.656	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DICKERSON,JOHN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: COLLE III

DATE/TIME OF TEST: 07/31/2025 02:27:13

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.132	02:22
Breath Volume	2.781	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.130	02:25
Breath Volume	2.464	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 08/15/2025 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024028434
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/04/2025 04:14:06

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.078	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.086	04:09
Breath Volume	2.488	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:11
Subject Sample # 2	0.091	04:12
Breath Volume	1.878	04:12
Air Blank	0.000	04:13
Control Test	0.078	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 07/11/2025 20:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: SHORT

DATE/TIME OF TEST: 07/10/2025 20:59:29

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.079	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.265	20:54
Breath Volume	2.089	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.259	20:57
Breath Volume	2.667	20:57
Air Blank	0.000	20:58
Control Test	0.077	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 07/26/2025 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTOX,STERLING,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STOKES,BRAYDEN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 07/13/2025 00:52:44

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.078	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.184	00:47
Breath Volume	2.855	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.191	00:50
Breath Volume	2.277	00:50
Air Blank	0.000	00:51
Control Test	0.077	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 07/26/2025 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON II,AUTIS,E
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: JIN,KEVIN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MANUEL

DATE/TIME OF TEST: 07/18/2025 02:11:45

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.000	02:07
Breath Volume	4.101	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.000	02:10
Breath Volume	4.851	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 07/26/2025 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES,LYNNE,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: ESCOBAR,ESTEBAN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SLADE

DATE/TIME OF TEST: 07/26/2025 22:08:38

Diagnostics Check	OK	22:01	
Air Blank	0.000	22:01	
Control Test	0.078	22:02	
Air Blank	0.000	22:02	
Subject Sample	# 1 0.188	22:03	
Breath Volume	1.562	22:03	
Air Blank	0.000	22:04	
Air Blank	0.000	22:06	
Subject Sample	# 2 0.192	22:06	
Breath Volume	1.316	22:06	
Air Blank	0.000	22:07	
Control Test	0.070	22:07	Control Outside Tolerance
Air Blank	0.000	22:08	

Data Download Date/Time: 08/27/2025 18:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES,LYNNE,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: ESCOBAR,ESTEBAN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SLADE

DATE/TIME OF TEST: 07/26/2025 22:19:01

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:11	
Control Test	0.077	22:11	
Air Blank	0.000	22:12	
Subject Sample # 1	0.189	22:14	
Breath Volume	1.558	22:14	
Air Blank	0.000	22:15	
Air Blank	0.000	22:16	
Subject Sample # 2	0.185	22:17	
Breath Volume	1.285	22:17	
Air Blank	0.000	22:18	
Control Test	0.073	22:18	Control Outside Tolerance
Air Blank	0.000	22:19	

Data Download Date/Time: 08/27/2025 18:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CATALINE

DATE/TIME OF TEST: 07/27/2025 02:25:30

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.078	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.153	02:20
Breath Volume	2.019	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.149	02:23
Breath Volume	1.316	02:23
Air Blank	0.000	02:24
Control Test	0.077	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 08/27/2025 18:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSBOURNE,DANTRELL,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MESHES,GREGORY
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/27/2025 04:29:45

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.077	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.184	04:24
Breath Volume	3.117	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.179	04:27
Breath Volume	2.613	04:27
Air Blank	0.000	04:28
Control Test	0.078	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 08/27/2025 18:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODFREY,BRANDON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GODFREY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 07/05/2025 17:53:37

Diagnostics Check	OK	17:46
Air Blank	0.000	17:47
Control Test	0.078	17:47
Air Blank	0.000	17:48
Subject Sample # 1	0.271	17:48
Breath Volume	1.160	17:48
Air Blank	0.000	17:49
Air Blank	0.000	17:51
Subject Sample # 2	0.271	17:51
Breath Volume	1.531	17:51
Air Blank	0.000	17:52
Control Test	0.075	17:53
Air Blank	0.000	17:53
Diagnostics Check	OK	17:53

Data Download Date/Time: 07/11/2025 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODFREY,BRANDON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GODFREY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/06/2025 09:28:16

Diagnostics Check	OK	09:22
Air Blank	0.000	09:22
Control Test	0.079	09:22
Air Blank	0.000	09:23
Subject Sample # 1	0.180	09:23
Breath Volume	1.859	09:23
Air Blank	0.000	09:24
Air Blank	0.000	09:26
Subject Sample # 2	0.165	09:26
Breath Volume	1.429	09:26
Air Blank	0.000	09:27
Control Test	0.076	09:27
Air Blank	0.000	09:28
Diagnostics Check	OK	09:28

Data Download Date/Time: 07/11/2025 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESTEK,JOHN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MARESCA

DATE/TIME OF TEST: 07/09/2025 18:31:39

Diagnostics Check	OK	18:20	
Air Blank	0.000	18:20	
Control Test	0.078	18:20	
Air Blank	0.000	18:21	
Subject Sample	# 1 0.000	18:24	Volume Not Met
Breath Volume	0.773	18:24	
Air Blank	0.000	18:25	
Air Blank	0.000	18:27	
Subject Sample	# 2 0.000	18:30	Volume Not Met
Breath Volume	0.300	18:30	
Air Blank	0.000	18:30	
Control Test	0.075	18:31	
Air Blank	0.000	18:31	
Diagnostics Check	OK	18:31	

Data Download Date/Time: 07/11/2025 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:01

SUBJECT'S LAST NAME: SCOGGINS

DATE/TIME OF TEST: 07/12/2025 16:28:06

Diagnostics Check	OK	16:21
Air Blank	0.000	16:22
Control Test	0.078	16:22
Air Blank	0.000	16:22
Subject Sample # 1	0.249	16:23
Breath Volume	3.308	16:23
Air Blank	0.000	16:24
Air Blank	0.000	16:25
Subject Sample # 2	0.238	16:26
Breath Volume	3.578	16:26
Air Blank	0.000	16:27
Control Test	0.075	16:27
Air Blank	0.000	16:28
Diagnostics Check	OK	16:28

Data Download Date/Time: 08/13/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: MORIN

DATE/TIME OF TEST: 07/12/2025 21:25:26

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.078	21:20
Air Blank	0.000	21:20
Subject Sample # 1	0.123	21:21
Breath Volume	1.902	21:21
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.113	21:23
Breath Volume	1.351	21:23
Air Blank	0.000	21:24
Control Test	0.077	21:24
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 08/13/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: RICKER

DATE/TIME OF TEST: 07/13/2025 04:17:18

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.077	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.000	04:13
Breath Volume	1.160	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.000	04:15
Breath Volume	1.175	04:15
Air Blank	0.000	04:16
Control Test	0.077	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 08/13/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 07/19/2025 01:15:42

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.077	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.190	01:11
Breath Volume	1.445	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.186	01:14
Breath Volume	1.371	01:14
Air Blank	0.000	01:14
Control Test	0.076	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 08/13/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: CANTOR GARCIA

DATE/TIME OF TEST: 07/19/2025 23:25:20

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:18	
Control Test	0.078	23:18	
Air Blank	0.000	23:19	
Subject Sample	# 1 0.155	23:22	Volume Not Met
Breath Volume	0.042	23:22	
Air Blank	0.000	23:22	
Air Blank	0.000	23:24	
Subject Sample	# 2	23:24	Sequence Aborted
Breath Volume	0.000	23:24	
Air Blank	0.000	23:25	

Data Download Date/Time: 08/13/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEEL,TOMMY,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BISHOP,NOLAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 07/21/2025 06:27:38

Diagnostics Check	OK	06:20
Air Blank	0.000	06:21
Control Test	0.078	06:21
Air Blank	0.000	06:22
Subject Sample # 1	0.094	06:23
Breath Volume	1.199	06:23
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample # 2	0.095	06:26
Breath Volume	1.617	06:26
Air Blank	0.000	06:26
Control Test	0.077	06:27
Air Blank	0.000	06:27
Diagnostics Check	OK	06:27

Data Download Date/Time: 08/13/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 14:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:21

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/21/2025 14:55:57

Diagnostics Check	OK	14:49
Air Blank	0.000	14:50
Control Test	0.077	14:50
Air Blank	0.000	14:50
Subject Sample # 1	0.229	14:51
Breath Volume	2.027	14:51
Air Blank	0.000	14:52
Air Blank	0.000	14:53
Subject Sample # 2	0.230	14:54
Breath Volume	1.660	14:54
Air Blank	0.000	14:55
Control Test	0.077	14:55
Air Blank	0.000	14:55
Diagnostics Check	OK	14:55

Data Download Date/Time: 08/13/2025 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 07/21/2025 19:28:51

Diagnostics Check	OK	19:16	
Air Blank	0.000	19:17	
Control Test	0.078	19:17	
Air Blank	0.000	19:17	
Subject Sample	# 1 0.166	19:21	Volume Not Met
Breath Volume	0.304	19:21	
Air Blank	0.000	19:21	
Air Blank	0.000	19:23	
Subject Sample	# 2 0.173	19:23	
Breath Volume	1.589	19:23	
Air Blank	0.000	19:24	
Air Blank	0.000	19:26	
Subject Sample	# 3 0.175	19:27	
Breath Volume	1.511	19:27	
Air Blank	0.000	19:27	
Control Test	0.076	19:28	
Air Blank	0.000	19:28	
Diagnostics Check	OK	19:28	

Data Download Date/Time: 08/13/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFORD,DERIC,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUDD,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/26/2025 00:02:54

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.078	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.273	23:57
Breath Volume	1.335	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.264	00:01
Breath Volume	1.746	00:01
Air Blank	0.000	00:01
Control Test	0.077	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 08/13/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORCE,TAYLOR,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FORCE,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: CIOBAN

DATE/TIME OF TEST: 07/27/2025 01:15:04

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.078	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.115	01:10
Breath Volume	3.042	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.116	01:13
Breath Volume	2.601	01:13
Air Blank	0.000	01:14
Control Test	0.078	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:15

Data Download Date/Time: 08/13/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: POLK

DATE/TIME OF TEST: 07/01/2025 22:18:35

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.081	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.197	22:14
Breath Volume	2.070	22:14
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.193	22:16
Breath Volume	2.015	22:16
Air Blank	0.000	22:17
Control Test	0.078	22:17
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 07/11/2025 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,RICHARD,F
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: TENNYSON,CHASE
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: MENDOZA IBARRA

DATE/TIME OF TEST: 07/04/2025 19:30:43

Diagnostics Check	OK	19:20
Air Blank	0.000	19:21
Control Test	0.081	19:21
Air Blank	0.000	19:22
Subject Sample # 1	0.210	19:22
Breath Volume	1.867	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.234	19:26
Breath Volume	2.082	19:26
Air Blank	0.000	19:26
Air Blank	0.000	19:28
Subject Sample # 3	0.228	19:28
Breath Volume	3.632	19:28
Air Blank	0.000	19:29
Control Test	0.078	19:30
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 07/11/2025 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLINGSWORTH, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HOLLINGSWORTH, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/04/2025 20:42:01

Diagnostics Check	OK	20:32	
Air Blank	0.000	20:32	
Control Test	0.080	20:33	
Air Blank	0.000	20:33	
Subject Sample	# 1 0.143	20:36	Volume Not Met
Breath Volume	0.464	20:36	
Air Blank	0.000	20:37	
Air Blank	0.000	20:39	
Subject Sample	# 2 0.140	20:40	
Breath Volume	1.417	20:40	
Air Blank		20:40	RFI Detect
Air Blank	0.000	20:41	
Control Test	0.078	20:41	
Air Blank	0.000	20:41	
Diagnostics Check	OK	20:42	

Data Download Date/Time: 07/11/2025 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLINGSWORTH, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HOLLINGSWORTH, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/04/2025 21:05:09

Diagnostics Check	OK	20:57
Air Blank	0.000	20:58
Control Test	0.080	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.137	21:00
Breath Volume	2.480	21:00
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.131	21:03
Breath Volume	2.746	21:03
Air Blank	0.000	21:04
Control Test	0.078	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 07/11/2025 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LAND,M
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: CARRIER

DATE/TIME OF TEST: 07/05/2025 17:48:32

Diagnostics Check	OK	17:45	
Air Blank	0.000	17:46	
Control Test	0.082	17:46	
Air Blank	0.000	17:47	
Subject Sample	# 1	17:47	Subject Test Refused
Breath Volume	0.000	17:47	
Air Blank	0.000	17:47	
Control Test	0.081	17:47	
Air Blank	0.000	17:48	
Diagnostics Check	OK	17:48	

Data Download Date/Time: 07/11/2025 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAHR,DANIEL,E
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: WATERS,DEVON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: BUI
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: DODD

DATE/TIME OF TEST: 07/06/2025 16:14:03

Diagnostics Check	OK	16:06
Air Blank	0.000	16:06
Control Test	0.081	16:07
Air Blank	0.000	16:07
Subject Sample # 1	0.163	16:08
Breath Volume	1.480	16:08
Air Blank	0.000	16:08
Air Blank	0.000	16:10
Subject Sample # 2	0.163	16:12
Breath Volume	1.328	16:12
Air Blank	0.000	16:13
Control Test	0.080	16:13
Air Blank	0.000	16:13
Diagnostics Check	OK	16:14

Data Download Date/Time: 07/11/2025 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAVES,TIMOTHY,F
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: CROCKER,JOSHUA
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: CASSON

DATE/TIME OF TEST: 07/08/2025 17:50:19

Diagnostics Check	OK	17:47	
Air Blank	0.000	17:47	
Control Test	0.081	17:47	
Air Blank	0.000	17:48	
Subject Sample	# 1	17:48	Subject Test Refused
Breath Volume	0.000	17:48	
Air Blank	0.000	17:49	
Control Test	0.080	17:49	
Air Blank	0.000	17:50	
Diagnostics Check	OK	17:50	

Data Download Date/Time: 07/11/2025 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 06:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: TRIBBLE

DATE/TIME OF TEST: 07/09/2025 06:44:58

Diagnostics Check	OK	06:38
Air Blank	0.000	06:38
Control Test	0.081	06:38
Air Blank	0.000	06:39
Subject Sample # 1	0.094	06:40
Breath Volume	1.996	06:40
Air Blank	0.000	06:40
Air Blank	0.000	06:42
Subject Sample # 2	0.090	06:43
Breath Volume	2.003	06:43
Air Blank	0.000	06:44
Control Test	0.081	06:44
Air Blank	0.000	06:44
Diagnostics Check	OK	06:44

Data Download Date/Time: 07/11/2025 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLENDER, MONICA, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MILLENDER, MONICA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:13

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/12/2025 14:39:57

Diagnostics Check	OK	14:33
Air Blank	0.000	14:34
Control Test	0.080	14:34
Air Blank	0.000	14:35
Subject Sample # 1	0.000	14:35
Breath Volume	3.070	14:35
Air Blank	0.000	14:36
Air Blank	0.000	14:38
Subject Sample # 2	0.000	14:38
Breath Volume	1.226	14:38
Air Blank	0.000	14:39
Control Test	0.081	14:39
Air Blank	0.000	14:39
Diagnostics Check	OK	14:39

Data Download Date/Time: 08/13/2025 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLENDER, MONICA, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MILLENDER, MONICA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: ROBY

DATE/TIME OF TEST: 07/17/2025 21:24:29

Diagnostics Check	OK	21:17
Air Blank	0.000	21:18
Control Test	0.082	21:18
Air Blank	0.000	21:19
Subject Sample # 1	0.100	21:19
Breath Volume	2.312	21:19
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.106	21:22
Breath Volume	2.343	21:22
Air Blank	0.000	21:23
Control Test	0.081	21:23
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 08/13/2025 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: SAMUELS

DATE/TIME OF TEST: 07/12/2025 02:04:23

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.080	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.207	01:58
Breath Volume	1.281	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.201	02:02
Breath Volume	1.707	02:02
Air Blank	0.000	02:03
Control Test	0.077	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 08/15/2025 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,BROOKE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: ROBLERO GODINEZ

DATE/TIME OF TEST: 07/12/2025 04:28:50

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.235	04:24
Breath Volume	1.421	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.230	04:27
Breath Volume	1.398	04:27
Air Blank	0.000	04:27
Control Test	0.078	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 08/15/2025 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: MIGUEL RIVERA

DATE/TIME OF TEST: 07/13/2025 02:49:53

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.117	02:45
Breath Volume	2.359	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.107	02:48
Breath Volume	1.750	02:48
Air Blank	0.000	02:48
Control Test	0.077	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 08/15/2025 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: VILLATORO SALMERON

DATE/TIME OF TEST: 07/13/2025 03:39:25

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.079	03:36	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:38	
Control Test	0.078	03:38	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 08/15/2025 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ESCALANTE DIAZ

DATE/TIME OF TEST: 07/13/2025 04:20:44

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.079	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.207	04:14
Breath Volume	1.843	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.204	04:18
Breath Volume	2.289	04:18
Air Blank	0.000	04:19
Control Test	0.078	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 08/15/2025 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BENNETT,JONATHAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 07/18/2025 08:55:17

Diagnostics Check	OK	08:47
Air Blank	0.000	08:47
Control Test	0.080	08:48
Air Blank	0.000	08:48
Subject Sample # 1	0.199	08:50
Breath Volume	2.226	08:50
Air Blank	0.000	08:51
Air Blank	0.000	08:53
Subject Sample # 2	0.203	08:53
Breath Volume	1.621	08:53
Air Blank	0.000	08:54
Control Test	0.076	08:54
Air Blank	0.000	08:55
Diagnostics Check	OK	08:55

Data Download Date/Time: 08/15/2025 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: OLIVO HERNANDEZ

DATE/TIME OF TEST: 07/18/2025 20:50:40

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.079	20:44
Air Blank	0.000	20:44
Subject Sample # 1	0.193	20:46
Breath Volume	2.300	20:46
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample # 2	0.185	20:49
Breath Volume	2.492	20:49
Air Blank	0.000	20:49
Control Test	0.078	20:50
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 08/15/2025 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 07/18/2025 21:26:47

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.122	21:21
Breath Volume	1.472	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.114	21:25
Breath Volume	1.824	21:25
Air Blank	0.000	21:25
Control Test	0.078	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: PERALTA

DATE/TIME OF TEST: 07/19/2025 00:46:15

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.139	00:41
Breath Volume	2.882	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:43
Subject Sample # 2	0.138	00:44
Breath Volume	1.941	00:44
Air Blank	0.000	00:45
Control Test	0.078	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,BROOKE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: VERMA

DATE/TIME OF TEST: 07/19/2025 02:21:03

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.214	02:14
Breath Volume	1.628	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.202	02:19
Breath Volume	1.890	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:21

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARABALLO,CRISTIAN
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: DELEON MEJIA

DATE/TIME OF TEST: 07/19/2025 06:25:35

Diagnostics Check	OK	06:11	
Air Blank	0.000	06:11	
Control Test	0.079	06:12	
Air Blank	0.000	06:12	
Subject Sample	# 1 0.210	06:14	
Breath Volume	1.406	06:14	
Air Blank	0.000	06:15	
Air Blank	0.000	06:17	
Subject Sample	# 2 0.174	06:20	Volume Not Met
Breath Volume	0.703	06:20	
Air Blank	0.000	06:21	
Air Blank	0.000	06:22	
Subject Sample	# 3 0.200	06:23	
Breath Volume	1.785	06:23	
Air Blank	0.000	06:24	
Control Test	0.076	06:24	
Air Blank	0.000	06:25	
Diagnostics Check	OK	06:25	

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: CASTRO LUCAS

DATE/TIME OF TEST: 07/20/2025 01:23:36

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.208	01:18
Breath Volume	1.414	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.210	01:21
Breath Volume	1.710	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: LOPRETO TZEP

DATE/TIME OF TEST: 07/20/2025 01:49:24

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.078	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.117	01:44
Breath Volume	1.472	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:46
Subject Sample # 2	0.114	01:47
Breath Volume	1.152	01:47
Air Blank	0.000	01:48
Control Test	0.078	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RENTERIA,ELIZABETH
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: SOWA

DATE/TIME OF TEST: 07/23/2025 04:25:25

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.165	04:20
Breath Volume	2.472	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.165	04:23
Breath Volume	1.609	04:23
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FARDELLA,BLAKE
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: CINTO RAMIREZ

DATE/TIME OF TEST: 07/24/2025 01:40:04

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.148	01:35
Breath Volume	1.343	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.151	01:38
Breath Volume	2.207	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:40

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HLUCH,MARIO
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:53

SUBJECT'S LAST NAME: FERGUS

DATE/TIME OF TEST: 07/24/2025 15:58:58

Diagnostics Check	OK	15:52
Air Blank	0.000	15:53
Control Test	0.079	15:53
Air Blank	0.000	15:53
Subject Sample # 1	0.011	15:54
Breath Volume	2.785	15:54
Air Blank	0.000	15:54
Air Blank	0.000	15:56
Subject Sample # 2	0.011	15:57
Breath Volume	2.863	15:57
Air Blank	0.000	15:58
Control Test	0.079	15:58
Air Blank	0.000	15:58
Diagnostics Check	OK	15:58

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/24/2025 20:10:56

Diagnostics Check	OK	19:59	
Air Blank	0.000	19:59	
Control Test	0.079	20:00	
Air Blank	0.000	20:00	
Subject Sample	# 1 0.149	20:03	Volume Not Met
Breath Volume	0.269	20:03	
Air Blank	0.000	20:04	
Air Blank	0.000	20:06	
Subject Sample	# 2	20:09	Subject Test Refused
Breath Volume	0.023	20:09	
Air Blank	0.000	20:10	
Control Test	0.077	20:10	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROBINSON, DAVIS
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: SUNJIC

DATE/TIME OF TEST: 07/25/2025 04:21:53

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.000	04:17
Breath Volume	1.289	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.000	04:20
Breath Volume	2.582	04:20
Air Blank	0.000	04:20
Control Test	0.079	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HARVEY,ANTHONY
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MUSALLAM

DATE/TIME OF TEST: 07/25/2025 23:00:37

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.080	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:56	No Sample Provided
Breath Volume	0.000	22:56	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample	# 2	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	22:59	
Control Test	0.080	23:00	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: EDWARDS,DARREL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: MACINTOSH

DATE/TIME OF TEST: 07/26/2025 17:32:23

Diagnostics Check	OK	17:25
Air Blank	0.000	17:26
Control Test	0.080	17:26
Air Blank	0.000	17:27
Subject Sample # 1	0.159	17:27
Breath Volume	2.660	17:27
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.158	17:30
Breath Volume	2.734	17:30
Air Blank	0.000	17:31
Control Test	0.079	17:31
Air Blank	0.000	17:32
Diagnostics Check	OK	17:32

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LONDONO,JULIAN
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: SAVAGE

DATE/TIME OF TEST: 07/26/2025 18:34:59

Diagnostics Check	OK	18:28
Air Blank	0.000	18:29
Control Test	0.079	18:29
Air Blank	0.000	18:29
Subject Sample # 1	0.000	18:30
Breath Volume	2.562	18:30
Air Blank	0.000	18:31
Air Blank	0.000	18:33
Subject Sample # 2	0.000	18:33
Breath Volume	2.777	18:33
Air Blank	0.000	18:34
Control Test	0.079	18:34
Air Blank	0.000	18:34
Diagnostics Check	OK	18:34

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXANDER,JACE
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: SWISS

DATE/TIME OF TEST: 07/26/2025 19:56:39

Diagnostics Check	OK	19:48
Air Blank	0.000	19:48
Control Test	0.079	19:48
Air Blank	0.000	19:49
Subject Sample # 1	0.183	19:50
Breath Volume	1.699	19:50
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.181	19:54
Breath Volume	1.835	19:54
Air Blank	0.000	19:55
Control Test	0.078	19:56
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 07/27/2025 00:09:25

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.079	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.000	00:04
Breath Volume	1.679	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.000	00:08
Breath Volume	1.523	00:08
Air Blank	0.000	00:08
Control Test	0.080	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: ALFARO

DATE/TIME OF TEST: 07/27/2025 03:10:50

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.077	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.147	03:06
Breath Volume	2.593	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.144	03:09
Breath Volume	2.660	03:09
Air Blank	0.000	03:09
Control Test	0.077	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 07/27/2025 06:38:26

Diagnostics Check	OK	06:22	
Air Blank	0.000	06:22	
Control Test	0.079	06:22	
Air Blank	0.000	06:23	
Subject Sample	# 1 0.251	06:26	Volume Not Met
Breath Volume	0.171	06:26	
Air Blank	0.000	06:27	
Air Blank	0.000	06:29	
Subject Sample	# 2 0.236	06:31	
Breath Volume	1.164	06:31	
Air Blank	0.000	06:31	
Air Blank	0.000	06:33	
Subject Sample	# 3 0.247	06:36	
Breath Volume	1.390	06:36	
Air Blank	0.000	06:37	
Control Test	0.075	06:37	
Air Blank	0.000	06:38	
Diagnostics Check	OK	06:38	

Data Download Date/Time: 08/15/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: EITEL

DATE/TIME OF TEST: 07/30/2025 05:00:49

Diagnostics Check	OK	04:54
Air Blank	0.000	04:54
Control Test	0.080	04:55
Air Blank	0.000	04:55
Subject Sample # 1	0.000	04:56
Breath Volume	1.128	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.000	04:59
Breath Volume	2.238	04:59
Air Blank	0.000	04:59
Control Test	0.079	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 08/15/2025 13:50