Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ADAMS,JUSTIN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 07/01/2025 01:18:16

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.081	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.175	01:13
Breath Volume		1.839	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.178	01:16
Breath Volume		1.269	01:16
Air Blank		0.000	01:17
Control Test		0.080	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 07/03/2025 04:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RAMIREZ,A
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MOSES

DATE/TIME OF TEST: 07/01/2025 04:03:40

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.079	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.110	03:59
Breath Volume		1.406	03:59
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.107	04:02
Breath Volume		1.257	04:02
Air Blank		0.000	04:02
Control Test		0.079	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 07/03/2025 04:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ADAMS,JUSTIN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/01/2025 04:36:57

Diagnostics Check		OK	04:29
Air Blank		0.000	04:29
Control Test		0.080	04:29
Air Blank		0.000	04:30
Subject Sample	# 1	0.137	04:32
Breath Volume		2.417	04:32
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.139	04:35
Breath Volume		2.445	04:35
Air Blank		0.000	04:36
Control Test		0.080	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 07/03/2025 04:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MALDONADO,K.J.
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 07/02/2025 02:06:47

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.000	02:02
Breath Volume		2.996	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.000	02:05
Breath Volume		2.152	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 07/03/2025 04:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SCHNEIDER,NICOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: ABDULLAH

DATE/TIME OF TEST: 07/02/2025 23:43:42

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.079	23:34
Air Blank		0.000	23:35
Subject Sample	# 1	0.096	23:38
Breath Volume		2.230	23:38
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.082	23:42
Breath Volume		1.839	23:42
Air Blank		0.000	23:42
Control Test		0.082	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 07/03/2025 04:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: MARTIN VEGA

DATE/TIME OF TEST: 07/03/2025 03:43:23

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.184	03:38
Breath Volume		1.832	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.182	03:41
Breath Volume		1.773	03:41
Air Blank		0.000	03:42
Control Test		0.080	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 07/03/2025 04:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: INOSTROZA

DATE/TIME OF TEST: 07/04/2025 00:49:53

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.080	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.144	00:45
Breath Volume		3.492	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.142	00:48
Breath Volume		2.835	00:48
Air Blank		0.000	00:48
Control Test		0.080	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 08/29/2025 04:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ROSE,MICHAEL
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/04/2025 04:49:51

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.080	04:44
Air Blank		0.000	04:44
Subject Sample	# 1	0.111	04:45
Breath Volume		1.847	04:45
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.115	04:48
Breath Volume		3.078	04:48
Air Blank		0.000	04:48
Control Test		0.080	04:49
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 08/29/2025 04:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TURNER,DYLAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: GONZALEZ BERRIOS

DATE/TIME OF TEST: 07/05/2025 02:22:59

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.080	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.091	02:18
Breath Volume		1.328	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.084	02:21
Breath Volume		1.203	02:21
Air Blank		0.000	02:22
Control Test		0.081	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 08/29/2025 04:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ADAMS,JUSTIN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 07/05/2025 02:38:21

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.079	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.097	02:34
Breath Volume		2.562	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.095	02:36
Breath Volume		2.640	02:36
Air Blank		0.000	02:37
Control Test		0.077	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 08/29/2025 04:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FARRELL,STEPHANIE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: CETOUTE

DATE/TIME OF TEST: 07/05/2025 03:39:46

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:28	
Control Test		0.079	03:29	
Air Blank		0.000	03:29	
Subject Sample	# 1	0.042	03:32	Volume Not Met
Breath Volume		0.167	03:32	
Air Blank		0.000	03:33	
Air Blank		0.000	03:35	
Subject Sample	# 2	0.042	03:38	Volume Not Met
Breath Volume		0.238	03:38	
Air Blank		0.000	03:38	
Control Test		0.083	03:39	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 08/29/2025 04:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ARMSTRONG,ISAAC
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 07/05/2025 03:52:24

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.084	03:48
Breath Volume		1.933	03:48
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.080	03:50
Breath Volume		1.300	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 08/29/2025 04:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FRITSCH,ANNIKA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: OLIVERA

DATE/TIME OF TEST: 07/05/2025 04:17:49

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.252	04:13
Breath Volume		2.117	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.249	04:16
Breath Volume		1.769	04:16
Air Blank		0.000	04:16
Control Test		0.079	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 08/29/2025 04:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ADAMS,JUSTIN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: COFFMAN-ACOSTA

DATE/TIME OF TEST: 07/05/2025 04:35:43

Diagnostics Check		OK	04:30	
Air Blank		0.000	04:31	
Control Test		0.078	04:31	
Air Blank		0.000	04:32	
Subject Sample	# 1	0.245	04:32	
Breath Volume		3.054	04:32	
Air Blank		0.000	04:33	
Air Blank			04:35	Ambient Fail
Air Blank		0.000	04:35	

Data Download Date/Time: 08/29/2025 04:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ADAMS,JUSTIN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: COFFMAN-ACOSTA

DATE/TIME OF TEST: 07/05/2025 04:43:41

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.078	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.247	04:39
Breath Volume		2.488	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.255	04:42
Breath Volume		2.570	04:42
Air Blank		0.000	04:42
Control Test		0.079	04:43
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UCELO,MEGAN,N
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GUILLEN-LARA,FERNANDO
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 10:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: WHITAKER

DATE/TIME OF TEST: 07/05/2025 11:29:27

Diagnostics Check		OK	11:23
Air Blank		0.000	11:23
Control Test		0.080	11:24
Air Blank		0.000	11:24
Subject Sample	# 1	0.000	11:25
Breath Volume		1.355	11:25
Air Blank		0.000	11:25
Air Blank		0.000	11:27
Subject Sample	# 2	0.000	11:28
Breath Volume		2.214	11:28
Air Blank		0.000	11:28
Control Test		0.080	11:28
Air Blank		0.000	11:29
Diagnostics Check		OK	11:29

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERVAIS,ROSEMARY,C
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KOEPPEL,ADAM
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 11:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: NUNNERY

DATE/TIME OF TEST: 07/05/2025 14:15:45

Diagnostics Check		OK	14:08
Air Blank		0.000	14:09
Control Test		0.080	14:09
Air Blank		0.000	14:10
Subject Sample	# 1	0.000	14:11
Breath Volume		3.285	14:11
Air Blank		0.000	14:11
Air Blank		0.000	14:13
Subject Sample	# 2	0.000	14:14
Breath Volume		4.304	14:14
Air Blank		0.000	14:14
Control Test		0.078	14:15
Air Blank		0.000	14:15
Diagnostics Check		OK	14:15

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MANNS,HUNTER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 07/06/2025 04:34:28

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:21	
Control Test		0.080	04:21	
Air Blank		0.000	04:22	
Subject Sample	# 1	0.146	04:25	Volume Not Met
Breath Volume		0.136	04:25	
Air Blank		0.000	04:26	
Air Blank		0.000	04:28	
Subject Sample	# 2	0.137	04:29	
Breath Volume		1.269	04:29	
Air Blank		0.000	04:29	
Air Blank		0.000	04:31	
Subject Sample	# 3	0.139	04:33	
Breath Volume		1.164	04:33	
Air Blank		0.000	04:33	
Control Test		0.080	04:33	
Air Blank		0.000	04:34	
Diagnostics Check		OK	04:34	

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FRITSCH,A
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: MULHOLLAND

DATE/TIME OF TEST: 07/07/2025 00:10:38

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.080	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.179	00:06
Breath Volume		3.179	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.180	00:09
Breath Volume		2.316	00:09
Air Blank		0.000	00:09
Control Test		0.080	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANDOVAL,R
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BACHILLER RODRIGUEZ

DATE/TIME OF TEST: 07/07/2025 04:25:11

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:13	
Control Test		0.079	04:13	
Air Blank		0.000	04:14	
Subject Sample	# 1	0.081	04:17	Volume Not Met
Breath Volume		0.109	04:17	
Air Blank		0.000	04:17	
Air Blank		0.000	04:19	
Subject Sample	# 2	0.088	04:20	
Breath Volume		1.718	04:20	
Air Blank		0.000	04:21	
Air Blank		0.000	04:22	
Subject Sample	# 3	0.086	04:23	
Breath Volume		1.285	04:23	
Air Blank		0.000	04:24	
Control Test		0.080	04:24	
Air Blank		0.000	04:25	
Diagnostics Check		OK	04:25	

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FRITSCH,A
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: MENDEZ ROBLERO

DATE/TIME OF TEST: 07/10/2025 00:07:57

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.080	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.210	00:02
Breath Volume		2.214	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.203	00:06
Breath Volume		2.132	00:06
Air Blank		0.000	00:07
Control Test		0.080	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LAYNE,NOAH
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: ELIAS

DATE/TIME OF TEST: 07/10/2025 02:29:49

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.059	02:24
Breath Volume		1.910	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.050	02:28
Breath Volume		1.144	02:28
Air Blank		0.000	02:28
Control Test		0.079	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KELLER,K
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/11/2025 01:03:35

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.080	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.191	00:58
Breath Volume		1.863	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.194	01:02
Breath Volume		2.046	01:02
Air Blank		0.000	01:02
Control Test		0.080	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMENS,ADAM,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KERR,ANDREW
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 08:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: GALLEGO

DATE/TIME OF TEST: 07/11/2025 11:30:41

Diagnostics Check		OK	11:24
Air Blank		0.000	11:25
Control Test		0.080	11:25
Air Blank		0.000	11:25
Subject Sample	# 1	0.161	11:26
Breath Volume		3.675	11:26
Air Blank		0.000	11:26
Air Blank		0.000	11:28
Subject Sample	# 2	0.161	11:29
Breath Volume		2.878	11:29
Air Blank		0.000	11:29
Control Test		0.079	11:30
Air Blank		0.000	11:30
Diagnostics Check		OK	11:30

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISSEY,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: DESARDEN,ANTHONY
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: MORALES JUL

DATE/TIME OF TEST: 07/13/2025 07:57:26

Diagnostics Check		OK	07:47	
Air Blank		0.000	07:47	
Control Test		0.079	07:48	
Air Blank		0.000	07:48	
Subject Sample	# 1	0.175	07:51	Volume Not Met
Breath Volume		0.468	07:51	
Air Blank		0.000	07:52	
Air Blank		0.000	07:54	
Subject Sample	# 2		07:55	Subject Test Refused
Breath Volume		1.164	07:55	
Air Blank		0.000	07:56	
Control Test		0.080	07:56	
Air Blank		0.000	07:57	
Diagnostics Check		OK	07:57	

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ADAMS,JUSTIN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HICKMAN

DATE/TIME OF TEST: 07/15/2025 03:13:08

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.181	03:08
Breath Volume		1.394	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.185	03:11
Breath Volume		1.621	03:11
Air Blank		0.000	03:12
Control Test		0.080	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KLINE,C
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 07/16/2025 23:21:08

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.080	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.145	23:16
Breath Volume		1.796	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:18
Subject Sample	# 2	0.140	23:19
Breath Volume		1.253	23:19
Air Blank		0.000	23:20
Control Test		0.080	23:20
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BERRIOS,J
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 07/17/2025 23:54:31

Diagnostics Check		OK	23:48
Air Blank		0.000	23:48
Control Test		0.079	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.115	23:50
Breath Volume		2.988	23:50
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.119	23:53
Breath Volume		2.054	23:53
Air Blank		0.000	23:53
Control Test		0.080	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SHIPLEY,D
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: RIKER

DATE/TIME OF TEST: 07/18/2025 03:10:44

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.142	03:06
Breath Volume		3.468	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.134	03:09
Breath Volume		2.933	03:09
Air Blank		0.000	03:09
Control Test		0.080	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HAYES,K
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ITURRIOZ

DATE/TIME OF TEST: 07/18/2025 04:49:54

Diagnostics Check		OK	04:35	
Air Blank		0.000	04:36	
Control Test		0.079	04:36	
Air Blank		0.000	04:37	
Subject Sample	# 1	0.216	04:40	Volume Not Met
Breath Volume		0.234	04:40	
Air Blank		0.000	04:40	
Air Blank		0.000	04:42	
Subject Sample	# 2	0.216	04:44	
Breath Volume		1.710	04:44	
Air Blank		0.000	04:45	
Air Blank		0.000	04:47	
Subject Sample	# 3	0.209	04:48	
Breath Volume		1.496	04:48	
Air Blank		0.000	04:49	
Control Test		0.081	04:49	
Air Blank		0.000	04:49	
Diagnostics Check		OK	04:49	

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BANOS,F
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: CARDEN

DATE/TIME OF TEST: 07/19/2025 00:46:49

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.079	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.134	00:42
Breath Volume		1.488	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.124	00:45
Breath Volume		1.882	00:45
Air Blank		0.000	00:45
Control Test		0.080	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FRITSCH,A
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CASTRO ANDRADE

DATE/TIME OF TEST: 07/19/2025 01:22:47

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.079	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.161	01:18
Breath Volume		1.179	01:18
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.154	01:21
Breath Volume		1.707	01:21
Air Blank		0.000	01:21
Control Test		0.081	01:22
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MALDONADO,K.J.
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BRANT

DATE/TIME OF TEST: 07/19/2025 01:55:42

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.124	01:51
Breath Volume		2.324	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.118	01:54
Breath Volume		1.746	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMENS,ADAM,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FRITSCH,A
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DE MRCO

DATE/TIME OF TEST: 07/20/2025 02:17:36

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:15	
Control Test		0.079	02:15	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:16	
Control Test		0.079	02:17	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 08/29/2025 04:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMENS,ADAM,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SENECAL,J
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 07/20/2025 03:23:18

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.245	03:18
Breath Volume		1.347	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.246	03:21
Breath Volume		1.109	03:21
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMNENS,ADAM,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LAYNE,NOAH
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: FARR

DATE/TIME OF TEST: 07/20/2025 03:48:55

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:46	
Control Test		0.079	03:46	
Air Blank		0.000	03:47	
Subject Sample	# 1		03:47	Subject Test Refused
Breath Volume		0.000	03:47	
Air Blank		0.000	03:48	
Control Test		0.080	03:48	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMENS,ADAM,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BITTIOFER,P
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: KOEPKE

DATE/TIME OF TEST: 07/20/2025 04:10:52

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.080	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.128	04:05
Breath Volume		1.410	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.133	04:09
Breath Volume		2.183	04:09
Air Blank		0.000	04:09
Control Test		0.080	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DEVERNOE,M
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MASTANDUNO

DATE/TIME OF TEST: 07/21/2025 02:31:54

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:17	
Control Test		0.080	02:18	
Air Blank		0.000	02:18	
Subject Sample	# 1	0.180	02:21	Volume Not Met
Breath Volume		1.042	02:21	
Air Blank		0.000	02:22	
Air Blank		0.000	02:24	
Subject Sample	# 2	0.185	02:26	
Breath Volume		1.378	02:26	
Air Blank		0.000	02:27	
Air Blank		0.000	02:29	
Subject Sample	# 3	0.180	02:30	
Breath Volume		1.468	02:30	
Air Blank		0.000	02:31	
Control Test		0.080	02:31	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:31	

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLUMSTEIN,R
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: FIGUEROA RUBIO

DATE/TIME OF TEST: 07/21/2025 04:17:25

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.080	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.049	04:12
Breath Volume		1.660	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.053	04:15
Breath Volume		1.355	04:15
Air Blank		0.000	04:16
Control Test		0.080	04:16
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ADAMS,JUSTIN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: ROLLE

DATE/TIME OF TEST: 07/23/2025 22:09:19

Diagnostics Check		OK	22:05	
Air Blank		0.000	22:06	
Control Test		0.079	22:06	
Air Blank		0.000	22:06	
Subject Sample	# 1		22:07	Slope Not Met
Breath Volume		1.003	22:07	
Air Blank		0.000	22:08	
Control Test		0.080	22:08	
Air Blank		0.000	22:09	
Diagnostics Check		OK	22:09	

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ADAMS,JUSTIN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: ROLLE

DATE/TIME OF TEST: 07/23/2025 22:19:08

Diagnostics Check		OK	22:12
Air Blank		0.000	22:12
Control Test		0.080	22:13
Air Blank		0.000	22:13
Subject Sample	# 1	0.311	22:14
Breath Volume		1.980	22:14
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.320	22:17
Breath Volume		2.367	22:17
Air Blank		0.000	22:18
Control Test		0.080	22:18
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CHESNUT,BRYAN
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/26/2025 02:23:53

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.080	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.114	02:19
Breath Volume		2.195	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.106	02:22
Breath Volume		1.582	02:22
Air Blank		0.000	02:22
Control Test		0.080	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CAEZ-ROJAS,J
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: MATA PADILLA

DATE/TIME OF TEST: 07/27/2025 22:00:26

Diagnostics Check		OK	21:51	
Air Blank		0.000	21:52	
Control Test		0.080	21:52	
Air Blank		0.000	21:53	
Subject Sample	# 1	0.225	21:54	
Breath Volume		2.363	21:54	
Air Blank		0.000	21:55	
Air Blank		0.000	21:56	
Subject Sample	# 2	0.204	21:57	
Breath Volume		3.378	21:57	
Air Blank		0.000	21:58	
Air Blank			21:59	Ambient Fail
Air Blank		0.000	22:00	

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CAEZ-ROJAS,J
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: MATA PADILLA

DATE/TIME OF TEST: 07/27/2025 22:09:48

Diagnostics Check		OK	22:03
Air Blank		0.000	22:03
Control Test		0.080	22:04
Air Blank		0.000	22:04
Subject Sample	# 1	0.211	22:05
Breath Volume		3.339	22:05
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.206	22:08
Breath Volume		3.859	22:08
Air Blank		0.000	22:08
Control Test		0.080	22:09
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SMITH,LOGAN
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/27/2025 22:59:09

Diagnostics Check		OK	22:52
Air Blank		0.000	22:53
Control Test		0.079	22:53
Air Blank		0.000	22:53
Subject Sample	# 1	0.084	22:54
Breath Volume		2.351	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.084	22:57
Breath Volume		2.269	22:57
Air Blank		0.000	22:58
Control Test		0.079	22:58
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RAMIREZ,ANDRES
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CINTRON

DATE/TIME OF TEST: 07/29/2025 00:14:58

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.079	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.323	00:09
Breath Volume		2.437	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.319	00:13
Breath Volume		1.968	00:13
Air Blank		0.000	00:14
Control Test		0.079	00:14
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ARMOUR,R
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 07/29/2025 04:56:01

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.079	04:49
Air Blank		0.000	04:50
Subject Sample	# 1	0.311	04:51
Breath Volume		1.855	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:53
Subject Sample	# 2	0.317	04:54
Breath Volume		2.031	04:54
Air Blank		0.000	04:55
Control Test		0.079	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:56

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ANDREWS,G
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/30/2025 05:51:28

Diagnostics Check		OK	05:37	
Air Blank		0.000	05:37	
Control Test		0.080	05:38	
Air Blank		0.000	05:38	
Subject Sample	# 1	0.086	05:41	Volume Not Met
Breath Volume		0.339	05:41	
Air Blank		0.000	05:42	
Air Blank		0.000	05:44	
Subject Sample	# 2	0.109	05:46	
Breath Volume		1.167	05:46	
Air Blank		0.000	05:46	
Air Blank		0.000	05:48	
Subject Sample	# 3	0.108	05:49	
Breath Volume		1.445	05:49	
Air Blank		0.000	05:50	
Control Test		0.082	05:50	
Air Blank		0.000	05:51	
Diagnostics Check		OK	05:51	

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ARMSTRONG,J
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: CABELL

DATE/TIME OF TEST: 07/31/2025 03:49:52

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.080	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.097	03:45
Breath Volume		1.320	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:47
Subject Sample	# 2	0.099	03:48
Breath Volume		1.355	03:48
Air Blank		0.000	03:48
Control Test		0.081	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 08/29/2025 04:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HAMM,M
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: D'AMICO

DATE/TIME OF TEST: 07/12/2025 00:19:41

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.080	00:14
Air Blank		0.000	00:14
Subject Sample	# 1	0.189	00:15
Breath Volume		1.945	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.187	00:18
Breath Volume		1.648	00:18
Air Blank		0.000	00:18
Control Test		0.080	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 07/23/2025 21:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ-SIFONTES,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: BENITEZ-SIFONTES,RAFAEL
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: ADMIN
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 07/16/2025 16:43:33

Diagnostics Check		OK	16:37
Air Blank		0.000	16:37
Control Test		0.080	16:37
Air Blank		0.000	16:38
Subject Sample	# 1	0.000	16:39
Breath Volume		3.765	16:39
Air Blank		0.000	16:39
Air Blank		0.000	16:41
Subject Sample	# 2	0.000	16:42
Breath Volume		3.753	16:42
Air Blank		0.000	16:42
Control Test		0.080	16:42
Air Blank		0.000	16:43
Diagnostics Check		OK	16:43

Data Download Date/Time: 07/23/2025 21:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSAS,CARLY,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: ROSAS,CARLY
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 06/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: DANIELSON

DATE/TIME OF TEST: 07/19/2025 19:09:15

Diagnostics Check		OK	19:03
Air Blank		0.000	19:03
Control Test		0.080	19:04
Air Blank		0.000	19:04
Subject Sample	# 1	0.000	19:05
Breath Volume		1.304	19:05
Air Blank		0.000	19:05
Air Blank		0.000	19:07
Subject Sample	# 2	0.000	19:07
Breath Volume		2.066	19:07
Air Blank		0.000	19:08
Control Test		0.081	19:08
Air Blank		0.000	19:09
Diagnostics Check		OK	19:09

Data Download Date/Time: 07/23/2025 21:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCCRANN,SEAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 07/05/2025 00:12:01

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:01	
Control Test		0.080	00:01	
Air Blank		0.000	00:01	
Subject Sample	# 1	0.136	00:05	Volume Not Met
Breath Volume		0.843	00:05	
Air Blank		0.000	00:05	
Air Blank		0.000	00:07	
Subject Sample	# 2		00:10	No Sample Provided
Breath Volume		0.000	00:10	
Air Blank		0.000	00:11	
Control Test		0.078	00:11	
Air Blank		0.000	00:11	
Diagnostics Check		OK	00:12	

Data Download Date/Time: 07/18/2025 12:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCCRANN,SEAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 07/05/2025 00:24:10

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.080	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.146	00:19
Breath Volume		1.570	00:19
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.142	00:22
Breath Volume		1.328	00:22
Air Blank		0.000	00:23
Control Test		0.077	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 07/18/2025 12:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,GABRIEL,J
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: PEREZ,GABRIEL
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: DIXON VALDEZ

DATE/TIME OF TEST: 07/10/2025 00:05:34

Diagnostics Check		OK	00:02	
Air Blank		0.000	00:03	
Control Test		0.080	00:03	
Air Blank		0.000	00:03	
Subject Sample	# 1		00:04	RFI Detect
Breath Volume		0.000	00:04	
Air Blank		0.000	00:04	
Control Test		0.080	00:04	
Air Blank		0.000	00:05	
Diagnostics Check		OK	00:05	

Data Download Date/Time: 07/18/2025 12:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,GABRIEL,J
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: PEREZ,GABRIEL
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 06/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: DIXON VALDEZ

DATE/TIME OF TEST: 07/10/2025 00:17:04

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.080	00:10
Air Blank		0.000	00:10
Subject Sample	# 1	0.142	00:12
Breath Volume		1.316	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.140	00:15
Breath Volume		1.500	00:15
Air Blank		0.000	00:16
Control Test		0.079	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:17

Data Download Date/Time: 07/18/2025 12:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRELSFORD,SEAN,M
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: STACY,C
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: BRELSFORD

DATE/TIME OF TEST: 07/26/2025 18:48:43

Diagnostics Check		OK	18:45	
Air Blank		0.000	18:46	
Control Test		0.080	18:46	
Air Blank		0.000	18:47	
Subject Sample	# 1		18:47	Subject Test Refused
Breath Volume		0.000	18:47	
Air Blank		0.000	18:47	
Control Test		0.080	18:48	
Air Blank		0.000	18:48	
Diagnostics Check		OK	18:48	

Data Download Date/Time: 08/27/2025 17:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,ERIC,G
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: CAMPOS,JARED
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: DONALDSON

DATE/TIME OF TEST: 07/29/2025 13:12:24

Diagnostics Check		OK	13:05
Air Blank		0.000	13:06
Control Test		0.080	13:06
Air Blank		0.000	13:07
Subject Sample	# 1	0.000	13:07
Breath Volume		1.343	13:07
Air Blank		0.000	13:08
Air Blank		0.000	13:10
Subject Sample	# 2	0.000	13:11
Breath Volume		1.726	13:11
Air Blank		0.000	13:11
Control Test		0.079	13:11
Air Blank		0.000	13:12
Diagnostics Check		OK	13:12

Data Download Date/Time: 08/27/2025 17:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,E,L
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: GRAHAM,E
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 07/30/2025 01:15:52

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.080	01:10
Air Blank		0.000	01:10
Subject Sample	# 1	0.125	01:11
Breath Volume		1.675	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:13
Subject Sample	# 2	0.124	01:14
Breath Volume		1.722	01:14
Air Blank		0.000	01:14
Control Test		0.080	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 08/27/2025 17:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,STEVEN,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: LAYNE,DALLAS
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: O'NEIL

DATE/TIME OF TEST: 07/30/2025 18:21:45

Diagnostics Check	OK	18:20	
Air Blank	0.000	18:20	
Control Test	0.081	18:20	
Air Blank		18:21	RFI Detect
Air Blank	0.000	18:21	

Data Download Date/Time: 08/27/2025 17:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,STEVEN,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: LAYNE,DALLAS
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: O'NEIL

DATE/TIME OF TEST: 07/30/2025 18:38:48

Diagnostics Check		OK	18:24	
Air Blank		0.000	18:25	
Control Test		0.080	18:25	
Air Blank		0.000	18:25	
Subject Sample	# 1		18:28	No Sample Provided
Breath Volume		0.000	18:28	
Air Blank		0.000	18:29	
Air Blank		0.000	18:31	
Subject Sample	# 2	0.000	18:34	
Breath Volume		1.113	18:34	
Air Blank		0.000	18:34	
Air Blank		0.000	18:36	
Subject Sample	# 3	0.000	18:37	
Breath Volume		2.140	18:37	
Air Blank		0.000	18:37	
Control Test		0.079	18:38	
Air Blank		0.000	18:38	
Diagnostics Check		OK	18:38	

Data Download Date/Time: 08/27/2025 17:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WATERS,BRODY
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 06/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4033455775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: VELASQUEZ MIRANDA

DATE/TIME OF TEST: 07/03/2025 23:43:20

Diagnostics Check		OK	23:36
Air Blank		0.000	23:36
Control Test		0.080	23:37
Air Blank		0.000	23:37
Subject Sample	# 1	0.159	23:38
Breath Volume		2.039	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.154	23:41
Breath Volume		1.558	23:41
Air Blank		0.000	23:42
Control Test		0.079	23:42
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 07/27/2025 21:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOVELL,ADAM
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 06/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4033455775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: BLACKNEY

DATE/TIME OF TEST: 07/05/2025 23:37:33

Diagnostics Check		OK	23:31
Air Blank		0.000	23:32
Control Test		0.079	23:32
Air Blank		0.000	23:32
Subject Sample	# 1	0.155	23:33
Breath Volume		1.417	23:33
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.153	23:36
Breath Volume		1.644	23:36
Air Blank		0.000	23:36
Control Test		0.079	23:36
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 07/27/2025 21:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: DAWSON,TYLER
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 06/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4033455775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 07/15/2025 22:04:20

Diagnostics Check		OK	21:58
Air Blank		0.000	21:58
Control Test		0.079	21:58
Air Blank		0.000	21:59
Subject Sample	# 1	0.261	21:59
Breath Volume		1.222	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.253	22:02
Breath Volume		1.410	22:02
Air Blank		0.000	22:03
Control Test		0.079	22:03
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 07/27/2025 21:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WATERS,BRODY
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 06/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4033455775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: PAULIN

DATE/TIME OF TEST: 07/18/2025 00:04:39

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:01	
Control Test		0.080	00:01	
Air Blank		0.000	00:02	
Subject Sample	# 1		00:03	Subject Test Refused
Breath Volume		0.000	00:03	
Air Blank		0.000	00:03	
Control Test		0.079	00:04	
Air Blank		0.000	00:04	
Diagnostics Check		OK	00:04	

Data Download Date/Time: 07/27/2025 21:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: TAYLOR,OCTAVIOUS
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 06/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4033455775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: WHITACRE

DATE/TIME OF TEST: 07/19/2025 04:23:27

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.079	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.099	04:18
Breath Volume		1.671	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.097	04:21
Breath Volume		2.085	04:21
Air Blank		0.000	04:22
Control Test		0.078	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 07/27/2025 21:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BAUTISTA,JAVIER
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 06/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4033455775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/20/2025 23:23:32

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:21	
Control Test		0.079	23:21	
Air Blank		0.000	23:22	
Subject Sample	# 1		23:22	Subject Test Refused
Breath Volume		0.000	23:22	
Air Blank		0.000	23:22	
Control Test		0.079	23:22	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 07/27/2025 21:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WATERS,BRODY
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 06/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4033455775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 07/27/2025 00:20:31

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.079	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.111	00:15
Breath Volume		1.726	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.108	00:18
Breath Volume		2.574	00:18
Air Blank		0.000	00:19
Control Test		0.079	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 07/27/2025 21:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARRIZA,DANIEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: HOOPER

DATE/TIME OF TEST: 07/01/2025 19:03:43

Diagnostics Check		OK	18:57
Air Blank		0.000	18:58
Control Test		0.081	18:58
Air Blank		0.000	18:59
Subject Sample	# 1	0.000	18:59
Breath Volume		3.691	18:59
Air Blank		0.000	18:59
Air Blank		0.000	19:01
Subject Sample	# 2	0.000	19:02
Breath Volume		4.117	19:02
Air Blank		0.000	19:02
Control Test		0.082	19:03
Air Blank		0.000	19:03
Diagnostics Check		OK	19:03

Data Download Date/Time: 07/09/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 07/01/2025 22:14:31

Diagnostics Check		OK	22:09	
Air Blank		0.000	22:09	
Control Test		0.082	22:10	
Air Blank		0.000	22:10	
Subject Sample	# 1		22:13	Subject Test Refused
Breath Volume		0.000	22:13	
Air Blank		0.000	22:13	
Control Test		0.083	22:13	
Air Blank		0.000	22:14	
Diagnostics Check		OK	22:14	

Data Download Date/Time: 07/09/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLLOWELL,JACKSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: BUNGERT

DATE/TIME OF TEST: 07/02/2025 02:46:29

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.081	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.132	02:42
Breath Volume		2.031	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.130	02:44
Breath Volume		2.179	02:44
Air Blank		0.000	02:45
Control Test		0.081	02:45
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 07/09/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PENNAMON MUHAMMAD,DANDRE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Female

OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: VENUTI

DATE/TIME OF TEST: 07/04/2025 20:26:48

Diagnostics Check		OK	20:23	
Air Blank		0.000	20:24	
Control Test		0.082	20:24	
Air Blank		0.000	20:25	
Subject Sample	# 1		20:25	Subject Test Refused
Breath Volume		0.000	20:25	
Air Blank		0.000	20:25	
Control Test		0.082	20:26	
Air Blank		0.000	20:26	
Diagnostics Check		OK	20:26	

Data Download Date/Time: 07/09/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 07/05/2025 04:32:14

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.082	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.000	04:27
Breath Volume		2.218	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.000	04:30
Breath Volume		2.878	04:30
Air Blank		0.000	04:31
Control Test		0.082	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 07/09/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,JOHN,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GUTIERREZ,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: NGO

DATE/TIME OF TEST: 07/05/2025 20:00:44

Diagnostics Check		OK	19:54
Air Blank		0.000	19:55
Control Test		0.082	19:55
Air Blank		0.000	19:55
Subject Sample	# 1	0.157	19:56
Breath Volume		1.878	19:56
Air Blank		0.000	19:56
Air Blank		0.000	19:58
Subject Sample	# 2	0.155	19:59
Breath Volume		1.890	19:59
Air Blank		0.000	19:59
Control Test		0.082	20:00
Air Blank		0.000	20:00
Diagnostics Check		OK	20:00

Data Download Date/Time: 07/09/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STOFFAN,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: BORRERO

DATE/TIME OF TEST: 07/07/2025 03:27:26

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.082	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.174	03:22
Breath Volume		1.351	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.177	03:25
Breath Volume		1.660	03:25
Air Blank		0.000	03:26
Control Test		0.082	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 07/09/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAVIS,BRITT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 11:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: LAGANA

DATE/TIME OF TEST: 07/08/2025 12:26:47

Diagnostics Check		OK	12:24	
Air Blank		0.000	12:25	
Control Test		0.082	12:25	
Air Blank		0.000	12:26	
Subject Sample	# 1		12:26	Sequence Aborted
Breath Volume		0.000	12:26	
Air Blank		0.000	12:26	

Data Download Date/Time: 07/09/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: MOCK

DATE/TIME OF TEST: 07/10/2025 03:46:40

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:43	
Control Test		0.083	03:44	
Air Blank		0.000	03:44	
Subject Sample	# 1		03:45	Subject Test Refused
Breath Volume		0.000	03:45	
Air Blank		0.000	03:45	
Control Test		0.082	03:46	
Air Blank		0.000	03:46	
Diagnostics Check		OK	03:46	

Data Download Date/Time: 08/13/2025 09:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: KRISKO

DATE/TIME OF TEST: 07/10/2025 04:34:24

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:31	
Control Test		0.081	04:31	
Air Blank		0.000	04:32	
Subject Sample	# 1		04:33	Subject Test Refused
Breath Volume		0.000	04:33	
Air Blank		0.000	04:33	
Control Test		0.082	04:33	
Air Blank		0.000	04:34	
Diagnostics Check		OK	04:34	

Data Download Date/Time: 08/13/2025 09:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROSS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: ZABALA-PARRA

DATE/TIME OF TEST: 07/10/2025 23:51:04

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.082	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.147	23:46
Breath Volume		1.660	23:46
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.147	23:49
Breath Volume		1.816	23:49
Air Blank		0.000	23:50
Control Test		0.082	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:51

Data Download Date/Time: 08/13/2025 09:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 07/11/2025 04:13:34

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:10	
Control Test		0.082	04:10	
Air Blank		0.000	04:11	
Subject Sample	# 1		04:12	Subject Test Refused
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	
Control Test		0.082	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 08/13/2025 09:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUSCH,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 08:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 07/11/2025 09:51:48

Diagnostics Check		OK	09:45
Air Blank		0.000	09:46
Control Test		0.083	09:46
Air Blank		0.000	09:47
Subject Sample	# 1	0.000	09:47
Breath Volume		3.667	09:47
Air Blank		0.000	09:47
Air Blank		0.000	09:49
Subject Sample	# 2	0.000	09:50
Breath Volume		3.136	09:50
Air Blank		0.000	09:50
Control Test		0.082	09:51
Air Blank		0.000	09:51
Diagnostics Check		OK	09:51

Data Download Date/Time: 08/13/2025 09:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JENKINS,DEVIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 14:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:56

SUBJECT'S LAST NAME: DESPAIN

DATE/TIME OF TEST: 07/14/2025 15:22:55

Diagnostics Check		OK	15:19	
Air Blank		0.000	15:20	
Control Test		0.082	15:20	
Air Blank		0.000	15:21	
Subject Sample	# 1		15:21	Subject Test Refused
Breath Volume		0.000	15:21	
Air Blank		0.000	15:22	
Control Test		0.082	15:22	
Air Blank		0.000	15:22	
Diagnostics Check		OK	15:22	

Data Download Date/Time: 08/13/2025 09:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTINO,ERIK,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CUTINO,ERIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: KERADMAN

DATE/TIME OF TEST: 07/15/2025 22:41:10

Diagnostics Check		OK	22:34
Air Blank		0.000	22:35
Control Test		0.081	22:35
Air Blank		0.000	22:36
Subject Sample	# 1	0.234	22:36
Breath Volume		2.605	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.233	22:39
Breath Volume		2.238	22:39
Air Blank		0.000	22:40
Control Test		0.080	22:40
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 08/13/2025 09:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GOSNELL

DATE/TIME OF TEST: 07/17/2025 01:44:02

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:41	
Control Test		0.081	01:41	
Air Blank		0.000	01:42	
Subject Sample	# 1		01:42	Subject Test Refused
Breath Volume		0.000	01:42	
Air Blank		0.000	01:43	
Control Test		0.082	01:43	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 08/13/2025 09:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: VONGKOMJAN

DATE/TIME OF TEST: 07/17/2025 04:31:38

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.083	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.208	04:27
Breath Volume		2.050	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.206	04:30
Breath Volume		2.160	04:30
Air Blank		0.000	04:30
Control Test		0.082	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 08/13/2025 09:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 07/17/2025 20:01:24

Diagnostics Check		OK	19:55
Air Blank		0.000	19:55
Control Test		0.082	19:56
Air Blank		0.000	19:56
Subject Sample	# 1	0.000	19:57
Breath Volume		2.289	19:57
Air Blank		0.000	19:57
Air Blank		0.000	19:59
Subject Sample	# 2	0.000	20:00
Breath Volume		1.875	20:00
Air Blank		0.000	20:00
Control Test		0.082	20:00
Air Blank		0.000	20:01
Diagnostics Check		OK	20:01

Data Download Date/Time: 08/13/2025 09:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSEN,DUSTIN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANSEN,DUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: FREYTES

DATE/TIME OF TEST: 07/18/2025 00:18:01

Diagnostics Check		OK	00:11
Air Blank		0.000	00:12
Control Test		0.082	00:12
Air Blank		0.000	00:12
Subject Sample	# 1	0.000	00:13
Breath Volume		1.410	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.000	00:16
Breath Volume		2.476	00:16
Air Blank		0.000	00:17
Control Test		0.082	00:17
Air Blank		0.000	00:17
Diagnostics Check		OK	00:18

Data Download Date/Time: 08/13/2025 09:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TAVERNARI,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/19/2025 02:06:35

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.082	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.234	02:01
Breath Volume		2.335	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.235	02:05
Breath Volume		2.156	02:05
Air Blank		0.000	02:05
Control Test		0.082	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 08/13/2025 09:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES GONZALEZ,PEDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REYES GONZALEZ,PEDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/19/2025 02:31:16

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.081	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.225	02:26
Breath Volume		2.039	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.218	02:29
Breath Volume		1.253	02:29
Air Blank		0.000	02:30
Control Test		0.082	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 08/13/2025 09:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RENAKER,DAVID
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 07/19/2025 04:02:58

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.082	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.224	03:58
Breath Volume		1.898	03:58
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.219	04:01
Breath Volume		1.832	04:01
Air Blank		0.000	04:02
Control Test		0.082	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 08/13/2025 09:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONZON GARCIA,ADEL,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MONZON GARCIA,ADEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 11:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 07/20/2025 01:15:40

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:11	
Control Test		0.083	01:12	
Air Blank		0.000	01:12	
Subject Sample	# 1		01:14	Subject Test Refused
Breath Volume		0.000	01:14	
Air Blank		0.000	01:14	
Control Test		0.082	01:15	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 08/13/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUZMAN,REINA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: GEDGAUDAS

DATE/TIME OF TEST: 07/20/2025 01:41:44

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:38	
Control Test		0.083	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:40	Subject Test Refused
Breath Volume		0.000	01:40	
Air Blank		0.000	01:40	
Control Test		0.082	01:41	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 08/13/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MESSENGER,SEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 07/20/2025 02:49:52

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.083	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.242	02:45
Breath Volume		2.492	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:47
Subject Sample	# 2	0.259	02:48
Breath Volume		2.457	02:48
Air Blank		0.000	02:48
Control Test		0.082	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 08/13/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WELCH,JAIME
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: QUEZADA MURILLO

DATE/TIME OF TEST: 07/20/2025 22:51:49

Diagnostics Check		OK	22:45
Air Blank		0.000	22:45
Control Test		0.082	22:45
Air Blank		0.000	22:46
Subject Sample	# 1	0.154	22:47
Breath Volume		3.335	22:47
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.150	22:50
Breath Volume		2.441	22:50
Air Blank		0.000	22:50
Control Test		0.082	22:51
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 08/13/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLLOWELL,JACKSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MARTINS

DATE/TIME OF TEST: 07/21/2025 02:58:30

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:56	
Control Test		0.082	02:56	
Air Blank		0.000	02:56	
Subject Sample	# 1		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:57	
Control Test		0.082	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 08/13/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 13:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: AUGER

DATE/TIME OF TEST: 07/22/2025 14:57:27

Diagnostics Check		OK	14:51
Air Blank		0.000	14:51
Control Test		0.082	14:51
Air Blank		0.000	14:52
Subject Sample	# 1	0.000	14:52
Breath Volume		2.085	14:52
Air Blank		0.000	14:53
Air Blank		0.000	14:55
Subject Sample	# 2	0.000	14:56
Breath Volume		1.675	14:56
Air Blank		0.000	14:56
Control Test		0.081	14:56
Air Blank		0.000	14:57
Diagnostics Check		OK	14:57

Data Download Date/Time: 08/13/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: MORLEY

DATE/TIME OF TEST: 07/22/2025 20:45:58

Diagnostics Check		OK	20:40
Air Blank		0.000	20:40
Control Test		0.082	20:40
Air Blank		0.000	20:41
Subject Sample	# 1	0.000	20:41
Breath Volume		2.816	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.000	20:44
Breath Volume		2.421	20:44
Air Blank		0.000	20:45
Control Test		0.081	20:45
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 08/13/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 07/23/2025 04:14:08

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:11	
Control Test		0.082	04:11	
Air Blank		0.000	04:12	
Subject Sample	# 1		04:12	Subject Test Refused
Breath Volume		0.000	04:12	
Air Blank		0.000	04:13	
Control Test		0.082	04:13	
Air Blank		0.000	04:14	
Diagnostics Check		OK	04:14	

Data Download Date/Time: 08/13/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEAL II,LOUIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: NELSON II

DATE/TIME OF TEST: 07/23/2025 18:55:53

Diagnostics Check		OK	18:52	
Air Blank		0.000	18:53	
Control Test		0.082	18:53	
Air Blank		0.000	18:54	
Subject Sample	# 1		18:54	Subject Test Refused
Breath Volume		0.000	18:54	
Air Blank		0.000	18:54	
Control Test		0.082	18:55	
Air Blank		0.000	18:55	
Diagnostics Check		OK	18:55	

Data Download Date/Time: 08/13/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLLOWELL,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: RIVERA-LEANDRY

DATE/TIME OF TEST: 07/24/2025 00:36:55

Diagnostics Check		OK	00:28
Air Blank		0.000	00:29
Control Test		0.082	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.190	00:31
Breath Volume		1.585	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:33
Subject Sample	# 2	0.188	00:35
Breath Volume		2.164	00:35
Air Blank		0.000	00:36
Control Test		0.080	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 08/13/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NUSSBAUM,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/24/2025 05:03:55

Diagnostics Check		OK	04:50	
Air Blank		0.000	04:51	
Control Test		0.083	04:51	
Air Blank		0.000	04:52	
Subject Sample	# 1	0.165	04:55	Volume Not Met
Breath Volume		0.566	04:55	
Air Blank		0.000	04:55	
Air Blank		0.000	04:57	
Subject Sample	# 2	0.170	04:58	
Breath Volume		2.019	04:58	
Air Blank		0.000	04:59	
Air Blank		0.000	05:01	
Subject Sample	# 3	0.166	05:02	
Breath Volume		2.117	05:02	
Air Blank		0.000	05:03	
Control Test		0.081	05:03	
Air Blank		0.000	05:03	
Diagnostics Check		OK	05:03	

Data Download Date/Time: 08/13/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ANDERSON,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: KIRKMAN

DATE/TIME OF TEST: 07/24/2025 23:07:17

Diagnostics Check		OK	23:04	
Air Blank		0.000	23:04	
Control Test		0.082	23:05	
Air Blank		0.000	23:05	
Subject Sample	# 1		23:05	Subject Test Refused
Breath Volume		0.000	23:05	
Air Blank		0.000	23:06	
Control Test		0.082	23:06	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 08/13/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK,ML,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YANICK,ML
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: EVA

DATE/TIME OF TEST: 07/25/2025 05:04:56

Diagnostics Check		OK	05:02	
Air Blank		0.000	05:02	
Control Test		0.082	05:02	
Air Blank		0.000	05:03	
Subject Sample	# 1		05:03	Subject Test Refused
Breath Volume		0.000	05:03	
Air Blank		0.000	05:04	
Control Test		0.082	05:04	
Air Blank		0.000	05:04	
Diagnostics Check		OK	05:04	

Data Download Date/Time: 08/13/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAPES,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: DARNOLD

DATE/TIME OF TEST: 07/25/2025 18:23:21

Diagnostics Check		OK	18:17
Air Blank		0.000	18:17
Control Test		0.082	18:17
Air Blank		0.000	18:18
Subject Sample	# 1	0.074	18:18
Breath Volume		2.113	18:18
Air Blank		0.000	18:19
Air Blank		0.000	18:21
Subject Sample	# 2	0.078	18:21
Breath Volume		2.179	18:21
Air Blank		0.000	18:22
Control Test		0.082	18:22
Air Blank		0.000	18:23
Diagnostics Check		OK	18:23

Data Download Date/Time: 08/13/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRINGTON,CODY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FARRINGTON,CODY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/25/2025 23:17:37

Diagnostics Check		OK	23:10
Air Blank		0.000	23:10
Control Test		0.083	23:10
Air Blank		0.000	23:11
Subject Sample	# 1	0.249	23:12
Breath Volume		1.375	23:12
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.246	23:16
Breath Volume		1.542	23:16
Air Blank		0.000	23:16
Control Test		0.081	23:17
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 08/13/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUHLMAIER,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 07/26/2025 00:55:43

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.082	00:47
Air Blank		0.000	00:47
Subject Sample	# 1	0.124	00:49
Breath Volume		1.484	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.116	00:54
Breath Volume		1.558	00:54
Air Blank		0.000	00:54
Control Test		0.081	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 08/13/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VALENTI,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 07/26/2025 19:41:41

Diagnostics Check		OK	19:35
Air Blank		0.000	19:35
Control Test		0.082	19:36
Air Blank		0.000	19:36
Subject Sample	# 1	0.011	19:37
Breath Volume		1.156	19:37
Air Blank		0.000	19:37
Air Blank		0.000	19:39
Subject Sample	# 2	0.011	19:40
Breath Volume		1.699	19:40
Air Blank		0.000	19:40
Control Test		0.082	19:41
Air Blank		0.000	19:41
Diagnostics Check		OK	19:41

Data Download Date/Time: 08/13/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BOATENG,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: LUNCEFORD

DATE/TIME OF TEST: 07/27/2025 19:37:29

Diagnostics Check		OK	19:27
Air Blank		0.000	19:28
Control Test		0.082	19:28
Air Blank		0.000	19:28
Subject Sample	# 1	0.189	19:31
Breath Volume		1.472	19:31
Air Blank		0.000	19:31
Air Blank		0.000	19:33
Subject Sample	# 2	0.186	19:35
Breath Volume		1.523	19:35
Air Blank		0.000	19:36
Control Test		0.080	19:36
Air Blank		0.000	19:37
Diagnostics Check		OK	19:37

Data Download Date/Time: 08/13/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/27/2025 23:22:16

Diagnostics Check		OK	23:15
Air Blank		0.000	23:15
Control Test		0.083	23:15
Air Blank		0.000	23:16
Subject Sample	# 1	0.114	23:17
Breath Volume		1.492	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.122	23:20
Breath Volume		1.269	23:20
Air Blank		0.000	23:21
Control Test		0.081	23:21
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 08/13/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KIRKENDALL,JESSIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 07/28/2025 20:38:37

Diagnostics Check		OK	20:32
Air Blank		0.000	20:32
Control Test		0.082	20:33
Air Blank		0.000	20:33
Subject Sample	# 1	0.113	20:34
Breath Volume		1.656	20:34
Air Blank		0.000	20:34
Air Blank		0.000	20:36
Subject Sample	# 2	0.107	20:37
Breath Volume		1.785	20:37
Air Blank		0.000	20:37
Control Test		0.081	20:38
Air Blank		0.000	20:38
Diagnostics Check		OK	20:38

Data Download Date/Time: 08/13/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,S,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAZIME,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MCNEW

DATE/TIME OF TEST: 07/29/2025 23:23:18

Diagnostics Check		OK	23:16
Air Blank		0.000	23:17
Control Test		0.082	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.000	23:18
Breath Volume		3.781	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.000	23:21
Breath Volume		2.468	23:21
Air Blank		0.000	23:22
Control Test		0.082	23:22
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 08/13/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNISTER,STEPHEN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BANNISTER,STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: TOBIN

DATE/TIME OF TEST: 07/30/2025 20:00:51

Diagnostics Check		OK	19:52
Air Blank		0.000	19:52
Control Test		0.082	19:53
Air Blank		0.000	19:53
Subject Sample	# 1	0.000	19:55
Breath Volume		1.804	19:55
Air Blank		0.000	19:55
Air Blank		0.000	19:57
Subject Sample	# 2	0.000	19:59
Breath Volume		1.515	19:59
Air Blank		0.000	19:59
Control Test		0.081	20:00
Air Blank		0.000	20:00
Diagnostics Check		OK	20:00

Data Download Date/Time: 08/13/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ALVAREZ,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: HAMMOCK

DATE/TIME OF TEST: 07/31/2025 00:20:20

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:17	
Control Test		0.083	00:18	
Air Blank		0.000	00:18	
Subject Sample	# 1		00:18	Subject Test Refused
Breath Volume		0.000	00:18	
Air Blank		0.000	00:19	
Control Test		0.082	00:19	
Air Blank		0.000	00:20	
Diagnostics Check		OK	00:20	

Data Download Date/Time: 08/13/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LOCKE,ALEXIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: RIVERA JR

DATE/TIME OF TEST: 07/31/2025 05:18:09

Diagnostics Check		OK	05:11
Air Blank		0.000	05:11
Control Test		0.082	05:12
Air Blank		0.000	05:12
Subject Sample	# 1	0.000	05:13
Breath Volume		2.324	05:13
Air Blank		0.000	05:13
Air Blank		0.000	05:15
Subject Sample	# 2	0.000	05:16
Breath Volume		1.828	05:16
Air Blank		0.000	05:17
Control Test		0.081	05:17
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 08/13/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 06:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: GONZALEZ RIVERA

DATE/TIME OF TEST: 07/31/2025 07:53:06

Diagnostics Check		OK	07:45
Air Blank		0.000	07:46
Control Test		0.082	07:46
Air Blank		0.000	07:47
Subject Sample	# 1	0.206	07:48
Breath Volume		1.867	07:48
Air Blank		0.000	07:49
Air Blank		0.000	07:51
Subject Sample	# 2	0.206	07:51
Breath Volume		1.789	07:51
Air Blank		0.000	07:52
Control Test		0.081	07:52
Air Blank		0.000	07:53
Diagnostics Check		OK	07:53

Data Download Date/Time: 08/13/2025 09:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GIARDINO,CARL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/01/2025 02:12:14

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.080	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.218	02:07
Breath Volume		2.906	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.233	02:10
Breath Volume		3.148	02:10
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 07/09/2025 10:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/05/2025 04:36:37

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:34	
Control Test		0.080	04:34	
Air Blank		0.000	04:35	
Subject Sample	# 1		04:35	Subject Test Refused
Breath Volume		0.000	04:35	
Air Blank		0.000	04:35	
Control Test		0.080	04:36	
Air Blank		0.000	04:36	
Diagnostics Check		OK	04:36	

Data Download Date/Time: 07/09/2025 10:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,SCOTT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CRYDERMAN,KEVIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: CASTLEBERRY

DATE/TIME OF TEST: 07/05/2025 07:41:10

Diagnostics Check		OK	07:36	
Air Blank		0.000	07:37	
Control Test		0.080	07:37	
Air Blank		0.000	07:38	
Subject Sample	# 1		07:39	Subject Test Refused
Breath Volume		0.000	07:39	
Air Blank		0.000	07:40	
Control Test		0.081	07:40	
Air Blank		0.000	07:41	
Diagnostics Check		OK	07:41	

Data Download Date/Time: 07/09/2025 10:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: BURR

DATE/TIME OF TEST: 07/07/2025 05:54:52

Diagnostics Check		OK	05:48
Air Blank		0.000	05:48
Control Test		0.079	05:48
Air Blank		0.000	05:49
Subject Sample	# 1	0.211	05:49
Breath Volume		2.542	05:49
Air Blank		0.000	05:50
Air Blank		0.000	05:52
Subject Sample	# 2	0.206	05:53
Breath Volume		2.507	05:53
Air Blank		0.000	05:53
Control Test		0.080	05:54
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 07/09/2025 10:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENTLEY,TYLER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BENTLEY,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: ELING

DATE/TIME OF TEST: 07/07/2025 20:29:12

Diagnostics Check		OK	20:26	
Air Blank		0.000	20:26	
Control Test		0.080	20:27	
Air Blank		0.000	20:27	
Subject Sample	# 1		20:27	Subject Test Refused
Breath Volume		0.000	20:27	
Air Blank		0.000	20:28	
Control Test		0.079	20:28	
Air Blank		0.000	20:29	
Diagnostics Check		OK	20:29	

Data Download Date/Time: 07/09/2025 10:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSTER,ROBERT,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROSTER,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: STRAIT

DATE/TIME OF TEST: 07/07/2025 20:41:52

Diagnostics Check		OK	20:35
Air Blank		0.000	20:36
Control Test		0.080	20:36
Air Blank		0.000	20:37
Subject Sample	# 1	0.237	20:37
Breath Volume		2.195	20:37
Air Blank		0.000	20:38
Air Blank		0.000	20:39
Subject Sample	# 2	0.241	20:40
Breath Volume		2.277	20:40
Air Blank		0.000	20:40
Control Test		0.079	20:41
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 07/09/2025 10:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DRISH,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 16:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: THIBAULT

DATE/TIME OF TEST: 07/10/2025 17:17:07

Diagnostics Check		OK	17:10
Air Blank		0.000	17:11
Control Test		0.080	17:11
Air Blank		0.000	17:11
Subject Sample	# 1	0.324	17:12
Breath Volume		1.988	17:12
Air Blank		0.000	17:13
Air Blank		0.000	17:15
Subject Sample	# 2	0.327	17:15
Breath Volume		1.382	17:15
Air Blank		0.000	17:16
Control Test		0.079	17:16
Air Blank		0.000	17:17
Diagnostics Check		OK	17:17

Data Download Date/Time: 08/13/2025 10:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSTER,ROBERT,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROSTER,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 13:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:09

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 07/11/2025 15:35:11

Diagnostics Check		OK	15:32	
Air Blank		0.000	15:32	
Control Test		0.080	15:33	
Air Blank		0.000	15:33	
Subject Sample	# 1		15:33	Subject Test Refused
Breath Volume		0.000	15:33	
Air Blank		0.000	15:34	
Control Test		0.080	15:34	
Air Blank		0.000	15:35	
Diagnostics Check		OK	15:35	

Data Download Date/Time: 08/13/2025 10:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/15/2025 05:39:49

Diagnostics Check		OK	05:31
Air Blank		0.000	05:32
Control Test		0.081	05:32
Air Blank		0.000	05:33
Subject Sample	# 1	0.133	05:33
Breath Volume		1.132	05:33
Air Blank		0.000	05:34
Air Blank		0.000	05:36
Subject Sample	# 2	0.133	05:38
Breath Volume		1.558	05:38
Air Blank		0.000	05:38
Control Test		0.080	05:39
Air Blank		0.000	05:39
Diagnostics Check		OK	05:39

Data Download Date/Time: 08/13/2025 10:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JENKINS,DEVIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: TOKARZ

DATE/TIME OF TEST: 07/15/2025 16:58:53

Diagnostics Check		OK	16:56	
Air Blank		0.000	16:56	
Control Test		0.080	16:56	
Air Blank		0.000	16:57	
Subject Sample	# 1		16:57	Subject Test Refused
Breath Volume		0.000	16:57	
Air Blank		0.000	16:57	
Control Test		0.080	16:58	
Air Blank		0.000	16:58	
Diagnostics Check		OK	16:58	

Data Download Date/Time: 08/13/2025 10:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/16/2025 05:24:16

Diagnostics Check		OK	05:13
Air Blank		0.000	05:14
Control Test		0.080	05:14
Air Blank		0.000	05:14
Subject Sample	# 1	0.252	05:15
Breath Volume		2.773	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:17
Subject Sample	# 2	0.223	05:18
Breath Volume		2.234	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:21
Subject Sample	# 3	0.260	05:22
Breath Volume		3.445	05:22
Air Blank		0.000	05:23
Control Test		0.080	05:23
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 08/13/2025 10:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEELER,B,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEELER,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 15:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 07/16/2025 16:00:28

Diagnostics Check		OK	15:54
Air Blank		0.000	15:55
Control Test		0.079	15:55
Air Blank		0.000	15:55
Subject Sample	# 1	0.000	15:56
Breath Volume		1.277	15:56
Air Blank		0.000	15:56
Air Blank		0.000	15:58
Subject Sample	# 2	0.000	15:59
Breath Volume		2.730	15:59
Air Blank		0.000	15:59
Control Test		0.080	15:59
Air Blank		0.000	16:00
Diagnostics Check		OK	16:00

Data Download Date/Time: 08/13/2025 10:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARRIT,KEVIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: YOVAN

DATE/TIME OF TEST: 07/19/2025 02:51:58

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:48	
Control Test		0.079	02:49	
Air Blank		0.000	02:49	
Subject Sample	# 1		02:50	Subject Test Refused
Breath Volume		0.000	02:50	
Air Blank		0.000	02:51	
Control Test		0.079	02:51	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 08/13/2025 10:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORING,BLAINE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,HOWARD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: SHANNON

DATE/TIME OF TEST: 07/20/2025 07:42:52

Diagnostics Check		OK	07:35
Air Blank		0.000	07:36
Control Test		0.080	07:36
Air Blank		0.000	07:37
Subject Sample	# 1	0.225	07:37
Breath Volume		4.058	07:37
Air Blank		0.000	07:38
Air Blank		0.000	07:40
Subject Sample	# 2	0.220	07:41
Breath Volume		4.476	07:41
Air Blank		0.000	07:41
Control Test		0.079	07:42
Air Blank		0.000	07:42
Diagnostics Check		OK	07:42

Data Download Date/Time: 08/13/2025 10:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 11:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: BURK

DATE/TIME OF TEST: 07/25/2025 12:12:37

Diagnostics Check		OK	12:06
Air Blank		0.000	12:06
Control Test		0.079	12:06
Air Blank		0.000	12:07
Subject Sample	# 1	0.000	12:08
Breath Volume		1.992	12:08
Air Blank		0.000	12:08
Air Blank		0.000	12:10
Subject Sample	# 2	0.000	12:11
Breath Volume		2.128	12:11
Air Blank		0.000	12:11
Control Test		0.080	12:12
Air Blank		0.000	12:12
Diagnostics Check		OK	12:12

Data Download Date/Time: 08/13/2025 10:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/25/2025 23:47:06

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.000	23:42
Breath Volume		2.332	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.000	23:45
Breath Volume		3.449	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 08/13/2025 10:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEWS,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 07/27/2025 21:46:26

Diagnostics Check		OK	21:43	
Air Blank		0.000	21:44	
Control Test		0.080	21:44	
Air Blank		0.000	21:44	
Subject Sample	# 1		21:45	Subject Test Refused
Breath Volume		0.000	21:45	
Air Blank		0.000	21:45	
Control Test		0.079	21:45	
Air Blank		0.000	21:46	
Diagnostics Check		OK	21:46	

Data Download Date/Time: 08/13/2025 10:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARRIZA,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: BIRD

DATE/TIME OF TEST: 07/29/2025 02:46:03

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.079	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.180	02:41
Breath Volume		2.218	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.181	02:44
Breath Volume		2.308	02:44
Air Blank		0.000	02:45
Control Test		0.079	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:46

Data Download Date/Time: 08/13/2025 10:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: JENNINGS JR

DATE/TIME OF TEST: 07/29/2025 04:04:22

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.140	03:59
Breath Volume		1.832	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.139	04:02
Breath Volume		2.250	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 08/13/2025 10:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: BARLOW

DATE/TIME OF TEST: 07/08/2025 14:22:22

Diagnostics Check		OK	14:16
Air Blank		0.000	14:16
Control Test		0.079	14:17
Air Blank		0.000	14:17
Subject Sample	# 1	0.000	14:18
Breath Volume		1.250	14:18
Air Blank		0.000	14:18
Air Blank		0.000	14:20
Subject Sample	# 2	0.000	14:20
Breath Volume		1.113	14:20
Air Blank		0.000	14:21
Control Test		0.078	14:21
Air Blank		0.000	14:22
Diagnostics Check		OK	14:22

Data Download Date/Time: 08/06/2025 16:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 07/11/2025 10:12:46

Diagnostics Check		OK	10:06
Air Blank		0.000	10:07
Control Test		0.079	10:07
Air Blank		0.000	10:08
Subject Sample	# 1	0.000	10:08
Breath Volume		2.046	10:08
Air Blank		0.000	10:08
Air Blank		0.000	10:10
Subject Sample	# 2	0.000	10:11
Breath Volume		2.093	10:11
Air Blank		0.000	10:11
Control Test		0.078	10:12
Air Blank		0.000	10:12
Diagnostics Check		OK	10:12

Data Download Date/Time: 08/06/2025 16:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: VANCOTT

DATE/TIME OF TEST: 07/20/2025 00:55:53

Diagnostics Check		OK	00:49
Air Blank		0.000	00:50
Control Test		0.079	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.130	00:51
Breath Volume		1.648	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:53
Subject Sample	# 2	0.125	00:54
Breath Volume		1.617	00:54
Air Blank		0.000	00:54
Control Test		0.078	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 08/06/2025 16:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 07/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/24/2025 08:46:23

Diagnostics Check		OK	08:40
Air Blank		0.000	08:40
Control Test		0.079	08:41
Air Blank		0.000	08:41
Subject Sample	# 1	0.000	08:42
Breath Volume		1.738	08:42
Air Blank		0.000	08:42
Air Blank		0.000	08:44
Subject Sample	# 2	0.000	08:45
Breath Volume		2.273	08:45
Air Blank		0.000	08:45
Control Test		0.079	08:45
Air Blank		0.000	08:46
Diagnostics Check		OK	08:46

Data Download Date/Time: 08/06/2025 16:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: WEBER,BRENDEN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 07/05/2025 00:48:03

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:40	
Control Test		0.077	00:40	
Air Blank		0.000	00:40	
Subject Sample	# 1	0.242	00:44	Volume Not Met
Breath Volume		0.066	00:44	
Air Blank		0.000	00:44	
Air Blank		0.000	00:46	
Subject Sample	# 2		00:46	Subject Test Refused
Breath Volume		0.000	00:46	
Air Blank		0.000	00:47	
Control Test		0.073	00:47	Control Outside Tolerance
Air Blank		0.000	00:48	

Data Download Date/Time: 07/13/2025 10:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,DYLAN,T
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ARNUDTS,GEORGE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: GATTRELL

DATE/TIME OF TEST: 07/06/2025 03:47:44

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.084	03:43
Breath Volume		3.179	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.088	03:46
Breath Volume		3.210	03:46
Air Blank		0.000	03:46
Control Test		0.078	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 07/13/2025 10:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: STANTON,EILEEN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/07/2025 03:27:27

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.079	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.130	03:22
Breath Volume		4.183	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.121	03:25
Breath Volume		2.289	03:25
Air Blank		0.000	03:26
Control Test		0.078	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 07/13/2025 10:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MATHEWS,ARNOLD
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: BRAVO

DATE/TIME OF TEST: 07/13/2025 03:28:46

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.080	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.186	03:24
Breath Volume		2.511	03:24
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.183	03:27
Breath Volume		2.296	03:27
Air Blank		0.000	03:27
Control Test		0.079	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 07/13/2025 10:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: O'CONNELL,WILLIAM
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 07/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 07/19/2025 03:08:32

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.081	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.110	03:03
Breath Volume		3.785	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.107	03:06
Breath Volume		3.960	03:06
Air Blank		0.000	03:07
Control Test		0.078	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 08/22/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: KAPLAN,MICHAEL
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 07/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 07/20/2025 22:56:53

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.080	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.252	22:51
Breath Volume		2.730	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.245	22:55
Breath Volume		2.386	22:55
Air Blank		0.000	22:55
Control Test		0.078	22:56
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 08/22/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,JUSTIN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: FRANCA SANTOS,RAFAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: WADLEIGH

DATE/TIME OF TEST: 07/22/2025 17:50:32

Diagnostics Check		OK	17:39
Air Blank		0.000	17:40
Control Test		0.080	17:40
Air Blank		0.000	17:41
Subject Sample	# 1	0.172	17:41
Breath Volume		1.105	17:41
Air Blank		0.000	17:42
Air Blank		0.000	17:44
Subject Sample	# 2	0.207	17:45
Breath Volume		1.343	17:45
Air Blank		0.000	17:46
Air Blank		0.000	17:47
Subject Sample	# 3	0.174	17:48
Breath Volume		1.742	17:48
Air Blank		0.000	17:49
Control Test		0.078	17:49
Air Blank		0.000	17:50
Diagnostics Check		OK	17:50

Data Download Date/Time: 08/22/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MATHEWS,ARNOLD
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 07/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: JESS

DATE/TIME OF TEST: 07/26/2025 22:15:32

Diagnostics Check		OK	22:07
Air Blank		0.000	22:07
Control Test		0.080	22:08
Air Blank		0.000	22:08
Subject Sample	# 1	0.227	22:09
Breath Volume		2.085	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.216	22:13
Breath Volume		1.367	22:13
Air Blank		0.000	22:14
Control Test		0.078	22:14
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 08/22/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CLEMENTS,ZACHARY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 07/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: SCHUSTER

DATE/TIME OF TEST: 07/28/2025 03:49:55

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.080	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.181	03:45
Breath Volume		3.171	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.172	03:48
Breath Volume		1.917	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 08/22/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: BRANUM

DATE/TIME OF TEST: 07/18/2025 19:46:32

Diagnostics Check		OK	19:40
Air Blank		0.000	19:40
Control Test		0.078	19:41
Air Blank		0.000	19:41
Subject Sample	# 1	0.195	19:42
Breath Volume		1.796	19:42
Air Blank		0.000	19:42
Air Blank		0.000	19:44
Subject Sample	# 2	0.196	19:44
Breath Volume		1.687	19:44
Air Blank		0.000	19:45
Control Test		0.078	19:45
Air Blank		0.000	19:46
Diagnostics Check		OK	19:46

Data Download Date/Time: 07/28/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BRIGGS

DATE/TIME OF TEST: 07/18/2025 22:28:12

Diagnostics Check		OK	22:22
Air Blank		0.000	22:22
Control Test		0.078	22:22
Air Blank		0.000	22:23
Subject Sample	# 1	0.175	22:23
Breath Volume		1.878	22:23
Air Blank		0.000	22:24
Air Blank		0.000	22:26
Subject Sample	# 2	0.165	22:26
Breath Volume		2.101	22:26
Air Blank		0.000	22:27
Control Test		0.078	22:27
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 07/28/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:38

SUBJECT'S LAST NAME: SCHAWRTZ

DATE/TIME OF TEST: 07/22/2025 14:23:09

Diagnostics Check		OK	14:16
Air Blank		0.000	14:16
Control Test		0.078	14:16
Air Blank		0.000	14:17
Subject Sample	# 1	0.088	14:18
Breath Volume		1.601	14:18
Air Blank		0.000	14:18
Air Blank		0.000	14:20
Subject Sample	# 2	0.079	14:21
Breath Volume		1.308	14:21
Air Blank		0.000	14:22
Control Test		0.078	14:22
Air Blank		0.000	14:23
Diagnostics Check		OK	14:23

Data Download Date/Time: 07/28/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: CONSENT,CONSENT
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 07/23/2025 21:46:09

Diagnostics Check		OK	21:38
Air Blank		0.000	21:39
Control Test		0.078	21:39
Air Blank		0.000	21:40
Subject Sample	# 1	0.109	21:41
Breath Volume		2.785	21:41
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.111	21:44
Breath Volume		4.109	21:44
Air Blank		0.000	21:45
Control Test		0.078	21:45
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 07/28/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: GOBAO

DATE/TIME OF TEST: 07/25/2025 12:23:01

Diagnostics Check		OK	12:17
Air Blank		0.000	12:17
Control Test		0.079	12:18
Air Blank		0.000	12:18
Subject Sample	# 1	0.000	12:18
Breath Volume		1.828	12:18
Air Blank		0.000	12:19
Air Blank		0.000	12:21
Subject Sample	# 2	0.000	12:21
Breath Volume		1.820	12:21
Air Blank		0.000	12:22
Control Test		0.079	12:22
Air Blank		0.000	12:22
Diagnostics Check		OK	12:23

Data Download Date/Time: 07/28/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: ENGELMEYER

DATE/TIME OF TEST: 07/10/2025 14:36:58

Diagnostics Check		OK	14:30
Air Blank		0.000	14:30
Control Test		0.079	14:30
Air Blank		0.000	14:31
Subject Sample	# 1	0.264	14:32
Breath Volume		1.187	14:32
Air Blank		0.000	14:33
Air Blank		0.000	14:34
Subject Sample	# 2	0.271	14:35
Breath Volume		1.285	14:35
Air Blank		0.000	14:36
Control Test		0.077	14:36
Air Blank		0.000	14:36
Diagnostics Check		OK	14:36

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: NEELD

DATE/TIME OF TEST: 07/11/2025 02:04:08

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.079	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.138	01:59
Breath Volume		1.617	01:59
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.133	02:02
Breath Volume		1.546	02:02
Air Blank		0.000	02:03
Control Test		0.077	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: BROUGHTON

DATE/TIME OF TEST: 07/11/2025 04:04:00

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.079	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:02	Subject Test Refused
Breath Volume		0.000	04:02	
Air Blank		0.000	04:03	
Control Test		0.079	04:03	
Air Blank		0.000	04:03	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: KRIEGL

DATE/TIME OF TEST: 07/12/2025 04:01:23

Diagnostics Check		OK	03:54
Air Blank		0.000	03:55
Control Test		0.079	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.107	03:56
Breath Volume		3.578	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.105	03:59
Breath Volume		3.253	03:59
Air Blank		0.000	04:00
Control Test		0.078	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HASKEW,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: GUZZETTI

DATE/TIME OF TEST: 07/12/2025 06:23:12

Diagnostics Check		OK	06:16
Air Blank		0.000	06:17
Control Test		0.079	06:17
Air Blank		0.000	06:17
Subject Sample	# 1	0.021	06:18
Breath Volume		1.242	06:18
Air Blank		0.000	06:19
Air Blank		0.000	06:21
Subject Sample	# 2	0.021	06:21
Breath Volume		1.421	06:21
Air Blank		0.000	06:22
Control Test		0.079	06:22
Air Blank		0.000	06:23
Diagnostics Check		OK	06:23

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SEARS,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GOODRIDGE

DATE/TIME OF TEST: 07/13/2025 00:37:51

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.079	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.117	00:33
Breath Volume		1.527	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:35
Subject Sample	# 2	0.120	00:36
Breath Volume		1.820	00:36
Air Blank		0.000	00:36
Control Test		0.078	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: RAMIREZ LASCANO

DATE/TIME OF TEST: 07/13/2025 22:17:41

Diagnostics Check		OK	22:11
Air Blank		0.000	22:11
Control Test		0.078	22:12
Air Blank		0.000	22:12
Subject Sample	# 1	0.157	22:13
Breath Volume		3.457	22:13
Air Blank		0.000	22:13
Air Blank		0.000	22:15
Subject Sample	# 2	0.159	22:16
Breath Volume		3.597	22:16
Air Blank		0.000	22:16
Control Test		0.078	22:17
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: VIJIL RODRIGUEZ

DATE/TIME OF TEST: 07/14/2025 01:51:33

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.076	01:47
Breath Volume		1.617	01:47
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.073	01:49
Breath Volume		1.648	01:49
Air Blank		0.000	01:50
Control Test		0.078	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: RADFORD

DATE/TIME OF TEST: 07/17/2025 00:57:33

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.079	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.204	00:52
Breath Volume		2.230	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.212	00:55
Breath Volume		1.406	00:55
Air Blank		0.000	00:56
Control Test		0.079	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/17/2025 21:41:19

Diagnostics Check		OK	21:29	
Air Blank		0.000	21:30	
Control Test		0.079	21:30	
Air Blank		0.000	21:31	
Subject Sample	# 1		21:34	No Sample Provided
Breath Volume		0.000	21:34	
Air Blank		0.000	21:34	
Air Blank		0.000	21:36	
Subject Sample	# 2		21:39	No Sample Provided
Breath Volume		0.000	21:39	
Air Blank		0.000	21:40	
Control Test		0.080	21:40	
Air Blank		0.000	21:41	
Diagnostics Check		OK	21:41	

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Neptune Beach PD
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: FUNEZ HERNANDEZ

DATE/TIME OF TEST: 07/18/2025 05:04:16

Diagnostics Check		OK	04:57
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.202	04:59
Breath Volume		1.667	04:59
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.189	05:02
Breath Volume		1.585	05:02
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: GUERRIER

DATE/TIME OF TEST: 07/19/2025 00:07:22

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:04	
Control Test		0.079	00:05	
Air Blank		0.000	00:05	
Subject Sample	# 1		00:05	Subject Test Refused
Breath Volume		0.000	00:05	
Air Blank		0.000	00:06	
Control Test		0.080	00:06	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PENDRY

DATE/TIME OF TEST: 07/19/2025 01:51:53

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.079	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.148	01:47
Breath Volume		1.605	01:47
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.143	01:50
Breath Volume		1.484	01:50
Air Blank		0.000	01:50
Control Test		0.079	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,S,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SATHER,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 07/19/2025 18:37:30

Diagnostics Check		OK	18:30
Air Blank		0.000	18:30
Control Test		0.079	18:30
Air Blank		0.000	18:31
Subject Sample	# 1	0.254	18:32
Breath Volume		2.601	18:32
Air Blank		0.000	18:33
Air Blank		0.000	18:35
Subject Sample	# 2	0.252	18:35
Breath Volume		2.910	18:35
Air Blank		0.000	18:36
Control Test		0.078	18:36
Air Blank		0.000	18:37
Diagnostics Check		OK	18:37

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 07/20/2025 01:16:31

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.079	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.307	01:08
Breath Volume		2.429	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.277	01:11
Breath Volume		1.230	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 3	0.291	01:14
Breath Volume		1.933	01:14
Air Blank		0.000	01:15
Control Test		0.078	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 07/24/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 07/20/2025 03:27:48

Diagnostics Check		OK	03:24	
Air Blank		0.000	03:25	
Control Test		0.079	03:25	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:26	Subject Test Refused
Breath Volume		0.000	03:26	
Air Blank		0.000	03:26	
Control Test		0.079	03:27	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:27	

Data Download Date/Time: 07/24/2025 15:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: GUERRERO GARCIA

DATE/TIME OF TEST: 07/20/2025 20:05:34

Diagnostics Check		OK	19:58
Air Blank		0.000	19:59
Control Test		0.079	19:59
Air Blank		0.000	19:59
Subject Sample	# 1	0.223	20:00
Breath Volume		1.878	20:00
Air Blank		0.000	20:01
Air Blank		0.000	20:03
Subject Sample	# 2	0.231	20:03
Breath Volume		1.960	20:03
Air Blank		0.000	20:04
Control Test		0.078	20:04
Air Blank		0.000	20:05
Diagnostics Check		OK	20:05

Data Download Date/Time: 07/24/2025 15:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: SHEETS

DATE/TIME OF TEST: 07/21/2025 01:32:59

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.079	01:27
Air Blank		0.000	01:28
Subject Sample	# 1	0.073	01:28
Breath Volume		3.417	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:30
Subject Sample	# 2	0.073	01:31
Breath Volume		3.925	01:31
Air Blank		0.000	01:32
Control Test		0.079	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 07/24/2025 15:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BERNARD,O
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 07/22/2025 00:59:49

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:57	
Control Test		0.079	00:57	
Air Blank		0.000	00:58	
Subject Sample	# 1		00:58	Subject Test Refused
Breath Volume		0.000	00:58	
Air Blank		0.000	00:58	
Control Test		0.079	00:59	
Air Blank		0.000	00:59	
Diagnostics Check		OK	00:59	

Data Download Date/Time: 07/24/2025 15:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/24/2025 03:04:06

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.116	02:58
Breath Volume		1.281	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.108	03:02
Breath Volume		1.347	03:02
Air Blank		0.000	03:03
Control Test		0.079	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 07/24/2025 15:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:09

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 07/24/2025 06:33:00

Diagnostics Check		OK	06:30	
Air Blank		0.000	06:30	
Control Test		0.079	06:30	
Air Blank		0.000	06:31	
Subject Sample	# 1		06:31	Subject Test Refused
Breath Volume		0.000	06:31	
Air Blank		0.000	06:32	
Control Test		0.079	06:32	
Air Blank		0.000	06:32	
Diagnostics Check		OK	06:33	

Data Download Date/Time: 07/24/2025 15:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,R,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 07/26/2025 02:14:22

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:11	
Control Test		0.079	02:12	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:13	
Control Test		0.079	02:13	
Air Blank		0.000	02:14	
Diagnostics Check		OK	02:14	

Data Download Date/Time: 08/21/2025 15:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 07/28/2025 02:38:14

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.076	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.227	02:33
Breath Volume		1.292	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:35
Subject Sample	# 2	0.235	02:36
Breath Volume		1.769	02:36
Air Blank		0.000	02:37
Control Test		0.078	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 08/21/2025 15:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUKIANOV,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: LUKIANOV,ALEXANDER
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MONJE

DATE/TIME OF TEST: 07/20/2025 00:13:36

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.079	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.118	00:08
Breath Volume		3.164	00:08
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.117	00:12
Breath Volume		3.113	00:12
Air Blank		0.000	00:12
Control Test		0.079	00:13
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 08/27/2025 11:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SZE

DATE/TIME OF TEST: 07/02/2025 03:58:38

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.081	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.091	03:53
Breath Volume		1.886	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.092	03:56
Breath Volume		1.480	03:56
Air Blank		0.000	03:57
Control Test		0.079	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 07/10/2025 14:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: WEEKS

DATE/TIME OF TEST: 07/05/2025 04:20:35

Diagnostics Check		OK	04:14
Air Blank		0.000	04:14
Control Test		0.081	04:14
Air Blank		0.000	04:15
Subject Sample	# 1	0.141	04:15
Breath Volume		2.601	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.135	04:18
Breath Volume		2.535	04:18
Air Blank		0.000	04:19
Control Test		0.079	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 07/10/2025 14:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,I
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: MESSA

DATE/TIME OF TEST: 07/05/2025 05:45:25

Diagnostics Check		OK	05:35
Air Blank		0.000	05:36
Control Test		0.081	05:36
Air Blank		0.000	05:37
Subject Sample	# 1	0.151	05:39
Breath Volume		1.238	05:39
Air Blank		0.000	05:40
Air Blank		0.000	05:41
Subject Sample	# 2	0.136	05:43
Breath Volume		1.343	05:43
Air Blank		0.000	05:44
Control Test		0.077	05:44
Air Blank		0.000	05:45
Diagnostics Check		OK	05:45

Data Download Date/Time: 07/10/2025 14:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/07/2025 03:53:40

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:51	
Control Test		0.080	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 1		03:52	Subject Test Refused
Breath Volume		0.000	03:52	
Air Blank		0.000	03:52	
Control Test		0.080	03:53	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 07/10/2025 14:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: RIGDON

DATE/TIME OF TEST: 07/11/2025 02:35:13

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.081	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.193	02:30
Breath Volume		3.781	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 2	0.189	02:33
Breath Volume		3.855	02:33
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 07/23/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PEMBROKE

DATE/TIME OF TEST: 07/12/2025 03:26:22

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.080	03:15	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:18	No Sample Provided
Breath Volume		0.000	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 2	0.120	03:24	Volume Not Met
Breath Volume		0.261	03:24	
Air Blank		0.000	03:25	
Control Test		0.081	03:25	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 07/23/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PEMBROKE

DATE/TIME OF TEST: 07/12/2025 03:40:45

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:29	
Control Test		0.079	03:29	
Air Blank		0.000	03:30	
Subject Sample	# 1	0.149	03:33	Volume Not Met
Breath Volume		0.183	03:33	
Air Blank		0.000	03:34	
Air Blank		0.000	03:35	
Subject Sample	# 2	0.109	03:38	Volume Not Met
Breath Volume		0.414	03:38	
Air Blank		0.000	03:39	
Control Test		0.075	03:40	
Air Blank		0.000	03:40	
Diagnostics Check		OK	03:40	

Data Download Date/Time: 07/23/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: CONE

DATE/TIME OF TEST: 07/12/2025 23:01:59

Diagnostics Check		OK	22:54
Air Blank		0.000	22:55
Control Test		0.080	22:55
Air Blank		0.000	22:56
Subject Sample	# 1	0.109	22:57
Breath Volume		1.812	22:57
Air Blank		0.000	22:58
Air Blank		0.000	22:59
Subject Sample	# 2	0.108	23:00
Breath Volume		1.531	23:00
Air Blank		0.000	23:01
Control Test		0.075	23:01
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 07/23/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: VILLANUEVA FLORES

DATE/TIME OF TEST: 07/14/2025 03:40:08

Diagnostics Check		OK	03:37	
Air Blank		0.000	03:37	
Control Test		0.080	03:38	
Air Blank		0.000	03:38	
Subject Sample	# 1		03:38	Subject Test Refused
Breath Volume		0.000	03:38	
Air Blank		0.000	03:39	
Control Test		0.080	03:39	
Air Blank		0.000	03:40	
Diagnostics Check		OK	03:40	

Data Download Date/Time: 07/23/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LEE,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: BRUGGNER

DATE/TIME OF TEST: 07/15/2025 20:04:52

Diagnostics Check		OK	20:01	
Air Blank		0.000	20:02	
Control Test		0.081	20:02	
Air Blank		0.000	20:03	
Subject Sample	# 1		20:03	Subject Test Refused
Breath Volume		0.000	20:03	
Air Blank		0.000	20:03	
Control Test		0.081	20:04	
Air Blank		0.000	20:04	
Diagnostics Check		OK	20:04	

Data Download Date/Time: 07/23/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: STRAYER

DATE/TIME OF TEST: 07/17/2025 00:46:48

Diagnostics Check		OK	00:43	
Air Blank		0.000	00:44	
Control Test		0.081	00:44	
Air Blank		0.000	00:45	
Subject Sample	# 1		00:45	Subject Test Refused
Breath Volume		0.000	00:45	
Air Blank		0.000	00:45	
Control Test		0.080	00:46	
Air Blank		0.000	00:46	
Diagnostics Check		OK	00:46	

Data Download Date/Time: 07/23/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ALEXANDRE,D
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 07/18/2025 02:09:35

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:07	
Control Test		0.081	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:08	
Control Test		0.081	02:08	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 07/23/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVN',J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HAMMETT

DATE/TIME OF TEST: 07/20/2025 01:24:46

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:22	
Control Test		0.082	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:23	
Control Test		0.080	01:24	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 07/23/2025 12:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: RAMO

DATE/TIME OF TEST: 07/20/2025 21:18:28

Diagnostics Check		OK	21:08
Air Blank		0.000	21:08
Control Test		0.081	21:09
Air Blank		0.000	21:09
Subject Sample	# 1	0.232	21:10
Breath Volume		1.925	21:10
Air Blank		0.000	21:11
Air Blank		0.000	21:13
Subject Sample	# 2	0.203	21:13
Breath Volume		1.394	21:13
Air Blank		0.000	21:14
Air Blank		0.000	21:16
Subject Sample	# 3	0.211	21:16
Breath Volume		1.746	21:16
Air Blank		0.000	21:17
Control Test		0.077	21:17
Air Blank		0.000	21:18
Diagnostics Check		OK	21:18

Data Download Date/Time: 07/23/2025 12:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: BARBOSA DA SILVA

DATE/TIME OF TEST: 07/21/2025 01:06:05

Diagnostics Check		OK	01:03	
Air Blank		0.000	01:03	
Control Test		0.081	01:03	
Air Blank		0.000	01:04	
Subject Sample	# 1		01:04	Subject Test Refused
Breath Volume		0.000	01:04	
Air Blank		0.000	01:05	
Control Test		0.081	01:05	
Air Blank		0.000	01:06	
Diagnostics Check		OK	01:06	

Data Download Date/Time: 07/23/2025 12:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: ASHLEY

DATE/TIME OF TEST: 07/22/2025 00:01:05

Diagnostics Check		OK	23:54
Air Blank		0.000	23:55
Control Test		0.081	23:55
Air Blank		0.000	23:56
Subject Sample	# 1	0.115	23:56
Breath Volume		1.183	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.110	23:59
Breath Volume		1.871	23:59
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 07/23/2025 12:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: PRETZELL

DATE/TIME OF TEST: 07/23/2025 01:11:39

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.081	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.115	01:06
Breath Volume		2.261	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.117	01:09
Breath Volume		2.765	01:09
Air Blank		0.000	01:10
Control Test		0.079	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 07/23/2025 12:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MATTIACE,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: TYSON

DATE/TIME OF TEST: 07/02/2025 01:16:32

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.175	01:11
Breath Volume		2.195	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.173	01:15
Breath Volume		2.039	01:15
Air Blank		0.000	01:15
Control Test		0.078	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BUI,L
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: HOLLIS

DATE/TIME OF TEST: 07/02/2025 02:14:12

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.079	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.142	02:09
Breath Volume		2.714	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.143	02:12
Breath Volume		2.472	02:12
Air Blank		0.000	02:13
Control Test		0.079	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 07/03/2025 01:39:38

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:31	
Control Test		0.079	01:31	
Air Blank		0.000	01:32	
Subject Sample	# 1		01:35	No Sample Provided
Breath Volume		0.000	01:35	
Air Blank		0.000	01:35	
Air Blank		0.000	01:37	
Subject Sample	# 2		01:38	Subject Test Refused
Breath Volume		0.000	01:38	
Air Blank		0.000	01:38	
Control Test		0.079	01:39	
Air Blank		0.000	01:39	
Diagnostics Check		OK	01:39	

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ERNEST,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HUGGINS,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:39

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 07/03/2025 13:09:42

Diagnostics Check		OK	13:06	
Air Blank		0.000	13:06	
Control Test		0.078	13:06	
Air Blank		0.000	13:07	
Subject Sample	# 1		13:08	Subject Test Refused
Breath Volume		0.000	13:08	
Air Blank		0.000	13:08	
Control Test		0.079	13:09	
Air Blank		0.000	13:09	
Diagnostics Check		OK	13:09	

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ROY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GUTHRIE

DATE/TIME OF TEST: 07/05/2025 02:29:15

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.079	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.045	02:24
Breath Volume		1.578	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.043	02:27
Breath Volume		1.378	02:27
Air Blank		0.000	02:28
Control Test		0.079	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ROY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HAUSLER,T
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: BUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: KERR

DATE/TIME OF TEST: 07/05/2025 19:20:27

Diagnostics Check		OK	19:14
Air Blank		0.000	19:14
Control Test		0.079	19:15
Air Blank		0.000	19:15
Subject Sample	# 1	0.053	19:16
Breath Volume		2.253	19:16
Air Blank		0.000	19:16
Air Blank		0.000	19:18
Subject Sample	# 2	0.048	19:19
Breath Volume		2.242	19:19
Air Blank		0.000	19:19
Control Test		0.079	19:19
Air Blank		0.000	19:20
Diagnostics Check		OK	19:20

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ROY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: NASWORTH,MATTHEW
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: BUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 07/05/2025 22:41:03

Diagnostics Check		OK	22:34
Air Blank		0.000	22:34
Control Test		0.079	22:35
Air Blank		0.000	22:35
Subject Sample	# 1	0.094	22:36
Breath Volume		2.132	22:36
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.098	22:39
Breath Volume		2.792	22:39
Air Blank		0.000	22:40
Control Test		0.079	22:40
Air Blank		0.000	22:40
Diagnostics Check		OK	22:41

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ROY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GANN,J
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: SANCHEZ AVILA

DATE/TIME OF TEST: 07/06/2025 05:49:59

Diagnostics Check		OK	05:43
Air Blank		0.000	05:43
Control Test		0.078	05:44
Air Blank		0.000	05:44
Subject Sample	# 1	0.184	05:45
Breath Volume		3.109	05:45
Air Blank		0.000	05:46
Air Blank		0.000	05:47
Subject Sample	# 2	0.183	05:48
Breath Volume		3.246	05:48
Air Blank		0.000	05:49
Control Test		0.078	05:49
Air Blank		0.000	05:49
Diagnostics Check		OK	05:49

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 07/08/2025 19:28:04

Diagnostics Check		OK	19:18	
Air Blank		0.000	19:18	
Control Test		0.079	19:19	
Air Blank		0.000	19:19	
Subject Sample	# 1	0.262	19:22	Volume Not Met
Breath Volume		0.066	19:22	
Air Blank		0.000	19:23	
Air Blank		0.000	19:25	
Subject Sample	# 2		19:26	Subject Test Refused
Breath Volume		0.000	19:26	
Air Blank		0.000	19:27	
Control Test		0.078	19:27	
Air Blank		0.000	19:27	
Diagnostics Check		OK	19:28	

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/12/2025 03:40:27

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.079	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.127	03:35
Breath Volume		2.746	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.117	03:38
Breath Volume		2.351	03:38
Air Blank		0.000	03:39
Control Test		0.079	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CORNEJO ZEPEDA

DATE/TIME OF TEST: 07/12/2025 21:39:47

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:38	
Control Test	0.079	21:39	
Air Blank		21:39	RFI Detect
Air Blank	0.000	21:39	

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CORNEJOZEPEDA

DATE/TIME OF TEST: 07/12/2025 21:49:21

Diagnostics Check		OK	21:42
Air Blank		0.000	21:43
Control Test		0.079	21:43
Air Blank		0.000	21:43
Subject Sample	# 1	0.267	21:44
Breath Volume		2.761	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.257	21:47
Breath Volume		1.789	21:47
Air Blank		0.000	21:48
Control Test		0.079	21:48
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 07/13/2025 02:52:09

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:44	
Control Test		0.080	02:44	
Air Blank		0.000	02:44	
Subject Sample	# 1		02:48	No Sample Provided
Breath Volume		0.000	02:48	
Air Blank		0.000	02:48	
Air Blank		0.000	02:50	
Subject Sample	# 2		02:50	Subject Test Refused
Breath Volume		0.000	02:50	
Air Blank		0.000	02:51	
Control Test		0.080	02:51	
Air Blank		0.000	02:52	
Diagnostics Check		OK	02:52	

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: ALVARENGA

DATE/TIME OF TEST: 07/13/2025 04:00:35

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:52	
Control Test		0.080	03:52	
Air Blank		0.000	03:53	
Subject Sample	# 1		03:56	No Sample Provided
Breath Volume		0.000	03:56	
Air Blank		0.000	03:56	
Air Blank		0.000	03:58	
Subject Sample	# 2		03:59	Subject Test Refused
Breath Volume		0.000	03:59	
Air Blank		0.000	03:59	
Control Test		0.079	04:00	
Air Blank		0.000	04:00	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: BUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: HICKOX

DATE/TIME OF TEST: 07/13/2025 19:11:38

Diagnostics Check		OK	19:03	
Air Blank		0.000	19:03	
Control Test		0.079	19:03	
Air Blank		0.000	19:04	
Subject Sample	# 1		19:07	No Sample Provided
Breath Volume		0.000	19:07	
Air Blank		0.000	19:08	
Air Blank		0.000	19:10	
Subject Sample	# 2		19:10	Subject Test Refused
Breath Volume		0.000	19:10	
Air Blank		0.000	19:10	
Control Test		0.079	19:11	
Air Blank		0.000	19:11	
Diagnostics Check		OK	19:11	

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: DEVANEY

DATE/TIME OF TEST: 07/17/2025 04:09:28

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.147	04:04
Breath Volume		4.199	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.148	04:08
Breath Volume		3.441	04:08
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BURKES

DATE/TIME OF TEST: 07/18/2025 03:14:44

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.144	03:10
Breath Volume		1.953	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.147	03:13
Breath Volume		2.128	03:13
Air Blank		0.000	03:13
Control Test		0.080	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILSON,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: BAYE

DATE/TIME OF TEST: 07/19/2025 02:45:35

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:37	
Control Test		0.079	02:37	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:41	No Sample Provided
Breath Volume		0.000	02:41	
Air Blank		0.000	02:41	
Air Blank		0.000	02:43	
Subject Sample	# 2		02:44	Subject Test Refused
Breath Volume		0.000	02:44	
Air Blank		0.000	02:44	
Control Test		0.079	02:45	
Air Blank		0.000	02:45	
Diagnostics Check		OK	02:45	

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KELLER,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: STARLING,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:14

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 07/19/2025 13:46:08

Diagnostics Check		OK	13:40
Air Blank		0.000	13:40
Control Test		0.080	13:40
Air Blank		0.000	13:41
Subject Sample	# 1	0.000	13:41
Breath Volume		3.765	13:41
Air Blank		0.000	13:42
Air Blank		0.000	13:44
Subject Sample	# 2	0.000	13:44
Breath Volume		3.871	13:44
Air Blank		0.000	13:45
Control Test		0.080	13:45
Air Blank		0.000	13:46
Diagnostics Check		OK	13:46

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/20/2025 21:48:38

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:36	
Control Test		0.079	21:36	
Air Blank		0.000	21:36	
Subject Sample	# 1	0.129	21:38	
Breath Volume		1.335	21:38	
Air Blank		0.000	21:38	
Air Blank		0.000	21:40	
Subject Sample	# 2		21:44	No Sample Provided
Breath Volume		0.000	21:44	
Air Blank		0.000	21:44	
Air Blank		0.000	21:46	
Subject Sample	# 3	0.128	21:47	
Breath Volume		1.359	21:47	
Air Blank		0.000	21:47	
Control Test		0.079	21:48	
Air Blank		0.000	21:48	
Diagnostics Check		OK	21:48	

Data Download Date/Time: 07/29/2025 18:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JORDAN,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: KERR

DATE/TIME OF TEST: 07/20/2025 22:25:12

Diagnostics Check		OK	22:16	
Air Blank		0.000	22:16	
Control Test		0.079	22:17	
Air Blank		0.000	22:17	
Subject Sample	# 1		22:20	No Sample Provided
Breath Volume		0.000	22:20	
Air Blank		0.000	22:21	
Air Blank		0.000	22:23	
Subject Sample	# 2		22:23	Subject Test Refused
Breath Volume		0.000	22:23	
Air Blank		0.000	22:24	
Control Test		0.079	22:24	
Air Blank		0.000	22:25	
Diagnostics Check		OK	22:25	

Data Download Date/Time: 07/29/2025 18:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BOIVIN,N
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SEPULREDA

DATE/TIME OF TEST: 07/22/2025 03:44:28

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.079	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.186	03:39
Breath Volume		2.242	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.179	03:42
Breath Volume		1.476	03:42
Air Blank		0.000	03:43
Control Test		0.079	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 07/29/2025 18:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CORNELIUS

DATE/TIME OF TEST: 07/22/2025 19:46:59

Diagnostics Check		OK	19:40
Air Blank		0.000	19:40
Control Test		0.078	19:41
Air Blank		0.000	19:41
Subject Sample	# 1	0.239	19:42
Breath Volume		1.539	19:42
Air Blank		0.000	19:43
Air Blank		0.000	19:44
Subject Sample	# 2	0.255	19:45
Breath Volume		2.019	19:45
Air Blank		0.000	19:46
Control Test		0.078	19:46
Air Blank		0.000	19:46
Diagnostics Check		OK	19:46

Data Download Date/Time: 07/29/2025 18:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KELLER,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HYSLER,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: DORF

DATE/TIME OF TEST: 07/23/2025 11:11:45

Diagnostics Check		OK	11:05
Air Blank		0.000	11:05
Control Test		0.079	11:05
Air Blank		0.000	11:06
Subject Sample	# 1	0.135	11:07
Breath Volume		3.921	11:07
Air Blank		0.000	11:07
Air Blank		0.000	11:09
Subject Sample	# 2	0.136	11:10
Breath Volume		3.996	11:10
Air Blank		0.000	11:10
Control Test		0.079	11:11
Air Blank		0.000	11:11
Diagnostics Check		OK	11:11

Data Download Date/Time: 07/29/2025 18:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILSON,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: WILBER

DATE/TIME OF TEST: 07/26/2025 01:43:19

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.080	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.072	01:38
Breath Volume		1.734	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.071	01:41
Breath Volume		1.792	01:41
Air Blank		0.000	01:42
Control Test		0.079	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 07/29/2025 18:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: FRANTZ

DATE/TIME OF TEST: 07/26/2025 03:12:23

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.079	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.000	03:08
Breath Volume		2.074	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.000	03:11
Breath Volume		2.328	03:11
Air Blank		0.000	03:11
Control Test		0.079	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 07/29/2025 18:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: ADDISON

DATE/TIME OF TEST: 07/27/2025 03:16:17

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.094	03:11
Breath Volume		1.183	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.095	03:14
Breath Volume		1.882	03:14
Air Blank		0.000	03:15
Control Test		0.079	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 07/29/2025 18:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: CORDOVA GOMEZ

DATE/TIME OF TEST: 07/27/2025 04:50:03

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.080	04:43
Air Blank		0.000	04:44
Subject Sample	# 1	0.128	04:45
Breath Volume		2.140	04:45
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.128	04:48
Breath Volume		2.125	04:48
Air Blank		0.000	04:49
Control Test		0.079	04:49
Air Blank		0.000	04:49
Diagnostics Check		OK	04:50

Data Download Date/Time: 07/29/2025 18:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,ROBERT,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: VANCE,NICHOLAS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ERTEL

DATE/TIME OF TEST: 07/02/2025 03:37:35

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.078	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.159	03:32
Breath Volume		3.632	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.160	03:36
Breath Volume		3.695	03:36
Air Blank		0.000	03:36
Control Test		0.078	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 07/25/2025 13:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: ROSENBERGER

DATE/TIME OF TEST: 07/26/2025 03:22:29

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.079	03:17	
Air Blank		0.000	03:17	
Subject Sample	# 1	0.166	03:18	
Breath Volume		1.453	03:18	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 2		03:21	Subject Test Refused
Breath Volume		0.000	03:21	
Air Blank		0.000	03:21	
Control Test		0.078	03:21	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 08/29/2025 13:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/16/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: HASSEY

DATE/TIME OF TEST: 07/06/2025 01:42:18

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.077	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.122	01:37
Breath Volume		3.113	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.121	01:40
Breath Volume		3.187	01:40
Air Blank		0.000	01:41
Control Test		0.077	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 07/10/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PEACE

DATE/TIME OF TEST: 07/06/2025 03:27:31

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.143	03:22
Breath Volume		1.851	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.142	03:25
Breath Volume		1.199	03:25
Air Blank		0.000	03:26
Control Test		0.078	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 07/10/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 06/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 07/03/2025 00:43:05

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.078	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.097	00:38
Breath Volume		4.707	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.090	00:41
Breath Volume		1.718	00:41
Air Blank		0.000	00:42
Control Test		0.078	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 07/16/2025 12:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 06/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: KRAMER

DATE/TIME OF TEST: 07/10/2025 02:41:15

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.077	02:35
Air Blank		0.000	02:36
Subject Sample	# 1	0.262	02:36
Breath Volume		2.523	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.267	02:39
Breath Volume		3.230	02:39
Air Blank		0.000	02:40
Control Test		0.079	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 07/16/2025 12:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 06/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 07/13/2025 00:55:08

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.077	00:49
Air Blank		0.000	00:50
Subject Sample	# 1	0.169	00:50
Breath Volume		1.933	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.170	00:53
Breath Volume		2.050	00:53
Air Blank		0.000	00:54
Control Test		0.078	00:54
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 07/16/2025 12:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 07/19/2025 23:52:48

Diagnostics Check		OK	23:46
Air Blank		0.000	23:47
Control Test		0.078	23:47
Air Blank		0.000	23:48
Subject Sample	# 1	0.000	23:48
Breath Volume		3.042	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.000	23:51
Breath Volume		3.277	23:51
Air Blank		0.000	23:51
Control Test		0.078	23:52
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 08/07/2025 13:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LISS,K,C
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: RABLE,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: INTELLI

DATE/TIME OF TEST: 07/27/2025 01:46:30

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.079	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.287	01:42
Breath Volume		2.562	01:42
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.284	01:44
Breath Volume		2.609	01:44
Air Blank		0.000	01:45
Control Test		0.078	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 08/16/2025 09:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 07/03/2025 02:17:47

Diagnostics Check		OK	02:11
Air Blank		0.000	02:12
Control Test		0.080	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.011	02:13
Breath Volume		1.753	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.012	02:16
Breath Volume		1.546	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 07/09/2025 12:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EULER,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 15:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:21

SUBJECT'S LAST NAME: JOSEPHIK

DATE/TIME OF TEST: 07/08/2025 16:07:26

Diagnostics Check		OK	16:00
Air Blank		0.000	16:01
Control Test		0.079	16:01
Air Blank		0.000	16:01
Subject Sample	# 1	0.000	16:02
Breath Volume		1.644	16:02
Air Blank		0.000	16:02
Air Blank		0.000	16:04
Subject Sample	# 2	0.000	16:06
Breath Volume		1.730	16:06
Air Blank		0.000	16:06
Control Test		0.078	16:06
Air Blank		0.000	16:07
Diagnostics Check		OK	16:07

Data Download Date/Time: 07/09/2025 12:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHRAMM,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: BURGUENO ABREU

DATE/TIME OF TEST: 07/08/2025 22:20:48

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:18	
Control Test		0.079	22:18	
Air Blank		0.000	22:19	
Subject Sample	# 1		22:19	Subject Test Refused
Breath Volume		0.000	22:19	
Air Blank		0.000	22:19	
Control Test		0.079	22:20	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:20	

Data Download Date/Time: 07/09/2025 12:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YANICK,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: HUSSAIN

DATE/TIME OF TEST: 07/12/2025 01:51:13

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.119	01:46
Breath Volume		1.175	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.131	01:49
Breath Volume		2.417	01:49
Air Blank		0.000	01:50
Control Test		0.078	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 08/13/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: PELAEZ

DATE/TIME OF TEST: 07/12/2025 21:27:46

Diagnostics Check		OK	21:24	
Air Blank		0.000	21:25	
Control Test		0.078	21:25	
Air Blank		0.000	21:26	
Subject Sample	# 1		21:26	Subject Test Refused
Breath Volume		0.000	21:26	
Air Blank		0.000	21:26	
Control Test		0.079	21:27	
Air Blank		0.000	21:27	
Diagnostics Check		OK	21:27	

Data Download Date/Time: 08/13/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHRAMM,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BUNN

DATE/TIME OF TEST: 07/13/2025 01:47:26

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.080	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.314	01:43
Breath Volume		2.855	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.314	01:45
Breath Volume		2.121	01:45
Air Blank		0.000	01:46
Control Test		0.078	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 08/13/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KILLIAN,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: KOSIOR

DATE/TIME OF TEST: 07/13/2025 18:40:44

Diagnostics Check		OK	18:34
Air Blank		0.000	18:35
Control Test		0.079	18:35
Air Blank		0.000	18:36
Subject Sample	# 1	0.213	18:36
Breath Volume		2.355	18:36
Air Blank		0.000	18:36
Air Blank		0.000	18:38
Subject Sample	# 2	0.213	18:39
Breath Volume		2.230	18:39
Air Blank		0.000	18:39
Control Test		0.078	18:40
Air Blank		0.000	18:40
Diagnostics Check		OK	18:40

Data Download Date/Time: 08/13/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,B,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BJURHOLM,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: HARUNI

DATE/TIME OF TEST: 07/16/2025 19:15:11

Diagnostics Check		OK	19:09
Air Blank		0.000	19:09
Control Test		0.079	19:09
Air Blank		0.000	19:10
Subject Sample	# 1	0.180	19:10
Breath Volume		1.605	19:10
Air Blank		0.000	19:11
Air Blank		0.000	19:13
Subject Sample	# 2	0.182	19:13
Breath Volume		1.777	19:13
Air Blank		0.000	19:14
Control Test		0.078	19:14
Air Blank		0.000	19:15
Diagnostics Check		OK	19:15

Data Download Date/Time: 08/13/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: BALLANCE

DATE/TIME OF TEST: 07/17/2025 19:33:14

Diagnostics Check		OK	19:26
Air Blank		0.000	19:26
Control Test		0.079	19:26
Air Blank		0.000	19:27
Subject Sample	# 1	0.254	19:28
Breath Volume		1.230	19:28
Air Blank		0.000	19:28
Air Blank		0.000	19:30
Subject Sample	# 2	0.251	19:31
Breath Volume		1.347	19:31
Air Blank		0.000	19:32
Control Test		0.078	19:32
Air Blank		0.000	19:33
Diagnostics Check		OK	19:33

Data Download Date/Time: 08/13/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KILIAN,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 07/22/2025 19:50:15

Diagnostics Check		OK	19:44
Air Blank		0.000	19:44
Control Test		0.079	19:44
Air Blank		0.000	19:45
Subject Sample	# 1	0.123	19:45
Breath Volume		1.433	19:45
Air Blank		0.000	19:46
Air Blank		0.000	19:48
Subject Sample	# 2	0.122	19:48
Breath Volume		1.593	19:48
Air Blank		0.000	19:49
Control Test		0.079	19:49
Air Blank		0.000	19:50
Diagnostics Check		OK	19:50

Data Download Date/Time: 08/13/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: STEPHENSON

DATE/TIME OF TEST: 07/22/2025 23:41:38

Diagnostics Check		OK	23:34
Air Blank		0.000	23:35
Control Test		0.079	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.065	23:37
Breath Volume		1.472	23:37
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.063	23:40
Breath Volume		1.242	23:40
Air Blank		0.000	23:40
Control Test		0.079	23:41
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 08/13/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BJURHOLM,JACOB,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BJURHOLM,JACOB
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: BAIRD

DATE/TIME OF TEST: 07/26/2025 19:44:47

Diagnostics Check		OK	19:41	
Air Blank		0.000	19:42	
Control Test		0.079	19:42	
Air Blank		0.000	19:43	
Subject Sample	# 1		19:43	Subject Test Refused
Breath Volume		0.000	19:43	
Air Blank		0.000	19:43	
Control Test		0.077	19:44	
Air Blank		0.000	19:44	
Diagnostics Check		OK	19:44	

Data Download Date/Time: 08/13/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: GOINS

DATE/TIME OF TEST: 07/26/2025 22:06:14

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:03	
Control Test		0.079	22:04	
Air Blank		0.000	22:04	
Subject Sample	# 1		22:04	Subject Test Refused
Breath Volume		0.000	22:04	
Air Blank		0.000	22:05	
Control Test		0.078	22:05	
Air Blank		0.000	22:06	
Diagnostics Check		OK	22:06	

Data Download Date/Time: 08/13/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: TASLIM-MOGHADDAM

DATE/TIME OF TEST: 07/29/2025 20:54:56

Diagnostics Check		OK	20:49
Air Blank		0.000	20:49
Control Test		0.078	20:49
Air Blank		0.000	20:50
Subject Sample	# 1	0.025	20:50
Breath Volume		3.519	20:50
Air Blank		0.000	20:51
Air Blank		0.000	20:53
Subject Sample	# 2	0.024	20:53
Breath Volume		3.320	20:53
Air Blank		0.000	20:54
Control Test		0.078	20:54
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 08/13/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: MAHEU

DATE/TIME OF TEST: 07/05/2025 21:29:02

Diagnostics Check		OK	21:22
Air Blank		0.000	21:22
Control Test		0.080	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.173	21:24
Breath Volume		2.265	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:26
Subject Sample	# 2	0.173	21:27
Breath Volume		2.734	21:27
Air Blank		0.000	21:28
Control Test		0.080	21:28
Air Blank		0.000	21:28
Diagnostics Check		OK	21:29

Data Download Date/Time: 07/17/2025 19:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/06/2025 03:47:00

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.081	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.176	03:42
Breath Volume		1.843	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.166	03:45
Breath Volume		1.656	03:45
Air Blank		0.000	03:46
Control Test		0.081	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:47

Data Download Date/Time: 07/17/2025 19:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,BRIAN,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: PERRY,MADISON
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/08/2025 19:26:49

Diagnostics Check		OK	19:20
Air Blank		0.000	19:20
Control Test		0.081	19:20
Air Blank		0.000	19:21
Subject Sample	# 1	0.000	19:22
Breath Volume		2.320	19:22
Air Blank		0.000	19:22
Air Blank		0.000	19:24
Subject Sample	# 2	0.000	19:25
Breath Volume		2.449	19:25
Air Blank		0.000	19:25
Control Test		0.081	19:26
Air Blank		0.000	19:26
Diagnostics Check		OK	19:26

Data Download Date/Time: 07/17/2025 19:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LIGHTFOOT,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: NAGY

DATE/TIME OF TEST: 07/09/2025 21:15:17

Diagnostics Check		OK	21:09
Air Blank		0.000	21:09
Control Test		0.081	21:09
Air Blank		0.000	21:10
Subject Sample	# 1	0.071	21:10
Breath Volume		1.679	21:10
Air Blank		0.000	21:11
Air Blank		0.000	21:13
Subject Sample	# 2	0.069	21:13
Breath Volume		1.609	21:13
Air Blank		0.000	21:14
Control Test		0.082	21:14
Air Blank		0.000	21:15
Diagnostics Check		OK	21:15

Data Download Date/Time: 07/17/2025 19:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 07/10/2025 21:34:16

Diagnostics Check		OK	21:28
Air Blank		0.000	21:28
Control Test		0.081	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.000	21:29
Breath Volume		1.320	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.000	21:32
Breath Volume		1.191	21:32
Air Blank		0.000	21:33
Control Test		0.081	21:33
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 07/17/2025 19:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: BRODIE

DATE/TIME OF TEST: 07/11/2025 23:14:52

Diagnostics Check		OK	23:08
Air Blank		0.000	23:09
Control Test		0.081	23:09
Air Blank		0.000	23:09
Subject Sample	# 1	0.000	23:10
Breath Volume		4.757	23:10
Air Blank		0.000	23:10
Air Blank		0.000	23:12
Subject Sample	# 2	0.000	23:13
Breath Volume		4.871	23:13
Air Blank		0.000	23:13
Control Test		0.082	23:14
Air Blank		0.000	23:14
Diagnostics Check		OK	23:14

Data Download Date/Time: 07/17/2025 19:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MCKINNES

DATE/TIME OF TEST: 07/12/2025 03:02:34

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.081	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.131	02:57
Breath Volume		1.382	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.130	03:01
Breath Volume		1.605	03:01
Air Blank		0.000	03:01
Control Test		0.082	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 07/17/2025 19:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: HABNER

DATE/TIME OF TEST: 07/12/2025 19:52:23

Diagnostics Check		OK	19:40	
Air Blank		0.000	19:41	
Control Test		0.081	19:41	
Air Blank		0.000	19:42	
Subject Sample	# 1		19:45	No Sample Provided
Breath Volume		0.000	19:45	
Air Blank		0.000	19:45	
Air Blank		0.000	19:47	
Subject Sample	# 2		19:50	No Sample Provided
Breath Volume		0.000	19:50	
Air Blank		0.000	19:51	
Control Test		0.082	19:51	
Air Blank		0.000	19:52	
Diagnostics Check		OK	19:52	

Data Download Date/Time: 07/17/2025 19:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 07/13/2025 02:43:18

Diagnostics Check		OK	02:37
Air Blank		0.000	02:37
Control Test		0.081	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.179	02:38
Breath Volume		1.738	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.175	02:41
Breath Volume		1.730	02:41
Air Blank		0.000	02:42
Control Test		0.081	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 07/17/2025 19:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: DOROTEO

DATE/TIME OF TEST: 07/13/2025 03:39:33

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.082	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.088	03:35
Breath Volume		1.679	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.071	03:38
Breath Volume		1.210	03:38
Air Blank		0.000	03:38
Control Test		0.081	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 07/17/2025 19:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GIARDINO,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/15/2025 03:54:20

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.081	03:49
Air Blank		0.000	03:49
Subject Sample	# 1	0.176	03:49
Breath Volume		1.304	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.181	03:52
Breath Volume		1.730	03:52
Air Blank		0.000	03:53
Control Test		0.081	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 07/17/2025 19:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,BRIAN,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: FROST,BRIAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: ZAFAR

DATE/TIME OF TEST: 07/15/2025 17:53:49

Diagnostics Check		OK	17:46
Air Blank		0.000	17:47
Control Test		0.081	17:47
Air Blank		0.000	17:48
Subject Sample	# 1	0.066	17:48
Breath Volume		1.394	17:48
Air Blank		0.000	17:49
Air Blank		0.000	17:51
Subject Sample	# 2	0.065	17:52
Breath Volume		1.800	17:52
Air Blank		0.000	17:52
Control Test		0.081	17:53
Air Blank		0.000	17:53
Diagnostics Check		OK	17:53

Data Download Date/Time: 07/17/2025 19:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ABERNATHY

DATE/TIME OF TEST: 07/17/2025 23:48:48

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.081	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.000	23:44
Breath Volume		2.640	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.000	23:47
Breath Volume		2.539	23:47
Air Blank		0.000	23:47
Control Test		0.081	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 08/21/2025 21:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: VELASCO GIRON

DATE/TIME OF TEST: 07/20/2025 04:14:56

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.081	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.206	04:10
Breath Volume		2.445	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:12
Subject Sample	# 2	0.213	04:13
Breath Volume		3.601	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 08/21/2025 21:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KIDD,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: SANTOS SANTIAGO

DATE/TIME OF TEST: 07/31/2025 21:56:19

Diagnostics Check		OK	21:49
Air Blank		0.000	21:50
Control Test		0.081	21:50
Air Blank		0.000	21:50
Subject Sample	# 1	0.016	21:51
Breath Volume		2.039	21:51
Air Blank		0.000	21:52
Air Blank		0.000	21:54
Subject Sample	# 2	0.015	21:54
Breath Volume		1.519	21:54
Air Blank		0.000	21:55
Control Test		0.081	21:55
Air Blank		0.000	21:56
Diagnostics Check		OK	21:56

Data Download Date/Time: 08/21/2025 21:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,DOMINIQUE,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HOWARD,DOMINIQUE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 07/03/2025 16:23:20

Diagnostics Check		OK	16:10	
Air Blank		0.000	16:10	
Control Test		0.075	16:10	
Air Blank		0.000	16:11	
Subject Sample	# 1	0.325	16:12	
Breath Volume		1.507	16:12	
Air Blank		0.000	16:13	
Air Blank		0.000	16:15	
Subject Sample	# 2	0.289	16:18	Volume Not Met
Breath Volume		0.683	16:18	
Air Blank		0.000	16:19	
Air Blank		0.000	16:20	
Subject Sample	# 3	0.327	16:21	
Breath Volume		2.632	16:21	
Air Blank		0.000	16:22	
Control Test		0.076	16:22	
Air Blank		0.000	16:23	
Diagnostics Check		OK	16:23	

Data Download Date/Time: 07/25/2025 13:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMBETA,JOHN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LAMBETA,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 07/03/2025 20:37:15

Diagnostics Check		OK	20:31
Air Blank		0.000	20:31
Control Test		0.080	20:32
Air Blank		0.000	20:32
Subject Sample	# 1	0.000	20:33
Breath Volume		2.093	20:33
Air Blank		0.000	20:33
Air Blank		0.000	20:35
Subject Sample	# 2	0.000	20:35
Breath Volume		1.425	20:35
Air Blank		0.000	20:36
Control Test		0.078	20:36
Air Blank		0.000	20:37
Diagnostics Check		OK	20:37

Data Download Date/Time: 07/25/2025 13:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY,GRANT,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY,GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: ALFARO GONZALEZ

DATE/TIME OF TEST: 07/03/2025 23:34:22

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.079	23:28
Air Blank		0.000	23:29
Subject Sample	# 1	0.150	23:29
Breath Volume		2.437	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.144	23:32
Breath Volume		2.746	23:32
Air Blank		0.000	23:33
Control Test		0.077	23:33
Air Blank		0.000	23:34
Diagnostics Check		OK	23:34

Data Download Date/Time: 07/25/2025 13:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GONZALEZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: CARADONNA

DATE/TIME OF TEST: 07/06/2025 22:23:34

Diagnostics Check		OK	22:17
Air Blank		0.000	22:18
Control Test		0.081	22:18
Air Blank		0.000	22:18
Subject Sample	# 1	0.000	22:19
Breath Volume		2.824	22:19
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.000	22:22
Breath Volume		3.464	22:22
Air Blank		0.000	22:22
Control Test		0.080	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 07/25/2025 13:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: MARTINEZ FERNANDEZ

DATE/TIME OF TEST: 07/12/2025 04:23:08

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.080	04:17
Air Blank		0.000	04:17
Subject Sample	# 1	0.105	04:18
Breath Volume		1.792	04:18
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.105	04:21
Breath Volume		1.808	04:21
Air Blank		0.000	04:22
Control Test		0.079	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 07/25/2025 13:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIXSON,JEREMY,C
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BRIGGS,PRESTON
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:27

SUBJECT'S LAST NAME: GEISINGER

DATE/TIME OF TEST: 07/12/2025 14:26:40

Diagnostics Check		OK	14:19
Air Blank		0.000	14:20
Control Test		0.078	14:20
Air Blank		0.000	14:21
Subject Sample	# 1	0.241	14:21
Breath Volume		1.257	14:21
Air Blank		0.000	14:22
Air Blank		0.000	14:24
Subject Sample	# 2	0.241	14:25
Breath Volume		1.335	14:25
Air Blank		0.000	14:25
Control Test		0.079	14:26
Air Blank		0.000	14:26
Diagnostics Check		OK	14:26

Data Download Date/Time: 07/25/2025 13:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTO,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: COGLETON

DATE/TIME OF TEST: 07/17/2025 22:02:13

Diagnostics Check		OK	21:59	
Air Blank		0.000	21:59	
Control Test		0.079	22:00	
Air Blank		0.000	22:00	
Subject Sample	# 1		22:00	Subject Test Refused
Breath Volume		0.000	22:00	
Air Blank		0.000	22:01	
Control Test		0.079	22:01	
Air Blank		0.000	22:02	
Diagnostics Check		OK	22:02	

Data Download Date/Time: 07/25/2025 13:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 07/26/2025 03:14:39

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.000	03:10
Breath Volume		3.019	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.000	03:13
Breath Volume		2.515	03:13
Air Blank		0.000	03:13
Control Test		0.080	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 08/29/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GERAIDI,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BEASON

DATE/TIME OF TEST: 07/26/2025 03:48:54

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.207	03:43
Breath Volume		1.265	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.222	03:47
Breath Volume		1.925	03:47
Air Blank		0.000	03:48
Control Test		0.080	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 08/29/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIZZO,HAYLEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RIZZO,HAYLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: AGOSTO MONTANEZ

DATE/TIME OF TEST: 07/26/2025 18:15:47

Diagnostics Check		OK	18:09
Air Blank		0.000	18:09
Control Test		0.080	18:10
Air Blank		0.000	18:10
Subject Sample	# 1	0.000	18:11
Breath Volume		1.886	18:11
Air Blank		0.000	18:11
Air Blank		0.000	18:13
Subject Sample	# 2	0.000	18:14
Breath Volume		1.351	18:14
Air Blank		0.000	18:14
Control Test		0.079	18:15
Air Blank		0.000	18:15
Diagnostics Check		OK	18:15

Data Download Date/Time: 08/29/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: FARINELLI

DATE/TIME OF TEST: 07/27/2025 21:02:05

Diagnostics Check		OK	20:55
Air Blank		0.000	20:56
Control Test		0.080	20:56
Air Blank		0.000	20:56
Subject Sample	# 1	0.136	20:57
Breath Volume		2.140	20:57
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.124	21:00
Breath Volume		1.207	21:00
Air Blank		0.000	21:01
Control Test		0.079	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 08/29/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ-GUILLEN,MARCOS,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LOPEZ-GUILLEN,MARCOS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: LAYSON

DATE/TIME OF TEST: 07/31/2025 19:51:07

Diagnostics Check		OK	19:44
Air Blank		0.000	19:45
Control Test		0.080	19:45
Air Blank		0.000	19:46
Subject Sample	# 1	0.234	19:46
Breath Volume		1.859	19:46
Air Blank		0.000	19:47
Air Blank		0.000	19:49
Subject Sample	# 2	0.233	19:49
Breath Volume		2.093	19:49
Air Blank		0.000	19:50
Control Test		0.080	19:50
Air Blank		0.000	19:51
Diagnostics Check		OK	19:51

Data Download Date/Time: 08/29/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: RICHARDSON,LOYD
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/06/2025 02:01:18

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.079	01:55
Air Blank		0.000	01:56
Subject Sample	# 1	0.165	01:56
Breath Volume		3.457	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.166	01:59
Breath Volume		2.906	01:59
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 07/21/2025 10:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: CHERUBINO,NICHOLAS
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/06/2025 23:00:28

Diagnostics Check		OK	22:53
Air Blank		0.000	22:53
Control Test		0.079	22:53
Air Blank		0.000	22:54
Subject Sample	# 1	0.150	22:56
Breath Volume		1.804	22:56
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.149	22:58
Breath Volume		1.593	22:58
Air Blank		0.000	22:59
Control Test		0.078	22:59
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 07/21/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SANTIGAO DIAZ

DATE/TIME OF TEST: 07/07/2025 01:08:05

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.078	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.210	01:03
Breath Volume		1.519	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.213	01:06
Breath Volume		1.609	01:06
Air Blank		0.000	01:07
Control Test		0.079	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 07/21/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIVECOAT,JUSTIN,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FIVECOAT,JUSTIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/18/2025 15:21:43

Diagnostics Check	OK	15:20	
Air Blank		15:21	RFI Detect
Air Blank	0.000	15:21	

Data Download Date/Time: 07/21/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIVECOAT,JUSTIN,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FIVECOAT,JUSTIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/18/2025 15:23:58

Diagnostics Check	OK	15:23	
Air Blank		15:23	RFI Detect
Air Blank	0.000	15:23	

Data Download Date/Time: 07/21/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIVECOAT,JUSTIN,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FIVECOAT,JUSTIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/18/2025 15:33:47

Diagnostics Check		OK	15:25
Air Blank		0.000	15:25
Control Test		0.076	15:26
Air Blank		0.000	15:26
Subject Sample	# 1	0.359	15:28
Breath Volume		1.125	15:28
Air Blank		0.000	15:29
Air Blank		0.000	15:30
Subject Sample	# 2	0.350	15:32
Breath Volume		2.765	15:32
Air Blank		0.000	15:32
Control Test		0.077	15:33
Air Blank		0.000	15:33
Diagnostics Check		OK	15:33

Data Download Date/Time: 07/21/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: BRYCE

DATE/TIME OF TEST: 07/25/2025 04:15:46

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:13	
Control Test		0.079	04:13	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:14	Subject Test Refused
Breath Volume		0.000	04:14	
Air Blank		0.000	04:14	
Control Test		0.079	04:15	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 08/25/2025 13:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUX,SAGE,E
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: LOUX,SAGE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BASILE

DATE/TIME OF TEST: 07/25/2025 23:13:10

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.080	23:07
Air Blank		0.000	23:07
Subject Sample	# 1	0.000	23:08
Breath Volume		1.644	23:08
Air Blank		0.000	23:09
Air Blank		0.000	23:11
Subject Sample	# 2	0.000	23:11
Breath Volume		1.597	23:11
Air Blank		0.000	23:12
Control Test		0.079	23:12
Air Blank		0.000	23:13
Diagnostics Check		OK	23:13

Data Download Date/Time: 08/25/2025 13:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,RYAN,N
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: CLARK,RYAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: SLATTERY

DATE/TIME OF TEST: 07/26/2025 00:54:10

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.079	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.255	00:49
Breath Volume		1.925	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:51
Subject Sample	# 2	0.269	00:52
Breath Volume		1.652	00:52
Air Blank		0.000	00:53
Control Test		0.079	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 08/25/2025 13:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: SETTLE

DATE/TIME OF TEST: 07/26/2025 01:35:50

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.079	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:34	
Control Test		0.080	01:35	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 08/25/2025 13:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,RYAN,N
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HILL,CORNELIUS
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: DAVENPORT

DATE/TIME OF TEST: 07/26/2025 03:36:57

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.187	03:31
Breath Volume		1.863	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.180	03:35
Breath Volume		1.250	03:35
Air Blank		0.000	03:36
Control Test		0.078	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 08/25/2025 13:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,RYAN,N
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: CLARK,RYAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: ROMANO

DATE/TIME OF TEST: 07/26/2025 23:41:54

Diagnostics Check		OK	23:36	
Air Blank		0.000	23:37	
Control Test		0.080	23:37	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:40	Subject Test Refused
Breath Volume		0.000	23:40	
Air Blank		0.000	23:40	
Control Test		0.080	23:41	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 08/25/2025 13:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARDNER,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 07/27/2025 21:50:48

Diagnostics Check		OK	21:44
Air Blank		0.000	21:45
Control Test		0.080	21:45
Air Blank		0.000	21:45
Subject Sample	# 1	0.168	21:46
Breath Volume		2.304	21:46
Air Blank		0.000	21:47
Air Blank		0.000	21:48
Subject Sample	# 2	0.170	21:49
Breath Volume		2.363	21:49
Air Blank		0.000	21:49
Control Test		0.080	21:50
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 08/25/2025 13:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CELESTINO

DATE/TIME OF TEST: 07/01/2025 00:01:10

Diagnostics Check		OK	23:54
Air Blank		0.000	23:54
Control Test		0.079	23:55
Air Blank		0.000	23:55
Subject Sample	# 1	0.266	23:56
Breath Volume		1.480	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 2	0.266	23:59
Breath Volume		1.453	23:59
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 07/29/2025 05:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,KEVIN,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: MCHARGUE

DATE/TIME OF TEST: 07/01/2025 04:15:17

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:04	
Control Test		0.080	04:04	
Air Blank		0.000	04:05	
Subject Sample	# 1		04:08	No Sample Provided
Breath Volume		0.000	04:08	
Air Blank		0.000	04:08	
Air Blank		0.000	04:10	
Subject Sample	# 2		04:13	No Sample Provided
Breath Volume		0.000	04:13	
Air Blank		0.000	04:14	
Control Test		0.080	04:14	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 07/29/2025 05:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,MALACHI,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RANDOLPH,MALACHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: ALANIZ

DATE/TIME OF TEST: 07/01/2025 04:38:18

Diagnostics Check		OK	04:31
Air Blank		0.000	04:32
Control Test		0.081	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.113	04:33
Breath Volume		1.667	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.109	04:36
Breath Volume		1.882	04:36
Air Blank		0.000	04:37
Control Test		0.080	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 07/29/2025 05:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/02/2025 02:43:01

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:28	
Control Test		0.081	02:28	
Air Blank		0.000	02:29	
Subject Sample	# 1	0.213	02:30	
Breath Volume		1.539	02:30	
Air Blank		0.000	02:31	
Air Blank		0.000	02:32	
Subject Sample	# 2	0.211	02:35	Slope Not Level
Breath Volume		1.113	02:35	
Air Blank		0.000	02:36	
Air Blank		0.000	02:38	
Subject Sample	# 3		02:41	No Sample Provided
Breath Volume		0.000	02:41	
Air Blank		0.000	02:42	
Control Test		0.079	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 07/29/2025 05:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO,DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: YEPES

DATE/TIME OF TEST: 07/03/2025 04:36:28

Diagnostics Check		OK	04:24	
Air Blank		0.000	04:25	
Control Test		0.081	04:25	
Air Blank		0.000	04:25	
Subject Sample	# 1		04:29	No Sample Provided
Breath Volume		0.000	04:29	
Air Blank		0.000	04:29	
Air Blank		0.000	04:31	
Subject Sample	# 2		04:34	No Sample Provided
Breath Volume		0.000	04:34	
Air Blank		0.000	04:35	
Control Test		0.081	04:35	
Air Blank			04:35	RFI Detect
Air Blank		0.000	04:36	

Data Download Date/Time: 07/29/2025 05:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/03/2025 23:27:56

Diagnostics Check		OK	23:14	
Air Blank		0.000	23:14	
Control Test		0.081	23:14	
Air Blank		0.000	23:15	
Subject Sample	# 1		23:18	No Sample Provided
Breath Volume		0.000	23:18	
Air Blank		0.000	23:19	
Air Blank		0.000	23:21	
Subject Sample	# 2	0.214	23:22	
Breath Volume		1.562	23:22	
Air Blank		0.000	23:23	
Air Blank		0.000	23:25	
Subject Sample	# 3	0.213	23:26	
Breath Volume		1.769	23:26	
Air Blank		0.000	23:27	
Control Test		0.078	23:27	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 07/29/2025 05:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JULES,EMMANUELLO
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 07/04/2025 00:25:52

Diagnostics Check	OK	00:25	
Air Blank		00:25	RFI Detect
Air Blank	0.000	00:25	

Data Download Date/Time: 07/29/2025 05:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JULES,EMMANUELLO
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 07/04/2025 00:39:46

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:28	
Control Test		0.081	00:29	
Air Blank		0.000	00:29	
Subject Sample	# 1		00:32	No Sample Provided
Breath Volume		0.000	00:32	
Air Blank		0.000	00:33	
Air Blank		0.000	00:35	
Subject Sample	# 2		00:38	No Sample Provided
Breath Volume		0.000	00:38	
Air Blank		0.000	00:38	
Control Test		0.081	00:39	
Air Blank		0.000	00:39	
Diagnostics Check		OK	00:39	

Data Download Date/Time: 07/29/2025 05:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,RANDY,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WILSON,RANDY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: TORRES BONET

DATE/TIME OF TEST: 07/04/2025 04:39:00

Diagnostics Check		OK	04:27	
Air Blank		0.000	04:27	
Control Test		0.080	04:27	
Air Blank		0.000	04:28	
Subject Sample	# 1	0.127	04:31	Volume Not Met
Breath Volume		0.488	04:31	
Air Blank		0.000	04:32	
Air Blank		0.000	04:34	
Subject Sample	# 2	0.145	04:37	Volume Not Met
Breath Volume		0.183	04:37	
Air Blank		0.000	04:38	
Control Test		0.078	04:38	
Air Blank		0.000	04:38	
Diagnostics Check		OK	04:39	

Data Download Date/Time: 07/29/2025 05:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO,DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: GIBBS III

DATE/TIME OF TEST: 07/04/2025 17:22:21

Diagnostics Check		OK	17:15
Air Blank		0.000	17:16
Control Test		0.080	17:16
Air Blank		0.000	17:17
Subject Sample	# 1	0.000	17:17
Breath Volume		1.679	17:17
Air Blank		0.000	17:18
Air Blank		0.000	17:20
Subject Sample	# 2	0.000	17:20
Breath Volume		2.625	17:20
Air Blank		0.000	17:21
Control Test		0.079	17:21
Air Blank		0.000	17:22
Diagnostics Check		OK	17:22

Data Download Date/Time: 07/29/2025 05:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: MCEACHERN

DATE/TIME OF TEST: 07/04/2025 19:00:49

Diagnostics Check		OK	18:54
Air Blank		0.000	18:55
Control Test		0.081	18:55
Air Blank		0.000	18:56
Subject Sample	# 1	0.000	18:56
Breath Volume		2.734	18:56
Air Blank		0.000	18:56
Air Blank		0.000	18:58
Subject Sample	# 2	0.000	18:59
Breath Volume		2.503	18:59
Air Blank		0.000	18:59
Control Test		0.080	19:00
Air Blank		0.000	19:00
Diagnostics Check		OK	19:00

Data Download Date/Time: 07/29/2025 05:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: HERNDON II

DATE/TIME OF TEST: 07/04/2025 19:18:35

Diagnostics Check		OK	19:07	
Air Blank		0.000	19:07	
Control Test		0.079	19:07	
Air Blank		0.000	19:08	
Subject Sample	# 1		19:11	No Sample Provided
Breath Volume		0.000	19:11	
Air Blank		0.000	19:12	
Air Blank		0.000	19:14	
Subject Sample	# 2		19:17	No Sample Provided
Breath Volume		0.000	19:17	
Air Blank		0.000	19:17	
Control Test		0.080	19:17	
Air Blank		0.000	19:18	
Diagnostics Check		OK	19:18	

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: GUZMAN VAZQUEZ

DATE/TIME OF TEST: 07/04/2025 21:08:10

Diagnostics Check		OK	20:56	
Air Blank		0.000	20:57	
Control Test		0.081	20:57	
Air Blank		0.000	20:58	
Subject Sample	# 1		21:01	No Sample Provided
Breath Volume		0.000	21:01	
Air Blank		0.000	21:01	
Air Blank		0.000	21:03	
Subject Sample	# 2		21:06	No Sample Provided
Breath Volume		0.000	21:06	
Air Blank		0.000	21:07	
Control Test		0.081	21:07	
Air Blank		0.000	21:08	
Diagnostics Check		OK	21:08	

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: SALGUERO

DATE/TIME OF TEST: 07/05/2025 03:39:08

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.081	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.084	03:34
Breath Volume		2.335	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.081	03:37
Breath Volume		2.007	03:37
Air Blank		0.000	03:38
Control Test		0.081	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 07/05/2025 03:58:21

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.081	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.139	03:51
Breath Volume		1.441	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.139	03:56
Breath Volume		1.976	03:56
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JANIAK,MALIKHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:02

SUBJECT'S LAST NAME: FONTANEZ ALICEA

DATE/TIME OF TEST: 07/05/2025 16:38:18

Diagnostics Check		OK	16:32
Air Blank		0.000	16:32
Control Test		0.080	16:33
Air Blank		0.000	16:33
Subject Sample	# 1	0.000	16:34
Breath Volume		1.644	16:34
Air Blank		0.000	16:34
Air Blank		0.000	16:36
Subject Sample	# 2	0.000	16:36
Breath Volume		2.339	16:36
Air Blank		0.000	16:37
Control Test		0.081	16:37
Air Blank		0.000	16:38
Diagnostics Check		OK	16:38

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 07/06/2025 01:04:39

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:53	
Control Test		0.077	00:53	
Air Blank		0.000	00:54	
Subject Sample	# 1		00:57	No Sample Provided
Breath Volume		0.000	00:57	
Air Blank		0.000	00:58	
Air Blank		0.000	01:00	
Subject Sample	# 2		01:03	No Sample Provided
Breath Volume		0.000	01:03	
Air Blank		0.000	01:03	
Control Test		0.080	01:04	
Air Blank		0.000	01:04	
Diagnostics Check		OK	01:04	

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: RODRIGUEZ AQUINO

DATE/TIME OF TEST: 07/06/2025 02:01:44

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.081	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.000	01:57
Breath Volume		2.300	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.000	02:00
Breath Volume		2.707	02:00
Air Blank		0.000	02:00
Control Test		0.079	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: YOC

DATE/TIME OF TEST: 07/06/2025 18:56:04

Diagnostics Check		OK	18:44	
Air Blank		0.000	18:45	
Control Test		0.081	18:45	
Air Blank		0.000	18:45	
Subject Sample	# 1		18:49	No Sample Provided
Breath Volume		0.000	18:49	
Air Blank		0.000	18:49	
Air Blank		0.000	18:51	
Subject Sample	# 2		18:54	No Sample Provided
Breath Volume		0.000	18:54	
Air Blank		0.000	18:55	
Control Test		0.079	18:55	
Air Blank		0.000	18:55	
Diagnostics Check		OK	18:56	

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TOMB

DATE/TIME OF TEST: 07/07/2025 03:23:04

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.000	03:18
Breath Volume		2.121	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.000	03:21
Breath Volume		1.980	03:21
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:23

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO,DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/09/2025 03:44:59

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:34	
Control Test		0.081	03:34	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:38	No Sample Provided
Breath Volume		0.000	03:38	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2		03:43	No Sample Provided
Breath Volume		0.000	03:43	
Air Blank		0.000	03:44	
Control Test		0.080	03:44	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: ORTIZ ORTIZ

DATE/TIME OF TEST: 07/09/2025 23:09:31

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.081	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.068	23:04
Breath Volume		1.480	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.072	23:07
Breath Volume		1.871	23:07
Air Blank		0.000	23:08
Control Test		0.079	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MOHEREK

DATE/TIME OF TEST: 07/10/2025 02:18:44

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:07	
Control Test		0.081	02:08	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:11	No Sample Provided
Breath Volume		0.000	02:11	
Air Blank		0.000	02:12	
Air Blank		0.000	02:14	
Subject Sample	# 2		02:17	No Sample Provided
Breath Volume		0.000	02:17	
Air Blank		0.000	02:17	
Control Test		0.081	02:18	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: HARTLEY

DATE/TIME OF TEST: 07/11/2025 16:44:55

Diagnostics Check		OK	16:33	
Air Blank		0.000	16:33	
Control Test		0.079	16:34	
Air Blank		0.000	16:34	
Subject Sample	# 1		16:37	No Sample Provided
Breath Volume		0.000	16:37	
Air Blank		0.000	16:38	
Air Blank		0.000	16:40	
Subject Sample	# 2		16:43	No Sample Provided
Breath Volume		0.000	16:43	
Air Blank		0.000	16:43	
Control Test		0.080	16:44	
Air Blank		0.000	16:44	
Diagnostics Check		OK	16:44	

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: YEATER

DATE/TIME OF TEST: 07/12/2025 02:18:17

Diagnostics Check		OK	02:11
Air Blank		0.000	02:12
Control Test		0.081	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.129	02:13
Breath Volume		1.378	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.124	02:16
Breath Volume		1.640	02:16
Air Blank		0.000	02:17
Control Test		0.080	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JANIAK,MALIKHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MIKELL III

DATE/TIME OF TEST: 07/12/2025 03:16:02

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:05	
Control Test		0.079	03:05	
Air Blank		0.000	03:05	
Subject Sample	# 1		03:09	No Sample Provided
Breath Volume		0.000	03:09	
Air Blank		0.000	03:09	
Air Blank		0.000	03:11	
Subject Sample	# 2		03:14	No Sample Provided
Breath Volume		0.000	03:14	
Air Blank		0.000	03:15	
Control Test		0.079	03:15	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:16	

Data Download Date/Time: 07/29/2025 05:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 07/12/2025 23:54:14

Diagnostics Check		OK	23:47
Air Blank		0.000	23:48
Control Test		0.081	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.200	23:49
Breath Volume		2.019	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.199	23:52
Breath Volume		1.406	23:52
Air Blank		0.000	23:53
Control Test		0.080	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 07/29/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PERALTA,ALLISON
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: BUCK-KELLY

DATE/TIME OF TEST: 07/14/2025 00:18:46

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.080	00:13
Air Blank		0.000	00:13
Subject Sample	# 1	0.218	00:14
Breath Volume		1.753	00:14
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.214	00:17
Breath Volume		1.195	00:17
Air Blank		0.000	00:17
Control Test		0.079	00:18
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 07/29/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JANIAK,MALIKHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: NUSENKO

DATE/TIME OF TEST: 07/14/2025 13:52:18

Diagnostics Check		OK	13:45
Air Blank		0.000	13:46
Control Test		0.080	13:46
Air Blank		0.000	13:46
Subject Sample	# 1	0.102	13:47
Breath Volume		2.914	13:47
Air Blank		0.000	13:48
Air Blank		0.000	13:49
Subject Sample	# 2	0.101	13:50
Breath Volume		3.082	13:50
Air Blank		0.000	13:51
Control Test		0.080	13:51
Air Blank		0.000	13:52
Diagnostics Check		OK	13:52

Data Download Date/Time: 07/29/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:56

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 07/14/2025 15:27:37

Diagnostics Check		OK	15:21
Air Blank		0.000	15:21
Control Test		0.081	15:22
Air Blank		0.000	15:22
Subject Sample	# 1	0.000	15:23
Breath Volume		1.691	15:23
Air Blank		0.000	15:23
Air Blank		0.000	15:25
Subject Sample	# 2	0.000	15:26
Breath Volume		1.597	15:26
Air Blank		0.000	15:26
Control Test		0.080	15:27
Air Blank		0.000	15:27
Diagnostics Check		OK	15:27

Data Download Date/Time: 07/29/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 07/15/2025 02:32:19

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.081	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.000	02:28
Breath Volume		1.671	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.000	02:30
Breath Volume		1.882	02:30
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 07/29/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: HERKEL

DATE/TIME OF TEST: 07/16/2025 20:57:16

Diagnostics Check		OK	20:50
Air Blank		0.000	20:51
Control Test		0.081	20:51
Air Blank		0.000	20:52
Subject Sample	# 1	0.178	20:52
Breath Volume		1.546	20:52
Air Blank		0.000	20:53
Air Blank		0.000	20:55
Subject Sample	# 2	0.172	20:55
Breath Volume		1.781	20:55
Air Blank		0.000	20:56
Control Test		0.079	20:56
Air Blank		0.000	20:57
Diagnostics Check		OK	20:57

Data Download Date/Time: 07/29/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: POMMELLS

DATE/TIME OF TEST: 07/19/2025 01:17:30

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.080	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.181	01:12
Breath Volume		2.000	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.168	01:15
Breath Volume		2.007	01:15
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 07/29/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: WENGLASZ

DATE/TIME OF TEST: 07/19/2025 01:33:03

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:22	
Control Test		0.080	01:22	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:26	No Sample Provided
Breath Volume		0.000	01:26	
Air Blank		0.000	01:26	
Air Blank		0.000	01:28	
Subject Sample	# 2		01:31	No Sample Provided
Breath Volume		0.000	01:31	
Air Blank		0.000	01:32	
Control Test		0.081	01:32	
Air Blank		0.000	01:32	
Diagnostics Check		OK	01:33	

Data Download Date/Time: 07/29/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,MALACHI,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELLIDO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: ISSA

DATE/TIME OF TEST: 07/19/2025 06:00:58

Diagnostics Check		OK	05:53
Air Blank		0.000	05:53
Control Test		0.081	05:54
Air Blank		0.000	05:54
Subject Sample	# 1	0.157	05:56
Breath Volume		3.265	05:56
Air Blank		0.000	05:57
Air Blank		0.000	05:58
Subject Sample	# 2	0.150	05:59
Breath Volume		1.644	05:59
Air Blank		0.000	06:00
Control Test		0.078	06:00
Air Blank		0.000	06:00
Diagnostics Check		OK	06:00

Data Download Date/Time: 07/29/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,WALDRICK,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,WALDRICK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:13

SUBJECT'S LAST NAME: RINCON CASANOVA

DATE/TIME OF TEST: 07/19/2025 11:49:18

Diagnostics Check		OK	11:40	
Air Blank		0.000	11:41	
Control Test		0.080	11:41	
Air Blank		0.000	11:41	
Subject Sample	# 1		11:45	No Sample Provided
Breath Volume		0.000	11:45	
Air Blank		0.000	11:45	
Air Blank		0.000	11:47	
Subject Sample	# 2		11:47	Subject Test Refused
Breath Volume		0.000	11:47	
Air Blank		0.000	11:48	
Control Test		0.080	11:48	
Air Blank		0.000	11:49	
Diagnostics Check		OK	11:49	

Data Download Date/Time: 07/29/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: TOMKOW JR

DATE/TIME OF TEST: 07/19/2025 21:25:48

Diagnostics Check		OK	21:14	
Air Blank		0.000	21:14	
Control Test		0.080	21:15	
Air Blank		0.000	21:15	
Subject Sample	# 1		21:18	No Sample Provided
Breath Volume		0.000	21:18	
Air Blank		0.000	21:19	
Air Blank		0.000	21:21	
Subject Sample	# 2		21:24	No Sample Provided
Breath Volume		0.000	21:24	
Air Blank		0.000	21:24	
Control Test		0.081	21:25	
Air Blank		0.000	21:25	
Diagnostics Check		OK	21:25	

Data Download Date/Time: 07/29/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: CORDOVA RAMIREZ

DATE/TIME OF TEST: 07/20/2025 04:32:09

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.081	04:25
Air Blank		0.000	04:26
Subject Sample	# 1	0.156	04:27
Breath Volume		2.128	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:29
Subject Sample	# 2	0.149	04:30
Breath Volume		2.472	04:30
Air Blank		0.000	04:31
Control Test		0.079	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 07/29/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREWS,TYREE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: TORRES,ABRAHAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: GALFORD

DATE/TIME OF TEST: 07/20/2025 17:11:44

Diagnostics Check		OK	17:00	
Air Blank		0.000	17:00	
Control Test		0.079	17:01	
Air Blank		0.000	17:01	
Subject Sample	# 1		17:04	No Sample Provided
Breath Volume		0.000	17:04	
Air Blank		0.000	17:05	
Air Blank		0.000	17:07	
Subject Sample	# 2		17:10	No Sample Provided
Breath Volume		0.000	17:10	
Air Blank		0.000	17:10	
Control Test		0.079	17:11	
Air Blank		0.000	17:11	
Diagnostics Check		OK	17:11	

Data Download Date/Time: 07/29/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: JUAN PASCUAL

DATE/TIME OF TEST: 07/21/2025 01:43:35

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.128	01:38
Breath Volume		1.273	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.128	01:41
Breath Volume		1.105	01:41
Air Blank		0.000	01:42
Control Test		0.079	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 07/29/2025 05:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELLIDO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: PRIMITIVO III

DATE/TIME OF TEST: 07/21/2025 03:10:25

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:58	
Control Test		0.080	02:58	
Air Blank		0.000	02:59	
Subject Sample	# 1	0.089	03:02	Volume Not Met
Breath Volume		0.121	03:02	
Air Blank		0.000	03:03	
Air Blank		0.000	03:04	
Subject Sample	# 2	0.112	03:05	
Breath Volume		2.144	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 3	0.098	03:08	
Breath Volume		1.128	03:08	
Air Blank		0.000	03:09	
Control Test		0.079	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 07/29/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: CONTRERAS-RAMIREZ

DATE/TIME OF TEST: 07/22/2025 02:32:32

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.081	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.179	02:27
Breath Volume		1.906	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.175	02:30
Breath Volume		2.316	02:30
Air Blank		0.000	02:31
Control Test		0.078	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 07/29/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: SEDA

DATE/TIME OF TEST: 07/25/2025 02:21:37

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:10	
Control Test		0.080	02:10	
Air Blank		0.000	02:11	
Subject Sample	# 1		02:14	No Sample Provided
Breath Volume		0.000	02:14	
Air Blank		0.000	02:15	
Air Blank		0.000	02:17	
Subject Sample	# 2		02:20	No Sample Provided
Breath Volume		0.000	02:20	
Air Blank		0.000	02:20	
Control Test		0.080	02:21	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 07/29/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,LEONARDO,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BALL,GARY
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/25/2025 05:31:44

Diagnostics Check		OK	05:24
Air Blank		0.000	05:25
Control Test		0.081	05:25
Air Blank		0.000	05:26
Subject Sample	# 1	0.219	05:26
Breath Volume		2.976	05:26
Air Blank		0.000	05:27
Air Blank		0.000	05:29
Subject Sample	# 2	0.218	05:29
Breath Volume		2.382	05:29
Air Blank		0.000	05:30
Control Test		0.078	05:31
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 07/29/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MOWERY

DATE/TIME OF TEST: 07/25/2025 23:21:45

Diagnostics Check		OK	23:14
Air Blank		0.000	23:15
Control Test		0.079	23:15
Air Blank		0.000	23:15
Subject Sample	# 1	0.112	23:16
Breath Volume		4.082	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.108	23:20
Breath Volume		4.007	23:20
Air Blank		0.000	23:20
Control Test		0.078	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 07/29/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIELDS,NIKOLAUS,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANZALONE,ROBERT
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/26/2025 05:46:41

Diagnostics Check		OK	05:39
Air Blank		0.000	05:39
Control Test		0.080	05:39
Air Blank		0.000	05:40
Subject Sample	# 1	0.118	05:41
Breath Volume		3.070	05:41
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.117	05:45
Breath Volume		3.218	05:45
Air Blank		0.000	05:45
Control Test		0.079	05:46
Air Blank		0.000	05:46
Diagnostics Check		OK	05:46

Data Download Date/Time: 07/29/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIELDS,NIKOLAUS,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: SPILLER

DATE/TIME OF TEST: 07/26/2025 06:12:57

Diagnostics Check		OK	06:06
Air Blank		0.000	06:06
Control Test		0.080	06:07
Air Blank		0.000	06:07
Subject Sample	# 1	0.213	06:08
Breath Volume		2.632	06:08
Air Blank		0.000	06:08
Air Blank		0.000	06:10
Subject Sample	# 2	0.221	06:11
Breath Volume		2.492	06:11
Air Blank		0.000	06:12
Control Test		0.079	06:12
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 07/29/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DYE,BRANDON
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: MONTANEZ

DATE/TIME OF TEST: 07/26/2025 23:41:29

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.081	23:35
Air Blank		0.000	23:36
Subject Sample	# 1	0.131	23:36
Breath Volume		1.613	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.126	23:39
Breath Volume		1.925	23:39
Air Blank		0.000	23:40
Control Test		0.080	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 07/29/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: LA VECCHIO

DATE/TIME OF TEST: 07/27/2025 00:42:35

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.081	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.174	00:37
Breath Volume		1.882	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.174	00:40
Breath Volume		1.910	00:40
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 07/29/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLORES,MARIA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: CRUZ TORRES

DATE/TIME OF TEST: 07/27/2025 03:33:27

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.080	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.139	03:28
Breath Volume		2.890	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:30
Subject Sample	# 2	0.137	03:31
Breath Volume		2.140	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 07/29/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: ROMAN GONZALEZ

DATE/TIME OF TEST: 07/27/2025 05:10:11

Diagnostics Check		OK	05:03
Air Blank		0.000	05:04
Control Test		0.081	05:04
Air Blank		0.000	05:05
Subject Sample	# 1	0.078	05:05
Breath Volume		1.417	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:08
Subject Sample	# 2	0.077	05:08
Breath Volume		2.449	05:08
Air Blank		0.000	05:09
Control Test		0.080	05:09
Air Blank		0.000	05:10
Diagnostics Check		OK	05:10

Data Download Date/Time: 07/29/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: IACOBUCCI,JERE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: PARRA

DATE/TIME OF TEST: 07/27/2025 21:31:37

Diagnostics Check		OK	21:25	
Air Blank		0.000	21:25	
Control Test		0.080	21:26	
Air Blank		0.000	21:26	
Subject Sample	# 1	0.319	21:27	
Breath Volume		1.855	21:27	
Air Blank		0.000	21:28	
Air Blank		0.000	21:30	
Subject Sample	# 2	0.311	21:30	
Breath Volume		1.621	21:30	
Air Blank			21:30	RFI Detect
Air Blank		0.000	21:31	

Data Download Date/Time: 07/29/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: IACOBUCCI,JERE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: PARRA

DATE/TIME OF TEST: 07/28/2025 00:43:47

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.080	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.246	00:39
Breath Volume		1.335	00:39
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.246	00:42
Breath Volume		1.445	00:42
Air Blank		0.000	00:42
Control Test		0.079	00:43
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 07/29/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,PHILIP,P
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: KING,PHILIP
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 06/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/29/2025 01:06:08

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.081	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.000	01:01
Breath Volume		1.949	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.000	01:04
Breath Volume		1.792	01:04
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 07/29/2025 05:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: BAMACA-SALES

DATE/TIME OF TEST: 07/30/2025 18:23:17

Diagnostics Check		OK	18:16
Air Blank		0.000	18:16
Control Test		0.080	18:17
Air Blank		0.000	18:17
Subject Sample	# 1	0.217	18:18
Breath Volume		1.585	18:18
Air Blank		0.000	18:19
Air Blank		0.000	18:20
Subject Sample	# 2	0.210	18:21
Breath Volume		2.398	18:21
Air Blank		0.000	18:22
Control Test		0.080	18:22
Air Blank		0.000	18:23
Diagnostics Check		OK	18:23

Data Download Date/Time: 08/26/2025 08:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HOBBY

DATE/TIME OF TEST: 07/31/2025 03:38:11

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:27	
Control Test		0.080	03:27	
Air Blank		0.000	03:27	
Subject Sample	# 1		03:30	No Sample Provided
Breath Volume		0.000	03:30	
Air Blank		0.000	03:31	
Air Blank		0.000	03:33	
Subject Sample	# 2		03:36	No Sample Provided
Breath Volume		0.000	03:36	
Air Blank		0.000	03:37	
Control Test		0.078	03:37	
Air Blank		0.000	03:38	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 08/26/2025 08:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HOBBY

DATE/TIME OF TEST: 07/31/2025 04:26:14

Diagnostics Check		OK	04:14	
Air Blank		0.000	04:14	
Control Test		0.078	04:15	
Air Blank		0.000	04:15	
Subject Sample	# 1	0.163	04:18	Volume Not Met
Breath Volume		0.117	04:18	
Air Blank		0.000	04:19	
Air Blank		0.000	04:21	
Subject Sample	# 2	0.156	04:24	Volume Not Met
Breath Volume		0.164	04:24	
Air Blank		0.000	04:25	
Control Test		0.078	04:25	
Air Blank		0.000	04:26	
Diagnostics Check		OK	04:26	

Data Download Date/Time: 08/26/2025 08:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIELDS,NIKOLAUS,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DELMORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: MORALES-GREGORIO

DATE/TIME OF TEST: 07/31/2025 04:52:12

Diagnostics Check		OK	04:38	
Air Blank		0.000	04:38	
Control Test		0.077	04:39	
Air Blank		0.000	04:39	
Subject Sample	# 1	0.231	04:42	Volume Not Met
Breath Volume		0.878	04:42	
Air Blank		0.000	04:43	
Air Blank		0.000	04:45	
Subject Sample	# 2	0.209	04:47	
Breath Volume		1.183	04:47	
Air Blank		0.000	04:48	
Air Blank		0.000	04:49	
Subject Sample	# 3	0.216	04:50	
Breath Volume		1.203	04:50	
Air Blank		0.000	04:51	
Control Test		0.074	04:51	Control Outside Tolerance
Air Blank		0.000	04:52	

Data Download Date/Time: 08/26/2025 08:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIELDS,NIKOLAUS,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DELMORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: MORALES-GREGORIO

DATE/TIME OF TEST: 07/31/2025 05:28:08

Diagnostics Check		OK	05:21
Air Blank		0.000	05:22
Control Test		0.078	05:22
Air Blank		0.000	05:22
Subject Sample	# 1	0.209	05:23
Breath Volume		1.437	05:23
Air Blank		0.000	05:24
Air Blank		0.000	05:25
Subject Sample	# 2	0.209	05:26
Breath Volume		1.183	05:26
Air Blank		0.000	05:27
Control Test		0.078	05:27
Air Blank		0.000	05:28
Diagnostics Check		OK	05:28

Data Download Date/Time: 08/26/2025 08:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENMARK,HAROLD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SOTO,MAVERICK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 07/31/2025 19:59:57

Diagnostics Check		OK	19:53
Air Blank		0.000	19:53
Control Test		0.079	19:54
Air Blank		0.000	19:54
Subject Sample	# 1	0.264	19:55
Breath Volume		1.863	19:55
Air Blank		0.000	19:56
Air Blank		0.000	19:57
Subject Sample	# 2	0.262	19:58
Breath Volume		1.195	19:58
Air Blank		0.000	19:59
Control Test		0.078	19:59
Air Blank		0.000	19:59
Diagnostics Check		OK	19:59

Data Download Date/Time: 08/26/2025 08:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENMARK,HAROLD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: RAMOS-QUIAHUA

DATE/TIME OF TEST: 07/31/2025 20:12:11

Diagnostics Check		OK	20:04
Air Blank		0.000	20:04
Control Test		0.080	20:05
Air Blank		0.000	20:05
Subject Sample	# 1	0.098	20:07
Breath Volume		1.589	20:07
Air Blank		0.000	20:07
Air Blank		0.000	20:09
Subject Sample	# 2	0.095	20:10
Breath Volume		1.175	20:10
Air Blank		0.000	20:11
Control Test		0.078	20:11
Air Blank		0.000	20:12
Diagnostics Check		OK	20:12

Data Download Date/Time: 08/26/2025 08:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORTEZA,CRYSTAL,L
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: GEE,MICHAEL
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: DOOKIE

DATE/TIME OF TEST: 07/14/2025 20:53:05

Diagnostics Check		OK	20:45	
Air Blank		0.000	20:46	
Control Test		0.077	20:46	
Air Blank		0.000	20:47	
Subject Sample	# 1	0.223	20:47	
Breath Volume		2.058	20:47	
Air Blank		0.000	20:48	
Air Blank		0.000	20:50	
Subject Sample	# 2	0.214	20:51	
Breath Volume		2.160	20:51	
Air Blank			20:52	Purge Fail
Air Blank		0.000	20:53	

Data Download Date/Time: 09/04/2025 16:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORTEZA,CRYSTAL,L
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: GEE,MICHAEL
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: DOOKIE

DATE/TIME OF TEST: 07/14/2025 20:57:11

Diagnostics Check	OK	20:55	
Air Blank	0.000	20:56	
Control Test	0.073	20:56	Control Outside Tolerance
Air Blank	0.000	20:57	

Data Download Date/Time: 09/04/2025 16:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORTEZA,CRYSTAL,L
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: GEE,MICHAEL
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: DOOKIE

DATE/TIME OF TEST: 07/14/2025 21:06:41

Diagnostics Check		OK	20:59	
Air Blank		0.000	21:00	
Control Test		0.076	21:00	
Air Blank		0.000	21:01	
Subject Sample	# 1	0.222	21:01	
Breath Volume		2.453	21:01	
Air Blank		0.000	21:02	
Air Blank		0.000	21:04	
Subject Sample	# 2	0.217	21:05	
Breath Volume		2.585	21:05	
Air Blank		0.000	21:05	
Control Test		0.074	21:06	Control Outside Tolerance
Air Blank		0.000	21:06	

Data Download Date/Time: 09/04/2025 16:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 07/02/2025 03:00:34

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.078	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.000	02:56
Breath Volume		1.507	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.000	02:59
Breath Volume		1.800	02:59
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 07/10/2025 10:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HUSBAND

DATE/TIME OF TEST: 07/02/2025 03:58:48

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:56	
Control Test		0.079	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 1		03:57	Subject Test Refused
Breath Volume		0.000	03:57	
Air Blank		0.000	03:57	
Control Test		0.078	03:58	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:58	

Data Download Date/Time: 07/10/2025 10:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 07/03/2025 01:16:43

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:14	
Control Test		0.079	01:14	
Air Blank		0.000	01:15	
Subject Sample	# 1		01:15	Subject Test Refused
Breath Volume		0.000	01:15	
Air Blank		0.000	01:15	
Control Test		0.079	01:16	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:16	

Data Download Date/Time: 07/10/2025 10:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 07/03/2025 17:23:43

Diagnostics Check		OK	17:17
Air Blank		0.000	17:18
Control Test		0.078	17:18
Air Blank		0.000	17:18
Subject Sample	# 1	0.000	17:19
Breath Volume		2.437	17:19
Air Blank		0.000	17:19
Air Blank		0.000	17:21
Subject Sample	# 2	0.000	17:22
Breath Volume		2.578	17:22
Air Blank		0.000	17:22
Control Test		0.078	17:23
Air Blank		0.000	17:23
Diagnostics Check		OK	17:23

Data Download Date/Time: 07/10/2025 10:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 07/04/2025 01:31:16

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.078	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 1		01:29	Subject Test Refused
Breath Volume		0.000	01:29	
Air Blank		0.000	01:30	
Control Test		0.079	01:30	
Air Blank		0.000	01:31	
Diagnostics Check		OK	01:31	

Data Download Date/Time: 07/10/2025 10:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: GLISAN

DATE/TIME OF TEST: 07/04/2025 03:24:27

Diagnostics Check		OK	03:21	
Air Blank		0.000	03:22	
Control Test		0.079	03:22	
Air Blank		0.000	03:22	
Subject Sample	# 1		03:23	Subject Test Refused
Breath Volume		0.000	03:23	
Air Blank		0.000	03:23	
Control Test		0.079	03:23	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 07/10/2025 10:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: CLARE

DATE/TIME OF TEST: 07/04/2025 23:30:49

Diagnostics Check		OK	23:27	
Air Blank		0.000	23:28	
Control Test		0.078	23:28	
Air Blank		0.000	23:29	
Subject Sample	# 1		23:29	Subject Test Refused
Breath Volume		0.000	23:29	
Air Blank		0.000	23:29	
Control Test		0.079	23:30	
Air Blank		0.000	23:30	
Diagnostics Check		OK	23:30	

Data Download Date/Time: 07/10/2025 10:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/05/2025 02:07:21

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:04	
Control Test		0.080	02:05	
Air Blank		0.000	02:05	
Subject Sample	# 1		02:05	Subject Test Refused
Breath Volume		0.000	02:05	
Air Blank		0.000	02:06	
Control Test		0.079	02:06	
Air Blank		0.000	02:07	
Diagnostics Check		OK	02:07	

Data Download Date/Time: 07/10/2025 10:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 07/05/2025 04:12:23

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.079	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:05	No Sample Provided
Breath Volume		0.000	04:05	
Air Blank		0.000	04:05	
Air Blank		0.000	04:07	
Subject Sample	# 2		04:11	No Sample Provided
Breath Volume		0.000	04:11	
Air Blank		0.000	04:11	
Control Test		0.080	04:11	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:12	

Data Download Date/Time: 07/10/2025 10:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 07/05/2025 04:17:13

Diagnostics Check		OK	04:14	
Air Blank		0.000	04:14	
Control Test		0.078	04:15	
Air Blank		0.000	04:15	
Subject Sample	# 1		04:15	Subject Test Refused
Breath Volume		0.000	04:15	
Air Blank		0.000	04:16	
Control Test		0.078	04:16	
Air Blank		0.000	04:17	
Diagnostics Check		OK	04:17	

Data Download Date/Time: 07/10/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: PRYER

DATE/TIME OF TEST: 07/05/2025 05:12:58

Diagnostics Check		OK	05:04
Air Blank		0.000	05:05
Control Test		0.079	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.247	05:08
Breath Volume		1.683	05:08
Air Blank		0.000	05:08
Air Blank		0.000	05:10
Subject Sample	# 2	0.257	05:11
Breath Volume		2.117	05:11
Air Blank		0.000	05:12
Control Test		0.078	05:12
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 07/10/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: LUU

DATE/TIME OF TEST: 07/05/2025 06:12:01

Diagnostics Check		OK	05:57	
Air Blank		0.000	05:58	
Control Test		0.078	05:58	
Air Blank		0.000	05:58	
Subject Sample	# 1	0.251	06:00	No .020 agreement
Breath Volume		1.199	06:00	
Air Blank		0.000	06:01	
Air Blank		0.000	06:03	
Subject Sample	# 2	0.218	06:06	Slope Not Level
Breath Volume		1.367	06:06	
Air Blank		0.000	06:07	
Air Blank		0.000	06:08	
Subject Sample	# 3	0.226	06:10	No .020 agreement
Breath Volume		1.621	06:10	
Air Blank		0.000	06:11	
Control Test		0.076	06:11	
Air Blank		0.000	06:11	
Diagnostics Check		OK	06:12	

Data Download Date/Time: 07/10/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: LUU

DATE/TIME OF TEST: 07/05/2025 06:48:17

Diagnostics Check		OK	06:40	
Air Blank		0.000	06:41	
Control Test		0.078	06:41	
Air Blank		0.000	06:42	
Subject Sample	# 1	0.219	06:42	
Breath Volume		1.140	06:42	
Air Blank		0.000	06:43	
Air Blank		0.000	06:45	
Subject Sample	# 2		06:46	RFI Detect
Breath Volume		1.386	06:46	
Air Blank		0.000	06:47	
Control Test		0.078	06:47	
Air Blank		0.000	06:48	
Diagnostics Check		OK	06:48	

Data Download Date/Time: 07/10/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: LUU

DATE/TIME OF TEST: 07/05/2025 07:00:56

Diagnostics Check		OK	06:54
Air Blank		0.000	06:54
Control Test		0.078	06:55
Air Blank		0.000	06:55
Subject Sample	# 1	0.213	06:56
Breath Volume		1.820	06:56
Air Blank		0.000	06:56
Air Blank		0.000	06:58
Subject Sample	# 2	0.208	06:59
Breath Volume		1.214	06:59
Air Blank		0.000	07:00
Control Test		0.078	07:00
Air Blank		0.000	07:00
Diagnostics Check		OK	07:00

Data Download Date/Time: 07/10/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 07/05/2025 21:31:02

Diagnostics Check		OK	21:25
Air Blank		0.000	21:25
Control Test		0.079	21:25
Air Blank		0.000	21:26
Subject Sample	# 1	0.021	21:26
Breath Volume		1.867	21:26
Air Blank		0.000	21:27
Air Blank		0.000	21:29
Subject Sample	# 2	0.021	21:29
Breath Volume		1.500	21:29
Air Blank		0.000	21:30
Control Test		0.079	21:30
Air Blank		0.000	21:30
Diagnostics Check		OK	21:31

Data Download Date/Time: 07/10/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 07/05/2025 22:46:54

Diagnostics Check		OK	22:40
Air Blank		0.000	22:41
Control Test		0.078	22:41
Air Blank		0.000	22:41
Subject Sample	# 1	0.090	22:42
Breath Volume		2.230	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:44
Subject Sample	# 2	0.093	22:45
Breath Volume		1.976	22:45
Air Blank		0.000	22:45
Control Test		0.078	22:46
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 07/10/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: HARDWICK

DATE/TIME OF TEST: 07/06/2025 05:28:39

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.079	05:19
Air Blank		0.000	05:19
Subject Sample	# 1	0.227	05:20
Breath Volume		1.406	05:20
Air Blank		0.000	05:21
Air Blank		0.000	05:22
Subject Sample	# 2	0.201	05:23
Breath Volume		1.324	05:23
Air Blank		0.000	05:24
Air Blank		0.000	05:26
Subject Sample	# 3	0.208	05:27
Breath Volume		2.136	05:27
Air Blank		0.000	05:27
Control Test		0.078	05:28
Air Blank		0.000	05:28
Diagnostics Check		OK	05:28

Data Download Date/Time: 07/10/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: DAMPSON

DATE/TIME OF TEST: 07/06/2025 21:53:26

Diagnostics Check		OK	21:47
Air Blank		0.000	21:47
Control Test		0.079	21:47
Air Blank		0.000	21:48
Subject Sample	# 1	0.069	21:48
Breath Volume		1.187	21:48
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.067	21:51
Breath Volume		1.257	21:51
Air Blank		0.000	21:52
Control Test		0.078	21:52
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 07/10/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/07/2025 00:14:26

Diagnostics Check		OK	00:07
Air Blank		0.000	00:08
Control Test		0.078	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.135	00:09
Breath Volume		2.277	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.119	00:12
Breath Volume		1.554	00:12
Air Blank		0.000	00:13
Control Test		0.079	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 07/10/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: PROCTOR

DATE/TIME OF TEST: 07/08/2025 17:21:28

Diagnostics Check		OK	17:15
Air Blank		0.000	17:15
Control Test		0.078	17:16
Air Blank		0.000	17:16
Subject Sample	# 1	0.000	17:17
Breath Volume		1.558	17:17
Air Blank		0.000	17:17
Air Blank		0.000	17:19
Subject Sample	# 2	0.000	17:20
Breath Volume		1.597	17:20
Air Blank		0.000	17:20
Control Test		0.078	17:20
Air Blank		0.000	17:21
Diagnostics Check		OK	17:21

Data Download Date/Time: 07/10/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: BARBOUR

DATE/TIME OF TEST: 07/08/2025 18:32:04

Diagnostics Check		OK	18:25
Air Blank		0.000	18:26
Control Test		0.078	18:26
Air Blank		0.000	18:27
Subject Sample	# 1	0.075	18:27
Breath Volume		2.511	18:27
Air Blank		0.000	18:28
Air Blank		0.000	18:30
Subject Sample	# 2	0.072	18:30
Breath Volume		2.257	18:30
Air Blank		0.000	18:31
Control Test		0.078	18:31
Air Blank		0.000	18:31
Diagnostics Check		OK	18:32

Data Download Date/Time: 07/10/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SHAW,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/09/2025 04:58:18

Diagnostics Check		OK	04:51
Air Blank		0.000	04:52
Control Test		0.078	04:52
Air Blank		0.000	04:52
Subject Sample	# 1	0.000	04:53
Breath Volume		1.859	04:53
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.000	04:56
Breath Volume		1.339	04:56
Air Blank		0.000	04:57
Control Test		0.078	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 07/10/2025 10:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,R,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FUNES

DATE/TIME OF TEST: 07/24/2025 23:42:31

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.078	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.035	23:38
Breath Volume		1.578	23:38
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.036	23:40
Breath Volume		1.468	23:40
Air Blank		0.000	23:41
Control Test		0.078	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 08/20/2025 12:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,R,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ALEXANDRE,D
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/26/2025 00:19:26

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:11	
Control Test		0.078	00:11	
Air Blank		0.000	00:12	
Subject Sample	# 1		00:15	No Sample Provided
Breath Volume		0.000	00:15	
Air Blank		0.000	00:15	
Air Blank		0.000	00:17	
Subject Sample	# 2		00:18	Subject Test Refused
Breath Volume		0.000	00:18	
Air Blank		0.000	00:18	
Control Test		0.078	00:18	
Air Blank		0.000	00:19	
Diagnostics Check		OK	00:19	

Data Download Date/Time: 08/20/2025 12:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KLING,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ISKANDAR

DATE/TIME OF TEST: 07/26/2025 02:10:37

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:08	
Control Test		0.078	02:08	
Air Blank		0.000	02:09	
Subject Sample	# 1		02:09	Subject Test Refused
Breath Volume		0.000	02:09	
Air Blank		0.000	02:09	
Control Test		0.078	02:10	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 08/20/2025 12:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/26/2025 21:18:49

Diagnostics Check		OK	21:15	
Air Blank		0.000	21:16	
Control Test		0.079	21:16	
Air Blank		0.000	21:16	
Subject Sample	# 1		21:17	Subject Test Refused
Breath Volume		0.000	21:17	
Air Blank		0.000	21:17	
Control Test		0.078	21:18	
Air Blank		0.000	21:18	
Diagnostics Check		OK	21:18	

Data Download Date/Time: 08/20/2025 12:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ARVIN

DATE/TIME OF TEST: 07/27/2025 01:37:01

Diagnostics Check		OK	01:34	
Air Blank		0.000	01:34	
Control Test		0.078	01:34	
Air Blank		0.000	01:35	
Subject Sample	# 1		01:35	Subject Test Refused
Breath Volume		0.000	01:35	
Air Blank		0.000	01:36	
Control Test		0.078	01:36	
Air Blank		0.000	01:36	
Diagnostics Check		OK	01:37	

Data Download Date/Time: 08/20/2025 12:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:28

SUBJECT'S LAST NAME: TORRES TORRES

DATE/TIME OF TEST: 07/27/2025 15:55:10

Diagnostics Check		OK	15:48
Air Blank		0.000	15:48
Control Test		0.078	15:49
Air Blank		0.000	15:49
Subject Sample	# 1	0.091	15:50
Breath Volume		1.292	15:50
Air Blank		0.000	15:50
Air Blank		0.000	15:52
Subject Sample	# 2	0.088	15:53
Breath Volume		1.398	15:53
Air Blank		0.000	15:54
Control Test		0.078	15:54
Air Blank		0.000	15:55
Diagnostics Check		OK	15:55

Data Download Date/Time: 08/20/2025 12:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CHURCH

DATE/TIME OF TEST: 07/28/2025 02:07:09

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.080	02:02
Breath Volume		2.386	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:04
Subject Sample	# 2	0.081	02:05
Breath Volume		2.125	02:05
Air Blank		0.000	02:06
Control Test		0.078	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 08/20/2025 12:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGAN,D,O
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: LAGMAY

DATE/TIME OF TEST: 07/29/2025 17:48:09

Diagnostics Check		OK	17:40
Air Blank		0.000	17:41
Control Test		0.077	17:41
Air Blank		0.000	17:42
Subject Sample	# 1	0.000	17:43
Breath Volume		1.343	17:43
Air Blank		0.000	17:43
Air Blank		0.000	17:45
Subject Sample	# 2	0.000	17:46
Breath Volume		1.902	17:46
Air Blank		0.000	17:47
Control Test		0.077	17:47
Air Blank		0.000	17:48
Diagnostics Check		OK	17:48

Data Download Date/Time: 08/20/2025 12:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: OYLER

DATE/TIME OF TEST: 07/11/2025 04:46:59

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.077	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.203	04:42
Breath Volume		2.785	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:44
Subject Sample	# 2	0.212	04:45
Breath Volume		1.738	04:45
Air Blank		0.000	04:46
Control Test		0.077	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 08/08/2025 16:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PIERCE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: SUCH

DATE/TIME OF TEST: 07/13/2025 04:00:49

Diagnostics Check	OK	03:59	
Air Blank		04:00	Ambient Fail
Air Blank	0.000	04:00	

Data Download Date/Time: 08/08/2025 16:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PIERCE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: SUCH

DATE/TIME OF TEST: 07/13/2025 04:05:31

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:03	
Control Test		0.081	04:03	
Air Blank		0.000	04:04	
Subject Sample	# 1		04:04	Subject Test Refused
Breath Volume		0.000	04:04	
Air Blank		0.000	04:04	
Control Test		0.079	04:04	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 08/08/2025 16:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: THEMAS

DATE/TIME OF TEST: 07/14/2025 22:58:22

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:55	
Control Test		0.078	22:56	
Air Blank		0.000	22:56	
Subject Sample	# 1		22:57	Interferent Detect
Breath Volume		2.191	22:57	
Air Blank			22:57	Purge Fail
Air Blank		0.000	22:58	

Data Download Date/Time: 08/08/2025 16:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: THEMAS

DATE/TIME OF TEST: 07/14/2025 23:04:44

Diagnostics Check		OK	23:01	
Air Blank		0.000	23:01	
Control Test		0.079	23:01	
Air Blank		0.000	23:02	
Subject Sample	# 1		23:02	Interferent Detect
Breath Volume		2.078	23:02	
Air Blank		0.000	23:03	
Control Test		0.075	23:03	
Air Blank		0.000	23:04	
Diagnostics Check		Fail	23:04	Analytical Stability Test
Air Blank		0.000	23:04	

Data Download Date/Time: 08/08/2025 16:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: THEMAS

DATE/TIME OF TEST: 07/15/2025 01:03:39

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.046	00:58
Breath Volume		1.582	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.047	01:01
Breath Volume		1.765	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 08/08/2025 16:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROVENCHER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: PROVENCHER,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SEARLES

DATE/TIME OF TEST: 07/02/2025 00:49:46

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.082	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.000	00:45
Breath Volume		3.664	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.000	00:48
Breath Volume		4.265	00:48
Air Blank		0.000	00:48
Control Test		0.082	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 07/31/2025 00:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS,TYLER,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/08/2025 16:30:10

Diagnostics Check		OK	16:23
Air Blank		0.000	16:24
Control Test		0.081	16:24
Air Blank		0.000	16:25
Subject Sample	# 1	0.101	16:25
Breath Volume		1.441	16:25
Air Blank		0.000	16:26
Air Blank		0.000	16:28
Subject Sample	# 2	0.100	16:28
Breath Volume		1.531	16:28
Air Blank		0.000	16:29
Control Test		0.081	16:29
Air Blank		0.000	16:30
Diagnostics Check		OK	16:30

Data Download Date/Time: 07/31/2025 00:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROVENCHER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: PROVENCHER,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 07/11/2025 19:35:43

Diagnostics Check		OK	19:29
Air Blank		0.000	19:30
Control Test		0.080	19:30
Air Blank		0.000	19:31
Subject Sample	# 1	0.000	19:31
Breath Volume		3.406	19:31
Air Blank		0.000	19:31
Air Blank		0.000	19:33
Subject Sample	# 2	0.000	19:34
Breath Volume		3.468	19:34
Air Blank		0.000	19:34
Control Test		0.081	19:35
Air Blank		0.000	19:35
Diagnostics Check		OK	19:35

Data Download Date/Time: 07/31/2025 00:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLLER,LOGAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: KOLLER,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MOHR

DATE/TIME OF TEST: 07/12/2025 23:42:22

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:40	
Control Test		0.080	23:40	
Air Blank		0.000	23:41	
Subject Sample	# 1		23:41	Sequence Aborted
Breath Volume		0.000	23:41	
Air Blank		0.000	23:42	

Data Download Date/Time: 07/31/2025 00:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS,TYLER,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 07/21/2025 17:30:20

Diagnostics Check		OK	17:24
Air Blank		0.000	17:24
Control Test		0.081	17:25
Air Blank		0.000	17:25
Subject Sample	# 1	0.128	17:26
Breath Volume		2.402	17:26
Air Blank		0.000	17:26
Air Blank		0.000	17:28
Subject Sample	# 2	0.120	17:28
Breath Volume		1.746	17:28
Air Blank		0.000	17:29
Control Test		0.081	17:29
Air Blank		0.000	17:30
Diagnostics Check		OK	17:30

Data Download Date/Time: 07/31/2025 00:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,BRETT,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WAGNER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 07/08/2025 20:57:54

Diagnostics Check		OK	20:51
Air Blank		0.000	20:52
Control Test		0.079	20:52
Air Blank		0.000	20:52
Subject Sample	# 1	0.205	20:53
Breath Volume		3.484	20:53
Air Blank		0.000	20:53
Air Blank		0.000	20:55
Subject Sample	# 2	0.207	20:56
Breath Volume		3.453	20:56
Air Blank		0.000	20:56
Control Test		0.079	20:57
Air Blank		0.000	20:57
Diagnostics Check		OK	20:57

Data Download Date/Time: 08/08/2025 15:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREEMAN,AUSTIN,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FREEMAN,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/02/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 11:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:04

SUBJECT'S LAST NAME: CUEVAR RIVAS

DATE/TIME OF TEST: 07/20/2025 12:33:14

Diagnostics Check		OK	12:25
Air Blank		0.000	12:25
Control Test		0.079	12:25
Air Blank		0.000	12:26
Subject Sample	# 1	0.178	12:27
Breath Volume		1.113	12:27
Air Blank		0.000	12:28
Air Blank		0.000	12:29
Subject Sample	# 2	0.177	12:31
Breath Volume		1.718	12:31
Air Blank		0.000	12:32
Control Test		0.079	12:32
Air Blank		0.000	12:33
Diagnostics Check		OK	12:33

Data Download Date/Time: 08/08/2025 15:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,BRETT,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WAGNER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: HERCULES

DATE/TIME OF TEST: 07/25/2025 23:13:10

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.080	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.197	23:08
Breath Volume		2.394	23:08
Air Blank		0.000	23:09
Air Blank		0.000	23:10
Subject Sample	# 2	0.200	23:11
Breath Volume		3.195	23:11
Air Blank		0.000	23:12
Control Test		0.080	23:12
Air Blank		0.000	23:13
Diagnostics Check		OK	23:13

Data Download Date/Time: 08/08/2025 15:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,BRETT,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WAGNER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: HESLER

DATE/TIME OF TEST: 07/26/2025 06:13:57

Diagnostics Check		OK	06:05
Air Blank		0.000	06:06
Control Test		0.081	06:06
Air Blank		0.000	06:07
Subject Sample	# 1	0.156	06:09
Breath Volume		2.914	06:09
Air Blank		0.000	06:10
Air Blank		0.000	06:12
Subject Sample	# 2	0.162	06:12
Breath Volume		2.707	06:12
Air Blank		0.000	06:13
Control Test		0.078	06:13
Air Blank		0.000	06:13
Diagnostics Check		OK	06:13

Data Download Date/Time: 08/08/2025 15:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: BERG

DATE/TIME OF TEST: 07/28/2025 02:48:40

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:46	
Control Test		0.079	02:46	
Air Blank		0.000	02:46	
Subject Sample	# 1		02:47	Subject Test Refused
Breath Volume		0.000	02:47	
Air Blank		0.000	02:47	
Control Test		0.080	02:48	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 08/08/2025 15:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,RICHARD,S
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: YARBROUGH,JONATHAN
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BYMAN

DATE/TIME OF TEST: 07/18/2025 23:14:35

Diagnostics Check		OK	23:07
Air Blank		0.000	23:08
Control Test		0.081	23:08
Air Blank		0.000	23:09
Subject Sample	# 1	0.231	23:10
Breath Volume		1.675	23:10
Air Blank		0.000	23:10
Air Blank		0.000	23:12
Subject Sample	# 2	0.236	23:13
Breath Volume		1.281	23:13
Air Blank		0.000	23:13
Control Test		0.080	23:14
Air Blank		0.000	23:14
Diagnostics Check		OK	23:14

Data Download Date/Time: 07/21/2025 05:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WOODIN,STEPHEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: RUST

DATE/TIME OF TEST: 07/12/2025 19:31:02

Diagnostics Check		OK	19:24
Air Blank		0.000	19:24
Control Test		0.080	19:24
Air Blank		0.000	19:25
Subject Sample	# 1	0.000	19:26
Breath Volume		1.957	19:26
Air Blank		0.000	19:27
Air Blank		0.000	19:29
Subject Sample	# 2	0.000	19:29
Breath Volume		1.285	19:29
Air Blank		0.000	19:30
Control Test		0.078	19:30
Air Blank		0.000	19:30
Diagnostics Check		OK	19:31

Data Download Date/Time: 07/30/2025 16:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTZER,ETHAN,E
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MARTIN,BAILEY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/13/2025 17:11:54

Diagnostics Check	OK	17:10	
Air Blank	0.000	17:11	
Control Test	0.081	17:11	
Air Blank		17:11	RFI Detect
Air Blank		17:11	RFI Detect

Data Download Date/Time: 07/30/2025 16:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTZER,ETHAN,E
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MARTIN,BAILEY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/13/2025 17:17:30

Diagnostics Check		OK	17:14	
Air Blank		0.000	17:14	
Control Test		0.080	17:14	
Air Blank		0.000	17:15	
Subject Sample	# 1	0.000	17:16	
Breath Volume		2.925	17:16	
Air Blank			17:16	RFI Detect
Air Blank		0.000	17:16	
Control Test		0.081	17:16	
Air Blank		0.000	17:17	
Diagnostics Check		OK	17:17	

Data Download Date/Time: 07/30/2025 16:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTZER,ETHAN,E
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MARTIN,BAILEY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/13/2025 17:31:36

Diagnostics Check		OK	17:25
Air Blank		0.000	17:25
Control Test		0.079	17:26
Air Blank		0.000	17:26
Subject Sample	# 1	0.000	17:27
Breath Volume		3.050	17:27
Air Blank		0.000	17:27
Air Blank		0.000	17:29
Subject Sample	# 2	0.000	17:30
Breath Volume		3.972	17:30
Air Blank		0.000	17:30
Control Test		0.079	17:31
Air Blank		0.000	17:31
Diagnostics Check		OK	17:31

Data Download Date/Time: 07/30/2025 16:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: KOSESKI

DATE/TIME OF TEST: 07/18/2025 21:23:16

Diagnostics Check		OK	21:20	
Air Blank		0.000	21:20	
Control Test		0.081	21:21	
Air Blank		0.000	21:21	
Subject Sample	# 1		21:21	Subject Test Refused
Breath Volume		0.000	21:21	
Air Blank		0.000	21:22	
Control Test		0.081	21:22	
Air Blank		0.000	21:23	
Diagnostics Check		OK	21:23	

Data Download Date/Time: 07/30/2025 16:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: ARREDONDO-SANCHEZ

DATE/TIME OF TEST: 07/20/2025 19:40:33

Diagnostics Check		OK	19:34
Air Blank		0.000	19:34
Control Test		0.080	19:35
Air Blank		0.000	19:35
Subject Sample	# 1	0.213	19:36
Breath Volume		2.078	19:36
Air Blank		0.000	19:36
Air Blank		0.000	19:38
Subject Sample	# 2	0.203	19:39
Breath Volume		2.746	19:39
Air Blank		0.000	19:39
Control Test		0.080	19:39
Air Blank		0.000	19:40
Diagnostics Check		OK	19:40

Data Download Date/Time: 07/30/2025 16:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: KING,BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/20/2025 20:38:09

Diagnostics Check		OK	20:31
Air Blank		0.000	20:31
Control Test		0.079	20:32
Air Blank		0.000	20:32
Subject Sample	# 1	0.223	20:33
Breath Volume		1.492	20:33
Air Blank		0.000	20:34
Air Blank		0.000	20:35
Subject Sample	# 2	0.223	20:36
Breath Volume		1.429	20:36
Air Blank		0.000	20:37
Control Test		0.080	20:37
Air Blank		0.000	20:38
Diagnostics Check		OK	20:38

Data Download Date/Time: 07/30/2025 16:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GARDNER,JAVON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 07/22/2025 03:01:48

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.192	02:57
Breath Volume		2.718	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.202	03:00
Breath Volume		1.933	03:00
Air Blank		0.000	03:00
Control Test		0.079	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 07/30/2025 16:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CLARK,ADDISON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: CARRASQUILLO CRUZ

DATE/TIME OF TEST: 07/27/2025 22:38:03

Diagnostics Check		OK	22:31
Air Blank		0.000	22:32
Control Test		0.080	22:32
Air Blank		0.000	22:33
Subject Sample	# 1	0.146	22:33
Breath Volume		3.699	22:33
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 2	0.149	22:36
Breath Volume		3.347	22:36
Air Blank		0.000	22:37
Control Test		0.080	22:37
Air Blank		0.000	22:37
Diagnostics Check		OK	22:38

Data Download Date/Time: 07/30/2025 16:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 07/29/2025 23:10:07

Diagnostics Check		OK	23:04
Air Blank		0.000	23:04
Control Test		0.080	23:04
Air Blank		0.000	23:05
Subject Sample	# 1	0.000	23:05
Breath Volume		1.675	23:05
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.000	23:08
Breath Volume		1.449	23:08
Air Blank		0.000	23:09
Control Test		0.080	23:09
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 07/30/2025 16:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.