Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OVIEDO,ANDRES,J
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: OVIEDO,ANDRES
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: ZAMBRANO

DATE/TIME OF TEST: 08/02/2025 13:48:50

Diagnostics Check		OK	13:40	
Air Blank		0.000	13:41	
Control Test		0.075	13:41	
Air Blank		0.000	13:42	
Subject Sample	# 1	0.232	13:44	
Breath Volume		2.902	13:44	
Air Blank		0.000	13:45	
Air Blank		0.000	13:46	
Subject Sample	# 2	0.224	13:47	
Breath Volume		1.515	13:47	
Air Blank		0.000	13:47	
Control Test		0.059	13:48	Control Outside Tolerance
Air Blank		0.000	13:48	

Data Download Date/Time: 08/07/2025 05:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/03/2025 04:13:24

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.165	04:08
Breath Volume		1.527	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.160	04:11
Breath Volume		1.472	04:11
Air Blank		0.000	04:12
Control Test		0.078	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 08/07/2025 05:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: DERAS NAVARRO

DATE/TIME OF TEST: 08/03/2025 04:23:11

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:20	
Control Test		0.079	04:20	
Air Blank		0.000	04:21	
Subject Sample	# 1		04:21	Subject Test Refused
Breath Volume		0.000	04:21	
Air Blank		0.000	04:22	
Control Test		0.079	04:22	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 08/07/2025 05:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: HICKEY,J
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: ESPINALPEREZ

DATE/TIME OF TEST: 08/03/2025 18:43:05

Diagnostics Check		OK	18:40	
Air Blank		0.000	18:40	
Control Test		0.078	18:40	
Air Blank		0.000	18:41	
Subject Sample	# 1		18:41	Subject Test Refused
Breath Volume		0.000	18:41	
Air Blank		0.000	18:42	
Control Test		0.079	18:42	
Air Blank		0.000	18:43	
Diagnostics Check		OK	18:43	

Data Download Date/Time: 08/07/2025 05:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEBO,KEVIN,
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: FEBO,KEVIN
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: TURMERO

DATE/TIME OF TEST: 08/06/2025 06:34:45

Diagnostics Check		OK	06:28	
Air Blank		0.000	06:28	
Control Test		0.079	06:28	
Air Blank		0.000	06:29	
Subject Sample	# 1	0.159	06:30	
Breath Volume		3.023	06:30	
Air Blank		0.000	06:31	
Air Blank		0.000	06:33	
Subject Sample	# 2		06:33	RFI Detect
Breath Volume		0.000	06:33	
Air Blank		0.000	06:33	
Control Test		0.077	06:34	
Air Blank		0.000	06:34	
Diagnostics Check		OK	06:34	

Data Download Date/Time: 08/07/2025 05:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEBO,KEVIN,
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: FEBO,KEVIN
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: TURMERO

DATE/TIME OF TEST: 08/06/2025 06:42:55

Diagnostics Check		OK	06:36
Air Blank		0.000	06:37
Control Test		0.079	06:37
Air Blank		0.000	06:38
Subject Sample	# 1	0.150	06:38
Breath Volume		1.726	06:38
Air Blank		0.000	06:39
Air Blank		0.000	06:40
Subject Sample	# 2	0.145	06:41
Breath Volume		1.917	06:41
Air Blank		0.000	06:42
Control Test		0.078	06:42
Air Blank		0.000	06:42
Diagnostics Check		OK	06:42

Data Download Date/Time: 08/07/2025 05:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAPE,RYAN,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LAPE,RYAN
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 08/07/2025 23:20:24

Diagnostics Check		OK	23:17	
Air Blank		0.000	23:17	
Control Test		0.079	23:18	
Air Blank		0.000	23:18	
Subject Sample	# 1		23:19	Subject Test Refused
Breath Volume		0.000	23:19	
Air Blank		0.000	23:19	
Control Test		0.079	23:19	
Air Blank		0.000	23:20	
Diagnostics Check		OK	23:20	

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 08/08/2025 05:01:36

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.079	04:56
Air Blank		0.000	04:56
Subject Sample	# 1	0.169	04:57
Breath Volume		1.769	04:57
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 2	0.159	05:00
Breath Volume		1.246	05:00
Air Blank		0.000	05:00
Control Test		0.077	05:01
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: NANCE

DATE/TIME OF TEST: 08/09/2025 05:34:40

Diagnostics Check		OK	05:22	
Air Blank		0.000	05:23	
Control Test		0.079	05:23	
Air Blank		0.000	05:24	
Subject Sample	# 1	0.165	05:27	Volume Not Met
Breath Volume		0.085	05:27	
Air Blank		0.000	05:28	
Air Blank		0.000	05:29	
Subject Sample	# 2	0.132	05:32	Volume Not Met
Breath Volume		0.402	05:32	
Air Blank		0.000	05:33	
Control Test		0.079	05:34	
Air Blank		0.000	05:34	
Diagnostics Check		OK	05:34	

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: NANCE

DATE/TIME OF TEST: 08/09/2025 06:02:51

Diagnostics Check		OK	05:55
Air Blank		0.000	05:55
Control Test		0.077	05:56
Air Blank		0.000	05:56
Subject Sample	# 1	0.159	05:57
Breath Volume		1.355	05:57
Air Blank		0.000	05:58
Air Blank		0.000	06:00
Subject Sample	# 2	0.150	06:01
Breath Volume		1.863	06:01
Air Blank		0.000	06:01
Control Test		0.079	06:02
Air Blank		0.000	06:02
Diagnostics Check		OK	06:02

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OVIEDO,ANDRES,J
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: OVIEDO,ANDRES
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MONTEALEGRE

DATE/TIME OF TEST: 08/15/2025 00:28:13

Diagnostics Check		OK	00:25	
Air Blank		0.000	00:25	
Control Test		0.078	00:26	
Air Blank		0.000	00:26	
Subject Sample	# 1		00:26	Subject Test Refused
Breath Volume		0.000	00:26	
Air Blank		0.000	00:27	
Control Test		0.078	00:27	
Air Blank		0.000	00:28	
Diagnostics Check		OK	00:28	

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 08/15/2025 03:09:25

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:05	
Control Test		0.079	03:06	
Air Blank		0.000	03:06	
Subject Sample	# 1		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:08	
Control Test		0.079	03:08	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLARKE,M
ARRESTING OFFICER AGENCY: CORAL GABLES PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 08/17/2025 00:17:44

Diagnostics Check		OK	00:14	
Air Blank		0.000	00:15	
Control Test		0.079	00:15	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:16	Subject Test Refused
Breath Volume		0.000	00:16	
Air Blank		0.000	00:16	
Control Test		0.079	00:17	
Air Blank		0.000	00:17	
Diagnostics Check		OK	00:17	

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PEREZ,A
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: SIERRA AGUILA

DATE/TIME OF TEST: 08/17/2025 01:54:38

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.078	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.192	01:49
Breath Volume		2.406	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.191	01:53
Breath Volume		2.398	01:53
Air Blank		0.000	01:53
Control Test		0.079	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FRIAS,KEVIN
ARRESTING OFFICER AGENCY: MIAMI SPRINGS PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: SOLIS ZEPEDA

DATE/TIME OF TEST: 08/17/2025 03:49:34

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.079	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.206	03:45
Breath Volume		1.554	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.221	03:48
Breath Volume		2.160	03:48
Air Blank		0.000	03:48
Control Test		0.079	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GONZALES,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: AVILA LABRADA

DATE/TIME OF TEST: 08/17/2025 05:34:38

Diagnostics Check		OK	05:23
Air Blank		0.000	05:24
Control Test		0.078	05:24
Air Blank		0.000	05:25
Subject Sample	# 1	0.184	05:25
Breath Volume		2.265	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.163	05:30
Breath Volume		2.140	05:30
Air Blank		0.000	05:30
Air Blank		0.000	05:32
Subject Sample	# 3	0.158	05:33
Breath Volume		2.179	05:33
Air Blank		0.000	05:33
Control Test		0.077	05:34
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,ALEXIS,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GARCIA,ALEXIS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 08/18/2025 04:08:16

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.079	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.180	04:03
Breath Volume		1.828	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.178	04:06
Breath Volume		1.554	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PIULATS,M
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ARBELAEZ MONTOYA

DATE/TIME OF TEST: 08/22/2025 04:21:56

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:12	
Control Test		0.079	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1	0.285	04:16	Volume Not Met
Breath Volume		0.378	04:16	
Air Blank		0.000	04:17	
Air Blank		0.000	04:19	
Subject Sample	# 2		04:20	Subject Test Refused
Breath Volume		0.000	04:20	
Air Blank		0.000	04:21	
Control Test		0.076	04:21	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OCON,L
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 08/23/2025 07:42:48

Diagnostics Check		OK	07:39	
Air Blank		0.000	07:39	
Control Test		0.079	07:40	
Air Blank		0.000	07:40	
Subject Sample	# 1		07:41	Subject Test Refused
Breath Volume		0.000	07:41	
Air Blank		0.000	07:41	
Control Test		0.079	07:42	
Air Blank		0.000	07:42	
Diagnostics Check		OK	07:42	

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,WILFREDO,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ORTEGA,WILFREDO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: MUSTELL FERREIRO

DATE/TIME OF TEST: 08/24/2025 07:39:52

Diagnostics Check		OK	07:33
Air Blank		0.000	07:33
Control Test		0.079	07:33
Air Blank		0.000	07:34
Subject Sample	# 1	0.147	07:35
Breath Volume		1.734	07:35
Air Blank		0.000	07:36
Air Blank		0.000	07:37
Subject Sample	# 2	0.137	07:38
Breath Volume		1.652	07:38
Air Blank		0.000	07:38
Control Test		0.077	07:39
Air Blank		0.000	07:39
Diagnostics Check		OK	07:39

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: CRESPO

DATE/TIME OF TEST: 08/27/2025 01:43:24

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.336	01:38
Breath Volume		1.460	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.347	01:41
Breath Volume		1.804	01:41
Air Blank		0.000	01:42
Control Test		0.078	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,WILFREDO,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 08/28/2025 21:18:23

Diagnostics Check		OK	21:10
Air Blank		0.000	21:10
Control Test		0.078	21:11
Air Blank		0.000	21:11
Subject Sample	# 1	0.128	21:13
Breath Volume		1.242	21:13
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.127	21:16
Breath Volume		1.300	21:16
Air Blank		0.000	21:17
Control Test		0.078	21:17
Air Blank		0.000	21:18
Diagnostics Check		OK	21:18

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: SANABRIA,M
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: YATETORRES

DATE/TIME OF TEST: 08/29/2025 23:34:58

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.078	23:29
Air Blank		0.000	23:29
Subject Sample	# 1	0.184	23:30
Breath Volume		1.113	23:30
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.189	23:33
Breath Volume		1.769	23:33
Air Blank		0.000	23:34
Control Test		0.078	23:34
Air Blank		0.000	23:34
Diagnostics Check		OK	23:34

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VILLEGAS,S
ARRESTING OFFICER AGENCY: SURFSIDE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: URIETA

DATE/TIME OF TEST: 08/30/2025 03:36:44

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:32	
Control Test		0.078	03:32	
Air Blank		0.000	03:33	
Subject Sample	# 1		03:35	Subject Test Refused
Breath Volume		0.000	03:35	
Air Blank		0.000	03:35	
Control Test		0.079	03:36	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:36	

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: MASSO,J
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: LAY

DATE/TIME OF TEST: 08/30/2025 20:52:24

Diagnostics Check		OK	20:49	
Air Blank		0.000	20:49	
Control Test		0.079	20:50	
Air Blank		0.000	20:50	
Subject Sample	# 1		20:51	Subject Test Refused
Breath Volume		0.000	20:51	
Air Blank		0.000	20:51	
Control Test		0.079	20:51	
Air Blank		0.000	20:52	
Diagnostics Check		OK	20:52	

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEANDRE,GERALD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LEANDRE,GERALD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: LACAYO

DATE/TIME OF TEST: 08/31/2025 04:25:23

Diagnostics Check		OK	04:22	
Air Blank		0.000	04:23	
Control Test		0.079	04:23	
Air Blank		0.000	04:23	
Subject Sample	# 1		04:24	Subject Test Refused
Breath Volume		0.000	04:24	
Air Blank		0.000	04:24	
Control Test		0.079	04:24	
Air Blank		0.000	04:25	
Diagnostics Check		OK	04:25	

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MORENO,FRANCISCO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:02

SUBJECT'S LAST NAME: ZABALETA

DATE/TIME OF TEST: 08/31/2025 08:29:59

Diagnostics Check		OK	08:23
Air Blank		0.000	08:23
Control Test		0.079	08:23
Air Blank		0.000	08:24
Subject Sample	# 1	0.168	08:25
Breath Volume		1.609	08:25
Air Blank		0.000	08:26
Air Blank		0.000	08:27
Subject Sample	# 2	0.175	08:28
Breath Volume		1.890	08:28
Air Blank		0.000	08:29
Control Test		0.078	08:29
Air Blank		0.000	08:29
Diagnostics Check		OK	08:29

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIZ,JOAQUIN,G
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:36

SUBJECT'S LAST NAME: RAMIREZFRANCISCO

DATE/TIME OF TEST: 08/31/2025 09:04:57

Diagnostics Check		OK	09:01	
Air Blank		0.000	09:01	
Control Test		0.079	09:02	
Air Blank		0.000	09:02	
Subject Sample	# 1	0.180	09:03	
Breath Volume		1.800	09:03	
Air Blank			09:03	RFI Detect
Air Blank		0.000	09:04	
Control Test		0.079	09:04	
Air Blank		0.000	09:04	
Diagnostics Check		OK	09:04	

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIZ,JOAQUIN,G
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:36

SUBJECT'S LAST NAME: RAMIREZFRANCISCO

DATE/TIME OF TEST: 08/31/2025 09:14:05

Diagnostics Check		OK	09:07
Air Blank		0.000	09:08
Control Test		0.078	09:08
Air Blank		0.000	09:09
Subject Sample	# 1	0.170	09:09
Breath Volume		1.183	09:09
Air Blank		0.000	09:10
Air Blank		0.000	09:12
Subject Sample	# 2	0.177	09:12
Breath Volume		1.796	09:12
Air Blank		0.000	09:13
Control Test		0.079	09:13
Air Blank		0.000	09:14
Diagnostics Check		OK	09:14

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: ACOSTA GONZALEZ

DATE/TIME OF TEST: 08/31/2025 23:51:32

Diagnostics Check		OK	23:45
Air Blank		0.000	23:45
Control Test		0.079	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.000	23:47
Breath Volume		1.941	23:47
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.000	23:50
Breath Volume		1.726	23:50
Air Blank		0.000	23:50
Control Test		0.079	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 09/04/2025 05:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: FACEY

DATE/TIME OF TEST: 08/23/2025 00:44:25

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.079	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.221	00:39
Breath Volume		1.789	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.225	00:42
Breath Volume		1.800	00:42
Air Blank		0.000	00:43
Control Test		0.078	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 09/04/2025 05:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: AVILES CHALEN

DATE/TIME OF TEST: 08/24/2025 03:54:41

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.079	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.081	03:50
Breath Volume		1.183	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.084	03:53
Breath Volume		1.914	03:53
Air Blank		0.000	03:53
Control Test		0.079	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 09/04/2025 05:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: CALVO

DATE/TIME OF TEST: 08/31/2025 02:48:24

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.079	02:42
Air Blank		0.000	02:43
Subject Sample	# 1	0.124	02:43
Breath Volume		1.769	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.125	02:46
Breath Volume		1.199	02:46
Air Blank		0.000	02:47
Control Test		0.080	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 09/04/2025 05:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: KITCHEN

DATE/TIME OF TEST: 08/04/2025 00:30:00

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.078	00:16	
Air Blank		0.000	00:16	
Subject Sample	# 1	0.204	00:19	Slope Not Level
Breath Volume		1.378	00:19	
Air Blank		0.000	00:20	
Air Blank		0.000	00:22	
Subject Sample	# 2	0.207	00:24	
Breath Volume		1.585	00:24	
Air Blank		0.000	00:25	
Air Blank		0.000	00:27	
Subject Sample	# 3	0.196	00:28	
Breath Volume		1.457	00:28	
Air Blank		0.000	00:29	
Control Test		0.077	00:29	
Air Blank		0.000	00:29	
Diagnostics Check		OK	00:30	

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MAYS

DATE/TIME OF TEST: 08/06/2025 03:11:53

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.079	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.086	03:07
Breath Volume		2.031	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.092	03:10
Breath Volume		2.417	03:10
Air Blank		0.000	03:10
Control Test		0.078	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 08/07/2025 00:59:35

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:57	
Control Test		0.078	00:57	
Air Blank		0.000	00:58	
Subject Sample	# 1		00:58	Subject Test Refused
Breath Volume		0.000	00:58	
Air Blank		0.000	00:58	
Control Test		0.078	00:59	
Air Blank		0.000	00:59	
Diagnostics Check		OK	00:59	

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: RUSTOM

DATE/TIME OF TEST: 08/07/2025 02:27:46

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.079	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.106	02:23
Breath Volume		2.027	02:23
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.112	02:26
Breath Volume		3.375	02:26
Air Blank		0.000	02:26
Control Test		0.078	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: JARMAN,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/07/2025 23:31:04

Diagnostics Check		OK	23:24
Air Blank		0.000	23:25
Control Test		0.078	23:25
Air Blank		0.000	23:25
Subject Sample	# 1	0.186	23:26
Breath Volume		1.339	23:26
Air Blank		0.000	23:27
Air Blank		0.000	23:29
Subject Sample	# 2	0.188	23:29
Breath Volume		1.210	23:29
Air Blank		0.000	23:30
Control Test		0.078	23:30
Air Blank		0.000	23:30
Diagnostics Check		OK	23:31

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: SEARCY

DATE/TIME OF TEST: 08/08/2025 03:09:27

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:07	
Control Test		0.078	03:07	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:08	
Control Test		0.078	03:08	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/09/2025 00:35:01

Diagnostics Check		OK	00:32	
Air Blank		0.000	00:32	
Control Test		0.078	00:33	
Air Blank		0.000	00:33	
Subject Sample	# 1		00:33	Subject Test Refused
Breath Volume		0.000	00:33	
Air Blank		0.000	00:34	
Control Test		0.078	00:34	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:35	

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BATEH,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 08/09/2025 04:30:43

Diagnostics Check		OK	04:27	
Air Blank		0.000	04:28	
Control Test		0.078	04:28	
Air Blank		0.000	04:29	
Subject Sample	# 1		04:29	Subject Test Refused
Breath Volume		0.000	04:29	
Air Blank		0.000	04:29	
Control Test		0.079	04:30	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:30	

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WALLS

DATE/TIME OF TEST: 08/10/2025 01:37:15

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.078	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.079	01:32
Breath Volume		1.761	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:34
Subject Sample	# 2	0.079	01:35
Breath Volume		1.449	01:35
Air Blank		0.000	01:36
Control Test		0.078	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: DENICOLA

DATE/TIME OF TEST: 08/10/2025 04:15:44

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:13	
Control Test		0.078	04:13	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:14	Subject Test Refused
Breath Volume		0.000	04:14	
Air Blank		0.000	04:14	
Control Test		0.078	04:15	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ABELLANA

DATE/TIME OF TEST: 08/11/2025 03:29:36

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.097	03:24
Breath Volume		1.304	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.098	03:28
Breath Volume		1.386	03:28
Air Blank		0.000	03:28
Control Test		0.078	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARRAR,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:17

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/11/2025 08:44:17

Diagnostics Check		OK	08:38
Air Blank		0.000	08:38
Control Test		0.078	08:38
Air Blank		0.000	08:39
Subject Sample	# 1	0.290	08:39
Breath Volume		1.949	08:39
Air Blank		0.000	08:40
Air Blank		0.000	08:42
Subject Sample	# 2	0.296	08:42
Breath Volume		1.871	08:42
Air Blank		0.000	08:43
Control Test		0.078	08:43
Air Blank		0.000	08:44
Diagnostics Check		OK	08:44

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 08/14/2025 04:23:48

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.079	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.106	04:19
Breath Volume		2.089	04:19
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.103	04:22
Breath Volume		2.164	04:22
Air Blank		0.000	04:22
Control Test		0.078	04:23
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 08/14/2025 20:18:35

Diagnostics Check		OK	20:12
Air Blank		0.000	20:13
Control Test		0.078	20:13
Air Blank		0.000	20:13
Subject Sample	# 1	0.214	20:14
Breath Volume		1.867	20:14
Air Blank		0.000	20:14
Air Blank		0.000	20:16
Subject Sample	# 2	0.210	20:17
Breath Volume		1.664	20:17
Air Blank		0.000	20:17
Control Test		0.079	20:17
Air Blank		0.000	20:18
Diagnostics Check		OK	20:18

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KLING,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: TURLEY

DATE/TIME OF TEST: 08/15/2025 00:21:42

Diagnostics Check		OK	00:18	
Air Blank		0.000	00:19	
Control Test		0.078	00:19	
Air Blank		0.000	00:20	
Subject Sample	# 1		00:20	Subject Test Refused
Breath Volume		0.000	00:20	
Air Blank		0.000	00:20	
Control Test		0.078	00:21	
Air Blank		0.000	00:21	
Diagnostics Check		OK	00:21	

Data Download Date/Time: 08/19/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: JOHNSON,I
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WEEKS

DATE/TIME OF TEST: 08/15/2025 03:46:34

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:44	
Control Test		0.079	03:44	
Air Blank		0.000	03:45	
Subject Sample	# 1		03:45	Subject Test Refused
Breath Volume		0.000	03:45	
Air Blank		0.000	03:45	
Control Test		0.079	03:45	
Air Blank		0.000	03:46	
Diagnostics Check		OK	03:46	

Data Download Date/Time: 08/19/2025 13:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: BUSSEN

DATE/TIME OF TEST: 08/15/2025 21:07:23

Diagnostics Check		OK	21:01
Air Blank		0.000	21:01
Control Test		0.079	21:01
Air Blank		0.000	21:02
Subject Sample	# 1	0.193	21:02
Breath Volume		4.472	21:02
Air Blank		0.000	21:03
Air Blank		0.000	21:05
Subject Sample	# 2	0.188	21:05
Breath Volume		4.367	21:05
Air Blank		0.000	21:06
Control Test		0.077	21:06
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

Data Download Date/Time: 08/19/2025 13:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: SOSA RODRIGUEZ

DATE/TIME OF TEST: 08/16/2025 23:37:57

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.078	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 1		23:36	Subject Test Refused
Breath Volume		0.000	23:36	
Air Blank		0.000	23:37	
Control Test		0.079	23:37	
Air Blank		0.000	23:37	
Diagnostics Check		OK	23:37	

Data Download Date/Time: 08/19/2025 13:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/17/2025 00:31:32

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.079	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.135	00:26
Breath Volume		1.734	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.134	00:30
Breath Volume		2.207	00:30
Air Blank		0.000	00:30
Control Test		0.078	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 08/19/2025 13:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 08/17/2025 01:56:23

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.078	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.239	01:52
Breath Volume		2.738	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.242	01:54
Breath Volume		2.140	01:54
Air Blank		0.000	01:55
Control Test		0.078	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 08/19/2025 13:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: KINCHELOE

DATE/TIME OF TEST: 08/17/2025 03:52:13

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.078	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.106	03:47
Breath Volume		2.324	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.111	03:50
Breath Volume		2.457	03:50
Air Blank		0.000	03:51
Control Test		0.078	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 08/19/2025 13:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PATTERSON,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: LOPEZ GUEVARA

DATE/TIME OF TEST: 08/17/2025 04:30:33

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.078	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.077	04:26
Breath Volume		2.816	04:26
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.065	04:29
Breath Volume		1.128	04:29
Air Blank		0.000	04:29
Control Test		0.078	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 08/19/2025 13:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: SIMANGAN

DATE/TIME OF TEST: 08/17/2025 22:59:45

Diagnostics Check		OK	22:53
Air Blank		0.000	22:54
Control Test		0.078	22:54
Air Blank		0.000	22:54
Subject Sample	# 1	0.082	22:55
Breath Volume		1.996	22:55
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.082	22:58
Breath Volume		1.710	22:58
Air Blank		0.000	22:58
Control Test		0.077	22:59
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 08/19/2025 13:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: KNEALE

DATE/TIME OF TEST: 08/18/2025 03:57:07

Diagnostics Check		OK	03:54	
Air Blank		0.000	03:54	
Control Test		0.079	03:55	
Air Blank		0.000	03:55	
Subject Sample	# 1		03:55	Subject Test Refused
Breath Volume		0.000	03:55	
Air Blank		0.000	03:56	
Control Test		0.078	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 08/19/2025 13:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMSON,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: HACKEMEYER

DATE/TIME OF TEST: 08/18/2025 15:01:52

Diagnostics Check		OK	14:50
Air Blank		0.000	14:50
Control Test		0.079	14:50
Air Blank		0.000	14:51
Subject Sample	# 1	0.307	14:52
Breath Volume		1.199	14:52
Air Blank		0.000	14:53
Air Blank		0.000	14:54
Subject Sample	# 2	0.347	14:56
Breath Volume		2.429	14:56
Air Blank		0.000	14:57
Air Blank		0.000	14:59
Subject Sample	# 3	0.335	15:00
Breath Volume		2.343	15:00
Air Blank		0.000	15:00
Control Test		0.077	15:01
Air Blank		0.000	15:01
Diagnostics Check		OK	15:01

Data Download Date/Time: 08/19/2025 13:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 08/21/2025 10:59:26

Diagnostics Check		OK	10:56	
Air Blank		0.000	10:57	
Control Test		0.078	10:57	
Air Blank		0.000	10:57	
Subject Sample	# 1		10:58	Subject Test Refused
Breath Volume		0.000	10:58	
Air Blank		0.000	10:58	
Control Test		0.079	10:58	
Air Blank		0.000	10:59	
Diagnostics Check		OK	10:59	

Data Download Date/Time: 09/24/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 08/24/2025 00:07:50

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:05	
Control Test		0.079	00:05	
Air Blank		0.000	00:06	
Subject Sample	# 1		00:06	Subject Test Refused
Breath Volume		0.000	00:06	
Air Blank		0.000	00:06	
Control Test		0.079	00:07	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 09/24/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 08/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/27/2025 01:16:43

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:14	
Control Test		0.078	01:14	
Air Blank		0.000	01:15	
Subject Sample	# 1		01:15	Subject Test Refused
Breath Volume		0.000	01:15	
Air Blank		0.000	01:15	
Control Test		0.078	01:16	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:16	

Data Download Date/Time: 09/24/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,D,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 08/29/2025 05:22:26

Diagnostics Check		OK	05:14	
Air Blank		0.000	05:14	
Control Test		0.079	05:15	
Air Blank		0.000	05:15	
Subject Sample	# 1	0.138	05:18	
Breath Volume		2.222	05:18	
Air Blank		0.000	05:18	
Air Blank		0.000	05:20	
Subject Sample	# 2	0.146	05:21	
Breath Volume		2.320	05:21	
Air Blank			05:21	RFI Detect
Air Blank		0.000	05:22	

Data Download Date/Time: 09/24/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,D,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 08/29/2025 05:32:34

Diagnostics Check		OK	05:25
Air Blank		0.000	05:26
Control Test		0.078	05:26
Air Blank		0.000	05:27
Subject Sample	# 1	0.141	05:27
Breath Volume		2.062	05:27
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.137	05:31
Breath Volume		1.132	05:31
Air Blank		0.000	05:31
Control Test		0.078	05:31
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 09/24/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,R,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 08/31/2025 05:08:28

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.079	05:02
Air Blank		0.000	05:03
Subject Sample	# 1	0.206	05:03
Breath Volume		1.285	05:03
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.194	05:06
Breath Volume		2.042	05:06
Air Blank		0.000	05:07
Control Test		0.078	05:07
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 09/24/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,R,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 08/31/2025 23:53:50

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:53	
Control Test	0.000	23:53	RFI Detect
Air Blank	0.000	23:53	

Data Download Date/Time: 09/24/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DEVIN,K
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: ESCABAR,ANGEL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ENGLE

DATE/TIME OF TEST: 08/19/2025 01:46:32

Diagnostics Check		OK	01:39	
Air Blank		0.000	01:39	
Control Test		0.077	01:40	
Air Blank		0.000	01:40	
Subject Sample	# 1	0.000	01:42	
Breath Volume		2.609	01:42	
Air Blank		0.000	01:42	
Air Blank		0.000	01:44	
Subject Sample	# 2	0.000	01:45	
Breath Volume		2.519	01:45	
Air Blank		0.000	01:45	
Control Test		0.074	01:45	Control Outside Tolerance
Air Blank		0.000	01:46	

Data Download Date/Time: 09/14/2025 11:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: CALLAWAY,DANIEL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: EAVES

DATE/TIME OF TEST: 08/22/2025 20:58:35

Diagnostics Check		OK	20:52
Air Blank		0.000	20:52
Control Test		0.077	20:53
Air Blank		0.000	20:53
Subject Sample	# 1	0.000	20:54
Breath Volume		2.019	20:54
Air Blank		0.000	20:54
Air Blank		0.000	20:56
Subject Sample	# 2	0.000	20:57
Breath Volume		4.359	20:57
Air Blank		0.000	20:57
Control Test		0.077	20:57
Air Blank		0.000	20:58
Diagnostics Check		OK	20:58

Data Download Date/Time: 09/14/2025 11:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEELEY,SGT,V
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: MOZINGO,DEP
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BOX

DATE/TIME OF TEST: 08/27/2025 03:02:56

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.078	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.245	02:58
Breath Volume		1.113	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:00
Subject Sample	# 2	0.246	03:01
Breath Volume		1.117	03:01
Air Blank		0.000	03:01
Control Test		0.077	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 09/14/2025 11:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEELEY,SGT,V
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: MURRAY,DEP
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HEMPHILL

DATE/TIME OF TEST: 08/30/2025 01:17:11

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.079	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.216	01:12
Breath Volume		1.539	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.201	01:15
Breath Volume		2.093	01:15
Air Blank		0.000	01:16
Control Test		0.076	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 09/14/2025 11:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEELEY,SGT,V
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: MURRAY,DEP
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: LIVINGSTON

DATE/TIME OF TEST: 08/30/2025 21:48:09

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:39	
Control Test		0.078	21:39	
Air Blank		0.000	21:39	
Subject Sample	# 1		21:43	No Sample Provided
Breath Volume		0.000	21:43	
Air Blank		0.000	21:43	
Air Blank		0.000	21:45	
Subject Sample	# 2		21:46	Subject Test Refused
Breath Volume		0.000	21:46	
Air Blank		0.000	21:47	
Control Test		0.078	21:47	
Air Blank		0.000	21:48	
Diagnostics Check		OK	21:48	

Data Download Date/Time: 09/14/2025 11:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEELEY,CPL,V
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HADDOCK,S
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: RINER

DATE/TIME OF TEST: 08/02/2025 02:34:59

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:25	
Control Test		0.078	02:26	
Air Blank		0.000	02:26	
Subject Sample	# 1		02:29	No Sample Provided
Breath Volume		0.000	02:29	
Air Blank		0.000	02:30	
Air Blank		0.000	02:32	
Subject Sample	# 2		02:33	Subject Test Refused
Breath Volume		0.000	02:33	
Air Blank		0.000	02:34	
Control Test		0.079	02:34	
Air Blank		0.000	02:34	
Diagnostics Check		OK	02:34	

Data Download Date/Time: 08/25/2025 09:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DEVIN,K
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HADDOCK,SHANE
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SPINDLER

DATE/TIME OF TEST: 08/03/2025 23:23:35

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:21	
Control Test		0.078	23:21	
Air Blank		0.000	23:21	
Subject Sample	# 1		23:22	Subject Test Refused
Breath Volume		0.000	23:22	
Air Blank		0.000	23:22	
Control Test		0.079	23:22	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 08/25/2025 09:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEELEY,CPL,V
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,SGT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HEYMAN

DATE/TIME OF TEST: 08/08/2025 01:18:27

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.079	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.130	01:14
Breath Volume		1.820	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.128	01:16
Breath Volume		1.457	01:16
Air Blank		0.000	01:17
Control Test		0.077	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 08/25/2025 09:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HADDOCK',BRAIN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/08/2025 18:58:47

Diagnostics Check		OK	18:50	
Air Blank		0.000	18:50	
Control Test		0.079	18:50	
Air Blank		0.000	18:51	
Subject Sample	# 1		18:54	No Sample Provided
Breath Volume		0.000	18:54	
Air Blank		0.000	18:55	
Air Blank		0.000	18:57	
Subject Sample	# 2		18:57	Subject Test Refused
Breath Volume		0.000	18:57	
Air Blank		0.000	18:57	
Control Test		0.079	18:58	
Air Blank		0.000	18:58	
Diagnostics Check		OK	18:58	

Data Download Date/Time: 08/25/2025 09:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDINER,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: ESCOBAR,ANGEL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: NIELSEN OSMAR

DATE/TIME OF TEST: 08/15/2025 23:14:21

Diagnostics Check		OK	23:11	
Air Blank		0.000	23:12	
Control Test		0.079	23:12	
Air Blank		0.000	23:12	
Subject Sample	# 1		23:12	Subject Test Refused
Breath Volume		0.000	23:12	
Air Blank		0.000	23:13	
Control Test		0.079	23:13	
Air Blank		0.000	23:14	
Diagnostics Check		OK	23:14	

Data Download Date/Time: 08/25/2025 09:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAY,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MABIRE,TYLER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: SALTER

DATE/TIME OF TEST: 08/07/2025 05:09:21

Diagnostics Check		OK	05:01
Air Blank		0.000	05:02
Control Test		0.078	05:02
Air Blank		0.000	05:03
Subject Sample	# 1	0.000	05:04
Breath Volume		1.421	05:04
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.000	05:07
Breath Volume		1.152	05:07
Air Blank		0.000	05:08
Control Test		0.079	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 09/07/2025 21:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAY,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: NAY,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: FREEFOLD

DATE/TIME OF TEST: 08/07/2025 23:50:04

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.078	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.170	23:45
Breath Volume		3.042	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:47
Subject Sample	# 2	0.168	23:48
Breath Volume		2.902	23:48
Air Blank		0.000	23:49
Control Test		0.078	23:49
Air Blank		0.000	23:49
Diagnostics Check		OK	23:50

Data Download Date/Time: 09/07/2025 21:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORTENBERRY,ZACHARY,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FORTENBERRY,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 08/08/2025 23:13:44

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.078	23:07
Air Blank		0.000	23:07
Subject Sample	# 1	0.238	23:08
Breath Volume		3.109	23:08
Air Blank		0.000	23:09
Air Blank		0.000	23:11
Subject Sample	# 2	0.232	23:12
Breath Volume		1.683	23:12
Air Blank		0.000	23:12
Control Test		0.077	23:13
Air Blank		0.000	23:13
Diagnostics Check		OK	23:13

Data Download Date/Time: 09/07/2025 21:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ONEILL,BRAYDEN,P
BREATH TEST OPERATOR AGENCY: Milton PD
ARRESTING OFFICER: OLDFEILD,AUSTIN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: TURTLE

DATE/TIME OF TEST: 08/09/2025 22:14:02

Diagnostics Check		OK	22:07
Air Blank		0.000	22:07
Control Test		0.078	22:08
Air Blank		0.000	22:08
Subject Sample	# 1	0.202	22:09
Breath Volume		4.152	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:11
Subject Sample	# 2	0.191	22:12
Breath Volume		3.156	22:12
Air Blank		0.000	22:13
Control Test		0.077	22:13
Air Blank		0.000	22:13
Diagnostics Check		OK	22:14

Data Download Date/Time: 09/07/2025 21:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 08/10/2025 02:06:30

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:57	
Control Test		0.079	01:57	
Air Blank		0.000	01:58	
Subject Sample	# 1		02:01	No Sample Provided
Breath Volume		0.000	02:01	
Air Blank		0.000	02:01	
Air Blank		0.000	02:03	
Subject Sample	# 2		02:05	Subject Test Refused
Breath Volume		0.000	02:05	
Air Blank		0.000	02:05	
Control Test		0.078	02:05	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 09/07/2025 21:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALBRECHT,SYDNEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HENDRICKS

DATE/TIME OF TEST: 08/10/2025 04:00:06

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.080	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.125	03:55
Breath Volume		1.421	03:55
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.142	03:58
Breath Volume		1.429	03:58
Air Blank		0.000	03:59
Control Test		0.078	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 09/07/2025 21:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LEWIS,TIAUNNA
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 08:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:24

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/10/2025 10:56:13

Diagnostics Check		OK	10:49
Air Blank		0.000	10:50
Control Test		0.079	10:50
Air Blank		0.000	10:51
Subject Sample	# 1	0.018	10:51
Breath Volume		1.542	10:51
Air Blank		0.000	10:52
Air Blank		0.000	10:54
Subject Sample	# 2	0.019	10:54
Breath Volume		1.730	10:54
Air Blank		0.000	10:55
Control Test		0.078	10:55
Air Blank		0.000	10:56
Diagnostics Check		OK	10:56

Data Download Date/Time: 09/07/2025 21:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACKENZIE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACKENZIE,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PATINO ARROYO

DATE/TIME OF TEST: 08/13/2025 02:22:23

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:11	
Control Test		0.078	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:15	No Sample Provided
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Air Blank		0.000	02:17	
Subject Sample	# 2		02:21	No Sample Provided
Breath Volume		0.000	02:21	
Air Blank		0.000	02:21	
Control Test		0.079	02:21	
Air Blank		0.000	02:22	
Diagnostics Check		OK	02:22	

Data Download Date/Time: 09/07/2025 21:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUCE,HEATHER,N
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BRUCE,HEATHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 08/13/2025 18:31:02

Diagnostics Check		OK	18:24
Air Blank		0.000	18:25
Control Test		0.078	18:25
Air Blank		0.000	18:25
Subject Sample	# 1	0.244	18:26
Breath Volume		2.777	18:26
Air Blank		0.000	18:27
Air Blank		0.000	18:28
Subject Sample	# 2	0.251	18:29
Breath Volume		3.832	18:29
Air Blank		0.000	18:30
Control Test		0.078	18:30
Air Blank		0.000	18:30
Diagnostics Check		OK	18:31

Data Download Date/Time: 09/07/2025 21:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,STANLEY,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALBRECHT,SYDNEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/15/2025 04:03:48

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.159	03:58
Breath Volume		1.867	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.155	04:02
Breath Volume		1.187	04:02
Air Blank		0.000	04:02
Control Test		0.078	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 09/07/2025 21:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CASEY,SHELBY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 08/16/2025 01:46:10

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:43	
Control Test		0.079	01:44	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:45	
Control Test		0.079	01:45	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 09/07/2025 21:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAY,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: NAY,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SALTER

DATE/TIME OF TEST: 08/17/2025 02:41:21

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.078	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.115	02:36
Breath Volume		2.605	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.116	02:39
Breath Volume		3.207	02:39
Air Blank		0.000	02:40
Control Test		0.078	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 09/07/2025 21:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEASANTS,STEVEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PLEASANTS,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: GERMAIN

DATE/TIME OF TEST: 08/19/2025 01:09:23

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.077	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.136	01:04
Breath Volume		1.832	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.137	01:07
Breath Volume		1.750	01:07
Air Blank		0.000	01:08
Control Test		0.076	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 09/07/2025 21:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUISE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SCHIOWITZ

DATE/TIME OF TEST: 08/21/2025 22:36:48

Diagnostics Check		OK	22:29
Air Blank		0.000	22:30
Control Test		0.078	22:30
Air Blank		0.000	22:31
Subject Sample	# 1	0.132	22:31
Breath Volume		3.062	22:31
Air Blank		0.000	22:32
Air Blank		0.000	22:34
Subject Sample	# 2	0.123	22:35
Breath Volume		2.261	22:35
Air Blank		0.000	22:35
Control Test		0.077	22:36
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 09/07/2025 21:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MCGEE,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 08/21/2025 23:34:10

Diagnostics Check		OK	23:24	
Air Blank		0.000	23:25	
Control Test		0.078	23:25	
Air Blank		0.000	23:25	
Subject Sample	# 1	0.228	23:28	Volume Not Met
Breath Volume		0.625	23:28	
Air Blank		0.000	23:29	
Air Blank		0.000	23:31	
Subject Sample	# 2		23:32	Subject Test Refused
Breath Volume		0.050	23:32	
Air Blank		0.000	23:33	
Control Test		0.076	23:33	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 09/07/2025 21:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORTENBERRY,ZACHARY,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FORTENBERRY,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: SURIN

DATE/TIME OF TEST: 08/23/2025 01:55:56

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.078	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.228	01:51
Breath Volume		1.812	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:53
Subject Sample	# 2	0.223	01:54
Breath Volume		2.972	01:54
Air Blank		0.000	01:55
Control Test		0.078	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 09/07/2025 21:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,NOAH,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JABLONSKI,MARK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 08/24/2025 22:30:45

Diagnostics Check		OK	22:24
Air Blank		0.000	22:24
Control Test		0.079	22:24
Air Blank		0.000	22:25
Subject Sample	# 1	0.113	22:25
Breath Volume		3.093	22:25
Air Blank		0.000	22:26
Air Blank		0.000	22:28
Subject Sample	# 2	0.115	22:29
Breath Volume		2.929	22:29
Air Blank		0.000	22:29
Control Test		0.078	22:30
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 09/07/2025 21:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLTON,JUSTIN,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOLTON,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/25/2025 22:11:13

Diagnostics Check		OK	22:05	
Air Blank		0.000	22:06	
Control Test		0.078	22:06	
Air Blank		0.000	22:07	
Subject Sample	# 1	0.327	22:07	
Breath Volume		1.144	22:07	
Air Blank		0.000	22:08	
Air Blank			22:09	RFI Detect
Air Blank		0.000	22:10	
Control Test		0.078	22:10	
Air Blank		0.000	22:11	
Diagnostics Check		OK	22:11	

Data Download Date/Time: 09/07/2025 21:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: CURREY

DATE/TIME OF TEST: 08/27/2025 03:40:10

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.078	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.221	03:35
Breath Volume		1.613	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.219	03:38
Breath Volume		2.140	03:38
Air Blank		0.000	03:39
Control Test		0.077	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 09/07/2025 21:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,STANLEY,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WATERS,CALEB
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BRICKMAN

DATE/TIME OF TEST: 08/29/2025 04:56:23

Diagnostics Check		OK	04:53	
Air Blank		0.000	04:53	
Control Test		0.079	04:54	
Air Blank		0.000	04:54	
Subject Sample	# 1		04:55	Subject Test Refused
Breath Volume		0.000	04:55	
Air Blank		0.000	04:55	
Control Test		0.079	04:55	
Air Blank		0.000	04:56	
Diagnostics Check		OK	04:56	

Data Download Date/Time: 09/07/2025 21:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TEICHNER,ADAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 08/30/2025 00:52:29

Diagnostics Check		OK	00:43	
Air Blank		0.000	00:43	
Control Test		0.078	00:44	
Air Blank		0.000	00:44	
Subject Sample	# 1		00:47	No Sample Provided
Breath Volume		0.000	00:47	
Air Blank		0.000	00:48	
Air Blank		0.000	00:50	
Subject Sample	# 2		00:51	Subject Test Refused
Breath Volume		0.820	00:51	
Air Blank			00:52	RFI Detect
Air Blank		0.000	00:52	

Data Download Date/Time: 09/07/2025 21:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 08/30/2025 01:02:01

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.078	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.000	00:57
Breath Volume		2.996	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.000	01:00
Breath Volume		3.000	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:02

Data Download Date/Time: 09/07/2025 21:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAY,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: NAY,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 08/30/2025 02:07:28

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.078	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.183	02:02
Breath Volume		2.492	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.182	02:05
Breath Volume		2.429	02:05
Air Blank		0.000	02:06
Control Test		0.078	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 09/07/2025 21:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CANETE

DATE/TIME OF TEST: 08/30/2025 03:11:55

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:00	
Control Test		0.079	03:01	
Air Blank		0.000	03:01	
Subject Sample	# 1		03:04	No Sample Provided
Breath Volume		0.000	03:04	
Air Blank		0.000	03:05	
Air Blank		0.000	03:07	
Subject Sample	# 2		03:10	No Sample Provided
Breath Volume		0.000	03:10	
Air Blank		0.000	03:11	
Control Test		0.079	03:11	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 09/07/2025 21:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,BRENNON,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEE,BRENNON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: PRICHARD

DATE/TIME OF TEST: 08/31/2025 01:12:56

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.078	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.141	01:08
Breath Volume		3.214	01:08
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.130	01:11
Breath Volume		2.089	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 09/07/2025 21:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,BRIAN,T
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: OLDFIELD,AUSTIN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: NIXON

DATE/TIME OF TEST: 08/31/2025 20:03:30

Diagnostics Check		OK	19:54	
Air Blank		0.000	19:55	
Control Test		0.078	19:55	
Air Blank		0.000	19:56	
Subject Sample	# 1		19:59	No Sample Provided
Breath Volume		0.000	19:59	
Air Blank		0.000	19:59	
Air Blank		0.000	20:01	
Subject Sample	# 2		20:02	Subject Test Refused
Breath Volume		0.000	20:02	
Air Blank		0.000	20:02	
Control Test		0.078	20:02	
Air Blank		0.000	20:03	
Diagnostics Check		OK	20:03	

Data Download Date/Time: 09/07/2025 21:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: KITCHENS

DATE/TIME OF TEST: 08/31/2025 20:38:12

Diagnostics Check		OK	20:32	
Air Blank		0.000	20:33	
Control Test		0.079	20:33	
Air Blank		0.000	20:33	
Subject Sample	# 1		20:36	Subject Test Refused
Breath Volume		0.000	20:36	
Air Blank		0.000	20:37	
Control Test		0.079	20:37	
Air Blank		0.000	20:38	
Diagnostics Check		OK	20:38	

Data Download Date/Time: 09/07/2025 21:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 08/02/2025 04:03:43

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.082	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.098	03:59
Breath Volume		2.593	03:59
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.095	04:02
Breath Volume		2.187	04:02
Air Blank		0.000	04:02
Control Test		0.081	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 08/07/2025 04:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAY,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: NAY,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/06/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: STAGGS-SCHLOTT

DATE/TIME OF TEST: 08/02/2025 04:45:02

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.081	04:39
Air Blank		0.000	04:39
Subject Sample	# 1	0.135	04:40
Breath Volume		2.644	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.125	04:43
Breath Volume		1.992	04:43
Air Blank		0.000	04:44
Control Test		0.081	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:45

Data Download Date/Time: 08/07/2025 04:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 08/03/2025 03:39:24

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.082	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.060	03:34
Breath Volume		1.683	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.058	03:37
Breath Volume		2.191	03:37
Air Blank		0.000	03:38
Control Test		0.081	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 08/07/2025 04:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 08/03/2025 22:16:41

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.082	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.175	22:12
Breath Volume		1.882	22:12
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.166	22:15
Breath Volume		1.746	22:15
Air Blank		0.000	22:15
Control Test		0.080	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 08/07/2025 04:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSLEY,KYLE,
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOSLEY,KYLE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/04/2025 03:49:47

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:47	
Control Test		0.080	03:47	
Air Blank		0.000	03:48	
Subject Sample	# 1		03:48	Subject Test Refused
Breath Volume		0.000	03:48	
Air Blank		0.000	03:48	
Control Test		0.080	03:49	
Air Blank		0.000	03:49	
Diagnostics Check		OK	03:49	

Data Download Date/Time: 08/07/2025 04:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: KOVALCIK

DATE/TIME OF TEST: 08/05/2025 21:14:53

Diagnostics Check		OK	21:08
Air Blank		0.000	21:08
Control Test		0.081	21:08
Air Blank		0.000	21:09
Subject Sample	# 1	0.205	21:10
Breath Volume		2.234	21:10
Air Blank		0.000	21:11
Air Blank		0.000	21:12
Subject Sample	# 2	0.210	21:13
Breath Volume		1.976	21:13
Air Blank		0.000	21:13
Control Test		0.081	21:14
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 08/07/2025 04:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCARBOROUGH,TYLER,J
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: CULBRETH,J
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 09:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:36

SUBJECT'S LAST NAME: KNOX

DATE/TIME OF TEST: 08/08/2025 10:46:04

Diagnostics Check		OK	10:30	
Air Blank		0.000	10:30	
Control Test		0.080	10:31	
Air Blank		0.000	10:31	
Subject Sample	# 1	0.180	10:33	
Breath Volume		1.687	10:33	
Air Blank		0.000	10:34	
Air Blank		0.000	10:35	
Subject Sample	# 2		10:39	No Sample Provided
Breath Volume		0.000	10:39	
Air Blank		0.000	10:39	
Air Blank		0.000	10:41	
Subject Sample	# 3		10:44	No Sample Provided
Breath Volume		0.000	10:44	
Air Blank		0.000	10:45	
Control Test		0.080	10:45	
Air Blank		0.000	10:45	
Diagnostics Check		OK	10:46	

Data Download Date/Time: 09/30/2025 14:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/11/2025 20:45:56

Diagnostics Check		OK	20:39
Air Blank		0.000	20:39
Control Test		0.081	20:40
Air Blank		0.000	20:40
Subject Sample	# 1	0.300	20:41
Breath Volume		2.519	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:43
Subject Sample	# 2	0.280	20:44
Breath Volume		2.109	20:44
Air Blank		0.000	20:45
Control Test		0.081	20:45
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 09/30/2025 14:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSICK,ROBERT,C
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: MERRITT,JACOB
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: ROEDEL

DATE/TIME OF TEST: 08/22/2025 21:31:26

Diagnostics Check		OK	21:25
Air Blank		0.000	21:25
Control Test		0.081	21:25
Air Blank		0.000	21:26
Subject Sample	# 1	0.000	21:26
Breath Volume		1.945	21:26
Air Blank		0.000	21:27
Air Blank		0.000	21:29
Subject Sample	# 2	0.000	21:30
Breath Volume		1.636	21:30
Air Blank		0.000	21:30
Control Test		0.081	21:30
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 09/30/2025 14:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBIN,TRUETT,L
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: CORBIN,TRUETT
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 08/24/2025 01:06:38

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.081	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.182	01:01
Breath Volume		1.972	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.188	01:05
Breath Volume		1.710	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 09/30/2025 14:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 08/26/2025 21:57:56

Diagnostics Check		OK	21:51
Air Blank		0.000	21:51
Control Test		0.081	21:52
Air Blank		0.000	21:52
Subject Sample	# 1	0.217	21:53
Breath Volume		2.183	21:53
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.225	21:56
Breath Volume		2.734	21:56
Air Blank		0.000	21:57
Control Test		0.080	21:57
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 09/30/2025 14:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PARK,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 08/27/2025 04:46:25

Diagnostics Check		OK	04:39
Air Blank		0.000	04:40
Control Test		0.081	04:40
Air Blank		0.000	04:40
Subject Sample	# 1	0.075	04:41
Breath Volume		1.867	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.080	04:44
Breath Volume		2.054	04:44
Air Blank		0.000	04:45
Control Test		0.081	04:45
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 09/30/2025 14:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRISP,JESSICA,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CRISP,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: CLUNAN

DATE/TIME OF TEST: 08/30/2025 02:29:18

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.081	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.222	02:24
Breath Volume		1.886	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.216	02:27
Breath Volume		1.156	02:27
Air Blank		0.000	02:28
Control Test		0.080	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 09/30/2025 14:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 07/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 08/02/2025 02:25:22

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.082	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.065	02:20
Breath Volume		3.195	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.064	02:23
Breath Volume		3.648	02:23
Air Blank		0.000	02:24
Control Test		0.080	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 08/04/2025 13:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELTHROP,MATTHEW,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BELTHROP,MATTHEW
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 08/02/2025 05:57:36

Diagnostics Check		OK	05:48	
Air Blank		0.000	05:49	
Control Test		0.079	05:49	
Air Blank		0.000	05:50	
Subject Sample	# 1		05:53	No Sample Provided
Breath Volume		0.000	05:53	
Air Blank		0.000	05:53	
Air Blank		0.000	05:55	
Subject Sample	# 2		05:56	Subject Test Refused
Breath Volume		0.000	05:56	
Air Blank		0.000	05:56	
Control Test		0.079	05:57	
Air Blank		0.000	05:57	
Diagnostics Check		OK	05:57	

Data Download Date/Time: 08/04/2025 13:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELTHROP,MATTHEW,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BELTHROP,MATTHEW
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 08/03/2025 02:55:11

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.081	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.141	02:50
Breath Volume		1.488	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.152	02:53
Breath Volume		1.414	02:53
Air Blank		0.000	02:54
Control Test		0.081	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 08/04/2025 13:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHERMAN,D,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HERRELL,J
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/11/2025 20:32:42

Diagnostics Check		OK	20:23
Air Blank		0.000	20:23
Control Test		0.081	20:24
Air Blank		0.000	20:24
Subject Sample	# 1	0.175	20:25
Breath Volume		1.816	20:25
Air Blank		0.000	20:25
Air Blank		0.000	20:27
Subject Sample	# 2	0.152	20:28
Breath Volume		2.093	20:28
Air Blank		0.000	20:28
Air Blank		0.000	20:30
Subject Sample	# 3	0.164	20:31
Breath Volume		2.386	20:31
Air Blank		0.000	20:31
Control Test		0.080	20:32
Air Blank		0.000	20:32
Diagnostics Check		OK	20:32

Data Download Date/Time: 09/10/2025 10:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMUEL,P
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: PALMER,SAMUEL
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 14:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:26

SUBJECT'S LAST NAME: GALLARDO

DATE/TIME OF TEST: 08/17/2025 15:03:40

Diagnostics Check		OK	14:57
Air Blank		0.000	14:57
Control Test		0.081	14:58
Air Blank		0.000	14:58
Subject Sample	# 1	0.194	14:59
Breath Volume		1.929	14:59
Air Blank		0.000	14:59
Air Blank		0.000	15:01
Subject Sample	# 2	0.195	15:02
Breath Volume		1.945	15:02
Air Blank		0.000	15:02
Control Test		0.080	15:03
Air Blank		0.000	15:03
Diagnostics Check		OK	15:03

Data Download Date/Time: 09/10/2025 10:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,F,H
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALLEN,F
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: BERNIER

DATE/TIME OF TEST: 08/19/2025 21:13:25

Diagnostics Check		OK	21:10	
Air Blank		0.000	21:11	
Control Test		0.081	21:11	
Air Blank		0.000	21:11	
Subject Sample	# 1		21:12	Subject Test Refused
Breath Volume		0.000	21:12	
Air Blank		0.000	21:12	
Control Test		0.082	21:12	
Air Blank		0.000	21:13	
Diagnostics Check		OK	21:13	

Data Download Date/Time: 09/10/2025 10:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISON,ALLISON,K
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HENDERSON,JOSHUA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: FINCHUM

DATE/TIME OF TEST: 08/23/2025 21:33:05

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:24	
Control Test		0.082	21:24	
Air Blank		0.000	21:25	
Subject Sample	# 1	0.206	21:28	Volume Not Met
Breath Volume		0.812	21:28	
Air Blank		0.000	21:28	
Air Blank		0.000	21:30	
Subject Sample	# 2		21:31	Subject Test Refused
Breath Volume		0.257	21:31	
Air Blank		0.000	21:32	
Control Test		0.079	21:32	
Air Blank		0.000	21:33	
Diagnostics Check		OK	21:33	

Data Download Date/Time: 09/10/2025 10:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CADY,ELISE,N
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: TETZLAFF,JOSHUA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 11:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 08/28/2025 00:50:22

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:39	
Control Test		0.080	00:39	
Air Blank		0.000	00:40	
Subject Sample	# 1	0.276	00:43	Volume Not Met
Breath Volume		0.070	00:43	
Air Blank		0.000	00:43	
Air Blank		0.000	00:45	
Subject Sample	# 2	0.237	00:48	Volume Not Met
Breath Volume		0.066	00:48	
Air Blank		0.000	00:49	
Control Test		0.079	00:49	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 09/10/2025 10:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CADY,ELISE,N
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: TETZLAFF,JOSHUA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 11:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 08/28/2025 01:12:47

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:01	
Control Test		0.079	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1	0.264	01:05	Volume Not Met
Breath Volume		0.031	01:05	
Air Blank		0.000	01:06	
Air Blank		0.000	01:08	
Subject Sample	# 2	0.244	01:11	Volume Not Met
Breath Volume		0.070	01:11	
Air Blank		0.000	01:11	
Control Test		0.079	01:12	
Air Blank		0.000	01:12	
Diagnostics Check		OK	01:12	

Data Download Date/Time: 09/10/2025 10:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BUNCH,BRADFORD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 08/29/2025 23:33:39

Diagnostics Check		OK	23:27
Air Blank		0.000	23:27
Control Test		0.081	23:28
Air Blank		0.000	23:28
Subject Sample	# 1	0.186	23:29
Breath Volume		1.957	23:29
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.193	23:32
Breath Volume		1.648	23:32
Air Blank		0.000	23:32
Control Test		0.080	23:33
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 09/10/2025 10:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALNAKER,DILLON,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BALES,SOMMER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BONNER

DATE/TIME OF TEST: 08/30/2025 01:05:00

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.081	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.000	01:00
Breath Volume		4.394	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.000	01:03
Breath Volume		4.292	01:03
Air Blank		0.000	01:04
Control Test		0.081	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:05

Data Download Date/Time: 09/10/2025 10:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: OUTLAW

DATE/TIME OF TEST: 08/30/2025 23:23:39

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:11	
Control Test		0.080	23:11	
Air Blank		0.000	23:11	
Subject Sample	# 1	0.168	23:15	Volume Not Met
Breath Volume		1.125	23:15	
Air Blank		0.000	23:15	
Air Blank		0.000	23:17	
Subject Sample	# 2	0.156	23:18	
Breath Volume		1.410	23:18	
Air Blank		0.000	23:19	
Air Blank		0.000	23:21	
Subject Sample	# 3	0.162	23:22	
Breath Volume		1.320	23:22	
Air Blank		0.000	23:22	
Control Test		0.079	23:23	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 09/10/2025 10:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: EADS

DATE/TIME OF TEST: 08/30/2025 23:35:20

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.079	23:29
Air Blank		0.000	23:29
Subject Sample	# 1	0.201	23:30
Breath Volume		3.859	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.208	23:33
Breath Volume		4.187	23:33
Air Blank		0.000	23:34
Control Test		0.080	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 09/10/2025 10:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: SCHULLER

DATE/TIME OF TEST: 08/31/2025 02:41:51

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.080	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.213	02:36
Breath Volume		2.070	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.208	02:40
Breath Volume		3.250	02:40
Air Blank		0.000	02:40
Control Test		0.079	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 09/10/2025 10:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: NELSON,JADEN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: THORPE

DATE/TIME OF TEST: 08/31/2025 05:24:25

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.081	05:18
Air Blank		0.000	05:19
Subject Sample	# 1	0.160	05:20
Breath Volume		1.222	05:20
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.164	05:22
Breath Volume		3.558	05:22
Air Blank		0.000	05:23
Control Test		0.080	05:23
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 09/10/2025 10:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAK,STEVEN,M
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: RAHN,BRENNAN
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: HERNANDEZ EVANS

DATE/TIME OF TEST: 08/31/2025 06:48:07

Diagnostics Check		OK	06:45	
Air Blank		0.000	06:45	
Control Test		0.081	06:46	
Air Blank		0.000	06:46	
Subject Sample	# 1		06:46	Subject Test Refused
Breath Volume		0.000	06:46	
Air Blank		0.000	06:47	
Control Test		0.080	06:47	
Air Blank		0.000	06:48	
Diagnostics Check		OK	06:48	

Data Download Date/Time: 09/10/2025 10:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMOND,JAXSON,W
BREATH TEST OPERATOR AGENCY: Springfield PD
ARRESTING OFFICER: HAMMOND,JAXSON
ARRESTING OFFICER AGENCY: Springfield PD
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: BANNISTER

DATE/TIME OF TEST: 08/31/2025 22:23:41

Diagnostics Check		OK	22:16
Air Blank		0.000	22:17
Control Test		0.082	22:17
Air Blank		0.000	22:18
Subject Sample	# 1	0.100	22:19
Breath Volume		1.558	22:19
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.098	22:22
Breath Volume		1.875	22:22
Air Blank		0.000	22:22
Control Test		0.081	22:23
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 09/10/2025 10:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: ERWIN

DATE/TIME OF TEST: 08/01/2025 02:22:25

Diagnostics Check	OK	02:22	
Air Blank		02:22	RFI Detect
Air Blank		02:22	RFI Detect

Data Download Date/Time: 08/26/2025 17:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: ERWIN

DATE/TIME OF TEST: 08/01/2025 02:36:00

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:24	
Control Test		0.078	02:24	
Air Blank		0.000	02:25	
Subject Sample	# 1		02:28	No Sample Provided
Breath Volume		0.000	02:28	
Air Blank		0.000	02:28	
Air Blank		0.000	02:30	
Subject Sample	# 2	0.172	02:31	
Breath Volume		2.593	02:31	
Air Blank		0.000	02:32	
Air Blank		0.000	02:33	
Subject Sample	# 3	0.178	02:34	
Breath Volume		3.082	02:34	
Air Blank		0.000	02:35	
Control Test		0.079	02:35	
Air Blank		0.000	02:35	
Diagnostics Check		OK	02:36	

Data Download Date/Time: 08/26/2025 17:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: STALNAKER

DATE/TIME OF TEST: 08/06/2025 01:11:27

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.078	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:04	No Sample Provided
Breath Volume		0.000	01:04	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 2		01:10	Subject Test Refused
Breath Volume		0.000	01:10	
Air Blank		0.000	01:10	
Control Test		0.078	01:10	
Air Blank		0.000	01:11	
Diagnostics Check		OK	01:11	

Data Download Date/Time: 08/26/2025 17:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SCHULZ

DATE/TIME OF TEST: 08/09/2025 01:30:19

Diagnostics Check		OK	01:23	
Air Blank		0.000	01:23	
Control Test		0.079	01:24	
Air Blank		0.000	01:24	
Subject Sample	# 1	0.078	01:25	
Breath Volume		2.089	01:25	
Air Blank		0.000	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 2	0.072	01:28	
Breath Volume		1.378	01:28	
Air Blank		0.000	01:29	
Control Test		0.078	01:29	
Air Blank			01:29	RFI Detect
Air Blank		0.000	01:30	

Data Download Date/Time: 08/26/2025 17:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRACY,NICOLE,L
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SCHULZ

DATE/TIME OF TEST: 08/09/2025 01:47:19

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.079	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.076	01:42
Breath Volume		1.578	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.074	01:45
Breath Volume		1.664	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 08/26/2025 17:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: LOLLIE

DATE/TIME OF TEST: 08/09/2025 06:22:42

Diagnostics Check		OK	06:11	
Air Blank		0.000	06:11	
Control Test		0.079	06:12	
Air Blank		0.000	06:12	
Subject Sample	# 1		06:15	No Sample Provided
Breath Volume		0.000	06:15	
Air Blank		0.000	06:16	
Air Blank		0.000	06:18	
Subject Sample	# 2		06:21	No Sample Provided
Breath Volume		0.000	06:21	
Air Blank		0.000	06:21	
Control Test		0.078	06:22	
Air Blank		0.000	06:22	
Diagnostics Check		OK	06:22	

Data Download Date/Time: 08/26/2025 17:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: DAMON

DATE/TIME OF TEST: 08/10/2025 04:08:13

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.079	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.014	04:03
Breath Volume		1.378	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.013	04:06
Breath Volume		1.277	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 08/26/2025 17:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: PACK

DATE/TIME OF TEST: 08/10/2025 23:57:17

Diagnostics Check	OK	23:56	
Air Blank		23:56	RFI Detect
Air Blank	0.000	23:57	

Data Download Date/Time: 08/26/2025 17:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: PACK

DATE/TIME OF TEST: 08/11/2025 00:11:53

Diagnostics Check		OK	23:59	
Air Blank		0.000	00:00	
Control Test		0.079	00:00	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:04	No Sample Provided
Breath Volume		0.000	00:04	
Air Blank		0.000	00:04	
Air Blank		0.000	00:06	
Subject Sample	# 2	0.142	00:08	
Breath Volume		2.328	00:08	
Air Blank		0.000	00:08	
Air Blank			00:10	RFI Detect
Air Blank		0.000	00:10	
Control Test		0.078	00:11	
Air Blank		0.000	00:11	
Diagnostics Check		OK	00:11	

Data Download Date/Time: 08/26/2025 17:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: PACK

DATE/TIME OF TEST: 08/11/2025 00:41:33

Diagnostics Check		OK	00:33
Air Blank		0.000	00:33
Control Test		0.078	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.131	00:36
Breath Volume		1.644	00:36
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.128	00:40
Breath Volume		1.242	00:40
Air Blank		0.000	00:40
Control Test		0.078	00:40
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 08/26/2025 17:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 08/15/2025 02:10:14

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:59	
Control Test		0.079	01:59	
Air Blank		0.000	02:00	
Subject Sample	# 1		02:03	No Sample Provided
Breath Volume		0.000	02:03	
Air Blank		0.000	02:03	
Air Blank		0.000	02:05	
Subject Sample	# 2		02:08	No Sample Provided
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Control Test		0.079	02:09	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 08/26/2025 17:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: DAHL

DATE/TIME OF TEST: 08/23/2025 23:17:11

Diagnostics Check		OK	23:08
Air Blank		0.000	23:09
Control Test		0.078	23:09
Air Blank		0.000	23:10
Subject Sample	# 1	0.072	23:11
Breath Volume		2.281	23:11
Air Blank		0.000	23:11
Air Blank		0.000	23:13
Subject Sample	# 2	0.070	23:15
Breath Volume		2.046	23:15
Air Blank		0.000	23:16
Control Test		0.078	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 08/26/2025 17:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 08/29/2025 04:20:45

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:12	
Control Test		0.078	04:13	
Air Blank		0.000	04:13	
Subject Sample	# 1	0.185	04:15	
Breath Volume		3.769	04:15	
Air Blank		0.000	04:16	
Air Blank		0.000	04:18	
Subject Sample	# 2	0.180	04:19	
Breath Volume		2.781	04:19	
Air Blank		0.000	04:19	
Control Test		0.099	04:19	Interferent Detect
Air Blank			04:20	Purge Fail

Data Download Date/Time: 10/09/2025 15:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 08/29/2025 04:26:28

Diagnostics Check		OK	04:22	
Air Blank		0.000	04:23	
Control Test		0.079	04:23	
Air Blank		0.000	04:24	
Subject Sample	# 1		04:25	Interferent Detect
Breath Volume		2.250	04:25	
Air Blank		0.000	04:25	
Control Test		0.058	04:25	Interferent Detect
Air Blank		0.000	04:26	

Data Download Date/Time: 10/09/2025 15:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 08/29/2025 04:36:06

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.078	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.184	04:31
Breath Volume		2.875	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:33
Subject Sample	# 2	0.180	04:34
Breath Volume		1.996	04:34
Air Blank		0.000	04:35
Control Test		0.076	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 10/09/2025 15:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIF,HAMPTON,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 08/28/2025 00:57:37

Diagnostics Check		OK	00:48
Air Blank		0.000	00:49
Control Test		0.080	00:49
Air Blank		0.000	00:50
Subject Sample	# 1	0.189	00:53
Breath Volume		1.941	00:53
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.177	00:56
Breath Volume		1.753	00:56
Air Blank		0.000	00:56
Control Test		0.079	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 09/22/2025 11:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BRYAN,SAMUEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LIPSCOMB

DATE/TIME OF TEST: 08/28/2025 01:21:17

Diagnostics Check	OK	01:20	
Air Blank		01:20	RFI Detect
Air Blank	0.000	01:21	

Data Download Date/Time: 09/22/2025 11:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BRYAN,SAMUEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LIPSCOMB

DATE/TIME OF TEST: 08/28/2025 01:26:24

Diagnostics Check		OK	01:23	
Air Blank		0.000	01:24	
Control Test		0.081	01:24	
Air Blank		0.000	01:24	
Subject Sample	# 1		01:25	Subject Test Refused
Breath Volume		0.000	01:25	
Air Blank		0.000	01:25	
Control Test		0.080	01:25	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 09/22/2025 11:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,TRISTAN,T
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ROGERS,TRISTAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MERRILL

DATE/TIME OF TEST: 08/06/2025 04:09:18

Diagnostics Check		OK	04:05	
Air Blank		0.000	04:05	
Control Test		0.078	04:06	
Air Blank		0.000	04:06	
Subject Sample	# 1	0.241	04:07	
Breath Volume		1.648	04:07	
Air Blank			04:07	RFI Detect
Air Blank		0.000	04:08	
Control Test		0.080	04:08	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 08/26/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,TRISTAN,T
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ROGERS,TRISTAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MERRILL

DATE/TIME OF TEST: 08/06/2025 04:14:43

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:11	
Control Test		0.079	04:11	
Air Blank		0.000	04:12	
Subject Sample	# 1		04:13	Slope Not Met
Breath Volume		0.851	04:13	
Air Blank		0.000	04:13	
Control Test		0.079	04:14	
Air Blank		0.000	04:14	
Diagnostics Check		OK	04:14	

Data Download Date/Time: 08/26/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,TRISTAN,T
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ROGERS,TRISTAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MERRILL

DATE/TIME OF TEST: 08/06/2025 04:23:44

Diagnostics Check		OK	04:16	
Air Blank		0.000	04:17	
Control Test		0.079	04:17	
Air Blank		0.000	04:17	
Subject Sample	# 1	0.227	04:18	
Breath Volume		1.628	04:18	
Air Blank		0.000	04:19	
Air Blank		0.000	04:21	
Subject Sample	# 2		04:22	Subject Test Refused
Breath Volume		0.015	04:22	
Air Blank		0.000	04:22	
Control Test		0.078	04:23	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 08/26/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVER,SALVADOR,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: OLIVER,SALVADOR
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 08:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: STANGE

DATE/TIME OF TEST: 08/11/2025 08:41:18

Diagnostics Check		OK	08:35
Air Blank		0.000	08:35
Control Test		0.079	08:36
Air Blank		0.000	08:36
Subject Sample	# 1	0.000	08:37
Breath Volume		2.269	08:37
Air Blank		0.000	08:37
Air Blank		0.000	08:39
Subject Sample	# 2	0.000	08:39
Breath Volume		1.929	08:39
Air Blank		0.000	08:40
Control Test		0.079	08:40
Air Blank		0.000	08:41
Diagnostics Check		OK	08:41

Data Download Date/Time: 08/26/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIF,HAMPTON,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: BARRETO JIMENEZ

DATE/TIME OF TEST: 08/13/2025 21:40:10

Diagnostics Check		OK	21:30	
Air Blank		0.000	21:31	
Control Test		0.079	21:31	
Air Blank		0.000	21:31	
Subject Sample	# 1	0.153	21:35	Volume Not Met
Breath Volume		0.296	21:35	
Air Blank		0.000	21:35	
Air Blank		0.000	21:37	
Subject Sample	# 2		21:38	Subject Test Refused
Breath Volume		0.000	21:38	
Air Blank		0.000	21:39	
Control Test		0.078	21:39	
Air Blank		0.000	21:40	
Diagnostics Check		OK	21:40	

Data Download Date/Time: 08/26/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESCHAMPS,KASEN,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: JOHNSON,LAUREN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: HASTINGS

DATE/TIME OF TEST: 08/02/2025 23:43:32

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:38	
Control Test		0.080	23:38	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:41	Subject Test Refused
Breath Volume		0.000	23:41	
Air Blank		0.000	23:42	
Control Test		0.081	23:42	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 08/20/2025 08:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: FAIRTRACE,JENNIFER
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 13:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:38

SUBJECT'S LAST NAME: ASKUE

DATE/TIME OF TEST: 08/06/2025 14:05:29

Diagnostics Check		OK	13:58
Air Blank		0.000	13:59
Control Test		0.081	13:59
Air Blank		0.000	14:00
Subject Sample	# 1	0.115	14:00
Breath Volume		2.097	14:00
Air Blank		0.000	14:01
Air Blank		0.000	14:03
Subject Sample	# 2	0.099	14:03
Breath Volume		1.468	14:03
Air Blank		0.000	14:04
Control Test		0.080	14:04
Air Blank		0.000	14:05
Diagnostics Check		OK	14:05

Data Download Date/Time: 08/20/2025 08:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:22

SUBJECT'S LAST NAME: MERAS

DATE/TIME OF TEST: 08/07/2025 12:48:46

Diagnostics Check		OK	12:42
Air Blank		0.000	12:43
Control Test		0.080	12:43
Air Blank		0.000	12:43
Subject Sample	# 1	0.000	12:44
Breath Volume		3.250	12:44
Air Blank		0.000	12:44
Air Blank		0.000	12:46
Subject Sample	# 2	0.000	12:47
Breath Volume		2.937	12:47
Air Blank		0.000	12:47
Control Test		0.081	12:48
Air Blank		0.000	12:48
Diagnostics Check		OK	12:48

Data Download Date/Time: 08/20/2025 08:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,MARK,K
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ZIEGLER,ADAM
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 08:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:49

SUBJECT'S LAST NAME: ZORRILLA-RODRIGUEZ

DATE/TIME OF TEST: 08/10/2025 09:16:33

Diagnostics Check		OK	09:13	
Air Blank		0.000	09:14	
Control Test		0.080	09:14	
Air Blank		0.000	09:15	
Subject Sample	# 1		09:15	Subject Test Refused
Breath Volume		0.000	09:15	
Air Blank		0.000	09:15	
Control Test		0.080	09:15	
Air Blank		0.000	09:16	
Diagnostics Check		OK	09:16	

Data Download Date/Time: 08/20/2025 08:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEABOLT,BRADY,Z
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: WRIGHT,MARK
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 08/17/2025 21:47:55

Diagnostics Check		OK	21:45	
Air Blank		0.000	21:45	
Control Test		0.081	21:45	
Air Blank		0.000	21:46	
Subject Sample	# 1		21:46	Subject Test Refused
Breath Volume		0.000	21:46	
Air Blank		0.000	21:47	
Control Test		0.078	21:47	
Air Blank		0.000	21:47	
Diagnostics Check		OK	21:47	

Data Download Date/Time: 08/20/2025 08:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VARNDORE,RICHARD
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 08/20/2025 20:40:01

Diagnostics Check		OK	20:33
Air Blank		0.000	20:33
Control Test		0.081	20:34
Air Blank		0.000	20:34
Subject Sample	# 1	0.000	20:35
Breath Volume		4.246	20:35
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.000	20:38
Breath Volume		3.964	20:38
Air Blank		0.000	20:39
Control Test		0.080	20:39
Air Blank		0.000	20:39
Diagnostics Check		OK	20:40

Data Download Date/Time: 09/23/2025 12:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTTARI,LOGAN,H
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: JEAN,ISABELL
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 09:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:38

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/24/2025 11:07:14

Diagnostics Check		OK	10:58
Air Blank		0.000	10:58
Control Test		0.080	10:59
Air Blank		0.000	10:59
Subject Sample	# 1	0.032	11:01
Breath Volume		1.937	11:01
Air Blank		0.000	11:02
Air Blank		0.000	11:04
Subject Sample	# 2	0.031	11:05
Breath Volume		1.187	11:05
Air Blank		0.000	11:06
Control Test		0.080	11:06
Air Blank		0.000	11:07
Diagnostics Check		OK	11:07

Data Download Date/Time: 09/23/2025 12:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: ZELENKA

DATE/TIME OF TEST: 08/28/2025 22:49:15

Diagnostics Check		OK	22:42
Air Blank		0.000	22:42
Control Test		0.080	22:43
Air Blank		0.000	22:43
Subject Sample	# 1	0.146	22:44
Breath Volume		3.007	22:44
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 2	0.144	22:47
Breath Volume		3.023	22:47
Air Blank		0.000	22:48
Control Test		0.080	22:48
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 09/23/2025 12:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUEVARA,JIMMY,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: RENO,K
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 08/29/2025 23:21:47

Diagnostics Check		OK	23:15
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:16
Subject Sample	# 1	0.157	23:17
Breath Volume		2.203	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:19
Subject Sample	# 2	0.166	23:20
Breath Volume		1.917	23:20
Air Blank		0.000	23:20
Control Test		0.079	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 09/23/2025 12:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,MARK,K
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ADDINGTON,TODD
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 08/30/2025 00:33:00

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:30	
Control Test		0.080	00:30	
Air Blank		0.000	00:31	
Subject Sample	# 1		00:31	Subject Test Refused
Breath Volume		0.000	00:31	
Air Blank		0.000	00:32	
Control Test		0.079	00:32	
Air Blank		0.000	00:32	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 09/23/2025 12:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,DALTON,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: LEONARD,HAYDEN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: CASTRO GONZALEZ

DATE/TIME OF TEST: 08/03/2025 02:28:37

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.079	02:23
Air Blank		0.000	02:23
Subject Sample	# 1	0.121	02:24
Breath Volume		1.320	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.118	02:27
Breath Volume		1.812	02:27
Air Blank		0.000	02:27
Control Test		0.079	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 08/22/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,YAMILET,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MORALES,FABIAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:14

SUBJECT'S LAST NAME: HUMPHREY

DATE/TIME OF TEST: 08/09/2025 08:41:51

Diagnostics Check		OK	08:35
Air Blank		0.000	08:36
Control Test		0.079	08:36
Air Blank		0.000	08:36
Subject Sample	# 1	0.079	08:37
Breath Volume		1.980	08:37
Air Blank		0.000	08:38
Air Blank		0.000	08:39
Subject Sample	# 2	0.078	08:40
Breath Volume		2.121	08:40
Air Blank		0.000	08:40
Control Test		0.079	08:41
Air Blank		0.000	08:41
Diagnostics Check		OK	08:41

Data Download Date/Time: 08/22/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWTON,BRENNAN,P
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: CUNDLE,MORGAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 08/15/2025 18:11:19

Diagnostics Check		OK	18:05
Air Blank		0.000	18:05
Control Test		0.079	18:06
Air Blank		0.000	18:06
Subject Sample	# 1	0.000	18:07
Breath Volume		1.867	18:07
Air Blank		0.000	18:07
Air Blank		0.000	18:09
Subject Sample	# 2	0.000	18:09
Breath Volume		1.617	18:09
Air Blank		0.000	18:10
Control Test		0.080	18:10
Air Blank		0.000	18:11
Diagnostics Check		OK	18:11

Data Download Date/Time: 08/22/2025 10:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERZ,JUSTIN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: DAIGNEAULT,CODY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: ZERBY

DATE/TIME OF TEST: 08/16/2025 23:24:22

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.079	23:18
Air Blank		0.000	23:19
Subject Sample	# 1	0.111	23:19
Breath Volume		1.476	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.121	23:22
Breath Volume		2.015	23:22
Air Blank		0.000	23:23
Control Test		0.079	23:23
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 08/22/2025 10:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIGH,ADAM,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: QUAGLIA,ERIC
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 08/24/2025 00:38:41

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.079	00:36	
Air Blank		0.000	00:36	
Subject Sample	# 1		00:36	RFI Detect
Breath Volume		0.460	00:36	
Air Blank		0.000	00:37	
Control Test		0.080	00:37	
Air Blank			00:38	RFI Detect
Air Blank		0.000	00:38	

Data Download Date/Time: 09/29/2025 12:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIGH,ADAM,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: QUAGLIA,ERIC
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 08/24/2025 00:47:17

Diagnostics Check		OK	00:40
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.088	00:42
Breath Volume		2.464	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.090	00:45
Breath Volume		2.644	00:45
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 09/29/2025 12:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARTOFF,SCOTT,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: CHARTOFF,SCOTT
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RIESS

DATE/TIME OF TEST: 08/24/2025 03:24:20

Diagnostics Check	Fail	03:23	Temperature Regulation Test
Air Blank	0.000	03:24	

Data Download Date/Time: 09/29/2025 12:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARTOFF,SCOTT,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: CHARTOFF,SCOTT
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RIESS

DATE/TIME OF TEST: 08/24/2025 03:33:39

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.079	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.089	03:29
Breath Volume		2.253	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.079	03:32
Breath Volume		2.171	03:32
Air Blank		0.000	03:32
Control Test		0.078	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 09/29/2025 12:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS,RODNESHIA,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: LOUIS,RODNESHIA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: LIMA

DATE/TIME OF TEST: 08/27/2025 04:36:58

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.172	04:32
Breath Volume		4.125	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.176	04:35
Breath Volume		4.320	04:35
Air Blank		0.000	04:36
Control Test		0.079	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 09/29/2025 12:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMCZYK,STENLY,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: DIANGNEAULT,CODY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GEISEL

DATE/TIME OF TEST: 08/28/2025 23:17:02

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.000	23:12
Breath Volume		2.460	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.000	23:15
Breath Volume		2.500	23:15
Air Blank		0.000	23:16
Control Test		0.080	23:16
Air Blank		0.000	23:16
Diagnostics Check		OK	23:17

Data Download Date/Time: 09/29/2025 12:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTISTELLI,TAYLOR,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: CONNER,MICHAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/30/2025 03:24:30

Diagnostics Check		OK	03:21	
Air Blank		0.000	03:22	
Control Test		0.080	03:22	
Air Blank		0.000	03:22	
Subject Sample	# 1		03:24	Sequence Aborted
Breath Volume		0.000	03:24	
Air Blank		0.000	03:24	

Data Download Date/Time: 09/29/2025 12:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: M,NEILL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: DAMIS

DATE/TIME OF TEST: 08/25/2025 22:25:51

Diagnostics Check		OK	22:23	
Air Blank		0.000	22:23	
Control Test		0.082	22:23	
Air Blank		0.000	22:24	
Subject Sample	# 1		22:24	Subject Test Refused
Breath Volume		0.000	22:24	
Air Blank		0.000	22:24	
Control Test		0.081	22:25	
Air Blank		0.000	22:25	
Diagnostics Check		OK	22:25	

Data Download Date/Time: 09/29/2025 09:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,JOHN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRANDES,M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: PERIS

DATE/TIME OF TEST: 08/28/2025 16:21:22

Diagnostics Check		OK	16:18	
Air Blank		0.000	16:18	
Control Test		0.082	16:19	
Air Blank		0.000	16:19	
Subject Sample	# 1		16:20	Subject Test Refused
Breath Volume		0.000	16:20	
Air Blank		0.000	16:20	
Control Test		0.082	16:20	
Air Blank		0.000	16:21	
Diagnostics Check		OK	16:21	

Data Download Date/Time: 09/29/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ESTRELLA,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: HAYDEN

DATE/TIME OF TEST: 08/29/2025 20:13:45

Diagnostics Check		OK	20:10	
Air Blank		0.000	20:11	
Control Test		0.082	20:11	
Air Blank		0.000	20:12	
Subject Sample	# 1		20:12	Subject Test Refused
Breath Volume		0.000	20:12	
Air Blank		0.000	20:12	
Control Test		0.082	20:13	
Air Blank		0.000	20:13	
Diagnostics Check		OK	20:13	

Data Download Date/Time: 09/29/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GLINES,W
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: KETTLEWELL

DATE/TIME OF TEST: 08/30/2025 02:33:02

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.082	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.000	02:28
Breath Volume		2.863	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.000	02:31
Breath Volume		3.402	02:31
Air Blank		0.000	02:32
Control Test		0.082	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:33

Data Download Date/Time: 09/29/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,OSMAY,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PORTOCARRERO,CHRISTIAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 06:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: CASTRO LOPEZ

DATE/TIME OF TEST: 08/31/2025 08:54:12

Diagnostics Check		OK	08:47
Air Blank		0.000	08:48
Control Test		0.082	08:48
Air Blank		0.000	08:48
Subject Sample	# 1	0.059	08:49
Breath Volume		1.640	08:49
Air Blank		0.000	08:50
Air Blank		0.000	08:52
Subject Sample	# 2	0.056	08:52
Breath Volume		1.695	08:52
Air Blank		0.000	08:53
Control Test		0.082	08:53
Air Blank		0.000	08:54
Diagnostics Check		OK	08:54

Data Download Date/Time: 09/29/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NEILL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: GLASSETT

DATE/TIME OF TEST: 08/31/2025 20:54:29

Diagnostics Check		OK	20:48
Air Blank		0.000	20:48
Control Test		0.082	20:49
Air Blank		0.000	20:49
Subject Sample	# 1	0.222	20:49
Breath Volume		2.167	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.216	20:52
Breath Volume		2.796	20:52
Air Blank		0.000	20:53
Control Test		0.081	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 09/29/2025 09:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,GREGORY,W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,GREGORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: JIMENEZ LOPEZ

DATE/TIME OF TEST: 08/01/2025 20:47:18

Diagnostics Check		OK	20:41
Air Blank		0.000	20:41
Control Test		0.079	20:41
Air Blank		0.000	20:42
Subject Sample	# 1	0.136	20:42
Breath Volume		2.441	20:42
Air Blank		0.000	20:43
Air Blank		0.000	20:45
Subject Sample	# 2	0.142	20:45
Breath Volume		2.449	20:45
Air Blank		0.000	20:46
Control Test		0.079	20:46
Air Blank		0.000	20:47
Diagnostics Check		OK	20:47

Data Download Date/Time: 08/28/2025 09:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: BOBKO

DATE/TIME OF TEST: 08/01/2025 22:39:07

Diagnostics Check		OK	22:32
Air Blank		0.000	22:33
Control Test		0.081	22:33
Air Blank		0.000	22:34
Subject Sample	# 1	0.000	22:34
Breath Volume		3.339	22:34
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.000	22:37
Breath Volume		1.429	22:37
Air Blank		0.000	22:38
Control Test		0.081	22:38
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 08/28/2025 09:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KISMARTONI,LES
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/02/2025 04:15:09

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.080	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.126	04:10
Breath Volume		2.484	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.122	04:13
Breath Volume		2.246	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 08/28/2025 09:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TILAN,ALEX
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: MARCIANO

DATE/TIME OF TEST: 08/03/2025 01:29:11

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.080	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.082	01:24
Breath Volume		1.640	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.080	01:27
Breath Volume		1.667	01:27
Air Blank		0.000	01:28
Control Test		0.079	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 08/28/2025 09:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 08/03/2025 23:39:17

Diagnostics Check		OK	23:32
Air Blank		0.000	23:32
Control Test		0.081	23:32
Air Blank		0.000	23:33
Subject Sample	# 1	0.183	23:34
Breath Volume		1.898	23:34
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.190	23:37
Breath Volume		1.773	23:37
Air Blank		0.000	23:38
Control Test		0.082	23:38
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 08/28/2025 09:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LUCE,C
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 08/04/2025 12:16:59

Diagnostics Check		OK	12:10
Air Blank		0.000	12:10
Control Test		0.080	12:11
Air Blank		0.000	12:11
Subject Sample	# 1	0.176	12:12
Breath Volume		1.515	12:12
Air Blank		0.000	12:12
Air Blank		0.000	12:14
Subject Sample	# 2	0.180	12:15
Breath Volume		1.570	12:15
Air Blank		0.000	12:16
Control Test		0.080	12:16
Air Blank		0.000	12:16
Diagnostics Check		OK	12:16

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLADVANDI,OFC. M.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: MERICLE,OFC. C.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: TUR

DATE/TIME OF TEST: 08/07/2025 20:59:40

Diagnostics Check		OK	20:44	
Air Blank		0.000	20:45	
Control Test		0.079	20:45	
Air Blank		0.000	20:45	
Subject Sample	# 1	0.243	20:48	
Breath Volume		1.160	20:48	
Air Blank		0.000	20:49	
Air Blank		0.000	20:50	
Subject Sample	# 2	0.249	20:53	Volume Not Met
Breath Volume		0.230	20:53	
Air Blank		0.000	20:54	
Air Blank		0.000	20:56	
Subject Sample	# 3	0.228	20:58	
Breath Volume		1.191	20:58	
Air Blank		0.000	20:58	
Control Test		0.079	20:59	
Air Blank		0.000	20:59	
Diagnostics Check		OK	20:59	

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRIMAN,PAUL,I
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: WHISNANT

DATE/TIME OF TEST: 08/08/2025 00:28:09

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.080	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.280	00:23
Breath Volume		1.867	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:25
Subject Sample	# 2	0.275	00:26
Breath Volume		2.015	00:26
Air Blank		0.000	00:27
Control Test		0.078	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: SMITHBOWER

DATE/TIME OF TEST: 08/08/2025 19:14:42

Diagnostics Check		OK	19:08
Air Blank		0.000	19:09
Control Test		0.079	19:09
Air Blank		0.000	19:09
Subject Sample	# 1	0.110	19:10
Breath Volume		2.308	19:10
Air Blank		0.000	19:10
Air Blank		0.000	19:12
Subject Sample	# 2	0.108	19:13
Breath Volume		2.472	19:13
Air Blank		0.000	19:13
Control Test		0.079	19:14
Air Blank		0.000	19:14
Diagnostics Check		OK	19:14

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LARIMER

DATE/TIME OF TEST: 08/10/2025 03:25:56

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.153	03:20
Breath Volume		2.113	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.153	03:24
Breath Volume		1.808	03:24
Air Blank		0.000	03:25
Control Test		0.079	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CAMBRIA,STEVEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: RAPPUHN\

DATE/TIME OF TEST: 08/10/2025 03:36:27

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.080	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.145	03:31
Breath Volume		2.675	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.142	03:34
Breath Volume		3.289	03:34
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PETROWSKI,DAVID
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/10/2025 05:08:54

Diagnostics Check		OK	05:02
Air Blank		0.000	05:03
Control Test		0.080	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.123	05:04
Breath Volume		1.425	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:06
Subject Sample	# 2	0.130	05:07
Breath Volume		2.730	05:07
Air Blank		0.000	05:07
Control Test		0.079	05:08
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PEREZ,ORLANDO
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 08/11/2025 00:10:24

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.080	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.222	00:05
Breath Volume		2.140	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.209	00:08
Breath Volume		1.894	00:08
Air Blank		0.000	00:09
Control Test		0.078	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STANLEY,ANDY
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MAROT

DATE/TIME OF TEST: 08/12/2025 03:24:25

Diagnostics Check		OK	03:12	
Air Blank		0.000	03:13	
Control Test		0.080	03:13	
Air Blank		0.000	03:14	
Subject Sample	# 1	0.151	03:17	Volume Not Met
Breath Volume		0.214	03:17	
Air Blank		0.000	03:17	
Air Blank		0.000	03:19	
Subject Sample	# 2	0.167	03:22	Volume Not Met
Breath Volume		0.117	03:22	
Air Blank		0.000	03:23	
Control Test		0.078	03:23	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERSTEN,DANIEL,N
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: NERSTEN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 08/12/2025 11:23:22

Diagnostics Check		OK	11:16
Air Blank		0.000	11:17
Control Test		0.080	11:17
Air Blank		0.000	11:17
Subject Sample	# 1	0.000	11:18
Breath Volume		1.148	11:18
Air Blank		0.000	11:19
Air Blank		0.000	11:21
Subject Sample	# 2	0.000	11:22
Breath Volume		1.992	11:22
Air Blank		0.000	11:22
Control Test		0.080	11:22
Air Blank		0.000	11:23
Diagnostics Check		OK	11:23

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRIMAN,PAUL,I
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PETROWSKI,DAVID
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: HOUDE

DATE/TIME OF TEST: 08/13/2025 23:54:35

Diagnostics Check		OK	23:48
Air Blank		0.000	23:48
Control Test		0.079	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.194	23:50
Breath Volume		2.226	23:50
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.191	23:53
Breath Volume		2.164	23:53
Air Blank		0.000	23:53
Control Test		0.079	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRIMAN,PAUL,I
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PETROWSKI,DAVID
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: ROMAIN

DATE/TIME OF TEST: 08/15/2025 02:17:10

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.080	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.064	02:12
Breath Volume		1.832	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:14
Subject Sample	# 2	0.066	02:15
Breath Volume		1.542	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: BOSTICK

DATE/TIME OF TEST: 08/16/2025 00:21:52

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.080	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.000	00:17
Breath Volume		1.105	00:17
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.000	00:20
Breath Volume		1.144	00:20
Air Blank		0.000	00:20
Control Test		0.079	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,ANDREW,G
BREATH TEST OPERATOR AGENCY: Venice PD
ARRESTING OFFICER: NELSON,ANDREW
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: LEONARDO

DATE/TIME OF TEST: 08/16/2025 12:11:52

Diagnostics Check		OK	12:05
Air Blank		0.000	12:05
Control Test		0.079	12:06
Air Blank		0.000	12:06
Subject Sample	# 1	0.297	12:07
Breath Volume		2.878	12:07
Air Blank		0.000	12:08
Air Blank		0.000	12:09
Subject Sample	# 2	0.299	12:10
Breath Volume		2.882	12:10
Air Blank		0.000	12:10
Control Test		0.078	12:11
Air Blank		0.000	12:11
Diagnostics Check		OK	12:11

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STANLEY,ANDY
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/17/2025 04:14:55

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.081	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.076	04:10
Breath Volume		1.207	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:12
Subject Sample	# 2	0.073	04:13
Breath Volume		1.226	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: JONES,NICHOLAS
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: EMILY

DATE/TIME OF TEST: 08/17/2025 04:42:01

Diagnostics Check		OK	04:33
Air Blank		0.000	04:34
Control Test		0.081	04:34
Air Blank		0.000	04:34
Subject Sample	# 1	0.173	04:36
Breath Volume		1.644	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.177	04:40
Breath Volume		1.179	04:40
Air Blank		0.000	04:41
Control Test		0.079	04:41
Air Blank		0.000	04:41
Diagnostics Check		OK	04:42

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: TERAN

DATE/TIME OF TEST: 08/18/2025 04:30:24

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.078	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.174	04:25
Breath Volume		1.332	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.187	04:28
Breath Volume		1.566	04:28
Air Blank		0.000	04:29
Control Test		0.080	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 08/19/2025 20:27:55

Diagnostics Check		OK	20:18
Air Blank		0.000	20:19
Control Test		0.080	20:19
Air Blank		0.000	20:20
Subject Sample	# 1	0.000	20:22
Breath Volume		1.328	20:22
Air Blank		0.000	20:23
Air Blank		0.000	20:25
Subject Sample	# 2	0.000	20:26
Breath Volume		1.687	20:26
Air Blank		0.000	20:27
Control Test		0.079	20:27
Air Blank		0.000	20:27
Diagnostics Check		OK	20:27

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLAVANDI,OFC. M.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: FOLADVANDI,OFC. M.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: CHANEY JR

DATE/TIME OF TEST: 08/20/2025 22:42:26

Diagnostics Check		OK	22:34
Air Blank		0.000	22:34
Control Test		0.079	22:35
Air Blank		0.000	22:35
Subject Sample	# 1	0.111	22:37
Breath Volume		1.171	22:37
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.102	22:40
Breath Volume		2.707	22:40
Air Blank		0.000	22:41
Control Test		0.079	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: SGAMBELLONE

DATE/TIME OF TEST: 08/20/2025 22:52:17

Diagnostics Check		OK	22:46
Air Blank		0.000	22:46
Control Test		0.078	22:46
Air Blank		0.000	22:47
Subject Sample	# 1	0.170	22:47
Breath Volume		1.996	22:47
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.173	22:50
Breath Volume		2.105	22:50
Air Blank		0.000	22:51
Control Test		0.079	22:51
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLADVANDI,OFC. M.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: FOLADVANDI,OFC. M.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: KRAUTTER

DATE/TIME OF TEST: 08/21/2025 04:13:15

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.307	04:08
Breath Volume		1.468	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.311	04:11
Breath Volume		1.800	04:11
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: SCHORN

DATE/TIME OF TEST: 08/21/2025 05:44:21

Diagnostics Check		OK	05:37
Air Blank		0.000	05:38
Control Test		0.080	05:38
Air Blank		0.000	05:38
Subject Sample	# 1	0.196	05:39
Breath Volume		3.804	05:39
Air Blank		0.000	05:40
Air Blank		0.000	05:42
Subject Sample	# 2	0.204	05:42
Breath Volume		3.644	05:42
Air Blank		0.000	05:43
Control Test		0.080	05:43
Air Blank		0.000	05:44
Diagnostics Check		OK	05:44

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEATY,JARED,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BEATY,JARED
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: PINILLA SUAREZ

DATE/TIME OF TEST: 08/22/2025 02:59:03

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.027	02:54
Breath Volume		1.398	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.026	02:57
Breath Volume		1.574	02:57
Air Blank		0.000	02:58
Control Test		0.080	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:59

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRIMAN,PAUL,I
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: THORPE,ZACHARY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 08/23/2025 02:15:32

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.178	02:11
Breath Volume		3.449	02:11
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.171	02:14
Breath Volume		3.234	02:14
Air Blank		0.000	02:14
Control Test		0.079	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: SWEDE

DATE/TIME OF TEST: 08/23/2025 04:15:12

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:01	
Control Test		0.076	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1	0.188	04:03	
Breath Volume		1.562	04:03	
Air Blank		0.000	04:04	
Air Blank		0.000	04:06	
Subject Sample	# 2	0.161	04:09	Volume Not Met
Breath Volume		0.683	04:09	
Air Blank		0.000	04:10	
Air Blank		0.000	04:11	
Subject Sample	# 3	0.207	04:13	
Breath Volume		2.175	04:13	
Air Blank		0.000	04:14	
Control Test		0.078	04:14	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,JAMIE,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PETROWSKI,DAVID
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: DALBEM DE OLIVEIRA

DATE/TIME OF TEST: 08/24/2025 00:52:29

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.081	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.231	00:47
Breath Volume		2.441	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.219	00:50
Breath Volume		1.488	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: ALBARRAN SANCHEZ

DATE/TIME OF TEST: 08/24/2025 05:05:39

Diagnostics Check		OK	04:59
Air Blank		0.000	04:59
Control Test		0.080	05:00
Air Blank		0.000	05:00
Subject Sample	# 1	0.084	05:00
Breath Volume		4.070	05:00
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.084	05:04
Breath Volume		3.328	05:04
Air Blank		0.000	05:04
Control Test		0.079	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUGAN,KASEY,L
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: BERGSTAD,BRANDON
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: NEMBHARD

DATE/TIME OF TEST: 08/25/2025 05:18:00

Diagnostics Check		OK	05:04	
Air Blank		0.000	05:05	
Control Test		0.080	05:05	
Air Blank		0.000	05:06	
Subject Sample	# 1	0.134	05:09	Volume Not Met
Breath Volume		0.375	05:09	
Air Blank		0.000	05:09	
Air Blank		0.000	05:11	
Subject Sample	# 2	0.133	05:13	
Breath Volume		1.726	05:13	
Air Blank		0.000	05:13	
Air Blank		0.000	05:15	
Subject Sample	# 3	0.130	05:16	
Breath Volume		2.027	05:16	
Air Blank		0.000	05:17	
Control Test		0.078	05:17	
Air Blank		0.000	05:17	
Diagnostics Check		OK	05:18	

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTKO,MARIAM,L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NADLER,ISABELLA
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 08/27/2025 02:23:22

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.191	02:17
Breath Volume		1.296	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.196	02:21
Breath Volume		1.710	02:21
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: MERDA

DATE/TIME OF TEST: 08/28/2025 01:34:33

Diagnostics Check		OK	01:27
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.181	01:29
Breath Volume		2.210	01:29
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.177	01:32
Breath Volume		1.863	01:32
Air Blank		0.000	01:33
Control Test		0.078	01:33
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 08/28/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: SIZEMORE

DATE/TIME OF TEST: 08/28/2025 22:10:26

Diagnostics Check		OK	22:02
Air Blank		0.000	22:03
Control Test		0.080	22:03
Air Blank		0.000	22:04
Subject Sample	# 1	0.148	22:05
Breath Volume		1.687	22:05
Air Blank		0.000	22:06
Air Blank		0.000	22:07
Subject Sample	# 2	0.144	22:08
Breath Volume		2.218	22:08
Air Blank		0.000	22:09
Control Test		0.079	22:09
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 08/30/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: MYER

DATE/TIME OF TEST: 08/28/2025 22:48:30

Diagnostics Check		OK	22:40
Air Blank		0.000	22:41
Control Test		0.079	22:41
Air Blank		0.000	22:42
Subject Sample	# 1	0.033	22:43
Breath Volume		1.320	22:43
Air Blank		0.000	22:44
Air Blank		0.000	22:46
Subject Sample	# 2	0.032	22:46
Breath Volume		1.308	22:46
Air Blank		0.000	22:47
Control Test		0.078	22:47
Air Blank		0.000	22:48
Diagnostics Check		OK	22:48

Data Download Date/Time: 08/30/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FERNANDEZ,JESSICA
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MIDAS

DATE/TIME OF TEST: 08/28/2025 23:02:28

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.079	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.175	22:58
Breath Volume		2.304	22:58
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.170	23:00
Breath Volume		1.953	23:00
Air Blank		0.000	23:01
Control Test		0.079	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 08/30/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTKO,MARIAM,L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MANNING,ZACHARY
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: HANSBERGER

DATE/TIME OF TEST: 08/29/2025 22:33:56

Diagnostics Check		OK	22:27
Air Blank		0.000	22:28
Control Test		0.081	22:28
Air Blank		0.000	22:29
Subject Sample	# 1	0.058	22:29
Breath Volume		2.074	22:29
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.060	22:32
Breath Volume		2.144	22:32
Air Blank		0.000	22:33
Control Test		0.080	22:33
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 08/30/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTKO,MARIAM,L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BAKER,DUSTIN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: DIETCH

DATE/TIME OF TEST: 08/30/2025 01:49:08

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:46	
Control Test		0.081	01:46	
Air Blank		0.000	01:46	
Subject Sample	# 1		01:48	RFI Detect
Breath Volume		0.113	01:48	
Air Blank			01:48	RFI Detect
Air Blank		0.000	01:49	

Data Download Date/Time: 08/30/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTKO,MARIAM,L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BAKER,DUSTIN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: DIETCH

DATE/TIME OF TEST: 08/30/2025 02:06:16

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:54	
Control Test		0.080	01:55	
Air Blank		0.000	01:55	
Subject Sample	# 1	0.146	01:58	Volume Not Met
Breath Volume		0.703	01:58	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 2	0.171	02:01	
Breath Volume		1.167	02:01	
Air Blank		0.000	02:02	
Air Blank		0.000	02:04	
Subject Sample	# 3	0.185	02:04	
Breath Volume		1.371	02:04	
Air Blank		0.000	02:05	
Control Test		0.078	02:05	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 08/30/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTKO,MARIAM,L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SHACO,COURTNEY
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: RAMIREZ CARRILLO

DATE/TIME OF TEST: 08/30/2025 03:43:36

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.081	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.074	03:39
Breath Volume		1.113	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.074	03:42
Breath Volume		1.804	03:42
Air Blank		0.000	03:42
Control Test		0.080	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 08/30/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 08/30/2025 04:20:10

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:12	
Control Test		0.080	04:12	
Air Blank		0.000	04:12	
Subject Sample	# 1		04:16	No Sample Provided
Breath Volume		0.000	04:16	
Air Blank		0.000	04:16	
Air Blank		0.000	04:18	
Subject Sample	# 2		04:18	Subject Test Refused
Breath Volume		0.000	04:18	
Air Blank		0.000	04:19	
Control Test		0.077	04:19	
Air Blank		0.000	04:20	
Diagnostics Check		OK	04:20	

Data Download Date/Time: 08/30/2025 12:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPOTO,MATTHEW,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SPOTO,MATTHEW
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: FLYNN ALVAREZ

DATE/TIME OF TEST: 08/01/2025 20:33:28

Diagnostics Check		OK	20:24	
Air Blank		0.000	20:24	
Control Test		0.080	20:25	
Air Blank		0.000	20:25	
Subject Sample	# 1	0.170	20:26	
Breath Volume		1.304	20:26	
Air Blank		0.000	20:27	
Air Blank		0.000	20:29	
Subject Sample	# 2		20:31	Slope Not Met
Breath Volume		0.968	20:31	
Air Blank		0.000	20:32	
Control Test		0.081	20:32	
Air Blank		0.000	20:33	
Diagnostics Check		OK	20:33	

Data Download Date/Time: 08/27/2025 09:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPOTO,MATTHEW,R
BREATH TEST OPERATOR AGENCY: Dundee PD
ARRESTING OFFICER: SPOTO,MATTHEW
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: FLYNN ALVAREZ

DATE/TIME OF TEST: 08/01/2025 20:42:16

Diagnostics Check		OK	20:35
Air Blank		0.000	20:35
Control Test		0.080	20:36
Air Blank		0.000	20:36
Subject Sample	# 1	0.164	20:37
Breath Volume		1.527	20:37
Air Blank		0.000	20:38
Air Blank		0.000	20:40
Subject Sample	# 2	0.161	20:40
Breath Volume		1.660	20:40
Air Blank		0.000	20:41
Control Test		0.079	20:41
Air Blank		0.000	20:42
Diagnostics Check		OK	20:42

Data Download Date/Time: 08/27/2025 09:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: SUCHERMAN

DATE/TIME OF TEST: 08/04/2025 21:56:59

Diagnostics Check		OK	21:50
Air Blank		0.000	21:51
Control Test		0.081	21:51
Air Blank		0.000	21:51
Subject Sample	# 1	0.128	21:52
Breath Volume		1.281	21:52
Air Blank		0.000	21:53
Air Blank		0.000	21:54
Subject Sample	# 2	0.122	21:55
Breath Volume		1.414	21:55
Air Blank		0.000	21:56
Control Test		0.080	21:56
Air Blank		0.000	21:56
Diagnostics Check		OK	21:56

Data Download Date/Time: 08/27/2025 09:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: BULTMAN-POWERS

DATE/TIME OF TEST: 08/09/2025 00:49:53

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.081	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.186	00:45
Breath Volume		1.359	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:47
Subject Sample	# 2	0.187	00:48
Breath Volume		1.105	00:48
Air Blank		0.000	00:48
Control Test		0.080	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 08/27/2025 09:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/10/2025 04:13:59

Diagnostics Check		OK	03:59	
Air Blank		0.000	03:59	
Control Test		0.081	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 1	0.199	04:01	
Breath Volume		1.136	04:01	
Air Blank		0.000	04:01	
Air Blank		0.000	04:03	
Subject Sample	# 2	0.163	04:06	Volume Not Met
Breath Volume		0.250	04:06	
Air Blank		0.000	04:07	
Air Blank		0.000	04:09	
Subject Sample	# 3	0.186	04:12	Volume Not Met
Breath Volume		0.910	04:12	
Air Blank		0.000	04:13	
Control Test		0.079	04:13	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 08/27/2025 09:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 08/10/2025 17:22:58

Diagnostics Check		OK	17:13
Air Blank		0.000	17:14
Control Test		0.080	17:14
Air Blank		0.000	17:15
Subject Sample	# 1	0.220	17:15
Breath Volume		1.238	17:15
Air Blank		0.000	17:16
Air Blank		0.000	17:18
Subject Sample	# 2	0.241	17:18
Breath Volume		1.882	17:18
Air Blank		0.000	17:19
Air Blank		0.000	17:21
Subject Sample	# 3	0.232	17:21
Breath Volume		1.152	17:21
Air Blank		0.000	17:22
Control Test		0.080	17:22
Air Blank		0.000	17:22
Diagnostics Check		OK	17:22

Data Download Date/Time: 08/27/2025 09:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: CARDOSA

DATE/TIME OF TEST: 08/30/2025 00:04:51

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.081	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.024	00:00
Breath Volume		2.710	00:00
Air Blank		0.000	00:01
Air Blank		0.000	00:02
Subject Sample	# 2	0.027	00:03
Breath Volume		2.796	00:03
Air Blank		0.000	00:03
Control Test		0.080	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 09/26/2025 10:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: NEKOLOFF

DATE/TIME OF TEST: 08/30/2025 03:00:30

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.166	02:55
Breath Volume		1.820	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.170	02:58
Breath Volume		1.535	02:58
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 09/26/2025 10:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPOTO,MATTHEW,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SHAWGER,J
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/30/2025 03:53:26

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.227	03:48
Breath Volume		1.210	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.230	03:51
Breath Volume		1.539	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 09/26/2025 10:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPOTO,MATTHEW,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SPOTO,MATTHEW
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/30/2025 04:21:50

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.166	04:16
Breath Volume		1.121	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.184	04:20
Breath Volume		1.246	04:20
Air Blank		0.000	04:20
Control Test		0.079	04:21
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 09/26/2025 10:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPOTO,MATTHEW,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SPOTO,MATTHEW
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: NEVILS

DATE/TIME OF TEST: 08/31/2025 02:16:33

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:11	
Control Test		0.080	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.081	02:15	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 09/26/2025 10:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPOTO,MATTHEW,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SPOTO,MATTHEW
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: GREENWELL

DATE/TIME OF TEST: 08/31/2025 21:19:14

Diagnostics Check		OK	21:11
Air Blank		0.000	21:11
Control Test		0.081	21:12
Air Blank		0.000	21:12
Subject Sample	# 1	0.138	21:13
Breath Volume		3.054	21:13
Air Blank		0.000	21:14
Air Blank		0.000	21:16
Subject Sample	# 2	0.118	21:17
Breath Volume		1.167	21:17
Air Blank		0.000	21:18
Control Test		0.081	21:18
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 09/26/2025 10:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTKO,MARIAM,L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BIEKER,BRENT
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: GOINS

DATE/TIME OF TEST: 08/30/2025 21:13:03

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.080	21:05
Air Blank		0.000	21:05
Subject Sample	# 1	0.143	21:08
Breath Volume		1.402	21:08
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.139	21:11
Breath Volume		1.613	21:11
Air Blank		0.000	21:12
Control Test		0.080	21:12
Air Blank		0.000	21:12
Diagnostics Check		OK	21:13

Data Download Date/Time: 09/29/2025 18:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,RIDGE,
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: COLLINS,RIDGE
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: DECK

DATE/TIME OF TEST: 08/31/2025 03:12:35

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.150	03:08
Breath Volume		2.386	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.148	03:11
Breath Volume		1.882	03:11
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 09/29/2025 18:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 08/31/2025 03:24:28

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.134	03:19
Breath Volume		1.414	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.137	03:22
Breath Volume		2.042	03:22
Air Blank		0.000	03:23
Control Test		0.079	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 09/29/2025 18:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,RIDGE,
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: MARTINEZ,OSCAR
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: RAMOS DEL LLANO

DATE/TIME OF TEST: 08/31/2025 03:37:10

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.079	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.163	03:32
Breath Volume		1.976	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.167	03:35
Breath Volume		1.707	03:35
Air Blank		0.000	03:36
Control Test		0.079	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 09/29/2025 18:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,RIDGE,
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: JONES,NICHOLAS
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BASURTO MONDRAGON

DATE/TIME OF TEST: 08/31/2025 03:48:16

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.080	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.085	03:43
Breath Volume		1.148	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.087	03:46
Breath Volume		1.679	03:46
Air Blank		0.000	03:47
Control Test		0.078	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 09/29/2025 18:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

INSTRUMENT REGISTERED TO: BARTOW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: ROQUE,ALEXANDER
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: CENTENO

DATE/TIME OF TEST: 08/22/2025 23:11:32

Diagnostics Check		OK	22:59	
Air Blank		0.000	23:00	
Control Test		0.078	23:00	
Air Blank		0.000	23:01	
Subject Sample	# 1		23:04	No Sample Provided
Breath Volume		0.000	23:04	
Air Blank		0.000	23:04	
Air Blank		0.000	23:06	
Subject Sample	# 2		23:09	No Sample Provided
Breath Volume		0.000	23:09	
Air Blank		0.000	23:10	
Control Test		0.080	23:10	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 09/25/2025 17:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

INSTRUMENT REGISTERED TO: BARTOW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: ROQUE,ALEXANDER
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: CENTENO

DATE/TIME OF TEST: 08/22/2025 23:25:33

Diagnostics Check		OK	23:13	
Air Blank		0.000	23:14	
Control Test		0.078	23:14	
Air Blank		0.000	23:15	
Subject Sample	# 1	0.141	23:18	Volume Not Met
Breath Volume		0.285	23:18	
Air Blank		0.000	23:19	
Air Blank		0.000	23:20	
Subject Sample	# 2	0.141	23:23	Volume Not Met
Breath Volume		0.339	23:23	
Air Blank		0.000	23:24	
Control Test		0.076	23:24	
Air Blank		0.000	23:25	
Diagnostics Check		OK	23:25	

Data Download Date/Time: 09/25/2025 17:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

INSTRUMENT REGISTERED TO: BARTOW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: ANZALONE,ROBERT
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: SURRAY

DATE/TIME OF TEST: 08/27/2025 19:02:27

Diagnostics Check		OK	18:56
Air Blank		0.000	18:57
Control Test		0.079	18:57
Air Blank		0.000	18:57
Subject Sample	# 1	0.000	18:58
Breath Volume		1.203	18:58
Air Blank		0.000	18:58
Air Blank		0.000	19:00
Subject Sample	# 2	0.000	19:01
Breath Volume		1.253	19:01
Air Blank		0.000	19:01
Control Test		0.078	19:01
Air Blank		0.000	19:02
Diagnostics Check		OK	19:02

Data Download Date/Time: 09/25/2025 17:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS,TYLER,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/01/2025 22:04:48

Diagnostics Check		OK	21:58
Air Blank		0.000	21:59
Control Test		0.079	21:59
Air Blank		0.000	21:59
Subject Sample	# 1	0.061	22:00
Breath Volume		1.457	22:00
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.058	22:03
Breath Volume		1.933	22:03
Air Blank		0.000	22:03
Control Test		0.078	22:04
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 08/29/2025 11:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS,TYLER,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 08/16/2025 21:24:47

Diagnostics Check		OK	21:18
Air Blank		0.000	21:18
Control Test		0.078	21:18
Air Blank		0.000	21:19
Subject Sample	# 1	0.174	21:20
Breath Volume		1.929	21:20
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.169	21:23
Breath Volume		2.199	21:23
Air Blank		0.000	21:23
Control Test		0.078	21:24
Air Blank		0.000	21:24
Diagnostics Check		OK	21:24

Data Download Date/Time: 08/29/2025 11:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS,TYLER,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: PRAGER

DATE/TIME OF TEST: 08/22/2025 04:26:59

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.078	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.084	04:22
Breath Volume		1.582	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.082	04:25
Breath Volume		1.625	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 08/29/2025 11:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 08/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 08/30/2025 22:49:20

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:48	
Control Test	0.019	22:48	RFI Detect
Air Blank	0.000	22:49	

Data Download Date/Time: 09/30/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 08/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 08/30/2025 22:58:27

Diagnostics Check		OK	22:52
Air Blank		0.000	22:52
Control Test		0.077	22:52
Air Blank		0.000	22:53
Subject Sample	# 1	0.208	22:53
Breath Volume		3.074	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.207	22:56
Breath Volume		2.035	22:56
Air Blank		0.000	22:57
Control Test		0.079	22:57
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 09/30/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS,TYLER,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DOMINGUEZ,ANDREW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: REASONER

DATE/TIME OF TEST: 08/31/2025 02:39:21

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.078	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.133	02:35
Breath Volume		1.511	02:35
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.113	02:37
Breath Volume		1.812	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 09/30/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MOTZ,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: ONEILL

DATE/TIME OF TEST: 08/09/2025 00:16:12

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:05	
Control Test		0.080	00:05	
Air Blank		0.000	00:06	
Subject Sample	# 1	0.000	00:09	Volume Not Met
Breath Volume		0.265	00:09	
Air Blank		0.000	00:09	
Air Blank		0.000	00:11	
Subject Sample	# 2	0.000	00:14	Volume Not Met
Breath Volume		0.769	00:14	
Air Blank		0.000	00:15	
Control Test		0.079	00:15	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 08/22/2025 20:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KUBA,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: SERRUDO NARANJO

DATE/TIME OF TEST: 08/09/2025 03:12:30

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.077	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.146	03:07
Breath Volume		1.281	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.135	03:10
Breath Volume		1.687	03:10
Air Blank		0.000	03:11
Control Test		0.079	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 08/22/2025 20:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GAINVORS,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 08/09/2025 04:19:14

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.080	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.106	04:14
Breath Volume		1.363	04:14
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.101	04:17
Breath Volume		1.550	04:17
Air Blank		0.000	04:18
Control Test		0.080	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 08/22/2025 20:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,CLARENCE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KUBA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: MCGILLIVRAY

DATE/TIME OF TEST: 08/10/2025 02:34:53

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.079	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.137	02:30
Breath Volume		2.105	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.128	02:33
Breath Volume		1.902	02:33
Air Blank		0.000	02:33
Control Test		0.079	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 08/22/2025 20:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KUTELLA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:14

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 08/13/2025 14:04:56

Diagnostics Check		OK	13:58
Air Blank		0.000	13:58
Control Test		0.080	13:58
Air Blank		0.000	13:59
Subject Sample	# 1	0.000	13:59
Breath Volume		2.980	13:59
Air Blank		0.000	14:00
Air Blank		0.000	14:02
Subject Sample	# 2	0.000	14:03
Breath Volume		2.355	14:03
Air Blank		0.000	14:04
Control Test		0.080	14:04
Air Blank		0.000	14:04
Diagnostics Check		OK	14:04

Data Download Date/Time: 08/22/2025 20:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MAGOON,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: CUMMINGS

DATE/TIME OF TEST: 08/13/2025 17:14:47

Diagnostics Check		OK	17:06
Air Blank		0.000	17:07
Control Test		0.080	17:07
Air Blank		0.000	17:08
Subject Sample	# 1	0.063	17:09
Breath Volume		1.679	17:09
Air Blank		0.000	17:10
Air Blank		0.000	17:12
Subject Sample	# 2	0.063	17:13
Breath Volume		1.871	17:13
Air Blank		0.000	17:13
Control Test		0.079	17:14
Air Blank		0.000	17:14
Diagnostics Check		OK	17:14

Data Download Date/Time: 08/22/2025 20:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWTHORNE,OLIVIA,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KUBA,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: SCHULTZ

DATE/TIME OF TEST: 08/15/2025 00:12:58

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.150	00:08
Breath Volume		1.195	00:08
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.142	00:11
Breath Volume		1.191	00:11
Air Blank		0.000	00:12
Control Test		0.079	00:12
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 08/22/2025 20:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MARISSA,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KIRMELEWICZ,L
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 07/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MCFADZEN

DATE/TIME OF TEST: 08/19/2025 01:30:22

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:16	
Control Test		0.080	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1	0.175	01:17	
Breath Volume		2.046	01:17	
Air Blank		0.000	01:18	
Air Blank		0.000	01:20	
Subject Sample	# 2	0.129	01:23	Volume Not Met
Breath Volume		0.648	01:23	
Air Blank		0.000	01:23	
Air Blank		0.000	01:25	
Subject Sample	# 3	0.125	01:28	Volume Not Met
Breath Volume		0.257	01:28	
Air Blank		0.000	01:29	
Control Test		0.079	01:29	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:30	

Data Download Date/Time: 08/22/2025 20:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,K,R
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: SMITH,E
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 08/07/2025 00:11:46

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.080	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.085	00:07
Breath Volume		2.531	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.086	00:10
Breath Volume		2.988	00:10
Air Blank		0.000	00:10
Control Test		0.078	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 08/28/2025 13:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARVAJAL,JUAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: SANCE

DATE/TIME OF TEST: 08/30/2025 02:45:55

Diagnostics Check		OK	02:39
Air Blank		0.000	02:40
Control Test		0.082	02:40
Air Blank		0.000	02:41
Subject Sample	# 1	0.133	02:41
Breath Volume		2.027	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.134	02:44
Breath Volume		1.414	02:44
Air Blank		0.000	02:45
Control Test		0.082	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 09/01/2025 10:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PADILLA SANTILLAN,ALBERTO
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 08/31/2025 02:47:33

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:36	
Control Test		0.081	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 1		02:40	No Sample Provided
Breath Volume		0.000	02:40	
Air Blank		0.000	02:41	
Air Blank		0.000	02:43	
Subject Sample	# 2		02:46	No Sample Provided
Breath Volume		0.000	02:46	
Air Blank		0.000	02:46	
Control Test		0.082	02:46	
Air Blank		0.000	02:47	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 09/01/2025 10:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO,EDWIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 08/31/2025 21:58:45

Diagnostics Check		OK	21:48
Air Blank		0.000	21:48
Control Test		0.082	21:49
Air Blank		0.000	21:49
Subject Sample	# 1	0.224	21:53
Breath Volume		1.593	21:53
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.221	21:57
Breath Volume		1.554	21:57
Air Blank		0.000	21:57
Control Test		0.078	21:58
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 09/01/2025 10:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: FLOYD,D
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 08/09/2025 03:56:09

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.180	03:51
Breath Volume		2.515	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:53
Subject Sample	# 2	0.181	03:54
Breath Volume		2.750	03:54
Air Blank		0.000	03:55
Control Test		0.079	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 09/15/2025 22:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,THOMAS,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HRADECKY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: MCCABE

DATE/TIME OF TEST: 08/10/2025 00:43:37

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:41	
Control Test		0.080	00:41	
Air Blank		0.000	00:41	
Subject Sample	# 1		00:42	Subject Test Refused
Breath Volume		0.000	00:42	
Air Blank		0.000	00:42	
Control Test		0.079	00:43	
Air Blank		0.000	00:43	
Diagnostics Check		OK	00:43	

Data Download Date/Time: 09/15/2025 22:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAVIN,DEMERIS,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARRIS,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: CLARKSON

DATE/TIME OF TEST: 08/15/2025 21:18:48

Diagnostics Check		OK	21:15	
Air Blank		0.000	21:16	
Control Test		0.077	21:16	
Air Blank		0.000	21:17	
Subject Sample	# 1		21:17	Subject Test Refused
Breath Volume		0.000	21:17	
Air Blank		0.000	21:17	
Control Test		0.076	21:18	
Air Blank		0.000	21:18	
Diagnostics Check		OK	21:18	

Data Download Date/Time: 09/15/2025 22:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,PEDRO,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARRIS,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 08/16/2025 03:57:11

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.077	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.252	03:52
Breath Volume		3.027	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.250	03:55
Breath Volume		2.835	03:55
Air Blank		0.000	03:56
Control Test		0.078	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 09/15/2025 22:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,ABIGAIL,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HRADECKY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: CREIGHTON

DATE/TIME OF TEST: 08/18/2025 19:42:07

Diagnostics Check		OK	19:28	
Air Blank		0.000	19:29	
Control Test		0.080	19:29	
Air Blank		0.000	19:30	
Subject Sample	# 1	0.257	19:31	
Breath Volume		1.109	19:31	
Air Blank		0.000	19:32	
Air Blank		0.000	19:34	
Subject Sample	# 2	0.268	19:37	Volume Not Met
Breath Volume		0.738	19:37	
Air Blank		0.000	19:38	
Air Blank		0.000	19:40	
Subject Sample	# 3	0.251	19:40	
Breath Volume		2.000	19:40	
Air Blank		0.000	19:41	
Control Test		0.077	19:41	
Air Blank		0.000	19:42	
Diagnostics Check		OK	19:42	

Data Download Date/Time: 09/15/2025 22:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,ABIGAIL,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HOSKEY,H
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 08/24/2025 21:45:07

Diagnostics Check		OK	21:37
Air Blank		0.000	21:37
Control Test		0.080	21:37
Air Blank		0.000	21:38
Subject Sample	# 1	0.204	21:39
Breath Volume		1.648	21:39
Air Blank		0.000	21:40
Air Blank		0.000	21:41
Subject Sample	# 2	0.207	21:43
Breath Volume		1.769	21:43
Air Blank		0.000	21:44
Control Test		0.079	21:44
Air Blank		0.000	21:45
Diagnostics Check		OK	21:45

Data Download Date/Time: 09/15/2025 22:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,ABIGAIL,M
BREATH TEST OPERATOR AGENCY: Wildwood PD
ARRESTING OFFICER: BLASER,A
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 08/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: LUTTRELL

DATE/TIME OF TEST: 08/25/2025 01:18:56

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.168	01:14
Breath Volume		2.781	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.164	01:17
Breath Volume		3.125	01:17
Air Blank		0.000	01:18
Control Test		0.078	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 09/15/2025 22:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAUENHOFER,TIMOTHY,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FRAUENHOFER,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:38

SUBJECT'S LAST NAME: MCCASKILL

DATE/TIME OF TEST: 08/20/2025 10:06:36

Diagnostics Check		OK	10:00
Air Blank		0.000	10:01
Control Test		0.081	10:01
Air Blank		0.000	10:01
Subject Sample	# 1	0.000	10:02
Breath Volume		2.644	10:02
Air Blank		0.000	10:02
Air Blank		0.000	10:04
Subject Sample	# 2	0.000	10:05
Breath Volume		3.339	10:05
Air Blank		0.000	10:05
Control Test		0.081	10:05
Air Blank		0.000	10:06
Diagnostics Check		OK	10:06

Data Download Date/Time: 08/30/2025 20:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA,R,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WISE,CURITS A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 08/01/2025 20:51:13

Diagnostics Check		OK	20:44
Air Blank		0.000	20:45
Control Test		0.081	20:45
Air Blank		0.000	20:46
Subject Sample	# 1	0.143	20:46
Breath Volume		2.871	20:46
Air Blank		0.000	20:47
Air Blank		0.000	20:49
Subject Sample	# 2	0.147	20:49
Breath Volume		2.519	20:49
Air Blank		0.000	20:50
Control Test		0.080	20:50
Air Blank		0.000	20:51
Diagnostics Check		OK	20:51

Data Download Date/Time: 08/02/2025 13:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,GATLIN,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MARTINEZ,VERONICA
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/16/2025 20:19:56

Diagnostics Check		OK	20:13
Air Blank		0.000	20:13
Control Test		0.082	20:14
Air Blank		0.000	20:14
Subject Sample	# 1	0.173	20:15
Breath Volume		2.218	20:15
Air Blank		0.000	20:15
Air Blank		0.000	20:17
Subject Sample	# 2	0.174	20:18
Breath Volume		2.394	20:18
Air Blank		0.000	20:19
Control Test		0.081	20:19
Air Blank		0.000	20:19
Diagnostics Check		OK	20:19

Data Download Date/Time: 09/28/2025 15:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,MEGAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MARZICOLA,GABRIELLE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: PABLOPABLO

DATE/TIME OF TEST: 08/22/2025 01:24:03

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:20	
Control Test		0.082	01:20	
Air Blank		0.000	01:21	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.218	01:22	
Air Blank		0.000	01:23	
Control Test		0.082	01:23	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 09/28/2025 15:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTZ,BRANDON,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HEARD,AALIYAH
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: LEAL

DATE/TIME OF TEST: 08/01/2025 04:00:51

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:49	
Control Test		0.082	03:50	
Air Blank		0.000	03:50	
Subject Sample	# 1		03:53	No Sample Provided
Breath Volume		0.000	03:53	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 2		03:59	No Sample Provided
Breath Volume		0.000	03:59	
Air Blank		0.000	03:59	
Control Test		0.083	04:00	
Air Blank		0.000	04:00	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTZ,BRANDON,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: WARE

DATE/TIME OF TEST: 08/01/2025 04:39:29

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.083	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.107	04:35
Breath Volume		1.460	04:35
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.099	04:37
Breath Volume		1.292	04:37
Air Blank		0.000	04:38
Control Test		0.081	04:38
Air Blank		0.000	04:39
Diagnostics Check		OK	04:39

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORAN,SEAN,W
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: MORAN,SEAN
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: WITTMANN

DATE/TIME OF TEST: 08/01/2025 23:59:57

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.083	23:54
Air Blank		0.000	23:54
Subject Sample	# 1	0.307	23:55
Breath Volume		1.781	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.301	23:58
Breath Volume		2.429	23:58
Air Blank		0.000	23:59
Control Test		0.082	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORAN,SEAN,W
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: GOSLIN,TYLER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: DIDONATO

DATE/TIME OF TEST: 08/02/2025 01:50:50

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.082	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.185	01:43
Breath Volume		2.210	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.156	01:46
Breath Volume		2.050	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:48
Subject Sample	# 3	0.177	01:49
Breath Volume		3.277	01:49
Air Blank		0.000	01:49
Control Test		0.084	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREGER,TENLEY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: HARRIS,RYAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: GORBE

DATE/TIME OF TEST: 08/03/2025 17:09:51

Diagnostics Check		OK	17:01
Air Blank		0.000	17:02
Control Test		0.083	17:02
Air Blank		0.000	17:03
Subject Sample	# 1	0.148	17:05
Breath Volume		1.589	17:05
Air Blank		0.000	17:05
Air Blank		0.000	17:07
Subject Sample	# 2	0.146	17:08
Breath Volume		1.636	17:08
Air Blank		0.000	17:08
Control Test		0.082	17:09
Air Blank		0.000	17:09
Diagnostics Check		OK	17:09

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUMAKER,THOMAS,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SHUMAKER,THOMAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 08/04/2025 04:04:19

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:52	
Control Test		0.083	03:53	
Air Blank		0.000	03:53	
Subject Sample	# 1	0.218	03:56	Volume Not Met
Breath Volume		0.335	03:56	
Air Blank		0.000	03:57	
Air Blank		0.000	03:59	
Subject Sample	# 2	0.213	04:02	Volume Not Met
Breath Volume		0.566	04:02	
Air Blank		0.000	04:03	
Control Test		0.081	04:03	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUMAKER,THOMAS,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SHUMAKER,THOMAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 08/04/2025 05:32:32

Diagnostics Check		OK	05:23	
Air Blank		0.000	05:23	
Control Test		0.084	05:24	
Air Blank		0.000	05:24	
Subject Sample	# 1	0.201	05:27	Volume Not Met
Breath Volume		0.050	05:27	
Air Blank		0.000	05:28	
Air Blank		0.000	05:30	
Subject Sample	# 2		05:31	Subject Test Refused
Breath Volume		0.000	05:31	
Air Blank		0.000	05:31	
Control Test		0.080	05:31	
Air Blank		0.000	05:32	
Diagnostics Check		OK	05:32	

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUI,DANIEL,J
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RAINSBERGER,LISA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: TRAMMELL

DATE/TIME OF TEST: 08/09/2025 01:21:58

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.084	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.161	01:17
Breath Volume		1.335	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:19
Subject Sample	# 2	0.149	01:20
Breath Volume		1.144	01:20
Air Blank		0.000	01:21
Control Test		0.084	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RUCCI,MATTHEW
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 08/10/2025 03:04:48

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.084	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.168	03:00
Breath Volume		2.289	03:00
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.173	03:03
Breath Volume		2.406	03:03
Air Blank		0.000	03:03
Control Test		0.083	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RUCCI,MATTHEW
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: PAOLINI

DATE/TIME OF TEST: 08/10/2025 22:14:08

Diagnostics Check		OK	22:07
Air Blank		0.000	22:08
Control Test		0.082	22:08
Air Blank		0.000	22:09
Subject Sample	# 1	0.000	22:09
Breath Volume		3.503	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.000	22:12
Breath Volume		3.429	22:12
Air Blank		0.000	22:13
Control Test		0.083	22:13
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SALIFU,EZULDEEN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 08/10/2025 23:58:14

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.000	23:58	RFI Detect
Air Blank		23:58	RFI Detect

Data Download Date/Time: 08/28/2025 22:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SALIFU,EZULDEEN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 08/11/2025 00:07:26

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.083	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.073	00:02
Breath Volume		2.351	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.071	00:05
Breath Volume		1.238	00:05
Air Blank		0.000	00:06
Control Test		0.084	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GOSLIN,TYLER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 08:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:48

SUBJECT'S LAST NAME: FELICIANO

DATE/TIME OF TEST: 08/11/2025 09:15:11

Diagnostics Check		OK	09:08
Air Blank		0.000	09:08
Control Test		0.083	09:09
Air Blank		0.000	09:09
Subject Sample	# 1	0.279	09:10
Breath Volume		1.507	09:10
Air Blank		0.000	09:11
Air Blank		0.000	09:13
Subject Sample	# 2	0.262	09:13
Breath Volume		1.242	09:13
Air Blank		0.000	09:14
Control Test		0.081	09:14
Air Blank		0.000	09:15
Diagnostics Check		OK	09:15

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEROS,KOBE,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: VERA,EDWARD
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/12/2025 14:17:25

Diagnostics Check		OK	14:03	
Air Blank		0.000	14:03	
Control Test		0.083	14:04	
Air Blank		0.000	14:04	
Subject Sample	# 1	0.000	14:07	Volume Not Met
Breath Volume		0.718	14:07	
Air Blank		0.000	14:08	
Air Blank		0.000	14:10	
Subject Sample	# 2	0.000	14:11	
Breath Volume		1.250	14:11	
Air Blank		0.000	14:11	
Air Blank		0.000	14:13	
Subject Sample	# 3	0.000	14:16	
Breath Volume		1.945	14:16	
Air Blank		0.000	14:16	
Control Test		0.083	14:16	
Air Blank		0.000	14:17	
Diagnostics Check		OK	14:17	

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEROS,KOBE,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: VERA,EDWARD
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/12/2025 14:26:19

Diagnostics Check		OK	14:19
Air Blank		0.000	14:20
Control Test		0.082	14:20
Air Blank		0.000	14:20
Subject Sample	# 1	0.000	14:21
Breath Volume		1.949	14:21
Air Blank		0.000	14:21
Air Blank		0.000	14:23
Subject Sample	# 2	0.000	14:24
Breath Volume		1.972	14:24
Air Blank		0.000	14:25
Control Test		0.082	14:25
Air Blank		0.000	14:26
Diagnostics Check		OK	14:26

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,MIKO,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HARRIS,RYAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: AGE

DATE/TIME OF TEST: 08/13/2025 02:18:34

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.082	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.000	02:14
Breath Volume		2.128	02:14
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.000	02:17
Breath Volume		1.500	02:17
Air Blank		0.000	02:17
Control Test		0.083	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 08/14/2025 23:49:23

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.083	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.217	23:42
Breath Volume		2.339	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.221	23:47
Breath Volume		1.593	23:47
Air Blank		0.000	23:48
Control Test		0.082	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRONSON,TAYLOR,K
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BRONSON,TAYLOR
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 08/17/2025 01:28:23

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.084	01:22
Air Blank		0.000	01:23
Subject Sample	# 1	0.213	01:23
Breath Volume		3.617	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.202	01:26
Breath Volume		3.816	01:26
Air Blank		0.000	01:27
Control Test		0.083	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORAN,SEAN,W
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: MCELROY

DATE/TIME OF TEST: 08/17/2025 19:30:53

Diagnostics Check		OK	19:27	
Air Blank		0.000	19:28	
Control Test		0.083	19:28	
Air Blank		0.000	19:29	
Subject Sample	# 1		19:29	Subject Test Refused
Breath Volume		0.000	19:29	
Air Blank		0.000	19:29	
Control Test		0.083	19:30	
Air Blank		0.000	19:30	
Diagnostics Check		OK	19:30	

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MIGLIACCIO,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: KOELKER

DATE/TIME OF TEST: 08/19/2025 23:44:26

Diagnostics Check		OK	23:37
Air Blank		0.000	23:38
Control Test		0.083	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.173	23:39
Breath Volume		1.312	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.182	23:42
Breath Volume		1.347	23:42
Air Blank		0.000	23:43
Control Test		0.081	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: OCLAIR

DATE/TIME OF TEST: 08/20/2025 00:54:23

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.083	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.226	00:49
Breath Volume		2.363	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.207	00:52
Breath Volume		1.105	00:52
Air Blank		0.000	00:53
Control Test		0.082	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: 2230,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: BEEBE

DATE/TIME OF TEST: 08/23/2025 23:45:17

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.083	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.241	23:37
Breath Volume		2.632	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.262	23:40
Breath Volume		2.761	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 3	0.251	23:43
Breath Volume		2.890	23:43
Air Blank		0.000	23:44
Control Test		0.082	23:44
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 07/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: KNUDSEN

DATE/TIME OF TEST: 08/27/2025 01:33:58

Diagnostics Check		OK	01:27
Air Blank		0.000	01:28
Control Test		0.083	01:28
Air Blank		0.000	01:28
Subject Sample	# 1	0.190	01:29
Breath Volume		3.976	01:29
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.190	01:32
Breath Volume		3.226	01:32
Air Blank		0.000	01:33
Control Test		0.083	01:33
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 08/28/2025 22:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.