

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TORRES,CORINNE,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: RIS,DILLON

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/25/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 20:12

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: URBAN

DATE/TIME OF TEST: 08/30/2025 21:01:40

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.076	20:54
Air Blank	0.000	20:55
Subject Sample # 1	0.297	20:55
Breath Volume	2.082	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.280	20:59
Breath Volume	1.960	20:59
Air Blank	0.000	21:00
Control Test	0.075	21:01
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 09/30/2025 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAYSON,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SMITH,JAYSON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: COXE

DATE/TIME OF TEST: 08/04/2025 22:03:01

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.080	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.210	21:58
Breath Volume	2.015	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:00
Subject Sample # 2	0.207	22:01
Breath Volume	1.789	22:01
Air Blank	0.000	22:02
Control Test	0.079	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:03

Data Download Date/Time: 09/02/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 08/08/2025 02:08:34

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.000	02:04
Breath Volume	2.769	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.000	02:07
Breath Volume	2.171	02:07
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 09/02/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: DI SANTO

DATE/TIME OF TEST: 08/08/2025 21:01:26

Diagnostics Check	OK	20:45	
Air Blank	0.000	20:46	
Control Test	0.079	20:46	
Air Blank	0.000	20:47	
Subject Sample	# 1 0.199	20:50	Volume Not Met
Breath Volume	0.898	20:50	
Air Blank	0.000	20:51	
Air Blank	0.000	20:52	
Subject Sample	# 2 0.219	20:54	
Breath Volume	1.605	20:54	
Air Blank	0.000	20:54	
Air Blank	0.000	20:56	
Subject Sample	# 3 0.237	20:59	Volume Not Met
Breath Volume	0.597	20:59	
Air Blank	0.000	21:00	
Control Test	0.079	21:00	
Air Blank	0.000	21:01	
Diagnostics Check	OK	21:01	

Data Download Date/Time: 09/02/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: DI SANTO

DATE/TIME OF TEST: 08/08/2025 21:17:27

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:03	
Control Test	0.078	21:03	
Air Blank	0.000	21:04	
Subject Sample	# 1 0.208	21:05	No .020 agreement
Breath Volume	1.699	21:05	
Air Blank	0.000	21:06	
Air Blank	0.000	21:08	
Subject Sample	# 2 0.223	21:11	Volume Not Met
Breath Volume	0.820	21:11	
Air Blank	0.000	21:12	
Air Blank	0.000	21:14	
Subject Sample	# 3 0.235	21:15	No .020 agreement
Breath Volume	1.929	21:15	
Air Blank	0.000	21:16	
Control Test	0.078	21:16	
Air Blank	0.000	21:17	
Diagnostics Check	OK	21:17	

Data Download Date/Time: 09/02/2025 22:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GIROUX

DATE/TIME OF TEST: 08/09/2025 01:34:17

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.000	01:29
Breath Volume	2.714	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.000	01:32
Breath Volume	1.773	01:32
Air Blank	0.000	01:33
Control Test	0.078	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 09/02/2025 22:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: HERVAS ESPINOZA

DATE/TIME OF TEST: 08/10/2025 05:01:17

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:46	
Control Test	0.079	04:46	
Air Blank	0.000	04:47	
Subject Sample	# 1 0.144	04:50	Volume Not Met
Breath Volume	0.050	04:50	
Air Blank	0.000	04:50	
Air Blank	0.000	04:52	
Subject Sample	# 2 0.146	04:53	
Breath Volume	2.351	04:53	
Air Blank	0.000	04:54	
Air Blank	0.000	04:56	
Subject Sample	# 3 0.124	04:59	Volume Not Met
Breath Volume	0.246	04:59	
Air Blank	0.000	05:00	
Control Test	0.078	05:00	
Air Blank	0.000	05:01	
Diagnostics Check	OK	05:01	

Data Download Date/Time: 09/02/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: HERVAS ESPINOZA

DATE/TIME OF TEST: 08/10/2025 05:09:13

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.077	05:03
Air Blank	0.000	05:03
Subject Sample # 1	0.140	05:04
Breath Volume	2.140	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.125	05:07
Breath Volume	1.593	05:07
Air Blank	0.000	05:08
Control Test	0.078	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 09/02/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOX, GRAHAM, M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: FRISK, BEN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MCCOWAN

DATE/TIME OF TEST: 08/15/2025 02:57:14

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:54	
Control Test	0.079	02:55	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:56	
Control Test	0.079	02:56	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 09/02/2025 22:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: STARCEVICH

DATE/TIME OF TEST: 08/21/2025 01:45:18

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.079	01:43	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.079	01:44	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 09/02/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOX, GRAHAM, M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: CARRANZA, RAUL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: SYLVIA

DATE/TIME OF TEST: 08/28/2025 01:44:34

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.112	01:40
Breath Volume	3.574	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.108	01:42
Breath Volume	3.007	01:42
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 09/02/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:54

SUBJECT'S LAST NAME: ROMANO

DATE/TIME OF TEST: 08/28/2025 11:24:20

Diagnostics Check	OK	11:17
Air Blank	0.000	11:18
Control Test	0.079	11:18
Air Blank	0.000	11:19
Subject Sample # 1	0.000	11:19
Breath Volume	1.515	11:19
Air Blank	0.000	11:20
Air Blank	0.000	11:22
Subject Sample # 2	0.000	11:22
Breath Volume	1.203	11:22
Air Blank	0.000	11:23
Control Test	0.079	11:23
Air Blank	0.000	11:24
Diagnostics Check	OK	11:24

Data Download Date/Time: 09/02/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,GARY,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: CALDWELL,GARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: STATON

DATE/TIME OF TEST: 08/28/2025 16:52:30

Diagnostics Check	OK	16:45
Air Blank	0.000	16:46
Control Test	0.079	16:46
Air Blank	0.000	16:47
Subject Sample # 1	0.000	16:47
Breath Volume	2.027	16:47
Air Blank	0.000	16:48
Air Blank	0.000	16:50
Subject Sample # 2	0.000	16:51
Breath Volume	1.953	16:51
Air Blank	0.000	16:51
Control Test	0.079	16:51
Air Blank	0.000	16:52
Diagnostics Check	OK	16:52

Data Download Date/Time: 09/02/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 08/30/2025 02:15:38

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.118	02:11
Breath Volume	1.730	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.120	02:14
Breath Volume	1.644	02:14
Air Blank	0.000	02:14
Control Test	0.079	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 09/02/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ERVIN

DATE/TIME OF TEST: 08/31/2025 04:09:11

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.077	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.153	04:04
Breath Volume	2.441	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.148	04:07
Breath Volume	2.011	04:07
Air Blank	0.000	04:08
Control Test	0.077	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 09/02/2025 22:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CCSO K,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GAINVORS,CCSO
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: CIFUENTES-ESTRADA

DATE/TIME OF TEST: 08/23/2025 00:17:27

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.079	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.220	00:12
Breath Volume	1.125	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.211	00:15
Breath Volume	1.148	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 09/17/2025 22:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BEAUDETTE,K
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: ZIMNIE

DATE/TIME OF TEST: 08/23/2025 01:18:39

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:08	
Air Blank	0.000	01:08	
Subject Sample	# 1 0.072	01:11	Volume Not Met
Breath Volume	0.125	01:11	
Air Blank	0.000	01:12	
Air Blank	0.000	01:14	
Subject Sample	# 2	01:17	No Sample Provided
Breath Volume	0.000	01:17	
Air Blank	0.000	01:17	
Control Test	0.080	01:18	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 09/17/2025 22:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN, CLARENCE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FERNANDEZ, DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PETTENGILL

DATE/TIME OF TEST: 08/24/2025 02:37:48

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.110	02:33
Breath Volume	2.976	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.107	02:36
Breath Volume	2.058	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 09/17/2025 22:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KUBA,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 08/24/2025 02:58:16

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.066	02:53
Breath Volume	1.199	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.071	02:56
Breath Volume	1.421	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 09/17/2025 22:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WENTZ,T
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: CUTRONI

DATE/TIME OF TEST: 08/25/2025 00:58:47

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.191	00:54
Breath Volume	3.703	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.180	00:57
Breath Volume	3.300	00:57
Air Blank	0.000	00:57
Control Test	0.080	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 09/17/2025 22:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CARARO,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: CORREIA

DATE/TIME OF TEST: 08/26/2025 21:11:29

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.081	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.114	21:06
Breath Volume	4.515	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.108	21:10
Breath Volume	4.589	21:10
Air Blank	0.000	21:10
Control Test	0.081	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 09/17/2025 22:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WEIHS,K
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 08/30/2025 00:14:46

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.081	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.215	00:10
Breath Volume	1.105	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.212	00:13
Breath Volume	2.406	00:13
Air Blank	0.000	00:13
Control Test	0.081	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 09/17/2025 22:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: LIVINGTON

DATE/TIME OF TEST: 08/05/2025 03:46:51

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.082	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.175	03:41
Breath Volume	1.609	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.165	03:45
Breath Volume	1.917	03:45
Air Blank	0.000	03:45
Control Test	0.080	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 09/10/2025 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: DUFFY

DATE/TIME OF TEST: 08/07/2025 00:25:06

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.080	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.000	00:20
Breath Volume	1.125	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.000	00:23
Breath Volume	1.304	00:23
Air Blank	0.000	00:24
Control Test	0.081	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 09/10/2025 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DEUSEN,COLE,C
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VAN DEUSEN,COLE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: GUERRA FIGUEROA

DATE/TIME OF TEST: 08/09/2025 20:07:25

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.081	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.134	20:02
Breath Volume	2.117	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.129	20:05
Breath Volume	1.777	20:05
Air Blank	0.000	20:06
Control Test	0.079	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 09/10/2025 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASCKO,J,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LASCKO,J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:12

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/10/2025 14:40:05

Diagnostics Check	OK	14:34
Air Blank	0.000	14:34
Control Test	0.083	14:34
Air Blank	0.000	14:35
Subject Sample # 1	0.000	14:35
Breath Volume	2.027	14:35
Air Blank	0.000	14:36
Air Blank	0.000	14:38
Subject Sample # 2	0.000	14:38
Breath Volume	1.179	14:38
Air Blank	0.000	14:39
Control Test	0.080	14:39
Air Blank	0.000	14:40
Diagnostics Check	OK	14:40

Data Download Date/Time: 09/10/2025 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DEUSEN,COLE,C
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VAN DEUSEN,COLE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 08/10/2025 23:32:50

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.081	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.104	23:28
Breath Volume	4.406	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.095	23:31
Breath Volume	3.460	23:31
Air Blank	0.000	23:31
Control Test	0.076	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 09/10/2025 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELATORRE-CASTELLANO,OFC,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: DELATORRE-
CASTELLANO,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:57
GENDER: Female

OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: PLUTNICKI

DATE/TIME OF TEST: 08/11/2025 15:48:53

Diagnostics Check	OK	15:35	
Air Blank	0.000	15:36	
Control Test	0.080	15:36	
Air Blank	0.000	15:37	
Subject Sample	# 1 0.207	15:40	Volume Not Met
Breath Volume	0.203	15:40	
Air Blank	0.000	15:41	
Air Blank	0.000	15:42	
Subject Sample	# 2 0.228	15:43	
Breath Volume	1.226	15:43	
Air Blank	0.000	15:44	
Air Blank	0.000	15:46	
Subject Sample	# 3 0.227	15:47	
Breath Volume	1.468	15:47	
Air Blank	0.000	15:47	
Control Test	0.080	15:48	
Air Blank	0.000	15:48	
Diagnostics Check	OK	15:48	

Data Download Date/Time: 09/10/2025 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELCHNER,M,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KELCHNER,M
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 08/13/2025 12:13:40

Diagnostics Check	OK	12:06
Air Blank	0.000	12:06
Control Test	0.081	12:07
Air Blank	0.000	12:07
Subject Sample # 1	0.000	12:09
Breath Volume	1.136	12:09
Air Blank	0.000	12:09
Air Blank	0.000	12:11
Subject Sample # 2	0.000	12:12
Breath Volume	1.296	12:12
Air Blank	0.000	12:12
Control Test	0.081	12:13
Air Blank	0.000	12:13
Diagnostics Check	OK	12:13

Data Download Date/Time: 09/10/2025 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHANDLER,RYAN,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: TIMBROOK,RICKY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: GOMEZ JR

DATE/TIME OF TEST: 08/15/2025 02:33:24

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.203	02:28
Breath Volume	2.335	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.203	02:31
Breath Volume	2.222	02:31
Air Blank	0.000	02:32
Control Test	0.078	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 09/10/2025 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: LOPEZ PEREZ

DATE/TIME OF TEST: 08/13/2025 23:46:46

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.081	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.122	23:42
Breath Volume	2.046	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.124	23:45
Breath Volume	2.066	23:45
Air Blank	0.000	23:45
Control Test	0.081	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 08/15/2025 15:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 08/03/2025 20:44:07

Diagnostics Check	OK	20:37
Air Blank	0.000	20:38
Control Test	0.078	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.123	20:39
Breath Volume	2.500	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.120	20:42
Breath Volume	2.605	20:42
Air Blank	0.000	20:42
Control Test	0.078	20:43
Air Blank	0.000	20:43
Diagnostics Check	OK	20:44

Data Download Date/Time: 08/14/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HOLLEY

DATE/TIME OF TEST: 08/07/2025 03:11:56

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.086	03:07
Breath Volume	1.343	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.085	03:10
Breath Volume	2.625	03:10
Air Blank	0.000	03:10
Control Test	0.078	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 08/14/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,D,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 08/02/2025 02:46:58

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:39	
Control Test	0.081	02:40	
Air Blank	0.000	02:40	
Subject Sample # 1	0.121	02:41	
Breath Volume	1.582	02:41	
Air Blank	0.000	02:41	
Air Blank	0.000	02:43	
Subject Sample # 2	0.110	02:46	
Breath Volume	1.562	02:46	
Air Blank		02:46	RFI Detect
Air Blank	0.000	02:46	

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,D,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 08/02/2025 03:00:11

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.101	02:55
Breath Volume	1.210	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.117	02:58
Breath Volume	2.546	02:58
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARRAR,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 08/03/2025 08:22:15

Diagnostics Check	OK	08:15
Air Blank	0.000	08:15
Control Test	0.081	08:16
Air Blank	0.000	08:16
Subject Sample # 1	0.178	08:17
Breath Volume	2.097	08:17
Air Blank	0.000	08:18
Air Blank	0.000	08:20
Subject Sample # 2	0.172	08:20
Breath Volume	1.785	08:20
Air Blank	0.000	08:21
Control Test	0.079	08:21
Air Blank	0.000	08:22
Diagnostics Check	OK	08:22

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 08/04/2025 00:06:19

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.008	00:01
Breath Volume	2.484	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.009	00:04
Breath Volume	2.582	00:04
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CRUMADY

DATE/TIME OF TEST: 08/07/2025 02:27:14

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.118	02:22
Breath Volume	1.425	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.117	02:25
Breath Volume	1.433	02:25
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BATEH,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 08/07/2025 03:38:36

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.142	03:34
Breath Volume	2.265	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.134	03:37
Breath Volume	1.906	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/08/2025 01:01:17

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.080	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.157	00:55
Breath Volume	1.175	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:57
Subject Sample # 2	0.162	00:59
Breath Volume	1.429	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: BARWICK

DATE/TIME OF TEST: 08/08/2025 20:14:45

Diagnostics Check	OK	20:07
Air Blank	0.000	20:08
Control Test	0.080	20:08
Air Blank	0.000	20:08
Subject Sample # 1	0.177	20:10
Breath Volume	1.914	20:10
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.175	20:13
Breath Volume	1.339	20:13
Air Blank	0.000	20:13
Control Test	0.080	20:14
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: TILBURY

DATE/TIME OF TEST: 08/09/2025 04:18:48

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.170	04:14
Breath Volume	1.847	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.168	04:17
Breath Volume	1.710	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 08/10/2025 02:31:40

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.159	02:26
Breath Volume	2.027	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.164	02:30
Breath Volume	1.968	02:30
Air Blank	0.000	02:30
Control Test	0.081	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: GREATHOUSE

DATE/TIME OF TEST: 08/10/2025 04:57:12

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.145	04:52
Breath Volume	1.882	04:52
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.149	04:55
Breath Volume	1.363	04:55
Air Blank	0.000	04:56
Control Test	0.081	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: SHIFLETT

DATE/TIME OF TEST: 08/10/2025 16:38:38

Diagnostics Check	OK	16:35	
Air Blank	0.000	16:36	
Control Test	0.080	16:36	
Air Blank	0.000	16:37	
Subject Sample	# 1	16:37	Subject Test Refused
Breath Volume	0.000	16:37	
Air Blank	0.000	16:37	
Control Test	0.081	16:38	
Air Blank	0.000	16:38	
Diagnostics Check	OK	16:38	

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: STRASSENBERG

DATE/TIME OF TEST: 08/10/2025 17:09:42

Diagnostics Check	OK	17:02
Air Blank	0.000	17:03
Control Test	0.081	17:03
Air Blank	0.000	17:04
Subject Sample # 1	0.285	17:05
Breath Volume	1.542	17:05
Air Blank	0.000	17:05
Air Blank	0.000	17:07
Subject Sample # 2	0.280	17:08
Breath Volume	1.148	17:08
Air Blank	0.000	17:08
Control Test	0.080	17:09
Air Blank	0.000	17:09
Diagnostics Check	OK	17:09

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/10/2025 22:39:24

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.081	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.042	22:34
Breath Volume	1.722	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.041	22:37
Breath Volume	1.250	22:37
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DENSON

DATE/TIME OF TEST: 08/12/2025 02:55:49

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.081	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.029	02:48
Breath Volume	1.277	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.028	02:54
Breath Volume	1.562	02:54
Air Blank	0.000	02:54
Control Test	0.081	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ELLISON,H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: GELINAS

DATE/TIME OF TEST: 08/14/2025 05:02:27

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:57	
Control Test	0.081	04:58	
Air Blank	0.000	04:58	
Subject Sample	# 1	05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:01	
Control Test	0.081	05:01	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: ROGAN

DATE/TIME OF TEST: 08/15/2025 00:02:41

Diagnostics Check	OK	23:59	
Air Blank	0.000	00:00	
Control Test	0.081	00:00	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:01	Subject Test Refused
Breath Volume	0.000	00:01	
Air Blank	0.000	00:01	
Control Test	0.080	00:02	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KLING,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 08/15/2025 04:34:00

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.079	04:20	
Air Blank	0.000	04:20	
Subject Sample	# 1 0.125	04:23	Volume Not Met
Breath Volume	0.750	04:23	
Air Blank	0.000	04:24	
Air Blank	0.000	04:26	
Subject Sample	# 2 0.138	04:26	
Breath Volume	1.773	04:26	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample	# 3 0.133	04:32	Volume Not Met
Breath Volume	0.195	04:32	
Air Blank	0.000	04:33	
Control Test	0.081	04:33	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:34	

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KLING,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 08/15/2025 05:02:54

Diagnostics Check	OK	04:59	
Air Blank	0.000	05:00	
Control Test	0.080	05:00	
Air Blank	0.000	05:01	
Subject Sample	# 1	05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:01	
Control Test	0.080	05:02	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: BRITTON

DATE/TIME OF TEST: 08/15/2025 20:16:13

Diagnostics Check	OK	20:13	
Air Blank	0.000	20:13	
Control Test	0.081	20:14	
Air Blank	0.000	20:14	
Subject Sample	# 1	20:14	Subject Test Refused
Breath Volume	0.000	20:14	
Air Blank	0.000	20:15	
Control Test	0.080	20:15	
Air Blank	0.000	20:16	
Diagnostics Check	OK	20:16	

Data Download Date/Time: 08/20/2025 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GLISSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: LAUFENBERG

DATE/TIME OF TEST: 08/16/2025 02:54:10

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.081	02:52	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.080	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 08/20/2025 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: GODIN

DATE/TIME OF TEST: 08/16/2025 04:15:22

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.218	04:10
Breath Volume	1.667	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.222	04:13
Breath Volume	1.871	04:13
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 08/20/2025 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 08/16/2025 05:07:46

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:05	
Control Test	0.081	05:05	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:06	Subject Test Refused
Breath Volume	0.000	05:06	
Air Blank	0.000	05:06	
Control Test	0.081	05:07	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 08/20/2025 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 08/17/2025 02:11:16

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.212	02:06
Breath Volume	1.332	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.222	02:09
Breath Volume	1.714	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 08/20/2025 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: MARTIN,T
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 08/17/2025 05:22:39

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.081	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.174	05:18
Breath Volume	1.132	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.177	05:21
Breath Volume	1.128	05:21
Air Blank	0.000	05:21
Control Test	0.080	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 08/20/2025 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: TORKY

DATE/TIME OF TEST: 08/17/2025 22:06:09

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:03	
Control Test	0.081	22:04	
Air Blank	0.000	22:04	
Subject Sample	# 1	22:04	Subject Test Refused
Breath Volume	0.000	22:04	
Air Blank	0.000	22:05	
Control Test	0.081	22:05	
Air Blank	0.000	22:06	
Diagnostics Check	OK	22:06	

Data Download Date/Time: 08/20/2025 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 08/18/2025 01:24:38

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:22	
Control Test	0.081	01:22	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:23	
Control Test	0.082	01:24	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 08/20/2025 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 08/18/2025 02:53:22

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.081	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:52	
Control Test	0.081	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 08/20/2025 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEJEDA,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: TEJEDA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 08/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: URUTIA-MAYORGA

DATE/TIME OF TEST: 08/07/2025 01:48:30

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.158	01:43
Breath Volume	2.140	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.144	01:46
Breath Volume	1.304	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 09/04/2025 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVA,J,C
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: OLIVA,JUAN
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 08/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 08/09/2025 19:23:52

Diagnostics Check	OK	19:17
Air Blank	0.000	19:18
Control Test	0.079	19:18
Air Blank	0.000	19:19
Subject Sample # 1	0.166	19:19
Breath Volume	2.046	19:19
Air Blank	0.000	19:20
Air Blank	0.000	19:21
Subject Sample # 2	0.177	19:22
Breath Volume	2.132	19:22
Air Blank	0.000	19:22
Control Test	0.079	19:23
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 09/04/2025 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEJEDA,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: DOMINGUEZ,CARLOS
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 08/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: CARDENAS-GARRIDO

DATE/TIME OF TEST: 08/17/2025 02:25:22

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:22	
Control Test	0.079	02:23	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:23	Subject Test Refused
Breath Volume	0.000	02:23	
Air Blank	0.000	02:24	
Control Test	0.079	02:24	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 09/04/2025 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVA,JUAN,C
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: OLIVA,JUAN
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 08/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: IGLESIAS

DATE/TIME OF TEST: 08/23/2025 20:31:56

Diagnostics Check	OK	20:29	
Air Blank	0.000	20:29	
Control Test	0.080	20:29	
Air Blank	0.000	20:30	
Subject Sample	# 1	20:30	Subject Test Refused
Breath Volume	0.000	20:30	
Air Blank	0.000	20:31	
Control Test	0.079	20:31	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 09/04/2025 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RASHAUN,K
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: CAMPBELL,KEANDRA
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 07/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: SIMEON

DATE/TIME OF TEST: 08/03/2025 05:34:34

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:30	
Control Test	0.080	05:30	
Air Blank	0.000	05:31	
Subject Sample	# 1	05:33	Subject Test Refused
Breath Volume	0.000	05:33	
Air Blank	0.000	05:33	
Control Test	0.079	05:33	
Air Blank	0.000	05:34	
Diagnostics Check	OK	05:34	

Data Download Date/Time: 08/13/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL, EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: STIMPSON

DATE/TIME OF TEST: 08/02/2025 03:08:45

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.177	03:04
Breath Volume	3.222	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.177	03:07
Breath Volume	3.351	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 08/20/2025 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: CORTEZ

DATE/TIME OF TEST: 08/03/2025 03:50:50

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.230	03:46
Breath Volume	1.617	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:48
Subject Sample # 2	0.230	03:49
Breath Volume	1.402	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 08/20/2025 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 08/09/2025 02:43:34

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.135	02:39
Breath Volume	1.324	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.129	02:42
Breath Volume	1.457	02:42
Air Blank	0.000	02:42
Control Test	0.077	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 08/20/2025 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL, EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MATHEWS

DATE/TIME OF TEST: 08/09/2025 23:50:29

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.077	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.328	23:45
Breath Volume	2.082	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.330	23:48
Breath Volume	1.265	23:48
Air Blank	0.000	23:49
Control Test	0.078	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 08/20/2025 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 08/13/2025 17:08:15

Diagnostics Check	OK	17:02
Air Blank	0.000	17:02
Control Test	0.080	17:02
Air Blank	0.000	17:03
Subject Sample # 1	0.326	17:03
Breath Volume	2.160	17:03
Air Blank	0.000	17:04
Air Blank	0.000	17:06
Subject Sample # 2	0.341	17:06
Breath Volume	1.792	17:06
Air Blank	0.000	17:07
Control Test	0.079	17:07
Air Blank	0.000	17:08
Diagnostics Check	OK	17:08

Data Download Date/Time: 08/20/2025 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL, EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 08/15/2025 02:57:25

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.236	02:51
Breath Volume	1.429	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample # 2	0.241	02:55
Breath Volume	1.488	02:55
Air Blank	0.000	02:56
Control Test	0.078	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 08/20/2025 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/31/2025 06:43:51

Diagnostics Check	OK	06:37
Air Blank	0.000	06:37
Control Test	0.077	06:38
Air Blank	0.000	06:38
Subject Sample # 1	0.149	06:39
Breath Volume	2.531	06:39
Air Blank	0.000	06:39
Air Blank	0.000	06:41
Subject Sample # 2	0.140	06:42
Breath Volume	1.601	06:42
Air Blank	0.000	06:42
Control Test	0.078	06:43
Air Blank	0.000	06:43
Diagnostics Check	OK	06:43

Data Download Date/Time: 09/09/2025 17:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CANALES,ANGEL,L

BREATH TEST OPERATOR AGENCY: Indian River Co SO

ARRESTING OFFICER: HACKETT,BRADLEY

ARRESTING OFFICER AGENCY: Indian River Co SO

LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG500903

DRY GAS EXPIRY DATE: 01/09/2027

VIOLATION CODE: DUI

ARREST TIME: 02:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 08/01/2025 03:48:00

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.080	03:46	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 08/24/2025 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HISS,J,M

BREATH TEST OPERATOR AGENCY: Indian River Co SO

ARRESTING OFFICER: MARTINEZ,J

ARRESTING OFFICER AGENCY: Indian River Co SO

LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG500903

DRY GAS EXPIRY DATE: 01/09/2027

VIOLATION CODE: DUI

ARREST TIME: 19:23

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: RIGGIO

DATE/TIME OF TEST: 08/02/2025 01:30:57

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.130	01:26
Breath Volume	1.476	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.131	01:29
Breath Volume	1.617	01:29
Air Blank	0.000	01:30
Control Test	0.078	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 08/24/2025 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NEIL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CAIRNS

DATE/TIME OF TEST: 08/02/2025 03:52:31

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.080	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Control Test	0.079	03:51	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 08/24/2025 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PORTOCARRERO,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: KIRCHMANN

DATE/TIME OF TEST: 08/02/2025 19:52:44

Diagnostics Check	OK	19:46
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:47
Subject Sample # 1	0.000	19:48
Breath Volume	2.484	19:48
Air Blank	0.000	19:48
Air Blank	0.000	19:50
Subject Sample # 2	0.000	19:51
Breath Volume	2.496	19:51
Air Blank	0.000	19:51
Control Test	0.079	19:52
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 08/24/2025 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HISS,J,M

BREATH TEST OPERATOR AGENCY: Indian River Co SO

ARRESTING OFFICER: MAGILL,D

ARRESTING OFFICER AGENCY: Indian River Co SO

LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG500903

DRY GAS EXPIRY DATE: 01/09/2027

VIOLATION CODE: DUI

ARREST TIME: 02:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: GONZALEZ DUARTE

DATE/TIME OF TEST: 08/03/2025 04:46:07

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.143	04:41
Breath Volume	3.453	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.125	04:44
Breath Volume	1.597	04:44
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 08/24/2025 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KANE,JAMES,G

BREATH TEST OPERATOR AGENCY: Indian River Co SO

ARRESTING OFFICER: MARTINEZ,J

ARRESTING OFFICER AGENCY: Indian River Co SO

LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG500903

DRY GAS EXPIRY DATE: 01/09/2027

VIOLATION CODE: DUI

ARREST TIME: 21:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: REYES-SANCHEZ

DATE/TIME OF TEST: 08/03/2025 06:37:34

Diagnostics Check	OK	06:30
Air Blank	0.000	06:31
Control Test	0.080	06:31
Air Blank	0.000	06:32
Subject Sample # 1	0.061	06:32
Breath Volume	1.660	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:35
Subject Sample # 2	0.060	06:36
Breath Volume	1.441	06:36
Air Blank	0.000	06:36
Control Test	0.079	06:36
Air Blank	0.000	06:37
Diagnostics Check	OK	06:37

Data Download Date/Time: 08/24/2025 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALES,ANGEL,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MAGILL,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 08/06/2025 03:33:51

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.000	03:29
Breath Volume	1.738	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.000	03:32
Breath Volume	1.578	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 08/24/2025 12:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HISS,J,M

BREATH TEST OPERATOR AGENCY: Indian River Co SO

ARRESTING OFFICER: NEILL,M

ARRESTING OFFICER AGENCY: Indian River Co SO

LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: AG500903

DRY GAS EXPIRY DATE: 01/09/2027

VIOLATION CODE: DUI

ARREST TIME: 01:11

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MORALES SIEZAR

DATE/TIME OF TEST: 08/07/2025 02:13:54

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.081	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.174	02:09
Breath Volume	2.285	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.175	02:12
Breath Volume	2.847	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 08/24/2025 12:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HISS,J,M

BREATH TEST OPERATOR AGENCY: Indian River Co SO

ARRESTING OFFICER: SCOTT,C

ARRESTING OFFICER AGENCY: Indian River Co SO

LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG500903

DRY GAS EXPIRY DATE: 01/09/2027

VIOLATION CODE: DUI

ARREST TIME: 22:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: LAPORTE

DATE/TIME OF TEST: 08/07/2025 23:42:45

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:40	
Control Test	0.080	23:40	
Air Blank	0.000	23:41	
Subject Sample	# 1	23:41	Subject Test Refused
Breath Volume	0.000	23:41	
Air Blank	0.000	23:41	
Control Test	0.080	23:42	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 08/24/2025 12:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HISS,J,M

BREATH TEST OPERATOR AGENCY: Indian River Co SO

ARRESTING OFFICER: MARTINEZ,J

ARRESTING OFFICER AGENCY: Indian River Co SO

LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG500903

DRY GAS EXPIRY DATE: 01/09/2027

VIOLATION CODE: DUI

ARREST TIME: 23:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/08/2025 00:34:39

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:32	
Control Test	0.080	00:32	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:33	
Control Test	0.080	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 08/24/2025 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,BRITTANY,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CHOUNARD,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 12:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 08/08/2025 13:41:30

Diagnostics Check	OK	13:35
Air Blank	0.000	13:35
Control Test	0.080	13:35
Air Blank	0.000	13:36
Subject Sample # 1	0.191	13:36
Breath Volume	1.121	13:36
Air Blank	0.000	13:37
Air Blank	0.000	13:39
Subject Sample # 2	0.201	13:40
Breath Volume	1.957	13:40
Air Blank	0.000	13:40
Control Test	0.080	13:40
Air Blank	0.000	13:41
Diagnostics Check	OK	13:41

Data Download Date/Time: 08/24/2025 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: VENEGAS, S
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BOOTH

DATE/TIME OF TEST: 08/08/2025 21:39:55

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.080	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.061	21:35
Breath Volume	3.480	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.060	21:38
Breath Volume	3.324	21:38
Air Blank	0.000	21:39
Control Test	0.079	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 08/24/2025 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WICKERT,PAUL,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: WICKERT,PAUL
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: FONG

DATE/TIME OF TEST: 08/09/2025 00:08:00

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.103	00:03
Breath Volume	1.382	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.101	00:06
Breath Volume	1.843	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:08

Data Download Date/Time: 08/24/2025 12:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GRANDY,KEVIN,

BREATH TEST OPERATOR AGENCY: Indian River Co SO

ARRESTING OFFICER: GLADING,N

ARRESTING OFFICER AGENCY: Indian River Co SO

LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG500903

DRY GAS EXPIRY DATE: 01/09/2027

VIOLATION CODE: DUI

ARREST TIME: 15:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:52

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 08/09/2025 16:25:26

Diagnostics Check	OK	16:19
Air Blank	0.000	16:19
Control Test	0.081	16:20
Air Blank	0.000	16:20
Subject Sample # 1	0.000	16:21
Breath Volume	2.578	16:21
Air Blank	0.000	16:21
Air Blank	0.000	16:23
Subject Sample # 2	0.000	16:24
Breath Volume	2.843	16:24
Air Blank	0.000	16:24
Control Test	0.080	16:24
Air Blank	0.000	16:25
Diagnostics Check	OK	16:25

Data Download Date/Time: 08/24/2025 12:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SHUREN,T,T

BREATH TEST OPERATOR AGENCY: Indian River Co SO

ARRESTING OFFICER: HACKET,B

ARRESTING OFFICER AGENCY: Indian River Co SO

LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG500903

DRY GAS EXPIRY DATE: 01/09/2027

VIOLATION CODE: DUI

ARREST TIME: 22:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: BROGNANO

DATE/TIME OF TEST: 08/11/2025 23:29:42

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.081	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:28	
Control Test	0.080	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 08/24/2025 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUREN,T,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HACKETT,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: ROMEO

DATE/TIME OF TEST: 08/12/2025 23:48:11

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:45	
Control Test	0.080	23:46	
Air Blank	0.000	23:46	
Subject Sample	# 1	23:46	Subject Test Refused
Breath Volume	0.000	23:46	
Air Blank	0.000	23:47	
Control Test	0.080	23:47	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 08/24/2025 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUREN,T,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HACKETT,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 08/13/2025 04:17:17

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.079	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:10	No Sample Provided
Breath Volume	0.000	04:10	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample	# 2	04:15	No Sample Provided
Breath Volume	0.000	04:15	
Air Blank	0.000	04:16	
Control Test	0.080	04:16	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

Data Download Date/Time: 08/24/2025 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GRANDY,KEVIN,

BREATH TEST OPERATOR AGENCY: Indian River Co SO

ARRESTING OFFICER: BARKWELL,J

ARRESTING OFFICER AGENCY: Indian River Co SO

LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG500903

DRY GAS EXPIRY DATE: 01/09/2027

VIOLATION CODE: DUI

ARREST TIME: 08:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 08/14/2025 09:23:43

Diagnostics Check	OK	09:17
Air Blank	0.000	09:18
Control Test	0.080	09:18
Air Blank	0.000	09:19
Subject Sample # 1	0.000	09:19
Breath Volume	1.929	09:19
Air Blank	0.000	09:19
Air Blank	0.000	09:21
Subject Sample # 2	0.000	09:22
Breath Volume	2.132	09:22
Air Blank	0.000	09:22
Control Test	0.080	09:23
Air Blank	0.000	09:23
Diagnostics Check	OK	09:23

Data Download Date/Time: 08/24/2025 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CANALES,ANGEL,L

BREATH TEST OPERATOR AGENCY: Indian River Co SO

ARRESTING OFFICER: NEIL,M

ARRESTING OFFICER AGENCY: Indian River Co SO

LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG500903

DRY GAS EXPIRY DATE: 01/09/2027

VIOLATION CODE: DUI

ARREST TIME: 18:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 08/14/2025 20:12:11

Diagnostics Check	OK	20:00	
Air Blank	0.000	20:00	
Control Test	0.080	20:00	
Air Blank	0.000	20:01	
Subject Sample	# 1 0.090	20:02	
Breath Volume	3.734	20:02	
Air Blank	0.000	20:03	
Air Blank	0.000	20:04	
Subject Sample	# 2	20:08	No Sample Provided
Breath Volume	0.000	20:08	
Air Blank	0.000	20:08	
Air Blank	0.000	20:10	
Subject Sample	# 3	20:10	Subject Test Refused
Breath Volume	0.000	20:10	
Air Blank	0.000	20:11	
Control Test	0.080	20:11	
Air Blank	0.000	20:12	
Diagnostics Check	OK	20:12	

Data Download Date/Time: 08/24/2025 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALES,ANGEL,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NEILL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 08/14/2025 20:24:39

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:16	
Control Test	0.080	20:16	
Air Blank	0.000	20:16	
Subject Sample	# 1 0.084	20:18	
Breath Volume	2.792	20:18	
Air Blank	0.000	20:19	
Air Blank	0.000	20:21	
Subject Sample	# 2	20:23	Subject Test Refused
Breath Volume	0.000	20:23	
Air Blank	0.000	20:23	
Control Test	0.079	20:24	
Air Blank	0.000	20:24	
Diagnostics Check	OK	20:24	

Data Download Date/Time: 08/24/2025 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUREN,T,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MARTIN,Q
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: LIVELY

DATE/TIME OF TEST: 08/15/2025 22:35:35

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:33	
Control Test	0.080	22:33	
Air Blank	0.000	22:33	
Subject Sample	# 1	22:34	Subject Test Refused
Breath Volume	0.000	22:34	
Air Blank	0.000	22:34	
Control Test	0.080	22:35	
Air Blank	0.000	22:35	
Diagnostics Check	OK	22:35	

Data Download Date/Time: 08/24/2025 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BEAUMONT,C
ARRESTING OFFICER AGENCY: Indian River Shores
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 08/17/2025 03:31:41

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.158	03:26
Breath Volume	1.824	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.153	03:30
Breath Volume	2.070	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 08/24/2025 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RUDOLPH,B,D

BREATH TEST OPERATOR AGENCY: Indian River Co SO

ARRESTING OFFICER: MOON,C

ARRESTING OFFICER AGENCY: Indian River Co SO

LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG500903

DRY GAS EXPIRY DATE: 01/09/2027

VIOLATION CODE: DUI

ARREST TIME: 15:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: ROSEWARNE

DATE/TIME OF TEST: 08/18/2025 16:50:56

Diagnostics Check	OK	16:48	
Air Blank	0.000	16:48	
Control Test	0.081	16:48	
Air Blank	0.000	16:49	
Subject Sample	# 1	16:49	Subject Test Refused
Breath Volume	0.000	16:49	
Air Blank	0.000	16:50	
Control Test	0.081	16:50	
Air Blank	0.000	16:50	
Diagnostics Check	OK	16:50	

Data Download Date/Time: 08/24/2025 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALES,ANGEL,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KRAMER,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 08/18/2025 19:17:13

Diagnostics Check	OK	19:10
Air Blank	0.000	19:11
Control Test	0.081	19:11
Air Blank	0.000	19:12
Subject Sample # 1	0.102	19:12
Breath Volume	1.746	19:12
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.108	19:15
Breath Volume	1.851	19:15
Air Blank	0.000	19:16
Control Test	0.081	19:16
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 08/24/2025 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALES,ANGEL,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KRAMER,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: METELLUS

DATE/TIME OF TEST: 08/19/2025 22:12:16

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.080	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.000	22:07
Breath Volume	2.046	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.000	22:10
Breath Volume	3.386	22:10
Air Blank	0.000	22:11
Control Test	0.080	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 08/24/2025 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CONSALO, A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 08/19/2025 22:54:03

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.081	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.000	22:48
Breath Volume	1.863	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.000	22:52
Breath Volume	1.613	22:52
Air Blank	0.000	22:53
Control Test	0.081	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:54

Data Download Date/Time: 08/24/2025 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HACKETT,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: KUBSCH

DATE/TIME OF TEST: 08/22/2025 03:56:12

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.160	03:51
Breath Volume	2.402	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.143	03:54
Breath Volume	1.265	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 08/24/2025 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAZARES,MARGARITA,C

BREATH TEST OPERATOR AGENCY: Indian River Co SO

ARRESTING OFFICER: BARRIENTOS,S

ARRESTING OFFICER AGENCY: Indian River Co SO

LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG500903

DRY GAS EXPIRY DATE: 01/09/2027

VIOLATION CODE: DUI

ARREST TIME: 02:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HAYDEN

DATE/TIME OF TEST: 08/23/2025 03:42:03

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Subject Sample	# 1 0.167	03:37
Breath Volume	2.074	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample	# 2 0.163	03:40
Breath Volume	2.023	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:42

Data Download Date/Time: 08/24/2025 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALES,ANGEL,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: POLLACEK,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 08/23/2025 23:19:20

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:16	
Control Test	0.080	23:17	
Air Blank	0.000	23:17	
Subject Sample	# 1	23:17	Subject Test Refused
Breath Volume	0.000	23:17	
Air Blank	0.000	23:18	
Control Test	0.079	23:18	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 08/24/2025 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HERNANDEZ, SOLEDAD,

BREATH TEST OPERATOR AGENCY: Vero Beach PD

ARRESTING OFFICER: HERNANDEZ, SOLEDAD

ARRESTING OFFICER AGENCY: Vero Beach PD

LAST AGENCY INSPECTION: 07/11/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG500903

DRY GAS EXPIRY DATE: 01/09/2027

VIOLATION CODE: DUI

ARREST TIME: 01:47

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: DUNBAR

DATE/TIME OF TEST: 08/24/2025 02:30:28

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:28	
Control Test	0.080	02:28	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Control Test	0.079	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 08/24/2025 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCLJOSE,F
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CUAREZMA,MICHAEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: PADRO

DATE/TIME OF TEST: 08/15/2025 22:33:24

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.079	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.135	22:28
Breath Volume	2.886	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.132	22:31
Breath Volume	1.636	22:31
Air Blank	0.000	22:32
Control Test	0.078	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 09/18/2025 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BICKEL III

DATE/TIME OF TEST: 08/30/2025 03:17:58

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.209	03:13
Breath Volume	3.019	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample # 2	0.199	03:16
Breath Volume	1.949	03:16
Air Blank	0.000	03:17
Control Test	0.079	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 09/21/2025 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVER,ISAIAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SILVER,ISAIAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: FELDER

DATE/TIME OF TEST: 08/22/2025 22:40:12

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:35	
Control Test	0.081	22:35	
Air Blank	0.000	22:36	
Subject Sample # 1	0.000	22:36	
Breath Volume	1.433	22:36	
Air Blank	0.000	22:37	
Air Blank		22:38	RFI Detect
Air Blank	0.000	22:39	
Control Test	0.080	22:39	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 09/11/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVER,ISAIAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SILVER,ISAIAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: FELDER

DATE/TIME OF TEST: 08/22/2025 22:48:39

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.081	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.000	22:44
Breath Volume	2.664	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.000	22:47
Breath Volume	1.765	22:47
Air Blank	0.000	22:47
Control Test	0.081	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 09/11/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 08/30/2025 00:01:01

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.081	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.150	23:56
Breath Volume	1.343	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.147	23:59
Breath Volume	1.695	23:59
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:01

Data Download Date/Time: 09/26/2025 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,RITCHIE,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: SANTIAGO,EDGARDO
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250284
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: MENDEZ HERNANDEZ

DATE/TIME OF TEST: 08/02/2025 02:21:51

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.078	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.197	02:17
Breath Volume	3.214	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.194	02:20
Breath Volume	1.789	02:20
Air Blank	0.000	02:20
Control Test	0.078	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 08/08/2025 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,RITCHIE,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: OROURKE,ROCKY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250284
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/02/2025 03:37:03

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.135	03:32
Breath Volume	2.636	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.135	03:35
Breath Volume	2.347	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:37

Data Download Date/Time: 08/08/2025 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,RITCHIE,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: O'ROURKE,ROCKY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250284
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: BERTUBIN

DATE/TIME OF TEST: 08/29/2025 22:21:20

Diagnostics Check	OK	22:04	
Air Blank	0.000	22:05	
Control Test	0.080	22:05	
Air Blank	0.000	22:06	
Subject Sample	# 1 0.227	22:09	Volume Not Met
Breath Volume	0.183	22:09	
Air Blank	0.000	22:09	
Air Blank	0.000	22:11	
Subject Sample	# 2 0.259	22:14	
Breath Volume	1.707	22:14	
Air Blank	0.000	22:14	
Air Blank	0.000	22:16	
Subject Sample	# 3 0.228	22:19	Volume Not Met
Breath Volume	0.066	22:19	
Air Blank	0.000	22:20	
Control Test	0.077	22:20	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:21	

Data Download Date/Time: 08/31/2025 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,RITCHIE,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: O'ROURKE,ROCKY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250284
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: BERTUBIN

DATE/TIME OF TEST: 08/29/2025 23:43:37

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:29	
Control Test	0.079	23:29	
Air Blank	0.000	23:30	
Subject Sample	# 1 0.221	23:30	
Breath Volume	1.308	23:30	
Air Blank	0.000	23:31	
Air Blank	0.000	23:33	
Subject Sample	# 2 0.188	23:36	Volume Not Met
Breath Volume	0.488	23:36	
Air Blank	0.000	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 3 0.170	23:42	Volume Not Met
Breath Volume	0.281	23:42	
Air Blank	0.000	23:42	
Control Test	0.079	23:43	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 08/31/2025 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,MORGAN,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MOONEY,MORGAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 250284
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:04

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 08/30/2025 07:28:11

Diagnostics Check	OK	07:25	
Air Blank	0.000	07:25	
Control Test	0.078	07:25	
Air Blank	0.000	07:26	
Subject Sample	# 1	07:26	Subject Test Refused
Breath Volume	0.000	07:26	
Air Blank	0.000	07:27	
Control Test	0.079	07:27	
Air Blank	0.000	07:28	
Diagnostics Check	OK	07:28	

Data Download Date/Time: 08/31/2025 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NASH,JUSTIN,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NASH,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SANVIDGE

DATE/TIME OF TEST: 08/29/2025 22:32:09

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.108	22:27
Breath Volume	3.187	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.103	22:30
Breath Volume	2.625	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 09/03/2025 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,DONALD,R
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SELLERS,DONALD
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: PHELPS

DATE/TIME OF TEST: 08/30/2025 01:49:55

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.188	01:45
Breath Volume	1.421	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:47
Subject Sample # 2	0.176	01:48
Breath Volume	1.238	01:48
Air Blank	0.000	01:48
Control Test	0.079	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 09/03/2025 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NOVAK,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: CALLAHAN

DATE/TIME OF TEST: 08/01/2025 02:37:17

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.080	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 08/29/2025 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: MURIEL MOHAN

DATE/TIME OF TEST: 08/01/2025 03:26:52

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.080	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:22	No Sample Provided
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 2	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.080	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 08/29/2025 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,KYLE,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HOGAN,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/04/2025 00:27:20

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.210	00:22
Breath Volume	3.558	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:24
Subject Sample # 2	0.230	00:25
Breath Volume	3.402	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 08/29/2025 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAPORCO,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: TAPORCO,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 08/04/2025 16:22:02

Diagnostics Check	OK	16:18	
Air Blank	0.000	16:19	
Control Test	0.080	16:19	
Air Blank	0.000	16:19	
Subject Sample	# 1	16:21	Sequence Aborted
Breath Volume	0.000	16:21	
Air Blank	0.000	16:22	

Data Download Date/Time: 08/29/2025 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEWART, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SEWART, WILLIAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 08/10/2025 05:24:26

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:21	
Control Test	0.080	05:22	
Air Blank	0.000	05:22	
Subject Sample	# 1	05:23	Subject Test Refused
Breath Volume	0.000	05:23	
Air Blank	0.000	05:23	
Control Test	0.080	05:23	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 08/29/2025 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HERTEL,JOHN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: MARTINEZ ACOSTA

DATE/TIME OF TEST: 08/10/2025 23:34:04

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.182	23:29
Breath Volume	1.328	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.191	23:32
Breath Volume	2.410	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:34

Data Download Date/Time: 08/29/2025 23:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: SALAS

DATE/TIME OF TEST: 08/14/2025 00:11:25

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:03	
Control Test	0.080	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:07	No Sample Provided
Breath Volume	0.000	00:07	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample	# 2	00:10	Subject Test Refused
Breath Volume	0.000	00:10	
Air Blank	0.000	00:10	
Control Test	0.080	00:10	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 08/29/2025 23:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,KYLE,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ACE MORENO,MAKENNA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: WILBURN

DATE/TIME OF TEST: 08/17/2025 01:15:19

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:11	No Sample Provided
Breath Volume	0.000	01:11	
Air Blank	0.000	01:11	
Air Blank	0.000	01:13	
Subject Sample	# 2	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:14	
Control Test	0.080	01:14	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 08/29/2025 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:17

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 08/17/2025 07:45:33

Diagnostics Check	OK	07:38
Air Blank	0.000	07:39
Control Test	0.080	07:39
Air Blank	0.000	07:40
Subject Sample # 1	0.147	07:40
Breath Volume	1.371	07:40
Air Blank	0.000	07:41
Air Blank	0.000	07:43
Subject Sample # 2	0.138	07:43
Breath Volume	3.402	07:43
Air Blank	0.000	07:44
Control Test	0.079	07:44
Air Blank	0.000	07:45
Diagnostics Check	OK	07:45

Data Download Date/Time: 08/29/2025 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JASON,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: REGISTER,JASON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 06:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: MORENO FRANCO

DATE/TIME OF TEST: 08/17/2025 08:02:33

Diagnostics Check	OK	07:55
Air Blank	0.000	07:55
Control Test	0.080	07:56
Air Blank	0.000	07:56
Subject Sample # 1	0.087	07:58
Breath Volume	3.265	07:58
Air Blank	0.000	07:58
Air Blank	0.000	08:00
Subject Sample # 2	0.087	08:01
Breath Volume	2.988	08:01
Air Blank	0.000	08:01
Control Test	0.080	08:01
Air Blank	0.000	08:02
Diagnostics Check	OK	08:02

Data Download Date/Time: 08/29/2025 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HERTEL,JOHN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/19/2025 03:14:01

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.184	03:09
Breath Volume	1.964	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:11
Subject Sample # 2	0.184	03:12
Breath Volume	1.761	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:14

Data Download Date/Time: 08/29/2025 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDDY,STERLING,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: EDDY,STERLING
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 08/23/2025 22:30:34

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.079	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.260	22:25
Breath Volume	4.636	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.257	22:28
Breath Volume	4.378	22:28
Air Blank	0.000	22:29
Control Test	0.077	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 08/29/2025 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: SOUR

DATE/TIME OF TEST: 08/24/2025 02:21:14

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:18	
Control Test	0.079	02:19	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:20	
Control Test	0.079	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 08/29/2025 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHEINHARDT,JERALD,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RHEINHARDT,JERALD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: HOLL

DATE/TIME OF TEST: 08/26/2025 03:25:01

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.262	03:20
Breath Volume	2.144	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:22
Subject Sample # 2	0.254	03:23
Breath Volume	2.296	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:25

Data Download Date/Time: 08/29/2025 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: CRONIN

DATE/TIME OF TEST: 08/27/2025 00:34:06

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.090	00:29
Breath Volume	4.164	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.081	00:32
Breath Volume	1.921	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 08/29/2025 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: LIZARDO

DATE/TIME OF TEST: 08/28/2025 01:45:00

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.079	01:37	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:40	No Sample Provided
Breath Volume	0.000	01:40	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample	# 2	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.080	01:44	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 08/29/2025 23:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDDY,S,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LEE,AARON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: PRESSNELL-HUFF

DATE/TIME OF TEST: 08/28/2025 21:43:17

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.261	21:38
Breath Volume	1.808	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.258	21:41
Breath Volume	1.839	21:41
Air Blank	0.000	21:42
Control Test	0.079	21:42
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 08/29/2025 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO, WESTON, C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DOUGHTY, LAURIE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/29/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 14:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 08/31/2025 17:23:48

Diagnostics Check	OK	17:20	
Air Blank	0.000	17:21	
Control Test	0.080	17:21	
Air Blank	0.000	17:22	
Subject Sample	# 1	17:22	Subject Test Refused
Breath Volume	0.000	17:22	
Air Blank	0.000	17:22	
Control Test	0.079	17:23	
Air Blank	0.000	17:23	
Diagnostics Check	OK	17:23	

Data Download Date/Time: 09/14/2025 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRAGUE,JONATHON,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: MUSE,DAVID
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: BOOKER

DATE/TIME OF TEST: 08/04/2025 22:11:58

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.082	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.000	22:07
Breath Volume	1.679	22:07
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.000	22:10
Breath Volume	1.289	22:10
Air Blank	0.000	22:11
Control Test	0.082	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 08/26/2025 19:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,IRAY,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: EHRLICH,DYLAN
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 08/08/2025 16:53:11

Diagnostics Check	OK	16:50	
Air Blank	0.000	16:50	
Control Test	0.082	16:51	
Air Blank	0.000	16:51	
Subject Sample	# 1	16:51	Subject Test Refused
Breath Volume	0.000	16:51	
Air Blank	0.000	16:52	
Control Test	0.081	16:52	
Air Blank	0.000	16:53	
Diagnostics Check	OK	16:53	

Data Download Date/Time: 08/26/2025 19:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINNIFIELD,SAM,M
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: MINNIFIELD,SAM
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 08/10/2025 00:21:56

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.083	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:21	
Control Test	0.083	00:21	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 08/26/2025 19:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRAGUE,JONATHON,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: MUSE,DAVID
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: SARVER

DATE/TIME OF TEST: 08/20/2025 00:01:23

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.081	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.206	23:55
Breath Volume	3.824	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.188	23:59
Breath Volume	2.453	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 08/26/2025 19:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREENE,JOSHUA,D

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: GREENE,JOSHUA

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403342588

DRY GAS EXPIRY DATE: 05/09/2028

VIOLATION CODE: DUI

ARREST TIME: 03:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: CARREON

DATE/TIME OF TEST: 08/22/2025 04:09:16

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:06	
Control Test	0.082	04:07	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:08	
Control Test	0.082	04:08	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 08/26/2025 19:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,MATHEW,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: GREENE,JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: MELON ROSA

DATE/TIME OF TEST: 08/24/2025 01:46:14

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:40	
Control Test	0.082	01:40	
Air Blank	0.000	01:41	
Subject Sample # 1	0.180	01:44	
Breath Volume	1.878	01:44	
Air Blank		01:45	Purge Fail
Air Blank	0.000	01:46	

Data Download Date/Time: 08/26/2025 19:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,MATHEW,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: GREENE,JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: MELON ROSA

DATE/TIME OF TEST: 08/24/2025 01:49:51

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.068	01:49	Control Outside Tolerance
Air Blank	0.000	01:49	

Data Download Date/Time: 08/26/2025 19:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,MATHEW,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: GREENE,JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: MELON ROSA

DATE/TIME OF TEST: 08/24/2025 02:04:16

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.076	01:52	
Air Blank	0.000	01:52	
Subject Sample	# 1 0.175	01:55	Slope Not Level
Breath Volume	1.195	01:55	
Air Blank	0.000	01:56	
Air Blank	0.000	01:58	
Subject Sample	# 2 0.193	01:59	
Breath Volume	2.816	01:59	
Air Blank	0.000	02:00	
Air Blank	0.000	02:02	
Subject Sample	# 3 0.194	02:02	
Breath Volume	3.078	02:02	
Air Blank	0.000	02:03	
Control Test	0.076	02:03	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 08/26/2025 19:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRY,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: GREENE,JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 08/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GESSENDORF

DATE/TIME OF TEST: 08/27/2025 02:54:51

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:52	
Control Test	0.082	02:52	
Air Blank	0.000	02:53	
Subject Sample	# 1	02:53	Subject Test Refused
Breath Volume	0.000	02:53	
Air Blank	0.000	02:53	
Control Test	0.082	02:54	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 09/30/2025 20:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREENE,JOSHUA,D

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: GREENE,JOSHUA

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 08/25/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403342588

DRY GAS EXPIRY DATE: 05/09/2028

VIOLATION CODE: DUI

ARREST TIME: 18:18

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: GAMBLE

DATE/TIME OF TEST: 08/27/2025 19:11:06

Diagnostics Check	OK	19:04
Air Blank	0.000	19:05
Control Test	0.081	19:05
Air Blank	0.000	19:05
Subject Sample # 1	0.000	19:06
Breath Volume	1.835	19:06
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.000	19:09
Breath Volume	1.480	19:09
Air Blank	0.000	19:10
Control Test	0.082	19:10
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 09/30/2025 20:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GROSS,BENJAMIN,E

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: GREENE,JOSHUA

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 08/25/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403342588

DRY GAS EXPIRY DATE: 05/09/2028

VIOLATION CODE: DUI

ARREST TIME: 19:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 08/30/2025 20:32:10

Diagnostics Check	OK	20:25
Air Blank	0.000	20:26
Control Test	0.081	20:26
Air Blank	0.000	20:27
Subject Sample # 1	0.369	20:27
Breath Volume	1.164	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.364	20:30
Breath Volume	1.480	20:30
Air Blank	0.000	20:31
Control Test	0.081	20:31
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 09/30/2025 20:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: CONWAY

DATE/TIME OF TEST: 08/22/2025 22:52:35

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.081	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.066	22:48
Breath Volume	2.394	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.069	22:51
Breath Volume	1.863	22:51
Air Blank	0.000	22:51
Control Test	0.081	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 08/28/2025 07:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: MARTIRE

DATE/TIME OF TEST: 08/23/2025 03:33:33

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.126	03:28
Breath Volume	2.097	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.112	03:32
Breath Volume	1.261	03:32
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 08/28/2025 07:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: FILS-AIME,KERVINS
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 08/23/2025 02:34:16

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.173	02:29
Breath Volume	1.554	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.167	02:32
Breath Volume	2.128	02:32
Air Blank	0.000	02:33
Control Test	0.078	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 09/01/2025 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVES,JULIAN,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: CHAVES,JULIAN
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: FANUS JR

DATE/TIME OF TEST: 08/28/2025 17:07:10

Diagnostics Check	OK	16:56	
Air Blank	0.000	16:57	
Control Test	0.080	16:57	
Air Blank	0.000	16:57	
Subject Sample	# 1 0.169	17:01	Volume Not Met
Breath Volume	0.484	17:01	
Air Blank	0.000	17:01	
Air Blank	0.000	17:03	
Subject Sample	# 2	17:05	Subject Test Refused
Breath Volume	0.066	17:05	
Air Blank	0.000	17:06	
Control Test	0.077	17:06	
Air Blank	0.000	17:07	
Diagnostics Check	OK	17:07	

Data Download Date/Time: 09/01/2025 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: PINEDA,MIGUEL
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 08/29/2025 01:33:22

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.082	01:31	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:32	
Control Test	0.081	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 09/01/2025 18:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: AHMED

DATE/TIME OF TEST: 08/19/2025 03:08:34

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.175	03:03
Breath Volume	2.343	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.164	03:06
Breath Volume	1.976	03:06
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 09/18/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CARRASCO LOPEZ

DATE/TIME OF TEST: 08/22/2025 02:33:28

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.155	02:28
Breath Volume	1.175	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.156	02:31
Breath Volume	2.285	02:31
Air Blank	0.000	02:32
Control Test	0.077	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 09/18/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PETERS, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: JUMPE

DATE/TIME OF TEST: 08/22/2025 04:46:26

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.078	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.163	04:41
Breath Volume	1.117	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.161	04:44
Breath Volume	2.058	04:44
Air Blank	0.000	04:45
Control Test	0.078	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 09/18/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: DELFIN

DATE/TIME OF TEST: 08/23/2025 02:50:06

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.078	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.173	02:45
Breath Volume	2.414	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.176	02:48
Breath Volume	2.683	02:48
Air Blank	0.000	02:49
Control Test	0.078	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 09/18/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: DIAZ LUQUE

DATE/TIME OF TEST: 08/23/2025 05:05:20

Diagnostics Check	OK	04:53	
Air Blank	0.000	04:53	
Control Test	0.077	04:54	
Air Blank	0.000	04:54	
Subject Sample	# 1 0.154	04:57	Volume Not Met
Breath Volume	0.171	04:57	
Air Blank	0.000	04:58	
Air Blank	0.000	05:00	
Subject Sample	# 2 0.148	05:00	
Breath Volume	2.003	05:00	
Air Blank	0.000	05:01	
Air Blank	0.000	05:03	
Subject Sample	# 3 0.150	05:03	
Breath Volume	2.085	05:03	
Air Blank	0.000	05:04	
Control Test	0.076	05:04	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 09/18/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEDINA,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: COMM VEHIC
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 08/24/2025 00:03:36

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:52	
Control Test	0.078	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 2	23:56	No Sample Provided
Breath Volume	0.000	23:56	
Air Blank	0.000	23:57	
Air Blank	0.000	23:59	
Subject Sample	# 1	00:02	Volume Not Met
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Control Test	0.078	00:02	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 09/18/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: VAILLANCOURT

DATE/TIME OF TEST: 08/24/2025 03:59:41

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.078	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.125	03:55
Breath Volume	3.007	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.122	03:58
Breath Volume	2.292	03:58
Air Blank	0.000	03:58
Control Test	0.077	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 09/18/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PETERS,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: AMBROSIO

DATE/TIME OF TEST: 08/24/2025 04:20:17

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.078	04:14
Air Blank	0.000	04:14
Subject Sample # 1	0.209	04:15
Breath Volume	2.601	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.213	04:18
Breath Volume	2.628	04:18
Air Blank	0.000	04:19
Control Test	0.077	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 09/18/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: MAKOWIECKI

DATE/TIME OF TEST: 08/24/2025 04:34:26

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.077	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.117	04:29
Breath Volume	3.867	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.115	04:32
Breath Volume	2.906	04:32
Air Blank	0.000	04:33
Control Test	0.077	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 09/18/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ, C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: VANGSNESS

DATE/TIME OF TEST: 08/27/2025 01:47:31

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.177	01:42
Breath Volume	1.562	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.175	01:45
Breath Volume	1.707	01:45
Air Blank	0.000	01:46
Control Test	0.078	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 09/18/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ABROMAVICIUS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: MESINAS-SEGUNDO

DATE/TIME OF TEST: 08/30/2025 04:27:56

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.014	04:23
Breath Volume	1.824	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.013	04:26
Breath Volume	2.445	04:26
Air Blank	0.000	04:27
Control Test	0.078	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 09/18/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/31/2025 02:04:20

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.078	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.097	01:59
Breath Volume	1.585	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.096	02:02
Breath Volume	1.554	02:02
Air Blank	0.000	02:03
Control Test	0.078	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 09/18/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: STARKEY

DATE/TIME OF TEST: 08/31/2025 04:12:42

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.078	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.229	04:08
Breath Volume	3.542	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.231	04:11
Breath Volume	3.136	04:11
Air Blank	0.000	04:11
Control Test	0.076	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 09/18/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: CARBALLO LLANO

DATE/TIME OF TEST: 08/31/2025 05:35:02

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.078	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.229	05:29
Breath Volume	1.882	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.212	05:33
Breath Volume	1.640	05:33
Air Blank	0.000	05:34
Control Test	0.077	05:34
Air Blank	0.000	05:34
Diagnostics Check	OK	05:35

Data Download Date/Time: 09/18/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

INSTRUMENT REGISTERED TO: PATRICK AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,NASSEEM,Q
BREATH TEST OPERATOR AGENCY: Patrick AFB PD
ARRESTING OFFICER: KAZEK,CAMERON
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: PFISTER

DATE/TIME OF TEST: 08/30/2025 00:34:18

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:32	
Control Test	0.081	00:32	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:33	Sequence Aborted
Breath Volume	0.000	00:33	
Air Blank	0.000	00:34	

Data Download Date/Time: 09/26/2025 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUTNEL,JOHN,D
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: PEACOCK,JAMES
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: EBERHART

DATE/TIME OF TEST: 08/25/2025 04:08:16

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.131	04:03
Breath Volume	1.359	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.135	04:06
Breath Volume	1.371	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 09/30/2025 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,D
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: CEPHUS,JAMIN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 08/29/2025 23:26:01

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.079	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.229	23:21
Breath Volume	1.570	23:21
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.236	23:24
Breath Volume	1.371	23:24
Air Blank	0.000	23:25
Control Test	0.075	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:26

Data Download Date/Time: 09/30/2025 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD, MITCHELL, J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BATES, E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: TALBOT

DATE/TIME OF TEST: 08/01/2025 00:33:50

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.082	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1 0.162	00:25	Volume Not Met
Breath Volume	0.949	00:25	
Air Blank	0.000	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 2 0.170	00:29	
Breath Volume	2.824	00:29	
Air Blank	0.000	00:29	
Air Blank	0.000	00:31	
Subject Sample	# 3 0.169	00:32	
Breath Volume	1.628	00:32	
Air Blank	0.000	00:32	
Control Test	0.081	00:33	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 08/06/2025 00:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD, MITCHELL, J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RUFIN

DATE/TIME OF TEST: 08/01/2025 03:26:54

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.081	03:23	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.080	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 08/06/2025 00:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BATES,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: RIDDLE

DATE/TIME OF TEST: 08/02/2025 00:47:00

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.083	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:46	
Control Test	0.082	00:46	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 08/06/2025 00:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,MITCHEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 08/02/2025 01:25:32

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.082	01:20
Air Blank	0.000	01:20
Subject Sample # 1	0.141	01:21
Breath Volume	1.703	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.145	01:24
Breath Volume	1.574	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 08/06/2025 00:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MAZZULLA

DATE/TIME OF TEST: 08/02/2025 05:00:18

Diagnostics Check	OK	04:52
Air Blank	0.000	04:53
Control Test	0.082	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.166	04:55
Breath Volume	2.832	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.170	04:58
Breath Volume	2.339	04:58
Air Blank	0.000	04:59
Control Test	0.082	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 08/06/2025 00:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: LUCAS SANTOS

DATE/TIME OF TEST: 08/02/2025 05:22:52

Diagnostics Check	OK	05:17	
Air Blank	0.000	05:17	
Control Test	0.082	05:18	
Air Blank	0.000	05:18	
Subject Sample	# 1	05:21	Subject Test Refused
Breath Volume	0.000	05:21	
Air Blank	0.000	05:21	
Control Test	0.082	05:22	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

Data Download Date/Time: 08/06/2025 00:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUNINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: FREY

DATE/TIME OF TEST: 08/02/2025 22:57:51

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:46	
Control Test	0.083	22:46	
Air Blank	0.000	22:47	
Subject Sample	# 1	22:50	No Sample Provided
Breath Volume	0.000	22:50	
Air Blank	0.000	22:50	
Air Blank	0.000	22:52	
Subject Sample	# 2	0.000	22:53
Breath Volume	2.136	22:53	
Air Blank	0.000	22:53	
Air Blank	0.000	22:55	
Subject Sample	# 3	0.000	22:56
Breath Volume	2.386	22:56	
Air Blank	0.000	22:56	
Control Test	0.082	22:57	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 08/06/2025 00:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD, MITCHELL, J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/03/2025 03:38:51

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.082	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.084	03:34
Breath Volume	1.363	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:36
Subject Sample # 2	0.078	03:37
Breath Volume	1.375	03:37
Air Blank	0.000	03:37
Control Test	0.081	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 08/06/2025 00:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: FRAGA CORCHO

DATE/TIME OF TEST: 08/05/2025 21:09:19

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:01	
Control Test	0.082	21:01	
Air Blank	0.000	21:01	
Subject Sample	# 1	21:05	No Sample Provided
Breath Volume	0.000	21:05	
Air Blank	0.000	21:05	
Air Blank	0.000	21:07	
Subject Sample	# 2	21:07	Subject Test Refused
Breath Volume	0.000	21:07	
Air Blank	0.000	21:08	
Control Test	0.082	21:08	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:09	

Data Download Date/Time: 08/06/2025 00:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 08/06/2025 02:17:01

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.083	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.168	02:11
Breath Volume	2.585	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.169	02:15
Breath Volume	2.675	02:15
Air Blank	0.000	02:16
Control Test	0.083	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:17

Data Download Date/Time: 09/17/2025 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 08/07/2025 03:37:17

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.082	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.129	03:32
Breath Volume	3.957	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.130	03:35
Breath Volume	3.898	03:35
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 09/17/2025 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: RUPERT

DATE/TIME OF TEST: 08/07/2025 22:45:48

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.082	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.000	22:41
Breath Volume	3.000	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.000	22:44
Breath Volume	2.597	22:44
Air Blank	0.000	22:44
Control Test	0.082	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 09/17/2025 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WECKNER

DATE/TIME OF TEST: 08/08/2025 01:51:33

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.083	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.174	01:47
Breath Volume	2.914	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.182	01:49
Breath Volume	2.988	01:49
Air Blank	0.000	01:50
Control Test	0.082	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 09/17/2025 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 08/08/2025 03:19:00

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.081	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.108	03:14
Breath Volume	1.296	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.103	03:17
Breath Volume	1.515	03:17
Air Blank	0.000	03:18
Control Test	0.082	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:19

Data Download Date/Time: 09/17/2025 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/08/2025 22:57:33

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.082	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.000	22:53
Breath Volume	1.746	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.000	22:56
Breath Volume	1.269	22:56
Air Blank	0.000	22:56
Control Test	0.082	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 09/17/2025 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD, MITCHELL, J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD, M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: BOHANNON

DATE/TIME OF TEST: 08/09/2025 01:34:09

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.128	01:29
Breath Volume	1.699	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.131	01:32
Breath Volume	2.121	01:32
Air Blank	0.000	01:33
Control Test	0.081	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 09/17/2025 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: FAGAN

DATE/TIME OF TEST: 08/09/2025 02:46:01

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.081	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.134	02:41
Breath Volume	1.773	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.134	02:44
Breath Volume	1.843	02:44
Air Blank	0.000	02:45
Control Test	0.081	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:46

Data Download Date/Time: 09/17/2025 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: SCHLUTER

DATE/TIME OF TEST: 08/09/2025 04:02:57

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:59	
Control Test	0.081	03:59	
Air Blank	0.000	03:59	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Control Test	0.082	04:02	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 09/17/2025 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: ESCARDO

DATE/TIME OF TEST: 08/11/2025 21:32:07

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.081	21:21	
Air Blank	0.000	21:22	
Subject Sample	# 1	21:25	No Sample Provided
Breath Volume	0.000	21:25	
Air Blank	0.000	21:25	
Air Blank	0.000	21:27	
Subject Sample	# 2	21:30	No Sample Provided
Breath Volume	0.000	21:30	
Air Blank	0.000	21:31	
Control Test	0.080	21:31	
Air Blank	0.000	21:32	
Diagnostics Check	OK	21:32	

Data Download Date/Time: 09/17/2025 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAH,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: WOJTAS

DATE/TIME OF TEST: 08/12/2025 15:21:39

Diagnostics Check	OK	15:15
Air Blank	0.000	15:16
Control Test	0.081	15:16
Air Blank	0.000	15:17
Subject Sample # 1	0.000	15:17
Breath Volume	1.398	15:17
Air Blank	0.000	15:17
Air Blank	0.000	15:19
Subject Sample # 2	0.000	15:20
Breath Volume	1.593	15:20
Air Blank	0.000	15:20
Control Test	0.079	15:21
Air Blank	0.000	15:21
Diagnostics Check	OK	15:21

Data Download Date/Time: 09/17/2025 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: INCANDELA

DATE/TIME OF TEST: 08/14/2025 00:42:58

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.081	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.270	00:38
Breath Volume	2.648	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:40
Subject Sample # 2	0.271	00:41
Breath Volume	3.230	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 09/17/2025 18:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,AUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: WOLD

DATE/TIME OF TEST: 08/14/2025 01:39:00

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.080	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:38	
Control Test	0.081	01:38	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:39	

Data Download Date/Time: 09/17/2025 18:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 08/14/2025 02:10:49

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.082	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.014	02:06
Breath Volume	1.398	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.014	02:09
Breath Volume	1.656	02:09
Air Blank	0.000	02:09
Control Test	0.082	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 09/17/2025 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: NUESI

DATE/TIME OF TEST: 08/14/2025 02:43:38

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.113	02:38
Breath Volume	1.523	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.109	02:42
Breath Volume	1.699	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 08/14/2025 22:43:56

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.081	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.048	22:39
Breath Volume	1.546	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.045	22:42
Breath Volume	1.851	22:42
Air Blank	0.000	22:43
Control Test	0.082	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/15/2025 01:54:04

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.082	01:47
Air Blank	0.000	01:48
Subject Sample # 1	0.080	01:49
Breath Volume	1.492	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.085	01:52
Breath Volume	1.210	01:52
Air Blank	0.000	01:53
Control Test	0.082	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:54

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CARVALHO PEREIRA DA

DATE/TIME OF TEST: 08/16/2025 03:45:42

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.081	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Air Blank	0.000	03:38	
Subject Sample	# 2	0.200	03:40
Breath Volume	1.175	03:40	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 3	0.211	03:44
Breath Volume	2.140	03:44	
Air Blank	0.000	03:44	
Control Test	0.081	03:45	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,AUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/16/2025 05:23:43

Diagnostics Check	OK	05:17
Air Blank	0.000	05:18
Control Test	0.081	05:18
Air Blank	0.000	05:19
Subject Sample # 1	0.000	05:19
Breath Volume	1.957	05:19
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.000	05:22
Breath Volume	2.175	05:22
Air Blank	0.000	05:22
Control Test	0.081	05:23
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GROOME

DATE/TIME OF TEST: 08/16/2025 23:29:07

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.081	23:22
Air Blank	0.000	23:22
Subject Sample # 1	0.154	23:23
Breath Volume	1.195	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.142	23:27
Breath Volume	1.242	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 08/16/2025 23:55:11

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.149	23:50
Breath Volume	2.046	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.144	23:53
Breath Volume	1.558	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 08/17/2025 00:05:12

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.078	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.133	23:59
Breath Volume	2.066	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.127	00:03
Breath Volume	1.957	00:03
Air Blank	0.000	00:04
Control Test	0.078	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: ROSSIER

DATE/TIME OF TEST: 08/17/2025 02:42:51

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.112	02:37
Breath Volume	2.707	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.110	02:41
Breath Volume	1.558	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/17/2025 04:51:14

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.081	04:40	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:43	No Sample Provided
Breath Volume	0.000	04:43	
Air Blank	0.000	04:44	
Air Blank	0.000	04:46	
Subject Sample	# 2	0.124	04:46
Breath Volume	2.425	04:46	
Air Blank	0.000	04:47	
Air Blank	0.000	04:49	
Subject Sample	# 3	0.132	04:49
Breath Volume	1.425	04:49	
Air Blank	0.000	04:50	
Control Test	0.081	04:50	
Air Blank	0.000	04:51	
Diagnostics Check	OK	04:51	

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:28

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/20/2025 10:54:57

Diagnostics Check	OK	10:48
Air Blank	0.000	10:49
Control Test	0.082	10:49
Air Blank	0.000	10:50
Subject Sample # 1	0.094	10:50
Breath Volume	1.246	10:50
Air Blank	0.000	10:51
Air Blank	0.000	10:53
Subject Sample # 2	0.094	10:53
Breath Volume	1.175	10:53
Air Blank	0.000	10:54
Control Test	0.080	10:54
Air Blank	0.000	10:54
Diagnostics Check	OK	10:54

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: KRAKHOVETSKII

DATE/TIME OF TEST: 08/21/2025 00:11:30

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.081	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:04	No Sample Provided
Breath Volume	0.000	00:04	
Air Blank	0.000	00:04	
Air Blank	0.000	00:07	
Subject Sample	# 2	00:10	No Sample Provided
Breath Volume	0.000	00:10	
Air Blank	0.000	00:10	
Control Test	0.082	00:10	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: KRAHOVETSKII

DATE/TIME OF TEST: 08/21/2025 00:21:48

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.081	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.155	00:16
Breath Volume	1.597	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.147	00:20
Breath Volume	1.917	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: OKER

DATE/TIME OF TEST: 08/21/2025 00:55:09

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
Control Test	0.081	00:52	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:54	
Control Test	0.081	00:54	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:55	

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: PUGLIESE

DATE/TIME OF TEST: 08/21/2025 14:17:31

Diagnostics Check	OK	14:11
Air Blank	0.000	14:11
Control Test	0.082	14:12
Air Blank	0.000	14:12
Subject Sample # 1	0.000	14:13
Breath Volume	3.343	14:13
Air Blank	0.000	14:13
Air Blank	0.000	14:15
Subject Sample # 2	0.000	14:16
Breath Volume	3.804	14:16
Air Blank	0.000	14:16
Control Test	0.081	14:16
Air Blank	0.000	14:17
Diagnostics Check	OK	14:17

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 08/21/2025 23:39:02

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.081	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.193	23:34
Breath Volume	1.250	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 2	0.202	23:37
Breath Volume	1.359	23:37
Air Blank	0.000	23:38
Control Test	0.081	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:39

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: LEARN

DATE/TIME OF TEST: 08/22/2025 00:30:00

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.081	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.099	00:25
Breath Volume	2.031	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.096	00:28
Breath Volume	1.910	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:30

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MRIBIHA

DATE/TIME OF TEST: 08/22/2025 02:45:08

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:42	
Control Test	0.081	02:42	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Control Test	0.081	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/22/2025 03:50:50

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.082	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.000	03:46
Breath Volume	1.105	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.000	03:49
Breath Volume	2.539	03:49
Air Blank	0.000	03:49
Control Test	0.081	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 08/22/2025 21:53:52

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.082	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.000	21:49
Breath Volume	1.898	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.000	21:52
Breath Volume	1.757	21:52
Air Blank	0.000	21:52
Control Test	0.081	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,CLAYTON,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,CLAYTON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: SMORODINSKAYA

DATE/TIME OF TEST: 08/23/2025 22:16:30

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:05	
Control Test	0.081	22:05	
Air Blank	0.000	22:06	
Subject Sample	# 1	22:09	No Sample Provided
Breath Volume	0.000	22:09	
Air Blank	0.000	22:09	
Air Blank	0.000	22:11	
Subject Sample	# 2	22:15	No Sample Provided
Breath Volume	0.000	22:15	
Air Blank	0.000	22:15	
Control Test	0.081	22:15	
Air Blank	0.000	22:16	
Diagnostics Check	OK	22:16	

Data Download Date/Time: 09/17/2025 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,WHITNEY,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MORROW,J
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: MANCHESTER

DATE/TIME OF TEST: 08/24/2025 04:44:20

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.082	04:42	
Air Blank	0.000	04:42	
Subject Sample	# 1	04:42	Subject Test Refused
Breath Volume	0.000	04:42	
Air Blank	0.000	04:43	
Control Test	0.083	04:43	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:44	

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BAKER, D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: GOSNEY

DATE/TIME OF TEST: 08/25/2025 18:20:52

Diagnostics Check	OK	18:14
Air Blank	0.000	18:14
Control Test	0.082	18:15
Air Blank	0.000	18:15
Subject Sample # 1	0.000	18:16
Breath Volume	1.695	18:16
Air Blank	0.000	18:16
Air Blank	0.000	18:18
Subject Sample # 2	0.000	18:19
Breath Volume	1.343	18:19
Air Blank	0.000	18:19
Control Test	0.081	18:20
Air Blank	0.000	18:20
Diagnostics Check	OK	18:20

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: TISDALE

DATE/TIME OF TEST: 08/26/2025 09:54:32

Diagnostics Check	OK	09:48
Air Blank	0.000	09:48
Control Test	0.082	09:49
Air Blank	0.000	09:49
Subject Sample # 1	0.000	09:50
Breath Volume	2.078	09:50
Air Blank	0.000	09:50
Air Blank	0.000	09:52
Subject Sample # 2	0.000	09:53
Breath Volume	2.003	09:53
Air Blank	0.000	09:53
Control Test	0.082	09:53
Air Blank	0.000	09:54
Diagnostics Check	OK	09:54

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 08/26/2025 23:29:45

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.081	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:28	
Control Test	0.081	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHETEE,NICHOLAS,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 08/27/2025 11:26:53

Diagnostics Check	OK	11:20
Air Blank	0.000	11:21
Control Test	0.082	11:21
Air Blank	0.000	11:22
Subject Sample # 1	0.099	11:22
Breath Volume	1.718	11:22
Air Blank	0.000	11:23
Air Blank	0.000	11:24
Subject Sample # 2	0.097	11:25
Breath Volume	1.851	11:25
Air Blank	0.000	11:25
Control Test	0.081	11:26
Air Blank	0.000	11:26
Diagnostics Check	OK	11:26

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KUSHNIR,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: CRESPO

DATE/TIME OF TEST: 08/28/2025 23:08:59

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.087	23:04
Breath Volume	2.113	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.085	23:07
Breath Volume	1.371	23:07
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: DELFIN

DATE/TIME OF TEST: 08/29/2025 03:30:18

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:26	
Control Test	0.082	03:27	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:28	Subject Test Refused
Breath Volume	0.000	03:28	
Air Blank	0.000	03:29	
Control Test	0.081	03:29	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,ZHARIA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: COUGHLIN

DATE/TIME OF TEST: 08/29/2025 05:16:41

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:14	
Control Test	0.082	05:14	
Air Blank	0.000	05:14	
Subject Sample	# 1	05:15	Subject Test Refused
Breath Volume	0.000	05:15	
Air Blank	0.000	05:15	
Control Test	0.081	05:16	
Air Blank	0.000	05:16	
Diagnostics Check	OK	05:16	

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, MATTHEW, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 08/30/2025 01:27:21

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.081	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.225	01:22
Breath Volume	3.230	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.216	01:25
Breath Volume	3.316	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 08/30/2025 02:32:30

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.009	02:26
Breath Volume	1.195	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.007	02:31
Breath Volume	1.589	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 08/30/2025 03:19:57

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.079	03:15
Breath Volume	1.804	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.085	03:18
Breath Volume	2.058	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 08/30/2025 03:58:47

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.080	03:48	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:51	No Sample Provided
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Air Blank	0.000	03:54	
Subject Sample	# 2	03:57	No Sample Provided
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.081	03:58	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KUSHNIR,ALEXANDER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/30/2025 04:18:19

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.176	04:11
Breath Volume	2.187	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.177	04:16
Breath Volume	2.347	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BAKER, D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: COFFIELD

DATE/TIME OF TEST: 08/30/2025 18:59:53

Diagnostics Check	OK	18:51	
Air Blank	0.000	18:51	
Control Test	0.081	18:52	
Air Blank	0.000	18:52	
Subject Sample	# 1	18:55	No Sample Provided
Breath Volume	0.000	18:55	
Air Blank	0.000	18:56	
Air Blank	0.000	18:58	
Subject Sample	# 2	18:58	Subject Test Refused
Breath Volume	0.000	18:58	
Air Blank	0.000	18:58	
Control Test	0.081	18:59	
Air Blank	0.000	18:59	
Diagnostics Check	OK	18:59	

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/31/2025 03:59:10

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.082	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.000	03:54
Breath Volume	2.519	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.000	03:57
Breath Volume	2.000	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: COSTA

DATE/TIME OF TEST: 08/31/2025 04:30:16

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.081	04:26	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:28	Subject Test Refused
Breath Volume	0.000	04:28	
Air Blank	0.000	04:29	
Control Test	0.082	04:29	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JULIO,C
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: GUSTIN

DATE/TIME OF TEST: 08/31/2025 04:59:22

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.082	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.000	04:55
Breath Volume	3.468	04:55
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.000	04:57
Breath Volume	3.042	04:57
Air Blank	0.000	04:58
Control Test	0.081	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 08/31/2025 05:52:46

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:40	
Control Test	0.081	05:40	
Air Blank	0.000	05:41	
Subject Sample	# 1	05:44	No Sample Provided
Breath Volume	0.000	05:44	
Air Blank	0.000	05:45	
Air Blank	0.000	05:47	
Subject Sample	# 2	0.195	No .020 agreement
Breath Volume	1.425	05:48	
Air Blank	0.000	05:48	
Air Blank	0.000	05:50	
Subject Sample	# 3	0.219	No .020 agreement
Breath Volume	1.304	05:51	
Air Blank	0.000	05:51	
Control Test	0.081	05:52	
Air Blank	0.000	05:52	
Diagnostics Check	OK	05:52	

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 08/31/2025 05:58:55

Diagnostics Check	OK	05:55	
Air Blank	0.000	05:55	
Control Test	0.080	05:55	
Air Blank	0.000	05:56	
Subject Sample	# 1	05:57	Subject Test Refused
Breath Volume	0.000	05:57	
Air Blank	0.000	05:58	
Control Test	0.081	05:58	
Air Blank	0.000	05:58	
Diagnostics Check	OK	05:58	

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KUSHNIR,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 08/31/2025 22:41:39

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:38	
Control Test	0.081	22:38	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:40	Subject Test Refused
Breath Volume	0.000	22:40	
Air Blank	0.000	22:40	
Control Test	0.081	22:41	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 09/17/2025 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEL VALLE,M,

BREATH TEST OPERATOR AGENCY: Dade City PD

ARRESTING OFFICER: RIVERA,R

ARRESTING OFFICER AGENCY: Dade City PD

LAST AGENCY INSPECTION: 07/27/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A3

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 15:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: GARCIA ORTIZ

DATE/TIME OF TEST: 08/23/2025 16:41:12

Diagnostics Check	OK	16:34
Air Blank	0.000	16:35
Control Test	0.080	16:35
Air Blank	0.000	16:36
Subject Sample # 1	0.236	16:36
Breath Volume	2.425	16:36
Air Blank	0.000	16:37
Air Blank	0.000	16:39
Subject Sample # 2	0.232	16:39
Breath Volume	2.882	16:39
Air Blank	0.000	16:40
Control Test	0.079	16:40
Air Blank	0.000	16:41
Diagnostics Check	OK	16:41

Data Download Date/Time: 08/28/2025 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYON,KYLE,A
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: ALBERT,COLBY
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 08/23/2025 22:06:25

Diagnostics Check	OK	21:59
Air Blank	0.000	22:00
Control Test	0.081	22:00
Air Blank	0.000	22:00
Subject Sample # 1	0.260	22:01
Breath Volume	2.917	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.241	22:04
Breath Volume	2.218	22:04
Air Blank	0.000	22:05
Control Test	0.077	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 09/15/2025 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: REYES,I
ARRESTING OFFICER AGENCY: KEY BISCAYNE PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CHAPARRO VILLAMIZAR

DATE/TIME OF TEST: 08/01/2025 01:55:57

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.075	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1 0.180	01:47	Volume Not Met
Breath Volume	0.035	01:47	
Air Blank	0.000	01:48	
Air Blank	0.000	01:49	
Subject Sample	# 2 0.215	01:50	
Breath Volume	1.937	01:50	
Air Blank	0.000	01:51	
Air Blank	0.000	01:53	
Subject Sample	# 3 0.202	01:54	
Breath Volume	2.242	01:54	
Air Blank	0.000	01:54	
Control Test	0.074	01:55	Control Outside Tolerance
Air Blank	0.000	01:55	

Data Download Date/Time: 08/07/2025 03:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: REYES,I
ARRESTING OFFICER AGENCY: KEY BISCAYNE PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CHAPARRO VILLAZAR

DATE/TIME OF TEST: 08/01/2025 02:18:05

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:15	
Control Test	0.075	02:15	
Air Blank	0.000	02:16	
Subject Sample	# 1	02:16	Subject Test Refused
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Control Test	0.080	02:17	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 08/07/2025 03:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 08/08/2025 04:17:10

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:11	
Control Test	0.078	04:11	
Air Blank	0.000	04:12	
Subject Sample # 1	0.191	04:12	
Breath Volume	2.042	04:12	
Air Blank	0.000	04:13	
Air Blank	0.000	04:15	
Subject Sample # 2	0.179	04:15	
Breath Volume	2.019	04:15	
Air Blank	0.000	04:16	
Control Test	0.074	04:16	Control Outside Tolerance
Air Blank	0.000	04:17	

Data Download Date/Time: 09/04/2025 05:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 08/08/2025 04:23:04

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:20	
Control Test	0.078	04:20	
Air Blank	0.000	04:21	
Subject Sample	# 1 0.180	04:21	
Breath Volume	1.976	04:21	
Air Blank		04:22	Purge Fail
Air Blank	0.000	04:23	

Data Download Date/Time: 09/04/2025 05:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: ARGUELLES

DATE/TIME OF TEST: 08/08/2025 22:30:10

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.076	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.000	22:25
Breath Volume	2.476	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.000	22:28
Breath Volume	1.574	22:28
Air Blank	0.000	22:29
Control Test	0.077	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 09/04/2025 05:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BELTRAN

DATE/TIME OF TEST: 08/16/2025 01:53:23

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.078	01:51	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.080	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 09/04/2025 05:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 08/16/2025 23:06:42

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:04	
Control Test	0.076	23:04	
Air Blank	0.000	23:05	
Subject Sample	# 1	23:05	Subject Test Refused
Breath Volume	0.000	23:05	
Air Blank	0.000	23:05	
Control Test	0.078	23:06	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 09/04/2025 05:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMAD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: BEN KIMON

DATE/TIME OF TEST: 08/19/2025 15:51:35

Diagnostics Check	OK	15:45
Air Blank	0.000	15:46
Control Test	0.079	15:46
Air Blank	0.000	15:46
Subject Sample # 1	0.000	15:47
Breath Volume	1.878	15:47
Air Blank	0.000	15:47
Air Blank	0.000	15:49
Subject Sample # 2	0.000	15:50
Breath Volume	2.226	15:50
Air Blank	0.000	15:50
Control Test	0.075	15:50
Air Blank	0.000	15:51
Diagnostics Check	OK	15:51

Data Download Date/Time: 09/04/2025 05:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: HORLSON

DATE/TIME OF TEST: 08/23/2025 06:06:21

Diagnostics Check	OK	05:59	
Air Blank	0.000	05:59	
Control Test	0.078	06:00	
Air Blank	0.000	06:00	
Subject Sample # 1	0.146	06:01	
Breath Volume	1.304	06:01	
Air Blank	0.000	06:02	
Air Blank	0.000	06:04	
Subject Sample # 2	0.141	06:04	
Breath Volume	1.714	06:04	
Air Blank	0.000	06:05	
Control Test	0.074	06:05	Control Outside Tolerance
Air Blank	0.000	06:06	

Data Download Date/Time: 09/04/2025 05:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: HORLSON

DATE/TIME OF TEST: 08/23/2025 06:12:04

Diagnostics Check	OK	06:10	
Air Blank	0.000	06:11	
Control Test	0.080	06:11	Interferent Detect
Air Blank	0.000	06:12	

Data Download Date/Time: 09/04/2025 05:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: HORLSON

DATE/TIME OF TEST: 08/23/2025 06:20:34

Diagnostics Check	OK	06:13	
Air Blank	0.000	06:14	
Control Test	0.078	06:14	
Air Blank	0.000	06:15	
Subject Sample # 1	0.149	06:15	
Breath Volume	1.359	06:15	
Air Blank	0.000	06:16	
Air Blank	0.000	06:18	
Subject Sample # 2	0.141	06:18	
Breath Volume	1.605	06:18	
Air Blank	0.000	06:19	
Control Test	0.080	06:19	
Air Blank	0.000	06:19	
Diagnostics Check	Fail	06:20	Analytical Stability Test
Air Blank	0.000	06:20	

Data Download Date/Time: 09/04/2025 05:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PEREZ,MARCOS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 08/29/2025 00:51:08

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.075	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.152	00:45
Breath Volume	1.562	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.142	00:49
Breath Volume	1.472	00:49
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 09/04/2025 05:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: LONDONO

DATE/TIME OF TEST: 08/30/2025 01:09:27

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.079	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.062	01:05
Breath Volume	2.222	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.063	01:07
Breath Volume	2.320	01:07
Air Blank	0.000	01:08
Control Test	0.076	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 09/04/2025 05:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MELENDEZ CANO

DATE/TIME OF TEST: 08/31/2025 03:05:41

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.077	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:04	
Control Test	0.075	03:05	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 09/04/2025 05:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LUNN,JAMES
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 08/09/2025 02:56:51

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.078	02:54	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.078	02:56	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 09/04/2025 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: PEREZ JIMENEZ

DATE/TIME OF TEST: 08/31/2025 04:57:53

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:54	
Control Test	0.079	04:55	
Air Blank	0.000	04:55	
Subject Sample	# 1	04:56	Subject Test Refused
Breath Volume	0.000	04:56	
Air Blank	0.000	04:56	
Control Test	0.079	04:57	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 09/04/2025 03:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEACHMAN,MICHAEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PEACHMAN,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 08/02/2025 19:30:53

Diagnostics Check	OK	19:24
Air Blank	0.000	19:24
Control Test	0.079	19:25
Air Blank	0.000	19:25
Subject Sample # 1	0.223	19:26
Breath Volume	2.035	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:28
Subject Sample # 2	0.217	19:29
Breath Volume	1.914	19:29
Air Blank	0.000	19:29
Control Test	0.077	19:30
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 08/22/2025 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, SMITH
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: LOGAN

DATE/TIME OF TEST: 08/02/2025 21:47:39

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.120	21:42
Breath Volume	2.265	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.122	21:45
Breath Volume	2.625	21:45
Air Blank	0.000	21:46
Control Test	0.078	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 08/22/2025 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ESCOBAR,JUAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: SANTIAGO LAZARO

DATE/TIME OF TEST: 08/03/2025 18:24:23

Diagnostics Check	OK	18:17
Air Blank	0.000	18:18
Control Test	0.080	18:18
Air Blank	0.000	18:19
Subject Sample # 1	0.140	18:19
Breath Volume	2.601	18:19
Air Blank	0.000	18:20
Air Blank	0.000	18:22
Subject Sample # 2	0.138	18:22
Breath Volume	2.226	18:22
Air Blank	0.000	18:23
Control Test	0.079	18:23
Air Blank	0.000	18:24
Diagnostics Check	OK	18:24

Data Download Date/Time: 08/22/2025 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZUNIGA,S,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: ZUNIGA,S
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: RINCON HIDALGO

DATE/TIME OF TEST: 08/09/2025 03:12:26

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.194	03:07
Breath Volume	2.671	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample # 2	0.187	03:10
Breath Volume	2.304	03:10
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 08/22/2025 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 08/10/2025 00:52:36

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.150	00:47
Breath Volume	2.355	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.150	00:50
Breath Volume	2.492	00:50
Air Blank	0.000	00:51
Control Test	0.078	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 08/22/2025 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 08/10/2025 05:48:36

Diagnostics Check	OK	05:41
Air Blank	0.000	05:41
Control Test	0.080	05:41
Air Blank	0.000	05:42
Subject Sample # 1	0.131	05:43
Breath Volume	1.281	05:43
Air Blank	0.000	05:44
Air Blank	0.000	05:45
Subject Sample # 2	0.131	05:46
Breath Volume	2.277	05:46
Air Blank	0.000	05:47
Control Test	0.078	05:47
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time: 08/22/2025 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRARO,PAUL,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: FERRARO,PAUL
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 08/11/2025 02:01:54

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
		02:01	RFI Detect
Air Blank	0.000	02:01	

Data Download Date/Time: 08/22/2025 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: DIPASCALE,JEFFERY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 08/13/2025 03:08:26

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.178	03:03
Breath Volume	1.761	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:05
Subject Sample # 2	0.179	03:06
Breath Volume	1.984	03:06
Air Blank	0.000	03:07
Control Test	0.076	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 08/22/2025 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: SHAFER

DATE/TIME OF TEST: 08/15/2025 22:48:27

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.080	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.123	22:43
Breath Volume	2.144	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.123	22:46
Breath Volume	2.082	22:46
Air Blank	0.000	22:47
Control Test	0.078	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 08/22/2025 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: CAPARCO

DATE/TIME OF TEST: 08/16/2025 00:15:31

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.077	00:10
Breath Volume	2.117	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.084	00:13
Breath Volume	3.085	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 08/22/2025 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEON,BRIAN,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: STEWART,CAMERON
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: ESPINALES MEJIA

DATE/TIME OF TEST: 08/16/2025 05:27:05

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.080	05:20
Air Blank	0.000	05:21
Subject Sample # 1	0.222	05:22
Breath Volume	1.792	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:24
Subject Sample # 2	0.213	05:25
Breath Volume	1.777	05:25
Air Blank	0.000	05:26
Control Test	0.078	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 08/22/2025 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEON,BRIAN,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CRUZ,MIA
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: DIEGO JUAN

DATE/TIME OF TEST: 08/17/2025 04:13:24

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:07	
Control Test	0.080	04:07	
Air Blank	0.000	04:08	
Subject Sample	# 1 0.218	04:08	
Breath Volume	2.039	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:11	
Subject Sample	# 2 0.225	04:11	
Breath Volume	1.812	04:11	
Air Blank	0.000	04:12	
Control Test	0.079	04:12	
Air Blank		04:12	RFI Detect
Air Blank	0.000	04:13	

Data Download Date/Time: 08/22/2025 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,TORL,D
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BIRTO,D
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: BODEN

DATE/TIME OF TEST: 08/19/2025 12:59:34

Diagnostics Check	OK	12:53
Air Blank	0.000	12:53
Control Test	0.079	12:53
Air Blank	0.000	12:54
Subject Sample # 1	0.100	12:54
Breath Volume	2.957	12:54
Air Blank	0.000	12:55
Air Blank	0.000	12:57
Subject Sample # 2	0.094	12:57
Breath Volume	2.535	12:57
Air Blank	0.000	12:58
Control Test	0.078	12:58
Air Blank	0.000	12:59
Diagnostics Check	OK	12:59

Data Download Date/Time: 08/22/2025 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH,GAVIN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: ESPINOZA BALBUENA

DATE/TIME OF TEST: 08/20/2025 02:14:22

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.081	02:09
Breath Volume	1.246	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.086	02:12
Breath Volume	1.777	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 08/22/2025 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, GAVIN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: KELSOR

DATE/TIME OF TEST: 08/23/2025 03:51:34

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.000	03:47
Breath Volume	3.058	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.000	03:50
Breath Volume	2.742	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 09/23/2025 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ROSELLI LARA

DATE/TIME OF TEST: 08/24/2025 03:37:08

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.123	03:32
Breath Volume	2.835	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.125	03:35
Breath Volume	1.750	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 09/23/2025 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: GARRETT, JARED
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: FELPETO

DATE/TIME OF TEST: 08/28/2025 01:34:49

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.000	01:30
Breath Volume	1.902	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.000	01:33
Breath Volume	2.687	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 09/23/2025 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MARCHENA, KELMY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: REAVES

DATE/TIME OF TEST: 08/28/2025 22:54:00

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.081	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.000	22:49
Breath Volume	3.796	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.000	22:52
Breath Volume	3.859	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:54

Data Download Date/Time: 09/23/2025 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: NIEMI

DATE/TIME OF TEST: 08/31/2025 04:06:47

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.235	04:01
Breath Volume	1.718	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.247	04:04
Breath Volume	2.160	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 09/23/2025 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 08/31/2025 04:43:11

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.079	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.198	04:38
Breath Volume	2.757	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:40
Subject Sample # 2	0.183	04:41
Breath Volume	1.734	04:41
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 09/23/2025 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SEAY

DATE/TIME OF TEST: 08/16/2025 01:11:58

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.079	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.164	01:07
Breath Volume	2.078	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:09
Subject Sample # 2	0.160	01:10
Breath Volume	1.937	01:10
Air Blank	0.000	01:11
Control Test	0.076	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 08/27/2025 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 08/23/2025 02:10:24

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.079	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.096	02:05
Breath Volume	2.582	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.091	02:08
Breath Volume	2.519	02:08
Air Blank	0.000	02:09
Control Test	0.078	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 08/27/2025 13:46