Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MANDILE,JEREMY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: CHENEY

DATE/TIME OF TEST: 09/18/2025 00:11:18

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.148	00:06
Breath Volume		3.886	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.148	00:09
Breath Volume		3.953	00:09
Air Blank		0.000	00:10
Control Test		0.079	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 09/30/2025 20:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 09/18/2025 02:03:39

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.108	01:59
Breath Volume		1.578	01:59
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.109	02:02
Breath Volume		3.523	02:02
Air Blank		0.000	02:02
Control Test		0.080	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 09/30/2025 20:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SIMS,WALKER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: TOLKOV

DATE/TIME OF TEST: 09/20/2025 02:46:52

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.080	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.119	02:42
Breath Volume		3.136	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.127	02:45
Breath Volume		3.000	02:45
Air Blank		0.000	02:45
Control Test		0.080	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 09/30/2025 20:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MIGLIACCIO,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: GRIGG

DATE/TIME OF TEST: 09/20/2025 04:23:55

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.080	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.203	04:19
Breath Volume		5.179	04:19
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.193	04:22
Breath Volume		4.000	04:22
Air Blank		0.000	04:22
Control Test		0.079	04:23
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 09/30/2025 20:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 09/20/2025 21:01:58

Diagnostics Check		OK	20:55
Air Blank		0.000	20:56
Control Test		0.081	20:56
Air Blank		0.000	20:57
Subject Sample	# 1	0.096	20:57
Breath Volume		2.121	20:57
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.096	21:00
Breath Volume		2.320	21:00
Air Blank		0.000	21:01
Control Test		0.080	21:01
Air Blank		0.000	21:01
Diagnostics Check		OK	21:01

Data Download Date/Time: 09/30/2025 20:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DAWIDOWICZ,SARAH
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 08/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/03/2025 22:28:23

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:22
Subject Sample	# 1	0.196	22:23
Breath Volume		1.980	22:23
Air Blank		0.000	22:24
Air Blank		0.000	22:26
Subject Sample	# 2	0.186	22:26
Breath Volume		1.140	22:26
Air Blank		0.000	22:27
Control Test		0.080	22:27
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 09/28/2025 16:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ABRAHAM,ARTHUR
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 08/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 11:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:41

SUBJECT'S LAST NAME: WARKER

DATE/TIME OF TEST: 09/09/2025 12:10:43

Diagnostics Check		OK	12:01
Air Blank		0.000	12:02
Control Test		0.079	12:02
Air Blank		0.000	12:03
Subject Sample	# 1	0.000	12:04
Breath Volume		1.781	12:04
Air Blank		0.000	12:05
Air Blank		0.000	12:07
Subject Sample	# 2	0.000	12:09
Breath Volume		1.628	12:09
Air Blank		0.000	12:09
Control Test		0.078	12:10
Air Blank		0.000	12:10
Diagnostics Check		OK	12:10

Data Download Date/Time: 09/28/2025 16:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WEST,TIECHA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: VEYNOVICH

DATE/TIME OF TEST: 09/11/2025 23:46:30

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.079	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.226	23:42
Breath Volume		1.937	23:42
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.229	23:44
Breath Volume		1.890	23:44
Air Blank		0.000	23:45
Control Test		0.079	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 09/28/2025 16:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNAL,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Holly Hills PD
ARRESTING OFFICER: SOMEREVE,ERNEST
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 08/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 08:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 09/14/2025 09:27:48

Diagnostics Check		OK	09:21
Air Blank		0.000	09:21
Control Test		0.079	09:21
Air Blank		0.000	09:22
Subject Sample	# 1	0.103	09:22
Breath Volume		1.214	09:22
Air Blank		0.000	09:23
Air Blank		0.000	09:25
Subject Sample	# 2	0.105	09:26
Breath Volume		3.085	09:26
Air Blank		0.000	09:26
Control Test		0.079	09:27
Air Blank		0.000	09:27
Diagnostics Check		OK	09:27

Data Download Date/Time: 09/28/2025 16:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: QUELIZ,CARLOS
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 08/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 09/17/2025 21:47:40

Diagnostics Check		OK	21:40
Air Blank		0.000	21:41
Control Test		0.080	21:41
Air Blank		0.000	21:41
Subject Sample	# 1	0.070	21:42
Breath Volume		1.992	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.067	21:46
Breath Volume		1.953	21:46
Air Blank		0.000	21:46
Control Test		0.079	21:47
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 09/28/2025 16:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSH,WESLEY,T
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: PETTY,NYROBI
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 08/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: VYAS

DATE/TIME OF TEST: 09/27/2025 01:01:13

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.079	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.249	00:56
Breath Volume		3.156	00:56
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.262	00:59
Breath Volume		3.750	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 09/28/2025 16:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SWIDER,MATTHEW
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 08/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: LOVELADY

DATE/TIME OF TEST: 09/27/2025 04:17:01

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.079	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.209	04:12
Breath Volume		3.527	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.199	04:15
Breath Volume		3.328	04:15
Air Blank		0.000	04:16
Control Test		0.079	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:17

Data Download Date/Time: 09/28/2025 16:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,KEELAN,D
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: LLINAS,JOEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 09/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: BOURGEOIS

DATE/TIME OF TEST: 09/30/2025 21:07:27

Diagnostics Check		OK	21:01
Air Blank		0.000	21:02
Control Test		0.079	21:02
Air Blank		0.000	21:02
Subject Sample	# 1	0.000	21:03
Breath Volume		2.191	21:03
Air Blank		0.000	21:03
Air Blank		0.000	21:05
Subject Sample	# 2	0.000	21:06
Breath Volume		1.976	21:06
Air Blank		0.000	21:06
Control Test		0.078	21:06
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

Data Download Date/Time: 10/30/2025 19:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMINI,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Holly Hills PD
ARRESTING OFFICER: DIMINI,NICHOLAS
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 09/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 09/30/2025 22:11:20

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.079	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.179	22:06
Breath Volume		1.746	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.178	22:09
Breath Volume		1.441	22:09
Air Blank		0.000	22:10
Control Test		0.077	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 10/30/2025 19:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOONE,MICHAEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: HEATON

DATE/TIME OF TEST: 09/04/2025 23:25:01

Diagnostics Check		OK	23:22	
Air Blank		0.000	23:22	
Control Test		0.080	23:23	
Air Blank		0.000	23:23	
Subject Sample	# 1		23:23	Subject Test Refused
Breath Volume		0.000	23:23	
Air Blank		0.000	23:24	
Control Test		0.080	23:24	
Air Blank		0.000	23:24	
Diagnostics Check		OK	23:25	

Data Download Date/Time: 10/01/2025 10:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: MARROQUIN MARTINEZ

DATE/TIME OF TEST: 09/06/2025 00:33:56

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.080	00:28
Air Blank		0.000	00:28
Subject Sample	# 1	0.165	00:29
Breath Volume		1.820	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:31
Subject Sample	# 2	0.171	00:32
Breath Volume		1.687	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 10/01/2025 10:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DE LA CAMPA,BRYANT
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: VELASQUEZ TORREZ

DATE/TIME OF TEST: 09/06/2025 02:00:03

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.079	01:53
Air Blank		0.000	01:54
Subject Sample	# 1	0.165	01:55
Breath Volume		3.382	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.164	01:58
Breath Volume		1.687	01:58
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	02:00

Data Download Date/Time: 10/01/2025 10:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VAZQUEZ,JUSTIN
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: DAVADI

DATE/TIME OF TEST: 09/06/2025 02:15:40

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.112	02:11
Breath Volume		1.484	02:11
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.110	02:14
Breath Volume		2.191	02:14
Air Blank		0.000	02:14
Control Test		0.079	02:15
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 10/01/2025 10:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PORTORO,SHERRY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 09:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:01

SUBJECT'S LAST NAME: LINARES SANABRIA

DATE/TIME OF TEST: 09/07/2025 13:47:27

Diagnostics Check		OK	13:40
Air Blank		0.000	13:40
Control Test		0.080	13:41
Air Blank		0.000	13:41
Subject Sample	# 1	0.105	13:43
Breath Volume		1.460	13:43
Air Blank		0.000	13:43
Air Blank		0.000	13:45
Subject Sample	# 2	0.107	13:46
Breath Volume		1.507	13:46
Air Blank		0.000	13:46
Control Test		0.079	13:46
Air Blank		0.000	13:47
Diagnostics Check		OK	13:47

Data Download Date/Time: 10/01/2025 10:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BEAUREGARD,JEFF
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SHAVER

DATE/TIME OF TEST: 09/13/2025 00:26:12

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.080	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.147	00:21
Breath Volume		3.968	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.151	00:24
Breath Volume		4.093	00:24
Air Blank		0.000	00:25
Control Test		0.080	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 10/01/2025 10:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BARFIELD,COLE
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: SOHOM

DATE/TIME OF TEST: 09/13/2025 01:10:05

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.081	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.245	01:05
Breath Volume		1.355	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.249	01:08
Breath Volume		2.605	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:10

Data Download Date/Time: 10/01/2025 10:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: REYES MEJIA

DATE/TIME OF TEST: 09/13/2025 05:20:07

Diagnostics Check		OK	05:13
Air Blank		0.000	05:14
Control Test		0.081	05:14
Air Blank		0.000	05:15
Subject Sample	# 1	0.204	05:15
Breath Volume		2.722	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.189	05:18
Breath Volume		3.109	05:18
Air Blank		0.000	05:19
Control Test		0.080	05:19
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 10/01/2025 10:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: CHRISTOPHER

DATE/TIME OF TEST: 09/14/2025 23:53:50

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:46
Subject Sample	# 1	0.214	23:48
Breath Volume		2.046	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.205	23:52
Breath Volume		1.824	23:52
Air Blank		0.000	23:52
Control Test		0.079	23:53
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 10/01/2025 10:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOWE,R
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BLEIBLE

DATE/TIME OF TEST: 09/20/2025 01:39:24

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.080	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.093	01:34
Breath Volume		3.636	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.095	01:37
Breath Volume		4.035	01:37
Air Blank		0.000	01:38
Control Test		0.080	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 10/01/2025 10:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WRIGHT,KIERSTEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: ZITMAN BRICENO

DATE/TIME OF TEST: 09/20/2025 03:25:35

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.135	03:21
Breath Volume		2.285	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.137	03:24
Breath Volume		2.429	03:24
Air Blank		0.000	03:24
Control Test		0.080	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 10/01/2025 10:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PAEZ,ANTHONY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: NELSON JR.

DATE/TIME OF TEST: 09/20/2025 04:01:28

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.080	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.125	03:56
Breath Volume		1.296	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.124	04:00
Breath Volume		1.894	04:00
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 10/01/2025 10:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: ROSSI CASTRO

DATE/TIME OF TEST: 09/20/2025 04:36:49

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.242	04:32
Breath Volume		2.035	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:34
Subject Sample	# 2	0.238	04:35
Breath Volume		2.628	04:35
Air Blank		0.000	04:35
Control Test		0.080	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 10/01/2025 10:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROMERO,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: HARVARD

DATE/TIME OF TEST: 09/20/2025 23:15:25

Diagnostics Check		OK	23:07
Air Blank		0.000	23:08
Control Test		0.080	23:08
Air Blank		0.000	23:08
Subject Sample	# 1	0.052	23:10
Breath Volume		1.246	23:10
Air Blank		0.000	23:11
Air Blank		0.000	23:12
Subject Sample	# 2	0.053	23:13
Breath Volume		2.281	23:13
Air Blank		0.000	23:14
Control Test		0.080	23:14
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 10/01/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 09/21/2025 00:57:50

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.080	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.138	00:53
Breath Volume		2.941	00:53
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.140	00:56
Breath Volume		3.046	00:56
Air Blank		0.000	00:56
Control Test		0.079	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 10/01/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WRIGHT,KIERSTEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: CABAL

DATE/TIME OF TEST: 09/21/2025 02:20:06

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.081	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.215	02:15
Breath Volume		1.433	02:15
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.231	02:18
Breath Volume		2.531	02:18
Air Blank		0.000	02:19
Control Test		0.079	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 10/01/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: USTINOVA

DATE/TIME OF TEST: 09/21/2025 02:47:55

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.080	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.221	02:42
Breath Volume		1.238	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:44
Subject Sample	# 2	0.239	02:46
Breath Volume		2.175	02:46
Air Blank		0.000	02:47
Control Test		0.079	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 10/01/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 09/26/2025 01:46:45

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.080	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.179	01:42
Breath Volume		2.164	01:42
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.179	01:45
Breath Volume		1.945	01:45
Air Blank		0.000	01:45
Control Test		0.080	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 10/01/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MATEO,JOHNNY
ARRESTING OFFICER AGENCY: Lauderhill PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 09/26/2025 18:06:43

Diagnostics Check		OK	17:59
Air Blank		0.000	18:00
Control Test		0.079	18:00
Air Blank		0.000	18:01
Subject Sample	# 1	0.124	18:02
Breath Volume		1.335	18:02
Air Blank		0.000	18:02
Air Blank		0.000	18:04
Subject Sample	# 2	0.130	18:05
Breath Volume		1.367	18:05
Air Blank		0.000	18:05
Control Test		0.080	18:06
Air Blank		0.000	18:06
Diagnostics Check		OK	18:06

Data Download Date/Time: 10/01/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TOWNSEND,KIRK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:14

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 09/27/2025 07:40:56

Diagnostics Check		OK	07:35
Air Blank		0.000	07:35
Control Test		0.080	07:35
Air Blank		0.000	07:36
Subject Sample	# 1	0.042	07:36
Breath Volume		2.050	07:36
Air Blank		0.000	07:37
Air Blank		0.000	07:39
Subject Sample	# 2	0.037	07:39
Breath Volume		1.636	07:39
Air Blank		0.000	07:40
Control Test		0.080	07:40
Air Blank		0.000	07:40
Diagnostics Check		OK	07:40

Data Download Date/Time: 10/01/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: SAWH

DATE/TIME OF TEST: 09/29/2025 22:54:51

Diagnostics Check		OK	22:46
Air Blank		0.000	22:47
Control Test		0.080	22:47
Air Blank		0.000	22:48
Subject Sample	# 1	0.189	22:50
Breath Volume		1.312	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.201	22:53
Breath Volume		1.339	22:53
Air Blank		0.000	22:53
Control Test		0.079	22:54
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 10/01/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,SARAH,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BROWN,SARAH
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: KHAWAM JIMENEZ

DATE/TIME OF TEST: 09/26/2025 20:26:44

Diagnostics Check		OK	20:19
Air Blank		0.000	20:19
Control Test		0.081	20:19
Air Blank		0.000	20:20
Subject Sample	# 1	0.088	20:21
Breath Volume		2.332	20:21
Air Blank		0.000	20:22
Air Blank		0.000	20:24
Subject Sample	# 2	0.083	20:25
Breath Volume		1.632	20:25
Air Blank		0.000	20:25
Control Test		0.078	20:26
Air Blank		0.000	20:26
Diagnostics Check		OK	20:26

Data Download Date/Time: 11/10/2025 09:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELACRUZ,JONATHAN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPRINGMAN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/28/2025 22:20:17

Diagnostics Check		OK	22:13	
Air Blank		0.000	22:14	
Control Test		0.081	22:14	
Air Blank		0.000	22:15	
Subject Sample	# 1	0.242	22:16	
Breath Volume		2.847	22:16	
Air Blank		0.000	22:17	
Air Blank			22:18	RFI Detect
Air Blank		0.000	22:19	
Control Test		0.080	22:19	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:20	

Data Download Date/Time: 11/10/2025 09:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELACRUZ,JONATHAN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPRINGMAN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/28/2025 22:28:36

Diagnostics Check		OK	22:21
Air Blank		0.000	22:22
Control Test		0.080	22:22
Air Blank		0.000	22:23
Subject Sample	# 1	0.238	22:23
Breath Volume		2.562	22:23
Air Blank		0.000	22:24
Air Blank		0.000	22:26
Subject Sample	# 2	0.229	22:26
Breath Volume		2.660	22:26
Air Blank		0.000	22:27
Control Test		0.076	22:27
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 11/10/2025 09:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIERA,SCOTT,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: RIERA,SCOTT
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: ELLIOT

DATE/TIME OF TEST: 09/07/2025 17:27:24

Diagnostics Check		OK	17:20
Air Blank		0.000	17:21
Control Test		0.078	17:21
Air Blank		0.000	17:21
Subject Sample	# 1	0.234	17:22
Breath Volume		2.304	17:22
Air Blank		0.000	17:23
Air Blank		0.000	17:25
Subject Sample	# 2	0.237	17:25
Breath Volume		2.062	17:25
Air Blank		0.000	17:26
Control Test		0.076	17:26
Air Blank		0.000	17:27
Diagnostics Check		OK	17:27

Data Download Date/Time: 09/24/2025 06:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OWENS,JONATHAN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: OLIBRICE

DATE/TIME OF TEST: 09/01/2025 23:51:32

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:45	
Control Test		0.081	23:46	
Air Blank		0.000	23:46	
Subject Sample	# 1	0.210	23:49	Volume Not Met
Breath Volume		0.789	23:49	
Air Blank			23:50	Improper Sample
Air Blank			23:50	Purge Fail
Air Blank		0.000	23:51	

Data Download Date/Time: 09/15/2025 11:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OWENS,JONATHAN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: OLIBRICE

DATE/TIME OF TEST: 09/01/2025 23:59:05

Diagnostics Check		OK	23:53	
Air Blank		0.000	23:53	
Control Test		0.078	23:54	
Air Blank		0.000	23:54	
Subject Sample	# 1	0.217	23:57	Volume Not Met
Breath Volume		1.066	23:57	
Air Blank			23:58	Purge Fail
Air Blank		0.000	23:59	

Data Download Date/Time: 09/15/2025 11:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFOUR,HALEY,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CARDONE,DOUG
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: MANGAN

DATE/TIME OF TEST: 09/04/2025 23:44:20

Diagnostics Check		OK	23:32
Air Blank		0.000	23:33
Control Test		0.082	23:33
Air Blank		0.000	23:34
Subject Sample	# 1	0.289	23:35
Breath Volume		1.480	23:35
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.267	23:39
Breath Volume		2.335	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 3	0.290	23:42
Breath Volume		1.414	23:42
Air Blank		0.000	23:43
Control Test		0.080	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 09/27/2025 12:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFOUR,HALEY,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CORNER,TROY
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/14/2025 01:21:19

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.083	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.180	01:16
Breath Volume		1.710	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.199	01:19
Breath Volume		1.890	01:19
Air Blank		0.000	01:20
Control Test		0.081	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 09/27/2025 12:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFOUR,HALEY,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: EDMONDS,MELISSA
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: GAUSTAD

DATE/TIME OF TEST: 09/14/2025 03:40:57

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.082	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.081	03:35
Breath Volume		2.593	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.085	03:39
Breath Volume		2.441	03:39
Air Blank		0.000	03:40
Control Test		0.082	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 09/27/2025 12:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMINI,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Holly Hills PD
ARRESTING OFFICER: DIMINI,NICHOLAS
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: STUBER

DATE/TIME OF TEST: 09/20/2025 22:43:59

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.082	22:37
Air Blank		0.000	22:38
Subject Sample	# 1	0.046	22:39
Breath Volume		1.507	22:39
Air Blank		0.000	22:40
Air Blank		0.000	22:41
Subject Sample	# 2	0.048	22:42
Breath Volume		1.875	22:42
Air Blank		0.000	22:43
Control Test		0.081	22:43
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 09/27/2025 12:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMINI,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Holly Hills PD
ARRESTING OFFICER: DIMINI,NICHOLAS
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: AL-YASIRY

DATE/TIME OF TEST: 09/26/2025 02:29:42

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:26	
Control Test		0.083	02:27	
Air Blank		0.000	02:27	
Subject Sample	# 1		02:28	Subject Test Refused
Breath Volume		0.000	02:28	
Air Blank		0.000	02:28	
Control Test		0.083	02:29	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 09/27/2025 12:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,RAYMOND,S
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: MILLER,RAYMOND
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 08/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: BILAL

DATE/TIME OF TEST: 09/06/2025 04:47:55

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:40	
Control Test		0.078	04:40	
Air Blank		0.000	04:41	
Subject Sample	# 1		04:44	No Sample Provided
Breath Volume		0.000	04:44	
Air Blank		0.000	04:44	
Air Blank		0.000	04:46	
Subject Sample	# 2		04:47	Sequence Aborted
Breath Volume		0.000	04:47	
Air Blank		0.000	04:47	

Data Download Date/Time: 09/18/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,RAYMOND,S
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: MILLER,RAYMOND
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 08/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: BILAL

DATE/TIME OF TEST: 09/06/2025 05:24:49

Diagnostics Check		OK	05:15
Air Blank		0.000	05:15
Control Test		0.078	05:16
Air Blank		0.000	05:16
Subject Sample	# 1	0.198	05:19
Breath Volume		1.562	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.215	05:23
Breath Volume		2.082	05:23
Air Blank		0.000	05:23
Control Test		0.075	05:24
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 09/18/2025 09:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DOMINGUEZ,M
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: USAGA FERNANDEZ

DATE/TIME OF TEST: 09/06/2025 15:25:52

Diagnostics Check	OK	15:24	
Air Blank	0.000	15:24	
Control Test	0.073	15:25	Control Outside Tolerance
Air Blank	0.000	15:25	

Data Download Date/Time: 09/12/2025 11:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DOMINGUEZ,M
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: USAGA FERNANDEZ

DATE/TIME OF TEST: 09/06/2025 15:35:28

Diagnostics Check		OK	15:28
Air Blank		0.000	15:29
Control Test		0.079	15:29
Air Blank		0.000	15:30
Subject Sample	# 1	0.218	15:31
Breath Volume		1.785	15:31
Air Blank		0.000	15:31
Air Blank		0.000	15:33
Subject Sample	# 2	0.210	15:34
Breath Volume		1.738	15:34
Air Blank		0.000	15:34
Control Test		0.080	15:34
Air Blank		0.000	15:35
Diagnostics Check		OK	15:35

Data Download Date/Time: 09/12/2025 11:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FERNANDEZ,ANGEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/20/2026
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: ROPER

DATE/TIME OF TEST: 09/28/2025 00:40:01

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:36	
Control Test		0.079	00:36	
Air Blank		0.000	00:37	
Subject Sample	# 1		00:38	Subject Test Refused
Breath Volume		0.000	00:38	
Air Blank		0.000	00:39	
Control Test		0.079	00:39	
Air Blank		0.000	00:39	
Diagnostics Check		OK	00:40	

Data Download Date/Time: 10/30/2025 17:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,LEANDRA,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: DAVIS,LEANDRA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: SERRANO MARTINEZ

DATE/TIME OF TEST: 09/02/2025 01:15:02

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.157	01:10
Breath Volume		1.570	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.169	01:13
Breath Volume		1.777	01:13
Air Blank		0.000	01:14
Control Test		0.082	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:15

Data Download Date/Time: 09/19/2025 15:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: KIGHT,BRANDON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GONZALEZ RODRIGUEZ

DATE/TIME OF TEST: 09/05/2025 03:36:35

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.080	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.074	03:32
Breath Volume		1.675	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.073	03:35
Breath Volume		1.992	03:35
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 09/19/2025 15:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,ROBERT,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WRIGHT,JAMES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: CARRILLO

DATE/TIME OF TEST: 09/07/2025 04:00:53

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.080	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.205	03:56
Breath Volume		3.128	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.200	03:59
Breath Volume		3.523	03:59
Air Blank		0.000	03:59
Control Test		0.081	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 09/19/2025 15:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELCHER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MELCHER,ANDREW
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 15:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: GILES

DATE/TIME OF TEST: 09/08/2025 15:50:26

Diagnostics Check		OK	15:40	
Air Blank		0.000	15:40	
Control Test		0.081	15:40	
Air Blank		0.000	15:41	
Subject Sample	# 1	0.116	15:44	Volume Not Met
Breath Volume		0.902	15:44	
Air Blank		0.000	15:45	
Air Blank		0.000	15:47	
Subject Sample	# 2		15:49	Sequence Aborted
Breath Volume		0.539	15:49	
Air Blank		0.000	15:50	

Data Download Date/Time: 09/19/2025 15:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,ROBERT,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WRIGHT,J
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: HOSFORD

DATE/TIME OF TEST: 09/11/2025 04:17:52

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.079	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.204	04:13
Breath Volume		2.203	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.190	04:16
Breath Volume		1.839	04:16
Air Blank		0.000	04:16
Control Test		0.080	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 09/19/2025 15:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: KIGHT,BRANDON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: CERVANTES

DATE/TIME OF TEST: 09/13/2025 21:49:46

Diagnostics Check		OK	21:42
Air Blank		0.000	21:43
Control Test		0.082	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.189	21:44
Breath Volume		1.667	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.175	21:48
Breath Volume		1.867	21:48
Air Blank		0.000	21:48
Control Test		0.080	21:49
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 09/19/2025 15:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: KIGHT,BRANDON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: DORRIETY

DATE/TIME OF TEST: 09/14/2025 04:32:25

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.080	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.102	04:27
Breath Volume		2.101	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.109	04:30
Breath Volume		1.210	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 09/19/2025 15:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,ROBERT,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WRIGHT,JAMES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ZOVLONSKY

DATE/TIME OF TEST: 09/20/2025 01:18:40

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.155	01:14
Breath Volume		2.121	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.162	01:17
Breath Volume		1.800	01:17
Air Blank		0.000	01:17
Control Test		0.079	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 10/17/2025 15:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIFT,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: SWIFT,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 06:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: RYE

DATE/TIME OF TEST: 09/21/2025 07:34:11

Diagnostics Check		OK	07:27
Air Blank		0.000	07:28
Control Test		0.080	07:28
Air Blank		0.000	07:29
Subject Sample	# 1	0.000	07:29
Breath Volume		1.695	07:29
Air Blank		0.000	07:30
Air Blank		0.000	07:32
Subject Sample	# 2	0.000	07:32
Breath Volume		1.832	07:32
Air Blank		0.000	07:33
Control Test		0.079	07:33
Air Blank		0.000	07:34
Diagnostics Check		OK	07:34

Data Download Date/Time: 10/17/2025 15:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: WILLCUTS

DATE/TIME OF TEST: 09/25/2025 22:58:34

Diagnostics Check		OK	22:52
Air Blank		0.000	22:52
Control Test		0.078	22:53
Air Blank		0.000	22:53
Subject Sample	# 1	0.097	22:54
Breath Volume		1.917	22:54
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.097	22:57
Breath Volume		1.226	22:57
Air Blank		0.000	22:57
Control Test		0.079	22:57
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 10/17/2025 15:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAVERNITI,MAX,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COZAD,ANDREIA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 09/27/2025 02:03:12

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:52	
Control Test		0.079	01:52	
Air Blank		0.000	01:52	
Subject Sample	# 1	0.182	01:56	Volume Not Met
Breath Volume		0.128	01:56	
Air Blank		0.000	01:56	
Air Blank		0.000	01:58	
Subject Sample	# 2	0.178	02:01	Volume Not Met
Breath Volume		0.105	02:01	
Air Blank		0.000	02:02	
Control Test		0.078	02:02	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 10/17/2025 15:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ARAFA,YUSUF
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 06:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: WALLER

DATE/TIME OF TEST: 09/27/2025 07:23:33

Diagnostics Check		OK	07:15
Air Blank		0.000	07:15
Control Test		0.079	07:16
Air Blank		0.000	07:16
Subject Sample	# 1	0.164	07:18
Breath Volume		1.886	07:18
Air Blank		0.000	07:19
Air Blank		0.000	07:20
Subject Sample	# 2	0.182	07:21
Breath Volume		2.093	07:21
Air Blank		0.000	07:22
Control Test		0.078	07:22
Air Blank		0.000	07:23
Diagnostics Check		OK	07:23

Data Download Date/Time: 10/17/2025 15:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COLWELL,NICHOLAS
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SHIELDS

DATE/TIME OF TEST: 09/28/2025 04:24:14

Diagnostics Check		OK	04:17
Air Blank		0.000	04:18
Control Test		0.079	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.139	04:19
Breath Volume		1.917	04:19
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.137	04:22
Breath Volume		2.039	04:22
Air Blank		0.000	04:23
Control Test		0.080	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 10/17/2025 15:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINAMORE,ABIGAIL,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MENDEZ,VIOLET
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: GARCIA GONZALEZ

DATE/TIME OF TEST: 09/06/2025 01:58:41

Diagnostics Check		OK	01:51
Air Blank		0.000	01:52
Control Test		0.080	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.159	01:53
Breath Volume		1.746	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.154	01:57
Breath Volume		2.167	01:57
Air Blank		0.000	01:57
Control Test		0.078	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CASSITY

DATE/TIME OF TEST: 09/06/2025 04:18:03

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.144	04:13
Breath Volume		1.148	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.162	04:16
Breath Volume		2.523	04:16
Air Blank		0.000	04:17
Control Test		0.079	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:18

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,HANNAH,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ACOSTA,CARLOS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: AJANEL UTUY

DATE/TIME OF TEST: 09/06/2025 21:34:02

Diagnostics Check		OK	21:27
Air Blank		0.000	21:28
Control Test		0.080	21:28
Air Blank		0.000	21:28
Subject Sample	# 1	0.186	21:29
Breath Volume		2.984	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:31
Subject Sample	# 2	0.181	21:32
Breath Volume		3.171	21:32
Air Blank		0.000	21:33
Control Test		0.078	21:33
Air Blank		0.000	21:33
Diagnostics Check		OK	21:34

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BRYANT,MATTHEW
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: LUX MEJIA

DATE/TIME OF TEST: 09/07/2025 04:42:20

Diagnostics Check		OK	04:30	
Air Blank		0.000	04:31	
Control Test		0.080	04:31	
Air Blank		0.000	04:32	
Subject Sample	# 1	0.158	04:35	Volume Not Met
Breath Volume		0.292	04:35	
Air Blank		0.000	04:35	
Air Blank		0.000	04:37	
Subject Sample	# 2	0.149	04:40	Volume Not Met
Breath Volume		0.058	04:40	
Air Blank		0.000	04:41	
Control Test		0.078	04:41	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MARTINEZ TREMINIO

DATE/TIME OF TEST: 09/08/2025 01:58:20

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.080	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.162	01:53
Breath Volume		2.593	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.169	01:56
Breath Volume		2.957	01:56
Air Blank		0.000	01:57
Control Test		0.079	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DIPASCALE,JEFF
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BRISTOL

DATE/TIME OF TEST: 09/08/2025 03:11:07

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.198	03:06
Breath Volume		2.105	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.193	03:09
Breath Volume		1.589	03:09
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ANDUJAR,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: LOPEZ DE LEON

DATE/TIME OF TEST: 09/08/2025 20:07:49

Diagnostics Check		OK	20:01
Air Blank		0.000	20:01
Control Test		0.080	20:02
Air Blank		0.000	20:02
Subject Sample	# 1	0.063	20:03
Breath Volume		3.269	20:03
Air Blank		0.000	20:03
Air Blank		0.000	20:05
Subject Sample	# 2	0.060	20:06
Breath Volume		2.945	20:06
Air Blank		0.000	20:06
Control Test		0.079	20:07
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: HUTTON

DATE/TIME OF TEST: 09/09/2025 23:33:14

Diagnostics Check		OK	23:25
Air Blank		0.000	23:26
Control Test		0.080	23:26
Air Blank		0.000	23:27
Subject Sample	# 1	0.074	23:28
Breath Volume		2.726	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.075	23:31
Breath Volume		2.449	23:31
Air Blank		0.000	23:32
Control Test		0.079	23:32
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WHITUS,O
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: TOWNE

DATE/TIME OF TEST: 09/13/2025 00:13:15

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.080	00:07
Air Blank		0.000	00:08
Subject Sample	# 1	0.114	00:08
Breath Volume		4.234	00:08
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.113	00:11
Breath Volume		4.269	00:11
Air Blank		0.000	00:12
Control Test		0.079	00:12
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 09/13/2025 00:21:08

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.171	00:16
Breath Volume		2.875	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.172	00:19
Breath Volume		2.812	00:19
Air Blank		0.000	00:20
Control Test		0.079	00:20
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANPELT,J,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WHITUS,O
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: KOSZELAK

DATE/TIME OF TEST: 09/14/2025 01:45:49

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:41	
Control Test		0.080	01:41	
Air Blank		0.000	01:42	
Subject Sample	# 1		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:44	
Control Test		0.079	01:45	
Air Blank		0.000	01:45	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASH,CONNOR,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HASH,CONNOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SAVANT

DATE/TIME OF TEST: 09/14/2025 02:57:15

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.080	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.000	02:53
Breath Volume		1.886	02:53
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.000	02:55
Breath Volume		1.957	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLUESNER,D,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ANDUJAR,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: VALENTIN

DATE/TIME OF TEST: 09/14/2025 09:37:50

Diagnostics Check		OK	09:31
Air Blank		0.000	09:31
Control Test		0.080	09:31
Air Blank		0.000	09:32
Subject Sample	# 1	0.201	09:33
Breath Volume		2.054	09:33
Air Blank		0.000	09:33
Air Blank		0.000	09:35
Subject Sample	# 2	0.205	09:36
Breath Volume		2.214	09:36
Air Blank		0.000	09:36
Control Test		0.078	09:37
Air Blank		0.000	09:37
Diagnostics Check		OK	09:37

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: SANCHEZ RACET

DATE/TIME OF TEST: 09/14/2025 20:35:37

Diagnostics Check		OK	20:29
Air Blank		0.000	20:30
Control Test		0.081	20:30
Air Blank		0.000	20:30
Subject Sample	# 1	0.099	20:31
Breath Volume		1.562	20:31
Air Blank		0.000	20:31
Air Blank		0.000	20:33
Subject Sample	# 2	0.100	20:34
Breath Volume		2.074	20:34
Air Blank		0.000	20:34
Control Test		0.079	20:35
Air Blank		0.000	20:35
Diagnostics Check		OK	20:35

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,JOSHUA,E
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: NORRIS,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MOLINA DIAZ

DATE/TIME OF TEST: 09/19/2025 00:32:53

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.000	00:28
Breath Volume		3.394	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.000	00:31
Breath Volume		3.210	00:31
Air Blank		0.000	00:31
Control Test		0.078	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: SANTOS PARADA

DATE/TIME OF TEST: 09/19/2025 01:16:08

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.134	01:11
Breath Volume		1.109	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:13
Subject Sample	# 2	0.141	01:14
Breath Volume		1.437	01:14
Air Blank		0.000	01:15
Control Test		0.079	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/20/2025 00:14:15

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.079	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.161	00:09
Breath Volume		2.507	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.161	00:12
Breath Volume		2.542	00:12
Air Blank		0.000	00:13
Control Test		0.080	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: TIU OSORIO

DATE/TIME OF TEST: 09/20/2025 02:01:37

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.080	01:55
Air Blank		0.000	01:56
Subject Sample	# 1	0.124	01:56
Breath Volume		2.074	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.126	02:00
Breath Volume		2.453	02:00
Air Blank		0.000	02:00
Control Test		0.079	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: GARCIA REYES

DATE/TIME OF TEST: 09/20/2025 21:52:05

Diagnostics Check		OK	21:45
Air Blank		0.000	21:46
Control Test		0.080	21:46
Air Blank		0.000	21:47
Subject Sample	# 1	0.207	21:47
Breath Volume		1.257	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.222	21:50
Breath Volume		1.511	21:50
Air Blank		0.000	21:51
Control Test		0.079	21:51
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 09/22/2025 10:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRARO,PAUL,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH,G
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: LIUBAR

DATE/TIME OF TEST: 09/21/2025 03:01:05

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:57	
Control Test		0.080	02:57	
Air Blank		0.000	02:57	
Subject Sample	# 1		02:59	Subject Test Refused
Breath Volume		0.000	02:59	
Air Blank		0.000	03:00	
Control Test		0.080	03:00	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 09/22/2025 10:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JOEL,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KARSNICK,KEVIN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: VILLANUEVA AVILES

DATE/TIME OF TEST: 09/21/2025 04:24:05

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.080	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.133	04:19
Breath Volume		1.542	04:19
Air Blank		0.000	04:20
Air Blank		0.000	04:21
Subject Sample	# 2	0.135	04:22
Breath Volume		1.394	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:23
Diagnostics Check		OK	04:24

Data Download Date/Time: 09/22/2025 10:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,NATHANIEL,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DRUMMOND,FLOYD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A3
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:43

SUBJECT'S LAST NAME: RODRIQUEZ

DATE/TIME OF TEST: 09/22/2025 13:12:22

Diagnostics Check		OK	13:04
Air Blank		0.000	13:04
Control Test		0.079	13:05
Air Blank		0.000	13:05
Subject Sample	# 1	0.080	13:07
Breath Volume		1.839	13:07
Air Blank		0.000	13:07
Air Blank		0.000	13:09
Subject Sample	# 2	0.080	13:10
Breath Volume		1.980	13:10
Air Blank		0.000	13:11
Control Test		0.078	13:11
Air Blank		0.000	13:12
Diagnostics Check		OK	13:12

Data Download Date/Time: 10/21/2025 11:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A3
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: LOBO

DATE/TIME OF TEST: 09/22/2025 18:40:41

Diagnostics Check		OK	18:34
Air Blank		0.000	18:34
Control Test		0.079	18:35
Air Blank		0.000	18:35
Subject Sample	# 1	0.274	18:36
Breath Volume		2.003	18:36
Air Blank		0.000	18:36
Air Blank		0.000	18:38
Subject Sample	# 2	0.285	18:39
Breath Volume		3.054	18:39
Air Blank		0.000	18:39
Control Test		0.078	18:40
Air Blank		0.000	18:40
Diagnostics Check		OK	18:40

Data Download Date/Time: 10/21/2025 11:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHIAPPY,CHRISTOFER,J
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: SMITH,GAVIN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A3
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: CORNELIO DE LA CRUZ

DATE/TIME OF TEST: 09/26/2025 00:15:18

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.079	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.119	00:10
Breath Volume		2.625	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.122	00:13
Breath Volume		2.453	00:13
Air Blank		0.000	00:14
Control Test		0.078	00:14
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 10/21/2025 11:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BENITEZ,CARLOS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A3
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/27/2025 03:31:46

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.079	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.196	03:27
Breath Volume		2.050	03:27
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.203	03:30
Breath Volume		2.906	03:30
Air Blank		0.000	03:30
Control Test		0.079	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 10/21/2025 11:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A3
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 09/28/2025 06:24:36

Diagnostics Check		OK	06:18	
Air Blank		0.000	06:19	
Control Test		0.079	06:19	
Air Blank		0.000	06:20	
Subject Sample	# 1	0.160	06:20	
Breath Volume		2.847	06:20	
Air Blank		0.000	06:21	
Air Blank		0.000	06:22	
			06:23	Improper Sample
Air Blank		0.000	06:23	
Control Test		0.079	06:24	
Air Blank		0.000	06:24	
Diagnostics Check		OK	06:24	

Data Download Date/Time: 10/21/2025 11:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A3
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 09/28/2025 06:32:25

Diagnostics Check		OK	06:26
Air Blank		0.000	06:26
Control Test		0.079	06:27
Air Blank		0.000	06:27
Subject Sample	# 1	0.156	06:28
Breath Volume		3.171	06:28
Air Blank		0.000	06:28
Air Blank		0.000	06:30
Subject Sample	# 2	0.146	06:30
Breath Volume		2.824	06:30
Air Blank		0.000	06:31
Control Test		0.079	06:31
Air Blank		0.000	06:32
Diagnostics Check		OK	06:32

Data Download Date/Time: 10/21/2025 11:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JOEL,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ACOSTA,CARLOS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A3
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: FRANCISCO ESTEBAN

DATE/TIME OF TEST: 09/30/2025 23:52:18

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.079	23:46
Air Blank		0.000	23:46
Subject Sample	# 1	0.158	23:47
Breath Volume		3.335	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.158	23:50
Breath Volume		2.542	23:50
Air Blank		0.000	23:51
Control Test		0.077	23:51
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 10/21/2025 11:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSTIN,AMANDA,N
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: CLARK,TYLER
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: JUSTIN

DATE/TIME OF TEST: 09/21/2025 01:59:55

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.080	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.000	01:55
Breath Volume		1.371	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.000	01:58
Breath Volume		1.363	01:58
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 10/14/2025 14:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 08/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 16:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: CONSTANZA LOPEZ

DATE/TIME OF TEST: 09/02/2025 16:50:43

Diagnostics Check		OK	16:46	
Air Blank		0.000	16:47	
Control Test		0.080	16:47	
Air Blank		0.000	16:47	
Subject Sample	# 1		16:49	Subject Test Refused
Breath Volume		0.000	16:49	
Air Blank		0.000	16:49	
Control Test		0.081	16:50	
Air Blank		0.000	16:50	
Diagnostics Check		OK	16:50	

Data Download Date/Time: 09/30/2025 20:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,ALINA,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MENDEZ,ALINA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 08/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: LEIZEAR

DATE/TIME OF TEST: 09/12/2025 00:03:08

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:00	
Control Test		0.080	00:00	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:01	Subject Test Refused
Breath Volume		0.000	00:01	
Air Blank		0.000	00:02	
Control Test		0.080	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 09/30/2025 20:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,ALINA,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MENDEZ,ALINA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 08/30/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GONZALEZ PEREZ

DATE/TIME OF TEST: 09/27/2025 03:14:33

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:05	
Control Test		0.081	03:06	
Air Blank		0.000	03:06	
Subject Sample	# 1	0.152	03:09	Volume Not Met
Breath Volume		1.078	03:09	
Air Blank		0.000	03:10	
Air Blank		0.000	03:12	
Subject Sample	# 2		03:13	Subject Test Refused
Breath Volume		0.015	03:13	
Air Blank		0.000	03:13	
Control Test		0.079	03:13	
Air Blank		0.000	03:14	
Diagnostics Check		OK	03:14	

Data Download Date/Time: 09/30/2025 20:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,ALINA,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MENDEZ,ALINA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 08/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 09/28/2025 01:59:20

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:56	
Control Test		0.081	01:57	
Air Blank		0.000	01:57	
Subject Sample	# 1		01:57	Subject Test Refused
Breath Volume		0.000	01:57	
Air Blank		0.000	01:58	
Control Test		0.081	01:58	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 09/30/2025 20:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KYLE,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN,DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: KAZEK

DATE/TIME OF TEST: 09/19/2025 16:49:17

Diagnostics Check		OK	16:42
Air Blank		0.000	16:42
Control Test		0.080	16:43
Air Blank		0.000	16:43
Subject Sample	# 1	0.196	16:44
Breath Volume		3.769	16:44
Air Blank		0.000	16:45
Air Blank		0.000	16:46
Subject Sample	# 2	0.204	16:47
Breath Volume		3.812	16:47
Air Blank		0.000	16:48
Control Test		0.079	16:48
Air Blank		0.000	16:49
Diagnostics Check		OK	16:49

Data Download Date/Time: 10/27/2025 11:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: FELICIANO,JAIME
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 08/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: DOUGAN

DATE/TIME OF TEST: 09/04/2025 20:14:08

Diagnostics Check		OK	20:06
Air Blank		0.000	20:06
Control Test		0.077	20:07
Air Blank		0.000	20:07
Subject Sample	# 1	0.201	20:09
Breath Volume		1.988	20:09
Air Blank		0.000	20:10
Air Blank		0.000	20:11
Subject Sample	# 2	0.211	20:12
Breath Volume		2.132	20:12
Air Blank		0.000	20:13
Control Test		0.075	20:13
Air Blank		0.000	20:14
Diagnostics Check		OK	20:14

Data Download Date/Time: 09/25/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWEENEY,PARKER,L
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: CASTELANO,VINCENT
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 08/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: DYER

DATE/TIME OF TEST: 09/06/2025 00:53:23

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:50	
Control Test		0.076	00:51	
Air Blank		0.000	00:51	
Subject Sample	# 1		00:52	Subject Test Refused
Breath Volume		0.000	00:52	
Air Blank		0.000	00:52	
Control Test		0.076	00:52	
Air Blank		0.000	00:53	
Diagnostics Check		OK	00:53	

Data Download Date/Time: 09/25/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE,ADAM,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: RAINGE,MARIAH
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/08/2025 00:38:36

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.076	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.185	00:33
Breath Volume		3.433	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.174	00:36
Breath Volume		3.093	00:36
Air Blank		0.000	00:37
Control Test		0.076	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 09/25/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE,ADAM,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: LIGHTFOOT,BRITTANY
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: HAMMOND

DATE/TIME OF TEST: 09/12/2025 22:28:20

Diagnostics Check		OK	22:22
Air Blank		0.000	22:22
Control Test		0.077	22:22
Air Blank		0.000	22:23
Subject Sample	# 1	0.094	22:23
Breath Volume		1.488	22:23
Air Blank		0.000	22:24
Air Blank		0.000	22:26
Subject Sample	# 2	0.092	22:26
Breath Volume		1.496	22:26
Air Blank		0.000	22:27
Control Test		0.077	22:27
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 09/25/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE,ADAM,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: SMITH,JESSEE
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 08/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/13/2025 20:55:06

Diagnostics Check		OK	20:48
Air Blank		0.000	20:49
Control Test		0.076	20:49
Air Blank		0.000	20:49
Subject Sample	# 1	0.208	20:50
Breath Volume		2.277	20:50
Air Blank		0.000	20:51
Air Blank		0.000	20:52
Subject Sample	# 2	0.211	20:53
Breath Volume		1.820	20:53
Air Blank		0.000	20:54
Control Test		0.076	20:54
Air Blank		0.000	20:55
Diagnostics Check		OK	20:55

Data Download Date/Time: 09/25/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE,ADAM,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: GUTHRIE,ADAM
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 08/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 09/13/2025 22:41:25

Diagnostics Check		OK	22:34
Air Blank		0.000	22:35
Control Test		0.076	22:35
Air Blank		0.000	22:36
Subject Sample	# 1	0.283	22:36
Breath Volume		3.718	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.278	22:39
Breath Volume		3.082	22:39
Air Blank		0.000	22:40
Control Test		0.076	22:40
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 09/25/2025 09:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE,ADAM,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: BIGELOW,PAUL
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 09/30/2025 22:10:41

Diagnostics Check		OK	22:03
Air Blank		0.000	22:03
Control Test		0.076	22:04
Air Blank		0.000	22:04
Subject Sample	# 1	0.229	22:05
Breath Volume		1.253	22:05
Air Blank		0.000	22:06
Air Blank		0.000	22:07
Subject Sample	# 2	0.232	22:09
Breath Volume		1.648	22:09
Air Blank		0.000	22:09
Control Test		0.076	22:10
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 10/31/2025 09:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SASSO,SCOTT,A
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: VARNER,ETHAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403096062
DRY GAS EXPIRY DATE: 07/15/2027
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/16/2025 02:20:10

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.081	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.000	02:15
Breath Volume		1.394	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.000	02:18
Breath Volume		1.167	02:18
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 09/24/2025 17:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SASSO,SCOTT,A
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: VARNER,ETHAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403096062
DRY GAS EXPIRY DATE: 07/15/2027
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: MEHLER

DATE/TIME OF TEST: 09/22/2025 00:49:57

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.081	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.150	00:44
Breath Volume		1.835	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.147	00:48
Breath Volume		1.714	00:48
Air Blank		0.000	00:49
Control Test		0.079	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 09/24/2025 17:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSENDEN,J,
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: CIESLAK,A
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 09/28/2025 02:39:55

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.087	02:33
Breath Volume		1.152	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.085	02:38
Breath Volume		1.160	02:38
Air Blank		0.000	02:39
Control Test		0.081	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 10/23/2025 03:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINICIS,LOURDES,P
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DOMINICIS,LOURDES
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 09/04/2025 03:39:54

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.081	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.156	03:35
Breath Volume		2.652	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.152	03:38
Breath Volume		2.269	03:38
Air Blank		0.000	03:39
Control Test		0.082	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 09/30/2025 08:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARISSA,H
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: PILL,JEREMY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:54

SUBJECT'S LAST NAME: O'GRADY

DATE/TIME OF TEST: 09/08/2025 15:21:35

Diagnostics Check		OK	15:14
Air Blank		0.000	15:15
Control Test		0.082	15:15
Air Blank		0.000	15:16
Subject Sample	# 1	0.000	15:17
Breath Volume		1.484	15:17
Air Blank		0.000	15:17
Air Blank		0.000	15:19
Subject Sample	# 2	0.000	15:20
Breath Volume		1.328	15:20
Air Blank		0.000	15:20
Control Test		0.080	15:21
Air Blank		0.000	15:21
Diagnostics Check		OK	15:21

Data Download Date/Time: 09/30/2025 08:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAROCHE,CHARLES,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: OSTEEN,ASHLEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: OZEROVA HERPICH

DATE/TIME OF TEST: 09/16/2025 17:39:00

Diagnostics Check		OK	17:32	
Air Blank		0.000	17:32	
Control Test		0.081	17:33	
Air Blank		0.000	17:33	
Subject Sample	# 1	0.202	17:34	
Breath Volume		2.089	17:34	
Air Blank		0.000	17:35	
Air Blank		0.000	17:37	
Subject Sample	# 2		17:37	RFI Detect
Breath Volume		0.363	17:37	
Air Blank		0.000	17:38	
Control Test		0.080	17:38	
Air Blank		0.000	17:38	
Diagnostics Check		OK	17:39	

Data Download Date/Time: 09/30/2025 08:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAROCHE,CHARLES,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: OSTEEN,ASHLEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: OZEROVA HERPICH

DATE/TIME OF TEST: 09/16/2025 17:48:56

Diagnostics Check		OK	17:42
Air Blank		0.000	17:43
Control Test		0.081	17:43
Air Blank		0.000	17:43
Subject Sample	# 1	0.205	17:44
Breath Volume		1.679	17:44
Air Blank		0.000	17:45
Air Blank		0.000	17:47
Subject Sample	# 2	0.211	17:47
Breath Volume		1.531	17:47
Air Blank		0.000	17:48
Control Test		0.079	17:48
Air Blank		0.000	17:48
Diagnostics Check		OK	17:48

Data Download Date/Time: 09/30/2025 08:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIL,JOSE,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GIL,JOSE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 09/01/2025 03:48:07

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:45	
Control Test		0.082	03:45	
Air Blank		0.000	03:46	
Subject Sample	# 1		03:46	Subject Test Refused
Breath Volume		0.000	03:46	
Air Blank		0.000	03:47	
Control Test		0.081	03:47	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 09/16/2025 17:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRAW,KAELEE,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GUERRERO-MARTINEZ,KEVEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: MENNEMEIER

DATE/TIME OF TEST: 09/06/2025 04:29:41

Diagnostics Check		OK	04:26	
Air Blank		0.000	04:27	
Control Test		0.081	04:27	
Air Blank		0.000	04:28	
Subject Sample	# 1		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:28	
Control Test		0.082	04:29	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 09/16/2025 17:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: FERO FUENTES

DATE/TIME OF TEST: 09/12/2025 21:51:15

Diagnostics Check		OK	21:44
Air Blank		0.000	21:44
Control Test		0.081	21:45
Air Blank		0.000	21:45
Subject Sample	# 1	0.147	21:46
Breath Volume		2.285	21:46
Air Blank		0.000	21:46
Air Blank		0.000	21:48
Subject Sample	# 2	0.143	21:49
Breath Volume		2.914	21:49
Air Blank		0.000	21:50
Control Test		0.081	21:50
Air Blank		0.000	21:51
Diagnostics Check		OK	21:51

Data Download Date/Time: 09/16/2025 17:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRAW,KAELEE,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GUERRERO-MARTINEZ,KEVEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: RIEHL

DATE/TIME OF TEST: 09/20/2025 00:27:27

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:25	
Control Test		0.081	00:25	
Air Blank		0.000	00:25	
Subject Sample	# 1		00:26	Subject Test Refused
Breath Volume		0.000	00:26	
Air Blank		0.000	00:26	
Control Test		0.081	00:26	
Air Blank		0.000	00:27	
Diagnostics Check		OK	00:27	

Data Download Date/Time: 10/17/2025 11:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRERO-MARTINEZ,KEVEN,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GUERRERO-MARTINEZ,KEVEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: JIMENEZ LEON

DATE/TIME OF TEST: 09/21/2025 03:01:36

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:59	
Control Test		0.082	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:00	Subject Test Refused
Breath Volume		0.000	03:00	
Air Blank		0.000	03:00	
Control Test		0.081	03:01	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 10/17/2025 11:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRERO-MARTINEZ,KEVEN,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GUERRERO-MARTINEZ,KEVEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: GARLAND

DATE/TIME OF TEST: 09/25/2025 22:50:46

Diagnostics Check		OK	22:47	
Air Blank		0.000	22:48	
Control Test		0.081	22:48	
Air Blank		0.000	22:49	
Subject Sample	# 1		22:49	Subject Test Refused
Breath Volume		0.000	22:49	
Air Blank		0.000	22:49	
Control Test		0.080	22:50	
Air Blank		0.000	22:50	
Diagnostics Check		OK	22:50	

Data Download Date/Time: 10/17/2025 11:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRERO-MARTINEZ,KEVEN,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GUERRERO-MARTINEZ,KEVEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: DE ARMAS MONTERO

DATE/TIME OF TEST: 09/26/2025 02:59:06

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:56	
Control Test		0.083	02:57	
Air Blank		0.000	02:57	
Subject Sample	# 1		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:58	
Control Test		0.082	02:58	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 10/17/2025 11:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTER,PARIS,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: WINTER,PARIS
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: BUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: FALKENSTEIN

DATE/TIME OF TEST: 09/26/2025 21:52:02

Diagnostics Check		OK	21:45
Air Blank		0.000	21:45
Control Test		0.081	21:46
Air Blank		0.000	21:46
Subject Sample	# 1	0.058	21:47
Breath Volume		1.808	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.051	21:50
Breath Volume		1.121	21:50
Air Blank		0.000	21:51
Control Test		0.082	21:51
Air Blank		0.000	21:51
Diagnostics Check		OK	21:52

Data Download Date/Time: 10/17/2025 11:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: NEWTON-GATLING

DATE/TIME OF TEST: 09/27/2025 03:34:56

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.082	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.139	03:30
Breath Volume		1.515	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.137	03:33
Breath Volume		1.867	03:33
Air Blank		0.000	03:34
Control Test		0.081	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 10/17/2025 11:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: PULSFORD

DATE/TIME OF TEST: 09/28/2025 03:56:41

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:44	
Control Test		0.081	03:44	
Air Blank		0.000	03:45	
Subject Sample	# 1	0.220	03:46	No .020 agreement
Breath Volume		1.746	03:46	
Air Blank		0.000	03:47	
Air Blank		0.000	03:48	
Subject Sample	# 2	0.195	03:49	No .020 agreement
Breath Volume		1.402	03:49	
Air Blank		0.000	03:50	
Air Blank		0.000	03:52	
Subject Sample	# 3	0.164	03:55	Volume Not Met
Breath Volume		0.753	03:55	
Air Blank		0.000	03:55	
Control Test		0.081	03:56	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 10/17/2025 11:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: PULSFORD

DATE/TIME OF TEST: 09/28/2025 04:12:30

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:58	
Control Test		0.079	03:59	
Air Blank		0.000	03:59	
Subject Sample	# 1	0.206	04:02	Volume Not Met
Breath Volume		0.699	04:02	
Air Blank		0.000	04:03	
Air Blank		0.000	04:05	
Subject Sample	# 2	0.208	04:07	
Breath Volume		1.664	04:07	
Air Blank		0.000	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 3	0.210	04:10	
Breath Volume		3.542	04:10	
Air Blank		0.000	04:11	
Control Test		0.081	04:11	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:12	

Data Download Date/Time: 10/17/2025 11:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARRERO,GABRIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MENDEZ RIOS

DATE/TIME OF TEST: 09/01/2025 03:10:59

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.082	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.250	03:05
Breath Volume		1.347	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.256	03:09
Breath Volume		1.433	03:09
Air Blank		0.000	03:10
Control Test		0.082	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 09/01/2025 12:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARUNA,HEATHER
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 09/01/2025 22:01:02

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.083	21:55
Air Blank		0.000	21:55
Subject Sample	# 1	0.174	21:56
Breath Volume		2.593	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:58
Subject Sample	# 2	0.173	21:59
Breath Volume		1.808	21:59
Air Blank		0.000	22:00
Control Test		0.081	22:00
Air Blank		0.000	22:00
Diagnostics Check		OK	22:01

Data Download Date/Time: 10/16/2025 11:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA,AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: ALEXANDRE

DATE/TIME OF TEST: 09/01/2025 22:55:32

Diagnostics Check		OK	22:48
Air Blank		0.000	22:48
Control Test		0.082	22:49
Air Blank		0.000	22:49
Subject Sample	# 1	0.187	22:50
Breath Volume		1.257	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.184	22:53
Breath Volume		1.539	22:53
Air Blank		0.000	22:54
Control Test		0.081	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 10/16/2025 11:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIOS,JOSHUA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: THOMSON

DATE/TIME OF TEST: 09/02/2025 01:07:46

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.082	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.139	01:02
Breath Volume		1.628	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.130	01:06
Breath Volume		1.949	01:06
Air Blank		0.000	01:06
Control Test		0.082	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 10/16/2025 11:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRUITT,JAMES
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BARBOSACONCEICAO

DATE/TIME OF TEST: 09/02/2025 02:39:56

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:28	
Control Test		0.082	02:29	
Air Blank		0.000	02:29	
Subject Sample	# 1		02:32	No Sample Provided
Breath Volume		0.000	02:32	
Air Blank		0.000	02:33	
Air Blank		0.000	02:35	
Subject Sample	# 2		02:38	No Sample Provided
Breath Volume		0.000	02:38	
Air Blank		0.000	02:39	
Control Test		0.083	02:39	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 10/16/2025 11:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBBINS,SCARLET
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: KOLLAR

DATE/TIME OF TEST: 09/02/2025 05:29:04

Diagnostics Check		OK	05:21
Air Blank		0.000	05:21
Control Test		0.082	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.250	05:24
Breath Volume		1.921	05:24
Air Blank		0.000	05:25
Air Blank		0.000	05:26
Subject Sample	# 2	0.250	05:27
Breath Volume		1.617	05:27
Air Blank		0.000	05:28
Control Test		0.079	05:28
Air Blank		0.000	05:28
Diagnostics Check		OK	05:29

Data Download Date/Time: 10/16/2025 11:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRUITT,JAMES
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: VALLECILLOCASTILLO

DATE/TIME OF TEST: 09/02/2025 21:35:30

Diagnostics Check		OK	21:29
Air Blank		0.000	21:29
Control Test		0.083	21:30
Air Blank		0.000	21:30
Subject Sample	# 1	0.035	21:31
Breath Volume		2.019	21:31
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.034	21:33
Breath Volume		2.046	21:33
Air Blank		0.000	21:34
Control Test		0.082	21:34
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 10/16/2025 11:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COLON,ALBERT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: HERNANDEZPINEDA

DATE/TIME OF TEST: 09/03/2025 22:54:12

Diagnostics Check		OK	22:46
Air Blank		0.000	22:47
Control Test		0.082	22:47
Air Blank		0.000	22:47
Subject Sample	# 1	0.128	22:49
Breath Volume		2.781	22:49
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.122	22:52
Breath Volume		2.863	22:52
Air Blank		0.000	22:53
Control Test		0.080	22:53
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 10/16/2025 11:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BIRNBAUM,DAVID
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 09/05/2025 23:39:24

Diagnostics Check		OK	23:33
Air Blank		0.000	23:33
Control Test		0.083	23:33
Air Blank		0.000	23:34
Subject Sample	# 1	0.193	23:34
Breath Volume		2.386	23:34
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.195	23:37
Breath Volume		2.394	23:37
Air Blank		0.000	23:38
Control Test		0.082	23:38
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 10/16/2025 11:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: YIM,RANDY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: WARREN JR

DATE/TIME OF TEST: 09/06/2025 00:46:11

Diagnostics Check		OK	00:43	
Air Blank		0.000	00:43	
Control Test		0.083	00:44	
Air Blank		0.000	00:44	
Subject Sample	# 1		00:44	Subject Test Refused
Breath Volume		0.000	00:44	
Air Blank		0.000	00:45	
Control Test		0.082	00:45	
Air Blank		0.000	00:46	
Diagnostics Check		OK	00:46	

Data Download Date/Time: 10/16/2025 11:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTILLO,SERGIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/06/2025 02:06:05

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:55	
Control Test		0.083	01:55	
Air Blank		0.000	01:55	
Subject Sample	# 1		01:59	No Sample Provided
Breath Volume		0.000	01:59	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 2		02:04	No Sample Provided
Breath Volume		0.000	02:04	
Air Blank		0.000	02:05	
Control Test		0.083	02:05	
Air Blank		0.000	02:05	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 10/16/2025 11:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSA,BRANDON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MCGARITY

DATE/TIME OF TEST: 09/07/2025 02:30:22

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.083	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.180	02:25
Breath Volume		1.644	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.183	02:28
Breath Volume		2.046	02:28
Air Blank		0.000	02:29
Control Test		0.081	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 10/16/2025 11:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SILVA,LUIS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ESTRADA AYALA

DATE/TIME OF TEST: 09/07/2025 02:52:51

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.082	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.170	02:48
Breath Volume		2.234	02:48
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.168	02:51
Breath Volume		1.710	02:51
Air Blank		0.000	02:51
Control Test		0.080	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 10/16/2025 11:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PERKINS,DEVANTI
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: TIRADO

DATE/TIME OF TEST: 09/07/2025 05:21:55

Diagnostics Check		OK	05:14
Air Blank		0.000	05:15
Control Test		0.082	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.244	05:16
Breath Volume		1.457	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.244	05:20
Breath Volume		1.753	05:20
Air Blank		0.000	05:20
Control Test		0.081	05:21
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

Data Download Date/Time: 10/16/2025 11:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAENGCHOTE,PHANUDEJ
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: BARRUETAREY

DATE/TIME OF TEST: 09/07/2025 08:04:26

Diagnostics Check		OK	07:54
Air Blank		0.000	07:55
Control Test		0.083	07:55
Air Blank		0.000	07:56
Subject Sample	# 1	0.126	07:57
Breath Volume		1.144	07:57
Air Blank		0.000	07:58
Air Blank		0.000	08:00
Subject Sample	# 2	0.130	08:02
Breath Volume		1.265	08:02
Air Blank		0.000	08:03
Control Test		0.081	08:03
Air Blank		0.000	08:04
Diagnostics Check		OK	08:04

Data Download Date/Time: 10/16/2025 11:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BIRNBAUM,DAVID
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: LARRIEUX

DATE/TIME OF TEST: 09/07/2025 21:15:55

Diagnostics Check		OK	21:06
Air Blank		0.000	21:06
Control Test		0.083	21:06
Air Blank		0.000	21:07
Subject Sample	# 1	0.295	21:08
Breath Volume		1.625	21:08
Air Blank		0.000	21:09
Air Blank		0.000	21:10
Subject Sample	# 2	0.288	21:13
Breath Volume		2.214	21:13
Air Blank		0.000	21:14
Control Test		0.080	21:15
Air Blank		0.000	21:15
Diagnostics Check		OK	21:15

Data Download Date/Time: 10/16/2025 11:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DEVOSS,EZEKIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 09/08/2025 01:10:27

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:05	
Control Test		0.082	01:05	
Air Blank		0.000	01:06	
Subject Sample	# 1	0.255	01:09	Volume Not Met
Breath Volume		0.015	01:09	
Air Blank			01:09	Purge Fail
Air Blank		0.000	01:10	

Data Download Date/Time: 10/16/2025 11:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DEVOSS,EZEKIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 09/08/2025 01:21:43

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.077	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.249	01:15
Breath Volume		1.222	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:17
Subject Sample	# 2	0.252	01:19
Breath Volume		1.914	01:19
Air Blank		0.000	01:20
Control Test		0.077	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 10/16/2025 11:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GRISALES,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: MEJIAS

DATE/TIME OF TEST: 09/08/2025 02:22:39

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.172	02:18
Breath Volume		1.515	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.172	02:20
Breath Volume		1.292	02:20
Air Blank		0.000	02:21
Control Test		0.078	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 10/16/2025 11:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUFKIN,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: EDWARDS LING

DATE/TIME OF TEST: 09/08/2025 02:50:58

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.081	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.176	02:45
Breath Volume		2.054	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:47
Subject Sample	# 2	0.161	02:49
Breath Volume		1.332	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 10/16/2025 11:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,PEDRO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: ELDRIDGE

DATE/TIME OF TEST: 09/08/2025 04:50:12

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.081	04:44
Air Blank		0.000	04:44
Subject Sample	# 1	0.280	04:45
Breath Volume		1.410	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:47
Subject Sample	# 2	0.277	04:48
Breath Volume		1.195	04:48
Air Blank		0.000	04:49
Control Test		0.079	04:49
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 10/16/2025 11:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIXSON,QUENTERIA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: BILLIE

DATE/TIME OF TEST: 09/09/2025 02:14:15

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.082	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.217	02:09
Breath Volume		2.261	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:11
Subject Sample	# 2	0.214	02:12
Breath Volume		2.531	02:12
Air Blank		0.000	02:13
Control Test		0.082	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 10/16/2025 11:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA,AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/10/2025 23:35:13

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.083	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.000	23:30
Breath Volume		1.761	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.000	23:33
Breath Volume		2.832	23:33
Air Blank		0.000	23:34
Control Test		0.081	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 10/16/2025 11:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: YIM,RANDY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: CARIAS POLANCO

DATE/TIME OF TEST: 09/12/2025 00:38:22

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.083	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.119	00:33
Breath Volume		1.445	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.113	00:36
Breath Volume		1.464	00:36
Air Blank		0.000	00:37
Control Test		0.083	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 10/16/2025 11:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALTINERLIELMAS,ADEM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MOORE II

DATE/TIME OF TEST: 09/12/2025 01:31:44

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.082	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.018	01:27
Breath Volume		2.082	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.015	01:30
Breath Volume		1.628	01:30
Air Blank		0.000	01:30
Control Test		0.081	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GAUTAM,RAYNARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: GARCIA MORALES

DATE/TIME OF TEST: 09/12/2025 03:08:36

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.082	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.131	03:03
Breath Volume		1.167	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.131	03:06
Breath Volume		1.398	03:06
Air Blank		0.000	03:07
Control Test		0.081	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHABAZZ,ELIJAH
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: DIAZ ROSARIO

DATE/TIME OF TEST: 09/12/2025 19:51:55

Diagnostics Check		OK	19:40	
Air Blank		0.000	19:40	
Control Test		0.082	19:40	
Air Blank		0.000	19:41	
Subject Sample	# 1	0.103	19:44	Volume Not Met
Breath Volume		0.031	19:44	
Air Blank		0.000	19:45	
Air Blank		0.000	19:47	
Subject Sample	# 2	0.106	19:50	Volume Not Met
Breath Volume		0.070	19:50	
Air Blank		0.000	19:50	
Control Test		0.080	19:51	
Air Blank		0.000	19:51	
Diagnostics Check		OK	19:51	

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: MARTIN FABIAN

DATE/TIME OF TEST: 09/12/2025 23:09:53

Diagnostics Check		OK	23:03
Air Blank		0.000	23:04
Control Test		0.083	23:04
Air Blank		0.000	23:04
Subject Sample	# 1	0.182	23:05
Breath Volume		2.957	23:05
Air Blank		0.000	23:06
Air Blank		0.000	23:07
Subject Sample	# 2	0.172	23:08
Breath Volume		3.183	23:08
Air Blank		0.000	23:08
Control Test		0.081	23:09
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FOGARTY,FRANCIS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 09/14/2025 00:55:13

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:52	
Control Test		0.083	00:52	
Air Blank		0.000	00:53	
Subject Sample	# 1		00:53	Slope Not Met
Breath Volume		0.570	00:53	
Air Blank		0.000	00:54	
Control Test		0.082	00:54	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FOGARTY,FRANCIS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 09/14/2025 01:31:45

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.083	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.211	01:27
Breath Volume		1.214	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.199	01:30
Breath Volume		2.328	01:30
Air Blank		0.000	01:30
Control Test		0.082	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATOS,LUIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: RODRIGUEZ PALOMARES

DATE/TIME OF TEST: 09/14/2025 02:27:43

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.083	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.176	02:23
Breath Volume		2.257	02:23
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.175	02:25
Breath Volume		3.328	02:25
Air Blank		0.000	02:26
Control Test		0.082	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BURNETT,STEVEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: RICO

DATE/TIME OF TEST: 09/14/2025 09:37:05

Diagnostics Check		OK	09:29
Air Blank		0.000	09:30
Control Test		0.083	09:30
Air Blank		0.000	09:31
Subject Sample	# 1	0.146	09:31
Breath Volume		1.753	09:31
Air Blank		0.000	09:32
Air Blank		0.000	09:34
Subject Sample	# 2	0.134	09:35
Breath Volume		1.117	09:35
Air Blank		0.000	09:36
Control Test		0.081	09:36
Air Blank		0.000	09:37
Diagnostics Check		OK	09:37

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORDOVA,ODELIA
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: RAMIREZ FARCIET

DATE/TIME OF TEST: 09/14/2025 18:35:56

Diagnostics Check		OK	18:28	
Air Blank		0.000	18:29	
Control Test		0.082	18:29	
Air Blank		0.000	18:29	
Subject Sample	# 1	0.168	18:30	
Breath Volume		1.718	18:30	
Air Blank		0.000	18:31	
Air Blank		0.000	18:33	
Subject Sample	# 2		18:34	Subject Test Refused
Breath Volume		0.000	18:34	
Air Blank		0.000	18:34	
Control Test		0.080	18:35	
Air Blank		0.000	18:35	
Diagnostics Check		OK	18:35	

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIVINGSTON,TARIQUE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: CAVALCANTI

DATE/TIME OF TEST: 09/14/2025 19:20:01

Diagnostics Check		OK	19:12
Air Blank		0.000	19:12
Control Test		0.083	19:13
Air Blank		0.000	19:13
Subject Sample	# 1	0.138	19:15
Breath Volume		2.082	19:15
Air Blank		0.000	19:16
Air Blank		0.000	19:17
Subject Sample	# 2	0.143	19:18
Breath Volume		1.292	19:18
Air Blank		0.000	19:19
Control Test		0.079	19:19
Air Blank		0.000	19:19
Diagnostics Check		OK	19:20

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: BAMACA VASQUEZ

DATE/TIME OF TEST: 09/14/2025 22:30:43

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:18	
Control Test		0.083	22:18	
Air Blank		0.000	22:19	
Subject Sample	# 1	0.267	22:22	Volume Not Met
Breath Volume		0.257	22:22	
Air Blank		0.000	22:23	
Air Blank		0.000	22:25	
Subject Sample	# 2	0.263	22:25	
Breath Volume		1.406	22:25	
Air Blank		0.000	22:26	
Air Blank		0.000	22:28	
Subject Sample	# 3	0.254	22:28	
Breath Volume		1.984	22:28	
Air Blank		0.000	22:29	
Control Test		0.079	22:30	
Air Blank		0.000	22:30	
Diagnostics Check		OK	22:30	

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TURNER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: BOTERO

DATE/TIME OF TEST: 09/15/2025 00:48:09

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.082	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.207	00:43
Breath Volume		2.066	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:45
Subject Sample	# 2	0.209	00:46
Breath Volume		2.062	00:46
Air Blank		0.000	00:47
Control Test		0.081	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: THOMAS JR

DATE/TIME OF TEST: 09/15/2025 03:39:54

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.082	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.000	03:35
Breath Volume		1.843	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.000	03:38
Breath Volume		1.457	03:38
Air Blank		0.000	03:38
Control Test		0.080	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LARA,JORDAN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: OROZCOCORONADO

DATE/TIME OF TEST: 09/15/2025 23:35:23

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.083	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.179	23:30
Breath Volume		1.222	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.171	23:33
Breath Volume		2.128	23:33
Air Blank		0.000	23:34
Control Test		0.082	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARVAJAL,JUAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: GARCIA CABERA

DATE/TIME OF TEST: 09/19/2025 00:53:30

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.083	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.176	00:48
Breath Volume		1.164	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.175	00:51
Breath Volume		2.183	00:51
Air Blank		0.000	00:52
Control Test		0.082	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPENCER,JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:42

SUBJECT'S LAST NAME: FELO

DATE/TIME OF TEST: 09/19/2025 08:14:13

Diagnostics Check		OK	08:06
Air Blank		0.000	08:06
Control Test		0.083	08:07
Air Blank		0.000	08:07
Subject Sample	# 1	0.061	08:08
Breath Volume		2.808	08:08
Air Blank		0.000	08:09
Air Blank		0.000	08:11
Subject Sample	# 2	0.055	08:12
Breath Volume		2.289	08:12
Air Blank		0.000	08:13
Control Test		0.081	08:13
Air Blank		0.000	08:14
Diagnostics Check		OK	08:14

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA,AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: COSNER

DATE/TIME OF TEST: 09/20/2025 01:27:53

Diagnostics Check		OK	01:21
Air Blank		0.000	01:22
Control Test		0.082	01:22
Air Blank		0.000	01:23
Subject Sample	# 1	0.132	01:23
Breath Volume		1.367	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:25
Subject Sample	# 2	0.134	01:26
Breath Volume		2.617	01:26
Air Blank		0.000	01:26
Control Test		0.082	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SILVA,LUIS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ZELEDON

DATE/TIME OF TEST: 09/20/2025 03:43:59

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:32	
Control Test		0.083	03:32	
Air Blank		0.000	03:33	
Subject Sample	# 1	0.147	03:36	Volume Not Met
Breath Volume		0.433	03:36	
Air Blank		0.000	03:37	
Air Blank		0.000	03:38	
Subject Sample	# 2	0.145	03:39	
Breath Volume		1.175	03:39	
Air Blank		0.000	03:40	
Air Blank		0.000	03:41	
Subject Sample	# 3	0.151	03:42	
Breath Volume		1.230	03:42	
Air Blank		0.000	03:43	
Control Test		0.080	03:43	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIOS,JOSHUA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: ALVERNIA

DATE/TIME OF TEST: 09/20/2025 04:27:02

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.082	04:20
Air Blank		0.000	04:20
Subject Sample	# 1	0.119	04:21
Breath Volume		1.968	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:23
Subject Sample	# 2	0.116	04:25
Breath Volume		3.011	04:25
Air Blank		0.000	04:26
Control Test		0.081	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:27

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCMORROW,RYAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: KANHAI

DATE/TIME OF TEST: 09/21/2025 00:56:57

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.083	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.073	00:52
Breath Volume		1.453	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:54
Subject Sample	# 2	0.069	00:55
Breath Volume		2.597	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA,AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: MEHTA

DATE/TIME OF TEST: 09/21/2025 19:19:21

Diagnostics Check		OK	19:12
Air Blank		0.000	19:12
Control Test		0.083	19:13
Air Blank		0.000	19:13
Subject Sample	# 1	0.128	19:14
Breath Volume		1.367	19:14
Air Blank		0.000	19:14
Air Blank		0.000	19:16
Subject Sample	# 2	0.126	19:17
Breath Volume		1.492	19:17
Air Blank		0.000	19:18
Control Test		0.082	19:18
Air Blank		0.000	19:19
Diagnostics Check		OK	19:19

Data Download Date/Time: 10/16/2025 11:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARRERO,GABRIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: DEWAAL

DATE/TIME OF TEST: 09/21/2025 21:32:12

Diagnostics Check		OK	21:26
Air Blank		0.000	21:26
Control Test		0.083	21:26
Air Blank		0.000	21:27
Subject Sample	# 1	0.109	21:27
Breath Volume		3.339	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.110	21:30
Breath Volume		3.257	21:30
Air Blank		0.000	21:31
Control Test		0.082	21:31
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 10/16/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIOS,JOSHUA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MANGUAL

DATE/TIME OF TEST: 09/22/2025 02:03:44

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.082	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.121	01:59
Breath Volume		2.554	01:59
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.119	02:02
Breath Volume		3.589	02:02
Air Blank		0.000	02:02
Control Test		0.082	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 10/16/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIRISIO,ROBERT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/23/2025 00:33:42

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.083	00:28
Air Blank		0.000	00:28
Subject Sample	# 1	0.034	00:29
Breath Volume		3.687	00:29
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.030	00:32
Breath Volume		3.660	00:32
Air Blank		0.000	00:32
Control Test		0.080	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 10/16/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTILLO,SERGIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 09/25/2025 03:38:09

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.083	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.229	03:32
Breath Volume		1.199	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.227	03:36
Breath Volume		2.464	03:36
Air Blank		0.000	03:37
Control Test		0.082	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 10/16/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: YIM,RANDY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MEJIA SOLIS

DATE/TIME OF TEST: 09/25/2025 23:29:24

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.083	23:23
Air Blank		0.000	23:24
Subject Sample	# 1	0.199	23:24
Breath Volume		1.867	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.199	23:27
Breath Volume		1.753	23:27
Air Blank		0.000	23:28
Control Test		0.082	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 10/16/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CONTE,COREY
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 09/26/2025 22:02:07

Diagnostics Check		OK	21:55
Air Blank		0.000	21:55
Control Test		0.082	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.177	21:57
Breath Volume		1.109	21:57
Air Blank		0.000	21:57
Air Blank		0.000	21:59
Subject Sample	# 2	0.176	22:00
Breath Volume		2.382	22:00
Air Blank		0.000	22:01
Control Test		0.081	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 10/16/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: BAUTISTA VELASQUEZ

DATE/TIME OF TEST: 09/26/2025 23:00:14

Diagnostics Check		OK	22:53
Air Blank		0.000	22:54
Control Test		0.082	22:54
Air Blank		0.000	22:55
Subject Sample	# 1	0.219	22:55
Breath Volume		2.003	22:55
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.227	22:58
Breath Volume		1.320	22:58
Air Blank		0.000	22:59
Control Test		0.081	22:59
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 10/16/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARVAJAL,JUAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: GARCIA CARRILLO

DATE/TIME OF TEST: 09/27/2025 01:18:44

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.082	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.156	01:12
Breath Volume		1.898	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:14
Subject Sample	# 2	0.151	01:16
Breath Volume		1.601	01:16
Air Blank		0.000	01:17
Control Test		0.078	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 10/16/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOSA,EMMANUEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: GRIJALVA

DATE/TIME OF TEST: 09/27/2025 09:51:10

Diagnostics Check		OK	09:44
Air Blank		0.000	09:45
Control Test		0.082	09:45
Air Blank		0.000	09:46
Subject Sample	# 1	0.174	09:46
Breath Volume		2.085	09:46
Air Blank		0.000	09:47
Air Blank		0.000	09:48
Subject Sample	# 2	0.174	09:49
Breath Volume		1.742	09:49
Air Blank		0.000	09:50
Control Test		0.081	09:50
Air Blank		0.000	09:51
Diagnostics Check		OK	09:51

Data Download Date/Time: 10/16/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,CARISSA
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: RINKS

DATE/TIME OF TEST: 09/28/2025 02:29:56

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.082	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.158	02:25
Breath Volume		1.433	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:27
Subject Sample	# 2	0.143	02:28
Breath Volume		1.914	02:28
Air Blank		0.000	02:29
Control Test		0.081	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 10/16/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 09/28/2025 08:36:59

Diagnostics Check		OK	08:21	
Air Blank		0.000	08:22	
Control Test		0.082	08:22	
Air Blank		0.000	08:23	
Subject Sample	# 1	0.112	08:26	Volume Not Met
Breath Volume		0.218	08:26	
Air Blank		0.000	08:27	
Air Blank		0.000	08:28	
Subject Sample	# 2	0.117	08:31	
Breath Volume		1.578	08:31	
Air Blank		0.000	08:32	
Air Blank		0.000	08:33	
Subject Sample	# 3	0.118	08:35	
Breath Volume		1.894	08:35	
Air Blank		0.000	08:36	
Control Test		0.079	08:36	
Air Blank		0.000	08:36	
Diagnostics Check		OK	08:36	

Data Download Date/Time: 10/16/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PADILLA,ALBERTO
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: CANDELIER

DATE/TIME OF TEST: 09/28/2025 21:29:22

Diagnostics Check		OK	21:21
Air Blank		0.000	21:21
Control Test		0.082	21:22
Air Blank		0.000	21:22
Subject Sample	# 1	0.175	21:24
Breath Volume		2.882	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.167	21:27
Breath Volume		2.140	21:27
Air Blank		0.000	21:28
Control Test		0.078	21:28
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 10/16/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SEGURA,JEFFREY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 09:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: GUERRIER

DATE/TIME OF TEST: 09/29/2025 10:13:18

Diagnostics Check		OK	09:59	
Air Blank		0.000	09:59	
Control Test		0.082	09:59	
Air Blank		0.000	10:00	
Subject Sample	# 1	0.325	10:03	Volume Not Met
Breath Volume		0.679	10:03	
Air Blank		0.000	10:04	
Air Blank		0.000	10:06	
Subject Sample	# 2	0.348	10:08	
Breath Volume		2.011	10:08	
Air Blank		0.000	10:09	
Air Blank		0.000	10:10	
Subject Sample	# 3	0.337	10:11	
Breath Volume		1.179	10:11	
Air Blank		0.000	10:12	
Control Test		0.077	10:12	
Air Blank		0.000	10:13	
Diagnostics Check		OK	10:13	

Data Download Date/Time: 10/16/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEAT,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MARKUNAS,ANDREW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: HACKETT

DATE/TIME OF TEST: 09/06/2025 02:22:09

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.079	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1	0.168	02:17	
Breath Volume		1.484	02:17	
Air Blank		0.000	02:18	
Air Blank		0.000	02:20	
Subject Sample	# 2		02:21	Sequence Aborted
Breath Volume		0.039	02:21	
Air Blank		0.000	02:22	

Data Download Date/Time: 09/29/2025 11:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEAT,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MARKUNAS,ANDREW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: HACKETT

DATE/TIME OF TEST: 09/06/2025 02:31:28

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:24	
Control Test		0.078	02:25	
Air Blank		0.000	02:25	
Subject Sample	# 1	0.170	02:26	
Breath Volume		2.742	02:26	
Air Blank		0.000	02:27	
Air Blank		0.000	02:28	
Subject Sample	# 2		02:30	Sequence Aborted
Breath Volume		0.097	02:30	
Air Blank		0.000	02:31	

Data Download Date/Time: 09/29/2025 11:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSAS,CARLY,A
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: DEAR,KYLAN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 08/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 09/22/2025 20:36:21

Diagnostics Check		OK	20:33	
Air Blank		0.000	20:34	
Control Test		0.079	20:34	
Air Blank		0.000	20:35	
Subject Sample	# 1		20:35	Subject Test Refused
Breath Volume		0.000	20:35	
Air Blank			20:35	RFI Detect
Air Blank		0.000	20:36	

Data Download Date/Time: 09/29/2025 11:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,AMANDA,C
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: CAMPAGNONE,A
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/01/2025 03:55:46

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.150	03:51
Breath Volume		3.414	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:53
Subject Sample	# 2	0.139	03:54
Breath Volume		2.539	03:54
Air Blank		0.000	03:54
Control Test		0.081	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 09/25/2025 08:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAL,JACOB,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SOSA,AMBER
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 08:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:48

SUBJECT'S LAST NAME: OVERBECK

DATE/TIME OF TEST: 09/10/2025 09:44:32

Diagnostics Check		OK	09:38
Air Blank		0.000	09:39
Control Test		0.081	09:39
Air Blank		0.000	09:39
Subject Sample	# 1	0.000	09:40
Breath Volume		1.875	09:40
Air Blank		0.000	09:40
Air Blank		0.000	09:42
Subject Sample	# 2	0.000	09:43
Breath Volume		2.023	09:43
Air Blank		0.000	09:43
Control Test		0.079	09:43
Air Blank		0.000	09:44
Diagnostics Check		OK	09:44

Data Download Date/Time: 09/25/2025 08:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALABISE,THOMAS,A
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: HAYCOCK,CODY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 09/14/2025 00:12:04

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.223	00:07
Breath Volume		1.109	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.221	00:10
Breath Volume		1.656	00:10
Air Blank		0.000	00:11
Control Test		0.080	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:12

Data Download Date/Time: 09/25/2025 08:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALABISE,THOMAS,A
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: CLEMENTS,ZACHARY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MIELKE

DATE/TIME OF TEST: 09/21/2025 05:15:48

Diagnostics Check		OK	05:09
Air Blank		0.000	05:09
Control Test		0.080	05:10
Air Blank		0.000	05:10
Subject Sample	# 1	0.000	05:11
Breath Volume		1.832	05:11
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.000	05:14
Breath Volume		2.000	05:14
Air Blank		0.000	05:14
Control Test		0.080	05:15
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 09/25/2025 08:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,AMANDA,C
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: WOOD,MICHAEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 09/23/2025 21:16:12

Diagnostics Check		OK	21:10
Air Blank		0.000	21:10
Control Test		0.081	21:11
Air Blank		0.000	21:11
Subject Sample	# 1	0.000	21:12
Breath Volume		2.679	21:12
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.000	21:14
Breath Volume		2.714	21:14
Air Blank		0.000	21:15
Control Test		0.081	21:15
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 09/25/2025 08:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAKE,B,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: PAVLAKOS,P
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 11:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:29

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/28/2025 12:02:09

Diagnostics Check		OK	11:56	
Air Blank		0.000	11:57	
Control Test		0.081	11:57	
Air Blank		0.000	11:58	
Subject Sample	# 1	0.119	12:01	Volume Not Met
Breath Volume		0.886	12:01	
Air Blank			12:01	RFI Detect
Air Blank			12:01	RFI Detect
Air Blank			12:02	RFI Detect

Data Download Date/Time: 10/23/2025 08:15

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAKE,B,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: PAVLAKOS,P
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 11:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:29

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/28/2025 12:13:57

Diagnostics Check		OK	12:06
Air Blank		0.000	12:07
Control Test		0.078	12:07
Air Blank		0.000	12:07
Subject Sample	# 1	0.140	12:08
Breath Volume		1.488	12:08
Air Blank		0.000	12:09
Air Blank		0.000	12:11
Subject Sample	# 2	0.140	12:12
Breath Volume		1.496	12:12
Air Blank		0.000	12:13
Control Test		0.079	12:13
Air Blank		0.000	12:13
Diagnostics Check		OK	12:13

Data Download Date/Time: 10/23/2025 08:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYATT,WILLIAM,M
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DUMMER,LUKE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: MANSFIELD

DATE/TIME OF TEST: 09/02/2025 18:05:18

Diagnostics Check	OK	18:04	
Air Blank		18:04	RFI Detect
Air Blank		18:05	RFI Detect

Data Download Date/Time: 09/30/2025 09:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYATT,WILLIAM,M
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DUMMER,LUKE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: MANSFIELD

DATE/TIME OF TEST: 09/02/2025 18:07:48

Diagnostics Check	OK	18:07	
Air Blank		18:07	RFI Detect
Air Blank		18:07	RFI Detect

Data Download Date/Time: 09/30/2025 09:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYATT,WILLIAM,M
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DUMMER,LUKE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: MANSFIELD

DATE/TIME OF TEST: 09/02/2025 18:11:11

Diagnostics Check	OK	18:10	
Air Blank		18:11	RFI Detect
Air Blank		18:11	RFI Detect

Data Download Date/Time: 09/30/2025 09:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOUNTAIN,D,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: RIVERA,D
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/06/2025 04:08:29

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.079	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.189	04:04
Breath Volume		1.750	04:04
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.195	04:06
Breath Volume		2.765	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 09/30/2025 09:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 09/09/2025 00:17:32

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.081	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.168	00:12
Breath Volume		1.597	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.153	00:16
Breath Volume		1.574	00:16
Air Blank		0.000	00:16
Control Test		0.079	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 09/30/2025 09:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: NIGH,ALEXANDER
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: SHORT

DATE/TIME OF TEST: 09/12/2025 01:05:16

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.081	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.067	01:00
Breath Volume		2.312	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.058	01:03
Breath Volume		2.281	01:03
Air Blank		0.000	01:04
Control Test		0.080	01:04
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 09/30/2025 09:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEAT,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: STEPHENS,JONATHAN
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: SOMMERS

DATE/TIME OF TEST: 09/13/2025 21:22:23

Diagnostics Check		OK	21:16
Air Blank		0.000	21:16
Control Test		0.079	21:16
Air Blank		0.000	21:17
Subject Sample	# 1	0.009	21:17
Breath Volume		2.066	21:17
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.007	21:21
Breath Volume		3.316	21:21
Air Blank		0.000	21:21
Control Test		0.080	21:21
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 09/30/2025 09:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SILVESTRI,CODY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 09/13/2025 23:35:56

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:28	
Control Test		0.080	23:29	
Air Blank		0.000	23:29	
Subject Sample	# 1	0.167	23:32	Volume Not Met
Breath Volume		0.613	23:32	
Air Blank		0.000	23:33	
Air Blank		0.000	23:35	
Subject Sample	# 2		23:35	RFI Detect
Breath Volume		0.304	23:35	
Air Blank			23:35	RFI Detect
Air Blank			23:35	RFI Detect

Data Download Date/Time: 09/30/2025 09:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SILVESTRI,CODY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 09/13/2025 23:48:55

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:38	
Control Test		0.079	23:38	
Air Blank		0.000	23:39	
Subject Sample	# 1	0.179	23:42	Volume Not Met
Breath Volume		0.101	23:42	
Air Blank		0.000	23:43	
Air Blank		0.000	23:45	
Subject Sample	# 2		23:47	Subject Test Refused
Breath Volume		0.039	23:47	
Air Blank		0.000	23:48	
Control Test		0.078	23:48	
Air Blank		0.000	23:48	
Diagnostics Check		OK	23:48	

Data Download Date/Time: 09/30/2025 09:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEAT,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FRENCH,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 09/15/2025 03:02:03

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.156	02:57
Breath Volume		1.734	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.151	03:00
Breath Volume		1.667	03:00
Air Blank		0.000	03:01
Control Test		0.078	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:02

Data Download Date/Time: 09/30/2025 09:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: SYLVAIN

DATE/TIME OF TEST: 09/15/2025 21:05:43

Diagnostics Check		OK	20:59
Air Blank		0.000	21:00
Control Test		0.080	21:00
Air Blank		0.000	21:00
Subject Sample	# 1	0.080	21:01
Breath Volume		2.535	21:01
Air Blank		0.000	21:01
Air Blank		0.000	21:03
Subject Sample	# 2	0.078	21:04
Breath Volume		1.898	21:04
Air Blank		0.000	21:04
Control Test		0.078	21:05
Air Blank		0.000	21:05
Diagnostics Check		OK	21:05

Data Download Date/Time: 09/30/2025 09:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: RASMUSSEN,MATTHEW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 09/18/2025 20:20:06

Diagnostics Check		OK	20:13
Air Blank		0.000	20:13
Control Test		0.080	20:14
Air Blank		0.000	20:14
Subject Sample	# 1	0.157	20:15
Breath Volume		1.792	20:15
Air Blank		0.000	20:16
Air Blank		0.000	20:17
Subject Sample	# 2	0.151	20:18
Breath Volume		1.527	20:18
Air Blank		0.000	20:19
Control Test		0.079	20:19
Air Blank		0.000	20:20
Diagnostics Check		OK	20:20

Data Download Date/Time: 09/30/2025 09:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAROCHE,CHARLES,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ISOLA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: RAYA PEREZ

DATE/TIME OF TEST: 09/21/2025 06:03:55

Diagnostics Check		OK	05:57
Air Blank		0.000	05:57
Control Test		0.080	05:58
Air Blank		0.000	05:58
Subject Sample	# 1	0.194	05:59
Breath Volume		1.753	05:59
Air Blank		0.000	06:00
Air Blank		0.000	06:01
Subject Sample	# 2	0.202	06:02
Breath Volume		2.117	06:02
Air Blank		0.000	06:03
Control Test		0.078	06:03
Air Blank		0.000	06:03
Diagnostics Check		OK	06:03

Data Download Date/Time: 09/30/2025 09:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,TYRONE,P
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: LOGAN,TYRONE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:27

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 09/23/2025 09:27:27

Diagnostics Check		OK	09:21
Air Blank		0.000	09:21
Control Test		0.080	09:22
Air Blank		0.000	09:22
Subject Sample	# 1	0.055	09:23
Breath Volume		1.996	09:23
Air Blank		0.000	09:23
Air Blank		0.000	09:25
Subject Sample	# 2	0.058	09:26
Breath Volume		2.019	09:26
Air Blank		0.000	09:26
Control Test		0.079	09:26
Air Blank		0.000	09:27
Diagnostics Check		OK	09:27

Data Download Date/Time: 09/30/2025 09:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDOUX,JACOB,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SHARP,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/27/2025 05:43:53

Diagnostics Check		OK	05:31	
Air Blank		0.000	05:32	
Control Test		0.080	05:32	
Air Blank		0.000	05:33	
Subject Sample	# 1	0.218	05:33	No .020 agreement
Breath Volume		2.332	05:33	
Air Blank		0.000	05:34	
Air Blank		0.000	05:35	
Subject Sample	# 2	0.195	05:36	No .020 agreement
Breath Volume		1.136	05:36	
Air Blank		0.000	05:37	
Air Blank		0.000	05:39	
Subject Sample	# 3	0.213	05:42	Volume Not Met
Breath Volume		0.328	05:42	
Air Blank		0.000	05:42	
Control Test		0.077	05:43	
Air Blank		0.000	05:43	
Diagnostics Check		OK	05:43	

Data Download Date/Time: 09/30/2025 09:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FRITSCH,A
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/01/2025 03:01:51

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.083	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.000	02:57
Breath Volume		1.761	02:57
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.000	03:00
Breath Volume		1.691	03:00
Air Blank		0.000	03:00
Control Test		0.083	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 09/04/2025 05:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUROUMOUSIS,GEORGE,T
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KOUROUMOUSIS,GEORGE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 09/01/2025 20:13:41

Diagnostics Check		OK	20:07
Air Blank		0.000	20:07
Control Test		0.083	20:08
Air Blank		0.000	20:08
Subject Sample	# 1	0.378	20:09
Breath Volume		2.171	20:09
Air Blank		0.000	20:09
Air Blank		0.000	20:11
Subject Sample	# 2	0.373	20:12
Breath Volume		1.578	20:12
Air Blank		0.000	20:12
Control Test		0.082	20:13
Air Blank		0.000	20:13
Diagnostics Check		OK	20:13

Data Download Date/Time: 09/04/2025 05:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NAKAMURA,L
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: BRAY

DATE/TIME OF TEST: 09/01/2025 23:53:16

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.083	23:46
Air Blank		0.000	23:47
Subject Sample	# 1	0.139	23:48
Breath Volume		1.832	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.131	23:51
Breath Volume		2.410	23:51
Air Blank		0.000	23:52
Control Test		0.082	23:52
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 09/04/2025 05:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NORTHCUT,E
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 09/03/2025 03:27:56

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:24	
Control Test		0.083	03:24	
Air Blank		0.000	03:25	
Subject Sample	# 1		03:26	Subject Test Refused
Breath Volume		0.000	03:26	
Air Blank		0.000	03:27	
Control Test		0.082	03:27	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:27	

Data Download Date/Time: 09/04/2025 05:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANTIAGO,BRYAN
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 09/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: TAVAREZ

DATE/TIME OF TEST: 09/04/2025 21:52:52

Diagnostics Check		OK	21:46
Air Blank		0.000	21:46
Control Test		0.082	21:47
Air Blank		0.000	21:47
Subject Sample	# 1	0.195	21:48
Breath Volume		2.578	21:48
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.187	21:51
Breath Volume		2.722	21:51
Air Blank		0.000	21:51
Control Test		0.081	21:52
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 10/02/2025 16:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BITTIKOFER,PEGGY
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 09/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: BROTHERTON

DATE/TIME OF TEST: 09/06/2025 01:53:49

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.083	01:48
Air Blank		0.000	01:48
Subject Sample	# 1	0.097	01:49
Breath Volume		3.816	01:49
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.094	01:52
Breath Volume		3.621	01:52
Air Blank		0.000	01:52
Control Test		0.082	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 10/02/2025 16:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SMALLEY,R
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 09/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: LESPINACE

DATE/TIME OF TEST: 09/07/2025 02:32:10

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.083	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.180	02:27
Breath Volume		1.886	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.176	02:30
Breath Volume		1.406	02:30
Air Blank		0.000	02:31
Control Test		0.082	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 10/02/2025 16:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ROSE,M
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 09/07/2025 05:17:12

Diagnostics Check		OK	05:09
Air Blank		0.000	05:10
Control Test		0.083	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.000	05:12
Breath Volume		5.285	05:12
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.000	05:15
Breath Volume		2.675	05:15
Air Blank		0.000	05:16
Control Test		0.082	05:16
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 10/02/2025 16:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: ROSE,M
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 08:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:39

SUBJECT'S LAST NAME: SMOAK

DATE/TIME OF TEST: 09/13/2025 10:09:56

Diagnostics Check		OK	09:59
Air Blank		0.000	09:59
Control Test		0.082	09:59
Air Blank		0.000	10:00
Subject Sample	# 1	0.177	10:01
Breath Volume		2.562	10:01
Air Blank		0.000	10:01
Air Blank		0.000	10:03
Subject Sample	# 2	0.150	10:04
Breath Volume		2.253	10:04
Air Blank		0.000	10:05
Air Blank		0.000	10:07
Subject Sample	# 3	0.148	10:08
Breath Volume		2.625	10:08
Air Blank		0.000	10:09
Control Test		0.081	10:09
Air Blank		0.000	10:09
Diagnostics Check		OK	10:09

Data Download Date/Time: 10/02/2025 16:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: FRENCH,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 09/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 13:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: COSEO

DATE/TIME OF TEST: 09/13/2025 14:46:43

Diagnostics Check		OK	14:37	
Air Blank		0.000	14:38	
Control Test		0.082	14:38	
Air Blank		0.000	14:38	
Subject Sample	# 1	0.064	14:40	
Breath Volume		2.402	14:40	
Air Blank		0.000	14:41	
Air Blank		0.000	14:43	
Subject Sample	# 2	0.059	14:45	
Breath Volume		1.882	14:45	
Air Blank		0.000	14:45	
Control Test		0.081	14:45	
Air Blank			14:46	RFI Detect
Air Blank		0.000	14:46	

Data Download Date/Time: 10/02/2025 16:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: VARKANY SELMER,L
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 09/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: ALZUBAIDI

DATE/TIME OF TEST: 09/14/2025 14:30:43

Diagnostics Check		OK	14:23
Air Blank		0.000	14:24
Control Test		0.082	14:24
Air Blank		0.000	14:25
Subject Sample	# 1	0.171	14:26
Breath Volume		2.199	14:26
Air Blank		0.000	14:26
Air Blank		0.000	14:28
Subject Sample	# 2	0.186	14:29
Breath Volume		3.800	14:29
Air Blank		0.000	14:29
Control Test		0.082	14:30
Air Blank		0.000	14:30
Diagnostics Check		OK	14:30

Data Download Date/Time: 10/02/2025 16:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MASON,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:34

SUBJECT'S LAST NAME: GUCCIONE

DATE/TIME OF TEST: 09/05/2025 16:59:02

Diagnostics Check		OK	16:55	
Air Blank		0.000	16:56	
Control Test		0.078	16:56	
Air Blank		0.000	16:57	
Subject Sample	# 1		16:57	Subject Test Refused
Breath Volume		0.000	16:57	
Air Blank		0.000	16:58	
Control Test		0.078	16:58	
Air Blank		0.000	16:58	
Diagnostics Check		OK	16:59	

Data Download Date/Time: 09/17/2025 13:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.