Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: BLACKMAN

DATE/TIME OF TEST: 10/04/2025 00:58:46

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.078	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.095	00:54
Breath Volume		2.906	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:56
Subject Sample	# 2	0.095	00:57
Breath Volume		2.941	00:57
Air Blank		0.000	00:57
Control Test		0.077	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 11/05/2025 21:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALLTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MONTOYA

DATE/TIME OF TEST: 10/05/2025 02:21:12

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:17	
Control Test		0.078	02:18	
Air Blank		0.000	02:18	
Subject Sample	# 1		02:19	Subject Test Refused
Breath Volume		0.000	02:19	
Air Blank		0.000	02:20	
Control Test		0.076	02:20	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 11/05/2025 21:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MENDOZA SUAREZ

DATE/TIME OF TEST: 10/11/2025 01:08:18

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.078	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.197	01:03
Breath Volume		1.496	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:05
Subject Sample	# 2	0.211	01:06
Breath Volume		1.449	01:06
Air Blank		0.000	01:07
Control Test		0.078	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 11/05/2025 21:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/12/2025 01:08:37

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:56	
Control Test		0.078	00:57	
Air Blank		0.000	00:57	
Subject Sample	# 1		01:00	No Sample Provided
Breath Volume		0.000	01:00	
Air Blank		0.000	01:01	
Air Blank		0.000	01:03	
Subject Sample	# 2	0.343	01:03	
Breath Volume		1.847	01:03	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 3	0.336	01:06	
Breath Volume		1.632	01:06	
Air Blank		0.000	01:07	
Control Test		0.077	01:07	
Air Blank		0.000	01:08	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 11/05/2025 21:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: MONTEAGUDO HERNANDEZ

DATE/TIME OF TEST: 10/17/2025 23:31:35

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:29	
Control Test		0.079	23:29	
Air Blank		0.000	23:30	
Subject Sample	# 1		23:30	Subject Test Refused
Breath Volume		0.000	23:30	
Air Blank		0.000	23:30	
Control Test		0.079	23:30	
Air Blank		0.000	23:31	
Diagnostics Check		OK	23:31	

Data Download Date/Time: 11/05/2025 21:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: ALVAREZ AVILA

DATE/TIME OF TEST: 10/18/2025 22:30:20

Diagnostics Check		OK	22:26	
Air Blank		0.000	22:27	
Control Test		0.079	22:27	
Air Blank		0.000	22:28	
Subject Sample	# 1		22:28	Subject Test Refused
Breath Volume		0.000	22:28	
Air Blank		0.000	22:29	
Control Test		0.079	22:29	
Air Blank		0.000	22:30	
Diagnostics Check		OK	22:30	

Data Download Date/Time: 11/05/2025 21:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:44

SUBJECT'S LAST NAME: ARANGO BARBOZA

DATE/TIME OF TEST: 10/19/2025 07:09:10

Diagnostics Check		OK	07:05	
Air Blank		0.000	07:06	
Control Test		0.080	07:06	
Air Blank		0.000	07:07	
Subject Sample	# 1		07:07	Slope Not Met
Breath Volume		1.039	07:07	
Air Blank		0.000	07:08	
Control Test		0.080	07:08	
Air Blank		0.000	07:09	
Diagnostics Check		OK	07:09	

Data Download Date/Time: 11/05/2025 21:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:44

SUBJECT'S LAST NAME: ARANGO BARBOZA

DATE/TIME OF TEST: 10/19/2025 07:17:13

Diagnostics Check		OK	07:10
Air Blank		0.000	07:11
Control Test		0.079	07:11
Air Blank		0.000	07:12
Subject Sample	# 1	0.215	07:12
Breath Volume		2.351	07:12
Air Blank		0.000	07:13
Air Blank		0.000	07:15
Subject Sample	# 2	0.210	07:15
Breath Volume		1.636	07:15
Air Blank		0.000	07:16
Control Test		0.079	07:16
Air Blank		0.000	07:17
Diagnostics Check		OK	07:17

Data Download Date/Time: 11/05/2025 21:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: OJEDA

DATE/TIME OF TEST: 10/24/2025 02:40:35

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.261	02:35
Breath Volume		1.828	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.259	02:38
Breath Volume		3.230	02:38
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 11/05/2025 21:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROSSI,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: HURST

DATE/TIME OF TEST: 10/01/2025 02:38:19

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.078	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.177	02:33
Breath Volume		1.683	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.184	02:36
Breath Volume		1.941	02:36
Air Blank		0.000	02:37
Control Test		0.078	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 10/16/2025 14:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: WIGH

DATE/TIME OF TEST: 10/01/2025 20:43:10

Diagnostics Check		OK	20:40	
Air Blank		0.000	20:40	
Control Test		0.078	20:41	
Air Blank		0.000	20:41	
Subject Sample	# 1		20:41	Subject Test Refused
Breath Volume		0.000	20:41	
Air Blank		0.000	20:42	
Control Test		0.079	20:42	
Air Blank		0.000	20:43	
Diagnostics Check		OK	20:43	

Data Download Date/Time: 10/16/2025 14:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BATEH,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: NOLES

DATE/TIME OF TEST: 10/02/2025 02:48:39

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.078	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.081	02:43
Breath Volume		1.878	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.066	02:47
Breath Volume		1.769	02:47
Air Blank		0.000	02:47
Control Test		0.078	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 10/16/2025 14:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 10/03/2025 01:10:34

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:08	
Control Test		0.079	01:08	
Air Blank		0.000	01:09	
Subject Sample	# 1		01:09	Subject Test Refused
Breath Volume		0.000	01:09	
Air Blank		0.000	01:09	
Control Test		0.079	01:09	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 10/16/2025 14:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ELLISON,H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: RIVERA SANCHEZ

DATE/TIME OF TEST: 10/03/2025 23:25:17

Diagnostics Check		OK	23:16
Air Blank		0.000	23:17
Control Test		0.079	23:17
Air Blank		0.000	23:17
Subject Sample	# 1	0.077	23:18
Breath Volume		1.699	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.072	23:23
Breath Volume		2.332	23:23
Air Blank		0.000	23:24
Control Test		0.079	23:24
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 10/16/2025 14:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: OWENS,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MIMS

DATE/TIME OF TEST: 10/04/2025 02:08:49

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:06	
Control Test		0.079	02:06	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:07	Subject Test Refused
Breath Volume		0.000	02:07	
Air Blank		0.000	02:07	
Control Test		0.079	02:08	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 10/16/2025 14:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 10/04/2025 03:07:32

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:05	
Control Test		0.079	03:05	
Air Blank		0.000	03:06	
Subject Sample	# 1		03:06	RFI Detect
Breath Volume		0.000	03:06	
Air Blank			03:07	RFI Detect
Air Blank		0.000	03:07	

Data Download Date/Time: 10/16/2025 14:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 10/04/2025 03:16:40

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.078	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.100	03:12
Breath Volume		1.835	03:12
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.100	03:15
Breath Volume		2.007	03:15
Air Blank		0.000	03:15
Control Test		0.078	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 10/16/2025 14:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRERA,R,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 10/04/2025 21:43:47

Diagnostics Check		OK	21:37
Air Blank		0.000	21:37
Control Test		0.079	21:38
Air Blank		0.000	21:38
Subject Sample	# 1	0.000	21:39
Breath Volume		1.917	21:39
Air Blank		0.000	21:39
Air Blank		0.000	21:41
Subject Sample	# 2	0.000	21:42
Breath Volume		2.210	21:42
Air Blank		0.000	21:42
Control Test		0.079	21:43
Air Blank		0.000	21:43
Diagnostics Check		OK	21:43

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 10/05/2025 00:51:29

Diagnostics Check		OK	00:45	
Air Blank		0.000	00:45	
Control Test		0.078	00:46	
Air Blank		0.000	00:46	
Subject Sample	# 1	0.072	00:47	
Breath Volume		1.269	00:47	
Air Blank		0.000	00:48	
Air Blank		0.000	00:50	
Subject Sample	# 2	0.063	00:50	
Breath Volume		1.285	00:50	
Air Blank			00:51	RFI Detect
Air Blank		0.000	00:51	

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 10/05/2025 01:01:48

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.078	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.061	00:55
Breath Volume		1.132	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.063	01:00
Breath Volume		1.191	01:00
Air Blank		0.000	01:00
Control Test		0.078	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: OWENS,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 10/05/2025 03:07:43

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:05	
Control Test		0.078	03:05	
Air Blank		0.000	03:06	
Subject Sample	# 1		03:06	Subject Test Refused
Breath Volume		0.000	03:06	
Air Blank		0.000	03:06	
Control Test		0.079	03:07	
Air Blank		0.000	03:07	
Diagnostics Check		OK	03:07	

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRERA,R,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 10/05/2025 05:24:44

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.078	05:18
Air Blank		0.000	05:19
Subject Sample	# 1	0.201	05:20
Breath Volume		1.984	05:20
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.193	05:23
Breath Volume		2.089	05:23
Air Blank		0.000	05:23
Control Test		0.078	05:24
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: GRINDSTEAD

DATE/TIME OF TEST: 10/05/2025 18:37:15

Diagnostics Check		OK	18:34	
Air Blank		0.000	18:34	
Control Test		0.078	18:35	
Air Blank		0.000	18:35	
Subject Sample	# 1		18:35	Subject Test Refused
Breath Volume		0.000	18:35	
Air Blank		0.000	18:36	
Control Test		0.078	18:36	
Air Blank		0.000	18:37	
Diagnostics Check		OK	18:37	

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DUBOSE

DATE/TIME OF TEST: 10/05/2025 23:25:32

Diagnostics Check		OK	23:22	
Air Blank		0.000	23:23	
Control Test		0.079	23:23	
Air Blank		0.000	23:23	
Subject Sample	# 1		23:24	Subject Test Refused
Breath Volume		0.000	23:24	
Air Blank		0.000	23:24	
Control Test		0.078	23:24	
Air Blank		0.000	23:25	
Diagnostics Check		OK	23:25	

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CROFOOT

DATE/TIME OF TEST: 10/06/2025 03:27:21

Diagnostics Check		OK	03:24	
Air Blank		0.000	03:24	
Control Test		0.079	03:25	
Air Blank		0.000	03:25	
Subject Sample	# 1		03:26	Subject Test Refused
Breath Volume		0.000	03:26	
Air Blank		0.000	03:26	
Control Test		0.078	03:26	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:27	

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRERA,R,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 10/06/2025 04:44:11

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.078	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.201	04:39
Breath Volume		1.523	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:41
Subject Sample	# 2	0.195	04:42
Breath Volume		1.269	04:42
Air Blank		0.000	04:43
Control Test		0.078	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOHAMMED,A,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/06/2025 05:44:01

Diagnostics Check		OK	05:41	
Air Blank		0.000	05:41	
Control Test		0.079	05:41	
Air Blank		0.000	05:42	
Subject Sample	# 1		05:42	Subject Test Refused
Breath Volume		0.000	05:42	
Air Blank		0.000	05:43	
Control Test		0.078	05:43	
Air Blank		0.000	05:43	
Diagnostics Check		OK	05:44	

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LITTLE,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: CHAPPLE

DATE/TIME OF TEST: 10/06/2025 23:13:51

Diagnostics Check		OK	23:11	
Air Blank		0.000	23:11	
Control Test		0.079	23:11	
Air Blank		0.000	23:12	
Subject Sample	# 1		23:12	Subject Test Refused
Breath Volume		0.000	23:12	
Air Blank		0.000	23:12	
Control Test		0.079	23:13	
Air Blank		0.000	23:13	
Diagnostics Check		OK	23:13	

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON JR,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MOONEY

DATE/TIME OF TEST: 10/07/2025 02:35:07

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.078	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.209	02:30
Breath Volume		1.488	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.209	02:33
Breath Volume		1.105	02:33
Air Blank		0.000	02:34
Control Test		0.077	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 10/07/2025 06:08:38

Diagnostics Check		OK	06:05	
Air Blank		0.000	06:06	
Control Test		0.078	06:06	
Air Blank		0.000	06:06	
Subject Sample	# 1		06:07	Subject Test Refused
Breath Volume		0.000	06:07	
Air Blank		0.000	06:07	
Control Test		0.078	06:08	
Air Blank		0.000	06:08	
Diagnostics Check		OK	06:08	

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BATEH,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: KOFSKEY

DATE/TIME OF TEST: 10/09/2025 00:17:18

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.078	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.165	00:12
Breath Volume		1.847	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.158	00:15
Breath Volume		1.445	00:15
Air Blank		0.000	00:16
Control Test		0.078	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: ACUFF

DATE/TIME OF TEST: 10/09/2025 23:01:33

Diagnostics Check		OK	22:55
Air Blank		0.000	22:55
Control Test		0.078	22:56
Air Blank		0.000	22:56
Subject Sample	# 1	0.229	22:57
Breath Volume		1.398	22:57
Air Blank		0.000	22:57
Air Blank		0.000	22:59
Subject Sample	# 2	0.227	23:00
Breath Volume		2.121	23:00
Air Blank		0.000	23:00
Control Test		0.079	23:00
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKISON,B,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ROMEO

DATE/TIME OF TEST: 10/11/2025 03:08:26

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:05	
Control Test		0.079	03:06	
Air Blank		0.000	03:06	
Subject Sample	# 1		03:07	RFI Detect
Breath Volume		0.000	03:07	
Air Blank		0.000	03:07	
Control Test		0.080	03:07	
Air Blank		0.000	03:08	
Diagnostics Check		OK	03:08	

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKISON,B,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ROMEO

DATE/TIME OF TEST: 10/11/2025 03:18:07

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.162	03:12
Breath Volume		1.746	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:14
Subject Sample	# 2	0.165	03:16
Breath Volume		2.105	03:16
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ARGUDO,L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/11/2025 05:01:43

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.079	04:55
Air Blank		0.000	04:56
Subject Sample	# 1	0.131	04:57
Breath Volume		2.050	04:57
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 2	0.127	05:00
Breath Volume		1.285	05:00
Air Blank		0.000	05:00
Control Test		0.078	05:01
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 10/16/2025 14:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARTIN,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 10/12/2025 01:07:40

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.078	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.140	01:02
Breath Volume		1.527	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.149	01:06
Breath Volume		1.910	01:06
Air Blank		0.000	01:06
Control Test		0.079	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 10/16/2025 14:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: KEISTER

DATE/TIME OF TEST: 10/12/2025 03:24:48

Diagnostics Check		OK	03:21	
Air Blank		0.000	03:22	
Control Test		0.079	03:22	
Air Blank		0.000	03:23	
Subject Sample	# 1		03:23	Subject Test Refused
Breath Volume		0.000	03:23	
Air Blank		0.000	03:23	
Control Test		0.079	03:24	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 10/16/2025 14:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 10/13/2025 00:12:06

Diagnostics Check		OK	00:05
Air Blank		0.000	00:06
Control Test		0.079	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.068	00:07
Breath Volume		2.136	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.068	00:10
Breath Volume		1.722	00:10
Air Blank		0.000	00:11
Control Test		0.079	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 10/16/2025 14:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: GILES

DATE/TIME OF TEST: 10/13/2025 01:19:17

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:16	
Control Test		0.078	01:17	
Air Blank		0.000	01:17	
Subject Sample	# 1		01:17	Subject Test Refused
Breath Volume		0.000	01:17	
Air Blank		0.000	01:18	
Control Test		0.078	01:18	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 10/16/2025 14:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 10/13/2025 03:48:00

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.079	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.128	03:43
Breath Volume		2.687	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.139	03:46
Breath Volume		2.582	03:46
Air Blank		0.000	03:47
Control Test		0.079	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:48

Data Download Date/Time: 10/16/2025 14:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILAM,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: SCULLY

DATE/TIME OF TEST: 10/13/2025 06:35:56

Diagnostics Check		OK	06:33	
Air Blank		0.000	06:33	
Control Test		0.079	06:33	
Air Blank		0.000	06:34	
Subject Sample	# 1		06:34	Subject Test Refused
Breath Volume		0.000	06:34	
Air Blank		0.000	06:35	
Control Test		0.080	06:35	
Air Blank		0.000	06:35	
Diagnostics Check		OK	06:35	

Data Download Date/Time: 10/16/2025 14:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESUYO,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: GIGGEY

DATE/TIME OF TEST: 10/13/2025 22:17:35

Diagnostics Check		OK	22:11
Air Blank		0.000	22:11
Control Test		0.080	22:12
Air Blank		0.000	22:12
Subject Sample	# 1	0.158	22:13
Breath Volume		2.480	22:13
Air Blank		0.000	22:13
Air Blank		0.000	22:15
Subject Sample	# 2	0.157	22:16
Breath Volume		2.003	22:16
Air Blank		0.000	22:16
Control Test		0.079	22:16
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 10/16/2025 14:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEIS,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PASTRAN

DATE/TIME OF TEST: 10/13/2025 23:15:28

Diagnostics Check		OK	23:12	
Air Blank		0.000	23:13	
Control Test		0.079	23:13	
Air Blank		0.000	23:13	
Subject Sample	# 1		23:14	Subject Test Refused
Breath Volume		0.000	23:14	
Air Blank		0.000	23:14	
Control Test		0.080	23:14	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 10/16/2025 14:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 10/15/2025 01:06:55

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.073	01:02
Breath Volume		1.472	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.072	01:05
Breath Volume		1.546	01:05
Air Blank		0.000	01:06
Control Test		0.079	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 10/16/2025 14:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TUCKER,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/15/2025 07:59:49

Diagnostics Check		OK	07:56	
Air Blank		0.000	07:57	
Control Test		0.079	07:57	
Air Blank		0.000	07:58	
Subject Sample	# 1		07:58	Subject Test Refused
Breath Volume		0.000	07:58	
Air Blank		0.000	07:58	
Control Test		0.079	07:59	
Air Blank		0.000	07:59	
Diagnostics Check		OK	07:59	

Data Download Date/Time: 10/16/2025 14:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/15/2025 16:11:14

Diagnostics Check		OK	16:05
Air Blank		0.000	16:05
Control Test		0.079	16:06
Air Blank		0.000	16:06
Subject Sample	# 1	0.141	16:06
Breath Volume		1.195	16:06
Air Blank		0.000	16:07
Air Blank		0.000	16:09
Subject Sample	# 2	0.158	16:09
Breath Volume		1.136	16:09
Air Blank		0.000	16:10
Control Test		0.079	16:10
Air Blank		0.000	16:11
Diagnostics Check		OK	16:11

Data Download Date/Time: 10/16/2025 14:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: SCHULZ

DATE/TIME OF TEST: 10/16/2025 00:56:34

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.079	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.088	00:52
Breath Volume		2.671	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.087	00:55
Breath Volume		2.511	00:55
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 10/16/2025 14:12

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/16/2025 18:49:38

Diagnostics Check		OK	18:46	
Air Blank		0.000	18:47	
Control Test		0.080	18:47	
Air Blank		0.000	18:48	
Subject Sample	# 1		18:48	Subject Test Refused
Breath Volume		0.000	18:48	
Air Blank		0.000	18:48	
Control Test		0.079	18:49	
Air Blank		0.000	18:49	
Diagnostics Check		OK	18:49	

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,D,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 10/16/2025 20:10:40

Diagnostics Check		OK	20:07	
Air Blank		0.000	20:08	
Control Test		0.079	20:08	
Air Blank		0.000	20:09	
Subject Sample	# 1		20:09	Subject Test Refused
Breath Volume		0.000	20:09	
Air Blank		0.000	20:09	
Control Test		0.079	20:10	
Air Blank		0.000	20:10	
Diagnostics Check		OK	20:10	

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: CHISM

DATE/TIME OF TEST: 10/17/2025 04:20:02

Diagnostics Check		OK	04:16	
Air Blank		0.000	04:17	
Control Test		0.079	04:17	
Air Blank		0.000	04:18	
Subject Sample	# 1		04:18	Slope Not Met
Breath Volume		1.066	04:18	
Air Blank		0.000	04:19	
Control Test		0.080	04:19	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:20	

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: CHISM

DATE/TIME OF TEST: 10/17/2025 04:27:26

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.079	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.219	04:23
Breath Volume		1.707	04:23
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.206	04:25
Breath Volume		1.378	04:25
Air Blank		0.000	04:26
Control Test		0.078	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: CISNEROS

DATE/TIME OF TEST: 10/18/2025 02:59:45

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:57	
Control Test		0.079	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1		02:58	Subject Test Refused
Breath Volume		0.000	02:58	
Air Blank		0.000	02:58	
Control Test		0.079	02:59	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,I
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: HURST

DATE/TIME OF TEST: 10/18/2025 05:09:56

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.079	05:00
Air Blank		0.000	05:00
Subject Sample	# 1	0.105	05:02
Breath Volume		1.433	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.111	05:08
Breath Volume		1.246	05:08
Air Blank		0.000	05:09
Control Test		0.079	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: OWENS,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: BROCKUNIER

DATE/TIME OF TEST: 10/18/2025 21:22:03

Diagnostics Check		OK	21:19	
Air Blank		0.000	21:19	
Control Test		0.079	21:20	
Air Blank		0.000	21:20	
Subject Sample	# 1		21:20	Subject Test Refused
Breath Volume		0.000	21:20	
Air Blank		0.000	21:21	
Control Test		0.079	21:21	
Air Blank		0.000	21:21	
Diagnostics Check		OK	21:22	

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SHAW,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: BAGAZAHU

DATE/TIME OF TEST: 10/18/2025 22:25:53

Diagnostics Check		OK	22:19
Air Blank		0.000	22:20
Control Test		0.079	22:20
Air Blank		0.000	22:20
Subject Sample	# 1	0.264	22:21
Breath Volume		1.476	22:21
Air Blank		0.000	22:22
Air Blank		0.000	22:23
Subject Sample	# 2	0.276	22:24
Breath Volume		1.957	22:24
Air Blank		0.000	22:24
Control Test		0.077	22:25
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: SALEM

DATE/TIME OF TEST: 10/19/2025 02:49:49

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.078	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.140	02:45
Breath Volume		1.324	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:47
Subject Sample	# 2	0.132	02:48
Breath Volume		1.203	02:48
Air Blank		0.000	02:48
Control Test		0.078	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: BUSSON

DATE/TIME OF TEST: 10/19/2025 03:45:53

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.079	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.104	03:40
Breath Volume		1.296	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.102	03:44
Breath Volume		1.324	03:44
Air Blank		0.000	03:44
Control Test		0.078	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: THAO

DATE/TIME OF TEST: 10/19/2025 18:05:08

Diagnostics Check		OK	18:02	
Air Blank		0.000	18:02	
Control Test		0.079	18:03	
Air Blank		0.000	18:03	
Subject Sample	# 1		18:03	Subject Test Refused
Breath Volume		0.000	18:03	
Air Blank		0.000	18:04	
Control Test		0.079	18:04	
Air Blank		0.000	18:05	
Diagnostics Check		OK	18:05	

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BAKER,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: STITES

DATE/TIME OF TEST: 10/21/2025 02:30:27

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.079	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.191	02:25
Breath Volume		2.941	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.191	02:28
Breath Volume		2.984	02:28
Air Blank		0.000	02:29
Control Test		0.078	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLADDEN,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:51

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/21/2025 09:15:02

Diagnostics Check		OK	09:12	
Air Blank		0.000	09:12	
Control Test		0.079	09:12	
Air Blank		0.000	09:13	
Subject Sample	# 1		09:13	Subject Test Refused
Breath Volume		0.000	09:13	
Air Blank		0.000	09:14	
Control Test		0.079	09:14	
Air Blank		0.000	09:14	
Diagnostics Check		OK	09:15	

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLOW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: MCALISTER

DATE/TIME OF TEST: 10/23/2025 05:52:35

Diagnostics Check		OK	05:46
Air Blank		0.000	05:46
Control Test		0.080	05:47
Air Blank		0.000	05:47
Subject Sample	# 1	0.127	05:48
Breath Volume		1.281	05:48
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.121	05:51
Breath Volume		1.531	05:51
Air Blank		0.000	05:51
Control Test		0.079	05:51
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/24/2025 00:09:31

Diagnostics Check		OK	00:03
Air Blank		0.000	00:03
Control Test		0.079	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.157	00:04
Breath Volume		1.398	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.143	00:08
Breath Volume		2.269	00:08
Air Blank		0.000	00:08
Control Test		0.079	00:08
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: GREGERSEN

DATE/TIME OF TEST: 10/24/2025 04:55:15

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.080	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.000	04:51
Breath Volume		1.984	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.000	04:53
Breath Volume		1.984	04:53
Air Blank		0.000	04:54
Control Test		0.079	04:54
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: CARTWRIGHT

DATE/TIME OF TEST: 10/24/2025 23:53:37

Diagnostics Check		OK	23:47
Air Blank		0.000	23:47
Control Test		0.079	23:47
Air Blank		0.000	23:48
Subject Sample	# 1	0.072	23:49
Breath Volume		2.085	23:49
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.071	23:52
Breath Volume		1.937	23:52
Air Blank		0.000	23:52
Control Test		0.079	23:53
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MCLENDON

DATE/TIME OF TEST: 10/25/2025 02:10:06

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.079	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.100	02:05
Breath Volume		1.371	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.098	02:08
Breath Volume		1.441	02:08
Air Blank		0.000	02:09
Control Test		0.079	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BEHAN,H
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/25/2025 03:10:04

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:07	
Control Test		0.079	03:08	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:09	
Control Test		0.079	03:09	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: LIPSCOMB

DATE/TIME OF TEST: 10/25/2025 05:47:57

Diagnostics Check		OK	05:43	
Air Blank		0.000	05:43	
Control Test		0.079	05:43	
Air Blank		0.000	05:44	
Subject Sample	# 1	0.214	05:45	
Breath Volume		1.542	05:45	
Air Blank		0.000	05:45	
Air Blank			05:47	RFI Detect
Air Blank			05:47	RFI Detect
Air Blank		0.000	05:47	

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: LIPSCOMB

DATE/TIME OF TEST: 10/25/2025 05:57:40

Diagnostics Check		OK	05:51
Air Blank		0.000	05:51
Control Test		0.079	05:52
Air Blank		0.000	05:52
Subject Sample	# 1	0.199	05:53
Breath Volume		1.703	05:53
Air Blank		0.000	05:53
Air Blank		0.000	05:55
Subject Sample	# 2	0.209	05:56
Breath Volume		1.347	05:56
Air Blank		0.000	05:56
Control Test		0.079	05:57
Air Blank		0.000	05:57
Diagnostics Check		OK	05:57

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 10/26/2025 03:21:10

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.167	03:15
Breath Volume		1.144	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.169	03:19
Breath Volume		1.507	03:19
Air Blank		0.000	03:20
Control Test		0.079	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 11/21/2025 13:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 10/26/2025 20:08:44

Diagnostics Check		OK	20:02
Air Blank		0.000	20:03
Control Test		0.078	20:03
Air Blank		0.000	20:03
Subject Sample	# 1	0.098	20:04
Breath Volume		1.757	20:04
Air Blank		0.000	20:04
Air Blank		0.000	20:06
Subject Sample	# 2	0.091	20:07
Breath Volume		1.367	20:07
Air Blank		0.000	20:07
Control Test		0.078	20:08
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 11/21/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 10/26/2025 21:28:28

Diagnostics Check		OK	21:22
Air Blank		0.000	21:22
Control Test		0.078	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.295	21:23
Breath Volume		2.664	21:23
Air Blank		0.000	21:24
Air Blank		0.000	21:26
Subject Sample	# 2	0.300	21:26
Breath Volume		2.906	21:26
Air Blank		0.000	21:27
Control Test		0.078	21:27
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 11/21/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: MAI

DATE/TIME OF TEST: 10/28/2025 20:24:57

Diagnostics Check		OK	20:13	
Air Blank		0.000	20:13	
Control Test		0.079	20:13	
Air Blank		0.000	20:14	
Subject Sample	# 1	0.290	20:17	Volume Not Met
Breath Volume		0.355	20:17	
Air Blank		0.000	20:18	
Air Blank		0.000	20:20	
Subject Sample	# 2	0.291	20:20	No .020 agreement
Breath Volume		1.707	20:20	
Air Blank		0.000	20:21	
Air Blank		0.000	20:23	
Subject Sample	# 3	0.312	20:23	No .020 agreement
Breath Volume		1.539	20:23	
Air Blank		0.000	20:24	
Control Test		0.078	20:24	
Air Blank		0.000	20:24	
Diagnostics Check		OK	20:24	

Data Download Date/Time: 11/21/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: CALVIS

DATE/TIME OF TEST: 10/30/2025 02:16:50

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.148	02:12
Breath Volume		2.101	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.148	02:15
Breath Volume		2.179	02:15
Air Blank		0.000	02:15
Control Test		0.078	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 11/21/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 10/31/2025 01:36:44

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.079	01:28
Air Blank		0.000	01:28
Subject Sample	# 1	0.224	01:29
Breath Volume		2.144	01:29
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.186	01:32
Breath Volume		1.269	01:32
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 3	0.189	01:35
Breath Volume		1.574	01:35
Air Blank		0.000	01:35
Control Test		0.079	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 11/21/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LOFTIN

DATE/TIME OF TEST: 10/31/2025 23:25:21

Diagnostics Check		OK	23:22	
Air Blank		0.000	23:23	
Control Test		0.080	23:23	
Air Blank		0.000	23:23	
Subject Sample	# 1		23:24	Subject Test Refused
Breath Volume		0.000	23:24	
Air Blank		0.000	23:24	
Control Test		0.079	23:24	
Air Blank		0.000	23:25	
Diagnostics Check		OK	23:25	

Data Download Date/Time: 11/21/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,CODY,T
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WHITE,L
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/02/2025 21:28:47

Diagnostics Check		OK	21:22
Air Blank		0.000	21:23
Control Test		0.079	21:23
Air Blank		0.000	21:23
Subject Sample	# 1	0.000	21:24
Breath Volume		3.214	21:24
Air Blank		0.000	21:24
Air Blank		0.000	21:26
Subject Sample	# 2	0.000	21:27
Breath Volume		3.535	21:27
Air Blank		0.000	21:27
Control Test		0.077	21:28
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 10/30/2025 12:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DEVIN,K
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: JONES,ZACHARY
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 10/06/2025 01:12:18

Diagnostics Check		OK	01:09	
Air Blank		0.000	01:09	
Control Test		0.078	01:09	
Air Blank		0.000	01:10	
Subject Sample	# 1		01:10	Subject Test Refused
Breath Volume		0.000	01:10	
Air Blank		0.000	01:11	
Control Test		0.078	01:11	
Air Blank		0.000	01:12	
Diagnostics Check		OK	01:12	

Data Download Date/Time: 10/30/2025 12:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: GARDINER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/18/2025 01:22:08

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.213	01:17
Breath Volume		3.628	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:19
Subject Sample	# 2	0.210	01:20
Breath Volume		4.402	01:20
Air Blank		0.000	01:21
Control Test		0.078	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 10/30/2025 12:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DEVIN,K
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: SANDERS,TAYLOR
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 10/20/2025 01:13:49

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:02	
Control Test		0.079	01:02	
Air Blank		0.000	01:03	
Subject Sample	# 1		01:06	No Sample Provided
Breath Volume		0.000	01:06	
Air Blank		0.000	01:07	
Air Blank		0.000	01:08	
Subject Sample	# 2	0.099	01:12	Volume Not Met
Breath Volume		0.882	01:12	
Air Blank		0.000	01:12	
Control Test		0.079	01:13	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 10/30/2025 12:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DEVIN,K
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: SANDERS,TAYLOR
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 10/20/2025 01:25:11

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.078	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.097	01:19
Breath Volume		1.800	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.094	01:23
Breath Volume		1.410	01:23
Air Blank		0.000	01:24
Control Test		0.077	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 10/30/2025 12:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: NORSWORTHY,BEN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: COVINGTON

DATE/TIME OF TEST: 10/23/2025 21:37:19

Diagnostics Check		OK	21:30
Air Blank		0.000	21:31
Control Test		0.080	21:31
Air Blank		0.000	21:32
Subject Sample	# 1	0.214	21:32
Breath Volume		3.492	21:32
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.199	21:35
Breath Volume		1.855	21:35
Air Blank		0.000	21:36
Control Test		0.080	21:36
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 10/30/2025 12:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELBOGIS,DAVID,R
BREATH TEST OPERATOR AGENCY: Trenton PD
ARRESTING OFFICER: KRUSZUSKI,CONNOR
ARRESTING OFFICER AGENCY: Trenton PD
LAST AGENCY INSPECTION: 10/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: WASWICK

DATE/TIME OF TEST: 10/15/2025 05:24:43

Diagnostics Check		OK	05:21	
Air Blank		0.000	05:22	
Control Test		0.083	05:22	
Air Blank		0.000	05:22	
Subject Sample	# 1		05:23	RFI Detect
Breath Volume		0.000	05:23	
Air Blank		0.000	05:23	
Control Test		0.083	05:24	
Air Blank		0.000	05:24	
Diagnostics Check		OK	05:24	

Data Download Date/Time: 10/15/2025 18:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELBOGIS,DAVID,R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: KRUSZUSKI,CONOR
ARRESTING OFFICER AGENCY: Trenton PD
LAST AGENCY INSPECTION: 10/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: WASWICK

DATE/TIME OF TEST: 10/15/2025 05:33:36

Diagnostics Check		OK	05:27
Air Blank		0.000	05:27
Control Test		0.082	05:28
Air Blank		0.000	05:28
Subject Sample	# 1	0.102	05:29
Breath Volume		3.386	05:29
Air Blank		0.000	05:29
Air Blank		0.000	05:31
Subject Sample	# 2	0.105	05:32
Breath Volume		3.171	05:32
Air Blank		0.000	05:32
Control Test		0.083	05:33
Air Blank		0.000	05:33
Diagnostics Check		OK	05:33

Data Download Date/Time: 10/15/2025 18:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODKIN,JASON,M
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: GODKIN,JASON
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 10/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: MOLES

DATE/TIME OF TEST: 10/17/2025 02:22:09

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.077	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1	0.196	02:17	
Breath Volume		2.570	02:17	
Air Blank		0.000	02:18	
Air Blank		0.000	02:19	
Subject Sample	# 2	0.194	02:20	
Breath Volume		2.152	02:20	
Air Blank		0.000	02:20	
Control Test		0.077	02:21	
Air Blank			02:21	RFI Detect
Air Blank		0.000	02:22	

Data Download Date/Time: 11/26/2025 15:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELBOGIS,DAVID,R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: VANFLEET,BRIAN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 10/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 10/19/2025 19:31:39

Diagnostics Check		OK	19:25
Air Blank		0.000	19:25
Control Test		0.077	19:26
Air Blank		0.000	19:26
Subject Sample	# 1	0.258	19:27
Breath Volume		2.777	19:27
Air Blank		0.000	19:27
Air Blank		0.000	19:29
Subject Sample	# 2	0.260	19:30
Breath Volume		3.121	19:30
Air Blank		0.000	19:30
Control Test		0.078	19:31
Air Blank		0.000	19:31
Diagnostics Check		OK	19:31

Data Download Date/Time: 11/26/2025 15:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,BRANDEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GONZALEZ RODRIGUEZ

DATE/TIME OF TEST: 10/26/2025 04:27:03

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.079	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.109	04:22
Breath Volume		2.035	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.112	04:25
Breath Volume		1.710	04:25
Air Blank		0.000	04:26
Control Test		0.078	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:27

Data Download Date/Time: 11/26/2025 15:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CORMAN,ZACKERY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: FAULCONER

DATE/TIME OF TEST: 10/29/2025 05:05:49

Diagnostics Check		OK	04:56
Air Blank		0.000	04:56
Control Test		0.078	04:57
Air Blank		0.000	04:57
Subject Sample	# 1	0.215	05:00
Breath Volume		1.585	05:00
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.210	05:04
Breath Volume		1.558	05:04
Air Blank		0.000	05:04
Control Test		0.078	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 11/26/2025 15:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITTINGTON,CAROLINE,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 10/29/2025 19:48:06

Diagnostics Check		OK	19:41
Air Blank		0.000	19:42
Control Test		0.078	19:42
Air Blank		0.000	19:43
Subject Sample	# 1	0.155	19:43
Breath Volume		3.660	19:43
Air Blank		0.000	19:44
Air Blank		0.000	19:46
Subject Sample	# 2	0.149	19:46
Breath Volume		4.183	19:46
Air Blank		0.000	19:47
Control Test		0.078	19:47
Air Blank		0.000	19:48
Diagnostics Check		OK	19:48

Data Download Date/Time: 11/26/2025 15:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,BRANDEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 10/31/2025 01:27:10

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.078	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.187	01:22
Breath Volume		1.882	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.190	01:25
Breath Volume		1.953	01:25
Air Blank		0.000	01:26
Control Test		0.079	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 11/26/2025 15:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: SUTTON,R
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 10/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: OTHER
ARREST TIME: 13:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 10/31/2025 14:29:50

Diagnostics Check		OK	14:23
Air Blank		0.000	14:23
Control Test		0.079	14:24
Air Blank		0.000	14:24
Subject Sample	# 1	0.178	14:25
Breath Volume		2.136	14:25
Air Blank		0.000	14:25
Air Blank		0.000	14:27
Subject Sample	# 2	0.189	14:28
Breath Volume		1.937	14:28
Air Blank		0.000	14:28
Control Test		0.079	14:29
Air Blank		0.000	14:29
Diagnostics Check		OK	14:29

Data Download Date/Time: 11/26/2025 15:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,STANLEY,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: PUTNAM,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BOWYER

DATE/TIME OF TEST: 10/04/2025 00:26:00

Diagnostics Check		OK	00:23	
Air Blank		0.000	00:23	
Control Test		0.078	00:23	
Air Blank		0.000	00:24	
Subject Sample	# 1		00:24	Subject Test Refused
Breath Volume		0.000	00:24	
Air Blank		0.000	00:25	
Control Test		0.079	00:25	
Air Blank		0.000	00:25	
Diagnostics Check		OK	00:26	

Data Download Date/Time: 10/06/2025 00:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODFREY,BRANDON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WILSON,STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 10/04/2025 21:28:28

Diagnostics Check		OK	21:21	
Air Blank		0.000	21:22	
Control Test		0.079	21:22	
Air Blank		0.000	21:23	
Subject Sample	# 1	0.181	21:24	
Breath Volume		1.820	21:24	
Air Blank		0.000	21:25	
Air Blank		0.000	21:27	
Subject Sample	# 2	0.170	21:27	
Breath Volume		1.437	21:27	
Air Blank			21:28	RFI Detect
Air Blank		0.000	21:28	

Data Download Date/Time: 10/06/2025 00:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODFREY,BRANDON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WILSON,STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 10/04/2025 22:11:01

Diagnostics Check		OK	22:02
Air Blank		0.000	22:03
Control Test		0.079	22:03
Air Blank		0.000	22:04
Subject Sample	# 1	0.167	22:05
Breath Volume		1.214	22:05
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.177	22:09
Breath Volume		2.117	22:09
Air Blank		0.000	22:10
Control Test		0.079	22:10
Air Blank		0.000	22:10
Diagnostics Check		OK	22:11

Data Download Date/Time: 10/06/2025 00:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ONEILL,BRAYDEN,P
BREATH TEST OPERATOR AGENCY: Milton PD
ARRESTING OFFICER: ONEILL,BRAYDEN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 09/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HOSCH

DATE/TIME OF TEST: 10/05/2025 02:08:02

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.077	02:03
Breath Volume		1.703	02:03
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.080	02:06
Breath Volume		2.023	02:06
Air Blank		0.000	02:07
Control Test		0.079	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:08

Data Download Date/Time: 10/06/2025 00:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 10/08/2025 18:08:33

Diagnostics Check		OK	17:58
Air Blank		0.000	17:59
Control Test		0.077	17:59
Air Blank		0.000	18:00
Subject Sample	# 1	0.227	18:00
Breath Volume		2.375	18:00
Air Blank		0.000	18:01
Air Blank		0.000	18:03
Subject Sample	# 2	0.178	18:03
Breath Volume		1.488	18:03
Air Blank		0.000	18:04
Air Blank		0.000	18:06
Subject Sample	# 3	0.177	18:07
Breath Volume		1.851	18:07
Air Blank		0.000	18:07
Control Test		0.076	18:07
Air Blank		0.000	18:08
Diagnostics Check		OK	18:08

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPER,CADEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPER,CADEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 10/10/2025 17:49:55

Diagnostics Check	OK	17:48	
Air Blank		17:49	Ambient Fail
Air Blank	0.000	17:49	

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPER,CADEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPER,CADEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 10/10/2025 17:58:33

Diagnostics Check		OK	17:52
Air Blank		0.000	17:52
Control Test		0.078	17:52
Air Blank		0.000	17:53
Subject Sample	# 1	0.236	17:54
Breath Volume		2.003	17:54
Air Blank		0.000	17:54
Air Blank		0.000	17:56
Subject Sample	# 2	0.237	17:57
Breath Volume		1.382	17:57
Air Blank		0.000	17:57
Control Test		0.078	17:57
Air Blank		0.000	17:58
Diagnostics Check		OK	17:58

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORENE,AIDAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NORENE,AIDAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BUFFINGTON

DATE/TIME OF TEST: 10/11/2025 02:27:30

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.078	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.232	02:22
Breath Volume		2.085	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.214	02:25
Breath Volume		2.046	02:25
Air Blank		0.000	02:26
Control Test		0.078	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 10/11/2025 05:30:12

Diagnostics Check		OK	05:23
Air Blank		0.000	05:24
Control Test		0.079	05:24
Air Blank		0.000	05:25
Subject Sample	# 1	0.076	05:25
Breath Volume		3.808	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.077	05:28
Breath Volume		3.417	05:28
Air Blank		0.000	05:29
Control Test		0.079	05:29
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAVIS,KAYLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MCCULLOUGH

DATE/TIME OF TEST: 10/11/2025 21:14:33

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:12	
Control Test		0.079	21:12	
Air Blank		0.000	21:12	
Subject Sample	# 1		21:13	Subject Test Refused
Breath Volume		0.000	21:13	
Air Blank		0.000	21:13	
Control Test		0.078	21:13	
Air Blank		0.000	21:14	
Diagnostics Check		OK	21:14	

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORENE,AIDAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NORENE,AIDAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: PONTELLO

DATE/TIME OF TEST: 10/12/2025 06:07:55

Diagnostics Check		OK	06:01
Air Blank		0.000	06:01
Control Test		0.079	06:02
Air Blank		0.000	06:02
Subject Sample	# 1	0.063	06:03
Breath Volume		2.777	06:03
Air Blank		0.000	06:03
Air Blank		0.000	06:05
Subject Sample	# 2	0.058	06:06
Breath Volume		2.128	06:06
Air Blank		0.000	06:06
Control Test		0.079	06:07
Air Blank		0.000	06:07
Diagnostics Check		OK	06:07

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMATEER,TRACY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PALMATEER,TRACY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: POPELY

DATE/TIME OF TEST: 10/16/2025 03:57:43

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.177	03:52
Breath Volume		2.324	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.171	03:56
Breath Volume		2.089	03:56
Air Blank		0.000	03:56
Control Test		0.078	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: KASTER

DATE/TIME OF TEST: 10/17/2025 01:41:00

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.078	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.113	01:36
Breath Volume		3.257	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.110	01:39
Breath Volume		2.445	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:41

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WOODS,KENNETH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/17/2025 21:43:40

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:39	
Control Test		0.078	21:39	
Air Blank		0.000	21:39	
Subject Sample	# 1		21:42	Subject Test Refused
Breath Volume		0.000	21:42	
Air Blank		0.000	21:42	
Control Test		0.079	21:43	
Air Blank		0.000	21:43	
Diagnostics Check		OK	21:43	

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 12:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:36

SUBJECT'S LAST NAME: COWART

DATE/TIME OF TEST: 10/18/2025 14:03:54

Diagnostics Check		OK	13:57
Air Blank		0.000	13:57
Control Test		0.079	13:58
Air Blank		0.000	13:58
Subject Sample	# 1	0.009	13:59
Breath Volume		1.996	13:59
Air Blank		0.000	13:59
Air Blank		0.000	14:01
Subject Sample	# 2	0.009	14:02
Breath Volume		2.351	14:02
Air Blank		0.000	14:02
Control Test		0.079	14:03
Air Blank		0.000	14:03
Diagnostics Check		OK	14:03

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: DELEYER

DATE/TIME OF TEST: 10/20/2025 04:42:01

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:34	
Control Test		0.078	04:34	
Air Blank		0.000	04:34	
Subject Sample	# 1		04:37	No Sample Provided
Breath Volume		0.000	04:37	
Air Blank		0.000	04:38	
Air Blank		0.000	04:40	
Subject Sample	# 2		04:40	Subject Test Refused
Breath Volume		0.000	04:40	
Air Blank		0.000	04:41	
Control Test		0.078	04:41	
Air Blank		0.000	04:41	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: DELEYER

DATE/TIME OF TEST: 10/20/2025 04:50:13

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.078	04:44
Air Blank		0.000	04:44
Subject Sample	# 1	0.274	04:45
Breath Volume		1.367	04:45
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.289	04:48
Breath Volume		1.378	04:48
Air Blank		0.000	04:49
Control Test		0.078	04:49
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAY,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: NAY,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: TRUDEL

DATE/TIME OF TEST: 10/24/2025 22:54:36

Diagnostics Check		OK	22:46
Air Blank		0.000	22:46
Control Test		0.079	22:47
Air Blank		0.000	22:47
Subject Sample	# 1	0.132	22:48
Breath Volume		2.128	22:48
Air Blank		0.000	22:49
Air Blank		0.000	22:50
Subject Sample	# 2	0.129	22:53
Breath Volume		2.085	22:53
Air Blank		0.000	22:53
Control Test		0.078	22:54
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKWOOD,PATRICK,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BLACKWOOD,PATRICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CORCORAN

DATE/TIME OF TEST: 10/25/2025 01:47:13

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.078	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.187	01:42
Breath Volume		3.601	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.186	01:45
Breath Volume		2.972	01:45
Air Blank		0.000	01:46
Control Test		0.077	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAY,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JABLONSKI,MARK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: RHODES

DATE/TIME OF TEST: 10/26/2025 06:01:37

Diagnostics Check		OK	05:52
Air Blank		0.000	05:53
Control Test		0.079	05:53
Air Blank		0.000	05:54
Subject Sample	# 1	0.205	05:56
Breath Volume		3.136	05:56
Air Blank		0.000	05:57
Air Blank		0.000	05:59
Subject Sample	# 2	0.195	06:00
Breath Volume		3.386	06:00
Air Blank		0.000	06:00
Control Test		0.079	06:01
Air Blank		0.000	06:01
Diagnostics Check		OK	06:01

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUISE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 10/27/2025 01:13:55

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:05	
Control Test		0.078	01:06	
Air Blank		0.000	01:06	
Subject Sample	# 1		01:09	No Sample Provided
Breath Volume		0.000	01:09	
Air Blank		0.000	01:10	
Air Blank		0.000	01:12	
Subject Sample	# 2		01:12	Subject Test Refused
Breath Volume		0.000	01:12	
Air Blank		0.000	01:13	
Control Test		0.078	01:13	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODFREY,BRANDON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GODFREY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: SEATON

DATE/TIME OF TEST: 10/30/2025 02:05:46

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:01	
Control Test		0.079	02:01	
Air Blank		0.000	02:01	
Subject Sample	# 1		02:04	Subject Test Refused
Breath Volume		0.000	02:04	
Air Blank		0.000	02:04	
Control Test		0.079	02:05	
Air Blank		0.000	02:05	
Diagnostics Check		OK	02:05	

Data Download Date/Time: 11/05/2025 21:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 10/11/2025 19:40:37

Diagnostics Check		OK	19:37	
Air Blank		0.000	19:38	
Control Test		0.080	19:38	
Air Blank		0.000	19:38	
Subject Sample	# 1		19:39	Subject Test Refused
Breath Volume		0.000	19:39	
Air Blank		0.000	19:39	
Control Test		0.080	19:40	
Air Blank		0.000	19:40	
Diagnostics Check		OK	19:40	

Data Download Date/Time: 10/31/2025 15:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEW,REECE,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DEW,REECE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 14:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 10/17/2025 15:02:23

Diagnostics Check		OK	14:55
Air Blank		0.000	14:55
Control Test		0.081	14:56
Air Blank		0.000	14:56
Subject Sample	# 1	0.174	14:57
Breath Volume		2.425	14:57
Air Blank		0.000	14:58
Air Blank		0.000	15:00
Subject Sample	# 2	0.174	15:00
Breath Volume		2.957	15:00
Air Blank		0.000	15:01
Control Test		0.080	15:01
Air Blank		0.000	15:02
Diagnostics Check		OK	15:02

Data Download Date/Time: 10/31/2025 15:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,GABRIEL
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 10/24/2025 19:21:49

Diagnostics Check		OK	19:17	
Air Blank		0.000	19:18	
Control Test		0.081	19:18	
Air Blank		0.000	19:18	
Subject Sample	# 1		19:20	Subject Test Refused
Breath Volume		0.000	19:20	
Air Blank		0.000	19:20	
Control Test		0.081	19:21	
Air Blank		0.000	19:21	
Diagnostics Check		OK	19:21	

Data Download Date/Time: 10/31/2025 15:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE,SEVEN,J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SIPE,SEVEN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429602
DRY GAS EXPIRY DATE: 10/22/2026
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BARMARE

DATE/TIME OF TEST: 10/03/2025 23:29:39

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:20	
Control Test		0.079	23:20	
Air Blank		0.000	23:21	
Subject Sample	# 1	0.060	23:24	Volume Not Met
Breath Volume		0.949	23:24	
Air Blank		0.000	23:25	
Air Blank		0.000	23:27	
Subject Sample	# 2	0.064	23:27	
Breath Volume		1.582	23:27	
Air Blank			23:28	RFI Detect
Air Blank		0.000	23:28	
Control Test		0.079	23:29	
Air Blank		0.000	23:29	
Diagnostics Check		OK	23:29	

Data Download Date/Time: 10/30/2025 13:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE,SEVEN,J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SIPE,SEVEN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429602
DRY GAS EXPIRY DATE: 10/22/2026
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BARMARE

DATE/TIME OF TEST: 10/03/2025 23:39:52

Diagnostics Check		OK	23:31
Air Blank		0.000	23:31
Control Test		0.079	23:32
Air Blank		0.000	23:32
Subject Sample	# 1	0.056	23:33
Breath Volume		1.160	23:33
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.059	23:38
Breath Volume		1.726	23:38
Air Blank		0.000	23:38
Control Test		0.080	23:39
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 10/30/2025 13:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FURMAN,FREDDIE,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MUTLUTURK,FATIH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429602
DRY GAS EXPIRY DATE: 10/22/2026
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:23

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 10/04/2025 08:46:31

Diagnostics Check	OK	08:45	
Air Blank		08:46	RFI Detect
Air Blank	0.000	08:46	

Data Download Date/Time: 10/30/2025 13:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FURMAN,FREDDIE,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MUTLUTURK,FATIH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429602
DRY GAS EXPIRY DATE: 10/22/2026
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:23

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 10/04/2025 08:54:22

Diagnostics Check		OK	08:47
Air Blank		0.000	08:48
Control Test		0.078	08:48
Air Blank		0.000	08:49
Subject Sample	# 1	0.189	08:49
Breath Volume		2.816	08:49
Air Blank		0.000	08:50
Air Blank		0.000	08:52
Subject Sample	# 2	0.190	08:52
Breath Volume		1.828	08:52
Air Blank		0.000	08:53
Control Test		0.077	08:53
Air Blank		0.000	08:54
Diagnostics Check		OK	08:54

Data Download Date/Time: 10/30/2025 13:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429602
DRY GAS EXPIRY DATE: 10/22/2026
VIOLATION CODE: DUI
ARREST TIME: 12:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:09

SUBJECT'S LAST NAME: VEARA

DATE/TIME OF TEST: 10/07/2025 13:43:47

Diagnostics Check		OK	13:29	
Air Blank		0.000	13:30	
Control Test		0.078	13:30	
Air Blank		0.000	13:31	
Subject Sample	# 1	0.294	13:33	
Breath Volume		1.796	13:33	
Air Blank		0.000	13:34	
Air Blank		0.000	13:36	
Subject Sample	# 2	0.262	13:39	Volume Not Met
Breath Volume		0.285	13:39	
Air Blank		0.000	13:40	
Air Blank		0.000	13:41	
Subject Sample	# 3		13:42	Subject Test Refused
Breath Volume		0.000	13:42	
Air Blank		0.000	13:42	
Control Test		0.076	13:43	
Air Blank		0.000	13:43	
Diagnostics Check		OK	13:43	

Data Download Date/Time: 10/30/2025 13:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429602
DRY GAS EXPIRY DATE: 10/22/2026
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 10/08/2025 22:09:34

Diagnostics Check		OK	22:00
Air Blank		0.000	22:01
Control Test		0.078	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.100	22:03
Breath Volume		1.218	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.103	22:08
Breath Volume		1.347	22:08
Air Blank		0.000	22:08
Control Test		0.078	22:08
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

Data Download Date/Time: 10/30/2025 13:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE,SEVEN,J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SIPE,SEVEN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429602
DRY GAS EXPIRY DATE: 10/22/2026
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: KARAGIANNAKIS

DATE/TIME OF TEST: 10/19/2025 01:35:25

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.079	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:34	
Control Test		0.079	01:34	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 10/30/2025 13:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,JOHN,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MILLER,ZACHARY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429602
DRY GAS EXPIRY DATE: 10/22/2026
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 10/20/2025 05:27:49

Diagnostics Check		OK	05:20
Air Blank		0.000	05:21
Control Test		0.077	05:21
Air Blank		0.000	05:21
Subject Sample	# 1	0.188	05:23
Breath Volume		2.730	05:23
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.196	05:26
Breath Volume		2.640	05:26
Air Blank		0.000	05:26
Control Test		0.077	05:27
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 10/30/2025 13:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIF,HAMPTON,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/05/2025 01:28:23

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.081	01:26	
Air Blank		0.000	01:26	
Subject Sample	# 1		01:27	Subject Test Refused
Breath Volume		0.000	01:27	
Air Blank		0.000	01:27	
Control Test		0.081	01:27	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 10/13/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDENAS,ADAM,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TISDALE,J
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: PENTZ

DATE/TIME OF TEST: 10/05/2025 20:59:48

Diagnostics Check		OK	20:55	
Air Blank		0.000	20:55	
Control Test		0.081	20:55	
Air Blank		0.000	20:56	
Subject Sample	# 1		20:58	Subject Test Refused
Breath Volume		0.000	20:58	
Air Blank		0.000	20:58	
Control Test		0.081	20:59	
Air Blank		0.000	20:59	
Diagnostics Check		OK	20:59	

Data Download Date/Time: 10/13/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,JUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ELRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: SCHOFIELD

DATE/TIME OF TEST: 10/08/2025 16:10:10

Diagnostics Check		OK	16:03
Air Blank		0.000	16:03
Control Test		0.080	16:04
Air Blank		0.000	16:04
Subject Sample	# 1	0.000	16:05
Breath Volume		2.550	16:05
Air Blank		0.000	16:06
Air Blank		0.000	16:08
Subject Sample	# 2	0.000	16:08
Breath Volume		3.441	16:08
Air Blank		0.000	16:09
Control Test		0.082	16:09
Air Blank		0.000	16:10
Diagnostics Check		OK	16:10

Data Download Date/Time: 10/13/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERY,JOHN,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: AVERY,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 13:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:22

SUBJECT'S LAST NAME: JEFFERSON

DATE/TIME OF TEST: 10/10/2025 15:45:57

Diagnostics Check		OK	15:43	
Air Blank		0.000	15:43	
Control Test		0.081	15:43	
Air Blank		0.000	15:44	
Subject Sample	# 1		15:44	Subject Test Refused
Breath Volume		0.000	15:44	
Air Blank		0.000	15:45	
Control Test		0.080	15:45	
Air Blank		0.000	15:45	
Diagnostics Check		OK	15:45	

Data Download Date/Time: 10/13/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JOHN,AVERY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 10:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 10/12/2025 11:28:07

Diagnostics Check		OK	11:20
Air Blank		0.000	11:21
Control Test		0.080	11:21
Air Blank		0.000	11:21
Subject Sample	# 1	0.196	11:23
Breath Volume		1.515	11:23
Air Blank		0.000	11:24
Air Blank		0.000	11:25
Subject Sample	# 2	0.205	11:26
Breath Volume		1.574	11:26
Air Blank		0.000	11:27
Control Test		0.080	11:27
Air Blank		0.000	11:28
Diagnostics Check		OK	11:28

Data Download Date/Time: 10/13/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRABTREE,JEREMIAH,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CRABTREE,JEREMIAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: PETRILLO

DATE/TIME OF TEST: 10/12/2025 23:05:08

Diagnostics Check		OK	23:02	
Air Blank		0.000	23:02	
Control Test		0.081	23:02	
Air Blank		0.000	23:03	
Subject Sample	# 1		23:03	Subject Test Refused
Breath Volume		0.000	23:03	
Air Blank		0.000	23:04	
Control Test		0.080	23:04	
Air Blank		0.000	23:05	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 10/13/2025 13:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,JUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: TISDALE,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: FILS

DATE/TIME OF TEST: 10/14/2025 22:16:45

Diagnostics Check		OK	22:09
Air Blank		0.000	22:09
Control Test		0.080	22:10
Air Blank		0.000	22:10
Subject Sample	# 1	0.253	22:11
Breath Volume		1.425	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.258	22:15
Breath Volume		1.753	22:15
Air Blank		0.000	22:15
Control Test		0.080	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 11/21/2025 16:05

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,TRISTAN,T
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CASON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/13/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 10/22/2025 01:57:03

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.188	01:52
Breath Volume		1.359	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:54
Subject Sample	# 2	0.178	01:55
Breath Volume		1.140	01:55
Air Blank		0.000	01:56
Control Test		0.079	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:57

Data Download Date/Time: 11/21/2025 16:05

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAFFA,AMELIA,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BAFFA,AMELIA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: MATACCHIERO

DATE/TIME OF TEST: 10/02/2025 09:16:48

Diagnostics Check		OK	09:10
Air Blank		0.000	09:10
Control Test		0.079	09:11
Air Blank		0.000	09:11
Subject Sample	# 1	0.000	09:12
Breath Volume		2.289	09:12
Air Blank		0.000	09:12
Air Blank		0.000	09:14
Subject Sample	# 2	0.000	09:15
Breath Volume		2.085	09:15
Air Blank		0.000	09:15
Control Test		0.080	09:16
Air Blank		0.000	09:16
Diagnostics Check		OK	09:16

Data Download Date/Time: 10/17/2025 09:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: WHITE,NICKOLAUS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 11:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/04/2025 12:33:26

Diagnostics Check		OK	12:25
Air Blank		0.000	12:26
Control Test		0.078	12:26
Air Blank		0.000	12:27
Subject Sample	# 1	0.200	12:28
Breath Volume		2.222	12:28
Air Blank		0.000	12:29
Air Blank		0.000	12:31
Subject Sample	# 2	0.196	12:31
Breath Volume		2.218	12:31
Air Blank		0.000	12:32
Control Test		0.078	12:32
Air Blank		0.000	12:33
Diagnostics Check		OK	12:33

Data Download Date/Time: 10/17/2025 09:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SOWER,FRANK
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 11:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:56

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 10/09/2025 12:23:49

Diagnostics Check		OK	12:17
Air Blank		0.000	12:18
Control Test		0.079	12:18
Air Blank		0.000	12:18
Subject Sample	# 1	0.000	12:19
Breath Volume		2.796	12:19
Air Blank		0.000	12:19
Air Blank		0.000	12:21
Subject Sample	# 2	0.000	12:22
Breath Volume		3.242	12:22
Air Blank		0.000	12:22
Control Test		0.079	12:23
Air Blank		0.000	12:23
Diagnostics Check		OK	12:23

Data Download Date/Time: 10/17/2025 09:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TULLEY,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 10/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: MATACCHIERO

DATE/TIME OF TEST: 10/15/2025 09:30:26

Diagnostics Check		OK	09:24
Air Blank		0.000	09:24
Control Test		0.079	09:25
Air Blank		0.000	09:25
Subject Sample	# 1	0.000	09:26
Breath Volume		1.214	09:26
Air Blank		0.000	09:26
Air Blank		0.000	09:28
Subject Sample	# 2	0.000	09:29
Breath Volume		1.230	09:29
Air Blank		0.000	09:29
Control Test		0.080	09:29
Air Blank		0.000	09:30
Diagnostics Check		OK	09:30

Data Download Date/Time: 10/17/2025 09:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORRESTER,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARMAN,CODY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 10/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: LOCHER

DATE/TIME OF TEST: 10/19/2025 03:50:39

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:36	
Control Test		0.080	03:37	
Air Blank		0.000	03:37	
Subject Sample	# 1	0.176	03:38	
Breath Volume		1.933	03:38	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2		03:43	No Sample Provided
Breath Volume		0.000	03:43	
Air Blank		0.000	03:44	
Air Blank		0.000	03:46	
Subject Sample	# 3		03:49	No Sample Provided
Breath Volume		0.000	03:49	
Air Blank		0.000	03:49	
Control Test		0.079	03:50	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 11/21/2025 17:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIF,HAMPTON,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 10/13/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 10:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 10/28/2025 12:13:55

Diagnostics Check		OK	12:11	
Air Blank		0.000	12:11	
Control Test		0.079	12:11	
Air Blank		0.000	12:12	
Subject Sample	# 1		12:12	Subject Test Refused
Breath Volume		0.000	12:12	
Air Blank		0.000	12:13	
Control Test		0.079	12:13	
Air Blank		0.000	12:13	
Diagnostics Check		OK	12:13	

Data Download Date/Time: 11/21/2025 17:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,MARK,K
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: HERNANDEZ,RICARDO
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BALTAZAR

DATE/TIME OF TEST: 10/04/2025 00:09:16

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:06	
Control Test		0.080	00:07	
Air Blank		0.000	00:07	
Subject Sample	# 1		00:07	Subject Test Refused
Breath Volume		0.000	00:07	
Air Blank		0.000	00:08	
Control Test		0.080	00:08	
Air Blank		0.000	00:09	
Diagnostics Check		OK	00:09	

Data Download Date/Time: 10/24/2025 17:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,MARK,K
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: CABRERA,VINCENT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: LAW

DATE/TIME OF TEST: 10/05/2025 03:44:07

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.081	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.000	03:39
Breath Volume		2.332	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.000	03:42
Breath Volume		2.277	03:42
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 10/24/2025 17:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTTARI,LOGAN,H
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: DONNELLY,COLIN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 10/05/2025 06:17:29

Diagnostics Check		OK	06:10
Air Blank		0.000	06:11
Control Test		0.081	06:11
Air Blank		0.000	06:12
Subject Sample	# 1	0.147	06:12
Breath Volume		2.703	06:12
Air Blank		0.000	06:13
Air Blank		0.000	06:15
Subject Sample	# 2	0.136	06:15
Breath Volume		1.316	06:15
Air Blank		0.000	06:16
Control Test		0.080	06:16
Air Blank		0.000	06:17
Diagnostics Check		OK	06:17

Data Download Date/Time: 10/24/2025 17:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTTARI,LOGAN,H
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: HULL,ANTHONY
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/14/2025 07:44:44

Diagnostics Check		OK	07:31	
Air Blank		0.000	07:32	
Control Test		0.081	07:32	
Air Blank		0.000	07:33	
Subject Sample	# 1	0.107	07:36	Volume Not Met
Breath Volume		0.050	07:36	
Air Blank		0.000	07:36	
Air Blank		0.000	07:38	
Subject Sample	# 2	0.140	07:39	
Breath Volume		2.988	07:39	
Air Blank		0.000	07:40	
Air Blank		0.000	07:42	
Subject Sample	# 3	0.138	07:43	
Breath Volume		2.242	07:43	
Air Blank		0.000	07:43	
Control Test		0.080	07:44	
Air Blank		0.000	07:44	
Diagnostics Check		OK	07:44	

Data Download Date/Time: 10/24/2025 17:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 10/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 11:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:32

SUBJECT'S LAST NAME: FAIRBANKS

DATE/TIME OF TEST: 10/30/2025 12:00:13

Diagnostics Check		OK	11:52
Air Blank		0.000	11:53
Control Test		0.081	11:53
Air Blank		0.000	11:53
Subject Sample	# 1	0.011	11:55
Breath Volume		2.019	11:55
Air Blank		0.000	11:56
Air Blank		0.000	11:58
Subject Sample	# 2	0.011	11:58
Breath Volume		1.539	11:58
Air Blank		0.000	11:59
Control Test		0.080	11:59
Air Blank		0.000	12:00
Diagnostics Check		OK	12:00

Data Download Date/Time: 11/23/2025 19:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMCZYK,STENLY,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TOMCZYK,STENLY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 10/02/2025 20:16:28

Diagnostics Check		OK	20:10
Air Blank		0.000	20:10
Control Test		0.080	20:10
Air Blank		0.000	20:11
Subject Sample	# 1	0.000	20:11
Breath Volume		1.121	20:11
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.000	20:15
Breath Volume		1.503	20:15
Air Blank		0.000	20:15
Control Test		0.079	20:15
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 10/28/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDES,MARCO,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: ADAMS,THOMAS
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: POMPEY

DATE/TIME OF TEST: 10/04/2025 15:23:01

Diagnostics Check		OK	15:21	
Air Blank		0.000	15:21	
Control Test		0.079	15:21	
Air Blank		0.000	15:22	
Subject Sample	# 1		15:22	Sequence Aborted
Breath Volume		0.000	15:22	
Air Blank			15:23	Improper Sample

Data Download Date/Time: 10/28/2025 15:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDES,MARCO,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: ADAMS,THOMAS
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: POMPEY

DATE/TIME OF TEST: 10/04/2025 15:26:10

Diagnostics Check	OK	15:25	
Air Blank		15:25	Improper Sample
Air Blank		15:26	Improper Sample

Data Download Date/Time: 10/28/2025 15:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDES,MARCO,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: ADAMS,THOMAS
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: POMPEY

DATE/TIME OF TEST: 10/04/2025 15:30:11

Diagnostics Check	OK	15:29	
Air Blank		15:30	Improper Sample
Air Blank		15:30	Improper Sample

Data Download Date/Time: 10/28/2025 15:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMCZYK,STENLY,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TOMCZYK,STENLY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: LAPON

DATE/TIME OF TEST: 10/05/2025 01:36:54

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.080	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.000	01:31
Breath Volume		1.218	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.000	01:35
Breath Volume		1.390	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 10/28/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUAGLIA,ERIC,S
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: LAWTON,BRENNAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: EPLEY

DATE/TIME OF TEST: 10/10/2025 23:26:30

Diagnostics Check		OK	23:20
Air Blank		0.000	23:20
Control Test		0.079	23:21
Air Blank		0.000	23:21
Subject Sample	# 1	0.230	23:21
Breath Volume		3.210	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.237	23:24
Breath Volume		2.914	23:24
Air Blank		0.000	23:25
Control Test		0.079	23:25
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 10/28/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,DALTON,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HOWARD,DALTON
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: GAMEZ

DATE/TIME OF TEST: 10/23/2025 00:51:28

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.230	00:46
Breath Volume		2.277	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.220	00:49
Breath Volume		2.808	00:49
Air Blank		0.000	00:50
Control Test		0.078	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 10/28/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUAGLIA,ERIC,S
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: LAWTON,BRENNAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 10/24/2025 19:47:46

Diagnostics Check		OK	19:40
Air Blank		0.000	19:41
Control Test		0.080	19:41
Air Blank		0.000	19:42
Subject Sample	# 1	0.024	19:42
Breath Volume		2.664	19:42
Air Blank		0.000	19:43
Air Blank		0.000	19:45
Subject Sample	# 2	0.024	19:46
Breath Volume		2.554	19:46
Air Blank		0.000	19:46
Control Test		0.080	19:47
Air Blank		0.000	19:47
Diagnostics Check		OK	19:47

Data Download Date/Time: 10/28/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTISTELLI,TAYLOR,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APREA,STEFANO
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/26/2025 03:12:02

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.074	03:11	Control Outside Tolerance
Air Blank	0.000	03:12	

Data Download Date/Time: 10/28/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTISTELLI,TAYLOR,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APREA,STEFANO
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/26/2025 03:28:36

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:15	
Control Test		0.079	03:15	
Air Blank		0.000	03:16	
Subject Sample	# 1	0.205	03:19	Volume Not Met
Breath Volume		1.386	03:19	
Air Blank		0.000	03:20	
Air Blank		0.000	03:22	
Subject Sample	# 2	0.229	03:23	
Breath Volume		2.507	03:23	
Air Blank		0.000	03:24	
Air Blank		0.000	03:25	
Subject Sample	# 3	0.229	03:26	
Breath Volume		2.253	03:26	
Air Blank		0.000	03:27	
Control Test		0.078	03:28	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 10/28/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMCZYK,STENLY,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TOMCZYK,STENLY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: DRAIN

DATE/TIME OF TEST: 10/26/2025 03:44:16

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:31	
Control Test		0.078	03:32	
Air Blank		0.000	03:32	
Subject Sample	# 1	0.062	03:33	
Breath Volume		1.167	03:33	
Air Blank		0.000	03:34	
Air Blank		0.000	03:36	
Subject Sample	# 2	0.061	03:39	Volume Not Met
Breath Volume		0.785	03:39	
Air Blank		0.000	03:39	
Air Blank		0.000	03:41	
Subject Sample	# 3	0.061	03:42	
Breath Volume		1.824	03:42	
Air Blank		0.000	03:43	
Control Test		0.078	03:43	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 10/28/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: WARD,AUBREE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WILHOIT

DATE/TIME OF TEST: 10/28/2025 00:02:12

Diagnostics Check		OK	23:55
Air Blank		0.000	23:56
Control Test		0.079	23:56
Air Blank		0.000	23:56
Subject Sample	# 2	0.084	23:57
Breath Volume		3.156	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.085	00:00
Breath Volume		3.351	00:00
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 10/28/2025 15:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ORTEGA,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 10/02/2025 20:56:02

Diagnostics Check		OK	20:49
Air Blank		0.000	20:50
Control Test		0.083	20:50
Air Blank		0.000	20:51
Subject Sample	# 1	0.147	20:51
Breath Volume		1.968	20:51
Air Blank		0.000	20:52
Air Blank		0.000	20:54
Subject Sample	# 2	0.154	20:54
Breath Volume		2.210	20:54
Air Blank		0.000	20:55
Control Test		0.082	20:55
Air Blank		0.000	20:55
Diagnostics Check		OK	20:56

Data Download Date/Time: 10/14/2025 17:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BINNER,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 10/02/2025 21:57:52

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:55	
Control Test		0.082	21:55	
Air Blank		0.000	21:56	
Subject Sample	# 1		21:56	Subject Test Refused
Breath Volume		0.000	21:56	
Air Blank		0.000	21:56	
Control Test		0.082	21:57	
Air Blank		0.000	21:57	
Diagnostics Check		OK	21:57	

Data Download Date/Time: 10/14/2025 17:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SCHULTZ,Z
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: DONZELLA

DATE/TIME OF TEST: 10/03/2025 01:08:34

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.083	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.175	01:03
Breath Volume		2.800	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:05
Subject Sample	# 2	0.187	01:06
Breath Volume		3.828	01:06
Air Blank		0.000	01:07
Control Test		0.082	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 10/14/2025 17:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,BRITTANY,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MUNNI,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 09:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/03/2025 11:12:02

Diagnostics Check		OK	11:06
Air Blank		0.000	11:06
Control Test		0.083	11:06
Air Blank		0.000	11:07
Subject Sample	# 1	0.000	11:07
Breath Volume		1.269	11:07
Air Blank		0.000	11:08
Air Blank		0.000	11:10
Subject Sample	# 2	0.000	11:10
Breath Volume		2.117	11:10
Air Blank		0.000	11:11
Control Test		0.084	11:11
Air Blank		0.000	11:11
Diagnostics Check		OK	11:12

Data Download Date/Time: 10/14/2025 17:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KRAMER,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: ALFORD

DATE/TIME OF TEST: 10/04/2025 04:03:41

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.083	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.123	03:58
Breath Volume		1.648	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.114	04:02
Breath Volume		2.128	04:02
Air Blank		0.000	04:02
Control Test		0.082	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 10/14/2025 17:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,R,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SCHULTZ,Z
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CHICO

DATE/TIME OF TEST: 10/05/2025 01:54:00

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.083	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.158	01:48
Breath Volume		2.558	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.146	01:52
Breath Volume		2.722	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:54

Data Download Date/Time: 10/14/2025 17:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAZARES,MARGARITA,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRUMLEY,W
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BIONDI

DATE/TIME OF TEST: 10/05/2025 02:31:38

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.081	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.175	02:27
Breath Volume		3.117	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.161	02:29
Breath Volume		2.710	02:29
Air Blank		0.000	02:30
Control Test		0.081	02:31
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 10/14/2025 17:14

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CHOUNARD,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 06:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 10/05/2025 08:06:49

Diagnostics Check		OK	08:02	
Air Blank		0.000	08:02	
Control Test		0.082	08:02	
Air Blank		0.000	08:03	
Subject Sample	# 1		08:05	Subject Test Refused
Breath Volume		0.109	08:05	
Air Blank		0.000	08:05	
Control Test		0.081	08:06	
Air Blank		0.000	08:06	
Diagnostics Check		OK	08:06	

Data Download Date/Time: 10/14/2025 17:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DAVIS,MATTHEW
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 10/10/2025 01:35:08

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.083	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:33	Subject Test Refused
Breath Volume		0.000	01:33	
Air Blank		0.000	01:34	
Control Test		0.082	01:34	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 10/14/2025 17:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GLINES,W
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BOND

DATE/TIME OF TEST: 10/11/2025 22:22:47

Diagnostics Check		OK	22:20	
Air Blank		0.000	22:20	
Control Test		0.083	22:20	
Air Blank		0.000	22:21	
Subject Sample	# 1		22:21	Subject Test Refused
Breath Volume		0.000	22:21	
Air Blank		0.000	22:21	
Control Test		0.081	22:22	
Air Blank		0.000	22:22	
Diagnostics Check		OK	22:22	

Data Download Date/Time: 10/14/2025 17:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NEILL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 10/12/2025 04:35:18

Diagnostics Check		OK	04:32	
Air Blank		0.000	04:32	
Control Test		0.083	04:33	
Air Blank		0.000	04:33	
Subject Sample	# 1		04:33	Subject Test Refused
Breath Volume		0.000	04:33	
Air Blank		0.000	04:34	
Control Test		0.082	04:34	
Air Blank		0.000	04:35	
Diagnostics Check		OK	04:35	

Data Download Date/Time: 10/14/2025 17:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HACKETT,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: PUTNEY

DATE/TIME OF TEST: 10/13/2025 00:07:13

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:04	
Control Test		0.083	00:05	
Air Blank		0.000	00:05	
Subject Sample	# 1		00:05	Subject Test Refused
Breath Volume		0.000	00:05	
Air Blank		0.000	00:06	
Control Test		0.080	00:06	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 10/14/2025 17:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HACKETT,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 10/13/2025 04:17:16

Diagnostics Check		OK	04:14	
Air Blank		0.000	04:14	
Control Test		0.083	04:15	
Air Blank		0.000	04:15	
Subject Sample	# 1		04:15	Subject Test Refused
Breath Volume		0.000	04:15	
Air Blank		0.000	04:16	
Control Test		0.081	04:16	
Air Blank		0.000	04:17	
Diagnostics Check		OK	04:17	

Data Download Date/Time: 10/14/2025 17:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ESTRELLA,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: RODRIGUES

DATE/TIME OF TEST: 10/13/2025 05:45:58

Diagnostics Check		OK	05:39
Air Blank		0.000	05:40
Control Test		0.083	05:40
Air Blank		0.000	05:41
Subject Sample	# 1	0.000	05:41
Breath Volume		2.476	05:41
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.000	05:44
Breath Volume		2.820	05:44
Air Blank		0.000	05:45
Control Test		0.083	05:45
Air Blank		0.000	05:45
Diagnostics Check		OK	05:45

Data Download Date/Time: 10/14/2025 17:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CONSALO,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 10/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: ARNONE

DATE/TIME OF TEST: 10/15/2025 19:19:34

Diagnostics Check		OK	19:16	
Air Blank		0.000	19:17	
Control Test		0.083	19:17	
Air Blank		0.000	19:18	
Subject Sample	# 1		19:18	Subject Test Refused
Breath Volume		0.000	19:18	
Air Blank		0.000	19:18	
Control Test		0.080	19:18	
Air Blank		0.000	19:19	
Diagnostics Check		OK	19:19	

Data Download Date/Time: 11/25/2025 16:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 10/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 10/16/2025 23:22:30

Diagnostics Check		OK	23:16
Air Blank		0.000	23:16
Control Test		0.083	23:17
Air Blank		0.000	23:17
Subject Sample	# 1	0.093	23:18
Breath Volume		2.742	23:18
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.090	23:20
Breath Volume		1.863	23:20
Air Blank		0.000	23:21
Control Test		0.082	23:21
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 11/25/2025 16:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PITTMAN,M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 10/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/18/2025 04:05:00

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.082	04:02	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:03	Subject Test Refused
Breath Volume		0.000	04:03	
Air Blank		0.000	04:04	
Control Test		0.082	04:04	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 11/25/2025 16:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HACKETT,BRADLEY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 10/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: SINCLAIR

DATE/TIME OF TEST: 10/21/2025 00:12:01

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.083	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.285	00:07
Breath Volume		2.929	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.277	00:10
Breath Volume		3.285	00:10
Air Blank		0.000	00:11
Control Test		0.082	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:12

Data Download Date/Time: 11/25/2025 16:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 10/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: BYKOWSKY

DATE/TIME OF TEST: 10/21/2025 21:17:38

Diagnostics Check		OK	21:14	
Air Blank		0.000	21:15	
Control Test		0.082	21:15	
Air Blank		0.000	21:16	
Subject Sample	# 1		21:16	Subject Test Refused
Breath Volume		0.000	21:16	
Air Blank		0.000	21:16	
Control Test		0.081	21:17	
Air Blank		0.000	21:17	
Diagnostics Check		OK	21:17	

Data Download Date/Time: 11/25/2025 16:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GADACZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 10/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: MCINTYRE

DATE/TIME OF TEST: 10/21/2025 23:29:11

Diagnostics Check		OK	23:22
Air Blank		0.000	23:23
Control Test		0.083	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.264	23:24
Breath Volume		1.980	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:26
Subject Sample	# 2	0.253	23:27
Breath Volume		2.003	23:27
Air Blank		0.000	23:28
Control Test		0.081	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 11/25/2025 16:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SCOTT,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 10/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 10/31/2025 04:42:06

Diagnostics Check		OK	04:35
Air Blank		0.000	04:36
Control Test		0.083	04:36
Air Blank		0.000	04:37
Subject Sample	# 1	0.029	04:37
Breath Volume		4.074	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.028	04:40
Breath Volume		3.039	04:40
Air Blank		0.000	04:41
Control Test		0.083	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 11/25/2025 16:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: HUGGINS

DATE/TIME OF TEST: 10/04/2025 02:27:48

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.081	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.240	02:23
Breath Volume		1.964	02:23
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.250	02:26
Breath Volume		1.675	02:26
Air Blank		0.000	02:26
Control Test		0.081	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 10/29/2025 09:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/05/2025 03:33:46

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.082	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.085	03:28
Breath Volume		1.222	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.088	03:32
Breath Volume		1.472	03:32
Air Blank		0.000	03:32
Control Test		0.081	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 10/29/2025 09:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAWGER,JACOB,L
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SHAWGER,JACOB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: LARRY

DATE/TIME OF TEST: 10/11/2025 03:49:53

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:38	
Control Test		0.081	03:39	
Air Blank		0.000	03:39	
Subject Sample	# 1	0.076	03:42	Volume Not Met
Breath Volume		0.082	03:42	
Air Blank		0.000	03:43	
Air Blank		0.000	03:45	
Subject Sample	# 2	0.085	03:48	Volume Not Met
Breath Volume		0.082	03:48	
Air Blank		0.000	03:48	
Control Test		0.082	03:49	
Air Blank		0.000	03:49	
Diagnostics Check		OK	03:49	

Data Download Date/Time: 10/29/2025 09:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAWGER,JACOB,L
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SHAWGER,JACOB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: LARRY

DATE/TIME OF TEST: 10/11/2025 04:08:59

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:57	
Control Test		0.080	03:58	
Air Blank		0.000	03:58	
Subject Sample	# 1	0.080	04:01	Volume Not Met
Breath Volume		0.046	04:01	
Air Blank		0.000	04:02	
Air Blank		0.000	04:04	
Subject Sample	# 2	0.079	04:07	Volume Not Met
Breath Volume		0.308	04:07	
Air Blank		0.000	04:08	
Control Test		0.081	04:08	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 10/29/2025 09:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/18/2025 02:49:37

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.081	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.109	02:45
Breath Volume		1.656	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.106	02:48
Breath Volume		2.222	02:48
Air Blank		0.000	02:48
Control Test		0.081	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 10/29/2025 09:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: MACEDO PINEDA

DATE/TIME OF TEST: 10/19/2025 03:00:26

Diagnostics Check	OK	02:59	
Air Blank	0.000	03:00	
Control Test	0.000	03:00	RFI Detect
Air Blank		03:00	RFI Detect

Data Download Date/Time: 10/29/2025 09:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: MACEDO PINEDA

DATE/TIME OF TEST: 10/19/2025 03:09:18

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.081	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.186	03:04
Breath Volume		1.972	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.174	03:07
Breath Volume		1.351	03:07
Air Blank		0.000	03:08
Control Test		0.081	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 10/29/2025 09:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:09

SUBJECT'S LAST NAME: PINA SANTANDER

DATE/TIME OF TEST: 10/19/2025 12:35:28

Diagnostics Check		OK	12:29
Air Blank		0.000	12:29
Control Test		0.081	12:30
Air Blank		0.000	12:30
Subject Sample	# 1	0.244	12:31
Breath Volume		2.691	12:31
Air Blank		0.000	12:31
Air Blank		0.000	12:33
Subject Sample	# 2	0.250	12:33
Breath Volume		2.113	12:33
Air Blank		0.000	12:34
Control Test		0.080	12:34
Air Blank		0.000	12:35
Diagnostics Check		OK	12:35

Data Download Date/Time: 10/29/2025 09:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 10/22/2025 20:15:54

Diagnostics Check		OK	20:04	
Air Blank		0.000	20:04	
Control Test		0.081	20:05	
Air Blank		0.000	20:05	
Subject Sample	# 1	0.257	20:08	Volume Not Met
Breath Volume		0.203	20:08	
Air Blank		0.000	20:09	
Air Blank		0.000	20:11	
Subject Sample	# 2	0.280	20:14	Volume Not Met
Breath Volume		0.222	20:14	
Air Blank		0.000	20:14	
Control Test		0.080	20:15	
Air Blank		0.000	20:15	
Diagnostics Check		OK	20:15	

Data Download Date/Time: 10/29/2025 09:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 10/22/2025 21:01:17

Diagnostics Check		OK	20:49	
Air Blank		0.000	20:50	
Control Test		0.080	20:50	
Air Blank		0.000	20:50	
Subject Sample	# 1	0.259	20:54	Volume Not Met
Breath Volume		0.500	20:54	
Air Blank		0.000	20:54	
Air Blank		0.000	20:56	
Subject Sample	# 2	0.239	20:59	Volume Not Met
Breath Volume		0.148	20:59	
Air Blank		0.000	21:00	
Control Test		0.080	21:00	
Air Blank		0.000	21:01	
Diagnostics Check		OK	21:01	

Data Download Date/Time: 10/29/2025 09:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 10/23/2025 03:36:12

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.081	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.071	03:31
Breath Volume		3.027	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.069	03:34
Breath Volume		2.281	03:34
Air Blank		0.000	03:35
Control Test		0.080	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 10/29/2025 09:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: SEARS

DATE/TIME OF TEST: 10/23/2025 18:21:44

Diagnostics Check		OK	18:15
Air Blank		0.000	18:16
Control Test		0.081	18:16
Air Blank		0.000	18:16
Subject Sample	# 1	0.273	18:17
Breath Volume		2.054	18:17
Air Blank		0.000	18:17
Air Blank		0.000	18:19
Subject Sample	# 2	0.287	18:20
Breath Volume		1.335	18:20
Air Blank		0.000	18:20
Control Test		0.080	18:21
Air Blank		0.000	18:21
Diagnostics Check		OK	18:21

Data Download Date/Time: 10/29/2025 09:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITZGERALD,FRANK,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: HERMANN,MATTHEW
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 10/28/2025 15:46:15

Diagnostics Check		OK	15:40
Air Blank		0.000	15:40
Control Test		0.081	15:40
Air Blank		0.000	15:41
Subject Sample	# 1	0.000	15:42
Breath Volume		3.761	15:42
Air Blank		0.000	15:42
Air Blank		0.000	15:44
Subject Sample	# 2	0.000	15:44
Breath Volume		2.421	15:44
Air Blank		0.000	15:45
Control Test		0.081	15:45
Air Blank		0.000	15:46
Diagnostics Check		OK	15:46

Data Download Date/Time: 10/29/2025 09:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,XAVIER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CHAN,JORDAN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: WILT

DATE/TIME OF TEST: 10/02/2025 06:27:16

Diagnostics Check		OK	06:20
Air Blank		0.000	06:20
Control Test		0.079	06:20
Air Blank		0.000	06:21
Subject Sample	# 1	0.034	06:22
Breath Volume		1.964	06:22
Air Blank		0.000	06:23
Air Blank		0.000	06:25
Subject Sample	# 2	0.031	06:25
Breath Volume		1.785	06:25
Air Blank		0.000	06:26
Control Test		0.078	06:26
Air Blank		0.000	06:27
Diagnostics Check		OK	06:27

Data Download Date/Time: 10/31/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTKO,MARIAM,L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: LAWLOR

DATE/TIME OF TEST: 10/02/2025 21:38:23

Diagnostics Check		OK	21:31
Air Blank		0.000	21:32
Control Test		0.080	21:32
Air Blank		0.000	21:33
Subject Sample	# 1	0.000	21:33
Breath Volume		1.218	21:33
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.000	21:37
Breath Volume		1.253	21:37
Air Blank		0.000	21:37
Control Test		0.080	21:37
Air Blank		0.000	21:38
Diagnostics Check		OK	21:38

Data Download Date/Time: 10/31/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: DE IRAOLA CUEBAS

DATE/TIME OF TEST: 10/04/2025 01:40:54

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.080	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.196	01:36
Breath Volume		1.550	01:36
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.197	01:39
Breath Volume		1.433	01:39
Air Blank		0.000	01:40
Control Test		0.077	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 10/31/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: POKROVSKIY

DATE/TIME OF TEST: 10/04/2025 02:34:28

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.080	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.245	02:30
Breath Volume		1.167	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.257	02:32
Breath Volume		1.738	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 10/31/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDES,KAVEN,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MENDES,KAVEN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: VOLK

DATE/TIME OF TEST: 10/04/2025 06:02:38

Diagnostics Check		OK	05:55
Air Blank		0.000	05:56
Control Test		0.079	05:56
Air Blank		0.000	05:57
Subject Sample	# 1	0.127	05:57
Breath Volume		2.753	05:57
Air Blank		0.000	05:58
Air Blank		0.000	06:00
Subject Sample	# 2	0.119	06:01
Breath Volume		2.457	06:01
Air Blank		0.000	06:01
Control Test		0.079	06:02
Air Blank		0.000	06:02
Diagnostics Check		OK	06:02

Data Download Date/Time: 10/31/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYTH,GARY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SMYTH,GARY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: ARELLANOALVARADO

DATE/TIME OF TEST: 10/04/2025 06:19:34

Diagnostics Check		OK	06:13
Air Blank		0.000	06:13
Control Test		0.079	06:14
Air Blank		0.000	06:14
Subject Sample	# 1	0.043	06:15
Breath Volume		3.804	06:15
Air Blank		0.000	06:15
Air Blank		0.000	06:17
Subject Sample	# 2	0.042	06:18
Breath Volume		3.777	06:18
Air Blank		0.000	06:18
Control Test		0.080	06:18
Air Blank		0.000	06:19
Diagnostics Check		OK	06:19

Data Download Date/Time: 10/31/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUGAN,KASEY,L
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: PIETRON,MICHAEL
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: WELSH

DATE/TIME OF TEST: 10/05/2025 00:09:56

Diagnostics Check		OK	00:03
Air Blank		0.000	00:04
Control Test		0.080	00:04
Air Blank		0.000	00:04
Subject Sample	# 1	0.136	00:05
Breath Volume		1.621	00:05
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.143	00:08
Breath Volume		1.601	00:08
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 10/31/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUGAN,KASEY,L
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: ZAPPONE,ANTHONY
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 10/05/2025 00:22:28

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.081	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.055	00:16
Breath Volume		1.292	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.055	00:20
Breath Volume		1.226	00:20
Air Blank		0.000	00:21
Control Test		0.080	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 10/31/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KEUSCH,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: VASQUEZ GUZMZN

DATE/TIME OF TEST: 10/05/2025 06:18:19

Diagnostics Check		OK	06:06	
Air Blank		0.000	06:07	
Control Test		0.080	06:07	
Air Blank		0.000	06:08	
Subject Sample	# 1	0.202	06:11	Volume Not Met
Breath Volume		0.359	06:11	
Air Blank		0.000	06:11	
Air Blank		0.000	06:13	
Subject Sample	# 2		06:16	No Sample Provided
Breath Volume		0.000	06:16	
Air Blank		0.000	06:17	
Control Test		0.079	06:17	
Air Blank		0.000	06:18	
Diagnostics Check		OK	06:18	

Data Download Date/Time: 10/31/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PETROWSKI,DAVID
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: ANDERSEN

DATE/TIME OF TEST: 10/05/2025 17:44:48

Diagnostics Check		OK	17:38
Air Blank		0.000	17:39
Control Test		0.080	17:39
Air Blank		0.000	17:40
Subject Sample	# 1	0.284	17:40
Breath Volume		1.644	17:40
Air Blank		0.000	17:41
Air Blank		0.000	17:42
Subject Sample	# 2	0.290	17:43
Breath Volume		1.332	17:43
Air Blank		0.000	17:43
Control Test		0.079	17:44
Air Blank		0.000	17:44
Diagnostics Check		OK	17:44

Data Download Date/Time: 10/31/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/06/2025 00:08:03

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.081	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.178	00:03
Breath Volume		1.343	00:03
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.174	00:06
Breath Volume		1.796	00:06
Air Blank		0.000	00:07
Control Test		0.079	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:08

Data Download Date/Time: 10/31/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: GRISSOM

DATE/TIME OF TEST: 10/07/2025 15:04:14

Diagnostics Check		OK	14:57
Air Blank		0.000	14:58
Control Test		0.079	14:58
Air Blank		0.000	14:59
Subject Sample	# 1	0.150	14:59
Breath Volume		1.742	14:59
Air Blank		0.000	15:00
Air Blank		0.000	15:02
Subject Sample	# 2	0.159	15:02
Breath Volume		2.214	15:02
Air Blank		0.000	15:03
Control Test		0.080	15:03
Air Blank		0.000	15:04
Diagnostics Check		OK	15:04

Data Download Date/Time: 10/31/2025 14:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 10/08/2025 17:13:16

Diagnostics Check		OK	17:07
Air Blank		0.000	17:07
Control Test		0.081	17:07
Air Blank		0.000	17:08
Subject Sample	# 1	0.000	17:09
Breath Volume		2.191	17:09
Air Blank		0.000	17:09
Air Blank		0.000	17:11
Subject Sample	# 2	0.000	17:11
Breath Volume		2.062	17:11
Air Blank		0.000	17:12
Control Test		0.080	17:12
Air Blank		0.000	17:13
Diagnostics Check		OK	17:13

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: THOMA

DATE/TIME OF TEST: 10/08/2025 19:14:17

Diagnostics Check		OK	19:07
Air Blank		0.000	19:08
Control Test		0.079	19:08
Air Blank		0.000	19:09
Subject Sample	# 1	0.306	19:09
Breath Volume		2.230	19:09
Air Blank		0.000	19:10
Air Blank		0.000	19:12
Subject Sample	# 2	0.289	19:12
Breath Volume		1.218	19:12
Air Blank		0.000	19:13
Control Test		0.080	19:13
Air Blank		0.000	19:14
Diagnostics Check		OK	19:14

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: AKBULUT

DATE/TIME OF TEST: 10/10/2025 03:04:55

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.091	02:58
Breath Volume		1.468	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.102	03:03
Breath Volume		1.410	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTINEZ,OSCAR
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SERRA

DATE/TIME OF TEST: 10/11/2025 02:20:51

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.080	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.184	02:16
Breath Volume		2.171	02:16
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.188	02:19
Breath Volume		3.039	02:19
Air Blank		0.000	02:19
Control Test		0.080	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GOSVENER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: BRAKKEE

DATE/TIME OF TEST: 10/11/2025 19:30:06

Diagnostics Check		OK	19:27	
Air Blank		0.000	19:27	
Control Test		0.081	19:27	
Air Blank		0.000	19:28	
Subject Sample	# 1		19:28	Subject Test Refused
Breath Volume		0.000	19:28	
Air Blank		0.000	19:29	
Control Test		0.080	19:29	
Air Blank		0.000	19:30	
Diagnostics Check		OK	19:30	

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GOSVENER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: BRAKKEE

DATE/TIME OF TEST: 10/11/2025 19:38:06

Diagnostics Check		OK	19:31
Air Blank		0.000	19:32
Control Test		0.081	19:32
Air Blank		0.000	19:33
Subject Sample	# 1	0.000	19:33
Breath Volume		1.417	19:33
Air Blank		0.000	19:34
Air Blank		0.000	19:36
Subject Sample	# 2	0.000	19:36
Breath Volume		2.296	19:36
Air Blank		0.000	19:37
Control Test		0.081	19:37
Air Blank		0.000	19:38
Diagnostics Check		OK	19:38

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: YODER

DATE/TIME OF TEST: 10/12/2025 02:51:50

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.081	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.134	02:47
Breath Volume		1.429	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.135	02:50
Breath Volume		1.550	02:50
Air Blank		0.000	02:50
Control Test		0.080	02:51
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTKO,MARIAM,L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTINEZ,OSCAR
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: YANG

DATE/TIME OF TEST: 10/12/2025 03:53:05

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.081	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.120	03:48
Breath Volume		1.417	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.122	03:51
Breath Volume		1.539	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GOSVENER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: GRANADOS SOTO

DATE/TIME OF TEST: 10/12/2025 04:39:47

Diagnostics Check		OK	04:24	
Air Blank		0.000	04:25	
Control Test		0.080	04:25	
Air Blank		0.000	04:26	
Subject Sample	# 1	0.134	04:29	
Breath Volume		2.175	04:29	
Air Blank		0.000	04:29	
Air Blank		0.000	04:31	
Subject Sample	# 2	0.131	04:34	Volume Not Met
Breath Volume		0.000	04:34	
Air Blank		0.000	04:35	
Air Blank		0.000	04:37	
Subject Sample	# 3	0.137	04:38	
Breath Volume		1.722	04:38	
Air Blank		0.000	04:38	
Control Test		0.080	04:39	
Air Blank		0.000	04:39	
Diagnostics Check		OK	04:39	

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: RANLETT,KALIEGH
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: NOWICKI

DATE/TIME OF TEST: 10/12/2025 04:56:22

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.080	04:51
Air Blank		0.000	04:51
Subject Sample	# 1	0.097	04:52
Breath Volume		2.527	04:52
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.090	04:54
Breath Volume		1.703	04:54
Air Blank		0.000	04:55
Control Test		0.080	04:55
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERR,CALEB,W
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: MORA,RUBEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: DACOSTA

DATE/TIME OF TEST: 10/12/2025 13:20:12

Diagnostics Check		OK	13:13
Air Blank		0.000	13:13
Control Test		0.080	13:14
Air Blank		0.000	13:14
Subject Sample	# 1	0.000	13:15
Breath Volume		2.304	13:15
Air Blank		0.000	13:16
Air Blank		0.000	13:18
Subject Sample	# 2	0.000	13:18
Breath Volume		2.410	13:18
Air Blank		0.000	13:19
Control Test		0.080	13:19
Air Blank		0.000	13:20
Diagnostics Check		OK	13:20

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERSTEN,DANIEL,N
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: NERSTEN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: DIGIULIO

DATE/TIME OF TEST: 10/12/2025 18:18:34

Diagnostics Check		OK	18:12
Air Blank		0.000	18:12
Control Test		0.079	18:13
Air Blank		0.000	18:13
Subject Sample	# 1	0.176	18:13
Breath Volume		2.058	18:13
Air Blank		0.000	18:14
Air Blank		0.000	18:16
Subject Sample	# 2	0.176	18:17
Breath Volume		2.472	18:17
Air Blank		0.000	18:17
Control Test		0.079	18:17
Air Blank		0.000	18:18
Diagnostics Check		OK	18:18

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,JOHAN,X
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ANGELINI,JENINE
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: PISKAC

DATE/TIME OF TEST: 10/14/2025 15:27:13

Diagnostics Check		OK	15:20
Air Blank		0.000	15:21
Control Test		0.080	15:21
Air Blank		0.000	15:22
Subject Sample	# 1	0.137	15:22
Breath Volume		2.687	15:22
Air Blank		0.000	15:23
Air Blank		0.000	15:25
Subject Sample	# 2	0.141	15:25
Breath Volume		2.000	15:25
Air Blank		0.000	15:26
Control Test		0.080	15:26
Air Blank		0.000	15:27
Diagnostics Check		OK	15:27

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: UNDERHILL,TIM
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: LANCE

DATE/TIME OF TEST: 10/15/2025 18:19:36

Diagnostics Check		OK	18:08
Air Blank		0.000	18:09
Control Test		0.080	18:09
Air Blank		0.000	18:10
Subject Sample	# 1	0.257	18:10
Breath Volume		2.121	18:10
Air Blank		0.000	18:11
Air Blank		0.000	18:13
Subject Sample	# 2	0.233	18:14
Breath Volume		1.796	18:14
Air Blank		0.000	18:14
Air Blank		0.000	18:16
Subject Sample	# 3	0.250	18:18
Breath Volume		2.773	18:18
Air Blank		0.000	18:18
Control Test		0.079	18:19
Air Blank		0.000	18:19
Diagnostics Check		OK	18:19

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: LOPEZ VIEYRA

DATE/TIME OF TEST: 10/15/2025 21:31:41

Diagnostics Check		OK	21:25
Air Blank		0.000	21:25
Control Test		0.080	21:25
Air Blank		0.000	21:26
Subject Sample	# 1	0.149	21:26
Breath Volume		1.925	21:26
Air Blank		0.000	21:27
Air Blank		0.000	21:29
Subject Sample	# 2	0.143	21:30
Breath Volume		1.136	21:30
Air Blank		0.000	21:30
Control Test		0.080	21:31
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GOSVENER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: TAPPER

DATE/TIME OF TEST: 10/15/2025 22:14:09

Diagnostics Check		OK	22:08
Air Blank		0.000	22:08
Control Test		0.080	22:08
Air Blank		0.000	22:09
Subject Sample	# 1	0.014	22:09
Breath Volume		2.429	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.014	22:12
Breath Volume		2.363	22:12
Air Blank		0.000	22:13
Control Test		0.081	22:13
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: JONES,NICK
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/16/2025 01:55:23

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.220	01:50
Breath Volume		1.414	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.205	01:53
Breath Volume		1.656	01:53
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/16/2025 02:25:41

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.081	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.176	02:21
Breath Volume		1.628	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.173	02:24
Breath Volume		1.882	02:24
Air Blank		0.000	02:24
Control Test		0.081	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: FOLTZ

DATE/TIME OF TEST: 10/16/2025 21:30:05

Diagnostics Check		OK	21:23
Air Blank		0.000	21:24
Control Test		0.080	21:24
Air Blank		0.000	21:24
Subject Sample	# 1	0.182	21:25
Breath Volume		1.781	21:25
Air Blank		0.000	21:26
Air Blank		0.000	21:28
Subject Sample	# 2	0.192	21:28
Breath Volume		2.281	21:28
Air Blank		0.000	21:29
Control Test		0.080	21:29
Air Blank		0.000	21:30
Diagnostics Check		OK	21:30

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: JONES,NICHOLAS
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: DICHIARA

DATE/TIME OF TEST: 10/17/2025 00:59:47

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.080	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.257	00:55
Breath Volume		3.898	00:55
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.258	00:58
Breath Volume		3.957	00:58
Air Blank		0.000	00:58
Control Test		0.079	00:59
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ANGELINI,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:13

SUBJECT'S LAST NAME: KRUEGER

DATE/TIME OF TEST: 10/17/2025 15:03:32

Diagnostics Check		OK	14:56
Air Blank		0.000	14:57
Control Test		0.080	14:57
Air Blank		0.000	14:58
Subject Sample	# 1	0.241	14:59
Breath Volume		2.304	14:59
Air Blank		0.000	14:59
Air Blank		0.000	15:01
Subject Sample	# 2	0.240	15:01
Breath Volume		2.316	15:01
Air Blank		0.000	15:02
Control Test		0.080	15:02
Air Blank		0.000	15:03
Diagnostics Check		OK	15:03

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: CARMAN

DATE/TIME OF TEST: 10/18/2025 01:52:36

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.081	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.151	01:47
Breath Volume		1.261	01:47
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.144	01:51
Breath Volume		1.734	01:51
Air Blank		0.000	01:51
Control Test		0.079	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: KEY

DATE/TIME OF TEST: 10/18/2025 22:00:50

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.081	21:55
Air Blank		0.000	21:56
Subject Sample	# 1	0.007	21:56
Breath Volume		1.390	21:56
Air Blank		0.000	21:56
Air Blank		0.000	21:58
Subject Sample	# 2	0.007	21:59
Breath Volume		1.847	21:59
Air Blank		0.000	21:59
Control Test		0.081	22:00
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: SIEDLECKI

DATE/TIME OF TEST: 10/18/2025 23:52:45

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:46	
Control Test		0.079	23:46	
Air Blank		0.000	23:47	
Subject Sample	# 1	0.208	23:47	
Breath Volume		1.414	23:47	
Air Blank		0.000	23:48	
Air Blank		0.000	23:50	
Subject Sample	# 2		23:51	Subject Test Refused
Breath Volume		0.000	23:51	
Air Blank		0.000	23:51	
Control Test		0.080	23:52	
Air Blank		0.000	23:52	
Diagnostics Check		OK	23:52	

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CAMBRIA,STEVEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DOWNS

DATE/TIME OF TEST: 10/19/2025 03:43:27

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:38
Subject Sample	# 1	0.152	03:39
Breath Volume		2.562	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.143	03:41
Breath Volume		1.148	03:41
Air Blank		0.000	03:42
Control Test		0.080	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CHAN,JORDAN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/19/2025 05:21:57

Diagnostics Check		OK	05:15
Air Blank		0.000	05:15
Control Test		0.080	05:16
Air Blank		0.000	05:16
Subject Sample	# 1	0.014	05:17
Breath Volume		2.800	05:17
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.013	05:20
Breath Volume		3.292	05:20
Air Blank		0.000	05:21
Control Test		0.080	05:21
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OPATZ,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: OPATZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: SEARS

DATE/TIME OF TEST: 10/19/2025 18:57:30

Diagnostics Check		OK	18:51
Air Blank		0.000	18:51
Control Test		0.080	18:51
Air Blank		0.000	18:52
Subject Sample	# 1	0.363	18:52
Breath Volume		4.101	18:52
Air Blank		0.000	18:53
Air Blank		0.000	18:55
Subject Sample	# 2	0.357	18:56
Breath Volume		3.496	18:56
Air Blank		0.000	18:56
Control Test		0.079	18:56
Air Blank		0.000	18:57
Diagnostics Check		OK	18:57

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CAMBRIA,STEVEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: LANGHAM

DATE/TIME OF TEST: 10/23/2025 01:40:17

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.081	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.179	01:35
Breath Volume		2.207	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.177	01:38
Breath Volume		2.972	01:38
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: MANNING,ZACHARY
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 10/24/2025 21:38:18

Diagnostics Check		OK	21:31
Air Blank		0.000	21:32
Control Test		0.079	21:32
Air Blank		0.000	21:33
Subject Sample	# 1	0.186	21:33
Breath Volume		1.972	21:33
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 2	0.197	21:36
Breath Volume		2.953	21:36
Air Blank		0.000	21:37
Control Test		0.081	21:37
Air Blank		0.000	21:38
Diagnostics Check		OK	21:38

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,JOHAN,X
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ROWELL,JOSHUA
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: SAUTER

DATE/TIME OF TEST: 10/25/2025 16:42:59

Diagnostics Check		OK	16:34
Air Blank		0.000	16:35
Control Test		0.081	16:35
Air Blank		0.000	16:36
Subject Sample	# 1	0.215	16:37
Breath Volume		1.160	16:37
Air Blank		0.000	16:38
Air Blank		0.000	16:39
Subject Sample	# 2	0.235	16:41
Breath Volume		2.070	16:41
Air Blank		0.000	16:42
Control Test		0.079	16:42
Air Blank		0.000	16:42
Diagnostics Check		OK	16:42

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BURLEY,SHAUN
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: NYKAMP

DATE/TIME OF TEST: 10/25/2025 20:58:33

Diagnostics Check		OK	20:51
Air Blank		0.000	20:52
Control Test		0.080	20:52
Air Blank		0.000	20:52
Subject Sample	# 1	0.275	20:53
Breath Volume		1.507	20:53
Air Blank		0.000	20:54
Air Blank		0.000	20:56
Subject Sample	# 2	0.271	20:57
Breath Volume		1.605	20:57
Air Blank		0.000	20:57
Control Test		0.080	20:57
Air Blank		0.000	20:58
Diagnostics Check		OK	20:58

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEATY,JARED,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BEATY,JARED
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: SIEGEL

DATE/TIME OF TEST: 10/25/2025 23:35:54

Diagnostics Check		OK	23:29
Air Blank		0.000	23:30
Control Test		0.080	23:30
Air Blank		0.000	23:30
Subject Sample	# 1	0.146	23:31
Breath Volume		2.781	23:31
Air Blank		0.000	23:32
Air Blank		0.000	23:33
Subject Sample	# 2	0.150	23:34
Breath Volume		2.144	23:34
Air Blank		0.000	23:35
Control Test		0.080	23:35
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEATY,JARED,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GOSVENER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 10/26/2025 00:54:38

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.079	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.205	00:50
Breath Volume		2.265	00:50
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.194	00:53
Breath Volume		2.312	00:53
Air Blank		0.000	00:53
Control Test		0.080	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTKO,MARIAM,L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MOATES

DATE/TIME OF TEST: 10/27/2025 00:19:44

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:05	
Control Test		0.081	00:05	
Air Blank		0.000	00:05	
Subject Sample	# 1	0.114	00:07	
Breath Volume		1.187	00:07	
Air Blank		0.000	00:07	
Air Blank		0.000	00:09	
Subject Sample	# 2	0.093	00:12	Volume Not Met
Breath Volume		0.433	00:12	
Air Blank		0.000	00:13	
Air Blank		0.000	00:15	
Subject Sample	# 3	0.110	00:18	Volume Not Met
Breath Volume		0.535	00:18	
Air Blank		0.000	00:18	
Control Test		0.079	00:19	
Air Blank		0.000	00:19	
Diagnostics Check		OK	00:19	

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/27/2025 01:15:17

Diagnostics Check		OK	00:59	
Air Blank		0.000	00:59	
Control Test		0.079	00:59	
Air Blank		0.000	01:00	
Subject Sample	# 1	0.147	01:03	Volume Not Met
Breath Volume		0.332	01:03	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 2	0.171	01:08	
Breath Volume		1.183	01:08	
Air Blank		0.000	01:08	
Air Blank		0.000	01:10	
Subject Sample	# 3	0.158	01:13	Volume Not Met
Breath Volume		0.343	01:13	
Air Blank		0.000	01:14	
Control Test		0.078	01:14	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CHAN,JORDAN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: CHIVALAN HERNANDEZ

DATE/TIME OF TEST: 10/27/2025 21:38:32

Diagnostics Check		OK	21:32
Air Blank		0.000	21:32
Control Test		0.080	21:32
Air Blank		0.000	21:33
Subject Sample	# 1	0.104	21:34
Breath Volume		2.031	21:34
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 2	0.101	21:37
Breath Volume		2.105	21:37
Air Blank		0.000	21:37
Control Test		0.080	21:37
Air Blank		0.000	21:38
Diagnostics Check		OK	21:38

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: PEREZ RAMOS

DATE/TIME OF TEST: 10/29/2025 02:12:56

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.081	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.232	02:08
Breath Volume		2.039	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:10
Subject Sample	# 2	0.218	02:11
Breath Volume		2.058	02:11
Air Blank		0.000	02:12
Control Test		0.078	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERSTEN,DANIEL,N
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: NERSTEN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 10/30/2025 11:26:43

Diagnostics Check		OK	11:16
Air Blank		0.000	11:16
Control Test		0.081	11:16
Air Blank		0.000	11:17
Subject Sample	# 1	0.138	11:20
Breath Volume		1.484	11:20
Air Blank		0.000	11:20
Air Blank		0.000	11:22
Subject Sample	# 2	0.136	11:25
Breath Volume		1.777	11:25
Air Blank		0.000	11:25
Control Test		0.080	11:26
Air Blank		0.000	11:26
Diagnostics Check		OK	11:26

Data Download Date/Time: 10/31/2025 14:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTINEZ,OSCAR
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: PELICOVELASQUEZ

DATE/TIME OF TEST: 10/31/2025 02:23:38

Diagnostics Check		OK	02:12	
Air Blank		0.000	02:12	
Control Test		0.080	02:13	
Air Blank		0.000	02:13	
Subject Sample	# 1		02:16	No Sample Provided
Breath Volume		0.000	02:16	
Air Blank		0.000	02:17	
Air Blank		0.000	02:19	
Subject Sample	# 2		02:22	No Sample Provided
Breath Volume		0.000	02:22	
Air Blank		0.000	02:22	
Control Test		0.080	02:23	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 10/31/2025 14:44

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

INSTRUMENT REGISTERED TO: BARTOW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: ROQUE,ALEXANDER
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 10/05/2025 22:53:30

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:42	
Control Test		0.079	22:42	
Air Blank		0.000	22:43	
Subject Sample	# 1		22:46	No Sample Provided
Breath Volume		0.000	22:46	
Air Blank		0.000	22:46	
Air Blank		0.000	22:48	
Subject Sample	# 2		22:52	No Sample Provided
Breath Volume		0.000	22:52	
Air Blank		0.000	22:52	
Control Test		0.080	22:52	
Air Blank		0.000	22:53	
Diagnostics Check		OK	22:53	

Data Download Date/Time: 10/19/2025 09:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

INSTRUMENT REGISTERED TO: BARTOW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: ROQUE,ALEXANDER
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 10/10/2025 01:10:16

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.079	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.166	01:04
Breath Volume		2.285	01:04
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.164	01:08
Breath Volume		1.968	01:08
Air Blank		0.000	01:09
Control Test		0.077	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 10/19/2025 09:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

INSTRUMENT REGISTERED TO: BARTOW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAGGETT,TYLER,S
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: ANZALONE,ROBERT
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 10/19/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: MCCALL-HILL

DATE/TIME OF TEST: 10/30/2025 23:40:57

Diagnostics Check		OK	23:33
Air Blank		0.000	23:34
Control Test		0.079	23:34
Air Blank		0.000	23:34
Subject Sample	# 1	0.000	23:36
Breath Volume		1.855	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.000	23:39
Breath Volume		1.335	23:39
Air Blank		0.000	23:40
Control Test		0.079	23:40
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 11/16/2025 08:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 11:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:37

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 10/01/2025 12:04:01

Diagnostics Check		OK	12:01	
Air Blank		0.000	12:01	
Control Test		0.081	12:01	
Air Blank		0.000	12:02	
Subject Sample	# 1		12:02	Subject Test Refused
Breath Volume		0.000	12:02	
Air Blank		0.000	12:03	
Control Test		0.080	12:03	
Air Blank		0.000	12:03	
Diagnostics Check		OK	12:04	

Data Download Date/Time: 10/15/2025 16:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 14:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:44

SUBJECT'S LAST NAME: MAZZOLA

DATE/TIME OF TEST: 10/01/2025 15:17:27

Diagnostics Check		OK	15:11
Air Blank		0.000	15:11
Control Test		0.081	15:12
Air Blank		0.000	15:12
Subject Sample	# 1	0.000	15:13
Breath Volume		1.488	15:13
Air Blank		0.000	15:13
Air Blank		0.000	15:15
Subject Sample	# 2	0.000	15:16
Breath Volume		1.968	15:16
Air Blank		0.000	15:16
Control Test		0.081	15:16
Air Blank		0.000	15:17
Diagnostics Check		OK	15:17

Data Download Date/Time: 10/15/2025 16:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 12:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 10/02/2025 12:52:35

Diagnostics Check		OK	12:46
Air Blank		0.000	12:46
Control Test		0.080	12:46
Air Blank		0.000	12:47
Subject Sample	# 1	0.163	12:47
Breath Volume		1.937	12:47
Air Blank		0.000	12:48
Air Blank		0.000	12:50
Subject Sample	# 2	0.171	12:50
Breath Volume		1.660	12:50
Air Blank		0.000	12:51
Control Test		0.080	12:51
Air Blank		0.000	12:52
Diagnostics Check		OK	12:52

Data Download Date/Time: 10/15/2025 16:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:46

SUBJECT'S LAST NAME: FARINA

DATE/TIME OF TEST: 10/02/2025 15:13:35

Diagnostics Check		OK	15:10	
Air Blank		0.000	15:10	
Control Test		0.080	15:11	
Air Blank		0.000	15:11	
Subject Sample	# 1		15:12	Subject Test Refused
Breath Volume		0.000	15:12	
Air Blank		0.000	15:12	
Control Test		0.080	15:12	
Air Blank		0.000	15:13	
Diagnostics Check		OK	15:13	

Data Download Date/Time: 10/15/2025 16:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TYNDAL,ERICK,D
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: ABRAHAM,MARLO
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/23/2025 22:45:18

Diagnostics Check		OK	22:38
Air Blank		0.000	22:38
Control Test		0.078	22:39
Air Blank		0.000	22:39
Subject Sample	# 1	0.244	22:40
Breath Volume		1.351	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.252	22:43
Breath Volume		1.898	22:43
Air Blank		0.000	22:44
Control Test		0.077	22:44
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 10/31/2025 14:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EDWARD,J
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: PAYNE,KEVIN
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 09/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 10/26/2025 04:00:34

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:58	
Control Test		0.079	03:58	
Air Blank		0.000	03:59	
Subject Sample	# 1		03:59	Subject Test Refused
Breath Volume		0.000	03:59	
Air Blank		0.000	03:59	
Control Test		0.078	03:59	
Air Blank		0.000	04:00	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 10/31/2025 14:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARDY,BRIAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: ROBLES ANAYA

DATE/TIME OF TEST: 10/17/2025 03:45:53

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.081	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.196	03:41
Breath Volume		1.339	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:43
Subject Sample	# 2	0.202	03:44
Breath Volume		2.179	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 11/11/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ-CRESPO,GIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: MOCHRIE

DATE/TIME OF TEST: 10/17/2025 20:23:49

Diagnostics Check		OK	20:17
Air Blank		0.000	20:17
Control Test		0.079	20:17
Air Blank		0.000	20:18
Subject Sample	# 1	0.231	20:18
Breath Volume		2.421	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.233	20:22
Breath Volume		1.898	20:22
Air Blank		0.000	20:22
Control Test		0.081	20:23
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 11/11/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA,AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: WILLIAMFOY

DATE/TIME OF TEST: 10/18/2025 01:33:06

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.081	01:27
Air Blank		0.000	01:27
Subject Sample	# 1	0.143	01:28
Breath Volume		1.828	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.142	01:31
Breath Volume		2.253	01:31
Air Blank		0.000	01:32
Control Test		0.081	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 11/11/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARRASQUILLO RAMOS,KARLOS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female

OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: TOLBERT

DATE/TIME OF TEST: 10/18/2025 05:52:24

Diagnostics Check		OK	05:46
Air Blank		0.000	05:46
Control Test		0.082	05:46
Air Blank		0.000	05:47
Subject Sample	# 1	0.152	05:47
Breath Volume		1.257	05:47
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.153	05:50
Breath Volume		1.156	05:50
Air Blank		0.000	05:51
Control Test		0.081	05:51
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

Data Download Date/Time: 11/11/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BYNOE,OLAJUWON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/19/2025 21:43:27

Diagnostics Check		OK	21:35
Air Blank		0.000	21:35
Control Test		0.081	21:36
Air Blank		0.000	21:36
Subject Sample	# 1	0.243	21:38
Breath Volume		2.332	21:38
Air Blank		0.000	21:38
Air Blank		0.000	21:40
Subject Sample	# 2	0.224	21:41
Breath Volume		1.199	21:41
Air Blank		0.000	21:42
Control Test		0.078	21:42
Air Blank		0.000	21:43
Diagnostics Check		OK	21:43

Data Download Date/Time: 11/11/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SILVA,LUIS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MOISE

DATE/TIME OF TEST: 10/20/2025 01:40:38

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:27	
Control Test		0.081	01:28	
Air Blank		0.000	01:28	
Subject Sample	# 1	0.171	01:31	Volume Not Met
Breath Volume		0.218	01:31	
Air Blank		0.000	01:32	
Air Blank		0.000	01:34	
Subject Sample	# 2	0.165	01:35	
Breath Volume		1.187	01:35	
Air Blank		0.000	01:36	
Air Blank		0.000	01:38	
Subject Sample	# 3	0.181	01:39	
Breath Volume		1.156	01:39	
Air Blank		0.000	01:39	
Control Test		0.079	01:40	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 11/11/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: OCASIO

DATE/TIME OF TEST: 10/21/2025 02:22:10

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.081	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.139	02:17
Breath Volume		3.132	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.139	02:20
Breath Volume		3.339	02:20
Air Blank		0.000	02:21
Control Test		0.081	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 11/11/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSE,JESUS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: AGUILAR LOPEZ

DATE/TIME OF TEST: 10/24/2025 02:28:42

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.079	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.214	02:24
Breath Volume		1.886	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.213	02:27
Breath Volume		2.250	02:27
Air Blank		0.000	02:27
Control Test		0.078	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 11/11/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LYLE,BRYAN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MUZ COLORADO

DATE/TIME OF TEST: 10/24/2025 03:03:19

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.130	02:58
Breath Volume		2.906	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.127	03:01
Breath Volume		1.933	03:01
Air Blank		0.000	03:02
Control Test		0.078	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 11/11/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CONTE,COREY
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 10/25/2025 01:02:04

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.079	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.018	00:57
Breath Volume		2.210	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.016	01:00
Breath Volume		2.128	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:02

Data Download Date/Time: 11/11/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ-VARGAS,HUMBERTO
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: DELACRUZ

DATE/TIME OF TEST: 10/25/2025 02:02:53

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.079	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.181	01:58
Breath Volume		1.609	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:00
Subject Sample	# 2	0.168	02:01
Breath Volume		1.335	02:01
Air Blank		0.000	02:01
Control Test		0.077	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 11/11/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RUIZ,ANDRES
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: RESTINO

DATE/TIME OF TEST: 10/25/2025 02:34:30

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.078	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.151	02:30
Breath Volume		2.082	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.157	02:32
Breath Volume		2.695	02:32
Air Blank		0.000	02:33
Control Test		0.076	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 11/11/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TRENDAFILOV,NICHOLAS
ARRESTING OFFICER AGENCY: OTHER PRIVATE
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CUESTAAROCA

DATE/TIME OF TEST: 10/25/2025 03:49:38

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.078	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.214	03:45
Breath Volume		3.480	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.204	03:48
Breath Volume		3.496	03:48
Air Blank		0.000	03:48
Control Test		0.078	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 11/11/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALICEA GALLOZA,MARK
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: LEGER

DATE/TIME OF TEST: 10/25/2025 05:15:56

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.079	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.256	05:10
Breath Volume		2.390	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.242	05:14
Breath Volume		1.265	05:14
Air Blank		0.000	05:15
Control Test		0.078	05:15
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 11/11/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORALES,YANGEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ARIUS

DATE/TIME OF TEST: 10/25/2025 23:03:12

Diagnostics Check		OK	22:56
Air Blank		0.000	22:57
Control Test		0.079	22:57
Air Blank		0.000	22:57
Subject Sample	# 1	0.135	22:58
Breath Volume		1.773	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.146	23:01
Breath Volume		2.054	23:01
Air Blank		0.000	23:02
Control Test		0.077	23:02
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 11/11/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ST KITTS,SHAMORI
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: JEREZMASAQUIZA

DATE/TIME OF TEST: 10/26/2025 01:16:21

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.078	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.211	01:08
Breath Volume		1.187	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.232	01:11
Breath Volume		2.347	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 3	0.208	01:14
Breath Volume		1.507	01:14
Air Blank		0.000	01:15
Control Test		0.078	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 11/11/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,CRAIG
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: CASTILLO JR

DATE/TIME OF TEST: 10/26/2025 03:27:03

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:16	
Control Test		0.078	03:16	
Air Blank		0.000	03:16	
Subject Sample	# 1		03:20	No Sample Provided
Breath Volume		0.000	03:20	
Air Blank		0.000	03:20	
Air Blank		0.000	03:22	
Subject Sample	# 2		03:25	No Sample Provided
Breath Volume		0.000	03:25	
Air Blank		0.000	03:26	
Control Test		0.079	03:26	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:27	

Data Download Date/Time: 11/11/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANIELS,DAMON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/26/2025 04:37:50

Diagnostics Check		OK	04:31
Air Blank		0.000	04:32
Control Test		0.079	04:32
Air Blank		0.000	04:32
Subject Sample	# 1	0.156	04:33
Breath Volume		2.992	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:35
Subject Sample	# 2	0.158	04:36
Breath Volume		3.085	04:36
Air Blank		0.000	04:36
Control Test		0.079	04:37
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 11/11/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: ZAYASNIEVES

DATE/TIME OF TEST: 10/27/2025 01:01:12

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.113	00:56
Breath Volume		5.210	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.109	00:59
Breath Volume		4.765	00:59
Air Blank		0.000	01:00
Control Test		0.080	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 11/11/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TRENDAFILOV,NICHOLAS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: RAMIREZRAMIREZ

DATE/TIME OF TEST: 10/27/2025 02:39:30

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.079	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.244	02:34
Breath Volume		1.773	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.263	02:37
Breath Volume		1.339	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 11/11/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JULES II,DELINCE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MANZO

DATE/TIME OF TEST: 10/28/2025 23:46:21

Diagnostics Check		OK	23:33	
Air Blank		0.000	23:33	
Control Test		0.079	23:34	
Air Blank		0.000	23:34	
Subject Sample	# 1	0.295	23:37	Volume Not Met
Breath Volume		0.367	23:37	
Air Blank		0.000	23:38	
Air Blank		0.000	23:40	
Subject Sample	# 2	0.296	23:40	
Breath Volume		1.636	23:40	
Air Blank		0.000	23:41	
Air Blank		0.000	23:43	
Subject Sample	# 3	0.305	23:44	
Breath Volume		1.675	23:44	
Air Blank		0.000	23:45	
Control Test		0.078	23:45	
Air Blank		0.000	23:46	
Diagnostics Check		OK	23:46	

Data Download Date/Time: 11/11/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 10/30/2025 22:13:06

Diagnostics Check		OK	22:07
Air Blank		0.000	22:07
Control Test		0.080	22:07
Air Blank		0.000	22:08
Subject Sample	# 1	0.000	22:08
Breath Volume		2.503	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:11
Subject Sample	# 2	0.000	22:11
Breath Volume		2.269	22:11
Air Blank		0.000	22:12
Control Test		0.079	22:12
Air Blank		0.000	22:13
Diagnostics Check		OK	22:13

Data Download Date/Time: 11/11/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DURAN,ALFREDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ESPINOZA HERNANDEZ

DATE/TIME OF TEST: 10/03/2025 01:37:55

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.084	01:32
Air Blank		0.000	01:33
Subject Sample	# 1	0.152	01:33
Breath Volume		3.542	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.152	01:36
Breath Volume		3.496	01:36
Air Blank		0.000	01:37
Control Test		0.083	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 10/16/2025 10:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BENAVIDES NARANJO,JORGE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/05/2025 03:50:55

Diagnostics Check		OK	03:39	
Air Blank		0.000	03:40	
Control Test		0.083	03:40	
Air Blank		0.000	03:40	
Subject Sample	# 1	0.164	03:43	Slope Not Level
Breath Volume		2.355	03:43	
Air Blank		0.000	03:44	
Air Blank		0.000	03:46	
Subject Sample	# 2		03:49	No Sample Provided
Breath Volume		0.000	03:49	
Air Blank		0.000	03:50	
Control Test		0.081	03:50	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 10/16/2025 10:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRADFORD,JACOB
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 10/06/2025 05:25:48

Diagnostics Check		OK	05:19
Air Blank		0.000	05:19
Control Test		0.083	05:20
Air Blank		0.000	05:20
Subject Sample	# 1	0.152	05:21
Breath Volume		2.898	05:21
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.161	05:24
Breath Volume		2.125	05:24
Air Blank		0.000	05:24
Control Test		0.082	05:25
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 10/16/2025 10:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WOMACK,ANITA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: VARDY

DATE/TIME OF TEST: 10/12/2025 02:02:38

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.084	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.187	01:58
Breath Volume		1.640	01:58
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.183	02:01
Breath Volume		1.359	02:01
Air Blank		0.000	02:01
Control Test		0.082	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 10/16/2025 10:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO,EDWIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: CUNDIFF

DATE/TIME OF TEST: 10/17/2025 03:26:15

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.083	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.299	03:21
Breath Volume		2.667	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.293	03:24
Breath Volume		2.812	03:24
Air Blank		0.000	03:25
Control Test		0.082	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 11/11/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JULES II,DELINCE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: CIOCCA

DATE/TIME OF TEST: 10/17/2025 19:54:17

Diagnostics Check		OK	19:42	
Air Blank		0.000	19:43	
Control Test		0.083	19:43	
Air Blank		0.000	19:44	
Subject Sample	# 1	0.211	19:47	Volume Not Met
Breath Volume		0.050	19:47	
Air Blank		0.000	19:47	
Air Blank		0.000	19:49	
Subject Sample	# 2		19:52	No Sample Provided
Breath Volume		0.000	19:52	
Air Blank		0.000	19:53	
Control Test		0.082	19:53	
Air Blank		0.000	19:54	
Diagnostics Check		OK	19:54	

Data Download Date/Time: 11/11/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SILVA,LUIS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: WILSON JR

DATE/TIME OF TEST: 10/18/2025 01:28:09

Diagnostics Check		OK	01:21
Air Blank		0.000	01:22
Control Test		0.084	01:22
Air Blank		0.000	01:23
Subject Sample	# 1	0.172	01:23
Breath Volume		3.000	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.177	01:26
Breath Volume		2.750	01:26
Air Blank		0.000	01:27
Control Test		0.082	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 11/11/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JULES,DELINCE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/18/2025 22:08:39

Diagnostics Check		OK	22:02
Air Blank		0.000	22:03
Control Test		0.082	22:03
Air Blank		0.000	22:03
Subject Sample	# 1	0.295	22:04
Breath Volume		2.875	22:04
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.297	22:07
Breath Volume		2.828	22:07
Air Blank		0.000	22:07
Control Test		0.082	22:08
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 11/11/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE,ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 10/19/2025 00:27:55

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.084	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.174	00:23
Breath Volume		1.480	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.173	00:26
Breath Volume		1.191	00:26
Air Blank		0.000	00:27
Control Test		0.083	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 11/11/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMOS,KARLOS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: REYESJIMENEZ

DATE/TIME OF TEST: 10/19/2025 05:32:40

Diagnostics Check		OK	05:25
Air Blank		0.000	05:26
Control Test		0.083	05:26
Air Blank		0.000	05:27
Subject Sample	# 1	0.154	05:27
Breath Volume		1.707	05:27
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.148	05:31
Breath Volume		1.472	05:31
Air Blank		0.000	05:31
Control Test		0.082	05:32
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 11/11/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA,AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: SUKHRAM

DATE/TIME OF TEST: 10/19/2025 21:13:56

Diagnostics Check		OK	21:02	
Air Blank		0.000	21:02	
Control Test		0.084	21:03	
Air Blank		0.000	21:03	
Subject Sample	# 1	0.210	21:06	Volume Not Met
Breath Volume		0.148	21:06	
Air Blank		0.000	21:07	
Air Blank		0.000	21:09	
Subject Sample	# 2	0.207	21:12	Volume Not Met
Breath Volume		0.199	21:12	
Air Blank		0.000	21:13	
Control Test		0.083	21:13	Interferent Detect
Air Blank		0.000	21:13	

Data Download Date/Time: 11/11/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA,AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: SUKHRAM

DATE/TIME OF TEST: 10/19/2025 21:28:39

Diagnostics Check		OK	21:17	
Air Blank		0.000	21:17	
Control Test		0.082	21:17	
Air Blank		0.000	21:18	
Subject Sample	# 1	0.214	21:21	Volume Not Met
Breath Volume		0.164	21:21	
Air Blank		0.000	21:22	
Air Blank		0.000	21:24	
Subject Sample	# 2	0.192	21:27	Volume Not Met
Breath Volume		0.238	21:27	
Air Blank		0.000	21:27	
Control Test		0.080	21:28	
Air Blank		0.000	21:28	
Diagnostics Check		OK	21:28	

Data Download Date/Time: 11/11/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BECKER,THOMAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 10/19/2025 23:55:36

Diagnostics Check		OK	23:49
Air Blank		0.000	23:49
Control Test		0.082	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.261	23:51
Breath Volume		1.714	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.259	23:54
Breath Volume		1.218	23:54
Air Blank		0.000	23:54
Control Test		0.081	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 11/11/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAAVEDRA,NIKOLAS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: ALVARADOOROZCO

DATE/TIME OF TEST: 10/20/2025 03:39:39

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.084	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.203	03:34
Breath Volume		2.500	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.207	03:38
Breath Volume		3.019	03:38
Air Blank		0.000	03:38
Control Test		0.083	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 11/11/2025 09:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HEBEN,KEVIN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: REHBACHRAMIREZ

DATE/TIME OF TEST: 10/20/2025 21:48:08

Diagnostics Check		OK	21:40
Air Blank		0.000	21:41
Control Test		0.082	21:41
Air Blank		0.000	21:41
Subject Sample	# 1	0.166	21:43
Breath Volume		3.292	21:43
Air Blank		0.000	21:44
Air Blank		0.000	21:46
Subject Sample	# 2	0.165	21:46
Breath Volume		3.242	21:46
Air Blank		0.000	21:47
Control Test		0.081	21:47
Air Blank		0.000	21:48
Diagnostics Check		OK	21:48

Data Download Date/Time: 11/11/2025 09:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CIESIELSKI,BRANDON
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MATEO MATEO

DATE/TIME OF TEST: 10/21/2025 01:48:52

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:43	
Control Test		0.084	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 1	0.369	01:44	
Breath Volume		1.921	01:44	
Air Blank		0.000	01:45	
Air Blank		0.000	01:47	
Subject Sample	# 2		01:47	Slope Not Met
Breath Volume		0.980	01:47	
Air Blank		0.000	01:48	
Control Test		0.083	01:48	Interferent Detect
Air Blank		0.000	01:48	

Data Download Date/Time: 11/11/2025 09:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CIESIELSKI,BRANDON
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MATEO MATEO

DATE/TIME OF TEST: 10/21/2025 02:26:28

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.082	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.351	02:22
Breath Volume		1.773	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.344	02:24
Breath Volume		1.269	02:24
Air Blank		0.000	02:25
Control Test		0.082	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 11/11/2025 09:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: VILLEGAS

DATE/TIME OF TEST: 10/21/2025 22:50:19

Diagnostics Check		OK	22:44
Air Blank		0.000	22:44
Control Test		0.085	22:45
Air Blank		0.000	22:45
Subject Sample	# 1	0.000	22:46
Breath Volume		1.304	22:46
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.000	22:48
Breath Volume		1.414	22:48
Air Blank		0.000	22:49
Control Test		0.084	22:49
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 11/11/2025 09:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSE,JESUS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: APARICIOREYES

DATE/TIME OF TEST: 10/22/2025 21:14:25

Diagnostics Check		OK	21:08
Air Blank		0.000	21:08
Control Test		0.083	21:08
Air Blank		0.000	21:09
Subject Sample	# 1	0.189	21:10
Breath Volume		2.828	21:10
Air Blank		0.000	21:10
Air Blank		0.000	21:12
Subject Sample	# 2	0.184	21:12
Breath Volume		1.203	21:12
Air Blank		0.000	21:13
Control Test		0.079	21:13
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 11/11/2025 09:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,KANARVIS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: NATHOO

DATE/TIME OF TEST: 10/23/2025 01:06:31

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:54	
Control Test		0.084	00:54	
Air Blank		0.000	00:55	
Subject Sample	# 1		00:58	No Sample Provided
Breath Volume		0.000	00:58	
Air Blank		0.000	00:58	
Air Blank		0.000	01:00	
Subject Sample	# 2	0.202	01:01	
Breath Volume		1.160	01:01	
Air Blank		0.000	01:02	
Air Blank		0.000	01:04	
Subject Sample	# 3	0.196	01:05	
Breath Volume		1.304	01:05	
Air Blank		0.000	01:05	
Control Test		0.081	01:05	
Air Blank		0.000	01:06	
Diagnostics Check		OK	01:06	

Data Download Date/Time: 11/11/2025 09:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSE,JESUS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: AGUILAR LOPEZ

DATE/TIME OF TEST: 10/24/2025 01:51:18

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:45	
Control Test		0.085	01:45	
Air Blank		0.000	01:46	
Subject Sample	# 1	0.225	01:46	
Breath Volume		2.085	01:46	
Air Blank		0.000	01:47	
Air Blank		0.000	01:49	
Subject Sample	# 2	0.223	01:49	
Breath Volume		1.933	01:49	
Air Blank		0.000	01:50	
Control Test		0.083	01:50	Interferent Detect
Air Blank		0.000	01:51	

Data Download Date/Time: 11/11/2025 09:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSE,JESUS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: AGUILAR LOPEZ

DATE/TIME OF TEST: 10/24/2025 01:59:02

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:53	
Control Test		0.082	01:53	
Air Blank		0.000	01:54	
Subject Sample	# 1	0.227	01:54	
Breath Volume		2.097	01:54	
Air Blank		0.000	01:55	
Air Blank		0.000	01:57	
Subject Sample	# 2	0.219	01:57	
Breath Volume		1.804	01:57	
Air Blank		0.000	01:58	
Control Test		0.084	01:58	Interferent Detect
Air Blank		0.000	01:59	

Data Download Date/Time: 11/11/2025 09:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: RUCKER

DATE/TIME OF TEST: 10/25/2025 03:11:39

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.083	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.127	03:07
Breath Volume		1.398	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.121	03:10
Breath Volume		1.410	03:10
Air Blank		0.000	03:10
Control Test		0.083	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 11/11/2025 09:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GALLOZA,ALICEA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ARANGO GALLEGO

DATE/TIME OF TEST: 10/17/2025 02:24:45

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.080	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.149	02:20
Breath Volume		1.308	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.153	02:23
Breath Volume		1.402	02:23
Air Blank		0.000	02:23
Control Test		0.080	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 10/25/2025 16:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO,DAVID
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: GABIREL ZACARIAS

DATE/TIME OF TEST: 10/19/2025 21:21:20

Diagnostics Check		OK	21:15
Air Blank		0.000	21:15
Control Test		0.080	21:15
Air Blank		0.000	21:16
Subject Sample	# 1	0.256	21:16
Breath Volume		3.566	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.258	21:19
Breath Volume		3.011	21:19
Air Blank		0.000	21:20
Control Test		0.079	21:20
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 10/25/2025 16:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BIRNBAUM,DAVID
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: SINGAM

DATE/TIME OF TEST: 10/20/2025 00:47:15

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.173	00:41
Breath Volume		1.402	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.182	00:45
Breath Volume		2.292	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 10/25/2025 16:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: APONTE,MIGUEL
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: SALCEDO

DATE/TIME OF TEST: 10/23/2025 03:23:03

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.080	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.204	03:18
Breath Volume		2.804	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:20
Subject Sample	# 2	0.205	03:21
Breath Volume		3.667	03:21
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:23

Data Download Date/Time: 10/25/2025 16:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORENO,PAYTON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 10/24/2025 04:48:47

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.080	04:40
Air Blank		0.000	04:41
Subject Sample	# 1	0.175	04:41
Breath Volume		1.910	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.178	04:47
Breath Volume		1.468	04:47
Air Blank		0.000	04:47
Control Test		0.079	04:48
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 10/25/2025 16:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCKUHEN,ANDREW
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: DEVERNEY

DATE/TIME OF TEST: 10/25/2025 01:55:12

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.182	01:50
Breath Volume		1.578	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:52
Subject Sample	# 2	0.181	01:53
Breath Volume		1.898	01:53
Air Blank		0.000	01:54
Control Test		0.078	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 10/25/2025 16:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALMAZAN,KATHERINE
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: RAMIREZ PEREZ

DATE/TIME OF TEST: 10/25/2025 02:20:33

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.080	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.183	02:15
Breath Volume		1.449	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:17
Subject Sample	# 2	0.171	02:18
Breath Volume		1.175	02:18
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 10/25/2025 16:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: DIXON JR

DATE/TIME OF TEST: 10/26/2025 00:51:51

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:38	
Control Test		0.080	00:38	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:42	No Sample Provided
Breath Volume		0.000	00:42	
Air Blank		0.000	00:42	
Air Blank		0.000	00:44	
Subject Sample	# 2	0.056	00:46	
Breath Volume		1.566	00:46	
Air Blank		0.000	00:47	
Air Blank		0.000	00:49	
Subject Sample	# 3	0.055	00:50	
Breath Volume		1.179	00:50	
Air Blank		0.000	00:50	
Control Test		0.079	00:51	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 11/11/2025 10:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALDES,BRYAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 10/26/2025 02:41:28

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.080	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.170	02:36
Breath Volume		1.121	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.190	02:39
Breath Volume		1.234	02:39
Air Blank		0.000	02:40
Control Test		0.079	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 11/11/2025 10:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SANCHEZ,ISABELLA
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: RAMIREZ MARTINEZ

DATE/TIME OF TEST: 10/26/2025 03:38:55

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.096	03:34
Breath Volume		2.582	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:36
Subject Sample	# 2	0.095	03:37
Breath Volume		2.761	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 11/11/2025 10:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,CRAIG
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/27/2025 00:43:20

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:38	
Control Test		0.080	00:38	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:41	Subject Test Refused
Breath Volume		0.000	00:41	
Air Blank		0.000	00:42	
Control Test		0.079	00:42	
Air Blank		0.000	00:43	
Diagnostics Check		OK	00:43	

Data Download Date/Time: 11/11/2025 10:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARVAJAL,JUAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: MARIN RUIZ

DATE/TIME OF TEST: 10/27/2025 02:11:40

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:57	
Control Test		0.081	01:58	
Air Blank		0.000	01:58	
Subject Sample	# 1	0.146	02:01	
Breath Volume		1.324	02:01	
Air Blank		0.000	02:01	
Air Blank		0.000	02:03	
Subject Sample	# 2	0.126	02:06	Volume Not Met
Breath Volume		0.898	02:06	
Air Blank		0.000	02:07	
Air Blank		0.000	02:09	
Subject Sample	# 3	0.140	02:10	
Breath Volume		1.757	02:10	
Air Blank		0.000	02:10	
Control Test		0.077	02:11	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 11/11/2025 10:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ CERVERA,ADRIAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male

OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: RAMIREZ GOMEZ

DATE/TIME OF TEST: 10/29/2025 03:43:44

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.081	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.222	03:39
Breath Volume		1.843	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.215	03:42
Breath Volume		1.394	03:42
Air Blank		0.000	03:42
Control Test		0.080	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 11/11/2025 10:04

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENN,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: DEPRIEST,T
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: HEROLD

DATE/TIME OF TEST: 10/04/2025 05:35:48

Diagnostics Check		OK	05:32	
Air Blank		0.000	05:33	
Control Test		0.080	05:33	
Air Blank		0.000	05:34	
Subject Sample	# 1		05:34	Subject Test Refused
Breath Volume		0.000	05:34	
Air Blank		0.000	05:34	
Control Test		0.080	05:35	
Air Blank		0.000	05:35	
Diagnostics Check		OK	05:35	

Data Download Date/Time: 11/03/2025 11:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: MASSENGILL

DATE/TIME OF TEST: 10/06/2025 18:47:42

Diagnostics Check		OK	18:40
Air Blank		0.000	18:40
Control Test		0.081	18:41
Air Blank		0.000	18:41
Subject Sample	# 1	0.298	18:43
Breath Volume		1.808	18:43
Air Blank		0.000	18:43
Air Blank		0.000	18:45
Subject Sample	# 2	0.286	18:46
Breath Volume		1.671	18:46
Air Blank		0.000	18:46
Control Test		0.081	18:47
Air Blank		0.000	18:47
Diagnostics Check		OK	18:47

Data Download Date/Time: 11/03/2025 11:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BRAVE,S
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: REMME

DATE/TIME OF TEST: 10/09/2025 02:50:49

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.079	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.205	02:46
Breath Volume		3.191	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.211	02:49
Breath Volume		3.042	02:49
Air Blank		0.000	02:49
Control Test		0.081	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 11/03/2025 11:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,RODERICK,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/10/2025 18:27:47

Diagnostics Check		OK	18:19
Air Blank		0.000	18:19
Control Test		0.080	18:20
Air Blank		0.000	18:20
Subject Sample	# 1	0.101	18:22
Breath Volume		1.835	18:22
Air Blank		0.000	18:22
Air Blank		0.000	18:24
Subject Sample	# 2	0.096	18:26
Breath Volume		1.812	18:26
Air Blank		0.000	18:26
Control Test		0.081	18:27
Air Blank		0.000	18:27
Diagnostics Check		OK	18:27

Data Download Date/Time: 11/03/2025 11:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,RODERICK,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WHITE,S
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 10/11/2025 02:14:12

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:11	
Control Test		0.080	02:12	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:13	
Control Test		0.080	02:13	
Air Blank		0.000	02:14	
Diagnostics Check		OK	02:14	

Data Download Date/Time: 11/03/2025 11:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,RODERICK,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: REILLY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 10/12/2025 21:52:56

Diagnostics Check		OK	21:50	
Air Blank		0.000	21:50	
Control Test		0.079	21:50	
Air Blank		0.000	21:51	
Subject Sample	# 1		21:51	Subject Test Refused
Breath Volume		0.000	21:51	
Air Blank		0.000	21:52	
Control Test		0.079	21:52	
Air Blank		0.000	21:52	
Diagnostics Check		OK	21:52	

Data Download Date/Time: 11/03/2025 11:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MALESKE,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 10/15/2025 20:23:18

Diagnostics Check		OK	20:17
Air Blank		0.000	20:17
Control Test		0.080	20:17
Air Blank		0.000	20:18
Subject Sample	# 1	0.036	20:18
Breath Volume		1.468	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.036	20:21
Breath Volume		1.574	20:21
Air Blank		0.000	20:22
Control Test		0.079	20:22
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 11/03/2025 11:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BOGGS,Z
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BORJA

DATE/TIME OF TEST: 10/16/2025 00:20:11

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:17	
Control Test		0.080	00:17	
Air Blank		0.000	00:18	
Subject Sample	# 1		00:18	RFI Detect
Breath Volume		0.000	00:18	
Air Blank		0.000	00:19	
Control Test		0.078	00:19	
Air Blank		0.000	00:20	
Diagnostics Check		OK	00:20	

Data Download Date/Time: 11/03/2025 11:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BOGGS,Z
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BORJA

DATE/TIME OF TEST: 10/16/2025 00:29:16

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.080	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.197	00:24
Breath Volume		1.699	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.190	00:27
Breath Volume		1.441	00:27
Air Blank		0.000	00:28
Control Test		0.080	00:28
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 11/03/2025 11:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSELL,TRISTAN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MACK,RACHAEL
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 10/21/2025 23:54:57

Diagnostics Check		OK	23:52	
Air Blank		0.000	23:52	
Control Test		0.082	23:52	
Air Blank		0.000	23:53	
Subject Sample	# 1		23:53	Subject Test Refused
Breath Volume		0.000	23:53	
Air Blank		0.000	23:54	
Control Test		0.080	23:54	
Air Blank		0.000	23:54	
Diagnostics Check		OK	23:54	

Data Download Date/Time: 11/03/2025 11:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA',L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SURIEL,E
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: SALYER

DATE/TIME OF TEST: 10/22/2025 03:23:13

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.080	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.095	03:18
Breath Volume		2.640	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.096	03:21
Breath Volume		2.484	03:21
Air Blank		0.000	03:22
Control Test		0.078	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 11/03/2025 11:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,KELVIN,
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: VEGA,KELVIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 10/14/2025 00:23:27

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.143	00:18
Breath Volume		2.808	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.143	00:21
Breath Volume		3.250	00:21
Air Blank		0.000	00:22
Control Test		0.078	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 12/03/2025 14:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,R,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PAGLIAROLI,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DAVILA HERNANDEZ

DATE/TIME OF TEST: 10/27/2025 02:26:00

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.081	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.143	02:21
Breath Volume		1.796	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:23
Subject Sample	# 2	0.136	02:24
Breath Volume		1.523	02:24
Air Blank		0.000	02:25
Control Test		0.080	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:26

Data Download Date/Time: 12/01/2025 10:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPISE,STEPHEN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SANDERS,S
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 10/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 10/27/2025 22:39:31

Diagnostics Check		OK	22:33
Air Blank		0.000	22:33
Control Test		0.081	22:33
Air Blank		0.000	22:34
Subject Sample	# 1	0.162	22:34
Breath Volume		2.496	22:34
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.159	22:37
Breath Volume		2.625	22:37
Air Blank		0.000	22:38
Control Test		0.082	22:38
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 12/01/2025 10:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDERE,ANTONIO,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WENNLUND,F
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 10/30/2025 03:15:09

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.114	03:10
Breath Volume		1.695	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.108	03:13
Breath Volume		1.453	03:13
Air Blank		0.000	03:14
Control Test		0.081	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 12/01/2025 10:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARON,ZACHARY,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ESPIN,T
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 10/31/2025 01:46:46

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.082	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.000	01:42
Breath Volume		1.359	01:42
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.000	01:45
Breath Volume		2.429	01:45
Air Blank		0.000	01:45
Control Test		0.081	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 12/01/2025 10:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,R,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HILL,G
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SANOJA SALAZAR

DATE/TIME OF TEST: 10/31/2025 03:40:54

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.081	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.143	03:36
Breath Volume		2.937	03:36
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.142	03:39
Breath Volume		2.921	03:39
Air Blank		0.000	03:39
Control Test		0.082	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 12/01/2025 10:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARON,ZACHARY,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WENNLUND,F
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 10/31/2025 04:04:02

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:00	
Control Test		0.081	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 1		04:02	Subject Test Refused
Breath Volume		0.000	04:02	
Air Blank		0.000	04:03	
Control Test		0.081	04:03	
Air Blank		0.000	04:03	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 12/01/2025 10:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/29/2025 23:01:42

Diagnostics Check		OK	22:54
Air Blank		0.000	22:55
Control Test		0.082	22:55
Air Blank		0.000	22:56
Subject Sample	# 1	0.041	22:57
Breath Volume		3.320	22:57
Air Blank		0.000	22:57
Air Blank		0.000	22:59
Subject Sample	# 2	0.042	23:00
Breath Volume		3.191	23:00
Air Blank		0.000	23:00
Control Test		0.083	23:01
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 11/27/2025 20:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SASSO

DATE/TIME OF TEST: 10/30/2025 04:31:55

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.083	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.000	04:27
Breath Volume		2.578	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.000	04:30
Breath Volume		1.449	04:30
Air Blank		0.000	04:30
Control Test		0.082	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 11/27/2025 20:43

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: RODRIGUEZ,O
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: PAGAN NEGRON

DATE/TIME OF TEST: 10/04/2025 04:18:47

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.266	04:09
Breath Volume		2.425	04:09
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.242	04:12
Breath Volume		1.160	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 3	0.234	04:17
Breath Volume		1.734	04:17
Air Blank		0.000	04:17
Control Test		0.079	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 10/29/2025 13:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 10/04/2025 20:13:34

Diagnostics Check		OK	20:07
Air Blank		0.000	20:07
Control Test		0.081	20:08
Air Blank		0.000	20:08
Subject Sample	# 1	0.136	20:08
Breath Volume		3.468	20:08
Air Blank		0.000	20:09
Air Blank		0.000	20:11
Subject Sample	# 2	0.137	20:11
Breath Volume		3.617	20:11
Air Blank		0.000	20:12
Control Test		0.079	20:12
Air Blank		0.000	20:13
Diagnostics Check		OK	20:13

Data Download Date/Time: 10/29/2025 13:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 10/05/2025 03:34:14

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.080	03:21	
Air Blank		0.000	03:21	
Subject Sample	# 1	0.177	03:25	Volume Not Met
Breath Volume		0.644	03:25	
Air Blank		0.000	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 2	0.208	03:28	
Breath Volume		1.511	03:28	
Air Blank		0.000	03:29	
Air Blank		0.000	03:30	
Subject Sample	# 3	0.195	03:32	
Breath Volume		1.218	03:32	
Air Blank		0.000	03:33	
Control Test		0.075	03:33	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 10/29/2025 13:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HOWARD,WILLIAM
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: OTERO ROSADO

DATE/TIME OF TEST: 10/06/2025 01:42:17

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.150	01:37
Breath Volume		1.632	01:37
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.147	01:40
Breath Volume		1.902	01:40
Air Blank		0.000	01:41
Control Test		0.081	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 10/29/2025 13:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: MIRANDA-OTERO

DATE/TIME OF TEST: 10/06/2025 04:24:59

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:15	
Control Test		0.080	04:16	
Air Blank		0.000	04:16	
Subject Sample	# 1	0.182	04:19	Volume Not Met
Breath Volume		0.000	04:19	
Air Blank		0.000	04:20	
Air Blank		0.000	04:22	
Subject Sample	# 2		04:23	Subject Test Refused
Breath Volume		0.035	04:23	
Air Blank		0.000	04:24	
Control Test		0.078	04:24	
Air Blank		0.000	04:24	
Diagnostics Check		OK	04:24	

Data Download Date/Time: 10/29/2025 13:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 10/09/2025 01:14:26

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.080	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1	0.066	01:03	
Breath Volume		1.851	01:03	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 2	0.051	01:09	Volume Not Met
Breath Volume		0.945	01:09	
Air Blank		0.000	01:10	
Air Blank		0.000	01:12	
Subject Sample	# 3	0.065	01:12	
Breath Volume		1.164	01:12	
Air Blank		0.000	01:13	
Control Test		0.076	01:13	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 10/29/2025 13:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HUSSAIN

DATE/TIME OF TEST: 10/10/2025 01:12:44

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.081	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.235	01:07
Breath Volume		1.226	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.226	01:10
Breath Volume		1.453	01:10
Air Blank		0.000	01:11
Control Test		0.078	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 10/29/2025 13:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,DERICK,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOPEZ,DERICK
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 10/12/2025 05:52:35

Diagnostics Check		OK	05:45
Air Blank		0.000	05:46
Control Test		0.080	05:46
Air Blank		0.000	05:46
Subject Sample	# 1	0.000	05:47
Breath Volume		1.328	05:47
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.000	05:51
Breath Volume		1.273	05:51
Air Blank		0.000	05:51
Control Test		0.080	05:51
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

Data Download Date/Time: 10/29/2025 13:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,C,B
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: ROSA,D
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: QUINTANA MOLINA

DATE/TIME OF TEST: 10/15/2025 22:51:54

Diagnostics Check		OK	22:45	
Air Blank		0.000	22:46	
Control Test		0.080	22:46	
Air Blank		0.000	22:47	
Subject Sample	# 1	0.179	22:47	
Breath Volume		3.601	22:47	
Air Blank		0.000	22:48	
Air Blank		0.000	22:50	
Subject Sample	# 2	0.173	22:50	
Breath Volume		3.722	22:50	
Air Blank		0.000	22:51	
Control Test		0.078	22:51	RFI Detect
Air Blank		0.000	22:51	

Data Download Date/Time: 10/29/2025 13:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,DERICK,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOPEZ,DERICK
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: DALTON

DATE/TIME OF TEST: 10/16/2025 20:05:13

Diagnostics Check		OK	19:58
Air Blank		0.000	19:59
Control Test		0.080	19:59
Air Blank		0.000	19:59
Subject Sample	# 1	0.209	20:00
Breath Volume		1.566	20:00
Air Blank		0.000	20:01
Air Blank		0.000	20:03
Subject Sample	# 2	0.201	20:03
Breath Volume		1.546	20:03
Air Blank		0.000	20:04
Control Test		0.080	20:04
Air Blank		0.000	20:05
Diagnostics Check		OK	20:05

Data Download Date/Time: 10/29/2025 13:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,DERICK,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PARRA,URIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: OSPINO-HINCAPIE

DATE/TIME OF TEST: 10/16/2025 23:37:43

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.080	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.240	23:32
Breath Volume		1.121	23:32
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.244	23:36
Breath Volume		1.203	23:36
Air Blank		0.000	23:36
Control Test		0.079	23:37
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 10/29/2025 13:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEREIRA,T
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: RAMOS-TORRES

DATE/TIME OF TEST: 10/17/2025 19:48:28

Diagnostics Check		OK	19:42
Air Blank		0.000	19:42
Control Test		0.080	19:42
Air Blank		0.000	19:43
Subject Sample	# 1	0.087	19:43
Breath Volume		2.218	19:43
Air Blank		0.000	19:44
Air Blank		0.000	19:46
Subject Sample	# 2	0.082	19:46
Breath Volume		1.246	19:46
Air Blank		0.000	19:47
Control Test		0.080	19:47
Air Blank		0.000	19:48
Diagnostics Check		OK	19:48

Data Download Date/Time: 10/29/2025 13:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HOWARD,WILLIAM
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: SUAREZ-MIRANDA

DATE/TIME OF TEST: 10/18/2025 02:15:58

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:13	
Control Test		0.079	02:13	
Air Blank		0.000	02:14	
Subject Sample	# 1		02:14	Subject Test Refused
Breath Volume		0.000	02:14	
Air Blank		0.000	02:15	
Control Test		0.079	02:15	
Air Blank		0.000	02:15	
Diagnostics Check		OK	02:15	

Data Download Date/Time: 10/29/2025 13:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,CORINA,M
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: GARCIA,CORINA
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: ALAMILLA FRANCISCO

DATE/TIME OF TEST: 10/18/2025 02:26:52

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.183	02:21
Breath Volume		1.976	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.180	02:25
Breath Volume		2.179	02:25
Air Blank		0.000	02:25
Control Test		0.079	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 10/29/2025 13:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HOWARD,WILLIAM
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: MERRCADO PAGAN

DATE/TIME OF TEST: 10/19/2025 00:13:16

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.081	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.145	00:08
Breath Volume		1.253	00:08
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.150	00:11
Breath Volume		1.312	00:11
Air Blank		0.000	00:12
Control Test		0.080	00:12
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 10/29/2025 13:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: REICH,P
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: JOYCE

DATE/TIME OF TEST: 10/19/2025 22:27:41

Diagnostics Check		OK	22:20
Air Blank		0.000	22:20
Control Test		0.080	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.077	22:22
Breath Volume		1.136	22:22
Air Blank		0.000	22:22
Air Blank		0.000	22:24
Subject Sample	# 2	0.085	22:26
Breath Volume		2.363	22:26
Air Blank		0.000	22:26
Control Test		0.080	22:27
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 10/29/2025 13:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HOWARD,WILLIAM
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: SEGARRA RUIZ

DATE/TIME OF TEST: 10/20/2025 04:51:42

Diagnostics Check		OK	04:45
Air Blank		0.000	04:45
Control Test		0.080	04:45
Air Blank		0.000	04:46
Subject Sample	# 1	0.123	04:47
Breath Volume		1.382	04:47
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.122	04:50
Breath Volume		1.472	04:50
Air Blank		0.000	04:50
Control Test		0.080	04:51
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 10/29/2025 13:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEREIRA,T
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: HERNANDEZ CUEBAS

DATE/TIME OF TEST: 10/27/2025 18:35:58

Diagnostics Check		OK	18:27
Air Blank		0.000	18:27
Control Test		0.081	18:28
Air Blank		0.000	18:28
Subject Sample	# 1	0.000	18:30
Breath Volume		1.726	18:30
Air Blank		0.000	18:30
Air Blank		0.000	18:32
Subject Sample	# 2	0.000	18:34
Breath Volume		1.339	18:34
Air Blank		0.000	18:35
Control Test		0.080	18:35
Air Blank		0.000	18:35
Diagnostics Check		OK	18:35

Data Download Date/Time: 10/29/2025 13:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH,GARRETT,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WORTH,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: HEMPHILL

DATE/TIME OF TEST: 10/01/2025 21:06:18

Diagnostics Check		OK	21:00
Air Blank		0.000	21:00
Control Test		0.078	21:00
Air Blank		0.000	21:01
Subject Sample	# 1	0.172	21:01
Breath Volume		2.687	21:01
Air Blank		0.000	21:02
Air Blank		0.000	21:04
Subject Sample	# 2	0.177	21:04
Breath Volume		3.144	21:04
Air Blank		0.000	21:05
Control Test		0.077	21:05
Air Blank		0.000	21:06
Diagnostics Check		OK	21:06

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MEYER,SAMUEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 10/04/2025 02:17:53

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.078	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.205	02:13
Breath Volume		2.882	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.197	02:16
Breath Volume		3.070	02:16
Air Blank		0.000	02:16
Control Test		0.076	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MALCOLM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: WILSON,MALCOLM
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/05/2025 10:06:09

Diagnostics Check		OK	10:00
Air Blank		0.000	10:00
Control Test		0.077	10:01
Air Blank		0.000	10:01
Subject Sample	# 1	0.000	10:01
Breath Volume		2.070	10:01
Air Blank		0.000	10:02
Air Blank		0.000	10:04
Subject Sample	# 2	0.000	10:04
Breath Volume		2.316	10:04
Air Blank		0.000	10:05
Control Test		0.077	10:05
Air Blank		0.000	10:06
Diagnostics Check		OK	10:06

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCWILLIAMS,JACOB,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MCWILLIAMS,JACOB
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 10/10/2025 04:28:09

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.077	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.063	04:23
Breath Volume		1.789	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.055	04:26
Breath Volume		1.175	04:26
Air Blank		0.000	04:27
Control Test		0.077	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONCIVAIS,DELFINO,
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: MORRIS,E.
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 10/03/2025 03:19:01

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:16	
Control Test		0.077	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:17	Subject Test Refused
Breath Volume		0.000	03:17	
Air Blank		0.000	03:18	
Control Test		0.078	03:18	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONCIVAIS,DELFINO,
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: MONCIVAIS,DELFINO
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 10/11/2025 00:55:11

Diagnostics Check		OK	00:51	
Air Blank		0.000	00:52	
Control Test		0.077	00:52	
Air Blank		0.000	00:53	
Subject Sample	# 1		00:53	Subject Test Refused
Breath Volume		0.000	00:53	
Air Blank		0.000	00:54	
Control Test		0.077	00:54	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,RONALD,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: EVANS,R
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: LATHAM

DATE/TIME OF TEST: 10/15/2025 02:47:23

Diagnostics Check		OK	02:39
Air Blank		0.000	02:40
Control Test		0.077	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.129	02:42
Breath Volume		2.582	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.127	02:45
Breath Volume		2.296	02:45
Air Blank		0.000	02:46
Control Test		0.077	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH,GARRETT,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WORTH,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: WELLS 2

DATE/TIME OF TEST: 10/15/2025 19:33:56

Diagnostics Check		OK	19:31	
Air Blank		0.000	19:31	
Control Test		0.077	19:31	
Air Blank		0.000	19:32	
Subject Sample	# 1		19:32	Subject Test Refused
Breath Volume		0.000	19:32	
Air Blank		0.000	19:33	
Control Test		0.078	19:33	
Air Blank		0.000	19:33	
Diagnostics Check		OK	19:33	

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,RONALD,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: EVANS,R
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BLATCH

DATE/TIME OF TEST: 10/16/2025 04:23:06

Diagnostics Check		OK	04:16	
Air Blank		0.000	04:17	
Control Test		0.078	04:17	
Air Blank		0.000	04:17	
Subject Sample	# 1	0.147	04:18	
Breath Volume		3.050	04:18	
Air Blank		0.000	04:19	
Air Blank		0.000	04:21	
Subject Sample	# 2	0.144	04:22	
Breath Volume		2.945	04:22	
Air Blank		0.000	04:22	
			04:22	RFI Detect
Air Blank			04:23	RFI Detect

Data Download Date/Time: 10/31/2025 10:34

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,RONALD,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: EVANS,R
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: BLATCH

DATE/TIME OF TEST: 10/16/2025 04:42:32

Diagnostics Check		OK	04:35
Air Blank		0.000	04:36
Control Test		0.077	04:36
Air Blank		0.000	04:36
Subject Sample	# 1	0.143	04:37
Breath Volume		3.046	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.142	04:40
Breath Volume		2.988	04:40
Air Blank		0.000	04:41
Control Test		0.077	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COPPOCK,JOEL,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: COPPOCK,JOEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:16

SUBJECT'S LAST NAME: LATSON

DATE/TIME OF TEST: 10/17/2025 14:47:45

Diagnostics Check		OK	14:40
Air Blank		0.000	14:41
Control Test		0.077	14:41
Air Blank		0.000	14:42
Subject Sample	# 1	0.102	14:42
Breath Volume		1.476	14:42
Air Blank		0.000	14:43
Air Blank		0.000	14:45
Subject Sample	# 2	0.088	14:46
Breath Volume		1.687	14:46
Air Blank		0.000	14:46
Control Test		0.078	14:47
Air Blank		0.000	14:47
Diagnostics Check		OK	14:47

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COPPOCK,JOEL,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: CRUSE,DALE
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: ELLINGHAM

DATE/TIME OF TEST: 10/17/2025 15:23:19

Diagnostics Check		OK	15:16
Air Blank		0.000	15:17
Control Test		0.077	15:17
Air Blank		0.000	15:17
Subject Sample	# 1	0.326	15:18
Breath Volume		2.050	15:18
Air Blank		0.000	15:19
Air Blank		0.000	15:21
Subject Sample	# 2	0.327	15:21
Breath Volume		2.808	15:21
Air Blank		0.000	15:22
Control Test		0.077	15:22
Air Blank		0.000	15:23
Diagnostics Check		OK	15:23

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCWILLIAMS,JACOB,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MCWILLIAMS,JACOB
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 10/18/2025 22:17:33

Diagnostics Check		OK	22:06	
Air Blank		0.000	22:06	
Control Test		0.078	22:06	
Air Blank		0.000	22:07	
Subject Sample	# 1		22:10	No Sample Provided
Breath Volume		0.000	22:10	
Air Blank		0.000	22:11	
Air Blank		0.000	22:13	
Subject Sample	# 2		22:16	No Sample Provided
Breath Volume		0.000	22:16	
Air Blank		0.000	22:16	
Control Test		0.078	22:16	
Air Blank		0.000	22:17	
Diagnostics Check		OK	22:17	

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MOORE,ISRAEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: WHITEFIELD

DATE/TIME OF TEST: 10/18/2025 22:51:25

Diagnostics Check		OK	22:44
Air Blank		0.000	22:45
Control Test		0.078	22:45
Air Blank		0.000	22:45
Subject Sample	# 1	0.000	22:47
Breath Volume		1.687	22:47
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.000	22:50
Breath Volume		1.339	22:50
Air Blank		0.000	22:50
Control Test		0.077	22:50
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO,KENNETH,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BRABHAM,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: JOHNSON JR

DATE/TIME OF TEST: 10/22/2025 01:13:04

Diagnostics Check		OK	00:58	
Air Blank		0.000	00:59	
Control Test		0.077	00:59	
Air Blank		0.000	01:00	
Subject Sample	# 1		01:03	No Sample Provided
Breath Volume		0.000	01:03	
Air Blank		0.000	01:03	
Air Blank		0.000	01:05	
Subject Sample	# 2	0.171	01:06	
Breath Volume		1.808	01:06	
Air Blank		0.000	01:07	
Air Blank		0.000	01:08	
Subject Sample	# 3	0.160	01:11	
Breath Volume		1.703	01:11	
Air Blank		0.000	01:12	
Control Test		0.077	01:12	
Air Blank		0.000	01:12	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: SOSA,KYLE
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 10/23/2025 22:13:39

Diagnostics Check		OK	22:06
Air Blank		0.000	22:07
Control Test		0.077	22:07
Air Blank		0.000	22:08
Subject Sample	# 1	0.177	22:08
Breath Volume		1.757	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:11
Subject Sample	# 2	0.181	22:12
Breath Volume		1.906	22:12
Air Blank		0.000	22:12
Control Test		0.077	22:13
Air Blank		0.000	22:13
Diagnostics Check		OK	22:13

Data Download Date/Time: 10/31/2025 10:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,KALEB,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MORRIS,ERIC
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/26/2025 02:32:13

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.078	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.216	02:27
Breath Volume		1.535	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.213	02:30
Breath Volume		1.398	02:30
Air Blank		0.000	02:31
Control Test		0.077	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 10/31/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,DILLON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MORRIS,ERIC
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: ORDONEZ RAMOS

DATE/TIME OF TEST: 10/26/2025 07:03:27

Diagnostics Check		OK	06:55
Air Blank		0.000	06:56
Control Test		0.078	06:56
Air Blank		0.000	06:57
Subject Sample	# 1	0.142	06:58
Breath Volume		2.113	06:58
Air Blank		0.000	06:58
Air Blank		0.000	07:00
Subject Sample	# 2	0.144	07:01
Breath Volume		1.976	07:01
Air Blank		0.000	07:02
Control Test		0.077	07:02
Air Blank		0.000	07:03
Diagnostics Check		OK	07:03

Data Download Date/Time: 10/31/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAMER,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: DELUCA,CHRISTIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 10/28/2025 00:53:57

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:51	
Control Test		0.077	00:51	
Air Blank		0.000	00:52	
Subject Sample	# 1		00:52	Subject Test Refused
Breath Volume		0.000	00:52	
Air Blank		0.000	00:53	
Control Test		0.078	00:53	
Air Blank		0.000	00:53	
Diagnostics Check		OK	00:53	

Data Download Date/Time: 10/31/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,R,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: EVANS,R
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: TRUONG

DATE/TIME OF TEST: 10/28/2025 01:10:52

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.078	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.121	01:05
Breath Volume		2.230	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.104	01:09
Breath Volume		1.167	01:09
Air Blank		0.000	01:09
Control Test		0.078	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 10/31/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,R,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: EVANS,R
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 10/30/2025 01:33:13

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.077	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.110	01:28
Breath Volume		2.355	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:30
Subject Sample	# 2	0.107	01:31
Breath Volume		2.753	01:31
Air Blank		0.000	01:32
Control Test		0.076	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 10/31/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.