

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,BRYAN,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MARTINI,BRYAN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/02/2025 00:05:58

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.079	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.112	00:01
Breath Volume	2.539	00:01
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.116	00:04
Breath Volume	2.957	00:04
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNALLY,TYLER,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: MCNALLY,TYLER
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: D532014870970

DATE/TIME OF TEST: 10/03/2025 02:53:56

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.079	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Control Test	0.079	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: PALACIO

DATE/TIME OF TEST: 10/04/2025 03:27:08

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:23	No Sample Provided
Breath Volume	0.000	03:23	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 2	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.079	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: GELVIN LEONEL

DATE/TIME OF TEST: 10/04/2025 22:29:20

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.079	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.227	22:24
Breath Volume	2.765	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.221	22:27
Breath Volume	2.238	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: RAFFERTY

DATE/TIME OF TEST: 10/06/2025 05:07:49

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.079	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.076	05:03
Breath Volume	2.769	05:03
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.076	05:06
Breath Volume	3.289	05:06
Air Blank	0.000	05:06
Control Test	0.079	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: JULIEN

DATE/TIME OF TEST: 10/07/2025 02:10:11

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:58	
Control Test	0.080	01:58	
Air Blank	0.000	01:58	
Subject Sample	# 1 0.058	02:02	Volume Not Met
Breath Volume	0.714	02:02	
Air Blank	0.000	02:02	
Air Blank	0.000	02:04	
Subject Sample	# 2 0.063	02:05	
Breath Volume	1.988	02:05	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample	# 3 0.063	02:08	
Breath Volume	1.222	02:08	
Air Blank	0.000	02:09	
Control Test	0.079	02:09	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 10/15/2025 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURENE,BIKENDY,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TURENE,BIKENDY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SOTHERN

DATE/TIME OF TEST: 10/08/2025 22:01:31

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:44	
Control Test	0.079	21:45	
Air Blank	0.000	21:45	
Subject Sample	# 1	21:48	No Sample Provided
Breath Volume	0.000	21:48	
Air Blank	0.000	21:49	
Air Blank	0.000	21:51	
Subject Sample	# 2	0.147 21:54	
Breath Volume	1.171	21:54	
Air Blank	0.000	21:54	
Air Blank	0.000	21:56	
Subject Sample	# 3	0.132 21:59	Volume Not Met
Breath Volume	0.218	21:59	
Air Blank	0.000	22:00	
Control Test	0.078	22:00	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 10/15/2025 19:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURENE,BIKENDY,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TURENE,BIKENDY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SOTHERN

DATE/TIME OF TEST: 10/08/2025 22:23:53

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:08	
Control Test	0.077	22:09	
Air Blank	0.000	22:09	
Subject Sample	# 1 0.134	22:12	Volume Not Met
Breath Volume	0.820	22:12	
Air Blank	0.000	22:13	
Air Blank	0.000	22:15	
Subject Sample	# 2 0.133	22:18	
Breath Volume	1.695	22:18	
Air Blank	0.000	22:18	
Air Blank	0.000	22:20	
Subject Sample	# 3 0.142	22:22	
Breath Volume	2.007	22:22	
Air Blank	0.000	22:22	
Control Test	0.077	22:23	
Air Blank	0.000	22:23	
Diagnostics Check	OK	22:23	

Data Download Date/Time: 10/15/2025 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNALLY, TYLER, S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: BURKE, L
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: SANTIAGO SANTIAGO

DATE/TIME OF TEST: 10/10/2025 22:16:46

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.114	22:12
Breath Volume	3.421	22:12
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.119	22:15
Breath Volume	2.964	22:15
Air Blank	0.000	22:15
Control Test	0.078	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 10/15/2025 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: GAMMON

DATE/TIME OF TEST: 10/11/2025 15:17:14

Diagnostics Check	OK	15:08
Air Blank	0.000	15:09
Control Test	0.079	15:09
Air Blank	0.000	15:09
Subject Sample # 1	0.189	15:11
Breath Volume	2.632	15:11
Air Blank	0.000	15:12
Air Blank	0.000	15:14
Subject Sample # 2	0.185	15:15
Breath Volume	2.363	15:15
Air Blank	0.000	15:16
Control Test	0.076	15:16
Air Blank	0.000	15:17
Diagnostics Check	OK	15:17

Data Download Date/Time: 10/15/2025 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 10/11/2025 20:38:12

Diagnostics Check	OK	20:33	
Air Blank	0.000	20:33	
Control Test	0.079	20:34	
Air Blank	0.000	20:34	
Subject Sample	# 1	20:36	Subject Test Refused
Breath Volume	0.000	20:36	
Air Blank	0.000	20:37	
Control Test	0.079	20:37	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 10/15/2025 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNALLY, TYLER, S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: BURKE, L
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 10/12/2025 02:41:51

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.117	02:37
Breath Volume	1.460	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.132	02:40
Breath Volume	2.261	02:40
Air Blank	0.000	02:40
Control Test	0.077	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 10/15/2025 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNALLY, TYLER, S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: MCNALLY, T
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 10/12/2025 03:14:51

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:07	
Control Test	0.078	03:07	
Air Blank	0.000	03:08	
Subject Sample	# 1 0.215	03:09	
Breath Volume	1.210	03:09	
Air Blank	0.000	03:10	
Air Blank	0.000	03:12	
Subject Sample	# 2	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Control Test	0.077	03:14	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 10/15/2025 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: BUSSEY

DATE/TIME OF TEST: 10/12/2025 23:11:26

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:03	
Control Test	0.079	23:03	
Air Blank	0.000	23:04	
Subject Sample	# 1 0.163	23:07	Volume Not Met
Breath Volume	0.273	23:07	
Air Blank	0.000	23:07	
Air Blank	0.000	23:09	
Subject Sample	# 2	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:10	
Control Test	0.078	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 10/15/2025 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 11:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: SANCHEZ DIAZ

DATE/TIME OF TEST: 10/05/2025 11:43:19

Diagnostics Check	OK	11:36
Air Blank	0.000	11:37
Control Test	0.080	11:37
Air Blank	0.000	11:38
Subject Sample # 1	0.145	11:38
Breath Volume	2.640	11:38
Air Blank	0.000	11:39
Air Blank	0.000	11:41
Subject Sample # 2	0.156	11:41
Breath Volume	2.179	11:41
Air Blank	0.000	11:42
Control Test	0.079	11:42
Air Blank	0.000	11:43
Diagnostics Check	OK	11:43

Data Download Date/Time: 10/30/2025 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIROGA,DIEGO,O
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ALTIDOR,J
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: DELVA

DATE/TIME OF TEST: 10/11/2025 04:16:43

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.081	04:09	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:12	No Sample Provided
Breath Volume	0.000	04:12	
Air Blank	0.000	04:13	
Air Blank	0.000	04:15	
Subject Sample	# 2	04:15	Subject Test Refused
Breath Volume	0.000	04:15	
Air Blank	0.000	04:15	
Control Test	0.080	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 10/30/2025 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INNOCENT,L,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: HILSON,R
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 10/12/2025 01:19:30

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.232	01:14
Breath Volume	1.453	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.240	01:17
Breath Volume	1.257	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 10/30/2025 05:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIROGA,DIEGO,O
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ALTIDOR,J
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: VICENTE DOMINGO

DATE/TIME OF TEST: 10/17/2025 04:13:22

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.225	04:03
Breath Volume	2.484	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.203	04:08
Breath Volume	1.417	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 3	0.205	04:11
Breath Volume	2.277	04:11
Air Blank	0.000	04:12
Control Test	0.075	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 10/30/2025 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: BARKER,AUSTIN
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: VILLALONA

DATE/TIME OF TEST: 10/29/2025 20:04:48

Diagnostics Check	OK	19:54	
Air Blank	0.000	19:55	
Control Test	0.080	19:55	
Air Blank	0.000	19:56	
Subject Sample	# 1 0.000	19:59	Volume Not Met
Breath Volume	0.203	19:59	
Air Blank	0.000	19:59	
Air Blank	0.000	20:01	
Subject Sample	# 2	20:03	Subject Test Refused
Breath Volume	0.203	20:03	
Air Blank	0.000	20:03	
Control Test	0.080	20:04	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:04	

Data Download Date/Time: 10/30/2025 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIBERNARDINO,BRANDON,M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CADENA,ISSAC
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/04/2025 21:33:38

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:20	
Control Test	0.083	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1 0.246	21:23	
Breath Volume	1.859	21:23	
Air Blank	0.000	21:23	
Air Blank	0.000	21:25	
Subject Sample	# 2 0.207	21:28	Volume Not Met
Breath Volume	0.300	21:28	
Air Blank	0.000	21:29	
Air Blank	0.000	21:31	
Subject Sample	# 3 0.249	21:31	
Breath Volume	2.039	21:31	
Air Blank	0.000	21:32	
Control Test	0.077	21:32	
Air Blank	0.000	21:33	
Diagnostics Check	OK	21:33	

Data Download Date/Time: 10/10/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRIENTOS,KENNETH,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: FARRAJ,SANAD
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/05/2025 03:05:36

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.082	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.080	02:59
Breath Volume	1.113	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:01
Subject Sample # 2	0.077	03:03
Breath Volume	1.128	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 10/10/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: HUAPILLA,MIGUEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HENDRY

DATE/TIME OF TEST: 10/06/2025 19:47:44

Diagnostics Check	OK	19:40
Air Blank	0.000	19:41
Control Test	0.082	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.213	19:42
Breath Volume	3.039	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.208	19:45
Breath Volume	2.710	19:45
Air Blank	0.000	19:46
Control Test	0.082	19:47
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 10/10/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIBERNARDINO,BRANDON,M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DIBERNARDINO,BRANDON
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: VELASQUEZ PEREZ

DATE/TIME OF TEST: 10/09/2025 23:57:38

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.082	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.215	23:52
Breath Volume	2.953	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.206	23:55
Breath Volume	2.625	23:55
Air Blank	0.000	23:56
Control Test	0.082	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 10/10/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRIENTOS,KENNETH,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: FARRAJ,SANAD
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 10/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: STEPHAN

DATE/TIME OF TEST: 10/18/2025 19:59:27

Diagnostics Check	OK	19:58	
Air Blank		19:58	Sequence Aborted
Air Blank	0.000	19:59	

Data Download Date/Time: 11/14/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRIENTOS,KENNETH,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: FARRAJ,SANAD
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 10/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: STEPHAN

DATE/TIME OF TEST: 10/18/2025 20:25:49

Diagnostics Check	OK	20:17
Air Blank	0.000	20:18
Control Test	0.083	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.111	20:19
Breath Volume	1.667	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.117	20:23
Breath Volume	1.328	20:23
Air Blank	0.000	20:24
Control Test	0.083	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 11/14/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRIENTOS,KENNETH,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CADENA,ISAAC
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 10/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: SHANLEY

DATE/TIME OF TEST: 10/19/2025 21:14:12

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:11	
Control Test	0.083	21:11	
Air Blank	0.000	21:12	
Subject Sample	# 1	21:12	Subject Test Refused
Breath Volume	0.000	21:12	
Air Blank	0.000	21:13	
Control Test	0.083	21:13	
Air Blank	0.000	21:14	
Diagnostics Check	OK	21:14	

Data Download Date/Time: 11/14/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, LAURA,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BLANCO, DAVIAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 10/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:28

SUBJECT'S LAST NAME: MEDINA FELLOVE

DATE/TIME OF TEST: 10/25/2025 13:06:33

Diagnostics Check	OK	12:59
Air Blank	0.000	13:00
Control Test	0.083	13:00
Air Blank	0.000	13:01
Subject Sample # 1	0.147	13:01
Breath Volume	1.453	13:01
Air Blank	0.000	13:02
Air Blank	0.000	13:04
Subject Sample # 2	0.140	13:04
Breath Volume	1.652	13:04
Air Blank	0.000	13:05
Control Test	0.080	13:05
Air Blank	0.000	13:06
Diagnostics Check	OK	13:06

Data Download Date/Time: 11/14/2025 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: GOSA,RONALD
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 10/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: SALERA

DATE/TIME OF TEST: 10/25/2025 23:09:30

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:07	
Control Test	0.083	23:07	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:08	
Control Test	0.083	23:08	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 11/14/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: PEREZ DELGADO

DATE/TIME OF TEST: 10/07/2025 01:36:24

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.142	01:31
Breath Volume	2.464	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.139	01:34
Breath Volume	3.042	01:34
Air Blank	0.000	01:35
Control Test	0.076	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 10/17/2025 09:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 10/15/2025 19:51:21

Diagnostics Check	OK	19:45
Air Blank	0.000	19:45
Control Test	0.081	19:45
Air Blank	0.000	19:46
Subject Sample # 1	0.124	19:46
Breath Volume	2.054	19:46
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.120	19:49
Breath Volume	3.628	19:49
Air Blank	0.000	19:50
Control Test	0.080	19:50
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 10/17/2025 09:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/17/2025 00:50:11

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.081	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.000	00:45
Breath Volume	1.937	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.000	00:48
Breath Volume	2.847	00:48
Air Blank	0.000	00:49
Control Test	0.077	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 10/17/2025 09:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SALAS

DATE/TIME OF TEST: 10/18/2025 01:30:47

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.218	01:25
Breath Volume	1.640	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.215	01:29
Breath Volume	1.425	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 11/10/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: DAWES

DATE/TIME OF TEST: 10/18/2025 02:41:33

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.081	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.229	02:36
Breath Volume	1.507	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.246	02:39
Breath Volume	1.582	02:39
Air Blank	0.000	02:40
Control Test	0.078	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 11/10/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORIARTY,QUINN,D
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: ILLAMES,MARTIN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MOYA MATA

DATE/TIME OF TEST: 10/18/2025 04:30:50

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.040	04:26
Breath Volume	2.289	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.039	04:29
Breath Volume	3.039	04:29
Air Blank	0.000	04:29
Control Test	0.079	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 11/10/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/19/2025 04:23:50

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.082	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1 0.070	04:13	
Breath Volume	2.648	04:13	
Air Blank	0.000	04:13	
Air Blank	0.000	04:15	
Subject Sample	# 2	04:18	No Sample Provided
Breath Volume	0.000	04:18	
Air Blank	0.000	04:19	
Air Blank	0.000	04:21	
Subject Sample	# 3 0.065	04:22	
Breath Volume	1.218	04:22	
Air Blank	0.000	04:22	
Control Test	0.081	04:23	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

Data Download Date/Time: 11/10/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 10/19/2025 19:25:30

Diagnostics Check	OK	19:19
Air Blank	0.000	19:19
Control Test	0.081	19:19
Air Blank	0.000	19:20
Subject Sample # 1	0.163	19:20
Breath Volume	2.597	19:20
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.157	19:23
Breath Volume	2.554	19:23
Air Blank	0.000	19:24
Control Test	0.077	19:24
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 11/10/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 10/19/2025 23:16:35

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:14	
Control Test	0.080	23:15	
Air Blank		23:15	Purge Fail
Air Blank		23:16	Purge Fail

Data Download Date/Time: 11/10/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 10/20/2025 18:43:31

Diagnostics Check	OK	18:37
Air Blank	0.000	18:37
Control Test	0.081	18:37
Air Blank	0.000	18:38
Subject Sample # 1	0.000	18:39
Breath Volume	1.902	18:39
Air Blank	0.000	18:39
Air Blank	0.000	18:41
Subject Sample # 2	0.000	18:42
Breath Volume	2.917	18:42
Air Blank	0.000	18:42
Control Test	0.079	18:42
Air Blank	0.000	18:43
Diagnostics Check	OK	18:43

Data Download Date/Time: 11/10/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLEHER,RICHARD,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SAVAGLIO,JAMES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: LEVATO

DATE/TIME OF TEST: 10/24/2025 13:56:32

Diagnostics Check	OK	13:49
Air Blank	0.000	13:49
Control Test	0.082	13:50
Air Blank	0.000	13:50
Subject Sample # 1	0.266	13:51
Breath Volume	2.171	13:51
Air Blank	0.000	13:52
Air Blank	0.000	13:53
Subject Sample # 2	0.252	13:54
Breath Volume	2.113	13:54
Air Blank	0.000	13:55
Control Test	0.079	13:55
Air Blank	0.000	13:56
Diagnostics Check	OK	13:56

Data Download Date/Time: 11/10/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: CANFIELD

DATE/TIME OF TEST: 10/24/2025 23:54:21

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.081	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.000	23:50
Breath Volume	2.503	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.000	23:52
Breath Volume	2.039	23:52
Air Blank	0.000	23:53
Control Test	0.079	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 11/10/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BONNET,GREGORY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 10/25/2025 01:27:49

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.082	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.135	01:23
Breath Volume	2.984	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:25
Subject Sample # 2	0.135	01:26
Breath Volume	3.425	01:26
Air Blank	0.000	01:26
Control Test	0.077	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 11/10/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPRINGMAN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: PALMA

DATE/TIME OF TEST: 10/25/2025 01:37:18

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.207	01:32
Breath Volume	1.800	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.191	01:35
Breath Volume	1.394	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 11/10/2025 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: BAES

DATE/TIME OF TEST: 10/26/2025 06:53:47

Diagnostics Check	OK	06:49	
Air Blank	0.000	06:49	
Control Test	0.081	06:50	
Air Blank	0.000	06:50	
Subject Sample	# 1	06:52	Subject Test Refused
Breath Volume	0.000	06:52	
Air Blank	0.000	06:52	
Control Test	0.081	06:53	
Air Blank	0.000	06:53	
Diagnostics Check	OK	06:53	

Data Download Date/Time: 11/10/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILSON,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: SPEARS

DATE/TIME OF TEST: 10/03/2025 05:27:03

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.081	05:21
Air Blank	0.000	05:21
Subject Sample # 1	0.186	05:22
Breath Volume	1.769	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:24
Subject Sample # 2	0.189	05:25
Breath Volume	1.664	05:25
Air Blank	0.000	05:26
Control Test	0.080	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:27

Data Download Date/Time: 10/23/2025 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/04/2025 04:42:39

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.000	04:38
Breath Volume	1.289	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.000	04:41
Breath Volume	1.542	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 10/23/2025 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BRESSLER

DATE/TIME OF TEST: 10/05/2025 01:33:05

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.116	01:28
Breath Volume	1.988	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.113	01:31
Breath Volume	1.421	01:31
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 10/23/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: ORBE

DATE/TIME OF TEST: 10/05/2025 19:30:44

Diagnostics Check	OK	19:23
Air Blank	0.000	19:24
Control Test	0.080	19:24
Air Blank	0.000	19:25
Subject Sample # 1	0.299	19:25
Breath Volume	4.984	19:25
Air Blank	0.000	19:26
Air Blank	0.000	19:28
Subject Sample # 2	0.296	19:28
Breath Volume	5.222	19:28
Air Blank	0.000	19:29
Control Test	0.079	19:30
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 10/23/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: LOPEZ MORENTE

DATE/TIME OF TEST: 10/06/2025 03:11:11

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.106	03:06
Breath Volume	3.296	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:08
Subject Sample # 2	0.107	03:09
Breath Volume	2.824	03:09
Air Blank	0.000	03:10
Control Test	0.078	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 10/23/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: JANASEK

DATE/TIME OF TEST: 10/07/2025 01:11:46

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.080	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.198	01:07
Breath Volume	2.441	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.195	01:10
Breath Volume	2.777	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 10/23/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 10/11/2025 03:56:07

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.172	03:51
Breath Volume	2.253	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample # 2	0.166	03:54
Breath Volume	2.113	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 10/23/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY, K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: STEIN

DATE/TIME OF TEST: 10/13/2025 19:16:11

Diagnostics Check	OK	19:07	
Air Blank	0.000	19:08	
Control Test	0.080	19:08	
Air Blank	0.000	19:08	
Subject Sample	# 1	19:12	No Sample Provided
Breath Volume	0.000	19:12	
Air Blank	0.000	19:12	
Air Blank	0.000	19:14	
Subject Sample	# 2	19:14	Subject Test Refused
Breath Volume	0.000	19:14	
Air Blank	0.000	19:15	
Control Test	0.080	19:15	
Air Blank	0.000	19:16	
Diagnostics Check	OK	19:16	

Data Download Date/Time: 10/23/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MARTIN,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/14/2025 23:50:10

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:42	
Control Test	0.080	23:42	
Air Blank	0.000	23:42	
Subject Sample	# 1	23:46	No Sample Provided
Breath Volume	0.000	23:46	
Air Blank	0.000	23:46	
Air Blank	0.000	23:48	
Subject Sample	# 2	23:48	Subject Test Refused
Breath Volume	0.000	23:48	
Air Blank	0.000	23:49	
Control Test	0.080	23:49	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

Data Download Date/Time: 10/23/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: RASH

DATE/TIME OF TEST: 10/15/2025 00:04:58

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.080	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.151	00:00
Breath Volume	1.753	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:02
Subject Sample # 2	0.150	00:03
Breath Volume	2.050	00:03
Air Blank	0.000	00:04
Control Test	0.079	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 10/23/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: DRAUGHON,B
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: VIRUET

DATE/TIME OF TEST: 10/17/2025 04:34:54

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.081	04:24	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:27	No Sample Provided
Breath Volume	0.000	04:27	
Air Blank	0.000	04:28	
Air Blank	0.000	04:30	
Subject Sample	# 2	04:33	No Sample Provided
Breath Volume	0.000	04:33	
Air Blank	0.000	04:33	
Control Test	0.080	04:34	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

Data Download Date/Time: 10/23/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: ALACK

DATE/TIME OF TEST: 10/17/2025 23:13:29

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.081	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.000	23:08
Breath Volume	1.718	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.000	23:12
Breath Volume	3.351	23:12
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 10/23/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY, K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: MEZQUITE

DATE/TIME OF TEST: 10/18/2025 19:34:31

Diagnostics Check	OK	19:26	
Air Blank	0.000	19:26	
Control Test	0.079	19:26	
Air Blank	0.000	19:27	
Subject Sample	# 1	19:30	No Sample Provided
Breath Volume	0.000	19:30	
Air Blank	0.000	19:30	
Air Blank	0.000	19:32	
Subject Sample	# 2	19:33	Subject Test Refused
Breath Volume	0.000	19:33	
Air Blank	0.000	19:33	
Control Test	0.080	19:33	
Air Blank	0.000	19:34	
Diagnostics Check	OK	19:34	

Data Download Date/Time: 10/23/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: LOWRY

DATE/TIME OF TEST: 10/18/2025 23:40:43

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:32	
Control Test	0.081	23:32	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:36	No Sample Provided
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Air Blank	0.000	23:39	
Subject Sample	# 2	23:39	Subject Test Refused
Breath Volume	0.000	23:39	
Air Blank	0.000	23:39	
Control Test	0.081	23:40	
Air Blank	0.000	23:40	
Diagnostics Check	OK	23:40	

Data Download Date/Time: 10/23/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DANFORD

DATE/TIME OF TEST: 10/19/2025 03:34:15

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.081	03:21	
Air Blank	0.000	03:22	
Subject Sample # 1	0.094	03:23	
Breath Volume	1.679	03:23	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample # 2	0.103	03:29	Volume Not Met
Breath Volume	0.230	03:29	
Air Blank	0.000	03:30	
Air Blank	0.000	03:31	
Subject Sample # 3	0.091	03:32	
Breath Volume	1.710	03:32	
Air Blank	0.000	03:33	
Control Test	0.078	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 10/23/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MCQUEEN,G
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: MULANAX

DATE/TIME OF TEST: 10/19/2025 03:48:45

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.275	03:43
Breath Volume	2.050	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.277	03:47
Breath Volume	3.042	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 10/23/2025 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ERNEST,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: LEMERAND

DATE/TIME OF TEST: 10/19/2025 07:59:38

Diagnostics Check	OK	07:56	
Air Blank	0.000	07:57	
Control Test	0.080	07:57	
Air Blank	0.000	07:58	
Subject Sample	# 1	07:58	Subject Test Refused
Breath Volume	0.000	07:58	
Air Blank	0.000	07:58	
Control Test	0.080	07:59	
Air Blank	0.000	07:59	
Diagnostics Check	OK	07:59	

Data Download Date/Time: 10/23/2025 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: NECESSARY

DATE/TIME OF TEST: 10/19/2025 20:19:15

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:11	
Control Test	0.080	20:11	
Air Blank	0.000	20:12	
Subject Sample	# 1	20:15	No Sample Provided
Breath Volume	0.000	20:15	
Air Blank	0.000	20:15	
Air Blank	0.000	20:17	
Subject Sample	# 2	20:17	Subject Test Refused
Breath Volume	0.000	20:17	
Air Blank	0.000	20:18	
Control Test	0.080	20:18	
Air Blank	0.000	20:19	
Diagnostics Check	OK	20:19	

Data Download Date/Time: 10/23/2025 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: LUKE

DATE/TIME OF TEST: 10/20/2025 04:03:42

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.094	03:58
Breath Volume	1.218	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.094	04:02
Breath Volume	1.914	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 10/23/2025 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 10/21/2025 19:10:01

Diagnostics Check	OK	19:03
Air Blank	0.000	19:03
Control Test	0.080	19:03
Air Blank	0.000	19:04
Subject Sample # 1	0.287	19:04
Breath Volume	2.234	19:04
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.297	19:08
Breath Volume	1.523	19:08
Air Blank	0.000	19:09
Control Test	0.079	19:09
Air Blank	0.000	19:09
Diagnostics Check	OK	19:10

Data Download Date/Time: 10/23/2025 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: CHARTIER

DATE/TIME OF TEST: 10/23/2025 01:25:45

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.080	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:21	No Sample Provided
Breath Volume	0.000	01:21	
Air Blank	0.000	01:22	
Air Blank	0.000	01:24	
Subject Sample	# 2	01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:24	
Control Test	0.080	01:25	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 10/23/2025 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILSON,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: WISNIEWSKI

DATE/TIME OF TEST: 10/30/2025 22:09:57

Diagnostics Check	OK	22:01	
Air Blank	0.000	22:01	
Control Test	0.081	22:02	
Air Blank	0.000	22:02	
Subject Sample	# 1	22:05	No Sample Provided
Breath Volume	0.000	22:05	
Air Blank	0.000	22:06	
Air Blank	0.000	22:08	
Subject Sample	# 2	22:08	Subject Test Refused
Breath Volume	0.000	22:08	
Air Blank	0.000	22:09	
Control Test	0.081	22:09	
Air Blank	0.000	22:09	
Diagnostics Check	OK	22:09	

Data Download Date/Time: 11/30/2025 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MAROTO,VALERY
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 10/17/2025 23:59:22

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.104	23:54
Breath Volume	2.292	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.107	23:57
Breath Volume	2.761	23:57
Air Blank	0.000	23:58
Control Test	0.081	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 10/18/2025 00:34:52

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.105	00:30
Breath Volume	2.656	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:32
Subject Sample # 2	0.104	00:33
Breath Volume	2.257	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: KAY

DATE/TIME OF TEST: 10/18/2025 01:06:58

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.015	01:02
Breath Volume	1.808	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.014	01:05
Breath Volume	2.328	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GABRIEL, JACKYME
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: DUMEL

DATE/TIME OF TEST: 10/19/2025 01:29:05

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.139	01:24
Breath Volume	1.808	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:26
Subject Sample # 2	0.146	01:27
Breath Volume	1.921	01:27
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:29

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOLLIS,JARVIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:19

SUBJECT'S LAST NAME: DACELIN

DATE/TIME OF TEST: 10/21/2025 09:09:40

Diagnostics Check	OK	09:03
Air Blank	0.000	09:03
Control Test	0.081	09:04
Air Blank	0.000	09:04
Subject Sample # 1	0.141	09:05
Breath Volume	1.632	09:05
Air Blank	0.000	09:05
Air Blank	0.000	09:07
Subject Sample # 2	0.141	09:08
Breath Volume	1.460	09:08
Air Blank	0.000	09:08
Control Test	0.080	09:09
Air Blank	0.000	09:09
Diagnostics Check	OK	09:09

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: ALABRE

DATE/TIME OF TEST: 10/22/2025 02:08:40

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.198	02:04
Breath Volume	1.429	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.206	02:07
Breath Volume	1.414	02:07
Air Blank	0.000	02:07
Control Test	0.079	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VALENCIA, IVAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: FERREIRA DE MOURA

DATE/TIME OF TEST: 10/22/2025 18:04:43

Diagnostics Check	OK	17:52	
Air Blank	0.000	17:53	
Control Test	0.080	17:53	
Air Blank	0.000	17:54	
Subject Sample	# 1 0.290	17:55	
Breath Volume	1.628	17:55	
Air Blank	0.000	17:55	
Air Blank	0.000	17:57	
Subject Sample	# 2	18:00	No Sample Provided
Breath Volume	0.000	18:00	
Air Blank	0.000	18:01	
Air Blank	0.000	18:03	
Subject Sample	# 3	18:03	Subject Test Refused
Breath Volume	0.000	18:03	
Air Blank	0.000	18:03	
Control Test	0.080	18:04	
Air Blank	0.000	18:04	
Diagnostics Check	OK	18:04	

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: FRANTZ

DATE/TIME OF TEST: 10/22/2025 23:14:05

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.156	23:09
Breath Volume	1.238	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:11
Subject Sample # 2	0.161	23:12
Breath Volume	1.316	23:12
Air Blank	0.000	23:13
Control Test	0.081	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: JEROME

DATE/TIME OF TEST: 10/23/2025 00:39:24

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.080	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.345	00:31
Breath Volume	1.285	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.368	00:34
Breath Volume	1.164	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:36
Subject Sample # 3	0.372	00:37
Breath Volume	1.132	00:37
Air Blank	0.000	00:38
Control Test	0.078	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALKER,MASON
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: LLERENA

DATE/TIME OF TEST: 10/23/2025 21:09:22

Diagnostics Check	OK	21:01
Air Blank	0.000	21:02
Control Test	0.080	21:02
Air Blank	0.000	21:03
Subject Sample # 1	0.143	21:03
Breath Volume	1.894	21:03
Air Blank	0.000	21:04
Air Blank	0.000	21:05
Subject Sample # 2	0.138	21:07
Breath Volume	2.183	21:07
Air Blank	0.000	21:08
Control Test	0.080	21:08
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEFILIPPIS, MICHAEL
ARRESTING OFFICER AGENCY: Manalapan Pub Safety
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 10/23/2025 21:57:23

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.080	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.085	21:52
Breath Volume	1.707	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.086	21:55
Breath Volume	2.023	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,BROOKE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 10/24/2025 00:04:22

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.080	23:58
Air Blank	0.000	23:59
Subject Sample # 2	0.189	23:59
Breath Volume	1.253	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.183	00:02
Breath Volume	1.734	00:02
Air Blank	0.000	00:03
Control Test	0.080	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: MATIAS FRANCISCO

DATE/TIME OF TEST: 10/24/2025 00:21:48

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.080	00:14	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:17	No Sample Provided
Breath Volume	0.000	00:17	
Air Blank	0.000	00:18	
Air Blank	0.000	00:20	
Subject Sample	# 2	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:20	
Control Test	0.080	00:21	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: OWEN,MATTHEW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: TOMALO

DATE/TIME OF TEST: 10/24/2025 17:03:23

Diagnostics Check	OK	16:56
Air Blank	0.000	16:56
Control Test	0.081	16:57
Air Blank	0.000	16:57
Subject Sample # 1	0.255	16:58
Breath Volume	2.914	16:58
Air Blank	0.000	16:59
Air Blank	0.000	17:01
Subject Sample # 2	0.253	17:01
Breath Volume	2.914	17:01
Air Blank	0.000	17:02
Control Test	0.079	17:02
Air Blank	0.000	17:03
Diagnostics Check	OK	17:03

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: NICHOLAS RAMON

DATE/TIME OF TEST: 10/24/2025 23:45:51

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.081	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.165	23:41
Breath Volume	1.652	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.164	23:44
Breath Volume	1.601	23:44
Air Blank	0.000	23:44
Control Test	0.079	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: CUTLER

DATE/TIME OF TEST: 10/28/2025 03:41:33

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.210	03:36
Breath Volume	2.371	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.216	03:39
Breath Volume	2.406	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: MIGUEL RIVERA

DATE/TIME OF TEST: 10/28/2025 05:15:34

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:08
Subject Sample # 1	0.180	05:10
Breath Volume	2.078	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.182	05:13
Breath Volume	1.570	05:13
Air Blank	0.000	05:14
Control Test	0.078	05:14
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GARCIA,ISRAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: LUIS

DATE/TIME OF TEST: 10/28/2025 16:54:04

Diagnostics Check	OK	16:49	
Air Blank	0.000	16:49	
Control Test	0.081	16:49	
Air Blank	0.000	16:50	
Subject Sample	# 1	16:52	Subject Test Refused
Breath Volume	0.000	16:52	
Air Blank	0.000	16:53	
Control Test	0.080	16:53	
Air Blank	0.000	16:53	
Diagnostics Check	OK	16:54	

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: LAURISSAINT

DATE/TIME OF TEST: 10/29/2025 02:56:22

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.000	02:51
Breath Volume	1.156	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.000	02:54
Breath Volume	1.257	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 10/30/2025 00:29:49

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.080	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.176	00:25
Breath Volume	1.339	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.177	00:28
Breath Volume	2.199	00:28
Air Blank	0.000	00:28
Control Test	0.079	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: GAO

DATE/TIME OF TEST: 10/30/2025 23:29:33

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:18	
Control Test	0.081	23:18	
Air Blank	0.000	23:19	
Subject Sample	# 1 0.000	23:22	Volume Not Met
Breath Volume	0.046	23:22	
Air Blank	0.000	23:22	
Air Blank	0.000	23:24	
Subject Sample	# 2 0.000	23:25	
Breath Volume	2.640	23:25	
Air Blank	0.000	23:25	
Air Blank	0.000	23:27	
Subject Sample	# 3 0.000	23:28	
Breath Volume	1.148	23:28	
Air Blank	0.000	23:28	
Control Test	0.082	23:28	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE, DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: VEGERANO

DATE/TIME OF TEST: 10/31/2025 02:35:46

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.132	02:31
Breath Volume	1.480	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.126	02:34
Breath Volume	1.238	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: KULL

DATE/TIME OF TEST: 10/31/2025 22:10:39

Diagnostics Check	OK	22:04
Air Blank	0.000	22:05
Control Test	0.081	22:05
Air Blank	0.000	22:05
Subject Sample # 1	0.110	22:06
Breath Volume	2.332	22:06
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.112	22:09
Breath Volume	2.886	22:09
Air Blank	0.000	22:09
Control Test	0.081	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 11/14/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024028434
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: SEREDA

DATE/TIME OF TEST: 10/26/2025 03:31:25

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.078	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.204	03:26
Breath Volume	1.167	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.207	03:29
Breath Volume	1.636	03:29
Air Blank	0.000	03:30
Control Test	0.077	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 11/09/2025 00:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMINI,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Holly Hills PD
ARRESTING OFFICER: DIMINI,NICHOLAS
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: CLEMMONS

DATE/TIME OF TEST: 10/08/2025 22:10:43

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.082	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.186	22:05
Breath Volume	1.585	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:07
Subject Sample # 2	0.184	22:08
Breath Volume	1.210	22:08
Air Blank	0.000	22:09
Control Test	0.078	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 10/31/2025 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,KRISTOPHER,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HALL,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/12/2025 03:30:50

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.082	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.126	03:24
Breath Volume	1.171	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.120	03:29
Breath Volume	1.242	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 10/31/2025 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JORDAN],C
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: LEWIS,JORDAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: DIPILLO

DATE/TIME OF TEST: 10/28/2025 03:42:40

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.081	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.281	03:34
Breath Volume	3.210	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.246	03:37
Breath Volume	3.105	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 3	0.261	03:40
Breath Volume	3.480	03:40
Air Blank	0.000	03:41
Control Test	0.078	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 10/31/2025 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM, LEANDRA,
BREATH TEST OPERATOR AGENCY: FDLE
ARRESTING OFFICER: TEST, TEST
ARRESTING OFFICER AGENCY: FDLE
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429602
DRY GAS EXPIRY DATE: 10/22/2026
VIOLATION CODE: SYS CHECK
ARREST TIME: 11:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 10/01/2025 12:01:32

Diagnostics Check	OK	11:58	
Air Blank	0.000	11:58	
Control Test	0.078	11:59	
Air Blank	0.000	11:59	
Subject Sample	# 1	11:59	Slope Not Met
Breath Volume	1.875	11:59	
Air Blank	0.000	12:00	
Control Test	0.077	12:00	
Air Blank	0.000	12:01	
Diagnostics Check	OK	12:01	

Data Download Date/Time: 10/01/2025 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,HARRY,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GREEN,HARRY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PETERSEN

DATE/TIME OF TEST: 10/22/2025 03:27:18

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.077	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.140	03:22
Breath Volume	3.394	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.139	03:25
Breath Volume	3.988	03:25
Air Blank	0.000	03:26
Control Test	0.077	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 11/24/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLENDER, MONICA, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALBRITTON, ROY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CUPU

DATE/TIME OF TEST: 10/25/2025 03:16:49

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.124	03:12
Breath Volume	1.871	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.130	03:15
Breath Volume	2.398	03:15
Air Blank	0.000	03:15
Control Test	0.078	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 11/24/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLENDER, MONICA, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK, J
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 11:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:48

SUBJECT'S LAST NAME: SMILEY

DATE/TIME OF TEST: 10/25/2025 23:49:36

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.078	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.204	23:44
Breath Volume	1.960	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.200	23:47
Breath Volume	1.863	23:47
Air Blank	0.000	23:48
Control Test	0.078	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 11/24/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLENDER, MONICA, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MILLENDER, MONICA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: ABARCA ESTRADA

DATE/TIME OF TEST: 10/26/2025 02:34:16

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.078	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.124	02:29
Breath Volume	3.042	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.122	02:32
Breath Volume	1.285	02:32
Air Blank	0.000	02:33
Control Test	0.078	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 11/24/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 10/12/2025 04:47:34

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.000	04:47	Control Outside Tolerance
Air Blank	0.000	04:47	

Data Download Date/Time: 11/06/2025 04:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 10/12/2025 04:58:46

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.077	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.150	04:53
Breath Volume	1.363	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.146	04:57
Breath Volume	1.253	04:57
Air Blank	0.000	04:57
Control Test	0.075	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 11/06/2025 04:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: DUKES

DATE/TIME OF TEST: 10/18/2025 04:07:17

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:53	
Control Test	0.078	03:54	
Air Blank	0.000	03:54	
Subject Sample	# 1 0.106	03:57	Volume Not Met
Breath Volume	0.496	03:57	
Air Blank	0.000	03:58	
Air Blank	0.000	04:00	
Subject Sample	# 2 0.097	04:02	
Breath Volume	1.191	04:02	
Air Blank	0.000	04:03	
Air Blank	0.000	04:04	
Subject Sample	# 3 0.104	04:05	
Breath Volume	1.402	04:05	
Air Blank	0.000	04:06	
Control Test	0.077	04:06	
Air Blank	0.000	04:07	
Diagnostics Check	OK	04:07	

Data Download Date/Time: 11/06/2025 04:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: CONROY

DATE/TIME OF TEST: 10/23/2025 01:30:26

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.078	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.257	01:21
Breath Volume	1.355	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.225	01:25
Breath Volume	1.949	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 3	0.239	01:28
Breath Volume	1.285	01:28
Air Blank	0.000	01:29
Control Test	0.076	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 11/06/2025 04:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: RICHARD

DATE/TIME OF TEST: 10/25/2025 02:45:11

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.077	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.075	02:40
Breath Volume	1.386	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.072	02:43
Breath Volume	1.363	02:43
Air Blank	0.000	02:44
Control Test	0.077	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 11/06/2025 04:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ESTRADA TORRES

DATE/TIME OF TEST: 10/02/2025 01:42:06

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.081	01:30	
Air Blank	0.000	01:31	
Subject Sample	# 1 0.168	01:34	Volume Not Met
Breath Volume	0.082	01:34	
Air Blank	0.000	01:35	
Air Blank	0.000	01:37	
Subject Sample	# 2 0.172	01:40	Volume Not Met
Breath Volume	0.160	01:40	
Air Blank	0.000	01:41	
Control Test	0.077	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 10/02/2025 23:11:13

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:08	
Control Test	0.079	23:09	
Air Blank	0.000	23:09	
Subject Sample	# 1	23:09	Subject Test Refused
Breath Volume	0.000	23:09	
Air Blank	0.000	23:10	
Control Test	0.080	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/03/2025 01:37:08

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.081	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.203	01:32
Breath Volume	1.414	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.205	01:35
Breath Volume	1.691	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MEJIA PEREZ

DATE/TIME OF TEST: 10/03/2025 04:54:17

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.081	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.147	04:49
Breath Volume	1.230	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.152	04:52
Breath Volume	2.250	04:52
Air Blank	0.000	04:53
Control Test	0.078	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN, JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: LEFLER

DATE/TIME OF TEST: 10/03/2025 22:04:01

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.000	21:59
Breath Volume	2.000	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.000	22:02
Breath Volume	1.667	22:02
Air Blank	0.000	22:03
Control Test	0.081	22:03
Air Blank	0.000	22:03
Diagnostics Check	OK	22:04

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: GODINEZ AGUILAR

DATE/TIME OF TEST: 10/04/2025 00:08:40

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.081	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.180	00:03
Breath Volume	2.804	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.181	00:06
Breath Volume	1.890	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIER, MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: GUTIERREZ GARCIA

DATE/TIME OF TEST: 10/04/2025 03:08:51

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.077	03:04
Breath Volume	2.125	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:06
Subject Sample # 2	0.076	03:07
Breath Volume	2.062	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRATEREAUX,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: INSIXIENGMAI

DATE/TIME OF TEST: 10/04/2025 06:01:15

Diagnostics Check	OK	05:54
Air Blank	0.000	05:54
Control Test	0.080	05:55
Air Blank	0.000	05:55
Subject Sample # 1	0.149	05:56
Breath Volume	1.546	05:56
Air Blank	0.000	05:57
Air Blank	0.000	05:59
Subject Sample # 2	0.149	05:59
Breath Volume	1.140	05:59
Air Blank	0.000	06:00
Control Test	0.079	06:00
Air Blank	0.000	06:01
Diagnostics Check	OK	06:01

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRATEREAUX,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: CENOPHAT

DATE/TIME OF TEST: 10/04/2025 06:26:41

Diagnostics Check	OK	06:19
Air Blank	0.000	06:19
Control Test	0.080	06:19
Air Blank	0.000	06:20
Subject Sample # 1	0.165	06:21
Breath Volume	1.812	06:21
Air Blank	0.000	06:22
Air Blank	0.000	06:24
Subject Sample # 2	0.177	06:24
Breath Volume	2.207	06:24
Air Blank	0.000	06:25
Control Test	0.079	06:26
Air Blank	0.000	06:26
Diagnostics Check	OK	06:26

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: VAZQUEZ SANCHEZ

DATE/TIME OF TEST: 10/04/2025 06:58:04

Diagnostics Check	OK	06:51
Air Blank	0.000	06:51
Control Test	0.078	06:51
Air Blank	0.000	06:52
Subject Sample # 1	0.172	06:53
Breath Volume	3.261	06:53
Air Blank	0.000	06:54
Air Blank	0.000	06:55
Subject Sample # 2	0.173	06:56
Breath Volume	3.355	06:56
Air Blank	0.000	06:57
Control Test	0.079	06:57
Air Blank	0.000	06:57
Diagnostics Check	OK	06:58

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VASQUEZ, JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/04/2025 21:46:55

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.080	21:40
Air Blank	0.000	21:40
Subject Sample # 1	0.185	21:42
Breath Volume	2.554	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:44
Subject Sample # 2	0.185	21:45
Breath Volume	2.007	21:45
Air Blank	0.000	21:45
Control Test	0.077	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: RAMOS MENDEZ

DATE/TIME OF TEST: 10/05/2025 06:05:46

Diagnostics Check	OK	05:58
Air Blank	0.000	05:59
Control Test	0.081	05:59
Air Blank	0.000	06:00
Subject Sample # 1	0.248	06:01
Breath Volume	2.039	06:01
Air Blank	0.000	06:01
Air Blank	0.000	06:03
Subject Sample # 2	0.239	06:03
Breath Volume	1.445	06:03
Air Blank	0.000	06:04
Control Test	0.077	06:05
Air Blank	0.000	06:05
Diagnostics Check	OK	06:05

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: D'ANDREA,ADAM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:36

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 10/05/2025 08:28:01

Diagnostics Check	OK	08:20
Air Blank	0.000	08:20
Control Test	0.080	08:21
Air Blank	0.000	08:21
Subject Sample # 1	0.220	08:22
Breath Volume	1.214	08:22
Air Blank	0.000	08:23
Air Blank	0.000	08:24
Subject Sample # 2	0.201	08:26
Breath Volume	2.316	08:26
Air Blank	0.000	08:27
Control Test	0.079	08:27
Air Blank	0.000	08:27
Diagnostics Check	OK	08:28

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: OVALLE MAYORGA

DATE/TIME OF TEST: 10/06/2025 01:30:17

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.183	01:25
Breath Volume	2.324	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.169	01:28
Breath Volume	1.496	01:28
Air Blank	0.000	01:29
Control Test	0.077	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PATINO,GABRIELA
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: VARNELL

DATE/TIME OF TEST: 10/06/2025 17:02:53

Diagnostics Check	OK	16:59	
Air Blank	0.000	16:59	
Control Test	0.080	17:00	
Air Blank	0.000	17:00	
Subject Sample	# 1	17:01	Interferent Detect
Breath Volume	1.800	17:01	
Air Blank	0.000	17:01	
Control Test	0.080	17:02	
Air Blank	0.000	17:02	
Diagnostics Check	OK	17:02	

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PATINO,GABRIELA
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: VARNELL

DATE/TIME OF TEST: 10/06/2025 17:23:41

Diagnostics Check	OK	17:20	
Air Blank	0.000	17:20	
Control Test	0.081	17:20	
Air Blank	0.000	17:21	
Subject Sample	# 1	17:21	Interferent Detect
Breath Volume	2.207	17:21	
Air Blank	0.000	17:22	
Control Test	0.080	17:23	
Air Blank	0.000	17:23	
Diagnostics Check	OK	17:23	

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: BRAVO GARCIA

DATE/TIME OF TEST: 10/07/2025 00:26:36

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.204	00:21
Breath Volume	3.613	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.200	00:24
Breath Volume	2.914	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: GARCIA FELIPE

DATE/TIME OF TEST: 10/07/2025 03:33:55

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.171	03:29
Breath Volume	1.671	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.178	03:32
Breath Volume	2.652	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PROPHETE,STEVE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: BIEN-AIME

DATE/TIME OF TEST: 10/09/2025 04:57:26

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.081	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.307	04:52
Breath Volume	1.859	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.301	04:55
Breath Volume	1.308	04:55
Air Blank	0.000	04:56
Control Test	0.078	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FISHBAUGH, CHARLES
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: SILVERMAN

DATE/TIME OF TEST: 10/10/2025 00:28:15

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.208	00:23
Breath Volume	1.566	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:25
Subject Sample # 2	0.206	00:26
Breath Volume	2.042	00:26
Air Blank	0.000	00:27
Control Test	0.079	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ORTIZ,JOSE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: EDWARD

DATE/TIME OF TEST: 10/10/2025 02:32:57

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.032	02:27
Breath Volume	1.441	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.028	02:31
Breath Volume	1.109	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/11/2025 05:38:25

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.080	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.094	05:33
Breath Volume	1.531	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:35
Subject Sample # 2	0.084	05:36
Breath Volume	1.492	05:36
Air Blank	0.000	05:37
Control Test	0.079	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CLARK,GRANT
ARRESTING OFFICER AGENCY: Atlantis PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: OROZCO

DATE/TIME OF TEST: 10/11/2025 07:49:02

Diagnostics Check	OK	07:42
Air Blank	0.000	07:43
Control Test	0.080	07:43
Air Blank	0.000	07:44
Subject Sample # 1	0.136	07:44
Breath Volume	1.750	07:44
Air Blank	0.000	07:45
Air Blank	0.000	07:47
Subject Sample # 2	0.135	07:47
Breath Volume	1.820	07:47
Air Blank	0.000	07:48
Control Test	0.080	07:48
Air Blank	0.000	07:48
Diagnostics Check	OK	07:49

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: COZ-GOMEZ

DATE/TIME OF TEST: 10/12/2025 02:24:31

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.190	02:19
Breath Volume	2.015	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.190	02:22
Breath Volume	2.101	02:22
Air Blank	0.000	02:23
Control Test	0.078	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: LUCAS PEREZ

DATE/TIME OF TEST: 10/12/2025 06:12:45

Diagnostics Check	OK	06:06
Air Blank	0.000	06:06
Control Test	0.080	06:06
Air Blank	0.000	06:07
Subject Sample # 1	0.187	06:07
Breath Volume	1.671	06:07
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample # 2	0.183	06:10
Breath Volume	1.800	06:10
Air Blank	0.000	06:11
Control Test	0.080	06:12
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/12/2025 21:48:21

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.082	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.169	21:42
Breath Volume	1.382	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:44
Subject Sample # 2	0.177	21:46
Breath Volume	1.621	21:46
Air Blank	0.000	21:47
Control Test	0.079	21:47
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALSTON,DEVIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ANGELI

DATE/TIME OF TEST: 10/13/2025 04:56:26

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.130	04:51
Breath Volume	1.406	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.134	04:54
Breath Volume	2.730	04:54
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: VASQUEZ BERDUO

DATE/TIME OF TEST: 10/13/2025 05:12:57

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.079	05:07
Air Blank	0.000	05:07
Subject Sample # 1	0.181	05:08
Breath Volume	1.945	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:10
Subject Sample # 2	0.178	05:11
Breath Volume	2.324	05:11
Air Blank	0.000	05:11
Control Test	0.079	05:12
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROSAS, DESIREE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 10/13/2025 23:30:15

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.196	23:25
Breath Volume	2.160	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:27
Subject Sample # 2	0.211	23:28
Breath Volume	3.332	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: EBERE,HALEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: MENDEZ-GONZALEZ

DATE/TIME OF TEST: 10/14/2025 01:37:54

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.161	01:33
Breath Volume	1.882	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.155	01:36
Breath Volume	1.710	01:36
Air Blank	0.000	01:36
Control Test	0.078	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEFILIPPIS, MICHEAL
ARRESTING OFFICER AGENCY: Manalapan Pub Safety
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 10/14/2025 21:09:22

Diagnostics Check	OK	21:02
Air Blank	0.000	21:03
Control Test	0.080	21:03
Air Blank	0.000	21:04
Subject Sample # 1	0.183	21:04
Breath Volume	1.910	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.185	21:07
Breath Volume	2.078	21:07
Air Blank	0.000	21:08
Control Test	0.079	21:08
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RHYMER,TRAVIS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:16

SUBJECT'S LAST NAME: ESPINOZA MARTINEZ

DATE/TIME OF TEST: 10/15/2025 10:59:43

Diagnostics Check	OK	10:50
Air Blank	0.000	10:51
Control Test	0.081	10:51
Air Blank	0.000	10:52
Subject Sample # 1	0.196	10:53
Breath Volume	2.667	10:53
Air Blank	0.000	10:54
Air Blank	0.000	10:55
Subject Sample # 2	0.189	10:57
Breath Volume	1.539	10:57
Air Blank	0.000	10:58
Control Test	0.078	10:59
Air Blank	0.000	10:59
Diagnostics Check	OK	10:59

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: DOUANGTAVANH

DATE/TIME OF TEST: 10/16/2025 01:56:00

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.161	01:51
Breath Volume	1.218	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:53
Subject Sample # 2	0.153	01:54
Breath Volume	1.105	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:56

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DEBERLE

DATE/TIME OF TEST: 10/16/2025 02:50:06

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.134	02:45
Breath Volume	1.539	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample # 2	0.147	02:48
Breath Volume	1.281	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/17/2025 04:56:27

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.000	04:52
Breath Volume	1.367	04:52
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.000	04:55
Breath Volume	1.554	04:55
Air Blank	0.000	04:55
Control Test	0.081	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TIANO,GIOVANNI
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:52

SUBJECT'S LAST NAME: NEMERSON

DATE/TIME OF TEST: 10/17/2025 12:36:07

Diagnostics Check	OK	12:24	
Air Blank	0.000	12:25	
Control Test	0.081	12:25	
Air Blank	0.000	12:26	
Subject Sample	# 1 0.000	12:29	Volume Not Met
Breath Volume	0.316	12:29	
Air Blank	0.000	12:29	
Air Blank	0.000	12:31	
Subject Sample	# 2	12:34	Subject Test Refused
Breath Volume	1.066	12:34	
Air Blank	0.000	12:35	
Control Test	0.080	12:35	
Air Blank	0.000	12:36	
Diagnostics Check	OK	12:36	

Data Download Date/Time: 10/17/2025 14:57