

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLACE,ELSTON,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MENA,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/31/2025 04:43:14

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:31	
Control Test	0.077	04:32	
Air Blank	0.000	04:32	
Subject Sample	# 1 0.128	04:35	Volume Not Met
Breath Volume	0.226	04:35	
Air Blank	0.000	04:36	
Air Blank	0.000	04:38	
Subject Sample	# 2 0.104	04:41	Volume Not Met
Breath Volume	0.289	04:41	
Air Blank	0.000	04:42	
Control Test	0.075	04:42	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

Data Download Date/Time: 11/25/2025 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALLACE,ELSTON,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MENA,JOSE

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402732434

DRY GAS EXPIRY DATE: 04/28/2026

VIOLATION CODE: DUI

ARREST TIME: 03:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/31/2025 04:54:50

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:50	
Control Test	0.077	04:50	
Air Blank	0.000	04:51	
Subject Sample	# 1	04:53	Subject Test Refused
Breath Volume	0.445	04:53	
Air Blank	0.000	04:53	
Control Test	0.077	04:54	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

Data Download Date/Time: 11/25/2025 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ANGEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: CABEZAS ROMERO

DATE/TIME OF TEST: 10/30/2025 22:02:55

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:51	
Control Test	0.079	21:51	
Air Blank	0.000	21:52	
Subject Sample	# 1 0.217	21:55	Slope Not Level
Breath Volume	1.863	21:55	
Air Blank	0.000	21:56	
Air Blank	0.000	21:57	
Subject Sample	# 2 0.224	21:58	
Breath Volume	2.414	21:58	
Air Blank	0.000	21:59	
Air Blank	0.000	22:00	
Subject Sample	# 3 0.229	22:01	
Breath Volume	2.468	22:01	
Air Blank	0.000	22:02	
Control Test	0.076	22:02	
Air Blank	0.000	22:02	
Diagnostics Check	OK	22:02	

Data Download Date/Time: 11/20/2025 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 12:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:23

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 10/02/2025 12:49:45

Diagnostics Check	OK	12:43
Air Blank	0.000	12:44
Control Test	0.082	12:44
Air Blank	0.000	12:45
Subject Sample # 1	0.000	12:45
Breath Volume	2.691	12:45
Air Blank	0.000	12:46
Air Blank	0.000	12:48
Subject Sample # 2	0.000	12:48
Breath Volume	2.207	12:48
Air Blank	0.000	12:48
Control Test	0.081	12:49
Air Blank	0.000	12:49
Diagnostics Check	OK	12:49

Data Download Date/Time: 10/06/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVA,JUAN,C
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: OLIVA,JUAN
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 09/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MILIANS

DATE/TIME OF TEST: 10/05/2025 22:46:43

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.081	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.236	22:42
Breath Volume	2.429	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.228	22:44
Breath Volume	3.121	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 10/09/2025 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES, EMMANUEL,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: TORRES, EMMANUEL
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 10/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/13/2025 04:33:46

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.082	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.119	04:29
Breath Volume	2.867	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:31
Subject Sample # 2	0.116	04:32
Breath Volume	2.804	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 11/05/2025 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TORRES, EMMANUEL,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: TORRES, EMMANUEL

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 10/09/2025

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 403255643

DRY GAS EXPIRY DATE: 01/29/2028

VIOLATION CODE: DUI

ARREST TIME: 00:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: MARCANO

DATE/TIME OF TEST: 10/15/2025 01:53:02

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:50	
Control Test	0.083	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:52	
Control Test	0.082	01:52	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 11/05/2025 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVA,JUAN,C
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: OLIVA,JUAN
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 10/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MORALESDUARTE

DATE/TIME OF TEST: 10/19/2025 23:10:15

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.082	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.190	23:05
Breath Volume	1.847	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.191	23:08
Breath Volume	1.859	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 11/05/2025 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEJEDA,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: TEJEDA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 10/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: NUNEZ-BARRETO

DATE/TIME OF TEST: 10/25/2025 05:23:32

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.082	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.183	05:18
Breath Volume	4.292	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.180	05:21
Breath Volume	3.871	05:21
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 11/05/2025 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES, EMMANUEL,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: TORRES, EMMANUEL
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 10/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 10/25/2025 23:46:53

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:44	
Control Test	0.082	23:44	
Air Blank	0.000	23:45	
Subject Sample	# 1	23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank	0.000	23:45	
Control Test	0.081	23:46	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 11/05/2025 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAGOS,GUILLERMO,E
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: EATON,MORTON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: OSCAR JR

DATE/TIME OF TEST: 10/07/2025 19:13:14

Diagnostics Check	OK	19:09	
Air Blank	0.000	19:10	
Control Test	0.078	19:10	
Air Blank	0.000	19:11	
Subject Sample	# 1	19:11	Subject Test Refused
Breath Volume	0.000	19:11	
Air Blank	0.000	19:12	
Control Test	0.078	19:12	
Air Blank	0.000	19:13	
Diagnostics Check	OK	19:13	

Data Download Date/Time: 10/16/2025 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,KELVIN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CUAREZMA,MICHAEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: QUINTANA PRADO

DATE/TIME OF TEST: 10/09/2025 02:04:42

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:02	
Control Test	0.079	02:02	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:03	Subject Test Refused
Breath Volume	0.000	02:03	
Air Blank	0.000	02:03	
Control Test	0.078	02:04	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 10/16/2025 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRIDO,JEYNY,S
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE,RENATO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: PICKETT

DATE/TIME OF TEST: 10/11/2025 05:19:26

Diagnostics Check	OK	05:16	
Air Blank	0.000	05:17	
Control Test	0.078	05:17	
Air Blank	0.000	05:17	
Subject Sample	# 1	05:18	Subject Test Refused
Breath Volume	0.000	05:18	
Air Blank	0.000	05:18	
Control Test	0.077	05:18	
Air Blank	0.000	05:19	
Diagnostics Check	OK	05:19	

Data Download Date/Time: 10/16/2025 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAURE CASCARET, RENATO, A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE CASCARET, RENATO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/12/2025 22:59:05

Diagnostics Check	OK	22:55	
Air Blank	0.000	22:56	
Control Test	0.078	22:56	
Air Blank	0.000	22:57	
Subject Sample	# 1	22:57	Subject Test Refused
Breath Volume	0.000	22:57	
Air Blank	0.000	22:58	
Control Test	0.078	22:58	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:59	

Data Download Date/Time: 10/16/2025 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDOSO,MARCOS,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: HECHAVARIA,EDENIS
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 10/19/2025 03:00:57

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:58	
Control Test	0.078	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	03:00	
Control Test	0.077	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 11/13/2025 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRANDON,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MILLER,BRANDON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 10/19/2025 03:11:58

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.186	03:07
Breath Volume	1.261	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.187	03:10
Breath Volume	1.464	03:10
Air Blank	0.000	03:11
Control Test	0.077	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 11/13/2025 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,KELVIN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MARRERO,GABRIEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: SORIANO MARTINEZ

DATE/TIME OF TEST: 10/20/2025 06:10:10

Diagnostics Check	OK	06:07	
Air Blank	0.000	06:07	
Control Test	0.078	06:08	
Air Blank	0.000	06:08	
Subject Sample	# 1	06:08	Subject Test Refused
Breath Volume	0.000	06:08	
Air Blank	0.000	06:09	
Control Test	0.077	06:09	
Air Blank	0.000	06:10	
Diagnostics Check	OK	06:10	

Data Download Date/Time: 11/13/2025 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: VALDES,YSABELLA
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: CICILIA RODRIGUEZ

DATE/TIME OF TEST: 10/21/2025 04:51:28

Diagnostics Check	OK	04:48	
Air Blank	0.000	04:48	
Control Test	0.078	04:49	
Air Blank	0.000	04:49	
Subject Sample	# 1	04:50	Subject Test Refused
Breath Volume	0.000	04:50	
Air Blank	0.000	04:50	
Control Test	0.078	04:50	
Air Blank	0.000	04:51	
Diagnostics Check	OK	04:51	

Data Download Date/Time: 11/13/2025 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAZARYAN,LILIT,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MARTINEZ,DANILO
ARRESTING OFFICER AGENCY: Miami-Dade School PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 10/27/2025 16:36:14

Diagnostics Check	OK	16:31	
Air Blank	0.000	16:32	
Control Test	0.076	16:32	
Air Blank	0.000	16:33	
Subject Sample	# 1	16:34	Subject Test Refused
Breath Volume	0.000	16:34	
Air Blank	0.000	16:35	
Control Test	0.077	16:35	
Air Blank	0.000	16:36	
Diagnostics Check	OK	16:36	

Data Download Date/Time: 11/13/2025 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: VALDES,YSABELLA
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: PORTAL

DATE/TIME OF TEST: 10/31/2025 00:48:03

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.078	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.214	00:42
Breath Volume	1.281	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.211	00:46
Breath Volume	1.316	00:46
Air Blank	0.000	00:47
Control Test	0.078	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:48

Data Download Date/Time: 11/13/2025 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 10/12/2025 19:40:18

Diagnostics Check	OK	19:29	
Air Blank	0.000	19:29	
Control Test	0.080	19:30	
Air Blank	0.000	19:30	
Subject Sample	# 1	19:33	No Sample Provided
Breath Volume	0.000	19:33	
Air Blank	0.000	19:34	
Air Blank	0.000	19:36	
Subject Sample	# 2	19:38	Subject Test Refused
Breath Volume	0.000	19:38	
Air Blank	0.000	19:39	
Control Test	0.078	19:39	
Air Blank	0.000	19:40	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 11/13/2025 03:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 10/12/2025 19:51:34

Diagnostics Check	OK	19:43
Air Blank	0.000	19:44
Control Test	0.079	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.130	19:46
Breath Volume	2.546	19:46
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.131	19:49
Breath Volume	2.812	19:49
Air Blank	0.000	19:50
Control Test	0.080	19:50
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 11/13/2025 03:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ZARATE,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: REYES MESA

DATE/TIME OF TEST: 10/17/2025 04:29:17

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.138	04:24
Breath Volume	1.460	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:26
Subject Sample # 2	0.132	04:27
Breath Volume	1.253	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 11/13/2025 03:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKEWELL,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: BAKEWELL,ZACHARY
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: PATINO

DATE/TIME OF TEST: 10/22/2025 00:05:50

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.079	00:03	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:04	
Control Test	0.077	00:05	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 11/13/2025 03:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERAPHIN,RANDY,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PAUL,DEJARDINS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: BELIARD

DATE/TIME OF TEST: 10/22/2025 18:33:01

Diagnostics Check	OK	18:31	
Air Blank	0.000	18:32	
Control Test	0.070	18:32	Control Outside Tolerance
Air Blank	0.000	18:33	

Data Download Date/Time: 11/13/2025 03:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERAPHIN,RANDY,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PAUL,DEJARDINS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: BELIARD

DATE/TIME OF TEST: 10/22/2025 18:43:32

Diagnostics Check	OK	18:36
Air Blank	0.000	18:36
Control Test	0.079	18:37
Air Blank	0.000	18:37
Subject Sample # 1	0.250	18:38
Breath Volume	2.125	18:38
Air Blank	0.000	18:39
Air Blank	0.000	18:41
Subject Sample # 2	0.262	18:41
Breath Volume	1.960	18:41
Air Blank	0.000	18:42
Control Test	0.077	18:42
Air Blank	0.000	18:43
Diagnostics Check	OK	18:43

Data Download Date/Time: 11/13/2025 03:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ZARATE,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BRIME

DATE/TIME OF TEST: 10/23/2025 03:44:48

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.080	03:34	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:37	No Sample Provided
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 2	03:43	No Sample Provided
Breath Volume	0.000	03:43	
Air Blank	0.000	03:43	
Control Test	0.080	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 11/13/2025 03:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ZARATE,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BRIME

DATE/TIME OF TEST: 10/23/2025 04:04:29

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1 0.074	04:00	
Breath Volume	1.312	04:00	
Air Blank	0.000	04:00	
Air Blank	0.000	04:02	
Subject Sample	# 2	04:03	RFI Detect
Breath Volume	0.000	04:03	
Air Blank	0.000	04:03	
Control Test	0.079	04:03	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 11/13/2025 03:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ZARATE,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BRIME

DATE/TIME OF TEST: 10/23/2025 04:14:02

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.069	04:08
Breath Volume	1.175	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.071	04:12
Breath Volume	1.984	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:14

Data Download Date/Time: 11/13/2025 03:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUELLO MERCADO, HAROLD,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ, CELINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: ALSINA

DATE/TIME OF TEST: 10/25/2025 23:24:18

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.080	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.140	23:19
Breath Volume	1.929	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.144	23:22
Breath Volume	1.593	23:22
Air Blank	0.000	23:23
Control Test	0.078	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 11/13/2025 03:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,CRISTIAN,X
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERRO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:48

SUBJECT'S LAST NAME: PEREZ FERRARI

DATE/TIME OF TEST: 10/01/2025 11:14:27

Diagnostics Check	OK	11:08
Air Blank	0.000	11:08
Control Test	0.080	11:09
Air Blank	0.000	11:09
Subject Sample # 1	0.299	11:09
Breath Volume	1.472	11:09
Air Blank	0.000	11:10
Air Blank	0.000	11:12
Subject Sample # 2	0.310	11:12
Breath Volume	1.871	11:12
Air Blank	0.000	11:13
Control Test	0.078	11:13
Air Blank	0.000	11:14
Diagnostics Check	OK	11:14

Data Download Date/Time: 10/02/2025 02:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MALDONADO,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ECHEVARRIALIENS

DATE/TIME OF TEST: 10/07/2025 00:37:56

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:25	
Control Test	0.082	00:26	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:29	No Sample Provided
Breath Volume	0.000	00:29	
Air Blank	0.000	00:30	
Air Blank	0.000	00:32	
Subject Sample	# 2	00:33	
Breath Volume	3.234	00:33	
Air Blank	0.000	00:33	
Air Blank	0.000	00:35	
Subject Sample	# 3	00:36	
Breath Volume	3.828	00:36	
Air Blank	0.000	00:36	
Control Test	0.080	00:37	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 11/13/2025 04:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MALDONADO,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 10/12/2025 23:03:41

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.082	22:53	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:56	No Sample Provided
Breath Volume	0.000	22:56	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample	# 2	23:02	No Sample Provided
Breath Volume	0.000	23:02	
Air Blank	0.000	23:02	
Control Test	0.081	23:03	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 11/13/2025 04:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GONZALEZ,L
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CAPDEVILA

DATE/TIME OF TEST: 10/13/2025 00:21:35

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.082	00:10	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:14	No Sample Provided
Breath Volume	0.000	00:14	
Air Blank	0.000	00:15	
Air Blank	0.000	00:17	
Subject Sample	# 2	00:20	No Sample Provided
Breath Volume	0.000	00:20	
Air Blank	0.000	00:20	
Control Test	0.081	00:20	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 11/13/2025 04:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGGINS,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ACOSTA,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: GARCIA MIRABAL

DATE/TIME OF TEST: 10/18/2025 20:31:57

Diagnostics Check	OK	20:22
Air Blank	0.000	20:23
Control Test	0.081	20:23
Air Blank	0.000	20:24
Subject Sample # 1	0.119	20:26
Breath Volume	1.261	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.106	20:30
Breath Volume	2.894	20:30
Air Blank	0.000	20:31
Control Test	0.079	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 11/13/2025 04:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGGINS,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ST. AUBYN,J
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 10/24/2025 17:32:52

Diagnostics Check	OK	17:29	
Air Blank	0.000	17:29	
Control Test	0.081	17:29	
Air Blank	0.000	17:30	
Subject Sample	# 1	17:31	Subject Test Refused
Breath Volume	0.000	17:31	
Air Blank	0.000	17:31	
Control Test	0.080	17:32	
Air Blank	0.000	17:32	
Diagnostics Check	OK	17:32	

Data Download Date/Time: 11/13/2025 04:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE JESUS,JULIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DE JESUS,JULIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: SIFUENTES ROMERO

DATE/TIME OF TEST: 10/25/2025 23:45:32

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.081	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.171	23:41
Breath Volume	2.382	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.176	23:43
Breath Volume	3.992	23:43
Air Blank	0.000	23:44
Control Test	0.079	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 11/13/2025 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MEJIA RESTREPO

DATE/TIME OF TEST: 10/23/2025 04:07:02

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.007	04:02
Breath Volume	1.503	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.000	04:05
Breath Volume	1.843	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:07

Data Download Date/Time: 11/05/2025 21:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:33

SUBJECT'S LAST NAME: CENTENO GUTIERREZ

DATE/TIME OF TEST: 10/25/2025 07:59:31

Diagnostics Check	OK	07:53
Air Blank	0.000	07:53
Control Test	0.080	07:54
Air Blank	0.000	07:54
Subject Sample # 1	0.077	07:54
Breath Volume	3.457	07:54
Air Blank	0.000	07:55
Air Blank	0.000	07:57
Subject Sample # 2	0.067	07:57
Breath Volume	1.343	07:57
Air Blank	0.000	07:58
Control Test	0.080	07:58
Air Blank	0.000	07:59
Diagnostics Check	OK	07:59

Data Download Date/Time: 11/05/2025 21:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMSON,ROBERT,H
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: ASBURY,BRIAN
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: MALTEZ

DATE/TIME OF TEST: 10/01/2025 09:38:28

Diagnostics Check	OK	09:35	
Air Blank	0.000	09:36	
Control Test	0.079	09:36	
Air Blank	0.000	09:36	
Subject Sample	# 1	09:37	Subject Test Refused
Breath Volume	0.000	09:37	
Air Blank	0.000	09:37	
Control Test	0.080	09:37	
Air Blank	0.000	09:38	
Diagnostics Check	OK	09:38	

Data Download Date/Time: 10/02/2025 04:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/02/2025 03:13:06

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.079	03:11	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:12	
Control Test	0.080	03:12	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 10/02/2025 04:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SILVA,ANTHONY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: MASULLO FERNANDEZ

DATE/TIME OF TEST: 10/03/2025 02:46:59

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.079	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.079	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 11/13/2025 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORENO,FRANCISCO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:32

SUBJECT'S LAST NAME: VILLAFUERTE

DATE/TIME OF TEST: 10/05/2025 09:56:44

Diagnostics Check	OK	09:52	
Air Blank	0.000	09:53	
Control Test	0.080	09:53	
Air Blank	0.000	09:54	
Subject Sample # 1	0.187	09:54	
Breath Volume	1.511	09:54	
Air Blank		09:55	RFI Detect
Air Blank	0.000	09:55	
Control Test	0.080	09:56	
Air Blank	0.000	09:56	
Diagnostics Check	OK	09:56	

Data Download Date/Time: 11/13/2025 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORENO,FRANCISCO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:32

SUBJECT'S LAST NAME: VILLAFUERTE

DATE/TIME OF TEST: 10/05/2025 10:05:53

Diagnostics Check	OK	09:59
Air Blank	0.000	09:59
Control Test	0.079	10:00
Air Blank	0.000	10:00
Subject Sample # 1	0.185	10:01
Breath Volume	1.546	10:01
Air Blank	0.000	10:01
Air Blank	0.000	10:03
Subject Sample # 2	0.192	10:04
Breath Volume	2.003	10:04
Air Blank	0.000	10:04
Control Test	0.077	10:05
Air Blank	0.000	10:05
Diagnostics Check	OK	10:05

Data Download Date/Time: 11/13/2025 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OVIEDO,ANDRES,J
BREATH TEST OPERATOR AGENCY: Medley PD
ARRESTING OFFICER: OVIEDO,ANDRES
ARRESTING OFFICER AGENCY: Medley PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: CARDOSO RODRIGUEZ

DATE/TIME OF TEST: 10/11/2025 03:09:08

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.078	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.167	03:04
Breath Volume	1.753	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:06
Subject Sample # 2	0.183	03:07
Breath Volume	1.695	03:07
Air Blank	0.000	03:08
Control Test	0.076	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 11/13/2025 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,WILFREDO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ORTEGA,WILFREDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:52

SUBJECT'S LAST NAME: SUAREZ TRUJILLO

DATE/TIME OF TEST: 10/16/2025 09:24:25

Diagnostics Check	OK	09:20	
Air Blank	0.000	09:21	
Control Test	0.079	09:21	
Air Blank	0.000	09:22	
Subject Sample	# 1	09:22	Subject Test Refused
Breath Volume	1.101	09:22	
Air Blank	0.000	09:23	
Control Test	0.080	09:23	
Air Blank	0.000	09:24	
Diagnostics Check	OK	09:24	

Data Download Date/Time: 11/13/2025 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OVIEDO,ANDRES,J
BREATH TEST OPERATOR AGENCY: Medley PD
ARRESTING OFFICER: OVIEDO,ANDRES
ARRESTING OFFICER AGENCY: Medley PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: LOPEZ GARCIA

DATE/TIME OF TEST: 10/18/2025 01:40:25

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:37	
Control Test	0.077	01:38	
Air Blank	0.000	01:38	
Subject Sample	# 1	01:39	RFI Detect
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.078	01:39	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 11/13/2025 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OVIEDO,ANDRES,J
BREATH TEST OPERATOR AGENCY: Medley PD
ARRESTING OFFICER: OVIEDO,ANDRES
ARRESTING OFFICER AGENCY: Medley PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: LOPEZ GARCIA

DATE/TIME OF TEST: 10/18/2025 01:47:52

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.079	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.177	01:43
Breath Volume	3.195	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:45
Subject Sample # 2	0.179	01:46
Breath Volume	2.792	01:46
Air Blank	0.000	01:46
Control Test	0.078	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 11/13/2025 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OVIEDO,ANDRES,J
BREATH TEST OPERATOR AGENCY: Medley PD
ARRESTING OFFICER: OVIEDO,ANDRES
ARRESTING OFFICER AGENCY: Medley PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 10/19/2025 03:37:06

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.159	03:32
Breath Volume	1.589	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.145	03:35
Breath Volume	3.851	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 11/13/2025 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,WILFREDO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ORTEGA,WILFREDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: DUQUE

DATE/TIME OF TEST: 10/19/2025 08:17:32

Diagnostics Check	OK	08:11
Air Blank	0.000	08:11
Control Test	0.080	08:12
Air Blank	0.000	08:12
Subject Sample # 1	0.092	08:13
Breath Volume	4.046	08:13
Air Blank	0.000	08:13
Air Blank	0.000	08:15
Subject Sample # 2	0.091	08:15
Breath Volume	4.417	08:15
Air Blank	0.000	08:16
Control Test	0.077	08:16
Air Blank	0.000	08:17
Diagnostics Check	OK	08:17

Data Download Date/Time: 11/13/2025 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLAVERDE,ORLANDO,R
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: GALLO,J
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: BEDOYA

DATE/TIME OF TEST: 10/21/2025 20:51:00

Diagnostics Check	OK	20:48	
Air Blank	0.000	20:48	
Control Test	0.078	20:48	
Air Blank	0.000	20:49	
Subject Sample	# 1	20:49	Subject Test Refused
Breath Volume	0.000	20:49	
Air Blank	0.000	20:50	
Control Test	0.079	20:50	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:51	

Data Download Date/Time: 11/13/2025 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,WILFREDO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ORTEGA,WILFREDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: TINOCO

DATE/TIME OF TEST: 10/26/2025 07:33:03

Diagnostics Check	OK	07:30	
Air Blank	0.000	07:30	
Control Test	0.079	07:30	
Air Blank	0.000	07:31	
Subject Sample	# 1	07:31	Subject Test Refused
Breath Volume	0.000	07:31	
Air Blank	0.000	07:32	
Control Test	0.079	07:32	
Air Blank	0.000	07:32	
Diagnostics Check	OK	07:33	

Data Download Date/Time: 11/13/2025 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: GALINDO BERMUDEZ

DATE/TIME OF TEST: 10/26/2025 22:46:32

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:44	
Control Test	0.079	22:44	
Air Blank	0.000	22:44	
Subject Sample	# 1	22:45	Subject Test Refused
Breath Volume	0.000	22:45	
Air Blank	0.000	22:45	
Control Test	0.079	22:45	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 11/13/2025 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PARKER,TOURCHE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: CONSUEGRA

DATE/TIME OF TEST: 10/28/2025 00:49:07

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:39	
Control Test	0.080	00:40	
Air Blank	0.000	00:40	
Subject Sample	# 1 0.085	00:42	
Breath Volume	1.625	00:42	
Air Blank	0.000	00:43	
Air Blank	0.000	00:45	
Subject Sample	# 2	00:47	Subject Test Refused
Breath Volume	0.652	00:47	
Air Blank	0.000	00:48	
Control Test	0.077	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 11/13/2025 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: BUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/31/2025 20:34:06

Diagnostics Check	OK	20:31	
Air Blank	0.000	20:31	
Control Test	0.078	20:31	
Air Blank	0.000	20:32	
Subject Sample	# 1	20:32	Subject Test Refused
Breath Volume	0.000	20:32	
Air Blank	0.000	20:33	
Control Test	0.079	20:33	
Air Blank	0.000	20:34	
Diagnostics Check	OK	20:34	

Data Download Date/Time: 11/13/2025 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,HENRIQUE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: COOPER,HENRIQUE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: QUIROGA

DATE/TIME OF TEST: 10/01/2025 05:50:33

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.080	05:44
Air Blank	0.000	05:45
Subject Sample # 1	0.179	05:45
Breath Volume	2.136	05:45
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.187	05:48
Breath Volume	2.464	05:48
Air Blank	0.000	05:49
Control Test	0.078	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 10/19/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSTROV,VANESSA,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: GONZALEZ,ALEXANDER
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 10/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 10/28/2025 19:29:22

Diagnostics Check	OK	19:19	
Air Blank	0.000	19:20	
Control Test	0.079	19:20	
Air Blank	0.000	19:21	
Subject Sample # 1	0.414	19:21	
Breath Volume	1.390	19:21	
Air Blank	0.000	19:22	
Air Blank	0.000	19:23	
Subject Sample # 2	0.391	19:24	
Breath Volume	1.621	19:24	
Air Blank	0.000	19:25	
Air Blank	0.000	19:27	
Subject Sample # 3	0.398	19:27	
Breath Volume	1.679	19:27	
Air Blank	0.000	19:28	
Control Test	0.073	19:28	Control Outside Tolerance
Air Blank	0.000	19:29	

Data Download Date/Time: 11/30/2025 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,S,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAZIME,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:21

SUBJECT'S LAST NAME: LUKE

DATE/TIME OF TEST: 10/04/2025 07:49:28

Diagnostics Check	OK	07:42
Air Blank	0.000	07:42
Control Test	0.077	07:43
Air Blank	0.000	07:43
Subject Sample # 1	0.089	07:44
Breath Volume	1.589	07:44
Air Blank	0.000	07:45
Air Blank	0.000	07:47
Subject Sample # 2	0.091	07:47
Breath Volume	1.523	07:47
Air Blank	0.000	07:48
Control Test	0.078	07:48
Air Blank	0.000	07:49
Diagnostics Check	OK	07:49

Data Download Date/Time: 10/15/2025 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 07:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: THORNE

DATE/TIME OF TEST: 10/18/2025 08:11:01

Diagnostics Check	OK	08:05	
Air Blank	0.000	08:05	
Control Test	0.079	08:05	
Air Blank	0.000	08:06	
Subject Sample	# 1 0.038	08:07	
Breath Volume	1.425	08:07	
Air Blank	0.000	08:07	
Air Blank	0.000	08:09	
Subject Sample	# 2	08:09	Subject Test Refused
Breath Volume	0.000	08:09	
Air Blank	0.000	08:10	
Control Test	0.080	08:10	
Air Blank	0.000	08:10	
Diagnostics Check	OK	08:11	

Data Download Date/Time: 11/12/2025 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: FITZ

DATE/TIME OF TEST: 10/23/2025 16:36:49

Diagnostics Check	OK	16:30
Air Blank	0.000	16:31
Control Test	0.079	16:31
Air Blank	0.000	16:32
Subject Sample # 1	0.000	16:32
Breath Volume	1.953	16:32
Air Blank	0.000	16:32
Air Blank	0.000	16:34
Subject Sample # 2	0.000	16:35
Breath Volume	1.878	16:35
Air Blank	0.000	16:35
Control Test	0.079	16:36
Air Blank	0.000	16:36
Diagnostics Check	OK	16:36

Data Download Date/Time: 11/13/2025 23:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUSTMAN,CAYLA,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DUSTMAN,CAYLA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 10/02/2025 00:53:03

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.082	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.220	00:47
Breath Volume	1.890	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.208	00:51
Breath Volume	2.136	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:53

Data Download Date/Time: 10/21/2025 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUSTMAN,CAYLA,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DUSTMAN,CAYLA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: LESLIE

DATE/TIME OF TEST: 10/06/2025 21:04:46

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:49	
Control Test	0.083	20:50	
Air Blank	0.000	20:50	
Subject Sample	# 1	20:53	No Sample Provided
Breath Volume	0.000	20:53	
Air Blank	0.000	20:54	
Air Blank	0.000	20:56	
Subject Sample	# 2	0.262	20:59
Breath Volume	1.988	20:59	
Air Blank	0.000	20:59	
Air Blank	0.000	21:01	
Subject Sample	# 3	0.268	21:03
Breath Volume	1.437	21:03	
Air Blank	0.000	21:03	
Control Test	0.080	21:04	
Air Blank	0.000	21:04	
Diagnostics Check	OK	21:04	

Data Download Date/Time: 10/21/2025 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUSTMAN,CAYLA,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DUSTMAN,CAYLA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 10/12/2025 03:15:46

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.081	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.162	03:11
Breath Volume	2.863	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.157	03:14
Breath Volume	2.500	03:14
Air Blank	0.000	03:14
Control Test	0.079	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 10/21/2025 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEBRETON,QUINN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROBIN,GASTON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: REGAN

DATE/TIME OF TEST: 10/24/2025 19:33:38

Diagnostics Check	OK	19:30	
Air Blank	0.000	19:31	
Control Test	0.081	19:31	
Air Blank	0.000	19:32	
Subject Sample	# 1	19:32	Subject Test Refused
Breath Volume	0.000	19:32	
Air Blank	0.000	19:32	
Control Test	0.082	19:33	
Air Blank	0.000	19:33	
Diagnostics Check	OK	19:33	

Data Download Date/Time: 11/19/2025 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUSTMAN,CAYLA,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DUSTMAN,CAYLA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: PADRICK

DATE/TIME OF TEST: 10/25/2025 03:52:42

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.082	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.178	03:48
Breath Volume	1.925	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.179	03:51
Breath Volume	3.539	03:51
Air Blank	0.000	03:51
Control Test	0.083	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 11/19/2025 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUSTMAN,CAYLA,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WILLIAMS,ERIC
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 10/26/2025 05:13:31

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.082	05:08
Air Blank	0.000	05:08
Subject Sample # 1	0.197	05:08
Breath Volume	1.148	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.209	05:11
Breath Volume	2.078	05:11
Air Blank	0.000	05:12
Control Test	0.083	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 11/19/2025 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOUSSAINT,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PURSLEY,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: ZIELINKI

DATE/TIME OF TEST: 10/02/2025 20:20:22

Diagnostics Check	OK	20:13
Air Blank	0.000	20:13
Control Test	0.080	20:13
Air Blank	0.000	20:14
Subject Sample # 1	0.131	20:15
Breath Volume	1.882	20:15
Air Blank	0.000	20:16
Air Blank	0.000	20:17
Subject Sample # 2	0.134	20:18
Breath Volume	2.101	20:18
Air Blank	0.000	20:19
Control Test	0.081	20:19
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ TRIANA,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOTTLE,G
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: BARTOLO BOLANOS

DATE/TIME OF TEST: 10/03/2025 19:31:46

Diagnostics Check	OK	19:24
Air Blank	0.000	19:25
Control Test	0.080	19:25
Air Blank	0.000	19:26
Subject Sample # 1	0.228	19:26
Breath Volume	2.535	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.232	19:30
Breath Volume	1.699	19:30
Air Blank	0.000	19:30
Control Test	0.079	19:31
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUGG,MEGAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BYERS,N
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MULHERN

DATE/TIME OF TEST: 10/04/2025 04:09:33

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:55	
Control Test	0.081	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:59	No Sample Provided
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Air Blank	0.000	04:01	
Subject Sample	# 2	0.126	
Breath Volume	2.019	04:02	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample	# 3	0.114	Volume Not Met
Breath Volume	0.539	04:07	
Air Blank	0.000	04:08	
Control Test	0.080	04:08	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUGG,MEGAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BYERS,N
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MULHERN

DATE/TIME OF TEST: 10/04/2025 04:29:09

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.079	04:27	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:27	Subject Test Refused
Breath Volume	0.000	04:27	
Air Blank	0.000	04:28	
Control Test	0.080	04:28	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPISE,STEPHEN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VERREAULT,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/05/2025 02:56:35

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.151	02:52
Breath Volume	3.406	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.147	02:54
Breath Volume	3.117	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAGBY,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WISE,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MACHINRUENES

DATE/TIME OF TEST: 10/08/2025 03:51:08

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.081	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.211	03:46
Breath Volume	1.605	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:48
Subject Sample # 2	0.216	03:49
Breath Volume	1.757	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,T,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENNINI,L
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/10/2025 01:14:26

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.151	01:09
Breath Volume	1.937	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.153	01:12
Breath Volume	1.863	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL, ADAM, J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HAMPTON, M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: FORTINA

DATE/TIME OF TEST: 10/10/2025 01:41:16

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.081	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.157	01:36
Breath Volume	2.460	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.152	01:39
Breath Volume	2.832	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL, ADAM, J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TROY, H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: PICHARDO

DATE/TIME OF TEST: 10/10/2025 02:14:54

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.203	02:10
Breath Volume	2.730	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample # 2	0.200	02:13
Breath Volume	3.007	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONSTABLE,AVRY,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WEEKANDER,WEEKANDER
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: KANIA

DATE/TIME OF TEST: 10/10/2025 19:10:14

Diagnostics Check	OK	19:03
Air Blank	0.000	19:04
Control Test	0.081	19:04
Air Blank	0.000	19:04
Subject Sample # 1	0.029	19:05
Breath Volume	1.769	19:05
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.029	19:08
Breath Volume	1.890	19:08
Air Blank	0.000	19:09
Control Test	0.081	19:09
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDERE,ANTONIO,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HAMPTON,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/11/2025 04:14:09

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.080	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:10	No Sample Provided
Breath Volume	0.000	04:10	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample	# 2	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:13	
Control Test	0.080	04:13	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,JAMES,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CCSO,CCSO
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 10/11/2025 16:49:02

Diagnostics Check	OK	16:42
Air Blank	0.000	16:42
Control Test	0.081	16:43
Air Blank	0.000	16:43
Subject Sample # 1	0.153	16:44
Breath Volume	3.152	16:44
Air Blank	0.000	16:44
Air Blank	0.000	16:46
Subject Sample # 2	0.158	16:47
Breath Volume	2.535	16:47
Air Blank	0.000	16:48
Control Test	0.080	16:48
Air Blank	0.000	16:48
Diagnostics Check	OK	16:49

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOUSSAINT,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WESTRIVER,V
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: CONNOLLY

DATE/TIME OF TEST: 10/11/2025 19:34:51

Diagnostics Check	OK	19:24
Air Blank	0.000	19:24
Control Test	0.081	19:25
Air Blank	0.000	19:25
Subject Sample # 1	0.297	19:26
Breath Volume	1.855	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.264	19:29
Breath Volume	1.273	19:29
Air Blank	0.000	19:30
Air Blank	0.000	19:32
Subject Sample # 3	0.282	19:33
Breath Volume	2.050	19:33
Air Blank	0.000	19:33
Control Test	0.080	19:34
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACUNA,V,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PRUCHNIK,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: MARTINEZ HERNANDEZ

DATE/TIME OF TEST: 10/11/2025 22:11:12

Diagnostics Check	OK	22:04
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.280	22:06
Breath Volume	1.257	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.278	22:09
Breath Volume	1.312	22:09
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,R,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GARRETT,N
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: OQUENDO CEDENO

DATE/TIME OF TEST: 10/12/2025 01:31:32

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.155	01:26
Breath Volume	3.625	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.153	01:29
Breath Volume	3.402	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,R,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HOWELL,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: FERNANDEZ VELASQUEZ

DATE/TIME OF TEST: 10/12/2025 03:53:25

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.081	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Control Test	0.079	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAGBY,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SHADLEY,O
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: ALEJANDRO

DATE/TIME OF TEST: 10/13/2025 01:38:57

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.081	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.125	01:34
Breath Volume	2.277	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.126	01:37
Breath Volume	2.335	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTT,RYAN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROHRAFF,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/14/2025 23:44:49

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.081	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.076	23:40
Breath Volume	1.457	23:40
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.075	23:43
Breath Volume	1.589	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDERE,ANTONIO,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SOLIS,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: JUDY

DATE/TIME OF TEST: 10/16/2025 00:24:23

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.081	00:18	
Air Blank	0.000	00:18	
Subject Sample	# 1 0.228	00:19	
Breath Volume	2.214	00:19	
Air Blank	0.000	00:20	
Air Blank	0.000	00:21	
Subject Sample	# 2	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Control Test	0.080	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIETY,G,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TROY,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: COLIN

DATE/TIME OF TEST: 10/18/2025 02:28:29

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:13	
Control Test	0.082	02:13	
Air Blank	0.000	02:13	
Subject Sample	# 1 0.165	02:15	
Breath Volume	1.289	02:15	
Air Blank	0.000	02:16	
Air Blank	0.000	02:17	
Subject Sample	# 2 0.162	02:21	Volume Not Met
Breath Volume	0.066	02:21	
Air Blank	0.000	02:21	
Air Blank	0.000	02:23	
Subject Sample	# 3 0.157	02:26	Volume Not Met
Breath Volume	0.097	02:26	
Air Blank	0.000	02:27	
Control Test	0.078	02:27	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIETY,G,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TROY,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: COLIN

DATE/TIME OF TEST: 10/18/2025 02:44:50

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.079	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 1 0.153	02:37	Volume Not Met
Breath Volume	0.082	02:37	
Air Blank	0.000	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 2 0.139	02:43	Volume Not Met
Breath Volume	0.753	02:43	
Air Blank	0.000	02:43	
Control Test	0.077	02:44	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIETY,G,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TROY,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: COLIN

DATE/TIME OF TEST: 10/18/2025 02:53:27

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.078	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Control Test	0.079	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EWERT,ZACHARY,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENNINI,L
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ROPER

DATE/TIME OF TEST: 10/18/2025 05:12:37

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.081	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.000	05:08
Breath Volume	2.546	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.000	05:11
Breath Volume	2.144	05:11
Air Blank	0.000	05:11
Control Test	0.080	05:12
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWANSON,CONNOR,E
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: AHRENS,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 10/19/2025 00:34:05

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.080	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.201	00:29
Breath Volume	3.195	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:31
Subject Sample # 2	0.194	00:32
Breath Volume	1.808	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIXON,CONNOR,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: REIFF,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 011
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: GORMLEY

DATE/TIME OF TEST: 10/19/2025 02:06:11

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.080	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:05	
Control Test	0.080	02:05	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUGG,MEGAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PALMISANO,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: BLAISE

DATE/TIME OF TEST: 10/19/2025 04:43:48

Diagnostics Check	OK	04:42	
Air Blank		04:43	RFI Detect
Air Blank	0.000	04:43	

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUGG,MEGAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PALMISANO,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: BLAISE

DATE/TIME OF TEST: 10/19/2025 04:52:41

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.080	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.167	04:48
Breath Volume	1.597	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.155	04:51
Breath Volume	1.425	04:51
Air Blank	0.000	04:51
Control Test	0.081	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACUNA,V,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENNINI,L
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: JUAREZ LAZARO

DATE/TIME OF TEST: 10/23/2025 22:02:53

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.080	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.137	21:58
Breath Volume	3.664	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.137	22:01
Breath Volume	3.468	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONSTABLE,AVRY,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LAMBROS,S
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/24/2025 21:46:10

Diagnostics Check	OK	21:39
Air Blank	0.000	21:40
Control Test	0.081	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.174	21:41
Breath Volume	1.324	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.161	21:44
Breath Volume	1.648	21:44
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAGBY,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ESPIN,T
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: SHIELDS

DATE/TIME OF TEST: 10/24/2025 22:18:26

Diagnostics Check	OK	22:11
Air Blank	0.000	22:12
Control Test	0.080	22:12
Air Blank	0.000	22:13
Subject Sample # 1	0.083	22:13
Breath Volume	2.992	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.085	22:16
Breath Volume	2.523	22:16
Air Blank	0.000	22:17
Control Test	0.079	22:17
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 10/25/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYMAN,PATRICK,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TARR,LOGAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: CORRERA-BUILES

DATE/TIME OF TEST: 10/13/2025 01:50:05

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.162	01:45
Breath Volume	3.058	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.162	01:48
Breath Volume	1.996	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 10/26/2025 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BUONPANE,V
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:08

SUBJECT'S LAST NAME: TERMINI

DATE/TIME OF TEST: 10/24/2025 14:38:41

Diagnostics Check	OK	14:31
Air Blank	0.000	14:32
Control Test	0.079	14:32
Air Blank	0.000	14:33
Subject Sample # 1	0.000	14:34
Breath Volume	2.203	14:34
Air Blank	0.000	14:34
Air Blank	0.000	14:36
Subject Sample # 2	0.000	14:37
Breath Volume	1.621	14:37
Air Blank	0.000	14:37
Control Test	0.079	14:38
Air Blank	0.000	14:38
Diagnostics Check	OK	14:38

Data Download Date/Time: 10/26/2025 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERRE,KRENID,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MARZICOLA,GABRIELLE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: PAXTORCHAVEZ

DATE/TIME OF TEST: 10/25/2025 00:11:08

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.186	00:06
Breath Volume	1.417	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.179	00:09
Breath Volume	1.191	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 10/26/2025 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYMAN,PATRICK,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ALLEN,JUSTIN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: SEBASTIAN

DATE/TIME OF TEST: 10/25/2025 18:05:00

Diagnostics Check	OK	17:58
Air Blank	0.000	17:59
Control Test	0.079	17:59
Air Blank	0.000	18:00
Subject Sample # 1	0.174	18:00
Breath Volume	1.601	18:00
Air Blank	0.000	18:01
Air Blank	0.000	18:03
Subject Sample # 2	0.171	18:03
Breath Volume	3.000	18:03
Air Blank	0.000	18:04
Control Test	0.079	18:04
Air Blank	0.000	18:04
Diagnostics Check	OK	18:05

Data Download Date/Time: 10/26/2025 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERRE,KRENID,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ESPIN,TINA
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/26/2025 04:28:20

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.114	04:23
Breath Volume	1.492	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.117	04:26
Breath Volume	2.128	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 10/26/2025 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN, CLARENCE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER, CPL
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: GOSLER

DATE/TIME OF TEST: 10/23/2025 18:43:19

Diagnostics Check	OK	18:36
Air Blank	0.000	18:37
Control Test	0.080	18:37
Air Blank	0.000	18:37
Subject Sample # 1	0.148	18:38
Breath Volume	3.035	18:38
Air Blank	0.000	18:39
Air Blank	0.000	18:41
Subject Sample # 2	0.150	18:41
Breath Volume	2.425	18:41
Air Blank	0.000	18:42
Control Test	0.080	18:42
Air Blank	0.000	18:43
Diagnostics Check	OK	18:43

Data Download Date/Time: 11/26/2025 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWTHORNE, OLIVIA, J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GIRARD, J
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 10/26/2025 01:14:27

Diagnostics Check	Fail	01:13	DSP Test
Air Blank	0.000	01:14	

Data Download Date/Time: 11/26/2025 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHANDLER,ADMIN,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ADMIN,ADMIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: ADMIN
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 10/27/2025 21:23:27

Diagnostics Check	Fail	21:22	DSP Test
Air Blank	0.000	21:23	

Data Download Date/Time: 11/26/2025 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADMIN,ADMIN,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ADMIN,ADMIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: ADMIN
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 10/27/2025 21:30:44

Diagnostics Check	OK	21:25
Air Blank	0.000	21:25
Control Test	0.080	21:25
Air Blank	0.000	21:26
Subject Sample # 1	0.000	21:26
Breath Volume	1.828	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.000	21:29
Breath Volume	1.406	21:29
Air Blank	0.000	21:29
Control Test	0.079	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 11/26/2025 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADMIN,ADMIN,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ADMIN,ADMIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: ADMIN
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 10/27/2025 21:37:45

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.079	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.000	21:33
Breath Volume	1.187	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.000	21:36
Breath Volume	1.519	21:36
Air Blank	0.000	21:36
Control Test	0.079	21:37
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 11/26/2025 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADMIN,ADMIN,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ADMIN,ADMIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: ADMIN
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 10/27/2025 21:44:52

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.079	21:39
Air Blank	0.000	21:40
Subject Sample # 1	0.000	21:40
Breath Volume	1.152	21:40
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.000	21:43
Breath Volume	1.437	21:43
Air Blank	0.000	21:43
Control Test	0.080	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 11/26/2025 20:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: HERNANDEZ,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 10/04/2025 01:21:56

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.078	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.114	01:17
Breath Volume	1.980	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.118	01:20
Breath Volume	1.933	01:20
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 10/31/2025 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCALERA,MARIANA,Y
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: PILZ,ALEXANDER
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 10/12/2025 01:05:44

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:54	
Control Test	0.078	00:54	
Air Blank	0.000	00:55	
Subject Sample	# 1 0.190	00:58	Volume Not Met
Breath Volume	0.140	00:58	
Air Blank	0.000	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 2 0.157	01:04	Volume Not Met
Breath Volume	0.042	01:04	
Air Blank	0.000	01:04	
Control Test	0.078	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 10/31/2025 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIVER,ROBERT,D
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: NASH,CHRISTOPHER
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: ADAIR

DATE/TIME OF TEST: 10/15/2025 18:57:40

Diagnostics Check	OK	18:51
Air Blank	0.000	18:51
Control Test	0.078	18:52
Air Blank	0.000	18:52
Subject Sample # 1	0.270	18:53
Breath Volume	3.136	18:53
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.267	18:56
Breath Volume	2.910	18:56
Air Blank	0.000	18:56
Control Test	0.078	18:57
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 10/27/2025 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUCKHALT,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Liberty County SO
ARRESTING OFFICER: HADDOCK,MAVERICK
ARRESTING OFFICER AGENCY: Liberty County SO
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: URBINA

DATE/TIME OF TEST: 10/19/2025 21:30:47

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.079	21:26	
Air Blank	0.000	21:26	
Subject Sample # 1	0.168	21:27	
Breath Volume	1.355	21:27	
Air Blank	0.000	21:27	
Air Blank		21:29	RFI Detect
Air Blank	0.000	21:29	
Control Test	0.079	21:30	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

Data Download Date/Time: 10/23/2025 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUCKHALT,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Liberty County SO
ARRESTING OFFICER: HADDOCK,MAVERICK
ARRESTING OFFICER AGENCY: Liberty County SO
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: URBINA

DATE/TIME OF TEST: 10/19/2025 21:40:07

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.079	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.181	21:35
Breath Volume	1.871	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:37
Subject Sample # 2	0.173	21:38
Breath Volume	2.054	21:38
Air Blank	0.000	21:39
Control Test	0.075	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 10/23/2025 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUCHOCKI,RAYMOND,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LONG,KATHY
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: VELASQUEZ LOPEZ

DATE/TIME OF TEST: 10/11/2025 21:07:23

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.080	21:01
Air Blank	0.000	21:01
Subject Sample # 1	0.177	21:02
Breath Volume	1.179	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.184	21:05
Breath Volume	1.425	21:05
Air Blank	0.000	21:06
Control Test	0.080	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 10/30/2025 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELPS,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: PHELPS,WILLIAM
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: IVY

DATE/TIME OF TEST: 10/01/2025 01:31:29

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.082	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.230	01:27
Breath Volume	1.949	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.240	01:29
Breath Volume	1.859	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 10/22/2025 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWEENEY,NELSON,I
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HEDGES,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:33

SUBJECT'S LAST NAME: SWEENEY

DATE/TIME OF TEST: 10/09/2025 14:14:43

Diagnostics Check	OK	14:08
Air Blank	0.000	14:08
Control Test	0.082	14:09
Air Blank	0.000	14:09
Subject Sample # 1	0.000	14:10
Breath Volume	1.207	14:10
Air Blank	0.000	14:10
Air Blank	0.000	14:12
Subject Sample # 2	0.000	14:13
Breath Volume	1.980	14:13
Air Blank	0.000	14:13
Control Test	0.082	14:14
Air Blank	0.000	14:14
Diagnostics Check	OK	14:14

Data Download Date/Time: 10/22/2025 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: KNOX

DATE/TIME OF TEST: 10/11/2025 01:13:42

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.082	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.112	01:09
Breath Volume	2.582	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.109	01:12
Breath Volume	2.050	01:12
Air Blank	0.000	01:12
Control Test	0.081	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 10/22/2025 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELPS, WILLIAM, D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: PHELPS, WILLIAM
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 10/16/2025 01:21:51

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.143	01:17
Breath Volume	1.343	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.129	01:20
Breath Volume	1.550	01:20
Air Blank	0.000	01:20
Control Test	0.082	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 10/22/2025 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOULTON, MAURICE, W
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MOULTON, MAURICE
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 10/16/2025 02:49:13

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.079	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.180	02:44
Breath Volume	2.167	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.182	02:47
Breath Volume	1.671	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 10/22/2025 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: SMITH,ALEXANDER
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 10/18/2025 03:08:12

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.151	03:03
Breath Volume	2.613	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:05
Subject Sample # 2	0.148	03:06
Breath Volume	3.328	03:06
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 10/22/2025 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEX,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: ARNOLD,MAURICE
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 10/19/2025 01:09:20

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.082	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.000	01:04
Breath Volume	3.726	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.000	01:07
Breath Volume	3.757	01:07
Air Blank	0.000	01:08
Control Test	0.082	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 10/22/2025 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: HAILSTOCK

DATE/TIME OF TEST: 10/19/2025 01:26:48

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.082	01:20
Air Blank	0.000	01:20
Subject Sample # 1	0.110	01:21
Breath Volume	3.324	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.109	01:25
Breath Volume	2.351	01:25
Air Blank	0.000	01:25
Control Test	0.082	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 10/22/2025 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,AARON,C
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CLARK,AARON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 10/19/2025 04:17:21

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.082	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.108	04:12
Breath Volume	3.535	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.109	04:15
Breath Volume	1.128	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 10/22/2025 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD, N
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 10/24/2025 19:16:50

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.082	19:10
Air Blank	0.000	19:11
Subject Sample # 1	0.150	19:12
Breath Volume	2.589	19:12
Air Blank	0.000	19:12
Air Blank	0.000	19:14
Subject Sample # 2	0.168	19:15
Breath Volume	2.687	19:15
Air Blank	0.000	19:15
Control Test	0.082	19:16
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

Data Download Date/Time: 11/10/2025 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON, COLIN, P
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: NELSON, COLIN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: COENS

DATE/TIME OF TEST: 10/25/2025 04:44:45

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.082	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.236	04:35
Breath Volume	1.214	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.211	04:38
Breath Volume	1.210	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 3	0.209	04:43
Breath Volume	1.398	04:43
Air Blank	0.000	04:43
Control Test	0.081	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 11/10/2025 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELPS, WILLIAM, D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: PHELPS, WILLIAM
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/25/2025 21:57:27

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.082	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.119	21:53
Breath Volume	2.015	21:53
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.115	21:55
Breath Volume	2.445	21:55
Air Blank	0.000	21:56
Control Test	0.081	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 11/10/2025 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTON,JEREMY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BLOUNT,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: MCDUFFIE

DATE/TIME OF TEST: 10/27/2025 20:44:02

Diagnostics Check	OK	20:37
Air Blank	0.000	20:38
Control Test	0.081	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.286	20:39
Breath Volume	1.726	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.271	20:42
Breath Volume	1.773	20:42
Air Blank	0.000	20:43
Control Test	0.081	20:43
Air Blank	0.000	20:43
Diagnostics Check	OK	20:44

Data Download Date/Time: 11/10/2025 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,R,R
BREATH TEST OPERATOR AGENCY: Midway PD
ARRESTING OFFICER: CRUZ,R
ARRESTING OFFICER AGENCY: Midway PD
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: LIBRADO

DATE/TIME OF TEST: 10/25/2025 02:08:22

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.174	02:03
Breath Volume	1.457	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.175	02:06
Breath Volume	1.960	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 11/18/2025 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COPELAND,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Quincy PD
ARRESTING OFFICER: GUTSHALL,CADEN
ARRESTING OFFICER AGENCY: Quincy PD
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MARTIN-AGUILAR

DATE/TIME OF TEST: 10/25/2025 04:27:10

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.089	04:22
Breath Volume	1.875	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.085	04:25
Breath Volume	1.449	04:25
Air Blank	0.000	04:26
Control Test	0.082	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 11/18/2025 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASS,JEFFREY,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GLASS,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: HOUSNI

DATE/TIME OF TEST: 10/12/2025 22:10:47

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:06	
Control Test	0.081	22:06	
Air Blank	0.000	22:07	
Subject Sample	# 1	22:09	Interferent Detect
Breath Volume	1.304	22:09	
Air Blank		22:10	Purge Fail
Air Blank		22:10	Purge Fail

Data Download Date/Time: 10/23/2025 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,ADRIAN,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MYERS,ADRIAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: FAIR

DATE/TIME OF TEST: 10/16/2025 10:57:23

Diagnostics Check	OK	10:56	
Air Blank	0.000	10:56	
		10:56	RFI Detect
Air Blank	0.000	10:57	

Data Download Date/Time: 10/23/2025 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,ADRIAN,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MYERS,ADRIAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: FAIR

DATE/TIME OF TEST: 10/16/2025 11:06:47

Diagnostics Check	OK	11:00
Air Blank	0.000	11:00
Control Test	0.081	11:01
Air Blank	0.000	11:01
Subject Sample # 1	0.000	11:02
Breath Volume	2.570	11:02
Air Blank	0.000	11:02
Air Blank	0.000	11:04
Subject Sample # 2	0.000	11:05
Breath Volume	2.734	11:05
Air Blank	0.000	11:05
Control Test	0.082	11:06
Air Blank	0.000	11:06
Diagnostics Check	OK	11:06

Data Download Date/Time: 10/23/2025 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMOS,KARLOS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: PINZONROJAS

DATE/TIME OF TEST: 10/01/2025 04:24:08

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.082	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.155	04:19
Breath Volume	3.058	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.142	04:22
Breath Volume	1.316	04:22
Air Blank	0.000	04:23
Control Test	0.082	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 10/16/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALDES,BRYAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/03/2025 02:57:32

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.083	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.261	02:52
Breath Volume	3.089	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.248	02:56
Breath Volume	3.847	02:56
Air Blank	0.000	02:56
Control Test	0.082	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 10/16/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BECKER,THOMAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BRENS

DATE/TIME OF TEST: 10/04/2025 01:42:06

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.082	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.173	01:37
Breath Volume	2.738	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.170	01:40
Breath Volume	2.910	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 10/16/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THOMAS,BENJAMIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: NECHEPURENKO

DATE/TIME OF TEST: 10/09/2025 02:03:13

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:53	
Control Test	0.082	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1 0.176	01:57	Volume Not Met
Breath Volume	0.671	01:57	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 2	02:01	Subject Test Refused
Breath Volume	0.539	02:01	
Air Blank	0.000	02:02	
Control Test	0.081	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 10/16/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA APONTE,ADRIEL
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: FREID

DATE/TIME OF TEST: 10/09/2025 06:01:20

Diagnostics Check	OK	05:55
Air Blank	0.000	05:56
Control Test	0.082	05:56
Air Blank	0.000	05:56
Subject Sample # 1	0.000	05:57
Breath Volume	2.917	05:57
Air Blank	0.000	05:57
Air Blank	0.000	05:59
Subject Sample # 2	0.000	05:59
Breath Volume	3.281	05:59
Air Blank	0.000	06:00
Control Test	0.083	06:00
Air Blank	0.000	06:01
Diagnostics Check	OK	06:01

Data Download Date/Time: 10/16/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEREZ,ADRIAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: VALDEZCABRERA

DATE/TIME OF TEST: 10/10/2025 04:03:56

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.082	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.159	03:59
Breath Volume	3.812	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.164	04:02
Breath Volume	4.101	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 10/16/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PASSARELLA,ANTHONY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 10/11/2025 02:35:39

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.082	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.192	02:31
Breath Volume	1.171	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.188	02:34
Breath Volume	1.886	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 10/16/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BARKIN

DATE/TIME OF TEST: 10/12/2025 03:43:01

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.083	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.000	03:38
Breath Volume	2.914	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.000	03:41
Breath Volume	3.789	03:41
Air Blank	0.000	03:42
Control Test	0.083	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:43

Data Download Date/Time: 10/16/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ-
VARGAS,HUMBERTO
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GUZMANROSARIO

DATE/TIME OF TEST: 10/12/2025 23:51:55

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.082	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.134	23:47
Breath Volume	1.523	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.134	23:50
Breath Volume	1.261	23:50
Air Blank	0.000	23:51
Control Test	0.082	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 10/16/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOEL,KENSON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: BLANC

DATE/TIME OF TEST: 10/25/2025 22:45:38

Diagnostics Check	OK	22:39
Air Blank	0.000	22:39
Control Test	0.080	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.056	22:41
Breath Volume	2.742	22:41
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.055	22:44
Breath Volume	2.437	22:44
Air Blank	0.000	22:44
Control Test	0.079	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: FLORESTAL

DATE/TIME OF TEST: 10/26/2025 02:10:55

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.080	01:57	
Air Blank	0.000	01:57	
Subject Sample	# 1 0.116	02:00	Volume Not Met
Breath Volume	0.707	02:00	
Air Blank	0.000	02:01	
Air Blank	0.000	02:03	
Subject Sample	# 2 0.119	02:03	
Breath Volume	1.589	02:03	
Air Blank	0.000	02:04	
Air Blank	0.000	02:06	
Subject Sample	# 3 0.114	02:09	Volume Not Met
Breath Volume	0.691	02:09	
Air Blank	0.000	02:10	
Control Test	0.080	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: FLORESTAL

DATE/TIME OF TEST: 10/26/2025 02:19:56

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.115	02:14
Breath Volume	1.234	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.115	02:18
Breath Volume	2.066	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: HAMAD

DATE/TIME OF TEST: 10/26/2025 18:01:31

Diagnostics Check	OK	17:55
Air Blank	0.000	17:55
Control Test	0.080	17:56
Air Blank	0.000	17:56
Subject Sample # 1	0.000	17:57
Breath Volume	1.867	17:57
Air Blank	0.000	17:57
Air Blank	0.000	17:59
Subject Sample # 2	0.000	18:00
Breath Volume	1.394	18:00
Air Blank	0.000	18:00
Control Test	0.080	18:00
Air Blank	0.000	18:01
Diagnostics Check	OK	18:01

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIAZREYES,GABRIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: RIOSRIBERA

DATE/TIME OF TEST: 10/27/2025 00:26:01

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.126	00:21
Breath Volume	2.269	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.122	00:24
Breath Volume	1.996	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:26

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALICEAGALLOZA,MARK
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 10/27/2025 02:00:14

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.195	01:55
Breath Volume	2.472	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.178	01:58
Breath Volume	1.144	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALDES,BRYAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 10/27/2025 03:26:48

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.207	03:22
Breath Volume	1.882	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.208	03:25
Breath Volume	1.511	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MANLEY

DATE/TIME OF TEST: 10/27/2025 04:41:51

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.155	04:37
Breath Volume	1.183	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.154	04:40
Breath Volume	1.914	04:40
Air Blank	0.000	04:40
Control Test	0.079	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERAAPONTE,ADRIEL
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: GREGG

DATE/TIME OF TEST: 10/28/2025 01:58:29

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.201	01:54
Breath Volume	2.082	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.195	01:57
Breath Volume	2.339	01:57
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAURENT,WILFRID
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:43

SUBJECT'S LAST NAME: FEBRES

DATE/TIME OF TEST: 10/28/2025 15:17:15

Diagnostics Check	OK	15:11
Air Blank	0.000	15:11
Control Test	0.079	15:12
Air Blank	0.000	15:12
Subject Sample # 1	0.312	15:13
Breath Volume	1.863	15:13
Air Blank	0.000	15:13
Air Blank	0.000	15:15
Subject Sample # 2	0.301	15:15
Breath Volume	1.390	15:15
Air Blank	0.000	15:16
Control Test	0.078	15:16
Air Blank	0.000	15:17
Diagnostics Check	OK	15:17

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALMANZAR,JANCARLOS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: AREVALO

DATE/TIME OF TEST: 10/29/2025 04:20:07

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:17	
Control Test	0.080	04:17	
Air Blank	0.000	04:18	
Subject Sample	# 1	04:18	Subject Test Refused
Breath Volume	0.000	04:18	
Air Blank	0.000	04:19	
Control Test	0.080	04:19	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,ANTONIO
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: BARKO

DATE/TIME OF TEST: 10/29/2025 21:12:44

Diagnostics Check	OK	21:06
Air Blank	0.000	21:06
Control Test	0.079	21:06
Air Blank	0.000	21:07
Subject Sample # 1	0.194	21:08
Breath Volume	1.179	21:08
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.203	21:11
Breath Volume	1.347	21:11
Air Blank	0.000	21:11
Control Test	0.079	21:12
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,CRAIG
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: MAC GREGOR

DATE/TIME OF TEST: 10/30/2025 05:19:28

Diagnostics Check	OK	05:10
Air Blank	0.000	05:10
Control Test	0.078	05:11
Air Blank	0.000	05:11
Subject Sample # 1	0.185	05:12
Breath Volume	4.070	05:12
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.148	05:14
Breath Volume	1.664	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 3	0.169	05:18
Breath Volume	1.554	05:18
Air Blank	0.000	05:18
Control Test	0.079	05:18
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALDES,BRYAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: KONG

DATE/TIME OF TEST: 10/30/2025 21:28:01

Diagnostics Check	OK	21:16	
Air Blank	0.000	21:17	
Control Test	0.079	21:17	
Air Blank	0.000	21:17	
Subject Sample	# 1	21:21	No Sample Provided
Breath Volume	0.000	21:21	
Air Blank	0.000	21:21	
Air Blank	0.000	21:23	
Subject Sample	# 2	21:26	No Sample Provided
Breath Volume	0.000	21:26	
Air Blank	0.000	21:27	
Control Test	0.080	21:27	
Air Blank	0.000	21:27	
Diagnostics Check	OK	21:28	

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALDES,BRYAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: KONG

DATE/TIME OF TEST: 10/30/2025 22:03:44

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:51	
Control Test	0.079	21:51	
Air Blank	0.000	21:52	
Subject Sample	# 1 0.183	21:55	Volume Not Met
Breath Volume	0.171	21:55	
Air Blank	0.000	21:55	
Air Blank	0.000	21:57	
Subject Sample	# 2 0.202	21:59	No .020 agreement
Breath Volume	1.351	21:59	
Air Blank	0.000	21:59	
Air Blank	0.000	22:01	
Subject Sample	# 3 0.224	22:02	No .020 agreement
Breath Volume	1.542	22:02	
Air Blank	0.000	22:02	
Control Test	0.079	22:03	
Air Blank	0.000	22:03	
Diagnostics Check	OK	22:03	

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALDES,BRYAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: KONG

DATE/TIME OF TEST: 10/30/2025 22:12:27

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.080	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.209	22:07
Breath Volume	1.156	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.222	22:10
Breath Volume	1.777	22:10
Air Blank	0.000	22:11
Control Test	0.079	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNSON,DEREK
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 10/31/2025 07:53:53

Diagnostics Check	OK	07:48
Air Blank	0.000	07:48
Control Test	0.080	07:48
Air Blank	0.000	07:49
Subject Sample # 1	0.000	07:49
Breath Volume	1.640	07:49
Air Blank	0.000	07:50
Air Blank	0.000	07:52
Subject Sample # 2	0.000	07:52
Breath Volume	1.460	07:52
Air Blank	0.000	07:52
Control Test	0.081	07:53
Air Blank	0.000	07:53
Diagnostics Check	OK	07:53

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMOS,KARLOS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/31/2025 11:18:52

Diagnostics Check	OK	11:13
Air Blank	0.000	11:13
Control Test	0.081	11:13
Air Blank	0.000	11:14
Subject Sample # 1	0.000	11:14
Breath Volume	2.910	11:14
Air Blank	0.000	11:15
Air Blank	0.000	11:17
Subject Sample # 2	0.000	11:17
Breath Volume	3.441	11:17
Air Blank	0.000	11:17
Control Test	0.080	11:18
Air Blank	0.000	11:18
Diagnostics Check	OK	11:18

Data Download Date/Time: 11/11/2025 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JEFFREY,W
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS,TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 10/02/2025 04:12:15

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.079	04:00	
Air Blank	0.000	04:01	
Subject Sample	# 1 0.144	04:04	Volume Not Met
Breath Volume	1.066	04:04	
Air Blank	0.000	04:05	
Air Blank	0.000	04:06	
Subject Sample	# 2 0.160	04:07	
Breath Volume	1.300	04:07	
Air Blank	0.000	04:07	
Air Blank	0.000	04:09	
Subject Sample	# 3 0.158	04:10	
Breath Volume	1.113	04:10	
Air Blank	0.000	04:11	
Control Test	0.078	04:11	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 10/31/2025 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONNEMANN, MARK, W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CLEMENTS, T
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: SATTERFIELD

DATE/TIME OF TEST: 10/03/2025 03:20:02

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.088	03:15
Breath Volume	2.449	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.085	03:18
Breath Volume	2.683	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:20

Data Download Date/Time: 10/31/2025 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONS,DEPUTY,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GERARDI,DEPUTY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 10/04/2025 04:18:31

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.078	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.093	04:13
Breath Volume	2.039	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.100	04:16
Breath Volume	3.324	04:16
Air Blank	0.000	04:17
Control Test	0.078	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 10/31/2025 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONS,DEPUTY,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GERARDI,DEPUTY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 10/06/2025 03:42:45

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.140	03:37
Breath Volume	1.882	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.131	03:41
Breath Volume	1.703	03:41
Air Blank	0.000	03:41
Control Test	0.077	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 10/31/2025 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARABALLO,D,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GERARDI,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: WEISS

DATE/TIME OF TEST: 10/10/2025 01:32:39

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.000	01:28
Breath Volume	3.496	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.000	01:31
Breath Volume	3.660	01:31
Air Blank	0.000	01:31
Control Test	0.077	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 10/31/2025 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONS,DEPUTY,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CLEMENTS,T
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: BOYINGTON

DATE/TIME OF TEST: 10/10/2025 04:48:30

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.079	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.073	04:43
Breath Volume	1.207	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.076	04:46
Breath Volume	1.171	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 10/31/2025 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JEFFREY,W
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS,TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/11/2025 01:00:15

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.121	00:55
Breath Volume	2.875	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.112	00:58
Breath Volume	2.886	00:58
Air Blank	0.000	00:59
Control Test	0.078	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 10/31/2025 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JEFFREY,W
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: MCCONNELL,JEFFREY
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: AHRENS

DATE/TIME OF TEST: 10/11/2025 02:37:43

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.259	02:33
Breath Volume	2.589	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.258	02:36
Breath Volume	2.539	02:36
Air Blank	0.000	02:36
Control Test	0.077	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 10/31/2025 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JEFFREY,W
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS,TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: WOODARD

DATE/TIME OF TEST: 10/11/2025 05:09:16

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.078	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.098	05:04
Breath Volume	2.042	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.096	05:07
Breath Volume	2.265	05:07
Air Blank	0.000	05:08
Control Test	0.079	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 10/31/2025 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONNEMANN,M,W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CLEMENTS,T
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 10/12/2025 04:24:35

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.077	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.094	04:19
Breath Volume	1.148	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.091	04:23
Breath Volume	1.351	04:23
Air Blank	0.000	04:23
Control Test	0.076	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 10/31/2025 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CLEMEENTS,T
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 10/24/2025 01:21:59

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.079	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.046	01:17
Breath Volume	1.390	01:17
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.058	01:20
Breath Volume	1.359	01:20
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 10/31/2025 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASPER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CLEMENTS,T
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: BATTON

DATE/TIME OF TEST: 10/24/2025 21:57:36

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.079	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.129	21:52
Breath Volume	3.652	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.130	21:56
Breath Volume	4.109	21:56
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 10/31/2025 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JEFFREY,W
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS,TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: ARCIDIACONO

DATE/TIME OF TEST: 10/25/2025 01:26:54

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.079	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.170	01:22
Breath Volume	2.699	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.159	01:25
Breath Volume	2.445	01:25
Air Blank	0.000	01:25
Control Test	0.078	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 10/31/2025 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JEFFREY,W
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: MCCONNELL,JEFFREY
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 10/25/2025 04:07:43

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.079	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.000	04:03
Breath Volume	2.890	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.000	04:06
Breath Volume	2.671	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 10/31/2025 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/28/2025 02:15:43

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.017	02:10
Breath Volume	1.757	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.017	02:14
Breath Volume	2.125	02:14
Air Blank	0.000	02:14
Control Test	0.078	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 10/31/2025 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 09/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CASILLAS

DATE/TIME OF TEST: 10/02/2025 01:36:53

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.190	01:32
Breath Volume	3.058	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.181	01:35
Breath Volume	2.253	01:35
Air Blank	0.000	01:36
Control Test	0.076	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 10/22/2025 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 09/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: RENTERIA

DATE/TIME OF TEST: 10/11/2025 03:51:02

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.077	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.100	03:46
Breath Volume	2.500	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.101	03:49
Breath Volume	2.660	03:49
Air Blank	0.000	03:50
Control Test	0.078	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:51

Data Download Date/Time: 10/22/2025 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 09/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 10/12/2025 02:49:44

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.077	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.092	02:45
Breath Volume	2.070	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.098	02:48
Breath Volume	3.085	02:48
Air Blank	0.000	02:48
Control Test	0.078	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 10/22/2025 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 09/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 10/13/2025 02:02:17

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.079	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.153	01:56
Breath Volume	2.558	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.166	02:00
Breath Volume	2.785	02:00
Air Blank	0.000	02:01
Control Test	0.078	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 10/22/2025 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WERDA,JACOB,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BARNES,NOAH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 09/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 10/22/2025 01:38:35

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:28	
Control Test	0.077	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 1 0.170	01:30	
Breath Volume	1.160	01:30	
Air Blank	0.000	01:30	
Air Blank	0.000	01:32	
Subject Sample	# 2 0.144	01:34	
Breath Volume	1.519	01:34	
Air Blank	0.000	01:35	
Air Blank	0.000	01:37	
Subject Sample	# 3	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.076	01:38	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 10/22/2025 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 10/26/2025 03:13:15

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.077	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.162	03:08
Breath Volume	3.621	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.158	03:11
Breath Volume	2.902	03:11
Air Blank	0.000	03:12
Control Test	0.076	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 11/20/2025 16:45