Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPA,KYLE,T
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: KRUPA,KYLE
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: PALMIERI

DATE/TIME OF TEST: 01/21/2022 19:18:37

Diagnostics Check		OK	19:12
Air Blank		0.000	19:12
Control Test		0.077	19:13
Air Blank		0.000	19:13
Subject Sample	# 1	0.353	19:13
Breath Volume		1.835	19:13
Air Blank		0.000	19:14
Air Blank		0.000	19:16
Subject Sample	# 2	0.360	19:17
Breath Volume		1.820	19:17
Air Blank		0.000	19:17
Control Test		0.078	19:18
Air Blank		0.000	19:18
Diagnostics Check		OK	19:18

Data Download Date/Time: 02/02/2022 22:08

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DIMELLA,PABLO
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MACINTOSH

DATE/TIME OF TEST: 01/22/2022 03:00:38

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.148	02:56
Breath Volume		4.265	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.142	02:59
Breath Volume		4.226	02:59
Air Blank		0.000	02:59
Control Test		0.078	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 02/02/2022 22:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SERRANO,JIMMY
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: BONNAIRE

DATE/TIME OF TEST: 01/22/2022 04:53:15

Diagnostics Check		OK	04:45
Air Blank		0.000	04:45
Control Test		0.080	04:45
Air Blank		0.000	04:46
Subject Sample	# 1	0.142	04:48
Breath Volume		2.546	04:48
Air Blank		0.000	04:49
Air Blank		0.000	04:51
Subject Sample	# 2	0.142	04:51
Breath Volume		1.378	04:51
Air Blank		0.000	04:52
Control Test		0.080	04:52
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 02/02/2022 22:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KAISER,DAVID
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: HERNKE

DATE/TIME OF TEST: 01/26/2022 04:06:44

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.079	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.254	04:01
Breath Volume		2.386	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.256	04:05
Breath Volume		2.562	04:05
Air Blank		0.000	04:05
Control Test		0.079	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 02/02/2022 22:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CHAMBERS,AARON
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: MCNELIS

DATE/TIME OF TEST: 01/27/2022 03:30:58

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.079	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.216	03:26
Breath Volume		2.679	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.203	03:29
Breath Volume		2.164	03:29
Air Blank		0.000	03:30
Control Test		0.080	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 02/02/2022 22:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,ERIC,G
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: FERRELL,RYAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: DOMINGUEZ JARAMILLO

DATE/TIME OF TEST: 01/02/2022 20:45:10

Diagnostics Check		OK	20:36
Air Blank		0.000	20:37
Control Test		0.080	20:37
Air Blank		0.000	20:38
Subject Sample	# 1	0.151	20:40
Breath Volume		1.738	20:40
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.152	20:43
Breath Volume		1.417	20:43
Air Blank		0.000	20:44
Control Test		0.077	20:44
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 01/16/2022 13:56

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,ERIC,G
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: CHANEY,TONYA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SCHMITZ

DATE/TIME OF TEST: 01/05/2022 03:44:17

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.187	03:39
Breath Volume		2.101	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.187	03:42
Breath Volume		2.726	03:42
Air Blank		0.000	03:43
Control Test		0.079	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 01/16/2022 13:56

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,ERIC,G
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: KINSEY,RONALD
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: SHOUP

DATE/TIME OF TEST: 01/15/2022 01:30:27

Diagnostics Check		OK	01:28	
Air Blank		0.000	01:28	
Control Test		0.080	01:28	
Air Blank		0.000	01:29	
Subject Sample	# 1		01:29	RFI Detect
Breath Volume		0.000	01:29	
Air Blank			01:29	RFI Detect
Air Blank		0.000	01:30	

Data Download Date/Time: 01/16/2022 13:56

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,ERIC,G
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: KINSEY,RONALD
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: SHOUP

DATE/TIME OF TEST: 01/15/2022 01:49:12

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.224	01:44
Breath Volume		2.363	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.223	01:47
Breath Volume		2.464	01:47
Air Blank		0.000	01:48
Control Test		0.079	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 01/16/2022 13:56

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,ERIC,G
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: SEVERNS,BRIAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 01/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: MICKLER

DATE/TIME OF TEST: 01/21/2022 19:58:53

Diagnostics Check		OK	19:50
Air Blank		0.000	19:50
Control Test		0.080	19:50
Air Blank		0.000	19:51
Subject Sample	# 1	0.170	19:53
Breath Volume		1.855	19:53
Air Blank		0.000	19:54
Air Blank		0.000	19:55
Subject Sample	# 2	0.167	19:57
Breath Volume		2.402	19:57
Air Blank		0.000	19:57
Control Test		0.078	19:58
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 02/13/2022 13:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,ERIC,G
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: GANNON,TIM
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 01/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: BEESON

DATE/TIME OF TEST: 01/23/2022 23:19:25

Diagnostics Check		OK	23:09
Air Blank		0.000	23:10
Control Test		0.080	23:10
Air Blank		0.000	23:10
Subject Sample	# 1	0.170	23:11
Breath Volume		3.363	23:11
Air Blank		0.000	23:12
Air Blank		0.000	23:13
Subject Sample	# 2	0.147	23:14
Breath Volume		1.113	23:14
Air Blank		0.000	23:15
Air Blank		0.000	23:16
Subject Sample	# 3	0.171	23:17
Breath Volume		3.042	23:17
Air Blank		0.000	23:18
Control Test		0.081	23:18
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 02/13/2022 13:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: ANKENBAUER

DATE/TIME OF TEST: 01/06/2022 22:02:45

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.081	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.000	21:58
Breath Volume		1.347	21:58
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.000	22:01
Breath Volume		2.062	22:01
Air Blank		0.000	22:01
Control Test		0.081	22:02
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 01/07/2022 18:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: ELIZALDE HUITRON

DATE/TIME OF TEST: 01/07/2022 01:12:05

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.081	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.116	01:07
Breath Volume		1.187	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.115	01:10
Breath Volume		2.332	01:10
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:12

Data Download Date/Time: 01/07/2022 18:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 01/07/2022 04:34:40

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.081	04:29
Air Blank		0.000	04:29
Subject Sample	# 1	0.153	04:30
Breath Volume		2.082	04:30
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.149	04:33
Breath Volume		2.027	04:33
Air Blank		0.000	04:33
Control Test		0.081	04:34
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 01/07/2022 18:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: SAROUKOS

DATE/TIME OF TEST: 01/15/2022 05:01:33

Diagnostics Check		OK	04:54
Air Blank		0.000	04:54
Control Test		0.081	04:54
Air Blank		0.000	04:55
Subject Sample	# 1	0.125	04:57
Breath Volume		1.851	04:57
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 2	0.120	05:00
Breath Volume		1.789	05:00
Air Blank		0.000	05:00
Control Test		0.081	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 02/17/2022 04:46

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: PANESIS

DATE/TIME OF TEST: 01/20/2022 05:01:05

Diagnostics Check		OK	04:56	
Air Blank		0.000	04:57	
Control Test		0.081	04:57	
Air Blank		0.000	04:57	
Subject Sample	# 1		04:59	Subject Test Refused
Breath Volume		0.000	04:59	
Air Blank		0.000	05:00	
Control Test		0.081	05:00	
Air Blank		0.000	05:01	
Diagnostics Check		OK	05:01	

Data Download Date/Time: 02/17/2022 04:46

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: PROUT

DATE/TIME OF TEST: 01/26/2022 23:42:17

Diagnostics Check		OK	23:36
Air Blank		0.000	23:36
Control Test		0.081	23:37
Air Blank		0.000	23:37
Subject Sample	# 1	0.000	23:38
Breath Volume		2.433	23:38
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.000	23:40
Breath Volume		2.546	23:40
Air Blank		0.000	23:41
Control Test		0.081	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 02/17/2022 04:47

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 01/01/2022 01:42:56

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.080	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.219	01:38
Breath Volume		3.613	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.212	01:41
Breath Volume		3.664	01:41
Air Blank		0.000	01:42
Control Test		0.079	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 01/06/2022 22:59

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: LARSSON

DATE/TIME OF TEST: 01/01/2022 03:24:36

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.079	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.110	03:20
Breath Volume		1.363	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.108	03:23
Breath Volume		1.894	03:23
Air Blank		0.000	03:23
Control Test		0.079	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 01/06/2022 22:59

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: FERBEE

DATE/TIME OF TEST: 01/01/2022 04:44:05

Diagnostics Check		OK	04:41	
Air Blank		0.000	04:41	
Control Test		0.079	04:42	
Air Blank		0.000	04:42	
Subject Sample	# 1		04:42	Subject Test Refused
Breath Volume		0.000	04:42	
Air Blank		0.000	04:43	
Control Test		0.079	04:43	
Air Blank		0.000	04:43	
Diagnostics Check		OK	04:44	

Data Download Date/Time: 01/06/2022 22:59

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: WHITLEY

DATE/TIME OF TEST: 01/02/2022 00:03:14

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:00	
Control Test		0.078	00:00	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:01	Subject Test Refused
Breath Volume		0.000	00:01	
Air Blank		0.000	00:02	
Control Test		0.078	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 01/06/2022 22:59

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MCARTHUR

DATE/TIME OF TEST: 01/02/2022 04:17:26

Diagnostics Check		OK	04:14	
Air Blank		0.000	04:14	
Control Test		0.079	04:15	
Air Blank		0.000	04:15	
Subject Sample	# 1		04:16	Subject Test Refused
Breath Volume		0.000	04:16	
Air Blank		0.000	04:16	
Control Test		0.078	04:16	
Air Blank		0.000	04:17	
Diagnostics Check		OK	04:17	

Data Download Date/Time: 01/06/2022 23:00

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILMAN,DUSTIN,D
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: MCDONALD,HALEY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GAUTHIER

DATE/TIME OF TEST: 01/01/2022 02:57:01

Diagnostics Check		OK	02:54	
Air Blank		0.000	02:54	
Control Test		0.081	02:54	
Air Blank		0.000	02:55	
Subject Sample	# 1		02:55	Subject Test Refused
Breath Volume		0.000	02:55	
Air Blank		0.000	02:56	
Control Test		0.080	02:56	
Air Blank		0.000	02:56	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 01/01/2022 21:39

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBLINO,NICK,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: DUBLINO,NICK
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:59

SUBJECT'S LAST NAME: ARAGON

DATE/TIME OF TEST: 01/06/2022 08:27:23

Diagnostics Check		OK	08:21
Air Blank		0.000	08:21
Control Test		0.080	08:21
Air Blank		0.000	08:22
Subject Sample	# 1	0.000	08:23
Breath Volume		1.878	08:23
Air Blank		0.000	08:23
Air Blank		0.000	08:25
Subject Sample	# 2	0.000	08:26
Breath Volume		2.019	08:26
Air Blank		0.000	08:26
Control Test		0.080	08:26
Air Blank		0.000	08:27
Diagnostics Check		OK	08:27

Data Download Date/Time: 02/02/2022 11:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,KALEN,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: HOLMES,KALEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: BILALOVSKI

DATE/TIME OF TEST: 01/06/2022 09:25:07

Diagnostics Check		OK	09:19
Air Blank		0.000	09:19
Control Test		0.081	09:20
Air Blank		0.000	09:20
Subject Sample	# 1	0.000	09:20
Breath Volume		1.265	09:20
Air Blank		0.000	09:21
Air Blank		0.000	09:23
Subject Sample	# 2	0.000	09:23
Breath Volume		1.605	09:23
Air Blank		0.000	09:24
Control Test		0.081	09:24
Air Blank		0.000	09:25
Diagnostics Check		OK	09:25

Data Download Date/Time: 02/02/2022 11:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,CHAD,B
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: SMITH,CHAD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:39

SUBJECT'S LAST NAME: ABDULLAH

DATE/TIME OF TEST: 01/06/2022 14:09:00

Diagnostics Check		OK	14:02
Air Blank		0.000	14:02
Control Test		0.081	14:03
Air Blank		0.000	14:03
Subject Sample	# 1	0.000	14:04
Breath Volume		1.292	14:04
Air Blank		0.000	14:04
Air Blank		0.000	14:06
Subject Sample	# 2	0.000	14:07
Breath Volume		1.789	14:07
Air Blank		0.000	14:08
Control Test		0.081	14:08
Air Blank		0.000	14:08
Diagnostics Check		OK	14:09

Data Download Date/Time: 02/02/2022 11:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,CHAD,B
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: SMITH,CHAD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/06/2022 15:06:23

Diagnostics Check		OK	14:59
Air Blank		0.000	15:00
Control Test		0.081	15:00
Air Blank		0.000	15:01
Subject Sample	# 1	0.000	15:01
Breath Volume		1.542	15:01
Air Blank		0.000	15:02
Air Blank		0.000	15:04
Subject Sample	# 2	0.000	15:05
Breath Volume		1.785	15:05
Air Blank		0.000	15:05
Control Test		0.081	15:05
Air Blank		0.000	15:06
Diagnostics Check		OK	15:06

Data Download Date/Time: 02/02/2022 11:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: LOPEZ GREGORIO

DATE/TIME OF TEST: 01/19/2022 22:48:21

Diagnostics Check		OK	22:41
Air Blank		0.000	22:42
Control Test		0.079	22:42
Air Blank		0.000	22:43
Subject Sample	# 1	0.278	22:43
Breath Volume		2.335	22:43
Air Blank		0.000	22:44
Air Blank		0.000	22:46
Subject Sample	# 2	0.279	22:46
Breath Volume		2.152	22:46
Air Blank		0.000	22:47
Control Test		0.077	22:47
Air Blank		0.000	22:48
Diagnostics Check		OK	22:48

Data Download Date/Time: 02/02/2022 11:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,GEORGE,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: GONZALEZ,GEORGE
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: ZARAZUA

DATE/TIME OF TEST: 01/20/2022 10:21:48

Diagnostics Check		OK	10:16
Air Blank		0.000	10:16
Control Test		0.080	10:16
Air Blank		0.000	10:17
Subject Sample	# 1	0.000	10:17
Breath Volume		1.468	10:17
Air Blank		0.000	10:18
Air Blank		0.000	10:20
Subject Sample	# 2	0.000	10:20
Breath Volume		1.582	10:20
Air Blank		0.000	10:20
Control Test		0.080	10:21
Air Blank		0.000	10:21
Diagnostics Check		OK	10:21

Data Download Date/Time: 02/02/2022 11:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,GEORGE,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: GONZALEZ,GEORGE
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: DESIR

DATE/TIME OF TEST: 01/20/2022 10:46:17

Diagnostics Check		OK	10:40
Air Blank		0.000	10:40
Control Test		0.081	10:41
Air Blank		0.000	10:41
Subject Sample	# 1	0.000	10:42
Breath Volume		1.640	10:42
Air Blank		0.000	10:42
Air Blank		0.000	10:44
Subject Sample	# 2	0.000	10:44
Breath Volume		1.562	10:44
Air Blank		0.000	10:45
Control Test		0.081	10:45
Air Blank		0.000	10:46
Diagnostics Check		OK	10:46

Data Download Date/Time: 02/02/2022 11:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,GEORGE,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: GONZALEZ,GEORGE
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/20/2022 11:05:33

Diagnostics Check		OK	10:59
Air Blank		0.000	11:00
Control Test		0.081	11:00
Air Blank		0.000	11:00
Subject Sample	# 1	0.000	11:01
Breath Volume		2.070	11:01
Air Blank		0.000	11:01
Air Blank		0.000	11:03
Subject Sample	# 2	0.000	11:04
Breath Volume		1.640	11:04
Air Blank		0.000	11:04
Control Test		0.080	11:04
Air Blank		0.000	11:05
Diagnostics Check		OK	11:05

Data Download Date/Time: 02/02/2022 11:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,GEORGE,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: GONZALEZ,GEORGE
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 01/20/2022 11:22:09

Diagnostics Check		OK	11:15
Air Blank		0.000	11:16
Control Test		0.081	11:16
Air Blank		0.000	11:17
Subject Sample	# 1	0.037	11:17
Breath Volume		1.246	11:17
Air Blank		0.000	11:18
Air Blank		0.000	11:20
Subject Sample	# 2	0.037	11:20
Breath Volume		1.156	11:20
Air Blank		0.000	11:21
Control Test		0.081	11:21
Air Blank		0.000	11:22
Diagnostics Check		OK	11:22

Data Download Date/Time: 02/02/2022 11:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,GEORGE,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: GONZALEZ,GEORGE
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: PUPO

DATE/TIME OF TEST: 01/20/2022 14:59:31

Diagnostics Check		OK	14:53
Air Blank		0.000	14:54
Control Test		0.081	14:54
Air Blank		0.000	14:55
Subject Sample	# 1	0.000	14:55
Breath Volume		1.718	14:55
Air Blank		0.000	14:55
Air Blank		0.000	14:57
Subject Sample	# 2	0.000	14:58
Breath Volume		1.765	14:58
Air Blank		0.000	14:58
Control Test		0.080	14:58
Air Blank		0.000	14:59
Diagnostics Check		OK	14:59

Data Download Date/Time: 02/02/2022 11:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,GEORGE,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: GONZALEZ,GEORGE
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 01/21/2022 09:37:16

Diagnostics Check		OK	09:31
Air Blank		0.000	09:31
Control Test		0.081	09:32
Air Blank		0.000	09:32
Subject Sample	# 1	0.000	09:33
Breath Volume		1.699	09:33
Air Blank		0.000	09:33
Air Blank		0.000	09:35
Subject Sample	# 2	0.000	09:35
Breath Volume		1.597	09:35
Air Blank		0.000	09:36
Control Test		0.080	09:36
Air Blank		0.000	09:37
Diagnostics Check		OK	09:37

Data Download Date/Time: 02/02/2022 11:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,GEORGE,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: GONZALEZ,GEORGE
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: MC ERLEAN

DATE/TIME OF TEST: 01/21/2022 09:52:17

Diagnostics Check		OK	09:46
Air Blank		0.000	09:46
Control Test		0.080	09:47
Air Blank		0.000	09:47
Subject Sample	# 1	0.000	09:48
Breath Volume		1.851	09:48
Air Blank		0.000	09:48
Air Blank		0.000	09:50
Subject Sample	# 2	0.000	09:50
Breath Volume		2.253	09:50
Air Blank		0.000	09:51
Control Test		0.081	09:51
Air Blank		0.000	09:52
Diagnostics Check		OK	09:52

Data Download Date/Time: 02/02/2022 11:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 15:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: CHAVEZ-ROBLES

DATE/TIME OF TEST: 01/26/2022 15:47:06

Diagnostics Check		OK	15:41
Air Blank		0.000	15:41
Control Test		0.081	15:41
Air Blank		0.000	15:42
Subject Sample	# 1	0.000	15:42
Breath Volume		2.066	15:42
Air Blank		0.000	15:43
Air Blank		0.000	15:45
Subject Sample	# 2	0.000	15:45
Breath Volume		2.109	15:45
Air Blank		0.000	15:46
Control Test		0.080	15:46
Air Blank		0.000	15:47
Diagnostics Check		OK	15:47

Data Download Date/Time: 02/02/2022 11:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: MCCROSKEY,SCOTT
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 01/01/2022 22:18:30

Diagnostics Check		OK	22:10
Air Blank		0.000	22:11
Control Test		0.081	22:11
Air Blank		0.000	22:12
Subject Sample	# 1	0.242	22:13
Breath Volume		1.777	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.247	22:16
Breath Volume		1.542	22:16
Air Blank		0.000	22:17
Control Test		0.079	22:17
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 01/28/2022 14:59

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: GREEN,KATELYN
ARRESTING OFFICER AGENCY: TITUSVILLE PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/02/2022 21:15:26

Diagnostics Check	OK	21:14	
Air Blank	0.000	21:14	
		21:14	RFI Detect
Air Blank	0.000	21:15	

Data Download Date/Time: 01/28/2022 14:59

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: GREEN,KATELYN
ARRESTING OFFICER AGENCY: TITUSVILLE PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/02/2022 21:32:48

Diagnostics Check		OK	21:18	
Air Blank		0.000	21:18	
Control Test		0.081	21:18	
Air Blank		0.000	21:19	
Subject Sample	# 1	0.190	21:21	
Breath Volume		1.265	21:21	
Air Blank		0.000	21:22	
Air Blank		0.000	21:24	
Subject Sample	# 2	0.171	21:27	Volume Not Met
Breath Volume		0.910	21:27	
Air Blank		0.000	21:28	
Air Blank		0.000	21:30	
Subject Sample	# 3	0.187	21:31	
Breath Volume		1.250	21:31	
Air Blank		0.000	21:31	
Control Test		0.079	21:32	
Air Blank		0.000	21:32	
Diagnostics Check		OK	21:32	

Data Download Date/Time: 01/28/2022 15:00

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: WINSTEAD,JUSTIN
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 01/07/2022 00:14:06

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.081	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.133	00:09
Breath Volume		1.218	00:09
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.147	00:12
Breath Volume		3.078	00:12
Air Blank		0.000	00:13
Control Test		0.080	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 01/28/2022 15:00

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: BACHELLER,ALEXANDRIA
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 01/07/2022 03:47:24

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.079	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.145	03:42
Breath Volume		2.468	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.140	03:45
Breath Volume		2.093	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 01/28/2022 15:00

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANG,MATTHEW,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: YANICK,MICHAEL
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 01/10/2022 01:36:13

Diagnostics Check		OK	01:28
Air Blank		0.000	01:29
Control Test		0.081	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.176	01:31
Breath Volume		2.960	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:33
Subject Sample	# 2	0.183	01:34
Breath Volume		2.882	01:34
Air Blank		0.000	01:35
Control Test		0.081	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 01/28/2022 15:00

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: JONES,DAKOTA
ARRESTING OFFICER AGENCY: TITUSVILLE PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 01/16/2022 03:15:43

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:11	
Control Test		0.081	03:12	
Air Blank		0.000	03:12	
Subject Sample	# 1	0.040	03:14	
Breath Volume		1.234	03:14	
Air Blank			03:15	Purge Fail
Air Blank		0.000	03:15	

Data Download Date/Time: 01/28/2022 15:00

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: JONES,DAKOTA
ARRESTING OFFICER AGENCY: TITUSVILLE PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 01/16/2022 03:19:37

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.068	03:19	Control Outside Tolerance
Air Blank	0.000	03:19	

Data Download Date/Time: 01/28/2022 15:00

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: JONES,DAKOTA
ARRESTING OFFICER AGENCY: TITUSVILLE PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 01/16/2022 03:23:09

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.069	03:22	Control Outside Tolerance
Air Blank	0.000	03:23	

Data Download Date/Time: 01/28/2022 15:00

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: JONES,DAKOTA
ARRESTING OFFICER AGENCY: TITUSVILLE PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 01/16/2022 03:34:53

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.072	03:34	Control Outside Tolerance
Air Blank	0.000	03:34	

Data Download Date/Time: 01/28/2022 15:00

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: SADLER,JASON
ARRESTING OFFICER AGENCY: MELBOURNE BEACH PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MCELROY

DATE/TIME OF TEST: 01/20/2022 03:17:40

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:10	
Control Test		0.075	03:10	
Air Blank		0.000	03:11	
Subject Sample	# 1	0.136	03:12	
Breath Volume		1.187	03:12	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 2	0.145	03:16	
Breath Volume		1.933	03:16	
Air Blank			03:16	Purge Fail
Air Blank			03:17	Purge Fail

Data Download Date/Time: 01/28/2022 15:00

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: SADLER,JASON
ARRESTING OFFICER AGENCY: MELBOURNE BEACH PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MCELROY

DATE/TIME OF TEST: 01/20/2022 03:22:28

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.051	03:21	Control Outside Tolerance
Air Blank	0.000	03:22	

Data Download Date/Time: 01/28/2022 15:00

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: SADLER,JASON
ARRESTING OFFICER AGENCY: MELBOURNE BEACH PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MCELROY

DATE/TIME OF TEST: 01/20/2022 03:26:10

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.057	03:25	Control Outside Tolerance
Air Blank	0.000	03:26	

Data Download Date/Time: 01/28/2022 15:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANG,MATTHEW,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: LEWIS,BREEAN
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BLACKETER

DATE/TIME OF TEST: 01/23/2022 01:41:36

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.076	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.098	01:37
Breath Volume		2.371	01:37
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.100	01:40
Breath Volume		2.906	01:40
Air Blank		0.000	01:40
Control Test		0.076	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 01/28/2022 15:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANG,MATTHEW,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: MORILLO,RUANNY
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: LOPEZ-MENDEZ

DATE/TIME OF TEST: 01/23/2022 04:01:26

Diagnostics Check		OK	03:52
Air Blank		0.000	03:53
Control Test		0.078	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.242	03:56
Breath Volume		1.574	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.236	03:59
Breath Volume		1.898	03:59
Air Blank		0.000	04:00
Control Test		0.078	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 01/28/2022 15:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,BRADLEY,
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: MCLAUGHLIN,BRADLEY
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 01/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: DATTILO

DATE/TIME OF TEST: 01/24/2022 22:47:31

Diagnostics Check		OK	22:40
Air Blank		0.000	22:40
Control Test		0.081	22:41
Air Blank		0.000	22:41
Subject Sample	# 1	0.320	22:42
Breath Volume		1.863	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.308	22:45
Breath Volume		2.046	22:45
Air Blank		0.000	22:46
Control Test		0.076	22:46
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 01/28/2022 15:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBIN,GASTON,A
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: PRICE,C
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 01/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: SPARKS

DATE/TIME OF TEST: 01/27/2022 04:32:25

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.081	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.211	04:27
Breath Volume		1.753	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.210	04:30
Breath Volume		1.996	04:30
Air Blank		0.000	04:31
Control Test		0.081	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 01/28/2022 15:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 01/01/2022 02:27:01

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:24	
Control Test		0.081	02:24	
Air Blank		0.000	02:25	
Subject Sample	# 1		02:25	Subject Test Refused
Breath Volume		0.000	02:25	
Air Blank		0.000	02:26	
Control Test		0.081	02:26	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:27	

Data Download Date/Time: 01/13/2022 10:33

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: LOPEZ AGUILAR

DATE/TIME OF TEST: 01/01/2022 04:27:55

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.081	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.103	04:23
Breath Volume		2.312	04:23
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.106	04:26
Breath Volume		2.394	04:26
Air Blank		0.000	04:26
Control Test		0.080	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 01/13/2022 10:33

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: COUSER

DATE/TIME OF TEST: 01/01/2022 05:13:47

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:11	
Control Test		0.081	05:11	
Air Blank		0.000	05:12	
Subject Sample	# 1		05:12	Subject Test Refused
Breath Volume		0.000	05:12	
Air Blank		0.000	05:12	
Control Test		0.080	05:13	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 01/13/2022 10:33

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,P,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 01/01/2022 18:43:20

Diagnostics Check		OK	18:36
Air Blank		0.000	18:36
Control Test		0.081	18:36
Air Blank		0.000	18:37
Subject Sample	# 1	0.149	18:37
Breath Volume		2.156	18:37
Air Blank		0.000	18:38
Air Blank		0.000	18:40
Subject Sample	# 2	0.140	18:41
Breath Volume		2.359	18:41
Air Blank		0.000	18:42
Control Test		0.079	18:42
Air Blank		0.000	18:43
Diagnostics Check		OK	18:43

Data Download Date/Time: 01/13/2022 10:33

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: FINLEY

DATE/TIME OF TEST: 01/01/2022 20:39:24

Diagnostics Check		OK	20:36	
Air Blank		0.000	20:36	
Control Test		0.081	20:37	
Air Blank		0.000	20:37	
Subject Sample	# 1		20:38	Subject Test Refused
Breath Volume		0.000	20:38	
Air Blank		0.000	20:38	
Control Test		0.081	20:38	
Air Blank		0.000	20:39	
Diagnostics Check		OK	20:39	

Data Download Date/Time: 01/13/2022 10:33

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 01/01/2022 23:54:31

Diagnostics Check		OK	23:47
Air Blank		0.000	23:47
Control Test		0.080	23:48
Air Blank		0.000	23:48
Subject Sample	# 1	0.246	23:49
Breath Volume		2.671	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.231	23:52
Breath Volume		3.593	23:52
Air Blank		0.000	23:53
Control Test		0.080	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 01/13/2022 10:33

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: DIAZ PERZ

DATE/TIME OF TEST: 01/02/2022 02:05:51

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:03	
Control Test		0.079	02:03	
Air Blank		0.000	02:04	
Subject Sample	# 1		02:04	Subject Test Refused
Breath Volume		0.000	02:04	
Air Blank		0.000	02:04	
Control Test		0.080	02:05	
Air Blank		0.000	02:05	
Diagnostics Check		OK	02:05	

Data Download Date/Time: 01/13/2022 10:33

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/03/2022 00:55:53

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:53	
Control Test		0.082	00:53	
Air Blank		0.000	00:54	
Subject Sample	# 1		00:54	Subject Test Refused
Breath Volume		0.000	00:54	
Air Blank		0.000	00:54	
Control Test		0.081	00:55	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 01/13/2022 10:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: SEATON

DATE/TIME OF TEST: 01/07/2022 04:50:14

Diagnostics Check		OK	04:47	
Air Blank		0.000	04:47	
Control Test		0.082	04:48	
Air Blank		0.000	04:48	
Subject Sample	# 1		04:48	Subject Test Refused
Breath Volume		0.000	04:48	
Air Blank		0.000	04:49	
Control Test		0.081	04:49	
Air Blank		0.000	04:50	
Diagnostics Check		OK	04:50	

Data Download Date/Time: 01/13/2022 10:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: WAJDA

DATE/TIME OF TEST: 01/07/2022 22:57:16

Diagnostics Check		OK	22:54	
Air Blank		0.000	22:54	
Control Test		0.081	22:55	
Air Blank		0.000	22:55	
Subject Sample	# 1		22:55	Subject Test Refused
Breath Volume		0.000	22:55	
Air Blank		0.000	22:56	
Control Test		0.081	22:56	
Air Blank		0.000	22:57	
Diagnostics Check		OK	22:57	

Data Download Date/Time: 01/13/2022 10:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: CRUMBIE

DATE/TIME OF TEST: 01/08/2022 06:01:36

Diagnostics Check		OK	05:58	
Air Blank		0.000	05:59	
Control Test		0.079	05:59	
Air Blank		0.000	06:00	
Subject Sample	# 1		06:00	Subject Test Refused
Breath Volume		0.000	06:00	
Air Blank		0.000	06:00	
Control Test		0.079	06:00	
Air Blank		0.000	06:01	
Diagnostics Check		OK	06:01	

Data Download Date/Time: 01/13/2022 10:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: BYRD SIMMONS

DATE/TIME OF TEST: 01/08/2022 21:46:35

Diagnostics Check		OK	21:43	
Air Blank		0.000	21:44	
Control Test		0.081	21:44	
Air Blank		0.000	21:45	
Subject Sample	# 1		21:45	Subject Test Refused
Breath Volume		0.000	21:45	
Air Blank		0.000	21:45	
Control Test		0.082	21:45	
Air Blank		0.000	21:46	
Diagnostics Check		OK	21:46	

Data Download Date/Time: 01/13/2022 10:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/09/2022 02:40:54

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:38	
Control Test		0.081	02:38	
Air Blank		0.000	02:39	
Subject Sample	# 1		02:39	Subject Test Refused
Breath Volume		0.000	02:39	
Air Blank		0.000	02:39	
Control Test		0.080	02:40	
Air Blank		0.000	02:40	
Diagnostics Check		OK	02:40	

Data Download Date/Time: 01/13/2022 10:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: VARGAS TORRES

DATE/TIME OF TEST: 01/09/2022 04:07:28

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.082	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.138	04:02
Breath Volume		2.863	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:04
Subject Sample	# 2	0.135	04:05
Breath Volume		2.074	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 01/13/2022 10:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: COLLERD

DATE/TIME OF TEST: 01/09/2022 22:21:22

Diagnostics Check		OK	22:16	
Air Blank		0.000	22:17	
Control Test		0.080	22:17	
Air Blank		0.000	22:18	
Subject Sample	# 1		22:19	Subject Test Refused
Breath Volume		0.000	22:19	
Air Blank		0.000	22:20	
Control Test		0.082	22:20	
Air Blank		0.000	22:21	
Diagnostics Check		OK	22:21	

Data Download Date/Time: 01/13/2022 10:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: WESTMORELAND

DATE/TIME OF TEST: 01/11/2022 02:44:54

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.081	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.149	02:40
Breath Volume		1.820	02:40
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.151	02:43
Breath Volume		1.632	02:43
Air Blank		0.000	02:43
Control Test		0.081	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 01/13/2022 10:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: WARE

DATE/TIME OF TEST: 01/12/2022 17:22:49

Diagnostics Check		OK	17:16
Air Blank		0.000	17:16
Control Test		0.081	17:17
Air Blank		0.000	17:17
Subject Sample	# 1	0.000	17:18
Breath Volume		4.476	17:18
Air Blank		0.000	17:18
Air Blank		0.000	17:20
Subject Sample	# 2	0.000	17:21
Breath Volume		3.625	17:21
Air Blank		0.000	17:21
Control Test		0.080	17:22
Air Blank		0.000	17:22
Diagnostics Check		OK	17:22

Data Download Date/Time: 01/13/2022 10:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MOORMAN

DATE/TIME OF TEST: 01/22/2022 03:57:25

Diagnostics Check		OK	03:54	
Air Blank		0.000	03:54	
Control Test		0.082	03:55	
Air Blank		0.000	03:55	
Subject Sample	# 1		03:56	Subject Test Refused
Breath Volume		0.000	03:56	
Air Blank		0.000	03:56	
Control Test		0.082	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 02/05/2022 17:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: GISLASON

DATE/TIME OF TEST: 01/23/2022 00:13:33

Diagnostics Check		OK	00:09	
Air Blank		0.000	00:10	
Control Test		0.081	00:10	
Air Blank		0.000	00:11	
Subject Sample	# 1		00:12	Subject Test Refused
Breath Volume		0.000	00:12	
Air Blank		0.000	00:12	
Control Test		0.081	00:12	
Air Blank		0.000	00:13	
Diagnostics Check		OK	00:13	

Data Download Date/Time: 02/05/2022 17:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: DOMINGO MALDONADO

DATE/TIME OF TEST: 01/23/2022 05:16:46

Diagnostics Check		OK	05:13	
Air Blank		0.000	05:14	
Control Test		0.082	05:14	
Air Blank		0.000	05:15	
Subject Sample	# 1		05:15	Subject Test Refused
Breath Volume		0.000	05:15	
Air Blank		0.000	05:15	
Control Test		0.082	05:16	
Air Blank		0.000	05:16	
Diagnostics Check		OK	05:16	

Data Download Date/Time: 02/05/2022 17:10

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 01/23/2022 22:48:09

Diagnostics Check		OK	22:40
Air Blank		0.000	22:41
Control Test		0.082	22:41
Air Blank		0.000	22:42
Subject Sample	# 1	0.000	22:42
Breath Volume		1.582	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.000	22:46
Breath Volume		1.816	22:46
Air Blank		0.000	22:47
Control Test		0.080	22:47
Air Blank		0.000	22:48
Diagnostics Check		OK	22:48

Data Download Date/Time: 02/05/2022 17:10

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 01/25/2022 22:51:49

Diagnostics Check		OK	22:44
Air Blank		0.000	22:45
Control Test		0.081	22:45
Air Blank		0.000	22:46
Subject Sample	# 1	0.213	22:46
Breath Volume		2.238	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.207	22:50
Breath Volume		1.917	22:50
Air Blank		0.000	22:50
Control Test		0.080	22:51
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 02/05/2022 17:10

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IKRAM,Y
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: WESTCOTT

DATE/TIME OF TEST: 01/26/2022 01:23:25

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.137	01:18
Breath Volume		2.953	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.122	01:21
Breath Volume		1.164	01:21
Air Blank		0.000	01:22
Control Test		0.080	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 02/07/2022 11:08

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: RADCLIFFE

DATE/TIME OF TEST: 01/26/2022 02:38:00

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:35	
Control Test		0.081	02:35	
Air Blank		0.000	02:36	
Subject Sample	# 1		02:36	Subject Test Refused
Breath Volume		0.000	02:36	
Air Blank		0.000	02:37	
Control Test		0.081	02:37	
Air Blank		0.000	02:37	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 02/07/2022 11:08

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 01/26/2022 03:45:40

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.081	03:43	
Air Blank		0.000	03:44	
Subject Sample	# 1		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:44	
Control Test		0.081	03:45	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 02/07/2022 11:08

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/27/2022 03:21:15

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:18	
Control Test		0.081	03:19	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:19	Subject Test Refused
Breath Volume		0.000	03:19	
Air Blank		0.000	03:20	
Control Test		0.082	03:20	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 02/07/2022 11:08

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 01/28/2022 04:23:12

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.082	04:17
Air Blank		0.000	04:17
Subject Sample	# 1	0.133	04:18
Breath Volume		1.765	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:20
Subject Sample	# 2	0.132	04:21
Breath Volume		3.234	04:21
Air Blank		0.000	04:22
Control Test		0.081	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 02/07/2022 11:08

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 01/29/2022 04:27:55

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.082	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.089	04:23
Breath Volume		2.535	04:23
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.090	04:26
Breath Volume		4.359	04:26
Air Blank		0.000	04:26
Control Test		0.082	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 02/07/2022 11:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: ALTAMIRANO VENEGAS

DATE/TIME OF TEST: 01/29/2022 05:09:06

Diagnostics Check		OK	05:01
Air Blank		0.000	05:02
Control Test		0.081	05:02
Air Blank		0.000	05:03
Subject Sample	# 1	0.171	05:04
Breath Volume		1.335	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:06
Subject Sample	# 2	0.176	05:07
Breath Volume		1.804	05:07
Air Blank		0.000	05:08
Control Test		0.080	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 02/07/2022 11:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 01/30/2022 05:32:58

Diagnostics Check		OK	05:30	
Air Blank		0.000	05:30	
Control Test		0.082	05:30	
Air Blank		0.000	05:31	
Subject Sample	# 1		05:31	Subject Test Refused
Breath Volume		0.000	05:31	
Air Blank		0.000	05:32	
Control Test		0.083	05:32	
Air Blank		0.000	05:32	
Diagnostics Check		OK	05:32	

Data Download Date/Time: 02/07/2022 11:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: SHARP

DATE/TIME OF TEST: 01/31/2022 02:18:14

Diagnostics Check		OK	02:11
Air Blank		0.000	02:12
Control Test		0.082	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.200	02:13
Breath Volume		2.527	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.209	02:16
Breath Volume		2.332	02:16
Air Blank		0.000	02:17
Control Test		0.082	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 02/07/2022 11:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 01/31/2022 22:40:52

Diagnostics Check		OK	22:34
Air Blank		0.000	22:34
Control Test		0.082	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.181	22:36
Breath Volume		1.750	22:36
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.180	22:39
Breath Volume		2.152	22:39
Air Blank		0.000	22:39
Control Test		0.081	22:40
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 02/07/2022 11:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 01/06/2022 03:12:01

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:09	
Control Test		0.076	03:10	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:10	Subject Test Refused
Breath Volume		0.000	03:10	
Air Blank		0.000	03:11	
Control Test		0.077	03:11	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 01/19/2022 15:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 01/07/2022 00:44:57

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:42	
Control Test		0.077	00:42	
Air Blank		0.000	00:43	
Subject Sample	# 1		00:43	Subject Test Refused
Breath Volume		0.000	00:43	
Air Blank		0.000	00:44	
Control Test		0.076	00:44	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 01/19/2022 15:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REQUENA,CARLOS,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROCHA,A
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: BREAN

DATE/TIME OF TEST: 01/08/2022 19:06:44

Diagnostics Check		OK	19:03	
Air Blank		0.000	19:04	
Control Test		0.077	19:04	
Air Blank		0.000	19:05	
Subject Sample	# 1		19:05	Subject Test Refused
Breath Volume		0.000	19:05	
Air Blank		0.000	19:05	
Control Test		0.077	19:06	
Air Blank		0.000	19:06	
Diagnostics Check		OK	19:06	

Data Download Date/Time: 01/19/2022 15:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REQUENA,CARLOS,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GOSIER,M
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/09/2022 02:14:41

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:09	
Control Test		0.077	02:09	
Air Blank		0.000	02:09	
Subject Sample	# 1	0.196	02:13	Volume Not Met
Breath Volume		0.019	02:13	
Air Blank			02:13	Purge Fail
Air Blank			02:14	Purge Fail

Data Download Date/Time: 01/19/2022 15:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REQUENA,CARLOS,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GOSIER,M
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/09/2022 02:22:05

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.076	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1	0.237	02:20	Volume Not Met
Breath Volume		0.164	02:20	
Air Blank			02:21	Purge Fail
Air Blank		0.000	02:22	

Data Download Date/Time: 01/19/2022 15:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REQUENA,CARLOS,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GOSIER,M
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/09/2022 02:36:43

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:23	
Control Test		0.076	02:24	
Air Blank		0.000	02:24	
Subject Sample	# 1	0.229	02:25	
Breath Volume		1.531	02:25	
Air Blank		0.000	02:26	
Air Blank		0.000	02:28	
Subject Sample	# 2	0.226	02:31	Volume Not Met
Breath Volume		0.093	02:31	
Air Blank		0.000	02:32	
Air Blank		0.000	02:34	
Subject Sample	# 3	0.219	02:35	
Breath Volume		1.398	02:35	
Air Blank		0.000	02:35	
Control Test		0.077	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:36	

Data Download Date/Time: 01/19/2022 15:54

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: JUDICE

DATE/TIME OF TEST: 01/12/2022 16:12:11

Diagnostics Check		OK	16:06
Air Blank		0.000	16:06
Control Test		0.078	16:07
Air Blank		0.000	16:07
Subject Sample	# 1	0.000	16:07
Breath Volume		2.433	16:07
Air Blank		0.000	16:08
Air Blank		0.000	16:10
Subject Sample	# 2	0.000	16:10
Breath Volume		2.207	16:10
Air Blank		0.000	16:11
Control Test		0.079	16:11
Air Blank		0.000	16:12
Diagnostics Check		OK	16:12

Data Download Date/Time: 01/19/2022 15:54

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: BRZOZOWSKI

DATE/TIME OF TEST: 01/13/2022 18:20:15

Diagnostics Check		OK	18:17	
Air Blank		0.000	18:17	
Control Test		0.077	18:18	
Air Blank		0.000	18:18	
Subject Sample	# 1		18:18	Subject Test Refused
Breath Volume		0.000	18:18	
Air Blank		0.000	18:19	
Control Test		0.076	18:19	
Air Blank		0.000	18:20	
Diagnostics Check		OK	18:20	

Data Download Date/Time: 01/19/2022 15:54

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: HERRICK

DATE/TIME OF TEST: 01/15/2022 00:26:40

Diagnostics Check		OK	00:23	
Air Blank		0.000	00:24	
Control Test		0.078	00:24	
Air Blank		0.000	00:25	
Subject Sample	# 1		00:25	Subject Test Refused
Breath Volume		0.000	00:25	
Air Blank		0.000	00:25	
Control Test		0.077	00:26	
Air Blank		0.000	00:26	
Diagnostics Check		OK	00:26	

Data Download Date/Time: 01/19/2022 15:54

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 01/15/2022 03:38:16

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.078	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.238	03:33
Breath Volume		1.496	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.224	03:36
Breath Volume		1.703	03:36
Air Blank		0.000	03:37
Control Test		0.079	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 01/19/2022 15:54

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCHUK,J,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: SHARMA

DATE/TIME OF TEST: 01/17/2022 22:58:38

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:56	
Control Test		0.078	22:56	
Air Blank		0.000	22:57	
Subject Sample	# 1		22:57	Subject Test Refused
Breath Volume		0.000	22:57	
Air Blank		0.000	22:57	
Control Test		0.077	22:58	
Air Blank		0.000	22:58	
Diagnostics Check		OK	22:58	

Data Download Date/Time: 01/19/2022 15:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094981
DRY GAS EXPIRY DATE: 02/22/2022
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: HULMES

DATE/TIME OF TEST: 01/27/2022 19:34:50

Diagnostics Check		OK	19:32	
Air Blank		0.000	19:32	
Control Test		0.080	19:32	
Air Blank		0.000	19:33	
Subject Sample	# 1		19:33	Subject Test Refused
Breath Volume		0.000	19:33	
Air Blank		0.000	19:33	
Control Test		0.080	19:34	
Air Blank		0.000	19:34	
Diagnostics Check		OK	19:34	

Data Download Date/Time: 02/24/2022 13:48

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094981
DRY GAS EXPIRY DATE: 02/22/2022
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/28/2022 02:21:32

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:19	
Control Test		0.080	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 1		02:20	Subject Test Refused
Breath Volume		0.000	02:20	
Air Blank		0.000	02:20	
Control Test		0.080	02:20	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 02/24/2022 13:48

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: LOPEZ,ROBERT
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094981
DRY GAS EXPIRY DATE: 02/22/2022
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 01/29/2022 01:16:16

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.079	01:10
Air Blank		0.000	01:10
Subject Sample	# 1	0.195	01:11
Breath Volume		1.613	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.187	01:14
Breath Volume		1.554	01:14
Air Blank		0.000	01:15
Control Test		0.080	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 02/24/2022 13:48

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: COCKFIELD

DATE/TIME OF TEST: 01/01/2022 00:02:47

Diagnostics Check		OK	23:55
Air Blank		0.000	23:55
Control Test		0.080	23:55
Air Blank		0.000	23:56
Subject Sample	# 2	0.198	23:57
Breath Volume		2.488	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.202	00:01
Breath Volume		2.890	00:01
Air Blank		0.000	00:01
Control Test		0.079	00:02
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 01/13/2022 10:26

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: FIORENTINO

DATE/TIME OF TEST: 01/01/2022 05:55:19

Diagnostics Check		OK	05:49
Air Blank		0.000	05:49
Control Test		0.080	05:49
Air Blank		0.000	05:50
Subject Sample	# 1	0.213	05:50
Breath Volume		1.542	05:50
Air Blank		0.000	05:51
Air Blank		0.000	05:53
Subject Sample	# 2	0.213	05:53
Breath Volume		1.812	05:53
Air Blank		0.000	05:54
Control Test		0.080	05:54
Air Blank		0.000	05:55
Diagnostics Check		OK	05:55

Data Download Date/Time: 01/13/2022 10:26

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 01/02/2022 02:54:18

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.081	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.026	02:49
Breath Volume		2.667	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.027	02:52
Breath Volume		2.410	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 01/13/2022 10:26

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: CLOTAIRE

DATE/TIME OF TEST: 01/04/2022 00:53:40

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.082	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.000	00:49
Breath Volume		1.914	00:49
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.000	00:52
Breath Volume		1.218	00:52
Air Blank		0.000	00:52
Control Test		0.081	00:53
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 01/13/2022 10:26

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 01/06/2022 03:17:25

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.081	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.000	03:12
Breath Volume		1.429	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.000	03:16
Breath Volume		1.207	03:16
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 01/13/2022 10:26

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TIMMONS,B
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 01/07/2022 04:05:58

Diagnostics Check		OK	03:59
Air Blank		0.000	03:59
Control Test		0.081	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.100	04:01
Breath Volume		3.113	04:01
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.098	04:04
Breath Volume		3.093	04:04
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 01/13/2022 10:26

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: PRESTON

DATE/TIME OF TEST: 01/08/2022 04:44:08

Diagnostics Check		OK	04:37
Air Blank		0.000	04:38
Control Test		0.081	04:38
Air Blank		0.000	04:39
Subject Sample	# 1	0.212	04:39
Breath Volume		1.437	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.217	04:42
Breath Volume		2.375	04:42
Air Blank		0.000	04:43
Control Test		0.080	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 01/13/2022 10:26

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SWANSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 01/09/2022 04:37:27

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.081	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.112	04:31
Breath Volume		1.140	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.110	04:35
Breath Volume		1.144	04:35
Air Blank		0.000	04:36
Control Test		0.080	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 01/13/2022 10:26

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: GOODE,D
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 01/09/2022 06:27:52

Diagnostics Check		OK	06:18
Air Blank		0.000	06:18
Control Test		0.080	06:19
Air Blank		0.000	06:19
Subject Sample	# 1	0.186	06:20
Breath Volume		1.449	06:20
Air Blank		0.000	06:21
Air Blank		0.000	06:23
Subject Sample	# 2	0.183	06:26
Breath Volume		1.972	06:26
Air Blank		0.000	06:26
Control Test		0.080	06:27
Air Blank		0.000	06:27
Diagnostics Check		OK	06:27

Data Download Date/Time: 01/13/2022 10:26

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/13/2022 20:24:40

Diagnostics Check		OK	20:18
Air Blank		0.000	20:18
Control Test		0.081	20:19
Air Blank		0.000	20:19
Subject Sample	# 1	0.125	20:20
Breath Volume		3.375	20:20
Air Blank		0.000	20:20
Air Blank		0.000	20:22
Subject Sample	# 2	0.125	20:23
Breath Volume		3.488	20:23
Air Blank		0.000	20:23
Control Test		0.081	20:24
Air Blank		0.000	20:24
Diagnostics Check		OK	20:24

Data Download Date/Time: 02/05/2022 16:54

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: GUIFARRO DIAZ

DATE/TIME OF TEST: 01/13/2022 21:35:21

Diagnostics Check		OK	21:32	
Air Blank		0.000	21:32	
Control Test		0.081	21:33	
Air Blank		0.000	21:33	
Subject Sample	# 1		21:33	Subject Test Refused
Breath Volume		0.000	21:33	
Air Blank		0.000	21:34	
Control Test		0.080	21:34	
Air Blank		0.000	21:35	
Diagnostics Check		OK	21:35	

Data Download Date/Time: 02/07/2022 11:13

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: MARTINEZ EUCEDA

DATE/TIME OF TEST: 01/14/2022 02:04:37

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.081	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.068	02:00
Breath Volume		1.128	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.065	02:03
Breath Volume		1.261	02:03
Air Blank		0.000	02:03
Control Test		0.080	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 02/07/2022 11:13

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: COMM VEHIC
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/14/2022 20:24:30

Diagnostics Check		OK	20:17
Air Blank		0.000	20:18
Control Test		0.081	20:18
Air Blank		0.000	20:19
Subject Sample	# 1	0.106	20:20
Breath Volume		2.746	20:20
Air Blank		0.000	20:20
Air Blank		0.000	20:22
Subject Sample	# 2	0.105	20:22
Breath Volume		1.914	20:22
Air Blank		0.000	20:23
Control Test		0.080	20:23
Air Blank		0.000	20:24
Diagnostics Check		OK	20:24

Data Download Date/Time: 02/07/2022 11:13

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/15/2022 03:40:00

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.081	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.011	03:35
Breath Volume		2.421	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.010	03:38
Breath Volume		1.871	03:38
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:40

Data Download Date/Time: 02/07/2022 11:13

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 01/15/2022 05:27:10

Diagnostics Check		OK	05:11	
Air Blank		0.000	05:12	
Control Test		0.081	05:12	
Air Blank		0.000	05:12	
Subject Sample	# 1	0.178	05:14	
Breath Volume		1.214	05:14	
Air Blank		0.000	05:15	
Air Blank		0.000	05:16	
Subject Sample	# 2	0.152	05:19	Volume Not Met
Breath Volume		0.074	05:19	
Air Blank		0.000	05:20	
Air Blank		0.000	05:22	
Subject Sample	# 3	0.153	05:25	Volume Not Met
Breath Volume		0.019	05:25	
Air Blank		0.000	05:26	
Control Test		0.078	05:26	
Air Blank		0.000	05:27	
Diagnostics Check		OK	05:27	

Data Download Date/Time: 02/07/2022 11:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: ELMORE

DATE/TIME OF TEST: 01/16/2022 00:15:40

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.081	00:10
Air Blank		0.000	00:10
Subject Sample	# 1	0.240	00:11
Breath Volume		1.937	00:11
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.242	00:13
Breath Volume		3.593	00:13
Air Blank		0.000	00:14
Control Test		0.080	00:15
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 02/07/2022 11:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SEKEROVIC

DATE/TIME OF TEST: 01/16/2022 03:42:18

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.081	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.000	03:37
Breath Volume		1.957	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.000	03:40
Breath Volume		1.449	03:40
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 02/07/2022 11:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTHRAN,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FLORY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/16/2022 07:23:40

Diagnostics Check		OK	07:17
Air Blank		0.000	07:17
Control Test		0.081	07:17
Air Blank		0.000	07:18
Subject Sample	# 1	0.172	07:19
Breath Volume		2.558	07:19
Air Blank		0.000	07:19
Air Blank		0.000	07:21
Subject Sample	# 2	0.179	07:22
Breath Volume		2.242	07:22
Air Blank		0.000	07:22
Control Test		0.080	07:23
Air Blank		0.000	07:23
Diagnostics Check		OK	07:23

Data Download Date/Time: 02/07/2022 11:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 01/16/2022 19:30:23

Diagnostics Check		OK	19:27	
Air Blank		0.000	19:27	
Control Test		0.081	19:28	
Air Blank		0.000	19:28	
Subject Sample	# 1		19:29	Subject Test Refused
Breath Volume		0.000	19:29	
Air Blank		0.000	19:29	
Control Test		0.080	19:29	
Air Blank		0.000	19:30	
Diagnostics Check		OK	19:30	

Data Download Date/Time: 02/07/2022 11:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: PARROTT

DATE/TIME OF TEST: 01/16/2022 22:58:24

Diagnostics Check		OK	22:52
Air Blank		0.000	22:52
Control Test		0.081	22:52
Air Blank		0.000	22:53
Subject Sample	# 1	0.102	22:53
Breath Volume		2.503	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.093	22:56
Breath Volume		1.210	22:56
Air Blank		0.000	22:57
Control Test		0.081	22:57
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 02/07/2022 11:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 01/17/2022 04:33:01

Diagnostics Check		OK	04:30	
Air Blank		0.000	04:30	
Control Test		0.082	04:30	
Air Blank		0.000	04:31	
Subject Sample	# 1		04:31	Subject Test Refused
Breath Volume		0.000	04:31	
Air Blank		0.000	04:32	
Control Test		0.081	04:32	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:33	

Data Download Date/Time: 02/07/2022 11:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 01/17/2022 22:36:22

Diagnostics Check		OK	22:33	
Air Blank		0.000	22:33	
Control Test		0.081	22:34	
Air Blank		0.000	22:34	
Subject Sample	# 1		22:34	Subject Test Refused
Breath Volume		0.000	22:34	
Air Blank		0.000	22:35	
Control Test		0.081	22:35	
Air Blank		0.000	22:36	
Diagnostics Check		OK	22:36	

Data Download Date/Time: 02/07/2022 11:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: PROCTOR

DATE/TIME OF TEST: 01/17/2022 23:32:04

Diagnostics Check		OK	23:29	
Air Blank		0.000	23:29	
Control Test		0.081	23:30	
Air Blank		0.000	23:30	
Subject Sample	# 1		23:30	Subject Test Refused
Breath Volume		0.000	23:30	
Air Blank		0.000	23:31	
Control Test		0.081	23:31	
Air Blank		0.000	23:31	
Diagnostics Check		OK	23:32	

Data Download Date/Time: 02/07/2022 11:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: NOLASCO GARCIA

DATE/TIME OF TEST: 01/18/2022 03:13:36

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.082	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.108	03:09
Breath Volume		2.781	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.104	03:12
Breath Volume		3.531	03:12
Air Blank		0.000	03:12
Control Test		0.081	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 02/07/2022 11:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: VYAS

DATE/TIME OF TEST: 01/18/2022 04:51:41

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:49	
Control Test		0.082	04:49	
Air Blank		0.000	04:50	
Subject Sample	# 1		04:50	Subject Test Refused
Breath Volume		0.000	04:50	
Air Blank		0.000	04:50	
Control Test		0.082	04:51	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 02/07/2022 11:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: DELOACH

DATE/TIME OF TEST: 01/18/2022 22:24:45

Diagnostics Check		OK	22:18
Air Blank		0.000	22:19
Control Test		0.081	22:19
Air Blank		0.000	22:19
Subject Sample	# 1	0.106	22:20
Breath Volume		1.816	22:20
Air Blank		0.000	22:20
Air Blank		0.000	22:22
Subject Sample	# 2	0.105	22:23
Breath Volume		1.585	22:23
Air Blank		0.000	22:23
Control Test		0.080	22:24
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 02/07/2022 11:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARTELT,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MCCLANAHAN

DATE/TIME OF TEST: 01/19/2022 05:38:23

Diagnostics Check		OK	05:30
Air Blank		0.000	05:31
Control Test		0.082	05:31
Air Blank		0.000	05:32
Subject Sample	# 1	0.201	05:33
Breath Volume		1.222	05:33
Air Blank		0.000	05:34
Air Blank		0.000	05:36
Subject Sample	# 2	0.193	05:36
Breath Volume		1.156	05:36
Air Blank		0.000	05:37
Control Test		0.080	05:37
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 02/07/2022 11:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 01/19/2022 22:03:58

Diagnostics Check		OK	22:01	
Air Blank		0.000	22:01	
Control Test		0.081	22:01	
Air Blank		0.000	22:02	
Subject Sample	# 1		22:02	Subject Test Refused
Breath Volume		0.000	22:02	
Air Blank		0.000	22:03	
Control Test		0.081	22:03	
Air Blank		0.000	22:03	
Diagnostics Check		OK	22:03	

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: COATS

DATE/TIME OF TEST: 01/19/2022 23:34:21

Diagnostics Check		OK	23:27
Air Blank		0.000	23:28
Control Test		0.081	23:28
Air Blank		0.000	23:29
Subject Sample	# 1	0.061	23:29
Breath Volume		1.785	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.060	23:32
Breath Volume		1.484	23:32
Air Blank		0.000	23:33
Control Test		0.080	23:33
Air Blank		0.000	23:34
Diagnostics Check		OK	23:34

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: GARCIA-CHAJ

DATE/TIME OF TEST: 01/20/2022 00:48:59

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.081	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.124	00:44
Breath Volume		1.757	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:46
Subject Sample	# 2	0.132	00:47
Breath Volume		2.488	00:47
Air Blank		0.000	00:48
Control Test		0.081	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: KASCHAK

DATE/TIME OF TEST: 01/20/2022 01:59:01

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.081	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.133	01:54
Breath Volume		3.515	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:56
Subject Sample	# 2	0.129	01:57
Breath Volume		3.363	01:57
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:59

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/21/2022 01:28:54

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.080	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 1		01:27	Subject Test Refused
Breath Volume		0.000	01:27	
Air Blank		0.000	01:27	
Control Test		0.080	01:28	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/21/2022 05:11:14

Diagnostics Check		OK	05:04
Air Blank		0.000	05:05
Control Test		0.081	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.148	05:06
Breath Volume		1.984	05:06
Air Blank		0.000	05:07
Air Blank		0.000	05:09
Subject Sample	# 2	0.154	05:09
Breath Volume		1.500	05:09
Air Blank		0.000	05:10
Control Test		0.080	05:10
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DILLARD

DATE/TIME OF TEST: 01/22/2022 03:23:22

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:20	
Control Test		0.082	03:21	
Air Blank		0.000	03:21	
Subject Sample	# 1		03:21	Subject Test Refused
Breath Volume		0.000	03:21	
Air Blank		0.000	03:22	
Control Test		0.081	03:22	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COTTON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ORTIZ PENA

DATE/TIME OF TEST: 01/23/2022 05:07:37

Diagnostics Check		OK	05:00
Air Blank		0.000	05:01
Control Test		0.081	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.177	05:02
Breath Volume		2.843	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.182	05:05
Breath Volume		3.417	05:05
Air Blank		0.000	05:06
Control Test		0.081	05:07
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RIBEIRO

DATE/TIME OF TEST: 01/24/2022 03:07:36

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.081	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.216	03:02
Breath Volume		1.503	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.229	03:05
Breath Volume		1.945	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TAYLOR,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 01/25/2022 04:29:37

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.081	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.181	04:25
Breath Volume		2.156	04:25
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.194	04:28
Breath Volume		1.468	04:28
Air Blank		0.000	04:28
Control Test		0.080	04:29
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETEERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: TEMPLE

DATE/TIME OF TEST: 01/25/2022 21:53:23

Diagnostics Check		OK	21:46
Air Blank		0.000	21:47
Control Test		0.081	21:47
Air Blank		0.000	21:48
Subject Sample	# 1	0.029	21:48
Breath Volume		1.773	21:48
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.029	21:51
Breath Volume		1.468	21:51
Air Blank		0.000	21:52
Control Test		0.081	21:52
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 01/25/2022 23:31:01

Diagnostics Check		OK	23:24
Air Blank		0.000	23:25
Control Test		0.082	23:25
Air Blank		0.000	23:25
Subject Sample	# 1	0.153	23:26
Breath Volume		2.253	23:26
Air Blank		0.000	23:27
Air Blank		0.000	23:28
Subject Sample	# 2	0.151	23:29
Breath Volume		2.308	23:29
Air Blank		0.000	23:30
Control Test		0.080	23:30
Air Blank		0.000	23:30
Diagnostics Check		OK	23:31

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 01/26/2022 03:20:25

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.081	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.187	03:15
Breath Volume		1.480	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:17
Subject Sample	# 2	0.179	03:18
Breath Volume		1.902	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 01/27/2022 02:04:17

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.081	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.110	01:59
Breath Volume		1.484	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.116	02:02
Breath Volume		2.414	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: NORDAN

DATE/TIME OF TEST: 01/27/2022 04:06:06

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.082	04:03	
Air Blank		0.000	04:04	
Subject Sample	# 1		04:04	Subject Test Refused
Breath Volume		0.000	04:04	
Air Blank		0.000	04:05	
Control Test		0.081	04:05	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 02/07/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: LUX MORALES

DATE/TIME OF TEST: 01/27/2022 06:18:19

Diagnostics Check		OK	06:12
Air Blank		0.000	06:12
Control Test		0.081	06:12
Air Blank		0.000	06:13
Subject Sample	# 1	0.192	06:13
Breath Volume		1.984	06:13
Air Blank		0.000	06:14
Air Blank		0.000	06:16
Subject Sample	# 2	0.179	06:16
Breath Volume		1.468	06:16
Air Blank		0.000	06:17
Control Test		0.081	06:17
Air Blank		0.000	06:18
Diagnostics Check		OK	06:18

Data Download Date/Time: 02/07/2022 11:16

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: DIAZ COLON

DATE/TIME OF TEST: 01/27/2022 23:40:20

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:37	
Control Test		0.081	23:38	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:38	Subject Test Refused
Breath Volume		0.000	23:38	
Air Blank		0.000	23:39	
Control Test		0.081	23:39	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 02/07/2022 11:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 01/28/2022 01:56:30

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.081	01:50
Air Blank		0.000	01:51
Subject Sample	# 1	0.000	01:51
Breath Volume		2.222	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.000	01:55
Breath Volume		1.957	01:55
Air Blank		0.000	01:55
Control Test		0.081	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 02/07/2022 11:16

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: VANDER-HEYDEN

DATE/TIME OF TEST: 01/28/2022 04:40:50

Diagnostics Check		OK	04:37	
Air Blank		0.000	04:38	
Control Test		0.081	04:38	
Air Blank		0.000	04:39	
Subject Sample	# 1		04:39	Subject Test Refused
Breath Volume		0.000	04:39	
Air Blank		0.000	04:39	
Control Test		0.081	04:40	
Air Blank		0.000	04:40	
Diagnostics Check		OK	04:40	

Data Download Date/Time: 02/07/2022 11:21

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTHRAN,D,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: TOMPKINS

DATE/TIME OF TEST: 01/28/2022 18:47:17

Diagnostics Check		OK	18:40
Air Blank		0.000	18:40
Control Test		0.082	18:41
Air Blank		0.000	18:41
Subject Sample	# 1	0.169	18:42
Breath Volume		2.386	18:42
Air Blank		0.000	18:43
Air Blank		0.000	18:44
Subject Sample	# 2	0.163	18:45
Breath Volume		2.234	18:45
Air Blank		0.000	18:46
Control Test		0.081	18:46
Air Blank		0.000	18:47
Diagnostics Check		OK	18:47

Data Download Date/Time: 02/07/2022 11:21

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: DORWARD

DATE/TIME OF TEST: 01/28/2022 23:22:45

Diagnostics Check		OK	23:16
Air Blank		0.000	23:16
Control Test		0.081	23:17
Air Blank		0.000	23:17
Subject Sample	# 1	0.048	23:18
Breath Volume		2.378	23:18
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.045	23:21
Breath Volume		1.988	23:21
Air Blank		0.000	23:21
Control Test		0.080	23:22
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 02/07/2022 11:21

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: SPURLIN

DATE/TIME OF TEST: 01/29/2022 05:24:30

Diagnostics Check		OK	05:21	
Air Blank		0.000	05:21	
Control Test		0.081	05:22	
Air Blank		0.000	05:22	
Subject Sample	# 1		05:23	Subject Test Refused
Breath Volume		0.000	05:23	
Air Blank		0.000	05:23	
Control Test		0.081	05:23	
Air Blank		0.000	05:24	
Diagnostics Check		OK	05:24	

Data Download Date/Time: 02/07/2022 11:21

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/30/2022 21:54:19

Diagnostics Check		OK	21:48
Air Blank		0.000	21:48
Control Test		0.082	21:49
Air Blank		0.000	21:49
Subject Sample	# 1	0.000	21:50
Breath Volume		2.894	21:50
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.000	21:52
Breath Volume		2.726	21:52
Air Blank		0.000	21:53
Control Test		0.082	21:53
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 02/07/2022 11:21

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,P,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WARD,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/31/2022 16:37:43

Diagnostics Check		OK	16:31
Air Blank		0.000	16:31
Control Test		0.081	16:31
Air Blank		0.000	16:32
Subject Sample	# 1	0.000	16:32
Breath Volume		2.613	16:32
Air Blank		0.000	16:33
Air Blank		0.000	16:35
Subject Sample	# 2	0.000	16:36
Breath Volume		2.554	16:36
Air Blank		0.000	16:36
Control Test		0.081	16:37
Air Blank		0.000	16:37
Diagnostics Check		OK	16:37

Data Download Date/Time: 02/07/2022 11:21

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: WHITE JR

DATE/TIME OF TEST: 01/01/2022 02:45:12

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.079	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.121	02:40
Breath Volume		2.835	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.124	02:43
Breath Volume		3.378	02:43
Air Blank		0.000	02:44
Control Test		0.079	02:44
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 01/27/2022 19:17

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: SHAW JR

DATE/TIME OF TEST: 01/02/2022 01:31:54

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.240	01:23
Breath Volume		1.101	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.268	01:26
Breath Volume		2.761	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 3	0.264	01:30
Breath Volume		1.460	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 01/27/2022 19:17

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DULIAKAS

DATE/TIME OF TEST: 01/03/2022 02:28:41

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.079	02:23
Breath Volume		1.203	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.077	02:27
Breath Volume		1.839	02:27
Air Blank		0.000	02:27
Control Test		0.080	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 01/27/2022 19:17

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: ARMSTRONG II

DATE/TIME OF TEST: 01/04/2022 02:13:37

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:05	
Control Test		0.081	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:09	No Sample Provided
Breath Volume		0.000	02:09	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 2		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:12	
Control Test		0.080	02:13	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 01/27/2022 19:17

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: BRETT

DATE/TIME OF TEST: 01/04/2022 21:43:59

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:35	
Control Test		0.080	21:36	
Air Blank		0.000	21:36	
Subject Sample	# 1		21:39	No Sample Provided
Breath Volume		0.000	21:39	
Air Blank		0.000	21:40	
Air Blank		0.000	21:42	
Subject Sample	# 2		21:42	Subject Test Refused
Breath Volume		0.000	21:42	
Air Blank		0.000	21:43	
Control Test		0.080	21:43	
Air Blank		0.000	21:43	
Diagnostics Check		OK	21:43	

Data Download Date/Time: 01/27/2022 19:17

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HAIRE,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: OSTEEN

DATE/TIME OF TEST: 01/05/2022 18:05:21

Diagnostics Check		OK	17:59
Air Blank		0.000	17:59
Control Test		0.080	17:59
Air Blank		0.000	18:00
Subject Sample	# 1	0.089	18:00
Breath Volume		2.160	18:00
Air Blank		0.000	18:01
Air Blank		0.000	18:03
Subject Sample	# 2	0.090	18:03
Breath Volume		1.695	18:03
Air Blank		0.000	18:04
Control Test		0.080	18:04
Air Blank		0.000	18:05
Diagnostics Check		OK	18:05

Data Download Date/Time: 01/27/2022 19:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: TAVARES

DATE/TIME OF TEST: 01/08/2022 01:29:49

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.080	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.159	01:25
Breath Volume		2.812	01:25
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.166	01:28
Breath Volume		2.519	01:28
Air Blank		0.000	01:28
Control Test		0.080	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 01/27/2022 19:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAERTZ,MARK,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HINSON,N/A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: HINNEBURG

DATE/TIME OF TEST: 01/08/2022 21:36:44

Diagnostics Check		OK	21:28	
Air Blank		0.000	21:28	
Control Test		0.080	21:28	
Air Blank		0.000	21:29	
Subject Sample	# 1		21:32	No Sample Provided
Breath Volume		0.000	21:32	
Air Blank		0.000	21:33	
Air Blank		0.000	21:35	
Subject Sample	# 2		21:35	Subject Test Refused
Breath Volume		0.000	21:35	
Air Blank		0.000	21:35	
Control Test		0.080	21:36	
Air Blank		0.000	21:36	
Diagnostics Check		OK	21:36	

Data Download Date/Time: 01/27/2022 19:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 01/09/2022 01:47:25

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.080	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.077	01:42
Breath Volume		1.605	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.077	01:45
Breath Volume		1.730	01:45
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 01/27/2022 19:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECK,SHEILA,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 01/11/2022 20:09:42

Diagnostics Check		OK	20:01	
Air Blank		0.000	20:01	
Control Test		0.080	20:01	
Air Blank		0.000	20:02	
Subject Sample	# 1		20:05	No Sample Provided
Breath Volume		0.000	20:05	
Air Blank		0.000	20:06	
Air Blank		0.000	20:08	
Subject Sample	# 2		20:08	Subject Test Refused
Breath Volume		0.000	20:08	
Air Blank		0.000	20:08	
Control Test		0.080	20:09	
Air Blank		0.000	20:09	
Diagnostics Check		OK	20:09	

Data Download Date/Time: 01/27/2022 19:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAERTZ,MARK,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ZENTNER

DATE/TIME OF TEST: 01/13/2022 01:47:21

Diagnostics Check		OK	01:34	
Air Blank		0.000	01:34	
Control Test		0.080	01:35	
Air Blank		0.000	01:35	
Subject Sample	# 1	0.148	01:38	Volume Not Met
Breath Volume		0.191	01:38	
Air Blank		0.000	01:39	
Air Blank		0.000	01:41	
Subject Sample	# 2	0.167	01:41	No .020 agreement
Breath Volume		2.203	01:41	
Air Blank		0.000	01:42	
Air Blank		0.000	01:44	
Subject Sample	# 3	0.136	01:45	No .020 agreement
Breath Volume		1.238	01:45	
Air Blank		0.000	01:46	
Control Test		0.080	01:46	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 01/27/2022 19:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAERTZ,MARK,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ZENTNER

DATE/TIME OF TEST: 01/13/2022 02:10:44

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.081	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.156	02:04
Breath Volume		2.730	02:04
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.152	02:09
Breath Volume		2.710	02:09
Air Blank		0.000	02:09
Control Test		0.079	02:10
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 01/27/2022 19:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECK,SHEILA,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BIERCE

DATE/TIME OF TEST: 01/15/2022 01:23:29

Diagnostics Check		OK	01:14
Air Blank		0.000	01:15
Control Test		0.081	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.181	01:17
Breath Volume		1.308	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.181	01:21
Breath Volume		1.812	01:21
Air Blank		0.000	01:22
Control Test		0.080	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 01/27/2022 19:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDERS,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: STARK

DATE/TIME OF TEST: 01/15/2022 03:39:26

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:30	
Control Test		0.081	03:30	
Air Blank		0.000	03:31	
Subject Sample	# 1		03:34	No Sample Provided
Breath Volume		0.000	03:34	
Air Blank		0.000	03:34	
Air Blank		0.000	03:36	
Subject Sample	# 2		03:38	Subject Test Refused
Breath Volume		0.000	03:38	
Air Blank		0.000	03:38	
Control Test		0.080	03:38	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 01/27/2022 19:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECK,SHEILA,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 01/16/2022 01:06:03

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:02	
Control Test		0.081	01:03	
Air Blank		0.000	01:03	
Subject Sample	# 1	0.166	01:04	
Breath Volume		1.367	01:04	
Air Blank			01:04	RFI Detect
Air Blank		0.000	01:05	
Control Test		0.080	01:05	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:06	

Data Download Date/Time: 01/27/2022 19:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECK,S,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 01/16/2022 03:33:21

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:17	
Control Test		0.080	03:17	
Air Blank		0.000	03:18	
Subject Sample	# 1	0.141	03:20	
Breath Volume		1.304	03:20	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 2	0.120	03:26	Volume Not Met
Breath Volume		0.894	03:26	
Air Blank		0.000	03:26	
Air Blank		0.000	03:28	
Subject Sample	# 3	0.120	03:31	Volume Not Met
Breath Volume		0.812	03:31	
Air Blank		0.000	03:32	
Control Test		0.079	03:32	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:33	

Data Download Date/Time: 01/27/2022 19:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDERS,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 01/17/2022 00:05:40

Diagnostics Check		OK	23:56	
Air Blank		0.000	23:56	
Control Test		0.081	23:57	
Air Blank		0.000	23:57	
Subject Sample	# 1		00:00	No Sample Provided
Breath Volume		0.000	00:00	
Air Blank		0.000	00:01	
Air Blank		0.000	00:03	
Subject Sample	# 2		00:04	Subject Test Refused
Breath Volume		0.000	00:04	
Air Blank		0.000	00:04	
Control Test		0.081	00:05	
Air Blank		0.000	00:05	
Diagnostics Check		OK	00:05	

Data Download Date/Time: 01/27/2022 19:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 01/20/2022 22:02:16

Diagnostics Check		OK	21:55
Air Blank		0.000	21:55
Control Test		0.080	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.252	21:57
Breath Volume		2.484	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.239	22:00
Breath Volume		1.242	22:00
Air Blank		0.000	22:01
Control Test		0.079	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 01/27/2022 19:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECK,S,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: GILLEY

DATE/TIME OF TEST: 01/21/2022 02:19:13

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:10	
Control Test		0.080	02:11	
Air Blank		0.000	02:11	
Subject Sample	# 1		02:14	No Sample Provided
Breath Volume		0.000	02:14	
Air Blank		0.000	02:15	
Air Blank		0.000	02:17	
Subject Sample	# 2		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:18	
Control Test		0.080	02:18	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 01/27/2022 19:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SUBIRATS

DATE/TIME OF TEST: 01/22/2022 00:52:20

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.146	00:47
Breath Volume		3.355	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:49
Subject Sample	# 2	0.147	00:50
Breath Volume		2.449	00:50
Air Blank		0.000	00:51
Control Test		0.081	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 01/27/2022 19:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: PADGETT

DATE/TIME OF TEST: 01/22/2022 20:54:27

Diagnostics Check		OK	20:47
Air Blank		0.000	20:48
Control Test		0.080	20:48
Air Blank		0.000	20:48
Subject Sample	# 1	0.000	20:49
Breath Volume		2.007	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.000	20:53
Breath Volume		2.000	20:53
Air Blank		0.000	20:53
Control Test		0.080	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 01/27/2022 19:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MAJOR

DATE/TIME OF TEST: 01/23/2022 03:10:40

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:02	
Control Test		0.081	03:02	
Air Blank		0.000	03:03	
Subject Sample	# 1		03:06	No Sample Provided
Breath Volume		0.000	03:06	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 2		03:09	Subject Test Refused
Breath Volume		0.000	03:09	
Air Blank		0.000	03:09	
Control Test		0.081	03:10	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 01/27/2022 19:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAERTZ,MARK,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: BOTELHO

DATE/TIME OF TEST: 01/23/2022 20:20:37

Diagnostics Check		OK	20:17	
Air Blank		0.000	20:18	
Control Test		0.081	20:18	
Air Blank		0.000	20:18	
Subject Sample	# 1	0.127	20:19	
Breath Volume		2.644	20:19	
Air Blank			20:20	RFI Detect
Air Blank			20:20	RFI Detect
Air Blank		0.000	20:20	

Data Download Date/Time: 01/27/2022 19:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAERTZ,MARK,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: BOTELHO

DATE/TIME OF TEST: 01/23/2022 20:29:13

Diagnostics Check		OK	20:22
Air Blank		0.000	20:22
Control Test		0.080	20:22
Air Blank		0.000	20:23
Subject Sample	# 1	0.129	20:24
Breath Volume		2.640	20:24
Air Blank		0.000	20:25
Air Blank		0.000	20:26
Subject Sample	# 2	0.127	20:27
Breath Volume		2.121	20:27
Air Blank		0.000	20:28
Control Test		0.080	20:28
Air Blank		0.000	20:29
Diagnostics Check		OK	20:29

Data Download Date/Time: 01/27/2022 19:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAERTZ,MARK,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: WOODRUM

DATE/TIME OF TEST: 01/23/2022 23:29:06

Diagnostics Check		OK	23:19	
Air Blank		0.000	23:20	
Control Test		0.080	23:20	
Air Blank		0.000	23:21	
Subject Sample	# 1		23:24	No Sample Provided
Breath Volume		0.000	23:24	
Air Blank		0.000	23:24	
Air Blank		0.000	23:26	
Subject Sample	# 2		23:27	Subject Test Refused
Breath Volume		0.000	23:27	
Air Blank		0.000	23:28	
Control Test		0.081	23:28	
Air Blank		0.000	23:29	
Diagnostics Check		OK	23:29	

Data Download Date/Time: 01/27/2022 19:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 01/25/2022 21:28:38

Diagnostics Check		OK	21:20	
Air Blank		0.000	21:20	
Control Test		0.081	21:20	
Air Blank		0.000	21:21	
Subject Sample	# 1		21:24	No Sample Provided
Breath Volume		0.000	21:24	
Air Blank		0.000	21:24	
Air Blank		0.000	21:26	
Subject Sample	# 2		21:27	Subject Test Refused
Breath Volume		0.000	21:27	
Air Blank		0.000	21:27	
Control Test		0.081	21:28	
Air Blank		0.000	21:28	
Diagnostics Check		OK	21:28	

Data Download Date/Time: 01/27/2022 19:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECK,SHEILA,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FLORY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: TISDELLE

DATE/TIME OF TEST: 01/28/2022 20:01:52

Diagnostics Check		OK	19:55
Air Blank		0.000	19:56
Control Test		0.080	19:56
Air Blank		0.000	19:56
Subject Sample	# 1	0.333	19:57
Breath Volume		2.742	19:57
Air Blank		0.000	19:58
Air Blank		0.000	19:59
Subject Sample	# 2	0.328	20:00
Breath Volume		2.847	20:00
Air Blank		0.000	20:00
Control Test		0.079	20:01
Air Blank		0.000	20:01
Diagnostics Check		OK	20:01

Data Download Date/Time: 02/28/2022 09:47

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GUNN

DATE/TIME OF TEST: 01/29/2022 03:34:35

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:26	
Control Test		0.080	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 1		03:30	No Sample Provided
Breath Volume		0.000	03:30	
Air Blank		0.000	03:30	
Air Blank		0.000	03:32	
Subject Sample	# 2		03:33	Subject Test Refused
Breath Volume		0.000	03:33	
Air Blank		0.000	03:33	
Control Test		0.080	03:34	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 02/28/2022 09:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSSMAN,R,K
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: NEWTON,N
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:46

SUBJECT'S LAST NAME: CANN

DATE/TIME OF TEST: 01/29/2022 14:20:01

Diagnostics Check		OK	14:11	
Air Blank		0.000	14:11	
Control Test		0.081	14:12	
Air Blank		0.000	14:12	
Subject Sample	# 1		14:15	No Sample Provided
Breath Volume		0.000	14:15	
Air Blank		0.000	14:16	
Air Blank		0.000	14:18	
Subject Sample	# 2		14:18	Subject Test Refused
Breath Volume		0.000	14:18	
Air Blank		0.000	14:19	
Control Test		0.081	14:19	
Air Blank		0.000	14:19	
Diagnostics Check		OK	14:20	

Data Download Date/Time: 02/28/2022 09:47

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDERS,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: COMPTON

DATE/TIME OF TEST: 01/29/2022 23:09:46

Diagnostics Check		OK	23:03
Air Blank		0.000	23:03
Control Test		0.080	23:04
Air Blank		0.000	23:04
Subject Sample	# 1	0.176	23:05
Breath Volume		3.304	23:05
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.181	23:08
Breath Volume		2.359	23:08
Air Blank		0.000	23:08
Control Test		0.080	23:09
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 02/28/2022 09:47

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECK,S,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DEES

DATE/TIME OF TEST: 01/30/2022 03:02:44

Diagnostics Check		OK	02:54	
Air Blank		0.000	02:54	
Control Test		0.081	02:55	
Air Blank		0.000	02:55	
Subject Sample	# 1		02:58	No Sample Provided
Breath Volume		0.000	02:58	
Air Blank		0.000	02:59	
Air Blank		0.000	03:01	
Subject Sample	# 2		03:01	Subject Test Refused
Breath Volume		0.000	03:01	
Air Blank		0.000	03:01	
Control Test		0.080	03:02	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 02/28/2022 09:47

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,ALEXANDER,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: PEREZ MANUEL

DATE/TIME OF TEST: 01/20/2022 01:35:12

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.218	01:30
Breath Volume		1.746	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.215	01:33
Breath Volume		2.226	01:33
Air Blank		0.000	01:34
Control Test		0.080	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 01/31/2022 10:37

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ERIC,R
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 01/22/2022 00:23:26

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.080	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.088	00:19
Breath Volume		2.054	00:19
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.086	00:21
Breath Volume		2.515	00:21
Air Blank		0.000	00:22
Control Test		0.080	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 01/31/2022 10:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,ALEXANDER,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: THOMAS,ALEXANDER
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: LITTMAN

DATE/TIME OF TEST: 01/30/2022 23:37:15

Diagnostics Check		OK	23:34	
Air Blank		0.000	23:34	
Control Test		0.081	23:35	
Air Blank		0.000	23:35	
Subject Sample	# 1		23:35	Subject Test Refused
Breath Volume		0.000	23:35	
Air Blank		0.000	23:36	
Control Test		0.081	23:36	
Air Blank		0.000	23:37	
Diagnostics Check		OK	23:37	

Data Download Date/Time: 01/31/2022 10:38

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: SCHAFER

DATE/TIME OF TEST: 01/01/2022 02:19:47

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:11	
Control Test		0.079	02:12	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:15	No Sample Provided
Breath Volume		0.000	02:15	
Air Blank		0.000	02:16	
Air Blank		0.000	02:18	
Subject Sample	# 2		02:18	Subject Test Refused
Breath Volume		0.000	02:18	
Air Blank		0.000	02:18	
Control Test		0.079	02:19	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 01/07/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTENSEN,S,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KOBONIN,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: MOSES

DATE/TIME OF TEST: 01/05/2022 05:09:58

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.079	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.142	05:05
Breath Volume		1.265	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:07
Subject Sample	# 2	0.153	05:08
Breath Volume		1.718	05:08
Air Blank		0.000	05:09
Control Test		0.078	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 01/07/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:37

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 01/05/2022 10:05:38

Diagnostics Check		OK	09:58
Air Blank		0.000	09:59
Control Test		0.078	09:59
Air Blank		0.000	09:59
Subject Sample	# 1	0.000	10:01
Breath Volume		1.855	10:01
Air Blank		0.000	10:01
Air Blank		0.000	10:03
Subject Sample	# 2	0.000	10:04
Breath Volume		3.238	10:04
Air Blank		0.000	10:04
Control Test		0.079	10:05
Air Blank		0.000	10:05
Diagnostics Check		OK	10:05

Data Download Date/Time: 01/07/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: JUSTICE

DATE/TIME OF TEST: 01/06/2022 02:12:02

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.078	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.000	02:07
Breath Volume		2.593	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.000	02:10
Breath Volume		2.722	02:10
Air Blank		0.000	02:11
Control Test		0.078	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:12

Data Download Date/Time: 01/07/2022 11:31

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: GUGINO

DATE/TIME OF TEST: 01/07/2022 00:33:56

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.079	00:28
Air Blank		0.000	00:28
Subject Sample	# 1	0.095	00:29
Breath Volume		1.867	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.097	00:32
Breath Volume		1.718	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 01/07/2022 11:31

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MATTHYS

DATE/TIME OF TEST: 01/07/2022 01:07:40

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.158	01:03
Breath Volume		1.101	01:03
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.168	01:06
Breath Volume		1.179	01:06
Air Blank		0.000	01:06
Control Test		0.078	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 01/07/2022 11:31

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/07/2022 14:22:18

Diagnostics Check		OK	14:16
Air Blank		0.000	14:16
Control Test		0.080	14:17
Air Blank		0.000	14:17
Subject Sample	# 1	0.000	14:18
Breath Volume		3.617	14:18
Air Blank		0.000	14:18
Air Blank		0.000	14:20
Subject Sample	# 2	0.000	14:20
Breath Volume		3.722	14:20
Air Blank		0.000	14:21
Control Test		0.080	14:21
Air Blank		0.000	14:22
Diagnostics Check		OK	14:22

Data Download Date/Time: 02/03/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: MUNLEY

DATE/TIME OF TEST: 01/08/2022 05:25:58

Diagnostics Check		OK	05:16
Air Blank		0.000	05:17
Control Test		0.080	05:17
Air Blank		0.000	05:17
Subject Sample	# 1	0.147	05:18
Breath Volume		1.234	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:20
Subject Sample	# 2	0.169	05:21
Breath Volume		2.144	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:23
Subject Sample	# 3	0.157	05:24
Breath Volume		1.917	05:24
Air Blank		0.000	05:25
Control Test		0.077	05:25
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 02/03/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: EBERT

DATE/TIME OF TEST: 01/14/2022 01:44:38

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.080	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.130	01:40
Breath Volume		1.679	01:40
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.128	01:43
Breath Volume		2.121	01:43
Air Blank		0.000	01:43
Control Test		0.078	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 02/03/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: GONSALVES

DATE/TIME OF TEST: 01/15/2022 04:25:16

Diagnostics Check		OK	04:16	
Air Blank		0.000	04:17	
Control Test		0.080	04:17	
Air Blank		0.000	04:18	
Subject Sample	# 1		04:21	No Sample Provided
Breath Volume		0.000	04:21	
Air Blank		0.000	04:21	
Air Blank		0.000	04:23	
Subject Sample	# 2		04:23	Subject Test Refused
Breath Volume		0.000	04:23	
Air Blank		0.000	04:24	
Control Test		0.080	04:24	
Air Blank		0.000	04:25	
Diagnostics Check		OK	04:25	

Data Download Date/Time: 02/03/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PRIDE

DATE/TIME OF TEST: 01/16/2022 02:44:27

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.080	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.000	02:40
Breath Volume		2.210	02:40
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.000	02:43
Breath Volume		1.980	02:43
Air Blank		0.000	02:43
Control Test		0.079	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 02/03/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: DORING

DATE/TIME OF TEST: 01/16/2022 18:24:27

Diagnostics Check		OK	18:16	
Air Blank		0.000	18:16	
Control Test		0.079	18:16	
Air Blank		0.000	18:17	
Subject Sample	# 1		18:20	No Sample Provided
Breath Volume		0.000	18:20	
Air Blank		0.000	18:20	
Air Blank		0.000	18:22	
Subject Sample	# 2		18:23	Subject Test Refused
Breath Volume		0.000	18:23	
Air Blank		0.000	18:23	
Control Test		0.079	18:23	
Air Blank		0.000	18:24	
Diagnostics Check		OK	18:24	

Data Download Date/Time: 02/03/2022 11:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: RAWLS

DATE/TIME OF TEST: 01/16/2022 21:40:08

Diagnostics Check		OK	21:37	
Air Blank		0.000	21:37	
Control Test		0.079	21:38	
Air Blank		0.000	21:38	
Subject Sample	# 1		21:38	RFI Detect
Breath Volume		0.000	21:38	
Air Blank		0.000	21:39	
Control Test		0.080	21:39	
Air Blank		0.000	21:40	
Diagnostics Check		OK	21:40	

Data Download Date/Time: 02/03/2022 11:16

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: RAWLS

DATE/TIME OF TEST: 01/16/2022 21:57:23

Diagnostics Check		OK	21:50
Air Blank		0.000	21:51
Control Test		0.079	21:51
Air Blank		0.000	21:52
Subject Sample	# 1	0.000	21:52
Breath Volume		1.644	21:52
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.000	21:56
Breath Volume		1.464	21:56
Air Blank		0.000	21:56
Control Test		0.080	21:56
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 02/03/2022 11:16

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:59

SUBJECT'S LAST NAME: HUELL

DATE/TIME OF TEST: 01/19/2022 18:25:45

Diagnostics Check		OK	18:19
Air Blank		0.000	18:20
Control Test		0.080	18:20
Air Blank		0.000	18:21
Subject Sample	# 1	0.000	18:21
Breath Volume		1.367	18:21
Air Blank		0.000	18:21
Air Blank		0.000	18:23
Subject Sample	# 2	0.000	18:24
Breath Volume		2.402	18:24
Air Blank		0.000	18:24
Control Test		0.079	18:25
Air Blank		0.000	18:25
Diagnostics Check		OK	18:25

Data Download Date/Time: 02/03/2022 11:16

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:26

SUBJECT'S LAST NAME: RUGOFF

DATE/TIME OF TEST: 01/21/2022 09:35:30

Diagnostics Check		OK	09:32	
Air Blank		0.000	09:33	
Control Test		0.079	09:33	
Air Blank		0.000	09:34	
Subject Sample	# 1		09:35	Sequence Aborted
Breath Volume		0.000	09:35	
Air Blank		0.000	09:35	

Data Download Date/Time: 02/03/2022 11:16

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 01/30/2022 19:11:58

Diagnostics Check		OK	19:08	
Air Blank		0.000	19:08	
Control Test		0.080	19:09	
Air Blank		0.000	19:09	
Subject Sample	# 1		19:10	Subject Test Refused
Breath Volume		0.000	19:10	
Air Blank		0.000	19:11	
Control Test		0.080	19:11	
Air Blank		0.000	19:11	
Diagnostics Check		OK	19:11	

Data Download Date/Time: 02/03/2022 11:16

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: O'HANLON

DATE/TIME OF TEST: 01/30/2022 21:26:39

Diagnostics Check		OK	21:20
Air Blank		0.000	21:20
Control Test		0.080	21:21
Air Blank		0.000	21:21
Subject Sample	# 1	0.113	21:22
Breath Volume		2.773	21:22
Air Blank		0.000	21:22
Air Blank		0.000	21:24
Subject Sample	# 2	0.120	21:25
Breath Volume		2.804	21:25
Air Blank		0.000	21:25
Control Test		0.080	21:26
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 02/03/2022 11:16

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BISHOP,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: LOCKETT

DATE/TIME OF TEST: 01/31/2022 22:56:42

Diagnostics Check		OK	22:53	
Air Blank		0.000	22:54	
Control Test		0.080	22:54	
Air Blank		0.000	22:54	
Subject Sample	# 1		22:55	Subject Test Refused
Breath Volume		0.000	22:55	
Air Blank		0.000	22:55	
Control Test		0.079	22:56	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 02/03/2022 11:16

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: MACARTHUR

DATE/TIME OF TEST: 01/01/2022 00:17:09

Diagnostics Check		OK	00:01	
Air Blank		0.000	00:02	
Control Test		0.080	00:02	
Air Blank		0.000	00:02	
Subject Sample	# 1	0.101	00:04	
Breath Volume		1.820	00:04	
Air Blank		0.000	00:05	
Air Blank		0.000	00:06	
Subject Sample	# 2	0.096	00:10	Volume Not Met
Breath Volume		0.070	00:10	
Air Blank		0.000	00:10	
Air Blank		0.000	00:12	
Subject Sample	# 3	0.089	00:15	Volume Not Met
Breath Volume		0.312	00:15	
Air Blank		0.000	00:16	
Control Test		0.079	00:16	
Air Blank		0.000	00:17	
Diagnostics Check		OK	00:17	

Data Download Date/Time: 01/25/2022 15:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PICKETT,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: FARRIS

DATE/TIME OF TEST: 01/01/2022 01:47:52

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.080	01:42
Breath Volume		2.164	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.061	01:46
Breath Volume		1.902	01:46
Air Blank		0.000	01:46
Control Test		0.079	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 01/25/2022 15:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,J,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GOTHE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/02/2022 00:18:51

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.080	00:12
Air Blank		0.000	00:13
Subject Sample	# 1	0.000	00:14
Breath Volume		1.394	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.000	00:17
Breath Volume		1.382	00:17
Air Blank		0.000	00:17
Control Test		0.079	00:18
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 01/25/2022 15:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/03/2022 03:14:28

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.000	03:09
Breath Volume		3.621	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.000	03:13
Breath Volume		3.617	03:13
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 01/25/2022 15:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIPPINS,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WEIGLE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: WERNER

DATE/TIME OF TEST: 01/08/2022 21:56:14

Diagnostics Check		OK	21:49
Air Blank		0.000	21:49
Control Test		0.080	21:50
Air Blank		0.000	21:50
Subject Sample	# 1	0.130	21:51
Breath Volume		2.656	21:51
Air Blank		0.000	21:52
Air Blank		0.000	21:54
Subject Sample	# 2	0.129	21:54
Breath Volume		2.480	21:54
Air Blank		0.000	21:55
Control Test		0.078	21:55
Air Blank		0.000	21:56
Diagnostics Check		OK	21:56

Data Download Date/Time: 01/25/2022 15:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JACKSON,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 01/12/2022 00:47:06

Diagnostics Check		OK	00:36	
Air Blank		0.000	00:37	
Control Test		0.081	00:37	
Air Blank		0.000	00:38	
Subject Sample	# 1	0.103	00:41	Volume Not Met
Breath Volume		0.089	00:41	
Air Blank		0.000	00:41	
Air Blank		0.000	00:43	
Subject Sample	# 2		00:45	Subject Test Refused
Breath Volume		0.015	00:45	
Air Blank		0.000	00:46	
Control Test		0.080	00:46	
Air Blank		0.000	00:47	
Diagnostics Check		OK	00:47	

Data Download Date/Time: 01/25/2022 15:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CARPENTER,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 01/13/2022 00:01:34

Diagnostics Check		OK	23:58	
Air Blank		0.000	23:59	
Control Test		0.080	23:59	
Air Blank		0.000	00:00	
Subject Sample	# 1		00:00	Subject Test Refused
Breath Volume		0.000	00:00	
Air Blank		0.000	00:00	
Control Test		0.080	00:00	
Air Blank		0.000	00:01	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 01/25/2022 15:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAL,JACOB,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WALTON,CHRIS
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:58

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 01/14/2022 13:31:00

Diagnostics Check		OK	13:27	
Air Blank		0.000	13:28	
Control Test		0.080	13:28	
Air Blank		0.000	13:28	
Subject Sample	# 1		13:29	Subject Test Refused
Breath Volume		0.000	13:29	
Air Blank		0.000	13:30	
Control Test		0.080	13:30	
Air Blank		0.000	13:30	
Diagnostics Check		OK	13:31	

Data Download Date/Time: 01/25/2022 15:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIPPINS,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: SCHWARTZ

DATE/TIME OF TEST: 01/14/2022 20:48:29

Diagnostics Check		OK	20:45	
Air Blank		0.000	20:46	
Control Test		0.081	20:46	
Air Blank		0.000	20:46	
Subject Sample	# 1		20:47	Subject Test Refused
Breath Volume		0.000	20:47	
Air Blank		0.000	20:47	
Control Test		0.080	20:47	
Air Blank		0.000	20:48	
Diagnostics Check		OK	20:48	

Data Download Date/Time: 01/25/2022 15:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 01/15/2022 05:29:36

Diagnostics Check		OK	05:23
Air Blank		0.000	05:23
Control Test		0.080	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.161	05:25
Breath Volume		3.035	05:25
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.159	05:28
Breath Volume		2.792	05:28
Air Blank		0.000	05:28
Control Test		0.079	05:29
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 01/25/2022 15:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAL,JACOB,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SMITH,BOBBY
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:46

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 01/15/2022 08:21:02

Diagnostics Check		OK	08:13
Air Blank		0.000	08:14
Control Test		0.079	08:14
Air Blank		0.000	08:14
Subject Sample	# 1	0.205	08:15
Breath Volume		2.097	08:15
Air Blank		0.000	08:15
Air Blank		0.000	08:17
Subject Sample	# 2	0.203	08:19
Breath Volume		3.070	08:19
Air Blank		0.000	08:20
Control Test		0.080	08:20
Air Blank		0.000	08:20
Diagnostics Check		OK	08:21

Data Download Date/Time: 01/25/2022 15:41

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNEAD,CODY,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: FITZPATRICK,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: NESMITH

DATE/TIME OF TEST: 01/15/2022 18:29:44

Diagnostics Check		OK	18:23
Air Blank		0.000	18:23
Control Test		0.079	18:23
Air Blank		0.000	18:24
Subject Sample	# 1	0.000	18:25
Breath Volume		4.152	18:25
Air Blank		0.000	18:25
Air Blank		0.000	18:27
Subject Sample	# 2	0.000	18:28
Breath Volume		4.253	18:28
Air Blank		0.000	18:28
Control Test		0.079	18:29
Air Blank		0.000	18:29
Diagnostics Check		OK	18:29

Data Download Date/Time: 01/25/2022 15:41

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JACKSON,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 01/15/2022 23:21:04

Diagnostics Check		OK	23:14
Air Blank		0.000	23:15
Control Test		0.080	23:15
Air Blank		0.000	23:15
Subject Sample	# 1	0.152	23:16
Breath Volume		1.203	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.166	23:19
Breath Volume		1.367	23:19
Air Blank		0.000	23:20
Control Test		0.080	23:20
Air Blank		0.000	23:20
Diagnostics Check		OK	23:21

Data Download Date/Time: 01/25/2022 15:41

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 01/16/2022 00:03:08

Diagnostics Check		OK	23:56
Air Blank		0.000	23:57
Control Test		0.079	23:57
Air Blank		0.000	23:57
Subject Sample	# 2	0.224	23:58
Breath Volume		2.000	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.218	00:01
Breath Volume		1.777	00:01
Air Blank		0.000	00:02
Control Test		0.079	00:02
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 01/25/2022 15:42

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/17/2022 01:07:27

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:05	
Control Test		0.081	01:05	
Air Blank		0.000	01:05	
Subject Sample	# 1		01:06	Subject Test Refused
Breath Volume		0.000	01:06	
Air Blank		0.000	01:06	
Control Test		0.081	01:06	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:07	

Data Download Date/Time: 01/25/2022 15:42

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIPPINS,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CARPENTER,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: MERCHANT

DATE/TIME OF TEST: 01/17/2022 19:02:21

Diagnostics Check		OK	18:56
Air Blank		0.000	18:56
Control Test		0.081	18:57
Air Blank		0.000	18:57
Subject Sample	# 1	0.208	18:57
Breath Volume		1.707	18:57
Air Blank		0.000	18:58
Air Blank		0.000	19:00
Subject Sample	# 2	0.207	19:00
Breath Volume		1.105	19:00
Air Blank		0.000	19:01
Control Test		0.080	19:01
Air Blank		0.000	19:02
Diagnostics Check		OK	19:02

Data Download Date/Time: 01/25/2022 15:42

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,J,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ROWELL

DATE/TIME OF TEST: 01/20/2022 00:36:49

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.081	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.000	00:32
Breath Volume		2.859	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.000	00:35
Breath Volume		3.226	00:35
Air Blank		0.000	00:35
Control Test		0.081	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 01/25/2022 15:42

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,J,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 01/29/2022 03:40:26

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.080	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.151	03:35
Breath Volume		2.582	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.161	03:38
Breath Volume		2.425	03:38
Air Blank		0.000	03:39
Control Test		0.079	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 03/01/2022 21:39

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIPPINS,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JACKSON,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: HARDEN

DATE/TIME OF TEST: 01/29/2022 04:32:09

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:29	
Control Test		0.081	04:30	
Air Blank		0.000	04:30	
Subject Sample	# 1		04:30	Subject Test Refused
Breath Volume		0.000	04:30	
Air Blank		0.000	04:31	
Control Test		0.081	04:31	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 03/01/2022 21:39

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WALTON,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:19

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 01/29/2022 08:47:45

Diagnostics Check		OK	08:40
Air Blank		0.000	08:41
Control Test		0.080	08:41
Air Blank		0.000	08:42
Subject Sample	# 1	0.155	08:42
Breath Volume		1.687	08:42
Air Blank		0.000	08:43
Air Blank		0.000	08:45
Subject Sample	# 2	0.146	08:46
Breath Volume		1.718	08:46
Air Blank		0.000	08:46
Control Test		0.080	08:47
Air Blank		0.000	08:47
Diagnostics Check		OK	08:47

Data Download Date/Time: 03/01/2022 21:39

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,J,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 01/30/2022 01:41:22

Diagnostics Check		OK	01:28	
Air Blank		0.000	01:29	
Control Test		0.081	01:29	
Air Blank		0.000	01:30	
Subject Sample	# 1	0.162	01:31	
Breath Volume		1.625	01:31	
Air Blank		0.000	01:32	
Air Blank		0.000	01:34	
Subject Sample	# 2		01:37	No Sample Provided
Breath Volume		0.000	01:37	
Air Blank		0.000	01:37	
Air Blank		0.000	01:39	
Subject Sample	# 3		01:40	Subject Test Refused
Breath Volume		0.000	01:40	
Air Blank		0.000	01:40	
Control Test		0.082	01:40	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 03/01/2022 21:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 01/05/2022 01:22:39

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:11	
Control Test		0.081	01:12	
Air Blank		0.000	01:12	
Subject Sample	# 1		01:15	No Sample Provided
Breath Volume		0.000	01:15	
Air Blank		0.000	01:16	
Air Blank		0.000	01:18	
Subject Sample	# 2		01:21	No Sample Provided
Breath Volume		0.000	01:21	
Air Blank		0.000	01:21	
Control Test		0.081	01:22	
Air Blank		0.000	01:22	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 01/11/2022 16:43

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: BRAGG

DATE/TIME OF TEST: 01/08/2022 00:44:19

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.081	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.103	00:39
Breath Volume		2.386	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.098	00:42
Breath Volume		2.089	00:42
Air Blank		0.000	00:43
Control Test		0.081	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 01/11/2022 16:43

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: ANDRADE

DATE/TIME OF TEST: 01/22/2022 21:25:49

Diagnostics Check		OK	21:18
Air Blank		0.000	21:19
Control Test		0.080	21:19
Air Blank		0.000	21:19
Subject Sample	# 1	0.177	21:21
Breath Volume		3.062	21:21
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.167	21:24
Breath Volume		2.433	21:24
Air Blank		0.000	21:24
Control Test		0.081	21:25
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

Data Download Date/Time: 02/15/2022 16:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ROCENDO

DATE/TIME OF TEST: 01/23/2022 03:28:44

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.081	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.223	03:24
Breath Volume		2.320	03:24
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.216	03:27
Breath Volume		2.375	03:27
Air Blank		0.000	03:27
Control Test		0.081	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 02/15/2022 16:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/23/2022 06:25:42

Diagnostics Check		OK	06:18
Air Blank		0.000	06:19
Control Test		0.080	06:19
Air Blank		0.000	06:20
Subject Sample	# 1	0.165	06:20
Breath Volume		1.507	06:20
Air Blank		0.000	06:21
Air Blank		0.000	06:23
Subject Sample	# 2	0.172	06:24
Breath Volume		1.960	06:24
Air Blank		0.000	06:24
Control Test		0.081	06:25
Air Blank		0.000	06:25
Diagnostics Check		OK	06:25

Data Download Date/Time: 02/15/2022 16:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: JENS

DATE/TIME OF TEST: 01/29/2022 18:14:07

Diagnostics Check		OK	18:07
Air Blank		0.000	18:07
Control Test		0.081	18:08
Air Blank		0.000	18:08
Subject Sample	# 1	0.219	18:09
Breath Volume		4.187	18:09
Air Blank		0.000	18:10
Air Blank		0.000	18:11
Subject Sample	# 2	0.217	18:12
Breath Volume		4.085	18:12
Air Blank		0.000	18:13
Control Test		0.082	18:13
Air Blank		0.000	18:14
Diagnostics Check		OK	18:14

Data Download Date/Time: 02/15/2022 16:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD,RONALD,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD,RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: BRIDWELL

DATE/TIME OF TEST: 01/01/2022 00:42:05

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.079	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.250	00:37
Breath Volume		4.488	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:39
Subject Sample	# 2	0.251	00:40
Breath Volume		3.839	00:40
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 01/19/2022 17:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,C,D
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,J
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 01/07/2022 00:44:47

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.078	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.173	00:39
Breath Volume		1.625	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.172	00:43
Breath Volume		1.425	00:43
Air Blank		0.000	00:43
Control Test		0.080	00:44
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 01/19/2022 17:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD,RONALD,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD,RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/09/2022 03:30:29

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.078	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.201	03:25
Breath Volume		2.933	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.212	03:28
Breath Volume		3.042	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 01/19/2022 17:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,C,D
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,J
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MURRETT

DATE/TIME OF TEST: 01/14/2022 03:06:58

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:59	
Control Test		0.079	02:59	
Air Blank		0.000	02:59	
Subject Sample	# 1		03:02	No Sample Provided
Breath Volume		0.000	03:02	
Air Blank		0.000	03:03	
Air Blank		0.000	03:05	
Subject Sample	# 2		03:05	Subject Test Refused
Breath Volume		0.000	03:05	
Air Blank		0.000	03:06	
Control Test		0.080	03:06	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 01/19/2022 17:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DUPONT

DATE/TIME OF TEST: 01/15/2022 04:06:37

Diagnostics Check		OK	03:59
Air Blank		0.000	03:59
Control Test		0.079	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.153	04:01
Breath Volume		1.214	04:01
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.145	04:05
Breath Volume		1.988	04:05
Air Blank		0.000	04:05
Control Test		0.081	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 01/19/2022 17:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: SEARS,C
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: HARDEE

DATE/TIME OF TEST: 01/15/2022 20:16:41

Diagnostics Check		OK	20:11	
Air Blank		0.000	20:12	
Control Test		0.079	20:12	
Air Blank		0.000	20:13	
Subject Sample	# 1	0.205	20:15	
Breath Volume		1.402	20:15	
Air Blank			20:15	Purge Fail
Air Blank			20:16	Purge Fail

Data Download Date/Time: 01/19/2022 17:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: SEARS,C
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: HARDEE

DATE/TIME OF TEST: 01/15/2022 20:26:45

Diagnostics Check		OK	20:20
Air Blank		0.000	20:20
Control Test		0.079	20:20
Air Blank		0.000	20:21
Subject Sample	# 1	0.211	20:22
Breath Volume		1.718	20:22
Air Blank		0.000	20:22
Air Blank		0.000	20:24
Subject Sample	# 2	0.203	20:25
Breath Volume		1.769	20:25
Air Blank		0.000	20:25
Control Test		0.079	20:26
Air Blank		0.000	20:26
Diagnostics Check		OK	20:26

Data Download Date/Time: 01/19/2022 17:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD,RONALD,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD,RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: SAENZ

DATE/TIME OF TEST: 01/21/2022 03:02:33

Diagnostics Check		OK	02:59	
Air Blank		0.000	03:00	
Control Test		0.079	03:00	
Air Blank		0.000	03:01	
Subject Sample	# 1		03:01	Subject Test Refused
Breath Volume		0.000	03:01	
Air Blank		0.000	03:01	
Control Test		0.079	03:01	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 02/23/2022 11:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD,RONALD,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD,RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 01/30/2022 02:51:58

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:49	
Control Test		0.080	02:49	
Air Blank		0.000	02:50	
Subject Sample	# 1		02:50	Subject Test Refused
Breath Volume		0.000	02:50	
Air Blank		0.000	02:51	
Control Test		0.080	02:51	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 02/23/2022 11:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD,RONALD,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD,RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: GOSSETT

DATE/TIME OF TEST: 01/31/2022 04:02:53

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:00	
Control Test		0.080	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 1		04:01	Subject Test Refused
Breath Volume		0.000	04:01	
Air Blank		0.000	04:01	
Control Test		0.080	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 02/23/2022 11:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEBSON,BRET,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: HEBSON,BRET
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 12/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 09:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: CARRENO

DATE/TIME OF TEST: 01/07/2022 10:30:28

Diagnostics Check		OK	10:24
Air Blank		0.000	10:24
Control Test		0.080	10:24
Air Blank		0.000	10:25
Subject Sample	# 1	0.000	10:26
Breath Volume		2.152	10:26
Air Blank		0.000	10:26
Air Blank		0.000	10:28
Subject Sample	# 2	0.000	10:29
Breath Volume		1.394	10:29
Air Blank		0.000	10:29
Control Test		0.080	10:29
Air Blank		0.000	10:30
Diagnostics Check		OK	10:30

Data Download Date/Time: 01/28/2022 10:27

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEBSON,BRET,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: HEBSON,BRET
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 12/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 11:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:52

SUBJECT'S LAST NAME: PETIT-FRERE

DATE/TIME OF TEST: 01/07/2022 12:19:15

Diagnostics Check		OK	12:13
Air Blank		0.000	12:13
Control Test		0.082	12:13
Air Blank		0.000	12:14
Subject Sample	# 1	0.000	12:14
Breath Volume		2.460	12:14
Air Blank		0.000	12:15
Air Blank		0.000	12:17
Subject Sample	# 2	0.000	12:17
Breath Volume		1.644	12:17
Air Blank		0.000	12:18
Control Test		0.082	12:18
Air Blank		0.000	12:19
Diagnostics Check		OK	12:19

Data Download Date/Time: 01/28/2022 10:27

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: MCCONNELL,JEFFREY
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 12/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A3
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: FINERFROCK

DATE/TIME OF TEST: 01/01/2022 01:01:16

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.078	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.201	00:56
Breath Volume		1.265	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:58
Subject Sample	# 2	0.186	00:59
Breath Volume		1.410	00:59
Air Blank		0.000	01:00
Control Test		0.078	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 01/13/2022 10:35

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,CPL. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: PARKER,OFC. J.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 12/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A3
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/07/2022 23:43:39

Diagnostics Check		OK	23:37
Air Blank		0.000	23:38
Control Test		0.079	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.000	23:39
Breath Volume		2.878	23:39
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.000	23:42
Breath Volume		3.308	23:42
Air Blank		0.000	23:42
Control Test		0.080	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 01/13/2022 10:36

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,CPL. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,CPL. A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 12/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A3
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: DEPALMA

DATE/TIME OF TEST: 01/08/2022 02:32:43

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.079	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.166	02:27
Breath Volume		1.597	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.185	02:31
Breath Volume		3.511	02:31
Air Blank		0.000	02:31
Control Test		0.079	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 01/13/2022 10:36

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: TACTUK,JOSEPH
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A3
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: BRAITHWAITE

DATE/TIME OF TEST: 01/09/2022 23:35:33

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.079	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.000	23:31
Breath Volume		1.476	23:31
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.000	23:34
Breath Volume		1.335	23:34
Air Blank		0.000	23:34
Control Test		0.079	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 01/13/2022 10:36

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENNELL,RUSSELL,M
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: PENNELL,RUSSELL
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 01/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A3
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: MONTALVO

DATE/TIME OF TEST: 01/20/2022 22:41:36

Diagnostics Check		OK	22:34
Air Blank		0.000	22:34
Control Test		0.078	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.183	22:36
Breath Volume		1.808	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.176	22:40
Breath Volume		1.250	22:40
Air Blank		0.000	22:40
Control Test		0.077	22:40
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 02/06/2022 10:21

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: COLLINS,THOMAS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A3
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: MCFARLAND

DATE/TIME OF TEST: 01/21/2022 19:44:40

Diagnostics Check		OK	19:38
Air Blank		0.000	19:38
Control Test		0.079	19:39
Air Blank		0.000	19:39
Subject Sample	# 1	0.000	19:40
Breath Volume		3.261	19:40
Air Blank		0.000	19:40
Air Blank		0.000	19:42
Subject Sample	# 2	0.000	19:43
Breath Volume		3.445	19:43
Air Blank		0.000	19:43
Control Test		0.078	19:44
Air Blank		0.000	19:44
Diagnostics Check		OK	19:44

Data Download Date/Time: 02/06/2022 10:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,S,
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: HAZIME,S
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 12/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: EVERITT

DATE/TIME OF TEST: 01/01/2022 04:09:34

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:06	
Control Test		0.078	04:07	
Air Blank		0.000	04:07	
Subject Sample	# 1		04:08	Subject Test Refused
Breath Volume		0.000	04:08	
Air Blank		0.000	04:08	
Control Test		0.080	04:08	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 02/09/2022 09:27

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON,JOHN,
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: HALABI,J
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 12/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:23

SUBJECT'S LAST NAME: SMALIOS

DATE/TIME OF TEST: 01/01/2022 06:52:15

Diagnostics Check		OK	06:43	
Air Blank		0.000	06:44	
Control Test		0.079	06:44	
Air Blank		0.000	06:45	
Subject Sample	# 1		06:48	No Sample Provided
Breath Volume		0.000	06:48	
Air Blank		0.000	06:48	
Air Blank		0.000	06:50	
Subject Sample	# 2		06:50	Subject Test Refused
Breath Volume		0.000	06:50	
Air Blank		0.000	06:51	
Control Test		0.078	06:51	
Air Blank		0.000	06:52	
Diagnostics Check		OK	06:52	

Data Download Date/Time: 02/09/2022 09:27

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC.S,
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: HAZIME,OFC S
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 12/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: PANTONI

DATE/TIME OF TEST: 01/06/2022 19:51:07

Diagnostics Check		OK	19:44
Air Blank		0.000	19:44
Control Test		0.081	19:45
Air Blank		0.000	19:45
Subject Sample	# 1	0.124	19:46
Breath Volume		3.457	19:46
Air Blank		0.000	19:47
Air Blank		0.000	19:49
Subject Sample	# 2	0.126	19:49
Breath Volume		3.519	19:49
Air Blank		0.000	19:50
Control Test		0.080	19:50
Air Blank		0.000	19:51
Diagnostics Check		OK	19:51

Data Download Date/Time: 02/09/2022 09:27

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,S,
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: HAZIME,S
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 12/01/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DEMETER

DATE/TIME OF TEST: 01/23/2022 04:36:17

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.080	04:30
Air Blank		0.000	04:30
Subject Sample	# 1	0.167	04:31
Breath Volume		3.476	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.158	04:34
Breath Volume		2.843	04:34
Air Blank		0.000	04:35
Control Test		0.079	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 02/09/2022 09:27

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: TABORDA

DATE/TIME OF TEST: 01/01/2022 03:45:00

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:32	
Control Test		0.082	03:32	
Air Blank		0.000	03:33	
Subject Sample	# 1	0.126	03:36	Volume Not Met
Breath Volume		1.062	03:36	
Air Blank		0.000	03:37	
Air Blank		0.000	03:38	
Subject Sample	# 2	0.135	03:39	
Breath Volume		1.355	03:39	
Air Blank		0.000	03:40	
Air Blank		0.000	03:41	
Subject Sample	# 3	0.133	03:43	
Breath Volume		1.421	03:43	
Air Blank		0.000	03:44	
Control Test		0.080	03:44	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 01/13/2022 11:33

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: YUHAS

DATE/TIME OF TEST: 01/03/2022 01:44:04

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.082	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.133	01:39
Breath Volume		3.890	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.124	01:42
Breath Volume		3.199	01:42
Air Blank		0.000	01:43
Control Test		0.080	01:43
Air Blank		0.000	01:43
Diagnostics Check		OK	01:44

Data Download Date/Time: 01/13/2022 11:33

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 01/05/2022 00:51:17

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.082	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.000	00:46
Breath Volume		2.195	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.000	00:49
Breath Volume		2.277	00:49
Air Blank		0.000	00:50
Control Test		0.083	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 01/13/2022 11:33

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: OPPER

DATE/TIME OF TEST: 01/08/2022 21:18:05

Diagnostics Check		OK	21:15	
Air Blank		0.000	21:15	
Control Test		0.083	21:15	
Air Blank		0.000	21:16	
Subject Sample	# 1		21:16	Subject Test Refused
Breath Volume		0.000	21:16	
Air Blank		0.000	21:17	
Control Test		0.083	21:17	
Air Blank		0.000	21:18	
Diagnostics Check		OK	21:18	

Data Download Date/Time: 01/13/2022 11:33

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: RAMOS-RIVERA

DATE/TIME OF TEST: 01/09/2022 03:43:17

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.082	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.135	03:38
Breath Volume		1.230	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:40
Subject Sample	# 2	0.135	03:41
Breath Volume		1.488	03:41
Air Blank		0.000	03:42
Control Test		0.083	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 01/13/2022 11:33

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 01/09/2022 05:31:28

Diagnostics Check		OK	05:28	
Air Blank		0.000	05:29	
Control Test		0.082	05:29	
Air Blank		0.000	05:29	
Subject Sample	# 1		05:30	Subject Test Refused
Breath Volume		0.000	05:30	
Air Blank		0.000	05:30	
Control Test		0.078	05:30	Interferent Detect
Air Blank		0.000	05:31	

Data Download Date/Time: 01/13/2022 11:34

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 01/09/2022 05:37:53

Diagnostics Check		OK	05:34	
Air Blank		0.000	05:35	
Control Test		0.082	05:35	
Air Blank		0.000	05:36	
Subject Sample	# 1		05:36	Subject Test Refused
Breath Volume		0.000	05:36	
Air Blank		0.000	05:36	
Control Test		0.082	05:37	
Air Blank		0.000	05:37	
Diagnostics Check		OK	05:37	

Data Download Date/Time: 01/13/2022 11:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: LITCHFIELD

DATE/TIME OF TEST: 01/11/2022 22:20:26

Diagnostics Check		OK	22:14
Air Blank		0.000	22:14
Control Test		0.083	22:14
Air Blank		0.000	22:15
Subject Sample	# 1	0.107	22:15
Breath Volume		2.734	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.106	22:18
Breath Volume		2.765	22:18
Air Blank		0.000	22:19
Control Test		0.084	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 01/13/2022 11:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CULALA

DATE/TIME OF TEST: 01/13/2022 00:07:13

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.083	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.173	00:02
Breath Volume		2.628	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.167	00:05
Breath Volume		2.589	00:05
Air Blank		0.000	00:06
Control Test		0.082	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 01/13/2022 11:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PORTAL

DATE/TIME OF TEST: 01/13/2022 03:43:11

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:40	
Control Test		0.082	03:41	
Air Blank		0.000	03:41	
Subject Sample	# 1		03:41	Subject Test Refused
Breath Volume		0.000	03:41	
Air Blank		0.000	03:42	
Control Test		0.083	03:42	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 01/13/2022 11:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: THROWER

DATE/TIME OF TEST: 01/15/2022 03:54:19

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.082	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.177	03:49
Breath Volume		1.425	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.184	03:52
Breath Volume		1.410	03:52
Air Blank		0.000	03:53
Control Test		0.082	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 02/05/2022 15:22

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOROZOV,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 01/15/2022 05:59:40

Diagnostics Check		OK	05:53
Air Blank		0.000	05:53
Control Test		0.083	05:54
Air Blank		0.000	05:54
Subject Sample	# 1	0.146	05:55
Breath Volume		1.222	05:55
Air Blank		0.000	05:55
Air Blank		0.000	05:57
Subject Sample	# 2	0.140	05:58
Breath Volume		1.878	05:58
Air Blank		0.000	05:58
Control Test		0.082	05:59
Air Blank		0.000	05:59
Diagnostics Check		OK	05:59

Data Download Date/Time: 02/05/2022 15:22

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 01/16/2022 23:27:28

Diagnostics Check		OK	23:24	
Air Blank		0.000	23:25	
Control Test		0.084	23:25	
Air Blank		0.000	23:25	
Subject Sample	# 1		23:26	Subject Test Refused
Breath Volume		0.000	23:26	
Air Blank		0.000	23:26	
Control Test		0.083	23:26	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LYON,L
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 01/17/2022 04:48:23

Diagnostics Check		OK	04:45	
Air Blank		0.000	04:46	
Control Test		0.084	04:46	
Air Blank		0.000	04:46	
Subject Sample	# 1		04:47	Subject Test Refused
Breath Volume		0.000	04:47	
Air Blank		0.000	04:47	
Control Test		0.084	04:47	
Air Blank		0.000	04:48	
Diagnostics Check		OK	04:48	

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: BROSSOIT

DATE/TIME OF TEST: 01/17/2022 23:02:24

Diagnostics Check		OK	22:59	
Air Blank		0.000	23:00	
Control Test		0.084	23:00	
Air Blank		0.000	23:00	
Subject Sample	# 1		23:01	Subject Test Refused
Breath Volume		0.000	23:01	
Air Blank		0.000	23:01	
Control Test		0.085	23:01	
Air Blank		0.000	23:02	
Diagnostics Check		OK	23:02	

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MEDDERS

DATE/TIME OF TEST: 01/18/2022 00:48:51

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:46	
Control Test		0.085	00:46	
Air Blank		0.000	00:47	
Subject Sample	# 1		00:47	Subject Test Refused
Breath Volume		0.000	00:47	
Air Blank		0.000	00:47	
Control Test		0.085	00:48	
Air Blank		0.000	00:48	
Diagnostics Check		OK	00:48	

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/18/2022 03:38:07

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.084	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.223	03:33
Breath Volume		1.523	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.227	03:36
Breath Volume		1.375	03:36
Air Blank		0.000	03:37
Control Test		0.084	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CADE

DATE/TIME OF TEST: 01/18/2022 19:47:25

Diagnostics Check		OK	19:40
Air Blank		0.000	19:41
Control Test		0.084	19:41
Air Blank		0.000	19:42
Subject Sample	# 1	0.254	19:42
Breath Volume		1.703	19:42
Air Blank		0.000	19:43
Air Blank		0.000	19:45
Subject Sample	# 2	0.260	19:45
Breath Volume		2.156	19:45
Air Blank		0.000	19:46
Control Test		0.085	19:46
Air Blank		0.000	19:47
Diagnostics Check		OK	19:47

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: PERI

DATE/TIME OF TEST: 01/18/2022 22:35:37

Diagnostics Check		OK	22:29
Air Blank		0.000	22:29
Control Test		0.085	22:30
Air Blank		0.000	22:30
Subject Sample	# 1	0.000	22:31
Breath Volume		1.660	22:31
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.000	22:34
Breath Volume		1.535	22:34
Air Blank		0.000	22:34
Control Test		0.084	22:35
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: GARDENER

DATE/TIME OF TEST: 01/20/2022 00:52:23

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.083	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.136	00:47
Breath Volume		2.308	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:49
Subject Sample	# 2	0.135	00:50
Breath Volume		1.828	00:50
Air Blank		0.000	00:51
Control Test		0.081	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: IBRAHIM

DATE/TIME OF TEST: 01/20/2022 02:33:27

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:31	
Control Test		0.083	02:31	
Air Blank		0.000	02:31	
Subject Sample	# 1		02:32	Subject Test Refused
Breath Volume		0.000	02:32	
Air Blank		0.000	02:32	
Control Test		0.083	02:32	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: OBROSKY

DATE/TIME OF TEST: 01/21/2022 21:53:04

Diagnostics Check		OK	21:46
Air Blank		0.000	21:46
Control Test		0.084	21:47
Air Blank		0.000	21:47
Subject Sample	# 1	0.106	21:48
Breath Volume		3.351	21:48
Air Blank		0.000	21:49
Air Blank		0.000	21:50
Subject Sample	# 2	0.101	21:51
Breath Volume		3.054	21:51
Air Blank		0.000	21:52
Control Test		0.083	21:52
Air Blank		0.000	21:52
Diagnostics Check		OK	21:53

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: ABLAN

DATE/TIME OF TEST: 01/22/2022 03:48:57

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:46	
Control Test		0.085	03:46	
Air Blank		0.000	03:47	
Subject Sample	# 1		03:47	Subject Test Refused
Breath Volume		0.000	03:47	
Air Blank		0.000	03:48	
Control Test		0.085	03:48	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ECHEVERRIAVIDAL,X
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/22/2022 06:04:58

Diagnostics Check		OK	06:01	
Air Blank		0.000	06:02	
Control Test		0.084	06:02	
Air Blank		0.000	06:03	
Subject Sample	# 1		06:03	Subject Test Refused
Breath Volume		0.000	06:03	
Air Blank		0.000	06:04	
Control Test		0.085	06:04	
Air Blank		0.000	06:04	
Diagnostics Check		OK	06:04	

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSEE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: WARREN-BRYSON

DATE/TIME OF TEST: 01/23/2022 04:04:14

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:53	
Control Test		0.084	03:53	
Air Blank		0.000	03:53	
Subject Sample	# 1		03:56	No Sample Provided
Breath Volume		0.000	03:56	
Air Blank		0.000	03:57	
Air Blank		0.000	03:59	
Subject Sample	# 2	0.126	04:02	Volume Not Met
Breath Volume		0.066	04:02	
Air Blank		0.000	04:03	
Control Test		0.085	04:03	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 01/23/2022 18:50:31

Diagnostics Check		OK	18:47	
Air Blank		0.000	18:47	
Control Test		0.083	18:47	
Air Blank		0.000	18:48	
Subject Sample	# 1		18:49	Subject Test Refused
Breath Volume		0.000	18:49	
Air Blank		0.000	18:49	
Control Test		0.085	18:49	
Air Blank		0.000	18:50	
Diagnostics Check		OK	18:50	

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 01/26/2022 03:07:22

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:04	
Control Test		0.083	03:05	
Air Blank		0.000	03:05	
Subject Sample	# 1		03:06	Subject Test Refused
Breath Volume		0.000	03:06	
Air Blank		0.000	03:06	
Control Test		0.084	03:06	
Air Blank		0.000	03:07	
Diagnostics Check		OK	03:07	

Data Download Date/Time: 02/05/2022 15:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: OLSEN

DATE/TIME OF TEST: 01/26/2022 22:04:57

Diagnostics Check		OK	22:02	
Air Blank		0.000	22:02	
Control Test		0.083	22:02	
Air Blank		0.000	22:03	
Subject Sample	# 1		22:03	Subject Test Refused
Breath Volume		0.000	22:03	
Air Blank		0.000	22:04	
Control Test		0.083	22:04	
Air Blank		0.000	22:04	
Diagnostics Check		OK	22:04	

Data Download Date/Time: 02/05/2022 15:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 01/28/2022 04:36:51

Diagnostics Check		OK	04:34	
Air Blank		0.000	04:34	
Control Test		0.084	04:34	
Air Blank		0.000	04:35	
Subject Sample	# 1		04:35	Subject Test Refused
Breath Volume		0.000	04:35	
Air Blank		0.000	04:35	
Control Test		0.084	04:36	
Air Blank		0.000	04:36	
Diagnostics Check		OK	04:36	

Data Download Date/Time: 02/05/2022 15:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: POUST

DATE/TIME OF TEST: 01/30/2022 04:04:33

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.086	04:03	Control Outside Tolerance
Air Blank	0.000	04:04	

Data Download Date/Time: 02/05/2022 15:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,LACEY,M
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: MOLINA,MOISES
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: GARCIA CANO

DATE/TIME OF TEST: 01/07/2022 21:47:17

Diagnostics Check		OK	21:41
Air Blank		0.000	21:41
Control Test		0.080	21:41
Air Blank		0.000	21:42
Subject Sample	# 1	0.129	21:42
Breath Volume		1.960	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.127	21:45
Breath Volume		2.113	21:45
Air Blank		0.000	21:46
Control Test		0.080	21:46
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 02/07/2022 14:11

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 01/08/2022 00:45:53

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.008	00:41
Breath Volume		2.816	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.007	00:44
Breath Volume		3.406	00:44
Air Blank		0.000	00:44
Control Test		0.081	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 02/07/2022 14:12

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 01/01/2022 02:23:15

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.080	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.113	02:18
Breath Volume		1.300	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.122	02:21
Breath Volume		1.453	02:21
Air Blank		0.000	02:22
Control Test		0.078	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 01/07/2022 11:44

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/01/2022 03:25:16

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:17	
Control Test		0.079	03:17	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:21	No Sample Provided
Breath Volume		0.000	03:21	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 2		03:23	Subject Test Refused
Breath Volume		0.000	03:23	
Air Blank		0.000	03:24	
Control Test		0.079	03:24	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 01/07/2022 11:45

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0812108A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: SHIRGER-RAMAGLIA

DATE/TIME OF TEST: 01/02/2022 18:16:42

Diagnostics Check		OK	18:10
Air Blank		0.000	18:10
Control Test		0.080	18:11
Air Blank		0.000	18:11
Subject Sample	# 1	0.136	18:12
Breath Volume		1.816	18:12
Air Blank		0.000	18:12
Air Blank		0.000	18:14
Subject Sample	# 2	0.135	18:15
Breath Volume		2.101	18:15
Air Blank		0.000	18:15
Control Test		0.079	18:16
Air Blank		0.000	18:16
Diagnostics Check		OK	18:16

Data Download Date/Time: 01/07/2022 11:45

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,R
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LAORENZA

DATE/TIME OF TEST: 01/05/2022 01:02:30

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.202	00:57
Breath Volume		3.007	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.204	01:00
Breath Volume		3.855	01:00
Air Blank		0.000	01:01
Control Test		0.078	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 01/07/2022 11:45

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/05/2022 01:22:05

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.077	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.140	01:17
Breath Volume		2.714	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.141	01:20
Breath Volume		2.597	01:20
Air Blank		0.000	01:21
Control Test		0.077	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 01/07/2022 11:45

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 01/05/2022 01:46:17

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.161	01:41
Breath Volume		1.339	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:43
Subject Sample	# 2	0.163	01:44
Breath Volume		1.117	01:44
Air Blank		0.000	01:45
Control Test		0.077	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 01/07/2022 11:45

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: WEBB JR

DATE/TIME OF TEST: 01/14/2022 23:28:30

Diagnostics Check		OK	23:21
Air Blank		0.000	23:22
Control Test		0.081	23:22
Air Blank		0.000	23:22
Subject Sample	# 1	0.103	23:24
Breath Volume		2.710	23:24
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.102	23:26
Breath Volume		2.617	23:26
Air Blank		0.000	23:27
Control Test		0.080	23:27
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 02/03/2022 10:45

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 01/15/2022 00:51:08

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.081	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.000	00:46
Breath Volume		4.410	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.000	00:49
Breath Volume		5.023	00:49
Air Blank		0.000	00:50
Control Test		0.080	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 02/03/2022 10:45

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: IGNACIO

DATE/TIME OF TEST: 01/15/2022 04:07:54

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.080	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.152	04:03
Breath Volume		1.578	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.139	04:06
Breath Volume		1.496	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 02/03/2022 10:45

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SATCHER

DATE/TIME OF TEST: 01/16/2022 04:18:34

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.079	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.187	04:14
Breath Volume		1.464	04:14
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.175	04:17
Breath Volume		1.460	04:17
Air Blank		0.000	04:17
Control Test		0.078	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 02/03/2022 10:45

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 01/18/2022 04:16:38

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.155	04:12
Breath Volume		1.625	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.147	04:15
Breath Volume		1.871	04:15
Air Blank		0.000	04:15
Control Test		0.080	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 02/03/2022 10:45

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWEIGER,J,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WRIGHT,N
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:37

SUBJECT'S LAST NAME: KEYSER

DATE/TIME OF TEST: 01/18/2022 12:21:09

Diagnostics Check		OK	12:13
Air Blank		0.000	12:14
Control Test		0.081	12:14
Air Blank		0.000	12:15
Subject Sample	# 1	0.000	12:16
Breath Volume		1.156	12:16
Air Blank		0.000	12:17
Air Blank		0.000	12:19
Subject Sample	# 2	0.000	12:19
Breath Volume		1.644	12:19
Air Blank		0.000	12:20
Control Test		0.080	12:20
Air Blank		0.000	12:21
Diagnostics Check		OK	12:21

Data Download Date/Time: 02/03/2022 10:45

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHAA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 01/20/2022 23:34:17

Diagnostics Check		OK	23:27
Air Blank		0.000	23:28
Control Test		0.081	23:28
Air Blank		0.000	23:28
Subject Sample	# 1	0.000	23:29
Breath Volume		1.195	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.000	23:32
Breath Volume		1.195	23:32
Air Blank		0.000	23:33
Control Test		0.080	23:33
Air Blank		0.000	23:34
Diagnostics Check		OK	23:34

Data Download Date/Time: 02/03/2022 10:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: COURCHENE

DATE/TIME OF TEST: 01/21/2022 01:41:02

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.056	01:36
Breath Volume		2.144	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.060	01:39
Breath Volume		2.277	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:41

Data Download Date/Time: 02/03/2022 10:45

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:26

SUBJECT'S LAST NAME: ROGOFF

DATE/TIME OF TEST: 01/21/2022 10:01:32

Diagnostics Check		OK	09:55
Air Blank		0.000	09:55
Control Test		0.080	09:55
Air Blank		0.000	09:56
Subject Sample	# 1	0.139	09:56
Breath Volume		2.375	09:56
Air Blank		0.000	09:57
Air Blank		0.000	09:59
Subject Sample	# 2	0.129	10:00
Breath Volume		1.988	10:00
Air Blank		0.000	10:00
Control Test		0.079	10:00
Air Blank		0.000	10:01
Diagnostics Check		OK	10:01

Data Download Date/Time: 02/03/2022 10:45

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: BRAKEFIELD

DATE/TIME OF TEST: 01/21/2022 18:42:48

Diagnostics Check		OK	18:30	
Air Blank		0.000	18:30	
Control Test		0.080	18:31	
Air Blank		0.000	18:31	
Subject Sample	# 1		18:34	No Sample Provided
Breath Volume		0.000	18:34	
Air Blank		0.000	18:35	
Air Blank		0.000	18:37	
Subject Sample	# 2	0.227	18:38	
Breath Volume		2.503	18:38	
Air Blank		0.000	18:38	
Air Blank		0.000	18:40	
Subject Sample	# 3	0.230	18:41	
Breath Volume		2.375	18:41	
Air Blank		0.000	18:41	
Control Test		0.078	18:42	
Air Blank		0.000	18:42	
Diagnostics Check		OK	18:42	

Data Download Date/Time: 02/03/2022 10:45

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: ZIMMERMAN

DATE/TIME OF TEST: 01/21/2022 22:58:26

Diagnostics Check		OK	22:52
Air Blank		0.000	22:53
Control Test		0.080	22:53
Air Blank		0.000	22:53
Subject Sample	# 1	0.000	22:54
Breath Volume		1.980	22:54
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.000	22:57
Breath Volume		1.675	22:57
Air Blank		0.000	22:57
Control Test		0.079	22:57
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 02/03/2022 10:46

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: SAHAGIAN IV

DATE/TIME OF TEST: 01/22/2022 00:55:42

Diagnostics Check		OK	00:49
Air Blank		0.000	00:50
Control Test		0.080	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.168	00:51
Breath Volume		2.031	00:51
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.165	00:54
Breath Volume		1.906	00:54
Air Blank		0.000	00:54
Control Test		0.080	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 02/03/2022 10:46

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: BESHARA

DATE/TIME OF TEST: 01/22/2022 03:51:58

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:43	
Control Test		0.080	03:44	
Air Blank		0.000	03:44	
Subject Sample	# 1		03:47	No Sample Provided
Breath Volume		0.000	03:47	
Air Blank		0.000	03:48	
Air Blank		0.000	03:50	
Subject Sample	# 2		03:50	Subject Test Refused
Breath Volume		0.000	03:50	
Air Blank		0.000	03:51	
Control Test		0.080	03:51	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 02/03/2022 10:46

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: FERMAN

DATE/TIME OF TEST: 01/23/2022 01:41:16

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.081	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.170	01:36
Breath Volume		2.097	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.171	01:39
Breath Volume		2.031	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 02/03/2022 10:46

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:22

SUBJECT'S LAST NAME: DENT

DATE/TIME OF TEST: 01/26/2022 09:49:51

Diagnostics Check		OK	09:42
Air Blank		0.000	09:42
Control Test		0.079	09:43
Air Blank		0.000	09:43
Subject Sample	# 1	0.141	09:45
Breath Volume		2.445	09:45
Air Blank		0.000	09:45
Air Blank		0.000	09:47
Subject Sample	# 2	0.146	09:48
Breath Volume		2.046	09:48
Air Blank		0.000	09:48
Control Test		0.078	09:49
Air Blank		0.000	09:49
Diagnostics Check		OK	09:49

Data Download Date/Time: 02/03/2022 10:46

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: FAJARDO

DATE/TIME OF TEST: 01/27/2022 00:20:44

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.081	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.247	00:16
Breath Volume		1.449	00:16
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.234	00:19
Breath Volume		1.535	00:19
Air Blank		0.000	00:19
Control Test		0.079	00:20
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 02/03/2022 10:46

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: SHEPARD

DATE/TIME OF TEST: 01/27/2022 01:07:51

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.080	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.165	01:03
Breath Volume		2.277	01:03
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.171	01:06
Breath Volume		1.722	01:06
Air Blank		0.000	01:06
Control Test		0.079	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 02/03/2022 10:46

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 01/27/2022 02:17:02

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:09	
Control Test		0.080	02:09	
Air Blank		0.000	02:09	
Subject Sample	# 1		02:12	No Sample Provided
Breath Volume		0.000	02:12	
Air Blank		0.000	02:13	
Air Blank		0.000	02:15	
Subject Sample	# 2		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:16	
Control Test		0.080	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 02/03/2022 10:46

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: ELLSBERRY

DATE/TIME OF TEST: 01/27/2022 20:48:46

Diagnostics Check		OK	20:36	
Air Blank		0.000	20:36	
Control Test		0.081	20:37	
Air Blank		0.000	20:37	
Subject Sample	# 1		20:40	No Sample Provided
Breath Volume		0.000	20:40	
Air Blank		0.000	20:41	
Air Blank		0.000	20:43	
Subject Sample	# 2	0.261	20:44	
Breath Volume		2.035	20:44	
Air Blank		0.000	20:44	
Air Blank		0.000	20:46	
Subject Sample	# 3	0.277	20:47	
Breath Volume		1.628	20:47	
Air Blank		0.000	20:47	
Control Test		0.079	20:48	
Air Blank		0.000	20:48	
Diagnostics Check		OK	20:48	

Data Download Date/Time: 02/03/2022 10:46

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: SCHARES

DATE/TIME OF TEST: 01/27/2022 23:12:26

Diagnostics Check		OK	23:05
Air Blank		0.000	23:05
Control Test		0.080	23:06
Air Blank		0.000	23:06
Subject Sample	# 1	0.118	23:07
Breath Volume		2.753	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.113	23:10
Breath Volume		1.476	23:10
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 02/03/2022 10:46

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 01/27/2022 23:38:24

Diagnostics Check		OK	23:32
Air Blank		0.000	23:32
Control Test		0.080	23:33
Air Blank		0.000	23:33
Subject Sample	# 1	0.061	23:34
Breath Volume		1.402	23:34
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.061	23:36
Breath Volume		1.269	23:36
Air Blank		0.000	23:37
Control Test		0.080	23:37
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 02/03/2022 10:47

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HOUTEN,JONATHAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: SAURI

DATE/TIME OF TEST: 01/28/2022 00:05:46

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.081	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.172	00:01
Breath Volume		2.679	00:01
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.171	00:04
Breath Volume		2.281	00:04
Air Blank		0.000	00:04
Control Test		0.078	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 02/03/2022 10:47

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DAVID,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: ROBERTS,DAVID
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: GANISIN

DATE/TIME OF TEST: 01/11/2022 17:09:51

Diagnostics Check		OK	17:03
Air Blank		0.000	17:03
Control Test		0.081	17:04
Air Blank		0.000	17:04
Subject Sample	# 1	0.000	17:05
Breath Volume		4.023	17:05
Air Blank		0.000	17:05
Air Blank		0.000	17:07
Subject Sample	# 2	0.000	17:08
Breath Volume		3.792	17:08
Air Blank		0.000	17:08
Control Test		0.081	17:09
Air Blank		0.000	17:09
Diagnostics Check		OK	17:09

Data Download Date/Time: 02/23/2022 15:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: KERANS

DATE/TIME OF TEST: 01/13/2022 21:35:09

Diagnostics Check		OK	21:22	
Air Blank		0.000	21:22	
Control Test		0.080	21:22	
Air Blank		0.000	21:23	
Subject Sample	# 1		21:26	No Sample Provided
Breath Volume		0.000	21:26	
Air Blank		0.000	21:27	
Air Blank		0.000	21:29	
Subject Sample	# 2	0.098	21:29	
Breath Volume		1.710	21:29	
Air Blank		0.000	21:30	
Air Blank		0.000	21:32	
Subject Sample	# 3	0.097	21:33	
Breath Volume		2.707	21:33	
Air Blank		0.000	21:34	
Control Test		0.080	21:34	
Air Blank		0.000	21:35	
Diagnostics Check		OK	21:35	

Data Download Date/Time: 02/23/2022 15:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: FERQUERON

DATE/TIME OF TEST: 01/14/2022 00:39:04

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.081	00:33
Air Blank		0.000	00:33
Subject Sample	# 1	0.151	00:34
Breath Volume		1.738	00:34
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.150	00:37
Breath Volume		2.085	00:37
Air Blank		0.000	00:38
Control Test		0.081	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:39

Data Download Date/Time: 02/23/2022 15:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: CORBET

DATE/TIME OF TEST: 01/14/2022 20:43:12

Diagnostics Check		OK	20:40	
Air Blank		0.000	20:40	
Control Test		0.081	20:40	
Air Blank		0.000	20:41	
Subject Sample	# 1		20:41	Subject Test Refused
Breath Volume		0.000	20:41	
Air Blank		0.000	20:42	
Control Test		0.081	20:42	
Air Blank		0.000	20:43	
Diagnostics Check		OK	20:43	

Data Download Date/Time: 02/23/2022 15:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MIHAL

DATE/TIME OF TEST: 01/15/2022 01:33:28

Diagnostics Check		OK	01:23	
Air Blank		0.000	01:24	
Control Test		0.081	01:24	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:28	No Sample Provided
Breath Volume		0.000	01:28	
Air Blank		0.000	01:28	
Air Blank		0.000	01:30	
Subject Sample	# 2		01:32	Subject Test Refused
Breath Volume		0.000	01:32	
Air Blank		0.000	01:32	
Control Test		0.081	01:32	
Air Blank		0.000	01:33	
Diagnostics Check		OK	01:33	

Data Download Date/Time: 02/23/2022 15:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GARLAND

DATE/TIME OF TEST: 01/15/2022 04:53:49

Diagnostics Check		OK	04:47
Air Blank		0.000	04:48
Control Test		0.081	04:48
Air Blank		0.000	04:48
Subject Sample	# 1	0.138	04:49
Breath Volume		2.933	04:49
Air Blank		0.000	04:49
Air Blank		0.000	04:51
Subject Sample	# 2	0.130	04:52
Breath Volume		1.828	04:52
Air Blank		0.000	04:52
Control Test		0.080	04:53
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 02/23/2022 15:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: BEVIS

DATE/TIME OF TEST: 01/17/2022 20:30:08

Diagnostics Check		OK	20:26	
Air Blank		0.000	20:26	
Control Test		0.080	20:27	
Air Blank		0.000	20:27	
Subject Sample	# 1		20:28	Subject Test Refused
Breath Volume		0.000	20:28	
Air Blank		0.000	20:29	
Control Test		0.081	20:29	
Air Blank		0.000	20:30	
Diagnostics Check		OK	20:30	

Data Download Date/Time: 02/23/2022 15:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 01/21/2022 00:17:12

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.080	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.202	00:12
Breath Volume		1.234	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.198	00:15
Breath Volume		1.429	00:15
Air Blank		0.000	00:16
Control Test		0.080	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 02/23/2022 15:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: FLOOD GARABI

DATE/TIME OF TEST: 01/22/2022 01:16:03

Diagnostics Check	OK	01:15	
Air Blank		01:15	RFI Detect
Air Blank	0.000	01:16	

Data Download Date/Time: 02/23/2022 15:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: FLOOD GARABI

DATE/TIME OF TEST: 01/22/2022 01:20:42

Diagnostics Check		OK	01:17	
Air Blank		0.000	01:18	
Control Test		0.080	01:18	
Air Blank		0.000	01:18	
Subject Sample	# 1		01:19	Subject Test Refused
Breath Volume		0.000	01:19	
Air Blank		0.000	01:19	
Control Test		0.080	01:20	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:20	

Data Download Date/Time: 02/23/2022 15:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 01/22/2022 03:43:27

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:32	
Control Test		0.081	03:32	
Air Blank		0.000	03:33	
Subject Sample	# 1		03:36	No Sample Provided
Breath Volume		0.175	03:36	
Air Blank		0.000	03:36	
Air Blank		0.000	03:38	
Subject Sample	# 2	0.197	03:41	Volume Not Met
Breath Volume		0.800	03:41	
Air Blank		0.000	03:42	
Control Test		0.082	03:42	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 02/23/2022 15:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALAMEDA,JORGE,L
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: MATHENY,DEVIN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 01/26/2022 18:23:48

Diagnostics Check		OK	18:17
Air Blank		0.000	18:18
Control Test		0.079	18:18
Air Blank		0.000	18:18
Subject Sample	# 1	0.200	18:19
Breath Volume		3.718	18:19
Air Blank		0.000	18:19
Air Blank		0.000	18:21
Subject Sample	# 2	0.194	18:22
Breath Volume		1.757	18:22
Air Blank		0.000	18:22
Control Test		0.079	18:23
Air Blank		0.000	18:23
Diagnostics Check		OK	18:23

Data Download Date/Time: 02/23/2022 15:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 01/30/2022 00:49:53

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:47	
Control Test		0.080	00:47	
Air Blank		0.000	00:47	
Subject Sample	# 1		00:48	Subject Test Refused
Breath Volume		0.000	00:48	
Air Blank		0.000	00:48	
Control Test		0.081	00:49	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 02/23/2022 15:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 01/01/2022 03:04:33

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.081	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.137	03:00
Breath Volume		1.484	03:00
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.138	03:03
Breath Volume		1.175	03:03
Air Blank		0.000	03:03
Control Test		0.081	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 01/07/2022 15:56

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JEFFERSON,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LONG,JEFFERSON
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 01/01/2022 04:03:04

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.082	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.122	03:58
Breath Volume		1.128	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:00
Subject Sample	# 2	0.119	04:01
Breath Volume		3.539	04:01
Air Blank		0.000	04:02
Control Test		0.081	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:03

Data Download Date/Time: 01/07/2022 15:57

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: BECHTOL

DATE/TIME OF TEST: 01/02/2022 06:26:14

Diagnostics Check		OK	06:20
Air Blank		0.000	06:20
Control Test		0.082	06:20
Air Blank		0.000	06:21
Subject Sample	# 1	0.116	06:21
Breath Volume		1.710	06:21
Air Blank		0.000	06:22
Air Blank		0.000	06:24
Subject Sample	# 2	0.119	06:24
Breath Volume		3.375	06:24
Air Blank		0.000	06:25
Control Test		0.080	06:25
Air Blank		0.000	06:26
Diagnostics Check		OK	06:26

Data Download Date/Time: 01/07/2022 15:57

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: HOWE

DATE/TIME OF TEST: 01/04/2022 21:58:02

Diagnostics Check		OK	21:51
Air Blank		0.000	21:52
Control Test		0.081	21:52
Air Blank		0.000	21:52
Subject Sample	# 1	0.154	21:53
Breath Volume		2.457	21:53
Air Blank		0.000	21:54
Air Blank		0.000	21:55
Subject Sample	# 2	0.152	21:56
Breath Volume		1.429	21:56
Air Blank		0.000	21:57
Control Test		0.080	21:57
Air Blank		0.000	21:57
Diagnostics Check		OK	21:58

Data Download Date/Time: 01/07/2022 15:57

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIZZO,HAYLEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RIZZO,HAYLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: PULLEY

DATE/TIME OF TEST: 01/05/2022 05:41:04

Diagnostics Check		OK	05:38	
Air Blank		0.000	05:38	
Control Test		0.081	05:39	
Air Blank		0.000	05:39	
Subject Sample	# 1		05:39	Subject Test Refused
Breath Volume		0.000	05:39	
Air Blank		0.000	05:40	
Control Test		0.080	05:40	
Air Blank		0.000	05:40	
Diagnostics Check		OK	05:41	

Data Download Date/Time: 01/07/2022 15:57

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,JESUS,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FIGUEROA,JESUS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/06/2022 00:58:05

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:54	
Control Test		0.081	00:55	
Air Blank		0.000	00:55	
Subject Sample	# 1		00:56	Subject Test Refused
Breath Volume		0.000	00:56	
Air Blank		0.000	00:57	
Control Test		0.081	00:57	
Air Blank		0.000	00:58	
Diagnostics Check		OK	00:58	

Data Download Date/Time: 01/07/2022 15:57

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: KAMMERER

DATE/TIME OF TEST: 01/08/2022 02:33:39

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.336	02:28
Breath Volume		2.328	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.319	02:32
Breath Volume		1.476	02:32
Air Blank		0.000	02:32
Control Test		0.081	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 02/24/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 01/15/2022 04:58:18

Diagnostics Check		OK	04:53	
Air Blank		0.000	04:53	
Control Test		0.081	04:53	
Air Blank		0.000	04:54	
Subject Sample	# 1	0.222	04:54	
Breath Volume		2.593	04:54	
Air Blank		0.000	04:55	
Air Blank			04:56	RFI Detect
Air Blank		0.000	04:57	
Control Test		0.082	04:57	
Air Blank		0.000	04:58	
Diagnostics Check		OK	04:58	

Data Download Date/Time: 02/24/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 01/15/2022 05:06:32

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.081	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.221	05:01
Breath Volume		1.785	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.218	05:04
Breath Volume		1.339	05:04
Air Blank		0.000	05:05
Control Test		0.081	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 02/24/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 01/18/2022 03:37:13

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:29	
Control Test		0.081	03:29	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:32	No Sample Provided
Breath Volume		0.000	03:32	
Air Blank		0.000	03:33	
Air Blank		0.000	03:35	
Subject Sample	# 2		03:35	Subject Test Refused
Breath Volume		0.000	03:35	
Air Blank		0.000	03:36	
Control Test		0.082	03:36	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 02/24/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: CASSON

DATE/TIME OF TEST: 01/18/2022 21:56:20

Diagnostics Check		OK	21:47
Air Blank		0.000	21:48
Control Test		0.082	21:48
Air Blank		0.000	21:49
Subject Sample	# 1	0.147	21:51
Breath Volume		2.703	21:51
Air Blank		0.000	21:52
Air Blank		0.000	21:53
Subject Sample	# 2	0.135	21:54
Breath Volume		1.703	21:54
Air Blank		0.000	21:55
Control Test		0.082	21:55
Air Blank		0.000	21:56
Diagnostics Check		OK	21:56

Data Download Date/Time: 02/24/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: KOSCIELICKI

DATE/TIME OF TEST: 01/19/2022 20:38:55

Diagnostics Check		OK	20:35	
Air Blank		0.000	20:35	
Control Test		0.082	20:36	
Air Blank		0.000	20:36	
Subject Sample	# 1		20:37	Subject Test Refused
Breath Volume		0.000	20:37	
Air Blank		0.000	20:38	
Control Test		0.081	20:38	
Air Blank		0.000	20:38	
Diagnostics Check		OK	20:38	

Data Download Date/Time: 02/24/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARQUES,JASON,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FIVECOAT,JUSTIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:39

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 01/21/2022 09:07:39

Diagnostics Check		OK	09:01
Air Blank		0.000	09:02
Control Test		0.082	09:02
Air Blank		0.000	09:02
Subject Sample	# 1	0.000	09:03
Breath Volume		1.906	09:03
Air Blank		0.000	09:03
Air Blank		0.000	09:05
Subject Sample	# 2	0.000	09:06
Breath Volume		2.191	09:06
Air Blank		0.000	09:06
Control Test		0.081	09:07
Air Blank		0.000	09:07
Diagnostics Check		OK	09:07

Data Download Date/Time: 02/24/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/21/2022 23:00:31

Diagnostics Check		OK	22:54
Air Blank		0.000	22:54
Control Test		0.080	22:54
Air Blank		0.000	22:55
Subject Sample	# 1	0.176	22:55
Breath Volume		2.464	22:55
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.164	22:58
Breath Volume		2.132	22:58
Air Blank		0.000	22:59
Control Test		0.080	22:59
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 02/24/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: FUENTES RANGEL

DATE/TIME OF TEST: 01/24/2022 00:43:31

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.082	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.200	00:38
Breath Volume		2.257	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.198	00:41
Breath Volume		2.691	00:41
Air Blank		0.000	00:42
Control Test		0.079	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 02/24/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: TSOLIS

DATE/TIME OF TEST: 01/24/2022 15:02:27

Diagnostics Check		OK	14:55
Air Blank		0.000	14:55
Control Test		0.082	14:56
Air Blank		0.000	14:56
Subject Sample	# 1	0.064	14:57
Breath Volume		1.593	14:57
Air Blank		0.000	14:57
Air Blank		0.000	14:59
Subject Sample	# 2	0.058	15:00
Breath Volume		1.132	15:00
Air Blank		0.000	15:01
Control Test		0.083	15:01
Air Blank		0.000	15:02
Diagnostics Check		OK	15:02

Data Download Date/Time: 02/24/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,JESUS,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FIGUEROA,JESUS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: VICENTE

DATE/TIME OF TEST: 01/24/2022 21:02:48

Diagnostics Check		OK	20:59	
Air Blank		0.000	21:00	
Control Test		0.081	21:00	
Air Blank		0.000	21:01	
Subject Sample	# 1		21:01	Subject Test Refused
Breath Volume		0.000	21:01	
Air Blank		0.000	21:01	
Control Test		0.081	21:02	
Air Blank		0.000	21:02	
Diagnostics Check		OK	21:02	

Data Download Date/Time: 02/24/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 01/27/2022 21:57:55

Diagnostics Check		OK	21:50
Air Blank		0.000	21:51
Control Test		0.081	21:51
Air Blank		0.000	21:51
Subject Sample	# 1	0.117	21:52
Breath Volume		1.613	21:52
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.121	21:56
Breath Volume		1.808	21:56
Air Blank		0.000	21:57
Control Test		0.081	21:57
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 02/24/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIZZO,HAYLEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RIZZO,HAYLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: CARSWELL

DATE/TIME OF TEST: 01/27/2022 23:56:42

Diagnostics Check		OK	23:49
Air Blank		0.000	23:49
Control Test		0.081	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.000	23:51
Breath Volume		1.707	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.000	23:55
Breath Volume		3.351	23:55
Air Blank		0.000	23:55
Control Test		0.082	23:56
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 02/24/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: KISSER

DATE/TIME OF TEST: 01/28/2022 15:05:28

Diagnostics Check		OK	14:57
Air Blank		0.000	14:57
Control Test		0.081	14:58
Air Blank		0.000	14:58
Subject Sample	# 1	0.213	14:59
Breath Volume		1.300	14:59
Air Blank		0.000	14:59
Air Blank		0.000	15:01
Subject Sample	# 2	0.232	15:03
Breath Volume		2.667	15:03
Air Blank		0.000	15:04
Control Test		0.082	15:04
Air Blank		0.000	15:05
Diagnostics Check		OK	15:05

Data Download Date/Time: 02/24/2022 18:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,AUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WELLS,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 01/20/2022 03:47:19

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.084	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.162	03:42
Breath Volume		1.410	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.161	03:45
Breath Volume		1.996	03:45
Air Blank		0.000	03:46
Control Test		0.083	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 01/26/2022 21:16

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: OCHSENHIRT

DATE/TIME OF TEST: 01/03/2022 18:32:56

Diagnostics Check	OK	18:31	
Air Blank		18:32	Ambient Fail
Air Blank	0.000	18:32	

Data Download Date/Time: 02/04/2022 18:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: ROYCE,JAMES
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: OCHSENHIRT

DATE/TIME OF TEST: 01/03/2022 18:40:57

Diagnostics Check		OK	18:36	
Air Blank		0.000	18:36	
Control Test		0.081	18:36	
Air Blank		0.000	18:37	
Subject Sample	# 1		18:40	Sequence Aborted
Breath Volume		0.050	18:40	
Air Blank		0.000	18:40	

Data Download Date/Time: 02/04/2022 18:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: OCHSENHIRT

DATE/TIME OF TEST: 01/03/2022 18:54:39

Diagnostics Check		OK	18:43
Air Blank		0.000	18:44
Control Test		0.079	18:44
Air Blank		0.000	18:44
Subject Sample	# 1	0.266	18:46
Breath Volume		1.156	18:46
Air Blank		0.000	18:46
Air Blank		0.000	18:48
Subject Sample	# 2	0.204	18:49
Breath Volume		1.464	18:49
Air Blank		0.000	18:50
Air Blank		0.000	18:52
Subject Sample	# 3	0.219	18:52
Breath Volume		1.832	18:52
Air Blank		0.000	18:53
Control Test		0.078	18:54
Air Blank		0.000	18:54
Diagnostics Check		OK	18:54

Data Download Date/Time: 02/04/2022 18:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: SHEARER

DATE/TIME OF TEST: 01/07/2022 21:08:46

Diagnostics Check		OK	21:01
Air Blank		0.000	21:01
Control Test		0.081	21:02
Air Blank		0.000	21:02
Subject Sample	# 1	0.206	21:03
Breath Volume		2.031	21:03
Air Blank		0.000	21:04
Air Blank		0.000	21:06
Subject Sample	# 2	0.196	21:07
Breath Volume		1.628	21:07
Air Blank		0.000	21:07
Control Test		0.079	21:08
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 02/04/2022 18:59

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: BERNAL,A
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/13/2022 03:21:39

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.081	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.215	03:16
Breath Volume		1.585	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.214	03:20
Breath Volume		2.039	03:20
Air Blank		0.000	03:20
Control Test		0.080	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 02/04/2022 18:59

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: BERNAL,A
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/17/2022 23:23:16

Diagnostics Check		OK	23:16
Air Blank		0.000	23:16
Control Test		0.081	23:17
Air Blank		0.000	23:17
Subject Sample	# 1	0.163	23:18
Breath Volume		1.871	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.165	23:21
Breath Volume		2.906	23:21
Air Blank		0.000	23:22
Control Test		0.079	23:22
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 02/04/2022 18:59

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WALTER,CRAIG
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 12/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ORTEGA RIVERA

DATE/TIME OF TEST: 01/27/2022 02:17:47

Diagnostics Check		OK	02:12	
Air Blank		0.000	02:13	
Control Test		0.081	02:13	
Air Blank		0.000	02:13	
Subject Sample	# 1	0.187	02:16	
Breath Volume		1.613	02:16	
Air Blank			02:17	Purge Fail
Air Blank		0.000	02:17	

Data Download Date/Time: 02/04/2022 18:59

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WALTER,CRAIG
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 12/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ORTEGA RIVERA

DATE/TIME OF TEST: 01/27/2022 02:30:52

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:19	
Control Test		0.079	02:20	
Air Blank		0.000	02:20	
Subject Sample	# 1	0.183	02:23	Volume Not Met
Breath Volume		0.339	02:23	
Air Blank		0.000	02:24	
Air Blank		0.000	02:26	
Subject Sample	# 2	0.172	02:29	Volume Not Met
Breath Volume		0.019	02:29	
Air Blank		0.000	02:29	
Control Test		0.079	02:30	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 02/04/2022 18:59

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATTS,SHANNON,N
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: BERNAL,ANTHONY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: AGUILAR CASTILLO

DATE/TIME OF TEST: 01/28/2022 01:58:25

Diagnostics Check		OK	01:51
Air Blank		0.000	01:51
Control Test		0.081	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.133	01:53
Breath Volume		1.839	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.142	01:56
Breath Volume		1.835	01:56
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 02/04/2022 18:59

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 01/13/2022 22:21:19

Diagnostics Check		OK	22:16	
Air Blank		0.000	22:17	
Control Test		0.079	22:17	
Air Blank		0.000	22:18	
Subject Sample	# 1		22:19	Subject Test Refused
Breath Volume		0.000	22:19	
Air Blank		0.000	22:20	
Control Test		0.080	22:20	
Air Blank		0.000	22:21	
Diagnostics Check		OK	22:21	

Data Download Date/Time: 02/09/2022 15:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARTHY,EVAN,V
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MCCARTHY,EVAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: JUTTING JR

DATE/TIME OF TEST: 01/14/2022 21:55:11

Diagnostics Check		OK	21:49
Air Blank		0.000	21:49
Control Test		0.081	21:49
Air Blank		0.000	21:50
Subject Sample	# 1	0.000	21:50
Breath Volume		3.308	21:50
Air Blank		0.000	21:51
Air Blank		0.000	21:53
Subject Sample	# 2	0.000	21:53
Breath Volume		3.503	21:53
Air Blank		0.000	21:54
Control Test		0.081	21:54
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 01/19/2022 12:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVEZ,BRYAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP K
ARRESTING OFFICER: VALDIVIA OLAY,MIGUEL
ARRESTING OFFICER AGENCY: FHP TROOP K
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: DESIR

DATE/TIME OF TEST: 01/16/2022 07:02:01

Diagnostics Check		OK	06:58	
Air Blank		0.000	06:59	
Control Test		0.082	06:59	
Air Blank		0.000	06:59	
Subject Sample	# 1		07:00	Subject Test Refused
Breath Volume		0.000	07:00	
Air Blank		0.000	07:01	
Control Test		0.083	07:01	
Air Blank		0.000	07:01	
Diagnostics Check		OK	07:02	

Data Download Date/Time: 01/26/2022 10:08

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RIGGALL,NICHOLS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 01/01/2022 02:48:18

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.081	02:42
Air Blank		0.000	02:43
Subject Sample	# 1	0.192	02:43
Breath Volume		2.687	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.196	02:46
Breath Volume		2.964	02:46
Air Blank		0.000	02:47
Control Test		0.080	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 01/11/2022 16:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 01/01/2022 03:00:10

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.080	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.101	02:54
Breath Volume		1.230	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.096	02:58
Breath Volume		1.304	02:58
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 01/11/2022 16:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROPST,JOHN,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: KING,CHARLES
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 01/01/2022 10:40:37

Diagnostics Check		OK	10:32
Air Blank		0.000	10:32
Control Test		0.081	10:32
Air Blank		0.000	10:33
Subject Sample	# 1	0.000	10:34
Breath Volume		1.304	10:34
Air Blank		0.000	10:35
Air Blank		0.000	10:37
Subject Sample	# 2	0.000	10:39
Breath Volume		1.242	10:39
Air Blank		0.000	10:39
Control Test		0.080	10:40
Air Blank		0.000	10:40
Diagnostics Check		OK	10:40

Data Download Date/Time: 01/11/2022 16:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: MESTIZA MELCHOR

DATE/TIME OF TEST: 01/02/2022 02:49:45

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.080	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.112	02:45
Breath Volume		3.222	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.105	02:48
Breath Volume		2.453	02:48
Air Blank		0.000	02:48
Control Test		0.081	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 01/11/2022 16:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: ANDRADE LARICS

DATE/TIME OF TEST: 01/02/2022 03:30:46

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.080	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.138	03:26
Breath Volume		2.394	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:28
Subject Sample	# 2	0.142	03:29
Breath Volume		1.777	03:29
Air Blank		0.000	03:29
Control Test		0.080	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 01/11/2022 16:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: PIKE

DATE/TIME OF TEST: 01/03/2022 04:35:00

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.081	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.241	04:30
Breath Volume		3.390	04:30
Air Blank		0.000	04:31
Air Blank		0.000	04:32
Subject Sample	# 2	0.244	04:33
Breath Volume		3.476	04:33
Air Blank		0.000	04:34
Control Test		0.080	04:34
Air Blank		0.000	04:34
Diagnostics Check		OK	04:35

Data Download Date/Time: 01/11/2022 16:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROPST,JOHN,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: NEWBERRY

DATE/TIME OF TEST: 01/03/2022 17:50:00

Diagnostics Check		OK	17:38	
Air Blank		0.000	17:39	
Control Test		0.080	17:39	
Air Blank		0.000	17:39	
Subject Sample	# 1		17:43	No Sample Provided
Breath Volume		0.000	17:43	
Air Blank		0.000	17:43	
Air Blank		0.000	17:45	
Subject Sample	# 2		17:48	No Sample Provided
Breath Volume		0.000	17:48	
Air Blank		0.000	17:49	
Control Test		0.080	17:49	
Air Blank		0.000	17:49	
Diagnostics Check		OK	17:50	

Data Download Date/Time: 01/11/2022 16:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAYNE,ABIGAIL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 01/04/2022 21:01:33

Diagnostics Check		OK	20:53
Air Blank		0.000	20:54
Control Test		0.080	20:54
Air Blank		0.000	20:55
Subject Sample	# 1	0.151	20:56
Breath Volume		1.593	20:56
Air Blank		0.000	20:57
Air Blank		0.000	20:58
Subject Sample	# 2	0.135	21:00
Breath Volume		1.218	21:00
Air Blank		0.000	21:00
Control Test		0.078	21:00
Air Blank		0.000	21:01
Diagnostics Check		OK	21:01

Data Download Date/Time: 01/11/2022 16:01

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORA,AARON,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MORA,AARON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MAGOUIRK

DATE/TIME OF TEST: 01/06/2022 00:44:51

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:41	
Control Test		0.081	00:41	
Air Blank		0.000	00:41	
Subject Sample	# 1		00:43	Subject Test Refused
Breath Volume		0.000	00:43	
Air Blank		0.000	00:43	
Control Test		0.082	00:44	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 01/11/2022 16:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 01/06/2022 20:53:34

Diagnostics Check		OK	20:47
Air Blank		0.000	20:47
Control Test		0.081	20:47
Air Blank		0.000	20:48
Subject Sample	# 1	0.112	20:49
Breath Volume		2.199	20:49
Air Blank		0.000	20:49
Air Blank		0.000	20:51
Subject Sample	# 2	0.107	20:52
Breath Volume		1.429	20:52
Air Blank		0.000	20:52
Control Test		0.080	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 01/11/2022 16:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,JOLNETH,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLON,JOLNETH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: GUARDADO-LOPEZ

DATE/TIME OF TEST: 01/07/2022 01:47:57

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:37	
Control Test		0.081	01:37	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:41	No Sample Provided
Breath Volume		0.000	01:41	
Air Blank		0.000	01:41	
Air Blank		0.000	01:43	
Subject Sample	# 2		01:46	No Sample Provided
Breath Volume		0.000	01:46	
Air Blank		0.000	01:47	
Control Test		0.082	01:47	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 01/11/2022 16:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAYNE,ABIGAIL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 01/08/2022 03:48:03

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:45	
Control Test		0.081	03:46	
Air Blank		0.000	03:46	
Subject Sample	# 1		03:46	Subject Test Refused
Breath Volume		0.000	03:46	
Air Blank		0.000	03:47	
Control Test		0.081	03:47	
Air Blank		0.000	03:47	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 01/11/2022 16:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JUDAH,JOSHUA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/09/2022 00:58:25

Diagnostics Check		OK	00:51
Air Blank		0.000	00:52
Control Test		0.081	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.154	00:53
Breath Volume		1.339	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.154	00:56
Breath Volume		1.187	00:56
Air Blank		0.000	00:57
Control Test		0.081	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 01/11/2022 16:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTS,JAMES,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BISPO,STEPHANIE
ARRESTING OFFICER AGENCY: Davenport PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/09/2022 06:27:51

Diagnostics Check		OK	06:21
Air Blank		0.000	06:21
Control Test		0.081	06:22
Air Blank		0.000	06:22
Subject Sample	# 1	0.123	06:23
Breath Volume		1.898	06:23
Air Blank		0.000	06:23
Air Blank		0.000	06:25
Subject Sample	# 2	0.125	06:26
Breath Volume		2.480	06:26
Air Blank		0.000	06:26
Control Test		0.081	06:27
Air Blank		0.000	06:27
Diagnostics Check		OK	06:27

Data Download Date/Time: 01/11/2022 16:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: DUNKLE

DATE/TIME OF TEST: 01/09/2022 19:58:42

Diagnostics Check		OK	19:49
Air Blank		0.000	19:50
Control Test		0.081	19:50
Air Blank		0.000	19:50
Subject Sample	# 1	0.227	19:52
Breath Volume		1.140	19:52
Air Blank		0.000	19:52
Air Blank		0.000	19:54
Subject Sample	# 2	0.217	19:57
Breath Volume		1.441	19:57
Air Blank		0.000	19:57
Control Test		0.081	19:58
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 01/11/2022 16:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHIRSTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHIRSTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: MERA ASIPALI

DATE/TIME OF TEST: 01/10/2022 03:41:42

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.081	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.138	03:37
Breath Volume		2.039	03:37
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.138	03:40
Breath Volume		1.808	03:40
Air Blank		0.000	03:40
Control Test		0.081	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 01/11/2022 16:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ORR

DATE/TIME OF TEST: 01/11/2022 00:17:19

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:06	
Control Test		0.080	00:06	
Air Blank		0.000	00:07	
Subject Sample	# 1	0.155	00:10	Volume Not Met
Breath Volume		0.035	00:10	
Air Blank		0.000	00:10	
Air Blank		0.000	00:12	
Subject Sample	# 2	0.189	00:15	Volume Not Met
Breath Volume		0.179	00:15	
Air Blank		0.000	00:16	
Control Test		0.079	00:16	
Air Blank		0.000	00:17	
Diagnostics Check		OK	00:17	

Data Download Date/Time: 01/11/2022 16:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MCLAUCHLIN

DATE/TIME OF TEST: 01/11/2022 03:29:03

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:18	
Control Test		0.082	03:18	
Air Blank		0.000	03:18	
Subject Sample	# 1		03:22	No Sample Provided
Breath Volume		0.000	03:22	
Air Blank		0.000	03:22	
Air Blank		0.000	03:24	
Subject Sample	# 2		03:27	No Sample Provided
Breath Volume		0.000	03:27	
Air Blank		0.000	03:28	
Control Test		0.081	03:28	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 01/11/2022 16:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: BEZADA

DATE/TIME OF TEST: 01/11/2022 04:51:56

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:41	
Control Test		0.081	04:41	
Air Blank		0.000	04:41	
Subject Sample	# 1		04:44	No Sample Provided
Breath Volume		0.000	04:44	
Air Blank		0.000	04:45	
Air Blank		0.000	04:47	
Subject Sample	# 2		04:50	No Sample Provided
Breath Volume		0.000	04:50	
Air Blank		0.000	04:51	
Control Test		0.081	04:51	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 01/11/2022 16:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/11/2022 22:05:58

Diagnostics Check		OK	21:54	
Air Blank		0.000	21:55	
Control Test		0.081	21:55	
Air Blank		0.000	21:55	
Subject Sample	# 1		21:59	No Sample Provided
Breath Volume		0.000	21:59	
Air Blank		0.000	21:59	
Air Blank		0.000	22:01	
Subject Sample	# 2		22:04	No Sample Provided
Breath Volume		0.000	22:04	
Air Blank		0.000	22:05	
Control Test		0.081	22:05	
Air Blank		0.000	22:05	
Diagnostics Check		OK	22:05	

Data Download Date/Time: 02/16/2022 08:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HORNER

DATE/TIME OF TEST: 01/12/2022 04:14:35

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.144	04:09
Breath Volume		1.199	04:09
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.141	04:13
Breath Volume		1.148	04:13
Air Blank		0.000	04:13
Control Test		0.078	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 02/16/2022 08:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHIRSTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHIRSTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 01/14/2022 19:48:06

Diagnostics Check		OK	19:36	
Air Blank		0.000	19:37	
Control Test		0.081	19:37	
Air Blank		0.000	19:38	
Subject Sample	# 1		19:41	No Sample Provided
Breath Volume		0.000	19:41	
Air Blank		0.000	19:41	
Air Blank		0.000	19:43	
Subject Sample	# 2		19:46	No Sample Provided
Breath Volume		0.000	19:46	
Air Blank		0.000	19:47	
Control Test		0.082	19:47	
Air Blank		0.000	19:48	
Diagnostics Check		OK	19:48	

Data Download Date/Time: 02/16/2022 08:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: RAMIREZ AVIA

DATE/TIME OF TEST: 01/14/2022 21:26:35

Diagnostics Check		OK	21:20
Air Blank		0.000	21:20
Control Test		0.081	21:21
Air Blank		0.000	21:21
Subject Sample	# 1	0.127	21:22
Breath Volume		1.132	21:22
Air Blank		0.000	21:22
Air Blank		0.000	21:24
Subject Sample	# 2	0.121	21:25
Breath Volume		1.722	21:25
Air Blank		0.000	21:25
Control Test		0.081	21:26
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 02/16/2022 08:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRSTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/15/2022 00:18:20

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:07	
Control Test		0.081	00:07	
Air Blank		0.000	00:08	
Subject Sample	# 1		00:11	No Sample Provided
Breath Volume		0.000	00:11	
Air Blank		0.000	00:11	
Air Blank		0.000	00:13	
Subject Sample	# 2		00:16	No Sample Provided
Breath Volume		0.000	00:16	
Air Blank		0.000	00:17	
Control Test		0.082	00:17	
Air Blank		0.000	00:18	
Diagnostics Check		OK	00:18	

Data Download Date/Time: 02/16/2022 08:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RIGGALL,NICK
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: FRANCISCO-ARTURO

DATE/TIME OF TEST: 01/15/2022 00:52:11

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.082	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.191	00:47
Breath Volume		2.070	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.184	00:50
Breath Volume		1.839	00:50
Air Blank		0.000	00:51
Control Test		0.081	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 02/16/2022 08:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: RUIZ PAGAN

DATE/TIME OF TEST: 01/15/2022 02:34:59

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.082	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.142	02:30
Breath Volume		2.554	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.141	02:33
Breath Volume		2.839	02:33
Air Blank		0.000	02:34
Control Test		0.081	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 02/16/2022 08:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: PAULY

DATE/TIME OF TEST: 01/15/2022 05:19:17

Diagnostics Check		OK	05:07	
Air Blank		0.000	05:08	
Control Test		0.082	05:08	
Air Blank		0.000	05:09	
Subject Sample	# 1		05:12	No Sample Provided
Breath Volume		0.000	05:12	
Air Blank		0.000	05:12	
Air Blank		0.000	05:14	
Subject Sample	# 2		05:17	No Sample Provided
Breath Volume		0.000	05:17	
Air Blank		0.000	05:18	
Control Test		0.082	05:18	
Air Blank		0.000	05:19	
Diagnostics Check		OK	05:19	

Data Download Date/Time: 02/16/2022 08:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 01/15/2022 23:49:18

Diagnostics Check		OK	23:34	
Air Blank		0.000	23:35	
Control Test		0.081	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 1	0.174	23:36	
Breath Volume		1.105	23:36	
Air Blank		0.000	23:37	
Air Blank		0.000	23:39	
Subject Sample	# 2	0.160	23:42	Slope Not Level
Breath Volume		1.371	23:42	
Air Blank		0.000	23:43	
Air Blank		0.000	23:44	
Subject Sample	# 3	0.168	23:47	
Breath Volume		1.429	23:47	
Air Blank		0.000	23:48	
Control Test		0.079	23:48	
Air Blank		0.000	23:49	
Diagnostics Check		OK	23:49	

Data Download Date/Time: 02/16/2022 08:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: AVONCE

DATE/TIME OF TEST: 01/16/2022 01:26:52

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.080	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:19	No Sample Provided
Breath Volume		0.000	01:19	
Air Blank		0.000	01:20	
Air Blank		0.000	01:22	
Subject Sample	# 2		01:25	No Sample Provided
Breath Volume		0.000	01:25	
Air Blank		0.000	01:25	
Control Test		0.081	01:26	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 02/16/2022 08:02

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DEL CASTILLO SUAREZ

DATE/TIME OF TEST: 01/16/2022 03:21:57

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.081	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.065	03:17
Breath Volume		2.769	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.064	03:20
Breath Volume		2.839	03:20
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 02/16/2022 08:03

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,LEONARDO,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DE LA CRUZ,LEONARDO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: ADAMES

DATE/TIME OF TEST: 01/17/2022 05:14:11

Diagnostics Check		OK	05:11	
Air Blank		0.000	05:11	
Control Test		0.082	05:12	
Air Blank		0.000	05:12	
Subject Sample	# 1		05:12	Subject Test Refused
Breath Volume		0.000	05:12	
Air Blank		0.000	05:13	
Control Test		0.082	05:13	
Air Blank		0.000	05:14	
Diagnostics Check		OK	05:14	

Data Download Date/Time: 02/16/2022 08:03

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: FRIERSON

DATE/TIME OF TEST: 01/17/2022 18:58:39

Diagnostics Check		OK	18:47	
Air Blank		0.000	18:47	
Control Test		0.081	18:47	
Air Blank		0.000	18:48	
Subject Sample	# 1	0.200	18:51	Volume Not Met
Breath Volume		0.800	18:51	
Air Blank		0.000	18:52	
Air Blank		0.000	18:53	
Subject Sample	# 2	0.184	18:57	Volume Not Met
Breath Volume		0.691	18:57	
Air Blank		0.000	18:57	
Control Test		0.080	18:58	
Air Blank		0.000	18:58	
Diagnostics Check		OK	18:58	

Data Download Date/Time: 02/16/2022 08:03

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: TAGURI

DATE/TIME OF TEST: 01/19/2022 19:18:44

Diagnostics Check		OK	19:08
Air Blank		0.000	19:08
Control Test		0.081	19:09
Air Blank		0.000	19:09
Subject Sample	# 1	0.133	19:11
Breath Volume		1.238	19:11
Air Blank		0.000	19:12
Air Blank		0.000	19:14
Subject Sample	# 2	0.124	19:17
Breath Volume		1.375	19:17
Air Blank		0.000	19:17
Control Test		0.080	19:18
Air Blank		0.000	19:18
Diagnostics Check		OK	19:18

Data Download Date/Time: 02/16/2022 08:03

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 01/20/2022 22:39:02

Diagnostics Check		OK	22:31
Air Blank		0.000	22:32
Control Test		0.080	22:32
Air Blank		0.000	22:32
Subject Sample	# 1	0.181	22:33
Breath Volume		1.566	22:33
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 2	0.182	22:37
Breath Volume		1.515	22:37
Air Blank		0.000	22:38
Control Test		0.081	22:38
Air Blank		0.000	22:38
Diagnostics Check		OK	22:39

Data Download Date/Time: 02/16/2022 08:03

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: KISKINIS

DATE/TIME OF TEST: 01/20/2022 23:54:54

Diagnostics Check		OK	23:43	
Air Blank		0.000	23:44	
Control Test		0.078	23:44	
Air Blank		0.000	23:44	
Subject Sample	# 1		23:47	No Sample Provided
Breath Volume		0.000	23:47	
Air Blank		0.000	23:48	
Air Blank		0.000	23:50	
Subject Sample	# 2		23:53	No Sample Provided
Breath Volume		0.000	23:53	
Air Blank		0.000	23:54	
Control Test		0.079	23:54	
Air Blank		0.000	23:54	
Diagnostics Check		OK	23:54	

Data Download Date/Time: 02/16/2022 08:03

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON JR,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: STEPHENS,TONY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/21/2022 21:37:40

Diagnostics Check		OK	21:30
Air Blank		0.000	21:30
Control Test		0.079	21:31
Air Blank		0.000	21:31
Subject Sample	# 1	0.122	21:33
Breath Volume		1.464	21:33
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.119	21:36
Breath Volume		1.644	21:36
Air Blank		0.000	21:36
Control Test		0.079	21:37
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 02/16/2022 08:04

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH,PATTY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BLACKMAN,GERIG
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:48

SUBJECT'S LAST NAME: BUDDECKE

DATE/TIME OF TEST: 01/22/2022 12:38:03

Diagnostics Check		OK	12:30
Air Blank		0.000	12:31
Control Test		0.080	12:31
Air Blank		0.000	12:31
Subject Sample	# 1	0.224	12:32
Breath Volume		1.484	12:32
Air Blank		0.000	12:32
Air Blank		0.000	12:34
Subject Sample	# 2	0.210	12:36
Breath Volume		1.589	12:36
Air Blank		0.000	12:37
Control Test		0.080	12:37
Air Blank		0.000	12:37
Diagnostics Check		OK	12:38

Data Download Date/Time: 02/16/2022 08:04

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTOSH,HALLIE,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: HASKINS

DATE/TIME OF TEST: 01/23/2022 02:05:53

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:54	
Control Test		0.080	01:55	
Air Blank		0.000	01:55	
Subject Sample	# 1		01:58	No Sample Provided
Breath Volume		0.000	01:58	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 2		02:04	No Sample Provided
Breath Volume		0.000	02:04	
Air Blank		0.000	02:04	
Control Test		0.078	02:05	
Air Blank		0.000	02:05	
Diagnostics Check		OK	02:05	

Data Download Date/Time: 02/16/2022 08:04

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARBANAS,SHANE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: TRINIDAD

DATE/TIME OF TEST: 01/24/2022 00:10:25

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.079	00:05
Air Blank		0.000	00:05
Subject Sample	# 1	0.186	00:06
Breath Volume		3.425	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.185	00:08
Breath Volume		2.773	00:08
Air Blank		0.000	00:09
Control Test		0.077	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 02/16/2022 08:04

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: CORREA ZULETA

DATE/TIME OF TEST: 01/25/2022 22:33:09

Diagnostics Check		OK	22:26
Air Blank		0.000	22:26
Control Test		0.081	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.168	22:28
Breath Volume		1.707	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.183	22:31
Breath Volume		2.394	22:31
Air Blank		0.000	22:32
Control Test		0.080	22:32
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 02/16/2022 08:04

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 01/26/2022 03:52:08

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.081	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:45	No Sample Provided
Breath Volume		0.000	03:45	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 2		03:50	No Sample Provided
Breath Volume		0.000	03:50	
Air Blank		0.000	03:51	
Control Test		0.080	03:51	
Air Blank		0.000	03:52	
Diagnostics Check		OK	03:52	

Data Download Date/Time: 02/16/2022 08:04

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,RICHARD,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SAFIEDDINE,JAMAL
ARRESTING OFFICER AGENCY: Davenport PD
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: ZARBA

DATE/TIME OF TEST: 01/26/2022 20:06:42

Diagnostics Check		OK	20:00
Air Blank		0.000	20:00
Control Test		0.080	20:01
Air Blank		0.000	20:01
Subject Sample	# 1	0.109	20:02
Breath Volume		3.503	20:02
Air Blank		0.000	20:02
Air Blank		0.000	20:04
Subject Sample	# 2	0.109	20:05
Breath Volume		3.289	20:05
Air Blank		0.000	20:05
Control Test		0.080	20:06
Air Blank		0.000	20:06
Diagnostics Check		OK	20:06

Data Download Date/Time: 02/16/2022 08:04

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 01/27/2022 01:39:28

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:25	
Control Test		0.078	01:26	
Air Blank		0.000	01:26	
Subject Sample	# 1		01:29	No Sample Provided
Breath Volume		0.000	01:29	
Air Blank		0.000	01:30	
Air Blank		0.000	01:32	
Subject Sample	# 2	0.187	01:34	
Breath Volume		1.382	01:34	
Air Blank		0.000	01:35	
Air Blank		0.000	01:37	
Subject Sample	# 3	0.179	01:38	
Breath Volume		1.136	01:38	
Air Blank		0.000	01:38	
Control Test		0.077	01:38	
Air Blank		0.000	01:39	
Diagnostics Check		OK	01:39	

Data Download Date/Time: 02/16/2022 08:04

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: STINE

DATE/TIME OF TEST: 01/28/2022 23:25:50

Diagnostics Check		OK	23:14	
Air Blank		0.000	23:14	
Control Test		0.081	23:15	
Air Blank		0.000	23:15	
Subject Sample	# 1		23:18	No Sample Provided
Breath Volume		0.000	23:18	
Air Blank		0.000	23:19	
Air Blank		0.000	23:21	
Subject Sample	# 2		23:24	No Sample Provided
Breath Volume		0.000	23:24	
Air Blank		0.000	23:24	
Control Test		0.081	23:25	
Air Blank		0.000	23:25	
Diagnostics Check		OK	23:25	

Data Download Date/Time: 02/16/2022 08:04

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: WATHEN

DATE/TIME OF TEST: 01/29/2022 03:11:09

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.081	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.000	03:06
Breath Volume		2.367	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.000	03:09
Breath Volume		2.898	03:09
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 02/16/2022 08:05

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WELLDEN,BRADLEY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: FINK

DATE/TIME OF TEST: 01/29/2022 03:19:34

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.081	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.142	03:15
Breath Volume		3.238	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.139	03:18
Breath Volume		2.597	03:18
Air Blank		0.000	03:18
Control Test		0.081	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 02/16/2022 08:05

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: AMY

DATE/TIME OF TEST: 01/30/2022 04:12:37

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.081	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.140	04:08
Breath Volume		1.671	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.139	04:11
Breath Volume		1.808	04:11
Air Blank		0.000	04:11
Control Test		0.081	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 02/16/2022 08:05

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROPST,JOHN,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: VILLAGRAN,DANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:19

SUBJECT'S LAST NAME: TREJO

DATE/TIME OF TEST: 01/30/2022 13:51:22

Diagnostics Check		OK	13:44
Air Blank		0.000	13:44
Control Test		0.082	13:45
Air Blank		0.000	13:45
Subject Sample	# 1	0.163	13:46
Breath Volume		1.718	13:46
Air Blank		0.000	13:47
Air Blank		0.000	13:49
Subject Sample	# 2	0.161	13:49
Breath Volume		1.777	13:49
Air Blank		0.000	13:50
Control Test		0.082	13:50
Air Blank		0.000	13:51
Diagnostics Check		OK	13:51

Data Download Date/Time: 02/16/2022 08:05

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROPST,JOHN,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 01/30/2022 16:26:31

Diagnostics Check		OK	16:15	
Air Blank		0.000	16:15	
Control Test		0.082	16:15	
Air Blank		0.000	16:16	
Subject Sample	# 1		16:19	No Sample Provided
Breath Volume		0.000	16:19	
Air Blank		0.000	16:20	
Air Blank		0.000	16:22	
Subject Sample	# 2		16:25	No Sample Provided
Breath Volume		0.000	16:25	
Air Blank		0.000	16:25	
Control Test		0.082	16:25	
Air Blank		0.000	16:26	
Diagnostics Check		OK	16:26	

Data Download Date/Time: 02/16/2022 08:05

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/30/2022 19:55:56

Diagnostics Check		OK	19:49
Air Blank		0.000	19:49
Control Test		0.081	19:50
Air Blank		0.000	19:50
Subject Sample	# 1	0.244	19:51
Breath Volume		2.316	19:51
Air Blank		0.000	19:51
Air Blank		0.000	19:53
Subject Sample	# 2	0.241	19:54
Breath Volume		3.031	19:54
Air Blank		0.000	19:55
Control Test		0.079	19:55
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 02/16/2022 08:05

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: HENRIES

DATE/TIME OF TEST: 01/30/2022 21:44:56

Diagnostics Check		OK	21:38
Air Blank		0.000	21:38
Control Test		0.081	21:38
Air Blank		0.000	21:39
Subject Sample	# 1	0.138	21:40
Breath Volume		1.625	21:40
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.145	21:43
Breath Volume		1.964	21:43
Air Blank		0.000	21:44
Control Test		0.081	21:44
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 02/16/2022 08:05

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MOATES,AUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/30/2022 22:15:48

Diagnostics Check		OK	22:03
Air Blank		0.000	22:04
Control Test		0.082	22:04
Air Blank		0.000	22:04
Subject Sample	# 1	0.247	22:05
Breath Volume		1.117	22:05
Air Blank		0.000	22:06
Air Blank		0.000	22:07
Subject Sample	# 2	0.291	22:11
Breath Volume		2.957	22:11
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 3	0.273	22:14
Breath Volume		1.960	22:14
Air Blank		0.000	22:14
Control Test		0.076	22:15
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 02/16/2022 08:05

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: SKYBERG

DATE/TIME OF TEST: 01/30/2022 22:46:35

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:35	
Control Test		0.080	22:36	
Air Blank		0.000	22:36	
Subject Sample	# 1		22:39	No Sample Provided
Breath Volume		0.000	22:39	
Air Blank		0.000	22:40	
Air Blank		0.000	22:42	
Subject Sample	# 2		22:45	No Sample Provided
Breath Volume		0.000	22:45	
Air Blank		0.000	22:45	
Control Test		0.081	22:46	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 02/16/2022 08:05

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BAGGETT,TYLER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: ANGERSON

DATE/TIME OF TEST: 01/31/2022 02:05:23

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:03	
Control Test		0.082	02:03	
Air Blank		0.000	02:03	
Subject Sample	# 1		02:04	Subject Test Refused
Breath Volume		0.000	02:04	
Air Blank		0.000	02:04	
Control Test		0.082	02:04	
Air Blank		0.000	02:05	
Diagnostics Check		OK	02:05	

Data Download Date/Time: 02/16/2022 08:05

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,RICHARD,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: VANNOSTRAND

DATE/TIME OF TEST: 01/31/2022 22:16:44

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.080	22:11
Air Blank		0.000	22:11
Subject Sample	# 1	0.194	22:12
Breath Volume		3.031	22:12
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.180	22:15
Breath Volume		1.808	22:15
Air Blank		0.000	22:15
Control Test		0.080	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 02/16/2022 08:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: TASKA,TIERRA
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 01/21/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 01/26/2022 22:30:18

Diagnostics Check		OK	22:23
Air Blank		0.000	22:24
Control Test		0.077	22:24
Air Blank		0.000	22:24
Subject Sample	# 1	0.247	22:25
Breath Volume		1.855	22:25
Air Blank		0.000	22:26
Air Blank		0.000	22:28
Subject Sample	# 2	0.244	22:28
Breath Volume		1.628	22:28
Air Blank		0.000	22:29
Control Test		0.077	22:29
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 02/28/2022 12:03

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.