

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAUDLE,PAUL,A

BREATH TEST OPERATOR AGENCY: Sumter County SO

ARRESTING OFFICER: DAVENPORT,J

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 181632

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 00:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 04/01/2021 01:53:26

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.138	01:48
Breath Volume	1.632	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.145	01:51
Breath Volume	1.351	01:51
Air Blank	0.000	01:52
Control Test	0.077	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 04/28/2021 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOLLIS,TERENCE,L

BREATH TEST OPERATOR AGENCY: Sumter County SO

ARRESTING OFFICER: LETSON,R

ARRESTING OFFICER AGENCY: Sumter County SO

LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 181632

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 00:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: COSTELLO

DATE/TIME OF TEST: 04/03/2021 02:03:45

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.058	01:58
Breath Volume	1.445	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.059	02:02
Breath Volume	1.980	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 04/28/2021 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCASTER,CODY,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: UNEBERG,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:49

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 04/10/2021 09:22:51

Diagnostics Check	OK	09:09	
Air Blank	0.000	09:10	
Control Test	0.080	09:10	
Air Blank	0.000	09:10	
Subject Sample	# 1 0.130	09:12	
Breath Volume	1.324	09:12	
Air Blank	0.000	09:13	
Air Blank	0.000	09:15	
Subject Sample	# 2 0.133	09:18	Volume Not Met
Breath Volume	0.050	09:18	
Air Blank	0.000	09:18	
Air Blank	0.000	09:20	
Subject Sample	# 3 0.129	09:21	
Breath Volume	1.601	09:21	
Air Blank	0.000	09:21	
Control Test	0.077	09:22	
Air Blank	0.000	09:22	
Diagnostics Check	OK	09:22	

Data Download Date/Time: 04/28/2021 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 04/12/2021 21:43:01

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.080	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.054	21:38
Breath Volume	1.703	21:38
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.054	21:41
Breath Volume	1.632	21:41
Air Blank	0.000	21:42
Control Test	0.080	21:42
Air Blank	0.000	21:42
Diagnostics Check	OK	21:43

Data Download Date/Time: 04/28/2021 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 181632
BREATH TEST OPERATOR: SMITH,WYATT,M	DRY GAS EXPIRY DATE: 10/05/2021
BREATH TEST OPERATOR AGENCY: Sumter County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: SOUDERS,JARED	ARREST TIME: 22:55
ARRESTING OFFICER AGENCY: Sumter County SO	GENDER: Female
LAST AGENCY INSPECTION: 03/29/2021	OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: BORGMAN	DATE/TIME OF TEST: 04/13/2021	00:17:51
------------------------------	-------------------------------	----------

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:04	
Control Test	0.079	00:04	
Air Blank	0.000	00:05	
Subject Sample	# 1 0.031	00:08	Volume Not Met
Breath Volume	1.015	00:08	
Air Blank	0.000	00:09	
Air Blank	0.000	00:10	
Subject Sample	# 2 0.031	00:13	
Breath Volume	1.699	00:13	
Air Blank	0.000	00:13	
Air Blank	0.000	00:15	
Subject Sample	# 3 0.028	00:16	
Breath Volume	1.246	00:16	
Air Blank	0.000	00:16	
Control Test	0.077	00:17	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 04/28/2021 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ISHALL,JALEN,L

BREATH TEST OPERATOR AGENCY: Sumter County SO

ARRESTING OFFICER: NORDLE,LACY

ARRESTING OFFICER AGENCY: Sumter County SO

LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 181632

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 20:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 04/14/2021 21:38:13

Diagnostics Check	OK	21:28
Air Blank	0.000	21:29
Control Test	0.079	21:29
Air Blank	0.000	21:29
Subject Sample # 1	0.000	21:32
Breath Volume	2.085	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.000	21:36
Breath Volume	2.812	21:36
Air Blank	0.000	21:37
Control Test	0.078	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 04/28/2021 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHALL,JALEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: COOK,MICHAEL
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MACDONALD

DATE/TIME OF TEST: 04/15/2021 01:51:30

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.030	01:46
Breath Volume	1.824	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.034	01:49
Breath Volume	2.378	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 04/28/2021 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ISHALL,JALEN,L

BREATH TEST OPERATOR AGENCY: Sumter County SO

ARRESTING OFFICER: HELM,DEREK

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 181632

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 00:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: OZKUL

DATE/TIME OF TEST: 04/16/2021 05:03:35

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:01	
Control Test	0.079	05:01	
Air Blank	0.000	05:02	
Subject Sample	# 1	05:02	Subject Test Refused
Breath Volume	0.000	05:02	
Air Blank	0.000	05:02	
Control Test	0.079	05:02	
Air Blank	0.000	05:03	
Diagnostics Check	OK	05:03	

Data Download Date/Time: 04/28/2021 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,WYATT,M

BREATH TEST OPERATOR AGENCY: Sumter County SO

ARRESTING OFFICER: KELLY,W

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 181632

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 18:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/16/2021 19:40:13

Diagnostics Check	OK	19:33
Air Blank	0.000	19:33
Control Test	0.080	19:34
Air Blank	0.000	19:34
Subject Sample # 1	0.137	19:35
Breath Volume	1.679	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.136	19:38
Breath Volume	1.738	19:38
Air Blank	0.000	19:39
Control Test	0.079	19:39
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

Data Download Date/Time: 04/28/2021 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRACKETT,ALAN,M

BREATH TEST OPERATOR AGENCY: Sumter County SO

ARRESTING OFFICER: KJEKSTAD,ERIC

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 181632

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 09:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 11:06

SUBJECT'S LAST NAME: MYLES

DATE/TIME OF TEST: 04/21/2021 11:34:56

Diagnostics Check	OK	11:28
Air Blank	0.000	11:28
Control Test	0.079	11:29
Air Blank	0.000	11:29
Subject Sample # 1	0.082	11:30
Breath Volume	2.843	11:30
Air Blank	0.000	11:30
Air Blank	0.000	11:32
Subject Sample # 2	0.083	11:33
Breath Volume	3.160	11:33
Air Blank	0.000	11:34
Control Test	0.079	11:34
Air Blank	0.000	11:34
Diagnostics Check	OK	11:34

Data Download Date/Time: 04/28/2021 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 181632
BREATH TEST OPERATOR: PEREZ,DENISE,	DRY GAS EXPIRY DATE: 10/05/2021
BREATH TEST OPERATOR AGENCY: Sumter County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ROSS,CALEB	ARREST TIME: 19:54
ARRESTING OFFICER AGENCY: Sumter County SO	GENDER: Male
LAST AGENCY INSPECTION: 03/29/2021	OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: HUGGINS	DATE/TIME OF TEST: 04/25/2021 20:37:05
------------------------------	--

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:34	
Control Test	0.081	20:34	
Air Blank	0.000	20:35	
Subject Sample	# 1	20:35	Subject Test Refused
Breath Volume	0.000	20:35	
Air Blank	0.000	20:36	
Control Test	0.080	20:36	
Air Blank	0.000	20:37	
Diagnostics Check	OK	20:37	

Data Download Date/Time: 04/28/2021 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ, DENISE,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LARGE, MICHAEL
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: SUSSNER

DATE/TIME OF TEST: 04/25/2021 21:22:42

Diagnostics Check	OK	21:19	
Air Blank	0.000	21:20	
Control Test	0.080	21:20	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:21	Subject Test Refused
Breath Volume	0.000	21:21	
Air Blank	0.000	21:21	
Control Test	0.081	21:22	
Air Blank	0.000	21:22	
Diagnostics Check	OK	21:22	

Data Download Date/Time: 04/28/2021 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,WYATT,M

BREATH TEST OPERATOR AGENCY: Sumter County SO

ARRESTING OFFICER: LETSON,REBECCA

ARRESTING OFFICER AGENCY: Sumter County SO

LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 181632

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 02:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/28/2021 04:10:26

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.000	04:06
Breath Volume	2.449	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.000	04:09
Breath Volume	2.351	04:09
Air Blank	0.000	04:09
Control Test	0.078	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 04/28/2021 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLON,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: DILLON,JUSTIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: GEDDES

DATE/TIME OF TEST: 04/02/2021 01:19:13

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.081	01:06	
Air Blank	0.000	01:06	
Subject Sample	# 1 0.108	01:10	Volume Not Met
Breath Volume	0.796	01:10	
Air Blank	0.000	01:10	
Air Blank	0.000	01:12	
Subject Sample	# 2 0.127	01:13	
Breath Volume	1.761	01:13	
Air Blank	0.000	01:14	
Air Blank	0.000	01:15	
Subject Sample	# 3 0.127	01:17	
Breath Volume	1.226	01:17	
Air Blank	0.000	01:18	
Control Test	0.080	01:18	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 04/13/2021 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 04/09/2021 04:15:54

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.121	04:11
Breath Volume	1.343	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.135	04:14
Breath Volume	1.585	04:14
Air Blank	0.000	04:14
Control Test	0.079	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 04/13/2021 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,DENZEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALEXANDER,DENZEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:22

SUBJECT'S LAST NAME: SENSANBAUGH

DATE/TIME OF TEST: 04/10/2021 10:51:17

Diagnostics Check	OK	10:48	
Air Blank	0.000	10:48	
Control Test	0.079	10:49	
Air Blank	0.000	10:49	
Subject Sample	# 1	10:49	Subject Test Refused
Breath Volume	0.000	10:49	
Air Blank	0.000	10:50	
Control Test	0.079	10:50	
Air Blank	0.000	10:51	
Diagnostics Check	OK	10:51	

Data Download Date/Time: 04/13/2021 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CARLTON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LITZKOW,CARLTON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/11/2021 01:55:04

Diagnostics Check	OK	01:54	
Air Blank		01:54	RFI Detect
Air Blank	0.000	01:55	

Data Download Date/Time: 04/13/2021 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CARLTON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LITZKOW,CARLTON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/11/2021 02:05:26

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.153	02:00
Breath Volume	1.128	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:02
Subject Sample # 2	0.145	02:03
Breath Volume	1.300	02:03
Air Blank	0.000	02:04
Control Test	0.075	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 04/13/2021 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,JONATHON,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BENNETT,JONATHON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 04/15/2021 18:35:10

Diagnostics Check	OK	18:34	
Air Blank	0.000	18:34	
Control Test	0.000	18:34	RFI Detect
Air Blank	0.000	18:35	

Data Download Date/Time: 05/18/2021 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,JONATHON,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BENNETT,JONATHON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 04/15/2021 18:40:04

Diagnostics Check	OK	18:37	
Air Blank	0.000	18:37	
Control Test	0.079	18:37	
Air Blank	0.000	18:38	
Subject Sample	# 1	18:38	Subject Test Refused
Breath Volume	0.000	18:38	
Air Blank	0.000	18:39	
Control Test	0.079	18:39	
Air Blank	0.000	18:39	
Diagnostics Check	OK	18:40	

Data Download Date/Time: 05/18/2021 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BENJAMIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: GONZALEZ,KAITLIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/16/2021 05:27:30

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:22
Subject Sample # 1	0.000	05:23
Breath Volume	1.113	05:23
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.000	05:26
Breath Volume	1.777	05:26
Air Blank	0.000	05:26
Control Test	0.078	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 05/18/2021 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,DENZEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALEXANDER,DENZEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: DOW

DATE/TIME OF TEST: 04/20/2021 00:29:31

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:27	
Control Test	0.079	00:27	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:28	
Control Test	0.079	00:28	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 05/18/2021 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLE,MADDISON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BIERMAN,CODY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: CLEPPER

DATE/TIME OF TEST: 04/25/2021 21:31:07

Diagnostics Check	OK	21:19	
Air Blank	0.000	21:20	
Control Test	0.080	21:20	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:24	No Sample Provided
Breath Volume	0.000	21:24	
Air Blank	0.000	21:24	
Air Blank	0.000	21:26	
Subject Sample	# 2	21:29	No Sample Provided
Breath Volume	0.000	21:29	
Air Blank	0.000	21:30	
Control Test	0.080	21:30	
Air Blank	0.000	21:31	
Diagnostics Check	OK	21:31	

Data Download Date/Time: 05/18/2021 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITMAN,AARON,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MUNDT,BRANDON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/01/2021 19:12:44

Diagnostics Check	OK	19:06
Air Blank	0.000	19:06
Control Test	0.081	19:07
Air Blank	0.000	19:07
Subject Sample # 1	0.000	19:08
Breath Volume	5.250	19:08
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 2	0.000	19:11
Breath Volume	4.816	19:11
Air Blank	0.000	19:11
Control Test	0.079	19:12
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time: 04/28/2021 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO, WESTON, C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MORENO, WESTON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 07:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:48

SUBJECT'S LAST NAME: DOWNS

DATE/TIME OF TEST: 04/02/2021 08:18:06

Diagnostics Check	OK	08:11
Air Blank	0.000	08:11
Control Test	0.081	08:11
Air Blank	0.000	08:12
Subject Sample # 1	0.073	08:13
Breath Volume	3.804	08:13
Air Blank	0.000	08:14
Air Blank	0.000	08:15
Subject Sample # 2	0.070	08:16
Breath Volume	3.972	08:16
Air Blank	0.000	08:17
Control Test	0.079	08:17
Air Blank	0.000	08:18
Diagnostics Check	OK	08:18

Data Download Date/Time: 04/28/2021 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: OGLESBY

DATE/TIME OF TEST: 04/02/2021 20:31:26

Diagnostics Check	OK	20:22	
Air Blank	0.000	20:22	
Control Test	0.082	20:23	
Air Blank	0.000	20:23	
Subject Sample	# 1	20:26	No Sample Provided
Breath Volume	0.000	20:26	
Air Blank	0.000	20:27	
Air Blank	0.000	20:29	
Subject Sample	# 2	20:30	Subject Test Refused
Breath Volume	0.000	20:30	
Air Blank	0.000	20:30	
Control Test	0.082	20:30	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 04/28/2021 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWEN,NATHANIEL,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: JOHNS,W
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: SHELBY

DATE/TIME OF TEST: 04/02/2021 23:44:46

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.081	23:39
Air Blank	0.000	23:39
Subject Sample # 1	0.197	23:40
Breath Volume	2.519	23:40
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.195	23:43
Breath Volume	2.140	23:43
Air Blank	0.000	23:43
Control Test	0.081	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 04/28/2021 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITMAN,AARON,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FULFORD,MYLON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ARELLANO

DATE/TIME OF TEST: 04/10/2021 00:54:33

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.177	00:49
Breath Volume	2.484	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.175	00:52
Breath Volume	2.468	00:52
Air Blank	0.000	00:53
Control Test	0.076	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 04/28/2021 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUS,CHRISTINA,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUS,CHRISTINA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: SHIRLEY

DATE/TIME OF TEST: 04/11/2021 03:47:16

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.118	03:42
Breath Volume	2.535	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.115	03:45
Breath Volume	1.523	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 04/28/2021 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: VERDUGHT

DATE/TIME OF TEST: 04/13/2021 01:45:19

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:37	
Control Test	0.080	01:37	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:41	No Sample Provided
Breath Volume	0.000	01:41	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample	# 2	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.080	01:44	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 04/28/2021 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 15:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: TUERS

DATE/TIME OF TEST: 04/13/2021 17:03:06

Diagnostics Check	OK	16:56
Air Blank	0.000	16:56
Control Test	0.081	16:56
Air Blank	0.000	16:57
Subject Sample # 1	0.232	16:58
Breath Volume	2.933	16:58
Air Blank	0.000	16:59
Air Blank	0.000	17:00
Subject Sample # 2	0.224	17:01
Breath Volume	3.964	17:01
Air Blank	0.000	17:02
Control Test	0.077	17:02
Air Blank	0.000	17:03
Diagnostics Check	OK	17:03

Data Download Date/Time: 04/28/2021 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRANNON,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SMALL,BRANNON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: YARDLEY

DATE/TIME OF TEST: 04/15/2021 05:58:00

Diagnostics Check	OK	05:52
Air Blank	0.000	05:52
Control Test	0.081	05:52
Air Blank	0.000	05:53
Subject Sample # 1	0.000	05:53
Breath Volume	3.132	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.000	05:56
Breath Volume	1.570	05:56
Air Blank	0.000	05:57
Control Test	0.078	05:57
Air Blank	0.000	05:57
Diagnostics Check	OK	05:58

Data Download Date/Time: 04/28/2021 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRANNON,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SMALL,BRANNON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:16

SUBJECT'S LAST NAME: MYLENBUSCH

DATE/TIME OF TEST: 04/23/2021 15:49:11

Diagnostics Check	OK	15:42
Air Blank	0.000	15:43
Control Test	0.080	15:43
Air Blank	0.000	15:44
Subject Sample # 1	0.167	15:44
Breath Volume	2.355	15:44
Air Blank	0.000	15:45
Air Blank	0.000	15:47
Subject Sample # 2	0.165	15:47
Breath Volume	1.898	15:47
Air Blank	0.000	15:48
Control Test	0.076	15:48
Air Blank	0.000	15:49
Diagnostics Check	OK	15:49

Data Download Date/Time: 04/28/2021 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERTENS, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MERTENS, WILLIAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: EDNEY

DATE/TIME OF TEST: 04/23/2021 22:20:53

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:11	
Control Test	0.080	22:12	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:15	No Sample Provided
Breath Volume	0.000	22:15	
Air Blank	0.000	22:16	
Air Blank	0.000	22:18	
Subject Sample	# 2	22:19	Subject Test Refused
Breath Volume	0.000	22:19	
Air Blank	0.000	22:19	
Control Test	0.080	22:20	
Air Blank	0.000	22:20	
Diagnostics Check	OK	22:20	

Data Download Date/Time: 04/28/2021 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILDE,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: SANCHEZ ROLDAN

DATE/TIME OF TEST: 04/29/2021 23:07:51

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.080	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.000	23:03
Breath Volume	1.261	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.000	23:06
Breath Volume	1.308	23:06
Air Blank	0.000	23:06
Control Test	0.080	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 05/27/2021 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILDE,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CHRISTEN

DATE/TIME OF TEST: 04/30/2021 02:22:47

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.152	02:18
Breath Volume	3.886	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.158	02:21
Breath Volume	4.097	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 05/27/2021 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACBETH,MATTHEW,R
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: EAKINS,SIERRA
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 04/01/2021 19:01:06

Diagnostics Check	OK	18:52
Air Blank	0.000	18:53
Control Test	0.078	18:53
Air Blank	0.000	18:53
Subject Sample # 1	0.218	18:55
Breath Volume	1.945	18:55
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.199	18:59
Breath Volume	1.484	18:59
Air Blank	0.000	19:00
Control Test	0.078	19:00
Air Blank	0.000	19:01
Diagnostics Check	OK	19:01

Data Download Date/Time: 04/26/2021 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SIMMONS,JUSTIN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: KALOHELANI

DATE/TIME OF TEST: 04/11/2021 03:43:41

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.078	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.251	03:39
Breath Volume	2.613	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.243	03:42
Breath Volume	2.390	03:42
Air Blank	0.000	03:42
Control Test	0.077	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 04/26/2021 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JAMIE,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WOODS,LINDSEY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: LINFORD-CIRCE

DATE/TIME OF TEST: 04/01/2021 04:33:39

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:28	
Control Test	0.081	04:28	
Air Blank	0.000	04:29	
Subject Sample	# 1 0.172	04:29	
Breath Volume	4.503	04:29	
Air Blank	0.000	04:30	
Air Blank	0.000	04:32	
Subject Sample	# 2 0.170	04:32	
Breath Volume	4.289	04:32	
Air Blank		04:33	RFI Detect
Air Blank	0.000	04:33	

Data Download Date/Time: 04/14/2021 21:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JAMIE,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WOODS,LINDSEY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: LINFORD-CIRCE

DATE/TIME OF TEST: 04/01/2021 04:42:50

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.082	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.168	04:38
Breath Volume	4.562	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:40
Subject Sample # 2	0.163	04:41
Breath Volume	3.398	04:41
Air Blank	0.000	04:41
Control Test	0.081	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 04/14/2021 21:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMAHON,KENNETH,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: TENA,MICHAEL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ZAVALA

DATE/TIME OF TEST: 04/03/2021 01:53:36

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.083	01:43	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:46	No Sample Provided
Breath Volume	0.000	01:46	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 2	01:52	No Sample Provided
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.082	01:53	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 04/14/2021 21:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMAHON,KENNETH,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: TENA,MICHAEL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ZAVALA

DATE/TIME OF TEST: 04/03/2021 02:19:13

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.082	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.129	02:14
Breath Volume	1.175	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample # 2	0.128	02:17
Breath Volume	2.968	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 04/14/2021 21:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARACLE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MARACLE,MATTHEW
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: HERBERT

DATE/TIME OF TEST: 04/13/2021 14:27:07

Diagnostics Check	OK	14:21
Air Blank	0.000	14:21
Control Test	0.081	14:22
Air Blank	0.000	14:22
Subject Sample # 1	0.000	14:22
Breath Volume	1.636	14:22
Air Blank	0.000	14:23
Air Blank	0.000	14:25
Subject Sample # 2	0.000	14:25
Breath Volume	1.722	14:25
Air Blank	0.000	14:26
Control Test	0.081	14:26
Air Blank	0.000	14:27
Diagnostics Check	OK	14:27

Data Download Date/Time: 04/14/2021 21:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JAMIE,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PINO,RICHARD
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 10:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:57

SUBJECT'S LAST NAME: BURTCHELL

DATE/TIME OF TEST: 04/15/2021 23:04:15

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:02	
Control Test	0.081	23:03	
Air Blank	0.000	23:03	
Subject Sample	# 1	23:03	Sequence Aborted
Breath Volume	0.000	23:03	
Air Blank	0.000	23:04	

Data Download Date/Time: 05/26/2021 20:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JAMIE,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PINO,RICHARD
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: BURTCHELL

DATE/TIME OF TEST: 04/15/2021 23:24:06

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.081	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.000	23:19
Breath Volume	2.972	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.000	23:22
Breath Volume	2.914	23:22
Air Blank	0.000	23:23
Control Test	0.082	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 05/26/2021 20:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFOUR,NOAH,I
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DUFOUR,NOAH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 04/17/2021 01:53:51

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:53	
Control Test	0.000	01:53	RFI Detect
Air Blank	0.000	01:53	

Data Download Date/Time: 05/26/2021 20:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFOUR,NOAH,I
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DUFOUR,NOAH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 04/17/2021 02:09:53

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.081	01:57	
Air Blank	0.000	01:57	
Subject Sample	# 1	02:00	No Sample Provided
Breath Volume	0.000	02:00	
Air Blank	0.000	02:01	
Air Blank	0.000	02:03	
Subject Sample	# 2	0.259 02:04	No .020 agreement
Breath Volume	1.914	02:04	
Air Blank	0.000	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 3	0.232 02:08	No .020 agreement
Breath Volume	1.453	02:08	
Air Blank	0.000	02:08	
Control Test	0.079	02:09	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 05/26/2021 20:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: HANKINS

DATE/TIME OF TEST: 04/17/2021 02:33:41

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.033	02:29
Breath Volume	2.683	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.035	02:32
Breath Volume	2.414	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 05/26/2021 20:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINDBLOM,ROBERT,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LINDBLOM,ROBERT
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 10:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: LOWERY

DATE/TIME OF TEST: 04/17/2021 11:43:21

Diagnostics Check	OK	11:37
Air Blank	0.000	11:37
Control Test	0.081	11:37
Air Blank	0.000	11:38
Subject Sample # 1	0.000	11:39
Breath Volume	4.132	11:39
Air Blank	0.000	11:39
Air Blank	0.000	11:41
Subject Sample # 2	0.000	11:42
Breath Volume	4.835	11:42
Air Blank	0.000	11:42
Control Test	0.081	11:42
Air Blank	0.000	11:43
Diagnostics Check	OK	11:43

Data Download Date/Time: 05/26/2021 20:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/17/2021 22:17:53

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.082	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.000	22:11
Breath Volume	1.613	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.000	22:16
Breath Volume	1.164	22:16
Air Blank	0.000	22:16
Control Test	0.082	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 05/26/2021 20:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ARGENT,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: CAMACHO FELICIANO

DATE/TIME OF TEST: 04/18/2021 01:37:12

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.175	01:29
Breath Volume	2.683	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:31
Subject Sample # 2	0.151	01:32
Breath Volume	1.468	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 3	0.173	01:35
Breath Volume	3.171	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 05/26/2021 20:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,CHASE,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COLLINS,CHASE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 04/18/2021 22:08:09

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:02	
Control Test	0.081	22:03	
Air Blank	0.000	22:03	
Subject Sample	# 1 0.000	22:04	
Breath Volume	3.964	22:04	
Air Blank	0.000	22:04	
Air Blank	0.000	22:07	
Subject Sample	# 2 0.000	22:07	
Breath Volume	2.855	22:07	
Air Blank		22:07	RFI Detect
Air Blank		22:08	RFI Detect

Data Download Date/Time: 05/26/2021 20:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFKJEN,THOMAS,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MADISE,GRAMBLING
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 12:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 04/24/2021 13:06:22

Diagnostics Check	OK	13:00
Air Blank	0.000	13:00
Control Test	0.082	13:01
Air Blank	0.000	13:01
Subject Sample # 1	0.000	13:02
Breath Volume	2.425	13:02
Air Blank	0.000	13:02
Air Blank	0.000	13:04
Subject Sample # 2	0.000	13:05
Breath Volume	2.484	13:05
Air Blank	0.000	13:05
Control Test	0.081	13:05
Air Blank	0.000	13:06
Diagnostics Check	OK	13:06

Data Download Date/Time: 05/26/2021 20:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,BRENDAN,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BROWN,BRENDAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: IGNACIO

DATE/TIME OF TEST: 04/25/2021 03:35:28

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.302	03:28
Breath Volume	1.187	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.317	03:33
Breath Volume	1.171	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 05/26/2021 20:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALON,DENNIS,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HEDDEN,JOSEPH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 11:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:36

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 04/26/2021 13:02:39

Diagnostics Check	OK	12:56
Air Blank	0.000	12:56
Control Test	0.081	12:57
Air Blank	0.000	12:57
Subject Sample # 1	0.231	12:58
Breath Volume	2.050	12:58
Air Blank	0.000	12:58
Air Blank	0.000	13:00
Subject Sample # 2	0.234	13:01
Breath Volume	2.015	13:01
Air Blank	0.000	13:01
Control Test	0.079	13:02
Air Blank	0.000	13:02
Diagnostics Check	OK	13:02

Data Download Date/Time: 05/26/2021 20:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERTHIAUME,WAYNE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BERTHIAUME,WAYNE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: FOUNTAIN

DATE/TIME OF TEST: 04/28/2021 02:17:45

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.286	02:12
Breath Volume	2.414	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.275	02:16
Breath Volume	1.511	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 05/26/2021 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLANDA, RICARDO, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HOLANDA, RICARDO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 04/18/2021 06:06:53

Diagnostics Check	OK	05:52	
Air Blank	0.000	05:52	
Control Test	0.080	05:52	
Air Blank	0.000	05:53	
Subject Sample	# 1 0.168	05:56	Volume Not Met
Breath Volume	0.031	05:56	
Air Blank	0.000	05:57	
Air Blank	0.000	05:58	
Subject Sample	# 2 0.180	05:59	
Breath Volume	1.132	05:59	
Air Blank	0.000	06:00	
Air Blank	0.000	06:02	
Subject Sample	# 3 0.182	06:05	Volume Not Met
Breath Volume	0.050	06:05	
Air Blank	0.000	06:05	
Control Test	0.079	06:06	
Air Blank	0.000	06:06	
Diagnostics Check	OK	06:06	

Data Download Date/Time: 05/10/2021 21:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLANDA, RICARDO, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HOLANDA, RICARDO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 04/18/2021 06:23:59

Diagnostics Check	OK	06:10	
Air Blank	0.000	06:10	
Control Test	0.079	06:10	
Air Blank	0.000	06:11	
Subject Sample	# 1 0.187	06:14	Volume Not Met
Breath Volume	0.984	06:14	
Air Blank	0.000	06:15	
Air Blank	0.000	06:17	
Subject Sample	# 2 0.183	06:18	
Breath Volume	1.460	06:18	
Air Blank	0.000	06:19	
Air Blank	0.000	06:21	
Subject Sample	# 3 0.188	06:22	
Breath Volume	1.199	06:22	
Air Blank	0.000	06:23	
Control Test	0.079	06:23	
Air Blank	0.000	06:23	
Diagnostics Check	OK	06:23	

Data Download Date/Time: 05/10/2021 21:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MCGUIRE,HOWARD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 04/21/2021 22:07:08

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.081	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.000	22:02
Breath Volume	3.222	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.000	22:05
Breath Volume	1.421	22:05
Air Blank	0.000	22:06
Control Test	0.082	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 05/10/2021 21:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNETT,JAMES,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BARNETT,JAMES
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: MOSBLECH

DATE/TIME OF TEST: 04/23/2021 21:08:52

Diagnostics Check	OK	21:02
Air Blank	0.000	21:02
Control Test	0.081	21:02
Air Blank	0.000	21:03
Subject Sample # 1	0.095	21:04
Breath Volume	4.976	21:04
Air Blank	0.000	21:04
Air Blank	0.000	21:06
Subject Sample # 2	0.092	21:07
Breath Volume	5.007	21:07
Air Blank	0.000	21:07
Control Test	0.081	21:08
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

Data Download Date/Time: 05/10/2021 21:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIRES, DAVID,
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: RAMIRES, DAVID
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: STIVERS

DATE/TIME OF TEST: 04/03/2021 05:53:22

Diagnostics Check	OK	05:46
Air Blank	0.000	05:47
Control Test	0.082	05:47
Air Blank	0.000	05:47
Subject Sample # 1	0.210	05:48
Breath Volume	1.160	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.206	05:51
Breath Volume	1.382	05:51
Air Blank	0.000	05:52
Control Test	0.079	05:52
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 04/14/2021 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARCUM,JOSEPH,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARCUM,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 15:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: PASTUCHA

DATE/TIME OF TEST: 04/15/2021 16:18:56

Diagnostics Check	OK	16:12
Air Blank	0.000	16:13
Control Test	0.084	16:13
Air Blank	0.000	16:13
Subject Sample # 1	0.000	16:14
Breath Volume	1.695	16:14
Air Blank	0.000	16:14
Air Blank	0.000	16:16
Subject Sample # 2	0.000	16:17
Breath Volume	1.464	16:17
Air Blank	0.000	16:18
Control Test	0.083	16:18
Air Blank	0.000	16:18
Diagnostics Check	OK	16:18

Data Download Date/Time: 05/11/2021 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMPSON,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAMPSON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 04/16/2021 21:45:25

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.083	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.144	21:40
Breath Volume	2.335	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.142	21:43
Breath Volume	2.312	21:43
Air Blank	0.000	21:44
Control Test	0.080	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 05/11/2021 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTHE,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: TAUER

DATE/TIME OF TEST: 04/18/2021 00:33:37

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:29	
Control Test	0.082	00:29	
Air Blank	0.000	00:29	
Subject Sample # 1	0.195	00:30	
Breath Volume	2.621	00:30	
Air Blank	0.000	00:31	
Air Blank		00:33	Ambient Fail
Air Blank	0.000	00:33	

Data Download Date/Time: 05/11/2021 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTHE,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: TAUER

DATE/TIME OF TEST: 04/18/2021 00:42:35

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.081	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.190	00:38
Breath Volume	3.066	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.191	00:41
Breath Volume	2.906	00:41
Air Blank	0.000	00:41
Control Test	0.081	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 05/11/2021 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMPSON,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAMPSON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: SKRZEC

DATE/TIME OF TEST: 04/18/2021 16:34:18

Diagnostics Check	OK	16:27
Air Blank	0.000	16:28
Control Test	0.083	16:28
Air Blank	0.000	16:29
Subject Sample # 1	0.000	16:29
Breath Volume	2.531	16:29
Air Blank	0.000	16:30
Air Blank	0.000	16:32
Subject Sample # 2	0.000	16:32
Breath Volume	2.476	16:32
Air Blank	0.000	16:33
Control Test	0.081	16:33
Air Blank	0.000	16:34
Diagnostics Check	OK	16:34

Data Download Date/Time: 05/11/2021 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: BIGGS

DATE/TIME OF TEST: 04/21/2021 00:54:38

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.083	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.150	00:50
Breath Volume	1.660	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.153	00:53
Breath Volume	1.765	00:53
Air Blank	0.000	00:53
Control Test	0.082	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 05/11/2021 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: LARAY

DATE/TIME OF TEST: 04/22/2021 05:53:30

Diagnostics Check	OK	05:50	
Air Blank	0.000	05:51	
Control Test	0.083	05:51	
Air Blank	0.000	05:51	
Subject Sample	# 1	05:52	Subject Test Refused
Breath Volume	0.000	05:52	
Air Blank	0.000	05:52	
Control Test	0.083	05:52	
Air Blank	0.000	05:53	
Diagnostics Check	OK	05:53	

Data Download Date/Time: 05/11/2021 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 04/25/2021 00:42:02

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.083	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.029	00:37
Breath Volume	1.703	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:39
Subject Sample # 2	0.027	00:40
Breath Volume	3.265	00:40
Air Blank	0.000	00:41
Control Test	0.082	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:42

Data Download Date/Time: 05/11/2021 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES, CONSTANCE, J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES, CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 04/27/2021 02:24:30

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.082	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.169	02:19
Breath Volume	2.011	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.164	02:23
Breath Volume	1.316	02:23
Air Blank	0.000	02:23
Control Test	0.082	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 05/11/2021 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER, LANDON, L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NASH, DAVID
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: PHIPPS

DATE/TIME OF TEST: 04/02/2021 22:35:43

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.082	22:29
Air Blank	0.000	22:29
Subject Sample # 1	0.142	22:30
Breath Volume	3.765	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.140	22:34
Breath Volume	4.074	22:34
Air Blank	0.000	22:34
Control Test	0.079	22:35
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 04/14/2021 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/04/2021 19:54:35

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:50	
Control Test	0.082	19:51	
Air Blank	0.000	19:51	
Subject Sample	# 1	19:53	RFI Detect
Breath Volume	0.000	19:53	
Air Blank	0.000	19:53	
Control Test	0.082	19:53	
Air Blank	0.000	19:54	
Diagnostics Check	OK	19:54	

Data Download Date/Time: 04/14/2021 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/04/2021 19:59:29

Diagnostics Check	OK	19:56	
Air Blank	0.000	19:57	
Control Test	0.082	19:57	
Air Blank	0.000	19:57	
Subject Sample	# 1	19:58	RFI Detect
Breath Volume	0.000	19:58	
Air Blank	0.000	19:58	
Control Test	0.082	19:58	
Air Blank	0.000	19:59	
Diagnostics Check	OK	19:59	

Data Download Date/Time: 04/14/2021 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASN,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/04/2021 20:02:40

Diagnostics Check	OK	20:00	
Air Blank	0.000	20:01	
Control Test	0.082	20:01	
Air Blank	0.000	20:02	
Subject Sample	# 1	20:02	RFI Detect
Breath Volume	0.000	20:02	
Air Blank		20:02	RFI Detect
Air Blank		20:02	RFI Detect

Data Download Date/Time: 04/14/2021 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/04/2021 20:11:09

Diagnostics Check	OK	20:04
Air Blank	0.000	20:04
Control Test	0.081	20:05
Air Blank	0.000	20:05
Subject Sample # 1	0.119	20:06
Breath Volume	2.000	20:06
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 2	0.118	20:09
Breath Volume	2.304	20:09
Air Blank	0.000	20:10
Control Test	0.079	20:10
Air Blank	0.000	20:11
Diagnostics Check	OK	20:11

Data Download Date/Time: 04/14/2021 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEENEY,ERIC,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FEENEY,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BROGDON

DATE/TIME OF TEST: 04/05/2021 00:28:43

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.082	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:27	
Control Test	0.082	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 04/14/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 12:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: HALLFORD

DATE/TIME OF TEST: 04/05/2021 13:32:33

Diagnostics Check	OK	13:26
Air Blank	0.000	13:26
Control Test	0.082	13:27
Air Blank	0.000	13:27
Subject Sample # 1	0.036	13:27
Breath Volume	1.746	13:27
Air Blank	0.000	13:28
Air Blank	0.000	13:30
Subject Sample # 2	0.036	13:31
Breath Volume	2.671	13:31
Air Blank	0.000	13:31
Control Test	0.081	13:31
Air Blank	0.000	13:32
Diagnostics Check	OK	13:32

Data Download Date/Time: 04/14/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/06/2021 20:33:37

Diagnostics Check	OK	20:26
Air Blank	0.000	20:27
Control Test	0.081	20:27
Air Blank	0.000	20:27
Subject Sample # 1	0.176	20:28
Breath Volume	1.433	20:28
Air Blank	0.000	20:29
Air Blank	0.000	20:30
Subject Sample # 2	0.181	20:31
Breath Volume	1.996	20:31
Air Blank	0.000	20:32
Control Test	0.080	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 04/14/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLANCK,ARIEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LIEB,JOHN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/08/2021 06:13:27

Diagnostics Check	OK	06:07
Air Blank	0.000	06:07
Control Test	0.080	06:07
Air Blank	0.000	06:08
Subject Sample # 1	0.148	06:08
Breath Volume	3.562	06:08
Air Blank	0.000	06:09
Air Blank	0.000	06:11
Subject Sample # 2	0.149	06:11
Breath Volume	3.660	06:11
Air Blank	0.000	06:12
Control Test	0.080	06:12
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

Data Download Date/Time: 04/14/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYBRAND,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LYBRAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: GUILLEN

DATE/TIME OF TEST: 04/08/2021 20:01:30

Diagnostics Check	OK	19:58	
Air Blank	0.000	19:58	
Control Test	0.081	19:59	
Air Blank	0.000	19:59	
Subject Sample	# 1	20:00	Subject Test Refused
Breath Volume	0.000	20:00	
Air Blank	0.000	20:00	
Control Test	0.081	20:00	
Air Blank	0.000	20:01	
Diagnostics Check	OK	20:01	

Data Download Date/Time: 04/14/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELTON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELTON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: GARCIA MARTINEZ

DATE/TIME OF TEST: 04/09/2021 22:34:04

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:30	
Control Test	0.081	22:31	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:32	Subject Test Refused
Breath Volume	0.000	22:32	
Air Blank	0.000	22:33	
Control Test	0.081	22:33	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:34	

Data Download Date/Time: 04/14/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: ENGLE

DATE/TIME OF TEST: 04/10/2021 02:16:18

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.147	02:11
Breath Volume	1.417	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.153	02:14
Breath Volume	1.238	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 04/14/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELTON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CORBIN,BRITTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 04/10/2021 23:05:21

Diagnostics Check	OK	22:58
Air Blank	0.000	22:59
Control Test	0.081	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.207	23:00
Breath Volume	1.464	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.204	23:03
Breath Volume	1.878	23:03
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 04/14/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: HOLLOMAN

DATE/TIME OF TEST: 04/10/2021 23:58:44

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.080	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.267	23:54
Breath Volume	2.281	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.249	23:57
Breath Volume	2.078	23:57
Air Blank	0.000	23:57
Control Test	0.078	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 04/14/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELTON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELTON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 04/11/2021 02:09:59

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.080	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.181	02:05
Breath Volume	3.921	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:07
Subject Sample # 2	0.175	02:08
Breath Volume	3.562	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 04/14/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM, TONY, J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LICHTY, JAKE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: PEAGLER

DATE/TIME OF TEST: 04/11/2021 06:32:14

Diagnostics Check	OK	06:25
Air Blank	0.000	06:26
Control Test	0.080	06:26
Air Blank	0.000	06:27
Subject Sample # 1	0.180	06:27
Breath Volume	2.535	06:27
Air Blank	0.000	06:28
Air Blank	0.000	06:30
Subject Sample # 2	0.171	06:30
Breath Volume	1.882	06:30
Air Blank	0.000	06:31
Control Test	0.077	06:31
Air Blank	0.000	06:32
Diagnostics Check	OK	06:32

Data Download Date/Time: 04/14/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER, LONDON, L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LAPEE, WYATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: CLENNAL

DATE/TIME OF TEST: 04/13/2021 20:40:46

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:38	
Control Test	0.081	20:38	
Air Blank	0.000	20:39	
Subject Sample	# 1	20:39	Subject Test Refused
Breath Volume	0.000	20:39	
Air Blank	0.000	20:39	
Control Test	0.080	20:40	
Air Blank	0.000	20:40	
Diagnostics Check	OK	20:40	

Data Download Date/Time: 04/14/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 04/15/2021 01:08:30

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.081	01:06	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 05/12/2021 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,KEITH,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,KEITH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: FELIPE LOPEZ

DATE/TIME OF TEST: 04/15/2021 01:56:13

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.081	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.125	01:51
Breath Volume	3.140	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:53
Subject Sample # 2	0.123	01:54
Breath Volume	2.292	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 05/12/2021 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRISTOL,JARRET,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BRISTOL,JARRET
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WILES

DATE/TIME OF TEST: 04/16/2021 23:10:06

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:07	
Control Test	0.082	23:07	
Air Blank	0.000	23:08	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:09	
Control Test	0.081	23:09	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:10	

Data Download Date/Time: 05/12/2021 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARRINGTON,IAN,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPENCE,JUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: PUROHIT

DATE/TIME OF TEST: 04/18/2021 02:55:17

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1 0.287	02:48	Volume Not Met
Breath Volume	0.105	02:48	
Air Blank	0.000	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 2	02:53	No Sample Provided
Breath Volume	0.000	02:53	
Air Blank	0.000	02:54	
Control Test	0.079	02:54	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 05/12/2021 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLEOD,JOEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCLEOD,JOEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 04/22/2021 03:39:09

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:36	
Control Test	0.082	03:37	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Control Test	0.082	03:38	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 05/12/2021 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,ROBERT,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LIEB,JOHN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 04/22/2021 22:52:09

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:43	
Control Test	0.082	22:43	
Air Blank	0.000	22:44	
Subject Sample	# 1 0.074	22:47	Volume Not Met
Breath Volume	0.101	22:47	
Air Blank	0.000	22:48	
Air Blank	0.000	22:49	
Subject Sample	# 2	22:50	Subject Test Refused
Breath Volume	0.125	22:50	
Air Blank	0.000	22:51	
Control Test	0.079	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 05/12/2021 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: GELLIS

DATE/TIME OF TEST: 04/24/2021 01:12:27

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:09	
Control Test	0.081	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:11	Subject Test Refused
Breath Volume	0.000	01:11	
Air Blank	0.000	01:11	
Control Test	0.080	01:11	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 05/12/2021 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ARMITAGE,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 04/24/2021 01:27:17

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:24	
Control Test	0.081	01:25	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:25	Subject Test Refused
Breath Volume	0.000	01:25	
Air Blank	0.000	01:26	
Control Test	0.080	01:26	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

Data Download Date/Time: 05/12/2021 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELTON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELTON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: KIMSEY

DATE/TIME OF TEST: 04/24/2021 02:02:38

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.080	01:59	
Air Blank	0.000	01:59	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.597	02:01	
Air Blank	0.000	02:01	
Control Test	0.081	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 05/12/2021 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/24/2021 02:47:37

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:45	
Control Test	0.081	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Control Test	0.080	02:47	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 05/12/2021 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARQUIS,NORMAND,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARQUIS,NORMAND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 09:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 04/24/2021 11:00:43

Diagnostics Check	OK	10:54
Air Blank	0.000	10:55
Control Test	0.081	10:55
Air Blank	0.000	10:56
Subject Sample # 1	0.000	10:56
Breath Volume	1.828	10:56
Air Blank	0.000	10:56
Air Blank	0.000	10:58
Subject Sample # 2	0.000	10:59
Breath Volume	1.203	10:59
Air Blank	0.000	10:59
Control Test	0.079	11:00
Air Blank	0.000	11:00
Diagnostics Check	OK	11:00

Data Download Date/Time: 05/12/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: BINDON

DATE/TIME OF TEST: 04/24/2021 21:29:42

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.080	21:23
Air Blank	0.000	21:23
Subject Sample # 1	0.135	21:24
Breath Volume	1.632	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.142	21:28
Breath Volume	1.714	21:28
Air Blank	0.000	21:28
Control Test	0.078	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 05/12/2021 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MLYNARSKI, NIKLAS, E
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MLYNARSKI, NIKLAS
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MATHEWS

DATE/TIME OF TEST: 04/25/2021 02:46:12

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.183	02:41
Breath Volume	2.625	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:43
Subject Sample # 2	0.173	02:44
Breath Volume	2.179	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 05/12/2021 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LICHTY,JAKE,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LICHTY,JAKE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: LEMONS

DATE/TIME OF TEST: 04/25/2021 06:14:42

Diagnostics Check	OK	06:11	
Air Blank	0.000	06:11	
Control Test	0.081	06:12	
Air Blank	0.000	06:12	
Subject Sample	# 1	06:13	Subject Test Refused
Breath Volume	0.000	06:13	
Air Blank	0.000	06:13	
Control Test	0.082	06:14	
Air Blank	0.000	06:14	
Diagnostics Check	OK	06:14	

Data Download Date/Time: 05/12/2021 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,SHELBY,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARROLL,SHELBY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 04/25/2021 22:15:10

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:11	
Control Test	0.080	22:11	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:13	Subject Test Refused
Breath Volume	0.000	22:13	
Air Blank	0.000	22:14	
Control Test	0.081	22:14	
Air Blank	0.000	22:15	
Diagnostics Check	OK	22:15	

Data Download Date/Time: 05/12/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,TODD,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MASON,NICHOLAS
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 04/28/2021 00:13:00

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.081	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.000	00:08
Breath Volume	4.089	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.000	00:11
Breath Volume	4.304	00:11
Air Blank	0.000	00:12
Control Test	0.079	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:13

Data Download Date/Time: 05/12/2021 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ARMITAGE,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 04/29/2021 04:28:07

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:25	
Control Test	0.081	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:27	
Control Test	0.080	04:27	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

Data Download Date/Time: 05/12/2021 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ECCOLS,BRANDON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 04/29/2021 23:11:30

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:08	
Control Test	0.080	23:09	
Air Blank	0.000	23:09	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:10	
Control Test	0.081	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 05/12/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 04/30/2021 01:47:06

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.081	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.182	01:42
Breath Volume	1.363	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.173	01:45
Breath Volume	1.835	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 05/12/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: HERLIHY

DATE/TIME OF TEST: 04/30/2021 16:42:06

Diagnostics Check	OK	16:35	
Air Blank	0.000	16:36	
Control Test	0.075	16:36	
Air Blank	0.000	16:37	
Subject Sample # 1	0.118	16:37	
Breath Volume	3.539	16:37	
Air Blank	0.000	16:38	
Air Blank	0.000	16:40	
Subject Sample # 2	0.116	16:40	
Breath Volume	3.625	16:40	
Air Blank	0.000	16:41	
Control Test	0.074	16:41	Control Outside Tolerance
Air Blank	0.000	16:42	

Data Download Date/Time: 05/12/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: HERLIHY

DATE/TIME OF TEST: 04/30/2021 16:50:03

Diagnostics Check	OK	16:43
Air Blank	0.000	16:44
Control Test	0.076	16:44
Air Blank	0.000	16:45
Subject Sample # 1	0.116	16:45
Breath Volume	2.289	16:45
Air Blank	0.000	16:46
Air Blank	0.000	16:47
Subject Sample # 2	0.110	16:48
Breath Volume	2.894	16:48
Air Blank	0.000	16:49
Control Test	0.075	16:49
Air Blank	0.000	16:49
Diagnostics Check	OK	16:50

Data Download Date/Time: 05/12/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,ALFRED,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HUGHES,ALFRED
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 04/10/2021 22:44:22

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.079	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.166	22:39
Breath Volume	1.187	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.170	22:42
Breath Volume	1.355	22:42
Air Blank	0.000	22:43
Control Test	0.075	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 05/04/2021 21:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:35

SUBJECT'S LAST NAME: OGLES

DATE/TIME OF TEST: 04/22/2021 12:59:35

Diagnostics Check	OK	12:56	
Air Blank	0.000	12:57	
Control Test	0.080	12:57	
Air Blank	0.000	12:58	
Subject Sample	# 1	12:58	Subject Test Refused
Breath Volume	0.000	12:58	
Air Blank	0.000	12:58	
Control Test	0.079	12:58	
Air Blank	0.000	12:59	
Diagnostics Check	OK	12:59	

Data Download Date/Time: 05/04/2021 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONTRAGER,WILLIAM,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BONTRAGER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: GERMAN

DATE/TIME OF TEST: 04/03/2021 05:44:05

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:40	
Control Test	0.080	05:41	
Air Blank	0.000	05:41	
Subject Sample	# 1 0.180	05:42	
Breath Volume	3.605	05:42	
Air Blank		05:42	RFI Detect
Air Blank	0.000	05:43	
Control Test	0.079	05:43	
Air Blank	0.000	05:44	
Diagnostics Check	OK	05:44	

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONTRAGER,WILLIAM,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BONTRAGER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: GERMAN

DATE/TIME OF TEST: 04/03/2021 05:52:05

Diagnostics Check	OK	05:45
Air Blank	0.000	05:46
Control Test	0.079	05:46
Air Blank	0.000	05:47
Subject Sample # 1	0.183	05:47
Breath Volume	3.343	05:47
Air Blank	0.000	05:48
Air Blank	0.000	05:49
Subject Sample # 2	0.184	05:50
Breath Volume	3.453	05:50
Air Blank	0.000	05:51
Control Test	0.078	05:51
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 04/03/2021 18:24:15

Diagnostics Check	OK	18:17
Air Blank	0.000	18:18
Control Test	0.081	18:18
Air Blank	0.000	18:19
Subject Sample # 1	0.246	18:19
Breath Volume	2.660	18:19
Air Blank	0.000	18:20
Air Blank	0.000	18:22
Subject Sample # 2	0.243	18:22
Breath Volume	2.945	18:22
Air Blank	0.000	18:23
Control Test	0.075	18:23
Air Blank	0.000	18:24
Diagnostics Check	OK	18:24

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONTRAGER,WILLIAM,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BONTRAGER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/03/2021 23:47:44

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.233	23:42
Breath Volume	3.437	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.231	23:46
Breath Volume	3.039	23:46
Air Blank	0.000	23:46
Control Test	0.077	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: WHITFIELD

DATE/TIME OF TEST: 04/04/2021 23:16:03

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.146	23:11
Breath Volume	2.175	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:13
Subject Sample # 2	0.141	23:14
Breath Volume	1.355	23:14
Air Blank	0.000	23:15
Control Test	0.075	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:16

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: MILLER,MARK
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 04/08/2021 03:52:44

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:48	
Subject Sample	# 1 0.231	03:51	Volume Not Met
Breath Volume	0.828	03:51	
Air Blank		03:52	Purge Fail
Air Blank	0.000	03:52	

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: MILLER,MARK
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 04/08/2021 03:55:37

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.073	03:54	Control Outside Tolerance
Air Blank	0.000	03:55	

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: MILLER,MARK
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 04/08/2021 04:08:57

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.078	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1 0.204	04:01	Volume Not Met
Breath Volume	0.933	04:01	
Air Blank	0.000	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 2 0.224	04:07	Volume Not Met
Breath Volume	0.582	04:07	
Air Blank	0.000	04:07	
Control Test	0.074	04:08	Control Outside Tolerance
Air Blank	0.000	04:08	

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 04/08/2021 23:19:02

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:16	
Control Test	0.079	23:16	
Air Blank	0.000	23:17	
Subject Sample	# 1	23:17	Subject Test Refused
Breath Volume	0.000	23:17	
Air Blank	0.000	23:18	
Control Test	0.079	23:18	
Air Blank	0.000	23:18	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: VASQUEZ JERONIMO

DATE/TIME OF TEST: 04/09/2021 01:41:55

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.079	01:39	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:40	Slope Not Met
Breath Volume	1.382	01:40	
Air Blank	0.000	01:40	
Control Test	0.078	01:41	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: VASQUEZ JERONIMO

DATE/TIME OF TEST: 04/09/2021 01:48:38

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.078	01:45	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:46	Subject Test Refused
Breath Volume	0.769	01:46	
Air Blank	0.000	01:47	
Control Test	0.078	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY, MATHEW, C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SORTO, ALEXANDRA
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 04/10/2021 19:39:15

Diagnostics Check	OK	19:36	
Air Blank	0.000	19:36	
Control Test	0.078	19:37	
Air Blank	0.000	19:37	
Subject Sample	# 1	19:37	Subject Test Refused
Breath Volume	0.000	19:37	
Air Blank	0.000	19:38	
Control Test	0.079	19:38	
Air Blank	0.000	19:39	
Diagnostics Check	OK	19:39	

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY, MATHEW, C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SORTO, ALEXANDRA
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 04/10/2021 19:47:44

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.079	19:42
Air Blank	0.000	19:42
Subject Sample # 1	0.165	19:43
Breath Volume	2.777	19:43
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.152	19:46
Breath Volume	2.578	19:46
Air Blank	0.000	19:46
Control Test	0.078	19:47
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: RIX,ARIANA
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: NICHOLAS

DATE/TIME OF TEST: 04/12/2021 17:30:01

Diagnostics Check	OK	17:23
Air Blank	0.000	17:23
Control Test	0.078	17:23
Air Blank	0.000	17:24
Subject Sample # 1	0.302	17:25
Breath Volume	1.515	17:25
Air Blank	0.000	17:26
Air Blank	0.000	17:27
Subject Sample # 2	0.292	17:28
Breath Volume	1.878	17:28
Air Blank	0.000	17:29
Control Test	0.077	17:29
Air Blank	0.000	17:29
Diagnostics Check	OK	17:30

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: ROSENDALE

DATE/TIME OF TEST: 04/13/2021 16:32:58

Diagnostics Check	OK	16:25
Air Blank	0.000	16:26
Control Test	0.079	16:26
Air Blank	0.000	16:26
Subject Sample # 1	0.199	16:27
Breath Volume	2.152	16:27
Air Blank	0.000	16:28
Air Blank	0.000	16:30
Subject Sample # 2	0.201	16:31
Breath Volume	1.160	16:31
Air Blank	0.000	16:32
Control Test	0.075	16:32
Air Blank	0.000	16:32
Diagnostics Check	OK	16:32

Data Download Date/Time: 05/13/2021 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DANDRA,M
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: STEPHENS,MAURICE
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: CHILDS

DATE/TIME OF TEST: 04/17/2021 01:46:23

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.112	01:41
Breath Volume	1.363	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.125	01:44
Breath Volume	2.199	01:44
Air Blank	0.000	01:45
Control Test	0.076	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 05/13/2021 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DENMARK

DATE/TIME OF TEST: 04/17/2021 02:57:28

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.158	02:52
Breath Volume	2.621	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.146	02:55
Breath Volume	2.398	02:55
Air Blank	0.000	02:56
Control Test	0.075	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 05/13/2021 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: LUCIANO

DATE/TIME OF TEST: 04/17/2021 22:09:58

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:08	
Control Test	0.078	22:08	
Air Blank	0.000	22:09	
Subject Sample	# 1	22:09	Sequence Aborted
Breath Volume	0.000	22:09	
Air Blank	0.000	22:09	

Data Download Date/Time: 05/13/2021 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LUCIANO

DATE/TIME OF TEST: 04/17/2021 22:23:11

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:20	
Control Test	0.079	22:21	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:21	Subject Test Refused
Breath Volume	0.000	22:21	
Air Blank	0.000	22:22	
Control Test	0.078	22:22	
Air Blank	0.000	22:23	
Diagnostics Check	OK	22:23	

Data Download Date/Time: 05/13/2021 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,STUART,P
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: KERNAN,MITCHELL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/18/2021 06:26:27

Diagnostics Check	OK	06:17
Air Blank	0.000	06:18
Control Test	0.078	06:18
Air Blank	0.000	06:19
Subject Sample # 1	0.182	06:21
Breath Volume	1.292	06:21
Air Blank	0.000	06:22
Air Blank	0.000	06:24
Subject Sample # 2	0.194	06:24
Breath Volume	1.613	06:24
Air Blank	0.000	06:25
Control Test	0.075	06:25
Air Blank	0.000	06:26
Diagnostics Check	OK	06:26

Data Download Date/Time: 05/13/2021 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/21/2021 22:26:24

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.080	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.000	22:22
Breath Volume	2.437	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.000	22:24
Breath Volume	2.269	22:24
Air Blank	0.000	22:25
Control Test	0.080	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 05/13/2021 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MADONIA

DATE/TIME OF TEST: 04/21/2021 23:13:39

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:11	
Control Test	0.080	23:11	
Air Blank	0.000	23:11	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:12	
Control Test	0.080	23:13	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:13	

Data Download Date/Time: 05/13/2021 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: MADONIA

DATE/TIME OF TEST: 04/21/2021 23:41:23

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.080	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.036	23:36
Breath Volume	1.472	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:38
Subject Sample # 2	0.036	23:39
Breath Volume	1.355	23:39
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 05/13/2021 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/30/2021 21:52:13

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.000	21:47
Breath Volume	3.894	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.000	21:50
Breath Volume	1.949	21:50
Air Blank	0.000	21:51
Control Test	0.078	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 05/13/2021 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGRETTE,JOAN,P
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: MOSER,KENDRICK
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:07

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/07/2021 13:35:02

Diagnostics Check	OK	13:28
Air Blank	0.000	13:29
Control Test	0.081	13:29
Air Blank	0.000	13:30
Subject Sample # 1	0.172	13:30
Breath Volume	1.207	13:30
Air Blank	0.000	13:31
Air Blank	0.000	13:33
Subject Sample # 2	0.162	13:33
Breath Volume	2.015	13:33
Air Blank	0.000	13:34
Control Test	0.080	13:34
Air Blank	0.000	13:34
Diagnostics Check	OK	13:35

Data Download Date/Time: 04/27/2021 17:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AKINS,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: NASH,JARRED
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 04/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 04/30/2021 01:41:49

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.087	01:37
Breath Volume	3.242	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.081	01:40
Breath Volume	2.378	01:40
Air Blank	0.000	01:40
Control Test	0.076	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 05/18/2021 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1076482
BREATH TEST OPERATOR: COX,DAVID,S	DRY GAS EXPIRY DATE: 01/15/2022
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: COX,DAVID	ARREST TIME: 14:57
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Female
LAST AGENCY INSPECTION: 04/09/2021	OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: MONTROY	DATE/TIME OF TEST: 04/24/2021	16:27:14
------------------------------	-------------------------------	----------

Diagnostics Check	OK	16:21
Air Blank	0.000	16:21
Control Test	0.078	16:21
Air Blank	0.000	16:22
Subject Sample # 1	0.000	16:23
Breath Volume	1.941	16:23
Air Blank	0.000	16:23
Air Blank	0.000	16:25
Subject Sample # 2	0.000	16:25
Breath Volume	1.769	16:25
Air Blank	0.000	16:26
Control Test	0.077	16:26
Air Blank	0.000	16:27
Diagnostics Check	OK	16:27

Data Download Date/Time: 05/10/2021 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNABB,KENNETH,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCNABB,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 04/06/2021 04:13:55

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.082	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.000	04:09
Breath Volume	2.406	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.000	04:12
Breath Volume	4.585	04:12
Air Blank	0.000	04:12
Control Test	0.082	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 04/09/2021 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 10:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 04/11/2021 00:40:58

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.115	00:35
Breath Volume	3.675	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.115	00:39
Breath Volume	3.832	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 05/06/2021 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 04/12/2021 19:57:29

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.082	19:49
Air Blank	0.000	19:50
Subject Sample # 1	0.189	19:51
Breath Volume	1.281	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample # 2	0.188	19:55
Breath Volume	3.093	19:55
Air Blank	0.000	19:56
Control Test	0.081	19:56
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time: 05/06/2021 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YASUDA,SHOHEI,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: YASUDA,SHOHEI
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: GRUNDKOWSKI

DATE/TIME OF TEST: 04/23/2021 17:24:53

Diagnostics Check	OK	17:18
Air Blank	0.000	17:18
Control Test	0.081	17:18
Air Blank	0.000	17:19
Subject Sample # 1	0.200	17:19
Breath Volume	3.503	17:19
Air Blank	0.000	17:20
Air Blank	0.000	17:22
Subject Sample # 2	0.198	17:23
Breath Volume	3.539	17:23
Air Blank	0.000	17:23
Control Test	0.080	17:24
Air Blank	0.000	17:24
Diagnostics Check	OK	17:24

Data Download Date/Time: 05/06/2021 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 05:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: SOULES

DATE/TIME OF TEST: 04/24/2021 06:47:21

Diagnostics Check	OK	06:38
Air Blank	0.000	06:39
Control Test	0.082	06:39
Air Blank	0.000	06:40
Subject Sample # 1	0.160	06:42
Breath Volume	4.449	06:42
Air Blank	0.000	06:43
Air Blank	0.000	06:44
Subject Sample # 2	0.155	06:45
Breath Volume	4.390	06:45
Air Blank	0.000	06:46
Control Test	0.078	06:46
Air Blank	0.000	06:47
Diagnostics Check	OK	06:47

Data Download Date/Time: 05/06/2021 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITAKER,JOHNNY,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WHITAKER,JOHNNY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/29/2021 18:57:13

Diagnostics Check	OK	18:54	
Air Blank	0.000	18:54	
Control Test	0.080	18:55	
Air Blank	0.000	18:55	
Subject Sample	# 1	18:55	Subject Test Refused
Breath Volume	0.000	18:55	
Air Blank	0.000	18:56	
Control Test	0.080	18:56	
Air Blank	0.000	18:57	
Diagnostics Check	OK	18:57	

Data Download Date/Time: 05/10/2021 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMS,KYLE,N
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SIMMS,KYLE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: WINTERS

DATE/TIME OF TEST: 04/28/2021 01:10:22

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.076	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.153	01:05
Breath Volume	2.730	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.141	01:08
Breath Volume	1.269	01:08
Air Blank	0.000	01:09
Control Test	0.077	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 05/04/2021 02:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMLIN,BRETT,W
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CONNOR,MICHAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 04/30/2021 00:32:25

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.175	00:27
Breath Volume	3.984	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.173	00:30
Breath Volume	3.519	00:30
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 05/20/2021 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: JONES,CORTLAND
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: KNOX

DATE/TIME OF TEST: 04/01/2021 18:29:27

Diagnostics Check	OK	18:17	
Air Blank	0.000	18:18	
Control Test	0.081	18:18	
Air Blank	0.000	18:19	
Subject Sample	# 1 0.071	18:22	Volume Not Met
Breath Volume	0.378	18:22	
Air Blank	0.000	18:22	
Air Blank	0.000	18:24	
Subject Sample	# 2 0.075	18:27	Volume Not Met
Breath Volume	0.792	18:27	
Air Blank	0.000	18:28	
Control Test	0.081	18:28	
Air Blank	0.000	18:29	
Diagnostics Check	OK	18:29	

Data Download Date/Time: 04/28/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINERNEY,JESSE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: TOVAR

DATE/TIME OF TEST: 04/03/2021 19:52:55

Diagnostics Check	OK	19:46
Air Blank	0.000	19:46
Control Test	0.081	19:47
Air Blank	0.000	19:47
Subject Sample # 1	0.205	19:48
Breath Volume	1.441	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:50
Subject Sample # 2	0.195	19:51
Breath Volume	1.484	19:51
Air Blank	0.000	19:52
Control Test	0.080	19:52
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 04/28/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CORTLAND,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: PHIFER,SKYLAR
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: DRYSDALE

DATE/TIME OF TEST: 04/10/2021 00:43:30

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.103	00:38
Breath Volume	4.515	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.103	00:41
Breath Volume	4.429	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 04/28/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: VEGA,BRENNAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 04/10/2021 19:47:45

Diagnostics Check	OK	19:40
Air Blank	0.000	19:40
Control Test	0.080	19:41
Air Blank	0.000	19:41
Subject Sample # 1	0.198	19:42
Breath Volume	4.621	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.197	19:46
Breath Volume	4.289	19:46
Air Blank	0.000	19:46
Control Test	0.077	19:47
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 04/28/2021 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINERNEY,JESSE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: TORRES MARTIN

DATE/TIME OF TEST: 04/11/2021 06:06:39

Diagnostics Check	OK	06:00
Air Blank	0.000	06:00
Control Test	0.080	06:01
Air Blank	0.000	06:01
Subject Sample # 1	0.147	06:02
Breath Volume	2.816	06:02
Air Blank	0.000	06:02
Air Blank	0.000	06:04
Subject Sample # 2	0.144	06:05
Breath Volume	2.140	06:05
Air Blank	0.000	06:05
Control Test	0.079	06:06
Air Blank	0.000	06:06
Diagnostics Check	OK	06:06

Data Download Date/Time: 04/28/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEFLANG,DANIEL,B
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MASTRANGIOLA,NATALIE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: MILEY

DATE/TIME OF TEST: 04/13/2021 05:44:28

Diagnostics Check	OK	05:38
Air Blank	0.000	05:38
Control Test	0.080	05:39
Air Blank	0.000	05:39
Subject Sample # 1	0.135	05:40
Breath Volume	3.535	05:40
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.131	05:42
Breath Volume	3.941	05:42
Air Blank	0.000	05:43
Control Test	0.078	05:43
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 04/28/2021 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LAVIANO,JOHN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MOSZKOW

DATE/TIME OF TEST: 04/15/2021 04:20:58

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.081	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.223	04:13
Breath Volume	3.644	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 2	0.201	04:16
Breath Volume	1.609	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 3	0.215	04:19
Breath Volume	2.878	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 04/28/2021 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CORTLAND,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: DERRICK

DATE/TIME OF TEST: 04/16/2021 15:59:09

Diagnostics Check	OK	15:52
Air Blank	0.000	15:53
Control Test	0.081	15:53
Air Blank	0.000	15:53
Subject Sample # 1	0.128	15:54
Breath Volume	2.351	15:54
Air Blank	0.000	15:55
Air Blank	0.000	15:57
Subject Sample # 2	0.129	15:57
Breath Volume	2.039	15:57
Air Blank	0.000	15:58
Control Test	0.080	15:58
Air Blank	0.000	15:59
Diagnostics Check	OK	15:59

Data Download Date/Time: 04/28/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILKEY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GILKEY,JOSHUA
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: HUSKIN

DATE/TIME OF TEST: 04/18/2021 02:04:31

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.210	02:00
Breath Volume	2.335	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.211	02:02
Breath Volume	2.296	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 04/28/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINERNEY,JESSE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 04/18/2021 05:36:06

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.080	05:30
Air Blank	0.000	05:30
Subject Sample # 1	0.149	05:31
Breath Volume	2.929	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.142	05:34
Breath Volume	2.730	05:34
Air Blank	0.000	05:35
Control Test	0.079	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 04/28/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUES,RUI,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:12

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/18/2021 10:39:41

Diagnostics Check	OK	10:32
Air Blank	0.000	10:33
Control Test	0.080	10:33
Air Blank	0.000	10:34
Subject Sample # 1	0.204	10:34
Breath Volume	1.351	10:34
Air Blank	0.000	10:35
Air Blank	0.000	10:37
Subject Sample # 2	0.200	10:37
Breath Volume	2.019	10:37
Air Blank	0.000	10:38
Control Test	0.079	10:39
Air Blank	0.000	10:39
Diagnostics Check	OK	10:39

Data Download Date/Time: 04/28/2021 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,JON,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: ARNOLD,JON
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: SANGBUSH

DATE/TIME OF TEST: 04/18/2021 20:49:17

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:46	
Control Test	0.080	20:47	
Air Blank	0.000	20:47	
Subject Sample	# 1	20:47	Subject Test Refused
Breath Volume	0.000	20:47	
Air Blank	0.000	20:48	
Control Test	0.079	20:48	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

Data Download Date/Time: 04/28/2021 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CORTLAND,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: JONES,CORTLAND
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: CLEVELAND

DATE/TIME OF TEST: 04/22/2021 23:13:45

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.184	23:09
Breath Volume	1.648	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.179	23:12
Breath Volume	1.628	23:12
Air Blank	0.000	23:12
Control Test	0.080	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 04/28/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINERNEY,JESSE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: FERRO

DATE/TIME OF TEST: 04/27/2021 03:40:45

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.081	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.187	03:36
Breath Volume	2.300	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:38
Subject Sample # 2	0.190	03:39
Breath Volume	2.035	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 04/28/2021 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:39

SUBJECT'S LAST NAME: LASTER

DATE/TIME OF TEST: 04/07/2021 11:06:21

Diagnostics Check	OK	10:59
Air Blank	0.000	11:00
Control Test	0.080	11:00
Air Blank	0.000	11:00
Subject Sample # 1	0.089	11:01
Breath Volume	2.707	11:01
Air Blank	0.000	11:02
Air Blank	0.000	11:04
Subject Sample # 2	0.086	11:04
Breath Volume	2.472	11:04
Air Blank	0.000	11:05
Control Test	0.080	11:05
Air Blank	0.000	11:06
Diagnostics Check	OK	11:06

Data Download Date/Time: 04/27/2021 19:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,PAUL,G
BREATH TEST OPERATOR AGENCY: Gulf County SO
ARRESTING OFFICER: GRIMES,DENISE
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 03/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: NUNNERY

DATE/TIME OF TEST: 04/08/2021 22:54:56

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.079	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:53	Subject Test Refused
Breath Volume	0.000	22:53	
Air Blank	0.000	22:54	
Control Test	0.079	22:54	
Air Blank	0.000	22:54	
Diagnostics Check	OK	22:54	

Data Download Date/Time: 04/27/2021 19:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DESIATO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:16

SUBJECT'S LAST NAME: DESTINO

DATE/TIME OF TEST: 04/06/2021 11:42:33

Diagnostics Check	OK	11:36
Air Blank	0.000	11:36
Control Test	0.081	11:37
Air Blank	0.000	11:37
Subject Sample # 1	0.000	11:38
Breath Volume	4.300	11:38
Air Blank	0.000	11:38
Air Blank	0.000	11:40
Subject Sample # 2	0.000	11:41
Breath Volume	1.886	11:41
Air Blank	0.000	11:41
Control Test	0.081	11:41
Air Blank	0.000	11:42
Diagnostics Check	OK	11:42

Data Download Date/Time: 04/06/2021 16:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: RIVERA SILVA

DATE/TIME OF TEST: 04/06/2021 21:44:15

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.081	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.234	21:39
Breath Volume	2.003	21:39
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.220	21:42
Breath Volume	1.441	21:42
Air Blank	0.000	21:43
Control Test	0.079	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 05/04/2021 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: DINES

DATE/TIME OF TEST: 04/08/2021 03:29:07

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.130	03:23
Breath Volume	1.777	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.130	03:27
Breath Volume	2.417	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 05/04/2021 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: LEIVA RODRIGUEZ

DATE/TIME OF TEST: 04/09/2021 01:12:05

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.080	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.135	01:07
Breath Volume	1.648	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:09
Subject Sample # 2	0.134	01:10
Breath Volume	1.605	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 05/04/2021 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/10/2021 01:05:48

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.082	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 1 0.170	00:55	
Breath Volume	1.621	00:55	
Air Blank	0.000	00:56	
Air Blank	0.000	00:57	
Subject Sample	# 2	01:01	No Sample Provided
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Air Blank	0.000	01:03	
Subject Sample	# 3 0.176	01:04	
Breath Volume	1.691	01:04	
Air Blank	0.000	01:04	
Control Test	0.080	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 05/04/2021 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: CHESTARO THOMAS

DATE/TIME OF TEST: 04/10/2021 01:54:12

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.081	01:45	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:48	No Sample Provided
Breath Volume	0.000	01:48	
Air Blank	0.000	01:49	
Air Blank	0.000	01:51	
Subject Sample	# 2	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Control Test	0.081	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 05/04/2021 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIZENDINE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:52

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 04/10/2021 16:22:02

Diagnostics Check	OK	16:12
Air Blank	0.000	16:12
Control Test	0.081	16:13
Air Blank	0.000	16:13
Subject Sample # 1	0.238	16:14
Breath Volume	1.871	16:14
Air Blank	0.000	16:14
Air Blank	0.000	16:16
Subject Sample # 2	0.213	16:17
Breath Volume	1.109	16:17
Air Blank	0.000	16:17
Air Blank	0.000	16:19
Subject Sample # 3	0.216	16:20
Breath Volume	1.734	16:20
Air Blank	0.000	16:21
Control Test	0.080	16:21
Air Blank	0.000	16:21
Diagnostics Check	OK	16:22

Data Download Date/Time: 05/04/2021 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/10/2021 21:35:34

Diagnostics Check	OK	21:24	
Air Blank	0.000	21:24	
Control Test	0.081	21:25	
Air Blank	0.000	21:25	
Subject Sample	# 1 0.000	21:28	Volume Not Met
Breath Volume	0.015	21:28	
Air Blank	0.000	21:29	
Air Blank	0.000	21:31	
Subject Sample	# 2	21:34	Subject Test Refused
Breath Volume	0.000	21:34	
Air Blank	0.000	21:34	
Control Test	0.079	21:34	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:35	

Data Download Date/Time: 05/04/2021 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 04/10/2021 23:53:31

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.081	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.197	23:48
Breath Volume	2.816	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.200	23:51
Breath Volume	2.882	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 05/04/2021 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ATKINS,A
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: FRASER

DATE/TIME OF TEST: 04/11/2021 02:37:42

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.169	02:32
Breath Volume	1.183	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.166	02:36
Breath Volume	1.480	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 05/04/2021 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SAMAC

DATE/TIME OF TEST: 04/11/2021 03:58:23

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.101	03:53
Breath Volume	1.421	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.101	03:56
Breath Volume	1.410	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 05/04/2021 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KROUSE,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: BACA QUINTANA

DATE/TIME OF TEST: 04/11/2021 05:55:32

Diagnostics Check	OK	05:47
Air Blank	0.000	05:47
Control Test	0.082	05:47
Air Blank	0.000	05:48
Subject Sample # 1	0.151	05:49
Breath Volume	1.429	05:49
Air Blank	0.000	05:50
Air Blank	0.000	05:52
Subject Sample # 2	0.156	05:53
Breath Volume	2.894	05:53
Air Blank	0.000	05:54
Control Test	0.081	05:54
Air Blank	0.000	05:55
Diagnostics Check	OK	05:55

Data Download Date/Time: 05/04/2021 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: WEINER

DATE/TIME OF TEST: 04/11/2021 23:27:38

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.081	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.247	23:22
Breath Volume	2.746	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.230	23:26
Breath Volume	1.589	23:26
Air Blank	0.000	23:26
Control Test	0.081	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 05/04/2021 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: STRICKLAND JR

DATE/TIME OF TEST: 04/12/2021 01:57:10

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.134	01:52
Breath Volume	1.808	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.134	01:55
Breath Volume	1.519	01:55
Air Blank	0.000	01:56
Control Test	0.082	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 05/04/2021 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: SAVERAID

DATE/TIME OF TEST: 04/12/2021 04:51:24

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.082	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.126	04:46
Breath Volume	2.707	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.123	04:49
Breath Volume	3.121	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 05/04/2021 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 04/13/2021 02:03:48

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.081	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.208	01:59
Breath Volume	2.691	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:01
Subject Sample # 2	0.208	02:02
Breath Volume	2.621	02:02
Air Blank	0.000	02:02
Control Test	0.081	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 05/04/2021 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FUCH,D 72804
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:08

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 04/13/2021 13:38:16

Diagnostics Check	OK	13:28
Air Blank	0.000	13:29
Control Test	0.081	13:29
Air Blank	0.000	13:30
Subject Sample # 1	0.254	13:32
Breath Volume	1.511	13:32
Air Blank	0.000	13:33
Air Blank	0.000	13:34
Subject Sample # 2	0.262	13:36
Breath Volume	1.113	13:36
Air Blank	0.000	13:37
Control Test	0.080	13:37
Air Blank	0.000	13:38
Diagnostics Check	OK	13:38

Data Download Date/Time: 05/04/2021 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: DESIR

DATE/TIME OF TEST: 04/14/2021 03:33:10

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.082	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.000	03:28
Breath Volume	2.226	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.000	03:31
Breath Volume	2.238	03:31
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 05/04/2021 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 04/14/2021 23:49:16

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.306	23:44
Breath Volume	1.964	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.311	23:47
Breath Volume	1.777	23:47
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 05/04/2021 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HANSEN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 04/15/2021 10:40:09

Diagnostics Check	OK	10:24	
Air Blank	0.000	10:25	
Control Test	0.081	10:25	
Air Blank	0.000	10:26	
Subject Sample	# 1 0.159	10:29	Volume Not Met
Breath Volume	0.035	10:29	
Air Blank	0.000	10:29	
Air Blank	0.000	10:31	
Subject Sample	# 2 0.186	10:33	
Breath Volume	1.609	10:33	
Air Blank	0.000	10:34	
Air Blank	0.000	10:35	
Subject Sample	# 3 0.171	10:38	
Breath Volume	1.187	10:38	
Air Blank	0.000	10:39	
Control Test	0.079	10:39	
Air Blank	0.000	10:40	
Diagnostics Check	OK	10:40	

Data Download Date/Time: 05/04/2021 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: ARRUBLA

DATE/TIME OF TEST: 04/16/2021 03:21:20

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.081	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.000	03:17
Breath Volume	2.957	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.000	03:19
Breath Volume	2.625	03:19
Air Blank	0.000	03:20
Control Test	0.081	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 05/04/2021 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/16/2021 23:06:55

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.082	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.107	23:02
Breath Volume	3.718	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:04
Subject Sample # 2	0.099	23:05
Breath Volume	3.199	23:05
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 05/04/2021 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 04/17/2021 00:31:51

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.081	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.145	00:26
Breath Volume	1.371	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.154	00:30
Breath Volume	2.035	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 05/04/2021 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: VARGAS GONZALEZ

DATE/TIME OF TEST: 04/17/2021 00:45:49

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.243	00:40
Breath Volume	1.664	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.239	00:44
Breath Volume	2.019	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 05/04/2021 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JOHNSON,B
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: BUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: FETTIG

DATE/TIME OF TEST: 04/18/2021 00:23:09

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.082	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.080	00:18
Breath Volume	3.359	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.081	00:21
Breath Volume	3.078	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 05/04/2021 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 04/18/2021 02:47:51

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.081	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.184	02:43
Breath Volume	1.585	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.189	02:46
Breath Volume	1.882	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 05/04/2021 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/18/2021 02:56:50

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.082	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.081	02:56	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 05/04/2021 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/18/2021 04:55:24

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.119	04:49
Breath Volume	1.226	04:49
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.122	04:53
Breath Volume	1.699	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 05/04/2021 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: FISK

DATE/TIME OF TEST: 04/18/2021 06:12:26

Diagnostics Check	OK	06:05
Air Blank	0.000	06:06
Control Test	0.082	06:06
Air Blank	0.000	06:07
Subject Sample # 1	0.225	06:07
Breath Volume	2.535	06:07
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample # 2	0.227	06:10
Breath Volume	3.277	06:10
Air Blank	0.000	06:11
Control Test	0.080	06:11
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

Data Download Date/Time: 05/04/2021 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 04/19/2021 01:38:35

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.082	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.095	01:33
Breath Volume	1.113	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.095	01:37
Breath Volume	1.417	01:37
Air Blank	0.000	01:37
Control Test	0.081	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 05/04/2021 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: BURES

DATE/TIME OF TEST: 04/19/2021 04:29:17

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.081	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.218	04:24
Breath Volume	3.042	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.220	04:27
Breath Volume	3.093	04:27
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 05/04/2021 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: YOUNG JR

DATE/TIME OF TEST: 04/20/2021 02:32:04

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.082	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.141	02:27
Breath Volume	2.328	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:29
Subject Sample # 2	0.137	02:30
Breath Volume	1.144	02:30
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:32

Data Download Date/Time: 05/04/2021 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINCLAIR,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: TRENKER

DATE/TIME OF TEST: 04/21/2021 18:45:05

Diagnostics Check	OK	18:38
Air Blank	0.000	18:39
Control Test	0.082	18:39
Air Blank	0.000	18:39
Subject Sample # 1	0.000	18:40
Breath Volume	2.468	18:40
Air Blank	0.000	18:41
Air Blank	0.000	18:43
Subject Sample # 2	0.000	18:43
Breath Volume	2.375	18:43
Air Blank	0.000	18:44
Control Test	0.081	18:44
Air Blank	0.000	18:45
Diagnostics Check	OK	18:45

Data Download Date/Time: 05/04/2021 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: SALLOT

DATE/TIME OF TEST: 04/21/2021 22:39:06

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.082	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.190	22:34
Breath Volume	2.773	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.196	22:37
Breath Volume	2.812	22:37
Air Blank	0.000	22:38
Control Test	0.081	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 05/04/2021 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEWART,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: SENDLOSKY

DATE/TIME OF TEST: 04/22/2021 16:26:27

Diagnostics Check	OK	16:20	
Air Blank	0.000	16:21	
Control Test	0.082	16:21	
Air Blank	0.000	16:22	
Subject Sample	# 1	16:25	Subject Test Refused
Breath Volume	0.000	16:25	
Air Blank	0.000	16:25	
Control Test	0.081	16:25	
Air Blank	0.000	16:26	
Diagnostics Check	OK	16:26	

Data Download Date/Time: 05/04/2021 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/22/2021 23:43:30

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.000	23:39
Breath Volume	3.125	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.000	23:42
Breath Volume	2.773	23:42
Air Blank	0.000	23:42
Control Test	0.082	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 05/04/2021 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 04/23/2021 04:05:10

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.082	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.247	04:00
Breath Volume	2.589	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.245	04:03
Breath Volume	2.996	04:03
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 05/04/2021 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: BOREK

DATE/TIME OF TEST: 04/24/2021 02:36:46

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.081	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.008	02:31
Breath Volume	1.675	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.007	02:35
Breath Volume	1.253	02:35
Air Blank	0.000	02:35
Control Test	0.081	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 05/04/2021 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: MONTOYA-ANGULO

DATE/TIME OF TEST: 04/24/2021 03:05:11

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.090	03:00
Breath Volume	1.527	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.096	03:03
Breath Volume	1.382	03:03
Air Blank	0.000	03:04
Control Test	0.081	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 05/04/2021 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CZAPLEWAKI

DATE/TIME OF TEST: 04/24/2021 04:12:48

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.082	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.122	04:08
Breath Volume	2.832	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.123	04:11
Breath Volume	2.843	04:11
Air Blank	0.000	04:11
Control Test	0.081	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 05/04/2021 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NORMAN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:53

SUBJECT'S LAST NAME: PICHARDO

DATE/TIME OF TEST: 04/24/2021 11:20:25

Diagnostics Check	OK	11:14
Air Blank	0.000	11:14
Control Test	0.081	11:14
Air Blank	0.000	11:15
Subject Sample # 1	0.249	11:15
Breath Volume	1.640	11:15
Air Blank	0.000	11:16
Air Blank	0.000	11:18
Subject Sample # 2	0.245	11:18
Breath Volume	2.035	11:18
Air Blank	0.000	11:19
Control Test	0.080	11:19
Air Blank	0.000	11:20
Diagnostics Check	OK	11:20

Data Download Date/Time: 05/04/2021 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 04/24/2021 23:12:10

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.082	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.134	23:07
Breath Volume	3.773	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.133	23:10
Breath Volume	3.695	23:10
Air Blank	0.000	23:11
Control Test	0.081	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 05/04/2021 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 04/25/2021 01:23:44

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.081	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.159	01:18
Breath Volume	1.597	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.166	01:22
Breath Volume	1.949	01:22
Air Blank	0.000	01:22
Control Test	0.081	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 05/04/2021 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: PERDOMO

DATE/TIME OF TEST: 04/25/2021 04:04:01

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.190	03:59
Breath Volume	2.816	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.180	04:02
Breath Volume	1.867	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:04

Data Download Date/Time: 05/04/2021 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 04/26/2021 00:42:30

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.082	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.191	00:36
Breath Volume	1.484	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.189	00:40
Breath Volume	1.769	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 05/04/2021 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/26/2021 02:14:44

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.081	02:02	
Air Blank	0.000	02:02	
Subject Sample	# 1 0.128	02:05	Volume Not Met
Breath Volume	0.570	02:05	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample	# 2 0.137	02:09	
Breath Volume	1.824	02:09	
Air Blank	0.000	02:10	
Air Blank	0.000	02:12	
Subject Sample	# 3 0.133	02:13	
Breath Volume	1.585	02:13	
Air Blank	0.000	02:13	
Control Test	0.079	02:14	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 05/04/2021 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 04/27/2021 01:37:00

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.081	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.083	01:32
Breath Volume	1.390	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.086	01:35
Breath Volume	1.500	01:35
Air Blank	0.000	01:36
Control Test	0.082	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:37

Data Download Date/Time: 05/04/2021 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RUIZ,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/27/2021 20:48:11

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.082	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.094	20:42
Breath Volume	1.453	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.091	20:46
Breath Volume	1.695	20:46
Air Blank	0.000	20:47
Control Test	0.081	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 05/04/2021 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: BARROSO

DATE/TIME OF TEST: 04/27/2021 22:13:28

Diagnostics Check	OK	22:06
Air Blank	0.000	22:07
Control Test	0.081	22:07
Air Blank	0.000	22:08
Subject Sample # 1	0.253	22:08
Breath Volume	1.800	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.262	22:11
Breath Volume	2.226	22:11
Air Blank	0.000	22:12
Control Test	0.080	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 05/04/2021 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/29/2021 03:09:27

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.082	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.171	03:04
Breath Volume	1.359	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.178	03:07
Breath Volume	1.171	03:07
Air Blank	0.000	03:08
Control Test	0.081	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 05/04/2021 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: KHAN

DATE/TIME OF TEST: 04/30/2021 04:16:33

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.082	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.129	04:11
Breath Volume	1.289	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.129	04:15
Breath Volume	1.230	04:15
Air Blank	0.000	04:15
Control Test	0.081	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 05/04/2021 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: MCGRATH

DATE/TIME OF TEST: 04/30/2021 23:30:20

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.082	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.000	23:25
Breath Volume	2.335	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.000	23:28
Breath Volume	1.742	23:28
Air Blank	0.000	23:29
Control Test	0.081	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 05/04/2021 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANUS,GLEN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GANUS,GLEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: HUMERICKHOUSE

DATE/TIME OF TEST: 04/11/2021 22:49:46

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.079	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.094	22:45
Breath Volume	1.519	22:45
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.105	22:48
Breath Volume	1.402	22:48
Air Blank	0.000	22:48
Control Test	0.077	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 04/16/2021 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,RICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: MCDOWELL

DATE/TIME OF TEST: 04/16/2021 12:16:51

Diagnostics Check	OK	12:10
Air Blank	0.000	12:11
Control Test	0.079	12:11
Air Blank	0.000	12:12
Subject Sample # 1	0.000	12:12
Breath Volume	2.703	12:12
Air Blank	0.000	12:13
Air Blank	0.000	12:15
Subject Sample # 2	0.000	12:15
Breath Volume	2.554	12:15
Air Blank	0.000	12:15
Control Test	0.079	12:16
Air Blank	0.000	12:16
Diagnostics Check	OK	12:16

Data Download Date/Time: 05/21/2021 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/07/2021 18:19:58

Diagnostics Check	OK	18:13
Air Blank	0.000	18:14
Control Test	0.079	18:14
Air Blank	0.000	18:15
Subject Sample # 1	0.000	18:15
Breath Volume	2.292	18:15
Air Blank	0.000	18:16
Air Blank	0.000	18:18
Subject Sample # 2	0.000	18:18
Breath Volume	2.281	18:18
Air Blank	0.000	18:19
Control Test	0.080	18:19
Air Blank	0.000	18:19
Diagnostics Check	OK	18:19

Data Download Date/Time: 04/19/2021 18:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELM,DEREK,W
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HELM,DEREK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: PARTRIDGE

DATE/TIME OF TEST: 04/10/2021 00:04:17

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.078	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.237	23:58
Breath Volume	2.152	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.233	00:02
Breath Volume	1.343	00:02
Air Blank	0.000	00:03
Control Test	0.076	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 04/19/2021 18:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,PATRICK,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RIZZO,H
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RICKARD

DATE/TIME OF TEST: 04/10/2021 23:34:45

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.077	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.077	23:34	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 04/19/2021 18:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/11/2021 02:56:19

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.078	02:54	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:55	
Control Test	0.078	02:55	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 04/19/2021 18:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FIGUEROA,JESUS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 04/11/2021 19:31:09

Diagnostics Check	OK	19:24
Air Blank	0.000	19:25
Control Test	0.078	19:25
Air Blank	0.000	19:26
Subject Sample # 1	0.000	19:26
Breath Volume	1.281	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.000	19:29
Breath Volume	2.300	19:29
Air Blank	0.000	19:30
Control Test	0.078	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 04/19/2021 18:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 04/13/2021 21:46:48

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:44	
Control Test	0.078	21:44	
Air Blank	0.000	21:44	
Subject Sample	# 1	21:45	Subject Test Refused
Breath Volume	0.000	21:45	
Air Blank	0.000	21:45	
Control Test	0.075	21:46	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

Data Download Date/Time: 04/19/2021 18:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WILSON,KYLE
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CLINE

DATE/TIME OF TEST: 04/14/2021 04:25:23

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:20	
Control Test	0.078	04:21	
Air Blank	0.000	04:21	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:24	
Control Test	0.078	04:24	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 04/19/2021 18:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLIN,RICHARD,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TEMPLIN,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: MAGGIO

DATE/TIME OF TEST: 04/17/2021 06:37:12

Diagnostics Check	OK	06:30
Air Blank	0.000	06:31
Control Test	0.078	06:31
Air Blank	0.000	06:31
Subject Sample # 1	0.264	06:32
Breath Volume	2.476	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:34
Subject Sample # 2	0.270	06:35
Breath Volume	2.183	06:35
Air Blank	0.000	06:36
Control Test	0.076	06:36
Air Blank	0.000	06:37
Diagnostics Check	OK	06:37

Data Download Date/Time: 04/19/2021 18:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNNE,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BARNLEY,K
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: HRANYCZNY

DATE/TIME OF TEST: 04/17/2021 14:12:32

Diagnostics Check	OK	14:10	
Air Blank	0.000	14:11	
Control Test	0.074	14:11	Control Outside Tolerance
Air Blank	0.000	14:12	

Data Download Date/Time: 04/19/2021 18:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNNE,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BARNLEY,K
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: HRANYCZNY

DATE/TIME OF TEST: 04/17/2021 14:16:53

Diagnostics Check	OK	14:15	
Air Blank	0.000	14:15	
Control Test	0.225	14:15	Control Outside Tolerance
Air Blank	0.000	14:16	

Data Download Date/Time: 04/19/2021 18:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNNE,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BARNLEY,K
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: HRANYCZNY

DATE/TIME OF TEST: 04/17/2021 14:20:25

Diagnostics Check	OK	14:18	
Air Blank	0.000	14:19	
Control Test	0.218	14:19	Control Outside Tolerance
Air Blank	0.000	14:20	

Data Download Date/Time: 04/19/2021 18:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNNE,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BARNLEY,K
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: HRANYCZNY

DATE/TIME OF TEST: 04/17/2021 14:24:01

Diagnostics Check	OK	14:22	
Air Blank	0.000	14:22	
Control Test	0.212	14:23	Control Outside Tolerance
Air Blank	0.000	14:24	

Data Download Date/Time: 04/19/2021 18:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNNE,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BARNLEY,K
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: HRANYCZNY

DATE/TIME OF TEST: 04/17/2021 14:28:37

Diagnostics Check	OK	14:26	
Air Blank	0.000	14:27	
Control Test	0.193	14:27	Control Outside Tolerance
Air Blank	0.000	14:28	

Data Download Date/Time: 04/19/2021 18:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: SCHWARTZ

DATE/TIME OF TEST: 04/18/2021 16:01:09

Diagnostics Check	OK	15:54
Air Blank	0.000	15:55
Control Test	0.078	15:55
Air Blank	0.000	15:56
Subject Sample # 1	0.305	15:56
Breath Volume	1.941	15:56
Air Blank	0.000	15:57
Air Blank	0.000	15:59
Subject Sample # 2	0.299	15:59
Breath Volume	1.589	15:59
Air Blank	0.000	16:00
Control Test	0.076	16:00
Air Blank	0.000	16:01
Diagnostics Check	OK	16:01

Data Download Date/Time: 04/19/2021 18:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: KILDAY

DATE/TIME OF TEST: 04/21/2021 22:42:58

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.079	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.222	22:37
Breath Volume	2.710	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.220	22:41
Breath Volume	2.890	22:41
Air Blank	0.000	22:42
Control Test	0.079	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 05/12/2021 18:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,C
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:32

SUBJECT'S LAST NAME: AMBROSINO

DATE/TIME OF TEST: 04/24/2021 12:00:59

Diagnostics Check	OK	11:52
Air Blank	0.000	11:53
Control Test	0.080	11:53
Air Blank	0.000	11:54
Subject Sample # 1	0.000	11:56
Breath Volume	1.882	11:56
Air Blank	0.000	11:56
Air Blank	0.000	11:58
Subject Sample # 2	0.000	11:59
Breath Volume	1.261	11:59
Air Blank	0.000	12:00
Control Test	0.079	12:00
Air Blank	0.000	12:00
Diagnostics Check	OK	12:00

Data Download Date/Time: 05/12/2021 18:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/24/2021 21:18:23

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:15	
Control Test	0.080	21:16	
Air Blank	0.000	21:16	
Subject Sample	# 1	21:17	Subject Test Refused
Breath Volume	0.000	21:17	
Air Blank	0.000	21:17	
Control Test	0.080	21:17	
Air Blank	0.000	21:18	
Diagnostics Check	OK	21:18	

Data Download Date/Time: 05/12/2021 18:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: STROWDER

DATE/TIME OF TEST: 04/25/2021 21:47:26

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.180	21:42
Breath Volume	1.250	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.188	21:45
Breath Volume	1.746	21:45
Air Blank	0.000	21:46
Control Test	0.080	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 05/12/2021 18:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBEY,BRADLEY,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBEY,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/26/2021 13:14:04

Diagnostics Check	OK	13:10	
Air Blank	0.000	13:10	
Control Test	0.079	13:11	
Air Blank	0.000	13:11	
Subject Sample	# 1	13:12	Subject Test Refused
Breath Volume	0.000	13:12	
Air Blank	0.000	13:13	
Control Test	0.079	13:13	
Air Blank	0.000	13:13	
Diagnostics Check	OK	13:14	

Data Download Date/Time: 05/12/2021 18:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 04/27/2021 02:08:34

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.121	02:04
Breath Volume	2.347	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.124	02:06
Breath Volume	1.824	02:06
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 05/12/2021 18:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MUSTO,ANTHONY
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MENNELLA

DATE/TIME OF TEST: 04/30/2021 21:46:59

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.079	21:41
Air Blank	0.000	21:41
Subject Sample # 1	0.000	21:42
Breath Volume	1.953	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.000	21:45
Breath Volume	1.617	21:45
Air Blank	0.000	21:46
Control Test	0.080	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 05/12/2021 18:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN, GERALD, E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN, GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: WILHELM

DATE/TIME OF TEST: 04/24/2021 02:57:58

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.078	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.131	02:53
Breath Volume	1.828	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.125	02:56
Breath Volume	1.195	02:56
Air Blank	0.000	02:57
Control Test	0.078	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 04/26/2021 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN, GERALD, E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RODRIQUEZ, R
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GIBSON-SIROLI

DATE/TIME OF TEST: 04/25/2021 03:02:10

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.078	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.172	02:57
Breath Volume	1.730	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.158	03:00
Breath Volume	1.300	03:00
Air Blank	0.000	03:01
Control Test	0.077	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 04/26/2021 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOVER, DAVID, W
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: PARISH, C
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: SAULS

DATE/TIME OF TEST: 04/26/2021 19:24:22

Diagnostics Check	OK	19:23	
Air Blank	0.000	19:23	
		19:23	RFI Detect
Air Blank	0.000	19:24	

Data Download Date/Time: 05/24/2021 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOVER,DAVID,W
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: PARISH,C
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: SAULS

DATE/TIME OF TEST: 04/26/2021 19:32:20

Diagnostics Check	OK	19:26
Air Blank	0.000	19:26
Control Test	0.080	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.120	19:27
Breath Volume	2.000	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.112	19:30
Breath Volume	2.148	19:30
Air Blank	0.000	19:31
Control Test	0.079	19:31
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 05/24/2021 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/22/2021 23:40:46

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.146	23:35
Breath Volume	1.562	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.143	23:39
Breath Volume	1.492	23:39
Air Blank	0.000	23:39
Control Test	0.077	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 04/26/2021 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: RODRIGUEZ,RONALD
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: SORIANO

DATE/TIME OF TEST: 04/23/2021 21:54:23

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.095	21:49
Breath Volume	1.441	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:51
Subject Sample # 2	0.096	21:52
Breath Volume	1.523	21:52
Air Blank	0.000	21:53
Control Test	0.077	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 04/26/2021 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 04/25/2021 02:28:54

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.078	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.174	02:23
Breath Volume	2.117	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.163	02:27
Breath Volume	1.320	02:27
Air Blank	0.000	02:27
Control Test	0.078	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 04/26/2021 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALSH,SHANNON,L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: WALSH,S

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 15:03

GENDER: Female

OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: SPRAGUE

DATE/TIME OF TEST: 04/21/2021 17:40:27

Diagnostics Check	OK	17:37	
Air Blank	0.000	17:37	
Control Test	0.080	17:38	
Air Blank	0.000	17:38	
Subject Sample	# 1	17:39	Subject Test Refused
Breath Volume	0.000	17:39	
Air Blank	0.000	17:39	
Control Test	0.079	17:39	
Air Blank	0.000	17:40	
Diagnostics Check	OK	17:40	

Data Download Date/Time: 04/28/2021 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALSH,SHANNON,L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: WALSH,S

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 23:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/24/2021 00:15:25

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.081	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.138	00:10
Breath Volume	1.601	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.146	00:13
Breath Volume	2.078	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 04/28/2021 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALSH,SHANNON,L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: WALSH,SHANNON

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 17:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: CHOLOQUINGA ALOMOTO

DATE/TIME OF TEST: 04/24/2021 17:56:22

Diagnostics Check	OK	17:50
Air Blank	0.000	17:50
Control Test	0.081	17:50
Air Blank	0.000	17:51
Subject Sample # 1	0.268	17:51
Breath Volume	1.992	17:51
Air Blank	0.000	17:52
Air Blank	0.000	17:54
Subject Sample # 2	0.263	17:54
Breath Volume	2.066	17:54
Air Blank	0.000	17:55
Control Test	0.080	17:55
Air Blank	0.000	17:56
Diagnostics Check	OK	17:56

Data Download Date/Time: 04/28/2021 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAGUADOG,SHAUN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CLAYTON,STEVEN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: VICKERY

DATE/TIME OF TEST: 04/06/2021 21:25:06

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.080	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.197	21:20
Breath Volume	2.750	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.185	21:23
Breath Volume	1.957	21:23
Air Blank	0.000	21:24
Control Test	0.080	21:24
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 05/05/2021 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTCHINSON,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HUTCHINSON,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: MCNICHOLS

DATE/TIME OF TEST: 04/09/2021 00:31:52

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.013	00:27
Breath Volume	1.515	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.012	00:30
Breath Volume	2.269	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 05/05/2021 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN, WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BATTLES

DATE/TIME OF TEST: 04/09/2021 03:08:39

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.000	03:03
Breath Volume	2.074	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.000	03:07
Breath Volume	1.492	03:07
Air Blank	0.000	03:07
Control Test	0.081	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 05/05/2021 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTCHINSON,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HUTCHINSON,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 04/09/2021 04:16:29

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.143	04:10
Breath Volume	1.750	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.134	04:15
Breath Volume	1.960	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 05/05/2021 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER, WILLIAM, B
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: TURNER, WILLIAM
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 04/11/2021 06:31:03

Diagnostics Check	OK	06:23
Air Blank	0.000	06:23
Control Test	0.081	06:23
Air Blank	0.000	06:24
Subject Sample # 1	0.170	06:26
Breath Volume	3.964	06:26
Air Blank	0.000	06:27
Air Blank	0.000	06:29
Subject Sample # 2	0.162	06:29
Breath Volume	3.234	06:29
Air Blank	0.000	06:30
Control Test	0.077	06:30
Air Blank	0.000	06:30
Diagnostics Check	OK	06:31

Data Download Date/Time: 05/05/2021 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTCHINSON,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HUTCHINSON,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/13/2021 23:42:11

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.000	23:37
Breath Volume	2.359	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.000	23:40
Breath Volume	2.390	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 05/05/2021 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTCHINSON,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HUTCHINSON,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: SYKES

DATE/TIME OF TEST: 04/14/2021 02:25:56

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.080	02:23	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Control Test	0.080	02:25	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 05/05/2021 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAGUADOG,SHAUN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MAGUADOG,SHAUN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/17/2021 04:47:51

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.166	04:43
Breath Volume	2.722	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.166	04:46
Breath Volume	3.484	04:46
Air Blank	0.000	04:46
Control Test	0.080	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 05/05/2021 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: DOUVILLE

DATE/TIME OF TEST: 04/20/2021 17:26:39

Diagnostics Check	OK	17:19
Air Blank	0.000	17:20
Control Test	0.080	17:20
Air Blank	0.000	17:20
Subject Sample # 1	0.258	17:22
Breath Volume	1.421	17:22
Air Blank	0.000	17:22
Air Blank	0.000	17:24
Subject Sample # 2	0.256	17:25
Breath Volume	1.851	17:25
Air Blank	0.000	17:25
Control Test	0.081	17:26
Air Blank	0.000	17:26
Diagnostics Check	OK	17:26

Data Download Date/Time: 05/05/2021 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAGUADOG,SHAUN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: WALKER,MARINA
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 04/25/2021 02:42:17

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.000	02:37
Breath Volume	1.503	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.000	02:40
Breath Volume	2.152	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 05/05/2021 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAGUADOG,SHAUN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SMOLKIN,DMITRY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: FINCH

DATE/TIME OF TEST: 04/27/2021 01:01:16

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.081	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.034	00:56
Breath Volume	1.550	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.033	00:59
Breath Volume	1.921	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 05/05/2021 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 08:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: ALMEIDA

DATE/TIME OF TEST: 04/28/2021 09:22:24

Diagnostics Check	OK	09:16
Air Blank	0.000	09:16
Control Test	0.081	09:17
Air Blank	0.000	09:17
Subject Sample # 1	0.090	09:17
Breath Volume	3.812	09:17
Air Blank	0.000	09:18
Air Blank	0.000	09:20
Subject Sample # 2	0.091	09:20
Breath Volume	4.175	09:20
Air Blank	0.000	09:21
Control Test	0.080	09:21
Air Blank	0.000	09:22
Diagnostics Check	OK	09:22

Data Download Date/Time: 05/05/2021 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUGA,JOHN,E
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: KINGRY,HOBIE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/30/2021 19:44:35

Diagnostics Check	OK	19:35
Air Blank	0.000	19:35
Control Test	0.080	19:35
Air Blank	0.000	19:36
Subject Sample # 1	0.000	19:38
Breath Volume	1.675	19:38
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.000	19:43
Breath Volume	1.292	19:43
Air Blank	0.000	19:43
Control Test	0.080	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 05/05/2021 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: CASH

DATE/TIME OF TEST: 04/01/2021 04:04:47

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.075	04:00
Breath Volume	1.859	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.069	04:03
Breath Volume	1.878	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 04/05/2021 19:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: SALZWEDEL

DATE/TIME OF TEST: 04/01/2021 23:24:11

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.216	23:19
Breath Volume	1.570	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.211	23:22
Breath Volume	1.859	23:22
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 04/05/2021 19:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,CHEYANNE,K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ROSS,CHEYANNE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 04/03/2021 20:12:42

Diagnostics Check	OK	20:06
Air Blank	0.000	20:06
Control Test	0.080	20:07
Air Blank	0.000	20:07
Subject Sample # 1	0.000	20:08
Breath Volume	4.285	20:08
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.000	20:11
Breath Volume	4.328	20:11
Air Blank	0.000	20:11
Control Test	0.079	20:12
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 04/05/2021 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,DAVIS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RIOS,DAVIS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BALLETTTO

DATE/TIME OF TEST: 04/06/2021 00:57:42

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.127	00:53
Breath Volume	3.449	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.123	00:56
Breath Volume	3.085	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATO, MARK, P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCARLATO, FRANK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/06/2021 19:23:37

Diagnostics Check	OK	19:16
Air Blank	0.000	19:17
Control Test	0.081	19:17
Air Blank	0.000	19:18
Subject Sample # 1	0.000	19:19
Breath Volume	2.222	19:19
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.000	19:22
Breath Volume	1.902	19:22
Air Blank	0.000	19:22
Control Test	0.080	19:23
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 04/10/2021 02:13:40

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.032	02:09
Breath Volume	2.121	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.031	02:12
Breath Volume	1.687	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNOVA,CASSANDRA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CANNOVA,CASSANDRA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: PLAU

DATE/TIME OF TEST: 04/11/2021 00:46:53

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.000	00:42
Breath Volume	2.136	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.000	00:45
Breath Volume	1.894	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: SCHULTE

DATE/TIME OF TEST: 04/11/2021 21:40:33

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.080	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.098	21:35
Breath Volume	1.476	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.101	21:39
Breath Volume	1.992	21:39
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/12/2021 03:17:28

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:12	
Control Test	0.080	03:12	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:16	Subject Test Refused
Breath Volume	0.054	03:16	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: DUNNING

DATE/TIME OF TEST: 04/12/2021 21:08:01

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:01	
Control Test	0.080	21:01	
Air Blank	0.000	21:02	
Subject Sample # 1	0.000	21:02	
Breath Volume	3.386	21:02	
Air Blank	0.000	21:03	
Air Blank	0.000	21:05	
Subject Sample # 2	0.000	21:06	
Breath Volume	1.664	21:06	
Air Blank	0.000	21:06	
Control Test	0.081	21:07	
Air Blank		21:07	RFI Detect
Air Blank	0.000	21:08	

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: DUNNING

DATE/TIME OF TEST: 04/12/2021 21:17:38

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.080	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.000	21:13
Breath Volume	3.335	21:13
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.000	21:16
Breath Volume	1.937	21:16
Air Blank	0.000	21:16
Control Test	0.080	21:17
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCHWERTHOFFER, TARYN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: ORDIWAY

DATE/TIME OF TEST: 04/13/2021 20:08:31

Diagnostics Check	OK	20:02
Air Blank	0.000	20:02
Control Test	0.080	20:02
Air Blank	0.000	20:03
Subject Sample # 1	0.290	20:03
Breath Volume	2.535	20:03
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 2	0.294	20:07
Breath Volume	2.996	20:07
Air Blank	0.000	20:07
Control Test	0.079	20:07
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: AWADALLAH

DATE/TIME OF TEST: 04/15/2021 02:30:29

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.080	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.000	02:26
Breath Volume	2.167	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.000	02:29
Breath Volume	2.574	02:29
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, ALFONSO,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS, DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: CARAMANICA

DATE/TIME OF TEST: 04/16/2021 01:14:24

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.099	01:10
Breath Volume	3.550	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.097	01:12
Breath Volume	1.753	01:12
Air Blank	0.000	01:13
Control Test	0.079	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGIEL,ADRIAN,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CHESNUT,JONATHON
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 04/18/2021 05:50:21

Diagnostics Check	OK	05:47	
Air Blank	0.000	05:48	
Control Test	0.080	05:48	
Air Blank	0.000	05:48	
Subject Sample	# 1	05:49	Subject Test Refused
Breath Volume	0.000	05:49	
Air Blank	0.000	05:49	
Control Test	0.080	05:49	
Air Blank	0.000	05:50	
Diagnostics Check	OK	05:50	

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANDLEY,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON,WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: FATOLITIS

DATE/TIME OF TEST: 04/18/2021 18:19:04

Diagnostics Check	OK	18:11
Air Blank	0.000	18:12
Control Test	0.080	18:12
Air Blank	0.000	18:13
Subject Sample # 1	0.000	18:14
Breath Volume	1.976	18:14
Air Blank	0.000	18:14
Air Blank	0.000	18:16
Subject Sample # 2	0.000	18:17
Breath Volume	2.332	18:17
Air Blank	0.000	18:18
Control Test	0.081	18:18
Air Blank	0.000	18:18
Diagnostics Check	OK	18:19

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 04/20/2021 02:21:36

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.000	02:17
Breath Volume	1.464	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.000	02:20
Breath Volume	1.265	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,ALEXANDER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 04/20/2021 05:20:24

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.081	05:15
Air Blank	0.000	05:15
Subject Sample # 1	0.000	05:16
Breath Volume	3.269	05:16
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.000	05:19
Breath Volume	2.238	05:19
Air Blank	0.000	05:19
Control Test	0.080	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: POLINO

DATE/TIME OF TEST: 04/20/2021 23:49:33

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.070	23:45
Breath Volume	2.550	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.070	23:48
Breath Volume	2.019	23:48
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: FLAKE

DATE/TIME OF TEST: 04/24/2021 04:17:10

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.110	04:12
Breath Volume	3.945	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:14
Subject Sample # 2	0.109	04:15
Breath Volume	4.203	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: HERRICK

DATE/TIME OF TEST: 04/24/2021 19:51:52

Diagnostics Check	OK	19:45
Air Blank	0.000	19:45
Control Test	0.080	19:45
Air Blank	0.000	19:46
Subject Sample # 1	0.081	19:47
Breath Volume	1.285	19:47
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.085	19:50
Breath Volume	1.378	19:50
Air Blank	0.000	19:50
Control Test	0.080	19:51
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: BARLE

DATE/TIME OF TEST: 04/25/2021 03:30:05

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.080	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.243	03:25
Breath Volume	2.796	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:27
Subject Sample # 2	0.228	03:28
Breath Volume	2.410	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:30

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATO,MARK,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ANTONIDES,KARA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: LARRAIN

DATE/TIME OF TEST: 04/25/2021 14:27:29

Diagnostics Check	OK	14:21
Air Blank	0.000	14:21
Control Test	0.080	14:21
Air Blank	0.000	14:22
Subject Sample # 1	0.103	14:23
Breath Volume	2.082	14:23
Air Blank	0.000	14:23
Air Blank	0.000	14:25
Subject Sample # 2	0.101	14:26
Breath Volume	2.191	14:26
Air Blank	0.000	14:26
Control Test	0.080	14:26
Air Blank	0.000	14:27
Diagnostics Check	OK	14:27

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: MESQUITA VENTURA

DATE/TIME OF TEST: 04/26/2021 01:49:24

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.107	01:44
Breath Volume	3.273	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.107	01:47
Breath Volume	2.167	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANDLEY,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON,WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: VANCE

DATE/TIME OF TEST: 04/27/2021 09:16:49

Diagnostics Check	OK	09:08
Air Blank	0.000	09:08
Control Test	0.080	09:08
Air Blank	0.000	09:09
Subject Sample # 1	0.000	09:12
Breath Volume	1.363	09:12
Air Blank	0.000	09:12
Air Blank	0.000	09:14
Subject Sample # 2	0.000	09:15
Breath Volume	1.136	09:15
Air Blank	0.000	09:15
Control Test	0.079	09:16
Air Blank	0.000	09:16
Diagnostics Check	OK	09:16

Data Download Date/Time: 05/16/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 04/27/2021 19:16:22

Diagnostics Check	OK	19:09
Air Blank	0.000	19:10
Control Test	0.080	19:10
Air Blank	0.000	19:11
Subject Sample # 1	0.000	19:12
Breath Volume	3.132	19:12
Air Blank	0.000	19:12
Air Blank	0.000	19:14
Subject Sample # 2	0.000	19:15
Breath Volume	2.187	19:15
Air Blank	0.000	19:15
Control Test	0.080	19:15
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

Data Download Date/Time: 05/16/2021 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/28/2021 02:39:59

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.112	02:35
Breath Volume	2.011	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample # 2	0.117	02:38
Breath Volume	2.445	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 05/16/2021 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: MAARLEVELD

DATE/TIME OF TEST: 04/28/2021 22:03:26

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.081	21:58
Air Blank	0.000	21:58
Subject Sample # 1	0.087	21:58
Breath Volume	3.945	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.088	22:01
Breath Volume	3.011	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 05/16/2021 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FRITZ,TYLER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 04/29/2021 00:27:33

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.091	00:23
Breath Volume	1.132	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.095	00:26
Breath Volume	2.250	00:26
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 05/16/2021 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CARROLL,JILLIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: LIZARDI-FERNANDEZ

DATE/TIME OF TEST: 04/29/2021 01:09:19

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.000	01:05
Breath Volume	1.890	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.000	01:07
Breath Volume	2.746	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 05/16/2021 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,DAVIS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RIOS,DAVIS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HOLMAN

DATE/TIME OF TEST: 04/29/2021 02:04:01

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.181	01:59
Breath Volume	1.230	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.186	02:02
Breath Volume	2.292	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:04

Data Download Date/Time: 05/16/2021 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: DEL RIO

DATE/TIME OF TEST: 04/29/2021 02:13:34

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.093	02:07
Breath Volume	1.691	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.083	02:12
Breath Volume	1.296	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 05/16/2021 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCARLATO,FRANK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/29/2021 17:35:27

Diagnostics Check	OK	17:28
Air Blank	0.000	17:29
Control Test	0.079	17:29
Air Blank	0.000	17:29
Subject Sample # 1	0.052	17:30
Breath Volume	1.718	17:30
Air Blank	0.000	17:31
Air Blank	0.000	17:33
Subject Sample # 2	0.050	17:34
Breath Volume	1.804	17:34
Air Blank	0.000	17:34
Control Test	0.079	17:34
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

Data Download Date/Time: 05/16/2021 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: CALABRESE

DATE/TIME OF TEST: 04/29/2021 22:19:04

Diagnostics Check	OK	22:12
Air Blank	0.000	22:13
Control Test	0.079	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.097	22:14
Breath Volume	2.539	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.099	22:17
Breath Volume	2.476	22:17
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:19

Data Download Date/Time: 05/16/2021 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,CHEYANNE,K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ROSS,CHEYANNE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: GOODRICH

DATE/TIME OF TEST: 04/30/2021 21:41:24

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.119	21:36
Breath Volume	4.046	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.119	21:39
Breath Volume	4.171	21:39
Air Blank	0.000	21:40
Control Test	0.080	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 05/16/2021 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SEVERINOVSKY

DATE/TIME OF TEST: 04/11/2021 01:46:40

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.225	01:42
Breath Volume	2.824	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.223	01:45
Breath Volume	1.921	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 04/14/2021 16:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: ROBLERO DIAZ

DATE/TIME OF TEST: 04/14/2021 21:50:14

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.080	21:41
Air Blank	0.000	21:41
Subject Sample # 1	0.233	21:42
Breath Volume	1.296	21:42
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.260	21:45
Breath Volume	2.425	21:45
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 3	0.263	21:48
Breath Volume	2.449	21:48
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 05/31/2021 06:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: RUUD

DATE/TIME OF TEST: 04/29/2021 18:12:31

Diagnostics Check	OK	18:06	
Air Blank	0.000	18:06	
Control Test	0.079	18:07	
Air Blank	0.000	18:07	
Subject Sample	# 1 0.188	18:08	
Breath Volume	3.171	18:08	
Air Blank	0.000	18:08	
Air Blank	0.000	18:10	
Subject Sample	# 2 0.192	18:10	
Breath Volume	3.253	18:10	
Air Blank	0.000	18:11	
Control Test	0.079	18:11	
Air Blank		18:12	RFI Detect
Air Blank	0.000	18:12	

Data Download Date/Time: 05/31/2021 06:47