Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MCCONNELL,AUSTIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 11/09/2021 19:25:14

Diagnostics Check		OK	19:18
Air Blank		0.000	19:18
Control Test		0.078	19:18
Air Blank		0.000	19:19
Subject Sample	# 1	0.000	19:20
Breath Volume		1.703	19:20
Air Blank		0.000	19:20
Air Blank		0.000	19:22
Subject Sample	# 2	0.000	19:23
Breath Volume		2.089	19:23
Air Blank		0.000	19:24
Control Test		0.079	19:24
Air Blank		0.000	19:25
Diagnostics Check		OK	19:25

Data Download Date/Time: 12/02/2021 11:02

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAVIN,DEMERIS,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAMOS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: BROWER

DATE/TIME OF TEST: 11/11/2021 22:38:07

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:35	
Control Test		0.078	22:36	
Air Blank		0.000	22:36	
Subject Sample	# 1		22:36	Subject Test Refused
Breath Volume		0.000	22:36	
Air Blank		0.000	22:37	
Control Test		0.078	22:37	
Air Blank		0.000	22:38	
Diagnostics Check		OK	22:38	

Data Download Date/Time: 12/02/2021 11:02

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WYATT,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: CODY,DEONTA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 11/14/2021 22:18:05

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.079	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.191	22:13
Breath Volume		2.687	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:15
Subject Sample	# 2	0.186	22:16
Breath Volume		1.410	22:16
Air Blank		0.000	22:17
Control Test		0.079	22:17
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 12/02/2021 11:02

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALE,NICHOLAS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ALLEN,G
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/19/2021 06:52:06

Diagnostics Check		OK	06:47	
Air Blank		0.000	06:48	
Control Test		0.077	06:48	
Air Blank		0.000	06:48	
Subject Sample	# 1		06:50	Subject Test Refused
Breath Volume		0.000	06:50	
Air Blank		0.000	06:51	
Control Test		0.078	06:51	
Air Blank		0.000	06:52	
Diagnostics Check		OK	06:52	

Data Download Date/Time: 12/02/2021 11:02

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HIGHTOWER,TUCKER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: PARSONS

DATE/TIME OF TEST: 11/21/2021 00:41:15

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:33
Subject Sample	# 1	0.139	00:36
Breath Volume		1.226	00:36
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.142	00:39
Breath Volume		1.414	00:39
Air Blank		0.000	00:40
Control Test		0.076	00:40
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 12/02/2021 11:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LARGE,MICHAEL
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 11/21/2021 02:33:16

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.078	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.000	02:28
Breath Volume		1.550	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.000	02:31
Breath Volume		1.726	02:31
Air Blank		0.000	02:32
Control Test		0.079	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 12/02/2021 11:03

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALE,NICHOLAS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: STILLION,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: MAITLAND

DATE/TIME OF TEST: 11/21/2021 17:19:36

Diagnostics Check		OK	17:06	
Air Blank		0.000	17:07	
Control Test		0.078	17:07	
Air Blank		0.000	17:07	
Subject Sample	# 1	0.207	17:11	Volume Not Met
Breath Volume		0.105	17:11	
Air Blank		0.000	17:11	
Air Blank		0.000	17:13	
Subject Sample	# 2	0.236	17:14	
Breath Volume		1.679	17:14	
Air Blank		0.000	17:14	
Air Blank		0.000	17:16	
Subject Sample	# 3	0.228	17:18	
Breath Volume		1.710	17:18	
Air Blank		0.000	17:18	
Control Test		0.077	17:19	
Air Blank		0.000	17:19	
Diagnostics Check		OK	17:19	

Data Download Date/Time: 12/02/2021 11:03

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MEJIAS,JOSEPH
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 11/26/2021 22:59:07

Diagnostics Check		OK	22:56	
Air Blank		0.000	22:56	
Control Test		0.078	22:57	
Air Blank		0.000	22:57	
Subject Sample	# 1		22:57	Subject Test Refused
Breath Volume		0.000	22:57	
Air Blank		0.000	22:58	
Control Test		0.078	22:58	
Air Blank		0.000	22:59	
Diagnostics Check		OK	22:59	

Data Download Date/Time: 12/02/2021 11:03

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MEJIAS,JOSEPH
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: SHANE

DATE/TIME OF TEST: 11/26/2021 23:05:03

Diagnostics Check		OK	23:00	
Air Blank		0.000	23:01	
Control Test		0.079	23:01	
Air Blank		0.000	23:02	
Subject Sample	# 1		23:03	Subject Test Refused
Breath Volume		0.000	23:03	
Air Blank		0.000	23:04	
Control Test		0.078	23:04	
Air Blank		0.000	23:04	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 12/02/2021 11:03

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ROSARIO,FAUSTA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: OBANNER

DATE/TIME OF TEST: 11/27/2021 07:36:15

Diagnostics Check		OK	07:33	
Air Blank		0.000	07:33	
Control Test		0.079	07:34	
Air Blank		0.000	07:34	
Subject Sample	# 1		07:34	Subject Test Refused
Breath Volume		0.000	07:34	
Air Blank		0.000	07:35	
Control Test		0.078	07:35	
Air Blank		0.000	07:36	
Diagnostics Check		OK	07:36	

Data Download Date/Time: 12/02/2021 11:03

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP METZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GUINANE,JOHN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 11/28/2021 19:21:16

Diagnostics Check		OK	19:18	
Air Blank		0.000	19:18	
Control Test		0.078	19:19	
Air Blank		0.000	19:19	
Subject Sample	# 1		19:19	Subject Test Refused
Breath Volume		0.000	19:19	
Air Blank		0.000	19:20	
Control Test		0.078	19:20	
Air Blank		0.000	19:21	
Diagnostics Check		OK	19:21	

Data Download Date/Time: 12/02/2021 11:03

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATFIELD,KASON,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WARE,GLENN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: HATFIELD

DATE/TIME OF TEST: 11/29/2021 20:08:18

Diagnostics Check		OK	20:05	
Air Blank		0.000	20:06	
Control Test		0.078	20:06	
Air Blank		0.000	20:06	
Subject Sample	# 1		20:06	Subject Test Refused
Breath Volume		0.000	20:06	
Air Blank		0.000	20:07	
Control Test		0.078	20:07	
Air Blank		0.000	20:08	
Diagnostics Check		OK	20:08	

Data Download Date/Time: 12/02/2021 11:03

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAVIN,DEMERIS,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WELLMAN,ANDREW
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 11/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: PIERSON

DATE/TIME OF TEST: 11/30/2021 04:50:57

Diagnostics Check		OK	04:44
Air Blank		0.000	04:44
Control Test		0.078	04:45
Air Blank		0.000	04:45
Subject Sample	# 1	0.148	04:46
Breath Volume		3.820	04:46
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.143	04:49
Breath Volume		3.066	04:49
Air Blank		0.000	04:50
Control Test		0.078	04:50
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 12/02/2021 11:03

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALE,NICHOLAS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WARE,G
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/30/2021 16:13:29

Diagnostics Check		OK	16:07
Air Blank		0.000	16:07
Control Test		0.080	16:08
Air Blank		0.000	16:08
Subject Sample	# 1	0.000	16:09
Breath Volume		2.789	16:09
Air Blank		0.000	16:09
Air Blank		0.000	16:11
Subject Sample	# 2	0.000	16:12
Breath Volume		2.539	16:12
Air Blank		0.000	16:12
Control Test		0.080	16:12
Air Blank		0.000	16:13
Diagnostics Check		OK	16:13

Data Download Date/Time: 12/02/2021 11:03

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HOUTEN,JONATHAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BISHOP,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: SUMNER

DATE/TIME OF TEST: 11/06/2021 22:30:48

Diagnostics Check		OK	22:26	
Air Blank		0.000	22:26	
Control Test		0.078	22:27	
Air Blank		0.000	22:27	
Subject Sample	# 1		22:29	Subject Test Refused
Breath Volume		0.000	22:29	
Air Blank		0.000	22:29	
Control Test		0.078	22:30	
Air Blank		0.000	22:30	
Diagnostics Check		OK	22:30	

Data Download Date/Time: 11/12/2021 09:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BISHOP,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: LESTER

DATE/TIME OF TEST: 11/09/2021 20:17:29

Diagnostics Check		OK	20:11
Air Blank		0.000	20:11
Control Test		0.078	20:11
Air Blank		0.000	20:12
Subject Sample	# 1	0.217	20:13
Breath Volume		2.753	20:13
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.209	20:16
Breath Volume		1.941	20:16
Air Blank		0.000	20:16
Control Test		0.079	20:16
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

Data Download Date/Time: 11/12/2021 09:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: CAUSHI

DATE/TIME OF TEST: 11/10/2021 00:03:33

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.078	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.000	23:59
Breath Volume		3.636	23:59
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.000	00:02
Breath Volume		3.656	00:02
Air Blank		0.000	00:02
Control Test		0.078	00:02
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 11/12/2021 09:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: POOLE-BITTLER

DATE/TIME OF TEST: 11/10/2021 22:25:54

Diagnostics Check		OK	22:14	
Air Blank		0.000	22:15	
Control Test		0.078	22:15	
Air Blank		0.000	22:15	
Subject Sample	# 1		22:18	No Sample Provided
Breath Volume		0.000	22:18	
Air Blank		0.000	22:19	
Air Blank		0.000	22:21	
Subject Sample	# 2		22:24	No Sample Provided
Breath Volume		0.000	22:24	
Air Blank		0.000	22:25	
Control Test		0.078	22:25	
Air Blank		0.000	22:25	
Diagnostics Check		OK	22:25	

Data Download Date/Time: 11/12/2021 09:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 11/10/2021 23:44:26

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.079	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.000	23:40
Breath Volume		4.402	23:40
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.000	23:43
Breath Volume		4.296	23:43
Air Blank		0.000	23:43
Control Test		0.078	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 11/12/2021 09:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BISHOP,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GYIRASZIN III

DATE/TIME OF TEST: 11/11/2021 01:13:51

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.078	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.175	01:08
Breath Volume		2.203	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.164	01:12
Breath Volume		2.222	01:12
Air Blank		0.000	01:12
Control Test		0.077	01:13
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 11/12/2021 09:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: POWELL,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:03

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/04/2021 09:44:27

Diagnostics Check		OK	09:38
Air Blank		0.000	09:38
Control Test		0.081	09:39
Air Blank		0.000	09:39
Subject Sample	# 1	0.053	09:40
Breath Volume		3.253	09:40
Air Blank		0.000	09:40
Air Blank		0.000	09:42
Subject Sample	# 2	0.053	09:42
Breath Volume		1.585	09:42
Air Blank		0.000	09:43
Control Test		0.082	09:43
Air Blank		0.000	09:44
Diagnostics Check		OK	09:44

Data Download Date/Time: 12/22/2021 11:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HART,JOSEPH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: GOODMAN

DATE/TIME OF TEST: 11/04/2021 23:36:23

Diagnostics Check		OK	23:29
Air Blank		0.000	23:30
Control Test		0.081	23:30
Air Blank		0.000	23:31
Subject Sample	# 1	0.204	23:31
Breath Volume		3.246	23:31
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.201	23:34
Breath Volume		3.210	23:34
Air Blank		0.000	23:35
Control Test		0.081	23:35
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 12/22/2021 11:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MLYNARSKI,NIKLAS,E
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MLYNARSKI,NIKLAS
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: SOURJOHN

DATE/TIME OF TEST: 11/07/2021 03:43:22

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:40	
Control Test		0.082	03:40	
Air Blank		0.000	03:41	
Subject Sample	# 1		03:41	Subject Test Refused
Breath Volume		0.000	03:41	
Air Blank		0.000	03:42	
Control Test		0.081	03:42	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 12/22/2021 11:12

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIAGET,MISTY,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PIAGET,MISTY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: QUINTANA SANCHEZ

DATE/TIME OF TEST: 11/13/2021 19:15:56

Diagnostics Check		OK	19:09
Air Blank		0.000	19:10
Control Test		0.082	19:10
Air Blank		0.000	19:11
Subject Sample	# 1	0.174	19:11
Breath Volume		1.976	19:11
Air Blank		0.000	19:12
Air Blank		0.000	19:13
Subject Sample	# 2	0.193	19:14
Breath Volume		1.585	19:14
Air Blank		0.000	19:15
Control Test		0.082	19:15
Air Blank		0.000	19:15
Diagnostics Check		OK	19:15

Data Download Date/Time: 12/22/2021 11:12

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HART,JOSEPH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: KIMBLE

DATE/TIME OF TEST: 11/14/2021 03:26:14

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:18	
Control Test		0.083	03:18	
Air Blank		0.000	03:18	
Subject Sample	# 1		03:22	No Sample Provided
Breath Volume		0.000	03:22	
Air Blank		0.000	03:22	
Air Blank		0.000	03:24	
Subject Sample	# 2		03:24	Subject Test Refused
Breath Volume		0.000	03:24	
Air Blank		0.000	03:25	
Control Test		0.084	03:25	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 12/22/2021 11:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GLENECK,JORDAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: STUKES

DATE/TIME OF TEST: 11/14/2021 04:36:39

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.084	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.060	04:31
Breath Volume		1.328	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.057	04:35
Breath Volume		1.210	04:35
Air Blank		0.000	04:35
Control Test		0.081	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 12/22/2021 11:12

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GEOGHAGAN,ANDREW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: HOKE

DATE/TIME OF TEST: 11/14/2021 05:29:10

Diagnostics Check		OK	05:22
Air Blank		0.000	05:23
Control Test		0.082	05:23
Air Blank		0.000	05:24
Subject Sample	# 1	0.189	05:24
Breath Volume		1.664	05:24
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.193	05:27
Breath Volume		1.281	05:27
Air Blank		0.000	05:28
Control Test		0.082	05:28
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 12/22/2021 11:12

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GENTRY,DYLAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: PICKETT

DATE/TIME OF TEST: 11/14/2021 20:13:52

Diagnostics Check		OK	20:07
Air Blank		0.000	20:07
Control Test		0.082	20:07
Air Blank		0.000	20:08
Subject Sample	# 1	0.194	20:08
Breath Volume		1.906	20:08
Air Blank		0.000	20:09
Air Blank		0.000	20:11
Subject Sample	# 2	0.204	20:12
Breath Volume		1.832	20:12
Air Blank		0.000	20:12
Control Test		0.083	20:13
Air Blank		0.000	20:13
Diagnostics Check		OK	20:13

Data Download Date/Time: 12/22/2021 11:12

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PAPPAS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: IMM

DATE/TIME OF TEST: 11/15/2021 01:48:49

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.082	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.247	01:42
Breath Volume		1.875	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.248	01:47
Breath Volume		2.960	01:47
Air Blank		0.000	01:47
Control Test		0.082	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 12/22/2021 11:12

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,LARRY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GRUNDIN,GABREAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 11/16/2021 22:40:29

Diagnostics Check		OK	22:33
Air Blank		0.000	22:33
Control Test		0.084	22:33
Air Blank		0.000	22:34
Subject Sample	# 1	0.052	22:35
Breath Volume		3.234	22:35
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.049	22:38
Breath Volume		2.085	22:38
Air Blank		0.000	22:39
Control Test		0.081	22:39
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 12/22/2021 11:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ZUPOKFSKA-THOMPSON

DATE/TIME OF TEST: 11/17/2021 03:34:59

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.082	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.183	03:30
Breath Volume		1.718	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:32
Subject Sample	# 2	0.178	03:33
Breath Volume		2.007	03:33
Air Blank		0.000	03:34
Control Test		0.080	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 12/22/2021 11:12

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GLENECK,JORDAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: FUJAN

DATE/TIME OF TEST: 11/17/2021 21:43:02

Diagnostics Check		OK	21:35
Air Blank		0.000	21:35
Control Test		0.083	21:36
Air Blank		0.000	21:36
Subject Sample	# 1	0.214	21:38
Breath Volume		2.007	21:38
Air Blank		0.000	21:39
Air Blank		0.000	21:41
Subject Sample	# 2	0.212	21:41
Breath Volume		2.976	21:41
Air Blank		0.000	21:42
Control Test		0.080	21:42
Air Blank		0.000	21:42
Diagnostics Check		OK	21:43

Data Download Date/Time: 12/22/2021 11:12

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: MAPLES

DATE/TIME OF TEST: 11/18/2021 00:41:17

Diagnostics Check		OK	00:32	
Air Blank		0.000	00:33	
Control Test		0.081	00:33	
Air Blank		0.000	00:33	
Subject Sample	# 1		00:37	No Sample Provided
Breath Volume		0.000	00:37	
Air Blank		0.000	00:37	
Air Blank		0.000	00:39	
Subject Sample	# 2		00:39	Subject Test Refused
Breath Volume		0.000	00:39	
Air Blank		0.000	00:40	
Control Test		0.083	00:40	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 12/22/2021 11:12

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,LARRY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HACKETT,TRAVIS
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/20/2021 04:05:25

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.083	03:58
Air Blank		0.000	03:58
Subject Sample	# 1	0.132	03:59
Breath Volume		1.589	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.135	04:03
Breath Volume		1.632	04:03
Air Blank		0.000	04:04
Control Test		0.081	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 12/22/2021 11:13

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HACKETT,TRAVIS
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: LOPEZ JR

DATE/TIME OF TEST: 11/26/2021 04:19:20

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.082	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.158	04:13
Breath Volume		1.929	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:15
Subject Sample	# 2	0.165	04:17
Breath Volume		2.820	04:17
Air Blank		0.000	04:18
Control Test		0.082	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 12/22/2021 11:13

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/26/2021 23:23:58

Diagnostics Check		OK	23:15	
Air Blank		0.000	23:15	
Control Test		0.083	23:16	
Air Blank		0.000	23:16	
Subject Sample	# 1		23:19	No Sample Provided
Breath Volume		0.000	23:19	
Air Blank		0.000	23:20	
Air Blank		0.000	23:22	
Subject Sample	# 2		23:22	Subject Test Refused
Breath Volume		0.000	23:22	
Air Blank		0.000	23:23	
Control Test		0.083	23:23	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 12/22/2021 11:13

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURKE,RYAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 11/27/2021 01:21:58

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:13	
Control Test		0.083	01:13	
Air Blank		0.000	01:14	
Subject Sample	# 1		01:17	No Sample Provided
Breath Volume		0.000	01:17	
Air Blank		0.000	01:17	
Air Blank		0.000	01:19	
Subject Sample	# 2		01:20	Subject Test Refused
Breath Volume		0.000	01:20	
Air Blank		0.000	01:21	
Control Test		0.083	01:21	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 12/22/2021 11:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MEDINA SIERRA

DATE/TIME OF TEST: 11/27/2021 04:59:55

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.083	04:51
Air Blank		0.000	04:51
Subject Sample	# 1	0.140	04:52
Breath Volume		2.132	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.126	04:58
Breath Volume		1.226	04:58
Air Blank		0.000	04:58
Control Test		0.083	04:59
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 12/22/2021 11:13

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 11/04/2021 01:21:18

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.079	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.069	01:16
Breath Volume		3.203	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.066	01:19
Breath Volume		3.000	01:19
Air Blank		0.000	01:20
Control Test		0.079	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 12/14/2021 10:37

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: SEIBOLD

DATE/TIME OF TEST: 11/05/2021 01:38:30

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:33
Subject Sample	# 1	0.138	01:34
Breath Volume		1.855	01:34
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.146	01:37
Breath Volume		2.828	01:37
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 12/14/2021 10:37

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MOBLEY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 11/05/2021 23:26:09

Diagnostics Check		OK	23:20
Air Blank		0.000	23:20
Control Test		0.080	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.311	23:21
Breath Volume		2.488	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.294	23:24
Breath Volume		2.113	23:24
Air Blank		0.000	23:25
Control Test		0.081	23:25
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 12/14/2021 10:37

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: REGAN

DATE/TIME OF TEST: 11/11/2021 00:23:41

Diagnostics Check		OK	00:17
Air Blank		0.000	00:18
Control Test		0.080	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.067	00:19
Breath Volume		1.441	00:19
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.065	00:22
Breath Volume		1.308	00:22
Air Blank		0.000	00:22
Control Test		0.081	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 12/14/2021 10:38

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 11/12/2021 02:14:05

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:11	
Control Test		0.079	02:11	
Air Blank		0.000	02:11	
Subject Sample	# 1		02:13	Sequence Aborted
Breath Volume		0.000	02:13	
Air Blank		0.000	02:14	

Data Download Date/Time: 12/14/2021 10:38

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 11/12/2021 02:47:40

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:34	
Control Test		0.079	02:35	
Air Blank		0.000	02:35	
Subject Sample	# 1	0.126	02:37	
Breath Volume		1.527	02:37	
Air Blank		0.000	02:37	
Air Blank		0.000	02:39	
Subject Sample	# 2	0.123	02:43	Volume Not Met
Breath Volume		0.000	02:43	
Air Blank		0.000	02:43	
Air Blank		0.000	02:45	
Subject Sample	# 3	0.142	02:46	
Breath Volume		1.386	02:46	
Air Blank		0.000	02:46	
Control Test		0.078	02:47	
Air Blank		0.000	02:47	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 12/14/2021 10:38

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 11/13/2021 01:53:43

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:48
Subject Sample	# 1	0.120	01:49
Breath Volume		1.648	01:49
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.119	01:52
Breath Volume		1.484	01:52
Air Blank		0.000	01:52
Control Test		0.080	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 12/14/2021 10:38

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: MARTINEZ ALVARADO

DATE/TIME OF TEST: 11/14/2021 00:52:42

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.080	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.162	00:44
Breath Volume		1.933	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.185	00:47
Breath Volume		3.281	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 3	0.180	00:51
Breath Volume		4.031	00:51
Air Blank		0.000	00:51
Control Test		0.080	00:52
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 12/14/2021 10:38

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 11/24/2021 23:58:47

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.079	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.144	23:53
Breath Volume		2.140	23:53
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.154	23:57
Breath Volume		1.574	23:57
Air Blank		0.000	23:57
Control Test		0.078	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 12/14/2021 10:38

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: DE JESUS RODRIGUEZ

DATE/TIME OF TEST: 11/25/2021 02:36:30

Diagnostics Check		OK	02:29
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.080	02:31
Breath Volume		3.171	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.082	02:35
Breath Volume		3.558	02:35
Air Blank		0.000	02:35
Control Test		0.079	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 12/14/2021 10:38

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: VANDEMARK

DATE/TIME OF TEST: 11/27/2021 22:53:31

Diagnostics Check		OK	22:46
Air Blank		0.000	22:47
Control Test		0.081	22:47
Air Blank		0.000	22:48
Subject Sample	# 1	0.251	22:49
Breath Volume		2.375	22:49
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.239	22:52
Breath Volume		1.464	22:52
Air Blank		0.000	22:52
Control Test		0.080	22:52
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 12/14/2021 10:38

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: DUNKER

DATE/TIME OF TEST: 11/02/2021 05:41:57

Diagnostics Check		OK	05:39	
Air Blank		0.000	05:39	
Control Test		0.079	05:39	
Air Blank		0.000	05:40	
Subject Sample	# 1		05:40	Subject Test Refused
Breath Volume		0.000	05:40	
Air Blank		0.000	05:41	
Control Test		0.082	05:41	
Air Blank		0.000	05:41	
Diagnostics Check		OK	05:41	

Data Download Date/Time: 11/29/2021 12:14

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,RYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 12:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 11/03/2021 01:35:50

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.080	01:33	
Air Blank		0.000	01:34	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:34	
Control Test		0.080	01:35	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 11/29/2021 12:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERTENS,WILLIAM,
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HALLBERG,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MORALES DIAZ

DATE/TIME OF TEST: 11/07/2021 02:34:23

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.081	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.181	02:28
Breath Volume		1.730	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.189	02:32
Breath Volume		2.542	02:32
Air Blank		0.000	02:33
Control Test		0.077	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 11/29/2021 12:14

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERTENS,WILLIAM,
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: JOHNS,WILLIAM
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: BILLER

DATE/TIME OF TEST: 11/20/2021 00:11:43

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.080	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.190	00:07
Breath Volume		3.222	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.192	00:10
Breath Volume		3.777	00:10
Air Blank		0.000	00:10
Control Test		0.080	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 11/29/2021 12:14

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: FOTHERINGHAM

DATE/TIME OF TEST: 11/20/2021 02:25:24

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:22	
Control Test		0.080	02:22	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:24	Subject Test Refused
Breath Volume		0.000	02:24	
Air Blank		0.000	02:24	
Control Test		0.081	02:24	
Air Blank		0.000	02:25	
Diagnostics Check		OK	02:25	

Data Download Date/Time: 11/29/2021 12:15

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: SHORTER

DATE/TIME OF TEST: 11/24/2021 05:19:40

Diagnostics Check		OK	05:13
Air Blank		0.000	05:13
Control Test		0.080	05:13
Air Blank		0.000	05:14
Subject Sample	# 1	0.166	05:14
Breath Volume		3.820	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.164	05:18
Breath Volume		3.515	05:18
Air Blank		0.000	05:18
Control Test		0.081	05:19
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 11/29/2021 12:15

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: VERA

DATE/TIME OF TEST: 11/25/2021 03:12:29

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.104	03:07
Breath Volume		3.242	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.105	03:10
Breath Volume		3.410	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 11/29/2021 12:15

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERTENS,WILLIAM,
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STUART,IAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 11/26/2021 00:00:17

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.080	23:54
Air Blank		0.000	23:54
Subject Sample	# 1	0.228	23:55
Breath Volume		2.585	23:55
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.217	23:58
Breath Volume		2.195	23:58
Air Blank		0.000	23:59
Control Test		0.078	23:59
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 11/29/2021 12:15

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 11/28/2021 16:54:53

Diagnostics Check		OK	16:48
Air Blank		0.000	16:49
Control Test		0.080	16:49
Air Blank		0.000	16:49
Subject Sample	# 1	0.000	16:50
Breath Volume		2.667	16:50
Air Blank		0.000	16:50
Air Blank		0.000	16:52
Subject Sample	# 2	0.000	16:53
Breath Volume		1.964	16:53
Air Blank		0.000	16:53
Control Test		0.077	16:54
Air Blank		0.000	16:54
Diagnostics Check		OK	16:54

Data Download Date/Time: 11/29/2021 12:16

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 11/25/2021 23:47:44

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.080	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.214	23:42
Breath Volume		1.628	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.233	23:46
Breath Volume		1.671	23:46
Air Blank		0.000	23:46
Control Test		0.077	23:47
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 12/13/2021 13:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNETT,JAMES,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BARNETT,JAMES
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: GENTRY

DATE/TIME OF TEST: 11/06/2021 22:03:33

Diagnostics Check		OK	21:57
Air Blank		0.000	21:57
Control Test		0.082	21:58
Air Blank		0.000	21:58
Subject Sample	# 1	0.000	21:59
Breath Volume		1.343	21:59
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.000	22:02
Breath Volume		1.660	22:02
Air Blank		0.000	22:02
Control Test		0.083	22:02
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 11/10/2021 18:56

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 11/20/2021 03:33:32

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.081	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.175	03:28
Breath Volume		2.574	03:28
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.176	03:31
Breath Volume		2.570	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 12/13/2021 16:11

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: CLARK,COLTIN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: SRONGPRAPA

DATE/TIME OF TEST: 11/20/2021 05:17:49

Diagnostics Check		OK	05:11
Air Blank		0.000	05:11
Control Test		0.082	05:12
Air Blank		0.000	05:12
Subject Sample	# 1	0.191	05:13
Breath Volume		2.078	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:15
Subject Sample	# 2	0.187	05:16
Breath Volume		2.078	05:16
Air Blank		0.000	05:16
Control Test		0.080	05:17
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 12/13/2021 16:11

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORELLANA,ADALBERTO,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MILLENDER,MONICA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 15:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/25/2021 17:22:05

Diagnostics Check		OK	17:15
Air Blank		0.000	17:15
Control Test		0.080	17:16
Air Blank		0.000	17:16
Subject Sample	# 1	0.110	17:17
Breath Volume		1.507	17:17
Air Blank		0.000	17:18
Air Blank		0.000	17:20
Subject Sample	# 2	0.115	17:20
Breath Volume		1.394	17:20
Air Blank		0.000	17:21
Control Test		0.080	17:21
Air Blank		0.000	17:22
Diagnostics Check		OK	17:22

Data Download Date/Time: 12/13/2021 16:11

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: HARPOLE

DATE/TIME OF TEST: 11/27/2021 02:21:11

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.082	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.110	02:16
Breath Volume		3.125	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.110	02:19
Breath Volume		2.687	02:19
Air Blank		0.000	02:20
Control Test		0.079	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 12/13/2021 16:11

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BOBO

DATE/TIME OF TEST: 11/27/2021 04:46:58

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:43	
Control Test		0.079	04:43	
Air Blank		0.000	04:44	
Subject Sample	# 1		04:45	Subject Test Refused
Breath Volume		0.000	04:45	
Air Blank		0.000	04:46	
Control Test		0.081	04:46	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 12/13/2021 16:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNDT,DAVID,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MANNING,CARL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/06/2021 21:03:39

Diagnostics Check		OK	20:56
Air Blank		0.000	20:57
Control Test		0.082	20:57
Air Blank		0.000	20:58
Subject Sample	# 1	0.226	20:59
Breath Volume		2.687	20:59
Air Blank		0.000	20:59
Air Blank		0.000	21:01
Subject Sample	# 2	0.220	21:02
Breath Volume		2.769	21:02
Air Blank		0.000	21:02
Control Test		0.078	21:03
Air Blank		0.000	21:03
Diagnostics Check		OK	21:03

Data Download Date/Time: 11/10/2021 19:08

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NABULSI,T,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NABULSI,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 11/09/2021 21:13:49

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:11	
Control Test		0.082	21:11	
Air Blank		0.000	21:12	
Subject Sample	# 1		21:12	Subject Test Refused
Breath Volume		0.000	21:12	
Air Blank		0.000	21:12	
Control Test		0.081	21:13	
Air Blank		0.000	21:13	
Diagnostics Check		OK	21:13	

Data Download Date/Time: 11/10/2021 19:08

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUSTER,CORY,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GUSTER,CORY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/13/2021 01:37:35

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.080	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.154	01:32
Breath Volume		1.648	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.165	01:36
Breath Volume		2.652	01:36
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 12/13/2021 12:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEW,REECE,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: REGISTER

DATE/TIME OF TEST: 11/18/2021 03:09:48

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:07	
Control Test		0.080	03:07	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:08	
Control Test		0.081	03:09	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 12/13/2021 12:21

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,DARRAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MORGAN,DARRAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: BLALOCK

DATE/TIME OF TEST: 11/01/2021 23:43:53

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:32	
Control Test		0.079	23:32	
Air Blank		0.000	23:33	
Subject Sample	# 1	0.141	23:36	Volume Not Met
Breath Volume		0.089	23:36	
Air Blank		0.000	23:37	
Air Blank		0.000	23:39	
Subject Sample	# 2	0.165	23:42	Volume Not Met
Breath Volume		0.898	23:42	
Air Blank		0.000	23:42	
Control Test		0.079	23:43	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 11/22/2021 10:17

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLE,MADDISON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: FLAHERTY,Z
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/11/2021 06:06:49

Diagnostics Check		OK	05:57
Air Blank		0.000	05:58
Control Test		0.079	05:58
Air Blank		0.000	05:58
Subject Sample	# 1	0.222	05:59
Breath Volume		3.222	05:59
Air Blank		0.000	06:00
Air Blank		0.000	06:01
Subject Sample	# 2	0.194	06:02
Breath Volume		2.359	06:02
Air Blank		0.000	06:03
Air Blank		0.000	06:04
Subject Sample	# 3	0.211	06:05
Breath Volume		3.105	06:05
Air Blank		0.000	06:05
Control Test		0.078	06:06
Air Blank		0.000	06:06
Diagnostics Check		OK	06:06

Data Download Date/Time: 11/22/2021 10:17

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS JR,STEPHANIE,G
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: SAVONA,JOHN
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: DAHAN

DATE/TIME OF TEST: 11/13/2021 15:54:35

Diagnostics Check		OK	15:51	
Air Blank		0.000	15:52	
Control Test		0.079	15:52	
Air Blank		0.000	15:53	
Subject Sample	# 1		15:53	Subject Test Refused
Breath Volume		0.000	15:53	
Air Blank		0.000	15:53	
Control Test		0.079	15:53	
Air Blank		0.000	15:54	
Diagnostics Check		OK	15:54	

Data Download Date/Time: 11/22/2021 10:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,JOSHUA,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILSON,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: DRIESSE

DATE/TIME OF TEST: 11/19/2021 17:45:52

Diagnostics Check		OK	17:39
Air Blank		0.000	17:39
Control Test		0.080	17:40
Air Blank		0.000	17:40
Subject Sample	# 1	0.000	17:41
Breath Volume		2.097	17:41
Air Blank		0.000	17:41
Air Blank		0.000	17:43
Subject Sample	# 2	0.000	17:44
Breath Volume		2.441	17:44
Air Blank		0.000	17:44
Control Test		0.079	17:45
Air Blank		0.000	17:45
Diagnostics Check		OK	17:45

Data Download Date/Time: 11/22/2021 10:17

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCREE,G,L
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LARRY,S
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: HIGHLEY

DATE/TIME OF TEST: 11/24/2021 03:25:12

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.080	03:24	
Air Blank		03:24	RFI Detect
Air Blank	0.000	03:25	

Data Download Date/Time: 12/16/2021 10:29

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCREE,G,L
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LARRY,S
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: HIGHLEY

DATE/TIME OF TEST: 11/24/2021 03:29:43

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:27	
Control Test		0.080	03:27	
Air Blank		0.000	03:27	
Subject Sample	# 1		03:28	Subject Test Refused
Breath Volume		0.000	03:28	
Air Blank		0.000	03:28	
Control Test		0.081	03:29	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 12/16/2021 10:29

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,JONATHON,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BENNETT,JONATHON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: RICKETTS

DATE/TIME OF TEST: 11/25/2021 10:44:17

Diagnostics Check		OK	10:38
Air Blank		0.000	10:38
Control Test		0.079	10:38
Air Blank		0.000	10:39
Subject Sample	# 1	0.154	10:39
Breath Volume		3.394	10:39
Air Blank		0.000	10:40
Air Blank		0.000	10:42
Subject Sample	# 2	0.156	10:42
Breath Volume		2.742	10:42
Air Blank		0.000	10:43
Control Test		0.079	10:43
Air Blank		0.000	10:44
Diagnostics Check		OK	10:44

Data Download Date/Time: 12/16/2021 10:29

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,JORDI,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAEZ,JORDI
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PEREZ MENDEZ

DATE/TIME OF TEST: 11/26/2021 02:48:05

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.268	02:43
Breath Volume		1.621	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.266	02:46
Breath Volume		1.585	02:46
Air Blank		0.000	02:47
Control Test		0.079	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 12/16/2021 10:29

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON,M,L
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: NICHOLSON,M
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 10/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: BUSTER

DATE/TIME OF TEST: 11/04/2021 00:31:26

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:20	
Control Test		0.079	00:20	
Air Blank		0.000	00:21	
Subject Sample	# 1		00:24	No Sample Provided
Breath Volume		0.000	00:24	
Air Blank		0.000	00:24	
Air Blank		0.000	00:26	
Subject Sample	# 2		00:30	No Sample Provided
Breath Volume		0.000	00:30	
Air Blank		0.000	00:30	
Control Test		0.077	00:30	
Air Blank		0.000	00:31	
Diagnostics Check		OK	00:31	

Data Download Date/Time: 11/26/2021 15:26

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HARMAN

DATE/TIME OF TEST: 11/07/2021 03:31:41

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.079	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.258	03:27
Breath Volume		1.691	03:27
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.254	03:30
Breath Volume		1.250	03:30
Air Blank		0.000	03:30
Control Test		0.080	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 11/26/2021 15:26

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,DANIEL,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WATSON,CLAYTON
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 10/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: HERRING

DATE/TIME OF TEST: 11/20/2021 18:16:02

Diagnostics Check		OK	18:09
Air Blank		0.000	18:09
Control Test		0.078	18:09
Air Blank		0.000	18:10
Subject Sample	# 1	0.143	18:11
Breath Volume		2.589	18:11
Air Blank		0.000	18:11
Air Blank		0.000	18:13
Subject Sample	# 2	0.132	18:14
Breath Volume		3.074	18:14
Air Blank		0.000	18:15
Control Test		0.077	18:15
Air Blank		0.000	18:15
Diagnostics Check		OK	18:16

Data Download Date/Time: 11/29/2021 10:12

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON,M,
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: LUCKETTE,J
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 10/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: MCQUAY

DATE/TIME OF TEST: 11/24/2021 00:18:30

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.080	00:12
Air Blank		0.000	00:12
Subject Sample	# 1	0.114	00:14
Breath Volume		1.871	00:14
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.108	00:16
Breath Volume		1.128	00:16
Air Blank		0.000	00:17
Control Test		0.077	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 11/29/2021 10:12

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRS,RUSSELL,B
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: HENDERSON,CEDRICK
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 11/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/28/2021 04:29:13

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.167	04:24
Breath Volume		2.644	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.159	04:27
Breath Volume		1.816	04:27
Air Blank		0.000	04:28
Control Test		0.077	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 11/29/2021 10:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH JR,LORNE,J
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: SMITH JR,LORNE
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 10/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MANGULAS

DATE/TIME OF TEST: 11/04/2021 23:18:07

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.078	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.145	23:13
Breath Volume		1.281	23:13
Air Blank		0.000	23:14
Air Blank		0.000	23:15
Subject Sample	# 2	0.146	23:16
Breath Volume		2.335	23:16
Air Blank		0.000	23:17
Control Test		0.076	23:17
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 11/28/2021 23:49

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH JR,LORNE,J
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: SMITH JR,LORNE
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 10/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MADDALINO

DATE/TIME OF TEST: 11/06/2021 01:02:44

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.144	00:57
Breath Volume		1.800	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.133	01:01
Breath Volume		1.449	01:01
Air Blank		0.000	01:01
Control Test		0.079	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 11/28/2021 23:49

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRYDEN,JOHN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DRYDEN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: CANAZON

DATE/TIME OF TEST: 11/10/2021 00:49:00

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.079	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.000	00:44
Breath Volume		1.332	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.000	00:47
Breath Volume		1.605	00:47
Air Blank		0.000	00:48
Control Test		0.077	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:49

Data Download Date/Time: 11/28/2021 23:49

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,RYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 12:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 11/04/2021 01:35:26

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:32	
Control Test		0.081	01:32	
Air Blank		0.000	01:32	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:34	
Control Test		0.081	01:34	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 11/30/2021 21:21

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLS,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZAHCHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 11/04/2021 02:04:16

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:01	
Control Test		0.081	02:02	
Air Blank		0.000	02:02	
Subject Sample	# 1		02:02	Subject Test Refused
Breath Volume		0.000	02:02	
Air Blank		0.000	02:03	
Control Test		0.081	02:03	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 11/30/2021 21:21

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: CAPERTON

DATE/TIME OF TEST: 11/04/2021 03:57:18

Diagnostics Check		OK	03:54	
Air Blank		0.000	03:54	
Control Test		0.081	03:55	
Air Blank		0.000	03:55	
Subject Sample	# 1		03:55	RFI Detect
Breath Volume		0.000	03:55	
Air Blank		0.000	03:56	
Control Test		0.081	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 11/30/2021 21:21

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: CAPERTON

DATE/TIME OF TEST: 11/04/2021 04:04:36

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.081	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.000	04:00
Breath Volume		2.386	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.000	04:03
Breath Volume		2.523	04:03
Air Blank		0.000	04:03
Control Test		0.081	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 11/30/2021 21:21

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: DUBOSE

DATE/TIME OF TEST: 11/05/2021 00:28:16

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.081	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.000	00:23
Breath Volume		3.101	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.000	00:26
Breath Volume		3.714	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 11/30/2021 21:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 11/05/2021 22:53:19

Diagnostics Check		OK	22:48	
Air Blank		0.000	22:49	
Control Test		0.081	22:49	
Air Blank		0.000	22:50	
Subject Sample	# 1		22:51	Subject Test Refused
Breath Volume		0.000	22:51	
Air Blank		0.000	22:52	
Control Test		0.081	22:52	
Air Blank		0.000	22:53	
Diagnostics Check		OK	22:53	

Data Download Date/Time: 11/30/2021 21:21

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: HAMMONTREE

DATE/TIME OF TEST: 11/06/2021 03:11:32

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:00	
Control Test		0.081	03:00	
Air Blank		0.000	03:01	
Subject Sample	# 1		03:04	No Sample Provided
Breath Volume		0.000	03:04	
Air Blank		0.000	03:05	
Air Blank		0.000	03:07	
Subject Sample	# 2		03:10	No Sample Provided
Breath Volume		0.000	03:10	
Air Blank		0.000	03:10	
Control Test		0.081	03:10	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 11/30/2021 21:21

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: SHORT

DATE/TIME OF TEST: 11/06/2021 04:18:21

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.081	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.077	04:13
Breath Volume		4.964	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.079	04:16
Breath Volume		4.769	04:16
Air Blank		0.000	04:17
Control Test		0.080	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 11/30/2021 21:21

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYD,J,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOYD,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: GEARY

DATE/TIME OF TEST: 11/06/2021 18:15:13

Diagnostics Check		OK	18:03	
Air Blank		0.000	18:04	
Control Test		0.081	18:04	
Air Blank		0.000	18:05	
Subject Sample	# 1		18:08	No Sample Provided
Breath Volume		0.000	18:08	
Air Blank		0.000	18:08	
Air Blank		0.000	18:10	
Subject Sample	# 2		18:13	No Sample Provided
Breath Volume		0.000	18:13	
Air Blank		0.000	18:14	
Control Test		0.080	18:14	
Air Blank		0.000	18:15	
Diagnostics Check		OK	18:15	

Data Download Date/Time: 11/30/2021 21:21

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALEXANDER,LLOYD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: BOYER

DATE/TIME OF TEST: 11/06/2021 23:35:04

Diagnostics Check		OK	23:28
Air Blank		0.000	23:29
Control Test		0.080	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.215	23:30
Breath Volume		3.039	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.204	23:33
Breath Volume		3.683	23:33
Air Blank		0.000	23:34
Control Test		0.080	23:34
Air Blank		0.000	23:34
Diagnostics Check		OK	23:35

Data Download Date/Time: 11/30/2021 21:21

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: ECKLEY

DATE/TIME OF TEST: 11/07/2021 01:36:19

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.080	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.116	01:31
Breath Volume		1.886	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:33
Subject Sample	# 2	0.118	01:34
Breath Volume		1.445	01:34
Air Blank		0.000	01:35
Control Test		0.081	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 11/30/2021 21:21

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 11/10/2021 01:49:02

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:44	
Control Test		0.081	01:44	
Air Blank		0.000	01:45	
Subject Sample	# 1	0.000	01:45	
Breath Volume		1.863	01:45	
Air Blank		0.000	01:46	
Air Blank		0.000	01:48	
Subject Sample	# 2	0.000	01:48	
Breath Volume		1.847	01:48	
Air Blank			01:48	RFI Detect
Air Blank			01:49	RFI Detect

Data Download Date/Time: 11/30/2021 21:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 11/11/2021 02:14:37

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.080	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.000	02:10
Breath Volume		1.148	02:10
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.000	02:13
Breath Volume		1.664	02:13
Air Blank		0.000	02:13
Control Test		0.081	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 11/30/2021 21:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: MANDEL

DATE/TIME OF TEST: 11/12/2021 00:46:12

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.080	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.000	00:42
Breath Volume		1.976	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.000	00:44
Breath Volume		1.562	00:44
Air Blank		0.000	00:45
Control Test		0.080	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 11/30/2021 21:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEIB,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SEIB,ALEXANDER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: TRIPLETT

DATE/TIME OF TEST: 11/13/2021 00:31:01

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.080	00:24
Air Blank		0.000	00:24
Subject Sample	# 1	0.280	00:26
Breath Volume		2.539	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:28
Subject Sample	# 2	0.283	00:29
Breath Volume		2.824	00:29
Air Blank		0.000	00:30
Control Test		0.079	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:31

Data Download Date/Time: 11/30/2021 21:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: FENTON

DATE/TIME OF TEST: 11/13/2021 05:38:46

Diagnostics Check		OK	05:30
Air Blank		0.000	05:31
Control Test		0.080	05:31
Air Blank		0.000	05:32
Subject Sample	# 1	0.190	05:32
Breath Volume		2.453	05:32
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.184	05:37
Breath Volume		1.675	05:37
Air Blank		0.000	05:37
Control Test		0.079	05:38
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 11/30/2021 21:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DALGLEISH,KEVIN,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TIEDEMANN,NICHOLAS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 06:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:07

SUBJECT'S LAST NAME: NORDSTROM

DATE/TIME OF TEST: 11/13/2021 07:28:03

Diagnostics Check	OK	07:27	
Air Blank		07:27	RFI Detect
Air Blank	0.000	07:28	

Data Download Date/Time: 11/30/2021 21:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,JL,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: FARLEY

DATE/TIME OF TEST: 11/14/2021 02:08:41

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.081	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.181	02:03
Breath Volume		2.109	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.173	02:07
Breath Volume		1.332	02:07
Air Blank		0.000	02:07
Control Test		0.079	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 11/30/2021 21:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNDT,DAVID,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MUNDT,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: CAVE

DATE/TIME OF TEST: 11/14/2021 22:55:42

Diagnostics Check		OK	22:52	
Air Blank		0.000	22:53	
Control Test		0.081	22:53	
Air Blank		0.000	22:54	
Subject Sample	# 1		22:54	Subject Test Refused
Breath Volume		0.000	22:54	
Air Blank		0.000	22:54	
Control Test		0.081	22:55	
Air Blank		0.000	22:55	
Diagnostics Check		OK	22:55	

Data Download Date/Time: 11/30/2021 21:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESTEK,JOHN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: COOTS

DATE/TIME OF TEST: 11/15/2021 02:52:13

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.081	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.138	02:47
Breath Volume		1.878	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:49
Subject Sample	# 2	0.134	02:50
Breath Volume		2.246	02:50
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 11/30/2021 21:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUCKER,JEREMY,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DUCKER,JEREMY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: FORTAG

DATE/TIME OF TEST: 11/15/2021 03:43:47

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.081	03:37
Air Blank		0.000	03:38
Subject Sample	# 1	0.159	03:39
Breath Volume		3.238	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.163	03:42
Breath Volume		3.308	03:42
Air Blank		0.000	03:42
Control Test		0.080	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 11/30/2021 21:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,ANN,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BLACK,KURT
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: GRESSETT

DATE/TIME OF TEST: 11/15/2021 16:37:48

Diagnostics Check		OK	16:29	
Air Blank		0.000	16:29	
Control Test		0.080	16:29	
Air Blank		0.000	16:30	
Subject Sample	# 1	0.293	16:30	
Breath Volume		1.628	16:30	
Air Blank		0.000	16:31	
Air Blank		0.000	16:33	
Subject Sample	# 2		16:35	Slope Not Met
Breath Volume		1.328	16:35	
Air Blank		0.000	16:36	
Control Test		0.076	16:37	
Air Blank		0.000	16:37	
Diagnostics Check		OK	16:37	

Data Download Date/Time: 11/30/2021 21:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEI,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SEIB,ALEXANDER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: STACEY

DATE/TIME OF TEST: 11/18/2021 00:45:34

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.080	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.000	00:41
Breath Volume		2.457	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.000	00:44
Breath Volume		2.500	00:44
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 11/30/2021 21:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: FOLEY

DATE/TIME OF TEST: 11/18/2021 02:17:42

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:14	
Control Test		0.080	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:16	
Control Test		0.080	02:17	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 11/30/2021 21:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 11/19/2021 22:07:29

Diagnostics Check		OK	21:59
Air Blank		0.000	21:59
Control Test		0.081	22:00
Air Blank		0.000	22:00
Subject Sample	# 1	0.113	22:01
Breath Volume		1.253	22:01
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.108	22:05
Breath Volume		1.191	22:05
Air Blank		0.000	22:06
Control Test		0.081	22:06
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 11/30/2021 21:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERA,DANTE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AVERA,DANTE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: SHEFFIELD

DATE/TIME OF TEST: 11/19/2021 22:19:31

Diagnostics Check		OK	22:12
Air Blank		0.000	22:13
Control Test		0.081	22:13
Air Blank		0.000	22:13
Subject Sample	# 1	0.100	22:15
Breath Volume		2.117	22:15
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.095	22:17
Breath Volume		1.597	22:17
Air Blank		0.000	22:18
Control Test		0.080	22:18
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 11/30/2021 21:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFORD,DERIC,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WELFORD,DERIC
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: HANDY

DATE/TIME OF TEST: 11/20/2021 15:16:48

Diagnostics Check		OK	15:10
Air Blank		0.000	15:11
Control Test		0.080	15:11
Air Blank		0.000	15:12
Subject Sample	# 1	0.283	15:12
Breath Volume		2.875	15:12
Air Blank		0.000	15:13
Air Blank		0.000	15:14
Subject Sample	# 2	0.296	15:15
Breath Volume		2.542	15:15
Air Blank		0.000	15:15
Control Test		0.078	15:16
Air Blank		0.000	15:16
Diagnostics Check		OK	15:16

Data Download Date/Time: 11/30/2021 21:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORENE,AIDAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NORENE,AIDAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: MOFFO

DATE/TIME OF TEST: 11/20/2021 22:21:55

Diagnostics Check		OK	22:15
Air Blank		0.000	22:15
Control Test		0.081	22:16
Air Blank		0.000	22:16
Subject Sample	# 1	0.121	22:17
Breath Volume		2.578	22:17
Air Blank		0.000	22:18
Air Blank		0.000	22:19
Subject Sample	# 2	0.115	22:20
Breath Volume		1.941	22:20
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 11/30/2021 21:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WATERS,CALEB
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RICCIO

DATE/TIME OF TEST: 11/21/2021 03:58:44

Diagnostics Check	OK	03:58	
Air Blank		03:58	RFI Detect
Air Blank	0.000	03:58	

Data Download Date/Time: 11/30/2021 21:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WATERS,CALEB
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RICCIO

DATE/TIME OF TEST: 11/21/2021 04:02:52

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:00	
Control Test		0.081	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 1		04:01	Subject Test Refused
Breath Volume		0.000	04:01	
Air Blank		0.000	04:01	
Control Test		0.080	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 11/30/2021 21:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: ALLAIRE

DATE/TIME OF TEST: 11/21/2021 23:09:17

Diagnostics Check		OK	23:02
Air Blank		0.000	23:03
Control Test		0.080	23:03
Air Blank		0.000	23:04
Subject Sample	# 1	0.246	23:04
Breath Volume		1.410	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.250	23:07
Breath Volume		2.031	23:07
Air Blank		0.000	23:08
Control Test		0.078	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 11/30/2021 21:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEIB,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SEIB,ALEXANDER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: EDDINS

DATE/TIME OF TEST: 11/22/2021 05:59:51

Diagnostics Check		OK	05:53
Air Blank		0.000	05:54
Control Test		0.082	05:54
Air Blank		0.000	05:55
Subject Sample	# 1	0.000	05:55
Breath Volume		3.699	05:55
Air Blank		0.000	05:56
Air Blank		0.000	05:58
Subject Sample	# 2	0.000	05:58
Breath Volume		2.660	05:58
Air Blank		0.000	05:58
Control Test		0.080	05:59
Air Blank		0.000	05:59
Diagnostics Check		OK	05:59

Data Download Date/Time: 11/30/2021 21:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/22/2021 21:45:36

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:42	
Control Test		0.080	21:42	
Air Blank		0.000	21:42	
Subject Sample	# 1		21:44	Subject Test Refused
Breath Volume		0.000	21:44	
Air Blank		0.000	21:44	
Control Test		0.080	21:45	
Air Blank		0.000	21:45	
Diagnostics Check		OK	21:45	

Data Download Date/Time: 11/30/2021 21:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/23/2021 02:47:54

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.081	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.000	02:43
Breath Volume		1.734	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.000	02:46
Breath Volume		1.183	02:46
Air Blank		0.000	02:46
Control Test		0.081	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 11/30/2021 21:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEIB,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SEIB,ALEXANDER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: HUSTY

DATE/TIME OF TEST: 11/23/2021 03:23:23

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:18	
Control Test		0.079	03:18	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:22	Subject Test Refused
Breath Volume		0.000	03:22	
Air Blank		0.000	03:22	
Control Test		0.080	03:22	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 11/30/2021 21:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMS,KYLE,N
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SIMMS,KYLE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CRUMPTON

DATE/TIME OF TEST: 11/27/2021 00:51:54

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.081	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.094	00:47
Breath Volume		3.175	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:49
Subject Sample	# 2	0.095	00:50
Breath Volume		3.367	00:50
Air Blank		0.000	00:50
Control Test		0.081	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 11/30/2021 21:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: WALES

DATE/TIME OF TEST: 11/27/2021 02:12:12

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.081	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.072	02:07
Breath Volume		2.847	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.076	02:10
Breath Volume		3.832	02:10
Air Blank		0.000	02:11
Control Test		0.081	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 11/30/2021 21:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRANTHAM,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BARCLAY

DATE/TIME OF TEST: 11/27/2021 03:48:09

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.037	03:43
Breath Volume		3.531	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.035	03:46
Breath Volume		3.339	03:46
Air Blank		0.000	03:47
Control Test		0.079	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 11/30/2021 21:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAUVET,N
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 11/27/2021 04:10:20

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:59	
Control Test		0.080	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 1		04:03	No Sample Provided
Breath Volume		0.000	04:03	
Air Blank		0.000	04:03	
Air Blank		0.000	04:05	
Subject Sample	# 2		04:08	No Sample Provided
Breath Volume		0.000	04:08	
Air Blank		0.000	04:09	
Control Test		0.080	04:09	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:10	

Data Download Date/Time: 11/30/2021 21:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WEAVER,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: HEARN

DATE/TIME OF TEST: 11/27/2021 05:04:32

Diagnostics Check		OK	04:57
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.156	04:59
Breath Volume		1.355	04:59
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.148	05:02
Breath Volume		1.429	05:02
Air Blank		0.000	05:03
Control Test		0.080	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 11/30/2021 21:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 11/28/2021 04:07:07

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.081	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.106	04:02
Breath Volume		3.078	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:04
Subject Sample	# 2	0.111	04:05
Breath Volume		4.203	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 11/30/2021 21:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLE,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BEAUDRY

DATE/TIME OF TEST: 11/29/2021 01:06:38

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.000	01:02
Breath Volume		2.367	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.000	01:05
Breath Volume		2.222	01:05
Air Blank		0.000	01:05
Control Test		0.081	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 11/30/2021 21:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERA,DANTE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AVERA,DANTE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: BROTHERS

DATE/TIME OF TEST: 11/29/2021 02:23:32

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.231	02:18
Breath Volume		1.808	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.217	02:21
Breath Volume		1.871	02:21
Air Blank		0.000	02:22
Control Test		0.076	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 11/30/2021 21:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DANDRA,M
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: THOMAS,JOEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 11/23/2021 00:58:17

Diagnostics Check		OK	00:51
Air Blank		0.000	00:52
Control Test		0.081	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.195	00:53
Breath Volume		2.289	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.197	00:56
Breath Volume		1.980	00:56
Air Blank		0.000	00:57
Control Test		0.077	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 12/20/2021 15:38

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: DIAZ,A
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 11/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 11/25/2021 01:28:49

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.082	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.236	01:20
Breath Volume		1.464	01:20
Air Blank		0.000	01:22
Air Blank		0.000	01:23
Subject Sample	# 2	0.203	01:23
Breath Volume		1.281	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 3	0.232	01:27
Breath Volume		2.859	01:27
Air Blank		0.000	01:27
Control Test		0.076	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 12/20/2021 15:38

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DANDRA,M
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: THOMAS,JOEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 11/27/2021 04:05:24

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:57	
Control Test		0.081	03:57	
Air Blank		0.000	03:58	
Subject Sample	# 1		04:01	No Sample Provided
Breath Volume		0.000	04:01	
Air Blank		0.000	04:01	
Air Blank		0.000	04:03	
Subject Sample	# 2		04:03	Subject Test Refused
Breath Volume		0.000	04:03	
Air Blank		0.000	04:04	
Control Test		0.081	04:04	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 12/20/2021 15:38

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,DARRAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MORGAN,DARRAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: YINGWANE

DATE/TIME OF TEST: 11/08/2021 18:52:23

Diagnostics Check		OK	18:45
Air Blank		0.000	18:46
Control Test		0.081	18:46
Air Blank		0.000	18:47
Subject Sample	# 1	0.222	18:47
Breath Volume		1.316	18:47
Air Blank		0.000	18:48
Air Blank		0.000	18:50
Subject Sample	# 2	0.222	18:50
Breath Volume		1.820	18:50
Air Blank		0.000	18:51
Control Test		0.079	18:51
Air Blank		0.000	18:52
Diagnostics Check		OK	18:52

Data Download Date/Time: 11/29/2021 14:54

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 11/08/2021 19:23:21

Diagnostics Check		OK	19:17
Air Blank		0.000	19:17
Control Test		0.080	19:18
Air Blank		0.000	19:18
Subject Sample	# 1	0.130	19:18
Breath Volume		3.644	19:18
Air Blank		0.000	19:19
Air Blank		0.000	19:21
Subject Sample	# 2	0.123	19:21
Breath Volume		3.699	19:21
Air Blank		0.000	19:22
Control Test		0.078	19:22
Air Blank		0.000	19:23
Diagnostics Check		OK	19:23

Data Download Date/Time: 11/29/2021 14:54

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 11/10/2021 01:48:03

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:38	
Control Test		0.081	01:39	
Air Blank		0.000	01:39	
Subject Sample	# 1	0.144	01:42	Volume Not Met
Breath Volume		0.769	01:42	
Air Blank		0.000	01:43	
Air Blank		0.000	01:45	
Subject Sample	# 2		01:46	Subject Test Refused
Breath Volume		0.566	01:46	
Air Blank		0.000	01:47	
Control Test		0.077	01:47	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:48	

Data Download Date/Time: 11/29/2021 14:54

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAY,DARREL,E
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: KERNAN,MITCHELL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MCAULEY

DATE/TIME OF TEST: 11/12/2021 04:01:05

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:49	
Control Test		0.080	03:50	
Air Blank		0.000	03:50	
Subject Sample	# 1	0.139	03:53	Volume Not Met
Breath Volume		0.156	03:53	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 2	0.136	03:59	Volume Not Met
Breath Volume		0.550	03:59	
Air Blank		0.000	04:00	
Control Test		0.075	04:00	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 11/29/2021 14:54

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICAHRD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: GUTH

DATE/TIME OF TEST: 11/13/2021 21:03:42

Diagnostics Check		OK	21:00	
Air Blank		0.000	21:01	
Control Test		0.081	21:01	
Air Blank		0.000	21:02	
Subject Sample	# 1		21:02	Subject Test Refused
Breath Volume		0.000	21:02	
Air Blank		0.000	21:02	
Control Test		0.081	21:03	
Air Blank		0.000	21:03	
Diagnostics Check		OK	21:03	

Data Download Date/Time: 11/29/2021 14:54

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DANDRA,M
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: STEPHENS,MAURICE
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/18/2021 03:26:40

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.144	03:22
Breath Volume		1.488	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.140	03:25
Breath Volume		1.531	03:25
Air Blank		0.000	03:25
Control Test		0.077	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 11/29/2021 14:54

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YASUDA,SHOHEI,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: YASUDA,SHOHEI
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/01/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 06:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 11/19/2021 08:33:09

Diagnostics Check		OK	08:25
Air Blank		0.000	08:26
Control Test		0.082	08:26
Air Blank		0.000	08:27
Subject Sample	# 1	0.162	08:27
Breath Volume		3.378	08:27
Air Blank		0.000	08:28
Air Blank		0.000	08:30
Subject Sample	# 2	0.157	08:31
Breath Volume		1.960	08:31
Air Blank		0.000	08:32
Control Test		0.079	08:32
Air Blank		0.000	08:33
Diagnostics Check		OK	08:33

Data Download Date/Time: 12/03/2021 13:36

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 07:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 11/27/2021 07:53:57

Diagnostics Check		OK	07:45	
Air Blank		0.000	07:45	
Control Test		0.082	07:46	
Air Blank		0.000	07:46	
Subject Sample	# 1		07:49	No Sample Provided
Breath Volume		0.000	07:49	
Air Blank		0.000	07:50	
Air Blank		0.000	07:52	
Subject Sample	# 2		07:52	Subject Test Refused
Breath Volume		0.000	07:52	
Air Blank		0.000	07:53	
Control Test		0.082	07:53	
Air Blank		0.000	07:53	
Diagnostics Check		OK	07:53	

Data Download Date/Time: 12/03/2021 13:36

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WATSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: BARTLETT

DATE/TIME OF TEST: 11/06/2021 17:41:49

Diagnostics Check		OK	17:35
Air Blank		0.000	17:36
Control Test		0.080	17:36
Air Blank		0.000	17:36
Subject Sample	# 1	0.156	17:37
Breath Volume		4.250	17:37
Air Blank		0.000	17:37
Air Blank		0.000	17:39
Subject Sample	# 2	0.159	17:40
Breath Volume		3.738	17:40
Air Blank		0.000	17:40
Control Test		0.080	17:41
Air Blank		0.000	17:41
Diagnostics Check		OK	17:41

Data Download Date/Time: 12/18/2021 09:57

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: IRVING

DATE/TIME OF TEST: 11/07/2021 03:29:29

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:25	
Control Test		0.080	03:25	
Air Blank		0.000	03:26	
Subject Sample	# 1	0.214	03:28	
Breath Volume		2.347	03:28	
Air Blank			03:29	Purge Fail
Air Blank		0.000	03:29	

Data Download Date/Time: 12/18/2021 09:57

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: IRVING

DATE/TIME OF TEST: 11/07/2021 03:37:35

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.079	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.215	03:33
Breath Volume		2.570	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.216	03:36
Breath Volume		1.613	03:36
Air Blank		0.000	03:36
Control Test		0.080	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 12/18/2021 09:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: GIRAO

DATE/TIME OF TEST: 11/12/2021 03:07:49

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.079	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.175	03:03
Breath Volume		3.492	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.195	03:06
Breath Volume		3.652	03:06
Air Blank		0.000	03:06
Control Test		0.079	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 12/18/2021 09:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 11/21/2021 20:46:42

Diagnostics Check		OK	20:40
Air Blank		0.000	20:40
Control Test		0.079	20:41
Air Blank		0.000	20:41
Subject Sample	# 1	0.244	20:42
Breath Volume		2.929	20:42
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.250	20:45
Breath Volume		3.597	20:45
Air Blank		0.000	20:45
Control Test		0.079	20:46
Air Blank		0.000	20:46
Diagnostics Check		OK	20:46

Data Download Date/Time: 12/18/2021 09:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOKES,SHAYNE,R
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: PARTRIDGE,DYLAN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: PALADNINO

DATE/TIME OF TEST: 11/05/2021 13:09:50

Diagnostics Check		OK	13:03
Air Blank		0.000	13:04
Control Test		0.080	13:04
Air Blank		0.000	13:04
Subject Sample	# 1	0.286	13:05
Breath Volume		2.621	13:05
Air Blank		0.000	13:05
Air Blank		0.000	13:07
Subject Sample	# 2	0.269	13:08
Breath Volume		1.636	13:08
Air Blank		0.000	13:08
Control Test		0.079	13:09
Air Blank		0.000	13:09
Diagnostics Check		OK	13:09

Data Download Date/Time: 11/12/2021 13:24

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OJEDA,NELSON,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MENDEZ,JORGE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:16

SUBJECT'S LAST NAME: ROWLISON

DATE/TIME OF TEST: 11/08/2021 16:43:30

Diagnostics Check		OK	16:37
Air Blank		0.000	16:37
Control Test		0.081	16:38
Air Blank		0.000	16:38
Subject Sample	# 1	0.000	16:39
Breath Volume		2.046	16:39
Air Blank		0.000	16:39
Air Blank		0.000	16:41
Subject Sample	# 2	0.000	16:42
Breath Volume		2.238	16:42
Air Blank		0.000	16:42
Control Test		0.079	16:42
Air Blank		0.000	16:43
Diagnostics Check		OK	16:43

Data Download Date/Time: 11/12/2021 13:24

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: SZYMAN

DATE/TIME OF TEST: 11/11/2021 20:31:34

Diagnostics Check		OK	20:25
Air Blank		0.000	20:25
Control Test		0.080	20:26
Air Blank		0.000	20:26
Subject Sample	# 1	0.125	20:27
Breath Volume		1.863	20:27
Air Blank		0.000	20:27
Air Blank		0.000	20:29
Subject Sample	# 2	0.122	20:30
Breath Volume		1.613	20:30
Air Blank		0.000	20:30
Control Test		0.080	20:30
Air Blank		0.000	20:31
Diagnostics Check		OK	20:31

Data Download Date/Time: 11/12/2021 13:24

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PORTILLO

DATE/TIME OF TEST: 11/14/2021 03:24:45

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.242	03:16
Breath Volume		1.441	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.212	03:20
Breath Volume		2.015	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:22
Subject Sample	# 3	0.235	03:23
Breath Volume		1.742	03:23
Air Blank		0.000	03:23
Control Test		0.077	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 12/22/2021 09:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOKES,SHAYNE,R
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: STOKES,SHAYNE
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: SHARPE

DATE/TIME OF TEST: 11/19/2021 07:47:29

Diagnostics Check		OK	07:41
Air Blank		0.000	07:41
Control Test		0.080	07:41
Air Blank		0.000	07:42
Subject Sample	# 1	0.283	07:42
Breath Volume		2.753	07:42
Air Blank		0.000	07:43
Air Blank		0.000	07:45
Subject Sample	# 2	0.284	07:45
Breath Volume		2.324	07:45
Air Blank		0.000	07:46
Control Test		0.079	07:46
Air Blank		0.000	07:47
Diagnostics Check		OK	07:47

Data Download Date/Time: 12/22/2021 09:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYFORD,GREGG,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HAYFORD,GREGG
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: FUNK

DATE/TIME OF TEST: 11/19/2021 19:41:48

Diagnostics Check		OK	19:35
Air Blank		0.000	19:36
Control Test		0.080	19:36
Air Blank		0.000	19:36
Subject Sample	# 1	0.084	19:37
Breath Volume		1.464	19:37
Air Blank		0.000	19:38
Air Blank		0.000	19:39
Subject Sample	# 2	0.080	19:40
Breath Volume		1.320	19:40
Air Blank		0.000	19:40
Control Test		0.080	19:41
Air Blank		0.000	19:41
Diagnostics Check		OK	19:41

Data Download Date/Time: 12/22/2021 09:26

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: CASSEUS

DATE/TIME OF TEST: 11/26/2021 02:50:43

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.080	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.159	02:45
Breath Volume		1.476	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.158	02:49
Breath Volume		1.730	02:49
Air Blank		0.000	02:49
Control Test		0.080	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 12/22/2021 09:26

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HEERDEN,TY,C
BREATH TEST OPERATOR AGENCY: Gulf County SO
ARRESTING OFFICER: PAGE,JESSICA
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 10/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: ENDERLE

DATE/TIME OF TEST: 11/06/2021 20:47:31

Diagnostics Check		OK	20:44	
Air Blank		0.000	20:45	
Control Test		0.079	20:45	
Air Blank		0.000	20:45	
Subject Sample	# 1		20:46	Subject Test Refused
Breath Volume		0.000	20:46	
Air Blank		0.000	20:46	
Control Test		0.079	20:46	
Air Blank		0.000	20:47	
Diagnostics Check		OK	20:47	

Data Download Date/Time: 11/19/2021 12:11

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: BENJAMIN JR

DATE/TIME OF TEST: 11/06/2021 00:25:29

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.080	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.225	00:19
Breath Volume		1.316	00:19
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.230	00:23
Breath Volume		1.675	00:23
Air Blank		0.000	00:24
Control Test		0.080	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 12/09/2021 11:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: SINGKHAOPHET

DATE/TIME OF TEST: 11/07/2021 01:37:49

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.080	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.250	01:33
Breath Volume		1.574	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:35
Subject Sample	# 2	0.259	01:36
Breath Volume		1.550	01:36
Air Blank		0.000	01:36
Control Test		0.079	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 12/09/2021 11:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/07/2021 02:37:38

Diagnostics Check		OK	02:32	
Air Blank		0.000	02:32	
Control Test		0.080	02:33	
Air Blank		0.000	02:33	
Subject Sample	# 1		02:36	Subject Test Refused
Breath Volume		0.000	02:36	
Air Blank		0.000	02:36	
Control Test		0.081	02:37	
Air Blank		0.000	02:37	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 12/09/2021 11:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: VADEN III

DATE/TIME OF TEST: 11/08/2021 02:14:35

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.232	02:10
Breath Volume		1.261	02:10
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.227	02:13
Breath Volume		2.367	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 12/09/2021 11:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: GONZALEZ-MARSHALL

DATE/TIME OF TEST: 11/08/2021 02:33:56

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.081	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.065	02:29
Breath Volume		2.554	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.070	02:32
Breath Volume		1.394	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 12/09/2021 11:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MARTINEZ BERRIOS

DATE/TIME OF TEST: 11/13/2021 04:37:15

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.080	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.109	04:32
Breath Volume		1.277	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:34
Subject Sample	# 2	0.113	04:35
Breath Volume		3.062	04:35
Air Blank		0.000	04:36
Control Test		0.079	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 12/09/2021 11:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: DAVENPORT JR

DATE/TIME OF TEST: 11/13/2021 05:42:00

Diagnostics Check		OK	05:35
Air Blank		0.000	05:35
Control Test		0.079	05:36
Air Blank		0.000	05:36
Subject Sample	# 1	0.119	05:37
Breath Volume		3.878	05:37
Air Blank		0.000	05:38
Air Blank		0.000	05:39
Subject Sample	# 2	0.118	05:40
Breath Volume		2.578	05:40
Air Blank		0.000	05:41
Control Test		0.080	05:41
Air Blank		0.000	05:41
Diagnostics Check		OK	05:42

Data Download Date/Time: 12/09/2021 11:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: CRUZ GALBAN

DATE/TIME OF TEST: 11/14/2021 02:49:41

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.080	02:42
Air Blank		0.000	02:43
Subject Sample	# 1	0.191	02:44
Breath Volume		1.761	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.211	02:48
Breath Volume		2.066	02:48
Air Blank		0.000	02:48
Control Test		0.079	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 12/09/2021 11:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: DELGHAVI

DATE/TIME OF TEST: 11/21/2021 02:11:56

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.081	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.181	02:03
Breath Volume		1.664	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.204	02:07
Breath Volume		2.023	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 3	0.168	02:10
Breath Volume		1.335	02:10
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 12/09/2021 11:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: CAMARGO

DATE/TIME OF TEST: 11/25/2021 04:15:42

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.082	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.124	04:11
Breath Volume		3.054	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.125	04:14
Breath Volume		2.363	04:14
Air Blank		0.000	04:14
Control Test		0.081	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 12/09/2021 11:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VELAZQUEZ,B
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 11/25/2021 05:48:37

Diagnostics Check		OK	05:40
Air Blank		0.000	05:41
Control Test		0.081	05:41
Air Blank		0.000	05:42
Subject Sample	# 1	0.210	05:44
Breath Volume		3.269	05:44
Air Blank		0.000	05:44
Air Blank		0.000	05:46
Subject Sample	# 2	0.213	05:47
Breath Volume		3.234	05:47
Air Blank		0.000	05:47
Control Test		0.081	05:48
Air Blank		0.000	05:48
Diagnostics Check		OK	05:48

Data Download Date/Time: 12/09/2021 11:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JAROSZEWSKI,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: STRAWTER JR

DATE/TIME OF TEST: 11/25/2021 15:44:44

Diagnostics Check		OK	15:36
Air Blank		0.000	15:36
Control Test		0.080	15:36
Air Blank		0.000	15:37
Subject Sample	# 1	0.219	15:39
Breath Volume		1.429	15:39
Air Blank		0.000	15:39
Air Blank		0.000	15:41
Subject Sample	# 2	0.222	15:43
Breath Volume		1.371	15:43
Air Blank		0.000	15:43
Control Test		0.080	15:44
Air Blank		0.000	15:44
Diagnostics Check		OK	15:44

Data Download Date/Time: 12/09/2021 11:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NAYLOR,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: GONSALVES

DATE/TIME OF TEST: 11/28/2021 08:37:41

Diagnostics Check		OK	08:30
Air Blank		0.000	08:30
Control Test		0.081	08:30
Air Blank		0.000	08:31
Subject Sample	# 1	0.146	08:32
Breath Volume		3.894	08:32
Air Blank		0.000	08:33
Air Blank		0.000	08:35
Subject Sample	# 2	0.153	08:36
Breath Volume		3.750	08:36
Air Blank		0.000	08:36
Control Test		0.082	08:37
Air Blank		0.000	08:37
Diagnostics Check		OK	08:37

Data Download Date/Time: 12/09/2021 11:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACABEO,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: NAPPER

DATE/TIME OF TEST: 11/29/2021 04:08:44

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.081	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.208	04:03
Breath Volume		1.804	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.210	04:07
Breath Volume		1.527	04:07
Air Blank		0.000	04:07
Control Test		0.081	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 12/09/2021 11:58

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: YAKUBU

DATE/TIME OF TEST: 11/30/2021 04:56:34

Diagnostics Check		OK	04:48
Air Blank		0.000	04:49
Control Test		0.080	04:49
Air Blank		0.000	04:49
Subject Sample	# 1	0.212	04:51
Breath Volume		1.152	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.209	04:54
Breath Volume		2.789	04:54
Air Blank		0.000	04:55
Control Test		0.081	04:55
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 12/09/2021 11:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,JOHN,S
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: WINDHAM,AUSTIN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 11/14/2021 02:13:10

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:11	
Control Test		0.077	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:12	Sequence Aborted
Breath Volume		0.000	02:12	
Air Blank		0.000	02:13	

Data Download Date/Time: 12/20/2021 12:56

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,JOHN,S
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: DEESON,AARON
ARRESTING OFFICER AGENCY: Taylor County SO
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/18/2021 03:36:36

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.078	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.082	03:32
Breath Volume		4.269	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.083	03:35
Breath Volume		4.097	03:35
Air Blank		0.000	03:35
Control Test		0.077	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 12/20/2021 12:56

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402204068
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: VALDES DE LEON

DATE/TIME OF TEST: 11/07/2021 01:36:32

Diagnostics Check	OK	01:35	
Air Blank		01:36	RFI Detect
Air Blank	0.000	01:36	

Data Download Date/Time: 12/09/2021 11:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402204068
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: VALDES DE LEON

DATE/TIME OF TEST: 11/07/2021 01:42:55

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:39	
Control Test		0.082	01:39	
Air Blank		0.000	01:40	
Subject Sample	# 1		01:42	Subject Test Refused
Breath Volume		0.000	01:42	
Air Blank			01:42	RFI Detect
Air Blank		0.000	01:42	

Data Download Date/Time: 12/09/2021 11:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYATT,JESSICA,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RAMOS,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 11/02/2021 20:35:22

Diagnostics Check		OK	20:32	
Air Blank		0.000	20:32	
Control Test		0.078	20:33	
Air Blank		0.000	20:33	
Subject Sample	# 1		20:34	Subject Test Refused
Breath Volume		0.000	20:34	
Air Blank		0.000	20:34	
Control Test		0.078	20:34	
Air Blank		0.000	20:35	
Diagnostics Check		OK	20:35	

Data Download Date/Time: 11/18/2021 13:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYATT,JESSICA,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DRAKE,N
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 11/03/2021 03:08:44

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:06	
Control Test		0.080	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:07	Subject Test Refused
Breath Volume		0.000	03:07	
Air Blank		0.000	03:07	
Control Test		0.078	03:08	
Air Blank		0.000	03:08	
Diagnostics Check		OK	03:08	

Data Download Date/Time: 11/18/2021 13:08

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VESPER,MITCHELL
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/06/2021 23:05:46

Diagnostics Check		OK	22:50	
Air Blank		0.000	22:51	
Control Test		0.079	22:51	
Air Blank		0.000	22:52	
Subject Sample	# 1	0.085	22:55	Volume Not Met
Breath Volume		0.230	22:55	
Air Blank		0.000	22:56	
Air Blank		0.000	22:57	
Subject Sample	# 2	0.085	22:58	
Breath Volume		1.500	22:58	
Air Blank		0.000	22:59	
Air Blank		0.000	23:01	
Subject Sample	# 3	0.088	23:04	Volume Not Met
Breath Volume		0.359	23:04	
Air Blank		0.000	23:04	
Control Test		0.079	23:05	
Air Blank		0.000	23:05	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 11/18/2021 13:08

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: GASKINS

DATE/TIME OF TEST: 11/07/2021 19:29:13

Diagnostics Check		OK	19:22
Air Blank		0.000	19:23
Control Test		0.080	19:23
Air Blank		0.000	19:24
Subject Sample	# 1	0.000	19:24
Breath Volume		2.242	19:24
Air Blank		0.000	19:25
Air Blank		0.000	19:27
Subject Sample	# 2	0.000	19:27
Breath Volume		2.015	19:27
Air Blank		0.000	19:28
Control Test		0.079	19:28
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

Data Download Date/Time: 11/18/2021 13:08

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 11/11/2021 22:56:46

Diagnostics Check		OK	22:41	
Air Blank		0.000	22:41	
Control Test		0.078	22:41	
Air Blank		0.000	22:42	
Subject Sample	# 1	0.000	22:45	Volume Not Met
Breath Volume		0.015	22:45	
Air Blank		0.000	22:46	
Air Blank		0.000	22:48	
Subject Sample	# 2	0.157	22:50	
Breath Volume		1.656	22:50	
Air Blank		0.000	22:51	
Air Blank		0.000	22:53	
Subject Sample	# 3	0.177	22:55	
Breath Volume		1.222	22:55	
Air Blank		0.000	22:55	
Control Test		0.073	22:56	Control Outside Tolerance
Air Blank		0.000	22:56	

Data Download Date/Time: 11/18/2021 13:08

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 11/11/2021 23:58:06

Diagnostics Check		OK	23:43	
Air Blank		0.000	23:44	
Control Test		0.077	23:44	
Air Blank		0.000	23:45	
Subject Sample	# 1	0.148	23:48	Volume Not Met
Breath Volume		0.816	23:48	
Air Blank		0.000	23:48	
Air Blank		0.000	23:50	
Subject Sample	# 2	0.159	23:52	
Breath Volume		1.496	23:52	
Air Blank		0.000	23:53	
Air Blank		0.000	23:54	
Subject Sample	# 3	0.166	23:56	
Breath Volume		1.812	23:56	
Air Blank		0.000	23:57	
Control Test		0.076	23:57	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 11/18/2021 13:08

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FRANKLIN,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: GEIST

DATE/TIME OF TEST: 11/12/2021 01:45:07

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:37	
Control Test		0.077	01:37	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:41	No Sample Provided
Breath Volume		0.000	01:41	
Air Blank		0.000	01:41	
Air Blank		0.000	01:43	
Subject Sample	# 2		01:43	Subject Test Refused
Breath Volume		0.000	01:43	
Air Blank		0.000	01:44	
Control Test		0.077	01:44	
Air Blank		0.000	01:45	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 11/18/2021 13:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNNE,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BURKHART,Z
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: REEL

DATE/TIME OF TEST: 11/13/2021 17:33:44

Diagnostics Check		OK	17:25
Air Blank		0.000	17:26
Control Test		0.080	17:26
Air Blank		0.000	17:26
Subject Sample	# 1	0.000	17:28
Breath Volume		2.261	17:28
Air Blank		0.000	17:29
Air Blank		0.000	17:31
Subject Sample	# 2	0.000	17:32
Breath Volume		2.460	17:32
Air Blank		0.000	17:32
Control Test		0.078	17:33
Air Blank		0.000	17:33
Diagnostics Check		OK	17:33

Data Download Date/Time: 11/18/2021 13:08

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBEY,BRADLEY,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBEY,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 11/13/2021 19:13:39

Diagnostics Check		OK	19:01	
Air Blank		0.000	19:01	
Control Test		0.079	19:02	
Air Blank		0.000	19:02	
Subject Sample	# 1	0.205	19:03	
Breath Volume		1.238	19:03	
Air Blank		0.000	19:03	
Air Blank		0.000	19:05	
Subject Sample	# 2	0.198	19:08	Volume Not Met
Breath Volume		0.894	19:08	
Air Blank		0.000	19:09	
Air Blank		0.000	19:11	
Subject Sample	# 3	0.207	19:12	
Breath Volume		1.667	19:12	
Air Blank		0.000	19:12	
Control Test		0.075	19:13	
Air Blank		0.000	19:13	
Diagnostics Check		OK	19:13	

Data Download Date/Time: 11/18/2021 13:08

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 11/16/2021 23:29:48

Diagnostics Check		OK	23:21
Air Blank		0.000	23:22
Control Test		0.080	23:22
Air Blank		0.000	23:22
Subject Sample	# 1	0.000	23:24
Breath Volume		1.269	23:24
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.000	23:28
Breath Volume		1.472	23:28
Air Blank		0.000	23:28
Control Test		0.080	23:29
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 11/18/2021 13:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PETERS,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MCGRATH

DATE/TIME OF TEST: 11/17/2021 22:06:51

Diagnostics Check		OK	22:00
Air Blank		0.000	22:01
Control Test		0.079	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.000	22:02
Breath Volume		3.519	22:02
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.000	22:05
Breath Volume		3.234	22:05
Air Blank		0.000	22:05
Control Test		0.080	22:06
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 11/18/2021 13:08

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: GEIGER

DATE/TIME OF TEST: 11/21/2021 17:32:15

Diagnostics Check		OK	17:25
Air Blank		0.000	17:26
Control Test		0.079	17:26
Air Blank		0.000	17:27
Subject Sample	# 1	0.059	17:27
Breath Volume		1.265	17:27
Air Blank		0.000	17:28
Air Blank		0.000	17:30
Subject Sample	# 2	0.051	17:30
Breath Volume		1.527	17:30
Air Blank		0.000	17:31
Control Test		0.079	17:31
Air Blank		0.000	17:32
Diagnostics Check		OK	17:32

Data Download Date/Time: 12/04/2021 11:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FRANKLIN,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/22/2021 03:51:22

Diagnostics Check	OK	03:50	
Air Blank		03:50	RFI Detect
Air Blank	0.000	03:51	

Data Download Date/Time: 12/04/2021 11:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FRANKLIN,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/22/2021 04:00:25

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:55
Subject Sample	# 1	0.050	03:55
Breath Volume		1.675	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.050	03:58
Breath Volume		2.003	03:58
Air Blank		0.000	03:59
Control Test		0.079	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 12/04/2021 11:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 11/22/2021 06:27:08

Diagnostics Check		OK	06:21
Air Blank		0.000	06:21
Control Test		0.079	06:21
Air Blank		0.000	06:22
Subject Sample	# 1	0.147	06:22
Breath Volume		1.496	06:22
Air Blank		0.000	06:23
Air Blank		0.000	06:25
Subject Sample	# 2	0.152	06:25
Breath Volume		1.687	06:25
Air Blank		0.000	06:26
Control Test		0.079	06:26
Air Blank		0.000	06:27
Diagnostics Check		OK	06:27

Data Download Date/Time: 12/04/2021 11:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WHITAKER,G
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 11/28/2021 04:59:55

Diagnostics Check		OK	04:51
Air Blank		0.000	04:52
Control Test		0.079	04:52
Air Blank		0.000	04:53
Subject Sample	# 1	0.172	04:55
Breath Volume		1.824	04:55
Air Blank		0.000	04:56
Air Blank		0.000	04:57
Subject Sample	# 2	0.165	04:58
Breath Volume		1.242	04:58
Air Blank		0.000	04:59
Control Test		0.078	04:59
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 12/04/2021 11:22

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTIGLIONE,THOMAS,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CASTIGLIONE,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: FREDRICKSON

DATE/TIME OF TEST: 11/28/2021 17:51:22

Diagnostics Check		OK	17:44
Air Blank		0.000	17:44
Control Test		0.079	17:45
Air Blank		0.000	17:45
Subject Sample	# 1	0.105	17:46
Breath Volume		1.562	17:46
Air Blank		0.000	17:47
Air Blank		0.000	17:49
Subject Sample	# 2	0.113	17:49
Breath Volume		1.906	17:49
Air Blank		0.000	17:50
Control Test		0.079	17:50
Air Blank		0.000	17:51
Diagnostics Check		OK	17:51

Data Download Date/Time: 12/04/2021 11:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,PEDRO,L
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: QUINN,S
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: MOBERLEY

DATE/TIME OF TEST: 11/01/2021 21:33:42

Diagnostics Check		OK	21:30	
Air Blank		0.000	21:31	
Control Test		0.077	21:31	
Air Blank		0.000	21:32	
Subject Sample	# 1		21:32	Subject Test Refused
Breath Volume		0.000	21:32	
Air Blank		0.000	21:32	
Control Test		0.077	21:33	
Air Blank		0.000	21:33	
Diagnostics Check		OK	21:33	

Data Download Date/Time: 01/20/2022 13:50

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,PEDRO,L
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 11/05/2021 20:06:45

Diagnostics Check		OK	20:00
Air Blank		0.000	20:00
Control Test		0.078	20:01
Air Blank		0.000	20:01
Subject Sample	# 1	0.000	20:02
Breath Volume		2.562	20:02
Air Blank		0.000	20:02
Air Blank		0.000	20:04
Subject Sample	# 2	0.000	20:05
Breath Volume		2.031	20:05
Air Blank		0.000	20:05
Control Test		0.077	20:06
Air Blank		0.000	20:06
Diagnostics Check		OK	20:06

Data Download Date/Time: 01/20/2022 13:50

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: HON

DATE/TIME OF TEST: 11/06/2021 22:08:37

Diagnostics Check		OK	22:05	
Air Blank		0.000	22:06	
Control Test		0.077	22:06	
Air Blank		0.000	22:06	
Subject Sample	# 1		22:07	Subject Test Refused
Breath Volume		0.000	22:07	
Air Blank		0.000	22:07	
Control Test		0.077	22:08	
Air Blank		0.000	22:08	
Diagnostics Check		OK	22:08	

Data Download Date/Time: 01/20/2022 13:50

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,PEDRO,L
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TITUS,OFFICER
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HEFFLEY

DATE/TIME OF TEST: 11/07/2021 23:51:28

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.077	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.228	23:43
Breath Volume		2.375	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.205	23:46
Breath Volume		1.433	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 3	0.201	23:49
Breath Volume		2.671	23:49
Air Blank		0.000	23:50
Control Test		0.076	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 01/20/2022 13:50

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURO,JAMES,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HORTON,H
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: VITALE-MOLLOY

DATE/TIME OF TEST: 11/12/2021 17:52:55

Diagnostics Check		OK	17:49	
Air Blank		0.000	17:50	
Control Test		0.077	17:50	
Air Blank		0.000	17:51	
Subject Sample	# 1		17:51	Subject Test Refused
Breath Volume		0.000	17:51	
Air Blank		0.000	17:52	
Control Test		0.077	17:52	
Air Blank		0.000	17:52	
Diagnostics Check		OK	17:52	

Data Download Date/Time: 01/20/2022 13:50

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE-MOLLOY,GREGORY-DAVID,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HORTON,H
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: VITALE-MOLLOY

DATE/TIME OF TEST: 11/12/2021 18:02:41

Diagnostics Check		OK	17:56
Air Blank		0.000	17:56
Control Test		0.077	17:56
Air Blank		0.000	17:57
Subject Sample	# 1	0.193	17:58
Breath Volume		2.757	17:58
Air Blank		0.000	17:58
Air Blank		0.000	18:00
Subject Sample	# 2	0.177	18:01
Breath Volume		2.144	18:01
Air Blank		0.000	18:01
Control Test		0.076	18:02
Air Blank		0.000	18:02
Diagnostics Check		OK	18:02

Data Download Date/Time: 01/20/2022 13:50

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,JAMES,R
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DOISY

DATE/TIME OF TEST: 11/14/2021 01:31:32

Diagnostics Check		OK	01:24
Air Blank		0.000	01:24
Control Test		0.077	01:24
Air Blank		0.000	01:25
Subject Sample	# 1	0.140	01:26
Breath Volume		1.726	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.140	01:29
Breath Volume		2.246	01:29
Air Blank		0.000	01:30
Control Test		0.076	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 01/20/2022 13:50

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,JAMES,R
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 11/17/2021 23:49:25

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.078	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.000	23:44
Breath Volume		3.144	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.000	23:48
Breath Volume		2.734	23:48
Air Blank		0.000	23:48
Control Test		0.077	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 01/20/2022 13:50

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,PEDRO,L
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: LAUTHER

DATE/TIME OF TEST: 11/24/2021 23:29:28

Diagnostics Check		OK	23:22
Air Blank		0.000	23:22
Control Test		0.077	23:22
Air Blank		0.000	23:23
Subject Sample	# 1	0.192	23:24
Breath Volume		1.300	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.186	23:27
Breath Volume		2.289	23:27
Air Blank		0.000	23:28
Control Test		0.076	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 01/20/2022 13:50

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWLES,ANALISE,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: BOWLES

DATE/TIME OF TEST: 11/26/2021 17:07:34

Diagnostics Check		OK	17:00
Air Blank		0.000	17:01
Control Test		0.076	17:01
Air Blank		0.000	17:02
Subject Sample	# 1	0.204	17:02
Breath Volume		3.035	17:02
Air Blank		0.000	17:03
Air Blank		0.000	17:04
Subject Sample	# 2	0.211	17:05
Breath Volume		2.160	17:05
Air Blank		0.000	17:06
Control Test		0.076	17:06
Air Blank		0.000	17:07
Diagnostics Check		OK	17:07

Data Download Date/Time: 01/20/2022 13:50

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,JAMES,R
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: VALENCIA,E
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ARRINGTON

DATE/TIME OF TEST: 11/27/2021 02:35:36

Diagnostics Check		OK	02:32	
Air Blank		0.000	02:33	
Control Test		0.077	02:33	
Air Blank		0.000	02:33	
Subject Sample	# 1		02:34	Subject Test Refused
Breath Volume		0.000	02:34	
Air Blank		0.000	02:34	
Control Test		0.076	02:35	
Air Blank		0.000	02:35	
Diagnostics Check		OK	02:35	

Data Download Date/Time: 01/20/2022 13:50

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBEY,BRADLEY,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBEY,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 11/27/2021 17:02:21

Diagnostics Check		OK	16:50	
Air Blank		0.000	16:51	
Control Test		0.077	16:51	
Air Blank		0.000	16:52	
Subject Sample	# 1	0.149	16:55	Volume Not Met
Breath Volume		0.000	16:55	
Air Blank		0.000	16:55	
Air Blank		0.000	16:57	
Subject Sample	# 2	0.145	17:00	Volume Not Met
Breath Volume		0.898	17:00	
Air Blank		0.000	17:01	
Control Test		0.076	17:01	
Air Blank		0.000	17:02	
Diagnostics Check		OK	17:02	

Data Download Date/Time: 01/20/2022 13:50

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,JAMES,R
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERRER,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 11/28/2021 04:35:08

Diagnostics Check		OK	04:32	
Air Blank		0.000	04:32	
Control Test		0.077	04:32	
Air Blank		0.000	04:33	
Subject Sample	# 1		04:33	Subject Test Refused
Breath Volume		0.000	04:33	
Air Blank		0.000	04:34	
Control Test		0.077	04:34	
Air Blank		0.000	04:35	
Diagnostics Check		OK	04:35	

Data Download Date/Time: 01/20/2022 13:50

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,JAMES,R
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: PENNEY,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 11/28/2021 19:23:08

Diagnostics Check		OK	19:20	
Air Blank		0.000	19:20	
Control Test		0.077	19:20	
Air Blank		0.000	19:21	
Subject Sample	# 1		19:21	Subject Test Refused
Breath Volume		0.000	19:21	
Air Blank		0.000	19:22	
Control Test		0.077	19:22	
Air Blank		0.000	19:23	
Diagnostics Check		OK	19:23	

Data Download Date/Time: 01/20/2022 13:50

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,DAVID,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: BURTZLAFF,SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 11/02/2021 03:24:43

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.130	03:20
Breath Volume		2.027	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.125	03:23
Breath Volume		1.226	03:23
Air Blank		0.000	03:23
Control Test		0.081	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 11/30/2021 19:25

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: CARTER,B
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: SAJDAK

DATE/TIME OF TEST: 11/02/2021 16:26:55

Diagnostics Check		OK	16:20
Air Blank		0.000	16:21
Control Test		0.080	16:21
Air Blank		0.000	16:22
Subject Sample	# 1	0.185	16:22
Breath Volume		2.457	16:22
Air Blank		0.000	16:23
Air Blank		0.000	16:24
Subject Sample	# 2	0.196	16:25
Breath Volume		2.074	16:25
Air Blank		0.000	16:26
Control Test		0.082	16:26
Air Blank		0.000	16:26
Diagnostics Check		OK	16:26

Data Download Date/Time: 11/30/2021 19:25

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,JOHN,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: CALHOUN,JOHN
ARRESTING OFFICER AGENCY: Astatula PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: BODIN

DATE/TIME OF TEST: 11/04/2021 01:23:12

Diagnostics Check		OK	01:14	
Air Blank		0.000	01:15	
Control Test		0.080	01:15	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:19	No Sample Provided
Breath Volume		0.000	01:19	
Air Blank		0.000	01:19	
Air Blank		0.000	01:21	
Subject Sample	# 2		01:21	Subject Test Refused
Breath Volume		0.000	01:21	
Air Blank		0.000	01:22	
Control Test		0.081	01:22	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 11/30/2021 19:25

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: SCHMIDL

DATE/TIME OF TEST: 11/04/2021 23:44:58

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.081	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.250	23:40
Breath Volume		3.667	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:42
Subject Sample	# 2	0.240	23:43
Breath Volume		2.550	23:43
Air Blank		0.000	23:44
Control Test		0.081	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 11/30/2021 19:25

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ DIAZ,STEVEN,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: BAEZ DIAZ,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 11/05/2021 02:45:40

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:40	
Control Test		0.081	02:40	
Air Blank		0.000	02:40	
Subject Sample	# 1	0.137	02:44	Volume Not Met
Breath Volume		0.253	02:44	
Air Blank			02:44	Purge Fail
Air Blank			02:45	Purge Fail

Data Download Date/Time: 11/30/2021 19:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ DIAZ,STEVEN,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: BAEZ DIAZ,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 11/05/2021 02:53:37

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.080	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.134	02:49
Breath Volume		2.171	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:51
Subject Sample	# 2	0.128	02:52
Breath Volume		2.433	02:52
Air Blank		0.000	02:52
Control Test		0.077	02:53
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 11/30/2021 19:25

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BURTZLAFF,SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: LANHAM

DATE/TIME OF TEST: 11/07/2021 02:25:25

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:23	
Control Test		0.081	02:23	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:24	Subject Test Refused
Breath Volume		0.000	02:24	
Air Blank		0.000	02:24	
Control Test		0.080	02:24	
Air Blank		0.000	02:25	
Diagnostics Check		OK	02:25	

Data Download Date/Time: 11/30/2021 19:25

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: COE,MELODY
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 11/11/2021 00:48:06

Diagnostics Check		OK	00:40
Air Blank		0.000	00:41
Control Test		0.081	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.077	00:42
Breath Volume		1.136	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.077	00:46
Breath Volume		1.332	00:46
Air Blank		0.000	00:47
Control Test		0.081	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 11/30/2021 19:25

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: ABREU,KENIA
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: MICHELENA

DATE/TIME OF TEST: 11/11/2021 15:06:37

Diagnostics Check		OK	15:00
Air Blank		0.000	15:00
Control Test		0.081	15:01
Air Blank		0.000	15:01
Subject Sample	# 1	0.000	15:02
Breath Volume		2.496	15:02
Air Blank		0.000	15:02
Air Blank		0.000	15:04
Subject Sample	# 2	0.000	15:05
Breath Volume		2.210	15:05
Air Blank		0.000	15:05
Control Test		0.081	15:06
Air Blank		0.000	15:06
Diagnostics Check		OK	15:06

Data Download Date/Time: 11/30/2021 19:25

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 11/13/2021 04:16:54

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.081	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.178	04:12
Breath Volume		3.617	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.172	04:15
Breath Volume		3.464	04:15
Air Blank		0.000	04:15
Control Test		0.082	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 11/30/2021 19:25

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRANSUE,MICHAEL,C
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: TRANSUE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:28

SUBJECT'S LAST NAME: ALBIZU

DATE/TIME OF TEST: 11/13/2021 09:02:03

Diagnostics Check		OK	08:59	
Air Blank		0.000	08:59	
Control Test		0.081	08:59	
Air Blank		0.000	09:00	
Subject Sample	# 1		09:00	Subject Test Refused
Breath Volume		0.000	09:00	
Air Blank		0.000	09:01	
Control Test		0.081	09:01	
Air Blank		0.000	09:01	
Diagnostics Check		OK	09:02	

Data Download Date/Time: 11/30/2021 19:25

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,R,S
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: PENA,R
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: SCHLOSSER

DATE/TIME OF TEST: 11/14/2021 01:15:35

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.080	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.204	01:10
Breath Volume		4.734	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.188	01:14
Breath Volume		3.777	01:14
Air Blank		0.000	01:14
Control Test		0.081	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 11/30/2021 19:26

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORFEL,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SULLIVAN,Z
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: KIDD

DATE/TIME OF TEST: 11/17/2021 00:51:51

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:48	
Control Test		0.080	00:49	
Air Blank		0.000	00:49	
Subject Sample	# 1		00:50	Subject Test Refused
Breath Volume		0.000	00:50	
Air Blank		0.000	00:50	
Control Test		0.081	00:51	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 11/30/2021 19:26

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 11/21/2021 02:38:09

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.160	02:33
Breath Volume		1.871	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.154	02:36
Breath Volume		1.359	02:36
Air Blank		0.000	02:37
Control Test		0.080	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 11/30/2021 19:26

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PERALTA,BRYAN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: BEISEL

DATE/TIME OF TEST: 11/21/2021 03:11:32

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.157	03:06
Breath Volume		2.222	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.151	03:10
Breath Volume		1.597	03:10
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 11/30/2021 19:26

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BURTZLAFF,SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 11/21/2021 03:40:48

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.080	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.199	03:36
Breath Volume		4.687	03:36
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.198	03:39
Breath Volume		4.609	03:39
Air Blank		0.000	03:39
Control Test		0.080	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 11/30/2021 19:26

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINS,JOSHUA,T
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: DU,WESLEY
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 11/22/2021 01:50:22

Diagnostics Check		OK	01:47	
Air Blank		0.000	01:47	
Control Test		0.080	01:48	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:49	Subject Test Refused
Breath Volume		0.000	01:49	
Air Blank		0.000	01:49	
Control Test		0.080	01:49	
Air Blank		0.000	01:50	
Diagnostics Check		OK	01:50	

Data Download Date/Time: 11/30/2021 19:26

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,DAVID,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN,DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SALEM

DATE/TIME OF TEST: 11/25/2021 22:13:53

Diagnostics Check		OK	22:05
Air Blank		0.000	22:06
Control Test		0.081	22:06
Air Blank		0.000	22:07
Subject Sample	# 1	0.050	22:08
Breath Volume		1.707	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:11
Subject Sample	# 2	0.045	22:12
Breath Volume		1.664	22:12
Air Blank		0.000	22:12
Control Test		0.080	22:13
Air Blank		0.000	22:13
Diagnostics Check		OK	22:13

Data Download Date/Time: 11/30/2021 19:26

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: HEFFLER,GAVIN
ARRESTING OFFICER AGENCY: Umatilla PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: ZEHNER

DATE/TIME OF TEST: 11/26/2021 03:02:41

Diagnostics Check		OK	02:59	
Air Blank		0.000	03:00	
Control Test		0.080	03:00	
Air Blank		0.000	03:01	
Subject Sample	# 1		03:01	Subject Test Refused
Breath Volume		0.000	03:01	
Air Blank		0.000	03:01	
Control Test		0.080	03:02	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 11/30/2021 19:27

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/26/2021 22:01:00

Diagnostics Check		OK	21:54
Air Blank		0.000	21:54
Control Test		0.081	21:54
Air Blank		0.000	21:55
Subject Sample	# 1	0.149	21:56
Breath Volume		1.664	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:58
Subject Sample	# 2	0.155	21:59
Breath Volume		1.457	21:59
Air Blank		0.000	22:00
Control Test		0.081	22:00
Air Blank		0.000	22:00
Diagnostics Check		OK	22:01

Data Download Date/Time: 11/30/2021 19:27

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: PEREZ,CALEB
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: SPITULSKI

DATE/TIME OF TEST: 11/26/2021 22:39:43

Diagnostics Check		OK	22:31
Air Blank		0.000	22:31
Control Test		0.080	22:32
Air Blank		0.000	22:32
Subject Sample	# 1	0.294	22:34
Breath Volume		1.777	22:34
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.291	22:38
Breath Volume		2.410	22:38
Air Blank		0.000	22:38
Control Test		0.080	22:39
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 11/30/2021 19:27

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: CAMACHO,NICOLE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 11/26/2021 23:34:19

Diagnostics Check		OK	23:29	
Air Blank		0.000	23:29	
Control Test		0.081	23:30	
Air Blank		0.000	23:30	
Subject Sample	# 1		23:32	Subject Test Refused
Breath Volume		0.000	23:32	
Air Blank		0.000	23:33	
Control Test		0.080	23:33	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 11/30/2021 19:27

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BURTZLAFF,SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: VALE

DATE/TIME OF TEST: 11/30/2021 02:26:21

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:23	
Control Test		0.080	02:23	
Air Blank		0.000	02:24	
Subject Sample	# 1		02:24	Subject Test Refused
Breath Volume		0.000	02:24	
Air Blank		0.000	02:25	
Control Test		0.081	02:25	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 11/30/2021 19:27

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 11/13/2021 03:08:25

Diagnostics Check	OK	03:07	
Air Blank		03:07	RFI Detect
Air Blank	0.000	03:08	

Data Download Date/Time: 11/30/2021 19:15

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 11/13/2021 03:13:26

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:10	
Control Test		0.078	03:10	
Air Blank		0.000	03:11	
Subject Sample	# 1	0.195	03:11	
Breath Volume		2.957	03:11	
Air Blank			03:11	RFI Detect
Air Blank		0.000	03:12	
Control Test		0.078	03:12	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 11/30/2021 19:15

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 11/13/2021 03:15:18

Diagnostics Check	Fail	03:14	Internal Printer Test
Air Blank		03:15	RFI Detect

Data Download Date/Time: 11/30/2021 19:15

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MCCOLLUM

DATE/TIME OF TEST: 11/15/2021 02:53:02

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.135	02:48
Breath Volume		1.691	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:50
Subject Sample	# 2	0.127	02:51
Breath Volume		1.140	02:51
Air Blank		0.000	02:52
Control Test		0.078	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:53

Data Download Date/Time: 11/30/2021 19:16

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: BRULAND,E
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 11/08/2021 22:09:36

Diagnostics Check		OK	22:03
Air Blank		0.000	22:03
Control Test		0.080	22:04
Air Blank		0.000	22:04
Subject Sample	# 1	0.000	22:05
Breath Volume		3.109	22:05
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.000	22:08
Breath Volume		3.199	22:08
Air Blank		0.000	22:08
Control Test		0.081	22:09
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

Data Download Date/Time: 11/29/2021 01:31

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: NAVARRO,D
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 11/13/2021 04:00:21

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:55
Subject Sample	# 1	0.201	03:55
Breath Volume		2.214	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.198	03:58
Breath Volume		1.718	03:58
Air Blank		0.000	03:59
Control Test		0.080	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 11/29/2021 01:31

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLCHARRAN,D,C
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: HOBBS,K
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: RAGIN

DATE/TIME OF TEST: 11/17/2021 20:33:30

Diagnostics Check		OK	20:20	
Air Blank		0.000	20:21	
Control Test		0.080	20:21	
Air Blank		0.000	20:22	
Subject Sample	# 1	0.344	20:25	Volume Not Met
Breath Volume		0.031	20:25	
Air Blank		0.000	20:26	
Air Blank		0.000	20:27	
Subject Sample	# 2	0.390	20:28	
Breath Volume		1.730	20:28	
Air Blank		0.000	20:29	
Air Blank		0.000	20:31	
Subject Sample	# 3	0.397	20:31	
Breath Volume		1.886	20:31	
Air Blank		0.000	20:32	
Control Test		0.078	20:32	
Air Blank			20:33	RFI Detect
Air Blank		0.000	20:33	

Data Download Date/Time: 11/29/2021 01:31

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,DAVID,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN,DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/19/2021 01:48:22

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.080	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.227	01:41
Breath Volume		1.308	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.225	01:46
Breath Volume		1.308	01:46
Air Blank		0.000	01:47
Control Test		0.077	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 11/29/2021 01:31

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEUSER,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: HARPER,SUSAN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: RIVERA PEREZ

DATE/TIME OF TEST: 11/22/2021 23:32:41

Diagnostics Check		OK	23:25
Air Blank		0.000	23:25
Control Test		0.081	23:26
Air Blank		0.000	23:26
Subject Sample	# 1	0.152	23:27
Breath Volume		1.726	23:27
Air Blank		0.000	23:27
Air Blank		0.000	23:29
Subject Sample	# 2	0.155	23:31
Breath Volume		1.992	23:31
Air Blank		0.000	23:31
Control Test		0.079	23:32
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 11/29/2021 01:31

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,MATTHEW,L
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LEVENAR,T
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 11/06/2021 21:17:07

Diagnostics Check		OK	21:11
Air Blank		0.000	21:11
Control Test		0.081	21:11
Air Blank		0.000	21:12
Subject Sample	# 1	0.000	21:12
Breath Volume		3.691	21:12
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.000	21:15
Breath Volume		3.722	21:15
Air Blank		0.000	21:16
Control Test		0.080	21:16
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 12/17/2021 09:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: LAURENT

DATE/TIME OF TEST: 11/08/2021 00:18:58

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.080	00:13
Air Blank		0.000	00:13
Subject Sample	# 1	0.145	00:14
Breath Volume		1.750	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.149	00:17
Breath Volume		1.335	00:17
Air Blank		0.000	00:18
Control Test		0.079	00:18
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 12/17/2021 09:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHAM,MATTHEW,H
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SPUGNARDI,TIMOTHY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 11/09/2021 03:22:04

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.080	03:10	
Air Blank		0.000	03:11	
Subject Sample	# 1		03:14	No Sample Provided
Breath Volume		0.000	03:14	
Air Blank		0.000	03:15	
Air Blank		0.000	03:17	
Subject Sample	# 2	0.165	03:17	
Breath Volume		2.625	03:17	
Air Blank		0.000	03:18	
Air Blank		0.000	03:19	
Subject Sample	# 3	0.165	03:20	
Breath Volume		4.636	03:20	
Air Blank		0.000	03:21	
Control Test		0.078	03:21	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 12/17/2021 09:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,DAVID,E
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: JACOBSON,DAVID
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: AMEER

DATE/TIME OF TEST: 11/14/2021 02:24:37

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.080	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.121	02:20
Breath Volume		1.382	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.127	02:23
Breath Volume		1.800	02:23
Air Blank		0.000	02:23
Control Test		0.079	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 12/17/2021 09:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ZERBE

DATE/TIME OF TEST: 11/20/2021 02:37:20

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.179	02:32
Breath Volume		1.574	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:34
Subject Sample	# 2	0.167	02:35
Breath Volume		1.742	02:35
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 12/17/2021 09:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,JOSE,G
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: ROBLERO DE LEON

DATE/TIME OF TEST: 11/27/2021 21:41:31

Diagnostics Check		OK	21:35
Air Blank		0.000	21:35
Control Test		0.080	21:36
Air Blank		0.000	21:36
Subject Sample	# 1	0.252	21:37
Breath Volume		2.886	21:37
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.244	21:39
Breath Volume		2.648	21:39
Air Blank		0.000	21:40
Control Test		0.079	21:40
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 12/17/2021 09:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,T
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 11/30/2021 18:40:53

Diagnostics Check		OK	18:34
Air Blank		0.000	18:35
Control Test		0.081	18:35
Air Blank		0.000	18:35
Subject Sample	# 1	0.000	18:36
Breath Volume		2.976	18:36
Air Blank		0.000	18:37
Air Blank		0.000	18:39
Subject Sample	# 2	0.000	18:39
Breath Volume		2.503	18:39
Air Blank		0.000	18:39
Control Test		0.081	18:40
Air Blank		0.000	18:40
Diagnostics Check		OK	18:40

Data Download Date/Time: 12/17/2021 09:23

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: H616823993470

DATE/TIME OF TEST: 11/25/2021 03:51:07

Diagnostics Check		OK	03:47	
Air Blank		0.000	03:48	
Control Test		0.079	03:48	
Air Blank		0.000	03:48	
Subject Sample	# 1		03:49	RFI Detect
Breath Volume		0.023	03:49	
Air Blank		0.000	03:50	
Control Test		0.080	03:50	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 12/03/2021 12:21

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 11/25/2021 04:09:39

Diagnostics Check		OK	03:54	
Air Blank		0.000	03:54	
Control Test		0.078	03:55	
Air Blank		0.000	03:55	
Subject Sample	# 1	0.209	03:58	
Breath Volume		1.742	03:58	
Air Blank		0.000	03:59	
Air Blank		0.000	04:01	
Subject Sample	# 2	0.205	04:04	Volume Not Met
Breath Volume		0.355	04:04	
Air Blank		0.000	04:05	
Air Blank		0.000	04:07	
Subject Sample	# 3	0.208	04:08	
Breath Volume		1.359	04:08	
Air Blank		0.000	04:08	
Control Test		0.078	04:09	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 12/03/2021 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KLAUSNER,JAMIE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: LAVERGNE

DATE/TIME OF TEST: 11/02/2021 04:42:52

Diagnostics Check		OK	04:36
Air Blank		0.000	04:37
Control Test		0.080	04:37
Air Blank		0.000	04:38
Subject Sample	# 1	0.143	04:38
Breath Volume		1.980	04:38
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.141	04:41
Breath Volume		1.785	04:41
Air Blank		0.000	04:41
Control Test		0.080	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 11/12/2021 18:57

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,DAVIS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RIOS,DAVIS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 11/05/2021 23:52:09

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:47
Subject Sample	# 1	0.000	23:47
Breath Volume		3.820	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.000	23:50
Breath Volume		3.839	23:50
Air Blank		0.000	23:51
Control Test		0.080	23:51
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 11/12/2021 18:57

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,DAVIS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CHRABOT,NICHOLAS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: KOLB

DATE/TIME OF TEST: 11/06/2021 00:14:03

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.080	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.000	00:09
Breath Volume		1.683	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.000	00:12
Breath Volume		1.601	00:12
Air Blank		0.000	00:13
Control Test		0.080	00:13
Air Blank		0.000	00:13
Diagnostics Check		OK	00:14

Data Download Date/Time: 11/12/2021 18:57

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HARRISON,TRAVIS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: KOCHEL

DATE/TIME OF TEST: 11/08/2021 03:30:33

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.081	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.103	03:26
Breath Volume		2.261	03:26
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.094	03:29
Breath Volume		1.132	03:29
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 11/12/2021 18:57

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: TOURLAKES,ALEXANDER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: JEUNEGENS

DATE/TIME OF TEST: 11/10/2021 00:49:08

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.080	00:35	
Air Blank		0.000	00:36	
Subject Sample	# 1	0.166	00:39	Volume Not Met
Breath Volume		0.714	00:39	
Air Blank		0.000	00:40	
Air Blank		0.000	00:42	
Subject Sample	# 2	0.151	00:44	
Breath Volume		1.214	00:44	
Air Blank		0.000	00:44	
Air Blank		0.000	00:46	
Subject Sample	# 3	0.153	00:47	
Breath Volume		1.375	00:47	
Air Blank		0.000	00:48	
Control Test		0.079	00:48	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 11/12/2021 18:57

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: HOUGHTON

DATE/TIME OF TEST: 11/10/2021 03:17:40

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.176	03:12
Breath Volume		1.984	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.162	03:16
Breath Volume		1.167	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 11/12/2021 18:57

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: REID,JILLIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 11/11/2021 19:42:35

Diagnostics Check		OK	19:36
Air Blank		0.000	19:37
Control Test		0.080	19:37
Air Blank		0.000	19:38
Subject Sample	# 1	0.000	19:38
Breath Volume		2.367	19:38
Air Blank		0.000	19:38
Air Blank		0.000	19:40
Subject Sample	# 2	0.000	19:41
Breath Volume		2.269	19:41
Air Blank		0.000	19:41
Control Test		0.080	19:42
Air Blank		0.000	19:42
Diagnostics Check		OK	19:42

Data Download Date/Time: 11/12/2021 18:57

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: NOVACK,BRADLEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: NOTTENSTEINER

DATE/TIME OF TEST: 11/12/2021 23:19:00

Diagnostics Check		OK	23:12
Air Blank		0.000	23:12
Control Test		0.080	23:12
Air Blank		0.000	23:13
Subject Sample	# 1	0.098	23:14
Breath Volume		2.785	23:14
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.093	23:17
Breath Volume		2.632	23:17
Air Blank		0.000	23:18
Control Test		0.080	23:18
Air Blank		0.000	23:18
Diagnostics Check		OK	23:19

Data Download Date/Time: 12/13/2021 12:19

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: TOURLAKES,ALEXANDER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: SCHULTE

DATE/TIME OF TEST: 11/13/2021 00:18:11

Diagnostics Check		OK	00:09
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.160	00:12
Breath Volume		1.421	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.163	00:16
Breath Volume		2.402	00:16
Air Blank		0.000	00:17
Control Test		0.080	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 12/13/2021 12:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINEZ,FRANCISCO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: FINNEY

DATE/TIME OF TEST: 11/13/2021 01:34:35

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.080	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.231	01:30
Breath Volume		1.316	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.220	01:33
Breath Volume		1.156	01:33
Air Blank		0.000	01:33
Control Test		0.079	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 12/13/2021 12:19

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HARRISON,WILLIE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 11/14/2021 02:11:39

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.173	02:07
Breath Volume		2.410	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.177	02:10
Breath Volume		3.312	02:10
Air Blank		0.000	02:10
Control Test		0.080	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 12/13/2021 12:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: TOURLAKES,ALEXANDER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: BURGOS

DATE/TIME OF TEST: 11/14/2021 21:21:17

Diagnostics Check		OK	21:14
Air Blank		0.000	21:15
Control Test		0.080	21:15
Air Blank		0.000	21:16
Subject Sample	# 1	0.187	21:16
Breath Volume		1.453	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.175	21:19
Breath Volume		1.300	21:19
Air Blank		0.000	21:20
Control Test		0.079	21:20
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 12/13/2021 12:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CAPLE,JENNIFER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 11/16/2021 04:12:11

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:06
Subject Sample	# 1	0.163	04:07
Breath Volume		3.968	04:07
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.155	04:10
Breath Volume		3.828	04:10
Air Blank		0.000	04:11
Control Test		0.080	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 12/13/2021 12:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWTON,JENNA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GUEVARA,JACOB
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MALLOY

DATE/TIME OF TEST: 11/16/2021 23:49:12

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.080	23:43
Air Blank		0.000	23:43
Subject Sample	# 1	0.140	23:44
Breath Volume		2.878	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:46
Subject Sample	# 2	0.126	23:47
Breath Volume		2.609	23:47
Air Blank		0.000	23:48
Control Test		0.078	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 12/13/2021 12:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HARRISON,TRAVIS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: PURCELL

DATE/TIME OF TEST: 11/17/2021 03:07:48

Diagnostics Check		OK	02:59	
Air Blank		0.000	02:59	
Control Test		0.080	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1	0.000	03:03	Volume Not Met
Breath Volume		0.781	03:03	
Air Blank		0.000	03:03	
Air Blank		0.000	03:05	
Subject Sample	# 2		03:06	Subject Test Refused
Breath Volume		0.000	03:06	
Air Blank		0.000	03:06	
Control Test		0.080	03:07	
Air Blank		0.000	03:07	
Diagnostics Check		OK	03:07	

Data Download Date/Time: 12/13/2021 12:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HARRISON,TRAVIS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: PURCELL

DATE/TIME OF TEST: 11/17/2021 03:42:26

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.081	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.000	03:38
Breath Volume		3.160	03:38
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.000	03:41
Breath Volume		2.484	03:41
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 12/13/2021 12:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 11/18/2021 19:40:58

Diagnostics Check		OK	19:34
Air Blank		0.000	19:34
Control Test		0.080	19:35
Air Blank		0.000	19:35
Subject Sample	# 1	0.000	19:36
Breath Volume		2.171	19:36
Air Blank		0.000	19:36
Air Blank		0.000	19:38
Subject Sample	# 2	0.000	19:39
Breath Volume		2.230	19:39
Air Blank		0.000	19:40
Control Test		0.080	19:40
Air Blank		0.000	19:40
Diagnostics Check		OK	19:40

Data Download Date/Time: 12/13/2021 12:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,DAVIS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RIOS,DAVIS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SYFRETT

DATE/TIME OF TEST: 11/21/2021 04:02:14

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.013	03:57
Breath Volume		2.265	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.013	04:00
Breath Volume		1.976	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 12/13/2021 12:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LUBBERS,KAYLA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: ESPANA-ZAPATA

DATE/TIME OF TEST: 11/21/2021 18:41:54

Diagnostics Check		OK	18:33
Air Blank		0.000	18:34
Control Test		0.080	18:34
Air Blank		0.000	18:34
Subject Sample	# 1	0.133	18:35
Breath Volume		1.429	18:35
Air Blank		0.000	18:35
Air Blank		0.000	18:37
Subject Sample	# 2	0.142	18:40
Breath Volume		1.269	18:40
Air Blank		0.000	18:41
Control Test		0.080	18:41
Air Blank		0.000	18:41
Diagnostics Check		OK	18:41

Data Download Date/Time: 12/13/2021 12:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP,RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: TORRES CABRERA

DATE/TIME OF TEST: 11/21/2021 21:25:00

Diagnostics Check		OK	21:15
Air Blank		0.000	21:16
Control Test		0.080	21:16
Air Blank		0.000	21:17
Subject Sample	# 1	0.220	21:20
Breath Volume		1.437	21:20
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.219	21:23
Breath Volume		1.839	21:23
Air Blank		0.000	21:24
Control Test		0.079	21:24
Air Blank		0.000	21:24
Diagnostics Check		OK	21:25

Data Download Date/Time: 12/13/2021 12:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MOHLENCAMP,HOLLY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: FAUCETT

DATE/TIME OF TEST: 11/22/2021 01:00:05

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.079	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.000	00:55
Breath Volume		3.363	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.000	00:58
Breath Volume		3.253	00:58
Air Blank		0.000	00:59
Control Test		0.079	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 12/13/2021 12:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CARROLL,JILLIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: POLLOCK

DATE/TIME OF TEST: 11/22/2021 02:30:44

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.080	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.188	02:26
Breath Volume		1.386	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.178	02:29
Breath Volume		1.210	02:29
Air Blank		0.000	02:29
Control Test		0.080	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 12/13/2021 12:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CHRABOT,NICOLAS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: POLLOCK

DATE/TIME OF TEST: 11/22/2021 02:41:45

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.079	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.197	02:37
Breath Volume		1.519	02:37
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.202	02:40
Breath Volume		2.152	02:40
Air Blank		0.000	02:40
Control Test		0.079	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 12/13/2021 12:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO,CARLOS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: GALZAGORRY RODRIGUEZ

DATE/TIME OF TEST: 11/24/2021 01:23:40

Diagnostics Check		OK	01:16
Air Blank		0.000	01:16
Control Test		0.081	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.235	01:19
Breath Volume		3.718	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.230	01:22
Breath Volume		2.476	01:22
Air Blank		0.000	01:22
Control Test		0.079	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 12/13/2021 12:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATO,MARK,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SIMONDS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 11/24/2021 09:33:05

Diagnostics Check		OK	09:26
Air Blank		0.000	09:27
Control Test		0.081	09:27
Air Blank		0.000	09:28
Subject Sample	# 1	0.000	09:28
Breath Volume		1.753	09:28
Air Blank		0.000	09:29
Air Blank		0.000	09:31
Subject Sample	# 2	0.000	09:31
Breath Volume		1.851	09:31
Air Blank		0.000	09:32
Control Test		0.080	09:32
Air Blank		0.000	09:33
Diagnostics Check		OK	09:33

Data Download Date/Time: 12/13/2021 12:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUSER,SGT,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BRIDGES,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: BRUNER

DATE/TIME OF TEST: 11/26/2021 10:16:59

Diagnostics Check		OK	10:10
Air Blank		0.000	10:10
Control Test		0.080	10:11
Air Blank		0.000	10:11
Subject Sample	# 1	0.000	10:12
Breath Volume		2.843	10:12
Air Blank		0.000	10:12
Air Blank		0.000	10:14
Subject Sample	# 2	0.000	10:15
Breath Volume		2.824	10:15
Air Blank		0.000	10:16
Control Test		0.080	10:16
Air Blank		0.000	10:16
Diagnostics Check		OK	10:16

Data Download Date/Time: 12/13/2021 12:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MADURO,VICTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: BUSBEE

DATE/TIME OF TEST: 11/27/2021 04:19:06

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.080	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.177	04:14
Breath Volume		1.992	04:14
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.186	04:17
Breath Volume		2.160	04:17
Air Blank		0.000	04:18
Control Test		0.079	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 12/13/2021 12:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTRELL,ROBERT,V
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CANNOVA,CASSANDRA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: RODIGUEZ

DATE/TIME OF TEST: 11/27/2021 07:03:09

Diagnostics Check		OK	06:51	
Air Blank		0.000	06:52	
Control Test		0.080	06:52	
Air Blank		0.000	06:52	
Subject Sample	# 1	0.107	06:56	Volume Not Met
Breath Volume		1.089	06:56	
Air Blank		0.000	06:56	
Air Blank		0.000	06:58	
Subject Sample	# 2	0.095	07:01	Volume Not Met
Breath Volume		1.070	07:01	
Air Blank		0.000	07:02	
Control Test		0.078	07:02	
Air Blank		0.000	07:03	
Diagnostics Check		OK	07:03	

Data Download Date/Time: 12/13/2021 12:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESNUT,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SONG,GUANG
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: HULLETT

DATE/TIME OF TEST: 11/28/2021 20:53:23

Diagnostics Check		OK	20:46
Air Blank		0.000	20:46
Control Test		0.080	20:46
Air Blank		0.000	20:47
Subject Sample	# 1	0.258	20:48
Breath Volume		1.742	20:48
Air Blank		0.000	20:48
Air Blank		0.000	20:50
Subject Sample	# 2	0.240	20:51
Breath Volume		1.226	20:51
Air Blank		0.000	20:52
Control Test		0.079	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 12/13/2021 12:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESNUT,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CHESNUT,JONATHAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: TROMBLEY

DATE/TIME OF TEST: 11/30/2021 21:26:03

Diagnostics Check		OK	21:19
Air Blank		0.000	21:19
Control Test		0.081	21:19
Air Blank		0.000	21:20
Subject Sample	# 1	0.177	21:21
Breath Volume		2.628	21:21
Air Blank		0.000	21:22
Air Blank		0.000	21:23
Subject Sample	# 2	0.169	21:24
Breath Volume		1.128	21:24
Air Blank		0.000	21:25
Control Test		0.080	21:25
Air Blank		0.000	21:25
Diagnostics Check		OK	21:26

Data Download Date/Time: 12/13/2021 12:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON,WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 10/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 11/30/2021 23:27:15

Diagnostics Check		OK	23:20
Air Blank		0.000	23:21
Control Test		0.080	23:21
Air Blank		0.000	23:22
Subject Sample	# 1	0.176	23:22
Breath Volume		3.210	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.174	23:25
Breath Volume		3.253	23:25
Air Blank		0.000	23:26
Control Test		0.079	23:26
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 12/13/2021 12:20

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: BOS

DATE/TIME OF TEST: 11/21/2021 22:07:27

Diagnostics Check		OK	22:01	
Air Blank		0.000	22:02	
Control Test		0.079	22:02	
Air Blank		0.000	22:02	
Subject Sample	# 1	0.090	22:03	
Breath Volume		1.855	22:03	
Air Blank		0.000	22:03	
Air Blank		0.000	22:05	
Subject Sample	# 2	0.099	22:06	
Breath Volume		2.757	22:06	
Air Blank		0.000	22:06	
Control Test		0.073	22:07	RFI Detect
Air Blank			22:07	RFI Detect

Data Download Date/Time: 11/30/2021 19:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: BOS

DATE/TIME OF TEST: 11/21/2021 22:14:58

Diagnostics Check		OK	22:08
Air Blank		0.000	22:09
Control Test		0.079	22:09
Air Blank		0.000	22:10
Subject Sample	# 1	0.097	22:10
Breath Volume		2.019	22:10
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 2	0.093	22:13
Breath Volume		1.843	22:13
Air Blank		0.000	22:14
Control Test		0.079	22:14
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 11/30/2021 19:14

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: BLUME

DATE/TIME OF TEST: 11/26/2021 00:10:12

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.079	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.000	00:04
Breath Volume		2.531	00:04
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.000	00:08
Breath Volume		1.722	00:08
Air Blank		0.000	00:09
Control Test		0.078	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 11/30/2021 19:14

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEETZ,JOHN,W
BREATH TEST OPERATOR AGENCY: Calhoun County SO
ARRESTING OFFICER: SCHEETZ,SCHEETZ
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 10/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: KUPISZ

DATE/TIME OF TEST: 11/02/2021 12:59:45

Diagnostics Check		OK	12:53
Air Blank		0.000	12:53
Control Test		0.081	12:54
Air Blank		0.000	12:54
Subject Sample	# 1	0.000	12:55
Breath Volume		1.722	12:55
Air Blank		0.000	12:55
Air Blank		0.000	12:57
Subject Sample	# 2	0.000	12:58
Breath Volume		2.601	12:58
Air Blank		0.000	12:58
Control Test		0.082	12:59
Air Blank		0.000	12:59
Diagnostics Check		OK	12:59

Data Download Date/Time: 11/05/2021 08:34

Printed: 15 February 2022

COBRA ©1997-2014 CMI, Inc.