

# Florida Department of Law Enforcement

## Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD  
Time of Inspection: 22:09

Date of Inspection: 04/11/2021

Serial Number: 80-001306  
Software: 8100.27

| Check or Test                               | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201910D<br>Exp: 10/22/2021 | 0.08g/210L Test (g/210L)<br>Lot#:202007A<br>Exp: 07/21/2022 | 0.20g/210L Test (g/210L)<br>Lot#:202005A<br>Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:17919080A1<br>Exp: 08/05/2021 |
|----------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-----------------------------------------------------------------------------|
| 0.000                      | 0.047                                                       | 0.081                                                       | 0.000 / 0.201                                               | 0.082                                                                       |
| 0.000                      | 0.048                                                       | 0.081                                                       | 0.000 / 0.201                                               | 0.081                                                                       |
| 0.000                      | 0.048                                                       | 0.082                                                       | 0.000 / 0.202                                               | 0.082                                                                       |

Number of Simulators Used: 4

**Remarks:**

20: Control Outside Tolerance. *Detached and reattached hoses. Tightened simulator on jar. Reran test. -JL*

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Thomas C. Roberson*

THOMAS C ROBERSON

Signature and Printed Name

04/11/2021  
Date

# Florida Department of Law Enforcement

## Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD  
Time of Inspection: 00:37

Date of Inspection: 02/04/2021

Serial Number: 80-001306  
Software: 8100.27

| Check or Test                               | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201910D<br>Exp: 10/22/2021 | 0.08g/210L Test (g/210L)<br>Lot#:202007A<br>Exp: 07/21/2022 | 0.20g/210L Test (g/210L)<br>Lot#:202005A<br>Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:17919080A1<br>Exp: 08/05/2021 |
|----------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-----------------------------------------------------------------------------|
| 0.000                      | 0.047                                                       | 0.082                                                       | 0.198                                                       | 0.081                                                                       |
| 0.000                      | 0.047                                                       | 0.082                                                       | 0.201                                                       | 0.081                                                                       |
| 0.000                      | 0.048                                                       | 0.082                                                       | 0.201                                                       | 0.081                                                                       |

Number of Simulators Used: 4

**Remarks:**

Int Det: .

*Checked simulator seal and all connections. Re-run test. - JLR*

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Thomas C. Roberson*

THOMAS C ROBERSON

Signature and Printed Name

02/04/2021  
Date

# Florida Department of Law Enforcement

## Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD  
Time of Inspection: 14:03

Date of Inspection: 01/04/2021

Serial Number: 80-001306  
Software: 8100.27

| Check or Test                               | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201910D<br>Exp: 10/22/2021 | 0.08g/210L Test (g/210L)<br>Lot#:202007A<br>Exp: 07/21/2022 | 0.20g/210L Test (g/210L)<br>Lot#:202005A<br>Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:1298240<br>Exp: 03/19/2023 |
|----------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-------------------------------------------------------------|--------------------------------------------------------------------------|
| 0.000                      | 0.046                                                       | 0.081                                                       | 0.196                                                       | 0.085 / 0.082                                                            |
| 0.000                      | 0.047                                                       | 0.082                                                       | 0.198                                                       | 0.086 / 0.082                                                            |
| 0.000                      | 0.048                                                       | 0.082                                                       | 0.199                                                       | 0.086 / 0.082                                                            |

Number of Simulators Used: 4

Lot 17919080A1  
Exp 8/5/2021

**Remarks:**

08: Control Outside Tolerance.

Supply cylinder expended during inspection. Replaced cylinder and reran test.  
- JLR

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Thomas C. Roberson

THOMAS C ROBERSON

Signature and Printed Name

01/04/2021  
Date