

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO
Time of Inspection: 13:45

Date of Inspection: 09/03/2024

Serial Number: 80-000815
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
/ 0.000	0.047	0.076	0.187 / 0.187	
/ 0.000	0.047	0.076	0.187 / 0.187	
/ 0.000	0.048	0.077	0.188 / 0.187	

Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded. 00: Ambient Fail. 20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: OUT OF TOLERANCE FOR .20 TEST, TESTED TWICE.


AF/MA: Range Exceeded due to too much mouth Alcohol present. Rinsed mouth, restarted test, had another officer provide the Alcohol Free test, reapplied mouth alcohol and Passed test.

Ambient Fail: Repositioned fan to better ventilate room.

Out of tolerance: Called FDLE and spoke with Inspector Gutschow. He advised to check the connections on the Simulator and then run ten stability checks on the instrument. Stability checks were conducted and all ten Passed. Instrument was then re-inspected.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name PAUL A CAUDLE

09/03/2024
Date