



Agency Marion County SO

S/N 80-000813

Date In 12-06-23

DI Completion Date 12/7/2023

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By	ALL	Quality Checks	By	ALL	Date	12-07-23	Flow Calibration	By		Date														
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>127</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.230</u> (.228 - .278) 103 mm <u>.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks				Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																		
			<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202303K 03-29-25</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202303L 03-29-25</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202304C 04-05-25</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG325603 09-13-25</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202303K 03-29-25	0.080	MP5089	202303L 03-29-25	0.200	MP5090	202304C 04-05-25	0.080 DGS	N/A	AG325603 09-13-25				Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____			
Simulator	Serial #	Lot #/Exp																							
0.050	MP5088	202303K 03-29-25																							
0.080	MP5089	202303L 03-29-25																							
0.200	MP5090	202304C 04-05-25																							
0.080 DGS	N/A	AG325603 09-13-25																							

Calibration Adjustment				By _____	
Barometric Pressure Gauge _____ ID # _____					
Simulator	Serial #	Lot #	Expiration		
0.000		N/A	N/A		
0.040					
0.100					
0.200					
0.300					
0.080 DGS	N/A				
☐ Post Calibration Adjustment Stability Checks					
Simulator	Serial #	Lot #	Expiration		
0.050					
0.080					
0.200					
0.080 DGS	N/A				
Notes/Suggested Service: <u>During first inspection, called away from instrument during mouth alcohol test, no sample provided. Started a new inspection.(BS 12/7/23)</u> <u>Admin Review: Replaced abbreviations in stability checks (BS 12/11/23)</u> 					

Department Inspection		By BS _____	
Barometric Pressure ID# <u>28427</u>			
Gauge <u>1021</u>	Instrument <u>1020</u>		
Mouth Alcohol Solution Lot # <u>2022-A</u>			
Acetone Stock Solution Lot # <u>2023-B</u>			
Simulator	Serial Number		
0.000	MP5086		
Interferent	MP5087		
0.050	MP5088		
0.080	MP5089		
0.200	MP5090		
Attachments			
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment		☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____	
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use			
Israel Soto Digitally signed by Israel Soto Date: 2023.12.11 08:34:49 -05'00'	Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.12.12 15:46:27 -05'00'		
Tech Review / Date _____		Admin Review / Date _____	

80-000813 stability checks

12-07-23

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
12/07/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:54
Control Test	0.050	09:55
Air Blank	0.000	09:55
Control Test	0.050	09:56
Air Blank	0.000	09:57
Control Test	0.050	09:57
Air Blank	0.000	09:58
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
12/07/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:01
Control Test	0.079	10:02
Air Blank	0.000	10:03
Control Test	0.079	10:03
Air Blank	0.000	10:04
Control Test	0.079	10:05
Air Blank	0.000	10:05
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Wet

BS 12/11/2023

Operator's Signature

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
12/07/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:47
Control Test	0.199	09:47
Air Blank	0.000	09:48
Control Test	0.198	09:49
Air Blank	0.000	09:49
Control Test	0.197	09:50
Air Blank	0.000	09:50
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

Operator's Signature

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
12/07/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:08
Control Test	0.081	10:08
Air Blank	0.000	10:09
Control Test	0.081	10:09
Air Blank	0.000	10:09
Control Test	0.081	10:10
Air Blank	0.000	10:10
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Dry

BS 12/11/2023

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 11:09

Date of Inspection: 12/07/2023

Serial Number: 80-000813
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:

A F / M A: No Sample Provided. Non-compliance:NO SAMPLE PROVIDED AT MOUTH ALCOHOL TEST .

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BENJAMIN W SIDDOWAY
Signature and Printed Name

12/07/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 13:12

Date of Inspection: 12/07/2023

Serial Number: 80-000813
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG325603 Exp: 09/13/2025
0.000	0.050	0.079	0.198	0.080
0.000	0.051	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.051	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.051	0.080	0.198	0.079
0.000	0.050	0.080	0.198	0.079
Standard Deviations	0.0004	0.0004	0.0000	0.0005


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

A F / M A: No Sample Provided. Repeated test and gave sample.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 BENJAMIN W SIDDOWAY
Signature and Printed Name

12/07/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000813, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000813</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MARION COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/07/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:12</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Benjamin
Siddoway
Digitally signed by
Benjamin Siddoway
Date: 2023.12.07 15:15:26
-05'00'

12/07/2023

Date

BENJAMIN W SIDDOWAY,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Marion County Sheriff's OfficeS/N 80-000813Florida Department of
Law EnforcementDate In 09-22-2023 DI Completion Date 9/26/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By ALL	Quality Checks	By BS	Date 9/25/2023	Flow Calibration	By	Date																																																														
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: upon arrival 3 feet disconnected, 1 missing spacer during intake shelf completely disconnected		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>128</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks			☐ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																																
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202303K 3/29/2025</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L 3/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C 4/5/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5 4/5/2025</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	202303K 3/29/2025	0.080	MP5089	202303L 3/29/2025	0.200	MP5090	202304C 4/5/2025	0.080 DGS	N/A	06723080A5 4/5/2025	Maintenance By BS ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Pushed foot anchors back into case, replaced missing spacer foot, and screwed shelf back to case. (9/25/2023)</u>																																																	
Simulator	Serial #	Lot #/Exp																																																																			
0.050	MP5088	202303K 3/29/2025																																																																			
0.080	MP5089	202303L 3/29/2025																																																																			
0.200	MP5090	202304C 4/5/2025																																																																			
0.080 DGS	N/A	06723080A5 4/5/2025																																																																			
Calibration Adjustment By _____				Department Inspection By BS _____																																																																	
Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28662</u> Gauge <u>1015</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table> Attachments <table border="1"><tbody><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment</td><td>☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____</td></tr></tbody></table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use				Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____
Simulator	Serial #	Lot #	Expiration																																																																		
0.000		N/A	N/A																																																																		
0.040																																																																					
0.100																																																																					
0.200																																																																					
0.300																																																																					
0.080 DGS	N/A																																																																				
Simulator	Serial #	Lot #	Expiration																																																																		
0.050																																																																					
0.080																																																																					
0.200																																																																					
0.080 DGS	N/A																																																																				
Simulator	Serial Number																																																																				
0.000	MP5086																																																																				
Interferent	MP5087																																																																				
0.050	MP5088																																																																				
0.080	MP5089																																																																				
0.200	MP5090																																																																				
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																																				
Notes/Suggested Service: _____ _____ _____ _____ _____ _____				Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2023.10.09 13:33:22 -04'00' Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.10.10 09:49:50 -04'00' Tech Review / Date _____ Admin Review / Date _____																																																																	

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 14:26

Date of Inspection: 09/26/2023

Serial Number: 80-000813
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06723080A5 Exp: 04/05/2025
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.197	0.079


Standard Deviations	0.0005	0.0003	0.0005	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 BENJAMIN W SIDDOWAY
Signature and Printed Name

09/26/2023
Date

Stability Checks

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
09/25/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:47
Control Test	0.049	14:48
Air Blank	0.000	14:48
Control Test	0.049	14:49
Air Blank	0.000	14:50
Control Test	0.049	14:50
Air Blank	0.000	14:51
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Benjamin Siddaway
Operator's Signature

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
09/25/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:53
Control Test	0.079	14:54
Air Blank	0.000	14:55
Control Test	0.079	14:55
Air Blank	0.000	14:56
Control Test	0.078	14:57
Air Blank	0.000	14:57
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Benjamin Siddaway
Operator's Signature

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
09/25/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:59
Control Test	0.197	14:59
Air Blank	0.000	15:00
Control Test	0.196	15:01
Air Blank	0.000	15:01
Control Test	0.196	15:02
Air Blank	0.000	15:02
Control Test Stats		
Average	0.1963	
Std Dev	0.0006	
Rel Std Dev(%)	0.2941	

Benjamin Siddaway
Operator's Signature

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
09/25/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:04
Control Test	0.080	15:05
Air Blank	0.000	15:05
Control Test	0.080	15:05
Air Blank	0.000	15:06
Control Test	0.080	15:06
Air Blank	0.000	15:07
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Benjamin Siddaway
Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000813, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000813</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MARION COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>09/26/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:26</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Digitally signed by
Benjamin Siddoway
Date: 2023.09.26 16:05:04
-04'00'

Benjamin
Siddoway

09/26/2023

Date

BENJAMIN W SIDDOWAY,
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Marion County Sheriff's OfficeS/N 80-000813Florida Department of
Law EnforcementDate In 02-24-2023 DI Completion Date 05-02-2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By IS	Quality Checks	By IS	Date	03-02-2023	Flow Calibration	By	Date																											
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: Instrument was in a car during a car accident. Dry gas shelf was pulled off of instrument. Dry gas regulator cord broken. Dry gas shelf and feet included in box.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>131</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.531</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks																																	
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C 01-11-2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202201D 01-18-2024</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E 01-18-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803 10-25-2024</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01-11-2024	0.080	MP6292	202201D 01-18-2024	0.200	MP6293	202201E 01-18-2024	0.080 DGS	N/A	AG229803 10-25-2024			<table border="1"><thead><tr><th>Maintenance</th><th>By IS</th></tr></thead><tbody><tr><td>☒ Battery Replacement ☒ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Reattached dry gas shelf, screwed feet back into place 2-28. Replaced Internal Printer paper 3-02 Dry Gas Sensor Tare after replacing regulator 5-01</u></td><td></td></tr></tbody></table>		Maintenance	By IS	☒ Battery Replacement ☒ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Reattached dry gas shelf, screwed feet back into place 2-28. Replaced Internal Printer paper 3-02 Dry Gas Sensor Tare after replacing regulator 5-01</u>										
Simulator	Serial #	Lot #/Exp																																	
0.050	MP6291	202201C 01-11-2024																																	
0.080	MP6292	202201D 01-18-2024																																	
0.200	MP6293	202201E 01-18-2024																																	
0.080 DGS	N/A	AG229803 10-25-2024																																	
Maintenance	By IS																																		
☒ Battery Replacement ☒ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Reattached dry gas shelf, screwed feet back into place 2-28. Replaced Internal Printer paper 3-02 Dry Gas Sensor Tare after replacing regulator 5-01</u>																																			
Calibration Adjustment		By		Department Inspection		By IS																													
Barometric Pressure Gauge <u> </u> ID # <u> </u>				Barometric Pressure ID# <u>28662</u>																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					Gauge <u>1009</u> Instrument <u>1007</u>			
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
☐ Post Calibration Adjustment Stability Checks				Mouth Alcohol Solution Lot # <u>2022-A</u>																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					Acetone Stock Solution Lot # <u>2022-B</u>											
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Notes/Suggested Service: <u>An FDLE dry gas regulator was used for the Stability Checks. Stability Check values in nominal range. Regulator was replaced with newly ordered regulator (5-01) prior to Department Inspection. Extra Dry Stability Checks done with new regulator, values nominal. Battery replaced 5-01 prior to Department Inspection. IS</u>				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090																		
Simulator	Serial Number																																		
0.000	MP5086																																		
Interferent	MP5087																																		
0.050	MP5088																																		
0.080	MP5089																																		
0.200	MP5090																																		
				Attachments																															
				☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment		☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 x2 ☒ Other <u>Extra Dry Stability Checks</u>																													
				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																															
				Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2023.05.03 09:13:20 -04'00'		Phil Nicodemmo Digitally signed by Phil Nicodemmo Date: 2023.05.04 15:44:16 -04'00'																													
				Tech Review / Date		Admin Review / Date																													

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 11:59

Date of Inspection: 05/02/2023

Serial Number: 80-000813
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.048	0.077	0.199	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.077	0.198	0.078
0.000	0.048	0.077	0.198	0.078
0.000	0.048	0.077	0.198	0.078
0.000	0.048	0.077	0.198	0.078
0.000	0.048	0.077	0.199	0.078
0.000	0.048	0.077	0.199	0.078
0.000	0.048	0.077	0.199	0.078

Standard Deviations	0.0000	0.0004	0.0005	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto ISRAEL SOTO
Signature and Printed Name

05/02/2023
Date

Stability Checks

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
03/02/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:37
Control Test	0.049	13:38
Air Blank	0.000	13:39
Control Test	0.049	13:39
Air Blank	0.000	13:40
Control Test	0.048	13:41
Air Blank	0.000	13:41
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

Operator's Signature

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
03/02/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:44
Control Test	0.078	13:45
Air Blank	0.000	13:45
Control Test	0.077	13:46
Air Blank	0.000	13:46
Control Test	0.077	13:47
Air Blank	0.000	13:48
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

Operator's Signature

wet

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
03/02/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:49
Control Test	0.197	13:49
Air Blank	0.000	13:50
Control Test	0.197	13:50
Air Blank	0.000	13:51
Control Test	0.196	13:52
Air Blank	0.000	13:52
Control Test Stats		
Average	0.1967	
Std Dev	0.0006	
Rel Std Dev(%)	0.2936	

Operator's Signature

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
03/02/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:53
Control Test	0.079	13:54
Air Blank	0.000	13:54
Control Test	0.078	13:55
Air Blank	0.000	13:55
Control Test	0.079	13:55
Air Blank	0.000	13:56
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Operator's Signature

Dry



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000813, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000813</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MARION COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/02/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:59</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel Soto
Date: 2023.05.02 12:23:23 -04'00'

05/02/2023

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 13:17

Date of Inspection: 03/02/2023

Serial Number: 80-000813
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

N/A Compliance not determined

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

03/02/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 14:40

Date of Inspection: 05/01/2023

Serial Number: 80-000813
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

N/A Compliance not determined

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

05/01/2023
Date

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
05/01/2023
Software: 8100.27

Dry Stability checks
with new dry gas regulator

Test	g/210L	Time
Air Blank	0.000	14:43
Control Test	0.079	14:43
Air Blank	0.000	14:43
Control Test	0.079	14:44
Air Blank	0.000	14:44
Control Test	0.078	14:45
Air Blank	0.000	14:45
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Dry


Operator's Signature