

Alcohol Testing Program - Instrument Processing Sheet

Agency: FDLE

Instrument Serial Number: 80-000206

Date In: 9/16/2026

DI Completion Date: 9/18/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: WKP Date: 9/17/2026				Quality Checks By: WKP Date: 9/18/2026				Flow Adjustment By:			
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Instrument arrived with a keyboard cover. WKP 9/17/2026				☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: 199 ☒ Flow Verification (L/S) Flow Column #: ATP 105 32 mm 0.160 (.139-.169) 36 mm 0.175 (.156-.190) 53 mm 0.253 (.228-.278) 103 mm 0.535 (.447-.547)				Flow Column #: ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: ☐ Post Adjustment Verification (L/S) Flow Column #: 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)			
				☒ Barometric Pressure Check Gauge ID #: 28427 Gauge: 1015 Instrument: 1017							
				☒ Stability Checks							
				Simulator		Serial #		Lot#/Exp		Maintenance By: Date:	
				0.050		MP6291		202603A 3/3/2028		☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:	
0.080		MP6292		202603B 3/3/2028							
0.200		MP6293		202603D 3/4/2028							
0.080 DGS		N/A		28424080A3 11/5/2026							
Optical Bench Adjustment By:				Department Inspection By: WKP							
Barometric Pressure Gauge: ID#:				Barometric Pressure ID#: 28662							
Simulator		Serial #		Lot #		Expiration					
0.000				N/A		N/A					
0.040											
0.100											
0.200											
0.300											
0.080 DGS		N/A									
☐ Post Optical Bench Adjustment Stability Checks											
Simulator		Serial #		Lot #		Expiration					
0.050											
0.080											
0.200											
0.080 DGS		N/A									
Gauge ID #:				Gauge: 1016 Instrument: 1017							
Gauge: Instrument:				Mouth Alcohol Solution Lot #: 2025-D Exp: 9/25/2027 Acetone Stock Solution Lot #: 2025-B Exp: 9/22/2027							
Notes/Suggested Service: Admin correction: Added return method check mark. WKP 9/22/2026				Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment ☒ Average Calculations ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:							
				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use							
Kaitlyn Spearin Digitally signed by Kaitlyn Date: 2026.09.21 11:17:36 -04'00'				Isabella Smykowski Digitally signed by Isabella Smykowski Date: 2026.09.22 09:21:32 -04'00'							
Tech Review				Admin Review							

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓ FDLE Intoxilyzer - Alcohol Analyzer Model: 8000 09/18/2026 Software: 8100.27 SN 80-001206 Test g/210L Time Air Blank 0.000 12:04 Control Test 0.049 12:05 Air Blank 0.000 12:05 Control Test 0.049 12:06 Air Blank 0.000 12:06 Control Test 0.049 12:07 Air Blank 0.000 12:08 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>menpi</i>	✓ FDLE Intoxilyzer - Alcohol Analyzer Model: 8000 09/18/2026 Software: 8100.27 SN 80-001206 Test g/210L Time Air Blank 0.000 12:10 Control Test 0.077 12:10 Air Blank 0.000 12:11 Control Test 0.078 12:12 Air Blank 0.000 12:12 Control Test 0.078 12:13 Air Blank 0.000 12:13 Control Test Stats Average 0.0777 Std Dev 0.0016 Rel Std Dev(%) 0.7434 Operator's Signature <i>menpi</i>	✓ FDLE Intoxilyzer - Alcohol Analyzer Model: 8000 09/18/2026 Software: 8100.27 SN 80-001205 Test g/210L Time Air Blank 0.000 12:15 Control Test 0.197 12:16 Air Blank 0.000 12:16 Control Test 0.198 12:17 Air Blank 0.000 12:17 Control Test 0.195 12:18 Air Blank 0.000 12:19 Control Test Stats Average 0.1967 Std Dev 0.0015 Rel Std Dev(%) 0.7767 Operator's Signature <i>menpi</i>	✓ FDLE Intoxilyzer - Alcohol Analyzer Model: 8000 09/18/2026 Software: 8100.27 SN 80-101206 Test g/210L Time Air Blank 0.000 12:22 Control Test 0.080 12:21 Air Blank 0.000 12:21 Control Test 0.079 12:22 Air Blank 0.000 12:22 Control Test 0.081 12:22 Air Blank 0.000 12:23 Control Test Stats Average 0.0800 Std Dev 0.0010 Rel Std Dev(%) 1.2500 Operator's Signature <i>menpi</i>

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 15:23

Date of Inspection: 09/18/2026

Serial Number: 80-000206

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.077	0.197	0.081
0.000	0.049	0.076	0.198	0.080
0.000	0.048	0.077	0.197	0.080
0.000	0.049	0.077	0.198	0.080
0.000	0.049	0.077	0.197	0.080
0.000	0.049	0.077	0.198	0.080
0.000	0.049	0.077	0.198	0.080
0.000	0.049	0.077	0.197	0.080
0.000	0.049	0.077	0.197	0.079
0.000	0.048	0.077	0.198	0.080


Standard Deviations	0.0004	0.0003	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WEN-CHI K. PIERSON

Signature and Printed Name

09/18/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000206, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000206</u>	Average of $y \pm$ Uncertainty*
Owning Agency:	<u>FDLE</u>	0.049 g/ 210 L 0.004
Calibration Date:	<u>09/18/2026</u>	0.077 g/ 210 L 0.004
Calibration Time:	<u>15:23</u>	0.198 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi
Pierson
Digitally signed by Wen-Chi
Pierson
Date: 2026.09.21 09:04:20
-04'00'

09/18/2026

Date

WEN-CHI K PIERSON,
Department Inspector

FDLE/ATP Form 69 August 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality