

Alcohol Testing Program - Instrument Processing Sheet

Agency: Monroe County Sheriff's Office

Instrument Serial Number: 80-000867

Date In: 8/31/2026

DI Completion Date: 9/10/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>TDG</u> Date: <u>9/4/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>9/4/2026</u>		Flow Adjustment By: _____																
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>300</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm 0.148 (.139-.169) 36 mm 0.164 (.156-.190) 53 mm 0.253 (.228-.278) 103 mm 0.507 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1017</u> Instrument: <u>1010</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																
		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>28424080A3 11/5/2026</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6286	202603A 3/3/2028	0.080	MP6287	202603B 3/3/2028	0.200	MP6288	202603D 3/4/2028	0.080 DGS	N/A	28424080A3 11/5/2026	Maintenance By: <u>TDG</u> Date: <u>9/4/2026</u> ☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:	
Simulator	Serial #	Lot#/Exp																		
0.050	MP6286	202603A 3/3/2028																		
0.080	MP6287	202603B 3/3/2028																		
0.200	MP6288	202603D 3/4/2028																		
0.080 DGS	N/A	28424080A3 11/5/2026																		
Optical Bench Adjustment By: <u>TDG</u>		Department Inspection By: <u>TDG</u>																		
Barometric Pressure Gauge: <u>1019</u> ID#: <u>33364</u>		Barometric Pressure ID#: <u>33363</u>																		
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1018</u> Instrument: <u>1018</u>																
0.000	MP4863	N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-B</u> Exp: <u>9/25/2027</u>																
0.040	MP4864	25090	3/11/2027	Acetone Stock Solution Lot #: <u>2025-D</u> Exp: <u>9/22/2027</u>																
0.100	MP8229	24390	10/29/2026	Simulator	Serial Number															
0.200	MP8230	25020	1/14/2027	0.000	MP6284															
0.300	MP8231	24430	12/10/2026	Interferent	MP6285															
0.080 DGS	N/A	23725080A1	10/5/2027	0.050	MP6286															
☒ Post Optical Bench Adjustment Stability Checks				0.080	MP6287															
Simulator	Serial #	Lot #	Expiration	0.200	MP6288															
0.050	MP6286	202603A	3/3/2028	Attachments																
0.080	MP6287	202603B	3/3/2028	☒ Form 41	☒ Post-Stability Checks															
0.200	MP6288	202603D	3/4/2028	☒ Stability Checks	☐ Flow Adjustment															
0.080 DGS	N/A	AG526603	9/23/2027	☒ Calibration Certificate	☐ Form 40															
Gauge ID #: <u>33363</u>				☒ Optical Bench Adjustment	☐ Other:															
Gauge: <u>1017</u> Instrument: <u>1017</u>				☒ Average Calculations																
Notes/Suggested Service:				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																
Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.09.11 13:15:41 -04'00'				S. LeAndra Higginbotham Digitally signed by S. LeAndra Higginbotham Date: 2026.09.11 13:56:19 -04'00'																
Tech Review				Admin Review																

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																						
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																						
✓	✗	✓	✗																																																																																																																																																						
MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000867 09/04/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:51</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:52</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:55</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:51	Control Test	0.050	12:52	Air Blank	0.000	12:52	Control Test	0.050	12:53	Air Blank	0.000	12:53	Control Test	0.049	12:54	Air Blank	0.000	12:55	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000867 09/04/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:51</td></tr> <tr><td>Control Test</td><td>0.076</td><td>12:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:28</td></tr> <tr><td>Control Test</td><td>0.076</td><td>12:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr> <tr><td>Control Test</td><td>0.076</td><td>12:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:30</td></tr> <tr><td>Control Test</td><td>0.076</td><td>12:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0760</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:51	Control Test	0.076	12:27	Air Blank	0.000	12:28	Control Test	0.076	12:28	Air Blank	0.000	12:29	Control Test	0.076	12:29	Air Blank	0.000	12:30	Control Test	0.076	12:30	Air Blank	0.000	12:31	Control Test Stats			Average	0.0760		Std Dev	0.0010		Rel Std Dev(%)	0.0000		MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000867 09/04/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:35</td></tr> <tr><td>Control Test</td><td>0.199</td><td>12:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:36</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:37</td></tr> <tr><td>Control Test</td><td>0.199</td><td>12:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:38</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2906</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:35	Control Test	0.199	12:36	Air Blank	0.000	12:36	Control Test	0.198	12:37	Air Blank	0.000	12:37	Control Test	0.199	12:38	Air Blank	0.000	12:38	Control Test Stats			Average	0.1987		Std Dev	0.0006		Rel Std Dev(%)	0.2906		MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000867 09/04/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:44</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:45</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:45</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:46</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:44	Control Test	0.080	12:44	Air Blank	0.000	12:45	Control Test	0.079	12:45	Air Blank	0.000	12:45	Control Test	0.080	12:46	Air Blank	0.000	12:46	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	12:51																																																																																																																																																							
Control Test	0.050	12:52																																																																																																																																																							
Air Blank	0.000	12:52																																																																																																																																																							
Control Test	0.050	12:53																																																																																																																																																							
Air Blank	0.000	12:53																																																																																																																																																							
Control Test	0.049	12:54																																																																																																																																																							
Air Blank	0.000	12:55																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.0497																																																																																																																																																								
Std Dev	0.0006																																																																																																																																																								
Rel Std Dev(%)	1.1625																																																																																																																																																								
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	12:51																																																																																																																																																							
Control Test	0.076	12:27																																																																																																																																																							
Air Blank	0.000	12:28																																																																																																																																																							
Control Test	0.076	12:28																																																																																																																																																							
Air Blank	0.000	12:29																																																																																																																																																							
Control Test	0.076	12:29																																																																																																																																																							
Air Blank	0.000	12:30																																																																																																																																																							
Control Test	0.076	12:30																																																																																																																																																							
Air Blank	0.000	12:31																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.0760																																																																																																																																																								
Std Dev	0.0010																																																																																																																																																								
Rel Std Dev(%)	0.0000																																																																																																																																																								
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	12:35																																																																																																																																																							
Control Test	0.199	12:36																																																																																																																																																							
Air Blank	0.000	12:36																																																																																																																																																							
Control Test	0.198	12:37																																																																																																																																																							
Air Blank	0.000	12:37																																																																																																																																																							
Control Test	0.199	12:38																																																																																																																																																							
Air Blank	0.000	12:38																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.1987																																																																																																																																																								
Std Dev	0.0006																																																																																																																																																								
Rel Std Dev(%)	0.2906																																																																																																																																																								
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	12:44																																																																																																																																																							
Control Test	0.080	12:44																																																																																																																																																							
Air Blank	0.000	12:45																																																																																																																																																							
Control Test	0.079	12:45																																																																																																																																																							
Air Blank	0.000	12:45																																																																																																																																																							
Control Test	0.080	12:46																																																																																																																																																							
Air Blank	0.000	12:46																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.0797																																																																																																																																																								
Std Dev	0.0006																																																																																																																																																								
Rel Std Dev(%)	0.7247																																																																																																																																																								

```

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000
% RDS = 0.128
Std Dev = 0.02 Rel Std Dev =
Sol Val = 0.1905 mg/l or 0.040
% RDS = 0.892
Std Dev = 0.01 Rel Std Dev =
Sol Val = 0.4762 mg/l or 0.100
% Abs = 2.072

```

```
Std Dev = 0.02 Rel Std Dev = 0.85
Sol Val = 0.924 mg/l or 0.200 g/210L
% Abs = 3.914
Std Dev = 0.04 Rel Std Dev = 1.05
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.728
Std Dev = 0.00 Rel Std Dev = 0.03
Zero Order Coef = -237.65
First Order Coef = 2410.59
Second Order Coef = 23.78
Standard Deviation = 27.54251
-----
<<<<< CHANNEL 2 >>>>>
Sol Val = 0.000 mg/l or 0.000 g/210L
```

Sol. Val. = 1.4286 mg/L or 0.300 g/210L
% Abs = 5.728
Std Dev = 0.00 Std Dev = 0.03
Zero Order Coef = -297.65
First Order Coef = 2410.59
Second Order Coef = 23.78
Standard Deviation = 27.542501

```

First Order Coef = 2410.59
Second Order Coef = 23.78
Standard Deviation = 27.542501
-----
<<<< CHANNEL 2 >>>>
oi Val = 0.0000 mg/l or 0.000 g/210L

```

Std Dev = 0.01 Rel Std Dev = 7.19
 Impl Val = 0.1915 max for 0.040 n/210

Std Dev = 0.11 Rel. Std. Dev. = 0.46
1 Val = 0.4762 mg/l or 0.100 g/210L
Abs = 3.624
Std Dev = 0.01 Rel. Std. Dev. = 0.17
1 Val = 0.9524 mg/l or 0.200 g/210L
Abs = 6.913
Std Dev = 0.01 Rel. Std. Dev. = 0.18
1 Val = 1.4266 mg/l or 0.300 g/210L
Abs = 10.161
Std Dev = 0.02 Rel. Std. Dev. = 0.20

First Order Coef = 1315.90
Second Order Coef = 11.84
Standard Deviation = 9.41322

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1018
3 um H2O Adjust (mg/L*10,000) = 747
9 um H2O Adjust (mg/L*10,000) = 449
**** AUTO CAL PASS

By: TDG

Post-Adjustment Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083	≤0.003 of Wet																																				
MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000867 09/09/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:39</td></tr> <tr><td>Control Test</td><td>0.076</td><td>15:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:40</td></tr> <tr><td>Control Test</td><td>0.075</td><td>15:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:41</td></tr> <tr><td>Control Test</td><td>0.076</td><td>15:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:43</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0757</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev (%)</td><td>0.7630</td><td></td></tr> </tbody> </table> Operator's Signature: 	Test	g/210L	Time	Air Blank	0.000	15:39	Control Test	0.076	15:40	Air Blank	0.000	15:40	Control Test	0.075	15:41	Air Blank	0.000	15:41	Control Test	0.076	15:42	Air Blank	0.000	15:43	Control Test Stats			Average	0.0757		Std Dev	0.0006		Rel Std Dev (%)	0.7630		Simulator head was loose. Tightened and repeated. ML 9/9/26			
Test	g/210L	Time																																						
Air Blank	0.000	15:39																																						
Control Test	0.076	15:40																																						
Air Blank	0.000	15:40																																						
Control Test	0.075	15:41																																						
Air Blank	0.000	15:41																																						
Control Test	0.076	15:42																																						
Air Blank	0.000	15:43																																						
Control Test Stats																																								
Average	0.0757																																							
Std Dev	0.0006																																							
Rel Std Dev (%)	0.7630																																							

Post-Adjustment Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
☒	☒	☒	☒ ☒ ☒ ☒
MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000867 09/09/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:33 Control Test 0.050 15:33 Air Blank 0.000 15:34 Control Test 0.050 15:34 Air Blank 0.000 15:35 Control Test 0.049 15:36 Air Blank 0.000 15:36 Control Test Stats Average 0.0497 Std Dev 0.0006 Rel Std Dev(%) 1.1625 Operator's Signature 	MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000867 09/09/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:51 Control Test 0.077 15:52 Air Blank 0.000 15:52 Control Test 0.077 15:53 Air Blank 0.000 15:53 Control Test 0.077 15:54 Air Blank 0.000 15:55 Control Test Stats Average 0.0770 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 	MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000867 09/09/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:45 Control Test 0.200 15:46 Air Blank 0.000 15:46 Control Test 0.200 15:47 Air Blank 0.000 15:47 Control Test 0.199 15:48 Air Blank 0.000 15:49 Control Test Stats Average 0.1997 Std Dev 0.0006 Rel Std Dev(%) 0.2892 Operator's Signature 	MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000867 09/09/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:18 Control Test 0.080 15:18 Air Blank 0.000 15:19 Control Test 0.080 15:19 Air Blank 0.000 15:19 Control Test 0.080 15:20 Air Blank 0.000 15:20 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 15:27

Date of Inspection: 09/10/2026

Serial Number: 80-000867
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.050	0.076	0.197	0.080
0.000	0.049	0.076	0.198	0.080
0.000	0.049	0.076	0.197	0.080
0.000	0.049	0.075	0.197	0.079
0.000	0.049	0.075	0.197	0.080
0.000	0.049	0.076	0.197	0.080
0.000	0.049	0.075	0.197	0.080
0.000	0.049	0.075	0.197	0.080
0.000	0.049	0.075	0.197	0.080
0.000	0.049	0.075	0.197	0.080
0.000	0.050	0.076	0.197	0.080


Standard Deviations	0.0004	0.0005	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

09/10/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number **80-000867**, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-000867	Average of y $\pm$ Uncertainty*
Owning Agency:	MONROE COUNTY SO	0.049 g/ 210 L 0.004
Calibration Date:	09/10/2026	0.076 g/210 L 0.004
Calibration Time:	15:27	0.197 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2026.09.11 11:13:56
-04'00'

09/10/2026

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 August 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Alcohol Testing Program - Instrument Processing Sheet

Agency: Monroe County Sheriff's Office Instrument Serial Number: 80-000867

Date In: 5/28/2026 DI Completion Date: 6/10/2026 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>TDG</u> Date: <u>6/9/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>6/9/2026</u>		Flow Adjustment By: _____																																							
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>260</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm 0.152 (.139-.169) 36 mm 0.171 (.156-.190) 53 mm 0.257 (.228-.278) 103 mm 0.507 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1019</u> Instrument: <u>1013</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																																							
						<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> <tr> <td>0.050</td> <td>MP6286</td> <td>202406K 6/19/2026</td> <td rowspan="5"> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: </td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202406N 6/20/2026</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG526603</td> </tr> <tr> <td>9/23/2027</td> </tr> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP6286	202406K 6/19/2026	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:	0.080	MP6287	202603B 3/3/2028	0.200	MP6288	202406N 6/20/2026	0.080 DGS	N/A	AG526603	9/23/2027																		
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																																		
						0.050	MP6286	202406K 6/19/2026	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:																																		
						0.080	MP6287	202603B 3/3/2028																																			
0.200	MP6288	202406N 6/20/2026																																									
0.080 DGS	N/A	AG526603																																									
		9/23/2027																																									
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 50%;">Simulator</th> <th style="width: 50%;">Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																								
0.000		N/A	N/A																																								
0.040																																											
0.100																																											
0.200																																											
0.300																																											
0.080 DGS	N/A																																										
Simulator	Serial Number																																										
0.000	MP6284																																										
Interferent	MP6285																																										
0.050	MP6286																																										
0.080	MP6287																																										
0.200	MP6288																																										
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th colspan="2" style="background-color: #d3d3d3;">Attachments</th> </tr> <tr> <td style="width: 50%; vertical-align: top;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment </td> <td style="width: 50%; vertical-align: top;"> ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other: </td> </tr> </table>		Attachments		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment	☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																
Simulator	Serial #	Lot #	Expiration																																								
0.050																																											
0.080																																											
0.200																																											
0.080 DGS	N/A																																										
Attachments																																											
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment	☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																																										
Gauge ID #: _____ Gauge: _____ Instrument: _____		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																									
Notes/Suggested Service: Tech Review: Added check mark to 0.08 ARS Stability Check. (TDG 6/12/26)		<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 33%; text-align: center;"> Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.06.12 14:41:26 -04'00' </td> <td style="width: 33%; text-align: center;"> LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.06.18 19:00:06 -04'00' </td> <td style="width: 33%;"></td> </tr> <tr> <td style="text-align: center;">Tech Review</td> <td style="text-align: center;">Admin Review</td> <td></td> </tr> </table>		Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.06.12 14:41:26 -04'00'	LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.06.18 19:00:06 -04'00'		Tech Review	Admin Review																																			
Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.06.12 14:41:26 -04'00'	LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.06.18 19:00:06 -04'00'																																										
Tech Review	Admin Review																																										

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000867 06/09/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:52</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:53</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:54</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	14:52	Control Test	0.049	14:53	Air Blank	0.000	14:53	Control Test	0.049	14:54	Air Blank	0.000	14:54	Control Test	0.049	14:55	Air Blank	0.000	14:56	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000867 06/09/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:31</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:32</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:34</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	15:31	Control Test	0.078	15:32	Air Blank	0.000	15:32	Control Test	0.078	15:33	Air Blank	0.000	15:34	Control Test	0.078	15:34	Air Blank	0.000	15:35	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000867 06/09/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:38</td></tr> <tr><td>Control Test</td><td>0.196</td><td>14:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:39</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:40</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:42</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1967</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2935</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	14:38	Control Test	0.196	14:39	Air Blank	0.000	14:39	Control Test	0.197	14:40	Air Blank	0.000	14:40	Control Test	0.197	14:41	Air Blank	0.000	14:42	Control Test Stats			Average	0.1967		Std Dev	0.0006		Rel Std Dev(%)	0.2935		MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000867 06/09/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:16</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	14:14	Control Test	0.079	14:14	Air Blank	0.000	14:15	Control Test	0.078	14:15	Air Blank	0.000	14:16	Control Test	0.079	14:16	Air Blank	0.000	14:17	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:52																																																																																																																																																	
Control Test	0.049	14:53																																																																																																																																																	
Air Blank	0.000	14:53																																																																																																																																																	
Control Test	0.049	14:54																																																																																																																																																	
Air Blank	0.000	14:54																																																																																																																																																	
Control Test	0.049	14:55																																																																																																																																																	
Air Blank	0.000	14:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:31																																																																																																																																																	
Control Test	0.078	15:32																																																																																																																																																	
Air Blank	0.000	15:32																																																																																																																																																	
Control Test	0.078	15:33																																																																																																																																																	
Air Blank	0.000	15:34																																																																																																																																																	
Control Test	0.078	15:34																																																																																																																																																	
Air Blank	0.000	15:35																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:38																																																																																																																																																	
Control Test	0.196	14:39																																																																																																																																																	
Air Blank	0.000	14:39																																																																																																																																																	
Control Test	0.197	14:40																																																																																																																																																	
Air Blank	0.000	14:40																																																																																																																																																	
Control Test	0.197	14:41																																																																																																																																																	
Air Blank	0.000	14:42																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1967																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2935																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:14																																																																																																																																																	
Control Test	0.079	14:14																																																																																																																																																	
Air Blank	0.000	14:15																																																																																																																																																	
Control Test	0.078	14:15																																																																																																																																																	
Air Blank	0.000	14:16																																																																																																																																																	
Control Test	0.079	14:16																																																																																																																																																	
Air Blank	0.000	14:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 10:47

Date of Inspection: 06/10/2026

Serial Number: 80-000867
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.049	0.077	0.197	0.078
0.000	0.049	0.076	0.198	0.079
0.000	0.049	0.076	0.197	0.079
0.000	0.049	0.076	0.198	0.078
0.000	0.049	0.076	0.198	0.078
0.000	0.048	0.076	0.198	0.078
0.000	0.048	0.076	0.198	0.079
0.000	0.049	0.076	0.198	0.079
0.000	0.048	0.076	0.197	0.079
0.000	0.049	0.076	0.198	0.078


Standard Deviations	0.0004	0.0003	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/10/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000867, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000867</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MONROE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/10/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:47</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2026.06.11 09:14:38
-04'00'

06/10/2026

Date

TAYLOR D GUTSCHOW,

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Department Inspector

Service • Integrity • Respect • Quality