

Alcohol Testing Program - Instrument Processing Sheet

Agency: Coral Gables PD Instrument Serial Number: 80-000871

Date In: 8/26/2026 DI Completion Date: N/A ☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake By: <u>TDG</u> Date: <u>9/14/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>9/15/2026</u>		Flow Adjustment By: _____																															
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>218</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP101</u> 32 mm 0.160 (.139-.169) 36 mm 0.171 (.156-.190) 53 mm 0.246 (.228-.278) 103 mm 0.531 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1019</u> Instrument: <u>1017</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																															
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP6286</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP6287</td> <td>202603B</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP6288</td> <td>202603D</td> <td>☐ Other:</td> </tr> <tr> <td>3/4/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG526603</td> <td></td> </tr> <tr> <td>9/23/2027</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP6286	202603A	☐ Battery Replacement	3/3/2028	☐ Dry Gas Regulator Replacement	0.080	MP6287	202603B	☐ Tank Sensor Tare	3/3/2028	☐ Breath Tube Replacement	0.200	MP6288	202603D	☐ Other:	3/4/2028		0.080 DGS	N/A	AG526603		9/23/2027	
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																										
						0.050	MP6286	202603A	☐ Battery Replacement																										
								3/3/2028	☐ Dry Gas Regulator Replacement																										
0.080	MP6287	202603B	☐ Tank Sensor Tare																																
		3/3/2028	☐ Breath Tube Replacement																																
0.200	MP6288	202603D	☐ Other:																																
		3/4/2028																																	
0.080 DGS	N/A	AG526603																																	
		9/23/2027																																	
Optical Bench Adjustment By: _____		Department Inspection By: _____																																	
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: _____																																	
Simulator	Serial #	Lot #	Expiration	Gauge: _____ Instrument: _____																															
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: _____ Exp: _____																															
0.040				Acetone Stock Solution Lot #: _____ Exp: _____																															
0.100				Simulator	Serial Number																														
0.200				0.000																															
0.300				Interferent																															
0.080 DGS	N/A			0.050																															
☐ Post Optical Bench Adjustment Stability Checks				0.080																															
Simulator	Serial #	Lot #	Expiration	0.200																															
0.050				Attachments																															
0.080				☐ Form 41 ☒ Stability Checks ☐ Calibration Certificate ☐ Optical Bench Adjustment ☐ Average Calculations ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☒ Other: Return Material Authorization																															
0.200																																			
0.080 DGS	N/A																																		
Gauge ID #: _____ Gauge: _____ Instrument: _____																																			
Notes/Suggested Service: Does not analyze wet-bath stability checks. (TDG 9/15/26)				☐ Instrument Complies with Chapter 11D-8, FAC ☒ Instrument Does Not Comply with Chapter 11D-8, FAC																															
				☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use																															
				☐ Conduct an Agency Inspection Before Evidentiary Use																															
				Digitally signed by Kaitlyn Spearin Date: 2026.09.18 06:19:11 -04'00' Digitally signed by Isabella Smykowski Date: 2026.09.18 11:57:06 -04'00'																															
				Tech Review	Admin Review																														

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ☒ ≤0.003 of Wet ☒																																																																																																																																																
CORAL GABLES PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000871 09/15/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>15:11</td></tr><tr><td>Control Test</td><td>0.000</td><td>15:12</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:12</td></tr><tr><td>Control Test</td><td>0.000</td><td>15:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:13</td></tr><tr><td>Control Test</td><td>0.000</td><td>15:14</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:14</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0000</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	15:11	Control Test	0.000	15:12	Air Blank	0.000	15:12	Control Test	0.000	15:13	Air Blank	0.000	15:13	Control Test	0.000	15:14	Air Blank	0.000	15:14	Control Test Stats			Average	0.0000		Std Dev	0.0000		Rel Std Dev(%)	0.0000		CORAL GABLES PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000871 09/15/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>15:16</td></tr><tr><td>Control Test</td><td>0.000</td><td>15:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:17</td></tr><tr><td>Control Test</td><td>0.000</td><td>15:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:18</td></tr><tr><td>Control Test</td><td>0.000</td><td>15:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:19</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0000</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	15:16	Control Test	0.000	15:17	Air Blank	0.000	15:17	Control Test	0.000	15:18	Air Blank	0.000	15:18	Control Test	0.000	15:19	Air Blank	0.000	15:19	Control Test Stats			Average	0.0000		Std Dev	0.0000		Rel Std Dev(%)	0.0000		CORAL GABLES PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000871 09/15/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>15:21</td></tr><tr><td>Control Test</td><td>0.000</td><td>15:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:23</td></tr><tr><td>Control Test</td><td>0.000</td><td>15:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:24</td></tr><tr><td>Control Test</td><td>0.000</td><td>15:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:25</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0000</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	15:21	Control Test	0.000	15:22	Air Blank	0.000	15:23	Control Test	0.000	15:23	Air Blank	0.000	15:24	Control Test	0.000	15:24	Air Blank	0.000	15:25	Control Test Stats			Average	0.0000		Std Dev	0.0000		Rel Std Dev(%)	0.0000		CORAL GABLES PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000871 09/15/2026 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>13:25</td></tr><tr><td>Control Test</td><td>0.078</td><td>13:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:26</td></tr><tr><td>Control Test</td><td>0.078</td><td>13:26</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:27</td></tr><tr><td>Control Test</td><td>0.078</td><td>13:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:27</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0780</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	13:25	Control Test	0.078	13:25	Air Blank	0.000	13:26	Control Test	0.078	13:26	Air Blank	0.000	13:27	Control Test	0.078	13:27	Air Blank	0.000	13:27	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:11																																																																																																																																																	
Control Test	0.000	15:12																																																																																																																																																	
Air Blank	0.000	15:12																																																																																																																																																	
Control Test	0.000	15:13																																																																																																																																																	
Air Blank	0.000	15:13																																																																																																																																																	
Control Test	0.000	15:14																																																																																																																																																	
Air Blank	0.000	15:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0000																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:16																																																																																																																																																	
Control Test	0.000	15:17																																																																																																																																																	
Air Blank	0.000	15:17																																																																																																																																																	
Control Test	0.000	15:18																																																																																																																																																	
Air Blank	0.000	15:18																																																																																																																																																	
Control Test	0.000	15:19																																																																																																																																																	
Air Blank	0.000	15:19																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0000																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:21																																																																																																																																																	
Control Test	0.000	15:22																																																																																																																																																	
Air Blank	0.000	15:23																																																																																																																																																	
Control Test	0.000	15:23																																																																																																																																																	
Air Blank	0.000	15:24																																																																																																																																																	
Control Test	0.000	15:24																																																																																																																																																	
Air Blank	0.000	15:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0000																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:25																																																																																																																																																	
Control Test	0.078	13:25																																																																																																																																																	
Air Blank	0.000	13:26																																																																																																																																																	
Control Test	0.078	13:26																																																																																																																																																	
Air Blank	0.000	13:27																																																																																																																																																	
Control Test	0.078	13:27																																																																																																																																																	
Air Blank	0.000	13:27																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Return Material Authorization

Ship to:☒ CMI, Inc.☐ Enforcement ElectronicsShipment to repair facility authorized by: Christopher Naya on 9/17/2026Items Returned: ☒ Instrument ☐ Supplies ☐ Other Describe: _____Instrument Model: Intoxilyzer 8000 Serial Number: 80-000871

Bill To Address:

Coral Gables Police Department
Attn: Christopher Naya

Ship to Address:

Florida Department of Law Enforcement
Fort Myers Regional Operations Center
Attn: Taylor Gutschow
4700 Terminal Drive, Suite 1
Fort Myers, FL 33907

Reason for Return:

Instrument does not analyze wet-bath samples (no bubbles in simulator during sample analysis).

I require an estimate **BEFORE** any repairs will be authorized and/or conductedPlease contact: Christopher Naya Phone #: 1-786-443-6389Email: cnaya2@coralgables.comATP Contact Name: Taylor GutschowATP Email: TaylorGutschow@fdle.state.fl.us