FDLE
Florida Department of
Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Suwannee County SO Instrument Serial Number: 80-000780

Department Inspection:	Agency Inspection/Agency Contact:	Records Audit:
☐ Registration ☐ Request for Registration Attached ☐ Annual ☐ Inspection After Repair ☐ Other: _____ ☐ Barometric Pressure: _____ Gauge _____ Instrument ☐ Instrument Set Up Verified	☐ Agency Inspection Notification ☑ Agency Contact ☐ Other _____ Date of Notification/Contact: 11/27/2018 Agency Inspector: David Falgout	The following records were audited: ☐ Agency Inspection Reports ☐ Breath Test Instrument Repair/Maintenance Records ☐ Instrument Registration ☐ Dry Gas Standard Certificate(s) of Analysis ☐ Other: _____ ☐ Other: _____

Comments:

David Falgout- Suwannee County S.O. (works nights, can call after 6pm) 386-590-3127

His instrument uploaded here during its 2018 annual in November, has not been able to upload at all for him since. Has tried multiple lines, etc. May be modem issue. Is sending it here to be checked, if not then contact him and send to repair if it will not upload from here also.

Instrument/Area:	Equipment:	Supplies:
☐ Clean/Dry ☐ Secure ☐ Limited Access ☐ Other: _____	☐ Proper Number of Simulators ☐ Checked Simulators for Air Leak Resistant Seal and Proper Temperature ☐ Class A Glassware	☐ Distilled/Deionized Water ☐ Mouth Alcohol Solution ☐ Acetone Stock Solution ☐ Alcohol Reference Solution ☐ Dry Gas Standard ☐ Mouth Pieces

Comments:

Action:	Comments:
☐ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC	
☐ Remain in Evidentiary Use ☐ Return to/Place into Evidentiary Use ☐ Remove from Evidentiary Use ☐ Remain Out of Evidentiary Use	
☐ Conduct an Agency Inspection Before Evidentiary Use	

Signature of Alcohol Testing Program Staff Member 11/30 11/28/18 Date

Original – FDLE Copy – Agency

80-000780

SDC

RECEIVED
DEC 10 2018
FDLE
Alcohol Testing Program

Will not upload or keep Date and Time

Thanks

David

Cell # 386 590 3127

Dam

1/8/19