Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: COURT ORD
ARREST TIME: 13:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 05/13/2026 13:48:57

Diagnostics Check		OK	13:42
Air Blank		0.000	13:43
Control Test		0.080	13:43
Air Blank		0.000	13:43
Subject Sample	# 1	0.000	13:44
Breath Volume		1.378	13:44
Air Blank		0.000	13:45
Air Blank		0.000	13:47
Subject Sample	# 2	0.000	13:47
Breath Volume		1.132	13:47
Air Blank		0.000	13:48
Control Test		0.080	13:48
Air Blank		0.000	13:48
Diagnostics Check		OK	13:48

Data Download Date/Time: 06/15/2026 17:13

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: COURT ORD
ARREST TIME: 08:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: CRUZ RODRIGUEZ

DATE/TIME OF TEST: 05/14/2026 09:11:16

Diagnostics Check		OK	09:05
Air Blank		0.000	09:05
Control Test		0.081	09:06
Air Blank		0.000	09:06
Subject Sample	# 1	0.000	09:07
Breath Volume		1.917	09:07
Air Blank		0.000	09:07
Air Blank		0.000	09:09
Subject Sample	# 2	0.000	09:09
Breath Volume		2.292	09:09
Air Blank		0.000	09:10
Control Test		0.080	09:10
Air Blank		0.000	09:11
Diagnostics Check		OK	09:11

Data Download Date/Time: 06/15/2026 17:13

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: COURT ORD
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: RIVERA PENA

DATE/TIME OF TEST: 05/14/2026 13:11:55

Diagnostics Check		OK	13:06
Air Blank		0.000	13:06
Control Test		0.080	13:06
Air Blank		0.000	13:07
Subject Sample	# 1	0.000	13:07
Breath Volume		1.507	13:07
Air Blank		0.000	13:08
Air Blank		0.000	13:10
Subject Sample	# 2	0.000	13:10
Breath Volume		1.816	13:10
Air Blank		0.000	13:11
Control Test		0.081	13:11
Air Blank		0.000	13:11
Diagnostics Check		OK	13:11

Data Download Date/Time: 06/15/2026 17:13

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: COURT ORD
ARREST TIME: 12:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: STEELMAN

DATE/TIME OF TEST: 05/14/2026 13:22:02

Diagnostics Check		OK	13:16
Air Blank		0.000	13:16
Control Test		0.081	13:16
Air Blank		0.000	13:17
Subject Sample	# 1	0.000	13:17
Breath Volume		1.472	13:17
Air Blank		0.000	13:18
Air Blank		0.000	13:20
Subject Sample	# 2	0.000	13:20
Breath Volume		1.640	13:20
Air Blank		0.000	13:21
Control Test		0.080	13:21
Air Blank		0.000	13:21
Diagnostics Check		OK	13:22

Data Download Date/Time: 06/15/2026 17:13

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: COURT ORD
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: IGNACIO MATA

DATE/TIME OF TEST: 05/14/2026 13:32:11

Diagnostics Check		OK	13:26
Air Blank		0.000	13:26
Control Test		0.080	13:26
Air Blank		0.000	13:27
Subject Sample	# 1	0.000	13:28
Breath Volume		1.863	13:28
Air Blank		0.000	13:28
Air Blank		0.000	13:30
Subject Sample	# 2	0.000	13:30
Breath Volume		1.828	13:30
Air Blank		0.000	13:31
Control Test		0.079	13:31
Air Blank		0.000	13:32
Diagnostics Check		OK	13:32

Data Download Date/Time: 06/15/2026 17:13

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: COURT ORD
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: AYEUL

DATE/TIME OF TEST: 05/14/2026 13:42:30

Diagnostics Check		OK	13:36
Air Blank		0.000	13:36
Control Test		0.080	13:37
Air Blank		0.000	13:37
Subject Sample	# 1	0.000	13:38
Breath Volume		1.582	13:38
Air Blank		0.000	13:38
Air Blank		0.000	13:40
Subject Sample	# 2	0.000	13:41
Breath Volume		1.906	13:41
Air Blank		0.000	13:41
Control Test		0.079	13:41
Air Blank		0.000	13:42
Diagnostics Check		OK	13:42

Data Download Date/Time: 06/15/2026 17:13

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: COURT ORD
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 05/19/2026 12:54:21

Diagnostics Check		OK	12:48
Air Blank		0.000	12:48
Control Test		0.080	12:49
Air Blank		0.000	12:49
Subject Sample	# 1	0.000	12:50
Breath Volume		1.378	12:50
Air Blank		0.000	12:50
Air Blank		0.000	12:52
Subject Sample	# 2	0.000	12:53
Breath Volume		1.101	12:53
Air Blank		0.000	12:53
Control Test		0.080	12:53
Air Blank		0.000	12:54
Diagnostics Check		OK	12:54

Data Download Date/Time: 06/15/2026 17:14

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: QUINTANA,STEPHANIE
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: RIVERA-ESTRADA

DATE/TIME OF TEST: 05/22/2026 22:11:07

Diagnostics Check		OK	22:05
Air Blank		0.000	22:05
Control Test		0.081	22:05
Air Blank		0.000	22:06
Subject Sample	# 1	0.133	22:06
Breath Volume		1.328	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.139	22:09
Breath Volume		1.246	22:09
Air Blank		0.000	22:10
Control Test		0.079	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 06/15/2026 17:14

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ZIEGLER,JEFFERY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: CARABALLO MALDONADO

DATE/TIME OF TEST: 05/23/2026 00:03:23

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.080	23:58
Air Blank		0.000	23:58
Subject Sample	# 2	0.210	23:59
Breath Volume		3.152	23:59
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.207	00:01
Breath Volume		1.117	00:01
Air Blank		0.000	00:02
Control Test		0.080	00:02
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 06/15/2026 17:14

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ZIEGLER,JEFFERY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MOHAMMED

DATE/TIME OF TEST: 05/23/2026 03:49:58

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.149	03:45
Breath Volume		1.269	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.143	03:48
Breath Volume		1.332	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 06/15/2026 17:15

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ZIEGLER,JEFFERY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 05/23/2026 21:08:48

Diagnostics Check		OK	21:02
Air Blank		0.000	21:03
Control Test		0.081	21:03
Air Blank		0.000	21:04
Subject Sample	# 1	0.000	21:04
Breath Volume		1.929	21:04
Air Blank		0.000	21:05
Air Blank		0.000	21:07
Subject Sample	# 2	0.000	21:07
Breath Volume		1.660	21:07
Air Blank		0.000	21:07
Control Test		0.080	21:08
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 06/15/2026 17:16

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 05/23/2026 23:12:53

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.079	23:12	
Air Blank		23:12	RFI Detect
Air Blank	0.000	23:12	

Data Download Date/Time: 06/15/2026 18:31

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 05/23/2026 23:18:54

Diagnostics Check		OK	23:14	
Air Blank		0.000	23:14	
Control Test		0.078	23:14	
Air Blank		0.000	23:15	
Subject Sample	# 1		23:17	Subject Test Refused
Breath Volume		0.000	23:17	
Air Blank		0.000	23:18	
Control Test		0.078	23:18	
Air Blank		0.000	23:18	
Diagnostics Check		OK	23:18	

Data Download Date/Time: 06/15/2026 18:31

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/24/2026 00:29:13

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.079	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.302	00:23
Breath Volume		1.105	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.286	00:27
Breath Volume		1.410	00:27
Air Blank		0.000	00:28
Control Test		0.078	00:28
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 06/15/2026 18:31

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ZIEGLER,JEFFERY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: BRAND

DATE/TIME OF TEST: 05/24/2026 02:41:14

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:38	
Control Test		0.079	02:39	
Air Blank		0.000	02:39	
Subject Sample	# 1		02:39	Subject Test Refused
Breath Volume		0.000	02:39	
Air Blank		0.000	02:40	
Control Test		0.079	02:40	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 06/15/2026 18:31

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/24/2026 03:57:12

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.120	03:52
Breath Volume		1.339	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.120	03:55
Breath Volume		1.539	03:55
Air Blank		0.000	03:56
Control Test		0.079	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 06/15/2026 18:31

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ZIEGLER,JEFFREY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/24/2026 22:10:20

Diagnostics Check		OK	22:07	
Air Blank		0.000	22:08	
Control Test		0.080	22:08	
Air Blank		0.000	22:08	
Subject Sample	# 1		22:09	Sequence Aborted
Breath Volume		0.000	22:09	
Air Blank		0.000	22:10	

Data Download Date/Time: 06/15/2026 18:31

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ZIEGLER,JEFFERY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/24/2026 22:29:29

Diagnostics Check		OK	22:23
Air Blank		0.000	22:23
Control Test		0.079	22:23
Air Blank		0.000	22:24
Subject Sample	# 1	0.146	22:24
Breath Volume		1.453	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.138	22:28
Breath Volume		1.246	22:28
Air Blank		0.000	22:28
Control Test		0.079	22:28
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 06/15/2026 18:31

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RANEY,TARRINGTON
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: VELEZ RIVERA

DATE/TIME OF TEST: 05/25/2026 23:08:00

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:56	
Control Test		0.080	22:56	
Air Blank		0.000	22:57	
Subject Sample	# 1	0.211	23:00	Volume Not Met
Breath Volume		0.281	23:00	
Air Blank		0.000	23:00	
Air Blank		0.000	23:02	
Subject Sample	# 2	0.261	23:03	
Breath Volume		1.550	23:03	
Air Blank		0.000	23:03	
Air Blank		0.000	23:05	
Subject Sample	# 3	0.247	23:06	
Breath Volume		1.253	23:06	
Air Blank		0.000	23:07	
Control Test		0.072	23:07	Control Outside Tolerance
Air Blank		0.000	23:08	

Data Download Date/Time: 06/15/2026 18:31

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RANEY,TARRINTON
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: VELEZ RIVERA

DATE/TIME OF TEST: 05/25/2026 23:17:12

Diagnostics Check		OK	23:09
Air Blank		0.000	23:10
Control Test		0.078	23:10
Air Blank		0.000	23:11
Subject Sample	# 1	0.218	23:12
Breath Volume		1.238	23:12
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.215	23:15
Breath Volume		1.175	23:15
Air Blank		0.000	23:16
Control Test		0.078	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 06/15/2026 18:31

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ZIEGLER,JEFFREY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: TORTORA

DATE/TIME OF TEST: 05/27/2026 20:11:21

Diagnostics Check		OK	20:01	
Air Blank		0.000	20:02	
Control Test		0.081	20:02	
Air Blank		0.000	20:02	
Subject Sample	# 1		20:06	No Sample Provided
Breath Volume		0.000	20:06	
Air Blank		0.000	20:06	
Air Blank		0.000	20:08	
Subject Sample	# 2		20:10	Subject Test Refused
Breath Volume		0.000	20:10	
Air Blank		0.000	20:10	
Control Test		0.081	20:10	
Air Blank		0.000	20:11	
Diagnostics Check		OK	20:11	

Data Download Date/Time: 06/15/2026 18:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: COURT ORD
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 05/28/2026 08:33:03

Diagnostics Check		OK	08:27
Air Blank		0.000	08:27
Control Test		0.080	08:28
Air Blank		0.000	08:28
Subject Sample	# 1	0.000	08:28
Breath Volume		1.910	08:28
Air Blank		0.000	08:29
Air Blank		0.000	08:31
Subject Sample	# 2	0.000	08:31
Breath Volume		1.667	08:31
Air Blank		0.000	08:32
Control Test		0.080	08:32
Air Blank		0.000	08:32
Diagnostics Check		OK	08:33

Data Download Date/Time: 06/15/2026 18:32

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: COURT ORD
ARREST TIME: 08:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:59

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 05/28/2026 09:25:45

Diagnostics Check		OK	09:19
Air Blank		0.000	09:20
Control Test		0.081	09:20
Air Blank		0.000	09:21
Subject Sample	# 1	0.000	09:21
Breath Volume		1.273	09:21
Air Blank		0.000	09:22
Air Blank		0.000	09:24
Subject Sample	# 2	0.000	09:24
Breath Volume		1.132	09:24
Air Blank		0.000	09:24
Control Test		0.081	09:25
Air Blank		0.000	09:25
Diagnostics Check		OK	09:25

Data Download Date/Time: 06/15/2026 18:32

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: COURT ORD
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: AYALA DELGADO

DATE/TIME OF TEST: 05/28/2026 14:39:04

Diagnostics Check		OK	14:32
Air Blank		0.000	14:33
Control Test		0.081	14:33
Air Blank		0.000	14:34
Subject Sample	# 1	0.000	14:34
Breath Volume		1.289	14:34
Air Blank		0.000	14:34
Air Blank		0.000	14:36
Subject Sample	# 2	0.000	14:37
Breath Volume		1.554	14:37
Air Blank		0.000	14:38
Control Test		0.081	14:38
Air Blank		0.000	14:38
Diagnostics Check		OK	14:39

Data Download Date/Time: 06/15/2026 18:32

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: ROJAS-CONCEPCION

DATE/TIME OF TEST: 05/29/2026 08:26:44

Diagnostics Check		OK	08:20
Air Blank		0.000	08:21
Control Test		0.080	08:21
Air Blank		0.000	08:22
Subject Sample	# 1	0.000	08:22
Breath Volume		2.109	08:22
Air Blank		0.000	08:23
Air Blank		0.000	08:25
Subject Sample	# 2	0.000	08:25
Breath Volume		2.027	08:25
Air Blank		0.000	08:25
Control Test		0.081	08:26
Air Blank		0.000	08:26
Diagnostics Check		OK	08:26

Data Download Date/Time: 06/15/2026 18:32

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/29/2026 10:22:35

Diagnostics Check		OK	10:16
Air Blank		0.000	10:17
Control Test		0.080	10:17
Air Blank		0.000	10:17
Subject Sample	# 1	0.000	10:18
Breath Volume		1.433	10:18
Air Blank		0.000	10:18
Air Blank		0.000	10:20
Subject Sample	# 2	0.000	10:21
Breath Volume		1.113	10:21
Air Blank		0.000	10:21
Control Test		0.081	10:22
Air Blank		0.000	10:22
Diagnostics Check		OK	10:22

Data Download Date/Time: 06/15/2026 18:32

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:07

SUBJECT'S LAST NAME: VILCHES

DATE/TIME OF TEST: 05/29/2026 10:37:29

Diagnostics Check		OK	10:31
Air Blank		0.000	10:32
Control Test		0.081	10:32
Air Blank		0.000	10:32
Subject Sample	# 1	0.000	10:33
Breath Volume		2.074	10:33
Air Blank		0.000	10:33
Air Blank		0.000	10:35
Subject Sample	# 2	0.000	10:36
Breath Volume		2.148	10:36
Air Blank		0.000	10:36
Control Test		0.080	10:36
Air Blank		0.000	10:37
Diagnostics Check		OK	10:37

Data Download Date/Time: 06/15/2026 18:32

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:47

SUBJECT'S LAST NAME: ESPINAL

DATE/TIME OF TEST: 05/29/2026 11:19:21

Diagnostics Check		OK	11:13
Air Blank		0.000	11:13
Control Test		0.081	11:14
Air Blank		0.000	11:14
Subject Sample	# 1	0.000	11:15
Breath Volume		1.742	11:15
Air Blank		0.000	11:15
Air Blank		0.000	11:17
Subject Sample	# 2	0.000	11:18
Breath Volume		1.691	11:18
Air Blank		0.000	11:18
Control Test		0.081	11:18
Air Blank		0.000	11:19
Diagnostics Check		OK	11:19

Data Download Date/Time: 06/15/2026 18:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4032688835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:42

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 05/29/2026 12:08:07

Diagnostics Check		OK	12:02
Air Blank		0.000	12:02
Control Test		0.080	12:03
Air Blank		0.000	12:03
Subject Sample	# 1	0.000	12:03
Breath Volume		1.519	12:03
Air Blank		0.000	12:04
Air Blank		0.000	12:06
Subject Sample	# 2	0.000	12:06
Breath Volume		1.511	12:06
Air Blank		0.000	12:07
Control Test		0.081	12:07
Air Blank		0.000	12:08
Diagnostics Check		OK	12:08

Data Download Date/Time: 06/15/2026 18:32

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARMON'=,BRANDEN,E
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BAKER,WILLIAM
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: VERSULIEN

DATE/TIME OF TEST: 05/01/2026 20:23:51

Diagnostics Check		OK	20:21	
Air Blank		0.000	20:21	
Control Test		0.078	20:21	
Air Blank		0.000	20:22	
Subject Sample	# 1		20:22	Subject Test Refused
Breath Volume		0.000	20:22	
Air Blank		0.000	20:22	
Control Test		0.079	20:23	
Air Blank		0.000	20:23	
Diagnostics Check		OK	20:23	

Data Download Date/Time: 05/27/2026 10:55

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LEBER,JAMES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/01/2026 22:39:20

Diagnostics Check		OK	22:33
Air Blank		0.000	22:34
Control Test		0.080	22:34
Air Blank		0.000	22:34
Subject Sample	# 1	0.000	22:35
Breath Volume		2.558	22:35
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.000	22:37
Breath Volume		2.246	22:37
Air Blank		0.000	22:38
Control Test		0.077	22:38
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 05/27/2026 10:55

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MILLER,JOHN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MUCHER

DATE/TIME OF TEST: 05/03/2026 01:53:35

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.080	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.167	01:48
Breath Volume		1.605	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.169	01:52
Breath Volume		1.328	01:52
Air Blank		0.000	01:52
Control Test		0.077	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 05/27/2026 10:55

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MOORE,ISRAEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: CHEBAR LOZINSKY

DATE/TIME OF TEST: 05/03/2026 03:00:42

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.087	02:56
Breath Volume		1.664	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.077	02:59
Breath Volume		1.277	02:59
Air Blank		0.000	02:59
Control Test		0.077	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 05/27/2026 10:55

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MEYER,SAMUEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: VETANCOURTQUINTERO

DATE/TIME OF TEST: 05/03/2026 05:55:44

Diagnostics Check		OK	05:42	
Air Blank		0.000	05:43	
Control Test		0.079	05:43	
Air Blank		0.000	05:44	
Subject Sample	# 1	0.167	05:47	Volume Not Met
Breath Volume		0.179	05:47	
Air Blank		0.000	05:48	
Air Blank		0.000	05:49	
Subject Sample	# 2	0.148	05:50	
Breath Volume		1.699	05:50	
Air Blank		0.000	05:51	
Air Blank		0.000	05:53	
Subject Sample	# 3	0.154	05:54	
Breath Volume		1.804	05:54	
Air Blank		0.000	05:54	
Control Test		0.074	05:55	Control Outside Tolerance
Air Blank		0.000	05:55	

Data Download Date/Time: 05/27/2026 10:55

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAPANES,CODY,E
BREATH TEST OPERATOR AGENCY: Alachua PD
ARRESTING OFFICER: TAPANES,CODY
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: MESSER

DATE/TIME OF TEST: 05/03/2026 06:18:58

Diagnostics Check		OK	06:12
Air Blank		0.000	06:13
Control Test		0.079	06:13
Air Blank		0.000	06:14
Subject Sample	# 1	0.133	06:14
Breath Volume		3.250	06:14
Air Blank		0.000	06:15
Air Blank		0.000	06:17
Subject Sample	# 2	0.128	06:17
Breath Volume		3.468	06:17
Air Blank		0.000	06:18
Control Test		0.078	06:18
Air Blank		0.000	06:18
Diagnostics Check		OK	06:18

Data Download Date/Time: 05/27/2026 10:55

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,JORDI,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAEZ,JORDI
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 05/03/2026 23:12:32

Diagnostics Check		OK	23:04
Air Blank		0.000	23:05
Control Test		0.080	23:05
Air Blank		0.000	23:06
Subject Sample	# 1	0.214	23:08
Breath Volume		1.875	23:08
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.220	23:10
Breath Volume		1.996	23:10
Air Blank		0.000	23:11
Control Test		0.079	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 05/27/2026 10:55

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRABHAM,CHARLES,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BRABHAM,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 05/04/2026 21:19:29

Diagnostics Check		OK	21:08	
Air Blank		0.000	21:08	
Control Test		0.080	21:08	
Air Blank		0.000	21:09	
Subject Sample	# 1		21:12	No Sample Provided
Breath Volume		0.000	21:12	
Air Blank		0.000	21:12	
Air Blank		0.000	21:14	
Subject Sample	# 2		21:18	No Sample Provided
Breath Volume		0.000	21:18	
Air Blank		0.000	21:18	
Control Test		0.080	21:18	
Air Blank		0.000	21:19	
Diagnostics Check		OK	21:19	

Data Download Date/Time: 05/27/2026 10:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCWILLIAMS,JACOB,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MCWILLIAMS,JACOB
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: SADDINGTON

DATE/TIME OF TEST: 05/05/2026 22:26:15

Diagnostics Check		OK	22:23	
Air Blank		0.000	22:23	
Control Test		0.079	22:24	
Air Blank		0.000	22:24	
Subject Sample	# 1		22:24	Subject Test Refused
Breath Volume		0.000	22:24	
Air Blank		0.000	22:25	
Control Test		0.079	22:25	
Air Blank		0.000	22:26	
Diagnostics Check		OK	22:26	

Data Download Date/Time: 05/27/2026 10:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MILLER,JOHN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: MULLER

DATE/TIME OF TEST: 05/08/2026 04:23:35

Diagnostics Check		OK	04:17
Air Blank		0.000	04:18
Control Test		0.079	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.066	04:19
Breath Volume		2.667	04:19
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.061	04:22
Breath Volume		1.742	04:22
Air Blank		0.000	04:22
Control Test		0.078	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 05/27/2026 11:00

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,KALEB,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MORRIS,ERIC
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: JACKSON-INGRAM

DATE/TIME OF TEST: 05/11/2026 04:11:08

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.080	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.143	04:06
Breath Volume		1.812	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:08
Subject Sample	# 2	0.144	04:09
Breath Volume		1.761	04:09
Air Blank		0.000	04:10
Control Test		0.076	04:10
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 05/27/2026 11:00

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAMER,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MOORE,ISRAEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/12/2026 02:16:39

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.000	02:12
Breath Volume		1.222	02:12
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.000	02:15
Breath Volume		1.589	02:15
Air Blank		0.000	02:15
Control Test		0.078	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 05/27/2026 11:00

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRANDALL,DOUGLAS,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MURPHY,JAVON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SAHIBDEEN

DATE/TIME OF TEST: 05/15/2026 19:39:39

Diagnostics Check		OK	19:34	
Air Blank		0.000	19:34	
Control Test		0.080	19:34	
Air Blank		0.000	19:35	
Subject Sample	# 1		19:38	Subject Test Refused
Breath Volume		0.000	19:38	
Air Blank		0.000	19:38	
Control Test		0.080	19:39	
Air Blank		0.000	19:39	
Diagnostics Check		OK	19:39	

Data Download Date/Time: 05/27/2026 11:00

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,SKYLER,O
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LAWSON,SKYLER
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 10:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:01

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 05/18/2026 11:26:29

Diagnostics Check		OK	11:21	
Air Blank		0.000	11:22	
Control Test		0.080	11:22	
Air Blank		0.000	11:22	
Subject Sample	# 1		11:24	Subject Test Refused
Breath Volume		0.964	11:24	
Air Blank		0.000	11:25	
Control Test		0.080	11:25	
Air Blank		0.000	11:26	
Diagnostics Check		OK	11:26	

Data Download Date/Time: 05/27/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRABHAM,CHARLES,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GASTON,ALEX
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 05/19/2026 19:48:50

Diagnostics Check		OK	19:40
Air Blank		0.000	19:41
Control Test		0.079	19:41
Air Blank		0.000	19:41
Subject Sample	# 1	0.145	19:43
Breath Volume		1.843	19:43
Air Blank		0.000	19:44
Air Blank		0.000	19:46
Subject Sample	# 2	0.154	19:47
Breath Volume		2.175	19:47
Air Blank		0.000	19:47
Control Test		0.076	19:48
Air Blank		0.000	19:48
Diagnostics Check		OK	19:48

Data Download Date/Time: 05/27/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,KALEB,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MOORE,ISRAEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/23/2026 03:52:39

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.235	03:44
Breath Volume		1.437	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.259	03:47
Breath Volume		1.988	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 3	0.258	03:50
Breath Volume		1.242	03:50
Air Blank		0.000	03:51
Control Test		0.076	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 05/27/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRICKER,JEFFREY,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: FRICKER,JEFFREY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 10:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:58

SUBJECT'S LAST NAME: MCEACHIN

DATE/TIME OF TEST: 05/23/2026 11:26:49

Diagnostics Check		OK	11:20
Air Blank		0.000	11:21
Control Test		0.079	11:21
Air Blank		0.000	11:22
Subject Sample	# 1	0.000	11:22
Breath Volume		1.769	11:22
Air Blank		0.000	11:23
Air Blank		0.000	11:25
Subject Sample	# 2	0.000	11:25
Breath Volume		1.820	11:25
Air Blank		0.000	11:25
Control Test		0.077	11:26
Air Blank		0.000	11:26
Diagnostics Check		OK	11:26

Data Download Date/Time: 05/27/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KITCHEN,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KITCHEN,DAVID
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/25/2026 21:32:57

Diagnostics Check		OK	21:26
Air Blank		0.000	21:26
Control Test		0.080	21:27
Air Blank		0.000	21:27
Subject Sample	# 1	0.000	21:28
Breath Volume		1.121	21:28
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.000	21:31
Breath Volume		1.343	21:31
Air Blank		0.000	21:32
Control Test		0.079	21:32
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 05/27/2026 11:01

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAPANES,CODY,E
BREATH TEST OPERATOR AGENCY: Alachua PD
ARRESTING OFFICER: TAPANES,CODY
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 05/26/2026 23:40:37

Diagnostics Check		OK	23:30	
Air Blank		0.000	23:31	
Control Test		0.079	23:31	
Air Blank		0.000	23:32	
Subject Sample	# 1	0.154	23:35	Volume Not Met
Breath Volume		0.769	23:35	
Air Blank		0.000	23:36	
Air Blank		0.000	23:37	
Subject Sample	# 2		23:38	Subject Test Refused
Breath Volume		0.546	23:38	
Air Blank		0.000	23:39	
Control Test		0.074	23:40	Control Outside Tolerance
Air Blank		0.000	23:40	

Data Download Date/Time: 05/27/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MAHARAJ,KRISHNA
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: CYRUS

DATE/TIME OF TEST: 05/29/2026 04:43:18

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.079	04:36
Air Blank		0.000	04:37
Subject Sample	# 1	0.179	04:38
Breath Volume		1.695	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.179	04:41
Breath Volume		1.769	04:41
Air Blank		0.000	04:42
Control Test		0.077	04:42
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 06/26/2026 14:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEHRENS,JACOB,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GARDNER,TRAVEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/03/2026 04:04:32

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.080	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.187	04:00
Breath Volume		3.343	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.194	04:02
Breath Volume		3.273	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 05/20/2026 06:11

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,TRAVEN,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GARDNER,TRAVEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: GOMEZ AGUILARA

DATE/TIME OF TEST: 05/03/2026 21:51:30

Diagnostics Check		OK	21:46	
Air Blank		0.000	21:47	
Control Test		0.078	21:47	
Air Blank		0.000	21:48	
Subject Sample	# 1		21:50	Subject Test Refused
Breath Volume		0.000	21:50	
Air Blank		0.000	21:50	
Control Test		0.079	21:50	
Air Blank		0.000	21:51	
Diagnostics Check		OK	21:51	

Data Download Date/Time: 05/20/2026 06:11

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCRUGGS,JADEN,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SCRUGGS,JADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 05/05/2026 03:25:56

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.079	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.094	03:21
Breath Volume		1.515	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:23
Subject Sample	# 2	0.100	03:24
Breath Volume		1.570	03:24
Air Blank		0.000	03:25
Control Test		0.078	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 05/20/2026 06:11

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,KYLE,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HOGAN,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: GORDEY

DATE/TIME OF TEST: 05/05/2026 18:05:31

Diagnostics Check		OK	17:54	
Air Blank		0.000	17:54	
Control Test		0.079	17:54	
Air Blank		0.000	17:55	
Subject Sample	# 1	0.000	17:58	Volume Not Met
Breath Volume		0.144	17:58	
Air Blank		0.000	17:59	
Air Blank		0.000	18:01	
Subject Sample	# 2	0.000	18:04	Volume Not Met
Breath Volume		0.429	18:04	
Air Blank		0.000	18:04	
Control Test		0.080	18:04	
Air Blank		0.000	18:05	
Diagnostics Check		OK	18:05	

Data Download Date/Time: 05/20/2026 06:11

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCRUGGS,JADEN,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SCRUGGS,JADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: BROCKWELL

DATE/TIME OF TEST: 05/05/2026 19:22:42

Diagnostics Check		OK	19:19	
Air Blank		0.000	19:20	
Control Test		0.079	19:20	
Air Blank		0.000	19:21	
Subject Sample	# 1		19:21	Subject Test Refused
Breath Volume		0.000	19:21	
Air Blank		0.000	19:21	
Control Test		0.079	19:22	
Air Blank		0.000	19:22	
Diagnostics Check		OK	19:22	

Data Download Date/Time: 05/20/2026 06:11

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCRUGGS,JADEN,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SCRUGGS,JADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: BROCKWELL

DATE/TIME OF TEST: 05/05/2026 19:40:54

Diagnostics Check		OK	19:34
Air Blank		0.000	19:35
Control Test		0.079	19:35
Air Blank		0.000	19:36
Subject Sample	# 1	0.220	19:36
Breath Volume		1.527	19:36
Air Blank		0.000	19:37
Air Blank		0.000	19:39
Subject Sample	# 2	0.210	19:39
Breath Volume		1.988	19:39
Air Blank		0.000	19:39
Control Test		0.077	19:40
Air Blank		0.000	19:40
Diagnostics Check		OK	19:40

Data Download Date/Time: 05/20/2026 06:11

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNGE,JACOB,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RUNGE,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: DINUNZIO

DATE/TIME OF TEST: 05/07/2026 03:47:01

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.079	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.257	03:41
Breath Volume		2.226	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.256	03:45
Breath Volume		2.527	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:47

Data Download Date/Time: 05/20/2026 06:11

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,WESTON,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GRUNDIN,GABREAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: GOMEZ COLON

DATE/TIME OF TEST: 05/09/2026 01:35:08

Diagnostics Check		OK	01:23	
Air Blank		0.000	01:24	
Control Test		0.079	01:24	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:28	No Sample Provided
Breath Volume		0.000	01:28	
Air Blank		0.000	01:28	
Air Blank		0.000	01:30	
Subject Sample	# 2		01:33	No Sample Provided
Breath Volume		0.000	01:33	
Air Blank		0.000	01:34	
Control Test		0.079	01:34	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 05/20/2026 06:11

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,WESTON,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GRUNDIN,GABREAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: GOMEZ COLON

DATE/TIME OF TEST: 05/09/2026 01:43:23

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:38	
Control Test		0.078	01:38	
Air Blank		0.000	01:38	
Subject Sample	# 1	0.174	01:42	Volume Not Met
Breath Volume		0.195	01:42	
Air Blank			01:42	Purge Fail
Air Blank		0.000	01:43	

Data Download Date/Time: 05/20/2026 06:11

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,WESTON,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GRUNDIN,GABREAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: GOMEZ COLON

DATE/TIME OF TEST: 05/09/2026 02:34:42

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.077	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.156	02:29
Breath Volume		2.113	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.160	02:33
Breath Volume		3.222	02:33
Air Blank		0.000	02:33
Control Test		0.079	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 05/20/2026 06:11

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: CHAIRES

DATE/TIME OF TEST: 05/13/2026 01:03:30

Diagnostics Check		OK	00:49	
Air Blank		0.000	00:49	
Control Test		0.079	00:49	
Air Blank		0.000	00:50	
Subject Sample	# 1	0.189	00:50	
Breath Volume		1.835	00:50	
Air Blank		0.000	00:51	
Air Blank		0.000	00:53	
Subject Sample	# 2		00:56	No Sample Provided
Breath Volume		0.000	00:56	
Air Blank		0.000	00:56	
Air Blank		0.000	00:59	
Subject Sample	# 3		01:02	No Sample Provided
Breath Volume		0.000	01:02	
Air Blank		0.000	01:02	
Control Test		0.079	01:02	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 05/20/2026 06:11

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,KYLE,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: TRAVERS,MICHELLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: MCFARLANE

DATE/TIME OF TEST: 05/19/2026 01:23:45

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.079	01:15	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:19	No Sample Provided
Breath Volume		0.000	01:19	
Air Blank		0.000	01:20	
Air Blank		0.000	01:22	
Subject Sample	# 2		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:22	
Control Test		0.079	01:23	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 05/20/2026 06:11

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JASON,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: REGISTER,JASON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 05/19/2026 06:34:44

Diagnostics Check		OK	06:28
Air Blank		0.000	06:29
Control Test		0.079	06:29
Air Blank		0.000	06:30
Subject Sample	# 1	0.000	06:30
Breath Volume		4.722	06:30
Air Blank		0.000	06:30
Air Blank		0.000	06:32
Subject Sample	# 2	0.000	06:33
Breath Volume		4.484	06:33
Air Blank		0.000	06:33
Control Test		0.080	06:34
Air Blank		0.000	06:34
Diagnostics Check		OK	06:34

Data Download Date/Time: 05/20/2026 06:11

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARSEN,ROBERT,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GALANTINE,TAYLOR
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 05/19/2026 13:19:29

Diagnostics Check		OK	13:13
Air Blank		0.000	13:13
Control Test		0.078	13:13
Air Blank		0.000	13:14
Subject Sample	# 1	0.205	13:15
Breath Volume		2.199	13:15
Air Blank		0.000	13:15
Air Blank		0.000	13:17
Subject Sample	# 2	0.206	13:17
Breath Volume		2.871	13:17
Air Blank		0.000	13:18
Control Test		0.078	13:18
Air Blank		0.000	13:19
Diagnostics Check		OK	13:19

Data Download Date/Time: 05/20/2026 06:11

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETWEILER,CONNOR,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DETWEILER,CONNOR
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BOGEN

DATE/TIME OF TEST: 05/22/2026 04:56:19

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.078	04:51
Air Blank		0.000	04:51
Subject Sample	# 1	0.000	04:52
Breath Volume		1.273	04:52
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.000	04:54
Breath Volume		3.121	04:54
Air Blank		0.000	04:55
Control Test		0.079	04:55
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 06/16/2026 02:57

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SISK

DATE/TIME OF TEST: 05/25/2026 00:34:33

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:21	
Control Test		0.078	00:21	
Air Blank		0.000	00:22	
Subject Sample	# 1	0.307	00:24	
Breath Volume		2.222	00:24	
Air Blank		0.000	00:24	
Air Blank		0.000	00:26	
Subject Sample	# 2		00:29	No Sample Provided
Breath Volume		0.000	00:29	
Air Blank		0.000	00:30	
Air Blank		0.000	00:32	
Subject Sample	# 3	0.324	00:32	
Breath Volume		2.542	00:32	
Air Blank		0.000	00:33	
Control Test		0.079	00:33	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 06/16/2026 02:57

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: WAITES

DATE/TIME OF TEST: 05/26/2026 21:43:38

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:35	
Control Test		0.078	21:35	
Air Blank		0.000	21:36	
Subject Sample	# 1		21:39	No Sample Provided
Breath Volume		0.000	21:39	
Air Blank		0.000	21:40	
Air Blank		0.000	21:42	
Subject Sample	# 2		21:42	Subject Test Refused
Breath Volume		0.000	21:42	
Air Blank		0.000	21:42	
Control Test		0.078	21:43	
Air Blank		0.000	21:43	
Diagnostics Check		OK	21:43	

Data Download Date/Time: 06/16/2026 02:57

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: SCAGLIA

DATE/TIME OF TEST: 05/29/2026 00:01:35

Diagnostics Check		OK	23:55
Air Blank		0.000	23:55
Control Test		0.078	23:56
Air Blank		0.000	23:56
Subject Sample	# 2	0.212	23:57
Breath Volume		1.347	23:57
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 1	0.206	00:00
Breath Volume		1.488	00:00
Air Blank		0.000	00:00
Control Test		0.078	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 06/16/2026 02:57

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEWART,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SEWART,WILLIAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 05/30/2026 05:35:07

Diagnostics Check		OK	05:28
Air Blank		0.000	05:29
Control Test		0.079	05:29
Air Blank		0.000	05:30
Subject Sample	# 1	0.108	05:30
Breath Volume		1.347	05:30
Air Blank		0.000	05:31
Air Blank		0.000	05:33
Subject Sample	# 2	0.102	05:33
Breath Volume		1.210	05:33
Air Blank		0.000	05:34
Control Test		0.079	05:34
Air Blank		0.000	05:35
Diagnostics Check		OK	05:35

Data Download Date/Time: 06/16/2026 02:57

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GODFREY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: LIGHTBOURNE

DATE/TIME OF TEST: 05/30/2026 18:46:53

Diagnostics Check		OK	18:32	
Air Blank		0.000	18:32	
Control Test		0.079	18:33	
Air Blank		0.000	18:33	
Subject Sample	# 1	0.249	18:34	
Breath Volume		1.539	18:34	
Air Blank		0.000	18:34	
Air Blank		0.000	18:36	
Subject Sample	# 2		18:39	No Sample Provided
Breath Volume		0.000	18:39	
Air Blank		0.000	18:40	
Air Blank		0.000	18:42	
Subject Sample	# 3		18:45	No Sample Provided
Breath Volume		0.000	18:45	
Air Blank		0.000	18:45	
Control Test		0.079	18:46	
Air Blank		0.000	18:46	
Diagnostics Check		OK	18:46	

Data Download Date/Time: 06/16/2026 02:57

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 05/31/2026 03:33:01

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:30	
Control Test		0.079	03:30	
Air Blank		0.000	03:31	
Subject Sample	# 1		03:31	Subject Test Refused
Breath Volume		0.000	03:31	
Air Blank		0.000	03:32	
Control Test		0.079	03:32	
Air Blank		0.000	03:32	
Diagnostics Check		OK	03:33	

Data Download Date/Time: 06/16/2026 02:57

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEHRENS,JACOB,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VANNIEUWENHOVEN,SCOTT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:42

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/31/2026 08:09:20

Diagnostics Check		OK	08:03
Air Blank		0.000	08:03
Control Test		0.079	08:03
Air Blank		0.000	08:04
Subject Sample	# 1	0.034	08:04
Breath Volume		2.300	08:04
Air Blank		0.000	08:05
Air Blank		0.000	08:07
Subject Sample	# 2	0.032	08:07
Breath Volume		1.730	08:07
Air Blank		0.000	08:08
Control Test		0.078	08:08
Air Blank		0.000	08:09
Diagnostics Check		OK	08:09

Data Download Date/Time: 06/16/2026 02:57

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 04/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 05/05/2026 01:45:25

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.079	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.221	01:40
Breath Volume		1.394	01:40
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.221	01:43
Breath Volume		1.886	01:43
Air Blank		0.000	01:44
Control Test		0.080	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 05/08/2026 18:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,STEPHEN,T
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: MILLER,RAYMOND
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 04/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/06/2026 21:14:13

Diagnostics Check		OK	21:07
Air Blank		0.000	21:08
Control Test		0.079	21:08
Air Blank		0.000	21:08
Subject Sample	# 1	0.262	21:09
Breath Volume		3.113	21:09
Air Blank		0.000	21:10
Air Blank		0.000	21:12
Subject Sample	# 2	0.254	21:12
Breath Volume		2.648	21:12
Air Blank		0.000	21:13
Control Test		0.078	21:13
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 05/08/2026 18:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOURQUE,ESSENCE,B
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: BOURQUE,ESSENCE
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 04/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BUDWELL

DATE/TIME OF TEST: 05/07/2026 04:18:52

Diagnostics Check		OK	04:07	
Air Blank		0.000	04:07	
Control Test		0.080	04:08	
Air Blank		0.000	04:08	
Subject Sample	# 1		04:11	No Sample Provided
Breath Volume		0.000	04:11	
Air Blank		0.000	04:12	
Air Blank		0.000	04:14	
Subject Sample	# 2		04:17	No Sample Provided
Breath Volume		0.000	04:17	
Air Blank		0.000	04:17	
Control Test		0.080	04:18	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:18	

Data Download Date/Time: 05/08/2026 18:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGGAN,WESLEY,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: HAGGAN,WESLEY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 05/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: BAGEANT

DATE/TIME OF TEST: 05/15/2026 22:30:46

Diagnostics Check		OK	22:24
Air Blank		0.000	22:24
Control Test		0.079	22:25
Air Blank		0.000	22:25
Subject Sample	# 1	0.287	22:26
Breath Volume		2.437	22:26
Air Blank		0.000	22:26
Air Blank		0.000	22:28
Subject Sample	# 2	0.295	22:29
Breath Volume		1.605	22:29
Air Blank		0.000	22:29
Control Test		0.078	22:30
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 06/03/2026 13:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 05/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: ARREGUIN

DATE/TIME OF TEST: 05/25/2026 04:43:37

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.000	04:43	RFI Detect
Air Blank	0.000	04:43	

Data Download Date/Time: 06/03/2026 13:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 05/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ARREGUIN

DATE/TIME OF TEST: 05/25/2026 04:57:26

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.080	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.176	04:52
Breath Volume		2.339	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.168	04:55
Breath Volume		1.914	04:55
Air Blank		0.000	04:56
Control Test		0.079	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 06/03/2026 13:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDD,RACHEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUDD,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/01/2026 03:02:17

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.079	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.165	02:57
Breath Volume		2.730	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.159	03:00
Breath Volume		2.085	03:00
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 05/15/2026 17:38

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 05/02/2026 14:47:00

Diagnostics Check		OK	14:41
Air Blank		0.000	14:41
Control Test		0.080	14:41
Air Blank		0.000	14:42
Subject Sample	# 1	0.176	14:42
Breath Volume		3.484	14:42
Air Blank		0.000	14:43
Air Blank		0.000	14:45
Subject Sample	# 2	0.177	14:45
Breath Volume		3.386	14:45
Air Blank		0.000	14:46
Control Test		0.080	14:46
Air Blank		0.000	14:46
Diagnostics Check		OK	14:47

Data Download Date/Time: 05/15/2026 17:38

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLANDA,RICARDO,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HOLANDA,RICARDO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 05/02/2026 16:36:54

Diagnostics Check		OK	16:30
Air Blank		0.000	16:31
Control Test		0.081	16:31
Air Blank		0.000	16:32
Subject Sample	# 1	0.262	16:32
Breath Volume		3.125	16:32
Air Blank		0.000	16:33
Air Blank		0.000	16:35
Subject Sample	# 2	0.263	16:35
Breath Volume		3.382	16:35
Air Blank		0.000	16:35
Control Test		0.080	16:36
Air Blank		0.000	16:36
Diagnostics Check		OK	16:36

Data Download Date/Time: 05/15/2026 17:38

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,DEQWAUN,
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: CALLOWAY,H
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/03/2026 05:42:20

Diagnostics Check		OK	05:34
Air Blank		0.000	05:34
Control Test		0.081	05:34
Air Blank		0.000	05:35
Subject Sample	# 1	0.074	05:35
Breath Volume		1.238	05:35
Air Blank		0.000	05:36
Air Blank		0.000	05:38
Subject Sample	# 2	0.075	05:40
Breath Volume		1.851	05:40
Air Blank		0.000	05:41
Control Test		0.080	05:41
Air Blank		0.000	05:42
Diagnostics Check		OK	05:42

Data Download Date/Time: 05/15/2026 17:38

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAVES,TIMOTHY,F
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: NEMERGUT,JOHN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 05/05/2026 03:14:34

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.081	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.203	03:09
Breath Volume		2.476	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.214	03:13
Breath Volume		2.433	03:13
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 05/15/2026 17:38

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: GORMAN

DATE/TIME OF TEST: 05/07/2026 21:09:58

Diagnostics Check		OK	21:03
Air Blank		0.000	21:03
Control Test		0.079	21:04
Air Blank		0.000	21:04
Subject Sample	# 1	0.336	21:05
Breath Volume		2.984	21:05
Air Blank		0.000	21:05
Air Blank		0.000	21:07
Subject Sample	# 2	0.342	21:08
Breath Volume		2.121	21:08
Air Blank		0.000	21:09
Control Test		0.078	21:09
Air Blank		0.000	21:09
Diagnostics Check		OK	21:09

Data Download Date/Time: 05/15/2026 17:38

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES-GARZA,ERIC,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MORALES-GARZA,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/11/2026 03:06:58

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.081	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.123	03:02
Breath Volume		2.460	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.117	03:05
Breath Volume		3.027	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 05/15/2026 17:38

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: REYNOLDS,JUSTIN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 05/12/2026 17:13:06

Diagnostics Check		OK	17:05
Air Blank		0.000	17:06
Control Test		0.081	17:06
Air Blank		0.000	17:06
Subject Sample	# 1	0.113	17:08
Breath Volume		1.796	17:08
Air Blank		0.000	17:09
Air Blank		0.000	17:11
Subject Sample	# 2	0.113	17:11
Breath Volume		2.199	17:11
Air Blank		0.000	17:12
Control Test		0.080	17:12
Air Blank		0.000	17:13
Diagnostics Check		OK	17:13

Data Download Date/Time: 05/15/2026 17:38

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,JOANNA,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: NEVILLE,TAYLOR
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/13/2026 04:31:43

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:20	
Control Test		0.081	04:21	
Air Blank		0.000	04:21	
Subject Sample	# 1		04:24	No Sample Provided
Breath Volume		0.000	04:24	
Air Blank		0.000	04:25	
Air Blank		0.000	04:27	
Subject Sample	# 2		04:30	No Sample Provided
Breath Volume		0.000	04:30	
Air Blank		0.000	04:30	
Control Test		0.081	04:31	
Air Blank		0.000	04:31	
Diagnostics Check		OK	04:31	

Data Download Date/Time: 05/15/2026 17:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,JOANNA,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: NEVILLE,TAYLOR
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/13/2026 04:54:24

Diagnostics Check		OK	04:48
Air Blank		0.000	04:48
Control Test		0.081	04:48
Air Blank		0.000	04:49
Subject Sample	# 1	0.000	04:49
Breath Volume		1.843	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:52
Subject Sample	# 2	0.000	04:53
Breath Volume		1.207	04:53
Air Blank		0.000	04:53
Control Test		0.081	04:53
Air Blank		0.000	04:54
Diagnostics Check		OK	04:54

Data Download Date/Time: 05/15/2026 17:48

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: OVERTON

DATE/TIME OF TEST: 05/14/2026 21:59:27

Diagnostics Check		OK	21:53
Air Blank		0.000	21:53
Control Test		0.082	21:54
Air Blank		0.000	21:54
Subject Sample	# 1	0.000	21:55
Breath Volume		1.820	21:55
Air Blank		0.000	21:55
Air Blank		0.000	21:57
Subject Sample	# 2	0.000	21:58
Breath Volume		1.148	21:58
Air Blank		0.000	21:58
Control Test		0.081	21:58
Air Blank		0.000	21:59
Diagnostics Check		OK	21:59

Data Download Date/Time: 05/15/2026 17:48

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES-GARZA,ERIC,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MORALES-GARZA,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/18/2026 02:18:54

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.081	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.150	02:13
Breath Volume		1.234	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.156	02:17
Breath Volume		1.417	02:17
Air Blank		0.000	02:18
Control Test		0.081	02:18
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 06/17/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREEMAN,AUSTIN,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FREEMAN,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: MCKINZIE

DATE/TIME OF TEST: 05/19/2026 20:15:41

Diagnostics Check		OK	20:09
Air Blank		0.000	20:09
Control Test		0.081	20:09
Air Blank		0.000	20:10
Subject Sample	# 1	0.201	20:11
Breath Volume		1.679	20:11
Air Blank		0.000	20:11
Air Blank		0.000	20:13
Subject Sample	# 2	0.199	20:14
Breath Volume		1.824	20:14
Air Blank		0.000	20:14
Control Test		0.079	20:15
Air Blank		0.000	20:15
Diagnostics Check		OK	20:15

Data Download Date/Time: 06/17/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,JACK,H
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: WILLIAMS,JUSTIN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: FEICKERT

DATE/TIME OF TEST: 05/21/2026 02:47:50

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.080	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.202	02:41
Breath Volume		1.675	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:43
Subject Sample	# 2	0.221	02:46
Breath Volume		3.003	02:46
Air Blank		0.000	02:46
Control Test		0.080	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 06/17/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTER,DYLAN,F
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HARTER,DYLAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: MCNEIL

DATE/TIME OF TEST: 05/21/2026 22:16:35

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.081	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.000	22:12
Breath Volume		3.265	22:12
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.000	22:15
Breath Volume		3.316	22:15
Air Blank		0.000	22:15
Control Test		0.081	22:15
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 06/17/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,MADISON,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALBRITTON,ROY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: ROY

DATE/TIME OF TEST: 05/23/2026 07:29:41

Diagnostics Check		OK	07:23
Air Blank		0.000	07:24
Control Test		0.080	07:24
Air Blank		0.000	07:24
Subject Sample	# 1	0.110	07:25
Breath Volume		2.324	07:25
Air Blank		0.000	07:25
Air Blank		0.000	07:27
Subject Sample	# 2	0.106	07:28
Breath Volume		2.964	07:28
Air Blank		0.000	07:28
Control Test		0.080	07:29
Air Blank		0.000	07:29
Diagnostics Check		OK	07:29

Data Download Date/Time: 06/17/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,TIMYRON,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: YOUNG,TIMYRON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 07:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:41

SUBJECT'S LAST NAME: BARDALES

DATE/TIME OF TEST: 05/23/2026 08:07:52

Diagnostics Check		OK	08:01
Air Blank		0.000	08:02
Control Test		0.079	08:02
Air Blank		0.000	08:02
Subject Sample	# 1	0.221	08:03
Breath Volume		1.980	08:03
Air Blank		0.000	08:03
Air Blank		0.000	08:05
Subject Sample	# 2	0.223	08:06
Breath Volume		2.152	08:06
Air Blank		0.000	08:06
Control Test		0.079	08:07
Air Blank		0.000	08:07
Diagnostics Check		OK	08:07

Data Download Date/Time: 06/17/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,JACK,H
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: WHITE,JACK
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ASTUDILLO

DATE/TIME OF TEST: 05/27/2026 03:20:15

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.080	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.138	03:15
Breath Volume		3.183	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.135	03:18
Breath Volume		2.281	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 06/17/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT,COLTEN,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VAN PELT,COLTEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: LUTZ

DATE/TIME OF TEST: 05/28/2026 00:02:05

Diagnostics Check		OK	23:55
Air Blank		0.000	23:56
Control Test		0.081	23:56
Air Blank		0.000	23:56
Subject Sample	# 2	0.119	23:57
Breath Volume		2.660	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.121	00:00
Breath Volume		2.609	00:00
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:01
Diagnostics Check		OK	00:02

Data Download Date/Time: 06/17/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEASANTS,STEVEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VAN PELT,COLTEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MOROYOQUI

DATE/TIME OF TEST: 05/29/2026 04:16:33

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.088	04:12
Breath Volume		3.578	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.083	04:15
Breath Volume		2.476	04:15
Air Blank		0.000	04:15
Control Test		0.081	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 06/17/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLCOMBE,JOHN,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MCCLELLAND,GARRETT
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: RICHARD

DATE/TIME OF TEST: 05/29/2026 22:18:36

Diagnostics Check		OK	22:12
Air Blank		0.000	22:13
Control Test		0.080	22:13
Air Blank		0.000	22:13
Subject Sample	# 1	0.172	22:14
Breath Volume		2.316	22:14
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.187	22:17
Breath Volume		3.890	22:17
Air Blank		0.000	22:17
Control Test		0.079	22:18
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 06/17/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT,COLTEN,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PLEASANTS,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: PAYNE JR

DATE/TIME OF TEST: 05/30/2026 03:47:14

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.080	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.194	03:42
Breath Volume		1.648	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.185	03:45
Breath Volume		1.242	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 06/17/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEASANTS,STEVEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VAN PELT,COLTEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 05/31/2026 04:40:23

Diagnostics Check		OK	04:33
Air Blank		0.000	04:34
Control Test		0.080	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.075	04:35
Breath Volume		2.214	04:35
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.079	04:38
Breath Volume		4.226	04:38
Air Blank		0.000	04:39
Control Test		0.081	04:39
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 06/17/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES-GARZA,ERIC,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MORALES-GARZA,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: STANCIL

DATE/TIME OF TEST: 05/31/2026 21:42:36

Diagnostics Check		OK	21:36
Air Blank		0.000	21:36
Control Test		0.081	21:37
Air Blank		0.000	21:37
Subject Sample	# 1	0.195	21:38
Breath Volume		2.398	21:38
Air Blank		0.000	21:38
Air Blank		0.000	21:40
Subject Sample	# 2	0.207	21:41
Breath Volume		1.152	21:41
Air Blank		0.000	21:41
Control Test		0.079	21:42
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 06/17/2026 11:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIFT,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: SWIFT,MICHAEL
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 05/05/2026 23:28:58

Diagnostics Check		OK	23:14	
Air Blank		0.000	23:15	
Control Test		0.079	23:15	
Air Blank		0.000	23:16	
Subject Sample	# 1		23:19	No Sample Provided
Breath Volume		0.000	23:19	
Air Blank		0.000	23:19	
Air Blank		0.000	23:21	
Subject Sample	# 2	0.114	23:23	
Breath Volume		1.917	23:23	
Air Blank		0.000	23:24	
Air Blank		0.000	23:25	
Subject Sample	# 3	0.120	23:27	
Breath Volume		1.652	23:27	
Air Blank		0.000	23:28	
Control Test		0.078	23:28	
Air Blank		0.000	23:28	
Diagnostics Check		OK	23:28	

Data Download Date/Time: 06/02/2026 08:06

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,VICTOR,S
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: SWIFT,MICHAEL
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 05/08/2026 23:59:21

Diagnostics Check		OK	23:56	
Air Blank		0.000	23:56	
Control Test		0.078	23:57	
Air Blank		0.000	23:57	
Subject Sample	# 1		23:57	RFI Detect
Breath Volume		0.000	23:57	
Air Blank		0.000	23:58	
Control Test		0.080	23:58	
Air Blank		0.000	23:59	
Diagnostics Check		OK	23:59	

Data Download Date/Time: 06/02/2026 08:06

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIFT,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: SWIFT,MICHAEL
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 05/09/2026 00:07:46

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.152	00:02
Breath Volume		1.585	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.161	00:06
Breath Volume		3.378	00:06
Air Blank		0.000	00:06
Control Test		0.079	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 06/02/2026 08:06

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LITTLE,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/19/2026 22:02:09

Diagnostics Check		OK	21:58	
Air Blank		0.000	21:59	
Control Test		0.078	21:59	
Air Blank		0.000	22:00	
Subject Sample	# 1	0.136	22:00	
Breath Volume		3.527	22:00	
Air Blank			22:00	RFI Detect
Air Blank		0.000	22:01	
Control Test		0.079	22:01	
Air Blank		0.000	22:02	
Diagnostics Check		OK	22:02	

Data Download Date/Time: 06/02/2026 08:06

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LITTLE,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/19/2026 22:11:32

Diagnostics Check		OK	22:05
Air Blank		0.000	22:05
Control Test		0.077	22:06
Air Blank		0.000	22:06
Subject Sample	# 1	0.131	22:07
Breath Volume		3.406	22:07
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.130	22:09
Breath Volume		2.730	22:09
Air Blank		0.000	22:10
Control Test		0.078	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 06/02/2026 08:06

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONEY,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MALONEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 06/25/2008
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 05/24/2026 05:28:29

Diagnostics Check		OK	05:22
Air Blank		0.000	05:22
Control Test		0.080	05:22
Air Blank		0.000	05:23
Subject Sample	# 1	0.106	05:23
Breath Volume		1.386	05:23
Air Blank		0.000	05:24
Air Blank		0.000	05:26
Subject Sample	# 2	0.104	05:26
Breath Volume		2.027	05:26
Air Blank		0.000	05:27
Control Test		0.079	05:27
Air Blank		0.000	05:28
Diagnostics Check		OK	05:28

Data Download Date/Time: 06/02/2026 08:06

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINNIFIELD,SAM,M
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: MINNIFIELD,SAM
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 12:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: HOLLEY

DATE/TIME OF TEST: 05/09/2026 01:50:38

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.081	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.122	01:46
Breath Volume		3.347	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.109	01:49
Breath Volume		1.984	01:49
Air Blank		0.000	01:49
Control Test		0.081	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 05/28/2026 10:46

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGEN,SAMANTHA,N
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GREEN,JOSUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:06

SUBJECT'S LAST NAME: ECHOLS

DATE/TIME OF TEST: 05/09/2026 13:52:04

Diagnostics Check	OK	13:51	
Air Blank		13:51	RFI Detect
Air Blank		13:52	RFI Detect

Data Download Date/Time: 05/28/2026 10:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGEN,SAMANTHA,N
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GREEN,JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:06

SUBJECT'S LAST NAME: ECHOLS

DATE/TIME OF TEST: 05/09/2026 13:56:02

Diagnostics Check	OK	13:55	
Air Blank		13:55	RFI Detect
Air Blank	0.000	13:56	

Data Download Date/Time: 05/28/2026 10:50

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGEN,SAMANTHA,N
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GREEN,JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:06

SUBJECT'S LAST NAME: ECHOLS

DATE/TIME OF TEST: 05/09/2026 14:04:34

Diagnostics Check		OK	13:57	
Air Blank		0.000	13:58	
Control Test		0.081	13:58	
Air Blank		0.000	13:59	
Subject Sample	# 1	0.248	14:00	
Breath Volume		2.527	14:00	
Air Blank		0.000	14:01	
Air Blank		0.000	14:03	
Subject Sample	# 2	0.245	14:03	
Breath Volume		2.214	14:03	
Air Blank			14:04	RFI Detect
Air Blank			14:04	RFI Detect

Data Download Date/Time: 05/28/2026 10:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGEN,SAMANTHA,N
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GREEN,JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:09

SUBJECT'S LAST NAME: ECHOLS

DATE/TIME OF TEST: 05/09/2026 14:21:32

Diagnostics Check		OK	14:14
Air Blank		0.000	14:15
Control Test		0.079	14:15
Air Blank		0.000	14:15
Subject Sample	# 1	0.245	14:16
Breath Volume		2.449	14:16
Air Blank		0.000	14:17
Air Blank		0.000	14:18
Subject Sample	# 2	0.239	14:19
Breath Volume		2.339	14:19
Air Blank		0.000	14:20
Control Test		0.079	14:20
Air Blank		0.000	14:21
Diagnostics Check		OK	14:21

Data Download Date/Time: 05/28/2026 10:50

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,MICHELLE,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: REEDER,GRACEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 05/10/2026 01:19:49

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:17	
Control Test		0.080	01:17	
Air Blank		0.000	01:18	
Subject Sample	# 1		01:18	Subject Test Refused
Breath Volume		0.000	01:18	
Air Blank		0.000	01:18	
Control Test		0.081	01:19	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 05/28/2026 10:51

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 05/22/2026 21:38:41

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:36	
Control Test		0.081	21:36	
Air Blank		0.000	21:36	
Subject Sample	# 1		21:37	Subject Test Refused
Breath Volume		0.000	21:37	
Air Blank		0.000	21:37	
Control Test		0.081	21:38	
Air Blank		0.000	21:38	
Diagnostics Check		OK	21:38	

Data Download Date/Time: 05/28/2026 10:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGEN,SAMANTHA,N
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GREEN,JOSH
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: TARVER

DATE/TIME OF TEST: 05/23/2026 00:48:34

Diagnostics Check		OK	00:45	
Air Blank		0.000	00:45	
Control Test		0.081	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 1		00:47	Subject Test Refused
Breath Volume		0.000	00:47	
Air Blank		0.000	00:47	
Control Test		0.082	00:47	
Air Blank		0.000	00:48	
Diagnostics Check		OK	00:48	

Data Download Date/Time: 05/28/2026 10:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/23/2026 02:32:05

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.081	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.194	02:27
Breath Volume		3.277	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.194	02:30
Breath Volume		3.781	02:30
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 05/28/2026 10:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEADMON,KAMRYN,L
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: LEADMON,KAMRYN
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 05/23/2026 23:53:13

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:39	
Control Test		0.082	23:39	
Air Blank		0.000	23:40	
Subject Sample	# 1	0.094	23:43	Volume Not Met
Breath Volume		0.445	23:43	
Air Blank		0.000	23:44	
Air Blank		0.000	23:46	
Subject Sample	# 2	0.096	23:47	
Breath Volume		1.933	23:47	
Air Blank		0.000	23:48	
Air Blank		0.000	23:49	
Subject Sample	# 3	0.088	23:51	
Breath Volume		1.898	23:51	
Air Blank		0.000	23:52	
Control Test		0.078	23:52	
Air Blank		0.000	23:53	
Diagnostics Check		OK	23:53	

Data Download Date/Time: 05/28/2026 10:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALIEL,AMANDA,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEADEN,GAGE
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/24/2026 03:06:49

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.081	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.107	03:02
Breath Volume		2.371	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.104	03:05
Breath Volume		2.152	03:05
Air Blank		0.000	03:05
Control Test		0.081	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 05/28/2026 10:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,RYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GONZALVES TEIXEIRA

DATE/TIME OF TEST: 05/27/2026 02:16:28

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.082	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.131	02:11
Breath Volume		2.613	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.132	02:14
Breath Volume		2.628	02:14
Air Blank		0.000	02:15
Control Test		0.081	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 05/28/2026 10:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FISHER,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/21/2026 02:52:05

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.224	02:47
Breath Volume		1.593	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.228	02:50
Breath Volume		2.343	02:50
Air Blank		0.000	02:51
Control Test		0.078	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 06/09/2026 11:04

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MANOR,AUTUMN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: ROUSE

DATE/TIME OF TEST: 05/02/2026 00:45:57

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.082	00:39
Air Blank		0.000	00:39
Subject Sample	# 1	0.179	00:41
Breath Volume		1.964	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.170	00:44
Breath Volume		2.113	00:44
Air Blank		0.000	00:45
Control Test		0.080	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 05/12/2026 19:30

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GULLEDGE,JOHN,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MILLER,DANNY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: HAYWOOD

DATE/TIME OF TEST: 05/02/2026 21:55:38

Diagnostics Check		OK	21:52	
Air Blank		0.000	21:52	
Control Test		0.081	21:53	
Air Blank		0.000	21:53	
Subject Sample	# 1		21:54	Subject Test Refused
Breath Volume		0.000	21:54	
Air Blank		0.000	21:54	
Control Test		0.080	21:55	
Air Blank		0.000	21:55	
Diagnostics Check		OK	21:55	

Data Download Date/Time: 05/12/2026 19:30

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUISE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 05/04/2026 22:26:30

Diagnostics Check		OK	22:23	
Air Blank		0.000	22:23	
Control Test		0.082	22:24	
Air Blank		0.000	22:24	
Subject Sample	# 1		22:25	Subject Test Refused
Breath Volume		0.000	22:25	
Air Blank		0.000	22:25	
Control Test		0.081	22:25	
Air Blank		0.000	22:26	
Diagnostics Check		OK	22:26	

Data Download Date/Time: 05/12/2026 19:30

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAY,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: NAY,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: DATZ

DATE/TIME OF TEST: 05/05/2026 21:55:59

Diagnostics Check		OK	21:46
Air Blank		0.000	21:46
Control Test		0.081	21:47
Air Blank		0.000	21:47
Subject Sample	# 1	0.157	21:49
Breath Volume		3.042	21:49
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.156	21:54
Breath Volume		1.289	21:54
Air Blank		0.000	21:55
Control Test		0.081	21:55
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 05/12/2026 19:30

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUISE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SPICER

DATE/TIME OF TEST: 05/06/2026 03:04:02

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.082	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.215	02:59
Breath Volume		2.296	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:01
Subject Sample	# 2	0.218	03:02
Breath Volume		2.105	03:02
Air Blank		0.000	03:03
Control Test		0.081	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:04

Data Download Date/Time: 05/12/2026 19:30

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JABLONSKI,MARK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JABLONSKI,MARK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: HERRING

DATE/TIME OF TEST: 05/06/2026 04:54:50

Diagnostics Check		OK	04:47
Air Blank		0.000	04:47
Control Test		0.081	04:48
Air Blank		0.000	04:48
Subject Sample	# 1	0.102	04:50
Breath Volume		1.273	04:50
Air Blank		0.000	04:51
Air Blank		0.000	04:52
Subject Sample	# 2	0.104	04:53
Breath Volume		1.339	04:53
Air Blank		0.000	04:53
Control Test		0.079	04:54
Air Blank		0.000	04:54
Diagnostics Check		OK	04:54

Data Download Date/Time: 05/12/2026 19:30

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: POSEY,DYLAN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 04/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: MEEKS

DATE/TIME OF TEST: 05/06/2026 18:32:50

Diagnostics Check		OK	18:24
Air Blank		0.000	18:24
Control Test		0.081	18:25
Air Blank		0.000	18:25
Subject Sample	# 1	0.000	18:26
Breath Volume		1.410	18:26
Air Blank		0.000	18:27
Air Blank		0.000	18:29
Subject Sample	# 2	0.000	18:31
Breath Volume		1.148	18:31
Air Blank		0.000	18:31
Control Test		0.081	18:32
Air Blank		0.000	18:32
Diagnostics Check		OK	18:32

Data Download Date/Time: 05/12/2026 19:30

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,BLAKE,T
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: NEUMANN,JESSE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: COONTS

DATE/TIME OF TEST: 05/07/2026 03:17:32

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.082	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.000	03:13
Breath Volume		2.468	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.000	03:16
Breath Volume		1.742	03:16
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 05/12/2026 19:30

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACKENZIE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SCHULTZ,COBBY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 05/09/2026 00:24:53

Diagnostics Check		OK	00:10	
Air Blank		0.000	00:10	
Control Test		0.081	00:11	
Air Blank		0.000	00:11	
Subject Sample	# 1	0.126	00:14	Volume Not Met
Breath Volume		0.914	00:14	
Air Blank		0.000	00:15	
Air Blank		0.000	00:17	
Subject Sample	# 2	0.150	00:19	
Breath Volume		1.593	00:19	
Air Blank		0.000	00:20	
Air Blank		0.000	00:22	
Subject Sample	# 3	0.141	00:23	
Breath Volume		1.312	00:23	
Air Blank		0.000	00:23	
Control Test		0.079	00:24	
Air Blank		0.000	00:24	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 05/12/2026 19:30

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUISE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: FOUNTAIN

DATE/TIME OF TEST: 05/09/2026 03:47:27

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.077	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.123	03:43
Breath Volume		1.406	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.126	03:45
Breath Volume		1.457	03:45
Air Blank		0.000	03:46
Control Test		0.081	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 05/12/2026 19:30

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAVIS,KAYLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: MORRISSETTE

DATE/TIME OF TEST: 05/10/2026 20:53:25

Diagnostics Check		OK	20:46
Air Blank		0.000	20:47
Control Test		0.081	20:47
Air Blank		0.000	20:48
Subject Sample	# 1	0.230	20:48
Breath Volume		1.210	20:48
Air Blank		0.000	20:49
Air Blank		0.000	20:51
Subject Sample	# 2	0.234	20:51
Breath Volume		1.507	20:51
Air Blank		0.000	20:52
Control Test		0.078	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 05/12/2026 19:30

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SLATER,ETHAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 05/13/2026 22:30:24

Diagnostics Check		OK	22:24
Air Blank		0.000	22:24
Control Test		0.082	22:24
Air Blank		0.000	22:25
Subject Sample	# 1	0.000	22:26
Breath Volume		1.750	22:26
Air Blank		0.000	22:26
Air Blank		0.000	22:28
Subject Sample	# 2	0.000	22:29
Breath Volume		2.199	22:29
Air Blank		0.000	22:29
Control Test		0.082	22:29
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 06/02/2026 20:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACKENZIE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACKENZIE,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 05/14/2026 04:02:19

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.081	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.211	03:57
Breath Volume		1.480	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.206	04:00
Breath Volume		1.214	04:00
Air Blank		0.000	04:01
Control Test		0.081	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 06/02/2026 20:41

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIAUNNA,R
BREATH TEST OPERATOR AGENCY: Milton PD
ARRESTING OFFICER: POSEY,DYLAN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 08:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 05/14/2026 09:40:18

Diagnostics Check		OK	09:36	
Air Blank		0.000	09:36	
Control Test		0.081	09:37	
Air Blank		0.000	09:37	
Subject Sample	# 1		09:38	RFI Detect
Breath Volume		0.000	09:38	
Air Blank		0.000	09:39	
Control Test		0.081	09:39	
Air Blank		0.000	09:40	
Diagnostics Check		OK	09:40	

Data Download Date/Time: 06/02/2026 20:41

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIAUNNA,R
BREATH TEST OPERATOR AGENCY: Milton PD
ARRESTING OFFICER: POSEY,DYLAN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 08:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 05/14/2026 10:07:10

Diagnostics Check		OK	10:00
Air Blank		0.000	10:00
Control Test		0.082	10:01
Air Blank		0.000	10:01
Subject Sample	# 1	0.157	10:02
Breath Volume		1.855	10:02
Air Blank		0.000	10:03
Air Blank		0.000	10:04
Subject Sample	# 2	0.140	10:05
Breath Volume		1.113	10:05
Air Blank		0.000	10:06
Control Test		0.080	10:06
Air Blank		0.000	10:07
Diagnostics Check		OK	10:07

Data Download Date/Time: 06/02/2026 20:41

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,SHELBY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SCHULTZ,COBBY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 05/18/2026 21:19:29

Diagnostics Check		OK	21:13
Air Blank		0.000	21:13
Control Test		0.081	21:13
Air Blank		0.000	21:14
Subject Sample	# 1	0.345	21:14
Breath Volume		3.347	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.355	21:17
Breath Volume		3.632	21:17
Air Blank		0.000	21:18
Control Test		0.080	21:18
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 06/02/2026 20:41

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUISE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WENDELL

DATE/TIME OF TEST: 05/20/2026 03:59:19

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.190	03:54
Breath Volume		2.031	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.188	03:57
Breath Volume		1.812	03:57
Air Blank		0.000	03:58
Control Test		0.082	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 06/02/2026 20:41

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,SHELBY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: IRWIN,KEVIN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: KHUNE

DATE/TIME OF TEST: 05/23/2026 02:01:36

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:58	
Control Test		0.080	01:59	
Air Blank		0.000	01:59	
Subject Sample	# 1		02:00	RFI Detect
Breath Volume		0.000	02:00	
Air Blank		0.000	02:00	
Control Test		0.081	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 06/02/2026 20:41

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,SHELBY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: IRWIN,KEVIN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: KHUNE

DATE/TIME OF TEST: 05/23/2026 02:13:00

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:04	
Control Test		0.081	02:04	
Air Blank		0.000	02:05	
Subject Sample	# 1	0.197	02:06	
Breath Volume		1.535	02:06	
Air Blank		0.000	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 2		02:11	Subject Test Refused
Breath Volume		0.726	02:11	
Air Blank		0.000	02:12	
Control Test		0.080	02:12	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 06/02/2026 20:41

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACKENZIE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SCHULTZ,COBBY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: MENDOZA LANDERO

DATE/TIME OF TEST: 05/23/2026 21:59:25

Diagnostics Check	OK	21:58	
Air Blank		21:58	RFI Detect
Air Blank	0.000	21:59	

Data Download Date/Time: 06/02/2026 20:41

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACKENZIE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SCHULTZ,COBBY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: MENDOZA LANDERO

DATE/TIME OF TEST: 05/23/2026 22:15:19

Diagnostics Check		OK	22:00	
Air Blank		0.000	22:01	
Control Test		0.081	22:01	
Air Blank		0.000	22:02	
Subject Sample	# 1		22:05	No Sample Provided
Breath Volume		0.000	22:05	
Air Blank		0.000	22:05	
Air Blank		0.000	22:07	
Subject Sample	# 2	0.192	22:08	
Breath Volume		1.144	22:08	
Air Blank		0.000	22:09	
Air Blank		0.000	22:11	
Subject Sample	# 3	0.177	22:13	
Breath Volume		1.410	22:13	
Air Blank		0.000	22:14	
Control Test		0.080	22:14	
Air Blank		0.000	22:15	
Diagnostics Check		OK	22:15	

Data Download Date/Time: 06/02/2026 20:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,PATRICK,W
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: STEPHENSON,PATRICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: ABNEY

DATE/TIME OF TEST: 05/23/2026 22:45:27

Diagnostics Check		OK	22:39
Air Blank		0.000	22:39
Control Test		0.079	22:39
Air Blank		0.000	22:40
Subject Sample	# 1	0.161	22:40
Breath Volume		2.191	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.160	22:43
Breath Volume		1.394	22:43
Air Blank		0.000	22:44
Control Test		0.080	22:44
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 06/02/2026 20:41

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,BRIAN,T
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DEVENEY,ETHAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MCKINNON

DATE/TIME OF TEST: 05/24/2026 21:14:53

Diagnostics Check		OK	21:10	
Air Blank		0.000	21:11	
Control Test		0.081	21:11	
Air Blank		0.000	21:12	
Subject Sample	# 1		21:13	Subject Test Refused
Breath Volume		0.000	21:13	
Air Blank		0.000	21:13	
Control Test		0.080	21:14	
Air Blank		0.000	21:14	
Diagnostics Check		OK	21:14	

Data Download Date/Time: 06/02/2026 20:41

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,BRIAN,T
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SLATER,ETHAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: ALLDAY

DATE/TIME OF TEST: 05/24/2026 21:49:35

Diagnostics Check		OK	21:42
Air Blank		0.000	21:42
Control Test		0.081	21:42
Air Blank		0.000	21:43
Subject Sample	# 1	0.175	21:44
Breath Volume		1.285	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.175	21:48
Breath Volume		1.300	21:48
Air Blank		0.000	21:48
Control Test		0.080	21:48
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 06/02/2026 20:41

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,SHELBY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CASEY,SHELBY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HOLLEY

DATE/TIME OF TEST: 05/25/2026 02:10:09

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.081	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.214	02:04
Breath Volume		1.527	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.199	02:08
Breath Volume		1.628	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 06/02/2026 20:42

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/26/2026 22:24:53

Diagnostics Check		OK	22:20	
Air Blank		0.000	22:21	
Control Test		0.081	22:21	
Air Blank		0.000	22:22	
Subject Sample	# 1		22:23	Subject Test Refused
Breath Volume		0.000	22:23	
Air Blank		0.000	22:23	
Control Test		0.082	22:24	
Air Blank		0.000	22:24	
Diagnostics Check		OK	22:24	

Data Download Date/Time: 06/02/2026 20:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: GIGNAC

DATE/TIME OF TEST: 05/30/2026 19:52:53

Diagnostics Check		OK	19:45
Air Blank		0.000	19:45
Control Test		0.081	19:46
Air Blank		0.000	19:46
Subject Sample	# 1	0.168	19:47
Breath Volume		4.535	19:47
Air Blank		0.000	19:48
Air Blank		0.000	19:50
Subject Sample	# 2	0.161	19:51
Breath Volume		3.386	19:51
Air Blank		0.000	19:51
Control Test		0.078	19:52
Air Blank		0.000	19:52
Diagnostics Check		OK	19:52

Data Download Date/Time: 06/02/2026 20:42

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,SHANE,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TAYLOR,SHANE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 05/16/2026 04:42:55

Diagnostics Check		OK	04:36
Air Blank		0.000	04:37
Control Test		0.079	04:37
Air Blank		0.000	04:37
Subject Sample	# 1	0.177	04:38
Breath Volume		1.332	04:38
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.170	04:41
Breath Volume		1.195	04:41
Air Blank		0.000	04:42
Control Test		0.080	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 06/08/2026 07:40

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEDJESKE,JAMES,R
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MOLINA,DANIEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 05/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: HAUGHTEN

DATE/TIME OF TEST: 05/31/2026 03:44:30

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:31	
Control Test		0.079	03:32	
Air Blank		0.000	03:32	
Subject Sample	# 1	0.138	03:35	Volume Not Met
Breath Volume		0.734	03:35	
Air Blank		0.000	03:36	
Air Blank		0.000	03:38	
Subject Sample	# 2	0.131	03:39	
Breath Volume		2.136	03:39	
Air Blank		0.000	03:39	
Air Blank		0.000	03:41	
Subject Sample	# 3	0.129	03:42	
Breath Volume		1.195	03:42	
Air Blank		0.000	03:43	
Control Test		0.077	03:43	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 07/06/2026 00:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNSED,JAXON,D
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: JENKINS,HARDY
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ALSTON

DATE/TIME OF TEST: 05/01/2026 01:44:37

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.078	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.167	01:39
Breath Volume		2.648	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.169	01:43
Breath Volume		3.644	01:43
Air Blank		0.000	01:43
Control Test		0.077	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 06/09/2026 10:29

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,TAUREON,L
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: WATTS,DANIEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 05/01/2026 16:30:25

Diagnostics Check		OK	16:20	
Air Blank		0.000	16:21	
Control Test		0.079	16:21	
Air Blank		0.000	16:22	
Subject Sample	# 1	0.348	16:22	
Breath Volume		1.320	16:22	
Air Blank		0.000	16:23	
Air Blank		0.000	16:25	
Subject Sample	# 2	0.372	16:25	
Breath Volume		2.054	16:25	
Air Blank		0.000	16:26	
Air Blank		0.000	16:28	
Subject Sample	# 3	0.340	16:28	
Breath Volume		2.382	16:28	
Air Blank		0.000	16:29	
Control Test		0.074	16:29	Control Outside Tolerance
Air Blank		0.000	16:30	

Data Download Date/Time: 06/09/2026 10:29

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNSED,JAXON,D
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: JENKINS,HARDY
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: MCKNIGHT

DATE/TIME OF TEST: 05/09/2026 04:38:20

Diagnostics Check		OK	04:31
Air Blank		0.000	04:32
Control Test		0.078	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.213	04:33
Breath Volume		2.781	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.205	04:36
Breath Volume		2.351	04:36
Air Blank		0.000	04:37
Control Test		0.078	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 06/09/2026 10:29

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSEN,ERIC,M
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: HANSEN,ERIC
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 05/17/2026 00:59:03

Diagnostics Check		OK	00:53
Air Blank		0.000	00:53
Control Test		0.079	00:53
Air Blank		0.000	00:54
Subject Sample	# 1	0.101	00:54
Breath Volume		3.175	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.102	00:57
Breath Volume		3.500	00:57
Air Blank		0.000	00:58
Control Test		0.079	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:59

Data Download Date/Time: 06/09/2026 10:29

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSEN,ERIC,M
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: HANSEN,ERIC
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/18/2026 02:07:32

Diagnostics Check	OK	02:06	
Air Blank		02:07	RFI Detect
Air Blank	0.000	02:07	

Data Download Date/Time: 06/09/2026 10:29

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSEN,ERIC,M
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: HANSEN,ERIC
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/18/2026 02:15:28

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.078	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.199	02:10
Breath Volume		3.312	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.197	02:13
Breath Volume		2.480	02:13
Air Blank		0.000	02:14
Control Test		0.076	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 06/09/2026 10:29

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: THOMPSON,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SHAKES

DATE/TIME OF TEST: 05/09/2026 03:47:59

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.078	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.143	03:43
Breath Volume		2.558	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:45
Subject Sample	# 2	0.140	03:46
Breath Volume		2.421	03:46
Air Blank		0.000	03:47
Control Test		0.079	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 06/01/2026 11:42

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: BACKER

DATE/TIME OF TEST: 05/10/2026 18:16:02

Diagnostics Check		OK	18:01	
Air Blank		0.000	18:01	
Control Test		0.081	18:02	
Air Blank		0.000	18:02	
Subject Sample	# 1	0.184	18:05	
Breath Volume		1.625	18:05	
Air Blank		0.000	18:06	
Air Blank		0.000	18:08	
Subject Sample	# 2	0.152	18:11	Volume Not Met
Breath Volume		0.128	18:11	
Air Blank		0.000	18:12	
Air Blank		0.000	18:13	
Subject Sample	# 3		18:14	Subject Test Refused
Breath Volume		0.000	18:14	
Air Blank		0.000	18:15	
Control Test		0.076	18:15	
Air Blank		0.000	18:15	
Diagnostics Check		OK	18:16	

Data Download Date/Time: 06/01/2026 11:42

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 06:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: MARTINA

DATE/TIME OF TEST: 05/23/2026 06:49:30

Diagnostics Check	OK	06:48	
Air Blank	0.000	06:48	
Control Test	0.085	06:48	Interferent Detect
Air Blank	0.000	06:49	

Data Download Date/Time: 06/01/2026 11:42

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 06:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: MARTINA

DATE/TIME OF TEST: 05/23/2026 06:52:49

Diagnostics Check	OK	06:51	
Air Blank	0.000	06:51	
Control Test	0.075	06:52	Interferent Detect
Air Blank	0.000	06:52	

Data Download Date/Time: 06/01/2026 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 06:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: MARTINA

DATE/TIME OF TEST: 05/23/2026 07:01:29

Diagnostics Check		OK	06:54	
Air Blank		0.000	06:54	
Control Test		0.077	06:55	
Air Blank		0.000	06:55	
Subject Sample	# 1	0.066	06:56	
Breath Volume		1.246	06:56	
Air Blank		0.000	06:57	
Air Blank		0.000	06:59	
Subject Sample	# 2		06:59	Interferent Detect
Breath Volume		1.386	06:59	
Air Blank		0.000	07:00	
Control Test		0.079	07:00	
Air Blank		0.000	07:01	
Diagnostics Check		OK	07:01	

Data Download Date/Time: 06/01/2026 11:42

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DALLAS,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,DALLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 05/16/2026 02:40:17

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.080	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.130	02:35
Breath Volume		1.707	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.118	02:38
Breath Volume		1.234	02:38
Air Blank		0.000	02:39
Control Test		0.079	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 06/01/2026 11:38

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE,SEVEN,J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SIPE,SEVEN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: PEACOCK

DATE/TIME OF TEST: 05/22/2026 03:00:07

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.081	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.295	02:54
Breath Volume		2.171	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.282	02:58
Breath Volume		2.171	02:58
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 06/01/2026 11:38

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROOME,PATRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 11:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: JARRETT

DATE/TIME OF TEST: 05/24/2026 01:20:47

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.080	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.193	01:16
Breath Volume		3.847	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:18
Subject Sample	# 2	0.186	01:18
Breath Volume		3.449	01:18
Air Blank		0.000	01:19
Control Test		0.076	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 06/01/2026 11:38

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOME,PATRICK,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: THOMPSON,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 05/24/2026 04:42:53

Diagnostics Check		OK	04:35
Air Blank		0.000	04:36
Control Test		0.079	04:36
Air Blank		0.000	04:37
Subject Sample	# 1	0.160	04:37
Breath Volume		1.296	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.152	04:41
Breath Volume		1.562	04:41
Air Blank		0.000	04:41
Control Test		0.078	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 06/01/2026 11:38

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARDNER,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 05/22/2026 04:10:18

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.000	04:06
Breath Volume		3.277	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.000	04:08
Breath Volume		2.574	04:08
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 06/01/2026 09:36

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARDNER,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BRAVERMAN

DATE/TIME OF TEST: 05/23/2026 03:12:15

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:09	
Control Test		0.080	03:10	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:10	Subject Test Refused
Breath Volume		0.000	03:10	
Air Blank		0.000	03:11	
Control Test		0.079	03:11	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 06/01/2026 09:36

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUX,SAGE,E
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: LOUX,SAGE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: BOBAK

DATE/TIME OF TEST: 05/25/2026 20:53:56

Diagnostics Check		OK	20:51	
Air Blank		0.000	20:51	
Control Test		0.080	20:51	
Air Blank		0.000	20:52	
Subject Sample	# 1		20:52	Subject Test Refused
Breath Volume		0.000	20:52	
Air Blank		0.000	20:53	
Control Test		0.079	20:53	
Air Blank		0.000	20:53	
Diagnostics Check		OK	20:53	

Data Download Date/Time: 06/01/2026 09:36

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARDNER,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: RESULASKA

DATE/TIME OF TEST: 05/31/2026 03:28:00

Diagnostics Check		OK	03:24	
Air Blank		0.000	03:25	
Control Test		0.079	03:25	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:26	Subject Test Refused
Breath Volume		0.000	03:26	
Air Blank		0.000	03:27	
Control Test		0.079	03:27	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 06/01/2026 09:36

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: CARLEY

DATE/TIME OF TEST: 05/06/2026 03:59:02

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.184	03:53
Breath Volume		1.593	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.191	03:57
Breath Volume		2.156	03:57
Air Blank		0.000	03:58
Control Test		0.081	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:59

Data Download Date/Time: 05/06/2026 17:51

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,DALTON,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HOWARD,DALTON
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MCTEIGUE

DATE/TIME OF TEST: 05/06/2026 21:37:56

Diagnostics Check		OK	21:31
Air Blank		0.000	21:32
Control Test		0.081	21:32
Air Blank		0.000	21:33
Subject Sample	# 1	0.000	21:33
Breath Volume		3.371	21:33
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 2	0.000	21:36
Breath Volume		3.292	21:36
Air Blank		0.000	21:37
Control Test		0.081	21:37
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/09/2026 02:40:41

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.081	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.127	02:33
Breath Volume		1.484	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.127	02:38
Breath Volume		1.468	02:38
Air Blank		0.000	02:39
Control Test		0.077	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LONG,LUKE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: DUGGAN

DATE/TIME OF TEST: 05/09/2026 04:36:30

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.079	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.224	04:31
Breath Volume		4.105	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.238	04:34
Breath Volume		4.488	04:34
Air Blank		0.000	04:35
Control Test		0.079	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ETIENNE,EMMANUEL,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ETIENNE,EMMANUEL
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CERON DUQUE

DATE/TIME OF TEST: 05/10/2026 01:07:36

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.081	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.199	01:02
Breath Volume		2.988	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.196	01:05
Breath Volume		3.507	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CHACON,BLAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/10/2026 02:01:46

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.081	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.166	01:57
Breath Volume		3.562	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.167	02:00
Breath Volume		2.605	02:00
Air Blank		0.000	02:00
Control Test		0.080	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIERKOWSKI,CHRISTIAN,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SWIERKOWSKI,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CARBONE

DATE/TIME OF TEST: 05/10/2026 04:16:24

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.081	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.220	04:11
Breath Volume		2.558	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.217	04:14
Breath Volume		1.863	04:14
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CLEVELAND,LAUREN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: TREJO ALMEDAREZ

DATE/TIME OF TEST: 05/11/2026 01:10:29

Diagnostics Check		OK	01:06	
Air Blank		0.000	01:06	
Control Test		0.082	01:06	
Air Blank		0.000	01:07	
Subject Sample	# 1		01:08	Subject Test Refused
Breath Volume		0.246	01:08	
Air Blank		0.000	01:09	
Control Test		0.081	01:09	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: SHIELDS

DATE/TIME OF TEST: 05/11/2026 03:45:36

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.081	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.112	03:41
Breath Volume		3.207	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.106	03:44
Breath Volume		3.152	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DAIGNEAULT,TOBY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 05/13/2026 23:45:31

Diagnostics Check		OK	23:39
Air Blank		0.000	23:39
Control Test		0.081	23:39
Air Blank		0.000	23:40
Subject Sample	# 1	0.112	23:40
Breath Volume		1.785	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.113	23:43
Breath Volume		1.968	23:43
Air Blank		0.000	23:44
Control Test		0.080	23:44
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MORENO,YAIRO
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CASMINSKI

DATE/TIME OF TEST: 05/14/2026 01:28:59

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.081	01:26	
Air Blank		0.000	01:26	
Subject Sample	# 1	0.142	01:27	
Breath Volume		2.414	01:27	
Air Blank			01:27	RFI Detect
Air Blank		0.000	01:28	
Control Test		0.081	01:28	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTENS,ANTHONY,J
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HAMPTON,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: FAJARDO ARIZA

DATE/TIME OF TEST: 05/16/2026 05:25:56

Diagnostics Check		OK	05:13	
Air Blank		0.000	05:14	
Control Test		0.082	05:14	
Air Blank		0.000	05:15	
Subject Sample	# 1	0.166	05:18	Volume Not Met
Breath Volume		0.160	05:18	
Air Blank		0.000	05:19	
Air Blank		0.000	05:20	
Subject Sample	# 2	0.153	05:21	
Breath Volume		1.535	05:21	
Air Blank		0.000	05:22	
Air Blank		0.000	05:23	
Subject Sample	# 3	0.165	05:24	
Breath Volume		1.746	05:24	
Air Blank		0.000	05:25	
Control Test		0.081	05:25	
Air Blank		0.000	05:25	
Diagnostics Check		OK	05:25	

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CASTLE,ALICIA
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: KAY

DATE/TIME OF TEST: 05/18/2026 21:30:16

Diagnostics Check		OK	21:21	
Air Blank		0.000	21:21	
Control Test		0.080	21:22	
Air Blank		0.000	21:22	
Subject Sample	# 1	0.225	21:24	
Breath Volume		1.199	21:24	
Air Blank		0.000	21:24	
Air Blank		0.000	21:26	
Subject Sample	# 2		21:28	Subject Test Refused
Breath Volume		0.519	21:28	
Air Blank		0.000	21:29	
Control Test		0.079	21:29	
Air Blank		0.000	21:30	
Diagnostics Check		OK	21:30	

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLIPIN,AARON,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: KLIPIN,AARON
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 05/23/2026 03:33:08

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.081	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.086	03:27
Breath Volume		2.074	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.087	03:31
Breath Volume		1.937	03:31
Air Blank		0.000	03:32
Control Test		0.081	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CAPUTO,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 05/24/2026 04:53:21

Diagnostics Check		OK	04:47
Air Blank		0.000	04:47
Control Test		0.080	04:47
Air Blank		0.000	04:48
Subject Sample	# 1	0.134	04:48
Breath Volume		2.179	04:48
Air Blank		0.000	04:49
Air Blank		0.000	04:51
Subject Sample	# 2	0.142	04:51
Breath Volume		3.089	04:51
Air Blank		0.000	04:52
Control Test		0.080	04:52
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DALFREY,JORDAN,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: DALFREY,JORDAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: BUI
ARREST TIME: 16:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: SUNDQUIST

DATE/TIME OF TEST: 05/24/2026 18:19:42

Diagnostics Check		OK	18:13
Air Blank		0.000	18:13
Control Test		0.081	18:14
Air Blank		0.000	18:14
Subject Sample	# 1	0.099	18:15
Breath Volume		3.000	18:15
Air Blank		0.000	18:15
Air Blank		0.000	18:17
Subject Sample	# 2	0.098	18:18
Breath Volume		1.445	18:18
Air Blank		0.000	18:18
Control Test		0.081	18:19
Air Blank		0.000	18:19
Diagnostics Check		OK	18:19

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,JANAYSIA,J
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DORVILIEN,GEORGES
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: HARISS

DATE/TIME OF TEST: 05/25/2026 21:19:09

Diagnostics Check		OK	21:12
Air Blank		0.000	21:13
Control Test		0.080	21:13
Air Blank		0.000	21:13
Subject Sample	# 1	0.137	21:14
Breath Volume		1.222	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.150	21:17
Breath Volume		2.929	21:17
Air Blank		0.000	21:18
Control Test		0.080	21:18
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,JANAYSIA,J
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WORTH,DONOVAN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: BOATWRIGHT

DATE/TIME OF TEST: 05/26/2026 23:50:23

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.081	23:45
Air Blank		0.000	23:45
Subject Sample	# 1	0.000	23:46
Breath Volume		2.300	23:46
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.000	23:49
Breath Volume		2.203	23:49
Air Blank		0.000	23:49
Control Test		0.081	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,DALTON,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HOWARD,DALTON
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 05/29/2026 02:30:34

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:24	
Control Test		0.081	02:24	
Air Blank		0.000	02:25	
Subject Sample	# 1	0.089	02:25	
Breath Volume		1.539	02:25	
Air Blank		0.000	02:26	
Air Blank		0.000	02:28	
Subject Sample	# 2	0.090	02:28	
Breath Volume		1.757	02:28	
Air Blank		0.000	02:29	
Control Test		0.081	02:29	
Air Blank			02:30	RFI Detect
Air Blank		0.000	02:30	

Data Download Date/Time: 06/25/2026 18:01

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,DALTON,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: DAIGNEAULT,CODY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/29/2026 02:59:14

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.081	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.000	02:54
Breath Volume		2.324	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.000	02:57
Breath Volume		1.578	02:57
Air Blank		0.000	02:58
Control Test		0.081	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 06/25/2026 18:02

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,JANAYSIA,J
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HAHN,KYLE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: CONTI

DATE/TIME OF TEST: 05/29/2026 21:10:23

Diagnostics Check		OK	21:03
Air Blank		0.000	21:04
Control Test		0.081	21:04
Air Blank		0.000	21:05
Subject Sample	# 1	0.238	21:05
Breath Volume		3.558	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.239	21:08
Breath Volume		3.882	21:08
Air Blank		0.000	21:09
Control Test		0.080	21:09
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 06/25/2026 18:02

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCATO,SHAWN,V
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: BROCATO,SHAWN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: WELMAR OSVALDO

DATE/TIME OF TEST: 05/30/2026 05:17:58

Diagnostics Check		OK	05:11
Air Blank		0.000	05:11
Control Test		0.081	05:12
Air Blank		0.000	05:12
Subject Sample	# 1	0.123	05:13
Breath Volume		3.640	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.126	05:16
Breath Volume		2.695	05:16
Air Blank		0.000	05:17
Control Test		0.079	05:17
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 06/25/2026 18:02

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,ZACHARY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: PFEIFFER,KEVIN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MONTEJO ANTONIO

DATE/TIME OF TEST: 05/01/2026 23:37:36

Diagnostics Check		OK	23:30
Air Blank		0.000	23:31
Control Test		0.082	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.187	23:33
Breath Volume		2.050	23:33
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.192	23:35
Breath Volume		3.140	23:35
Air Blank		0.000	23:36
Control Test		0.081	23:37
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 05/06/2026 17:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMCZYK,NIKKI,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TOMCZYK,NIKKI
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 05/02/2026 03:51:39

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.082	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.080	03:47
Breath Volume		2.246	03:47
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.079	03:50
Breath Volume		2.308	03:50
Air Blank		0.000	03:50
Control Test		0.081	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 05/06/2026 17:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,DALTON,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HOWARD,DALTON
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: KETCHUM

DATE/TIME OF TEST: 05/02/2026 13:28:43

Diagnostics Check		OK	13:22
Air Blank		0.000	13:23
Control Test		0.082	13:23
Air Blank		0.000	13:23
Subject Sample	# 1	0.012	13:24
Breath Volume		4.203	13:24
Air Blank		0.000	13:24
Air Blank		0.000	13:26
Subject Sample	# 2	0.011	13:27
Breath Volume		4.332	13:27
Air Blank		0.000	13:27
Control Test		0.081	13:28
Air Blank		0.000	13:28
Diagnostics Check		OK	13:28

Data Download Date/Time: 05/06/2026 17:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KICKLIGHTER,KEVIN,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: KICKLIGHTER,KEVIN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: BERGREN

DATE/TIME OF TEST: 05/02/2026 20:38:53

Diagnostics Check		OK	20:32	
Air Blank		0.000	20:33	
Control Test		0.082	20:33	
Air Blank		0.000	20:34	
Subject Sample	# 1	0.062	20:34	
Breath Volume		2.011	20:34	
Air Blank		0.000	20:35	
Air Blank		0.000	20:37	
Subject Sample	# 2	0.062	20:37	
Breath Volume		1.441	20:37	
Air Blank		0.000	20:37	
Control Test		0.081	20:38	
Air Blank			20:38	RFI Detect
Air Blank			20:38	RFI Detect

Data Download Date/Time: 05/06/2026 17:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCATO,SHAWN,V
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: BROCATO,SHAWN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: GARCIA PORTILLO

DATE/TIME OF TEST: 05/02/2026 21:42:18

Diagnostics Check		OK	21:30	
Air Blank		0.000	21:31	
Control Test		0.082	21:31	
Air Blank		0.000	21:31	
Subject Sample	# 1	0.116	21:35	Volume Not Met
Breath Volume		0.257	21:35	
Air Blank		0.000	21:35	
Air Blank		0.000	21:37	
Subject Sample	# 2	0.119	21:40	Volume Not Met
Breath Volume		0.156	21:40	
Air Blank		0.000	21:41	
Control Test		0.078	21:41	
Air Blank		0.000	21:42	
Diagnostics Check		OK	21:42	

Data Download Date/Time: 05/06/2026 17:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIANO,GIOVANNI,P
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: CUADRADO,ESTEBAN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:51

SUBJECT'S LAST NAME: HERRERA MORALES

DATE/TIME OF TEST: 05/03/2026 08:38:02

Diagnostics Check	OK	08:37	
Air Blank		08:37	RFI Detect
Air Blank		08:38	RFI Detect

Data Download Date/Time: 05/06/2026 17:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIANO,GIOVANNI,P
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: CUADRADO,ESTEBAN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:51

SUBJECT'S LAST NAME: HERRERA MORALES

DATE/TIME OF TEST: 05/03/2026 08:49:27

Diagnostics Check		OK	08:42
Air Blank		0.000	08:43
Control Test		0.081	08:43
Air Blank		0.000	08:43
Subject Sample	# 1	0.110	08:44
Breath Volume		2.257	08:44
Air Blank		0.000	08:45
Air Blank		0.000	08:47
Subject Sample	# 2	0.107	08:47
Breath Volume		2.339	08:47
Air Blank		0.000	08:48
Control Test		0.081	08:48
Air Blank		0.000	08:49
Diagnostics Check		OK	08:49

Data Download Date/Time: 05/06/2026 17:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,ADAEL,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: CAPUTO,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BOUTCHIE

DATE/TIME OF TEST: 05/05/2026 01:50:39

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:47	
Control Test		0.081	01:47	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:50	Sequence Aborted
Breath Volume		0.000	01:50	
Air Blank		0.000	01:50	

Data Download Date/Time: 05/06/2026 17:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BLACK,EDWARD
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MIGUEL JOAQUIN

DATE/TIME OF TEST: 05/06/2026 01:24:50

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:21	
Control Test		0.082	01:21	
Air Blank		0.000	01:22	
Subject Sample	# 1	0.201	01:23	
Breath Volume		1.777	01:23	
Air Blank			01:24	Purge Fail
Air Blank			01:24	Purge Fail

Data Download Date/Time: 05/06/2026 17:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BLACK,EDWARD
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MIGUEL JOAQUIN

DATE/TIME OF TEST: 05/06/2026 01:29:43

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.031	01:29	Interferent Detect
Air Blank	0.000	01:29	

Data Download Date/Time: 05/06/2026 17:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: CARLEY

DATE/TIME OF TEST: 05/06/2026 02:44:02

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.042	02:43	Control Outside Tolerance
Air Blank	0.000	02:44	

Data Download Date/Time: 05/06/2026 17:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: CARLEY

DATE/TIME OF TEST: 05/06/2026 02:47:17

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.054	02:46	Control Outside Tolerance
Air Blank	0.000	02:47	

Data Download Date/Time: 05/06/2026 17:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: CARLEY

DATE/TIME OF TEST: 05/06/2026 03:03:56

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.055	03:03	Control Outside Tolerance
Air Blank	0.000	03:03	

Data Download Date/Time: 05/06/2026 17:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: CREED,KENDALL
ARRESTING OFFICER AGENCY: Port St. Joe PD
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403476659
DRY GAS EXPIRY DATE: 10/28/2028
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 05/31/2026 02:41:40

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.081	02:35
Air Blank		0.000	02:36
Subject Sample	# 1	0.217	02:37
Breath Volume		1.925	02:37
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.212	02:40
Breath Volume		2.757	02:40
Air Blank		0.000	02:40
Control Test		0.081	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 06/09/2026 14:50

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,SHANE,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TAYLOR,SHANE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: GAINEY

DATE/TIME OF TEST: 05/17/2026 00:24:44

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.078	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.165	00:20
Breath Volume		2.113	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.160	00:23
Breath Volume		2.261	00:23
Air Blank		0.000	00:23
Control Test		0.078	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 05/19/2026 11:25

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERSON,CAMERON,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ROGERSON,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: TADEN

DATE/TIME OF TEST: 05/17/2026 06:20:22

Diagnostics Check		OK	06:09
Air Blank		0.000	06:10
Control Test		0.077	06:10
Air Blank		0.000	06:10
Subject Sample	# 1	0.054	06:13
Breath Volume		1.167	06:13
Air Blank		0.000	06:13
Air Blank		0.000	06:15
Subject Sample	# 2	0.050	06:18
Breath Volume		1.171	06:18
Air Blank		0.000	06:19
Control Test		0.077	06:19
Air Blank		0.000	06:20
Diagnostics Check		OK	06:20

Data Download Date/Time: 06/08/2026 07:43

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 05/01/2026 23:12:22

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.079	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.000	23:07
Breath Volume		2.292	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.000	23:11
Breath Volume		2.093	23:11
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 06/01/2026 12:44

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MORALES,O
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/07/2026 02:58:09

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.078	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.132	02:53
Breath Volume		1.531	02:53
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.125	02:56
Breath Volume		2.152	02:56
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 06/01/2026 12:44

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: GALLIER

DATE/TIME OF TEST: 05/07/2026 23:34:46

Diagnostics Check		OK	23:25
Air Blank		0.000	23:25
Control Test		0.079	23:25
Air Blank		0.000	23:26
Subject Sample	# 1	0.187	23:26
Breath Volume		1.273	23:26
Air Blank		0.000	23:27
Air Blank		0.000	23:29
Subject Sample	# 2	0.216	23:29
Breath Volume		1.769	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 3	0.222	23:33
Breath Volume		1.718	23:33
Air Blank		0.000	23:33
Control Test		0.079	23:34
Air Blank		0.000	23:34
Diagnostics Check		OK	23:34

Data Download Date/Time: 06/01/2026 12:44

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STASIO,J,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LOCKE,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/08/2026 21:46:51

Diagnostics Check		OK	21:40
Air Blank		0.000	21:40
Control Test		0.079	21:41
Air Blank		0.000	21:41
Subject Sample	# 1	0.000	21:42
Breath Volume		1.851	21:42
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.000	21:45
Breath Volume		1.855	21:45
Air Blank		0.000	21:45
Control Test		0.080	21:46
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 06/01/2026 12:44

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PODRAZA,BRIAN,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GREER,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: BORMAN

DATE/TIME OF TEST: 05/09/2026 15:02:18

Diagnostics Check		OK	14:55
Air Blank		0.000	14:56
Control Test		0.080	14:56
Air Blank		0.000	14:57
Subject Sample	# 1	0.147	14:57
Breath Volume		1.324	14:57
Air Blank		0.000	14:58
Air Blank		0.000	15:00
Subject Sample	# 2	0.149	15:00
Breath Volume		2.031	15:00
Air Blank		0.000	15:01
Control Test		0.078	15:01
Air Blank		0.000	15:02
Diagnostics Check		OK	15:02

Data Download Date/Time: 06/01/2026 12:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,JAMES,B
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KELLY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 05/10/2026 02:23:47

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.154	02:19
Breath Volume		1.656	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.163	02:22
Breath Volume		2.015	02:22
Air Blank		0.000	02:22
Control Test		0.079	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 06/01/2026 12:44

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:36

SUBJECT'S LAST NAME: VERGARA POMPA

DATE/TIME OF TEST: 05/11/2026 10:05:47

Diagnostics Check		OK	09:58
Air Blank		0.000	09:59
Control Test		0.078	09:59
Air Blank		0.000	09:59
Subject Sample	# 1	0.242	10:01
Breath Volume		1.710	10:01
Air Blank		0.000	10:01
Air Blank		0.000	10:03
Subject Sample	# 2	0.239	10:04
Breath Volume		1.347	10:04
Air Blank		0.000	10:04
Control Test		0.078	10:05
Air Blank		0.000	10:05
Diagnostics Check		OK	10:05

Data Download Date/Time: 06/01/2026 12:44

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,EDELISA,N
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: VEGA,SASHA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: NEWLON

DATE/TIME OF TEST: 05/16/2026 05:31:17

Diagnostics Check		OK	05:28	
Air Blank		0.000	05:28	
Control Test		0.079	05:29	
Air Blank		0.000	05:29	
Subject Sample	# 1		05:29	Subject Test Refused
Breath Volume		0.000	05:29	
Air Blank		0.000	05:30	
Control Test		0.080	05:30	
Air Blank		0.000	05:31	
Diagnostics Check		OK	05:31	

Data Download Date/Time: 06/01/2026 12:44

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRO,TYLER,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FERRO,TYLER
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: HUFSTEDLER-BELSER

DATE/TIME OF TEST: 05/16/2026 20:39:55

Diagnostics Check		OK	20:32
Air Blank		0.000	20:33
Control Test		0.079	20:33
Air Blank		0.000	20:34
Subject Sample	# 1	0.153	20:34
Breath Volume		1.328	20:34
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.138	20:38
Breath Volume		1.625	20:38
Air Blank		0.000	20:38
Control Test		0.079	20:39
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 06/01/2026 12:44

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRO,TYLER,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FERRO,TYLER
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 05/20/2026 18:28:00

Diagnostics Check		OK	18:25	
Air Blank		0.000	18:25	
Control Test		0.079	18:25	
Air Blank		0.000	18:26	
Subject Sample	# 1		18:26	Subject Test Refused
Breath Volume		0.000	18:26	
Air Blank		0.000	18:27	
Control Test		0.079	18:27	
Air Blank		0.000	18:27	
Diagnostics Check		OK	18:28	

Data Download Date/Time: 06/01/2026 12:44

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: STORY

DATE/TIME OF TEST: 05/21/2026 23:05:11

Diagnostics Check		OK	23:01	
Air Blank		0.000	23:01	
Control Test		0.079	23:02	
Air Blank		0.000	23:02	
Subject Sample	# 1		23:03	Subject Test Refused
Breath Volume		0.000	23:03	
Air Blank		0.000	23:04	
Control Test		0.079	23:04	
Air Blank		0.000	23:05	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 06/01/2026 12:44

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MORALES,O
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: GRANDSHAW

DATE/TIME OF TEST: 05/22/2026 18:18:50

Diagnostics Check		OK	18:15	
Air Blank		0.000	18:16	
Control Test		0.079	18:16	
Air Blank		0.000	18:17	
Subject Sample	# 1		18:17	Subject Test Refused
Breath Volume		0.000	18:17	
Air Blank		0.000	18:17	
Control Test		0.080	18:18	
Air Blank		0.000	18:18	
Diagnostics Check		OK	18:18	

Data Download Date/Time: 06/01/2026 12:45

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOCKE,J,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LOCKE,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: MULCONREY

DATE/TIME OF TEST: 05/22/2026 18:36:39

Diagnostics Check		OK	18:28	
Air Blank		0.000	18:28	
Control Test		0.079	18:28	
Air Blank		0.000	18:29	
Subject Sample	# 1		18:32	No Sample Provided
Breath Volume		0.000	18:32	
Air Blank		0.000	18:33	
Air Blank		0.000	18:35	
Subject Sample	# 2		18:35	Subject Test Refused
Breath Volume		0.000	18:35	
Air Blank		0.000	18:35	
Control Test		0.079	18:36	
Air Blank		0.000	18:36	
Diagnostics Check		OK	18:36	

Data Download Date/Time: 06/01/2026 12:45

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MCCLUSKY

DATE/TIME OF TEST: 05/22/2026 23:38:18

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.079	23:36	
Air Blank		0.000	23:36	
Subject Sample	# 1		23:36	Subject Test Refused
Breath Volume		0.000	23:36	
Air Blank		0.000	23:37	
Control Test		0.079	23:37	
Air Blank		0.000	23:38	
Diagnostics Check		OK	23:38	

Data Download Date/Time: 06/01/2026 12:45

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MONTALVO MENDEZ

DATE/TIME OF TEST: 05/26/2026 23:58:56

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:47	
Control Test		0.079	23:48	
Air Blank		0.000	23:48	
Subject Sample	# 1	0.095	23:51	Volume Not Met
Breath Volume		0.558	23:51	
Air Blank		0.000	23:52	
Air Blank		0.000	23:54	
Subject Sample	# 2	0.072	23:57	Volume Not Met
Breath Volume		0.019	23:57	
Air Blank		0.000	23:58	
Control Test		0.079	23:58	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 06/01/2026 12:45

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,J,B
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KELLY,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/29/2026 03:02:09

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.115	02:57
Breath Volume		1.566	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.117	03:00
Breath Volume		1.972	03:00
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 06/01/2026 12:45

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FABY,ROBERT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 05/29/2026 23:07:52

Diagnostics Check		OK	23:02	
Air Blank		0.000	23:02	
Control Test		0.079	23:03	
Air Blank		0.000	23:03	
Subject Sample	# 1		23:06	Subject Test Refused
Breath Volume		0.285	23:06	
Air Blank		0.000	23:06	
Control Test		0.079	23:07	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 06/01/2026 12:45

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,STACY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FABY,ROBERT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: FERRO

DATE/TIME OF TEST: 05/31/2026 20:42:32

Diagnostics Check		OK	20:26	
Air Blank		0.000	20:26	
Control Test		0.079	20:26	
Air Blank		0.000	20:27	
Subject Sample	# 1	0.074	20:30	Volume Not Met
Breath Volume		0.175	20:30	
Air Blank		0.000	20:31	
Air Blank		0.000	20:32	
Subject Sample	# 2	0.090	20:35	
Breath Volume		1.304	20:35	
Air Blank		0.000	20:35	
Air Blank		0.000	20:37	
Subject Sample	# 3	0.072	20:40	Volume Not Met
Breath Volume		0.425	20:40	
Air Blank		0.000	20:41	
Control Test		0.078	20:41	
Air Blank		0.000	20:42	
Diagnostics Check		OK	20:42	

Data Download Date/Time: 06/01/2026 12:45

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSELL,TRISTAN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MUNSTER,CODY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 251683
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: GROVES

DATE/TIME OF TEST: 05/04/2026 22:22:29

Diagnostics Check		OK	22:19	
Air Blank		0.000	22:20	
Control Test		0.080	22:20	
Air Blank		0.000	22:20	
Subject Sample	# 1		22:21	Subject Test Refused
Breath Volume		0.000	22:21	
Air Blank		0.000	22:21	
Control Test		0.079	22:21	
Air Blank		0.000	22:22	
Diagnostics Check		OK	22:22	

Data Download Date/Time: 05/30/2026 09:15

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BINEGAR,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 251683
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: DODGE

DATE/TIME OF TEST: 05/07/2026 18:50:02

Diagnostics Check		OK	18:43
Air Blank		0.000	18:43
Control Test		0.079	18:44
Air Blank		0.000	18:44
Subject Sample	# 1	0.176	18:45
Breath Volume		2.417	18:45
Air Blank		0.000	18:46
Air Blank		0.000	18:47
Subject Sample	# 2	0.173	18:48
Breath Volume		2.476	18:48
Air Blank		0.000	18:49
Control Test		0.079	18:49
Air Blank		0.000	18:49
Diagnostics Check		OK	18:50

Data Download Date/Time: 05/30/2026 09:15

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JAYCOB,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HOLLOWAY,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 251683
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: LOVATO

DATE/TIME OF TEST: 05/08/2026 00:20:13

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.079	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.368	00:15
Breath Volume		1.933	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.364	00:18
Breath Volume		2.097	00:18
Air Blank		0.000	00:19
Control Test		0.079	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 05/30/2026 09:15

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JAYCOB,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WADDELL,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 251683
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: CARTA

DATE/TIME OF TEST: 05/08/2026 01:46:14

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:43	
Control Test		0.079	01:44	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:45	
Control Test		0.079	01:45	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 05/30/2026 09:15

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSELL,TRISTAN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: REDWAN,ALFREDO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 251683
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 05/10/2026 04:34:49

Diagnostics Check		OK	04:27
Air Blank		0.000	04:27
Control Test		0.079	04:27
Air Blank		0.000	04:28
Subject Sample	# 1	0.082	04:28
Breath Volume		1.195	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.078	04:33
Breath Volume		1.332	04:33
Air Blank		0.000	04:33
Control Test		0.078	04:34
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 05/30/2026 09:16

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JAYCOB,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WILSON,T
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 251683
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 05/17/2026 21:03:03

Diagnostics Check		OK	20:56
Air Blank		0.000	20:56
Control Test		0.079	20:57
Air Blank		0.000	20:57
Subject Sample	# 1	0.225	20:58
Breath Volume		3.570	20:58
Air Blank		0.000	20:59
Air Blank		0.000	21:00
Subject Sample	# 2	0.214	21:01
Breath Volume		2.433	21:01
Air Blank		0.000	21:02
Control Test		0.078	21:02
Air Blank		0.000	21:02
Diagnostics Check		OK	21:03

Data Download Date/Time: 05/30/2026 09:16

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNSEY,ROBERT,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LYONS,C
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 251683
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: MANCUSO

DATE/TIME OF TEST: 05/21/2026 01:13:33

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.079	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.213	01:08
Breath Volume		3.480	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.215	01:11
Breath Volume		2.980	01:11
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 05/30/2026 09:16

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,PEDRO,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SURIEL,E
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 04/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 251683
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 05/27/2026 17:23:06

Diagnostics Check		OK	17:20	
Air Blank		0.000	17:20	
Control Test		0.079	17:21	
Air Blank		0.000	17:21	
Subject Sample	# 1		17:21	Subject Test Refused
Breath Volume		0.000	17:21	
Air Blank		0.000	17:22	
Control Test		0.079	17:22	
Air Blank		0.000	17:23	
Diagnostics Check		OK	17:23	

Data Download Date/Time: 05/30/2026 09:16

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: WERNER

DATE/TIME OF TEST: 05/03/2026 05:35:34

Diagnostics Check		OK	05:32	
Air Blank		0.000	05:33	
Control Test		0.078	05:33	
Air Blank		0.000	05:33	
Subject Sample	# 1		05:34	Subject Test Refused
Breath Volume		0.000	05:34	
Air Blank		0.000	05:34	
Control Test		0.078	05:34	
Air Blank		0.000	05:35	
Diagnostics Check		OK	05:35	

Data Download Date/Time: 06/03/2026 14:18

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: COBIN

DATE/TIME OF TEST: 05/06/2026 01:29:51

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.078	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.000	01:24
Breath Volume		1.210	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.000	01:28
Breath Volume		1.457	01:28
Air Blank		0.000	01:28
Control Test		0.078	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 06/03/2026 14:18

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: BLOMQUIST

DATE/TIME OF TEST: 05/07/2026 00:58:58

Diagnostics Check		OK	00:55	
Air Blank		0.000	00:56	
Control Test		0.078	00:56	
Air Blank		0.000	00:57	
Subject Sample	# 1		00:57	Subject Test Refused
Breath Volume		0.000	00:57	
Air Blank		0.000	00:58	
Control Test		0.078	00:58	
Air Blank		0.000	00:58	
Diagnostics Check		OK	00:58	

Data Download Date/Time: 06/03/2026 14:18

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,DALE,E
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: TITUS,JOSHUA
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 09:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: MCCLELLAN

DATE/TIME OF TEST: 05/12/2026 10:54:18

Diagnostics Check		OK	10:51	
Air Blank		0.000	10:51	
Control Test		0.078	10:51	
Air Blank		0.000	10:52	
Subject Sample	# 1		10:52	Subject Test Refused
Breath Volume		0.000	10:52	
Air Blank		0.000	10:53	
Control Test		0.078	10:53	
Air Blank			10:53	RFI Detect
Air Blank		0.000	10:54	

Data Download Date/Time: 06/03/2026 14:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: TOXEN

DATE/TIME OF TEST: 05/12/2026 20:06:47

Diagnostics Check		OK	20:03	
Air Blank		0.000	20:04	
Control Test		0.078	20:04	
Air Blank		0.000	20:05	
Subject Sample	# 1		20:05	Subject Test Refused
Breath Volume		0.000	20:05	
Air Blank		0.000	20:05	
Control Test		0.078	20:06	
Air Blank		0.000	20:06	
Diagnostics Check		OK	20:06	

Data Download Date/Time: 06/03/2026 14:18

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEAL,JESSE,R
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KHUU,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BROGDON

DATE/TIME OF TEST: 05/15/2026 22:16:30

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.078	22:11
Air Blank		0.000	22:11
Subject Sample	# 1	0.217	22:12
Breath Volume		1.523	22:12
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.219	22:14
Breath Volume		1.441	22:14
Air Blank		0.000	22:15
Control Test		0.084	22:15
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 06/03/2026 14:18

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEAL,JESSE,R
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GETE,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LUPIEN

DATE/TIME OF TEST: 05/15/2026 22:45:07

Diagnostics Check		OK	22:40	
Air Blank		0.000	22:40	
Control Test		0.078	22:41	
Air Blank		0.000	22:41	
Subject Sample	# 1		22:43	Subject Test Refused
Breath Volume		0.000	22:43	
Air Blank		0.000	22:44	
Control Test		0.078	22:44	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 06/03/2026 14:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: ALDRIDGE

DATE/TIME OF TEST: 05/17/2026 01:49:03

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.079	01:42
Air Blank		0.000	01:43
Subject Sample	# 1	0.236	01:44
Breath Volume		2.214	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.237	01:47
Breath Volume		2.261	01:47
Air Blank		0.000	01:48
Control Test		0.078	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:49

Data Download Date/Time: 06/03/2026 14:18

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JEFFERY,P
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BALUFAS,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: VEGA HERNANDEZ

DATE/TIME OF TEST: 05/18/2026 00:54:04

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:43	
Control Test		0.078	00:43	
Air Blank		0.000	00:43	
Subject Sample	# 1		00:47	No Sample Provided
Breath Volume		0.000	00:47	
Air Blank		0.000	00:47	
Air Blank		0.000	00:49	
Subject Sample	# 2		00:52	No Sample Provided
Breath Volume		0.000	00:52	
Air Blank		0.000	00:53	
Control Test		0.078	00:53	
Air Blank		0.000	00:53	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 06/03/2026 14:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JEFFERY,P
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BALAFUS,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1982508A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/18/2026 01:05:58

Diagnostics Check		OK	00:57
Air Blank		0.000	00:58
Control Test		0.078	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.202	01:00
Breath Volume		1.187	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.190	01:04
Breath Volume		1.277	01:04
Air Blank		0.000	01:05
Control Test		0.078	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 06/03/2026 14:19

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLMAN JR,DAVID,E
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: SMINIMER,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/20/2026 12:57:40

Diagnostics Check		OK	12:51
Air Blank		0.000	12:51
Control Test		0.078	12:51
Air Blank		0.000	12:52
Subject Sample	# 1	0.029	12:53
Breath Volume		2.183	12:53
Air Blank		0.000	12:53
Air Blank		0.000	12:55
Subject Sample	# 2	0.028	12:56
Breath Volume		2.136	12:56
Air Blank		0.000	12:56
Control Test		0.077	12:57
Air Blank		0.000	12:57
Diagnostics Check		OK	12:57

Data Download Date/Time: 06/03/2026 14:19

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOVANOV,MICHAEL,N
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: JOVANOV,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: THURMOND

DATE/TIME OF TEST: 05/23/2026 00:26:23

Diagnostics Check		OK	00:19
Air Blank		0.000	00:20
Control Test		0.078	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.096	00:21
Breath Volume		3.859	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.096	00:24
Breath Volume		2.742	00:24
Air Blank		0.000	00:25
Control Test		0.078	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 06/03/2026 14:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOVANOV,MICHAEL,N
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: JOVANOV,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: BUI
ARREST TIME: 17:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: RIPPY

DATE/TIME OF TEST: 05/23/2026 19:27:31

Diagnostics Check		OK	19:21
Air Blank		0.000	19:21
Control Test		0.078	19:21
Air Blank		0.000	19:22
Subject Sample	# 1	0.118	19:22
Breath Volume		1.984	19:22
Air Blank		0.000	19:23
Air Blank		0.000	19:25
Subject Sample	# 2	0.127	19:26
Breath Volume		1.902	19:26
Air Blank		0.000	19:26
Control Test		0.077	19:26
Air Blank		0.000	19:27
Diagnostics Check		OK	19:27

Data Download Date/Time: 06/03/2026 14:19

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERAFINI,JOSHUA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TWITTY,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 05/24/2026 18:23:22

Diagnostics Check		OK	18:11	
Air Blank		0.000	18:12	
Control Test		0.077	18:12	
Air Blank		0.000	18:13	
Subject Sample	# 1		18:16	No Sample Provided
Breath Volume		0.000	18:16	
Air Blank		0.000	18:16	
Air Blank		0.000	18:18	
Subject Sample	# 2		18:21	No Sample Provided
Breath Volume		0.000	18:21	
Air Blank		0.000	18:22	
Control Test		0.078	18:22	
Air Blank		0.000	18:23	
Diagnostics Check		OK	18:23	

Data Download Date/Time: 06/03/2026 14:19

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: HESS,DREW
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: MOYNIHAN

DATE/TIME OF TEST: 05/25/2026 22:41:36

Diagnostics Check		OK	22:38	
Air Blank		0.000	22:39	
Control Test		0.078	22:39	
Air Blank		0.000	22:40	
Subject Sample	# 1		22:40	Subject Test Refused
Breath Volume		0.000	22:40	
Air Blank		0.000	22:40	
Control Test		0.078	22:40	
Air Blank		0.000	22:41	
Diagnostics Check		OK	22:41	

Data Download Date/Time: 06/03/2026 14:19

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PACKWOOD

DATE/TIME OF TEST: 05/26/2026 04:00:20

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.078	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.151	03:55
Breath Volume		1.867	03:55
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.159	03:58
Breath Volume		2.683	03:58
Air Blank		0.000	03:59
Control Test		0.078	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 06/03/2026 14:19

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/26/2026 22:42:40

Diagnostics Check		OK	22:39	
Air Blank		0.000	22:40	
Control Test		0.078	22:40	
Air Blank		0.000	22:41	
Subject Sample	# 1		22:41	Subject Test Refused
Breath Volume		0.000	22:41	
Air Blank		0.000	22:41	
Control Test		0.078	22:42	
Air Blank		0.000	22:42	
Diagnostics Check		OK	22:42	

Data Download Date/Time: 06/03/2026 14:19

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: COBIN

DATE/TIME OF TEST: 05/30/2026 00:27:38

Diagnostics Check		OK	00:22	
Air Blank		0.000	00:22	
Control Test		0.079	00:22	
Air Blank		0.000	00:23	
Subject Sample	# 1	0.000	00:23	
Breath Volume		1.949	00:23	
Air Blank		0.000	00:24	
Air Blank			00:26	RFI Detect
Air Blank		0.000	00:26	
Control Test		0.079	00:26	
Air Blank			00:27	RFI Detect
Air Blank		0.000	00:27	

Data Download Date/Time: 06/03/2026 14:19

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: COBIN

DATE/TIME OF TEST: 05/30/2026 00:49:23

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.079	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.000	00:45
Breath Volume		2.929	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.000	00:48
Breath Volume		2.625	00:48
Air Blank		0.000	00:48
Control Test		0.077	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 06/03/2026 14:19

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLMAN JR,DAVID,E
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WASYLYK,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 05/30/2026 18:26:34

Diagnostics Check		OK	18:20
Air Blank		0.000	18:20
Control Test		0.079	18:21
Air Blank		0.000	18:21
Subject Sample	# 1	0.000	18:22
Breath Volume		2.566	18:22
Air Blank		0.000	18:22
Air Blank		0.000	18:24
Subject Sample	# 2	0.000	18:25
Breath Volume		2.996	18:25
Air Blank		0.000	18:25
Control Test		0.079	18:25
Air Blank		0.000	18:26
Diagnostics Check		OK	18:26

Data Download Date/Time: 06/03/2026 14:19

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JEFFERY,P
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DAVIS,B
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 198250801A
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/30/2026 23:25:20

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.078	23:18
Air Blank		0.000	23:19
Subject Sample	# 1	0.203	23:20
Breath Volume		1.726	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.202	23:23
Breath Volume		1.519	23:23
Air Blank		0.000	23:24
Control Test		0.078	23:24
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 06/03/2026 14:19

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLMAN JR,DAVID,E
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DAVIS,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 11:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:18

SUBJECT'S LAST NAME: MASEK

DATE/TIME OF TEST: 05/31/2026 12:41:33

Diagnostics Check		OK	12:38	
Air Blank		0.000	12:38	
Control Test		0.077	12:39	
Air Blank		0.000	12:39	
Subject Sample	# 1		12:40	Subject Test Refused
Breath Volume		0.000	12:40	
Air Blank		0.000	12:40	
Control Test		0.077	12:40	
Air Blank		0.000	12:41	
Diagnostics Check		OK	12:41	

Data Download Date/Time: 06/03/2026 14:19

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOVANOV,MICHAEL,N
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: JOVANOV,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: RAUTENKRANZ

DATE/TIME OF TEST: 05/31/2026 22:51:12

Diagnostics Check		OK	22:44
Air Blank		0.000	22:45
Control Test		0.077	22:45
Air Blank		0.000	22:46
Subject Sample	# 1	0.098	22:46
Breath Volume		1.410	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.112	22:49
Breath Volume		2.675	22:49
Air Blank		0.000	22:50
Control Test		0.077	22:50
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 06/03/2026 14:19

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: DRIGGERS,DRIGGERS
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 05/16/2026 22:05:55

Diagnostics Check		OK	21:55
Air Blank		0.000	21:55
Control Test		0.080	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.130	21:59
Breath Volume		1.589	21:59
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.125	22:04
Breath Volume		1.812	22:04
Air Blank		0.000	22:05
Control Test		0.080	22:05
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 06/05/2026 05:55

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 05/25/2026 21:14:24

Diagnostics Check		OK	21:10	
Air Blank		0.000	21:11	
Control Test		0.079	21:11	
Air Blank		0.000	21:12	
Subject Sample	# 1		21:14	RFI Detect
Breath Volume		0.054	21:14	
Air Blank			21:14	RFI Detect
Air Blank			21:14	RFI Detect

Data Download Date/Time: 06/05/2026 05:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 05/25/2026 21:16:32

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:16	
		21:16	RFI Detect
Air Blank		21:16	RFI Detect

Data Download Date/Time: 06/05/2026 05:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BENOIT

DATE/TIME OF TEST: 05/27/2026 00:41:09

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:36	
Control Test		0.079	00:36	
Air Blank		0.000	00:37	
Subject Sample	# 1	0.105	00:37	
Breath Volume		1.425	00:37	
Air Blank		0.000	00:38	
Air Blank			00:39	RFI Detect
Air Blank		0.000	00:40	
Control Test		0.079	00:40	
Air Blank			00:40	RFI Detect
Air Blank		0.000	00:41	

Data Download Date/Time: 06/05/2026 05:55

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BENOIT

DATE/TIME OF TEST: 05/27/2026 00:42:58

Diagnostics Check	OK	00:42	
Air Blank		00:42	RFI Detect
Air Blank		00:42	RFI Detect

Data Download Date/Time: 06/05/2026 05:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 05/30/2026 03:08:54

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:06	
Control Test		0.080	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:07	RFI Detect
Breath Volume		0.000	03:07	
Air Blank		0.000	03:07	
Control Test		0.080	03:08	
Air Blank		0.000	03:08	
Diagnostics Check		OK	03:08	

Data Download Date/Time: 06/05/2026 05:55

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 05/30/2026 03:11:44

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.082	03:10	
Air Blank	0.000	03:11	
		03:11	RFI Detect
Air Blank		03:11	RFI Detect

Data Download Date/Time: 06/05/2026 05:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMAGGIO,MATTHEW,R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: DEMAGGIO,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:52

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/04/2026 18:20:24

Diagnostics Check		OK	18:13
Air Blank		0.000	18:13
Control Test		0.080	18:13
Air Blank		0.000	18:14
Subject Sample	# 1	0.125	18:15
Breath Volume		2.570	18:15
Air Blank		0.000	18:16
Air Blank		0.000	18:18
Subject Sample	# 2	0.135	18:18
Breath Volume		1.605	18:18
Air Blank		0.000	18:19
Control Test		0.079	18:19
Air Blank		0.000	18:20
Diagnostics Check		OK	18:20

Data Download Date/Time: 06/11/2026 21:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: PEREZ PASCUAL

DATE/TIME OF TEST: 05/09/2026 01:12:20

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.222	01:07
Breath Volume		4.347	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.219	01:10
Breath Volume		4.308	01:10
Air Blank		0.000	01:11
Control Test		0.077	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 06/11/2026 21:55

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELTE,BRENDEN,M
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: ROBBINS,CLIM
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: ROSA CRUZ

DATE/TIME OF TEST: 05/02/2026 04:35:27

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.079	04:29
Air Blank		0.000	04:29
Subject Sample	# 1	0.099	04:30
Breath Volume		2.578	04:30
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.095	04:33
Breath Volume		1.691	04:33
Air Blank		0.000	04:34
Control Test		0.079	04:34
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 05/26/2026 09:51

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELTE,BRENDEN,M
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: ROBBINS,CLIM
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/06/2026 22:55:17

Diagnostics Check		OK	22:48
Air Blank		0.000	22:49
Control Test		0.080	22:49
Air Blank		0.000	22:50
Subject Sample	# 1	0.259	22:50
Breath Volume		3.523	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.244	22:53
Breath Volume		2.472	22:53
Air Blank		0.000	22:54
Control Test		0.079	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 05/26/2026 09:51

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: OWENS,D
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 05/07/2026 20:29:17

Diagnostics Check		OK	20:25	
Air Blank		0.000	20:26	
Control Test		0.080	20:26	
Air Blank		0.000	20:27	
Subject Sample	# 1		20:27	Subject Test Refused
Breath Volume		0.000	20:27	
Air Blank		0.000	20:28	
Control Test		0.080	20:28	
Air Blank		0.000	20:29	
Diagnostics Check		OK	20:29	

Data Download Date/Time: 05/26/2026 09:51

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUM,SAVANNAH,K
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: FLOURONY,WESLEY
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: NICHOLSON

DATE/TIME OF TEST: 05/08/2026 02:07:39

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:02	
Control Test		0.079	02:02	
Air Blank		0.000	02:03	
Subject Sample	# 1		02:06	Subject Test Refused
Breath Volume		0.000	02:06	
Air Blank		0.000	02:06	
Control Test		0.080	02:07	
Air Blank		0.000	02:07	
Diagnostics Check		OK	02:07	

Data Download Date/Time: 05/26/2026 09:51

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN SWEET,B,
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: ALLEN SWEET,B
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/10/2026 02:38:46

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:36	
Control Test		0.079	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 1		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:37	
Control Test		0.079	02:38	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 05/26/2026 09:51

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISH,ALLYSSA,J
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: OBERHOLZER,BRIAN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:26

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/11/2026 10:53:12

Diagnostics Check		OK	10:47
Air Blank		0.000	10:47
Control Test		0.080	10:47
Air Blank		0.000	10:48
Subject Sample	# 1	0.150	10:48
Breath Volume		3.554	10:48
Air Blank		0.000	10:49
Air Blank		0.000	10:51
Subject Sample	# 2	0.145	10:51
Breath Volume		5.054	10:51
Air Blank		0.000	10:52
Control Test		0.080	10:52
Air Blank		0.000	10:53
Diagnostics Check		OK	10:53

Data Download Date/Time: 05/26/2026 09:51

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,ROBERT,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BRUINSMA,SETH
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: HEREDIA

DATE/TIME OF TEST: 05/12/2026 06:23:25

Diagnostics Check		OK	06:16
Air Blank		0.000	06:17
Control Test		0.080	06:17
Air Blank		0.000	06:18
Subject Sample	# 1	0.022	06:18
Breath Volume		2.539	06:18
Air Blank		0.000	06:19
Air Blank		0.000	06:21
Subject Sample	# 2	0.019	06:21
Breath Volume		1.304	06:21
Air Blank		0.000	06:22
Control Test		0.079	06:22
Air Blank		0.000	06:23
Diagnostics Check		OK	06:23

Data Download Date/Time: 05/26/2026 09:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILDEMAN-STUART,NATASH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: ALEZONO WEBB

DATE/TIME OF TEST: 05/13/2026 01:28:40

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:17	
Control Test		0.080	01:17	
Air Blank		0.000	01:18	
Subject Sample	# 1	0.069	01:21	Volume Not Met
Breath Volume		0.050	01:21	
Air Blank		0.000	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 2	0.078	01:26	Volume Not Met
Breath Volume		0.406	01:26	
Air Blank		0.000	01:27	
Control Test		0.079	01:28	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 05/26/2026 09:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILDEMAN-STUART,NATASH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: ALEZONO WEBB

DATE/TIME OF TEST: 05/13/2026 01:39:07

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.077	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.080	01:34
Breath Volume		2.136	01:34
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.075	01:37
Breath Volume		1.984	01:37
Air Blank		0.000	01:38
Control Test		0.077	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 05/26/2026 09:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: MENNECHY,LORENA
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 05/13/2026 22:08:02

Diagnostics Check		OK	21:59	
Air Blank		0.000	21:59	
Control Test		0.079	22:00	
Air Blank		0.000	22:00	
Subject Sample	# 1		22:03	No Sample Provided
Breath Volume		0.000	22:03	
Air Blank		0.000	22:04	
Air Blank		0.000	22:06	
Subject Sample	# 2		22:06	Subject Test Refused
Breath Volume		0.000	22:06	
Air Blank		0.000	22:07	
Control Test		0.079	22:07	
Air Blank		0.000	22:07	
Diagnostics Check		OK	22:08	

Data Download Date/Time: 05/26/2026 09:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: DOUGLAS,JORDAN
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 05/13/2026 22:56:30

Diagnostics Check		OK	22:49
Air Blank		0.000	22:49
Control Test		0.079	22:49
Air Blank		0.000	22:50
Subject Sample	# 1	0.236	22:51
Breath Volume		1.476	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.242	22:54
Breath Volume		1.179	22:54
Air Blank		0.000	22:55
Control Test		0.079	22:55
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 05/26/2026 09:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,SIAN,A
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: PITMAN,W
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: RIVERA-MANTILLA

DATE/TIME OF TEST: 05/15/2026 00:39:38

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:33
Subject Sample	# 1	0.207	00:35
Breath Volume		2.703	00:35
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.194	00:38
Breath Volume		1.691	00:38
Air Blank		0.000	00:38
Control Test		0.080	00:39
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 05/26/2026 09:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEFFLER,GAVIN,W
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: ROBBINS,CLIM
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MAPLE

DATE/TIME OF TEST: 05/16/2026 21:48:17

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:41	
Control Test		0.080	21:42	
Air Blank		0.000	21:42	
Subject Sample	# 1	0.214	21:45	Volume Not Met
Breath Volume		0.792	21:45	
Air Blank			21:46	Improper Sample
Air Blank		0.000	21:47	
Control Test		0.081	21:47	
Air Blank		0.000	21:48	
Diagnostics Check		OK	21:48	

Data Download Date/Time: 05/26/2026 09:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEFFLER,GAVIN,W
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: ROBBINS,CLIM
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MAPLE

DATE/TIME OF TEST: 05/16/2026 22:08:39

Diagnostics Check		OK	21:54	
Air Blank		0.000	21:55	
Control Test		0.078	21:55	
Air Blank		0.000	21:55	
Subject Sample	# 1	0.193	21:57	
Breath Volume		1.445	21:57	
Air Blank		0.000	21:58	
Air Blank		0.000	22:00	
Subject Sample	# 2	0.195	22:03	Volume Not Met
Breath Volume		0.230	22:03	
Air Blank		0.000	22:04	
Air Blank		0.000	22:05	
Subject Sample	# 3	0.184	22:07	
Breath Volume		1.382	22:07	
Air Blank		0.000	22:07	
Control Test		0.072	22:08	Control Outside Tolerance
Air Blank		0.000	22:08	

Data Download Date/Time: 05/26/2026 09:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEFFLER,GAVIN,W
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: CHOUDRI,HASHAAM
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BERNADIN

DATE/TIME OF TEST: 05/17/2026 02:14:53

Diagnostics Check		OK	02:12	
Air Blank		0.000	02:12	
Control Test		0.080	02:12	
Air Blank		0.000	02:13	
Subject Sample	# 1		02:13	Subject Test Refused
Breath Volume		0.000	02:13	
Air Blank		0.000	02:13	
Control Test		0.079	02:14	
Air Blank		0.000	02:14	
Diagnostics Check		OK	02:14	

Data Download Date/Time: 05/26/2026 09:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,STEPHEN,V
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: COK,STEVEN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/17/2026 04:55:49

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:53	
Control Test		0.080	04:53	
Air Blank		0.000	04:54	
Subject Sample	# 1		04:54	Subject Test Refused
Breath Volume		0.000	04:54	
Air Blank		0.000	04:54	
Control Test		0.080	04:55	
Air Blank		0.000	04:55	
Diagnostics Check		OK	04:55	

Data Download Date/Time: 05/26/2026 09:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERDT,ANDRU,J
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: ECKERDT,ANDRU
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SKOLITS

DATE/TIME OF TEST: 05/17/2026 21:57:51

Diagnostics Check		OK	21:51
Air Blank		0.000	21:52
Control Test		0.079	21:52
Air Blank		0.000	21:53
Subject Sample	# 1	0.000	21:53
Breath Volume		1.945	21:53
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.000	21:56
Breath Volume		1.800	21:56
Air Blank		0.000	21:56
Control Test		0.079	21:57
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 05/26/2026 09:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA JR,KEVIN,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: DUQUE,JOSHUA
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: HERNANDEZ JIMENEZ

DATE/TIME OF TEST: 05/19/2026 00:12:00

Diagnostics Check		OK	00:09	
Air Blank		0.000	00:09	
Control Test		0.080	00:09	
Air Blank		0.000	00:10	
Subject Sample	# 1		00:10	Subject Test Refused
Breath Volume		0.000	00:10	
Air Blank		0.000	00:11	
Control Test		0.079	00:11	
Air Blank		0.000	00:11	
Diagnostics Check		OK	00:12	

Data Download Date/Time: 05/26/2026 09:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN SWEET,B,
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: ALLEN SWEET,B
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: DAVENPORT

DATE/TIME OF TEST: 05/20/2026 00:54:46

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.080	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.153	00:50
Breath Volume		1.792	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:52
Subject Sample	# 2	0.154	00:53
Breath Volume		4.375	00:53
Air Blank		0.000	00:53
Control Test		0.077	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 05/26/2026 09:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURTZLAFF,SYDNEE,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: AKINS,GARY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 05/20/2026 01:08:47

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.078	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.105	01:04
Breath Volume		2.242	01:04
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.103	01:07
Breath Volume		2.261	01:07
Air Blank		0.000	01:07
Control Test		0.079	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 05/26/2026 09:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEFFLER,GAVIN,W
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: PETTIS,JESSE
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: GARREAU

DATE/TIME OF TEST: 05/21/2026 22:25:20

Diagnostics Check		OK	22:17
Air Blank		0.000	22:18
Control Test		0.080	22:18
Air Blank		0.000	22:19
Subject Sample	# 1	0.007	22:20
Breath Volume		1.140	22:20
Air Blank		0.000	22:20
Air Blank		0.000	22:22
Subject Sample	# 2	0.000	22:23
Breath Volume		1.976	22:23
Air Blank		0.000	22:24
Control Test		0.080	22:24
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 05/26/2026 09:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURTZLAFF,SYDNEE,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BURTZLAFF,SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: SHERWOOD

DATE/TIME OF TEST: 05/22/2026 01:28:55

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.079	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.197	01:24
Breath Volume		2.089	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.197	01:27
Breath Volume		2.152	01:27
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 05/26/2026 09:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRAY,CHANTAL,P
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: MURRAY,CHANTAL
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: CONTRERAS TARELO

DATE/TIME OF TEST: 05/23/2026 00:55:42

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.080	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.134	00:51
Breath Volume		2.316	00:51
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.129	00:54
Breath Volume		2.074	00:54
Air Blank		0.000	00:54
Control Test		0.079	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 05/26/2026 09:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRAY,CHANTAL,P
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: SULLIVAN,ALLISON
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: GONZALEZ GONZALEZ

DATE/TIME OF TEST: 05/23/2026 21:47:09

Diagnostics Check		OK	21:39
Air Blank		0.000	21:39
Control Test		0.080	21:40
Air Blank		0.000	21:40
Subject Sample	# 1	0.067	21:41
Breath Volume		1.546	21:41
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.066	21:45
Breath Volume		1.867	21:45
Air Blank		0.000	21:46
Control Test		0.080	21:46
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 05/26/2026 09:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKFIELD,ZACHARY,T
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: TUCKFIELD,ZACHARY
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 251683
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: BUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: DESKE

DATE/TIME OF TEST: 05/24/2026 19:36:26

Diagnostics Check		OK	19:33	
Air Blank		0.000	19:34	
Control Test		0.080	19:34	
Air Blank		0.000	19:34	
Subject Sample	# 1		19:35	Subject Test Refused
Breath Volume		0.000	19:35	
Air Blank		0.000	19:35	
Control Test		0.080	19:35	
Air Blank		0.000	19:36	
Diagnostics Check		OK	19:36	

Data Download Date/Time: 05/26/2026 09:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRAY,CHANTAL,P
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: MURRAY,CHANTAL
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: EASTMAN

DATE/TIME OF TEST: 05/25/2026 00:42:08

Diagnostics Check		OK	00:33	
Air Blank		0.000	00:33	
Control Test		0.079	00:34	
Air Blank		0.000	00:34	
Subject Sample	# 1		00:37	No Sample Provided
Breath Volume		0.000	00:37	
Air Blank		0.000	00:38	
Air Blank		0.000	00:40	
Subject Sample	# 2		00:40	Subject Test Refused
Breath Volume		0.000	00:40	
Air Blank		0.000	00:41	
Control Test		0.079	00:41	
Air Blank		0.000	00:42	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 05/26/2026 09:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES,MONIQUE,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BARNES,MONIQUE
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 05/25/2026 04:33:05

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.080	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.138	04:28
Breath Volume		1.132	04:28
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.127	04:31
Breath Volume		4.085	04:31
Air Blank		0.000	04:32
Control Test		0.080	04:32
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 05/26/2026 09:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: CARDOSO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: GARCIA GONZALEZ

DATE/TIME OF TEST: 05/26/2026 02:01:00

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:55	
Control Test		0.080	01:56	
Air Blank		0.000	01:56	
Subject Sample	# 1	0.173	01:57	
Breath Volume		2.796	01:57	
Air Blank		0.000	01:58	
Air Blank			01:59	RFI Detect
Air Blank		0.000	02:00	
Control Test		0.082	02:00	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 05/26/2026 09:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: CARDOSO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: GARCIA GONZALEZ

DATE/TIME OF TEST: 05/26/2026 02:06:17

Diagnostics Check		OK	02:03	
Air Blank		0.000	02:03	
Control Test		0.080	02:04	
Air Blank		0.000	02:04	
Subject Sample	# 1		02:04	Subject Test Refused
Breath Volume		0.000	02:04	
Air Blank		0.000	02:05	
Control Test		0.079	02:05	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 05/26/2026 09:54

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA JR,KEVIN,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: GARCIA JR,K
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 05/27/2026 23:29:09

Diagnostics Check		OK	23:21
Air Blank		0.000	23:21
Control Test		0.080	23:22
Air Blank		0.000	23:22
Subject Sample	# 1	0.110	23:24
Breath Volume		1.320	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:26
Subject Sample	# 2	0.114	23:27
Breath Volume		1.472	23:27
Air Blank		0.000	23:28
Control Test		0.080	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 06/29/2026 10:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRAY,CHANTAL,P
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: SULLIVAN,ALLISON
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CAMANO

DATE/TIME OF TEST: 05/29/2026 02:35:26

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.121	02:30
Breath Volume		1.773	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 2	0.106	02:33
Breath Volume		1.375	02:33
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 06/29/2026 10:36

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTRIDGE,JAMES,T
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: EASTRIDGE,JAMES
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 05/29/2026 22:30:14

Diagnostics Check		OK	22:23
Air Blank		0.000	22:23
Control Test		0.079	22:24
Air Blank		0.000	22:24
Subject Sample	# 1	0.145	22:25
Breath Volume		1.933	22:25
Air Blank		0.000	22:26
Air Blank		0.000	22:28
Subject Sample	# 2	0.145	22:28
Breath Volume		2.273	22:28
Air Blank		0.000	22:29
Control Test		0.078	22:29
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 06/29/2026 10:36

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: CHOUDRI,H
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 05/29/2026 23:01:14

Diagnostics Check		OK	22:49	
Air Blank		0.000	22:50	
Control Test		0.078	22:50	
Air Blank		0.000	22:50	
Subject Sample	# 1	0.077	22:54	Volume Not Met
Breath Volume		0.113	22:54	
Air Blank		0.000	22:54	
Air Blank		0.000	22:56	
Subject Sample	# 2	0.074	22:59	Volume Not Met
Breath Volume		0.144	22:59	
Air Blank		0.000	23:00	
Control Test		0.077	23:00	
Air Blank		0.000	23:01	
Diagnostics Check		OK	23:01	

Data Download Date/Time: 06/29/2026 10:36

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURTZLAFF,SYDNEE,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: SMITH,NICHOLAS
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 05/30/2026 01:22:10

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.080	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.073	01:17
Breath Volume		1.250	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:19
Subject Sample	# 2	0.084	01:20
Breath Volume		1.382	01:20
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 06/29/2026 10:36

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,STEPHEN,V
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: MENNECHEY,LORENA
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: MIDDLETON

DATE/TIME OF TEST: 05/30/2026 15:15:25

Diagnostics Check		OK	15:09
Air Blank		0.000	15:09
Control Test		0.080	15:09
Air Blank		0.000	15:10
Subject Sample	# 1	0.030	15:11
Breath Volume		1.320	15:11
Air Blank		0.000	15:11
Air Blank		0.000	15:13
Subject Sample	# 2	0.027	15:13
Breath Volume		1.437	15:13
Air Blank		0.000	15:14
Control Test		0.079	15:14
Air Blank		0.000	15:15
Diagnostics Check		OK	15:15

Data Download Date/Time: 06/29/2026 10:36

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN SWEET,B,
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: ALLEN SWEET,B
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: O'BRIEN

DATE/TIME OF TEST: 05/31/2026 02:24:13

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:21	
Control Test		0.080	02:22	
Air Blank		0.000	02:22	
Subject Sample	# 1		02:22	Subject Test Refused
Breath Volume		0.000	02:22	
Air Blank		0.000	02:23	
Control Test		0.080	02:23	
Air Blank		0.000	02:24	
Diagnostics Check		OK	02:24	

Data Download Date/Time: 06/29/2026 10:36

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: GREEN,E
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: VASTOLA

DATE/TIME OF TEST: 05/31/2026 03:30:22

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.080	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 1		03:23	No Sample Provided
Breath Volume		0.000	03:23	
Air Blank		0.000	03:23	
Air Blank		0.000	03:25	
Subject Sample	# 2		03:29	No Sample Provided
Breath Volume		0.000	03:29	
Air Blank		0.000	03:29	
Control Test		0.080	03:29	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 06/29/2026 10:36

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORFEL,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WORFEL,KYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: MCCANTS

DATE/TIME OF TEST: 05/31/2026 17:25:36

Diagnostics Check		OK	17:19
Air Blank		0.000	17:19
Control Test		0.079	17:19
Air Blank		0.000	17:20
Subject Sample	# 1	0.218	17:21
Breath Volume		1.128	17:21
Air Blank		0.000	17:21
Air Blank		0.000	17:23
Subject Sample	# 2	0.236	17:24
Breath Volume		1.433	17:24
Air Blank		0.000	17:24
Control Test		0.080	17:24
Air Blank		0.000	17:25
Diagnostics Check		OK	17:25

Data Download Date/Time: 06/29/2026 10:36

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOROS,CHUWAN,
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BOROS,CHUWAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/01/2026 05:46:04

Diagnostics Check		OK	05:42	
Air Blank		0.000	05:43	
Control Test		0.080	05:43	
Air Blank		0.000	05:44	
Subject Sample	# 1		05:44	Subject Test Refused
Breath Volume		0.000	05:44	
Air Blank		0.000	05:45	
Control Test		0.080	05:45	
Air Blank		0.000	05:45	
Diagnostics Check		OK	05:46	

Data Download Date/Time: 05/29/2026 12:48

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 05/02/2026 01:25:42

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.081	01:20
Air Blank		0.000	01:20
Subject Sample	# 1	0.149	01:21
Breath Volume		3.402	01:21
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.130	01:24
Breath Volume		1.300	01:24
Air Blank		0.000	01:24
Control Test		0.080	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 05/29/2026 12:48

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALNAKER,TEDDY,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: STALNAKER,TEDDY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: BALK-KRADEL

DATE/TIME OF TEST: 05/03/2026 16:36:40

Diagnostics Check		OK	16:29
Air Blank		0.000	16:30
Control Test		0.080	16:30
Air Blank		0.000	16:31
Subject Sample	# 1	0.159	16:32
Breath Volume		1.449	16:32
Air Blank		0.000	16:32
Air Blank		0.000	16:34
Subject Sample	# 2	0.158	16:35
Breath Volume		1.253	16:35
Air Blank		0.000	16:35
Control Test		0.080	16:36
Air Blank		0.000	16:36
Diagnostics Check		OK	16:36

Data Download Date/Time: 05/29/2026 12:48

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: MELENDEZ AGOSTO

DATE/TIME OF TEST: 05/05/2026 00:08:13

Diagnostics Check		OK	00:01
Air Blank		0.000	00:02
Control Test		0.080	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.165	00:03
Breath Volume		2.664	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.162	00:06
Breath Volume		2.816	00:06
Air Blank		0.000	00:07
Control Test		0.080	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 05/29/2026 12:48

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 05/06/2026 00:56:43

Diagnostics Check		OK	00:43	
Air Blank		0.000	00:44	
Control Test		0.080	00:44	
Air Blank		0.000	00:45	
Subject Sample	# 1	0.182	00:48	Slope Not Level
Breath Volume		1.777	00:48	
Air Blank		0.000	00:49	
Air Blank		0.000	00:50	
Subject Sample	# 2	0.181	00:52	
Breath Volume		1.488	00:52	
Air Blank		0.000	00:52	
Air Blank		0.000	00:54	
Subject Sample	# 3	0.178	00:55	
Breath Volume		1.355	00:55	
Air Blank		0.000	00:55	
Control Test		0.079	00:56	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 05/29/2026 12:48

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOROS,CHUWAN,
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HALE,CALBE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/08/2026 05:41:14

Diagnostics Check		OK	05:33
Air Blank		0.000	05:34
Control Test		0.078	05:34
Air Blank		0.000	05:35
Subject Sample	# 1	0.231	05:36
Breath Volume		1.773	05:36
Air Blank		0.000	05:36
Air Blank		0.000	05:38
Subject Sample	# 2	0.231	05:39
Breath Volume		1.394	05:39
Air Blank		0.000	05:40
Control Test		0.079	05:40
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

Data Download Date/Time: 05/29/2026 12:48

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MITCHELL,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: SLADE

DATE/TIME OF TEST: 05/09/2026 21:19:16

Diagnostics Check		OK	21:14	
Air Blank		0.000	21:15	
Control Test		0.080	21:15	
Air Blank		0.000	21:15	
Subject Sample	# 1		21:17	Subject Test Refused
Breath Volume		0.554	21:17	
Air Blank		0.000	21:18	
Control Test		0.080	21:18	
Air Blank		0.000	21:19	
Diagnostics Check		OK	21:19	

Data Download Date/Time: 05/29/2026 12:48

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUMBIE,JAMES,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MILLER,ZACHARY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: GRYSZOWKA

DATE/TIME OF TEST: 05/09/2026 22:46:31

Diagnostics Check		OK	22:40
Air Blank		0.000	22:40
Control Test		0.079	22:41
Air Blank		0.000	22:41
Subject Sample	# 1	0.089	22:42
Breath Volume		3.097	22:42
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.085	22:45
Breath Volume		1.886	22:45
Air Blank		0.000	22:45
Control Test		0.078	22:45
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 05/29/2026 12:48

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/10/2026 01:35:47

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.079	01:29
Air Blank		0.000	01:30
Subject Sample	# 1	0.197	01:31
Breath Volume		1.507	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.199	01:34
Breath Volume		1.332	01:34
Air Blank		0.000	01:34
Control Test		0.079	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 05/29/2026 12:48

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,R,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CRUZ,R
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: RICCI

DATE/TIME OF TEST: 05/13/2026 19:04:50

Diagnostics Check		OK	18:56
Air Blank		0.000	18:56
Control Test		0.080	18:57
Air Blank		0.000	18:57
Subject Sample	# 1	0.200	18:58
Breath Volume		1.570	18:58
Air Blank		0.000	18:59
Air Blank		0.000	19:01
Subject Sample	# 2	0.200	19:03
Breath Volume		2.445	19:03
Air Blank		0.000	19:03
Control Test		0.079	19:04
Air Blank		0.000	19:04
Diagnostics Check		OK	19:04

Data Download Date/Time: 05/29/2026 12:48

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIPF,RICHARD,K
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF,RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: SPINDLER

DATE/TIME OF TEST: 05/14/2026 18:03:36

Diagnostics Check		OK	17:54
Air Blank		0.000	17:54
Control Test		0.080	17:55
Air Blank		0.000	17:55
Subject Sample	# 1	0.225	17:58
Breath Volume		1.433	17:58
Air Blank		0.000	17:58
Air Blank		0.000	18:00
Subject Sample	# 2	0.227	18:02
Breath Volume		1.855	18:02
Air Blank		0.000	18:02
Control Test		0.079	18:03
Air Blank		0.000	18:03
Diagnostics Check		OK	18:03

Data Download Date/Time: 05/29/2026 12:48

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOBEL,EARL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: WRAY,J
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: MASONTE

DATE/TIME OF TEST: 05/14/2026 22:05:28

Diagnostics Check		OK	21:58
Air Blank		0.000	21:59
Control Test		0.081	21:59
Air Blank		0.000	22:00
Subject Sample	# 1	0.170	22:00
Breath Volume		1.867	22:00
Air Blank		0.000	22:01
Air Blank		0.000	22:03
Subject Sample	# 2	0.175	22:04
Breath Volume		1.851	22:04
Air Blank		0.000	22:04
Control Test		0.078	22:04
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 05/29/2026 12:48

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOBEL,EARL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: YOUNG,T.
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: STEPHENSON

DATE/TIME OF TEST: 05/15/2026 04:11:44

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.082	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.179	04:07
Breath Volume		3.558	04:07
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.180	04:10
Breath Volume		4.394	04:10
Air Blank		0.000	04:10
Control Test		0.080	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 05/29/2026 12:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMUEL,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: WRAY,JOHN
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 10:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: BRITT

DATE/TIME OF TEST: 05/18/2026 23:49:44

Diagnostics Check		OK	23:34	
Air Blank		0.000	23:35	
Control Test		0.080	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 1	0.182	23:39	Slope Not Level
Breath Volume		1.304	23:39	
Air Blank		0.000	23:39	
Air Blank		0.000	23:41	
Subject Sample	# 2	0.194	23:44	
Breath Volume		1.171	23:44	
Air Blank		0.000	23:45	
Air Blank		0.000	23:47	
Subject Sample	# 3	0.189	23:48	
Breath Volume		1.101	23:48	
Air Blank		0.000	23:48	
Control Test		0.080	23:49	
Air Blank		0.000	23:49	
Diagnostics Check		OK	23:49	

Data Download Date/Time: 05/29/2026 12:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSME,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HALE,CALEB
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 05/19/2026 01:13:52

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:01	
Control Test		0.080	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1	0.150	01:05	Volume Not Met
Breath Volume		0.597	01:05	
Air Blank		0.000	01:06	
Air Blank		0.000	01:07	
Subject Sample	# 2	0.193	01:08	
Breath Volume		1.496	01:08	
Air Blank		0.000	01:08	
Air Blank		0.000	01:10	
Subject Sample	# 3		01:12	Subject Test Refused
Breath Volume		0.214	01:12	
Air Blank		0.000	01:12	
Control Test		0.081	01:13	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 05/29/2026 12:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: KODER

DATE/TIME OF TEST: 05/22/2026 01:49:00

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:46	
Control Test		0.081	01:46	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:47	RFI Detect
Breath Volume		0.000	01:47	
Air Blank		0.000	01:48	
Control Test		0.080	01:48	
Air Blank		0.000	01:48	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 05/29/2026 12:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: KODER

DATE/TIME OF TEST: 05/22/2026 01:56:59

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.144	01:52
Breath Volume		3.871	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.146	01:55
Breath Volume		3.796	01:55
Air Blank		0.000	01:56
Control Test		0.080	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 05/29/2026 12:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOBEL,OFC. E.,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MILLER,OFC. Z.
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 05/23/2026 01:57:37

Diagnostics Check		OK	01:51
Air Blank		0.000	01:51
Control Test		0.080	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.208	01:53
Breath Volume		3.203	01:53
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.206	01:56
Breath Volume		2.480	01:56
Air Blank		0.000	01:56
Control Test		0.080	01:57
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 05/29/2026 12:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMUEL,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: PALMER,SAMUEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/26/2026 16:17:27

Diagnostics Check		OK	16:10
Air Blank		0.000	16:11
Control Test		0.080	16:11
Air Blank		0.000	16:12
Subject Sample	# 1	0.000	16:13
Breath Volume		2.195	16:13
Air Blank		0.000	16:13
Air Blank		0.000	16:15
Subject Sample	# 2	0.000	16:16
Breath Volume		2.753	16:16
Air Blank		0.000	16:16
Control Test		0.079	16:16
Air Blank		0.000	16:17
Diagnostics Check		OK	16:17

Data Download Date/Time: 05/29/2026 12:50

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 05/26/2026 23:29:07

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.080	23:23
Air Blank		0.000	23:24
Subject Sample	# 1	0.185	23:24
Breath Volume		3.144	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.187	23:27
Breath Volume		2.574	23:27
Air Blank		0.000	23:28
Control Test		0.080	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 05/29/2026 12:50

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,KYLE,F
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HOLMES,KYLE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: ELDRIDGE

DATE/TIME OF TEST: 05/30/2026 23:40:14

Diagnostics Check		OK	23:31
Air Blank		0.000	23:31
Control Test		0.080	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.171	23:35
Breath Volume		1.789	23:35
Air Blank		0.000	23:36
Air Blank		0.000	23:38
Subject Sample	# 2	0.163	23:38
Breath Volume		1.519	23:38
Air Blank		0.000	23:39
Control Test		0.079	23:39
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 06/30/2026 18:35

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FULLER,ZACHARY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: MLADUCKY

DATE/TIME OF TEST: 05/01/2026 04:04:36

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.083	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.155	03:59
Breath Volume		3.488	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.159	04:03
Breath Volume		3.906	04:03
Air Blank		0.000	04:03
Control Test		0.082	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 05/09/2026 11:20

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP,RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,CRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: ECKEL

DATE/TIME OF TEST: 05/01/2026 21:41:12

Diagnostics Check		OK	21:35
Air Blank		0.000	21:35
Control Test		0.082	21:35
Air Blank		0.000	21:36
Subject Sample	# 1	0.218	21:36
Breath Volume		3.035	21:36
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.217	21:39
Breath Volume		2.746	21:39
Air Blank		0.000	21:40
Control Test		0.081	21:40
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 05/09/2026 11:21

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HEWITT,HUNTER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SHAFER

DATE/TIME OF TEST: 05/02/2026 01:44:39

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:41	
Control Test		0.082	01:41	
Air Blank		0.000	01:42	
Subject Sample	# 1		01:43	Subject Test Refused
Breath Volume		0.000	01:43	
Air Blank		0.000	01:43	
Control Test		0.082	01:44	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 05/09/2026 11:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HEWITT,HUNTER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: SHAFER

DATE/TIME OF TEST: 05/02/2026 02:23:01

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.082	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.203	02:18
Breath Volume		1.718	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.216	02:21
Breath Volume		1.406	02:21
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:23

Data Download Date/Time: 05/09/2026 11:21

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: MACHADO PEREZ

DATE/TIME OF TEST: 05/02/2026 18:53:18

Diagnostics Check		OK	18:46
Air Blank		0.000	18:46
Control Test		0.082	18:47
Air Blank		0.000	18:47
Subject Sample	# 1	0.135	18:48
Breath Volume		1.328	18:48
Air Blank		0.000	18:49
Air Blank		0.000	18:50
Subject Sample	# 2	0.134	18:51
Breath Volume		2.058	18:51
Air Blank		0.000	18:52
Control Test		0.082	18:52
Air Blank		0.000	18:53
Diagnostics Check		OK	18:53

Data Download Date/Time: 05/09/2026 11:21

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOULIHAN,RILEY,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MCCLOUGHEN,KORY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: BLASSES

DATE/TIME OF TEST: 05/03/2026 19:47:59

Diagnostics Check		OK	19:41
Air Blank		0.000	19:42
Control Test		0.081	19:42
Air Blank		0.000	19:42
Subject Sample	# 1	0.219	19:43
Breath Volume		4.277	19:43
Air Blank		0.000	19:44
Air Blank		0.000	19:46
Subject Sample	# 2	0.209	19:46
Breath Volume		4.308	19:46
Air Blank		0.000	19:47
Control Test		0.081	19:47
Air Blank		0.000	19:47
Diagnostics Check		OK	19:47

Data Download Date/Time: 05/09/2026 11:21

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FULLER,ZACHARY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: CALLOWAY

DATE/TIME OF TEST: 05/05/2026 02:53:32

Diagnostics Check		OK	02:46
Air Blank		0.000	02:47
Control Test		0.082	02:47
Air Blank		0.000	02:48
Subject Sample	# 1	0.136	02:48
Breath Volume		1.957	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.142	02:52
Breath Volume		3.449	02:52
Air Blank		0.000	02:52
Control Test		0.081	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 05/09/2026 11:21

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: LOWMAN

DATE/TIME OF TEST: 05/05/2026 23:00:35

Diagnostics Check		OK	22:54
Air Blank		0.000	22:54
Control Test		0.082	22:55
Air Blank		0.000	22:55
Subject Sample	# 1	0.000	22:56
Breath Volume		4.539	22:56
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.000	22:59
Breath Volume		4.605	22:59
Air Blank		0.000	22:59
Control Test		0.082	23:00
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 05/09/2026 11:21

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRITZ,TYLER,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RODRIGUEZ,ALFONSO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:21

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 05/07/2026 14:48:23

Diagnostics Check		OK	14:45	
Air Blank		0.000	14:46	
Control Test		0.082	14:46	
Air Blank		0.000	14:46	
Subject Sample	# 1		14:47	Subject Test Refused
Breath Volume		0.000	14:47	
Air Blank		0.000	14:47	
Control Test		0.081	14:47	
Air Blank		0.000	14:48	
Diagnostics Check		OK	14:48	

Data Download Date/Time: 05/09/2026 11:21

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO PEREZ,CARLOS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 05/07/2026 23:45:00

Diagnostics Check		OK	23:38
Air Blank		0.000	23:39
Control Test		0.081	23:39
Air Blank		0.000	23:40
Subject Sample	# 1	0.112	23:40
Breath Volume		1.933	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.115	23:43
Breath Volume		3.808	23:43
Air Blank		0.000	23:44
Control Test		0.081	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:45

Data Download Date/Time: 05/09/2026 11:21

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINTER,ERIC,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALFONSO SANTANA,DAVID
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: TASKA JR

DATE/TIME OF TEST: 05/09/2026 22:28:56

Diagnostics Check		OK	22:21
Air Blank		0.000	22:22
Control Test		0.082	22:22
Air Blank		0.000	22:23
Subject Sample	# 1	0.219	22:24
Breath Volume		1.152	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.213	22:27
Breath Volume		1.183	22:27
Air Blank		0.000	22:28
Control Test		0.081	22:28
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 06/06/2026 11:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINTER,ERIC,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: NELSON,COLE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: KURKIMILIS

DATE/TIME OF TEST: 05/09/2026 22:42:30

Diagnostics Check		OK	22:36
Air Blank		0.000	22:36
Control Test		0.081	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.174	22:37
Breath Volume		2.519	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.175	22:41
Breath Volume		2.472	22:41
Air Blank		0.000	22:41
Control Test		0.080	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 06/06/2026 11:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINTER,ERIC,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: NELSON,COLE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/09/2026 22:50:30

Diagnostics Check		OK	22:44
Air Blank		0.000	22:44
Control Test		0.081	22:45
Air Blank		0.000	22:45
Subject Sample	# 1	0.156	22:45
Breath Volume		2.640	22:45
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.166	22:49
Breath Volume		3.066	22:49
Air Blank		0.000	22:49
Control Test		0.081	22:49
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 06/06/2026 11:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINTER,ERIC,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,CHRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 05/10/2026 03:48:32

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.081	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.159	03:44
Breath Volume		1.605	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.154	03:47
Breath Volume		1.593	03:47
Air Blank		0.000	03:47
Control Test		0.079	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 06/06/2026 11:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGATTI,BRADLEY,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALFONSO,CHRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/12/2026 01:35:58

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:22	
Control Test		0.082	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 1	0.142	01:26	Volume Not Met
Breath Volume		0.035	01:26	
Air Blank		0.000	01:27	
Air Blank		0.000	01:29	
Subject Sample	# 2	0.146	01:31	
Breath Volume		2.945	01:31	
Air Blank		0.000	01:31	
Air Blank		0.000	01:33	
Subject Sample	# 3	0.145	01:34	
Breath Volume		2.179	01:34	
Air Blank		0.000	01:35	
Control Test		0.080	01:35	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 06/06/2026 11:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGATTI,BRADLEY,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FRAZZETTO,ASHLEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: CUBBIN

DATE/TIME OF TEST: 05/12/2026 23:58:30

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:47	
Control Test		0.081	23:47	
Air Blank		0.000	23:48	
Subject Sample	# 1		23:51	No Sample Provided
Breath Volume		0.000	23:51	
Air Blank		0.000	23:52	
Air Blank		0.000	23:54	
Subject Sample	# 2		23:57	No Sample Provided
Breath Volume		0.000	23:57	
Air Blank		0.000	23:57	
Control Test		0.082	23:57	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 06/06/2026 11:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP,RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP,RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: RODRIGUEZ BRIANO

DATE/TIME OF TEST: 05/13/2026 21:45:57

Diagnostics Check		OK	21:39
Air Blank		0.000	21:39
Control Test		0.081	21:40
Air Blank		0.000	21:40
Subject Sample	# 1	0.149	21:41
Breath Volume		1.136	21:41
Air Blank		0.000	21:41
Air Blank		0.000	21:43
Subject Sample	# 2	0.143	21:44
Breath Volume		1.414	21:44
Air Blank		0.000	21:45
Control Test		0.082	21:45
Air Blank		0.000	21:45
Diagnostics Check		OK	21:45

Data Download Date/Time: 06/06/2026 11:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP,RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP,RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 05/21/2026 00:06:05

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.082	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.218	00:01
Breath Volume		1.917	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.208	00:04
Breath Volume		1.859	00:04
Air Blank		0.000	00:05
Control Test		0.082	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 06/06/2026 11:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO PEREZ,CARLOS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/21/2026 01:42:56

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.082	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.000	01:38
Breath Volume		2.898	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.000	01:41
Breath Volume		2.167	01:41
Air Blank		0.000	01:42
Control Test		0.082	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 06/06/2026 11:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANKOSKY,D,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FULLER,Z
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 05/25/2026 02:45:08

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.082	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.193	02:40
Breath Volume		3.781	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:42
Subject Sample	# 2	0.193	02:43
Breath Volume		3.804	02:43
Air Blank		0.000	02:44
Control Test		0.081	02:44
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 06/06/2026 11:49

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/03/2026 00:36:15

Diagnostics Check		OK	00:29
Air Blank		0.000	00:30
Control Test		0.079	00:30
Air Blank		0.000	00:30
Subject Sample	# 1	0.205	00:31
Breath Volume		1.500	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.204	00:34
Breath Volume		2.605	00:34
Air Blank		0.000	00:35
Control Test		0.078	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 05/29/2026 11:17

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: EWING

DATE/TIME OF TEST: 05/03/2026 22:03:46

Diagnostics Check		OK	21:57
Air Blank		0.000	21:57
Control Test		0.079	21:58
Air Blank		0.000	21:58
Subject Sample	# 1	0.000	21:59
Breath Volume		2.339	21:59
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.000	22:02
Breath Volume		2.761	22:02
Air Blank		0.000	22:02
Control Test		0.078	22:03
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 05/29/2026 11:17

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: ARNONE

DATE/TIME OF TEST: 05/16/2026 20:06:33

Diagnostics Check		OK	19:57
Air Blank		0.000	19:58
Control Test		0.079	19:58
Air Blank		0.000	19:59
Subject Sample	# 1	0.282	20:01
Breath Volume		2.023	20:01
Air Blank		0.000	20:02
Air Blank		0.000	20:04
Subject Sample	# 2	0.285	20:05
Breath Volume		1.945	20:05
Air Blank		0.000	20:05
Control Test		0.078	20:05
Air Blank		0.000	20:06
Diagnostics Check		OK	20:06

Data Download Date/Time: 05/29/2026 11:17

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASHLEY,ADAM,B
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: LASHLEY,ADAM
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: BUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: ARAGON MAZARIEGOS

DATE/TIME OF TEST: 05/03/2026 06:38:01

Diagnostics Check		OK	06:30
Air Blank		0.000	06:31
Control Test		0.080	06:31
Air Blank		0.000	06:31
Subject Sample	# 1	0.206	06:32
Breath Volume		3.132	06:32
Air Blank		0.000	06:33
Air Blank		0.000	06:35
Subject Sample	# 2	0.211	06:36
Breath Volume		3.835	06:36
Air Blank		0.000	06:37
Control Test		0.081	06:37
Air Blank		0.000	06:37
Diagnostics Check		OK	06:38

Data Download Date/Time: 05/21/2026 18:51

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,JOHNASIA,L
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: DIXON,AARON
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: TANTON

DATE/TIME OF TEST: 05/03/2026 19:43:19

Diagnostics Check	OK	19:42	
Air Blank	0.000	19:42	
Control Test	0.080	19:43	
Air Blank		19:43	Improper Sample
Air Blank		19:43	Improper Sample

Data Download Date/Time: 05/21/2026 18:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,JOHNASIA,L
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: DIXON,AARON
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: TANTOM

DATE/TIME OF TEST: 05/03/2026 19:46:17

Diagnostics Check	OK	19:45	
Air Blank		19:45	Improper Sample
Air Blank	0.000	19:46	

Data Download Date/Time: 05/21/2026 18:51

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,JOHNASIA,L
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: DIXON,AARON
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: TANTON

DATE/TIME OF TEST: 05/03/2026 19:49:48

Diagnostics Check	OK	19:49	
Air Blank		19:49	Improper Sample
Air Blank		19:49	Improper Sample

Data Download Date/Time: 05/21/2026 18:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,JOHNASIA,L
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: DIXON,AARON
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: TANTON

DATE/TIME OF TEST: 05/03/2026 19:54:40

Diagnostics Check	OK	19:54	
Air Blank		19:54	Improper Sample
Air Blank	0.000	19:54	

Data Download Date/Time: 05/21/2026 18:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,JOHNASIA,L
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: DIXON,AARON
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: TANTON

DATE/TIME OF TEST: 05/03/2026 20:00:33

Diagnostics Check	OK	19:59	
Air Blank		20:00	RFI Detect
Air Blank	0.000	20:00	

Data Download Date/Time: 05/21/2026 18:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,JOHNASIA,L
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: DIXON,AARON
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: TANTON

DATE/TIME OF TEST: 05/03/2026 20:16:56

Diagnostics Check		OK	20:13	
Air Blank		0.000	20:14	
Control Test		0.080	20:14	
Air Blank		0.000	20:15	
Subject Sample	# 1		20:15	RFI Detect
Breath Volume		0.488	20:15	
Air Blank		0.000	20:16	
Control Test		0.080	20:16	
Air Blank		0.000	20:16	
Diagnostics Check		OK	20:16	

Data Download Date/Time: 05/21/2026 18:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,JOHNASIA,L
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: DIXON,AARON
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: TANTON

DATE/TIME OF TEST: 05/03/2026 20:29:54

Diagnostics Check		OK	20:24
Air Blank		0.000	20:24
Control Test		0.079	20:24
Air Blank		0.000	20:25
Subject Sample	# 1	0.185	20:25
Breath Volume		2.039	20:25
Air Blank		0.000	20:26
Air Blank		0.000	20:28
Subject Sample	# 2	0.185	20:28
Breath Volume		1.449	20:28
Air Blank		0.000	20:28
Control Test		0.079	20:29
Air Blank		0.000	20:29
Diagnostics Check		OK	20:29

Data Download Date/Time: 05/21/2026 18:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASHLEY,ADAM,B
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: LASHLEY,ADAM
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: ALFONSO

DATE/TIME OF TEST: 05/07/2026 00:12:45

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.081	00:07
Air Blank		0.000	00:07
Subject Sample	# 1	0.176	00:08
Breath Volume		2.371	00:08
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.180	00:11
Breath Volume		2.394	00:11
Air Blank		0.000	00:11
Control Test		0.081	00:12
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 05/21/2026 18:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KORZEN,GREG,B
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: KORZEN,GREG
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 13:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 05/08/2026 15:16:46

Diagnostics Check		OK	15:10
Air Blank		0.000	15:10
Control Test		0.081	15:11
Air Blank		0.000	15:11
Subject Sample	# 1	0.142	15:12
Breath Volume		2.011	15:12
Air Blank		0.000	15:12
Air Blank		0.000	15:14
Subject Sample	# 2	0.142	15:15
Breath Volume		2.449	15:15
Air Blank		0.000	15:15
Control Test		0.080	15:16
Air Blank		0.000	15:16
Diagnostics Check		OK	15:16

Data Download Date/Time: 05/21/2026 18:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADORNO,SAMUEL,D
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: FASULO,DANIELLE
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: FREEMAN EBANKS

DATE/TIME OF TEST: 05/08/2026 22:41:04

Diagnostics Check		OK	22:35
Air Blank		0.000	22:35
Control Test		0.081	22:35
Air Blank		0.000	22:36
Subject Sample	# 1	0.171	22:36
Breath Volume		1.585	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.174	22:39
Breath Volume		2.847	22:39
Air Blank		0.000	22:40
Control Test		0.080	22:40
Air Blank		0.000	22:40
Diagnostics Check		OK	22:41

Data Download Date/Time: 05/21/2026 18:52

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,ESTEBAN,
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: CAHILL,CHRISTOPHER
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 05/11/2026 00:15:12

Diagnostics Check		OK	00:12	
Air Blank		0.000	00:12	
Control Test		0.081	00:13	
Air Blank		0.000	00:13	
Subject Sample	# 1		00:13	Subject Test Refused
Breath Volume		0.000	00:13	
Air Blank		0.000	00:14	
Control Test		0.080	00:14	
Air Blank		0.000	00:15	
Diagnostics Check		OK	00:15	

Data Download Date/Time: 05/21/2026 18:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: THORNBRUGH,ELAM
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 05/14/2026 06:42:26

Diagnostics Check		OK	06:36
Air Blank		0.000	06:36
Control Test		0.081	06:36
Air Blank		0.000	06:37
Subject Sample	# 1	0.151	06:38
Breath Volume		2.289	06:38
Air Blank		0.000	06:38
Air Blank		0.000	06:40
Subject Sample	# 2	0.144	06:40
Breath Volume		2.296	06:40
Air Blank		0.000	06:41
Control Test		0.080	06:41
Air Blank		0.000	06:42
Diagnostics Check		OK	06:42

Data Download Date/Time: 05/21/2026 18:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADORNO,SAMUEL,D
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: ADORNO,SAMUEL
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 05/18/2026 01:37:16

Diagnostics Check		OK	01:26	
Air Blank		0.000	01:26	
Control Test		0.080	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 1	0.171	01:30	Slope Not Level
Breath Volume		1.511	01:30	
Air Blank		0.000	01:31	
Air Blank		0.000	01:32	
Subject Sample	# 2		01:35	Subject Test Refused
Breath Volume		0.019	01:35	
Air Blank		0.000	01:36	
Control Test		0.077	01:36	
Air Blank		0.000	01:37	
Diagnostics Check		OK	01:37	

Data Download Date/Time: 05/21/2026 18:53

Printed: 03 August 2026

COBRA ©1997-2014 CMI, Inc.